

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: TAYLOR,T

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 17:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 02/03/2020 19:19:06

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:10 | |
| Air Blank | 0.000 | 19:10 | |
| Control Test | 0.077 | 19:10 | |
| Air Blank | 0.000 | 19:11 | |
| Subject Sample | # 1 | 19:14 | No Sample Provided |
| Breath Volume | 0.000 | 19:14 | |
| Air Blank | 0.000 | 19:14 | |
| Air Blank | 0.000 | 19:16 | |
| Subject Sample | # 2 | 19:17 | Subject Test Refused |
| Breath Volume | 0.000 | 19:17 | |
| Air Blank | 0.000 | 19:18 | |
| Control Test | 0.076 | 19:18 | |
| Air Blank | 0.000 | 19:19 | |
| Diagnostics Check | OK | 19:19 | |

Data Download Date/Time: 03/10/2020 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WIGGLESWORTH,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: TERHUNE

DATE/TIME OF TEST: 02/08/2020 18:47:26

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 18:43 | |
| Air Blank | 0.000 | 18:43 | |
| Control Test | 0.077 | 18:44 | |
| Air Blank | 0.000 | 18:44 | |
| Subject Sample | # 1 | 18:46 | Subject Test Refused |
| Breath Volume | 0.000 | 18:46 | |
| Air Blank | 0.000 | 18:46 | |
| Control Test | 0.076 | 18:46 | |
| Air Blank | 0.000 | 18:47 | |
| Diagnostics Check | OK | 18:47 | |

Data Download Date/Time: 03/10/2020 17:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH,DEVIN,A

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: ERDMAN,N

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 02/08/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 19:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 02/08/2020 20:03:40

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 20:00 | |
| Air Blank | 0.000 | 20:00 | |
| Control Test | 0.076 | 20:00 | |
| Air Blank | 0.000 | 20:01 | |
| Subject Sample | # 1 | 20:02 | Subject Test Refused |
| Breath Volume | 0.000 | 20:02 | |
| Air Blank | 0.000 | 20:02 | |
| Control Test | 0.077 | 20:03 | |
| Air Blank | 0.000 | 20:03 | |
| Diagnostics Check | OK | 20:03 | |

Data Download Date/Time: 03/10/2020 17:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CASON,CHRISTOPHER,N

BREATH TEST OPERATOR AGENCY: CCA Citrus

ARRESTING OFFICER: WIGGLSEWORTH,C

ARRESTING OFFICER AGENCY: Citrus County SO

LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13518080A5

DRY GAS EXPIRY DATE: 08/05/2020

VIOLATION CODE: DUI

ARREST TIME: 00:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/14/2020 01:23:10

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:20 | |
| Air Blank | 0.000 | 01:20 | |
| Control Test | 0.077 | 01:21 | |
| Air Blank | 0.000 | 01:21 | |
| Subject Sample | # 1 | 01:21 | Subject Test Refused |
| Breath Volume | 0.000 | 01:21 | |
| Air Blank | 0.000 | 01:22 | |
| Control Test | 0.077 | 01:22 | |
| Air Blank | 0.000 | 01:23 | |
| Diagnostics Check | OK | 01:23 | |

Data Download Date/Time: 03/10/2020 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT, JOSHUA, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CLOUD, TODD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: SCHILLING

DATE/TIME OF TEST: 02/14/2020 19:00:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:54 |
| Air Blank | 0.000 | 18:54 |
| Control Test | 0.077 | 18:55 |
| Air Blank | 0.000 | 18:55 |
| Subject Sample # 1 | 0.193 | 18:56 |
| Breath Volume | 2.078 | 18:56 |
| Air Blank | 0.000 | 18:56 |
| Air Blank | 0.000 | 18:58 |
| Subject Sample # 2 | 0.174 | 18:59 |
| Breath Volume | 2.234 | 18:59 |
| Air Blank | 0.000 | 18:59 |
| Control Test | 0.077 | 18:59 |
| Air Blank | 0.000 | 19:00 |
| Diagnostics Check | OK | 19:00 |

Data Download Date/Time: 03/10/2020 17:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 02/15/2020 03:16:58

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 03:14 | |
| Air Blank | 0.000 | 03:14 | |
| Control Test | 0.081 | 03:14 | |
| Air Blank | 0.000 | 03:15 | |
| Subject Sample | # 1 | 03:15 | Subject Test Refused |
| Breath Volume | 0.000 | 03:15 | |
| Air Blank | 0.000 | 03:16 | |
| Control Test | 0.081 | 03:16 | |
| Air Blank | 0.000 | 03:16 | |
| Diagnostics Check | OK | 03:16 | |

Data Download Date/Time: 03/04/2020 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS, ABNER, K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SIRMANS, ABNER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SCAFIDI

DATE/TIME OF TEST: 02/21/2020 22:57:43

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:50 |
| Air Blank | 0.000 | 22:50 |
| Control Test | 0.079 | 22:51 |
| Air Blank | 0.000 | 22:51 |
| Subject Sample # 1 | 0.165 | 22:52 |
| Breath Volume | 1.734 | 22:52 |
| Air Blank | 0.000 | 22:53 |
| Air Blank | 0.000 | 22:55 |
| Subject Sample # 2 | 0.169 | 22:56 |
| Breath Volume | 1.613 | 22:56 |
| Air Blank | 0.000 | 22:56 |
| Control Test | 0.078 | 22:57 |
| Air Blank | 0.000 | 22:57 |
| Diagnostics Check | OK | 22:57 |

Data Download Date/Time: 03/04/2020 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL, MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL, MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: IVEY

DATE/TIME OF TEST: 02/22/2020 00:10:34

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:03 |
| Air Blank | 0.000 | 00:03 |
| Control Test | 0.080 | 00:04 |
| Air Blank | 0.000 | 00:04 |
| Subject Sample # 1 | 0.174 | 00:05 |
| Breath Volume | 2.964 | 00:05 |
| Air Blank | 0.000 | 00:06 |
| Air Blank | 0.000 | 00:08 |
| Subject Sample # 2 | 0.177 | 00:08 |
| Breath Volume | 1.445 | 00:08 |
| Air Blank | 0.000 | 00:09 |
| Control Test | 0.080 | 00:09 |
| Air Blank | 0.000 | 00:10 |
| Diagnostics Check | OK | 00:10 |

Data Download Date/Time: 03/04/2020 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HAMBLEY

DATE/TIME OF TEST: 02/23/2020 01:26:33

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:23 | |
| Air Blank | 0.000 | 01:24 | |
| Control Test | 0.080 | 01:24 | |
| Air Blank | 0.000 | 01:24 | |
| Subject Sample | # 1 | 01:25 | Subject Test Refused |
| Breath Volume | 0.000 | 01:25 | |
| Air Blank | 0.000 | 01:25 | |
| Control Test | 0.080 | 01:25 | |
| Air Blank | 0.000 | 01:26 | |
| Diagnostics Check | OK | 01:26 | |

Data Download Date/Time: 03/04/2020 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CHAPIN,NATHAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 02/27/2020 02:08:38

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:05 | |
| Air Blank | 0.000 | 02:06 | |
| Control Test | 0.082 | 02:06 | |
| Air Blank | 0.000 | 02:06 | |
| Subject Sample | # 1 | 02:07 | Subject Test Refused |
| Breath Volume | 0.000 | 02:07 | |
| Air Blank | 0.000 | 02:07 | |
| Control Test | 0.082 | 02:08 | |
| Air Blank | 0.000 | 02:08 | |
| Diagnostics Check | OK | 02:08 | |

Data Download Date/Time: 03/04/2020 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CABRAL,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: VAN BUREN

DATE/TIME OF TEST: 02/21/2020 02:57:56

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:52 | |
| Air Blank | 0.000 | 02:52 | |
| Control Test | 0.080 | 02:53 | |
| Air Blank | 0.000 | 02:53 | |
| Subject Sample | # 1 | 02:56 | Subject Test Refused |
| Breath Volume | 0.000 | 02:56 | |
| Air Blank | 0.000 | 02:57 | |
| Control Test | 0.081 | 02:57 | |
| Air Blank | 0.000 | 02:57 | |
| Diagnostics Check | OK | 02:57 | |

Data Download Date/Time: 03/10/2020 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CARLTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 02/24/2020 01:36:53

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 01:25 | |
| Air Blank | 0.000 | 01:25 | |
| Control Test | 0.079 | 01:26 | |
| Air Blank | 0.000 | 01:26 | |
| Subject Sample | # 1 | 01:29 | No Sample Provided |
| Breath Volume | 0.000 | 01:29 | |
| Air Blank | 0.000 | 01:30 | |
| Air Blank | 0.000 | 01:32 | |
| Subject Sample | # 2 | 01:35 | No Sample Provided |
| Breath Volume | 0.000 | 01:35 | |
| Air Blank | 0.000 | 01:35 | |
| Control Test | 0.081 | 01:36 | |
| Air Blank | 0.000 | 01:36 | |
| Diagnostics Check | OK | 01:36 | |

Data Download Date/Time: 03/10/2020 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: CARLOS

DATE/TIME OF TEST: 02/01/2020 05:17:00

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:13 | |
| Air Blank | 0.000 | 05:14 | |
| Control Test | 0.081 | 05:14 | |
| Air Blank | 0.000 | 05:15 | |
| Subject Sample | # 1 | 05:15 | Subject Test Refused |
| Breath Volume | 0.000 | 05:15 | |
| Air Blank | 0.000 | 05:16 | |
| Control Test | 0.081 | 05:16 | |
| Air Blank | 0.000 | 05:16 | |
| Diagnostics Check | OK | 05:17 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SAVOY

DATE/TIME OF TEST: 02/03/2020 01:31:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:24 |
| Air Blank | 0.000 | 01:25 |
| Control Test | 0.082 | 01:25 |
| Air Blank | 0.000 | 01:26 |
| Subject Sample # 1 | 0.163 | 01:26 |
| Breath Volume | 2.714 | 01:26 |
| Air Blank | 0.000 | 01:27 |
| Air Blank | 0.000 | 01:29 |
| Subject Sample # 2 | 0.158 | 01:29 |
| Breath Volume | 2.730 | 01:29 |
| Air Blank | 0.000 | 01:30 |
| Control Test | 0.081 | 01:30 |
| Air Blank | 0.000 | 01:31 |
| Diagnostics Check | OK | 01:31 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,TONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILHELM,TONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GAFFORD

DATE/TIME OF TEST: 02/03/2020 01:55:16

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:49 |
| Air Blank | | 0.000 | 01:49 |
| Control Test | | 0.081 | 01:49 |
| Air Blank | | 0.000 | 01:50 |
| Subject Sample | # 1 | 0.094 | 01:50 |
| Breath Volume | | 1.816 | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Air Blank | | 0.000 | 01:53 |
| Subject Sample | # 2 | 0.086 | 01:53 |
| Breath Volume | | 1.929 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Control Test | | 0.080 | 01:54 |
| Air Blank | | 0.000 | 01:55 |
| Diagnostics Check | | OK | 01:55 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL,SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 02/03/2020 02:21:39

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:16 | |
| Air Blank | 0.000 | 02:17 | |
| Control Test | 0.081 | 02:17 | |
| Air Blank | 0.000 | 02:18 | |
| Subject Sample | # 1 | 02:20 | Subject Test Refused |
| Breath Volume | 0.000 | 02:20 | |
| Air Blank | 0.000 | 02:20 | |
| Control Test | 0.082 | 02:21 | |
| Air Blank | 0.000 | 02:21 | |
| Diagnostics Check | OK | 02:21 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, RAYMOND, S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMITH, JACKSON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/03/2020 23:37:56

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:34 | |
| Air Blank | 0.000 | 23:34 | |
| Control Test | 0.081 | 23:34 | |
| Air Blank | 0.000 | 23:35 | |
| Subject Sample | # 1 | 23:36 | Subject Test Refused |
| Breath Volume | 0.000 | 23:36 | |
| Air Blank | 0.000 | 23:37 | |
| Control Test | 0.081 | 23:37 | |
| Air Blank | 0.000 | 23:37 | |
| Diagnostics Check | OK | 23:37 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 02/07/2020 01:48:59

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:46 | |
| Air Blank | 0.000 | 01:46 | |
| Control Test | 0.081 | 01:46 | |
| Air Blank | 0.000 | 01:47 | |
| Subject Sample | # 1 | 01:47 | Subject Test Refused |
| Breath Volume | 0.000 | 01:47 | |
| Air Blank | 0.000 | 01:48 | |
| Control Test | 0.081 | 01:48 | |
| Air Blank | 0.000 | 01:48 | |
| Diagnostics Check | OK | 01:48 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 02/08/2020 03:04:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:56 |
| Air Blank | 0.000 | 02:57 |
| Control Test | 0.082 | 02:57 |
| Air Blank | 0.000 | 02:58 |
| Subject Sample # 1 | 0.151 | 02:59 |
| Breath Volume | 2.714 | 02:59 |
| Air Blank | 0.000 | 03:00 |
| Air Blank | 0.000 | 03:01 |
| Subject Sample # 2 | 0.149 | 03:02 |
| Breath Volume | 2.152 | 03:02 |
| Air Blank | 0.000 | 03:03 |
| Control Test | 0.082 | 03:03 |
| Air Blank | 0.000 | 03:03 |
| Diagnostics Check | OK | 03:04 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/09/2020 00:29:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:21 |
| Air Blank | 0.000 | 00:22 |
| Control Test | 0.080 | 00:22 |
| Air Blank | 0.000 | 00:23 |
| Subject Sample # 1 | 0.128 | 00:24 |
| Breath Volume | 1.578 | 00:24 |
| Air Blank | 0.000 | 00:24 |
| Air Blank | 0.000 | 00:26 |
| Subject Sample # 2 | 0.129 | 00:27 |
| Breath Volume | 1.558 | 00:27 |
| Air Blank | 0.000 | 00:28 |
| Control Test | 0.080 | 00:28 |
| Air Blank | 0.000 | 00:29 |
| Diagnostics Check | OK | 00:29 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/09/2020 05:04:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:58 |
| Air Blank | 0.000 | 04:58 |
| Control Test | 0.081 | 04:59 |
| Air Blank | 0.000 | 04:59 |
| Subject Sample # 1 | 0.057 | 05:00 |
| Breath Volume | 1.421 | 05:00 |
| Air Blank | 0.000 | 05:00 |
| Air Blank | 0.000 | 05:02 |
| Subject Sample # 2 | 0.058 | 05:03 |
| Breath Volume | 2.750 | 05:03 |
| Air Blank | 0.000 | 05:03 |
| Control Test | 0.077 | 05:04 |
| Air Blank | 0.000 | 05:04 |
| Diagnostics Check | OK | 05:04 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MONARCH,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TURNPAUGH

DATE/TIME OF TEST: 02/09/2020 21:28:55

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:20 | |
| Air Blank | 0.000 | 21:20 | |
| Control Test | 0.081 | 21:21 | |
| Air Blank | 0.000 | 21:21 | |
| Subject Sample | # 1 | 21:24 | No Sample Provided |
| Breath Volume | 0.000 | 21:24 | |
| Air Blank | 0.000 | 21:25 | |
| Air Blank | 0.000 | 21:27 | |
| Subject Sample | # 2 | 21:27 | Subject Test Refused |
| Breath Volume | 0.000 | 21:27 | |
| Air Blank | 0.000 | 21:28 | |
| Control Test | 0.081 | 21:28 | |
| Air Blank | 0.000 | 21:28 | |
| Diagnostics Check | OK | 21:28 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: FINLEY

DATE/TIME OF TEST: 02/12/2020 21:14:08

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:10 | |
| Air Blank | 0.000 | 21:11 | |
| Control Test | 0.081 | 21:11 | |
| Air Blank | 0.000 | 21:12 | |
| Subject Sample | # 1 | 21:12 | Subject Test Refused |
| Breath Volume | 0.000 | 21:12 | |
| Air Blank | 0.000 | 21:13 | |
| Control Test | 0.080 | 21:13 | |
| Air Blank | 0.000 | 21:14 | |
| Diagnostics Check | OK | 21:14 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MOLTER

DATE/TIME OF TEST: 02/13/2020 02:28:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:22 |
| Air Blank | 0.000 | 02:22 |
| Control Test | 0.080 | 02:23 |
| Air Blank | 0.000 | 02:23 |
| Subject Sample # 1 | 0.019 | 02:24 |
| Breath Volume | 2.386 | 02:24 |
| Air Blank | 0.000 | 02:24 |
| Air Blank | 0.000 | 02:26 |
| Subject Sample # 2 | 0.018 | 02:27 |
| Breath Volume | 2.375 | 02:27 |
| Air Blank | 0.000 | 02:27 |
| Control Test | 0.081 | 02:28 |
| Air Blank | 0.000 | 02:28 |
| Diagnostics Check | OK | 02:28 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMSON, TYLER, H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMSON, TYLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BARREIRA

DATE/TIME OF TEST: 02/14/2020 02:18:50

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:11 |
| Air Blank | 0.000 | 02:12 |
| Control Test | 0.082 | 02:12 |
| Air Blank | 0.000 | 02:13 |
| Subject Sample # 1 | 0.256 | 02:14 |
| Breath Volume | 2.574 | 02:14 |
| Air Blank | 0.000 | 02:15 |
| Air Blank | 0.000 | 02:16 |
| Subject Sample # 2 | 0.241 | 02:17 |
| Breath Volume | 2.164 | 02:17 |
| Air Blank | 0.000 | 02:17 |
| Control Test | 0.081 | 02:18 |
| Air Blank | 0.000 | 02:18 |
| Diagnostics Check | OK | 02:18 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SHOLLER

DATE/TIME OF TEST: 02/15/2020 03:32:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:25 |
| Air Blank | 0.000 | 03:25 |
| Control Test | 0.082 | 03:26 |
| Air Blank | 0.000 | 03:26 |
| Subject Sample # 1 | 0.205 | 03:27 |
| Breath Volume | 3.078 | 03:27 |
| Air Blank | 0.000 | 03:28 |
| Air Blank | 0.000 | 03:29 |
| Subject Sample # 2 | 0.196 | 03:30 |
| Breath Volume | 2.320 | 03:30 |
| Air Blank | 0.000 | 03:31 |
| Control Test | 0.081 | 03:31 |
| Air Blank | 0.000 | 03:31 |
| Diagnostics Check | OK | 03:32 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK, TAVISH, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK, TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BONILLA-OCHOA

DATE/TIME OF TEST: 02/16/2020 05:11:25

| | | | | |
|-------------------|-----|-------|-------|--------------------|
| Diagnostics Check | | OK | 05:00 | |
| Air Blank | | 0.000 | 05:00 | |
| Control Test | | 0.082 | 05:00 | |
| Air Blank | | 0.000 | 05:01 | |
| Subject Sample | # 1 | 0.141 | 05:04 | Volume Not Met |
| Breath Volume | | 0.246 | 05:04 | |
| Air Blank | | 0.000 | 05:05 | |
| Air Blank | | 0.000 | 05:06 | |
| Subject Sample | # 2 | | 05:10 | No Sample Provided |
| Breath Volume | | 0.000 | 05:10 | |
| Air Blank | | 0.000 | 05:10 | |
| Control Test | | 0.080 | 05:10 | |
| Air Blank | | 0.000 | 05:11 | |
| Diagnostics Check | | OK | 05:11 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,MAREIKE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DREW,MAREIKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: SALVADOR

DATE/TIME OF TEST: 02/16/2020 05:45:15

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 05:38 |
| Air Blank | 0.000 | 05:38 |
| Control Test | 0.081 | 05:38 |
| Air Blank | 0.000 | 05:39 |
| Subject Sample # 1 | 0.192 | 05:39 |
| Breath Volume | 1.402 | 05:39 |
| Air Blank | 0.000 | 05:40 |
| Air Blank | 0.000 | 05:42 |
| Subject Sample # 2 | 0.206 | 05:43 |
| Breath Volume | 4.621 | 05:43 |
| Air Blank | 0.000 | 05:44 |
| Control Test | 0.081 | 05:44 |
| Air Blank | 0.000 | 05:45 |
| Diagnostics Check | OK | 05:45 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: SEALE

DATE/TIME OF TEST: 02/16/2020 06:20:14

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 06:17 | |
| Air Blank | 0.000 | 06:17 | |
| Control Test | 0.081 | 06:18 | |
| Air Blank | 0.000 | 06:18 | |
| Subject Sample | # 1 | 06:18 | Subject Test Refused |
| Breath Volume | 0.000 | 06:18 | |
| Air Blank | 0.000 | 06:19 | |
| Control Test | 0.080 | 06:19 | |
| Air Blank | 0.000 | 06:20 | |
| Diagnostics Check | OK | 06:20 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK, TAVISH, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK, TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BALDER

DATE/TIME OF TEST: 02/17/2020 04:33:17

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 04:21 | |
| Air Blank | 0.000 | 04:22 | |
| Control Test | 0.080 | 04:22 | |
| Air Blank | 0.000 | 04:23 | |
| Subject Sample | # 1 | 04:26 | No Sample Provided |
| Breath Volume | 0.000 | 04:26 | |
| Air Blank | 0.000 | 04:26 | |
| Air Blank | 0.000 | 04:28 | |
| Subject Sample | # 2 | 04:31 | No Sample Provided |
| Breath Volume | 0.000 | 04:31 | |
| Air Blank | 0.000 | 04:32 | |
| Control Test | 0.081 | 04:32 | |
| Air Blank | 0.000 | 04:33 | |
| Diagnostics Check | OK | 04:33 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, RAYMOND, S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMITH, JACKSON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BALTAZAR CATALAN

DATE/TIME OF TEST: 02/19/2020 04:30:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:23 |
| Air Blank | 0.000 | 04:23 |
| Control Test | 0.080 | 04:23 |
| Air Blank | 0.000 | 04:24 |
| Subject Sample # 1 | 0.148 | 04:25 |
| Breath Volume | 3.078 | 04:25 |
| Air Blank | 0.000 | 04:26 |
| Air Blank | 0.000 | 04:28 |
| Subject Sample # 2 | 0.146 | 04:29 |
| Breath Volume | 2.804 | 04:29 |
| Air Blank | 0.000 | 04:29 |
| Control Test | 0.079 | 04:30 |
| Air Blank | 0.000 | 04:30 |
| Diagnostics Check | OK | 04:30 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PAPPAS,D
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: STATON

DATE/TIME OF TEST: 02/19/2020 07:03:02

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 06:54 | |
| Air Blank | 0.000 | 06:54 | |
| Control Test | 0.080 | 06:55 | |
| Air Blank | 0.000 | 06:55 | |
| Subject Sample | # 1 | 06:58 | No Sample Provided |
| Breath Volume | 0.000 | 06:58 | |
| Air Blank | 0.000 | 06:59 | |
| Air Blank | 0.000 | 07:01 | |
| Subject Sample | # 2 | 07:01 | Subject Test Refused |
| Breath Volume | 0.000 | 07:01 | |
| Air Blank | 0.000 | 07:02 | |
| Control Test | 0.080 | 07:02 | |
| Air Blank | 0.000 | 07:02 | |
| Diagnostics Check | OK | 07:03 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,MAREIKE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DREW,MAREIKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 02/20/2020 02:41:20

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 02:29 | |
| Air Blank | 0.000 | 02:30 | |
| Control Test | 0.080 | 02:30 | |
| Air Blank | 0.000 | 02:31 | |
| Subject Sample | # 1 | 02:34 | No Sample Provided |
| Breath Volume | 0.000 | 02:34 | |
| Air Blank | 0.000 | 02:34 | |
| Air Blank | 0.000 | 02:36 | |
| Subject Sample | # 2 | 02:39 | No Sample Provided |
| Breath Volume | 0.000 | 02:39 | |
| Air Blank | 0.000 | 02:40 | |
| Control Test | 0.081 | 02:40 | |
| Air Blank | 0.000 | 02:41 | |
| Diagnostics Check | OK | 02:41 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK, TAVISH, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK, TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LOPEZ CADENA

DATE/TIME OF TEST: 02/20/2020 02:47:01

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 02:46 | |
| Air Blank | | 02:46 | RFI Detect |
| Air Blank | 0.000 | 02:47 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK, TAVISH, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK, TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LOPEZ CADENA

DATE/TIME OF TEST: 02/20/2020 03:05:28

| | | | |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check | OK | 02:48 | |
| Air Blank | 0.000 | 02:49 | |
| Control Test | 0.080 | 02:49 | |
| Air Blank | 0.000 | 02:50 | |
| Subject Sample # 1 | 0.171 | 02:52 | |
| Breath Volume | 1.144 | 02:52 | |
| Air Blank | 0.000 | 02:53 | |
| Air Blank | 0.000 | 02:55 | |
| Subject Sample # 2 | 0.155 | 02:58 | Volume Not Met |
| Breath Volume | 0.535 | 02:58 | |
| Air Blank | 0.000 | 02:59 | |
| Air Blank | 0.000 | 03:00 | |
| Subject Sample # 3 | | 03:04 | No Sample Provided |
| Breath Volume | 0.000 | 03:04 | |
| Air Blank | 0.000 | 03:04 | |
| Control Test | 0.073 | 03:04 | Control Outside Tolerance |
| Air Blank | 0.000 | 03:05 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COOPER,TERRI
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 02/20/2020 12:51:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 12:44 |
| Air Blank | 0.000 | 12:45 |
| Control Test | 0.079 | 12:45 |
| Air Blank | 0.000 | 12:46 |
| Subject Sample # 1 | 0.363 | 12:46 |
| Breath Volume | 2.000 | 12:46 |
| Air Blank | 0.000 | 12:47 |
| Air Blank | 0.000 | 12:49 |
| Subject Sample # 2 | 0.351 | 12:49 |
| Breath Volume | 2.554 | 12:49 |
| Air Blank | 0.000 | 12:50 |
| Control Test | 0.075 | 12:50 |
| Air Blank | 0.000 | 12:51 |
| Diagnostics Check | OK | 12:51 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: SCHUBERT

DATE/TIME OF TEST: 02/21/2020 05:32:29

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:28 | |
| Air Blank | 0.000 | 05:29 | |
| Control Test | 0.081 | 05:29 | |
| Air Blank | 0.000 | 05:30 | |
| Subject Sample | # 1 | 05:31 | Subject Test Refused |
| Breath Volume | 0.000 | 05:31 | |
| Air Blank | 0.000 | 05:31 | |
| Control Test | 0.081 | 05:31 | |
| Air Blank | 0.000 | 05:32 | |
| Diagnostics Check | OK | 05:32 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,A,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMITH,JACKSON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: TRAHAN

DATE/TIME OF TEST: 02/22/2020 05:04:57

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 05:02 | |
| Air Blank | 0.000 | 05:02 | |
| Control Test | 0.082 | 05:02 | |
| Air Blank | 0.000 | 05:03 | |
| Subject Sample | # 1 | 05:03 | Subject Test Refused |
| Breath Volume | 0.000 | 05:03 | |
| Air Blank | 0.000 | 05:04 | |
| Control Test | 0.082 | 05:04 | |
| Air Blank | 0.000 | 05:04 | |
| Diagnostics Check | OK | 05:04 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRISTOL,JARRET,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BRISTOL,JARRET
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: BREEDEN

DATE/TIME OF TEST: 02/22/2020 11:19:03

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 11:11 |
| Air Blank | 0.000 | 11:12 |
| Control Test | 0.083 | 11:12 |
| Air Blank | 0.000 | 11:13 |
| Subject Sample # 1 | 0.120 | 11:14 |
| Breath Volume | 2.378 | 11:14 |
| Air Blank | 0.000 | 11:14 |
| Air Blank | 0.000 | 11:16 |
| Subject Sample # 2 | 0.130 | 11:17 |
| Breath Volume | 2.863 | 11:17 |
| Air Blank | 0.000 | 11:18 |
| Control Test | 0.082 | 11:18 |
| Air Blank | 0.000 | 11:18 |
| Diagnostics Check | OK | 11:19 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON, RAY, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BEMENT, JESSICA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 02/22/2020 23:54:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:45 |
| Air Blank | 0.000 | 23:46 |
| Control Test | 0.081 | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Subject Sample # 1 | 0.207 | 23:49 |
| Breath Volume | 1.320 | 23:49 |
| Air Blank | 0.000 | 23:49 |
| Air Blank | 0.000 | 23:51 |
| Subject Sample # 2 | 0.192 | 23:52 |
| Breath Volume | 2.902 | 23:52 |
| Air Blank | 0.000 | 23:53 |
| Control Test | 0.083 | 23:53 |
| Air Blank | 0.000 | 23:54 |
| Diagnostics Check | OK | 23:54 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/23/2020 01:24:16

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:17 |
| Air Blank | 0.000 | 01:18 |
| Control Test | 0.082 | 01:18 |
| Air Blank | 0.000 | 01:19 |
| Subject Sample # 1 | 0.199 | 01:19 |
| Breath Volume | 4.503 | 01:19 |
| Air Blank | 0.000 | 01:20 |
| Air Blank | 0.000 | 01:22 |
| Subject Sample # 2 | 0.198 | 01:22 |
| Breath Volume | 4.027 | 01:22 |
| Air Blank | 0.000 | 01:23 |
| Control Test | 0.082 | 01:23 |
| Air Blank | 0.000 | 01:24 |
| Diagnostics Check | OK | 01:24 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,IAN,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARRINGTON,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 02/23/2020 19:29:38

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:26 | |
| Air Blank | 0.000 | 19:26 | |
| Control Test | 0.081 | 19:27 | |
| Air Blank | 0.000 | 19:27 | |
| Subject Sample | # 1 | 19:28 | Subject Test Refused |
| Breath Volume | 0.000 | 19:28 | |
| Air Blank | 0.000 | 19:28 | |
| Control Test | 0.081 | 19:29 | |
| Air Blank | 0.000 | 19:29 | |
| Diagnostics Check | OK | 19:29 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON, RAY, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BEMENT, JESSICA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DEARING

DATE/TIME OF TEST: 02/23/2020 21:23:00

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:18 | |
| Air Blank | 0.000 | 21:19 | |
| Control Test | 0.081 | 21:19 | |
| Air Blank | 0.000 | 21:20 | |
| Subject Sample | # 1 | 21:21 | Subject Test Refused |
| Breath Volume | 0.000 | 21:21 | |
| Air Blank | 0.000 | 21:22 | |
| Control Test | 0.081 | 21:22 | |
| Air Blank | 0.000 | 21:22 | |
| Diagnostics Check | OK | 21:23 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,LANCE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ESPARANZA

DATE/TIME OF TEST: 02/23/2020 22:29:53

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:25 | |
| Air Blank | 0.000 | 22:26 | |
| Control Test | 0.081 | 22:26 | |
| Air Blank | 0.000 | 22:26 | |
| Subject Sample | # 1 | 22:28 | Subject Test Refused |
| Breath Volume | 0.000 | 22:28 | |
| Air Blank | 0.000 | 22:28 | |
| Control Test | 0.081 | 22:29 | |
| Air Blank | 0.000 | 22:29 | |
| Diagnostics Check | OK | 22:29 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PAPPAS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: HENN

DATE/TIME OF TEST: 02/24/2020 01:46:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:38 |
| Air Blank | 0.000 | 01:38 |
| Control Test | 0.081 | 01:39 |
| Air Blank | 0.000 | 01:39 |
| Subject Sample # 1 | 0.199 | 01:41 |
| Breath Volume | 1.292 | 01:41 |
| Air Blank | 0.000 | 01:42 |
| Air Blank | 0.000 | 01:44 |
| Subject Sample # 2 | 0.211 | 01:44 |
| Breath Volume | 1.343 | 01:44 |
| Air Blank | 0.000 | 01:45 |
| Control Test | 0.080 | 01:45 |
| Air Blank | 0.000 | 01:46 |
| Diagnostics Check | OK | 01:46 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STEPHENSON,DEBRA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 02/25/2020 02:11:36

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:03 | |
| Air Blank | 0.000 | 02:03 | |
| Control Test | 0.080 | 02:03 | |
| Air Blank | 0.000 | 02:04 | |
| Subject Sample | # 1 | 02:07 | No Sample Provided |
| Breath Volume | 0.000 | 02:07 | |
| Air Blank | 0.000 | 02:08 | |
| Air Blank | 0.000 | 02:10 | |
| Subject Sample | # 2 | 02:10 | Subject Test Refused |
| Breath Volume | 0.000 | 02:10 | |
| Air Blank | 0.000 | 02:10 | |
| Control Test | 0.080 | 02:10 | |
| Air Blank | 0.000 | 02:11 | |
| Diagnostics Check | OK | 02:11 | |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: ZERO TOL
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: LAIN

DATE/TIME OF TEST: 02/27/2020 23:38:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:32 |
| Air Blank | 0.000 | 23:32 |
| Control Test | 0.083 | 23:32 |
| Air Blank | 0.000 | 23:33 |
| Subject Sample # 1 | 0.115 | 23:33 |
| Breath Volume | 1.777 | 23:33 |
| Air Blank | 0.000 | 23:34 |
| Air Blank | 0.000 | 23:36 |
| Subject Sample # 2 | 0.127 | 23:36 |
| Breath Volume | 2.132 | 23:36 |
| Air Blank | 0.000 | 23:37 |
| Control Test | 0.081 | 23:37 |
| Air Blank | 0.000 | 23:38 |
| Diagnostics Check | OK | 23:38 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNHAM,J,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 02/28/2020 02:10:37

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:04 |
| Air Blank | 0.000 | 02:04 |
| Control Test | 0.083 | 02:04 |
| Air Blank | 0.000 | 02:05 |
| Subject Sample # 1 | 0.196 | 02:05 |
| Breath Volume | 1.296 | 02:05 |
| Air Blank | 0.000 | 02:06 |
| Air Blank | 0.000 | 02:08 |
| Subject Sample # 2 | 0.193 | 02:09 |
| Breath Volume | 1.179 | 02:09 |
| Air Blank | 0.000 | 02:09 |
| Control Test | 0.082 | 02:10 |
| Air Blank | 0.000 | 02:10 |
| Diagnostics Check | OK | 02:10 |

Data Download Date/Time: 02/28/2020 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALISON,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HEPBURN,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 08:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/03/2020 10:17:12

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 10:05 | |
| Air Blank | 0.000 | 10:06 | |
| Control Test | 0.081 | 10:06 | |
| Air Blank | 0.000 | 10:07 | |
| Subject Sample | # 1 | 10:10 | No Sample Provided |
| Breath Volume | 0.000 | 10:10 | |
| Air Blank | 0.000 | 10:10 | |
| Air Blank | 0.000 | 10:12 | |
| Subject Sample | # 2 | 10:15 | No Sample Provided |
| Breath Volume | 0.000 | 10:15 | |
| Air Blank | 0.000 | 10:16 | |
| Control Test | 0.082 | 10:16 | |
| Air Blank | 0.000 | 10:17 | |
| Diagnostics Check | OK | 10:17 | |

Data Download Date/Time: 02/14/2020 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: CANFIELD

DATE/TIME OF TEST: 02/10/2020 15:44:55

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 15:38 |
| Air Blank | 0.000 | 15:38 |
| Control Test | 0.080 | 15:38 |
| Air Blank | 0.000 | 15:39 |
| Subject Sample # 1 | 0.328 | 15:40 |
| Breath Volume | 2.429 | 15:40 |
| Air Blank | 0.000 | 15:40 |
| Air Blank | 0.000 | 15:42 |
| Subject Sample # 2 | 0.332 | 15:43 |
| Breath Volume | 1.843 | 15:43 |
| Air Blank | 0.000 | 15:43 |
| Control Test | 0.078 | 15:44 |
| Air Blank | 0.000 | 15:44 |
| Diagnostics Check | OK | 15:44 |

Data Download Date/Time: 02/14/2020 22:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DELORGE

DATE/TIME OF TEST: 02/11/2020 02:51:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:44 |
| Air Blank | 0.000 | 02:45 |
| Control Test | 0.081 | 02:45 |
| Air Blank | 0.000 | 02:46 |
| Subject Sample # 1 | 0.168 | 02:46 |
| Breath Volume | 2.703 | 02:46 |
| Air Blank | 0.000 | 02:47 |
| Air Blank | 0.000 | 02:49 |
| Subject Sample # 2 | 0.161 | 02:49 |
| Breath Volume | 2.359 | 02:49 |
| Air Blank | 0.000 | 02:50 |
| Control Test | 0.078 | 02:50 |
| Air Blank | 0.000 | 02:51 |
| Diagnostics Check | OK | 02:51 |

Data Download Date/Time: 02/14/2020 22:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: GILES JR

DATE/TIME OF TEST: 02/11/2020 20:31:59

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:25 |
| Air Blank | 0.000 | 20:26 |
| Control Test | 0.081 | 20:26 |
| Air Blank | 0.000 | 20:27 |
| Subject Sample # 1 | 0.000 | 20:27 |
| Breath Volume | 4.218 | 20:27 |
| Air Blank | 0.000 | 20:28 |
| Air Blank | 0.000 | 20:30 |
| Subject Sample # 2 | 0.000 | 20:30 |
| Breath Volume | 4.417 | 20:30 |
| Air Blank | 0.000 | 20:31 |
| Control Test | 0.080 | 20:31 |
| Air Blank | 0.000 | 20:31 |
| Diagnostics Check | OK | 20:31 |

Data Download Date/Time: 02/14/2020 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKER,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ACKER,AARON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 792976
DRY GAS EXPIRY DATE: 06/06/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 02/12/2020 03:40:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:30 |
| Air Blank | 0.000 | 03:30 |
| Control Test | 0.082 | 03:31 |
| Air Blank | 0.000 | 03:31 |
| Subject Sample # 1 | 0.168 | 03:33 |
| Breath Volume | 1.183 | 03:33 |
| Air Blank | 0.000 | 03:34 |
| Air Blank | 0.000 | 03:36 |
| Subject Sample # 2 | 0.171 | 03:38 |
| Breath Volume | 1.394 | 03:38 |
| Air Blank | 0.000 | 03:39 |
| Control Test | 0.080 | 03:39 |
| Air Blank | 0.000 | 03:40 |
| Diagnostics Check | OK | 03:40 |

Data Download Date/Time: 02/14/2020 22:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHELLEY,SAMUEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: OLSHEFSKIE

DATE/TIME OF TEST: 02/15/2020 01:13:35

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:05 |
| Air Blank | 0.000 | 01:06 |
| Control Test | 0.082 | 01:06 |
| Air Blank | 0.000 | 01:07 |
| Subject Sample # 1 | 0.095 | 01:08 |
| Breath Volume | 1.531 | 01:08 |
| Air Blank | 0.000 | 01:09 |
| Air Blank | 0.000 | 01:10 |
| Subject Sample # 2 | 0.110 | 01:12 |
| Breath Volume | 2.031 | 01:12 |
| Air Blank | 0.000 | 01:12 |
| Control Test | 0.081 | 01:12 |
| Air Blank | 0.000 | 01:13 |
| Diagnostics Check | OK | 01:13 |

Data Download Date/Time: 03/06/2020 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ZBYDENIEWSKI

DATE/TIME OF TEST: 02/16/2020 02:28:00

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 02:16 | |
| Air Blank | 0.000 | 02:17 | |
| Control Test | 0.082 | 02:17 | |
| Air Blank | 0.000 | 02:17 | |
| Subject Sample | # 1 | 02:21 | No Sample Provided |
| Breath Volume | 0.000 | 02:21 | |
| Air Blank | 0.000 | 02:21 | |
| Air Blank | 0.000 | 02:23 | |
| Subject Sample | # 2 | 02:26 | No Sample Provided |
| Breath Volume | 0.000 | 02:26 | |
| Air Blank | 0.000 | 02:27 | |
| Control Test | 0.082 | 02:27 | |
| Air Blank | 0.000 | 02:27 | |
| Diagnostics Check | OK | 02:28 | |

Data Download Date/Time: 03/06/2020 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 02/22/2020 02:44:24

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:37 |
| Air Blank | 0.000 | 02:37 |
| Control Test | 0.083 | 02:38 |
| Air Blank | 0.000 | 02:38 |
| Subject Sample # 1 | 0.133 | 02:39 |
| Breath Volume | 1.468 | 02:39 |
| Air Blank | 0.000 | 02:40 |
| Air Blank | 0.000 | 02:41 |
| Subject Sample # 2 | 0.137 | 02:42 |
| Breath Volume | 1.601 | 02:42 |
| Air Blank | 0.000 | 02:43 |
| Control Test | 0.082 | 02:43 |
| Air Blank | 0.000 | 02:44 |
| Diagnostics Check | OK | 02:44 |

Data Download Date/Time: 03/06/2020 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SELLERS, DONALD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: DANSBY

DATE/TIME OF TEST: 02/29/2020 23:47:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:37 |
| Air Blank | 0.000 | 23:38 |
| Control Test | 0.083 | 23:38 |
| Air Blank | 0.000 | 23:39 |
| Subject Sample # 1 | 0.252 | 23:39 |
| Breath Volume | 1.613 | 23:39 |
| Air Blank | 0.000 | 23:40 |
| Air Blank | 0.000 | 23:42 |
| Subject Sample # 2 | 0.277 | 23:42 |
| Breath Volume | 1.519 | 23:42 |
| Air Blank | 0.000 | 23:43 |
| Air Blank | 0.000 | 23:45 |
| Subject Sample # 3 | 0.271 | 23:45 |
| Breath Volume | 1.246 | 23:45 |
| Air Blank | 0.000 | 23:46 |
| Control Test | 0.081 | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Diagnostics Check | OK | 23:47 |

Data Download Date/Time: 03/06/2020 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: RAFFERTY,JOSHUA
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 02/01/2020 19:00:35

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:53 |
| Air Blank | 0.000 | 18:53 |
| Control Test | 0.082 | 18:54 |
| Air Blank | 0.000 | 18:54 |
| Subject Sample # 1 | 0.249 | 18:55 |
| Breath Volume | 1.242 | 18:55 |
| Air Blank | 0.000 | 18:55 |
| Air Blank | 0.000 | 18:57 |
| Subject Sample # 2 | 0.248 | 18:58 |
| Breath Volume | 1.378 | 18:58 |
| Air Blank | 0.000 | 18:59 |
| Control Test | 0.079 | 18:59 |
| Air Blank | 0.000 | 19:00 |
| Diagnostics Check | OK | 19:00 |

Data Download Date/Time: 02/14/2020 19:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,MIGUEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ROBLES,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 02/03/2020 23:37:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:30 |
| Air Blank | 0.000 | 23:30 |
| Control Test | 0.079 | 23:30 |
| Air Blank | 0.000 | 23:31 |
| Subject Sample # 1 | 0.168 | 23:32 |
| Breath Volume | 3.363 | 23:32 |
| Air Blank | 0.000 | 23:33 |
| Air Blank | 0.000 | 23:34 |
| Subject Sample # 2 | 0.164 | 23:35 |
| Breath Volume | 4.207 | 23:35 |
| Air Blank | 0.000 | 23:36 |
| Control Test | 0.078 | 23:36 |
| Air Blank | 0.000 | 23:37 |
| Diagnostics Check | OK | 23:37 |

Data Download Date/Time: 02/18/2020 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: RALEY,BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 02/04/2020 01:51:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:44 |
| Air Blank | 0.000 | 01:45 |
| Control Test | 0.079 | 01:45 |
| Air Blank | 0.000 | 01:46 |
| Subject Sample # 1 | 0.167 | 01:46 |
| Breath Volume | 1.683 | 01:46 |
| Air Blank | 0.000 | 01:47 |
| Air Blank | 0.000 | 01:49 |
| Subject Sample # 2 | 0.160 | 01:49 |
| Breath Volume | 1.488 | 01:49 |
| Air Blank | 0.000 | 01:50 |
| Control Test | 0.079 | 01:50 |
| Air Blank | 0.000 | 01:50 |
| Diagnostics Check | OK | 01:51 |

Data Download Date/Time: 02/18/2020 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,DAVID,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HUGHES,DAVID
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 02/09/2020 01:02:14

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:58 | |
| Air Blank | 0.000 | 00:59 | |
| Control Test | 0.081 | 00:59 | |
| Air Blank | 0.000 | 00:59 | |
| Subject Sample | # 1 | 01:00 | Subject Test Refused |
| Breath Volume | 0.000 | 01:00 | |
| Air Blank | 0.000 | 01:01 | |
| Control Test | 0.081 | 01:01 | |
| Air Blank | | 01:01 | RFI Detect |
| Air Blank | 0.000 | 01:02 | |

Data Download Date/Time: 02/18/2020 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DILLON,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GLINERT

DATE/TIME OF TEST: 02/11/2020 00:11:06

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:02 |
| Air Blank | 0.000 | 00:02 |
| Control Test | 0.079 | 00:02 |
| Air Blank | 0.000 | 00:03 |
| Subject Sample # 1 | 0.203 | 00:06 |
| Breath Volume | 2.105 | 00:06 |
| Air Blank | 0.000 | 00:06 |
| Air Blank | 0.000 | 00:08 |
| Subject Sample # 2 | 0.219 | 00:09 |
| Breath Volume | 1.851 | 00:09 |
| Air Blank | 0.000 | 00:10 |
| Control Test | 0.078 | 00:10 |
| Air Blank | 0.000 | 00:11 |
| Diagnostics Check | OK | 00:11 |

Data Download Date/Time: 02/18/2020 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/13/2020 01:54:49

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 01:54 | |
| Air Blank | | 01:54 | RFI Detect |
| Air Blank | 0.000 | 01:54 | |

Data Download Date/Time: 02/18/2020 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/13/2020 02:03:26

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:56 |
| Air Blank | 0.000 | 01:56 |
| Control Test | 0.080 | 01:57 |
| Air Blank | 0.000 | 01:57 |
| Subject Sample # 1 | 0.105 | 01:58 |
| Breath Volume | 1.667 | 01:58 |
| Air Blank | 0.000 | 01:59 |
| Air Blank | 0.000 | 02:01 |
| Subject Sample # 2 | 0.105 | 02:01 |
| Breath Volume | 2.437 | 02:01 |
| Air Blank | 0.000 | 02:02 |
| Control Test | 0.078 | 02:02 |
| Air Blank | 0.000 | 02:03 |
| Diagnostics Check | OK | 02:03 |

Data Download Date/Time: 02/18/2020 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTSCHER,REBECCA,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MYERS,BEN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: MOYER

DATE/TIME OF TEST: 02/15/2020 06:55:21

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 06:43 | |
| Air Blank | | 0.000 | 06:43 | |
| Control Test | | 0.081 | 06:44 | |
| Air Blank | | 0.000 | 06:44 | |
| Subject Sample | # 1 | 0.155 | 06:47 | Volume Not Met |
| Breath Volume | | 0.792 | 06:47 | |
| Air Blank | | 0.000 | 06:48 | |
| Air Blank | | 0.000 | 06:50 | |
| Subject Sample | # 2 | 0.201 | 06:50 | |
| Breath Volume | | 2.179 | 06:50 | |
| Air Blank | | 0.000 | 06:51 | |
| Air Blank | | 0.000 | 06:53 | |
| Subject Sample | # 3 | 0.202 | 06:53 | |
| Breath Volume | | 1.765 | 06:53 | |
| Air Blank | | 0.000 | 06:54 | |
| Control Test | | 0.078 | 06:54 | |
| Air Blank | | 0.000 | 06:55 | |
| Diagnostics Check | | OK | 06:55 | |

Data Download Date/Time: 02/18/2020 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/01/2020 00:32:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:25 |
| Air Blank | 0.000 | 00:26 |
| Control Test | 0.080 | 00:26 |
| Air Blank | 0.000 | 00:27 |
| Subject Sample # 1 | 0.176 | 00:27 |
| Breath Volume | 3.269 | 00:27 |
| Air Blank | 0.000 | 00:28 |
| Air Blank | 0.000 | 00:30 |
| Subject Sample # 2 | 0.174 | 00:30 |
| Breath Volume | 2.710 | 00:30 |
| Air Blank | 0.000 | 00:31 |
| Control Test | 0.080 | 00:31 |
| Air Blank | 0.000 | 00:32 |
| Diagnostics Check | OK | 00:32 |

Data Download Date/Time: 02/27/2020 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRAGUE,JONATHON,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SPRAGUE,JONATHON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: KUMFER

DATE/TIME OF TEST: 02/03/2020 03:13:27

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:06 |
| Air Blank | | 0.000 | 03:07 |
| Control Test | | 0.080 | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Subject Sample | # 1 | 0.082 | 03:08 |
| Breath Volume | | 4.289 | 03:08 |
| Air Blank | | 0.000 | 03:09 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 2 | 0.080 | 03:11 |
| Breath Volume | | 3.343 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Control Test | | 0.080 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Diagnostics Check | | OK | 03:13 |

Data Download Date/Time: 02/27/2020 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: STEPHENSON,PATRICK
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 02/05/2020 16:16:34

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 16:08 |
| Air Blank | | 0.000 | 16:09 |
| Control Test | | 0.080 | 16:09 |
| Air Blank | | 0.000 | 16:09 |
| Subject Sample | # 1 | 0.177 | 16:11 |
| Breath Volume | | 1.359 | 16:11 |
| Air Blank | | 0.000 | 16:12 |
| Air Blank | | 0.000 | 16:14 |
| Subject Sample | # 2 | 0.185 | 16:14 |
| Breath Volume | | 1.343 | 16:14 |
| Air Blank | | 0.000 | 16:15 |
| Control Test | | 0.078 | 16:15 |
| Air Blank | | 0.000 | 16:16 |
| Diagnostics Check | | OK | 16:16 |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,JODY,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: TOLES

DATE/TIME OF TEST: 02/08/2020 01:28:53

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:25 | |
| Air Blank | 0.000 | 01:26 | |
| Control Test | 0.081 | 01:26 | |
| Air Blank | 0.000 | 01:27 | |
| Subject Sample | # 1 | 01:27 | Subject Test Refused |
| Breath Volume | 0.000 | 01:27 | |
| Air Blank | 0.000 | 01:27 | |
| Control Test | 0.080 | 01:28 | |
| Air Blank | 0.000 | 01:28 | |
| Diagnostics Check | OK | 01:28 | |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,JODY,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURGESS,JOSH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HUTCHERSON

DATE/TIME OF TEST: 02/08/2020 03:46:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:40 |
| Air Blank | 0.000 | 03:40 |
| Control Test | 0.081 | 03:40 |
| Air Blank | 0.000 | 03:41 |
| Subject Sample # 1 | 0.120 | 03:41 |
| Breath Volume | 1.570 | 03:41 |
| Air Blank | 0.000 | 03:42 |
| Air Blank | 0.000 | 03:44 |
| Subject Sample # 2 | 0.118 | 03:44 |
| Breath Volume | 1.710 | 03:44 |
| Air Blank | 0.000 | 03:45 |
| Control Test | 0.080 | 03:45 |
| Air Blank | 0.000 | 03:46 |
| Diagnostics Check | OK | 03:46 |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,JODY,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOWARD,JODY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 02/09/2020 02:02:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:55 |
| Air Blank | 0.000 | 01:55 |
| Control Test | 0.080 | 01:56 |
| Air Blank | 0.000 | 01:56 |
| Subject Sample # 1 | 0.214 | 01:57 |
| Breath Volume | 2.230 | 01:57 |
| Air Blank | 0.000 | 01:58 |
| Air Blank | 0.000 | 02:00 |
| Subject Sample # 2 | 0.227 | 02:00 |
| Breath Volume | 2.433 | 02:00 |
| Air Blank | 0.000 | 02:01 |
| Control Test | 0.078 | 02:01 |
| Air Blank | 0.000 | 02:02 |
| Diagnostics Check | OK | 02:02 |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MOORE,JACOB
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: MALAMOS

DATE/TIME OF TEST: 02/11/2020 15:05:39

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 14:58 |
| Air Blank | 0.000 | 14:59 |
| Control Test | 0.081 | 14:59 |
| Air Blank | 0.000 | 15:00 |
| Subject Sample # 1 | 0.053 | 15:01 |
| Breath Volume | 1.445 | 15:01 |
| Air Blank | 0.000 | 15:01 |
| Air Blank | 0.000 | 15:03 |
| Subject Sample # 2 | 0.054 | 15:04 |
| Breath Volume | 2.023 | 15:04 |
| Air Blank | 0.000 | 15:04 |
| Control Test | 0.080 | 15:05 |
| Air Blank | 0.000 | 15:05 |
| Diagnostics Check | OK | 15:05 |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENRICH,KRISTEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 13:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: WESTBROOK

DATE/TIME OF TEST: 02/17/2020 14:22:29

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 14:15 |
| Air Blank | 0.000 | 14:16 |
| Control Test | 0.079 | 14:16 |
| Air Blank | 0.000 | 14:17 |
| Subject Sample # 1 | 0.191 | 14:17 |
| Breath Volume | 4.824 | 14:17 |
| Air Blank | 0.000 | 14:18 |
| Air Blank | 0.000 | 14:20 |
| Subject Sample # 2 | 0.180 | 14:20 |
| Breath Volume | 4.898 | 14:20 |
| Air Blank | 0.000 | 14:21 |
| Control Test | 0.079 | 14:21 |
| Air Blank | 0.000 | 14:22 |
| Diagnostics Check | OK | 14:22 |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DAVID,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPRAGUE,JONATHON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 02/22/2020 01:32:40

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:29 | |
| Air Blank | 0.000 | 01:30 | |
| Control Test | 0.081 | 01:30 | |
| Air Blank | 0.000 | 01:31 | |
| Subject Sample | # 1 | 01:31 | Subject Test Refused |
| Breath Volume | 0.000 | 01:31 | |
| Air Blank | 0.000 | 01:31 | |
| Control Test | 0.081 | 01:32 | |
| Air Blank | 0.000 | 01:32 | |
| Diagnostics Check | OK | 01:32 | |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/22/2020 20:47:30

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:38 |
| Air Blank | 0.000 | 20:39 |
| Control Test | 0.081 | 20:39 |
| Air Blank | 0.000 | 20:40 |
| Subject Sample # 1 | 0.315 | 20:40 |
| Breath Volume | 2.207 | 20:40 |
| Air Blank | 0.000 | 20:41 |
| Air Blank | 0.000 | 20:43 |
| Subject Sample # 2 | 0.325 | 20:45 |
| Breath Volume | 1.343 | 20:45 |
| Air Blank | 0.000 | 20:46 |
| Control Test | 0.078 | 20:46 |
| Air Blank | 0.000 | 20:47 |
| Diagnostics Check | OK | 20:47 |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,JON,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HURST,CARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 02/23/2020 02:57:47

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:51 |
| Air Blank | 0.000 | 02:51 |
| Control Test | 0.082 | 02:52 |
| Air Blank | 0.000 | 02:52 |
| Subject Sample # 1 | 0.125 | 02:53 |
| Breath Volume | 3.464 | 02:53 |
| Air Blank | 0.000 | 02:54 |
| Air Blank | 0.000 | 02:55 |
| Subject Sample # 2 | 0.129 | 02:56 |
| Breath Volume | 2.496 | 02:56 |
| Air Blank | 0.000 | 02:56 |
| Control Test | 0.080 | 02:57 |
| Air Blank | 0.000 | 02:57 |
| Diagnostics Check | OK | 02:57 |

Data Download Date/Time: 02/27/2020 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NOLAND

DATE/TIME OF TEST: 02/06/2020 03:59:36

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:53 |
| Air Blank | 0.000 | 03:53 |
| Control Test | 0.077 | 03:54 |
| Air Blank | 0.000 | 03:54 |
| Subject Sample # 1 | 0.120 | 03:55 |
| Breath Volume | 2.101 | 03:55 |
| Air Blank | 0.000 | 03:55 |
| Air Blank | 0.000 | 03:57 |
| Subject Sample # 2 | 0.121 | 03:57 |
| Breath Volume | 1.953 | 03:57 |
| Air Blank | 0.000 | 03:58 |
| Control Test | 0.077 | 03:58 |
| Air Blank | 0.000 | 03:59 |
| Diagnostics Check | OK | 03:59 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCFARLAND,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 02/08/2020 18:28:05

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 18:25 | |
| Air Blank | 0.000 | 18:25 | |
| Control Test | 0.079 | 18:25 | |
| Air Blank | 0.000 | 18:26 | |
| Subject Sample | # 1 | 18:26 | Subject Test Refused |
| Breath Volume | 0.000 | 18:26 | |
| Air Blank | 0.000 | 18:27 | |
| Control Test | 0.079 | 18:27 | |
| Air Blank | 0.000 | 18:28 | |
| Diagnostics Check | OK | 18:28 | |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HOLMAN

DATE/TIME OF TEST: 02/11/2020 23:05:08

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:58 |
| Air Blank | 0.000 | 22:58 |
| Control Test | 0.079 | 22:58 |
| Air Blank | 0.000 | 22:59 |
| Subject Sample # 1 | 0.245 | 23:00 |
| Breath Volume | 2.390 | 23:00 |
| Air Blank | 0.000 | 23:01 |
| Air Blank | 0.000 | 23:02 |
| Subject Sample # 2 | 0.247 | 23:03 |
| Breath Volume | 2.984 | 23:03 |
| Air Blank | 0.000 | 23:04 |
| Control Test | 0.077 | 23:04 |
| Air Blank | 0.000 | 23:05 |
| Diagnostics Check | OK | 23:05 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,BRIAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STEPHENS,JAKE
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/13/2020 20:47:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:40 |
| Air Blank | 0.000 | 20:41 |
| Control Test | 0.078 | 20:41 |
| Air Blank | 0.000 | 20:42 |
| Subject Sample # 1 | 0.068 | 20:43 |
| Breath Volume | 2.101 | 20:43 |
| Air Blank | 0.000 | 20:43 |
| Air Blank | 0.000 | 20:45 |
| Subject Sample # 2 | 0.060 | 20:46 |
| Breath Volume | 2.093 | 20:46 |
| Air Blank | 0.000 | 20:46 |
| Control Test | 0.076 | 20:47 |
| Air Blank | 0.000 | 20:47 |
| Diagnostics Check | OK | 20:47 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATIVIDAD,DANIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/14/2020 00:48:08

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:45 | |
| Air Blank | 0.000 | 00:45 | |
| Control Test | 0.079 | 00:46 | |
| Air Blank | 0.000 | 00:46 | |
| Subject Sample | # 1 | 00:46 | Subject Test Refused |
| Breath Volume | 0.000 | 00:46 | |
| Air Blank | 0.000 | 00:47 | |
| Control Test | 0.079 | 00:47 | |
| Air Blank | 0.000 | 00:48 | |
| Diagnostics Check | OK | 00:48 | |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: REGISTER

DATE/TIME OF TEST: 02/14/2020 21:56:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:50 |
| Air Blank | 0.000 | 21:50 |
| Control Test | 0.078 | 21:51 |
| Air Blank | 0.000 | 21:51 |
| Subject Sample # 1 | 0.230 | 21:52 |
| Breath Volume | 1.410 | 21:52 |
| Air Blank | 0.000 | 21:53 |
| Air Blank | 0.000 | 21:54 |
| Subject Sample # 2 | 0.244 | 21:54 |
| Breath Volume | 1.468 | 21:54 |
| Air Blank | 0.000 | 21:55 |
| Control Test | 0.077 | 21:56 |
| Air Blank | 0.000 | 21:56 |
| Diagnostics Check | OK | 21:56 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MUNCY

DATE/TIME OF TEST: 02/15/2020 01:33:12

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:24 | |
| Air Blank | 0.000 | 01:24 | |
| Control Test | 0.079 | 01:25 | |
| Air Blank | 0.000 | 01:25 | |
| Subject Sample | # 1 | 01:28 | No Sample Provided |
| Breath Volume | 0.000 | 01:28 | |
| Air Blank | 0.000 | 01:29 | |
| Air Blank | 0.000 | 01:31 | |
| Subject Sample | # 2 | 01:31 | Subject Test Refused |
| Breath Volume | 0.000 | 01:31 | |
| Air Blank | 0.000 | 01:32 | |
| Control Test | 0.080 | 01:32 | |
| Air Blank | 0.000 | 01:33 | |
| Diagnostics Check | OK | 01:33 | |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PLEMMONS,LAQUINTA
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HUTCHESON

DATE/TIME OF TEST: 02/15/2020 02:28:02

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Control Test | | 0.079 | 02:21 |
| Air Blank | | 0.000 | 02:22 |
| Subject Sample | # 1 | 0.151 | 02:23 |
| Breath Volume | | 2.292 | 02:23 |
| Air Blank | | 0.000 | 02:24 |
| Air Blank | | 0.000 | 02:25 |
| Subject Sample | # 2 | 0.153 | 02:26 |
| Breath Volume | | 2.761 | 02:26 |
| Air Blank | | 0.000 | 02:27 |
| Control Test | | 0.077 | 02:27 |
| Air Blank | | 0.000 | 02:27 |
| Diagnostics Check | | OK | 02:28 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: BREEN

DATE/TIME OF TEST: 02/15/2020 20:11:03

| | | | |
|--------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 20:04 | |
| Air Blank | 0.000 | 20:05 | |
| Control Test | 0.079 | 20:05 | |
| Air Blank | 0.000 | 20:06 | |
| Subject Sample # 1 | 0.033 | 20:07 | |
| Breath Volume | 3.976 | 20:07 | |
| Air Blank | 0.000 | 20:07 | |
| Air Blank | 0.000 | 20:09 | |
| Subject Sample # 2 | | 20:09 | Subject Test Refused |
| Breath Volume | 0.000 | 20:09 | |
| Air Blank | 0.000 | 20:10 | |
| Control Test | 0.077 | 20:10 | |
| Air Blank | 0.000 | 20:10 | |
| Diagnostics Check | OK | 20:11 | |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 02/16/2020 03:00:21

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:54 |
| Air Blank | 0.000 | 02:54 |
| Control Test | 0.080 | 02:54 |
| Air Blank | 0.000 | 02:55 |
| Subject Sample # 1 | 0.101 | 02:55 |
| Breath Volume | 2.644 | 02:55 |
| Air Blank | 0.000 | 02:56 |
| Air Blank | 0.000 | 02:58 |
| Subject Sample # 2 | 0.096 | 02:58 |
| Breath Volume | 2.582 | 02:58 |
| Air Blank | 0.000 | 02:59 |
| Control Test | 0.078 | 02:59 |
| Air Blank | 0.000 | 03:00 |
| Diagnostics Check | OK | 03:00 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCKLES,SCHIEFER,S
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: BUCKLES,SCHIEFER
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SERNA

DATE/TIME OF TEST: 02/16/2020 03:57:18

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:50 |
| Air Blank | 0.000 | 03:51 |
| Control Test | 0.079 | 03:51 |
| Air Blank | 0.000 | 03:52 |
| Subject Sample # 1 | 0.126 | 03:52 |
| Breath Volume | 1.308 | 03:52 |
| Air Blank | 0.000 | 03:53 |
| Air Blank | 0.000 | 03:55 |
| Subject Sample # 2 | 0.122 | 03:55 |
| Breath Volume | 2.789 | 03:55 |
| Air Blank | 0.000 | 03:56 |
| Control Test | 0.076 | 03:56 |
| Air Blank | 0.000 | 03:57 |
| Diagnostics Check | OK | 03:57 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/17/2020 00:31:12

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:24 |
| Air Blank | 0.000 | 00:25 |
| Control Test | 0.079 | 00:25 |
| Air Blank | 0.000 | 00:25 |
| Subject Sample # 1 | 0.103 | 00:26 |
| Breath Volume | 3.910 | 00:26 |
| Air Blank | 0.000 | 00:27 |
| Air Blank | 0.000 | 00:28 |
| Subject Sample # 2 | 0.099 | 00:29 |
| Breath Volume | 3.976 | 00:29 |
| Air Blank | 0.000 | 00:30 |
| Control Test | 0.076 | 00:30 |
| Air Blank | 0.000 | 00:31 |
| Diagnostics Check | OK | 00:31 |

Data Download Date/Time: 02/18/2020 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BOUSQUET

DATE/TIME OF TEST: 02/17/2020 01:35:32

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:28 |
| Air Blank | 0.000 | 01:29 |
| Control Test | 0.078 | 01:29 |
| Air Blank | 0.000 | 01:30 |
| Subject Sample # 1 | 0.113 | 01:30 |
| Breath Volume | 3.500 | 01:30 |
| Air Blank | 0.000 | 01:31 |
| Air Blank | 0.000 | 01:33 |
| Subject Sample # 2 | 0.114 | 01:33 |
| Breath Volume | 4.343 | 01:33 |
| Air Blank | 0.000 | 01:34 |
| Control Test | 0.077 | 01:34 |
| Air Blank | 0.000 | 01:35 |
| Diagnostics Check | OK | 01:35 |

Data Download Date/Time: 02/18/2020 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,J,J
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: DUPREE,A
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 02/17/2020 18:45:35

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:38 |
| Air Blank | 0.000 | 18:39 |
| Control Test | 0.078 | 18:39 |
| Air Blank | 0.000 | 18:40 |
| Subject Sample # 1 | 0.191 | 18:40 |
| Breath Volume | 3.179 | 18:40 |
| Air Blank | 0.000 | 18:41 |
| Air Blank | 0.000 | 18:43 |
| Subject Sample # 2 | 0.189 | 18:43 |
| Breath Volume | 3.894 | 18:43 |
| Air Blank | 0.000 | 18:44 |
| Control Test | 0.075 | 18:44 |
| Air Blank | 0.000 | 18:45 |
| Diagnostics Check | OK | 18:45 |

Data Download Date/Time: 02/18/2020 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 02/26/2020 16:38:15

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 16:32 |
| Air Blank | 0.000 | 16:32 |
| Control Test | 0.079 | 16:33 |
| Air Blank | 0.000 | 16:33 |
| Subject Sample # 1 | 0.000 | 16:34 |
| Breath Volume | 2.000 | 16:34 |
| Air Blank | 0.000 | 16:34 |
| Air Blank | 0.000 | 16:36 |
| Subject Sample # 2 | 0.000 | 16:36 |
| Breath Volume | 1.968 | 16:36 |
| Air Blank | 0.000 | 16:37 |
| Control Test | 0.077 | 16:37 |
| Air Blank | 0.000 | 16:38 |
| Diagnostics Check | OK | 16:38 |

Data Download Date/Time: 04/03/2020 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 02/26/2020 22:24:10

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 22:21 | |
| Air Blank | 0.000 | 22:21 | |
| Control Test | 0.079 | 22:22 | |
| Air Blank | 0.000 | 22:22 | |
| Subject Sample | # 1 | 22:22 | Subject Test Refused |
| Breath Volume | 0.000 | 22:22 | |
| Air Blank | 0.000 | 22:23 | |
| Control Test | 0.079 | 22:23 | |
| Air Blank | 0.000 | 22:24 | |
| Diagnostics Check | OK | 22:24 | |

Data Download Date/Time: 04/03/2020 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALDREE,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: STIRN

DATE/TIME OF TEST: 02/28/2020 19:04:46

| | | | |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK | 19:00 | |
| Air Blank | 0.000 | 19:00 | |
| Control Test | 0.081 | 19:01 | |
| Air Blank | 0.000 | 19:01 | |
| Subject Sample | # 1 0.330 | 19:03 | |
| Breath Volume | 1.925 | 19:03 | |
| Air Blank | | 19:04 | Purge Fail |
| Air Blank | 0.000 | 19:04 | |

Data Download Date/Time: 04/03/2020 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALDREE,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 915310
DRY GAS EXPIRY DATE: 03/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: STIRN

DATE/TIME OF TEST: 02/28/2020 19:24:09

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:15 |
| Air Blank | 0.000 | 19:15 |
| Control Test | 0.079 | 19:15 |
| Air Blank | 0.000 | 19:16 |
| Subject Sample # 1 | 0.319 | 19:16 |
| Breath Volume | 1.468 | 19:16 |
| Air Blank | 0.000 | 19:17 |
| Air Blank | 0.000 | 19:19 |
| Subject Sample # 2 | 0.328 | 19:22 |
| Breath Volume | 2.101 | 19:22 |
| Air Blank | 0.000 | 19:23 |
| Control Test | 0.078 | 19:23 |
| Air Blank | 0.000 | 19:24 |
| Diagnostics Check | OK | 19:24 |

Data Download Date/Time: 04/03/2020 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIN,ALBERT,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: QUATER,ANTHONY
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLEMAN MURILLO

DATE/TIME OF TEST: 02/01/2020 01:12:03

| | | | |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 01:11 | |
| Air Blank | | 01:11 | RFI Detect |
| Air Blank | | 01:12 | RFI Detect |

Data Download Date/Time: 02/13/2020 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIN,ALBERT,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: QUATER,ANTHONY
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLEMAN MURILLO

DATE/TIME OF TEST: 02/01/2020 01:21:39

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 01:15 | |
| Air Blank | 0.000 | 01:16 | |
| Control Test | 0.078 | 01:16 | |
| Air Blank | 0.000 | 01:17 | |
| Subject Sample | # 1 | 01:19 | RFI Detect |
| Breath Volume | 0.695 | 01:19 | |
| Air Blank | 0.000 | 01:20 | |
| Control Test | 0.079 | 01:21 | |
| Air Blank | 0.000 | 01:21 | |
| Diagnostics Check | OK | 01:21 | |

Data Download Date/Time: 02/13/2020 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIN,ALBERT,L
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: QUATER,ANTHONY
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COLEMAN MURILLO

DATE/TIME OF TEST: 02/01/2020 01:31:30

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:23 |
| Air Blank | 0.000 | 01:24 |
| Control Test | 0.076 | 01:24 |
| Air Blank | 0.000 | 01:25 |
| Subject Sample # 1 | 0.123 | 01:25 |
| Breath Volume | 3.410 | 01:25 |
| Air Blank | 0.000 | 01:26 |
| Air Blank | 0.000 | 01:27 |
| Subject Sample # 2 | 0.111 | 01:29 |
| Breath Volume | 2.363 | 01:29 |
| Air Blank | 0.000 | 01:30 |
| Control Test | 0.076 | 01:30 |
| Air Blank | 0.000 | 01:31 |
| Diagnostics Check | OK | 01:31 |

Data Download Date/Time: 02/13/2020 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE, RAYMOND, C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WHITE, RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/12/2020 20:02:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:55 |
| Air Blank | 0.000 | 19:56 |
| Control Test | 0.080 | 19:56 |
| Air Blank | 0.000 | 19:57 |
| Subject Sample # 1 | 0.226 | 19:58 |
| Breath Volume | 3.652 | 19:58 |
| Air Blank | 0.000 | 19:58 |
| Air Blank | 0.000 | 20:00 |
| Subject Sample # 2 | 0.206 | 20:01 |
| Breath Volume | 1.812 | 20:01 |
| Air Blank | 0.000 | 20:01 |
| Control Test | 0.078 | 20:02 |
| Air Blank | 0.000 | 20:02 |
| Diagnostics Check | OK | 20:02 |

Data Download Date/Time: 02/13/2020 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DAVID,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH,DAVID
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 02/19/2020 22:51:46

| | | | |
|-------------------|-----------|-------|------------------|
| Diagnostics Check | OK | 22:43 | |
| Air Blank | 0.000 | 22:44 | |
| Control Test | 0.079 | 22:44 | |
| Air Blank | 0.000 | 22:44 | |
| Subject Sample | # 1 0.170 | 22:48 | Volume Not Met |
| Breath Volume | 0.531 | 22:48 | |
| Air Blank | 0.000 | 22:48 | |
| Air Blank | 0.000 | 22:50 | |
| Subject Sample | # 2 | 22:51 | Sequence Aborted |
| Breath Volume | 0.000 | 22:51 | |
| Air Blank | 0.000 | 22:51 | |

Data Download Date/Time: 03/10/2020 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OPRY, WILLIAM, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: OPRY, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 02/08/2020 02:41:08

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:34 |
| Air Blank | 0.000 | 02:34 |
| Control Test | 0.079 | 02:35 |
| Air Blank | 0.000 | 02:35 |
| Subject Sample # 1 | 0.101 | 02:36 |
| Breath Volume | 1.167 | 02:36 |
| Air Blank | 0.000 | 02:37 |
| Air Blank | 0.000 | 02:39 |
| Subject Sample # 2 | 0.098 | 02:39 |
| Breath Volume | 1.878 | 02:39 |
| Air Blank | 0.000 | 02:40 |
| Control Test | 0.078 | 02:40 |
| Air Blank | 0.000 | 02:41 |
| Diagnostics Check | OK | 02:41 |

Data Download Date/Time: 02/14/2020 19:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OPRY, WILLIAM, J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: OPRY, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 02/08/2020 03:01:07

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:58 | |
| Air Blank | 0.000 | 02:58 | |
| Control Test | 0.078 | 02:58 | |
| Air Blank | 0.000 | 02:59 | |
| Subject Sample | # 1 | 02:59 | Subject Test Refused |
| Breath Volume | 0.000 | 02:59 | |
| Air Blank | 0.000 | 03:00 | |
| Control Test | 0.078 | 03:00 | |
| Air Blank | 0.000 | 03:01 | |
| Diagnostics Check | OK | 03:01 | |

Data Download Date/Time: 02/14/2020 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON, JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SOLOMON, JR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/17/2020 21:46:44

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:40 |
| Air Blank | 0.000 | 21:41 |
| Control Test | 0.079 | 21:41 |
| Air Blank | 0.000 | 21:42 |
| Subject Sample # 1 | 0.000 | 21:42 |
| Breath Volume | 1.617 | 21:42 |
| Air Blank | 0.000 | 21:43 |
| Air Blank | 0.000 | 21:45 |
| Subject Sample # 2 | 0.000 | 21:45 |
| Breath Volume | 2.132 | 21:45 |
| Air Blank | 0.000 | 21:45 |
| Control Test | 0.079 | 21:46 |
| Air Blank | 0.000 | 21:46 |
| Diagnostics Check | OK | 21:46 |

Data Download Date/Time: 03/05/2020 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JONATHAN, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/07/2020
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TURNEY

DATE/TIME OF TEST: 02/11/2020 18:59:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:51 |
| Air Blank | 0.000 | 18:51 |
| Control Test | 0.084 | 18:51 |
| Air Blank | 0.000 | 18:52 |
| Subject Sample # 1 | 0.192 | 18:54 |
| Breath Volume | 1.843 | 18:54 |
| Air Blank | 0.000 | 18:55 |
| Air Blank | 0.000 | 18:56 |
| Subject Sample # 2 | 0.188 | 18:57 |
| Breath Volume | 1.648 | 18:57 |
| Air Blank | 0.000 | 18:58 |
| Control Test | 0.079 | 18:59 |
| Air Blank | 0.000 | 18:59 |
| Diagnostics Check | OK | 18:59 |

Data Download Date/Time: 02/12/2020 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/07/2020
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: COTTLE

DATE/TIME OF TEST: 02/14/2020 07:40:42

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 07:33 |
| Air Blank | | 0.000 | 07:34 |
| Control Test | | 0.083 | 07:34 |
| Air Blank | | 0.000 | 07:34 |
| Subject Sample | # 1 | 0.228 | 07:35 |
| Breath Volume | | 4.050 | 07:35 |
| Air Blank | | 0.000 | 07:36 |
| Air Blank | | 0.000 | 07:38 |
| Subject Sample | # 2 | 0.230 | 07:38 |
| Breath Volume | | 3.929 | 07:38 |
| Air Blank | | 0.000 | 07:39 |
| Control Test | | 0.078 | 07:40 |
| Air Blank | | 0.000 | 07:40 |
| Diagnostics Check | | OK | 07:40 |

Data Download Date/Time: 03/02/2020 03:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/07/2020
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CEYDELI

DATE/TIME OF TEST: 02/22/2020 02:26:02

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:23 | |
| Air Blank | 0.000 | 02:23 | |
| Control Test | 0.084 | 02:23 | |
| Air Blank | 0.000 | 02:24 | |
| Subject Sample | # 1 | 02:24 | Subject Test Refused |
| Breath Volume | 0.000 | 02:24 | |
| Air Blank | 0.000 | 02:25 | |
| Control Test | 0.084 | 02:25 | |
| Air Blank | 0.000 | 02:25 | |
| Diagnostics Check | OK | 02:26 | |

Data Download Date/Time: 03/02/2020 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/07/2020
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BAXLEY

DATE/TIME OF TEST: 02/23/2020 20:53:52

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:45 |
| Air Blank | 0.000 | 20:46 |
| Control Test | 0.083 | 20:46 |
| Air Blank | 0.000 | 20:47 |
| Subject Sample # 1 | 0.099 | 20:47 |
| Breath Volume | 1.230 | 20:47 |
| Air Blank | 0.000 | 20:48 |
| Air Blank | 0.000 | 20:50 |
| Subject Sample # 2 | 0.104 | 20:52 |
| Breath Volume | 1.589 | 20:52 |
| Air Blank | 0.000 | 20:52 |
| Control Test | 0.082 | 20:53 |
| Air Blank | 0.000 | 20:53 |
| Diagnostics Check | OK | 20:53 |

Data Download Date/Time: 03/02/2020 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JONATHAN, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/07/2020
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: FORD II

DATE/TIME OF TEST: 02/28/2020 18:47:47

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:40 |
| Air Blank | 0.000 | 18:40 |
| Control Test | 0.084 | 18:41 |
| Air Blank | 0.000 | 18:41 |
| Subject Sample # 1 | 0.335 | 18:42 |
| Breath Volume | 4.187 | 18:42 |
| Air Blank | 0.000 | 18:43 |
| Air Blank | 0.000 | 18:45 |
| Subject Sample # 2 | 0.324 | 18:45 |
| Breath Volume | 3.414 | 18:45 |
| Air Blank | 0.000 | 18:46 |
| Control Test | 0.080 | 18:47 |
| Air Blank | 0.000 | 18:47 |
| Diagnostics Check | OK | 18:47 |

Data Download Date/Time: 03/02/2020 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/07/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/29/2020 04:18:30

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:09 |
| Air Blank | 0.000 | 04:10 |
| Control Test | 0.085 | 04:10 |
| Air Blank | 0.000 | 04:11 |
| Subject Sample # 1 | 0.179 | 04:13 |
| Breath Volume | 1.523 | 04:13 |
| Air Blank | 0.000 | 04:14 |
| Air Blank | 0.000 | 04:16 |
| Subject Sample # 2 | 0.178 | 04:16 |
| Breath Volume | 1.480 | 04:16 |
| Air Blank | 0.000 | 04:17 |
| Control Test | 0.081 | 04:17 |
| Air Blank | 0.000 | 04:18 |
| Diagnostics Check | OK | 04:18 |

Data Download Date/Time: 03/02/2020 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/07/2020
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/29/2020 17:02:51

| | | | |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK | 16:55 | |
| Air Blank | 0.000 | 16:55 | |
| Control Test | 0.083 | 16:56 | |
| Air Blank | 0.000 | 16:56 | |
| Subject Sample | # 1 0.266 | 16:57 | |
| Breath Volume | 1.191 | 16:57 | |
| Air Blank | 0.000 | 16:58 | |
| Air Blank | 0.000 | 17:00 | |
| Subject Sample | # 2 0.253 | 17:01 | |
| Breath Volume | 1.488 | 17:01 | |
| Air Blank | 0.000 | 17:01 | |
| Control Test | 0.080 | 17:02 | Interferent Detect |
| Air Blank | 0.000 | 17:02 | |

Data Download Date/Time: 03/02/2020 03:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TATRO,KERSTAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 02/07/2020 22:44:59

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:38 |
| Air Blank | 0.000 | 22:38 |
| Control Test | 0.080 | 22:38 |
| Air Blank | 0.000 | 22:39 |
| Subject Sample # 1 | 0.084 | 22:40 |
| Breath Volume | 1.367 | 22:40 |
| Air Blank | 0.000 | 22:41 |
| Air Blank | 0.000 | 22:42 |
| Subject Sample # 2 | 0.086 | 22:43 |
| Breath Volume | 2.289 | 22:43 |
| Air Blank | 0.000 | 22:44 |
| Control Test | 0.080 | 22:44 |
| Air Blank | 0.000 | 22:44 |
| Diagnostics Check | OK | 22:44 |

Data Download Date/Time: 02/10/2020 02:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CONDREAY

DATE/TIME OF TEST: 02/09/2020 04:03:53

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:57 |
| Air Blank | | 0.000 | 03:57 |
| Control Test | | 0.079 | 03:58 |
| Air Blank | | 0.000 | 03:58 |
| Subject Sample | # 1 | 0.245 | 03:59 |
| Breath Volume | | 2.027 | 03:59 |
| Air Blank | | 0.000 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Subject Sample | # 2 | 0.254 | 04:02 |
| Breath Volume | | 2.296 | 04:02 |
| Air Blank | | 0.000 | 04:02 |
| Control Test | | 0.080 | 04:03 |
| Air Blank | | 0.000 | 04:03 |
| Diagnostics Check | | OK | 04:03 |

Data Download Date/Time: 02/10/2020 02:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WHITE,TRAVIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: FRILOUX

DATE/TIME OF TEST: 02/09/2020 04:55:35

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:49 |
| Air Blank | | 0.000 | 04:49 |
| Control Test | | 0.079 | 04:50 |
| Air Blank | | 0.000 | 04:50 |
| Subject Sample | # 1 | 0.048 | 04:51 |
| Breath Volume | | 2.300 | 04:51 |
| Air Blank | | 0.000 | 04:51 |
| Air Blank | | 0.000 | 04:53 |
| Subject Sample | # 2 | 0.046 | 04:54 |
| Breath Volume | | 2.679 | 04:54 |
| Air Blank | | 0.000 | 04:54 |
| Control Test | | 0.079 | 04:54 |
| Air Blank | | 0.000 | 04:55 |
| Diagnostics Check | | OK | 04:55 |

Data Download Date/Time: 02/10/2020 02:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/10/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/13/2020 04:28:28

| | | | |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK | 04:22 | |
| Air Blank | 0.000 | 04:23 | |
| Control Test | 0.079 | 04:23 | |
| Air Blank | 0.000 | 04:23 | |
| Subject Sample | # 1 0.163 | 04:27 | Volume Not Met |
| Breath Volume | 0.992 | 04:27 | |
| Air Blank | | 04:27 | Purge Fail |
| Air Blank | 0.000 | 04:28 | |

Data Download Date/Time: 03/03/2020 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/13/2020 04:35:59

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 04:31 | |
| Air Blank | 0.000 | 04:31 | |
| Control Test | 0.078 | 04:32 | |
| Air Blank | 0.000 | 04:32 | |
| Subject Sample | # 1 | 04:35 | Subject Test Refused |
| Breath Volume | 1.167 | 04:35 | |
| Air Blank | | 04:35 | RFI Detect |
| Air Blank | 0.000 | 04:35 | |

Data Download Date/Time: 03/03/2020 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 02/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 02/28/2020 19:11:26

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:05 |
| Air Blank | 0.000 | 19:05 |
| Control Test | 0.080 | 19:06 |
| Air Blank | 0.000 | 19:06 |
| Subject Sample # 1 | 0.000 | 19:07 |
| Breath Volume | 4.183 | 19:07 |
| Air Blank | 0.000 | 19:07 |
| Air Blank | 0.000 | 19:09 |
| Subject Sample # 2 | 0.000 | 19:10 |
| Breath Volume | 3.355 | 19:10 |
| Air Blank | 0.000 | 19:10 |
| Control Test | 0.081 | 19:10 |
| Air Blank | 0.000 | 19:11 |
| Diagnostics Check | OK | 19:11 |

Data Download Date/Time: 03/03/2020 00:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: LOSEY

DATE/TIME OF TEST: 02/03/2020 23:52:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:45 |
| Air Blank | 0.000 | 23:46 |
| Control Test | 0.080 | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Subject Sample # 1 | 0.138 | 23:47 |
| Breath Volume | 1.558 | 23:47 |
| Air Blank | 0.000 | 23:48 |
| Air Blank | 0.000 | 23:50 |
| Subject Sample # 2 | 0.135 | 23:50 |
| Breath Volume | 1.824 | 23:50 |
| Air Blank | 0.000 | 23:51 |
| Control Test | 0.080 | 23:51 |
| Air Blank | 0.000 | 23:52 |
| Diagnostics Check | OK | 23:52 |

Data Download Date/Time: 02/04/2020 06:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: RAYA RAYA

DATE/TIME OF TEST: 02/04/2020 21:32:32

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:26 |
| Air Blank | 0.000 | 21:26 |
| Control Test | 0.080 | 21:27 |
| Air Blank | 0.000 | 21:27 |
| Subject Sample # 1 | 0.109 | 21:28 |
| Breath Volume | 2.925 | 21:28 |
| Air Blank | 0.000 | 21:28 |
| Air Blank | 0.000 | 21:30 |
| Subject Sample # 2 | 0.111 | 21:31 |
| Breath Volume | 3.035 | 21:31 |
| Air Blank | 0.000 | 21:31 |
| Control Test | 0.080 | 21:31 |
| Air Blank | 0.000 | 21:32 |
| Diagnostics Check | OK | 21:32 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ECKRODE,ALBERT
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SICES

DATE/TIME OF TEST: 02/05/2020 05:10:49

| | | | | |
|-------------------|-----|-------|-------|---------------|
| Diagnostics Check | | OK | 05:02 | |
| Air Blank | | 0.000 | 05:03 | |
| Control Test | | 0.080 | 05:03 | |
| Air Blank | | 0.000 | 05:04 | |
| Subject Sample | # 1 | 0.288 | 05:06 | |
| Breath Volume | | 1.902 | 05:06 | |
| Air Blank | | 0.000 | 05:07 | |
| Air Blank | | 0.000 | 05:08 | |
| Subject Sample | # 2 | | 05:09 | Slope Not Met |
| Breath Volume | | 0.625 | 05:09 | |
| Air Blank | | 0.000 | 05:09 | |
| Control Test | | 0.079 | 05:10 | |
| Air Blank | | 0.000 | 05:10 | |
| Diagnostics Check | | OK | 05:10 | |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELPH,JASON,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SELPH,JASON
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 02/06/2020 15:52:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 15:46 |
| Air Blank | 0.000 | 15:46 |
| Control Test | 0.079 | 15:46 |
| Air Blank | 0.000 | 15:47 |
| Subject Sample # 1 | 0.000 | 15:48 |
| Breath Volume | 2.816 | 15:48 |
| Air Blank | 0.000 | 15:48 |
| Air Blank | 0.000 | 15:50 |
| Subject Sample # 2 | 0.000 | 15:50 |
| Breath Volume | 2.746 | 15:50 |
| Air Blank | 0.000 | 15:51 |
| Control Test | 0.079 | 15:51 |
| Air Blank | 0.000 | 15:52 |
| Diagnostics Check | OK | 15:52 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: QUILES,ALEXANDER
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LEROY

DATE/TIME OF TEST: 02/12/2020 03:19:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:12 |
| Air Blank | 0.000 | 03:12 |
| Control Test | 0.079 | 03:13 |
| Air Blank | 0.000 | 03:13 |
| Subject Sample # 1 | 0.143 | 03:14 |
| Breath Volume | 2.445 | 03:14 |
| Air Blank | 0.000 | 03:15 |
| Air Blank | 0.000 | 03:17 |
| Subject Sample # 2 | 0.146 | 03:17 |
| Breath Volume | 1.632 | 03:17 |
| Air Blank | 0.000 | 03:18 |
| Control Test | 0.079 | 03:18 |
| Air Blank | 0.000 | 03:19 |
| Diagnostics Check | OK | 03:19 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/13/2020 19:19:55

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:13 |
| Air Blank | 0.000 | 19:14 |
| Control Test | 0.079 | 19:14 |
| Air Blank | 0.000 | 19:15 |
| Subject Sample # 1 | 0.000 | 19:15 |
| Breath Volume | 2.500 | 19:15 |
| Air Blank | 0.000 | 19:16 |
| Air Blank | 0.000 | 19:18 |
| Subject Sample # 2 | 0.000 | 19:18 |
| Breath Volume | 2.625 | 19:18 |
| Air Blank | 0.000 | 19:19 |
| Control Test | 0.080 | 19:19 |
| Air Blank | 0.000 | 19:19 |
| Diagnostics Check | OK | 19:19 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/14/2020 21:14:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:06 |
| Air Blank | 0.000 | 21:06 |
| Control Test | 0.080 | 21:07 |
| Air Blank | 0.000 | 21:07 |
| Subject Sample # 1 | 0.190 | 21:09 |
| Breath Volume | 1.480 | 21:09 |
| Air Blank | 0.000 | 21:09 |
| Air Blank | 0.000 | 21:11 |
| Subject Sample # 2 | 0.199 | 21:12 |
| Breath Volume | 1.777 | 21:12 |
| Air Blank | 0.000 | 21:13 |
| Control Test | 0.080 | 21:13 |
| Air Blank | 0.000 | 21:14 |
| Diagnostics Check | OK | 21:14 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SUHLING,TASHA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SOMMER

DATE/TIME OF TEST: 02/15/2020 01:40:13

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:33 |
| Air Blank | 0.000 | 01:34 |
| Control Test | 0.080 | 01:34 |
| Air Blank | 0.000 | 01:35 |
| Subject Sample # 1 | 0.094 | 01:35 |
| Breath Volume | 2.835 | 01:35 |
| Air Blank | 0.000 | 01:36 |
| Air Blank | 0.000 | 01:38 |
| Subject Sample # 2 | 0.097 | 01:38 |
| Breath Volume | 3.066 | 01:38 |
| Air Blank | 0.000 | 01:39 |
| Control Test | 0.080 | 01:39 |
| Air Blank | 0.000 | 01:40 |
| Diagnostics Check | OK | 01:40 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: QUILES,ALEXANDER
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: PINE

DATE/TIME OF TEST: 02/16/2020 18:42:57

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:36 |
| Air Blank | 0.000 | 18:37 |
| Control Test | 0.079 | 18:37 |
| Air Blank | 0.000 | 18:37 |
| Subject Sample # 1 | 0.169 | 18:38 |
| Breath Volume | 3.714 | 18:38 |
| Air Blank | 0.000 | 18:39 |
| Air Blank | 0.000 | 18:40 |
| Subject Sample # 2 | 0.159 | 18:41 |
| Breath Volume | 2.843 | 18:41 |
| Air Blank | 0.000 | 18:42 |
| Control Test | 0.079 | 18:42 |
| Air Blank | 0.000 | 18:42 |
| Diagnostics Check | OK | 18:42 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 02/17/2020 00:57:16

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:51 |
| Air Blank | 0.000 | 00:51 |
| Control Test | 0.080 | 00:51 |
| Air Blank | 0.000 | 00:52 |
| Subject Sample # 1 | 0.000 | 00:52 |
| Breath Volume | 2.390 | 00:52 |
| Air Blank | 0.000 | 00:53 |
| Air Blank | 0.000 | 00:55 |
| Subject Sample # 2 | 0.000 | 00:55 |
| Breath Volume | 2.507 | 00:55 |
| Air Blank | 0.000 | 00:56 |
| Control Test | 0.080 | 00:56 |
| Air Blank | 0.000 | 00:57 |
| Diagnostics Check | OK | 00:57 |

Data Download Date/Time: 02/26/2020 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/19/2020 01:26:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:20 |
| Air Blank | 0.000 | 01:20 |
| Control Test | 0.080 | 01:20 |
| Air Blank | 0.000 | 01:21 |
| Subject Sample # 1 | 0.053 | 01:21 |
| Breath Volume | 2.476 | 01:21 |
| Air Blank | 0.000 | 01:22 |
| Air Blank | 0.000 | 01:24 |
| Subject Sample # 2 | 0.052 | 01:24 |
| Breath Volume | 2.628 | 01:24 |
| Air Blank | 0.000 | 01:25 |
| Control Test | 0.080 | 01:25 |
| Air Blank | 0.000 | 01:26 |
| Diagnostics Check | OK | 01:26 |

Data Download Date/Time: 02/26/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: QUILES,ALEXANDER
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GRIMSLEY

DATE/TIME OF TEST: 02/20/2020 21:59:24

| | | | |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK | 21:56 | |
| Air Blank | 0.000 | 21:57 | |
| Control Test | 0.079 | 21:57 | |
| Air Blank | 0.000 | 21:58 | |
| Subject Sample | # 1 | 21:58 | Sequence Aborted |
| Breath Volume | 0.000 | 21:58 | |
| Air Blank | 0.000 | 21:59 | |

Data Download Date/Time: 02/26/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: RIVERA AQUINO

DATE/TIME OF TEST: 02/21/2020 23:18:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:11 |
| Air Blank | 0.000 | 23:12 |
| Control Test | 0.080 | 23:12 |
| Air Blank | 0.000 | 23:12 |
| Subject Sample # 1 | 0.056 | 23:13 |
| Breath Volume | 1.609 | 23:13 |
| Air Blank | 0.000 | 23:14 |
| Air Blank | 0.000 | 23:16 |
| Subject Sample # 2 | 0.058 | 23:17 |
| Breath Volume | 1.644 | 23:17 |
| Air Blank | 0.000 | 23:17 |
| Control Test | 0.080 | 23:18 |
| Air Blank | 0.000 | 23:18 |
| Diagnostics Check | OK | 23:18 |

Data Download Date/Time: 02/26/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MAGANA ANDRADE

DATE/TIME OF TEST: 02/23/2020 00:43:47

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:39 | |
| Air Blank | 0.000 | 00:40 | |
| Control Test | 0.080 | 00:40 | |
| Air Blank | 0.000 | 00:40 | |
| Subject Sample | # 1 | 00:42 | Subject Test Refused |
| Breath Volume | 0.000 | 00:42 | |
| Air Blank | 0.000 | 00:42 | |
| Control Test | 0.080 | 00:43 | |
| Air Blank | 0.000 | 00:43 | |
| Diagnostics Check | OK | 00:43 | |

Data Download Date/Time: 02/26/2020 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CASSERLY,STEPHEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 02/24/2020 21:35:51

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:29 |
| Air Blank | 0.000 | 21:30 |
| Control Test | 0.080 | 21:30 |
| Air Blank | 0.000 | 21:30 |
| Subject Sample # 1 | 0.173 | 21:31 |
| Breath Volume | 1.621 | 21:31 |
| Air Blank | 0.000 | 21:32 |
| Air Blank | 0.000 | 21:33 |
| Subject Sample # 2 | 0.155 | 21:34 |
| Breath Volume | 1.687 | 21:34 |
| Air Blank | 0.000 | 21:34 |
| Control Test | 0.078 | 21:35 |
| Air Blank | 0.000 | 21:35 |
| Diagnostics Check | OK | 21:35 |

Data Download Date/Time: 02/26/2020 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 02/26/2020 20:23:40

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:17 |
| Air Blank | 0.000 | 20:17 |
| Control Test | 0.079 | 20:18 |
| Air Blank | 0.000 | 20:18 |
| Subject Sample # 1 | 0.295 | 20:19 |
| Breath Volume | 1.871 | 20:19 |
| Air Blank | 0.000 | 20:19 |
| Air Blank | 0.000 | 20:21 |
| Subject Sample # 2 | 0.284 | 20:22 |
| Breath Volume | 1.839 | 20:22 |
| Air Blank | 0.000 | 20:22 |
| Control Test | 0.079 | 20:23 |
| Air Blank | 0.000 | 20:23 |
| Diagnostics Check | OK | 20:23 |

Data Download Date/Time: 03/13/2020 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CHOUINARD,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: RAYAAZ

DATE/TIME OF TEST: 02/27/2020 03:30:18

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:23 |
| Air Blank | 0.000 | 03:23 |
| Control Test | 0.079 | 03:23 |
| Air Blank | 0.000 | 03:24 |
| Subject Sample # 1 | 0.215 | 03:25 |
| Breath Volume | 2.203 | 03:25 |
| Air Blank | 0.000 | 03:26 |
| Air Blank | 0.000 | 03:28 |
| Subject Sample # 2 | 0.205 | 03:28 |
| Breath Volume | 1.835 | 03:28 |
| Air Blank | 0.000 | 03:29 |
| Control Test | 0.077 | 03:29 |
| Air Blank | 0.000 | 03:30 |
| Diagnostics Check | OK | 03:30 |

Data Download Date/Time: 03/13/2020 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/02/2020 04:03:14

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:56 |
| Air Blank | 0.000 | 03:56 |
| Control Test | 0.079 | 03:56 |
| Air Blank | 0.000 | 03:57 |
| Subject Sample # 1 | 0.199 | 03:58 |
| Breath Volume | 1.726 | 03:58 |
| Air Blank | 0.000 | 03:59 |
| Air Blank | 0.000 | 04:00 |
| Subject Sample # 2 | 0.186 | 04:01 |
| Breath Volume | 2.375 | 04:01 |
| Air Blank | 0.000 | 04:02 |
| Control Test | 0.078 | 04:02 |
| Air Blank | 0.000 | 04:03 |
| Diagnostics Check | OK | 04:03 |

Data Download Date/Time: 02/03/2020 22:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CZABAN,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: HOWELL,DANIEL
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 02/02/2020 14:37:58

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 14:30 |
| Air Blank | | 0.000 | 14:31 |
| Control Test | | 0.078 | 14:31 |
| Air Blank | | 0.000 | 14:31 |
| Subject Sample | # 1 | 0.225 | 14:33 |
| Breath Volume | | 2.988 | 14:33 |
| Air Blank | | 0.000 | 14:33 |
| Air Blank | | 0.000 | 14:35 |
| Subject Sample | # 2 | 0.212 | 14:36 |
| Breath Volume | | 2.480 | 14:36 |
| Air Blank | | 0.000 | 14:37 |
| Control Test | | 0.076 | 14:37 |
| Air Blank | | 0.000 | 14:37 |
| Diagnostics Check | | OK | 14:37 |

Data Download Date/Time: 02/03/2020 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALDREE,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/08/2020 10:17:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 10:10 |
| Air Blank | 0.000 | 10:11 |
| Control Test | 0.079 | 10:11 |
| Air Blank | 0.000 | 10:12 |
| Subject Sample # 1 | 0.108 | 10:12 |
| Breath Volume | 2.250 | 10:12 |
| Air Blank | 0.000 | 10:13 |
| Air Blank | 0.000 | 10:15 |
| Subject Sample # 2 | 0.106 | 10:15 |
| Breath Volume | 2.050 | 10:15 |
| Air Blank | 0.000 | 10:16 |
| Control Test | 0.079 | 10:16 |
| Air Blank | 0.000 | 10:17 |
| Diagnostics Check | OK | 10:17 |

Data Download Date/Time: 02/27/2020 18:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADFORD,CHARLES,T
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: WADFORD,CHARLES
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GREATHOUSE

DATE/TIME OF TEST: 02/22/2020 01:55:15

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:48 |
| Air Blank | 0.000 | 01:49 |
| Control Test | 0.079 | 01:49 |
| Air Blank | 0.000 | 01:50 |
| Subject Sample # 1 | 0.190 | 01:50 |
| Breath Volume | 1.808 | 01:50 |
| Air Blank | 0.000 | 01:51 |
| Air Blank | 0.000 | 01:53 |
| Subject Sample # 2 | 0.188 | 01:53 |
| Breath Volume | 1.265 | 01:53 |
| Air Blank | 0.000 | 01:54 |
| Control Test | 0.079 | 01:54 |
| Air Blank | 0.000 | 01:55 |
| Diagnostics Check | OK | 01:55 |

Data Download Date/Time: 02/27/2020 18:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: TOWNSEND, CLYDE
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 01/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: AVINS

DATE/TIME OF TEST: 02/23/2020 03:49:41

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 03:46 | |
| Air Blank | 0.000 | 03:47 | |
| Control Test | 0.080 | 03:47 | |
| Air Blank | 0.000 | 03:47 | |
| Subject Sample | # 1 | 03:48 | Subject Test Refused |
| Breath Volume | 0.000 | 03:48 | |
| Air Blank | 0.000 | 03:48 | |
| Control Test | 0.080 | 03:49 | |
| Air Blank | 0.000 | 03:49 | |
| Diagnostics Check | OK | 03:49 | |

Data Download Date/Time: 02/27/2020 18:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: HIMES-RUETH

DATE/TIME OF TEST: 02/18/2020 01:41:37

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:34 |
| Air Blank | | 0.000 | 01:35 |
| Control Test | | 0.078 | 01:35 |
| Air Blank | | 0.000 | 01:36 |
| Subject Sample | # 1 | 0.244 | 01:36 |
| Breath Volume | | 1.945 | 01:36 |
| Air Blank | | 0.000 | 01:37 |
| Air Blank | | 0.000 | 01:39 |
| Subject Sample | # 2 | 0.252 | 01:40 |
| Breath Volume | | 1.710 | 01:40 |
| Air Blank | | 0.000 | 01:40 |
| Control Test | | 0.077 | 01:41 |
| Air Blank | | 0.000 | 01:41 |
| Diagnostics Check | | OK | 01:41 |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 02/20/2020 01:25:57

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:20 | |
| Air Blank | 0.000 | 01:20 | |
| Control Test | 0.077 | 01:21 | |
| Air Blank | 0.000 | 01:21 | |
| Subject Sample | # 1 | 01:24 | Subject Test Refused |
| Breath Volume | 0.000 | 01:24 | |
| Air Blank | 0.000 | 01:25 | |
| Control Test | 0.077 | 01:25 | |
| Air Blank | 0.000 | 01:25 | |
| Diagnostics Check | OK | 01:25 | |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, COLLEEN, E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CASTIGLIONE, SALVATORE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HILDEBRAND

DATE/TIME OF TEST: 02/22/2020 00:47:51

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:40 |
| Air Blank | | 0.000 | 00:41 |
| Control Test | | 0.078 | 00:41 |
| Air Blank | | 0.000 | 00:41 |
| Subject Sample | # 1 | 0.132 | 00:43 |
| Breath Volume | | 1.464 | 00:43 |
| Air Blank | | 0.000 | 00:43 |
| Air Blank | | 0.000 | 00:45 |
| Subject Sample | # 2 | 0.139 | 00:46 |
| Breath Volume | | 1.949 | 00:46 |
| Air Blank | | 0.000 | 00:46 |
| Control Test | | 0.079 | 00:47 |
| Air Blank | | 0.000 | 00:47 |
| Diagnostics Check | | OK | 00:47 |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HELM,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 02/22/2020 04:16:47

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 04:13 | |
| Air Blank | 0.000 | 04:14 | |
| Control Test | 0.079 | 04:14 | |
| Air Blank | 0.000 | 04:14 | |
| Subject Sample | # 1 | 04:15 | Subject Test Refused |
| Breath Volume | 0.000 | 04:15 | |
| Air Blank | 0.000 | 04:15 | |
| Control Test | 0.079 | 04:16 | |
| Air Blank | 0.000 | 04:16 | |
| Diagnostics Check | OK | 04:16 | |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: GARIEPY

DATE/TIME OF TEST: 02/22/2020 22:39:20

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:33 |
| Air Blank | 0.000 | 22:33 |
| Control Test | 0.079 | 22:33 |
| Air Blank | 0.000 | 22:34 |
| Subject Sample # 1 | 0.076 | 22:34 |
| Breath Volume | 2.734 | 22:34 |
| Air Blank | 0.000 | 22:35 |
| Air Blank | 0.000 | 22:37 |
| Subject Sample # 2 | 0.070 | 22:37 |
| Breath Volume | 2.191 | 22:37 |
| Air Blank | 0.000 | 22:38 |
| Control Test | 0.079 | 22:38 |
| Air Blank | 0.000 | 22:39 |
| Diagnostics Check | OK | 22:39 |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SCHMIDT,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/22/2020 23:55:11

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:50 | |
| Air Blank | 0.000 | 23:51 | |
| Control Test | 0.077 | 23:51 | |
| Air Blank | 0.000 | 23:51 | |
| Subject Sample | # 1 | 23:53 | Subject Test Refused |
| Breath Volume | 0.000 | 23:53 | |
| Air Blank | 0.000 | 23:54 | |
| Control Test | 0.078 | 23:54 | |
| Air Blank | 0.000 | 23:55 | |
| Diagnostics Check | OK | 23:55 | |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,DANIEL,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,VINCENT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 02/27/2020 00:45:34

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:34 |
| Air Blank | 0.000 | 00:35 |
| Control Test | 0.077 | 00:35 |
| Air Blank | 0.000 | 00:36 |
| Subject Sample # 1 | 0.157 | 00:38 |
| Breath Volume | 1.164 | 00:38 |
| Air Blank | 0.000 | 00:39 |
| Air Blank | 0.000 | 00:40 |
| Subject Sample # 2 | 0.156 | 00:43 |
| Breath Volume | 1.757 | 00:43 |
| Air Blank | 0.000 | 00:44 |
| Control Test | 0.077 | 00:44 |
| Air Blank | 0.000 | 00:45 |
| Diagnostics Check | OK | 00:45 |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 02/28/2020 01:42:56

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:36 |
| Air Blank | 0.000 | 01:36 |
| Control Test | 0.079 | 01:36 |
| Air Blank | 0.000 | 01:37 |
| Subject Sample # 1 | 0.029 | 01:38 |
| Breath Volume | 1.242 | 01:38 |
| Air Blank | 0.000 | 01:38 |
| Air Blank | 0.000 | 01:40 |
| Subject Sample # 2 | 0.029 | 01:41 |
| Breath Volume | 1.437 | 01:41 |
| Air Blank | 0.000 | 01:42 |
| Control Test | 0.079 | 01:42 |
| Air Blank | 0.000 | 01:42 |
| Diagnostics Check | OK | 01:42 |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SCROGGINS

DATE/TIME OF TEST: 02/29/2020 01:57:51

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:53 | |
| Air Blank | 0.000 | 01:53 | |
| Control Test | 0.078 | 01:54 | |
| Air Blank | 0.000 | 01:54 | |
| Subject Sample | # 1 | 01:56 | Subject Test Refused |
| Breath Volume | 0.000 | 01:56 | |
| Air Blank | 0.000 | 01:56 | |
| Control Test | 0.079 | 01:57 | |
| Air Blank | 0.000 | 01:57 | |
| Diagnostics Check | OK | 01:57 | |

Data Download Date/Time: 03/31/2020 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WELLMAN,A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WRAY

DATE/TIME OF TEST: 02/18/2020 02:51:11

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 02:48 | |
| Air Blank | 0.000 | 02:48 | |
| Control Test | 0.078 | 02:49 | |
| Air Blank | 0.000 | 02:49 | |
| Subject Sample | # 1 | 02:49 | Subject Test Refused |
| Breath Volume | 0.000 | 02:49 | |
| Air Blank | 0.000 | 02:50 | |
| Control Test | 0.077 | 02:50 | |
| Air Blank | 0.000 | 02:51 | |
| Diagnostics Check | OK | 02:51 | |

Data Download Date/Time: 03/14/2020 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GOMAZ MENDEZ

DATE/TIME OF TEST: 02/19/2020 19:21:16

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:18 | |
| Air Blank | 0.000 | 19:18 | |
| Control Test | 0.077 | 19:19 | |
| Air Blank | 0.000 | 19:19 | |
| Subject Sample | # 1 | 19:19 | Subject Test Refused |
| Breath Volume | 0.000 | 19:19 | |
| Air Blank | 0.000 | 19:20 | |
| Control Test | 0.077 | 19:20 | |
| Air Blank | 0.000 | 19:21 | |
| Diagnostics Check | OK | 19:21 | |

Data Download Date/Time: 03/14/2020 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MEYERS,ERIC
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: ZEEMAN

DATE/TIME OF TEST: 02/20/2020 23:09:29

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 23:02 |
| Air Blank | | 0.000 | 23:02 |
| Control Test | | 0.077 | 23:02 |
| Air Blank | | 0.000 | 23:03 |
| Subject Sample | # 1 | 0.281 | 23:04 |
| Breath Volume | | 1.292 | 23:04 |
| Air Blank | | 0.000 | 23:05 |
| Air Blank | | 0.000 | 23:06 |
| Subject Sample | # 2 | 0.282 | 23:07 |
| Breath Volume | | 3.480 | 23:07 |
| Air Blank | | 0.000 | 23:08 |
| Control Test | | 0.075 | 23:08 |
| Air Blank | | 0.000 | 23:09 |
| Diagnostics Check | | OK | 23:09 |

Data Download Date/Time: 03/14/2020 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,ALAN,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LOMBARDI,ANTHONY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MATTIE

DATE/TIME OF TEST: 02/21/2020 04:27:33

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:20 |
| Air Blank | 0.000 | 04:21 |
| Control Test | 0.077 | 04:21 |
| Air Blank | 0.000 | 04:21 |
| Subject Sample # 1 | 0.159 | 04:22 |
| Breath Volume | 3.433 | 04:22 |
| Air Blank | 0.000 | 04:23 |
| Air Blank | 0.000 | 04:25 |
| Subject Sample # 2 | 0.163 | 04:25 |
| Breath Volume | 3.335 | 04:25 |
| Air Blank | 0.000 | 04:26 |
| Control Test | 0.077 | 04:26 |
| Air Blank | 0.000 | 04:27 |
| Diagnostics Check | OK | 04:27 |

Data Download Date/Time: 03/14/2020 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANE,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 02/22/2020 18:49:08

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:37 |
| Air Blank | 0.000 | 18:38 |
| Control Test | 0.078 | 18:38 |
| Air Blank | 0.000 | 18:39 |
| Subject Sample # 1 | 0.291 | 18:40 |
| Breath Volume | 1.964 | 18:40 |
| Air Blank | 0.000 | 18:41 |
| Air Blank | 0.000 | 18:42 |
| Subject Sample # 2 | 0.269 | 18:43 |
| Breath Volume | 1.160 | 18:43 |
| Air Blank | 0.000 | 18:44 |
| Air Blank | 0.000 | 18:46 |
| Subject Sample # 3 | 0.283 | 18:47 |
| Breath Volume | 1.789 | 18:47 |
| Air Blank | 0.000 | 18:48 |
| Control Test | 0.078 | 18:48 |
| Air Blank | 0.000 | 18:49 |
| Diagnostics Check | OK | 18:49 |

Data Download Date/Time: 03/14/2020 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: OTERO,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BADAL

DATE/TIME OF TEST: 02/23/2020 00:43:25

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 00:37 |
| Air Blank | 0.000 | 00:37 |
| Control Test | 0.079 | 00:37 |
| Air Blank | 0.000 | 00:38 |
| Subject Sample # 1 | 0.115 | 00:38 |
| Breath Volume | 1.941 | 00:38 |
| Air Blank | 0.000 | 00:39 |
| Air Blank | 0.000 | 00:41 |
| Subject Sample # 2 | 0.122 | 00:41 |
| Breath Volume | 2.164 | 00:41 |
| Air Blank | 0.000 | 00:42 |
| Control Test | 0.078 | 00:42 |
| Air Blank | 0.000 | 00:43 |
| Diagnostics Check | OK | 00:43 |

Data Download Date/Time: 03/14/2020 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: COLLINS,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MCMILLEN

DATE/TIME OF TEST: 02/24/2020 01:58:40

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:52 |
| Air Blank | 0.000 | 01:52 |
| Control Test | 0.078 | 01:52 |
| Air Blank | 0.000 | 01:53 |
| Subject Sample # 1 | 0.107 | 01:54 |
| Breath Volume | 2.960 | 01:54 |
| Air Blank | 0.000 | 01:54 |
| Air Blank | 0.000 | 01:56 |
| Subject Sample # 2 | 0.105 | 01:57 |
| Breath Volume | 1.980 | 01:57 |
| Air Blank | 0.000 | 01:57 |
| Control Test | 0.078 | 01:58 |
| Air Blank | 0.000 | 01:58 |
| Diagnostics Check | OK | 01:58 |

Data Download Date/Time: 03/14/2020 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: KUJAVA

DATE/TIME OF TEST: 02/01/2020 23:06:14

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:03 | |
| Air Blank | 0.000 | 23:03 | |
| Control Test | 0.079 | 23:04 | |
| Air Blank | 0.000 | 23:04 | |
| Subject Sample | # 1 | 23:04 | Subject Test Refused |
| Breath Volume | 0.000 | 23:04 | |
| Air Blank | 0.000 | 23:05 | |
| Control Test | 0.079 | 23:05 | |
| Air Blank | 0.000 | 23:06 | |
| Diagnostics Check | OK | 23:06 | |

Data Download Date/Time: 02/15/2020 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: KLEPACKI

DATE/TIME OF TEST: 02/02/2020 01:18:15

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:11 |
| Air Blank | | 0.000 | 01:11 |
| Control Test | | 0.079 | 01:12 |
| Air Blank | | 0.000 | 01:12 |
| Subject Sample | # 1 | 0.211 | 01:13 |
| Breath Volume | | 2.296 | 01:13 |
| Air Blank | | 0.000 | 01:14 |
| Air Blank | | 0.000 | 01:16 |
| Subject Sample | # 2 | 0.194 | 01:16 |
| Breath Volume | | 1.933 | 01:16 |
| Air Blank | | 0.000 | 01:17 |
| Control Test | | 0.080 | 01:17 |
| Air Blank | | 0.000 | 01:18 |
| Diagnostics Check | | OK | 01:18 |

Data Download Date/Time: 02/15/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: KELLY,WESLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:13

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/06/2020 13:39:28

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 13:33 |
| Air Blank | 0.000 | 13:33 |
| Control Test | 0.078 | 13:33 |
| Air Blank | 0.000 | 13:34 |
| Subject Sample # 1 | 0.081 | 13:34 |
| Breath Volume | 2.148 | 13:34 |
| Air Blank | 0.000 | 13:35 |
| Air Blank | 0.000 | 13:37 |
| Subject Sample # 2 | 0.076 | 13:37 |
| Breath Volume | 1.320 | 13:37 |
| Air Blank | 0.000 | 13:38 |
| Control Test | 0.078 | 13:38 |
| Air Blank | 0.000 | 13:39 |
| Diagnostics Check | OK | 13:39 |

Data Download Date/Time: 02/15/2020 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,BRENDA,R
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FLYNN,PATRICK
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DYAL

DATE/TIME OF TEST: 02/08/2020 21:12:43

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:05 |
| Air Blank | 0.000 | 21:06 |
| Control Test | 0.081 | 21:06 |
| Air Blank | 0.000 | 21:07 |
| Subject Sample # 1 | 0.071 | 21:08 |
| Breath Volume | 2.496 | 21:08 |
| Air Blank | 0.000 | 21:09 |
| Air Blank | 0.000 | 21:10 |
| Subject Sample # 2 | 0.072 | 21:11 |
| Breath Volume | 1.785 | 21:11 |
| Air Blank | 0.000 | 21:11 |
| Control Test | 0.080 | 21:12 |
| Air Blank | 0.000 | 21:12 |
| Diagnostics Check | OK | 21:12 |

Data Download Date/Time: 02/15/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MATA,R
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/09/2020 22:20:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:12 |
| Air Blank | 0.000 | 22:12 |
| Control Test | 0.080 | 22:12 |
| Air Blank | 0.000 | 22:13 |
| Subject Sample # 1 | 0.211 | 22:15 |
| Breath Volume | 1.980 | 22:15 |
| Air Blank | 0.000 | 22:16 |
| Air Blank | 0.000 | 22:18 |
| Subject Sample # 2 | 0.213 | 22:18 |
| Breath Volume | 1.480 | 22:18 |
| Air Blank | 0.000 | 22:19 |
| Control Test | 0.079 | 22:19 |
| Air Blank | 0.000 | 22:20 |
| Diagnostics Check | OK | 22:20 |

Data Download Date/Time: 02/15/2020 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: POWERS,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/12/2020 19:33:50

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 19:30 | |
| Air Blank | 0.000 | 19:31 | |
| Control Test | 0.079 | 19:31 | |
| Air Blank | 0.000 | 19:32 | |
| Subject Sample | # 1 | 19:32 | Subject Test Refused |
| Breath Volume | 0.000 | 19:32 | |
| Air Blank | 0.000 | 19:32 | |
| Control Test | 0.080 | 19:33 | |
| Air Blank | 0.000 | 19:33 | |
| Diagnostics Check | OK | 19:33 | |

Data Download Date/Time: 02/15/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GILBERT,JASON
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LUDWIG

DATE/TIME OF TEST: 02/15/2020 00:58:34

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:55 | |
| Air Blank | 0.000 | 00:55 | |
| Control Test | 0.080 | 00:56 | |
| Air Blank | 0.000 | 00:56 | |
| Subject Sample | # 1 | 00:57 | Subject Test Refused |
| Breath Volume | 0.000 | 00:57 | |
| Air Blank | 0.000 | 00:57 | |
| Control Test | 0.080 | 00:57 | |
| Air Blank | 0.000 | 00:58 | |
| Diagnostics Check | OK | 00:58 | |

Data Download Date/Time: 02/15/2020 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/15/2020 03:36:53

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:30 |
| Air Blank | 0.000 | 03:31 |
| Control Test | 0.080 | 03:31 |
| Air Blank | 0.000 | 03:32 |
| Subject Sample # 1 | 0.000 | 03:32 |
| Breath Volume | 2.484 | 03:32 |
| Air Blank | 0.000 | 03:33 |
| Air Blank | 0.000 | 03:35 |
| Subject Sample # 2 | 0.000 | 03:35 |
| Breath Volume | 2.093 | 03:35 |
| Air Blank | 0.000 | 03:35 |
| Control Test | 0.079 | 03:36 |
| Air Blank | 0.000 | 03:36 |
| Diagnostics Check | OK | 03:36 |

Data Download Date/Time: 02/15/2020 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIET,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: GALARZA

DATE/TIME OF TEST: 02/15/2020 06:35:40

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 06:32 | |
| Air Blank | 0.000 | 06:33 | |
| Control Test | 0.080 | 06:33 | |
| Air Blank | 0.000 | 06:34 | |
| Subject Sample | # 1 | 06:34 | Subject Test Refused |
| Breath Volume | 0.000 | 06:34 | |
| Air Blank | 0.000 | 06:34 | |
| Control Test | 0.080 | 06:35 | |
| Air Blank | 0.000 | 06:35 | |
| Diagnostics Check | OK | 06:35 | |

Data Download Date/Time: 02/15/2020 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 02/14/2020 00:43:50

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:40 | |
| Air Blank | 0.000 | 00:41 | |
| Control Test | 0.079 | 00:41 | |
| Air Blank | 0.000 | 00:42 | |
| Subject Sample | # 1 | 00:42 | Subject Test Refused |
| Breath Volume | 0.000 | 00:42 | |
| Air Blank | 0.000 | 00:42 | |
| Control Test | 0.079 | 00:43 | |
| Air Blank | 0.000 | 00:43 | |
| Diagnostics Check | OK | 00:43 | |

Data Download Date/Time: 02/21/2020 02:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: THEISEN

DATE/TIME OF TEST: 02/15/2020 00:17:53

| | | | | |
|-------------------|-----|-------|-------|---------------------------|
| Diagnostics Check | | OK | 00:05 | |
| Air Blank | | 0.000 | 00:06 | |
| Control Test | | 0.080 | 00:06 | |
| Air Blank | | 0.000 | 00:07 | |
| Subject Sample | # 1 | 0.211 | 00:10 | Volume Not Met |
| Breath Volume | | 1.046 | 00:10 | |
| Air Blank | | 0.000 | 00:10 | |
| Air Blank | | 0.000 | 00:12 | |
| Subject Sample | # 2 | 0.225 | 00:15 | Slope Not Level |
| Breath Volume | | 1.289 | 00:15 | |
| Air Blank | | 0.000 | 00:16 | |
| Control Test | | 0.080 | 00:16 | |
| Air Blank | | 0.000 | 00:17 | |
| Diagnostics Check | | Fail | 00:17 | Analytical Stability Test |
| Air Blank | | 0.000 | 00:17 | |

Data Download Date/Time: 02/21/2020 02:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,SHANNON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WINDER

DATE/TIME OF TEST: 02/04/2020 22:46:27

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:39 |
| Air Blank | 0.000 | 22:40 |
| Control Test | 0.078 | 22:40 |
| Air Blank | 0.000 | 22:41 |
| Subject Sample # 1 | 0.119 | 22:41 |
| Breath Volume | 1.683 | 22:41 |
| Air Blank | 0.000 | 22:42 |
| Air Blank | 0.000 | 22:44 |
| Subject Sample # 2 | 0.103 | 22:45 |
| Breath Volume | 1.562 | 22:45 |
| Air Blank | 0.000 | 22:45 |
| Control Test | 0.079 | 22:45 |
| Air Blank | 0.000 | 22:46 |
| Diagnostics Check | OK | 22:46 |

Data Download Date/Time: 02/25/2020 00:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WALSH,SHANNON,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: WALSH,SHANNON

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:27

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/16/2020 20:12:36

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:05 |
| Air Blank | 0.000 | 20:05 |
| Control Test | 0.080 | 20:06 |
| Air Blank | 0.000 | 20:06 |
| Subject Sample # 1 | 0.287 | 20:07 |
| Breath Volume | 1.968 | 20:07 |
| Air Blank | 0.000 | 20:08 |
| Air Blank | 0.000 | 20:09 |
| Subject Sample # 2 | 0.267 | 20:11 |
| Breath Volume | 2.492 | 20:11 |
| Air Blank | 0.000 | 20:11 |
| Control Test | 0.080 | 20:12 |
| Air Blank | 0.000 | 20:12 |
| Diagnostics Check | OK | 20:12 |

Data Download Date/Time: 02/25/2020 00:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, JONATHAN, B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: RICHARD, LATASHA
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 01/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/23/2020 02:55:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:47 |
| Air Blank | 0.000 | 02:47 |
| Control Test | 0.081 | 02:47 |
| Air Blank | 0.000 | 02:48 |
| Subject Sample # 1 | 0.178 | 02:50 |
| Breath Volume | 1.253 | 02:50 |
| Air Blank | 0.000 | 02:51 |
| Air Blank | 0.000 | 02:53 |
| Subject Sample # 2 | 0.163 | 02:53 |
| Breath Volume | 1.804 | 02:53 |
| Air Blank | 0.000 | 02:54 |
| Control Test | 0.079 | 02:54 |
| Air Blank | 0.000 | 02:55 |
| Diagnostics Check | OK | 02:55 |

Data Download Date/Time: 02/25/2020 00:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DAGOSTINO,STEFANO
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MORELLI

DATE/TIME OF TEST: 02/26/2020 00:07:56

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 00:05 | |
| Air Blank | 0.000 | 00:05 | |
| Control Test | 0.081 | 00:05 | |
| Air Blank | 0.000 | 00:06 | |
| Subject Sample | # 1 | 00:06 | Subject Test Refused |
| Breath Volume | 0.000 | 00:06 | |
| Air Blank | 0.000 | 00:07 | |
| Control Test | 0.081 | 00:07 | |
| Air Blank | 0.000 | 00:07 | |
| Diagnostics Check | OK | 00:07 | |

Data Download Date/Time: 03/19/2020 00:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/02/2020 01:41:16

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 01:32 | |
| Air Blank | 0.000 | 01:33 | |
| Control Test | 0.079 | 01:33 | |
| Air Blank | 0.000 | 01:34 | |
| Subject Sample | # 1 | 01:37 | No Sample Provided |
| Breath Volume | 0.000 | 01:37 | |
| Air Blank | 0.000 | 01:37 | |
| Air Blank | 0.000 | 01:39 | |
| Subject Sample | # 2 | 01:39 | Subject Test Refused |
| Breath Volume | 0.000 | 01:39 | |
| Air Blank | 0.000 | 01:40 | |
| Control Test | 0.080 | 01:40 | |
| Air Blank | 0.000 | 01:41 | |
| Diagnostics Check | OK | 01:41 | |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 02/02/2020 02:26:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:20 |
| Air Blank | 0.000 | 02:20 |
| Control Test | 0.080 | 02:21 |
| Air Blank | 0.000 | 02:21 |
| Subject Sample # 1 | 0.062 | 02:22 |
| Breath Volume | 2.238 | 02:22 |
| Air Blank | 0.000 | 02:23 |
| Air Blank | 0.000 | 02:24 |
| Subject Sample # 2 | 0.062 | 02:25 |
| Breath Volume | 1.453 | 02:25 |
| Air Blank | 0.000 | 02:25 |
| Control Test | 0.080 | 02:26 |
| Air Blank | 0.000 | 02:26 |
| Diagnostics Check | OK | 02:26 |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PRYSHCHEP

DATE/TIME OF TEST: 02/02/2020 04:28:06

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 04:16 | |
| Air Blank | | 0.000 | 04:16 | |
| Control Test | | 0.080 | 04:17 | |
| Air Blank | | 0.000 | 04:17 | |
| Subject Sample | # 1 | 0.097 | 04:20 | Volume Not Met |
| Breath Volume | | 0.351 | 04:20 | |
| Air Blank | | 0.000 | 04:21 | |
| Air Blank | | 0.000 | 04:23 | |
| Subject Sample | # 2 | 0.116 | 04:23 | |
| Breath Volume | | 1.441 | 04:23 | |
| Air Blank | | 0.000 | 04:24 | |
| Air Blank | | 0.000 | 04:26 | |
| Subject Sample | # 3 | 0.117 | 04:26 | |
| Breath Volume | | 1.785 | 04:26 | |
| Air Blank | | 0.000 | 04:27 | |
| Control Test | | 0.079 | 04:27 | |
| Air Blank | | 0.000 | 04:28 | |
| Diagnostics Check | | OK | 04:28 | |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/02/2020 04:47:57

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 04:35 | |
| Air Blank | 0.000 | 04:35 | |
| Control Test | 0.079 | 04:35 | |
| Air Blank | 0.000 | 04:36 | |
| Subject Sample | # 1 | 04:39 | No Sample Provided |
| Breath Volume | 0.000 | 04:39 | |
| Air Blank | 0.000 | 04:40 | |
| Air Blank | 0.000 | 04:42 | |
| Subject Sample | # 2 | 0.154 | 04:42 |
| Breath Volume | 3.582 | 04:42 | |
| Air Blank | 0.000 | 04:43 | |
| Air Blank | 0.000 | 04:45 | |
| Subject Sample | # 3 | 0.159 | 04:46 |
| Breath Volume | 4.308 | 04:46 | |
| Air Blank | 0.000 | 04:47 | |
| Control Test | 0.079 | 04:47 | |
| Air Blank | 0.000 | 04:47 | |
| Diagnostics Check | OK | 04:47 | |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 02/02/2020 21:03:01

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:56 |
| Air Blank | 0.000 | 20:56 |
| Control Test | 0.080 | 20:57 |
| Air Blank | 0.000 | 20:57 |
| Subject Sample # 1 | 0.184 | 20:58 |
| Breath Volume | 1.867 | 20:58 |
| Air Blank | 0.000 | 20:59 |
| Air Blank | 0.000 | 21:00 |
| Subject Sample # 2 | 0.196 | 21:01 |
| Breath Volume | 1.832 | 21:01 |
| Air Blank | 0.000 | 21:02 |
| Control Test | 0.080 | 21:02 |
| Air Blank | 0.000 | 21:02 |
| Diagnostics Check | OK | 21:03 |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 02/03/2020 22:26:33

| | | | |
|--------------------|-------|-------|------------|
| Diagnostics Check | OK | 22:22 | |
| Air Blank | 0.000 | 22:23 | |
| Control Test | 0.081 | 22:23 | |
| Air Blank | 0.000 | 22:24 | |
| Subject Sample # 1 | 0.197 | 22:25 | |
| Breath Volume | 1.621 | 22:25 | |
| Air Blank | | 22:25 | RFI Detect |
| Air Blank | 0.000 | 22:25 | |
| Control Test | 0.079 | 22:25 | |
| Air Blank | 0.000 | 22:26 | |
| Diagnostics Check | OK | 22:26 | |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 02/03/2020 22:34:45

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:27 |
| Air Blank | 0.000 | 22:28 |
| Control Test | 0.080 | 22:28 |
| Air Blank | 0.000 | 22:28 |
| Subject Sample # 1 | 0.206 | 22:30 |
| Breath Volume | 1.761 | 22:30 |
| Air Blank | 0.000 | 22:30 |
| Air Blank | 0.000 | 22:32 |
| Subject Sample # 2 | 0.207 | 22:33 |
| Breath Volume | 2.105 | 22:33 |
| Air Blank | 0.000 | 22:33 |
| Control Test | 0.080 | 22:34 |
| Air Blank | 0.000 | 22:34 |
| Diagnostics Check | OK | 22:34 |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: DUBOIS

DATE/TIME OF TEST: 02/04/2020 03:55:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 03:49 |
| Air Blank | 0.000 | 03:50 |
| Control Test | 0.081 | 03:50 |
| Air Blank | 0.000 | 03:50 |
| Subject Sample # 1 | 0.023 | 03:51 |
| Breath Volume | 1.347 | 03:51 |
| Air Blank | 0.000 | 03:51 |
| Air Blank | 0.000 | 03:53 |
| Subject Sample # 2 | 0.022 | 03:54 |
| Breath Volume | 2.007 | 03:54 |
| Air Blank | 0.000 | 03:54 |
| Control Test | 0.080 | 03:55 |
| Air Blank | 0.000 | 03:55 |
| Diagnostics Check | OK | 03:55 |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: WRENN

DATE/TIME OF TEST: 02/04/2020 21:31:15

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 21:22 | |
| Air Blank | 0.000 | 21:23 | |
| Control Test | 0.079 | 21:23 | |
| Air Blank | 0.000 | 21:24 | |
| Subject Sample | # 1 | 21:27 | No Sample Provided |
| Breath Volume | 0.000 | 21:27 | |
| Air Blank | 0.000 | 21:27 | |
| Air Blank | 0.000 | 21:29 | |
| Subject Sample | # 2 | 21:29 | Subject Test Refused |
| Breath Volume | 0.000 | 21:29 | |
| Air Blank | 0.000 | 21:30 | |
| Control Test | 0.080 | 21:30 | |
| Air Blank | 0.000 | 21:31 | |
| Diagnostics Check | OK | 21:31 | |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/05/2020 00:49:07

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:42 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.079 | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Subject Sample | # 1 | 0.200 | 00:44 |
| Breath Volume | | 1.382 | 00:44 |
| Air Blank | | 0.000 | 00:45 |
| Air Blank | | 0.000 | 00:47 |
| Subject Sample | # 2 | 0.195 | 00:47 |
| Breath Volume | | 1.417 | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Control Test | | 0.078 | 00:48 |
| Air Blank | | 0.000 | 00:49 |
| Diagnostics Check | | OK | 00:49 |

Data Download Date/Time: 02/05/2020 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GASCA

DATE/TIME OF TEST: 02/19/2020 01:30:58

| | | | |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK | 01:22 | |
| Air Blank | 0.000 | 01:22 | |
| Control Test | 0.077 | 01:23 | |
| Air Blank | 0.000 | 01:23 | |
| Subject Sample | # 1 0.135 | 01:26 | Volume Not Met |
| Breath Volume | 0.507 | 01:26 | |
| Air Blank | 0.000 | 01:27 | |
| Air Blank | 0.000 | 01:29 | |
| Subject Sample | # 2 | 01:29 | Subject Test Refused |
| Breath Volume | 0.000 | 01:29 | |
| Air Blank | 0.000 | 01:30 | |
| Control Test | 0.075 | 01:30 | |
| Air Blank | 0.000 | 01:30 | |
| Diagnostics Check | OK | 01:30 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGLIA,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BRIGLIA,NICHOLAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/19/2020 17:33:23

| | | | |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK | 17:25 | |
| Air Blank | 0.000 | 17:25 | |
| Control Test | 0.077 | 17:26 | |
| Air Blank | 0.000 | 17:26 | |
| Subject Sample | # 1 0.187 | 17:28 | |
| Breath Volume | 2.019 | 17:28 | |
| Air Blank | 0.000 | 17:29 | |
| Air Blank | 0.000 | 17:31 | |
| Subject Sample | # 2 0.181 | 17:31 | |
| Breath Volume | 1.742 | 17:31 | |
| Air Blank | 0.000 | 17:32 | |
| Control Test | 0.073 | 17:32 | Control Outside Tolerance |
| Air Blank | 0.000 | 17:33 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGLIA,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BRIGLIA,NICHOLAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/19/2020 17:43:32

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:35 |
| Air Blank | 0.000 | 17:35 |
| Control Test | 0.076 | 17:36 |
| Air Blank | 0.000 | 17:36 |
| Subject Sample # 1 | 0.200 | 17:38 |
| Breath Volume | 2.449 | 17:38 |
| Air Blank | 0.000 | 17:39 |
| Air Blank | 0.000 | 17:41 |
| Subject Sample # 2 | 0.200 | 17:41 |
| Breath Volume | 2.402 | 17:41 |
| Air Blank | 0.000 | 17:42 |
| Control Test | 0.075 | 17:42 |
| Air Blank | 0.000 | 17:43 |
| Diagnostics Check | OK | 17:43 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BUSHBY

DATE/TIME OF TEST: 02/21/2020 23:53:30

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:46 |
| Air Blank | 0.000 | 23:47 |
| Control Test | 0.077 | 23:47 |
| Air Blank | 0.000 | 23:47 |
| Subject Sample # 1 | 0.146 | 23:48 |
| Breath Volume | 2.621 | 23:48 |
| Air Blank | 0.000 | 23:49 |
| Air Blank | 0.000 | 23:51 |
| Subject Sample # 2 | 0.140 | 23:51 |
| Breath Volume | 2.171 | 23:51 |
| Air Blank | 0.000 | 23:52 |
| Control Test | 0.076 | 23:52 |
| Air Blank | 0.000 | 23:53 |
| Diagnostics Check | OK | 23:53 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISOPHR,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ROWLING

DATE/TIME OF TEST: 02/22/2020 22:02:10

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 21:52 |
| Air Blank | | 0.000 | 21:52 |
| Control Test | | 0.077 | 21:52 |
| Air Blank | | 0.000 | 21:53 |
| Subject Sample | # 1 | 0.204 | 21:54 |
| Breath Volume | | 1.425 | 21:54 |
| Air Blank | | 0.000 | 21:54 |
| Air Blank | | 0.000 | 21:56 |
| Subject Sample | # 2 | 0.183 | 21:57 |
| Breath Volume | | 1.121 | 21:57 |
| Air Blank | | 0.000 | 21:58 |
| Air Blank | | 0.000 | 21:59 |
| Subject Sample | # 3 | 0.184 | 22:00 |
| Breath Volume | | 1.371 | 22:00 |
| Air Blank | | 0.000 | 22:01 |
| Control Test | | 0.077 | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Diagnostics Check | | OK | 22:02 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,KEVIN,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MORRIS,KEVIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: LOPEZ-LOPEZ

DATE/TIME OF TEST: 02/23/2020 06:53:41

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 06:47 |
| Air Blank | 0.000 | 06:47 |
| Control Test | 0.077 | 06:47 |
| Air Blank | 0.000 | 06:48 |
| Subject Sample # 1 | 0.137 | 06:48 |
| Breath Volume | 3.648 | 06:48 |
| Air Blank | 0.000 | 06:49 |
| Air Blank | 0.000 | 06:51 |
| Subject Sample # 2 | 0.130 | 06:51 |
| Breath Volume | 3.300 | 06:51 |
| Air Blank | 0.000 | 06:52 |
| Control Test | 0.075 | 06:53 |
| Air Blank | 0.000 | 06:53 |
| Diagnostics Check | OK | 06:53 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSINGA,KYLE,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: OSINGA,KYLE
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 02/23/2020 17:39:55

| | | | |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK | 17:32 | |
| Air Blank | 0.000 | 17:33 | |
| Control Test | 0.076 | 17:33 | |
| Air Blank | 0.000 | 17:34 | |
| Subject Sample | # 1 0.202 | 17:34 | |
| Breath Volume | 2.472 | 17:34 | |
| Air Blank | 0.000 | 17:35 | |
| Air Blank | 0.000 | 17:37 | |
| Subject Sample | # 2 0.195 | 17:38 | |
| Breath Volume | 2.695 | 17:38 | |
| Air Blank | 0.000 | 17:38 | |
| Control Test | 0.074 | 17:39 | Control Outside Tolerance |
| Air Blank | | 17:39 | Ambient Fail |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBOIS,BROOKS,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUBOIS,BROOKS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 02/24/2020 19:31:09

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 19:19 | |
| Air Blank | 0.000 | 19:20 | |
| Control Test | 0.077 | 19:20 | |
| Air Blank | 0.000 | 19:21 | |
| Subject Sample | # 1 | 19:24 | No Sample Provided |
| Breath Volume | 0.000 | 19:24 | |
| Air Blank | 0.000 | 19:24 | |
| Air Blank | 0.000 | 19:26 | |
| Subject Sample | # 2 | 19:29 | No Sample Provided |
| Breath Volume | 0.000 | 19:29 | |
| Air Blank | 0.000 | 19:30 | |
| Control Test | 0.078 | 19:30 | Interferent Detect |
| Air Blank | 0.000 | 19:31 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GROSSMAN

DATE/TIME OF TEST: 02/25/2020 03:56:38

| | | | |
|--------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 03:45 | |
| Air Blank | 0.000 | 03:45 | |
| Control Test | 0.078 | 03:45 | |
| Air Blank | 0.000 | 03:46 | |
| Subject Sample # 1 | 0.154 | 03:46 | |
| Breath Volume | 2.589 | 03:46 | |
| Air Blank | 0.000 | 03:47 | |
| Air Blank | 0.000 | 03:49 | |
| Subject Sample # 2 | | 03:52 | No Sample Provided |
| Breath Volume | 0.000 | 03:52 | |
| Air Blank | 0.000 | 03:53 | |
| Air Blank | 0.000 | 03:55 | |
| Subject Sample # 3 | | 03:55 | Subject Test Refused |
| Breath Volume | 0.000 | 03:55 | |
| Air Blank | 0.000 | 03:55 | |
| Control Test | 0.077 | 03:56 | |
| Air Blank | 0.000 | 03:56 | |
| Diagnostics Check | OK | 03:56 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBOIS,BROOKS,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUBOIS,BROOKS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HAYSLIP

DATE/TIME OF TEST: 02/26/2020 17:18:44

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 17:14 | |
| Air Blank | 0.000 | 17:15 | |
| Control Test | 0.076 | 17:15 | |
| Air Blank | 0.000 | 17:16 | |
| Subject Sample | # 1 | 17:17 | Interferent Detect |
| Breath Volume | 3.394 | 17:17 | |
| Air Blank | 0.000 | 17:17 | |
| Control Test | 0.075 | 17:18 | Interferent Detect |
| Air Blank | 0.000 | 17:18 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBOIS,BROOKS,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUBOIS,BROOKS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HAYSLIP

DATE/TIME OF TEST: 02/26/2020 17:28:00

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:20 |
| Air Blank | 0.000 | 17:21 |
| Control Test | 0.076 | 17:21 |
| Air Blank | 0.000 | 17:22 |
| Subject Sample # 1 | 0.314 | 17:23 |
| Breath Volume | 2.765 | 17:23 |
| Air Blank | 0.000 | 17:23 |
| Air Blank | 0.000 | 17:25 |
| Subject Sample # 2 | 0.314 | 17:26 |
| Breath Volume | 2.824 | 17:26 |
| Air Blank | 0.000 | 17:27 |
| Control Test | 0.076 | 17:27 |
| Air Blank | 0.000 | 17:27 |
| Diagnostics Check | OK | 17:28 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GAAR

DATE/TIME OF TEST: 02/27/2020 01:45:54

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:39 |
| Air Blank | 0.000 | 01:39 |
| Control Test | 0.078 | 01:39 |
| Air Blank | 0.000 | 01:40 |
| Subject Sample # 1 | 0.166 | 01:41 |
| Breath Volume | 1.863 | 01:41 |
| Air Blank | 0.000 | 01:41 |
| Air Blank | 0.000 | 01:43 |
| Subject Sample # 2 | 0.162 | 01:44 |
| Breath Volume | 1.425 | 01:44 |
| Air Blank | 0.000 | 01:44 |
| Control Test | 0.077 | 01:45 |
| Air Blank | 0.000 | 01:45 |
| Diagnostics Check | OK | 01:45 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,KEVIN,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MORRIS,KEVIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:12

SUBJECT'S LAST NAME: KLINO

DATE/TIME OF TEST: 02/27/2020 09:45:24

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 09:40 | |
| Air Blank | 0.000 | 09:40 | |
| Control Test | 0.077 | 09:41 | |
| Air Blank | 0.000 | 09:41 | |
| Subject Sample | # 1 | 09:44 | Subject Test Refused |
| Breath Volume | 0.000 | 09:44 | |
| Air Blank | 0.000 | 09:44 | |
| Control Test | 0.077 | 09:44 | |
| Air Blank | 0.000 | 09:45 | |
| Diagnostics Check | OK | 09:45 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSKO,DAVID,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROSKO,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 07:55:12

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 07:54 | |
| Air Blank | | 07:54 | RFI Detect |
| Air Blank | 0.000 | 07:55 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSKO,DAVID,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROSKO,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 07:58:40

| | | | |
|-------------------|-------|-------|--------------|
| Diagnostics Check | OK | 07:56 | |
| Air Blank | 0.000 | 07:57 | |
| Control Test | 0.078 | 07:57 | |
| Air Blank | | 07:58 | Ambient Fail |
| Air Blank | 0.000 | 07:58 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSKO,DAVID,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROSKO,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 08:03:48

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 08:00 | |
| Air Blank | 0.000 | 08:00 | |
| Control Test | 0.078 | 08:01 | |
| Air Blank | 0.000 | 08:01 | |
| Subject Sample | # 1 | 08:02 | Interferent Detect |
| Breath Volume | 1.785 | 08:02 | |
| Air Blank | 0.000 | 08:02 | |
| Control Test | 0.077 | 08:03 | |
| Air Blank | 0.000 | 08:03 | |
| Diagnostics Check | OK | 08:03 | |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSKO,DAVID,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROSKO,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 02/28/2020 08:12:51

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 08:06 |
| Air Blank | 0.000 | 08:06 |
| Control Test | 0.077 | 08:06 |
| Air Blank | 0.000 | 08:07 |
| Subject Sample # 1 | 0.000 | 08:08 |
| Breath Volume | 2.394 | 08:08 |
| Air Blank | 0.000 | 08:08 |
| Air Blank | 0.000 | 08:10 |
| Subject Sample # 2 | 0.000 | 08:11 |
| Breath Volume | 2.000 | 08:11 |
| Air Blank | 0.000 | 08:11 |
| Control Test | 0.077 | 08:12 |
| Air Blank | 0.000 | 08:12 |
| Diagnostics Check | OK | 08:12 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: JACINTO

DATE/TIME OF TEST: 02/29/2020 07:48:12

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 07:41 |
| Air Blank | 0.000 | 07:41 |
| Control Test | 0.077 | 07:41 |
| Air Blank | 0.000 | 07:42 |
| Subject Sample # 1 | 0.155 | 07:43 |
| Breath Volume | 3.363 | 07:43 |
| Air Blank | 0.000 | 07:44 |
| Air Blank | 0.000 | 07:45 |
| Subject Sample # 2 | 0.152 | 07:46 |
| Breath Volume | 3.800 | 07:46 |
| Air Blank | 0.000 | 07:47 |
| Control Test | 0.076 | 07:47 |
| Air Blank | 0.000 | 07:48 |
| Diagnostics Check | OK | 07:48 |

Data Download Date/Time: 03/06/2020 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PAGUADA MEZA

DATE/TIME OF TEST: 02/05/2020 01:31:11

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:24 |
| Air Blank | 0.000 | 01:24 |
| Control Test | 0.078 | 01:24 |
| Air Blank | 0.000 | 01:25 |
| Subject Sample # 1 | 0.000 | 01:26 |
| Breath Volume | 4.296 | 01:26 |
| Air Blank | 0.000 | 01:27 |
| Air Blank | 0.000 | 01:29 |
| Subject Sample # 2 | 0.000 | 01:29 |
| Breath Volume | 4.550 | 01:29 |
| Air Blank | 0.000 | 01:30 |
| Control Test | 0.079 | 01:30 |
| Air Blank | 0.000 | 01:31 |
| Diagnostics Check | OK | 01:31 |

Data Download Date/Time: 02/07/2020 18:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: LOPEZ VICENTE

DATE/TIME OF TEST: 02/06/2020 02:03:28

| | | | | |
|-------------------|-----|-------|-------|----------------------|
| Diagnostics Check | | OK | 01:53 | |
| Air Blank | | 0.000 | 01:54 | |
| Control Test | | 0.079 | 01:54 | |
| Air Blank | | 0.000 | 01:54 | |
| Subject Sample | # 1 | 0.167 | 01:58 | Volume Not Met |
| Breath Volume | | 0.132 | 01:58 | |
| Air Blank | | 0.000 | 01:58 | |
| Air Blank | | 0.000 | 02:00 | |
| Subject Sample | # 2 | | 02:02 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:02 | |
| Air Blank | | 0.000 | 02:02 | |
| Control Test | | 0.078 | 02:02 | |
| Air Blank | | 0.000 | 02:03 | |
| Diagnostics Check | | OK | 02:03 | |

Data Download Date/Time: 02/07/2020 18:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPLAN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CAPLAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WHEATON

DATE/TIME OF TEST: 02/08/2020 00:00:55

| | | | |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK | 23:49 | |
| Air Blank | 0.000 | 23:50 | |
| Control Test | 0.079 | 23:50 | |
| Air Blank | 0.000 | 23:51 | |
| Subject Sample | # 1 0.148 | 23:54 | Volume Not Met |
| Breath Volume | 0.871 | 23:54 | |
| Air Blank | 0.000 | 23:54 | |
| Air Blank | 0.000 | 23:56 | |
| Subject Sample | # 2 | 23:59 | Subject Test Refused |
| Breath Volume | 1.289 | 23:59 | |
| Air Blank | 0.000 | 23:59 | |
| Control Test | 0.078 | 00:00 | |
| Air Blank | 0.000 | 00:00 | |
| Diagnostics Check | OK | 00:00 | |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 02/08/2020 03:45:49

| | | | |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK | 03:36 | |
| Air Blank | 0.000 | 03:36 | |
| Control Test | 0.078 | 03:37 | |
| Air Blank | 0.000 | 03:37 | |
| Subject Sample | # 1 0.170 | 03:39 | |
| Breath Volume | 1.132 | 03:39 | |
| Air Blank | 0.000 | 03:40 | |
| Air Blank | 0.000 | 03:41 | |
| Subject Sample | # 2 0.148 | 03:44 | |
| Breath Volume | 2.152 | 03:44 | |
| Air Blank | | 03:45 | Purge Fail |
| Air Blank | 0.000 | 03:45 | |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 02/08/2020 04:01:33

| | | | |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK | 03:50 | |
| Air Blank | 0.000 | 03:50 | |
| Control Test | 0.077 | 03:50 | |
| Air Blank | 0.000 | 03:51 | |
| Subject Sample | # 1 | 03:54 | No Sample Provided |
| Breath Volume | 0.000 | 03:54 | |
| Air Blank | 0.000 | 03:55 | |
| Air Blank | 0.000 | 03:57 | |
| Subject Sample | # 2 | 04:00 | No Sample Provided |
| Breath Volume | 0.000 | 04:00 | |
| Air Blank | 0.000 | 04:00 | |
| Control Test | 0.078 | 04:00 | |
| Air Blank | 0.000 | 04:01 | |
| Diagnostics Check | OK | 04:01 | |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEITFELD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: HEITFELD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: DAWES III

DATE/TIME OF TEST: 02/08/2020 17:41:32

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 17:33 |
| Air Blank | 0.000 | 17:33 |
| Control Test | 0.079 | 17:33 |
| Air Blank | 0.000 | 17:34 |
| Subject Sample # 1 | 0.326 | 17:36 |
| Breath Volume | 2.140 | 17:36 |
| Air Blank | 0.000 | 17:37 |
| Air Blank | 0.000 | 17:39 |
| Subject Sample # 2 | 0.334 | 17:39 |
| Breath Volume | 1.984 | 17:39 |
| Air Blank | 0.000 | 17:40 |
| Control Test | 0.075 | 17:40 |
| Air Blank | 0.000 | 17:41 |
| Diagnostics Check | OK | 17:41 |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS, WILLIE, V
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ANDRE, GAETGENS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/09/2020 07:47:25

| | | | |
|-------------------|-------|-------|-----------------|
| Diagnostics Check | OK | 07:45 | |
| Air Blank | 0.000 | 07:46 | |
| Control Test | 0.079 | 07:46 | |
| Air Blank | | 07:46 | Improper Sample |
| Air Blank | 0.000 | 07:47 | |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,WILLIE,V
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ANDRE,GAETGENS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/09/2020 07:56:33

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 07:50 |
| Air Blank | 0.000 | 07:50 |
| Control Test | 0.078 | 07:51 |
| Air Blank | 0.000 | 07:51 |
| Subject Sample # 1 | 0.107 | 07:52 |
| Breath Volume | 3.199 | 07:52 |
| Air Blank | 0.000 | 07:52 |
| Air Blank | 0.000 | 07:54 |
| Subject Sample # 2 | 0.105 | 07:54 |
| Breath Volume | 2.894 | 07:54 |
| Air Blank | 0.000 | 07:55 |
| Control Test | 0.077 | 07:55 |
| Air Blank | 0.000 | 07:56 |
| Diagnostics Check | OK | 07:56 |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,DARIUS,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: HOLMES,DARIUS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: DAWES

DATE/TIME OF TEST: 02/09/2020 20:03:16

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 19:57 |
| Air Blank | 0.000 | 19:57 |
| Control Test | 0.079 | 19:57 |
| Air Blank | 0.000 | 19:58 |
| Subject Sample # 1 | 0.354 | 19:58 |
| Breath Volume | 2.136 | 19:58 |
| Air Blank | 0.000 | 19:59 |
| Air Blank | 0.000 | 20:01 |
| Subject Sample # 2 | 0.335 | 20:01 |
| Breath Volume | 1.421 | 20:01 |
| Air Blank | 0.000 | 20:02 |
| Control Test | 0.078 | 20:02 |
| Air Blank | 0.000 | 20:03 |
| Diagnostics Check | OK | 20:03 |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A4
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: TAYON HERRERA

DATE/TIME OF TEST: 02/10/2020 04:07:02

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 04:00 |
| Air Blank | | 0.000 | 04:00 |
| Control Test | | 0.079 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Subject Sample | # 1 | 0.251 | 04:02 |
| Breath Volume | | 1.769 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Subject Sample | # 2 | 0.252 | 04:05 |
| Breath Volume | | 3.781 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Control Test | | 0.079 | 04:06 |
| Air Blank | | 0.000 | 04:06 |
| Diagnostics Check | | OK | 04:07 |

Data Download Date/Time: 03/27/2020 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 02/08/2020 00:55:44

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:48 |
| Air Blank | | 0.000 | 00:49 |
| Control Test | | 0.079 | 00:49 |
| Air Blank | | 0.000 | 00:50 |
| Subject Sample | # 1 | 0.220 | 00:50 |
| Breath Volume | | 1.726 | 00:50 |
| Air Blank | | 0.000 | 00:51 |
| Air Blank | | 0.000 | 00:53 |
| Subject Sample | # 2 | 0.222 | 00:54 |
| Breath Volume | | 1.378 | 00:54 |
| Air Blank | | 0.000 | 00:54 |
| Control Test | | 0.080 | 00:55 |
| Air Blank | | 0.000 | 00:55 |
| Diagnostics Check | | OK | 00:55 |

Data Download Date/Time: 02/21/2020 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 02/13/2020 01:44:49

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:38 |
| Air Blank | 0.000 | 01:38 |
| Control Test | 0.079 | 01:39 |
| Air Blank | 0.000 | 01:39 |
| Subject Sample # 1 | 0.164 | 01:40 |
| Breath Volume | 1.828 | 01:40 |
| Air Blank | 0.000 | 01:40 |
| Air Blank | 0.000 | 01:42 |
| Subject Sample # 2 | 0.158 | 01:43 |
| Breath Volume | 1.582 | 01:43 |
| Air Blank | 0.000 | 01:43 |
| Control Test | 0.080 | 01:44 |
| Air Blank | 0.000 | 01:44 |
| Diagnostics Check | OK | 01:44 |

Data Download Date/Time: 02/21/2020 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 02/15/2020 01:48:08

| | | | | |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check | | OK | 01:35 | |
| Air Blank | | 0.000 | 01:36 | |
| Control Test | | 0.080 | 01:36 | |
| Air Blank | | 0.000 | 01:37 | |
| Subject Sample | # 1 | 0.054 | 01:40 | Volume Not Met |
| Breath Volume | | 0.570 | 01:40 | |
| Air Blank | | 0.000 | 01:41 | |
| Air Blank | | 0.000 | 01:42 | |
| Subject Sample | # 2 | 0.072 | 01:43 | |
| Breath Volume | | 3.328 | 01:43 | |
| Air Blank | | 0.000 | 01:44 | |
| Air Blank | | 0.000 | 01:45 | |
| Subject Sample | # 3 | 0.060 | 01:46 | |
| Breath Volume | | 1.207 | 01:46 | |
| Air Blank | | 0.000 | 01:47 | |
| Control Test | | 0.077 | 01:47 | |
| Air Blank | | 0.000 | 01:48 | |
| Diagnostics Check | | OK | 01:48 | |

Data Download Date/Time: 02/21/2020 02:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 02/16/2020 02:22:13

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 02:15 |
| Air Blank | | 0.000 | 02:16 |
| Control Test | | 0.079 | 02:16 |
| Air Blank | | 0.000 | 02:17 |
| Subject Sample | # 1 | 0.148 | 02:17 |
| Breath Volume | | 2.265 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 2 | 0.151 | 02:20 |
| Breath Volume | | 1.949 | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Control Test | | 0.079 | 02:21 |
| Air Blank | | 0.000 | 02:22 |
| Diagnostics Check | | OK | 02:22 |

Data Download Date/Time: 02/21/2020 02:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PAQUETTE

DATE/TIME OF TEST: 02/16/2020 03:17:53

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:11 |
| Air Blank | | 0.000 | 03:11 |
| Control Test | | 0.078 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Subject Sample | # 1 | 0.201 | 03:12 |
| Breath Volume | | 1.210 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Air Blank | | 0.000 | 03:15 |
| Subject Sample | # 2 | 0.217 | 03:16 |
| Breath Volume | | 1.648 | 03:16 |
| Air Blank | | 0.000 | 03:16 |
| Control Test | | 0.080 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Diagnostics Check | | OK | 03:17 |

Data Download Date/Time: 02/21/2020 02:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KANE,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: NARDINI-FREYRE

DATE/TIME OF TEST: 02/01/2020 16:15:17

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 16:09 |
| Air Blank | | 0.000 | 16:09 |
| Control Test | | 0.080 | 16:09 |
| Air Blank | | 0.000 | 16:10 |
| Subject Sample | # 1 | 0.260 | 16:10 |
| Breath Volume | | 2.652 | 16:10 |
| Air Blank | | 0.000 | 16:11 |
| Air Blank | | 0.000 | 16:13 |
| Subject Sample | # 2 | 0.253 | 16:13 |
| Breath Volume | | 2.636 | 16:13 |
| Air Blank | | 0.000 | 16:14 |
| Control Test | | 0.079 | 16:14 |
| Air Blank | | 0.000 | 16:15 |
| Diagnostics Check | | OK | 16:15 |

Data Download Date/Time: 02/05/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: PALAZOLO

DATE/TIME OF TEST: 02/01/2020 22:24:07

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:16 |
| Air Blank | 0.000 | 22:17 |
| Control Test | 0.081 | 22:17 |
| Air Blank | 0.000 | 22:17 |
| Subject Sample # 1 | 0.106 | 22:18 |
| Breath Volume | 2.265 | 22:18 |
| Air Blank | 0.000 | 22:19 |
| Air Blank | 0.000 | 22:21 |
| Subject Sample # 2 | 0.102 | 22:22 |
| Breath Volume | 2.800 | 22:22 |
| Air Blank | 0.000 | 22:23 |
| Control Test | 0.081 | 22:23 |
| Air Blank | 0.000 | 22:24 |
| Diagnostics Check | OK | 22:24 |

Data Download Date/Time: 02/05/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: AMBRIOSO

DATE/TIME OF TEST: 02/01/2020 23:59:57

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:53 |
| Air Blank | 0.000 | 23:53 |
| Control Test | 0.080 | 23:53 |
| Air Blank | 0.000 | 23:54 |
| Subject Sample # 1 | 0.176 | 23:55 |
| Breath Volume | 3.195 | 23:55 |
| Air Blank | 0.000 | 23:55 |
| Air Blank | 0.000 | 23:57 |
| Subject Sample # 2 | 0.184 | 23:58 |
| Breath Volume | 3.878 | 23:58 |
| Air Blank | 0.000 | 23:59 |
| Control Test | 0.080 | 23:59 |
| Air Blank | 0.000 | 23:59 |
| Diagnostics Check | OK | 23:59 |

Data Download Date/Time: 02/05/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: REYES DE JESUS

DATE/TIME OF TEST: 02/02/2020 03:47:07

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.080 | 03:41 |
| Air Blank | | 0.000 | 03:42 |
| Subject Sample | # 1 | 0.125 | 03:42 |
| Breath Volume | | 2.484 | 03:42 |
| Air Blank | | 0.000 | 03:43 |
| Air Blank | | 0.000 | 03:45 |
| Subject Sample | # 2 | 0.127 | 03:45 |
| Breath Volume | | 3.082 | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.080 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Diagnostics Check | | OK | 03:47 |

Data Download Date/Time: 02/05/2020 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 02/06/2020 20:49:41

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 20:42 |
| Air Blank | 0.000 | 20:43 |
| Control Test | 0.080 | 20:43 |
| Air Blank | 0.000 | 20:43 |
| Subject Sample # 1 | 0.147 | 20:44 |
| Breath Volume | 2.589 | 20:44 |
| Air Blank | 0.000 | 20:45 |
| Air Blank | 0.000 | 20:47 |
| Subject Sample # 2 | 0.147 | 20:48 |
| Breath Volume | 2.781 | 20:48 |
| Air Blank | 0.000 | 20:48 |
| Control Test | 0.080 | 20:49 |
| Air Blank | 0.000 | 20:49 |
| Diagnostics Check | OK | 20:49 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CATANIA

DATE/TIME OF TEST: 02/07/2020 02:15:26

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:08 |
| Air Blank | 0.000 | 02:09 |
| Control Test | 0.080 | 02:09 |
| Air Blank | 0.000 | 02:09 |
| Subject Sample # 1 | 0.161 | 02:10 |
| Breath Volume | 2.710 | 02:10 |
| Air Blank | 0.000 | 02:11 |
| Air Blank | 0.000 | 02:13 |
| Subject Sample # 2 | 0.165 | 02:13 |
| Breath Volume | 2.664 | 02:13 |
| Air Blank | 0.000 | 02:14 |
| Control Test | 0.080 | 02:14 |
| Air Blank | 0.000 | 02:15 |
| Diagnostics Check | OK | 02:15 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRONK,MATTHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: WALDRUP

DATE/TIME OF TEST: 02/07/2020 10:23:58

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 10:17 |
| Air Blank | 0.000 | 10:18 |
| Control Test | 0.080 | 10:18 |
| Air Blank | 0.000 | 10:18 |
| Subject Sample # 1 | 0.000 | 10:19 |
| Breath Volume | 2.863 | 10:19 |
| Air Blank | 0.000 | 10:20 |
| Air Blank | 0.000 | 10:22 |
| Subject Sample # 2 | 0.000 | 10:22 |
| Breath Volume | 2.667 | 10:22 |
| Air Blank | 0.000 | 10:23 |
| Control Test | 0.079 | 10:23 |
| Air Blank | 0.000 | 10:23 |
| Diagnostics Check | OK | 10:23 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 02/08/2020 13:40:51

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 13:34 |
| Air Blank | 0.000 | 13:35 |
| Control Test | 0.081 | 13:35 |
| Air Blank | 0.000 | 13:35 |
| Subject Sample # 1 | 0.099 | 13:36 |
| Breath Volume | 2.539 | 13:36 |
| Air Blank | 0.000 | 13:37 |
| Air Blank | 0.000 | 13:38 |
| Subject Sample # 2 | 0.094 | 13:39 |
| Breath Volume | 1.386 | 13:39 |
| Air Blank | 0.000 | 13:39 |
| Control Test | 0.080 | 13:40 |
| Air Blank | 0.000 | 13:40 |
| Diagnostics Check | OK | 13:40 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,J,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHINKEL,M
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/11/2020 14:21:38

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 14:15 |
| Air Blank | 0.000 | 14:15 |
| Control Test | 0.080 | 14:15 |
| Air Blank | 0.000 | 14:16 |
| Subject Sample # 1 | 0.131 | 14:17 |
| Breath Volume | 1.761 | 14:17 |
| Air Blank | 0.000 | 14:17 |
| Air Blank | 0.000 | 14:19 |
| Subject Sample # 2 | 0.129 | 14:20 |
| Breath Volume | 1.839 | 14:20 |
| Air Blank | 0.000 | 14:20 |
| Control Test | 0.080 | 14:21 |
| Air Blank | 0.000 | 14:21 |
| Diagnostics Check | OK | 14:21 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/14/2020 22:02:31

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:55 |
| Air Blank | 0.000 | 21:56 |
| Control Test | 0.080 | 21:56 |
| Air Blank | 0.000 | 21:56 |
| Subject Sample # 1 | 0.189 | 21:57 |
| Breath Volume | 3.589 | 21:57 |
| Air Blank | 0.000 | 21:58 |
| Air Blank | 0.000 | 22:00 |
| Subject Sample # 2 | 0.189 | 22:01 |
| Breath Volume | 3.238 | 22:01 |
| Air Blank | 0.000 | 22:01 |
| Control Test | 0.080 | 22:01 |
| Air Blank | 0.000 | 22:02 |
| Diagnostics Check | OK | 22:02 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DEWEY

DATE/TIME OF TEST: 02/15/2020 02:56:56

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 02:50 |
| Air Blank | 0.000 | 02:50 |
| Control Test | 0.080 | 02:51 |
| Air Blank | 0.000 | 02:51 |
| Subject Sample # 1 | 0.173 | 02:52 |
| Breath Volume | 1.703 | 02:52 |
| Air Blank | 0.000 | 02:52 |
| Air Blank | 0.000 | 02:54 |
| Subject Sample # 2 | 0.163 | 02:55 |
| Breath Volume | 2.375 | 02:55 |
| Air Blank | 0.000 | 02:56 |
| Control Test | 0.080 | 02:56 |
| Air Blank | 0.000 | 02:56 |
| Diagnostics Check | OK | 02:56 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: OCHS,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: ZIOLKOWSKI

DATE/TIME OF TEST: 02/15/2020 18:59:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:53 |
| Air Blank | 0.000 | 18:53 |
| Control Test | 0.080 | 18:53 |
| Air Blank | 0.000 | 18:54 |
| Subject Sample # 1 | 0.000 | 18:54 |
| Breath Volume | 4.937 | 18:54 |
| Air Blank | 0.000 | 18:55 |
| Air Blank | 0.000 | 18:57 |
| Subject Sample # 2 | 0.000 | 18:57 |
| Breath Volume | 4.941 | 18:57 |
| Air Blank | 0.000 | 18:58 |
| Control Test | 0.081 | 18:58 |
| Air Blank | 0.000 | 18:59 |
| Diagnostics Check | OK | 18:59 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/17/2020 01:03:12

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:57 |
| Air Blank | | 0.000 | 00:57 |
| Control Test | | 0.079 | 00:57 |
| Air Blank | | 0.000 | 00:58 |
| Subject Sample | # 1 | 0.103 | 00:58 |
| Breath Volume | | 1.496 | 00:58 |
| Air Blank | | 0.000 | 00:59 |
| Air Blank | | 0.000 | 01:01 |
| Subject Sample | # 2 | 0.107 | 01:01 |
| Breath Volume | | 1.300 | 01:01 |
| Air Blank | | 0.000 | 01:02 |
| Control Test | | 0.080 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Diagnostics Check | | OK | 01:03 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LAGOW

DATE/TIME OF TEST: 02/17/2020 04:50:04

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 04:43 |
| Air Blank | 0.000 | 04:43 |
| Control Test | 0.080 | 04:44 |
| Air Blank | 0.000 | 04:44 |
| Subject Sample # 1 | 0.158 | 04:45 |
| Breath Volume | 2.535 | 04:45 |
| Air Blank | 0.000 | 04:46 |
| Air Blank | 0.000 | 04:48 |
| Subject Sample # 2 | 0.144 | 04:48 |
| Breath Volume | 2.761 | 04:48 |
| Air Blank | 0.000 | 04:49 |
| Control Test | 0.079 | 04:49 |
| Air Blank | 0.000 | 04:49 |
| Diagnostics Check | OK | 04:50 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGIEL,ADRIAN,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCHWIGK,GEORGE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: JAMES JR

DATE/TIME OF TEST: 02/18/2020 22:56:22

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:50 |
| Air Blank | 0.000 | 22:50 |
| Control Test | 0.080 | 22:50 |
| Air Blank | 0.000 | 22:51 |
| Subject Sample # 1 | 0.000 | 22:51 |
| Breath Volume | 1.648 | 22:51 |
| Air Blank | 0.000 | 22:52 |
| Air Blank | 0.000 | 22:54 |
| Subject Sample # 2 | 0.000 | 22:55 |
| Breath Volume | 1.101 | 22:55 |
| Air Blank | 0.000 | 22:55 |
| Control Test | 0.081 | 22:55 |
| Air Blank | 0.000 | 22:56 |
| Diagnostics Check | OK | 22:56 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:40:51

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 01:40 | |
| Air Blank | | 01:40 | RFI Detect |
| Air Blank | 0.000 | 01:40 | |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:43:36

| | | | |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK | 01:43 | |
| Air Blank | | 01:43 | RFI Detect |
| Air Blank | 0.000 | 01:43 | |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:47:33

| | | | |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK | 01:45 | |
| Air Blank | 0.000 | 01:45 | |
| Control Test | 0.080 | 01:46 | |
| Air Blank | 0.000 | 01:46 | |
| Subject Sample | # 1 0.193 | 01:47 | |
| Breath Volume | 2.382 | 01:47 | |
| Air Blank | | 01:47 | RFI Detect |
| Air Blank | | 01:47 | RFI Detect |
| Air Blank | | 01:47 | RFI Detect |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARRILLO,N,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SURA

DATE/TIME OF TEST: 02/25/2020 01:57:05

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 01:49 |
| Air Blank | 0.000 | 01:50 |
| Control Test | 0.080 | 01:50 |
| Air Blank | 0.000 | 01:50 |
| Subject Sample # 1 | 0.178 | 01:52 |
| Breath Volume | 1.539 | 01:52 |
| Air Blank | 0.000 | 01:53 |
| Air Blank | 0.000 | 01:55 |
| Subject Sample # 2 | 0.189 | 01:55 |
| Breath Volume | 1.781 | 01:55 |
| Air Blank | 0.000 | 01:56 |
| Control Test | 0.079 | 01:56 |
| Air Blank | 0.000 | 01:57 |
| Diagnostics Check | OK | 01:57 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK, MICHAEL, O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ, CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 02/27/2020 01:09:16

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 01:02 |
| Air Blank | | 0.000 | 01:02 |
| Control Test | | 0.080 | 01:03 |
| Air Blank | | 0.000 | 01:03 |
| Subject Sample | # 1 | 0.170 | 01:04 |
| Breath Volume | | 1.914 | 01:04 |
| Air Blank | | 0.000 | 01:05 |
| Air Blank | | 0.000 | 01:07 |
| Subject Sample | # 2 | 0.167 | 01:07 |
| Breath Volume | | 2.277 | 01:07 |
| Air Blank | | 0.000 | 01:08 |
| Control Test | | 0.080 | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Diagnostics Check | | OK | 01:09 |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,CHEYANNE,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ROSS,CHEYANNE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CAPIZZI

DATE/TIME OF TEST: 02/27/2020 23:22:41

| | | | |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK | 23:17 | |
| Air Blank | 0.000 | 23:17 | |
| Control Test | 0.080 | 23:18 | |
| Air Blank | 0.000 | 23:18 | |
| Subject Sample | # 1 | 23:21 | Subject Test Refused |
| Breath Volume | 0.019 | 23:21 | |
| Air Blank | 0.000 | 23:21 | |
| Control Test | 0.080 | 23:22 | |
| Air Blank | 0.000 | 23:22 | |
| Diagnostics Check | OK | 23:22 | |

Data Download Date/Time: 03/05/2020 08:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/29/2020 00:58:55

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 00:52 |
| Air Blank | | 0.000 | 00:52 |
| Control Test | | 0.080 | 00:53 |
| Air Blank | | 0.000 | 00:53 |
| Subject Sample | # 1 | 0.237 | 00:54 |
| Breath Volume | | 1.906 | 00:54 |
| Air Blank | | 0.000 | 00:55 |
| Air Blank | | 0.000 | 00:56 |
| Subject Sample | # 2 | 0.235 | 00:57 |
| Breath Volume | | 3.019 | 00:57 |
| Air Blank | | 0.000 | 00:58 |
| Control Test | | 0.080 | 00:58 |
| Air Blank | | 0.000 | 00:58 |
| Diagnostics Check | | OK | 00:58 |

Data Download Date/Time: 03/05/2020 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WETMORE

DATE/TIME OF TEST: 02/29/2020 23:45:08

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 23:38 |
| Air Blank | 0.000 | 23:39 |
| Control Test | 0.081 | 23:39 |
| Air Blank | 0.000 | 23:40 |
| Subject Sample # 1 | 0.063 | 23:40 |
| Breath Volume | 4.839 | 23:40 |
| Air Blank | 0.000 | 23:41 |
| Air Blank | 0.000 | 23:43 |
| Subject Sample # 2 | 0.063 | 23:43 |
| Breath Volume | 5.425 | 23:43 |
| Air Blank | 0.000 | 23:44 |
| Control Test | 0.081 | 23:44 |
| Air Blank | 0.000 | 23:45 |
| Diagnostics Check | OK | 23:45 |

Data Download Date/Time: 03/05/2020 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: DIETER

DATE/TIME OF TEST: 02/05/2020 22:50:10

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 22:43 |
| Air Blank | 0.000 | 22:44 |
| Control Test | 0.080 | 22:44 |
| Air Blank | 0.000 | 22:44 |
| Subject Sample # 1 | 0.032 | 22:45 |
| Breath Volume | 4.070 | 22:45 |
| Air Blank | 0.000 | 22:46 |
| Air Blank | 0.000 | 22:48 |
| Subject Sample # 2 | 0.031 | 22:48 |
| Breath Volume | 3.640 | 22:48 |
| Air Blank | 0.000 | 22:49 |
| Control Test | 0.080 | 22:49 |
| Air Blank | 0.000 | 22:50 |
| Diagnostics Check | OK | 22:50 |

Data Download Date/Time: 02/26/2020 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MONTANEZ,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: PENTTILA

DATE/TIME OF TEST: 02/29/2020 21:11:12

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 21:04 |
| Air Blank | 0.000 | 21:04 |
| Control Test | 0.081 | 21:05 |
| Air Blank | 0.000 | 21:05 |
| Subject Sample # 1 | 0.182 | 21:06 |
| Breath Volume | 5.191 | 21:06 |
| Air Blank | 0.000 | 21:07 |
| Air Blank | 0.000 | 21:08 |
| Subject Sample # 2 | 0.175 | 21:09 |
| Breath Volume | 3.664 | 21:09 |
| Air Blank | 0.000 | 21:10 |
| Control Test | 0.081 | 21:10 |
| Air Blank | 0.000 | 21:11 |
| Diagnostics Check | OK | 21:11 |

Data Download Date/Time: 03/26/2020 08:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,DARRYL,D
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: RODGERS,WILL
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 01/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/11/2020 18:43:39

| | | |
|--------------------|-------|-------|
| Diagnostics Check | OK | 18:37 |
| Air Blank | 0.000 | 18:38 |
| Control Test | 0.083 | 18:38 |
| Air Blank | 0.000 | 18:39 |
| Subject Sample # 1 | 0.000 | 18:39 |
| Breath Volume | 1.878 | 18:39 |
| Air Blank | 0.000 | 18:39 |
| Air Blank | 0.000 | 18:41 |
| Subject Sample # 2 | 0.000 | 18:42 |
| Breath Volume | 1.898 | 18:42 |
| Air Blank | 0.000 | 18:42 |
| Control Test | 0.082 | 18:43 |
| Air Blank | 0.000 | 18:43 |
| Diagnostics Check | OK | 18:43 |

Data Download Date/Time: 02/12/2020 03:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEETZ,JOHN,W
BREATH TEST OPERATOR AGENCY: Calhoun County SO
ARRESTING OFFICER: DILLARD,PAUL
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: MONTFORD

DATE/TIME OF TEST: 02/21/2020 07:56:25

| | | | |
|-------------------|-----|-------|-------|
| Diagnostics Check | | OK | 07:49 |
| Air Blank | | 0.000 | 07:50 |
| Control Test | | 0.083 | 07:50 |
| Air Blank | | 0.000 | 07:51 |
| Subject Sample | # 1 | 0.265 | 07:51 |
| Breath Volume | | 2.042 | 07:51 |
| Air Blank | | 0.000 | 07:52 |
| Air Blank | | 0.000 | 07:54 |
| Subject Sample | # 2 | 0.253 | 07:54 |
| Breath Volume | | 2.332 | 07:54 |
| Air Blank | | 0.000 | 07:55 |
| Control Test | | 0.083 | 07:55 |
| Air Blank | | 0.000 | 07:56 |
| Diagnostics Check | | OK | 07:56 |

Data Download Date/Time: 03/02/2020 03:07