Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEDE,BRENNEN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CLAUSEN

DATE/TIME OF TEST: 02/16/2020 03:44:58

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.081	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:43	Subject Test Refused
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,CORPORAL WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SURITA,TROOPER C.
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: JAQUIN

DATE/TIME OF TEST: 02/17/2020 20:35:24

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:31	
Control Test		0.081	20:32	
Air Blank		0.000	20:32	
Subject Sample	# 1		20:34	Subject Test Refused
Breath Volume		0.000	20:34	
Air Blank		0.000	20:34	
Control Test		0.081	20:34	
Air Blank		0.000	20:35	
Diagnostics Check		OK	20:35	

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 02/19/2020 02:15:32

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.082	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.133	02:11
Breath Volume		2.742	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.126	02:14
Breath Volume		1.492	02:14
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 02/22/2020 21:34:13

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.081	21:27
Air Blank		0.000	21:28
Subject Sample	# 1	0.000	21:29
Breath Volume		2.503	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.000	21:32
Breath Volume		1.542	21:32
Air Blank		0.000	21:33
Control Test		0.081	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,COURTNEY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: LAFLAMME

DATE/TIME OF TEST: 02/23/2020 02:42:54

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.081	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.149	02:38
Breath Volume		1.722	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:40
Subject Sample	# 2	0.158	02:41
Breath Volume		1.792	02:41
Air Blank		0.000	02:41
Control Test		0.081	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: REVELS

DATE/TIME OF TEST: 02/23/2020 03:12:19

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.161	03:08
Breath Volume		2.035	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.150	03:10
Breath Volume		1.234	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/23/2020 04:30:58

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:25	
Control Test		0.081	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:30	
Control Test		0.080	04:30	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 02/23/2020 05:12:52

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.123	05:08
Breath Volume		1.285	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:10
Subject Sample	# 2	0.120	05:11
Breath Volume		1.445	05:11
Air Blank		0.000	05:11
Control Test		0.080	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/23/2020 22:12:06

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.081	22:06
Air Blank		0.000	22:07
Subject Sample	# 1	0.265	22:07
Breath Volume		2.175	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.253	22:10
Breath Volume		2.273	22:10
Air Blank		0.000	22:11
Control Test		0.077	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: HEFFNER

DATE/TIME OF TEST: 02/24/2020 00:11:43

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.167	00:06
Breath Volume		2.050	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.165	00:10
Breath Volume		2.503	00:10
Air Blank		0.000	00:10
Control Test		0.081	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: SALTER

DATE/TIME OF TEST: 02/26/2020 07:11:43

Diagnostics Check		OK	07:03
Air Blank		0.000	07:03
Control Test		0.081	07:03
Air Blank		0.000	07:04
Subject Sample	# 1	0.097	07:06
Breath Volume		1.488	07:06
Air Blank		0.000	07:07
Air Blank		0.000	07:09
Subject Sample	# 2	0.107	07:10
Breath Volume		1.367	07:10
Air Blank		0.000	07:10
Control Test		0.080	07:11
Air Blank		0.000	07:11
Diagnostics Check		OK	07:11

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 02/26/2020 19:45:55

Diagnostics Check		OK	19:38
Air Blank		0.000	19:39
Control Test		0.080	19:39
Air Blank		0.000	19:40
Subject Sample	# 1	0.112	19:40
Breath Volume		2.195	19:40
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.114	19:44
Breath Volume		2.433	19:44
Air Blank		0.000	19:45
Control Test		0.079	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/27/2020 02:46:10

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.000	02:41
Breath Volume		2.144	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.000	02:44
Breath Volume		2.328	02:44
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNAVELY,KENT,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NORRIS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: NYGAARD

DATE/TIME OF TEST: 02/27/2020 15:04:53

Diagnostics Check		OK	14:55
Air Blank		0.000	14:56
Control Test		0.081	14:56
Air Blank		0.000	14:56
Subject Sample	# 1	0.182	14:57
Breath Volume		2.648	14:57
Air Blank		0.000	14:57
Air Blank		0.000	14:59
Subject Sample	# 2	0.161	15:00
Breath Volume		1.375	15:00
Air Blank		0.000	15:00
Air Blank		0.000	15:02
Subject Sample	# 3	0.175	15:03
Breath Volume		1.968	15:03
Air Blank		0.000	15:03
Control Test		0.081	15:04
Air Blank		0.000	15:04
Diagnostics Check		OK	15:04

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 02/28/2020 02:02:46

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.081	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.202	01:57
Breath Volume		1.894	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.195	02:01
Breath Volume		1.519	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 02/28/2020 21:31:25

Diagnostics Check		OK	21:24
Air Blank		0.000	21:25
Control Test		0.081	21:25
Air Blank		0.000	21:25
Subject Sample	# 1	0.155	21:26
Breath Volume		1.210	21:26
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.149	21:29
Breath Volume		1.371	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 02/29/2020 01:09:05

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.081	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.093	01:04
Breath Volume		2.867	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.095	01:07
Breath Volume		2.789	01:07
Air Blank		0.000	01:08
Control Test		0.081	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REQUENA,CARLOS,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: REQUENA,CARLOS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GRIJALVA

DATE/TIME OF TEST: 02/29/2020 02:37:10

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.150	02:32
Breath Volume		2.730	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.151	02:35
Breath Volume		1.835	02:35
Air Blank		0.000	02:36
Control Test		0.081	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG916303
DRY GAS EXPIRY DATE: 06/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: DIAZ JR

DATE/TIME OF TEST: 02/29/2020 04:02:19

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.057	03:57
Breath Volume		1.746	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.058	04:00
Breath Volume		1.605	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 03/06/2020 23:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: FLANNERY

DATE/TIME OF TEST: 02/01/2020 05:22:28

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:08	
Control Test		0.080	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 1	0.098	05:12	Volume Not Met
Breath Volume		0.101	05:12	
Air Blank		0.000	05:12	
Air Blank		0.000	05:14	
Subject Sample	# 2	0.106	05:17	
Breath Volume		1.406	05:17	
Air Blank		0.000	05:18	
Air Blank		0.000	05:19	
Subject Sample	# 3	0.100	05:20	
Breath Volume		1.250	05:20	
Air Blank		0.000	05:21	
Control Test		0.078	05:21	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 02/03/2020 11:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GILROY,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/03/2020 03:03:28

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.190	02:58
Breath Volume		3.082	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.192	03:01
Breath Volume		2.546	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 02/03/2020 11:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: GARRISON

DATE/TIME OF TEST: 02/03/2020 21:05:36

Diagnostics Check		OK	20:59
Air Blank		0.000	20:59
Control Test		0.080	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.261	21:00
Breath Volume		1.937	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.262	21:03
Breath Volume		1.652	21:03
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,GREGORY,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: HEDBLOM

DATE/TIME OF TEST: 02/06/2020 14:00:02

Diagnostics Check		OK	13:53
Air Blank		0.000	13:53
Control Test		0.079	13:54
Air Blank		0.000	13:54
Subject Sample	# 1	0.000	13:55
Breath Volume		1.968	13:55
Air Blank		0.000	13:55
Air Blank		0.000	13:57
Subject Sample	# 2	0.000	13:58
Breath Volume		1.695	13:58
Air Blank		0.000	13:59
Control Test		0.079	13:59
Air Blank		0.000	13:59
Diagnostics Check		OK	14:00

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: ROWSE

DATE/TIME OF TEST: 02/06/2020 22:29:40

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:26	
Control Test		0.080	22:26	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Control Test		0.080	22:29	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/07/2020 04:17:41

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.080	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:16	Subject Test Refused
Breath Volume		0.000	04:16	
Air Blank		0.000	04:16	
Control Test		0.080	04:17	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ROMBER

DATE/TIME OF TEST: 02/07/2020 21:13:11

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.080	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.080	21:08
Breath Volume		2.855	21:08
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.079	21:11
Breath Volume		2.761	21:11
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GOLDIE III

DATE/TIME OF TEST: 02/08/2020 02:35:00

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.124	02:30
Breath Volume		3.792	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.123	02:33
Breath Volume		2.328	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CANTERO DIAZ

DATE/TIME OF TEST: 02/13/2020 00:48:29

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.000	00:44
Breath Volume		2.207	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.000	00:47
Breath Volume		2.179	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: SHERIDAN

DATE/TIME OF TEST: 02/14/2020 16:57:48

Diagnostics Check		OK	16:50
Air Blank		0.000	16:50
Control Test		0.078	16:51
Air Blank		0.000	16:51
Subject Sample	# 1	0.155	16:52
Breath Volume		2.753	16:52
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.152	16:56
Breath Volume		2.589	16:56
Air Blank		0.000	16:56
Control Test		0.079	16:57
Air Blank		0.000	16:57
Diagnostics Check		OK	16:57

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: LONESS

DATE/TIME OF TEST: 02/18/2020 02:57:34

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS,M,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CONNORS

DATE/TIME OF TEST: 02/18/2020 03:10:52

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.079	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.080	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/22/2020 22:32:30

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.081	22:30	
Air Blank		22:31	Purge Fail
Air Blank		22:32	Purge Fail

Data Download Date/Time: 03/06/2020 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/22/2020 22:36:40

Diagnostics Check	OK	22:35	
Air Blank		22:36	Ambient Fail
Air Blank	0.000	22:36	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,MEGAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/22/2020 22:41:32

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:39	
Control Test		0.081	22:39	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.080	22:40	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: BLYDENBURGH

DATE/TIME OF TEST: 02/22/2020 23:19:53

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.081	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.000	23:15
Breath Volume		1.480	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.000	23:18
Breath Volume		1.953	23:18
Air Blank		0.000	23:18
Control Test		0.081	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/23/2020 00:06:39

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.080	00:04	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.080	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 02/23/2020 01:21:54

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.104	01:16
Breath Volume		1.882	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.104	01:20
Breath Volume		1.531	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BRILLANTES

DATE/TIME OF TEST: 02/23/2020 02:03:26

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.081	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Control Test		0.080	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LIPSCOMB

DATE/TIME OF TEST: 02/23/2020 02:38:37

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.185	02:33
Breath Volume		2.667	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.195	02:36
Breath Volume		1.984	02:36
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HEMPHILL

DATE/TIME OF TEST: 02/23/2020 02:49:10

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.142	02:44
Breath Volume		5.250	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:46
Subject Sample	# 2	0.145	02:47
Breath Volume		5.367	02:47
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG815602
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 02/23/2020 03:55:41

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.224	03:50
Breath Volume		4.691	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.207	03:54
Breath Volume		4.484	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 03/06/2020 20:11

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,ZACHARY,C
BREATH TEST OPERATOR AGENCY: Melbourne Beach PD
ARRESTING OFFICER: MARTIN,ZACHARY
ARRESTING OFFICER AGENCY: Melbourne Beach PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 02/03/2020 00:04:36

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.173	23:58
Breath Volume		1.796	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.158	00:03
Breath Volume		2.710	00:03
Air Blank		0.000	00:03
Control Test		0.078	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 02/25/2020 10:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: CRUZ,EDGRADO
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 02/01/2020 03:45:42

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.087	03:41
Breath Volume		4.375	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.086	03:44
Breath Volume		5.105	03:44
Air Blank		0.000	03:44
Control Test		0.083	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/24/2020 11:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/01/2020 04:57:06

Diagnostics Check		OK	04:54	
Air Blank		0.000	04:54	
Control Test		0.081	04:55	
Air Blank		0.000	04:55	
Subject Sample	# 1		04:55	Subject Test Refused
Breath Volume		0.000	04:55	
Air Blank		0.000	04:56	
Control Test		0.082	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 02/24/2020 11:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 02/01/2020 05:06:03

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.230	05:01
Breath Volume		2.289	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.226	05:04
Breath Volume		2.191	05:04
Air Blank		0.000	05:05
Control Test		0.082	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:06

Data Download Date/Time: 02/24/2020 11:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BILALOVSKI

DATE/TIME OF TEST: 02/03/2020 00:04:39

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.081	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	
Control Test		0.082	00:04	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 02/24/2020 11:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: FERMIN

DATE/TIME OF TEST: 02/05/2020 23:16:44

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.083	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.000	23:12
Breath Volume		1.695	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.000	23:15
Breath Volume		2.089	23:15
Air Blank		0.000	23:15
Control Test		0.081	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 02/24/2020 11:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CABRERA SANTIAGO

DATE/TIME OF TEST: 02/07/2020 03:04:13

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.082	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.034	02:59
Breath Volume		2.433	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.029	03:02
Breath Volume		2.300	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 02/24/2020 11:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 02/12/2020 16:22:46

Diagnostics Check		OK	16:16
Air Blank		0.000	16:17
Control Test		0.082	16:17
Air Blank		0.000	16:18
Subject Sample	# 1	0.000	16:18
Breath Volume		2.160	16:18
Air Blank		0.000	16:19
Air Blank		0.000	16:21
Subject Sample	# 2	0.000	16:21
Breath Volume		2.003	16:21
Air Blank		0.000	16:21
Control Test		0.083	16:22
Air Blank		0.000	16:22
Diagnostics Check		OK	16:22

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALFRED,BRANDON,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALFRED,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: LOPEZ BADILLO

DATE/TIME OF TEST: 02/15/2020 06:21:13

Diagnostics Check		OK	06:14
Air Blank		0.000	06:14
Control Test		0.081	06:14
Air Blank		0.000	06:15
Subject Sample	# 1	0.153	06:16
Breath Volume		1.898	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.157	06:19
Breath Volume		1.847	06:19
Air Blank		0.000	06:20
Control Test		0.082	06:20
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CHAD,B
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 02/16/2020 13:05:07

Diagnostics Check		OK	12:58
Air Blank		0.000	12:59
Control Test		0.082	12:59
Air Blank		0.000	13:00
Subject Sample	# 1	0.010	13:00
Breath Volume		1.605	13:00
Air Blank		0.000	13:01
Air Blank		0.000	13:03
Subject Sample	# 2	0.009	13:03
Breath Volume		1.464	13:03
Air Blank		0.000	13:04
Control Test		0.080	13:04
Air Blank		0.000	13:05
Diagnostics Check		OK	13:05

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 02/17/2020 02:05:35

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.082	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.145	02:01
Breath Volume		2.117	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.137	02:04
Breath Volume		1.851	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: OLIVO-FELICIANO

DATE/TIME OF TEST: 02/19/2020 23:30:04

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.082	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.144	23:25
Breath Volume		1.945	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.138	23:28
Breath Volume		1.796	23:28
Air Blank		0.000	23:29
Control Test		0.082	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:30

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,RAY,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WEBER,RAY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: KOEHLER

DATE/TIME OF TEST: 02/20/2020 10:03:28

Diagnostics Check		OK	09:57
Air Blank		0.000	09:58
Control Test		0.081	09:58
Air Blank		0.000	09:59
Subject Sample	# 1	0.000	09:59
Breath Volume		3.890	09:59
Air Blank		0.000	09:59
Air Blank		0.000	10:01
Subject Sample	# 2	0.000	10:02
Breath Volume		4.066	10:02
Air Blank		0.000	10:02
Control Test		0.078	10:02
Air Blank		0.000	10:03
Diagnostics Check		OK	10:03

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,RAY,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WEBER,RAY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: ROMAN ROSA

DATE/TIME OF TEST: 02/20/2020 10:17:22

Diagnostics Check		OK	10:11
Air Blank		0.000	10:12
Control Test		0.081	10:12
Air Blank		0.000	10:12
Subject Sample	# 1	0.000	10:13
Breath Volume		2.742	10:13
Air Blank		0.000	10:13
Air Blank		0.000	10:15
Subject Sample	# 2	0.000	10:16
Breath Volume		2.722	10:16
Air Blank		0.000	10:16
Control Test		0.084	10:16
Air Blank		0.000	10:17
Diagnostics Check		OK	10:17

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: DEEMER

DATE/TIME OF TEST: 02/20/2020 11:57:49

Diagnostics Check		OK	11:51
Air Blank		0.000	11:52
Control Test		0.082	11:52
Air Blank		0.000	11:53
Subject Sample	# 1	0.000	11:53
Breath Volume		1.687	11:53
Air Blank		0.000	11:54
Air Blank		0.000	11:56
Subject Sample	# 2	0.000	11:56
Breath Volume		1.832	11:56
Air Blank		0.000	11:56
Control Test		0.079	11:57
Air Blank		0.000	11:57
Diagnostics Check		OK	11:57

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,RAY,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WEBER,RAY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: HOKE

DATE/TIME OF TEST: 02/20/2020 13:28:25

Diagnostics Check		OK	13:22
Air Blank		0.000	13:23
Control Test		0.083	13:23
Air Blank		0.000	13:23
Subject Sample	# 1	0.000	13:24
Breath Volume		2.476	13:24
Air Blank		0.000	13:24
Air Blank		0.000	13:26
Subject Sample	# 2	0.000	13:27
Breath Volume		1.929	13:27
Air Blank		0.000	13:27
Control Test		0.082	13:27
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,RAY,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WEBER,RAY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: KARMARIE

DATE/TIME OF TEST: 02/20/2020 13:41:36

Diagnostics Check		OK	13:35
Air Blank		0.000	13:36
Control Test		0.082	13:36
Air Blank		0.000	13:36
Subject Sample	# 1	0.000	13:37
Breath Volume		1.335	13:37
Air Blank		0.000	13:37
Air Blank		0.000	13:39
Subject Sample	# 2	0.000	13:40
Breath Volume		1.136	13:40
Air Blank		0.000	13:40
Control Test		0.083	13:41
Air Blank		0.000	13:41
Diagnostics Check		OK	13:41

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JACKSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: YANEZ SOTO

DATE/TIME OF TEST: 02/20/2020 14:06:26

Diagnostics Check		OK	14:00
Air Blank		0.000	14:00
Control Test		0.083	14:01
Air Blank		0.000	14:01
Subject Sample	# 1	0.000	14:02
Breath Volume		3.292	14:02
Air Blank		0.000	14:02
Air Blank		0.000	14:04
Subject Sample	# 2	0.000	14:05
Breath Volume		2.875	14:05
Air Blank		0.000	14:05
Control Test		0.080	14:05
Air Blank		0.000	14:06
Diagnostics Check		OK	14:06

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KALLMEYER,DANIEL,K
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/21/2020 13:31:18

Diagnostics Check		OK	13:25
Air Blank		0.000	13:25
Control Test		0.082	13:26
Air Blank		0.000	13:26
Subject Sample	# 1	0.000	13:26
Breath Volume		1.683	13:26
Air Blank		0.000	13:27
Air Blank		0.000	13:29
Subject Sample	# 2	0.000	13:29
Breath Volume		1.527	13:29
Air Blank		0.000	13:30
Control Test		0.080	13:30
Air Blank		0.000	13:31
Diagnostics Check		OK	13:31

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,RAY,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: WEBER,RAY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/21/2020 16:58:48

Diagnostics Check		OK	16:52
Air Blank		0.000	16:53
Control Test		0.083	16:53
Air Blank		0.000	16:54
Subject Sample	# 1	0.007	16:54
Breath Volume		1.820	16:54
Air Blank		0.000	16:55
Air Blank		0.000	16:57
Subject Sample	# 2	0.000	16:57
Breath Volume		2.457	16:57
Air Blank		0.000	16:57
Control Test		0.082	16:58
Air Blank		0.000	16:58
Diagnostics Check		OK	16:58

Data Download Date/Time: 02/24/2020 11:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HAGER,KEVIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: RAMKISHUN

DATE/TIME OF TEST: 02/28/2020 15:00:35

Diagnostics Check		OK	14:54
Air Blank		0.000	14:55
Control Test		0.083	14:55
Air Blank		0.000	14:56
Subject Sample	# 1	0.000	14:56
Breath Volume		2.070	14:56
Air Blank		0.000	14:56
Air Blank		0.000	14:58
Subject Sample	# 2	0.000	14:59
Breath Volume		1.738	14:59
Air Blank		0.000	14:59
Control Test		0.083	14:59
Air Blank		0.000	15:00
Diagnostics Check		OK	15:00

Data Download Date/Time: 04/01/2020 16:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DWIGHT,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ROGERS,DWIGHT
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ORTIZ'TORRES

DATE/TIME OF TEST: 02/29/2020 00:06:05

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.084	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.028	00:01
Breath Volume		3.882	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.030	00:04
Breath Volume		2.957	00:04
Air Blank		0.000	00:05
Control Test		0.083	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 04/01/2020 16:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CHAD,B
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ARKUSZESKI,DAVID
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 02/29/2020 07:08:38

Diagnostics Check		OK	07:05	
Air Blank		0.000	07:06	
Control Test		0.083	07:06	
Air Blank		0.000	07:07	
Subject Sample	# 1		07:07	Subject Test Refused
Breath Volume		0.000	07:07	
Air Blank		0.000	07:07	
Control Test		0.083	07:08	
Air Blank		0.000	07:08	
Diagnostics Check		OK	07:08	

Data Download Date/Time: 04/01/2020 16:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: RODRIGUEZ ALVARADO

DATE/TIME OF TEST: 02/29/2020 21:22:23

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.083	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.176	21:14
Breath Volume		1.718	21:14
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.148	21:17
Breath Volume		1.933	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 3	0.167	21:20
Breath Volume		2.179	21:20
Air Blank		0.000	21:21
Control Test		0.083	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 04/01/2020 16:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORJAS,NOE,A
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CARROCCIO'

DATE/TIME OF TEST: 02/08/2020 01:54:03

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.186	01:46
Breath Volume		2.863	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.165	01:49
Breath Volume		2.335	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:51
Subject Sample	# 3	0.183	01:52
Breath Volume		2.746	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:54

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 02/08/2020 23:09:33

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.168	23:04
Breath Volume		2.078	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.171	23:07
Breath Volume		1.601	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: DUENAS

DATE/TIME OF TEST: 02/13/2020 19:15:04

Diagnostics Check		OK	19:08
Air Blank		0.000	19:08
Control Test		0.080	19:08
Air Blank		0.000	19:09
Subject Sample	# 1	0.113	19:10
Breath Volume		1.128	19:10
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.128	19:13
Breath Volume		1.187	19:13
Air Blank		0.000	19:14
Control Test		0.080	19:14
Air Blank		0.000	19:14
Diagnostics Check		OK	19:15

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: RUEHMANN

DATE/TIME OF TEST: 02/14/2020 11:10:16

Diagnostics Check		OK	11:03
Air Blank		0.000	11:03
Control Test		0.079	11:03
Air Blank		0.000	11:04
Subject Sample	# 1	0.013	11:06
Breath Volume		3.148	11:06
Air Blank		0.000	11:06
Air Blank		0.000	11:08
Subject Sample	# 2	0.012	11:08
Breath Volume		2.398	11:08
Air Blank		0.000	11:09
Control Test		0.079	11:09
Air Blank		0.000	11:10
Diagnostics Check		OK	11:10

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: DUFFES,DAVID
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/17/2020 00:29:08

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:24
Breath Volume		1.589	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:27
Breath Volume		1.953	00:27
Air Blank		0.000	00:28
Control Test		0.080	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/18/2020 04:25:30

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.000	04:20
Breath Volume		1.308	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.000	04:24
Breath Volume		1.320	04:24
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BERRY,VICTORIA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BULL

DATE/TIME OF TEST: 02/18/2020 04:43:02

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.079	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.054	04:38
Breath Volume		1.265	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.055	04:41
Breath Volume		1.156	04:41
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:43

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 02/22/2020 02:42:56

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.118	02:37
Breath Volume		2.148	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.106	02:41
Breath Volume		1.539	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 02/25/2020 13:44:51

Diagnostics Check		OK	13:38
Air Blank		0.000	13:39
Control Test		0.079	13:39
Air Blank		0.000	13:40
Subject Sample	# 1	0.208	13:40
Breath Volume		1.808	13:40
Air Blank		0.000	13:41
Air Blank		0.000	13:42
Subject Sample	# 2	0.220	13:43
Breath Volume		2.277	13:43
Air Blank		0.000	13:43
Control Test		0.079	13:44
Air Blank		0.000	13:44
Diagnostics Check		OK	13:44

Data Download Date/Time: 02/26/2020 16:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: NOLTE

DATE/TIME OF TEST: 02/26/2020 17:12:01

Diagnostics Check		OK	17:03
Air Blank		0.000	17:03
Control Test		0.079	17:04
Air Blank		0.000	17:04
Subject Sample	# 1	0.168	17:06
Breath Volume		1.437	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:09
Subject Sample	# 2	0.164	17:10
Breath Volume		1.152	17:10
Air Blank		0.000	17:11
Control Test		0.078	17:11
Air Blank		0.000	17:11
Diagnostics Check		OK	17:12

Data Download Date/Time: 03/28/2020 12:19

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 02/01/2020 21:35:50

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:33	
Control Test		0.079	21:33	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:34	Subject Test Refused
Breath Volume		0.000	21:34	
Air Blank		0.000	21:34	
Control Test		0.079	21:35	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 02/12/2020 14:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,ROCHELLE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,JAKE
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/08/2020 02:21:01

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.067	02:16
Breath Volume		3.214	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:18
Subject Sample	# 2	0.068	02:19
Breath Volume		3.312	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/12/2020 14:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCLENDON,R
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: PARKES

DATE/TIME OF TEST: 02/09/2020 20:07:50

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.078	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.124	20:03
Breath Volume		2.785	20:03
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.121	20:06
Breath Volume		2.058	20:06
Air Blank		0.000	20:06
Control Test		0.078	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 02/12/2020 14:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: SCHEIDER

DATE/TIME OF TEST: 02/09/2020 22:43:47

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.078	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.033	22:39
Breath Volume		1.531	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.032	22:42
Breath Volume		2.632	22:42
Air Blank		0.000	22:42
Control Test		0.079	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 02/12/2020 14:26

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECK,EMILY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/08/2020 03:51:49

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.000	03:47
Breath Volume		3.578	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.000	03:50
Breath Volume		3.156	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 02/13/2020 02:05

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/15/2020 01:30:25

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.079	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 03/18/2020 07:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: WINHOVEN

DATE/TIME OF TEST: 02/23/2020 03:36:06

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.173	03:31
Breath Volume		2.449	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.172	03:34
Breath Volume		2.359	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 03/18/2020 07:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ACEVEDO PERALTA,EMMANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/23/2020 04:17:48

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:14	
Control Test		0.080	04:15	
Air Blank		0.000	04:15	
Subject Sample	# 1		04:16	Subject Test Refused
Breath Volume		0.000	04:16	
Air Blank		0.000	04:16	
Control Test		0.079	04:17	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 03/18/2020 07:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ARELLANO SANCHEZ

DATE/TIME OF TEST: 02/29/2020 02:00:44

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.196	01:55
Breath Volume		2.773	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.177	01:59
Breath Volume		1.152	01:59
Air Blank		0.000	01:59
Control Test		0.080	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 03/18/2020 07:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: FUTCH

DATE/TIME OF TEST: 02/29/2020 02:24:12

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.133	02:19
Breath Volume		1.281	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.133	02:22
Breath Volume		1.421	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/18/2020 07:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,CHAZ,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 02/29/2020 05:36:07

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.149	05:31
Breath Volume		4.601	05:31
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.153	05:34
Breath Volume		2.523	05:34
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 03/18/2020 07:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/13/2020 23:34:10

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:31	
Control Test		0.079	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.080	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 02/27/2020 23:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 02/17/2020 19:19:54

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:17	
Control Test		0.079	19:17	
Air Blank		0.000	19:18	
Subject Sample	# 1		19:18	Subject Test Refused
Breath Volume		0.000	19:18	
Air Blank		0.000	19:19	
Control Test		0.080	19:19	
Air Blank		0.000	19:19	
Diagnostics Check		OK	19:19	

Data Download Date/Time: 02/27/2020 23:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: COCONUT CREEK PD
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: COCONUT CREEK PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094981
DRY GAS EXPIRY DATE: 02/22/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/27/2020 22:50:37

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.080	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.000	22:44
Breath Volume		1.109	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.000	22:49
Breath Volume		1.281	22:49
Air Blank		0.000	22:49
Control Test		0.079	22:50
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 02/27/2020 23:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: EGER

DATE/TIME OF TEST: 02/01/2020 21:33:09

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:30	
Control Test		0.079	21:30	
Air Blank		0.000	21:31	
Subject Sample	# 1		21:31	Subject Test Refused
Breath Volume		0.000	21:31	
Air Blank		0.000	21:32	
Control Test		0.078	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 02/13/2020 09:47

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FOREMAN

DATE/TIME OF TEST: 02/05/2020 02:18:38

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.190	02:13
Breath Volume		2.554	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.198	02:16
Breath Volume		2.921	02:16
Air Blank		0.000	02:17
Control Test		0.078	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 02/13/2020 09:47

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: YOUNG,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 02/05/2020 05:19:40

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.078	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.168	05:14
Breath Volume		3.003	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.174	05:18
Breath Volume		3.433	05:18
Air Blank		0.000	05:18
Control Test		0.078	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 02/13/2020 09:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHOWDHURY,P
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: CHEREVKO

DATE/TIME OF TEST: 02/06/2020 04:06:18

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.000	04:02
Breath Volume		3.757	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:04
Breath Volume		2.964	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 02/13/2020 09:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/08/2020 04:09:07

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.106	04:04
Breath Volume		2.074	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.109	04:07
Breath Volume		2.488	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/13/2020 09:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TURPIN,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 02/08/2020 18:37:21

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.078	18:31
Air Blank		0.000	18:31
Subject Sample	# 1	0.175	18:32
Breath Volume		3.812	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:35
Subject Sample	# 2	0.164	18:35
Breath Volume		2.652	18:35
Air Blank		0.000	18:36
Control Test		0.077	18:36
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 02/13/2020 09:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 02/10/2020 22:02:59

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.139	21:58
Breath Volume		1.406	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.136	22:01
Breath Volume		1.765	22:01
Air Blank		0.000	22:02
Control Test		0.077	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 02/13/2020 09:48

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHERENA,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PETERSEN

DATE/TIME OF TEST: 02/15/2020 03:53:15

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:45	
Control Test		0.079	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 1		03:49	No Sample Provided
Breath Volume		0.000	03:49	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/16/2020 04:36:36

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.078	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:35	
Control Test		0.078	04:36	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NOEORE,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: PAYTON

DATE/TIME OF TEST: 02/16/2020 07:36:34

Diagnostics Check		OK	07:30
Air Blank		0.000	07:30
Control Test		0.079	07:30
Air Blank		0.000	07:31
Subject Sample	# 1	0.100	07:32
Breath Volume		2.679	07:32
Air Blank		0.000	07:32
Air Blank		0.000	07:34
Subject Sample	# 2	0.097	07:34
Breath Volume		2.460	07:34
Air Blank		0.000	07:35
Control Test		0.079	07:35
Air Blank		0.000	07:36
Diagnostics Check		OK	07:36

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: VIETZE

DATE/TIME OF TEST: 02/17/2020 20:17:12

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.079	20:15	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:15	Subject Test Refused
Breath Volume		0.000	20:15	
Air Blank		0.000	20:16	
Control Test		0.078	20:16	
Air Blank		0.000	20:17	
Diagnostics Check		OK	20:17	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 02/17/2020 22:46:50

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:44	
Control Test		0.079	22:44	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Control Test		0.078	22:46	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRIFFIN,C
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 02/18/2020 04:44:29

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:42	
Control Test		0.079	04:42	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:43	
Control Test		0.078	04:43	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 02/19/2020 02:56:18

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:55	
Control Test		0.078	02:55	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: TRAYLOR

DATE/TIME OF TEST: 02/19/2020 17:07:40

Diagnostics Check		OK	17:01
Air Blank		0.000	17:01
Control Test		0.078	17:01
Air Blank		0.000	17:02
Subject Sample	# 1	0.000	17:03
Breath Volume		1.531	17:03
Air Blank		0.000	17:03
Air Blank		0.000	17:05
Subject Sample	# 2	0.000	17:06
Breath Volume		1.664	17:06
Air Blank		0.000	17:06
Control Test		0.079	17:07
Air Blank		0.000	17:07
Diagnostics Check		OK	17:07

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: YUKNA

DATE/TIME OF TEST: 02/22/2020 01:09:04

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.238	01:04
Breath Volume		2.527	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:06
Subject Sample	# 2	0.241	01:07
Breath Volume		2.097	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:09

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 02/22/2020 02:40:05

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.000	02:35
Breath Volume		2.769	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.000	02:38
Breath Volume		1.320	02:38
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/23/2020 02:25:01

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.296	02:20
Breath Volume		1.890	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.311	02:23
Breath Volume		2.652	02:23
Air Blank		0.000	02:24
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:25

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NASH,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/23/2020 06:26:41

Diagnostics Check		OK	06:23	
Air Blank		0.000	06:24	
Control Test		0.078	06:24	
Air Blank		0.000	06:25	
Subject Sample	# 1		06:25	Subject Test Refused
Breath Volume		0.000	06:25	
Air Blank		0.000	06:25	
Control Test		0.078	06:26	
Air Blank		0.000	06:26	
Diagnostics Check		OK	06:26	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/23/2020 22:56:14

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.079	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.211	22:51
Breath Volume		2.363	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.194	22:54
Breath Volume		2.019	22:54
Air Blank		0.000	22:55
Control Test		0.078	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 02/24/2020 05:45:01

Diagnostics Check		OK	05:39
Air Blank		0.000	05:39
Control Test		0.078	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.000	05:40
Breath Volume		2.339	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.000	05:43
Breath Volume		2.054	05:43
Air Blank		0.000	05:44
Control Test		0.078	05:44
Air Blank		0.000	05:44
Diagnostics Check		OK	05:45

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MOHISSEN

DATE/TIME OF TEST: 02/25/2020 04:31:15

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.079	04:29	
Air Blank		0.000	04:29	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:30	
Control Test		0.078	04:30	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: HARTER

DATE/TIME OF TEST: 02/27/2020 20:47:51

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.079	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.066	20:43
Breath Volume		1.574	20:43
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.067	20:46
Breath Volume		1.324	20:46
Air Blank		0.000	20:46
Control Test		0.079	20:47
Air Blank		0.000	20:47
Diagnostics Check		OK	20:47

Data Download Date/Time: 02/28/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: NEEDHAM

DATE/TIME OF TEST: 02/29/2020 01:39:00

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.229	01:34
Breath Volume		1.367	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.222	01:37
Breath Volume		1.191	01:37
Air Blank		0.000	01:38
Control Test		0.077	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/11/2020 12:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SANDIFORD,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: DAWKINS-THOMAS

DATE/TIME OF TEST: 02/29/2020 04:45:00

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.110	04:40
Breath Volume		1.417	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.109	04:43
Breath Volume		1.281	04:43
Air Blank		0.000	04:44
Control Test		0.078	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:45

Data Download Date/Time: 03/11/2020 12:31

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 02/01/2020 01:45:12

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:34	
Control Test		0.080	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:38	No Sample Provided
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 2		01:43	No Sample Provided
Breath Volume		0.000	01:43	
Air Blank		0.000	01:44	
Control Test		0.080	01:44	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 02/02/2020 04:43:49

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.189	04:39
Breath Volume		2.601	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.193	04:42
Breath Volume		2.136	04:42
Air Blank		0.000	04:42
Control Test		0.080	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: LUONG

DATE/TIME OF TEST: 02/03/2020 01:46:49

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.080	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:42	No Sample Provided
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Control Test		0.080	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TOLLICK,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LAMKIN

DATE/TIME OF TEST: 02/03/2020 02:00:14

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.246	01:54
Breath Volume		1.871	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.237	01:58
Breath Volume		2.433	01:58
Air Blank		0.000	01:59
Control Test		0.080	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CAHILL

DATE/TIME OF TEST: 02/07/2020 03:32:52

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.224	03:28
Breath Volume		2.214	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.239	03:31
Breath Volume		2.433	03:31
Air Blank		0.000	03:31
Control Test		0.081	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY,MARILYN,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 02/10/2020 03:05:15

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.169	03:00
Breath Volume		3.238	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.177	03:03
Breath Volume		3.074	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/11/2020 21:34:49

Diagnostics Check		OK	21:26
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:27
Subject Sample	# 1	0.142	21:29
Breath Volume		1.535	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.143	21:33
Breath Volume		1.378	21:33
Air Blank		0.000	21:33
Control Test		0.080	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY,M,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: VILORIA

DATE/TIME OF TEST: 02/13/2020 22:25:51

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.080	22:17	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:21	No Sample Provided
Breath Volume		0.000	22:21	
Air Blank		0.000	22:22	
Air Blank		0.000	22:24	
Subject Sample	# 2		22:24	Subject Test Refused
Breath Volume		0.000	22:24	
Air Blank		0.000	22:24	
Control Test		0.080	22:25	
Air Blank		0.000	22:25	
Diagnostics Check		OK	22:25	

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 02/14/2020 18:35:23

Diagnostics Check		OK	18:21	
Air Blank		0.000	18:22	
Control Test		0.080	18:22	
Air Blank		0.000	18:22	
Subject Sample	# 1		18:26	No Sample Provided
Breath Volume		0.000	18:26	
Air Blank		0.000	18:26	
Air Blank		0.000	18:28	
Subject Sample	# 2	0.159	18:29	
Breath Volume		2.007	18:29	
Air Blank		0.000	18:29	
Air Blank		0.000	18:31	
Subject Sample	# 3	0.147	18:33	
Breath Volume		1.578	18:33	
Air Blank		0.000	18:34	
Control Test		0.080	18:34	
Air Blank		0.000	18:35	
Diagnostics Check		OK	18:35	

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: SCHAUB

DATE/TIME OF TEST: 02/14/2020 21:38:13

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.081	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.247	21:33
Breath Volume		2.324	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:35
Subject Sample	# 2	0.261	21:36
Breath Volume		2.292	21:36
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PERALTA-REYES

DATE/TIME OF TEST: 02/15/2020 04:17:34

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.125	04:13
Breath Volume		3.175	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.116	04:16
Breath Volume		2.871	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 02/16/2020 02:18:58

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.031	02:09	Volume Not Met
Breath Volume		0.273	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.034	02:14	
Breath Volume		3.207	02:14	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 3	0.030	02:17	
Breath Volume		1.734	02:17	
Air Blank		0.000	02:18	
Control Test		0.081	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WHISENANT

DATE/TIME OF TEST: 02/17/2020 00:05:05

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:54	
Control Test		0.080	23:54	
Air Blank		0.000	23:54	
Subject Sample	# 2		23:58	No Sample Provided
Breath Volume		0.000	23:58	
Air Blank		0.000	23:58	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:03	No Sample Provided
Breath Volume		0.000	00:03	
Air Blank		0.000	00:04	
Control Test		0.080	00:04	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY,M,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 02/18/2020 02:41:20

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:29	
Subject Sample	# 1	0.156	02:31	
Breath Volume		1.332	02:31	
Air Blank		0.000	02:31	
Air Blank		0.000	02:33	
Subject Sample	# 2	0.158	02:36	Volume Not Met
Breath Volume		0.066	02:36	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 3	0.160	02:39	
Breath Volume		1.449	02:39	
Air Blank		0.000	02:40	
Control Test		0.079	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/21/2020 02:40:02

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.190	02:34
Breath Volume		3.730	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:36
Subject Sample	# 2	0.192	02:38
Breath Volume		3.250	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLOW,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,TRAVIS
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: LANGONE

DATE/TIME OF TEST: 02/22/2020 03:29:45

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.081	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.187	03:25
Breath Volume		2.972	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:27
Subject Sample	# 2	0.188	03:28
Breath Volume		2.816	03:28
Air Blank		0.000	03:28
Control Test		0.081	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 02/28/2020 00:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BARNHILL

DATE/TIME OF TEST: 02/29/2020 03:23:11

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.198	03:17
Breath Volume		2.722	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.188	03:21
Breath Volume		1.753	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 03/21/2020 15:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTZ,JARROD,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SWIM,G
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/29/2020 04:05:56

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.296	04:01
Breath Volume		2.343	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.291	04:04
Breath Volume		2.621	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 03/21/2020 15:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANELLA,E,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LUNSFORD,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: GIORDANO

DATE/TIME OF TEST: 02/29/2020 16:30:45

Diagnostics Check		OK	16:24
Air Blank		0.000	16:25
Control Test		0.081	16:25
Air Blank		0.000	16:25
Subject Sample	# 1	0.000	16:26
Breath Volume		1.320	16:26
Air Blank		0.000	16:26
Air Blank		0.000	16:28
Subject Sample	# 2	0.000	16:29
Breath Volume		1.679	16:29
Air Blank		0.000	16:29
Control Test		0.080	16:30
Air Blank		0.000	16:30
Diagnostics Check		OK	16:30

Data Download Date/Time: 03/21/2020 15:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,ALEX,J
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: THOMAS,ALEX
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: ALONGE

DATE/TIME OF TEST: 02/07/2020 00:49:29

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.079	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Control Test		0.079	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 02/28/2020 09:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,RENE,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BEST,R
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/07/2020 16:36:30

Diagnostics Check		OK	16:30
Air Blank		0.000	16:30
Control Test		0.079	16:31
Air Blank		0.000	16:31
Subject Sample	# 1	0.217	16:32
Breath Volume		1.628	16:32
Air Blank		0.000	16:32
Air Blank		0.000	16:34
Subject Sample	# 2	0.217	16:34
Breath Volume		1.207	16:34
Air Blank		0.000	16:35
Control Test		0.078	16:35
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 02/28/2020 09:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ERIC,D
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: JOSEPH,DAVID
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: SWASEY

DATE/TIME OF TEST: 02/13/2020 07:17:54

Diagnostics Check		OK	07:11
Air Blank		0.000	07:12
Control Test		0.079	07:12
Air Blank		0.000	07:12
Subject Sample	# 1	0.100	07:13
Breath Volume		1.101	07:13
Air Blank		0.000	07:14
Air Blank		0.000	07:15
Subject Sample	# 2	0.104	07:16
Breath Volume		1.671	07:16
Air Blank		0.000	07:16
Control Test		0.079	07:17
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

Data Download Date/Time: 02/28/2020 09:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ARDILA,MONICA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 035119080A
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: JARIDO

DATE/TIME OF TEST: 02/22/2020 01:05:20

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.151	01:00
Breath Volume		1.949	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.150	01:03
Breath Volume		2.589	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 02/28/2020 09:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 02/01/2020 04:01:12

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.097	03:56
Breath Volume		1.425	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.100	03:59
Breath Volume		1.300	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 02/01/2020 04:39:03

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.080	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.139	04:34
Breath Volume		1.406	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.132	04:37
Breath Volume		1.316	04:37
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:39

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 02/02/2020 02:42:25

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.081	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.173	02:36
Breath Volume		1.140	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.164	02:40
Breath Volume		2.000	02:40
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: VAZQUEZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: KINSER

DATE/TIME OF TEST: 02/02/2020 03:14:27

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.080	03:08
Breath Volume		1.484	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.081	03:13
Breath Volume		1.699	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 02/02/2020 04:15:42

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1	0.147	04:02	
Breath Volume		3.074	04:02	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2	0.131	04:08	Volume Not Met
Breath Volume		0.394	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:10	
Subject Sample	# 3	0.140	04:14	Volume Not Met
Breath Volume		0.914	04:14	
Air Blank		0.000	04:14	
Control Test		0.079	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 02/03/2020 00:01:46

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:59	
Control Test		0.081	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:00	Subject Test Refused
Breath Volume		0.000	00:00	
Air Blank		0.000	00:00	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,M,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 02/04/2020 01:03:18

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.085	00:58
Breath Volume		2.128	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.086	01:01
Breath Volume		2.039	01:01
Air Blank		0.000	01:02
Control Test		0.082	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,J,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/05/2020 22:54:04

Diagnostics Check		OK	22:47
Air Blank		0.000	22:48
Control Test		0.082	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.133	22:49
Breath Volume		1.953	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.136	22:52
Breath Volume		2.113	22:52
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:54

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ROYSTER

DATE/TIME OF TEST: 02/06/2020 02:02:04

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.190	01:57
Breath Volume		1.296	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.191	02:00
Breath Volume		1.585	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:02

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,M,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUFFMON,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: DELOSSANTOS

DATE/TIME OF TEST: 02/08/2020 04:05:26

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.082	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.163	04:00
Breath Volume		2.199	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.169	04:03
Breath Volume		1.933	04:03
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,DEANNE,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 02/08/2020 22:25:29

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.133	22:20
Breath Volume		1.703	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.133	22:24
Breath Volume		1.160	22:24
Air Blank		0.000	22:24
Control Test		0.080	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,W,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 02/09/2020 03:16:33

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.081	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.082	03:15	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGWELL,DESHONNA,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DAVIS,N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 02/09/2020 08:07:24

Diagnostics Check		OK	08:04	
Air Blank		0.000	08:04	
Control Test		0.081	08:05	
Air Blank		0.000	08:05	
Subject Sample	# 1		08:06	Subject Test Refused
Breath Volume		0.000	08:06	
Air Blank		0.000	08:06	
Control Test		0.081	08:06	
Air Blank		0.000	08:07	
Diagnostics Check		OK	08:07	

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGWELL,DESHONNA,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DAVIS,N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 02/09/2020 08:34:27

Diagnostics Check		OK	08:28	
Air Blank		0.000	08:28	
Control Test		0.080	08:28	
Air Blank		0.000	08:29	
Subject Sample	# 1	0.108	08:30	
Breath Volume		2.488	08:30	
Air Blank		0.000	08:30	
Air Blank		0.000	08:32	
Subject Sample	# 2	0.106	08:32	
Breath Volume		3.199	08:32	
Air Blank		0.000	08:33	
Control Test		0.071	08:33	Control Outside Tolerance
Air Blank		0.000	08:34	

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGWELL,DESHONNA,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DAVIS,N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 02/09/2020 09:06:29

Diagnostics Check		OK	09:00
Air Blank		0.000	09:00
Control Test		0.081	09:00
Air Blank		0.000	09:01
Subject Sample	# 1	0.105	09:01
Breath Volume		3.609	09:01
Air Blank		0.000	09:02
Air Blank		0.000	09:04
Subject Sample	# 2	0.105	09:04
Breath Volume		4.007	09:04
Air Blank		0.000	09:05
Control Test		0.080	09:05
Air Blank		0.000	09:06
Diagnostics Check		OK	09:06

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,M,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SANDIFER,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BLOUNT

DATE/TIME OF TEST: 02/09/2020 20:06:06

Diagnostics Check		OK	20:03	
Air Blank		0.000	20:03	
Control Test		0.081	20:04	
Air Blank		0.000	20:04	
Subject Sample	# 1		20:04	Subject Test Refused
Breath Volume		0.000	20:04	
Air Blank		0.000	20:05	
Control Test		0.082	20:05	
Air Blank		0.000	20:06	
Diagnostics Check		OK	20:06	

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ROJAS,H
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 02/11/2020 09:03:21

Diagnostics Check		OK	08:57
Air Blank		0.000	08:57
Control Test		0.081	08:58
Air Blank		0.000	08:58
Subject Sample	# 1	0.000	08:59
Breath Volume		3.281	08:59
Air Blank		0.000	08:59
Air Blank		0.000	09:01
Subject Sample	# 2	0.000	09:02
Breath Volume		1.953	09:02
Air Blank		0.000	09:02
Control Test		0.081	09:02
Air Blank		0.000	09:03
Diagnostics Check		OK	09:03

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,W,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 02/12/2020 21:02:38

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.081	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.000	20:58
Breath Volume		4.101	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.000	21:01
Breath Volume		3.523	21:01
Air Blank		0.000	21:01
Control Test		0.080	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,M,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CONE,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RENE VILLIA

DATE/TIME OF TEST: 02/12/2020 22:08:26

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.080	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1		22:07	Subject Test Refused
Breath Volume		0.000	22:07	
Air Blank		0.000	22:07	
Control Test		0.081	22:07	
Air Blank		0.000	22:08	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MARREL,EMILY
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/13/2020 12:04:24

Diagnostics Check		OK	11:57
Air Blank		0.000	11:58
Control Test		0.082	11:58
Air Blank		0.000	11:58
Subject Sample	# 1	0.000	11:59
Breath Volume		1.652	11:59
Air Blank		0.000	11:59
Air Blank		0.000	12:01
Subject Sample	# 2	0.000	12:03
Breath Volume		1.605	12:03
Air Blank		0.000	12:03
Control Test		0.081	12:03
Air Blank		0.000	12:04
Diagnostics Check		OK	12:04

Data Download Date/Time: 02/24/2020 12:51

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: QUICK,WILLFRED
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:23

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 02/13/2020 13:06:52

Diagnostics Check		OK	12:54	
Air Blank		0.000	12:55	
Control Test		0.081	12:55	
Air Blank		0.000	12:56	
Subject Sample	# 1	0.011	12:59	Volume Not Met
Breath Volume		0.796	12:59	
Air Blank		0.000	12:59	
Air Blank		0.000	13:01	
Subject Sample	# 2	0.011	13:02	
Breath Volume		1.460	13:02	
Air Blank		0.000	13:02	
Air Blank		0.000	13:04	
Subject Sample	# 3	0.010	13:05	
Breath Volume		1.148	13:05	
Air Blank		0.000	13:05	
Control Test		0.080	13:06	
Air Blank		0.000	13:06	
Diagnostics Check		OK	13:06	

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,E,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: SUITS

DATE/TIME OF TEST: 02/14/2020 18:20:53

Diagnostics Check		OK	18:13
Air Blank		0.000	18:13
Control Test		0.081	18:14
Air Blank		0.000	18:14
Subject Sample	# 1	0.151	18:15
Breath Volume		1.621	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.136	18:19
Breath Volume		1.453	18:19
Air Blank		0.000	18:19
Control Test		0.081	18:20
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DOSS

DATE/TIME OF TEST: 02/14/2020 23:23:56

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.081	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.000	23:19
Breath Volume		1.460	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.000	23:22
Breath Volume		1.328	23:22
Air Blank		0.000	23:23
Control Test		0.081	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 02/15/2020 03:47:46

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.081	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.000	03:40	Volume Not Met
Breath Volume		0.652	03:40	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 2	0.000	03:46	Volume Not Met
Breath Volume		0.765	03:46	
Air Blank		0.000	03:46	
Control Test		0.081	03:47	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: BADAL

DATE/TIME OF TEST: 02/15/2020 04:48:04

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:45	
Control Test		0.081	04:46	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:46	Subject Test Refused
Breath Volume		0.000	04:46	
Air Blank		0.000	04:47	
Control Test		0.080	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ENNIS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 02/15/2020 05:22:19

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.080	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1	0.139	05:15	Volume Not Met
Breath Volume		0.101	05:15	
Air Blank		0.000	05:15	
Air Blank		0.000	05:17	
Subject Sample	# 2	0.118	05:20	Volume Not Met
Breath Volume		0.101	05:20	
Air Blank		0.000	05:21	
Control Test		0.080	05:21	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: VAZQUEZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 02/15/2020 06:13:19

Diagnostics Check		OK	06:09	
Air Blank		0.000	06:10	
Control Test		0.081	06:10	
Air Blank		0.000	06:11	
Subject Sample	# 1		06:11	Subject Test Refused
Breath Volume		0.000	06:11	
Air Blank		0.000	06:12	
Control Test		0.081	06:12	
Air Blank		0.000	06:13	
Diagnostics Check		OK	06:13	

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 02/15/2020 23:33:42

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.116	23:28
Breath Volume		1.632	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.119	23:32
Breath Volume		1.414	23:32
Air Blank		0.000	23:32
Control Test		0.080	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BIONDI

DATE/TIME OF TEST: 02/16/2020 00:26:13

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.150	00:21
Breath Volume		2.566	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.158	00:24
Breath Volume		2.246	00:24
Air Blank		0.000	00:25
Control Test		0.081	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,G,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GRAY,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:47

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 02/16/2020 08:14:07

Diagnostics Check		OK	08:08
Air Blank		0.000	08:08
Control Test		0.080	08:08
Air Blank		0.000	08:09
Subject Sample	# 1	0.235	08:09
Breath Volume		2.644	08:09
Air Blank		0.000	08:10
Air Blank		0.000	08:12
Subject Sample	# 2	0.226	08:12
Breath Volume		2.437	08:12
Air Blank		0.000	08:13
Control Test		0.079	08:13
Air Blank		0.000	08:14
Diagnostics Check		OK	08:14

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGWELL,DESHONNA,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KNIGHT,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: HEER

DATE/TIME OF TEST: 02/17/2020 16:25:06

Diagnostics Check		OK	16:19
Air Blank		0.000	16:19
Control Test		0.081	16:19
Air Blank		0.000	16:20
Subject Sample	# 1	0.000	16:20
Breath Volume		2.050	16:20
Air Blank		0.000	16:21
Air Blank		0.000	16:23
Subject Sample	# 2	0.000	16:23
Breath Volume		2.027	16:23
Air Blank		0.000	16:24
Control Test		0.081	16:24
Air Blank		0.000	16:25
Diagnostics Check		OK	16:25

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,CATHERINE,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ROCHA

DATE/TIME OF TEST: 02/19/2020 05:05:45

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.081	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.000	05:01
Breath Volume		2.562	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.000	05:04
Breath Volume		2.175	05:04
Air Blank		0.000	05:04
Control Test		0.081	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,JD,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 02/19/2020 20:26:40

Diagnostics Check		OK	20:20
Air Blank		0.000	20:21
Control Test		0.082	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.000	20:22
Breath Volume		1.250	20:22
Air Blank		0.000	20:22
Air Blank		0.000	20:24
Subject Sample	# 2	0.000	20:25
Breath Volume		1.304	20:25
Air Blank		0.000	20:25
Control Test		0.081	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERON,W,G
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 02/22/2020 03:02:29

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.219	02:57
Breath Volume		2.847	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.219	03:01
Breath Volume		2.332	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENSA,M,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 01/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LUCE

DATE/TIME OF TEST: 02/23/2020 03:51:17

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.083	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.082	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 02/24/2020 12:52

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: BOLGER

DATE/TIME OF TEST: 02/25/2020 20:25:39

Diagnostics Check		OK	20:19
Air Blank		0.000	20:20
Control Test		0.081	20:20
Air Blank		0.000	20:20
Subject Sample	# 1	0.000	20:21
Breath Volume		1.453	20:21
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.000	20:24
Breath Volume		1.824	20:24
Air Blank		0.000	20:24
Control Test		0.081	20:25
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 03/21/2020 13:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUTZ,T,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SIEH,Z
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MCKIE

DATE/TIME OF TEST: 02/29/2020 02:08:51

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.165	02:04
Breath Volume		3.617	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.155	02:07
Breath Volume		3.910	02:07
Air Blank		0.000	02:07
Control Test		0.081	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 03/21/2020 13:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WISHART,D
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 02/02/2020 03:00:38

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.081	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.266	02:56
Breath Volume		3.097	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.254	02:59
Breath Volume		3.218	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/13/2020 02:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECERRA,JOSE
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: VASQUES

DATE/TIME OF TEST: 02/02/2020 18:38:36

Diagnostics Check		OK	18:31
Air Blank		0.000	18:31
Control Test		0.082	18:32
Air Blank		0.000	18:32
Subject Sample	# 1	0.141	18:34
Breath Volume		2.273	18:34
Air Blank		0.000	18:34
Air Blank		0.000	18:36
Subject Sample	# 2	0.138	18:37
Breath Volume		1.726	18:37
Air Blank		0.000	18:37
Control Test		0.081	18:38
Air Blank		0.000	18:38
Diagnostics Check		OK	18:38

Data Download Date/Time: 02/13/2020 02:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: ZERO TOL
ARREST TIME: 10:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: MOYES

DATE/TIME OF TEST: 02/03/2020 10:42:18

Diagnostics Check		OK	10:35
Air Blank		0.000	10:36
Control Test		0.081	10:36
Air Blank		0.000	10:37
Subject Sample	# 1	0.216	10:37
Breath Volume		2.140	10:37
Air Blank		0.000	10:38
Air Blank		0.000	10:40
Subject Sample	# 2	0.217	10:40
Breath Volume		2.179	10:40
Air Blank		0.000	10:41
Control Test		0.081	10:41
Air Blank		0.000	10:42
Diagnostics Check		OK	10:42

Data Download Date/Time: 02/13/2020 02:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GODINEZ,JORGE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/06/2020 18:16:33

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.081	18:10
Air Blank		0.000	18:11
Subject Sample	# 1	0.171	18:12
Breath Volume		1.968	18:12
Air Blank		0.000	18:12
Air Blank		0.000	18:14
Subject Sample	# 2	0.178	18:15
Breath Volume		1.882	18:15
Air Blank		0.000	18:15
Control Test		0.080	18:15
Air Blank		0.000	18:16
Diagnostics Check		OK	18:16

Data Download Date/Time: 02/13/2020 02:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HOLMES,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: COURT ORD
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: HATFIELD

DATE/TIME OF TEST: 02/07/2020 15:41:42

Diagnostics Check		OK	15:35
Air Blank		0.000	15:36
Control Test		0.080	15:36
Air Blank		0.000	15:36
Subject Sample	# 1	0.000	15:37
Breath Volume		1.328	15:37
Air Blank		0.000	15:37
Air Blank		0.000	15:39
Subject Sample	# 2	0.000	15:40
Breath Volume		1.347	15:40
Air Blank		0.000	15:40
Control Test		0.080	15:41
Air Blank		0.000	15:41
Diagnostics Check		OK	15:41

Data Download Date/Time: 02/13/2020 02:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GARRISON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 02/09/2020 04:18:52

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:16	
Control Test		0.081	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.081	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 02/13/2020 02:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECK,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 02/10/2020 03:27:30

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.081	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1	0.189	03:22	
Breath Volume		1.242	03:22	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Control Test		0.080	03:26	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 02/13/2020 02:59

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/13/2020 20:29:33

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.080	20:23
Air Blank		0.000	20:23
Subject Sample	# 1	0.177	20:24
Breath Volume		3.500	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.169	20:28
Breath Volume		2.917	20:28
Air Blank		0.000	20:28
Control Test		0.080	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 02/25/2020 09:47

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: COLAS

DATE/TIME OF TEST: 02/01/2020 00:04:12

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:52	
Control Test		0.081	23:52	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.172	23:56	Volume Not Met
Breath Volume		0.109	23:56	
Air Blank		0.000	23:57	
Air Blank		0.000	23:58	
Subject Sample	# 3	0.187	23:59	
Breath Volume		2.035	23:59	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.176	00:02	
Breath Volume		1.996	00:02	
Air Blank		0.000	00:03	
Control Test		0.078	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PICHE,RUSSELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/01/2020 00:17:49

Diagnostics Check		OK	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.042	00:13
Breath Volume		1.539	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.041	00:16
Breath Volume		1.144	00:16
Air Blank		0.000	00:16
Control Test		0.080	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/01/2020 00:39:44

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.211	00:35
Breath Volume		2.082	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.209	00:38
Breath Volume		2.691	00:38
Air Blank		0.000	00:38
Control Test		0.079	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 02/01/2020 02:33:20

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.099	02:28
Breath Volume		1.710	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.098	02:31
Breath Volume		2.097	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: ALIFF

DATE/TIME OF TEST: 02/01/2020 20:22:57

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:12	
Control Test		0.080	20:12	
Air Blank		0.000	20:13	
Subject Sample	# 1		20:16	No Sample Provided
Breath Volume		0.000	20:16	
Air Blank		0.000	20:16	
Air Blank		0.000	20:18	
Subject Sample	# 2		20:21	Subject Test Refused
Breath Volume		0.000	20:21	
Air Blank		0.000	20:22	
Control Test		0.080	20:22	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:22	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JACKSON,RONALD
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 02/02/2020 01:38:45

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.078	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.212	01:33
Breath Volume		1.859	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.201	01:37
Breath Volume		1.656	01:37
Air Blank		0.000	01:37
Control Test		0.077	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 02/02/2020 01:57:57

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.000	01:53
Breath Volume		1.132	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.000	01:56
Breath Volume		1.550	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALFRED,BRANDON,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALFRED,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/02/2020 03:08:39

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Control Test		0.080	03:08	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/02/2020 03:26:02

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.094	03:21
Breath Volume		1.875	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.089	03:24
Breath Volume		1.753	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:26

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRON,DAVID,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DANIEL,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: JEAN FRANCOIS

DATE/TIME OF TEST: 02/02/2020 08:35:46

Diagnostics Check		OK	08:28
Air Blank		0.000	08:29
Control Test		0.080	08:29
Air Blank		0.000	08:30
Subject Sample	# 1	0.234	08:31
Breath Volume		1.152	08:31
Air Blank		0.000	08:31
Air Blank		0.000	08:33
Subject Sample	# 2	0.240	08:34
Breath Volume		1.753	08:34
Air Blank		0.000	08:34
Control Test		0.079	08:35
Air Blank		0.000	08:35
Diagnostics Check		OK	08:35

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: PAWLOWSKI

DATE/TIME OF TEST: 02/03/2020 00:56:28

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.298	00:51
Breath Volume		3.195	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.292	00:54
Breath Volume		3.003	00:54
Air Blank		0.000	00:55
Control Test		0.076	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 02/03/2020 02:16:42

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.079	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:12	No Sample Provided
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 02/03/2020 23:00:22

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.081	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.116	22:55
Breath Volume		2.417	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.119	22:58
Breath Volume		2.347	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/05/2020 01:22:34

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.081	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.062	01:17
Breath Volume		1.835	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.064	01:21
Breath Volume		1.875	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 02/06/2020 01:47:04

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.197	01:42
Breath Volume		1.601	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.199	01:45
Breath Volume		3.472	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:47

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PICHE,RUSSELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: NERIO JR

DATE/TIME OF TEST: 02/07/2020 00:28:32

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.081	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:24
Breath Volume		4.148	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:27
Breath Volume		4.257	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MEYERS,ROCKY
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SORISA

DATE/TIME OF TEST: 02/08/2020 04:04:54

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.198	03:59
Breath Volume		2.164	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.183	04:03
Breath Volume		1.894	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MEYERS,ROCKY
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 02/08/2020 22:30:34

Diagnostics Check		OK	22:27	
Air Blank		0.000	22:28	
Control Test		0.080	22:28	
Air Blank		0.000	22:29	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.080	22:29	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: GULDEN

DATE/TIME OF TEST: 02/08/2020 23:22:38

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:08	
Control Test		0.080	23:09	
Air Blank		0.000	23:09	
Subject Sample	# 1	0.066	23:12	Volume Not Met
Breath Volume		0.160	23:12	
Air Blank		0.000	23:13	
Air Blank		0.000	23:15	
Subject Sample	# 2	0.228	23:18	
Breath Volume		1.832	23:18	
Air Blank		0.000	23:18	
Air Blank		0.000	23:20	
Subject Sample	# 3	0.228	23:21	
Breath Volume		1.195	23:21	
Air Blank		0.000	23:21	
Control Test		0.078	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLRED,COBY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GARCIA-ZAPATA

DATE/TIME OF TEST: 02/09/2020 00:30:16

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.179	00:24
Breath Volume		1.238	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.178	00:28
Breath Volume		2.089	00:28
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BUSTOS-RUIZ

DATE/TIME OF TEST: 02/09/2020 03:52:00

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.080	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:50	Subject Test Refused
Breath Volume		0.000	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BUSTOS-RUIZ

DATE/TIME OF TEST: 02/09/2020 04:23:17

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.081	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.099	04:16	Volume Not Met
Breath Volume		0.484	04:16	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2	0.108	04:21	Volume Not Met
Breath Volume		0.941	04:21	
Air Blank		0.000	04:22	
Control Test		0.079	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 02/10/2020 20:19:53

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.081	20:14
Air Blank		0.000	20:14
Subject Sample	# 1	0.117	20:15
Breath Volume		2.964	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:17
Subject Sample	# 2	0.115	20:18
Breath Volume		2.250	20:18
Air Blank		0.000	20:18
Control Test		0.079	20:19
Air Blank		0.000	20:19
Diagnostics Check		OK	20:19

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MANLEY

DATE/TIME OF TEST: 02/11/2020 04:01:10

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:58	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: JOHNSON-ALBRITTON

DATE/TIME OF TEST: 02/11/2020 23:32:07

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.000	23:27
Breath Volume		1.863	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.000	23:30
Breath Volume		2.238	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 02/25/2020 08:43

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/12/2020 00:33:18

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.234	00:28
Breath Volume		3.039	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.230	00:31
Breath Volume		3.171	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ERCOLINO

DATE/TIME OF TEST: 02/12/2020 03:20:22

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.126	03:15
Breath Volume		1.476	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.131	03:18
Breath Volume		1.945	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CID

DATE/TIME OF TEST: 02/12/2020 05:10:58

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:08	
Control Test		0.080	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:09	Subject Test Refused
Breath Volume		0.000	05:09	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: WUSTEFELD HUELSMAN

DATE/TIME OF TEST: 02/12/2020 23:17:36

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.078	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.214	23:12
Breath Volume		1.199	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.219	23:16
Breath Volume		1.242	23:16
Air Blank		0.000	23:16
Control Test		0.080	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/13/2020 01:58:38

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.193	01:53
Breath Volume		2.238	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.174	01:57
Breath Volume		1.421	01:57
Air Blank		0.000	01:57
Control Test		0.079	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLRED,COBY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: LAURIN

DATE/TIME OF TEST: 02/13/2020 22:08:17

Diagnostics Check		OK	22:01
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.157	22:03
Breath Volume		3.699	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.158	22:06
Breath Volume		3.558	22:06
Air Blank		0.000	22:07
Control Test		0.078	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALFRED,BRANDON,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALFRED,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DYER JR.

DATE/TIME OF TEST: 02/15/2020 00:51:41

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.174	00:46
Breath Volume		1.878	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.173	00:50
Breath Volume		2.660	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PICHE,RUSSELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: COSBY

DATE/TIME OF TEST: 02/15/2020 01:02:20

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.086	00:57
Breath Volume		1.527	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.088	01:00
Breath Volume		1.203	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: JOHNSON-ALBRITTON

DATE/TIME OF TEST: 02/15/2020 01:11:13

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.000	01:06
Breath Volume		1.707	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.000	01:09
Breath Volume		2.300	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/15/2020 03:40:02

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.000	03:39	RFI Detect
Air Blank	0.000	03:40	

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/15/2020 03:53:40

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.080	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:46	No Sample Provided
Breath Volume		0.000	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2		03:52	No Sample Provided
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.081	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSTALES,ALEXANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: WADDINGTON

DATE/TIME OF TEST: 02/15/2020 09:59:20

Diagnostics Check		OK	09:50
Air Blank		0.000	09:51
Control Test		0.080	09:51
Air Blank		0.000	09:51
Subject Sample	# 1	0.091	09:53
Breath Volume		2.195	09:53
Air Blank		0.000	09:54
Air Blank		0.000	09:56
Subject Sample	# 2	0.088	09:57
Breath Volume		1.941	09:57
Air Blank		0.000	09:58
Control Test		0.081	09:58
Air Blank		0.000	09:59
Diagnostics Check		OK	09:59

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 02/15/2020 22:51:17

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:48	
Control Test		0.080	22:48	
Air Blank		0.000	22:48	
Subject Sample	# 1	0.137	22:49	
Breath Volume		1.199	22:49	
Air Blank			22:49	RFI Detect
Air Blank		0.000	22:50	
Control Test		0.080	22:50	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 02/15/2020 23:00:51

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.080	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.123	22:56
Breath Volume		1.324	22:56
Air Blank		0.000	22:57
Air Blank		0.000	22:58
Subject Sample	# 2	0.125	22:59
Breath Volume		1.980	22:59
Air Blank		0.000	22:59
Control Test		0.080	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BOYLES

DATE/TIME OF TEST: 02/16/2020 01:30:30

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.141	01:25
Breath Volume		1.199	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.145	01:29
Breath Volume		1.417	01:29
Air Blank		0.000	01:29
Control Test		0.079	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PICHE,RUSSELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: PULIDO

DATE/TIME OF TEST: 02/16/2020 20:42:24

Diagnostics Check		OK	20:34
Air Blank		0.000	20:34
Control Test		0.080	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.196	20:36
Breath Volume		2.640	20:36
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.190	20:40
Breath Volume		2.273	20:40
Air Blank		0.000	20:41
Control Test		0.079	20:41
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/16/2020 21:15:03

Diagnostics Check		OK	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:10
Subject Sample	# 1	0.165	21:10
Breath Volume		1.597	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.164	21:13
Breath Volume		1.203	21:13
Air Blank		0.000	21:14
Control Test		0.080	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:15

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,MAVERICK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOTO,MAVERICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 02/17/2020 10:45:41

Diagnostics Check		OK	10:38
Air Blank		0.000	10:38
Control Test		0.080	10:38
Air Blank		0.000	10:39
Subject Sample	# 1	0.059	10:40
Breath Volume		2.574	10:40
Air Blank		0.000	10:40
Air Blank		0.000	10:42
Subject Sample	# 2	0.061	10:44
Breath Volume		3.031	10:44
Air Blank		0.000	10:44
Control Test		0.078	10:45
Air Blank		0.000	10:45
Diagnostics Check		OK	10:45

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PONS

DATE/TIME OF TEST: 02/18/2020 02:19:39

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.000	02:15
Breath Volume		2.652	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.000	02:18
Breath Volume		2.621	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: TIPPER

DATE/TIME OF TEST: 02/18/2020 22:16:05

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.025	22:11
Breath Volume		1.761	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.026	22:14
Breath Volume		2.359	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/20/2020 01:42:13

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.175	01:37
Breath Volume		1.914	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.178	01:40
Breath Volume		1.460	01:40
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: POUERIE URBAEZ

DATE/TIME OF TEST: 02/20/2020 12:14:36

Diagnostics Check		OK	12:08
Air Blank		0.000	12:08
Control Test		0.080	12:09
Air Blank		0.000	12:09
Subject Sample	# 1	0.202	12:10
Breath Volume		3.800	12:10
Air Blank		0.000	12:10
Air Blank		0.000	12:12
Subject Sample	# 2	0.217	12:13
Breath Volume		4.171	12:13
Air Blank		0.000	12:13
Control Test		0.080	12:14
Air Blank		0.000	12:14
Diagnostics Check		OK	12:14

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,ROBERT,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HAMILTON,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 02/21/2020 17:48:33

Diagnostics Check		OK	17:41
Air Blank		0.000	17:42
Control Test		0.080	17:42
Air Blank		0.000	17:42
Subject Sample	# 1	0.138	17:43
Breath Volume		1.519	17:43
Air Blank		0.000	17:44
Air Blank		0.000	17:46
Subject Sample	# 2	0.143	17:47
Breath Volume		2.113	17:47
Air Blank		0.000	17:47
Control Test		0.080	17:47
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 02/22/2020 01:20:07

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.081	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.174	01:15
Breath Volume		1.906	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.191	01:18
Breath Volume		2.410	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUABA,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GUABA,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: ISIDRO

DATE/TIME OF TEST: 02/22/2020 05:32:01

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:29	
Control Test		0.080	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 1		05:30	Subject Test Refused
Breath Volume		0.000	05:30	
Air Blank		0.000	05:31	
Control Test		0.081	05:31	
Air Blank		0.000	05:31	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: TOVILLA-ESPINOZA

DATE/TIME OF TEST: 02/23/2020 02:21:31

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.180	02:17
Breath Volume		2.347	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.175	02:20
Breath Volume		1.335	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/23/2020 05:19:45

Diagnostics Check		OK	05:16	
Air Blank		0.000	05:17	
Control Test		0.080	05:17	
Air Blank		0.000	05:18	
Subject Sample	# 1		05:18	Subject Test Refused
Breath Volume		0.000	05:18	
Air Blank		0.000	05:18	
Control Test		0.081	05:19	
Air Blank		0.000	05:19	
Diagnostics Check		OK	05:19	

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MENDEZ,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: POLEN

DATE/TIME OF TEST: 02/23/2020 06:16:37

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.080	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.046	06:12
Breath Volume		3.062	06:12
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.044	06:15
Breath Volume		2.500	06:15
Air Blank		0.000	06:15
Control Test		0.081	06:16
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALLRED,COBY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: HURT

DATE/TIME OF TEST: 02/23/2020 20:12:56

Diagnostics Check		OK	20:09	
Air Blank		0.000	20:10	
Control Test		0.081	20:10	
Air Blank		0.000	20:11	
Subject Sample	# 1		20:11	Subject Test Refused
Breath Volume		0.000	20:11	
Air Blank		0.000	20:12	
Control Test		0.081	20:12	
Air Blank		0.000	20:12	
Diagnostics Check		OK	20:12	

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 02/23/2020 21:13:44

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.081	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.211	21:09
Breath Volume		2.316	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.206	21:12
Breath Volume		1.902	21:12
Air Blank		0.000	21:12
Control Test		0.081	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WIRTH

DATE/TIME OF TEST: 02/23/2020 23:33:23

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.081	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.107	23:28
Breath Volume		1.347	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.115	23:31
Breath Volume		1.644	23:31
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WENNING

DATE/TIME OF TEST: 02/25/2020 03:31:07

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.000	03:26
Breath Volume		1.843	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.000	03:29
Breath Volume		1.476	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 02/25/2020 08:44

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ZIMMER,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: CEPEDA

DATE/TIME OF TEST: 02/26/2020 00:08:36

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.081	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		00:01	No Sample Provided
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2		00:07	No Sample Provided
Breath Volume		0.000	00:07	
Air Blank		0.000	00:07	
Control Test		0.081	00:08	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: 0046,PICHE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 02/26/2020 01:48:36

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.081	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 2		01:47	No Sample Provided
Breath Volume		0.000	01:47	
Air Blank		0.000	01:47	
Control Test		0.081	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: CARBAJAL-MORALES

DATE/TIME OF TEST: 02/26/2020 19:37:18

Diagnostics Check		OK	19:22	
Air Blank		0.000	19:22	
Control Test		0.081	19:23	
Air Blank		0.000	19:23	
Subject Sample	# 1	0.215	19:24	
Breath Volume		2.339	19:24	
Air Blank		0.000	19:25	
Air Blank		0.000	19:27	
Subject Sample	# 2		19:30	No Sample Provided
Breath Volume		0.000	19:30	
Air Blank		0.000	19:30	
Air Blank		0.000	19:32	
Subject Sample	# 3		19:35	No Sample Provided
Breath Volume		0.000	19:35	
Air Blank		0.000	19:36	
Control Test		0.080	19:36	
Air Blank		0.000	19:37	
Diagnostics Check		OK	19:37	

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/26/2020 23:28:01

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.081	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.080	23:23
Breath Volume		2.390	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:25
Subject Sample	# 2	0.081	23:26
Breath Volume		2.250	23:26
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:28

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 02/27/2020 00:00:45

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.262	23:52
Breath Volume		2.378	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.239	23:56
Breath Volume		1.726	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 3	0.243	23:59
Breath Volume		1.921	23:59
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALSTON,SONDRA,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: POSILLICO

DATE/TIME OF TEST: 02/28/2020 06:35:11

Diagnostics Check		OK	06:29
Air Blank		0.000	06:29
Control Test		0.080	06:29
Air Blank		0.000	06:30
Subject Sample	# 1	0.106	06:30
Breath Volume		2.410	06:30
Air Blank		0.000	06:31
Air Blank		0.000	06:33
Subject Sample	# 2	0.106	06:33
Breath Volume		2.031	06:33
Air Blank		0.000	06:34
Control Test		0.081	06:34
Air Blank		0.000	06:35
Diagnostics Check		OK	06:35

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 02/29/2020 02:36:31

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:29	No Sample Provided
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:35	
Control Test		0.081	02:35	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUABA,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GUABA,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 02/29/2020 04:27:30

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.175	04:22
Breath Volume		1.964	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.166	04:25
Breath Volume		1.175	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PICHE,RUSSELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HAINES

DATE/TIME OF TEST: 02/29/2020 04:42:34

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.082	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:41	
Control Test		0.082	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: SERLETIC

DATE/TIME OF TEST: 02/29/2020 21:16:04

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:12	
Control Test		0.082	21:12	
Air Blank		0.000	21:13	
Subject Sample	# 1		21:14	RFI Detect
Breath Volume		0.000	21:14	
Air Blank		0.000	21:15	
Control Test		0.082	21:15	
Air Blank		0.000	21:15	
Diagnostics Check		OK	21:16	

Data Download Date/Time: 03/27/2020 11:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: SERLETIC

DATE/TIME OF TEST: 02/29/2020 19:28:27

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:17	
Control Test		0.082	19:17	
Air Blank		0.000	19:18	
Subject Sample	# 1		19:21	No Sample Provided
Breath Volume		0.000	19:21	
Air Blank		0.000	19:21	
Air Blank		0.000	19:23	
Subject Sample	# 2		19:27	No Sample Provided
Breath Volume		0.000	19:27	
Air Blank		0.000	19:27	
Control Test		0.081	19:27	
Air Blank		0.000	19:28	
Diagnostics Check		OK	19:28	

Data Download Date/Time: 03/27/2020 11:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: REPOSH

DATE/TIME OF TEST: 02/29/2020 20:23:22

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:12	
Control Test		0.080	20:12	
Air Blank		0.000	20:13	
Subject Sample	# 1		20:16	No Sample Provided
Breath Volume		0.000	20:16	
Air Blank		0.000	20:16	
Air Blank		0.000	20:18	
Subject Sample	# 2		20:22	No Sample Provided
Breath Volume		0.000	20:22	
Air Blank		0.000	20:22	
Control Test		0.082	20:22	
Air Blank		0.000	20:23	
Diagnostics Check		OK	20:23	

Data Download Date/Time: 03/27/2020 11:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,RICHARD,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PICHE,RUSSELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: TOJIN CHIVALAN

DATE/TIME OF TEST: 02/29/2020 23:41:25

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.081	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.192	23:36
Breath Volume		3.039	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.208	23:39
Breath Volume		3.585	23:39
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 03/27/2020 11:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PENE MURCIA

DATE/TIME OF TEST: 02/29/2020 00:18:52

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.197	00:14
Breath Volume		1.441	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:16
Subject Sample	# 2	0.199	00:17
Breath Volume		1.902	00:17
Air Blank		0.000	00:17
Control Test		0.080	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 03/27/2020 11:18

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MAZURYK,MICHAEL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 02/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: HEYWARD

DATE/TIME OF TEST: 02/23/2020 01:51:43

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.078	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.190	01:47
Breath Volume		1.832	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.187	01:50
Breath Volume		2.312	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 03/19/2020 13:33

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,KENNETH,D
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: NICHOLS,KENNETH
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: SEKYI

DATE/TIME OF TEST: 02/20/2020 12:39:19

Diagnostics Check		OK	12:33
Air Blank		0.000	12:33
Control Test		0.081	12:34
Air Blank		0.000	12:34
Subject Sample	# 1	0.000	12:35
Breath Volume		2.273	12:35
Air Blank		0.000	12:35
Air Blank		0.000	12:37
Subject Sample	# 2	0.000	12:37
Breath Volume		1.828	12:37
Air Blank		0.000	12:38
Control Test		0.081	12:38
Air Blank		0.000	12:39
Diagnostics Check		OK	12:39

Data Download Date/Time: 02/27/2020 08:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEBSON,BRET,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: HEBSON,BRET
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 09:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:47

SUBJECT'S LAST NAME: MCELWEE

DATE/TIME OF TEST: 02/21/2020 10:20:37

Diagnostics Check		OK	10:14
Air Blank		0.000	10:14
Control Test		0.080	10:15
Air Blank		0.000	10:15
Subject Sample	# 1	0.000	10:16
Breath Volume		2.429	10:16
Air Blank		0.000	10:16
Air Blank		0.000	10:18
Subject Sample	# 2	0.000	10:19
Breath Volume		2.160	10:19
Air Blank		0.000	10:19
Control Test		0.082	10:19
Air Blank		0.000	10:20
Diagnostics Check		OK	10:20

Data Download Date/Time: 02/27/2020 08:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,KENNETH,D
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: NICHOLS,KENNETH
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:13

SUBJECT'S LAST NAME: LEAVITT

DATE/TIME OF TEST: 02/21/2020 13:56:16

Diagnostics Check		OK	13:49
Air Blank		0.000	13:50
Control Test		0.080	13:50
Air Blank		0.000	13:51
Subject Sample	# 1	0.000	13:51
Breath Volume		2.535	13:51
Air Blank		0.000	13:52
Air Blank		0.000	13:54
Subject Sample	# 2	0.000	13:54
Breath Volume		2.398	13:54
Air Blank		0.000	13:55
Control Test		0.082	13:55
Air Blank		0.000	13:56
Diagnostics Check		OK	13:56

Data Download Date/Time: 02/27/2020 08:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERGE,TIMOTHY
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 02/02/2020 01:51:20

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.077	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.151	01:46
Breath Volume		1.718	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.153	01:49
Breath Volume		2.250	01:49
Air Blank		0.000	01:50
Control Test		0.077	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 02/25/2020 22:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,AARON,R
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HAMILTON,BRANDON
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A6
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LANEY

DATE/TIME OF TEST: 02/08/2020 02:17:37

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.078	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1	0.207	02:13	
Breath Volume		3.437	02:13	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2	0.210	02:16	
Breath Volume		3.679	02:16	
Air Blank		0.000	02:16	
Control Test		0.078	02:17	
Air Blank			02:17	RFI Detect
Air Blank			02:17	RFI Detect

Data Download Date/Time: 02/25/2020 22:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: WLODARSKI

DATE/TIME OF TEST: 02/01/2020 21:47:45

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.080	21:39	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:42	No Sample Provided
Breath Volume		0.000	21:42	
Air Blank		0.000	21:43	
Air Blank		0.000	21:45	
Subject Sample	# 2		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:46	
Control Test		0.079	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 02/28/2020 08:27

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 02/03/2020 01:06:41

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:04	
Control Test		0.078	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:05	Subject Test Refused
Breath Volume		0.000	01:05	
Air Blank		0.000	01:05	
Control Test		0.079	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 02/28/2020 08:27

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/22/2020 01:32:23

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.079	01:30	
Air Blank		0.000	01:30	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:31	
Control Test		0.079	01:31	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 02/28/2020 08:27

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEER,M,T
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GEER,M
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: MILLAN

DATE/TIME OF TEST: 02/27/2020 09:48:06

Diagnostics Check		OK	09:45	
Air Blank		0.000	09:45	
Control Test		0.079	09:46	
Air Blank		0.000	09:46	
Subject Sample	# 1		09:46	Subject Test Refused
Breath Volume		0.000	09:46	
Air Blank		0.000	09:47	
Control Test		0.080	09:47	
Air Blank		0.000	09:48	
Diagnostics Check		OK	09:48	

Data Download Date/Time: 02/28/2020 08:27

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: WARWICK,ANDREW
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: JETER

DATE/TIME OF TEST: 02/29/2020 20:54:27

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.253	20:49
Breath Volume		2.468	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.259	20:52
Breath Volume		1.910	20:52
Air Blank		0.000	20:53
Control Test		0.079	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 04/01/2020 09:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/21/2020 20:22:41

Diagnostics Check		OK	20:16
Air Blank		0.000	20:17
Control Test		0.082	20:17
Air Blank		0.000	20:17
Subject Sample	# 1	0.099	20:18
Breath Volume		1.449	20:18
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.104	20:21
Breath Volume		1.750	20:21
Air Blank		0.000	20:21
Control Test		0.081	20:22
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 02/21/2020 22:05:25

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:02	
Control Test		0.082	22:02	
Air Blank		0.000	22:03	
Subject Sample	# 1		22:04	Sequence Aborted
Breath Volume		0.000	22:04	
Air Blank		0.000	22:05	

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 02/21/2020 22:31:12

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.082	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.118	22:26
Breath Volume		2.062	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.114	22:29
Breath Volume		1.265	22:29
Air Blank		0.000	22:30
Control Test		0.081	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: LEAHY

DATE/TIME OF TEST: 02/23/2020 01:54:47

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.082	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.000	01:50
Breath Volume		1.921	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.000	01:53
Breath Volume		2.132	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ALDANA-PENA

DATE/TIME OF TEST: 02/23/2020 02:26:11

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.169	02:21
Breath Volume		2.292	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.162	02:24
Breath Volume		1.589	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: RODGIGUEZ JR

DATE/TIME OF TEST: 02/23/2020 03:56:08

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.134	03:50
Breath Volume		1.949	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.128	03:54
Breath Volume		1.636	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANWORP,JUDSON,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANWORP,JUDSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 02/27/2020 03:03:31

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.082	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:02	Subject Test Refused
Breath Volume		0.000	03:02	
Air Blank		0.000	03:02	
Control Test		0.081	03:02	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: KOTSOVOLOS

DATE/TIME OF TEST: 02/27/2020 04:41:49

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.081	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.196	04:36
Breath Volume		1.980	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:38
Subject Sample	# 2	0.193	04:40
Breath Volume		1.660	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 02/28/2020 02:50:56

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.113	02:46
Breath Volume		1.656	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.110	02:49
Breath Volume		2.968	02:49
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/28/2020 03:18:07

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.081	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:13	No Sample Provided
Breath Volume		0.000	03:13	
Air Blank		0.000	03:14	
Air Blank		0.000	03:16	
Subject Sample	# 2		03:16	Subject Test Refused
Breath Volume		0.000	03:16	
Air Blank		0.000	03:17	
Control Test		0.081	03:17	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: NEILD

DATE/TIME OF TEST: 02/29/2020 01:04:47

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.183	00:59
Breath Volume		3.718	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.182	01:03
Breath Volume		3.187	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG831804
DRY GAS EXPIRY DATE: 11/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RIVERA ALOMAR

DATE/TIME OF TEST: 02/29/2020 02:29:58

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:19	No Sample Provided
Breath Volume		0.000	02:19	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2	0.132	02:24	
Breath Volume		1.511	02:24	
Air Blank		0.000	02:25	
Air Blank		0.000	02:27	
Subject Sample	# 3	0.129	02:28	
Breath Volume		2.644	02:28	
Air Blank		0.000	02:29	
Control Test		0.081	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 03/06/2020 23:24

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: RAMPONE

DATE/TIME OF TEST: 02/10/2020 00:49:15

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.077	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.119	00:44
Breath Volume		2.273	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.107	00:47
Breath Volume		1.214	00:47
Air Blank		0.000	00:48
Control Test		0.077	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 02/21/2020 21:27

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: FIORE

DATE/TIME OF TEST: 02/12/2020 03:34:00

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.075	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.270	03:29
Breath Volume		2.085	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.278	03:32
Breath Volume		2.074	03:32
Air Blank		0.000	03:33
Control Test		0.075	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:34

Data Download Date/Time: 02/21/2020 21:27

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIARS,KELSEY,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VIARS,KELSEY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/01/2020 02:03:17

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.079	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.164	01:58
Breath Volume		1.824	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.160	02:01
Breath Volume		1.660	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 02/03/2020 18:14

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VIARS,KELSEY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 02/01/2020 20:16:36

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.079	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.000	20:12
Breath Volume		1.945	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.000	20:15
Breath Volume		1.507	20:15
Air Blank		0.000	20:15
Control Test		0.080	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 02/03/2020 18:14

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 02/01/2020 23:37:56

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.079	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.195	23:32
Breath Volume		1.843	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.180	23:36
Breath Volume		1.242	23:36
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 02/03/2020 18:14

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIARS,KELSEY,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VIARS,KELSEY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 02/01/2020 23:54:20

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:43	
Control Test		0.080	23:43	
Air Blank		0.000	23:43	
Subject Sample	# 1	0.201	23:47	Volume Not Met
Breath Volume		0.238	23:47	
Air Blank		0.000	23:47	
Air Blank		0.000	23:49	
Subject Sample	# 2	0.214	23:52	Volume Not Met
Breath Volume		0.207	23:52	
Air Blank		0.000	23:53	
Control Test		0.079	23:53	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 02/03/2020 18:14

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIARS,KELSEY,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VIARS,KELSEY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A2
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 02/02/2020 00:06:29

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:57	
Control Test		0.079	23:57	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.223	00:01	Volume Not Met
Breath Volume		0.078	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 2		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.077	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 02/03/2020 18:14

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THEODORE,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 02/05/2020 23:00:42

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.080	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.166	22:55
Breath Volume		2.351	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.169	22:59
Breath Volume		1.285	22:59
Air Blank		0.000	22:59
Control Test		0.078	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 02/12/2020 23:20:31

Diagnostics Check		OK	23:12
Air Blank		0.000	23:13
Control Test		0.080	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.205	23:15
Breath Volume		2.597	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 2	0.204	23:18
Breath Volume		2.453	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,OFC. J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,OFC. M
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: TUTEIN

DATE/TIME OF TEST: 02/13/2020 16:42:03

Diagnostics Check		OK	16:35
Air Blank		0.000	16:35
Control Test		0.080	16:36
Air Blank		0.000	16:36
Subject Sample	# 1	0.257	16:37
Breath Volume		1.589	16:37
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.251	16:40
Breath Volume		1.804	16:40
Air Blank		0.000	16:41
Control Test		0.078	16:41
Air Blank		0.000	16:41
Diagnostics Check		OK	16:42

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: FAUNCE

DATE/TIME OF TEST: 02/13/2020 23:26:14

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.080	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.057	23:21
Breath Volume		1.867	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.055	23:24
Breath Volume		1.773	23:24
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 02/14/2020 01:13:18

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.106	01:08
Breath Volume		2.199	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.098	01:11
Breath Volume		1.742	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 02/14/2020 04:06:00

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.000	04:01
Breath Volume		4.222	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:04
Breath Volume		3.660	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:06

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: NUNEZ ESPINOZA

DATE/TIME OF TEST: 02/16/2020 02:02:44

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.080	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:55	No Sample Provided
Breath Volume		0.000	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2		02:01	No Sample Provided
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Control Test		0.080	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/17/2020 21:33:31

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.080	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.222	21:28
Breath Volume		1.613	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.231	21:31
Breath Volume		1.890	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 02/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 02/29/2020 20:24:14

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.080	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.145	20:19
Breath Volume		3.566	20:19
Air Blank		0.000	20:20
Air Blank		0.000	20:21
Subject Sample	# 2	0.146	20:22
Breath Volume		3.316	20:22
Air Blank		0.000	20:23
Control Test		0.080	20:23
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 03/24/2020 19:06

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 02/13/2020 01:59:05

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.081	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:58	
Control Test		0.080	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/19/2020 03:09:46

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.081	02:55	
Air Blank		0.000	02:55	
Subject Sample	# 1	0.149	02:58	Volume Not Met
Breath Volume		0.484	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:01	
Subject Sample	# 2	0.172	03:02	
Breath Volume		1.605	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 3	0.186	03:08	
Breath Volume		1.164	03:08	
Air Blank		0.000	03:08	
Control Test		0.080	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT-COOK,S,M
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: SHANNON,J.B.
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: RUTAN

DATE/TIME OF TEST: 02/22/2020 07:59:43

Diagnostics Check		OK	07:53
Air Blank		0.000	07:53
Control Test		0.082	07:54
Air Blank		0.000	07:54
Subject Sample	# 1	0.133	07:55
Breath Volume		2.777	07:55
Air Blank		0.000	07:55
Air Blank		0.000	07:57
Subject Sample	# 2	0.138	07:58
Breath Volume		2.500	07:58
Air Blank		0.000	07:58
Control Test		0.083	07:59
Air Blank		0.000	07:59
Diagnostics Check		OK	07:59

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT-COOK,S,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOREL-SEPULVEDA,M.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: LAUREANO-PARILLA

DATE/TIME OF TEST: 02/22/2020 09:47:13

Diagnostics Check		OK	09:40
Air Blank		0.000	09:41
Control Test		0.083	09:41
Air Blank		0.000	09:42
Subject Sample	# 1	0.047	09:42
Breath Volume		1.523	09:42
Air Blank		0.000	09:43
Air Blank		0.000	09:45
Subject Sample	# 2	0.050	09:45
Breath Volume		1.449	09:45
Air Blank		0.000	09:46
Control Test		0.083	09:46
Air Blank		0.000	09:47
Diagnostics Check		OK	09:47

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: FORTES

DATE/TIME OF TEST: 02/24/2020 01:28:35

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.082	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.081	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 02/24/2020 04:40:02

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.081	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.175	04:34
Breath Volume		1.566	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.168	04:38
Breath Volume		1.308	04:38
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:39
Diagnostics Check		OK	04:40

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: KAUFMAN

DATE/TIME OF TEST: 02/25/2020 02:36:58

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.082	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TUCKER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:51

SUBJECT'S LAST NAME: MAXEY

DATE/TIME OF TEST: 02/26/2020 16:18:18

Diagnostics Check		OK	16:11
Air Blank		0.000	16:11
Control Test		0.081	16:12
Air Blank		0.000	16:12
Subject Sample	# 1	0.164	16:13
Breath Volume		1.687	16:13
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.157	16:16
Breath Volume		1.187	16:16
Air Blank		0.000	16:17
Control Test		0.079	16:17
Air Blank		0.000	16:18
Diagnostics Check		OK	16:18

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 02/27/2020 03:03:43

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.082	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.113	02:59
Breath Volume		1.714	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.111	03:02
Breath Volume		1.437	03:02
Air Blank		0.000	03:02
Control Test		0.081	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GOUGH

DATE/TIME OF TEST: 02/27/2020 23:23:39

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.082	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.269	23:19
Breath Volume		2.074	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.260	23:22
Breath Volume		2.054	23:22
Air Blank		0.000	23:22
Control Test		0.083	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 02/28/2020 12:46

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHEELER,ADDISON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: ARDOIN

DATE/TIME OF TEST: 02/28/2020 21:38:23

Diagnostics Check		OK	21:32
Air Blank		0.000	21:32
Control Test		0.082	21:33
Air Blank		0.000	21:33
Subject Sample	# 1	0.076	21:33
Breath Volume		1.195	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.076	21:36
Breath Volume		1.937	21:36
Air Blank		0.000	21:37
Control Test		0.082	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 03/12/2020 11:56

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 02/29/2020 01:03:08

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.082	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Control Test		0.082	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 03/12/2020 11:56

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BUSATTA,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CODY

DATE/TIME OF TEST: 02/29/2020 03:16:40

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.083	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.083	03:16	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 03/12/2020 11:56

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ROCKWELL

DATE/TIME OF TEST: 02/29/2020 04:34:11

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.083	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.094	04:29
Breath Volume		1.195	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.082	04:32
Breath Volume		1.175	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 03/12/2020 11:56

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/29/2020 22:58:02

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.083	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:56	Subject Test Refused
Breath Volume		0.000	22:56	
Air Blank		0.000	22:57	
Control Test		0.083	22:57	
Air Blank		0.000	22:57	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 03/12/2020 11:56

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 02/01/2020 07:53:30

Diagnostics Check		OK	07:47
Air Blank		0.000	07:47
Control Test		0.080	07:47
Air Blank		0.000	07:48
Subject Sample	# 1	0.205	07:48
Breath Volume		1.640	07:48
Air Blank		0.000	07:49
Air Blank		0.000	07:51
Subject Sample	# 2	0.201	07:52
Breath Volume		1.570	07:52
Air Blank		0.000	07:52
Control Test		0.079	07:52
Air Blank		0.000	07:53
Diagnostics Check		OK	07:53

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 02/01/2020 17:47:59

Diagnostics Check		OK	17:41
Air Blank		0.000	17:42
Control Test		0.080	17:42
Air Blank		0.000	17:43
Subject Sample	# 1	0.000	17:43
Breath Volume		1.386	17:43
Air Blank		0.000	17:44
Air Blank		0.000	17:46
Subject Sample	# 2	0.000	17:46
Breath Volume		1.531	17:46
Air Blank		0.000	17:47
Control Test		0.081	17:47
Air Blank		0.000	17:47
Diagnostics Check		OK	17:47

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ALEXANDER,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 02/13/2020 10:01:55

Diagnostics Check		OK	09:55
Air Blank		0.000	09:55
Control Test		0.079	09:56
Air Blank		0.000	09:56
Subject Sample	# 1	0.196	09:57
Breath Volume		1.847	09:57
Air Blank		0.000	09:57
Air Blank		0.000	09:59
Subject Sample	# 2	0.196	10:00
Breath Volume		1.925	10:00
Air Blank		0.000	10:01
Control Test		0.080	10:01
Air Blank		0.000	10:01
Diagnostics Check		OK	10:01

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 02/15/2020 13:09:30

Diagnostics Check		OK	13:02
Air Blank		0.000	13:03
Control Test		0.080	13:03
Air Blank		0.000	13:04
Subject Sample	# 1	0.421	13:04
Breath Volume		1.753	13:04
Air Blank		0.000	13:05
Air Blank		0.000	13:07
Subject Sample	# 2	0.408	13:08
Breath Volume		1.445	13:08
Air Blank		0.000	13:08
Control Test		0.079	13:08
Air Blank		0.000	13:09
Diagnostics Check		OK	13:09

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANELLA,E,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LAMAR,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 02/15/2020 15:55:31

Diagnostics Check		OK	15:48
Air Blank		0.000	15:49
Control Test		0.080	15:49
Air Blank		0.000	15:49
Subject Sample	# 1	0.000	15:51
Breath Volume		1.335	15:51
Air Blank		0.000	15:51
Air Blank		0.000	15:53
Subject Sample	# 2	0.000	15:54
Breath Volume		2.011	15:54
Air Blank		0.000	15:54
Control Test		0.079	15:54
Air Blank		0.000	15:55
Diagnostics Check		OK	15:55

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WHISENANT

DATE/TIME OF TEST: 02/16/2020 21:25:43

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.080	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.109	21:21
Breath Volume		2.414	21:21
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.108	21:24
Breath Volume		1.550	21:24
Air Blank		0.000	21:24
Control Test		0.080	21:25
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STONE,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: TZUL

DATE/TIME OF TEST: 02/17/2020 00:02:12

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.186	23:57
Breath Volume		2.351	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.185	00:00
Breath Volume		2.945	00:00
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,JESSE,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MCREE

DATE/TIME OF TEST: 02/20/2020 01:24:46

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.104	01:19
Breath Volume		1.449	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.103	01:23
Breath Volume		2.007	01:23
Air Blank		0.000	01:23
Control Test		0.081	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,JESSE,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 02/20/2020 04:20:53

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.187	04:15
Breath Volume		1.164	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.195	04:19
Breath Volume		1.488	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,JESSE,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CARTMEL

DATE/TIME OF TEST: 02/20/2020 20:52:11

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:41	
Control Test		0.080	20:41	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:45	No Sample Provided
Breath Volume		0.000	20:45	
Air Blank		0.000	20:45	
Air Blank		0.000	20:47	
Subject Sample	# 2	0.138	20:50	Volume Not Met
Breath Volume		0.070	20:50	
Air Blank		0.000	20:51	
Control Test		0.080	20:51	
Air Blank		0.000	20:52	
Diagnostics Check		OK	20:52	

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAHEY,M,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: DESIKAN

DATE/TIME OF TEST: 02/23/2020 03:11:56

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.113	03:07
Breath Volume		2.367	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:09
Subject Sample	# 2	0.108	03:10
Breath Volume		2.234	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,JESSE,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A1
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: WILLICK

DATE/TIME OF TEST: 02/25/2020 03:48:53

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.169	03:44
Breath Volume		2.308	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:46
Subject Sample	# 2	0.162	03:47
Breath Volume		1.992	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/28/2020 00:39

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,ALBERT,M
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: COLLINS,ALBEERT
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ARELLANO

DATE/TIME OF TEST: 02/03/2020 00:11:50

Diagnostics Check	OK	00:11	
Air Blank		00:11	RFI Detect
Air Blank		00:11	RFI Detect

Data Download Date/Time: 02/29/2020 23:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,ALBERT,M
BREATH TEST OPERATOR AGENCY: FHP TROOP K
ARRESTING OFFICER: COLLINS,ALBERT
ARRESTING OFFICER AGENCY: FHP TROOP K
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ARELLANO

DATE/TIME OF TEST: 02/03/2020 00:20:39

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.084	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.172	00:16
Breath Volume		2.398	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.165	00:19
Breath Volume		1.882	00:19
Air Blank		0.000	00:19
Control Test		0.083	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 02/29/2020 23:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KNISELY JR

DATE/TIME OF TEST: 02/01/2020 03:35:03

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.081	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:34	
Control Test		0.081	03:34	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: DRISCOLL

DATE/TIME OF TEST: 02/01/2020 19:31:41

Diagnostics Check		OK	19:28	
Air Blank		0.000	19:29	
Control Test		0.080	19:29	
Air Blank		0.000	19:30	
Subject Sample	# 1		19:30	Subject Test Refused
Breath Volume		0.000	19:30	
Air Blank		0.000	19:30	
Control Test		0.080	19:31	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:31	

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/03/2020 02:10:11

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.082	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1	0.189	02:00	Volume Not Met
Breath Volume		0.894	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 2	0.206	02:05	
Breath Volume		1.574	02:05	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 3	0.192	02:08	
Breath Volume		1.230	02:08	
Air Blank		0.000	02:09	
Control Test		0.080	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: LORICK

DATE/TIME OF TEST: 02/03/2020 03:25:35

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.081	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.253	03:21
Breath Volume		1.414	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.259	03:24
Breath Volume		2.265	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: YOTT

DATE/TIME OF TEST: 02/06/2020 02:55:36

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.077	02:51
Breath Volume		2.906	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.078	02:54
Breath Volume		2.441	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 02/07/2020 21:04:40

Diagnostics Check		OK	21:01	
Air Blank		0.000	21:02	
Control Test		0.081	21:02	
Air Blank		0.000	21:03	
Subject Sample	# 1		21:03	Subject Test Refused
Breath Volume		0.000	21:03	
Air Blank		0.000	21:03	
Control Test		0.081	21:04	
Air Blank		0.000	21:04	
Diagnostics Check		OK	21:04	

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 02/08/2020 04:03:04

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.082	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:57	No Sample Provided
Breath Volume		0.000	03:57	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 2		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Control Test		0.081	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 02/15/2020 00:59:18

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.081	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.100	00:54
Breath Volume		2.589	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.097	00:57
Breath Volume		2.546	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ALJALIL

DATE/TIME OF TEST: 02/16/2020 01:02:19

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.082	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.161	00:55
Breath Volume		2.277	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.157	01:00
Breath Volume		2.230	01:00
Air Blank		0.000	01:01
Control Test		0.082	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN],CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HODNOVICH

DATE/TIME OF TEST: 02/20/2020 02:10:13

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.215	02:05
Breath Volume		2.175	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.213	02:08
Breath Volume		2.121	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 02/22/2020 02:53:30

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.195	02:49
Breath Volume		3.652	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.196	02:52
Breath Volume		3.593	02:52
Air Blank		0.000	02:52
Control Test		0.082	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HINSON

DATE/TIME OF TEST: 02/22/2020 03:02:44

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.167	02:57
Breath Volume		2.679	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.152	03:01
Breath Volume		2.578	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: WARD,LOGAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LAVENTURE

DATE/TIME OF TEST: 02/23/2020 19:48:38

Diagnostics Check		OK	19:40
Air Blank		0.000	19:41
Control Test		0.081	19:41
Air Blank		0.000	19:42
Subject Sample	# 1	0.236	19:42
Breath Volume		1.328	19:42
Air Blank		0.000	19:43
Air Blank		0.000	19:44
Subject Sample	# 2	0.230	19:47
Breath Volume		1.421	19:47
Air Blank		0.000	19:47
Control Test		0.081	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDIE,BRYAN,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: BALDIE,BRYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 02/25/2020 21:30:05

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:19	
Control Test		0.080	21:19	
Air Blank		0.000	21:19	
Subject Sample	# 1		21:23	No Sample Provided
Breath Volume		0.000	21:23	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2		21:28	No Sample Provided
Breath Volume		0.000	21:28	
Air Blank		0.000	21:29	
Control Test		0.081	21:29	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 02/26/2020 13:36

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASBURY,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ASBURY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: MAFFETT

DATE/TIME OF TEST: 02/27/2020 13:05:08

Diagnostics Check		OK	12:58
Air Blank		0.000	12:59
Control Test		0.082	12:59
Air Blank		0.000	13:00
Subject Sample	# 1	0.000	13:00
Breath Volume		1.250	13:00
Air Blank		0.000	13:01
Air Blank		0.000	13:03
Subject Sample	# 2	0.000	13:03
Breath Volume		2.554	13:03
Air Blank		0.000	13:04
Control Test		0.081	13:04
Air Blank		0.000	13:05
Diagnostics Check		OK	13:05

Data Download Date/Time: 03/11/2020 22:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 02/27/2020 22:27:54

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.081	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.203	22:23
Breath Volume		1.851	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:25
Subject Sample	# 2	0.209	22:26
Breath Volume		1.574	22:26
Air Blank		0.000	22:26
Control Test		0.081	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 03/11/2020 22:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: YANEZ

DATE/TIME OF TEST: 02/28/2020 04:07:36

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.082	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.128	04:03
Breath Volume		1.460	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.132	04:06
Breath Volume		1.503	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 03/11/2020 22:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/29/2020 01:25:24

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.160	01:20
Breath Volume		1.406	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.176	01:23
Breath Volume		1.476	01:23
Air Blank		0.000	01:24
Control Test		0.081	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 03/11/2020 22:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: TAYLOR-CULBRETH

DATE/TIME OF TEST: 02/29/2020 04:52:09

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.082	04:42	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:45	No Sample Provided
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Air Blank		0.000	04:48	
Subject Sample	# 2		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:51	
Control Test		0.081	04:51	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 03/11/2020 22:02

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REAY,ALEX,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HIXSON,JEREMY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: SWAN

DATE/TIME OF TEST: 02/01/2020 13:58:06

Diagnostics Check		OK	13:51
Air Blank		0.000	13:52
Control Test		0.080	13:52
Air Blank		0.000	13:53
Subject Sample	# 1	0.058	13:53
Breath Volume		1.515	13:53
Air Blank		0.000	13:54
Air Blank		0.000	13:56
Subject Sample	# 2	0.060	13:56
Breath Volume		1.316	13:56
Air Blank		0.000	13:57
Control Test		0.081	13:57
Air Blank		0.000	13:58
Diagnostics Check		OK	13:58

Data Download Date/Time: 02/13/2020 11:14

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MURRAY,K
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 02/02/2020 05:38:35

Diagnostics Check		OK	05:31
Air Blank		0.000	05:32
Control Test		0.077	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.121	05:33
Breath Volume		1.941	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.130	05:37
Breath Volume		2.933	05:37
Air Blank		0.000	05:37
Control Test		0.076	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 02/20/2020 20:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLTON,ERIK,J
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUGARD,J
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 02/04/2020 18:47:39

Diagnostics Check		OK	18:41
Air Blank		0.000	18:41
Control Test		0.075	18:42
Air Blank		0.000	18:42
Subject Sample	# 1	0.207	18:43
Breath Volume		2.226	18:43
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.203	18:46
Breath Volume		1.535	18:46
Air Blank		0.000	18:46
Control Test		0.075	18:47
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 02/20/2020 20:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLTON,ERIK,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: UMPIERRE,M
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/10/2020 23:20:05

Diagnostics Check		OK	23:17	
Air Blank		0.000	23:17	
Control Test		0.076	23:18	
Air Blank		0.000	23:18	
Subject Sample	# 1		23:18	Subject Test Refused
Breath Volume		0.000	23:18	
Air Blank		0.000	23:19	
Control Test		0.076	23:19	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 02/20/2020 20:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: MARCELLON,VENSON
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 02/14/2020 22:59:26

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.076	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.169	22:54
Breath Volume		2.667	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.150	22:57
Breath Volume		1.425	22:57
Air Blank		0.000	22:58
Control Test		0.076	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 02/20/2020 20:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLTON,ERIK,J
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: TORO,F
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/17/2020 19:34:15

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:30	
Control Test		0.076	19:30	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:32	Subject Test Refused
Breath Volume		0.750	19:32	
Air Blank		0.000	19:33	
Control Test		0.077	19:33	
Air Blank		0.000	19:34	
Diagnostics Check		OK	19:34	

Data Download Date/Time: 02/20/2020 20:16

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,ROCHELLE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SETTE

DATE/TIME OF TEST: 02/16/2020 02:14:55

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.262	02:10
Breath Volume		2.109	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.272	02:13
Breath Volume		1.675	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROHDE,A,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: D'ALESSIO,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BLEAN

DATE/TIME OF TEST: 02/16/2020 04:54:24

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.079	04:52	
Air Blank		0.000	04:52	
Subject Sample	# 1		04:53	Subject Test Refused
Breath Volume		0.000	04:53	
Air Blank		0.000	04:53	
Control Test		0.079	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 02/18/2020 02:59:52

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1	0.102	02:50	Volume Not Met
Breath Volume		0.093	02:50	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.132	02:54	
Breath Volume		1.496	02:54	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 3	0.134	02:58	
Breath Volume		1.386	02:58	
Air Blank		0.000	02:58	
Control Test		0.077	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SAAVEDRA

DATE/TIME OF TEST: 02/19/2020 00:38:55

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.078	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.155	00:34
Breath Volume		2.082	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:36
Subject Sample	# 2	0.156	00:37
Breath Volume		2.355	00:37
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/20/2020 00:47:36

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.147	00:42
Breath Volume		1.613	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.151	00:46
Breath Volume		1.550	00:46
Air Blank		0.000	00:46
Control Test		0.078	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANGLOIS,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 02/20/2020 03:39:48

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.049	03:35
Breath Volume		1.414	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.049	03:38
Breath Volume		2.109	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 02/23/2020 08:05:39

Diagnostics Check		OK	07:56
Air Blank		0.000	07:56
Control Test		0.079	07:56
Air Blank		0.000	07:57
Subject Sample	# 1	0.281	07:57
Breath Volume		4.085	07:57
Air Blank		0.000	07:58
Air Blank		0.000	08:00
Subject Sample	# 2	0.240	08:00
Breath Volume		1.800	08:00
Air Blank		0.000	08:01
Air Blank		0.000	08:03
Subject Sample	# 3	0.268	08:03
Breath Volume		1.746	08:03
Air Blank		0.000	08:04
Control Test		0.077	08:05
Air Blank		0.000	08:05
Diagnostics Check		OK	08:05

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,R,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: MARKANT

DATE/TIME OF TEST: 02/23/2020 18:17:52

Diagnostics Check		OK	18:14	
Air Blank		0.000	18:15	
Control Test		0.078	18:15	
Air Blank		0.000	18:15	
Subject Sample	# 1		18:16	Subject Test Refused
Breath Volume		0.000	18:16	
Air Blank		0.000	18:16	
Control Test		0.078	18:17	
Air Blank		0.000	18:17	
Diagnostics Check		OK	18:17	

Data Download Date/Time: 02/28/2020 12:20

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/29/2020 01:17:29

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.198	01:12
Breath Volume		2.183	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.199	01:15
Breath Volume		2.121	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 03/12/2020 11:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: UBEL

DATE/TIME OF TEST: 02/29/2020 05:20:50

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.137	05:16
Breath Volume		2.738	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:18
Subject Sample	# 2	0.131	05:19
Breath Volume		1.367	05:19
Air Blank		0.000	05:19
Control Test		0.079	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 03/12/2020 11:35

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JORDAN,WILFORD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: BENTZ

DATE/TIME OF TEST: 02/18/2020 18:57:55

Diagnostics Check	OK	18:56	
Air Blank	0.000	18:57	
Control Test	0.000	18:57	Control Outside Tolerance
Air Blank	0.000	18:57	

Data Download Date/Time: 03/10/2020 01:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JORDAN,WILFORD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: BENTZ

DATE/TIME OF TEST: 02/18/2020 19:02:27

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:01	
Control Test	0.000	19:01	Control Outside Tolerance
Air Blank	0.000	19:02	

Data Download Date/Time: 03/10/2020 01:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 02/08/2020 02:43:54

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.082	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.155	02:39
Breath Volume		2.269	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.149	02:42
Breath Volume		1.378	02:42
Air Blank		0.000	02:42
Control Test		0.081	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 02/29/2020 22:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EASTMAN,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/24/2020 04:56:23

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.081	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.116	04:52
Breath Volume		1.871	04:52
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.114	04:54
Breath Volume		1.679	04:54
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 02/29/2020 22:25

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KRYSHCHENKO,ILLYA
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/01/2020 22:45:42

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:43	
Control Test		0.077	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:44	Subject Test Refused
Breath Volume		0.000	22:44	
Air Blank		0.000	22:44	
Control Test		0.078	22:45	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/05/2020 22:25:40

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.078	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.106	22:21
Breath Volume		2.625	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.100	22:24
Breath Volume		2.496	22:24
Air Blank		0.000	22:24
Control Test		0.077	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATEL,AKHIL,M
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WELCH,KYLE
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/09/2020 20:18:44

Diagnostics Check		OK	20:06	
Air Blank		0.000	20:06	
Control Test		0.078	20:07	
Air Blank		0.000	20:07	
Subject Sample	# 1	0.142	20:08	
Breath Volume		1.136	20:08	
Air Blank		0.000	20:09	
Air Blank		0.000	20:11	
Subject Sample	# 2	0.124	20:14	Volume Not Met
Breath Volume		0.402	20:14	
Air Blank		0.000	20:14	
Air Blank		0.000	20:16	
Subject Sample	# 3	0.153	20:17	
Breath Volume		1.140	20:17	
Air Blank		0.000	20:17	
Control Test		0.075	20:18	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:18	

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATEL,AKHIL,M
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: LEARY,DANTE
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 02/10/2020 05:30:05

Diagnostics Check		OK	05:17	
Air Blank		0.000	05:18	
Control Test		0.078	05:18	
Air Blank		0.000	05:19	
Subject Sample	# 1	0.175	05:19	
Breath Volume		1.753	05:19	
Air Blank		0.000	05:19	
Air Blank		0.000	05:21	
Subject Sample	# 2	0.154	05:24	Volume Not Met
Breath Volume		1.050	05:24	
Air Blank		0.000	05:25	
Air Blank		0.000	05:27	
Subject Sample	# 3	0.147	05:28	
Breath Volume		1.933	05:28	
Air Blank		0.000	05:29	
Control Test		0.074	05:29	Control Outside Tolerance
Air Blank		0.000	05:30	

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: LAWRENCE,T
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 02/12/2020 02:59:30

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.077	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.098	02:55
Breath Volume		1.714	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.095	02:57
Breath Volume		1.195	02:57
Air Blank		0.000	02:58
Control Test		0.077	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: SCOTT,D
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HOSTUTTLER

DATE/TIME OF TEST: 02/12/2020 04:18:17

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.060	04:13
Breath Volume		2.757	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.057	04:16
Breath Volume		1.531	04:16
Air Blank		0.000	04:17
Control Test		0.077	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CALEB,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CROFUTT,BRITTANY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 12:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: GARCIA RESTO

DATE/TIME OF TEST: 02/14/2020 13:18:20

Diagnostics Check		OK	13:12
Air Blank		0.000	13:12
Control Test		0.078	13:13
Air Blank		0.000	13:13
Subject Sample	# 1	0.000	13:14
Breath Volume		1.835	13:14
Air Blank		0.000	13:14
Air Blank		0.000	13:16
Subject Sample	# 2	0.000	13:16
Breath Volume		1.523	13:16
Air Blank		0.000	13:17
Control Test		0.078	13:17
Air Blank		0.000	13:18
Diagnostics Check		OK	13:18

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KRYSCHENKO,I
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BOUTHIETTE

DATE/TIME OF TEST: 02/19/2020 22:20:51

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.078	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.149	22:16
Breath Volume		3.605	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.149	22:19
Breath Volume		4.324	22:19
Air Blank		0.000	22:19
Control Test		0.076	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 02/28/2020 16:04

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 02/01/2020 21:13:54

Diagnostics Check		OK	21:07
Air Blank		0.000	21:08
Control Test		0.080	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.069	21:09
Breath Volume		1.859	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.066	21:12
Breath Volume		1.574	21:12
Air Blank		0.000	21:12
Control Test		0.079	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 02/04/2020 10:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/02/2020 04:49:24

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.152	04:45
Breath Volume		1.757	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.145	04:47
Breath Volume		1.785	04:47
Air Blank		0.000	04:48
Control Test		0.078	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 02/04/2020 10:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,A.,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMART,J.
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 02/03/2020 00:36:39

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.256	00:28
Breath Volume		1.687	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.233	00:31
Breath Volume		1.812	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:33
Subject Sample	# 3	0.229	00:35
Breath Volume		2.566	00:35
Air Blank		0.000	00:35
Control Test		0.077	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 02/04/2020 10:34

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/01/2020 03:50:06

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.139	03:45
Breath Volume		2.574	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.138	03:48
Breath Volume		2.074	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MOJICA SANTIAGO

DATE/TIME OF TEST: 02/02/2020 01:11:49

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.073	01:07
Breath Volume		2.230	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.074	01:10
Breath Volume		2.753	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: POE

DATE/TIME OF TEST: 02/03/2020 01:49:18

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.080	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:45	No Sample Provided
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 2		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,JOSHUA,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PULASKI,DAMON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SAYERS

DATE/TIME OF TEST: 02/03/2020 03:33:20

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.000	03:29
Breath Volume		2.789	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.000	03:31
Breath Volume		2.511	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,RONALD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 02/03/2020 19:24:49

Diagnostics Check		OK	19:18
Air Blank		0.000	19:18
Control Test		0.081	19:19
Air Blank		0.000	19:19
Subject Sample	# 1	0.000	19:20
Breath Volume		2.359	19:20
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 2	0.000	19:23
Breath Volume		2.000	19:23
Air Blank		0.000	19:23
Control Test		0.080	19:24
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WIERMAN

DATE/TIME OF TEST: 02/04/2020 04:08:11

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.168	04:02
Breath Volume		2.621	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.173	04:06
Breath Volume		2.972	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,DANIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COMSTOCK,M
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: SZUCS

DATE/TIME OF TEST: 02/05/2020 03:05:56

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.185	03:01
Breath Volume		2.933	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.200	03:04
Breath Volume		5.292	03:04
Air Blank		0.000	03:05
Control Test		0.078	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NETTLES,TRISTAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: CHATTERSON

DATE/TIME OF TEST: 02/05/2020 07:56:17

Diagnostics Check		OK	07:48	
Air Blank		0.000	07:48	
Control Test		0.081	07:49	
Air Blank		0.000	07:49	
Subject Sample	# 1		07:52	No Sample Provided
Breath Volume		0.000	07:52	
Air Blank		0.000	07:53	
Air Blank			07:54	RFI Detect
Air Blank		0.000	07:55	
Control Test		0.081	07:55	
Air Blank		0.000	07:56	
Diagnostics Check		OK	07:56	

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NETTLES,TRISTAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: RAMETTA

DATE/TIME OF TEST: 02/08/2020 06:07:22

Diagnostics Check		OK	06:01
Air Blank		0.000	06:01
Control Test		0.080	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.217	06:02
Breath Volume		2.300	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.234	06:05
Breath Volume		3.335	06:05
Air Blank		0.000	06:06
Control Test		0.080	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: REDMOND

DATE/TIME OF TEST: 02/08/2020 22:15:50

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.080	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:07	No Sample Provided
Breath Volume		0.000	22:07	
Air Blank		0.000	22:08	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.199	22:11	
Breath Volume		2.335	22:11	
Air Blank		0.000	22:11	
Air Blank		0.000	22:13	
Subject Sample	# 3	0.185	22:14	
Breath Volume		2.269	22:14	
Air Blank		0.000	22:14	
Control Test		0.080	22:15	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,SUMMER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: NOVENSON

DATE/TIME OF TEST: 02/09/2020 04:30:45

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.162	04:26
Breath Volume		2.398	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.160	04:29
Breath Volume		1.378	04:29
Air Blank		0.000	04:29
Control Test		0.079	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,SUMMER,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,SUMMER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/09/2020 04:42:36

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:34	
Control Test		0.079	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 1		04:38	No Sample Provided
Breath Volume		0.000	04:38	
Air Blank		0.000	04:38	
Air Blank		0.000	04:40	
Subject Sample	# 2		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:41	
Control Test		0.080	04:42	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 02/11/2020 02:47:15

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.219	02:42
Breath Volume		5.171	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.221	02:45
Breath Volume		4.691	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 02/13/2020 11:07

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TRINIDAD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: COLLEY

DATE/TIME OF TEST: 02/12/2020 00:11:08

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.230	00:06
Breath Volume		3.671	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.219	00:09
Breath Volume		3.226	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 02/13/2020 11:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/14/2020 23:16:01

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:12	
Control Test		0.078	23:12	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Control Test		0.079	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 03/22/2020 15:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BLAHA

DATE/TIME OF TEST: 02/15/2020 01:27:01

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:23	
Control Test		0.080	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 03/22/2020 15:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/17/2020 00:48:50

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:39	No Sample Provided
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2	0.182	00:43	
Breath Volume		3.058	00:43	
Air Blank		0.000	00:44	
Air Blank		0.000	00:46	
Subject Sample	# 3	0.175	00:47	
Breath Volume		2.496	00:47	
Air Blank		0.000	00:47	
Control Test		0.078	00:48	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 03/22/2020 15:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BURCHAM

DATE/TIME OF TEST: 02/18/2020 21:34:50

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:29
Air Blank		0.000	21:29
Subject Sample	# 1	0.180	21:30
Breath Volume		2.503	21:30
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.174	21:33
Breath Volume		1.230	21:33
Air Blank		0.000	21:33
Control Test		0.079	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 03/22/2020 15:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FITTS,DEREK
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 02/19/2020 23:37:25

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.118	23:32
Breath Volume		1.292	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.110	23:35
Breath Volume		1.695	23:35
Air Blank		0.000	23:36
Control Test		0.076	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JOHANSEN,LLOYD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: TRAXEL

DATE/TIME OF TEST: 02/20/2020 05:26:47

Diagnostics Check		OK	05:11	
Air Blank		0.000	05:12	
Control Test		0.079	05:12	
Air Blank		0.000	05:12	
Subject Sample	# 1	0.159	05:16	Volume Not Met
Breath Volume		0.210	05:16	
Air Blank		0.000	05:16	
Air Blank		0.000	05:18	
Subject Sample	# 2	0.154	05:20	No .020 agreement
Breath Volume		1.992	05:20	
Air Blank		0.000	05:21	
Air Blank		0.000	05:23	
Subject Sample	# 3	0.133	05:24	No .020 agreement
Breath Volume		2.085	05:24	
Air Blank		0.000	05:25	
Control Test		0.075	05:26	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JOHANSEN,LLOYD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: TRAXEL

DATE/TIME OF TEST: 02/20/2020 06:07:09

Diagnostics Check		OK	05:58
Air Blank		0.000	05:58
Control Test		0.078	05:59
Air Blank		0.000	05:59
Subject Sample	# 1	0.129	06:01
Breath Volume		1.789	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:03
Subject Sample	# 2	0.132	06:05
Breath Volume		1.589	06:05
Air Blank		0.000	06:06
Control Test		0.076	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANKLIN,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FRANKLIN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/21/2020 04:58:16

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.080	04:50	
Air Blank		0.000	04:50	
Subject Sample	# 1		04:53	No Sample Provided
Breath Volume		0.000	04:53	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 2		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:57	
Control Test		0.080	04:57	
Air Blank		0.000	04:58	
Diagnostics Check		OK	04:58	

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 02/22/2020 00:16:11

Diagnostics Check		OK	00:07	
Air Blank		0.000	00:08	
Control Test		0.078	00:08	
Air Blank		0.000	00:08	
Subject Sample	# 1		00:12	No Sample Provided
Breath Volume		0.000	00:12	
Air Blank		0.000	00:12	
Air Blank		0.000	00:14	
Subject Sample	# 2		00:14	Subject Test Refused
Breath Volume		0.000	00:14	
Air Blank		0.000	00:15	
Control Test		0.080	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,TRISTAN,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NETTLES,TRISTAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 02/22/2020 23:56:45

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.141	23:52
Breath Volume		1.730	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.134	23:55
Breath Volume		1.417	23:55
Air Blank		0.000	23:55
Control Test		0.079	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 02/23/2020 18:01:50

Diagnostics Check		OK	17:55
Air Blank		0.000	17:55
Control Test		0.079	17:56
Air Blank		0.000	17:56
Subject Sample	# 1	0.133	17:57
Breath Volume		3.113	17:57
Air Blank		0.000	17:57
Air Blank		0.000	17:59
Subject Sample	# 2	0.134	18:00
Breath Volume		2.671	18:00
Air Blank		0.000	18:00
Control Test		0.078	18:01
Air Blank		0.000	18:01
Diagnostics Check		OK	18:01

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 02/24/2020 21:27:28

Diagnostics Check		OK	21:20
Air Blank		0.000	21:21
Control Test		0.079	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.117	21:22
Breath Volume		2.925	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.115	21:25
Breath Volume		2.765	21:25
Air Blank		0.000	21:26
Control Test		0.079	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BURROW

DATE/TIME OF TEST: 02/28/2020 03:58:36

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.079	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.212	03:54
Breath Volume		2.062	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.215	03:57
Breath Volume		2.093	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/28/2020 12:46:20

Diagnostics Check		OK	12:39
Air Blank		0.000	12:39
Control Test		0.080	12:39
Air Blank		0.000	12:40
Subject Sample	# 1	0.000	12:41
Breath Volume		1.933	12:41
Air Blank		0.000	12:42
Air Blank		0.000	12:44
Subject Sample	# 2	0.000	12:44
Breath Volume		1.566	12:44
Air Blank		0.000	12:45
Control Test		0.078	12:45
Air Blank		0.000	12:46
Diagnostics Check		OK	12:46

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,DANIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LAVALLE,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/29/2020 00:38:15

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.079	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.233	00:33
Breath Volume		2.867	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.233	00:36
Breath Volume		2.238	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 02/29/2020 18:50:50

Diagnostics Check		OK	18:45	
Air Blank		0.000	18:46	
Control Test		0.080	18:46	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:49	RFI Detect
Breath Volume		0.000	18:49	
Air Blank		0.000	18:49	
Control Test		0.081	18:50	
Air Blank		0.000	18:50	
Diagnostics Check		OK	18:50	

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 02/29/2020 19:06:19

Diagnostics Check		OK	18:52
Air Blank		0.000	18:53
Control Test		0.079	18:53
Air Blank		0.000	18:54
Subject Sample	# 1	0.166	18:55
Breath Volume		1.250	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:57
Subject Sample	# 2	0.140	18:59
Breath Volume		2.085	18:59
Air Blank		0.000	18:59
Air Blank		0.000	19:01
Subject Sample	# 3	0.144	19:04
Breath Volume		2.257	19:04
Air Blank		0.000	19:05
Control Test		0.080	19:05
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FITTS,DEREK
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BRUNEAU

DATE/TIME OF TEST: 02/29/2020 21:02:13

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.080	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.109	20:57
Breath Volume		2.992	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.106	21:00
Breath Volume		2.832	21:00
Air Blank		0.000	21:01
Control Test		0.078	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 03/22/2020 15:09

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,FRANK,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: QUINTANA,FRANK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: COTTA

DATE/TIME OF TEST: 02/21/2020 03:18:44

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank		03:18	RFI Detect
Air Blank	0.000	03:18	

Data Download Date/Time: 02/29/2020 23:17

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,FRANK,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: QUINTANA,FRANK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: COTTA

DATE/TIME OF TEST: 02/21/2020 03:26:27

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:21	
Control Test		0.079	03:22	
Air Blank		0.000	03:22	
Subject Sample	# 1	0.151	03:24	
Breath Volume		1.128	03:24	
Air Blank			03:25	Purge Fail
Air Blank			03:26	Purge Fail

Data Download Date/Time: 02/29/2020 23:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: DILTZ

DATE/TIME OF TEST: 02/01/2020 22:31:15

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.080	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.171	22:26
Breath Volume		1.480	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.180	22:29
Breath Volume		1.402	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 02/18/2020 23:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

INSTRUMENT REGISTERED TO: FHP TROOP D ORLANDO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,STEFAN,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 02/09/2020 00:42:03

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.081	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1	0.205	00:34	Volume Not Met
Breath Volume		0.171	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 2	0.218	00:37	No .020 agreement
Breath Volume		1.664	00:37	
Air Blank		0.000	00:37	
Air Blank		0.000	00:39	
Subject Sample	# 3	0.240	00:40	No .020 agreement
Breath Volume		2.214	00:40	
Air Blank		0.000	00:41	
Control Test		0.080	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 02/18/2020 23:08

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JARED,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: WESTBURY,MATTHEW
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: SCEE

DATE/TIME OF TEST: 02/16/2020 22:06:10

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.080	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.270	22:01
Breath Volume		3.515	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:03
Subject Sample	# 2	0.270	22:04
Breath Volume		3.867	22:04
Air Blank		0.000	22:05
Control Test		0.079	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 03/03/2020 03:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JARED,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: DAVIS,JARED
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 02/20/2020 23:32:19

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.080	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:31	
Control Test		0.080	23:31	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:32	

Data Download Date/Time: 03/03/2020 03:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIFFANY,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CULBRETH

DATE/TIME OF TEST: 02/22/2020 00:05:43

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
		00:05	RFI Detect
Air Blank	0.000	00:05	

Data Download Date/Time: 03/03/2020 03:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIFFANY,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CULBRETH

DATE/TIME OF TEST: 02/22/2020 00:12:07

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:09	
Control Test		0.080	00:10	
Air Blank		0.000	00:10	
Subject Sample	# 1		00:10	RFI Detect
Breath Volume		0.000	00:10	
Air Blank		0.000	00:11	
Control Test		0.080	00:11	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 03/03/2020 03:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TIFFANY,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CULBRETH

DATE/TIME OF TEST: 02/22/2020 00:22:40

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.146	00:18
Breath Volume		2.714	00:18
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.151	00:21
Breath Volume		2.886	00:21
Air Blank		0.000	00:21
Control Test		0.082	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 03/03/2020 03:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,KENNETH,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ROBERTS,KENNETH
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BARNARD

DATE/TIME OF TEST: 02/23/2020 20:09:49

Diagnostics Check		OK	20:03
Air Blank		0.000	20:04
Control Test		0.080	20:04
Air Blank		0.000	20:04
Subject Sample	# 1	0.062	20:05
Breath Volume		1.652	20:05
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.058	20:08
Breath Volume		1.414	20:08
Air Blank		0.000	20:08
Control Test		0.080	20:09
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 03/03/2020 03:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,MATTHEW,T
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: BUTLER,MATTHEW
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: HOLBROOK

DATE/TIME OF TEST: 02/28/2020 20:04:18

Diagnostics Check		OK	19:55
Air Blank		0.000	19:56
Control Test		0.081	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.103	19:58
Breath Volume		1.121	19:58
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.098	20:02
Breath Volume		1.285	20:02
Air Blank		0.000	20:03
Control Test		0.080	20:03
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 03/03/2020 03:41

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: SLATE

DATE/TIME OF TEST: 02/04/2020 21:34:50

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:29
Air Blank		0.000	21:29
Subject Sample	# 1	0.277	21:30
Breath Volume		1.562	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:32
Subject Sample	# 2	0.277	21:33
Breath Volume		1.343	21:33
Air Blank		0.000	21:33
Control Test		0.079	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 02/24/2020 13:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 02/07/2020 20:50:16

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.080	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.258	20:45
Breath Volume		3.640	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.249	20:48
Breath Volume		3.425	20:48
Air Blank		0.000	20:49
Control Test		0.080	20:49
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 02/24/2020 13:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/09/2020 20:42:42

Diagnostics Check		OK	20:34
Air Blank		0.000	20:35
Control Test		0.081	20:35
Air Blank		0.000	20:36
Subject Sample	# 1	0.186	20:37
Breath Volume		1.863	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:39
Subject Sample	# 2	0.179	20:41
Breath Volume		2.625	20:41
Air Blank		0.000	20:41
Control Test		0.078	20:42
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 02/24/2020 13:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: DOUGAN

DATE/TIME OF TEST: 02/12/2020 20:39:24

Diagnostics Check		OK	20:30	
Air Blank		0.000	20:30	
Control Test		0.080	20:30	
Air Blank		0.000	20:31	
Subject Sample	# 1	0.261	20:33	
Breath Volume		1.597	20:33	
Air Blank		0.000	20:34	
Air Blank		0.000	20:36	
Subject Sample	# 2	0.274	20:37	
Breath Volume		1.402	20:37	
Air Blank		0.000	20:38	
Control Test		0.077	20:38	
Air Blank			20:38	RFI Detect
Air Blank		0.000	20:39	

Data Download Date/Time: 02/24/2020 13:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: DOUGAN

DATE/TIME OF TEST: 02/12/2020 21:28:04

Diagnostics Check		OK	21:13	
Air Blank		0.000	21:13	
Control Test		0.079	21:14	
Air Blank		0.000	21:14	
Subject Sample	# 1	0.240	21:17	Slope Not Level
Breath Volume		1.121	21:17	
Air Blank		0.000	21:18	
Air Blank		0.000	21:20	
Subject Sample	# 2	0.258	21:23	
Breath Volume		1.105	21:23	
Air Blank		0.000	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 3	0.270	21:26	
Breath Volume		1.406	21:26	
Air Blank		0.000	21:27	
Control Test		0.077	21:27	
Air Blank		0.000	21:27	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 02/24/2020 13:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 02/17/2020 17:22:09

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.079	17:16
Air Blank		0.000	17:16
Subject Sample	# 1	0.330	17:17
Breath Volume		2.363	17:17
Air Blank		0.000	17:18
Air Blank		0.000	17:19
Subject Sample	# 2	0.319	17:20
Breath Volume		1.808	17:20
Air Blank		0.000	17:21
Control Test		0.079	17:21
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 02/24/2020 13:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODENZI,CAMERON,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GODENZI,CAMERON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SALMON

DATE/TIME OF TEST: 02/21/2020 21:19:14

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.072	21:14
Breath Volume		1.816	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.066	21:17
Breath Volume		1.621	21:17
Air Blank		0.000	21:18
Control Test		0.079	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 02/24/2020 13:13

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/29/2020 01:09:42

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.161	01:05
Breath Volume		2.199	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.155	01:08
Breath Volume		3.281	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 03/30/2020 13:58

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,JIMMIE,N
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DAGLEY,BRIE
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 01/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: JEANTY

DATE/TIME OF TEST: 02/06/2020 16:32:58

Diagnostics Check		OK	16:25
Air Blank		0.000	16:26
Control Test		0.079	16:26
Air Blank		0.000	16:27
Subject Sample	# 1	0.333	16:27
Breath Volume		1.347	16:27
Air Blank		0.000	16:28
Air Blank		0.000	16:30
Subject Sample	# 2	0.334	16:31
Breath Volume		1.335	16:31
Air Blank		0.000	16:32
Control Test		0.080	16:32
Air Blank		0.000	16:32
Diagnostics Check		OK	16:32

Data Download Date/Time: 02/24/2020 12:54

Printed: 22 April 2020

COBRA ©1997-2014 CMI, Inc.