# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: RIX

DATE/TIME OF TEST: 03/27/2020 23:15:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:08 |
| Air Blank | | 0.000 | 23:09 |
| Control Test | | 0.078 | 23:09 |
| Air Blank | | 0.000 | 23:10 |
| Subject Sample | # 1 | 0.246 | 23:10 |
| Breath Volume | | 1.617 | 23:10 |
| Air Blank | | 0.000 | 23:11 |
| Air Blank | | 0.000 | 23:13 |
| Subject Sample | # 2 | 0.253 | 23:13 |
| Breath Volume | | 1.773 | 23:13 |
| Air Blank | | 0.000 | 23:14 |
| Control Test | | 0.079 | 23:14 |
| Air Blank | | 0.000 | 23:15 |
| Diagnostics Check | | OK | 23:15 |

Data Download Date/Time: 04/21/2020 07:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 03/29/2020 23:31:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:23 |
| Air Blank | | 0.000 | 23:24 |
| Control Test | | 0.080 | 23:24 |
| Air Blank | | 0.000 | 23:25 |
| Subject Sample | # 1 | 0.211 | 23:25 |
| Breath Volume | | 1.773 | 23:25 |
| Air Blank | | 0.000 | 23:26 |
| Air Blank | | 0.000 | 23:27 |
| Subject Sample | # 2 | 0.223 | 23:29 |
| Breath Volume | | 1.796 | 23:29 |
| Air Blank | | 0.000 | 23:30 |
| Control Test | | 0.079 | 23:30 |
| Air Blank | | 0.000 | 23:31 |
| Diagnostics Check | | OK | 23:31 |

Data Download Date/Time: 04/21/2020 07:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/02/2020 18:16:05

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 18:14 | |
| Air Blank | 0.000 | 18:15 | |
| Control Test | 0.081 | 18:15 | |
| Air Blank | | 18:15 | RFI Detect |
| Air Blank | 0.000 | 18:16 | |

Data Download Date/Time: 03/13/2020 11:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/02/2020 18:24:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:18 |
| Air Blank | | 0.000 | 18:18 |
| Control Test | | 0.081 | 18:18 |
| Air Blank | | 0.000 | 18:19 |
| Subject Sample | # 1 | 0.000 | 18:19 |
| Breath Volume | | 1.796 | 18:19 |
| Air Blank | | 0.000 | 18:20 |
| Air Blank | | 0.000 | 18:22 |
| Subject Sample | # 2 | 0.000 | 18:23 |
| Breath Volume | | 1.179 | 18:23 |
| Air Blank | | 0.000 | 18:24 |
| Control Test | | 0.082 | 18:24 |
| Air Blank | | 0.000 | 18:24 |
| Diagnostics Check | | OK | 18:24 |

Data Download Date/Time: 03/13/2020 11:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/02/2020 20:21:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:15 |
| Air Blank | | 0.000 | 20:15 |
| Control Test | | 0.082 | 20:16 |
| Air Blank | | 0.000 | 20:16 |
| Subject Sample | # 1 | 0.378 | 20:17 |
| Breath Volume | | 2.054 | 20:17 |
| Air Blank | | 0.000 | 20:18 |
| Air Blank | | 0.000 | 20:19 |
| Subject Sample | # 2 | 0.383 | 20:20 |
| Breath Volume | | 3.152 | 20:20 |
| Air Blank | | 0.000 | 20:20 |
| Control Test | | 0.080 | 20:21 |
| Air Blank | | 0.000 | 20:21 |
| Diagnostics Check | | OK | 20:21 |

Data Download Date/Time: 03/13/2020 11:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 03/07/2020 00:38:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:31 |
| Air Blank | | 0.000 | 00:32 |
| Control Test | | 0.081 | 00:32 |
| Air Blank | | 0.000 | 00:32 |
| Subject Sample | # 1 | 0.098 | 00:33 |
| Breath Volume | | 2.796 | 00:33 |
| Air Blank | | 0.000 | 00:34 |
| Air Blank | | 0.000 | 00:35 |
| Subject Sample | # 2 | 0.103 | 00:36 |
| Breath Volume | | 2.746 | 00:36 |
| Air Blank | | 0.000 | 00:37 |
| Control Test | | 0.081 | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Diagnostics Check | | OK | 00:38 |

Data Download Date/Time: 03/13/2020 11:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: GARCIA VELAZQUEZ

DATE/TIME OF TEST: 03/07/2020 03:02:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:56 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.081 | 02:57 |
| Air Blank | | 0.000 | 02:57 |
| Subject Sample | # 1 | 0.263 | 02:58 |
| Breath Volume | | 2.792 | 02:58 |
| Air Blank | | 0.000 | 02:58 |
| Air Blank | | 0.000 | 03:00 |
| Subject Sample | # 2 | 0.249 | 03:01 |
| Breath Volume | | 2.472 | 03:01 |
| Air Blank | | 0.000 | 03:01 |
| Control Test | | 0.081 | 03:02 |
| Air Blank | | 0.000 | 03:02 |
| Diagnostics Check | | OK | 03:02 |

Data Download Date/Time: 03/13/2020 11:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: CUYUCH MARTIN

DATE/TIME OF TEST: 03/07/2020 19:17:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:10 |
| Air Blank | | 0.000 | 19:10 |
| Control Test | | 0.083 | 19:11 |
| Air Blank | | 0.000 | 19:11 |
| Subject Sample | # 1 | 0.176 | 19:12 |
| Breath Volume | | 2.691 | 19:12 |
| Air Blank | | 0.000 | 19:13 |
| Air Blank | | 0.000 | 19:15 |
| Subject Sample | # 2 | 0.172 | 19:15 |
| Breath Volume | | 2.722 | 19:15 |
| Air Blank | | 0.000 | 19:16 |
| Control Test | | 0.082 | 19:16 |
| Air Blank | | 0.000 | 19:17 |
| Diagnostics Check | | OK | 19:17 |

Data Download Date/Time: 03/13/2020 11:13

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MURRELL

DATE/TIME OF TEST: 03/21/2020 00:31:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:23 |
| Air Blank | | 0.000 | 00:24 |
| Control Test | | 0.080 | 00:24 |
| Air Blank | | 0.000 | 00:25 |
| Subject Sample | # 1 | 0.174 | 00:26 |
| Breath Volume | | 1.492 | 00:26 |
| Air Blank | | 0.000 | 00:27 |
| Air Blank | | 0.000 | 00:29 |
| Subject Sample | # 2 | 0.177 | 00:29 |
| Breath Volume | | 1.695 | 00:29 |
| Air Blank | | 0.000 | 00:30 |
| Control Test | | 0.079 | 00:30 |
| Air Blank | | 0.000 | 00:31 |
| Diagnostics Check | | OK | 00:31 |

Data Download Date/Time: 04/21/2020 07:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACKY,TAYLOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: RACKY

DATE/TIME OF TEST: 03/21/2020 10:37:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:30 |
| Air Blank | | 0.000 | 10:30 |
| Control Test | | 0.080 | 10:31 |
| Air Blank | | 0.000 | 10:31 |
| Subject Sample | # 1 | 0.165 | 10:32 |
| Breath Volume | | 1.863 | 10:32 |
| Air Blank | | 0.000 | 10:32 |
| Air Blank | | 0.000 | 10:34 |
| Subject Sample | # 2 | 0.163 | 10:35 |
| Breath Volume | | 1.585 | 10:35 |
| Air Blank | | 0.000 | 10:36 |
| Control Test | | 0.078 | 10:36 |
| Air Blank | | 0.000 | 10:37 |
| Diagnostics Check | | OK | 10:37 |

Data Download Date/Time: 04/21/2020 07:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHT,JAMES,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VAUGHT,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NAVARRETE

DATE/TIME OF TEST: 03/29/2020 04:53:06

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:49 | |
| Air Blank | | 0.000 | 04:50 | |
| Control Test | | 0.081 | 04:50 | |
| Air Blank | | 0.000 | 04:51 | |
| Subject Sample | # 1 | | 04:51 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:51 | |
| Air Blank | | 0.000 | 04:52 | |
| Control Test | | 0.081 | 04:52 | |
| Air Blank | | 0.000 | 04:53 | |
| Diagnostics Check | | OK | 04:53 | |

Data Download Date/Time: 04/21/2020 07:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: 6L03,WENTLAND
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 03/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 03/30/2020 04:30:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:23 |
| Air Blank | | 0.000 | 04:24 |
| Control Test | | 0.081 | 04:24 |
| Air Blank | | 0.000 | 04:25 |
| Subject Sample | # 1 | 0.145 | 04:26 |
| Breath Volume | | 1.335 | 04:26 |
| Air Blank | | 0.000 | 04:26 |
| Air Blank | | 0.000 | 04:28 |
| Subject Sample | # 2 | 0.153 | 04:29 |
| Breath Volume | | 2.777 | 04:29 |
| Air Blank | | 0.000 | 04:29 |
| Control Test | | 0.080 | 04:30 |
| Air Blank | | 0.000 | 04:30 |
| Diagnostics Check | | OK | 04:30 |

Data Download Date/Time: 04/21/2020 07:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CARPIN

DATE/TIME OF TEST: 03/01/2020 02:36:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:29 |
| Air Blank | | 0.000 | 02:29 |
| Control Test | | 0.080 | 02:30 |
| Air Blank | | 0.000 | 02:30 |
| Subject Sample | # 1 | 0.173 | 02:31 |
| Breath Volume | | 2.839 | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 2 | 0.169 | 02:34 |
| Breath Volume | | 2.972 | 02:34 |
| Air Blank | | 0.000 | 02:35 |
| Control Test | | 0.080 | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Diagnostics Check | | OK | 02:36 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CASTRO SANTIAGO

DATE/TIME OF TEST: 03/01/2020 03:00:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:54 |
| Air Blank | | 0.000 | 02:54 |
| Control Test | | 0.080 | 02:55 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 1 | 0.170 | 02:56 |
| Breath Volume | | 2.140 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 2 | 0.169 | 02:59 |
| Breath Volume | | 1.878 | 02:59 |
| Air Blank | | 0.000 | 02:59 |
| Control Test | | 0.080 | 03:00 |
| Air Blank | | 0.000 | 03:00 |
| Diagnostics Check | | OK | 03:00 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 03/02/2020 01:04:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:55 |
| Air Blank | | 0.000 | 00:56 |
| Control Test | | 0.081 | 00:56 |
| Air Blank | | 0.000 | 00:57 |
| Subject Sample | # 1 | 0.166 | 00:59 |
| Breath Volume | | 2.148 | 00:59 |
| Air Blank | | 0.000 | 00:59 |
| Air Blank | | 0.000 | 01:01 |
| Subject Sample | # 2 | 0.168 | 01:02 |
| Breath Volume | | 2.117 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Control Test | | 0.080 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Diagnostics Check | | OK | 01:04 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PEREZ CORTES

DATE/TIME OF TEST: 03/02/2020 02:19:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:11 | |
| Air Blank | | 0.000 | 02:11 | |
| Control Test | | 0.081 | 02:11 | |
| Air Blank | | 0.000 | 02:12 | |
| Subject Sample | # 1 | 0.127 | 02:15 | Volume Not Met |
| Breath Volume | | 0.800 | 02:15 | |
| Air Blank | | 0.000 | 02:16 | |
| Air Blank | | 0.000 | 02:18 | |
| Subject Sample | # 2 | | 02:18 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:18 | |
| Air Blank | | 0.000 | 02:18 | |
| Control Test | | 0.079 | 02:19 | |
| Air Blank | | 0.000 | 02:19 | |
| Diagnostics Check | | OK | 02:19 | |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: RODRIGUEZ POMPA

DATE/TIME OF TEST: 03/02/2020 21:53:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:46 |
| Air Blank | | 0.000 | 21:47 |
| Control Test | | 0.080 | 21:47 |
| Air Blank | | 0.000 | 21:48 |
| Subject Sample | # 1 | 0.273 | 21:49 |
| Breath Volume | | 2.167 | 21:49 |
| Air Blank | | 0.000 | 21:49 |
| Air Blank | | 0.000 | 21:51 |
| Subject Sample | # 2 | 0.284 | 21:52 |
| Breath Volume | | 1.781 | 21:52 |
| Air Blank | | 0.000 | 21:52 |
| Control Test | | 0.080 | 21:53 |
| Air Blank | | 0.000 | 21:53 |
| Diagnostics Check | | OK | 21:53 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 03/04/2020 03:58:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:52 |
| Air Blank | | 0.000 | 03:52 |
| Control Test | | 0.080 | 03:53 |
| Air Blank | | 0.000 | 03:53 |
| Subject Sample | # 1 | 0.229 | 03:53 |
| Breath Volume | | 1.277 | 03:53 |
| Air Blank | | 0.000 | 03:54 |
| Air Blank | | 0.000 | 03:56 |
| Subject Sample | # 2 | 0.234 | 03:56 |
| Breath Volume | | 1.847 | 03:56 |
| Air Blank | | 0.000 | 03:57 |
| Control Test | | 0.079 | 03:57 |
| Air Blank | | 0.000 | 03:58 |
| Diagnostics Check | | OK | 03:58 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUTAN,C
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CHISLEY

DATE/TIME OF TEST: 03/04/2020 04:33:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:25 |
| Air Blank | | 0.000 | 04:25 |
| Control Test | | 0.080 | 04:25 |
| Air Blank | | 0.000 | 04:26 |
| Subject Sample | # 1 | 0.158 | 04:28 |
| Breath Volume | | 1.691 | 04:28 |
| Air Blank | | 0.000 | 04:28 |
| Air Blank | | 0.000 | 04:30 |
| Subject Sample | # 2 | 0.152 | 04:31 |
| Breath Volume | | 1.445 | 04:31 |
| Air Blank | | 0.000 | 04:32 |
| Control Test | | 0.079 | 04:32 |
| Air Blank | | 0.000 | 04:33 |
| Diagnostics Check | | OK | 04:33 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 03/05/2020 23:53:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:42 | |
| Air Blank | | 0.000 | 23:42 | |
| Control Test | | 0.080 | 23:42 | |
| Air Blank | | 0.000 | 23:43 | |
| Subject Sample | # 1 | 0.140 | 23:46 | Volume Not Met |
| Breath Volume | | 0.359 | 23:46 | |
| Air Blank | | 0.000 | 23:47 | |
| Air Blank | | 0.000 | 23:48 | |
| Subject Sample | # 2 | 0.153 | 23:52 | Volume Not Met |
| Breath Volume | | 0.128 | 23:52 | |
| Air Blank | | 0.000 | 23:52 | |
| Control Test | | 0.079 | 23:53 | |
| Air Blank | | 0.000 | 23:53 | |
| Diagnostics Check | | OK | 23:53 | |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 03/06/2020 00:10:12

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:00 | |
| Air Blank | | 0.000 | 00:00 | |
| Control Test | | 0.079 | 00:01 | |
| Air Blank | | 0.000 | 00:01 | |
| Subject Sample | # 1 | 0.145 | 00:04 | Volume Not Met |
| Breath Volume | | 0.347 | 00:04 | |
| Air Blank | | 0.000 | 00:05 | |
| Air Blank | | 0.000 | 00:07 | |
| Subject Sample | # 2 | | 00:08 | Subject Test Refused |
| Breath Volume | | 0.312 | 00:08 | |
| Air Blank | | 0.000 | 00:09 | |
| Control Test | | 0.079 | 00:09 | |
| Air Blank | | 0.000 | 00:10 | |
| Diagnostics Check | | OK | 00:10 | |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 03/06/2020 22:20:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:14 |
| Air Blank | | 0.000 | 22:14 |
| Control Test | | 0.080 | 22:15 |
| Air Blank | | 0.000 | 22:15 |
| Subject Sample | # 1 | 0.000 | 22:16 |
| Breath Volume | | 3.269 | 22:16 |
| Air Blank | | 0.000 | 22:16 |
| Air Blank | | 0.000 | 22:18 |
| Subject Sample | # 2 | 0.000 | 22:19 |
| Breath Volume | | 3.093 | 22:19 |
| Air Blank | | 0.000 | 22:20 |
| Control Test | | 0.080 | 22:20 |
| Air Blank | | 0.000 | 22:20 |
| Diagnostics Check | | OK | 22:20 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: UNGERATHEN

DATE/TIME OF TEST: 03/07/2020 00:03:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:55 |
| Air Blank | | 0.000 | 23:56 |
| Control Test | | 0.079 | 23:56 |
| Air Blank | | 0.000 | 23:57 |
| Subject Sample | # 2 | 0.107 | 23:58 |
| Breath Volume | | 3.128 | 23:58 |
| Air Blank | | 0.000 | 23:59 |
| Air Blank | | 0.000 | 00:01 |
| Subject Sample | # 1 | 0.108 | 00:01 |
| Breath Volume | | 1.507 | 00:01 |
| Air Blank | | 0.000 | 00:02 |
| Control Test | | 0.080 | 00:02 |
| Air Blank | | 0.000 | 00:03 |
| Diagnostics Check | | OK | 00:03 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MICHAUD

DATE/TIME OF TEST: 03/07/2020 01:08:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:01 |
| Air Blank | | 0.000 | 01:02 |
| Control Test | | 0.080 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Subject Sample | # 1 | 0.070 | 01:03 |
| Breath Volume | | 2.410 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Air Blank | | 0.000 | 01:06 |
| Subject Sample | # 2 | 0.071 | 01:07 |
| Breath Volume | | 2.343 | 01:07 |
| Air Blank | | 0.000 | 01:07 |
| Control Test | | 0.081 | 01:07 |
| Air Blank | | 0.000 | 01:08 |
| Diagnostics Check | | OK | 01:08 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ASHMORE III

DATE/TIME OF TEST: 03/07/2020 03:54:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:48 |
| Air Blank | | 0.000 | 03:48 |
| Control Test | | 0.081 | 03:48 |
| Air Blank | | 0.000 | 03:49 |
| Subject Sample | # 1 | 0.240 | 03:50 |
| Breath Volume | | 1.289 | 03:50 |
| Air Blank | | 0.000 | 03:50 |
| Air Blank | | 0.000 | 03:52 |
| Subject Sample | # 2 | 0.245 | 03:53 |
| Breath Volume | | 1.855 | 03:53 |
| Air Blank | | 0.000 | 03:53 |
| Control Test | | 0.080 | 03:54 |
| Air Blank | | 0.000 | 03:54 |
| Diagnostics Check | | OK | 03:54 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DE LEON VELASQUEZ

DATE/TIME OF TEST: 03/08/2020 01:17:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Control Test | | 0.080 | 01:10 |
| Air Blank | | 0.000 | 01:11 |
| Subject Sample | # 1 | 0.195 | 01:12 |
| Breath Volume | | 1.562 | 01:12 |
| Air Blank | | 0.000 | 01:13 |
| Air Blank | | 0.000 | 01:14 |
| Subject Sample | # 2 | 0.192 | 01:15 |
| Breath Volume | | 1.507 | 01:15 |
| Air Blank | | 0.000 | 01:16 |
| Control Test | | 0.079 | 01:16 |
| Air Blank | | 0.000 | 01:17 |
| Diagnostics Check | | OK | 01:17 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MANNS

DATE/TIME OF TEST: 03/08/2020 04:16:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Control Test | | 0.080 | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Subject Sample | # 1 | 0.152 | 04:12 |
| Breath Volume | | 1.441 | 04:12 |
| Air Blank | | 0.000 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Subject Sample | # 2 | 0.149 | 04:15 |
| Breath Volume | | 1.398 | 04:15 |
| Air Blank | | 0.000 | 04:15 |
| Control Test | | 0.079 | 04:16 |
| Air Blank | | 0.000 | 04:16 |
| Diagnostics Check | | OK | 04:16 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 03/08/2020 04:45:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:37 |
| Air Blank | | 0.000 | 04:37 |
| Control Test | | 0.080 | 04:37 |
| Air Blank | | 0.000 | 04:38 |
| Subject Sample | # 1 | 0.200 | 04:40 |
| Breath Volume | | 1.664 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Air Blank | | 0.000 | 04:43 |
| Subject Sample | # 2 | 0.204 | 04:44 |
| Breath Volume | | 1.343 | 04:44 |
| Air Blank | | 0.000 | 04:44 |
| Control Test | | 0.079 | 04:45 |
| Air Blank | | 0.000 | 04:45 |
| Diagnostics Check | | OK | 04:45 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: STIMPHIL

DATE/TIME OF TEST: 03/08/2020 06:00:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:54 |
| Air Blank | | 0.000 | 05:55 |
| Control Test | | 0.079 | 05:55 |
| Air Blank | | 0.000 | 05:55 |
| Subject Sample | # 1 | 0.000 | 05:56 |
| Breath Volume | | 2.531 | 05:56 |
| Air Blank | | 0.000 | 05:56 |
| Air Blank | | 0.000 | 05:58 |
| Subject Sample | # 2 | 0.000 | 05:59 |
| Breath Volume | | 2.382 | 05:59 |
| Air Blank | | 0.000 | 05:59 |
| Control Test | | 0.080 | 05:59 |
| Air Blank | | 0.000 | 06:00 |
| Diagnostics Check | | OK | 06:00 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MAYSONET RODRIGUEZ

DATE/TIME OF TEST: 03/09/2020 04:27:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:19 |
| Air Blank | | 0.000 | 04:19 |
| Control Test | | 0.080 | 04:20 |
| Air Blank | | 0.000 | 04:20 |
| Subject Sample | # 1 | 0.169 | 04:22 |
| Breath Volume | | 1.933 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Air Blank | | 0.000 | 04:25 |
| Subject Sample | # 2 | 0.169 | 04:26 |
| Breath Volume | | 2.058 | 04:26 |
| Air Blank | | 0.000 | 04:26 |
| Control Test | | 0.079 | 04:27 |
| Air Blank | | 0.000 | 04:27 |
| Diagnostics Check | | OK | 04:27 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/09/2020 05:32:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:25 |
| Air Blank | | 0.000 | 05:26 |
| Control Test | | 0.080 | 05:26 |
| Air Blank | | 0.000 | 05:27 |
| Subject Sample | # 1 | 0.169 | 05:27 |
| Breath Volume | | 2.085 | 05:27 |
| Air Blank | | 0.000 | 05:28 |
| Air Blank | | 0.000 | 05:30 |
| Subject Sample | # 2 | 0.171 | 05:30 |
| Breath Volume | | 2.574 | 05:30 |
| Air Blank | | 0.000 | 05:31 |
| Control Test | | 0.079 | 05:31 |
| Air Blank | | 0.000 | 05:32 |
| Diagnostics Check | | OK | 05:32 |

Data Download Date/Time: 03/09/2020 18:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURROUGHS,H
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: FULLWOOD

DATE/TIME OF TEST: 03/11/2020 03:33:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:26 |
| Air Blank | | 0.000 | 03:26 |
| Control Test | | 0.080 | 03:27 |
| Air Blank | | 0.000 | 03:27 |
| Subject Sample | # 1 | 0.078 | 03:28 |
| Breath Volume | | 1.726 | 03:28 |
| Air Blank | | 0.000 | 03:29 |
| Air Blank | | 0.000 | 03:31 |
| Subject Sample | # 2 | 0.079 | 03:32 |
| Breath Volume | | 2.222 | 03:32 |
| Air Blank | | 0.000 | 03:32 |
| Control Test | | 0.080 | 03:33 |
| Air Blank | | 0.000 | 03:33 |
| Diagnostics Check | | OK | 03:33 |

Data Download Date/Time: 04/03/2020 14:02

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/11/2020 03:50:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:43 |
| Air Blank | | 0.000 | 03:44 |
| Control Test | | 0.080 | 03:44 |
| Air Blank | | 0.000 | 03:45 |
| Subject Sample | # 1 | 0.153 | 03:46 |
| Breath Volume | | 2.527 | 03:46 |
| Air Blank | | 0.000 | 03:46 |
| Air Blank | | 0.000 | 03:48 |
| Subject Sample | # 2 | 0.165 | 03:49 |
| Breath Volume | | 3.167 | 03:49 |
| Air Blank | | 0.000 | 03:49 |
| Control Test | | 0.080 | 03:50 |
| Air Blank | | 0.000 | 03:50 |
| Diagnostics Check | | OK | 03:50 |

Data Download Date/Time: 04/03/2020 14:02

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: HAYLES

DATE/TIME OF TEST: 03/12/2020 00:59:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:52 |
| Air Blank | | 0.000 | 00:53 |
| Control Test | | 0.080 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Subject Sample | # 1 | 0.186 | 00:55 |
| Breath Volume | | 2.738 | 00:55 |
| Air Blank | | 0.000 | 00:55 |
| Air Blank | | 0.000 | 00:57 |
| Subject Sample | # 2 | 0.177 | 00:58 |
| Breath Volume | | 2.917 | 00:58 |
| Air Blank | | 0.000 | 00:58 |
| Control Test | | 0.080 | 00:59 |
| Air Blank | | 0.000 | 00:59 |
| Diagnostics Check | | OK | 00:59 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JEREZ INFANTE

DATE/TIME OF TEST: 03/12/2020 02:39:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:32 |
| Air Blank | | 0.000 | 02:33 |
| Control Test | | 0.080 | 02:33 |
| Air Blank | | 0.000 | 02:33 |
| Subject Sample | # 1 | 0.000 | 02:35 |
| Breath Volume | | 2.187 | 02:35 |
| Air Blank | | 0.000 | 02:35 |
| Air Blank | | 0.000 | 02:37 |
| Subject Sample | # 2 | 0.000 | 02:38 |
| Breath Volume | | 1.160 | 02:38 |
| Air Blank | | 0.000 | 02:38 |
| Control Test | | 0.079 | 02:38 |
| Air Blank | | 0.000 | 02:39 |
| Diagnostics Check | | OK | 02:39 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIERMEK,T
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: BEST JR

DATE/TIME OF TEST: 03/13/2020 01:11:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:04 |
| Air Blank | | 0.000 | 01:04 |
| Control Test | | 0.080 | 01:04 |
| Air Blank | | 0.000 | 01:05 |
| Subject Sample | # 1 | 0.158 | 01:06 |
| Breath Volume | | 2.878 | 01:06 |
| Air Blank | | 0.000 | 01:06 |
| Air Blank | | 0.000 | 01:08 |
| Subject Sample | # 2 | 0.169 | 01:10 |
| Breath Volume | | 3.109 | 01:10 |
| Air Blank | | 0.000 | 01:10 |
| Control Test | | 0.080 | 01:11 |
| Air Blank | | 0.000 | 01:11 |
| Diagnostics Check | | OK | 01:11 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SKEENS II

DATE/TIME OF TEST: 03/13/2020 02:00:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Control Test | | 0.080 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Subject Sample | # 1 | 0.180 | 01:55 |
| Breath Volume | | 3.738 | 01:55 |
| Air Blank | | 0.000 | 01:56 |
| Air Blank | | 0.000 | 01:57 |
| Subject Sample | # 2 | 0.178 | 01:58 |
| Breath Volume | | 2.328 | 01:58 |
| Air Blank | | 0.000 | 01:59 |
| Control Test | | 0.079 | 01:59 |
| Air Blank | | 0.000 | 01:59 |
| Diagnostics Check | | OK | 02:00 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MARTIN JR

DATE/TIME OF TEST: 03/13/2020 03:14:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Control Test | | 0.080 | 03:08 |
| Air Blank | | 0.000 | 03:09 |
| Subject Sample | # 1 | 0.063 | 03:09 |
| Breath Volume | | 2.417 | 03:09 |
| Air Blank | | 0.000 | 03:10 |
| Air Blank | | 0.000 | 03:12 |
| Subject Sample | # 2 | 0.061 | 03:12 |
| Breath Volume | | 2.167 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.079 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Diagnostics Check | | OK | 03:14 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NELSEN,G
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DYKES

DATE/TIME OF TEST: 03/13/2020 21:24:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:18 |
| Air Blank | | 0.000 | 21:18 |
| Control Test | | 0.078 | 21:19 |
| Air Blank | | 0.000 | 21:19 |
| Subject Sample | # 1 | 0.239 | 21:20 |
| Breath Volume | | 1.347 | 21:20 |
| Air Blank | | 0.000 | 21:20 |
| Air Blank | | 0.000 | 21:22 |
| Subject Sample | # 2 | 0.227 | 21:23 |
| Breath Volume | | 2.171 | 21:23 |
| Air Blank | | 0.000 | 21:23 |
| Control Test | | 0.080 | 21:24 |
| Air Blank | | 0.000 | 21:24 |
| Diagnostics Check | | OK | 21:24 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GUESS

DATE/TIME OF TEST: 03/14/2020 03:39:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:32 |
| Air Blank | | 0.000 | 03:33 |
| Control Test | | 0.079 | 03:33 |
| Air Blank | | 0.000 | 03:34 |
| Subject Sample | # 1 | 0.061 | 03:34 |
| Breath Volume | | 1.335 | 03:34 |
| Air Blank | | 0.000 | 03:35 |
| Air Blank | | 0.000 | 03:37 |
| Subject Sample | # 2 | 0.061 | 03:37 |
| Breath Volume | | 1.359 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Control Test | | 0.078 | 03:38 |
| Air Blank | | 0.000 | 03:39 |
| Diagnostics Check | | OK | 03:39 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CROSSMAN II

DATE/TIME OF TEST: 03/14/2020 04:35:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:25 |
| Air Blank | | 0.000 | 04:26 |
| Control Test | | 0.080 | 04:26 |
| Air Blank | | 0.000 | 04:27 |
| Subject Sample | # 1 | 0.176 | 04:28 |
| Breath Volume | | 1.339 | 04:28 |
| Air Blank | | 0.000 | 04:29 |
| Air Blank | | 0.000 | 04:31 |
| Subject Sample | # 2 | 0.166 | 04:33 |
| Breath Volume | | 1.671 | 04:33 |
| Air Blank | | 0.000 | 04:34 |
| Control Test | | 0.078 | 04:35 |
| Air Blank | | 0.000 | 04:35 |
| Diagnostics Check | | OK | 04:35 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 03/15/2020 03:30:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:23 |
| Air Blank | | 0.000 | 03:24 |
| Control Test | | 0.080 | 03:24 |
| Air Blank | | 0.000 | 03:25 |
| Subject Sample | # 1 | 0.152 | 03:25 |
| Breath Volume | | 2.523 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Air Blank | | 0.000 | 03:28 |
| Subject Sample | # 2 | 0.148 | 03:28 |
| Breath Volume | | 2.316 | 03:28 |
| Air Blank | | 0.000 | 03:29 |
| Control Test | | 0.080 | 03:29 |
| Air Blank | | 0.000 | 03:30 |
| Diagnostics Check | | OK | 03:30 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 03/15/2020 04:45:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:37 |
| Air Blank | | 0.000 | 04:37 |
| Control Test | | 0.080 | 04:38 |
| Air Blank | | 0.000 | 04:38 |
| Subject Sample | # 1 | 0.119 | 04:40 |
| Breath Volume | | 1.214 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Air Blank | | 0.000 | 04:43 |
| Subject Sample | # 2 | 0.129 | 04:43 |
| Breath Volume | | 1.195 | 04:43 |
| Air Blank | | 0.000 | 04:44 |
| Control Test | | 0.079 | 04:44 |
| Air Blank | | 0.000 | 04:45 |
| Diagnostics Check | | OK | 04:45 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: LLOYD JR

DATE/TIME OF TEST: 03/15/2020 23:23:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:17 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.080 | 23:18 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 1 | 0.116 | 23:19 |
| Breath Volume | | 1.609 | 23:19 |
| Air Blank | | 0.000 | 23:20 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 2 | 0.114 | 23:22 |
| Breath Volume | | 2.589 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Control Test | | 0.080 | 23:23 |
| Air Blank | | 0.000 | 23:23 |
| Diagnostics Check | | OK | 23:23 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MCKINLEY

DATE/TIME OF TEST: 03/17/2020 20:22:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:15 |
| Air Blank | | 0.000 | 20:15 |
| Control Test | | 0.080 | 20:16 |
| Air Blank | | 0.000 | 20:16 |
| Subject Sample | # 1 | 0.364 | 20:17 |
| Breath Volume | | 1.367 | 20:17 |
| Air Blank | | 0.000 | 20:18 |
| Air Blank | | 0.000 | 20:20 |
| Subject Sample | # 2 | 0.359 | 20:20 |
| Breath Volume | | 2.187 | 20:20 |
| Air Blank | | 0.000 | 20:21 |
| Control Test | | 0.079 | 20:21 |
| Air Blank | | 0.000 | 20:21 |
| Diagnostics Check | | OK | 20:22 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/26/2020 20:32:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:23 |
| Air Blank | | 0.000 | 20:24 |
| Control Test | | 0.079 | 20:24 |
| Air Blank | | 0.000 | 20:25 |
| Subject Sample | # 1 | 0.124 | 20:27 |
| Breath Volume | | 1.761 | 20:27 |
| Air Blank | | 0.000 | 20:27 |
| Air Blank | | 0.000 | 20:29 |
| Subject Sample | # 2 | 0.120 | 20:30 |
| Breath Volume | | 1.316 | 20:30 |
| Air Blank | | 0.000 | 20:31 |
| Control Test | | 0.078 | 20:31 |
| Air Blank | | 0.000 | 20:32 |
| Diagnostics Check | | OK | 20:32 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNES AVERY,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 03/28/2020 00:43:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Control Test | | 0.079 | 00:38 |
| Air Blank | | 0.000 | 00:39 |
| Subject Sample | # 1 | 0.069 | 00:39 |
| Breath Volume | | 1.980 | 00:39 |
| Air Blank | | 0.000 | 00:40 |
| Air Blank | | 0.000 | 00:42 |
| Subject Sample | # 2 | 0.067 | 00:42 |
| Breath Volume | | 2.546 | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Control Test | | 0.080 | 00:43 |
| Air Blank | | 0.000 | 00:43 |
| Diagnostics Check | | OK | 00:43 |

Data Download Date/Time: 04/03/2020 14:03

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: OSVAR

DATE/TIME OF TEST: 03/01/2020 05:09:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:02 |
| Air Blank | | 0.000 | 05:03 |
| Control Test | | 0.081 | 05:03 |
| Air Blank | | 0.000 | 05:04 |
| Subject Sample | # 1 | 0.040 | 05:04 |
| Breath Volume | | 1.117 | 05:04 |
| Air Blank | | 0.000 | 05:05 |
| Air Blank | | 0.000 | 05:07 |
| Subject Sample | # 2 | 0.036 | 05:07 |
| Breath Volume | | 1.468 | 05:07 |
| Air Blank | | 0.000 | 05:08 |
| Control Test | | 0.079 | 05:08 |
| Air Blank | | 0.000 | 05:09 |
| Diagnostics Check | | OK | 05:09 |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MEGGINSON

DATE/TIME OF TEST: 03/03/2020 03:50:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:41 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.080 | 03:42 |
| Air Blank | | 0.000 | 03:42 |
| Subject Sample | # 1 | 0.176 | 03:45 |
| Breath Volume | | 2.429 | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Air Blank | | 0.000 | 03:48 |
| Subject Sample | # 2 | 0.172 | 03:49 |
| Breath Volume | | 3.289 | 03:49 |
| Air Blank | | 0.000 | 03:49 |
| Control Test | | 0.078 | 03:50 |
| Air Blank | | 0.000 | 03:50 |
| Diagnostics Check | | OK | 03:50 |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TAYLOR,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/04/2020 00:55:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:46 |
| Air Blank | | 0.000 | 00:46 |
| Control Test | | 0.080 | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Subject Sample | # 1 | 0.131 | 00:49 |
| Breath Volume | | 1.800 | 00:49 |
| Air Blank | | 0.000 | 00:50 |
| Air Blank | | 0.000 | 00:52 |
| Subject Sample | # 2 | 0.122 | 00:53 |
| Breath Volume | | 1.679 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Control Test | | 0.075 | 00:54 |
| Air Blank | | 0.000 | 00:55 |
| Diagnostics Check | | OK | 00:55 |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONADIE,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BONADIE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DEWITT

DATE/TIME OF TEST: 03/05/2020 23:30:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:24 |
| Air Blank | | 0.000 | 23:24 |
| Control Test | | 0.080 | 23:24 |
| Air Blank | | 0.000 | 23:25 |
| Subject Sample | # 1 | 0.214 | 23:25 |
| Breath Volume | | 2.187 | 23:25 |
| Air Blank | | 0.000 | 23:26 |
| Air Blank | | 0.000 | 23:28 |
| Subject Sample | # 2 | 0.234 | 23:28 |
| Breath Volume | | 3.414 | 23:28 |
| Air Blank | | 0.000 | 23:29 |
| Control Test | | 0.078 | 23:29 |
| Air Blank | | 0.000 | 23:30 |
| Diagnostics Check | | OK | 23:30 |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHERR,E
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: PENWELL

DATE/TIME OF TEST: 03/06/2020 21:53:37

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:50 | |
| Air Blank | | 0.000 | 21:50 | |
| Control Test | | 0.080 | 21:50 | |
| Air Blank | | 0.000 | 21:51 | |
| Subject Sample | # 1 | | 21:52 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:52 | |
| Air Blank | | 0.000 | 21:52 | |
| Control Test | | 0.079 | 21:52 | |
| Air Blank | | 0.000 | 21:53 | |
| Diagnostics Check | | OK | 21:53 | |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DALRYMPLE,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 03/06/2020 22:45:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:42 | |
| Air Blank | | 0.000 | 22:43 | |
| Control Test | | 0.079 | 22:43 | |
| Air Blank | | 0.000 | 22:44 | |
| Subject Sample | # 1 | | 22:44 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:44 | |
| Air Blank | | 0.000 | 22:45 | |
| Control Test | | 0.079 | 22:45 | |
| Air Blank | | 0.000 | 22:45 | |
| Diagnostics Check | | OK | 22:45 | |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 03/07/2020 22:48:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:41 |
| Air Blank | | 0.000 | 22:41 |
| Control Test | | 0.081 | 22:41 |
| Air Blank | | 0.000 | 22:42 |
| Subject Sample | # 1 | 0.195 | 22:44 |
| Breath Volume | | 1.457 | 22:44 |
| Air Blank | | 0.000 | 22:44 |
| Air Blank | | 0.000 | 22:46 |
| Subject Sample | # 2 | 0.196 | 22:47 |
| Breath Volume | | 1.484 | 22:47 |
| Air Blank | | 0.000 | 22:47 |
| Control Test | | 0.079 | 22:48 |
| Air Blank | | 0.000 | 22:48 |
| Diagnostics Check | | OK | 22:48 |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CASH

DATE/TIME OF TEST: 03/13/2020 00:39:04

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:36 | |
| Air Blank | | 0.000 | 00:36 | |
| Control Test | | 0.079 | 00:36 | |
| Air Blank | | 0.000 | 00:37 | |
| Subject Sample | # 1 | | 00:37 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:37 | |
| Air Blank | | 0.000 | 00:38 | |
| Control Test | | 0.079 | 00:38 | |
| Air Blank | | 0.000 | 00:38 | |
| Diagnostics Check | | OK | 00:39 | |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWING,BRANDON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 03/20/2020 02:37:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:34 | |
| Air Blank | | 0.000 | 02:34 | |
| Control Test | | 0.080 | 02:35 | |
| Air Blank | | 0.000 | 02:35 | |
| Subject Sample | # 1 | | 02:35 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:35 | |
| Air Blank | | 0.000 | 02:36 | |
| Control Test | | 0.080 | 02:36 | |
| Air Blank | | 0.000 | 02:37 | |
| Diagnostics Check | | OK | 02:37 | |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHMIDT,JONATHAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 03/26/2020 23:23:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.080 | 23:17 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 1 | 0.051 | 23:19 |
| Breath Volume | | 3.726 | 23:19 |
| Air Blank | | 0.000 | 23:19 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 2 | 0.050 | 23:22 |
| Breath Volume | | 2.156 | 23:22 |
| Air Blank | | 0.000 | 23:22 |
| Control Test | | 0.079 | 23:23 |
| Air Blank | | 0.000 | 23:23 |
| Diagnostics Check | | OK | 23:23 |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GORDON,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/31/2020 01:15:38

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:12 | |
| Air Blank | | 0.000 | 01:13 | |
| Control Test | | 0.079 | 01:13 | |
| Air Blank | | 0.000 | 01:13 | |
| Subject Sample | # 1 | | 01:14 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:14 | |
| Air Blank | | 0.000 | 01:14 | |
| Control Test | | 0.079 | 01:14 | |
| Air Blank | | 0.000 | 01:15 | |
| Diagnostics Check | | OK | 01:15 | |

Data Download Date/Time: 03/30/2020 10:47

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SORIA,KEITH
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CRICKENBERGER

DATE/TIME OF TEST: 03/05/2020 00:26:54

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:24 | |
| Air Blank | | 0.000 | 00:24 | |
| Control Test | | 0.080 | 00:24 | |
| Air Blank | | 0.000 | 00:25 | |
| Subject Sample | # 1 | | 00:25 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:25 | |
| Air Blank | | 0.000 | 00:25 | |
| Control Test | | 0.080 | 00:26 | |
| Air Blank | | 0.000 | 00:26 | |
| Diagnostics Check | | OK | 00:26 | |

Data Download Date/Time: 03/30/2020 16:46

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RAFALKO,TRAVIS
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 03/13/2020 13:51:12

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 13:48 | |
| Air Blank | | 0.000 | 13:48 | |
| Control Test | | 0.079 | 13:49 | |
| Air Blank | | 0.000 | 13:49 | |
| Subject Sample | # 1 | | 13:49 | Subject Test Refused |
| Breath Volume | | 0.000 | 13:49 | |
| Air Blank | | 0.000 | 13:50 | |
| Control Test | | 0.080 | 13:50 | |
| Air Blank | | 0.000 | 13:51 | |
| Diagnostics Check | | OK | 13:51 | |

Data Download Date/Time: 03/30/2020 16:46

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CALHOUN,KEVIN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 03/14/2020 02:57:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:51 |
| Air Blank | | 0.000 | 02:51 |
| Control Test | | 0.080 | 02:51 |
| Air Blank | | 0.000 | 02:52 |
| Subject Sample | # 1 | 0.244 | 02:53 |
| Breath Volume | | 2.667 | 02:53 |
| Air Blank | | 0.000 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 2 | 0.245 | 02:55 |
| Breath Volume | | 1.582 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.077 | 02:57 |
| Air Blank | | 0.000 | 02:57 |
| Diagnostics Check | | OK | 02:57 |

Data Download Date/Time: 03/30/2020 16:46

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1791908A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: RODRIGUEZ MIRANDA

DATE/TIME OF TEST: 03/15/2020 04:54:34

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 04:53 | |
| Air Blank | | 04:54 | RFI Detect |
| Air Blank | 0.000 | 04:54 | |

Data Download Date/Time: 03/29/2020 00:53

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1791908A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: RODRIGUEZ MIRANDA

DATE/TIME OF TEST: 03/15/2020 05:02:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:56 |
| Air Blank | | 0.000 | 04:56 |
| Control Test | | 0.080 | 04:57 |
| Air Blank | | 0.000 | 04:57 |
| Subject Sample | # 1 | 0.120 | 04:58 |
| Breath Volume | | 2.300 | 04:58 |
| Air Blank | | 0.000 | 04:58 |
| Air Blank | | 0.000 | 05:00 |
| Subject Sample | # 2 | 0.112 | 05:01 |
| Breath Volume | | 1.546 | 05:01 |
| Air Blank | | 0.000 | 05:01 |
| Control Test | | 0.079 | 05:02 |
| Air Blank | | 0.000 | 05:02 |
| Diagnostics Check | | OK | 05:02 |

Data Download Date/Time: 03/29/2020 00:53

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MACKLIN

DATE/TIME OF TEST: 03/05/2020 02:55:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:49 |
| Air Blank | | 0.000 | 02:49 |
| Control Test | | 0.079 | 02:50 |
| Air Blank | | 0.000 | 02:50 |
| Subject Sample | # 1 | 0.008 | 02:51 |
| Breath Volume | | 2.996 | 02:51 |
| Air Blank | | 0.000 | 02:51 |
| Air Blank | | 0.000 | 02:53 |
| Subject Sample | # 2 | 0.007 | 02:54 |
| Breath Volume | | 2.648 | 02:54 |
| Air Blank | | 0.000 | 02:54 |
| Control Test | | 0.079 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Diagnostics Check | | OK | 02:55 |

Data Download Date/Time: 03/25/2020 10:07

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAAVILAINEN,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAAVILAINEN,JOSEPH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ESCOBAR VILLALTA

DATE/TIME OF TEST: 03/06/2020 00:42:16

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:36 | |
| Air Blank | | 0.000 | 00:36 | |
| Control Test | | 0.079 | 00:37 | |
| Air Blank | | 0.000 | 00:37 | |
| Subject Sample | # 1 | 0.178 | 00:40 | |
| Breath Volume | | 2.234 | 00:40 | |
| Air Blank | | | 00:40 | RFI Detect |
| Air Blank | | 0.000 | 00:41 | |
| Control Test | | 0.080 | 00:41 | |
| Air Blank | | 0.000 | 00:42 | |
| Diagnostics Check | | OK | 00:42 | |

Data Download Date/Time: 03/25/2020 10:07

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAAVILAINEN,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAAVILAINEN,JOSEPH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ESCOBAR VILLALTA

DATE/TIME OF TEST: 03/06/2020 00:50:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:44 |
| Air Blank | | 0.000 | 00:44 |
| Control Test | | 0.078 | 00:45 |
| Air Blank | | 0.000 | 00:45 |
| Subject Sample | # 1 | 0.177 | 00:46 |
| Breath Volume | | 1.468 | 00:46 |
| Air Blank | | 0.000 | 00:46 |
| Air Blank | | 0.000 | 00:48 |
| Subject Sample | # 2 | 0.179 | 00:49 |
| Breath Volume | | 1.277 | 00:49 |
| Air Blank | | 0.000 | 00:49 |
| Control Test | | 0.078 | 00:50 |
| Air Blank | | 0.000 | 00:50 |
| Diagnostics Check | | OK | 00:50 |

Data Download Date/Time: 03/25/2020 10:07

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUMMINGS,GARY,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CUMMINGS,GARY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 03/07/2020 00:47:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:41 |
| Air Blank | | 0.000 | 00:41 |
| Control Test | | 0.080 | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Subject Sample | # 1 | 0.062 | 00:42 |
| Breath Volume | | 1.441 | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Air Blank | | 0.000 | 00:45 |
| Subject Sample | # 2 | 0.059 | 00:45 |
| Breath Volume | | 1.476 | 00:45 |
| Air Blank | | 0.000 | 00:46 |
| Control Test | | 0.080 | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Diagnostics Check | | OK | 00:47 |

Data Download Date/Time: 03/25/2020 10:07

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUMMINGS,GARY,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RUBI,JONATHAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 03/07/2020 01:07:02

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:04 | |
| Air Blank | | 0.000 | 01:04 | |
| Control Test | | 0.079 | 01:04 | |
| Air Blank | | 0.000 | 01:05 | |
| Subject Sample | # 1 | | 01:05 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:05 | |
| Air Blank | | 0.000 | 01:06 | |
| Control Test | | 0.080 | 01:06 | |
| Air Blank | | 0.000 | 01:06 | |
| Diagnostics Check | | OK | 01:07 | |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RIVERO RAMOS

DATE/TIME OF TEST: 03/09/2020 02:59:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:50 | |
| Air Blank | | 0.000 | 02:51 | |
| Control Test | | 0.080 | 02:51 | |
| Air Blank | | 0.000 | 02:52 | |
| Subject Sample | # 1 | | 02:55 | No Sample Provided |
| Breath Volume | | 0.000 | 02:55 | |
| Air Blank | | 0.000 | 02:55 | |
| Air Blank | | 0.000 | 02:57 | |
| Subject Sample | # 2 | | 02:58 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:58 | |
| Air Blank | | 0.000 | 02:58 | |
| Control Test | | 0.080 | 02:58 | |
| Air Blank | | 0.000 | 02:59 | |
| Diagnostics Check | | OK | 02:59 | |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 03/10/2020 21:47:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:41 |
| Air Blank | | 0.000 | 21:41 |
| Control Test | | 0.080 | 21:42 |
| Air Blank | | 0.000 | 21:42 |
| Subject Sample | # 1 | 0.000 | 21:43 |
| Breath Volume | | 3.929 | 21:43 |
| Air Blank | | 0.000 | 21:43 |
| Air Blank | | 0.000 | 21:45 |
| Subject Sample | # 2 | 0.000 | 21:46 |
| Breath Volume | | 1.492 | 21:46 |
| Air Blank | | 0.000 | 21:46 |
| Control Test | | 0.080 | 21:47 |
| Air Blank | | 0.000 | 21:47 |
| Diagnostics Check | | OK | 21:47 |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAAVILAINEN,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAAVILAINEN,JOSEPH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ROMAGNINO

DATE/TIME OF TEST: 03/11/2020 00:02:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:55 |
| Air Blank | | 0.000 | 23:56 |
| Control Test | | 0.080 | 23:56 |
| Air Blank | | 0.000 | 23:56 |
| Subject Sample | # 2 | 0.200 | 23:57 |
| Breath Volume | | 3.453 | 23:57 |
| Air Blank | | 0.000 | 23:58 |
| Air Blank | | 0.000 | 23:59 |
| Subject Sample | # 1 | 0.191 | 00:00 |
| Breath Volume | | 2.964 | 00:00 |
| Air Blank | | 0.000 | 00:01 |
| Control Test | | 0.079 | 00:01 |
| Air Blank | | 0.000 | 00:02 |
| Diagnostics Check | | OK | 00:02 |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: BERLIN

DATE/TIME OF TEST: 03/12/2020 16:35:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:26 |
| Air Blank | | 0.000 | 16:27 |
| Control Test | | 0.079 | 16:27 |
| Air Blank | | 0.000 | 16:27 |
| Subject Sample | # 1 | 0.185 | 16:29 |
| Breath Volume | | 3.625 | 16:29 |
| Air Blank | | 0.000 | 16:30 |
| Air Blank | | 0.000 | 16:31 |
| Subject Sample | # 2 | 0.173 | 16:33 |
| Breath Volume | | 2.535 | 16:33 |
| Air Blank | | 0.000 | 16:34 |
| Control Test | | 0.079 | 16:34 |
| Air Blank | | 0.000 | 16:35 |
| Diagnostics Check | | OK | 16:35 |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 03/13/2020 04:03:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:56 |
| Air Blank | | 0.000 | 03:57 |
| Control Test | | 0.079 | 03:57 |
| Air Blank | | 0.000 | 03:58 |
| Subject Sample | # 1 | 0.139 | 03:58 |
| Breath Volume | | 2.890 | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Air Blank | | 0.000 | 04:01 |
| Subject Sample | # 2 | 0.143 | 04:01 |
| Breath Volume | | 3.160 | 04:01 |
| Air Blank | | 0.000 | 04:02 |
| Control Test | | 0.078 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Diagnostics Check | | OK | 04:03 |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,E,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANTANA,E
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/13/2020 15:05:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 14:58 |
| Air Blank | | 0.000 | 14:58 |
| Control Test | | 0.078 | 14:58 |
| Air Blank | | 0.000 | 14:59 |
| Subject Sample | # 1 | 0.180 | 15:00 |
| Breath Volume | | 2.070 | 15:00 |
| Air Blank | | 0.000 | 15:01 |
| Air Blank | | 0.000 | 15:03 |
| Subject Sample | # 2 | 0.167 | 15:03 |
| Breath Volume | | 1.847 | 15:03 |
| Air Blank | | 0.000 | 15:04 |
| Control Test | | 0.078 | 15:04 |
| Air Blank | | 0.000 | 15:05 |
| Diagnostics Check | | OK | 15:05 |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/18/2020 00:58:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:51 |
| Air Blank | | 0.000 | 00:52 |
| Control Test | | 0.079 | 00:52 |
| Air Blank | | 0.000 | 00:53 |
| Subject Sample | # 1 | 0.210 | 00:53 |
| Breath Volume | | 3.203 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Air Blank | | 0.000 | 00:56 |
| Subject Sample | # 2 | 0.216 | 00:56 |
| Breath Volume | | 2.816 | 00:56 |
| Air Blank | | 0.000 | 00:57 |
| Control Test | | 0.078 | 00:57 |
| Air Blank | | 0.000 | 00:58 |
| Diagnostics Check | | OK | 00:58 |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 962587
DRY GAS EXPIRY DATE: 06/04/2021
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/20/2020 01:31:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:25 |
| Air Blank | | 0.000 | 01:25 |
| Control Test | | 0.080 | 01:25 |
| Air Blank | | 0.000 | 01:26 |
| Subject Sample | # 1 | 0.148 | 01:27 |
| Breath Volume | | 1.632 | 01:27 |
| Air Blank | | 0.000 | 01:27 |
| Air Blank | | 0.000 | 01:29 |
| Subject Sample | # 2 | 0.155 | 01:29 |
| Breath Volume | | 1.648 | 01:29 |
| Air Blank | | 0.000 | 01:30 |
| Control Test | | 0.078 | 01:30 |
| Air Blank | | 0.000 | 01:31 |
| Diagnostics Check | | OK | 01:31 |

Data Download Date/Time: 03/25/2020 10:08

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 03/07/2020 03:42:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:36 |
| Air Blank | | 0.000 | 03:36 |
| Control Test | | 0.081 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Subject Sample | # 1 | 0.136 | 03:38 |
| Breath Volume | | 1.148 | 03:38 |
| Air Blank | | 0.000 | 03:38 |
| Air Blank | | 0.000 | 03:40 |
| Subject Sample | # 2 | 0.130 | 03:41 |
| Breath Volume | | 1.242 | 03:41 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.082 | 03:42 |
| Air Blank | | 0.000 | 03:42 |
| Diagnostics Check | | OK | 03:42 |

Data Download Date/Time: 03/29/2020 01:19

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,BLAKE,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CARPENTER,BLAKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ALMANSOUR

DATE/TIME OF TEST: 03/08/2020 04:42:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:34 |
| Air Blank | | 0.000 | 04:34 |
| Control Test | | 0.080 | 04:34 |
| Air Blank | | 0.000 | 04:35 |
| Subject Sample | # 1 | 0.117 | 04:37 |
| Breath Volume | | 3.609 | 04:37 |
| Air Blank | | 0.000 | 04:38 |
| Air Blank | | 0.000 | 04:40 |
| Subject Sample | # 2 | 0.115 | 04:40 |
| Breath Volume | | 3.632 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Control Test | | 0.079 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Diagnostics Check | | OK | 04:42 |

Data Download Date/Time: 03/29/2020 01:19

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: MERCADO RIVERA

DATE/TIME OF TEST: 03/13/2020 22:31:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:25 |
| Air Blank | | 0.000 | 22:25 |
| Control Test | | 0.079 | 22:26 |
| Air Blank | | 0.000 | 22:26 |
| Subject Sample | # 1 | 0.000 | 22:27 |
| Breath Volume | | 3.078 | 22:27 |
| Air Blank | | 0.000 | 22:27 |
| Air Blank | | 0.000 | 22:29 |
| Subject Sample | # 2 | 0.000 | 22:30 |
| Breath Volume | | 3.324 | 22:30 |
| Air Blank | | 0.000 | 22:30 |
| Control Test | | 0.081 | 22:30 |
| Air Blank | | 0.000 | 22:31 |
| Diagnostics Check | | OK | 22:31 |

Data Download Date/Time: 03/29/2020 01:19

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/21/2020 05:05:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:58 |
| Air Blank | | 0.000 | 04:58 |
| Control Test | | 0.079 | 04:59 |
| Air Blank | | 0.000 | 04:59 |
| Subject Sample | # 1 | 0.221 | 05:00 |
| Breath Volume | | 1.171 | 05:00 |
| Air Blank | | 0.000 | 05:01 |
| Air Blank | | 0.000 | 05:02 |
| Subject Sample | # 2 | 0.223 | 05:04 |
| Breath Volume | | 1.730 | 05:04 |
| Air Blank | | 0.000 | 05:04 |
| Control Test | | 0.078 | 05:05 |
| Air Blank | | 0.000 | 05:05 |
| Diagnostics Check | | OK | 05:05 |

Data Download Date/Time: 03/29/2020 01:19

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,VICTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17618080A3
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 03/25/2020 22:29:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:18 |
| Air Blank | | 0.000 | 22:19 |
| Control Test | | 0.078 | 22:19 |
| Air Blank | | 0.000 | 22:20 |
| Subject Sample | # 1 | 0.247 | 22:22 |
| Breath Volume | | 1.695 | 22:22 |
| Air Blank | | 0.000 | 22:23 |
| Air Blank | | 0.000 | 22:24 |
| Subject Sample | # 2 | 0.224 | 22:25 |
| Breath Volume | | 1.589 | 22:25 |
| Air Blank | | 0.000 | 22:25 |
| Air Blank | | 0.000 | 22:27 |
| Subject Sample | # 3 | 0.240 | 22:28 |
| Breath Volume | | 1.988 | 22:28 |
| Air Blank | | 0.000 | 22:28 |
| Control Test | | 0.078 | 22:29 |
| Air Blank | | 0.000 | 22:29 |
| Diagnostics Check | | OK | 22:29 |

Data Download Date/Time: 03/29/2020 01:19

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: DORWARD

DATE/TIME OF TEST: 03/01/2020 05:03:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:55 |
| Air Blank | | 0.000 | 04:55 |
| Control Test | | 0.079 | 04:55 |
| Air Blank | | 0.000 | 04:56 |
| Subject Sample | # 1 | 0.185 | 04:56 |
| Breath Volume | | 1.617 | 04:56 |
| Air Blank | | 0.000 | 04:57 |
| Air Blank | | 0.000 | 04:59 |
| Subject Sample | # 2 | 0.187 | 05:01 |
| Breath Volume | | 2.335 | 05:01 |
| Air Blank | | 0.000 | 05:02 |
| Control Test | | 0.077 | 05:02 |
| Air Blank | | 0.000 | 05:03 |
| Diagnostics Check | | OK | 05:03 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DONOHUE

DATE/TIME OF TEST: 03/01/2020 21:23:45

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:20 | |
| Air Blank | | 0.000 | 21:21 | |
| Control Test | | 0.080 | 21:21 | |
| Air Blank | | 0.000 | 21:22 | |
| Subject Sample | # 1 | | 21:22 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:22 | |
| Air Blank | | 0.000 | 21:22 | |
| Control Test | | 0.079 | 21:23 | |
| Air Blank | | 0.000 | 21:23 | |
| Diagnostics Check | | OK | 21:23 | |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LORENZO LUCAS

DATE/TIME OF TEST: 03/02/2020 03:42:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:35 |
| Air Blank | | 0.000 | 03:36 |
| Control Test | | 0.079 | 03:36 |
| Air Blank | | 0.000 | 03:36 |
| Subject Sample | # 1 | 0.170 | 03:37 |
| Breath Volume | | 1.636 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Air Blank | | 0.000 | 03:39 |
| Subject Sample | # 2 | 0.168 | 03:40 |
| Breath Volume | | 2.113 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.079 | 03:41 |
| Air Blank | | 0.000 | 03:42 |
| Diagnostics Check | | OK | 03:42 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAEUSSLER,BRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HAEUSSLER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 03/02/2020 03:56:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:47 |
| Air Blank | | 0.000 | 03:48 |
| Control Test | | 0.080 | 03:48 |
| Air Blank | | 0.000 | 03:48 |
| Subject Sample | # 1 | 0.000 | 03:51 |
| Breath Volume | | 2.222 | 03:51 |
| Air Blank | | 0.000 | 03:52 |
| Air Blank | | 0.000 | 03:54 |
| Subject Sample | # 2 | 0.000 | 03:54 |
| Breath Volume | | 2.703 | 03:54 |
| Air Blank | | 0.000 | 03:55 |
| Control Test | | 0.079 | 03:55 |
| Air Blank | | 0.000 | 03:56 |
| Diagnostics Check | | OK | 03:56 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 03/03/2020 02:37:49

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:33 | |
| Air Blank | | 0.000 | 02:33 | |
| Control Test | | 0.079 | 02:34 | |
| Air Blank | | 0.000 | 02:34 | |
| Subject Sample | # 1 | | 02:36 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:36 | |
| Air Blank | | 0.000 | 02:36 | |
| Control Test | | 0.079 | 02:37 | |
| Air Blank | | 0.000 | 02:37 | |
| Diagnostics Check | | OK | 02:37 | |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VERICELLA

DATE/TIME OF TEST: 03/07/2020 00:58:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:51 |
| Air Blank | | 0.000 | 00:51 |
| Control Test | | 0.080 | 00:51 |
| Air Blank | | 0.000 | 00:52 |
| Subject Sample | # 1 | 0.108 | 00:53 |
| Breath Volume | | 1.695 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Air Blank | | 0.000 | 00:55 |
| Subject Sample | # 2 | 0.121 | 00:56 |
| Breath Volume | | 3.285 | 00:56 |
| Air Blank | | 0.000 | 00:57 |
| Control Test | | 0.080 | 00:57 |
| Air Blank | | 0.000 | 00:58 |
| Diagnostics Check | | OK | 00:58 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FLORES OLVERA

DATE/TIME OF TEST: 03/09/2020 04:07:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:00 |
| Air Blank | | 0.000 | 04:00 |
| Control Test | | 0.079 | 04:01 |
| Air Blank | | 0.000 | 04:01 |
| Subject Sample | # 1 | 0.169 | 04:02 |
| Breath Volume | | 2.101 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Subject Sample | # 2 | 0.162 | 04:05 |
| Breath Volume | | 1.976 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Control Test | | 0.079 | 04:06 |
| Air Blank | | 0.000 | 04:06 |
| Diagnostics Check | | OK | 04:07 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,R
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ARANGO

DATE/TIME OF TEST: 03/11/2020 03:53:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Control Test | | 0.079 | 03:47 |
| Air Blank | | 0.000 | 03:48 |
| Subject Sample | # 1 | 0.065 | 03:48 |
| Breath Volume | | 3.136 | 03:48 |
| Air Blank | | 0.000 | 03:49 |
| Air Blank | | 0.000 | 03:51 |
| Subject Sample | # 2 | 0.061 | 03:51 |
| Breath Volume | | 4.039 | 03:51 |
| Air Blank | | 0.000 | 03:52 |
| Control Test | | 0.077 | 03:52 |
| Air Blank | | 0.000 | 03:53 |
| Diagnostics Check | | OK | 03:53 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BLANK

DATE/TIME OF TEST: 03/13/2020 00:26:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:19 |
| Air Blank | | 0.000 | 00:20 |
| Control Test | | 0.080 | 00:20 |
| Air Blank | | 0.000 | 00:21 |
| Subject Sample | # 1 | 0.115 | 00:21 |
| Breath Volume | | 3.222 | 00:21 |
| Air Blank | | 0.000 | 00:22 |
| Air Blank | | 0.000 | 00:24 |
| Subject Sample | # 2 | 0.109 | 00:24 |
| Breath Volume | | 3.800 | 00:24 |
| Air Blank | | 0.000 | 00:25 |
| Control Test | | 0.080 | 00:25 |
| Air Blank | | 0.000 | 00:26 |
| Diagnostics Check | | OK | 00:26 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCHETT,JASON,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: FORESTER

DATE/TIME OF TEST: 03/13/2020 01:16:32

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:00 | |
| Air Blank | | 0.000 | 01:01 | |
| Control Test | | 0.079 | 01:01 | |
| Air Blank | | 0.000 | 01:02 | |
| Subject Sample | # 1 | 0.242 | 01:05 | Volume Not Met |
| Breath Volume | | 0.179 | 01:05 | |
| Air Blank | | 0.000 | 01:06 | |
| Air Blank | | 0.000 | 01:07 | |
| Subject Sample | # 2 | 0.246 | 01:09 | |
| Breath Volume | | 1.421 | 01:09 | |
| Air Blank | | 0.000 | 01:10 | |
| Air Blank | | 0.000 | 01:11 | |
| Subject Sample | # 3 | 0.219 | 01:14 | Volume Not Met |
| Breath Volume | | 1.046 | 01:14 | |
| Air Blank | | 0.000 | 01:15 | |
| Control Test | | 0.065 | 01:15 | Control Outside Tolerance |
| Air Blank | | 0.000 | 01:16 | |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: LANHAM

DATE/TIME OF TEST: 03/14/2020 01:05:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:01 | |
| Air Blank | | 0.000 | 01:02 | |
| Control Test | | 0.079 | 01:02 | |
| Air Blank | | 0.000 | 01:03 | |
| Subject Sample | # 1 | | 01:03 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:03 | |
| Air Blank | | 0.000 | 01:04 | |
| Control Test | | 0.079 | 01:04 | |
| Air Blank | | 0.000 | 01:05 | |
| Diagnostics Check | | OK | 01:05 | |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAEUSSLER,BRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HAEUSSLER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: DORMEVIL

DATE/TIME OF TEST: 03/15/2020 05:42:17

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:39 | |
| Air Blank | | 0.000 | 05:39 | |
| Control Test | | 0.080 | 05:40 | |
| Air Blank | | 0.000 | 05:40 | |
| Subject Sample | # 1 | | 05:40 | Subject Test Refused |
| Breath Volume | | 0.000 | 05:40 | |
| Air Blank | | 0.000 | 05:41 | |
| Control Test | | 0.079 | 05:41 | |
| Air Blank | | 0.000 | 05:42 | |
| Diagnostics Check | | OK | 05:42 | |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: BUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: LAVIOLETTE III

DATE/TIME OF TEST: 03/15/2020 19:50:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:44 |
| Air Blank | | 0.000 | 19:44 |
| Control Test | | 0.079 | 19:45 |
| Air Blank | | 0.000 | 19:45 |
| Subject Sample | # 1 | 0.055 | 19:46 |
| Breath Volume | | 2.457 | 19:46 |
| Air Blank | | 0.000 | 19:46 |
| Air Blank | | 0.000 | 19:48 |
| Subject Sample | # 2 | 0.054 | 19:49 |
| Breath Volume | | 2.230 | 19:49 |
| Air Blank | | 0.000 | 19:49 |
| Control Test | | 0.079 | 19:50 |
| Air Blank | | 0.000 | 19:50 |
| Diagnostics Check | | OK | 19:50 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: VAUGHN,RYAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 03/15/2020 22:04:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:57 |
| Air Blank | | 0.000 | 21:57 |
| Control Test | | 0.080 | 21:58 |
| Air Blank | | 0.000 | 21:58 |
| Subject Sample | # 1 | 0.299 | 21:59 |
| Breath Volume | | 3.410 | 21:59 |
| Air Blank | | 0.000 | 22:00 |
| Air Blank | | 0.000 | 22:01 |
| Subject Sample | # 2 | 0.296 | 22:02 |
| Breath Volume | | 3.007 | 22:02 |
| Air Blank | | 0.000 | 22:03 |
| Control Test | | 0.079 | 22:03 |
| Air Blank | | 0.000 | 22:04 |
| Diagnostics Check | | OK | 22:04 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 03/18/2020 03:01:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Control Test | | 0.079 | 02:55 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 1 | 0.138 | 02:56 |
| Breath Volume | | 3.984 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 2 | 0.121 | 02:59 |
| Breath Volume | | 1.718 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Control Test | | 0.076 | 03:00 |
| Air Blank | | 0.000 | 03:01 |
| Diagnostics Check | | OK | 03:01 |

Data Download Date/Time: 03/25/2020 15:40

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SYRE

DATE/TIME OF TEST: 03/23/2020 00:17:16

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 00:16 | |
| Air Blank | | 00:16 | Ambient Fail |
| Air Blank | 0.000 | 00:17 | |

Data Download Date/Time: 03/27/2020 09:32


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SYRE

DATE/TIME OF TEST: 03/23/2020 00:28:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:21 |
| Air Blank | | 0.000 | 00:21 |
| Control Test | | 0.081 | 00:22 |
| Air Blank | | 0.000 | 00:22 |
| Subject Sample | # 1 | 0.266 | 00:23 |
| Breath Volume | | 3.417 | 00:23 |
| Air Blank | | 0.000 | 00:24 |
| Air Blank | | 0.000 | 00:26 |
| Subject Sample | # 2 | 0.261 | 00:26 |
| Breath Volume | | 3.148 | 00:26 |
| Air Blank | | 0.000 | 00:27 |
| Control Test | | 0.080 | 00:28 |
| Air Blank | | 0.000 | 00:28 |
| Diagnostics Check | | OK | 00:28 |

Data Download Date/Time: 03/27/2020 09:32

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/25/2020 21:31:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:25 | |
| Air Blank | | 0.000 | 21:25 | |
| Control Test | | 0.080 | 21:25 | |
| Air Blank | | 0.000 | 21:26 | |
| Subject Sample | # 1 | 0.125 | 21:27 | |
| Breath Volume | | 2.402 | 21:27 | |
| Air Blank | | 0.000 | 21:27 | |
| Air Blank | | 0.000 | 21:29 | |
| Subject Sample | # 2 | | 21:30 | Interferent Detect |
| Breath Volume | | 2.093 | 21:30 | |
| Air Blank | | 0.000 | 21:30 | |
| Control Test | | 0.080 | 21:31 | |
| Air Blank | | 0.000 | 21:31 | |
| Diagnostics Check | | OK | 21:31 | |

Data Download Date/Time: 03/27/2020 09:32

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/25/2020 21:39:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:33 |
| Air Blank | | 0.000 | 21:33 |
| Control Test | | 0.079 | 21:33 |
| Air Blank | | 0.000 | 21:34 |
| Subject Sample | # 1 | 0.116 | 21:34 |
| Breath Volume | | 2.726 | 21:34 |
| Air Blank | | 0.000 | 21:35 |
| Air Blank | | 0.000 | 21:37 |
| Subject Sample | # 2 | 0.112 | 21:37 |
| Breath Volume | | 2.191 | 21:37 |
| Air Blank | | 0.000 | 21:38 |
| Control Test | | 0.080 | 21:38 |
| Air Blank | | 0.000 | 21:39 |
| Diagnostics Check | | OK | 21:39 |

Data Download Date/Time: 03/27/2020 09:32

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 03/04/2020 20:07:38

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:59 | |
| Air Blank | | 0.000 | 19:59 | |
| Control Test | | 0.077 | 19:59 | |
| Air Blank | | 0.000 | 20:00 | |
| Subject Sample | # 1 | | 20:03 | No Sample Provided |
| Breath Volume | | 0.000 | 20:03 | |
| Air Blank | | 0.000 | 20:04 | |
| Air Blank | | 0.000 | 20:06 | |
| Subject Sample | # 2 | | 20:06 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:06 | |
| Air Blank | | 0.000 | 20:06 | |
| Control Test | | 0.078 | 20:07 | |
| Air Blank | | 0.000 | 20:07 | |
| Diagnostics Check | | OK | 20:07 | |

Data Download Date/Time: 04/06/2020 01:44

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: LARSEN

DATE/TIME OF TEST: 03/12/2020 00:36:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:33 | |
| Air Blank | | 0.000 | 00:33 | |
| Control Test | | 0.080 | 00:34 | |
| Air Blank | | 0.000 | 00:34 | |
| Subject Sample | # 1 | | 00:35 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:35 | |
| Air Blank | | 0.000 | 00:35 | |
| Control Test | | 0.080 | 00:36 | |
| Air Blank | | 0.000 | 00:36 | |
| Diagnostics Check | | OK | 00:36 | |

Data Download Date/Time: 04/06/2020 01:44

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: LEAGUE

DATE/TIME OF TEST: 03/22/2020 22:24:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:15 |
| Air Blank | | 0.000 | 22:15 |
| Control Test | | 0.079 | 22:16 |
| Air Blank | | 0.000 | 22:16 |
| Subject Sample | # 1 | 0.223 | 22:18 |
| Breath Volume | | 1.535 | 22:18 |
| Air Blank | | 0.000 | 22:19 |
| Air Blank | | 0.000 | 22:20 |
| Subject Sample | # 2 | 0.209 | 22:22 |
| Breath Volume | | 1.195 | 22:22 |
| Air Blank | | 0.000 | 22:23 |
| Control Test | | 0.077 | 22:23 |
| Air Blank | | 0.000 | 22:24 |
| Diagnostics Check | | OK | 22:24 |

Data Download Date/Time: 04/06/2020 01:44

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/03/2020 00:01:08

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:55 | |
| Air Blank | | 0.000 | 23:55 | |
| Control Test | | 0.077 | 23:56 | |
| Air Blank | | 0.000 | 23:56 | |
| Subject Sample | # 2 | 0.000 | 23:57 | |
| Breath Volume | | 2.886 | 23:57 | |
| Air Blank | | 0.000 | 23:57 | |
| Air Blank | | 0.000 | 23:59 | |
| Subject Sample | # 1 | 0.000 | 00:00 | |
| Breath Volume | | 2.433 | 00:00 | |
| Air Blank | | 0.000 | 00:00 | |
| | | | 00:00 | RFI Detect |
| Air Blank | | 0.000 | 00:01 | |

Data Download Date/Time: 03/11/2020 18:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: KOVACH

DATE/TIME OF TEST: 03/07/2020 20:02:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:55 |
| Air Blank | | 0.000 | 19:56 |
| Control Test | | 0.078 | 19:56 |
| Air Blank | | 0.000 | 19:56 |
| Subject Sample | # 1 | 0.124 | 19:57 |
| Breath Volume | | 2.468 | 19:57 |
| Air Blank | | 0.000 | 19:58 |
| Air Blank | | 0.000 | 20:00 |
| Subject Sample | # 2 | 0.126 | 20:00 |
| Breath Volume | | 2.492 | 20:00 |
| Air Blank | | 0.000 | 20:01 |
| Control Test | | 0.077 | 20:01 |
| Air Blank | | 0.000 | 20:02 |
| Diagnostics Check | | OK | 20:02 |

Data Download Date/Time: 03/11/2020 18:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRIGGS,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 03/08/2020 03:10:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Control Test | | 0.078 | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Subject Sample | # 1 | 0.244 | 03:06 |
| Breath Volume | | 3.542 | 03:06 |
| Air Blank | | 0.000 | 03:06 |
| Air Blank | | 0.000 | 03:08 |
| Subject Sample | # 2 | 0.258 | 03:09 |
| Breath Volume | | 3.105 | 03:09 |
| Air Blank | | 0.000 | 03:10 |
| Control Test | | 0.081 | 03:10 |
| Air Blank | | 0.000 | 03:10 |
| Diagnostics Check | | OK | 03:10 |

Data Download Date/Time: 03/11/2020 18:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ARENA

DATE/TIME OF TEST: 03/08/2020 21:47:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:40 |
| Air Blank | | 0.000 | 21:41 |
| Control Test | | 0.079 | 21:41 |
| Air Blank | | 0.000 | 21:41 |
| Subject Sample | # 1 | 0.082 | 21:42 |
| Breath Volume | | 3.468 | 21:42 |
| Air Blank | | 0.000 | 21:43 |
| Air Blank | | 0.000 | 21:45 |
| Subject Sample | # 2 | 0.076 | 21:45 |
| Breath Volume | | 2.488 | 21:45 |
| Air Blank | | 0.000 | 21:46 |
| Control Test | | 0.078 | 21:46 |
| Air Blank | | 0.000 | 21:46 |
| Diagnostics Check | | OK | 21:47 |

Data Download Date/Time: 03/11/2020 19:04

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 03/17/2020 23:41:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:35 |
| Air Blank | | 0.000 | 23:35 |
| Control Test | | 0.075 | 23:36 |
| Air Blank | | 0.000 | 23:36 |
| Subject Sample | # 1 | 0.046 | 23:37 |
| Breath Volume | | 2.601 | 23:37 |
| Air Blank | | 0.000 | 23:37 |
| Air Blank | | 0.000 | 23:39 |
| Subject Sample | # 2 | 0.047 | 23:39 |
| Breath Volume | | 2.609 | 23:39 |
| Air Blank | | 0.000 | 23:40 |
| Control Test | | 0.075 | 23:40 |
| Air Blank | | 0.000 | 23:41 |
| Diagnostics Check | | OK | 23:41 |

Data Download Date/Time: 04/13/2020 20:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 03/20/2020 22:03:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:57 |
| Air Blank | | 0.000 | 21:58 |
| Control Test | | 0.077 | 21:58 |
| Air Blank | | 0.000 | 21:58 |
| Subject Sample | # 1 | 0.136 | 21:59 |
| Breath Volume | | 3.445 | 21:59 |
| Air Blank | | 0.000 | 22:00 |
| Air Blank | | 0.000 | 22:01 |
| Subject Sample | # 2 | 0.136 | 22:02 |
| Breath Volume | | 3.480 | 22:02 |
| Air Blank | | 0.000 | 22:02 |
| Control Test | | 0.076 | 22:03 |
| Air Blank | | 0.000 | 22:03 |
| Diagnostics Check | | OK | 22:03 |

Data Download Date/Time: 04/13/2020 20:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: LINDSAY

DATE/TIME OF TEST: 03/14/2020 01:53:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:46 |
| Air Blank | | 0.000 | 01:47 |
| Control Test | | 0.080 | 01:47 |
| Air Blank | | 0.000 | 01:47 |
| Subject Sample | # 1 | 0.303 | 01:48 |
| Breath Volume | | 2.527 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Air Blank | | 0.000 | 01:51 |
| Subject Sample | # 2 | 0.308 | 01:51 |
| Breath Volume | | 2.652 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Control Test | | 0.079 | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Diagnostics Check | | OK | 01:53 |

Data Download Date/Time: 03/25/2020 21:05

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/27/2020 23:04:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:57 |
| Air Blank | | 0.000 | 22:58 |
| Control Test | | 0.080 | 22:58 |
| Air Blank | | 0.000 | 22:59 |
| Subject Sample | # 1 | 0.246 | 22:59 |
| Breath Volume | | 1.234 | 22:59 |
| Air Blank | | 0.000 | 23:00 |
| Air Blank | | 0.000 | 23:02 |
| Subject Sample | # 2 | 0.241 | 23:03 |
| Breath Volume | | 1.343 | 23:03 |
| Air Blank | | 0.000 | 23:03 |
| Control Test | | 0.079 | 23:04 |
| Air Blank | | 0.000 | 23:04 |
| Diagnostics Check | | OK | 23:04 |

Data Download Date/Time: 04/04/2020 22:24

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/14/2020 01:07:19

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:57 | |
| Air Blank | | 0.000 | 00:58 | |
| Control Test | | 0.079 | 00:58 | |
| Air Blank | | 0.000 | 00:59 | |
| Subject Sample | # 1 | | 01:02 | No Sample Provided |
| Breath Volume | | 0.000 | 01:02 | |
| Air Blank | | 0.000 | 01:02 | |
| Air Blank | | 0.000 | 01:04 | |
| Subject Sample | # 2 | | 01:05 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:05 | |
| Air Blank | | 0.000 | 01:06 | |
| Control Test | | 0.080 | 01:06 | |
| Air Blank | | | 01:06 | RFI Detect |
| Air Blank | | 0.000 | 01:07 | |

Data Download Date/Time: 03/21/2020 13:26


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: NADAL

DATE/TIME OF TEST: 03/01/2020 01:58:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:52 |
| Air Blank | | 0.000 | 01:52 |
| Control Test | | 0.081 | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Subject Sample | # 1 | 0.242 | 01:53 |
| Breath Volume | | 2.121 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Air Blank | | 0.000 | 01:56 |
| Subject Sample | # 2 | 0.244 | 01:57 |
| Breath Volume | | 1.531 | 01:57 |
| Air Blank | | 0.000 | 01:57 |
| Control Test | | 0.081 | 01:58 |
| Air Blank | | 0.000 | 01:58 |
| Diagnostics Check | | OK | 01:58 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BIANCHI,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 03/04/2020 01:48:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:38 |
| Air Blank | | 0.000 | 01:39 |
| Control Test | | 0.078 | 01:39 |
| Air Blank | | 0.000 | 01:40 |
| Subject Sample | # 1 | 0.225 | 01:43 |
| Breath Volume | | 1.687 | 01:43 |
| Air Blank | | 0.000 | 01:44 |
| Air Blank | | 0.000 | 01:45 |
| Subject Sample | # 2 | 0.221 | 01:46 |
| Breath Volume | | 2.031 | 01:46 |
| Air Blank | | 0.000 | 01:47 |
| Control Test | | 0.077 | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Diagnostics Check | | OK | 01:48 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 03/07/2020 03:51:09

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:35 | |
| Air Blank | | 0.000 | 03:35 | |
| Control Test | | 0.080 | 03:35 | |
| Air Blank | | 0.000 | 03:36 | |
| Subject Sample | # 1 | 0.136 | 03:39 | Volume Not Met |
| Breath Volume | | 0.699 | 03:39 | |
| Air Blank | | 0.000 | 03:40 | |
| Air Blank | | 0.000 | 03:41 | |
| Subject Sample | # 2 | 0.175 | 03:43 | |
| Breath Volume | | 1.300 | 03:43 | |
| Air Blank | | 0.000 | 03:44 | |
| Air Blank | | 0.000 | 03:46 | |
| Subject Sample | # 3 | 0.152 | 03:49 | Volume Not Met |
| Breath Volume | | 0.039 | 03:49 | |
| Air Blank | | 0.000 | 03:50 | |
| Control Test | | 0.080 | 03:50 | |
| Air Blank | | 0.000 | 03:51 | |
| Diagnostics Check | | OK | 03:51 | |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 03/07/2020 04:03:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:54 |
| Air Blank | | 0.000 | 03:54 |
| Control Test | | 0.080 | 03:55 |
| Air Blank | | 0.000 | 03:55 |
| Subject Sample | # 1 | 0.164 | 03:57 |
| Breath Volume | | 1.496 | 03:57 |
| Air Blank | | 0.000 | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Subject Sample | # 2 | 0.167 | 04:01 |
| Breath Volume | | 1.304 | 04:01 |
| Air Blank | | 0.000 | 04:02 |
| Control Test | | 0.079 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Diagnostics Check | | OK | 04:03 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CAPOTE

DATE/TIME OF TEST: 03/12/2020 02:52:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:45 |
| Air Blank | | 0.000 | 02:45 |
| Control Test | | 0.079 | 02:46 |
| Air Blank | | 0.000 | 02:46 |
| Subject Sample | # 1 | 0.202 | 02:47 |
| Breath Volume | | 1.691 | 02:47 |
| Air Blank | | 0.000 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Subject Sample | # 2 | 0.202 | 02:50 |
| Breath Volume | | 1.972 | 02:50 |
| Air Blank | | 0.000 | 02:51 |
| Control Test | | 0.079 | 02:51 |
| Air Blank | | 0.000 | 02:52 |
| Diagnostics Check | | OK | 02:52 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/12/2020 04:00:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:57 | |
| Air Blank | | 0.000 | 03:58 | |
| Control Test | | 0.079 | 03:58 | |
| Air Blank | | 0.000 | 03:58 | |
| Subject Sample | # 1 | | 03:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:59 | |
| Air Blank | | 0.000 | 03:59 | |
| Control Test | | 0.077 | 03:59 | |
| Air Blank | | 0.000 | 04:00 | |
| Diagnostics Check | | OK | 04:00 | |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/13/2020 02:33:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:27 |
| Air Blank | | 0.000 | 02:27 |
| Control Test | | 0.079 | 02:28 |
| Air Blank | | 0.000 | 02:28 |
| Subject Sample | # 1 | 0.118 | 02:29 |
| Breath Volume | | 2.472 | 02:29 |
| Air Blank | | 0.000 | 02:29 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 2 | 0.118 | 02:32 |
| Breath Volume | | 2.632 | 02:32 |
| Air Blank | | 0.000 | 02:32 |
| Control Test | | 0.080 | 02:33 |
| Air Blank | | 0.000 | 02:33 |
| Diagnostics Check | | OK | 02:33 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: RODRIGUEZ RAMIREZ

DATE/TIME OF TEST: 03/13/2020 06:47:58

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 06:45 | |
| Air Blank | | 0.000 | 06:45 | |
| Control Test | | 0.080 | 06:45 | |
| Air Blank | | 0.000 | 06:46 | |
| Subject Sample | # 1 | | 06:46 | Subject Test Refused |
| Breath Volume | | 0.000 | 06:46 | |
| Air Blank | | 0.000 | 06:47 | |
| Control Test | | 0.078 | 06:47 | |
| Air Blank | | 0.000 | 06:47 | |
| Diagnostics Check | | OK | 06:47 | |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: RODRIGUEZ RAMIREZ

DATE/TIME OF TEST: 03/13/2020 06:58:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:50 |
| Air Blank | | 0.000 | 06:50 |
| Control Test | | 0.079 | 06:51 |
| Air Blank | | 0.000 | 06:51 |
| Subject Sample | # 1 | 0.189 | 06:52 |
| Breath Volume | | 1.792 | 06:52 |
| Air Blank | | 0.000 | 06:53 |
| Air Blank | | 0.000 | 06:55 |
| Subject Sample | # 2 | 0.182 | 06:56 |
| Breath Volume | | 1.542 | 06:56 |
| Air Blank | | 0.000 | 06:57 |
| Control Test | | 0.079 | 06:57 |
| Air Blank | | 0.000 | 06:58 |
| Diagnostics Check | | OK | 06:58 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/14/2020 04:27:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:24 | |
| Air Blank | | 0.000 | 04:25 | |
| Control Test | | 0.079 | 04:25 | |
| Air Blank | | 0.000 | 04:25 | |
| Subject Sample | # 1 | | 04:26 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:26 | |
| Air Blank | | 0.000 | 04:26 | |
| Control Test | | 0.077 | 04:26 | |
| Air Blank | | 0.000 | 04:27 | |
| Diagnostics Check | | OK | 04:27 | |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 03/15/2020 04:18:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:11 |
| Air Blank | | 0.000 | 04:11 |
| Control Test | | 0.078 | 04:12 |
| Air Blank | | 0.000 | 04:12 |
| Subject Sample | # 1 | 0.065 | 04:13 |
| Breath Volume | | 1.437 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Air Blank | | 0.000 | 04:16 |
| Subject Sample | # 2 | 0.064 | 04:17 |
| Breath Volume | | 2.253 | 04:17 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.079 | 04:18 |
| Air Blank | | 0.000 | 04:18 |
| Diagnostics Check | | OK | 04:18 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: RODRIGUEZ LOPEZ

DATE/TIME OF TEST: 03/15/2020 05:27:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:21 |
| Air Blank | | 0.000 | 05:21 |
| Control Test | | 0.078 | 05:21 |
| Air Blank | | 0.000 | 05:22 |
| Subject Sample | # 1 | 0.210 | 05:23 |
| Breath Volume | | 2.226 | 05:23 |
| Air Blank | | 0.000 | 05:23 |
| Air Blank | | 0.000 | 05:25 |
| Subject Sample | # 2 | 0.209 | 05:26 |
| Breath Volume | | 2.148 | 05:26 |
| Air Blank | | 0.000 | 05:26 |
| Control Test | | 0.079 | 05:27 |
| Air Blank | | 0.000 | 05:27 |
| Diagnostics Check | | OK | 05:27 |

Data Download Date/Time: 04/29/2020 16:35

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAMZE,NEZAR
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 09:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: CASTEEL

DATE/TIME OF TEST: 03/01/2020 11:15:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:09 |
| Air Blank | | 0.000 | 11:09 |
| Control Test | | 0.081 | 11:09 |
| Air Blank | | 0.000 | 11:10 |
| Subject Sample | # 1 | 0.051 | 11:11 |
| Breath Volume | | 2.558 | 11:11 |
| Air Blank | | 0.000 | 11:11 |
| Air Blank | | 0.000 | 11:13 |
| Subject Sample | # 2 | 0.048 | 11:14 |
| Breath Volume | | 2.992 | 11:14 |
| Air Blank | | 0.000 | 11:14 |
| Control Test | | 0.081 | 11:15 |
| Air Blank | | 0.000 | 11:15 |
| Diagnostics Check | | OK | 11:15 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CLARKE,RAKEM
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 03/02/2020 23:18:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:11 |
| Air Blank | | 0.000 | 23:11 |
| Control Test | | 0.080 | 23:12 |
| Air Blank | | 0.000 | 23:12 |
| Subject Sample | # 1 | 0.144 | 23:13 |
| Breath Volume | | 2.359 | 23:13 |
| Air Blank | | 0.000 | 23:13 |
| Air Blank | | 0.000 | 23:15 |
| Subject Sample | # 2 | 0.134 | 23:16 |
| Breath Volume | | 1.628 | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.080 | 23:17 |
| Air Blank | | 0.000 | 23:18 |
| Diagnostics Check | | OK | 23:18 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 03/02/2020 23:54:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:48 |
| Air Blank | | 0.000 | 23:48 |
| Control Test | | 0.079 | 23:49 |
| Air Blank | | 0.000 | 23:49 |
| Subject Sample | # 1 | 0.127 | 23:50 |
| Breath Volume | | 2.945 | 23:50 |
| Air Blank | | 0.000 | 23:50 |
| Air Blank | | 0.000 | 23:52 |
| Subject Sample | # 2 | 0.126 | 23:53 |
| Breath Volume | | 3.003 | 23:53 |
| Air Blank | | 0.000 | 23:53 |
| Control Test | | 0.080 | 23:54 |
| Air Blank | | 0.000 | 23:54 |
| Diagnostics Check | | OK | 23:54 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CARRIER

DATE/TIME OF TEST: 03/03/2020 01:49:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:43 |
| Air Blank | | 0.000 | 01:43 |
| Control Test | | 0.080 | 01:43 |
| Air Blank | | 0.000 | 01:44 |
| Subject Sample | # 1 | 0.137 | 01:45 |
| Breath Volume | | 1.542 | 01:45 |
| Air Blank | | 0.000 | 01:45 |
| Air Blank | | 0.000 | 01:47 |
| Subject Sample | # 2 | 0.122 | 01:48 |
| Breath Volume | | 1.468 | 01:48 |
| Air Blank | | 0.000 | 01:48 |
| Control Test | | 0.078 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Diagnostics Check | | OK | 01:49 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 03/03/2020 22:34:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:28 |
| Air Blank | | 0.000 | 22:28 |
| Control Test | | 0.080 | 22:29 |
| Air Blank | | 0.000 | 22:29 |
| Subject Sample | # 1 | 0.000 | 22:30 |
| Breath Volume | | 2.175 | 22:30 |
| Air Blank | | 0.000 | 22:30 |
| Air Blank | | 0.000 | 22:32 |
| Subject Sample | # 2 | 0.000 | 22:33 |
| Breath Volume | | 2.789 | 22:33 |
| Air Blank | | 0.000 | 22:33 |
| Control Test | | 0.079 | 22:34 |
| Air Blank | | 0.000 | 22:34 |
| Diagnostics Check | | OK | 22:34 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLIGAN,D
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:43

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/06/2020 09:20:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 09:14 |
| Air Blank | | 0.000 | 09:14 |
| Control Test | | 0.081 | 09:14 |
| Air Blank | | 0.000 | 09:15 |
| Subject Sample | # 1 | 0.101 | 09:15 |
| Breath Volume | | 1.757 | 09:15 |
| Air Blank | | 0.000 | 09:16 |
| Air Blank | | 0.000 | 09:18 |
| Subject Sample | # 2 | 0.103 | 09:18 |
| Breath Volume | | 2.199 | 09:18 |
| Air Blank | | 0.000 | 09:19 |
| Control Test | | 0.080 | 09:19 |
| Air Blank | | 0.000 | 09:20 |
| Diagnostics Check | | OK | 09:20 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGRO,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 03/07/2020 23:39:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:32 |
| Air Blank | | 0.000 | 23:33 |
| Control Test | | 0.079 | 23:33 |
| Air Blank | | 0.000 | 23:34 |
| Subject Sample | # 1 | 0.109 | 23:34 |
| Breath Volume | | 2.687 | 23:34 |
| Air Blank | | 0.000 | 23:35 |
| Air Blank | | 0.000 | 23:37 |
| Subject Sample | # 2 | 0.115 | 23:37 |
| Breath Volume | | 2.406 | 23:37 |
| Air Blank | | 0.000 | 23:38 |
| Control Test | | 0.079 | 23:38 |
| Air Blank | | 0.000 | 23:39 |
| Diagnostics Check | | OK | 23:39 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/10/2020 04:42:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:35 |
| Air Blank | | 0.000 | 04:36 |
| Control Test | | 0.080 | 04:36 |
| Air Blank | | 0.000 | 04:37 |
| Subject Sample | # 1 | 0.231 | 04:37 |
| Breath Volume | | 1.820 | 04:37 |
| Air Blank | | 0.000 | 04:38 |
| Air Blank | | 0.000 | 04:40 |
| Subject Sample | # 2 | 0.228 | 04:40 |
| Breath Volume | | 2.218 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Control Test | | 0.079 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Diagnostics Check | | OK | 04:42 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: VICTOR

DATE/TIME OF TEST: 03/11/2020 02:48:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Control Test | | 0.081 | 02:38 |
| Air Blank | | 0.000 | 02:38 |
| Subject Sample | # 1 | 0.167 | 02:40 |
| Breath Volume | | 1.140 | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Subject Sample | # 2 | 0.189 | 02:43 |
| Breath Volume | | 2.542 | 02:43 |
| Air Blank | | 0.000 | 02:44 |
| Air Blank | | 0.000 | 02:45 |
| Subject Sample | # 3 | 0.184 | 02:46 |
| Breath Volume | | 2.601 | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Control Test | | 0.080 | 02:47 |
| Air Blank | | 0.000 | 02:47 |
| Diagnostics Check | | OK | 02:48 |

Data Download Date/Time: 03/11/2020 12:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: COURT ORD
ARREST TIME: 14:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/11/2020 15:06:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:00 |
| Air Blank | | 0.000 | 15:00 |
| Control Test | | 0.080 | 15:01 |
| Air Blank | | 0.000 | 15:01 |
| Subject Sample | # 1 | 0.000 | 15:02 |
| Breath Volume | | 2.105 | 15:02 |
| Air Blank | | 0.000 | 15:02 |
| Air Blank | | 0.000 | 15:04 |
| Subject Sample | # 2 | 0.000 | 15:04 |
| Breath Volume | | 2.156 | 15:04 |
| Air Blank | | 0.000 | 15:05 |
| Control Test | | 0.080 | 15:05 |
| Air Blank | | 0.000 | 15:06 |
| Diagnostics Check | | OK | 15:06 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RIBEIRO-SANTAMARINA

DATE/TIME OF TEST: 03/11/2020 23:52:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:45 |
| Air Blank | | 0.000 | 23:46 |
| Control Test | | 0.080 | 23:46 |
| Air Blank | | 0.000 | 23:47 |
| Subject Sample | # 1 | 0.109 | 23:47 |
| Breath Volume | | 2.472 | 23:47 |
| Air Blank | | 0.000 | 23:48 |
| Air Blank | | 0.000 | 23:50 |
| Subject Sample | # 2 | 0.113 | 23:50 |
| Breath Volume | | 2.304 | 23:50 |
| Air Blank | | 0.000 | 23:51 |
| Control Test | | 0.080 | 23:51 |
| Air Blank | | 0.000 | 23:52 |
| Diagnostics Check | | OK | 23:52 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/12/2020 02:06:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:00 |
| Air Blank | | 0.000 | 02:00 |
| Control Test | | 0.080 | 02:00 |
| Air Blank | | 0.000 | 02:01 |
| Subject Sample | # 1 | 0.205 | 02:02 |
| Breath Volume | | 1.921 | 02:02 |
| Air Blank | | 0.000 | 02:02 |
| Air Blank | | 0.000 | 02:04 |
| Subject Sample | # 2 | 0.196 | 02:05 |
| Breath Volume | | 2.105 | 02:05 |
| Air Blank | | 0.000 | 02:05 |
| Control Test | | 0.079 | 02:06 |
| Air Blank | | 0.000 | 02:06 |
| Diagnostics Check | | OK | 02:06 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOAYZA,RICARDO
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: OCON GONZALEZ

DATE/TIME OF TEST: 03/13/2020 00:17:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:11 |
| Air Blank | | 0.000 | 00:11 |
| Control Test | | 0.081 | 00:12 |
| Air Blank | | 0.000 | 00:12 |
| Subject Sample | # 1 | 0.126 | 00:13 |
| Breath Volume | | 2.000 | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Subject Sample | # 2 | 0.121 | 00:16 |
| Breath Volume | | 2.347 | 00:16 |
| Air Blank | | 0.000 | 00:17 |
| Control Test | | 0.079 | 00:17 |
| Air Blank | | 0.000 | 00:17 |
| Diagnostics Check | | OK | 00:17 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 03/13/2020 01:05:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:58 |
| Air Blank | | 0.000 | 00:58 |
| Control Test | | 0.079 | 00:58 |
| Air Blank | | 0.000 | 00:59 |
| Subject Sample | # 1 | 0.198 | 01:00 |
| Breath Volume | | 1.277 | 01:00 |
| Air Blank | | 0.000 | 01:01 |
| Air Blank | | 0.000 | 01:03 |
| Subject Sample | # 2 | 0.217 | 01:04 |
| Breath Volume | | 1.515 | 01:04 |
| Air Blank | | 0.000 | 01:04 |
| Control Test | | 0.077 | 01:05 |
| Air Blank | | 0.000 | 01:05 |
| Diagnostics Check | | OK | 01:05 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: NOLASCO

DATE/TIME OF TEST: 03/13/2020 02:41:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.080 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Subject Sample | # 1 | 0.098 | 02:37 |
| Breath Volume | | 1.867 | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Air Blank | | 0.000 | 02:39 |
| Subject Sample | # 2 | 0.099 | 02:40 |
| Breath Volume | | 2.359 | 02:40 |
| Air Blank | | 0.000 | 02:40 |
| Control Test | | 0.079 | 02:41 |
| Air Blank | | 0.000 | 02:41 |
| Diagnostics Check | | OK | 02:41 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MATADOR

DATE/TIME OF TEST: 03/13/2020 03:46:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:39 |
| Air Blank | | 0.000 | 03:39 |
| Control Test | | 0.078 | 03:40 |
| Air Blank | | 0.000 | 03:40 |
| Subject Sample | # 1 | 0.175 | 03:41 |
| Breath Volume | | 1.527 | 03:41 |
| Air Blank | | 0.000 | 03:42 |
| Air Blank | | 0.000 | 03:43 |
| Subject Sample | # 2 | 0.176 | 03:44 |
| Breath Volume | | 1.187 | 03:44 |
| Air Blank | | 0.000 | 03:45 |
| Control Test | | 0.079 | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Diagnostics Check | | OK | 03:46 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: BESSETTE

DATE/TIME OF TEST: 03/13/2020 19:40:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:34 |
| Air Blank | | 0.000 | 19:34 |
| Control Test | | 0.080 | 19:35 |
| Air Blank | | 0.000 | 19:35 |
| Subject Sample | # 1 | 0.000 | 19:36 |
| Breath Volume | | 2.238 | 19:36 |
| Air Blank | | 0.000 | 19:36 |
| Air Blank | | 0.000 | 19:38 |
| Subject Sample | # 2 | 0.000 | 19:38 |
| Breath Volume | | 1.605 | 19:38 |
| Air Blank | | 0.000 | 19:39 |
| Control Test | | 0.080 | 19:39 |
| Air Blank | | 0.000 | 19:40 |
| Diagnostics Check | | OK | 19:40 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HEALY

DATE/TIME OF TEST: 03/14/2020 02:08:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:01 |
| Air Blank | | 0.000 | 02:01 |
| Control Test | | 0.079 | 02:02 |
| Air Blank | | 0.000 | 02:02 |
| Subject Sample | # 1 | 0.136 | 02:04 |
| Breath Volume | | 2.699 | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 2 | 0.131 | 02:07 |
| Breath Volume | | 2.242 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.078 | 02:08 |
| Air Blank | | 0.000 | 02:08 |
| Diagnostics Check | | OK | 02:08 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 03/14/2020 02:53:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Control Test | | 0.080 | 02:47 |
| Air Blank | | 0.000 | 02:48 |
| Subject Sample | # 1 | 0.127 | 02:48 |
| Breath Volume | | 1.867 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Air Blank | | 0.000 | 02:51 |
| Subject Sample | # 2 | 0.119 | 02:51 |
| Breath Volume | | 1.492 | 02:51 |
| Air Blank | | 0.000 | 02:52 |
| Control Test | | 0.079 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Diagnostics Check | | OK | 02:53 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 03/14/2020 04:27:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:20 |
| Air Blank | | 0.000 | 04:21 |
| Control Test | | 0.079 | 04:21 |
| Air Blank | | 0.000 | 04:21 |
| Subject Sample | # 1 | 0.151 | 04:22 |
| Breath Volume | | 3.054 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Air Blank | | 0.000 | 04:25 |
| Subject Sample | # 2 | 0.155 | 04:25 |
| Breath Volume | | 2.351 | 04:25 |
| Air Blank | | 0.000 | 04:26 |
| Control Test | | 0.079 | 04:26 |
| Air Blank | | 0.000 | 04:27 |
| Diagnostics Check | | OK | 04:27 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 03/14/2020 05:23:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:17 |
| Air Blank | | 0.000 | 05:17 |
| Control Test | | 0.079 | 05:18 |
| Air Blank | | 0.000 | 05:18 |
| Subject Sample | # 1 | 0.156 | 05:19 |
| Breath Volume | | 2.042 | 05:19 |
| Air Blank | | 0.000 | 05:19 |
| Air Blank | | 0.000 | 05:21 |
| Subject Sample | # 2 | 0.146 | 05:22 |
| Breath Volume | | 1.261 | 05:22 |
| Air Blank | | 0.000 | 05:23 |
| Control Test | | 0.080 | 05:23 |
| Air Blank | | 0.000 | 05:23 |
| Diagnostics Check | | OK | 05:23 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HIPPOLYTE,VENEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: GARCIADIAZ-TORRES

DATE/TIME OF TEST: 03/14/2020 05:42:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:30 | |
| Air Blank | | 0.000 | 05:31 | |
| Control Test | | 0.079 | 05:31 | |
| Air Blank | | 0.000 | 05:31 | |
| Subject Sample | # 1 | | 05:35 | No Sample Provided |
| Breath Volume | | 0.000 | 05:35 | |
| Air Blank | | 0.000 | 05:35 | |
| Air Blank | | 0.000 | 05:37 | |
| Subject Sample | # 2 | 0.079 | 05:40 | Volume Not Met |
| Breath Volume | | 0.070 | 05:40 | |
| Air Blank | | 0.000 | 05:41 | |
| Control Test | | 0.079 | 05:41 | |
| Air Blank | | 0.000 | 05:42 | |
| Diagnostics Check | | OK | 05:42 | |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HIPPOLYTE,VENEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: GARCIADIAZ TORRES

DATE/TIME OF TEST: 03/14/2020 05:55:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:46 |
| Air Blank | | 0.000 | 05:46 |
| Control Test | | 0.078 | 05:46 |
| Air Blank | | 0.000 | 05:47 |
| Subject Sample | # 1 | 0.083 | 05:49 |
| Breath Volume | | 1.335 | 05:49 |
| Air Blank | | 0.000 | 05:50 |
| Air Blank | | 0.000 | 05:51 |
| Subject Sample | # 2 | 0.078 | 05:53 |
| Breath Volume | | 1.332 | 05:53 |
| Air Blank | | 0.000 | 05:54 |
| Control Test | | 0.077 | 05:54 |
| Air Blank | | 0.000 | 05:54 |
| Diagnostics Check | | OK | 05:55 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGT. BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: MUDRA

DATE/TIME OF TEST: 03/14/2020 22:48:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:40 |
| Air Blank | | 0.000 | 22:40 |
| Control Test | | 0.079 | 22:40 |
| Air Blank | | 0.000 | 22:41 |
| Subject Sample | # 1 | 0.242 | 22:42 |
| Breath Volume | | 1.636 | 22:42 |
| Air Blank | | 0.000 | 22:43 |
| Air Blank | | 0.000 | 22:44 |
| Subject Sample | # 2 | 0.224 | 22:46 |
| Breath Volume | | 2.125 | 22:46 |
| Air Blank | | 0.000 | 22:47 |
| Control Test | | 0.078 | 22:47 |
| Air Blank | | 0.000 | 22:48 |
| Diagnostics Check | | OK | 22:48 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GRADDY

DATE/TIME OF TEST: 03/14/2020 23:56:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:49 |
| Air Blank | | 0.000 | 23:49 |
| Control Test | | 0.078 | 23:50 |
| Air Blank | | 0.000 | 23:50 |
| Subject Sample | # 1 | 0.199 | 23:51 |
| Breath Volume | | 1.246 | 23:51 |
| Air Blank | | 0.000 | 23:52 |
| Air Blank | | 0.000 | 23:54 |
| Subject Sample | # 2 | 0.192 | 23:54 |
| Breath Volume | | 1.121 | 23:54 |
| Air Blank | | 0.000 | 23:55 |
| Control Test | | 0.078 | 23:55 |
| Air Blank | | 0.000 | 23:56 |
| Diagnostics Check | | OK | 23:56 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGRO,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 03/15/2020 01:21:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:15 |
| Air Blank | | 0.000 | 01:15 |
| Control Test | | 0.079 | 01:16 |
| Air Blank | | 0.000 | 01:16 |
| Subject Sample | # 1 | 0.165 | 01:16 |
| Breath Volume | | 1.917 | 01:16 |
| Air Blank | | 0.000 | 01:17 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 2 | 0.162 | 01:19 |
| Breath Volume | | 2.003 | 01:19 |
| Air Blank | | 0.000 | 01:20 |
| Control Test | | 0.078 | 01:20 |
| Air Blank | | 0.000 | 01:21 |
| Diagnostics Check | | OK | 01:21 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MATTHIS,CASEY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JIMENEZ ROMERO

DATE/TIME OF TEST: 03/15/2020 02:19:14

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:16 | |
| Air Blank | | 0.000 | 02:16 | |
| Control Test | | 0.079 | 02:17 | |
| Air Blank | | 0.000 | 02:17 | |
| Subject Sample | # 1 | | 02:17 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:17 | |
| Air Blank | | 0.000 | 02:18 | |
| Control Test | | 0.079 | 02:18 | |
| Air Blank | | 0.000 | 02:19 | |
| Diagnostics Check | | OK | 02:19 | |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MATTHIS,CASEY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JIMENEZ ROMERO

DATE/TIME OF TEST: 03/15/2020 02:30:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:23 |
| Air Blank | | 0.000 | 02:23 |
| Control Test | | 0.079 | 02:24 |
| Air Blank | | 0.000 | 02:24 |
| Subject Sample | # 1 | 0.195 | 02:25 |
| Breath Volume | | 1.546 | 02:25 |
| Air Blank | | 0.000 | 02:25 |
| Air Blank | | 0.000 | 02:27 |
| Subject Sample | # 2 | 0.176 | 02:29 |
| Breath Volume | | 1.105 | 02:29 |
| Air Blank | | 0.000 | 02:29 |
| Control Test | | 0.078 | 02:30 |
| Air Blank | | 0.000 | 02:30 |
| Diagnostics Check | | OK | 02:30 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROBBINS,MARK
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/15/2020 03:50:21

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 03:48 | |
| Air Blank | 0.000 | 03:49 | |
| Control Test | 0.070 | 03:49 | Control Outside Tolerance |
| Air Blank | 0.000 | 03:50 | |

Data Download Date/Time: 04/09/2020 16:01


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROBBINS,MARK
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/15/2020 03:53:46

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 03:52 | |
| Air Blank | 0.000 | 03:52 | |
| Control Test | 0.073 | 03:53 | Control Outside Tolerance |
| Air Blank | 0.000 | 03:53 | |

Data Download Date/Time: 04/09/2020 16:01


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROBBINS,MARK
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/15/2020 04:07:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:01 |
| Air Blank | | 0.000 | 04:01 |
| Control Test | | 0.076 | 04:01 |
| Air Blank | | 0.000 | 04:02 |
| Subject Sample | # 1 | 0.168 | 04:03 |
| Breath Volume | | 2.292 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Air Blank | | 0.000 | 04:05 |
| Subject Sample | # 2 | 0.165 | 04:06 |
| Breath Volume | | 1.675 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Control Test | | 0.076 | 04:07 |
| Air Blank | | 0.000 | 04:07 |
| Diagnostics Check | | OK | 04:07 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SHOR

DATE/TIME OF TEST: 03/15/2020 04:51:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:44 |
| Air Blank | | 0.000 | 04:45 |
| Control Test | | 0.078 | 04:45 |
| Air Blank | | 0.000 | 04:45 |
| Subject Sample | # 1 | 0.274 | 04:46 |
| Breath Volume | | 1.777 | 04:46 |
| Air Blank | | 0.000 | 04:47 |
| Air Blank | | 0.000 | 04:49 |
| Subject Sample | # 2 | 0.264 | 04:49 |
| Breath Volume | | 1.843 | 04:49 |
| Air Blank | | 0.000 | 04:50 |
| Control Test | | 0.077 | 04:50 |
| Air Blank | | 0.000 | 04:51 |
| Diagnostics Check | | OK | 04:51 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 03/17/2020 00:15:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:08 |
| Air Blank | | 0.000 | 00:09 |
| Control Test | | 0.080 | 00:09 |
| Air Blank | | 0.000 | 00:10 |
| Subject Sample | # 1 | 0.175 | 00:10 |
| Breath Volume | | 2.562 | 00:10 |
| Air Blank | | 0.000 | 00:11 |
| Air Blank | | 0.000 | 00:13 |
| Subject Sample | # 2 | 0.166 | 00:13 |
| Breath Volume | | 1.484 | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Control Test | | 0.081 | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Diagnostics Check | | OK | 00:15 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 03/17/2020 21:47:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:41 |
| Air Blank | | 0.000 | 21:41 |
| Control Test | | 0.079 | 21:41 |
| Air Blank | | 0.000 | 21:42 |
| Subject Sample | # 1 | 0.000 | 21:43 |
| Breath Volume | | 3.933 | 21:43 |
| Air Blank | | 0.000 | 21:43 |
| Air Blank | | 0.000 | 21:45 |
| Subject Sample | # 2 | 0.000 | 21:46 |
| Breath Volume | | 4.738 | 21:46 |
| Air Blank | | 0.000 | 21:46 |
| Control Test | | 0.080 | 21:47 |
| Air Blank | | 0.000 | 21:47 |
| Diagnostics Check | | OK | 21:47 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLIGAN,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/20/2020 00:10:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:03 |
| Air Blank | | 0.000 | 00:04 |
| Control Test | | 0.080 | 00:04 |
| Air Blank | | 0.000 | 00:04 |
| Subject Sample | # 1 | 0.126 | 00:06 |
| Breath Volume | | 1.296 | 00:06 |
| Air Blank | | 0.000 | 00:06 |
| Air Blank | | 0.000 | 00:08 |
| Subject Sample | # 2 | 0.135 | 00:09 |
| Breath Volume | | 1.410 | 00:09 |
| Air Blank | | 0.000 | 00:09 |
| Control Test | | 0.078 | 00:10 |
| Air Blank | | 0.000 | 00:10 |
| Diagnostics Check | | OK | 00:10 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARINO,GREG
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 08:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 03/22/2020 09:55:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 09:48 |
| Air Blank | | 0.000 | 09:49 |
| Control Test | | 0.081 | 09:49 |
| Air Blank | | 0.000 | 09:50 |
| Subject Sample | # 1 | 0.184 | 09:51 |
| Breath Volume | | 2.160 | 09:51 |
| Air Blank | | 0.000 | 09:51 |
| Air Blank | | 0.000 | 09:53 |
| Subject Sample | # 2 | 0.178 | 09:54 |
| Breath Volume | | 2.417 | 09:54 |
| Air Blank | | 0.000 | 09:54 |
| Control Test | | 0.078 | 09:55 |
| Air Blank | | 0.000 | 09:55 |
| Diagnostics Check | | OK | 09:55 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEON,JOEL
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: STANDIFORD

DATE/TIME OF TEST: 03/27/2020 23:12:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:04 |
| Air Blank | | 0.000 | 23:04 |
| Control Test | | 0.080 | 23:05 |
| Air Blank | | 0.000 | 23:05 |
| Subject Sample | # 1 | 0.183 | 23:06 |
| Breath Volume | | 1.371 | 23:06 |
| Air Blank | | 0.000 | 23:06 |
| Air Blank | | 0.000 | 23:08 |
| Subject Sample | # 2 | 0.182 | 23:11 |
| Breath Volume | | 1.550 | 23:11 |
| Air Blank | | 0.000 | 23:12 |
| Control Test | | 0.080 | 23:12 |
| Air Blank | | 0.000 | 23:12 |
| Diagnostics Check | | OK | 23:12 |

Data Download Date/Time: 04/09/2020 16:01

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 03/07/2020 04:56:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:50 | |
| Air Blank | | 0.000 | 04:50 | |
| Control Test | | 0.082 | 04:51 | |
| Air Blank | | 0.000 | 04:51 | |
| Subject Sample | # 1 | 0.069 | 04:52 | |
| Breath Volume | | 1.828 | 04:52 | |
| Air Blank | | 0.000 | 04:52 | |
| Air Blank | | 0.000 | 04:54 | |
| Subject Sample | # 2 | 0.074 | 04:55 | |
| Breath Volume | | 1.648 | 04:55 | |
| Air Blank | | 0.000 | 04:55 | |
| Control Test | | 0.083 | 04:56 | |
| Air Blank | | | 04:56 | RFI Detect |
| Air Blank | | 0.000 | 04:56 | |

Data Download Date/Time: 03/24/2020 15:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MCFATTER

DATE/TIME OF TEST: 03/13/2020 23:54:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:48 |
| Air Blank | | 0.000 | 23:48 |
| Control Test | | 0.081 | 23:48 |
| Air Blank | | 0.000 | 23:49 |
| Subject Sample | # 1 | 0.143 | 23:49 |
| Breath Volume | | 1.652 | 23:49 |
| Air Blank | | 0.000 | 23:50 |
| Air Blank | | 0.000 | 23:52 |
| Subject Sample | # 2 | 0.137 | 23:52 |
| Breath Volume | | 1.648 | 23:52 |
| Air Blank | | 0.000 | 23:53 |
| Control Test | | 0.080 | 23:53 |
| Air Blank | | 0.000 | 23:54 |
| Diagnostics Check | | OK | 23:54 |

Data Download Date/Time: 03/24/2020 15:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRITT,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PIERCE,BRAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CHANCY

DATE/TIME OF TEST: 03/14/2020 00:50:35

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:47 | |
| Air Blank | | 0.000 | 00:48 | |
| Control Test | | 0.081 | 00:48 | |
| Air Blank | | 0.000 | 00:49 | |
| Subject Sample | # 1 | | 00:49 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:49 | |
| Air Blank | | 0.000 | 00:49 | |
| Control Test | | 0.081 | 00:49 | |
| Air Blank | | 0.000 | 00:50 | |
| Diagnostics Check | | OK | 00:50 | |

Data Download Date/Time: 03/24/2020 15:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: REXFORD,CHRISPOTHER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: JUSTICE

DATE/TIME OF TEST: 03/14/2020 02:00:37

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:57 | |
| Air Blank | | 0.000 | 01:57 | |
| Control Test | | 0.081 | 01:57 | |
| Air Blank | | 0.000 | 01:58 | |
| Subject Sample | # 1 | 0.100 | 01:59 | |
| Breath Volume | | 3.152 | 01:59 | |
| Air Blank | | | 01:59 | RFI Detect |
| Air Blank | | 0.000 | 01:59 | |
| Control Test | | 0.080 | 02:00 | |
| Air Blank | | 0.000 | 02:00 | |
| Diagnostics Check | | OK | 02:00 | |

Data Download Date/Time: 03/24/2020 15:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: REXFORD,CHRISPOTHER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: JUSTICE

DATE/TIME OF TEST: 03/14/2020 02:10:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:03 |
| Air Blank | | 0.000 | 02:04 |
| Control Test | | 0.081 | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Subject Sample | # 1 | 0.095 | 02:05 |
| Breath Volume | | 2.785 | 02:05 |
| Air Blank | | 0.000 | 02:06 |
| Air Blank | | 0.000 | 02:08 |
| Subject Sample | # 2 | 0.099 | 02:08 |
| Breath Volume | | 2.746 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Control Test | | 0.081 | 02:09 |
| Air Blank | | 0.000 | 02:10 |
| Diagnostics Check | | OK | 02:10 |

Data Download Date/Time: 03/24/2020 15:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 03/15/2020 21:37:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:34 | |
| Air Blank | | 0.000 | 21:34 | |
| Control Test | | 0.081 | 21:35 | |
| Air Blank | | 0.000 | 21:35 | |
| Subject Sample | # 1 | | 21:36 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:36 | |
| Air Blank | | 0.000 | 21:36 | |
| Control Test | | 0.080 | 21:37 | |
| Air Blank | | 0.000 | 21:37 | |
| Diagnostics Check | | OK | 21:37 | |

Data Download Date/Time: 03/24/2020 15:41

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: SMUTNY

DATE/TIME OF TEST: 03/01/2020 17:50:30

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 17:33 | |
| Air Blank | | 0.000 | 17:33 | |
| Control Test | | 0.080 | 17:33 | |
| Air Blank | | 0.000 | 17:34 | |
| Subject Sample | # 1 | 0.151 | 17:37 | Volume Not Met |
| Breath Volume | | 0.042 | 17:37 | |
| Air Blank | | 0.000 | 17:38 | |
| Air Blank | | 0.000 | 17:39 | |
| Subject Sample | # 2 | 0.209 | 17:43 | No .020 agreement |
| Breath Volume | | 2.230 | 17:43 | |
| Air Blank | | 0.000 | 17:43 | |
| Air Blank | | 0.000 | 17:45 | |
| Subject Sample | # 3 | 0.186 | 17:48 | No .020 agreement |
| Breath Volume | | 2.093 | 17:48 | |
| Air Blank | | 0.000 | 17:49 | |
| Control Test | | 0.078 | 17:49 | |
| Air Blank | | 0.000 | 17:50 | |
| Diagnostics Check | | OK | 17:50 | |

Data Download Date/Time: 03/11/2020 12:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: SMUTNY

DATE/TIME OF TEST: 03/01/2020 18:01:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:52 |
| Air Blank | | 0.000 | 17:53 |
| Control Test | | 0.077 | 17:53 |
| Air Blank | | 0.000 | 17:54 |
| Subject Sample | # 1 | 0.204 | 17:56 |
| Breath Volume | | 1.523 | 17:56 |
| Air Blank | | 0.000 | 17:57 |
| Air Blank | | 0.000 | 17:59 |
| Subject Sample | # 2 | 0.196 | 18:00 |
| Breath Volume | | 1.855 | 18:00 |
| Air Blank | | 0.000 | 18:00 |
| Control Test | | 0.076 | 18:01 |
| Air Blank | | 0.000 | 18:01 |
| Diagnostics Check | | OK | 18:01 |

Data Download Date/Time: 03/11/2020 12:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02618080A3
DRY GAS EXPIRY DATE: 03/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: STEPP

DATE/TIME OF TEST: 03/02/2020 03:06:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:59 |
| Air Blank | | 0.000 | 02:59 |
| Control Test | | 0.079 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Subject Sample | # 1 | 0.203 | 03:01 |
| Breath Volume | | 1.902 | 03:01 |
| Air Blank | | 0.000 | 03:02 |
| Air Blank | | 0.000 | 03:04 |
| Subject Sample | # 2 | 0.193 | 03:04 |
| Breath Volume | | 1.417 | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Control Test | | 0.079 | 03:05 |
| Air Blank | | 0.000 | 03:06 |
| Diagnostics Check | | OK | 03:06 |

Data Download Date/Time: 03/11/2020 12:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LOPEZ,JOSE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/01/2020 18:58:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 18:55 | |
| Air Blank | | 0.000 | 18:55 | |
| Control Test | | 0.081 | 18:55 | |
| Air Blank | | 0.000 | 18:56 | |
| Subject Sample | # 1 | | 18:56 | Subject Test Refused |
| Breath Volume | | 0.000 | 18:56 | |
| Air Blank | | 0.000 | 18:57 | |
| Control Test | | 0.081 | 18:57 | |
| Air Blank | | 0.000 | 18:58 | |
| Diagnostics Check | | OK | 18:58 | |

Data Download Date/Time: 03/15/2020 13:54

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CARRILLO,XAVIER
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WOODRUFF

DATE/TIME OF TEST: 03/07/2020 22:03:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:57 |
| Air Blank | | 0.000 | 21:57 |
| Control Test | | 0.081 | 21:57 |
| Air Blank | | 0.000 | 21:58 |
| Subject Sample | # 1 | 0.000 | 21:58 |
| Breath Volume | | 1.242 | 21:58 |
| Air Blank | | 0.000 | 21:59 |
| Air Blank | | 0.000 | 22:01 |
| Subject Sample | # 2 | 0.000 | 22:01 |
| Breath Volume | | 1.367 | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Control Test | | 0.081 | 22:02 |
| Air Blank | | 0.000 | 22:03 |
| Diagnostics Check | | OK | 22:03 |

Data Download Date/Time: 03/15/2020 13:54

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ELDA,I
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LANTINI,NICHOLAS
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RODRIGUEZSOSA

DATE/TIME OF TEST: 03/07/2020 23:25:05

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:15 | |
| Air Blank | | 0.000 | 23:15 | |
| Control Test | | 0.080 | 23:16 | |
| Air Blank | | 0.000 | 23:16 | |
| Subject Sample | # 1 | 0.166 | 23:18 | |
| Breath Volume | | 1.179 | 23:18 | |
| Air Blank | | 0.000 | 23:19 | |
| Air Blank | | 0.000 | 23:21 | |
| Subject Sample | # 2 | | 23:23 | Subject Test Refused |
| Breath Volume | | 0.777 | 23:23 | |
| Air Blank | | 0.000 | 23:24 | |
| Control Test | | 0.078 | 23:24 | |
| Air Blank | | 0.000 | 23:24 | |
| Diagnostics Check | | OK | 23:25 | |

Data Download Date/Time: 03/15/2020 13:54

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ELLISON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: POHL

DATE/TIME OF TEST: 03/01/2020 02:21:02

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:04 | |
| Air Blank | | 0.000 | 02:04 | |
| Control Test | | 0.081 | 02:04 | |
| Air Blank | | 0.000 | 02:05 | |
| Subject Sample | # 1 | 0.104 | 02:08 | Volume Not Met |
| Breath Volume | | 0.222 | 02:08 | |
| Air Blank | | 0.000 | 02:09 | |
| Air Blank | | 0.000 | 02:11 | |
| Subject Sample | # 2 | 0.111 | 02:13 | |
| Breath Volume | | 1.546 | 02:13 | |
| Air Blank | | 0.000 | 02:14 | |
| Air Blank | | 0.000 | 02:16 | |
| Subject Sample | # 3 | 0.081 | 02:19 | Volume Not Met |
| Breath Volume | | 0.867 | 02:19 | |
| Air Blank | | 0.000 | 02:20 | |
| Control Test | | 0.076 | 02:20 | |
| Air Blank | | 0.000 | 02:20 | |
| Diagnostics Check | | OK | 02:21 | |

Data Download Date/Time: 03/29/2020 03:22


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: POHL

DATE/TIME OF TEST: 03/01/2020 03:02:20

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 03:01 | |
| Air Blank | 0.000 | 03:01 | |
| Control Test | 0.075 | 03:01 | Interferent Detect |
| Air Blank | 0.000 | 03:02 | |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: POHK

DATE/TIME OF TEST: 03/01/2020 03:18:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:04 | |
| Air Blank | | 0.000 | 03:04 | |
| Control Test | | 0.077 | 03:05 | |
| Air Blank | | 0.000 | 03:05 | |
| Subject Sample | # 1 | 0.088 | 03:08 | Volume Not Met |
| Breath Volume | | 0.667 | 03:08 | |
| Air Blank | | 0.000 | 03:09 | |
| Air Blank | | 0.000 | 03:11 | |
| Subject Sample | # 2 | 0.075 | 03:12 | |
| Breath Volume | | 1.199 | 03:12 | |
| Air Blank | | 0.000 | 03:13 | |
| Air Blank | | 0.000 | 03:15 | |
| Subject Sample | # 3 | 0.073 | 03:16 | |
| Breath Volume | | 2.453 | 03:16 | |
| Air Blank | | 0.000 | 03:17 | |
| Control Test | | 0.070 | 03:17 | Control Outside Tolerance |
| Air Blank | | 0.000 | 03:18 | |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TEMS,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 03/01/2020 20:05:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:58 |
| Air Blank | | 0.000 | 19:58 |
| Control Test | | 0.078 | 19:58 |
| Air Blank | | 0.000 | 19:59 |
| Subject Sample | # 1 | 0.161 | 20:00 |
| Breath Volume | | 2.375 | 20:00 |
| Air Blank | | 0.000 | 20:00 |
| Air Blank | | 0.000 | 20:02 |
| Subject Sample | # 2 | 0.158 | 20:03 |
| Breath Volume | | 2.382 | 20:03 |
| Air Blank | | 0.000 | 20:04 |
| Control Test | | 0.079 | 20:04 |
| Air Blank | | 0.000 | 20:04 |
| Diagnostics Check | | OK | 20:05 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,MARDEE,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VAUGHAN,W
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: THAYER

DATE/TIME OF TEST: 03/01/2020 21:19:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:12 |
| Air Blank | | 0.000 | 21:13 |
| Control Test | | 0.080 | 21:13 |
| Air Blank | | 0.000 | 21:14 |
| Subject Sample | # 1 | 0.261 | 21:14 |
| Breath Volume | | 1.597 | 21:14 |
| Air Blank | | 0.000 | 21:15 |
| Air Blank | | 0.000 | 21:17 |
| Subject Sample | # 2 | 0.270 | 21:18 |
| Breath Volume | | 2.746 | 21:18 |
| Air Blank | | 0.000 | 21:18 |
| Control Test | | 0.079 | 21:19 |
| Air Blank | | 0.000 | 21:19 |
| Diagnostics Check | | OK | 21:19 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DELEON,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 03/01/2020 21:57:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:50 |
| Air Blank | | 0.000 | 21:51 |
| Control Test | | 0.080 | 21:51 |
| Air Blank | | 0.000 | 21:52 |
| Subject Sample | # 1 | 0.102 | 21:52 |
| Breath Volume | | 1.367 | 21:52 |
| Air Blank | | 0.000 | 21:53 |
| Air Blank | | 0.000 | 21:55 |
| Subject Sample | # 2 | 0.101 | 21:55 |
| Breath Volume | | 1.339 | 21:55 |
| Air Blank | | 0.000 | 21:56 |
| Control Test | | 0.080 | 21:56 |
| Air Blank | | 0.000 | 21:56 |
| Diagnostics Check | | OK | 21:57 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 08:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 03/04/2020 08:54:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:47 |
| Air Blank | | 0.000 | 08:47 |
| Control Test | | 0.080 | 08:48 |
| Air Blank | | 0.000 | 08:48 |
| Subject Sample | # 1 | 0.013 | 08:49 |
| Breath Volume | | 1.414 | 08:49 |
| Air Blank | | 0.000 | 08:49 |
| Air Blank | | 0.000 | 08:51 |
| Subject Sample | # 2 | 0.014 | 08:52 |
| Breath Volume | | 1.976 | 08:52 |
| Air Blank | | 0.000 | 08:53 |
| Control Test | | 0.080 | 08:53 |
| Air Blank | | 0.000 | 08:53 |
| Diagnostics Check | | OK | 08:54 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 03/04/2020 10:57:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:51 |
| Air Blank | | 0.000 | 10:51 |
| Control Test | | 0.080 | 10:52 |
| Air Blank | | 0.000 | 10:52 |
| Subject Sample | # 1 | 0.000 | 10:53 |
| Breath Volume | | 1.718 | 10:53 |
| Air Blank | | 0.000 | 10:53 |
| Air Blank | | 0.000 | 10:55 |
| Subject Sample | # 2 | 0.000 | 10:56 |
| Breath Volume | | 1.750 | 10:56 |
| Air Blank | | 0.000 | 10:56 |
| Control Test | | 0.081 | 10:56 |
| Air Blank | | 0.000 | 10:57 |
| Diagnostics Check | | OK | 10:57 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MAGOON,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 03/05/2020 03:19:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:12 |
| Control Test | | 0.080 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Subject Sample | # 1 | 0.142 | 03:14 |
| Breath Volume | | 1.695 | 03:14 |
| Air Blank | | 0.000 | 03:14 |
| Air Blank | | 0.000 | 03:16 |
| Subject Sample | # 2 | 0.135 | 03:17 |
| Breath Volume | | 1.898 | 03:17 |
| Air Blank | | 0.000 | 03:18 |
| Control Test | | 0.083 | 03:18 |
| Air Blank | | 0.000 | 03:18 |
| Diagnostics Check | | OK | 03:19 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,DEP,B
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JUDD,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MILOT

DATE/TIME OF TEST: 03/06/2020 18:53:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:46 |
| Air Blank | | 0.000 | 18:47 |
| Control Test | | 0.081 | 18:47 |
| Air Blank | | 0.000 | 18:47 |
| Subject Sample | # 1 | 0.078 | 18:48 |
| Breath Volume | | 2.652 | 18:48 |
| Air Blank | | 0.000 | 18:49 |
| Air Blank | | 0.000 | 18:51 |
| Subject Sample | # 2 | 0.070 | 18:51 |
| Breath Volume | | 2.796 | 18:51 |
| Air Blank | | 0.000 | 18:52 |
| Control Test | | 0.080 | 18:52 |
| Air Blank | | 0.000 | 18:53 |
| Diagnostics Check | | OK | 18:53 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,DEP,B
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JUDD,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MILOT

DATE/TIME OF TEST: 03/06/2020 23:00:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:53 |
| Air Blank | | 0.000 | 22:54 |
| Control Test | | 0.080 | 22:54 |
| Air Blank | | 0.000 | 22:54 |
| Subject Sample | # 1 | 0.000 | 22:55 |
| Breath Volume | | 2.492 | 22:55 |
| Air Blank | | 0.000 | 22:56 |
| Air Blank | | 0.000 | 22:58 |
| Subject Sample | # 2 | 0.000 | 22:58 |
| Breath Volume | | 2.496 | 22:58 |
| Air Blank | | 0.000 | 22:59 |
| Control Test | | 0.083 | 22:59 |
| Air Blank | | 0.000 | 22:59 |
| Diagnostics Check | | OK | 23:00 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANGELO,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GAVIA CHAVEZ

DATE/TIME OF TEST: 03/12/2020 02:58:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:55 | |
| Air Blank | | 0.000 | 02:56 | |
| Control Test | | 0.080 | 02:56 | |
| Air Blank | | 0.000 | 02:57 | |
| Subject Sample | # 1 | | 02:57 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:57 | |
| Air Blank | | 0.000 | 02:58 | |
| Control Test | | 0.080 | 02:58 | |
| Air Blank | | 0.000 | 02:58 | |
| Diagnostics Check | | OK | 02:58 | |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTSMANSORPHER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: HOLSKE

DATE/TIME OF TEST: 03/13/2020 23:48:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:33 | |
| Air Blank | | 0.000 | 23:34 | |
| Control Test | | 0.081 | 23:34 | |
| Air Blank | | 0.000 | 23:34 | |
| Subject Sample | # 1 | 0.209 | 23:38 | Volume Not Met |
| Breath Volume | | 1.117 | 23:38 | |
| Air Blank | | 0.000 | 23:39 | |
| Air Blank | | 0.000 | 23:40 | |
| Subject Sample | # 2 | 0.279 | 23:41 | |
| Breath Volume | | 1.894 | 23:41 | |
| Air Blank | | 0.000 | 23:42 | |
| Air Blank | | 0.000 | 23:44 | |
| Subject Sample | # 3 | 0.258 | 23:47 | Volume Not Met |
| Breath Volume | | 0.000 | 23:47 | |
| Air Blank | | | 23:48 | Purge Fail |
| Air Blank | | 0.000 | 23:48 | |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BARLEY,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: POPOVICH

DATE/TIME OF TEST: 03/15/2020 20:34:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:23 | |
| Air Blank | | 0.000 | 20:23 | |
| Control Test | | 0.079 | 20:23 | |
| Air Blank | | 0.000 | 20:24 | |
| Subject Sample | # 1 | | 20:27 | No Sample Provided |
| Breath Volume | | 0.000 | 20:27 | |
| Air Blank | | 0.000 | 20:28 | |
| Air Blank | | 0.000 | 20:29 | |
| Subject Sample | # 2 | | 20:33 | No Sample Provided |
| Breath Volume | | 0.000 | 20:33 | |
| Air Blank | | 0.000 | 20:33 | |
| Control Test | | 0.077 | 20:33 | |
| Air Blank | | 0.000 | 20:34 | |
| Diagnostics Check | | OK | 20:34 | |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WELCH,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CAPDAREST

DATE/TIME OF TEST: 03/16/2020 03:09:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:02 |
| Air Blank | | 0.000 | 03:03 |
| Control Test | | 0.078 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Subject Sample | # 1 | 0.139 | 03:04 |
| Breath Volume | | 2.261 | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Air Blank | | 0.000 | 03:07 |
| Subject Sample | # 2 | 0.139 | 03:07 |
| Breath Volume | | 1.839 | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Control Test | | 0.078 | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Diagnostics Check | | OK | 03:09 |

Data Download Date/Time: 03/29/2020 03:22


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FIORINO,RUSSELL
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: TIPPINS

DATE/TIME OF TEST: 03/16/2020 22:40:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:28 | |
| Air Blank | | 0.000 | 22:29 | |
| Control Test | | 0.081 | 22:29 | |
| Air Blank | | 0.000 | 22:30 | |
| Subject Sample | # 1 | 0.125 | 22:33 | Volume Not Met |
| Breath Volume | | 0.148 | 22:33 | |
| Air Blank | | 0.000 | 22:34 | |
| Air Blank | | 0.000 | 22:35 | |
| Subject Sample | # 2 | 0.132 | 22:38 | Volume Not Met |
| Breath Volume | | 0.601 | 22:38 | |
| Air Blank | | 0.000 | 22:39 | |
| Control Test | | 0.081 | 22:40 | |
| Air Blank | | 0.000 | 22:40 | |
| Diagnostics Check | | OK | 22:40 | |

Data Download Date/Time: 03/29/2020 03:22


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAYLOR,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: LUFT

DATE/TIME OF TEST: 03/16/2020 23:52:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:44 |
| Air Blank | | 0.000 | 23:45 |
| Control Test | | 0.080 | 23:45 |
| Air Blank | | 0.000 | 23:46 |
| Subject Sample | # 1 | 0.169 | 23:47 |
| Breath Volume | | 1.382 | 23:47 |
| Air Blank | | 0.000 | 23:48 |
| Air Blank | | 0.000 | 23:49 |
| Subject Sample | # 2 | 0.164 | 23:50 |
| Breath Volume | | 1.500 | 23:50 |
| Air Blank | | 0.000 | 23:51 |
| Control Test | | 0.080 | 23:51 |
| Air Blank | | 0.000 | 23:52 |
| Diagnostics Check | | OK | 23:52 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDD,DEP,B
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CORBIN,T
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/17/2020 09:02:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:54 |
| Air Blank | | 0.000 | 08:55 |
| Control Test | | 0.079 | 08:55 |
| Air Blank | | 0.000 | 08:55 |
| Subject Sample | # 1 | 0.043 | 08:56 |
| Breath Volume | | 3.046 | 08:56 |
| Air Blank | | 0.000 | 08:57 |
| Air Blank | | 0.000 | 08:59 |
| Subject Sample | # 2 | 0.042 | 09:00 |
| Breath Volume | | 2.429 | 09:00 |
| Air Blank | | 0.000 | 09:01 |
| Control Test | | 0.083 | 09:01 |
| Air Blank | | 0.000 | 09:01 |
| Diagnostics Check | | OK | 09:02 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAROSA,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ALVAREZ SARAT

DATE/TIME OF TEST: 03/18/2020 02:54:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:47 |
| Air Blank | | 0.000 | 02:48 |
| Control Test | | 0.080 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Subject Sample | # 1 | 0.194 | 02:50 |
| Breath Volume | | 1.859 | 02:50 |
| Air Blank | | 0.000 | 02:50 |
| Air Blank | | 0.000 | 02:52 |
| Subject Sample | # 2 | 0.186 | 02:53 |
| Breath Volume | | 2.332 | 02:53 |
| Air Blank | | 0.000 | 02:53 |
| Control Test | | 0.080 | 02:54 |
| Air Blank | | 0.000 | 02:54 |
| Diagnostics Check | | OK | 02:54 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PATTEN,JACOB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GRISHAM III

DATE/TIME OF TEST: 03/20/2020 02:42:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.079 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Subject Sample | # 1 | 0.144 | 02:37 |
| Breath Volume | | 2.671 | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Air Blank | | 0.000 | 02:40 |
| Subject Sample | # 2 | 0.143 | 02:40 |
| Breath Volume | | 2.609 | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Control Test | | 0.082 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Diagnostics Check | | OK | 02:42 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORDSTROM,DEPUTY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COTE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 03/21/2020 15:23:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:16 |
| Air Blank | | 0.000 | 15:17 |
| Control Test | | 0.080 | 15:17 |
| Air Blank | | 0.000 | 15:17 |
| Subject Sample | # 1 | 0.211 | 15:18 |
| Breath Volume | | 3.207 | 15:18 |
| Air Blank | | 0.000 | 15:19 |
| Air Blank | | 0.000 | 15:21 |
| Subject Sample | # 2 | 0.212 | 15:21 |
| Breath Volume | | 1.281 | 15:21 |
| Air Blank | | 0.000 | 15:22 |
| Control Test | | 0.076 | 15:22 |
| Air Blank | | 0.000 | 15:23 |
| Diagnostics Check | | OK | 15:23 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PUTMAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: LINDSTRAND

DATE/TIME OF TEST: 03/22/2020 22:02:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:55 |
| Air Blank | | 0.000 | 21:55 |
| Control Test | | 0.081 | 21:55 |
| Air Blank | | 0.000 | 21:56 |
| Subject Sample | # 1 | 0.127 | 21:57 |
| Breath Volume | | 3.007 | 21:57 |
| Air Blank | | 0.000 | 21:58 |
| Air Blank | | 0.000 | 21:59 |
| Subject Sample | # 2 | 0.127 | 22:00 |
| Breath Volume | | 3.246 | 22:00 |
| Air Blank | | 0.000 | 22:01 |
| Control Test | | 0.082 | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Diagnostics Check | | OK | 22:02 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VAUGHAN,W
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SNOKE

DATE/TIME OF TEST: 03/25/2020 01:22:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:15 |
| Air Blank | | 0.000 | 01:16 |
| Control Test | | 0.080 | 01:16 |
| Air Blank | | 0.000 | 01:17 |
| Subject Sample | # 1 | 0.097 | 01:17 |
| Breath Volume | | 2.226 | 01:17 |
| Air Blank | | 0.000 | 01:18 |
| Air Blank | | 0.000 | 01:20 |
| Subject Sample | # 2 | 0.096 | 01:21 |
| Breath Volume | | 2.304 | 01:21 |
| Air Blank | | 0.000 | 01:21 |
| Control Test | | 0.082 | 01:22 |
| Air Blank | | 0.000 | 01:22 |
| Diagnostics Check | | OK | 01:22 |

Data Download Date/Time: 03/29/2020 03:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KAKU

DATE/TIME OF TEST: 03/08/2020 04:26:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:20 |
| Air Blank | | 0.000 | 04:20 |
| Control Test | | 0.080 | 04:21 |
| Air Blank | | 0.000 | 04:21 |
| Subject Sample | # 1 | 0.155 | 04:22 |
| Breath Volume | | 2.960 | 04:22 |
| Air Blank | | 0.000 | 04:22 |
| Air Blank | | 0.000 | 04:24 |
| Subject Sample | # 2 | 0.151 | 04:24 |
| Breath Volume | | 3.140 | 04:24 |
| Air Blank | | 0.000 | 04:25 |
| Control Test | | 0.080 | 04:25 |
| Air Blank | | 0.000 | 04:26 |
| Diagnostics Check | | OK | 04:26 |

Data Download Date/Time: 03/11/2020 12:52

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: SAMITT

DATE/TIME OF TEST: 03/09/2020 22:47:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:41 |
| Air Blank | | 0.000 | 22:41 |
| Control Test | | 0.079 | 22:42 |
| Air Blank | | 0.000 | 22:42 |
| Subject Sample | # 1 | 0.093 | 22:43 |
| Breath Volume | | 2.140 | 22:43 |
| Air Blank | | 0.000 | 22:43 |
| Air Blank | | 0.000 | 22:45 |
| Subject Sample | # 2 | 0.090 | 22:46 |
| Breath Volume | | 2.359 | 22:46 |
| Air Blank | | 0.000 | 22:46 |
| Control Test | | 0.080 | 22:47 |
| Air Blank | | 0.000 | 22:47 |
| Diagnostics Check | | OK | 22:47 |

Data Download Date/Time: 03/11/2020 12:52

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLT,CATESSA,B
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: GONZALEZ NUNEZ

DATE/TIME OF TEST: 03/08/2020 21:08:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:05 | |
| Air Blank | | 0.000 | 21:06 | |
| Control Test | | 0.079 | 21:06 | |
| Air Blank | | 0.000 | 21:07 | |
| Subject Sample | # 1 | | 21:07 | Sequence Aborted |
| Breath Volume | | 0.000 | 21:07 | |
| Air Blank | | 0.000 | 21:08 | |

Data Download Date/Time: 03/31/2020 15:38

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLT,CATESSA,B
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BOGUSKY,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GONZALEZ NUNEZ

DATE/TIME OF TEST: 03/08/2020 21:27:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:21 |
| Air Blank | | 0.000 | 21:21 |
| Control Test | | 0.079 | 21:22 |
| Air Blank | | 0.000 | 21:22 |
| Subject Sample | # 1 | 0.120 | 21:23 |
| Breath Volume | | 3.085 | 21:23 |
| Air Blank | | 0.000 | 21:23 |
| Air Blank | | 0.000 | 21:25 |
| Subject Sample | # 2 | 0.122 | 21:25 |
| Breath Volume | | 2.589 | 21:25 |
| Air Blank | | 0.000 | 21:26 |
| Control Test | | 0.077 | 21:26 |
| Air Blank | | 0.000 | 21:27 |
| Diagnostics Check | | OK | 21:27 |

Data Download Date/Time: 03/31/2020 15:38


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORT,THOMAS,B
BREATH TEST OPERATOR AGENCY: Wauchula PD
ARRESTING OFFICER: RAGLIN,JEREMMY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 02/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 03/14/2020 02:48:35

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:31 | |
| Air Blank | | 0.000 | 02:32 | |
| Control Test | | 0.078 | 02:32 | |
| Air Blank | | 0.000 | 02:33 | |
| Subject Sample | # 1 | 0.112 | 02:36 | Volume Not Met |
| Breath Volume | | 0.261 | 02:36 | |
| Air Blank | | 0.000 | 02:36 | |
| Air Blank | | 0.000 | 02:38 | |
| Subject Sample | # 2 | 0.110 | 02:41 | |
| Breath Volume | | 1.132 | 02:41 | |
| Air Blank | | 0.000 | 02:42 | |
| Air Blank | | 0.000 | 02:43 | |
| Subject Sample | # 3 | 0.108 | 02:46 | Volume Not Met |
| Breath Volume | | 0.117 | 02:46 | |
| Air Blank | | 0.000 | 02:47 | |
| Control Test | | 0.077 | 02:47 | |
| Air Blank | | 0.000 | 02:48 | |
| Diagnostics Check | | OK | 02:48 | |

Data Download Date/Time: 03/31/2020 15:38


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: PACHECO SANCHEZ

DATE/TIME OF TEST: 03/02/2020 15:16:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:10 |
| Air Blank | | 0.000 | 15:10 |
| Control Test | | 0.079 | 15:11 |
| Air Blank | | 0.000 | 15:11 |
| Subject Sample | # 1 | 0.094 | 15:12 |
| Breath Volume | | 2.078 | 15:12 |
| Air Blank | | 0.000 | 15:12 |
| Air Blank | | 0.000 | 15:14 |
| Subject Sample | # 2 | 0.099 | 15:15 |
| Breath Volume | | 1.902 | 15:15 |
| Air Blank | | 0.000 | 15:15 |
| Control Test | | 0.079 | 15:16 |
| Air Blank | | 0.000 | 15:16 |
| Diagnostics Check | | OK | 15:16 |

Data Download Date/Time: 03/05/2020 10:14

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: VILCHEZ ORTIZ

DATE/TIME OF TEST: 03/07/2020 04:32:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:22 |
| Air Blank | | 0.000 | 04:22 |
| Control Test | | 0.079 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Subject Sample | # 1 | 0.208 | 04:24 |
| Breath Volume | | 1.660 | 04:24 |
| Air Blank | | 0.000 | 04:25 |
| Air Blank | | 0.000 | 04:27 |
| Subject Sample | # 2 | 0.209 | 04:30 |
| Breath Volume | | 1.878 | 04:30 |
| Air Blank | | 0.000 | 04:31 |
| Control Test | | 0.079 | 04:31 |
| Air Blank | | 0.000 | 04:32 |
| Diagnostics Check | | OK | 04:32 |

Data Download Date/Time: 04/07/2020 11:57

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LUND

DATE/TIME OF TEST: 03/08/2020 00:24:09

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:21 | |
| Air Blank | | 0.000 | 00:21 | |
| Control Test | | 0.080 | 00:22 | |
| Air Blank | | 0.000 | 00:22 | |
| Subject Sample | # 1 | | 00:22 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:22 | |
| Air Blank | | 0.000 | 00:23 | |
| Control Test | | 0.080 | 00:23 | |
| Air Blank | | 0.000 | 00:24 | |
| Diagnostics Check | | OK | 00:24 | |

Data Download Date/Time: 04/07/2020 11:57

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 03/11/2020 22:20:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:17 | |
| Air Blank | | 0.000 | 22:17 | |
| Control Test | | 0.079 | 22:17 | |
| Air Blank | | 0.000 | 22:18 | |
| Subject Sample | # 1 | | 22:19 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:19 | |
| Air Blank | | 0.000 | 22:19 | |
| Control Test | | 0.078 | 22:19 | |
| Air Blank | | 0.000 | 22:20 | |
| Diagnostics Check | | OK | 22:20 | |

Data Download Date/Time: 04/07/2020 11:57

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: SZELA

DATE/TIME OF TEST: 03/12/2020 12:03:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 12:00 | |
| Air Blank | | 0.000 | 12:00 | |
| Control Test | | 0.079 | 12:01 | |
| Air Blank | | 0.000 | 12:01 | |
| Subject Sample | # 1 | | 12:02 | Subject Test Refused |
| Breath Volume | | 0.000 | 12:02 | |
| Air Blank | | 0.000 | 12:02 | |
| Control Test | | 0.079 | 12:03 | |
| Air Blank | | 0.000 | 12:03 | |
| Diagnostics Check | | OK | 12:03 | |

Data Download Date/Time: 04/07/2020 11:57

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYLER,BRYAN,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DULL

DATE/TIME OF TEST: 03/13/2020 01:09:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:03 |
| Air Blank | | 0.000 | 01:03 |
| Control Test | | 0.079 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Subject Sample | # 1 | 0.162 | 01:04 |
| Breath Volume | | 2.085 | 01:04 |
| Air Blank | | 0.000 | 01:05 |
| Air Blank | | 0.000 | 01:07 |
| Subject Sample | # 2 | 0.165 | 01:07 |
| Breath Volume | | 2.585 | 01:07 |
| Air Blank | | 0.000 | 01:08 |
| Control Test | | 0.078 | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Diagnostics Check | | OK | 01:09 |

Data Download Date/Time: 04/07/2020 11:57

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 03/13/2020 18:42:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 18:29 | |
| Air Blank | | 0.000 | 18:30 | |
| Control Test | | 0.079 | 18:30 | |
| Air Blank | | 0.000 | 18:31 | |
| Subject Sample | # 1 | 0.152 | 18:34 | Volume Not Met |
| Breath Volume | | 0.570 | 18:34 | |
| Air Blank | | 0.000 | 18:35 | |
| Air Blank | | 0.000 | 18:36 | |
| Subject Sample | # 2 | 0.189 | 18:37 | |
| Breath Volume | | 2.414 | 18:37 | |
| Air Blank | | 0.000 | 18:37 | |
| Air Blank | | 0.000 | 18:39 | |
| Subject Sample | # 3 | 0.192 | 18:40 | |
| Breath Volume | | 2.351 | 18:40 | |
| Air Blank | | 0.000 | 18:41 | |
| Control Test | | 0.078 | 18:41 | |
| Air Blank | | 0.000 | 18:42 | |
| Diagnostics Check | | OK | 18:42 | |

Data Download Date/Time: 04/07/2020 11:57

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: PITA-JOYA

DATE/TIME OF TEST: 03/08/2020 21:31:51

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:27 | |
| Air Blank | | 0.000 | 21:28 | |
| Control Test | | 0.075 | 21:28 | |
| Air Blank | | 0.000 | 21:28 | |
| Subject Sample | # 1 | | 21:31 | Sequence Aborted |
| Breath Volume | | 0.000 | 21:31 | |
| Air Blank | | 0.000 | 21:31 | |

Data Download Date/Time: 03/31/2020 21:24

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A1
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: PITA-JOYA

DATE/TIME OF TEST: 03/08/2020 22:04:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:50 | |
| Air Blank | | 0.000 | 21:50 | |
| Control Test | | 0.078 | 21:51 | |
| Air Blank | | 0.000 | 21:51 | |
| Subject Sample | # 1 | 0.101 | 21:54 | Volume Not Met |
| Breath Volume | | 0.679 | 21:54 | |
| Air Blank | | 0.000 | 21:55 | |
| Air Blank | | 0.000 | 21:57 | |
| Subject Sample | # 2 | 0.132 | 21:57 | |
| Breath Volume | | 1.367 | 21:57 | |
| Air Blank | | 0.000 | 21:58 | |
| Air Blank | | 0.000 | 22:00 | |
| Subject Sample | # 3 | 0.106 | 22:03 | Volume Not Met |
| Breath Volume | | 0.687 | 22:03 | |
| Air Blank | | 0.000 | 22:04 | |
| Control Test | | 0.077 | 22:04 | |
| Air Blank | | 0.000 | 22:04 | |
| Diagnostics Check | | OK | 22:04 | |

Data Download Date/Time: 03/31/2020 21:24

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RALSTON

DATE/TIME OF TEST: 03/01/2020 03:07:41

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:03 | |
| Air Blank | | 0.000 | 03:04 | |
| Control Test | | 0.081 | 03:04 | |
| Air Blank | | 0.000 | 03:05 | |
| Subject Sample | # 1 | 0.177 | 03:05 | |
| Breath Volume | | 1.109 | 03:05 | |
| Air Blank | | | 03:06 | RFI Detect |
| Air Blank | | 0.000 | 03:06 | |
| Control Test | | 0.080 | 03:07 | |
| Air Blank | | 0.000 | 03:07 | |
| Diagnostics Check | | OK | 03:07 | |

Data Download Date/Time: 03/05/2020 12:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RALSTON

DATE/TIME OF TEST: 03/01/2020 03:09:46

| | | | |
|---|---|---|---|
| Diagnostics Check | Fail | 03:09 | Internal Printer Test |
| Air Blank | 0.000 | 03:09 | |

Data Download Date/Time: 03/05/2020 12:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RALSTON

DATE/TIME OF TEST: 03/01/2020 03:17:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:11 |
| Air Blank | | 0.000 | 03:11 |
| Control Test | | 0.080 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Subject Sample | # 1 | 0.171 | 03:13 |
| Breath Volume | | 1.812 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Air Blank | | 0.000 | 03:15 |
| Subject Sample | # 2 | 0.169 | 03:16 |
| Breath Volume | | 1.570 | 03:16 |
| Air Blank | | 0.000 | 03:16 |
| Control Test | | 0.079 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Diagnostics Check | | OK | 03:17 |

Data Download Date/Time: 03/05/2020 12:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: GONZALEZ DE LA ROSA

DATE/TIME OF TEST: 03/02/2020 04:39:58

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 04:38 | |
| Air Blank | 0.000 | 04:38 | |
| Control Test | 0.080 | 04:39 | |
| Air Blank | | 04:39 | RFI Detect |
| Air Blank | 0.000 | 04:39 | |

Data Download Date/Time: 03/05/2020 12:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: GONZALEZ DE LA ROSA

DATE/TIME OF TEST: 03/02/2020 04:47:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:41 |
| Air Blank | | 0.000 | 04:41 |
| Control Test | | 0.081 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Subject Sample | # 1 | 0.133 | 04:43 |
| Breath Volume | | 2.285 | 04:43 |
| Air Blank | | 0.000 | 04:44 |
| Air Blank | | 0.000 | 04:46 |
| Subject Sample | # 2 | 0.131 | 04:46 |
| Breath Volume | | 1.593 | 04:46 |
| Air Blank | | 0.000 | 04:47 |
| Control Test | | 0.079 | 04:47 |
| Air Blank | | 0.000 | 04:47 |
| Diagnostics Check | | OK | 04:47 |

Data Download Date/Time: 03/05/2020 12:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 183142
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CARLISLE

DATE/TIME OF TEST: 03/05/2020 01:24:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:13 |
| Air Blank | | 0.000 | 01:14 |
| Control Test | | 0.079 | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Subject Sample | # 1 | 0.230 | 01:15 |
| Breath Volume | | 2.625 | 01:15 |
| Air Blank | | 0.000 | 01:16 |
| Air Blank | | 0.000 | 01:18 |
| Subject Sample | # 2 | 0.182 | 01:19 |
| Breath Volume | | 1.722 | 01:19 |
| Air Blank | | 0.000 | 01:19 |
| Air Blank | | 0.000 | 01:21 |
| Subject Sample | # 3 | 0.229 | 01:22 |
| Breath Volume | | 3.195 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Control Test | | 0.079 | 01:23 |
| Air Blank | | 0.000 | 01:23 |
| Diagnostics Check | | OK | 01:24 |

Data Download Date/Time: 03/05/2020 12:59

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: HUERTA

DATE/TIME OF TEST: 03/07/2020 00:04:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:54 |
| Air Blank | | 0.000 | 23:55 |
| Control Test | | 0.080 | 23:55 |
| Air Blank | | 0.000 | 23:55 |
| Subject Sample | # 3 | 0.202 | 23:56 |
| Breath Volume | | 1.398 | 23:56 |
| Air Blank | | 0.000 | 23:57 |
| Air Blank | | 0.000 | 23:59 |
| Subject Sample | # 1 | 0.178 | 00:00 |
| Breath Volume | | 1.523 | 00:00 |
| Air Blank | | 0.000 | 00:00 |
| Air Blank | | 0.000 | 00:02 |
| Subject Sample | # 2 | 0.183 | 00:03 |
| Breath Volume | | 1.289 | 00:03 |
| Air Blank | | 0.000 | 00:03 |
| Control Test | | 0.079 | 00:04 |
| Air Blank | | 0.000 | 00:04 |
| Diagnostics Check | | OK | 00:04 |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTERO,LISSETTE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PAHUACHO

DATE/TIME OF TEST: 03/07/2020 01:10:18

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:56 | |
| Air Blank | | 0.000 | 00:56 | |
| Control Test | | 0.080 | 00:56 | |
| Air Blank | | 0.000 | 00:57 | |
| Subject Sample | # 1 | 0.067 | 01:00 | Volume Not Met |
| Breath Volume | | 0.042 | 01:00 | |
| Air Blank | | 0.000 | 01:01 | |
| Air Blank | | 0.000 | 01:02 | |
| Subject Sample | # 2 | 0.061 | 01:05 | |
| Breath Volume | | 1.156 | 01:05 | |
| Air Blank | | 0.000 | 01:05 | |
| Air Blank | | 0.000 | 01:07 | |
| Subject Sample | # 3 | 0.060 | 01:08 | |
| Breath Volume | | 1.289 | 01:08 | |
| Air Blank | | 0.000 | 01:09 | |
| Control Test | | 0.078 | 01:09 | |
| Air Blank | | 0.000 | 01:10 | |
| Diagnostics Check | | OK | 01:10 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 03/14/2020 03:47:34

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:35 | |
| Air Blank | | 0.000 | 03:36 | |
| Control Test | | 0.079 | 03:36 | |
| Air Blank | | 0.000 | 03:37 | |
| Subject Sample | # 1 | | 03:40 | No Sample Provided |
| Breath Volume | | 0.000 | 03:40 | |
| Air Blank | | 0.000 | 03:40 | |
| Air Blank | | 0.000 | 03:42 | |
| Subject Sample | # 2 | 0.207 | 03:45 | Volume Not Met |
| Breath Volume | | 0.960 | 03:45 | |
| Air Blank | | 0.000 | 03:46 | |
| Control Test | | 0.079 | 03:46 | |
| Air Blank | | 0.000 | 03:47 | |
| Diagnostics Check | | OK | 03:47 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: VIDAL,NERELYS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 03/14/2020 04:21:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:14 |
| Air Blank | | 0.000 | 04:14 |
| Control Test | | 0.078 | 04:15 |
| Air Blank | | 0.000 | 04:15 |
| Subject Sample | # 1 | 0.207 | 04:16 |
| Breath Volume | | 1.460 | 04:16 |
| Air Blank | | 0.000 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Subject Sample | # 2 | 0.215 | 04:19 |
| Breath Volume | | 2.781 | 04:19 |
| Air Blank | | 0.000 | 04:20 |
| Control Test | | 0.079 | 04:20 |
| Air Blank | | 0.000 | 04:21 |
| Diagnostics Check | | OK | 04:21 |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 183142
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SECKINGER

DATE/TIME OF TEST: 03/15/2020 00:19:25

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:04 | |
| Air Blank | | 0.000 | 00:05 | |
| Control Test | | 0.080 | 00:05 | |
| Air Blank | | 0.000 | 00:06 | |
| Subject Sample | # 1 | | 00:09 | No Sample Provided |
| Breath Volume | | 0.000 | 00:09 | |
| Air Blank | | 0.000 | 00:09 | |
| Air Blank | | 0.000 | 00:11 | |
| Subject Sample | # 2 | 0.107 | 00:13 | |
| Breath Volume | | 2.656 | 00:13 | |
| Air Blank | | 0.000 | 00:14 | |
| Air Blank | | 0.000 | 00:16 | |
| Subject Sample | # 3 | 0.108 | 00:17 | |
| Breath Volume | | 2.757 | 00:17 | |
| Air Blank | | 0.000 | 00:18 | |
| Control Test | | 0.079 | 00:18 | |
| Air Blank | | 0.000 | 00:19 | |
| Diagnostics Check | | OK | 00:19 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DUKES JR

DATE/TIME OF TEST: 03/15/2020 03:56:05

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:53 | |
| Air Blank | | 0.000 | 03:53 | |
| Control Test | | 0.080 | 03:53 | |
| Air Blank | | 0.000 | 03:54 | |
| Subject Sample | # 1 | | 03:54 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:54 | |
| Air Blank | | 0.000 | 03:55 | |
| Control Test | | 0.079 | 03:55 | |
| Air Blank | | 0.000 | 03:55 | |
| Diagnostics Check | | OK | 03:56 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GISTEA,VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 03/17/2020 00:31:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:24 |
| Air Blank | | 0.000 | 00:24 |
| Control Test | | 0.081 | 00:25 |
| Air Blank | | 0.000 | 00:25 |
| Subject Sample | # 1 | 0.125 | 00:26 |
| Breath Volume | | 3.554 | 00:26 |
| Air Blank | | 0.000 | 00:27 |
| Air Blank | | 0.000 | 00:28 |
| Subject Sample | # 2 | 0.126 | 00:29 |
| Breath Volume | | 3.941 | 00:29 |
| Air Blank | | 0.000 | 00:30 |
| Control Test | | 0.080 | 00:30 |
| Air Blank | | 0.000 | 00:31 |
| Diagnostics Check | | OK | 00:31 |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MUSCATELLO

DATE/TIME OF TEST: 03/20/2020 02:30:38

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:27 | |
| Air Blank | | 0.000 | 02:28 | |
| Control Test | | 0.080 | 02:28 | |
| Air Blank | | 0.000 | 02:29 | |
| Subject Sample | # 1 | | 02:29 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:29 | |
| Air Blank | | 0.000 | 02:29 | |
| Control Test | | 0.080 | 02:30 | |
| Air Blank | | 0.000 | 02:30 | |
| Diagnostics Check | | OK | 02:30 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 03/20/2020 18:29:05

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 18:26 | |
| Air Blank | | 0.000 | 18:26 | |
| Control Test | | 0.080 | 18:26 | |
| Air Blank | | 0.000 | 18:27 | |
| Subject Sample | # 1 | | 18:27 | Subject Test Refused |
| Breath Volume | | 0.000 | 18:27 | |
| Air Blank | | 0.000 | 18:28 | |
| Control Test | | 0.080 | 18:28 | |
| Air Blank | | 0.000 | 18:29 | |
| Diagnostics Check | | OK | 18:29 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 03/24/2020 01:43:54

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:31 | |
| Air Blank | | 0.000 | 01:32 | |
| Control Test | | 0.079 | 01:32 | |
| Air Blank | | 0.000 | 01:33 | |
| Subject Sample | # 1 | | 01:36 | No Sample Provided |
| Breath Volume | | 0.000 | 01:36 | |
| Air Blank | | 0.000 | 01:36 | |
| Air Blank | | 0.000 | 01:38 | |
| Subject Sample | # 2 | 0.129 | 01:39 | |
| Breath Volume | | 3.003 | 01:39 | |
| Air Blank | | 0.000 | 01:40 | |
| Air Blank | | 0.000 | 01:41 | |
| Subject Sample | # 3 | 0.127 | 01:42 | |
| Breath Volume | | 1.929 | 01:42 | |
| Air Blank | | 0.000 | 01:42 | |
| Control Test | | 0.080 | 01:43 | |
| Air Blank | | 0.000 | 01:43 | |
| Diagnostics Check | | OK | 01:43 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: VALDES,DANIEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:49

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/26/2020 09:16:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 09:10 |
| Air Blank | | 0.000 | 09:10 |
| Control Test | | 0.079 | 09:10 |
| Air Blank | | 0.000 | 09:11 |
| Subject Sample | # 1 | 0.000 | 09:12 |
| Breath Volume | | 2.144 | 09:12 |
| Air Blank | | 0.000 | 09:12 |
| Air Blank | | 0.000 | 09:14 |
| Subject Sample | # 2 | 0.000 | 09:15 |
| Breath Volume | | 1.816 | 09:15 |
| Air Blank | | 0.000 | 09:15 |
| Control Test | | 0.079 | 09:15 |
| Air Blank | | 0.000 | 09:16 |
| Diagnostics Check | | OK | 09:16 |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CRISTEA,VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 03/26/2020 23:25:33

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:21 | |
| Air Blank | | 0.000 | 23:21 | |
| Control Test | | 0.079 | 23:22 | |
| Air Blank | | 0.000 | 23:22 | |
| Subject Sample | # 1 | | 23:24 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:24 | |
| Air Blank | | 0.000 | 23:24 | |
| Control Test | | 0.079 | 23:24 | |
| Air Blank | | 0.000 | 23:25 | |
| Diagnostics Check | | OK | 23:25 | |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 03/28/2020 03:17:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:09 |
| Air Blank | | 0.000 | 03:09 |
| Control Test | | 0.080 | 03:10 |
| Air Blank | | 0.000 | 03:10 |
| Subject Sample | # 1 | 0.145 | 03:11 |
| Breath Volume | | 1.484 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Air Blank | | 0.000 | 03:14 |
| Subject Sample | # 2 | 0.129 | 03:15 |
| Breath Volume | | 1.562 | 03:15 |
| Air Blank | | 0.000 | 03:16 |
| Control Test | | 0.079 | 03:16 |
| Air Blank | | 0.000 | 03:17 |
| Diagnostics Check | | OK | 03:17 |

Data Download Date/Time: 04/06/2020 12:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/29/2020 01:24:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:17 |
| Air Blank | | 0.000 | 01:18 |
| Control Test | | 0.080 | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 1 | 0.257 | 01:19 |
| Breath Volume | | 1.652 | 01:19 |
| Air Blank | | 0.000 | 01:20 |
| Air Blank | | 0.000 | 01:22 |
| Subject Sample | # 2 | 0.255 | 01:22 |
| Breath Volume | | 1.468 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Control Test | | 0.079 | 01:23 |
| Air Blank | | 0.000 | 01:24 |
| Diagnostics Check | | OK | 01:24 |

Data Download Date/Time: 04/06/2020 12:55


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,ALYSSA,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,ALYSSA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 03/01/2020 03:39:41

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:34 | |
| Air Blank | | 0.000 | 03:35 | |
| Control Test | | 0.081 | 03:35 | |
| Air Blank | | 0.000 | 03:36 | |
| Subject Sample | # 1 | 0.122 | 03:37 | |
| Breath Volume | | 1.218 | 03:37 | |
| Air Blank | | | 03:38 | RFI Detect |
| Air Blank | | 0.000 | 03:38 | |
| Control Test | | 0.082 | 03:39 | |
| Air Blank | | 0.000 | 03:39 | |
| Diagnostics Check | | OK | 03:39 | |

Data Download Date/Time: 03/25/2020 15:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,ALYSSA,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,ALYSSA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 03/01/2020 03:47:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:41 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.080 | 03:42 |
| Air Blank | | 0.000 | 03:42 |
| Subject Sample | # 1 | 0.124 | 03:43 |
| Breath Volume | | 1.656 | 03:43 |
| Air Blank | | 0.000 | 03:43 |
| Air Blank | | 0.000 | 03:45 |
| Subject Sample | # 2 | 0.134 | 03:46 |
| Breath Volume | | 2.851 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Control Test | | 0.080 | 03:47 |
| Air Blank | | 0.000 | 03:47 |
| Diagnostics Check | | OK | 03:47 |

Data Download Date/Time: 03/25/2020 15:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,CALVIN,K
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PALMATEER,CALVIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 03/01/2020 05:09:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:02 |
| Air Blank | | 0.000 | 05:02 |
| Control Test | | 0.080 | 05:03 |
| Air Blank | | 0.000 | 05:03 |
| Subject Sample | # 1 | 0.189 | 05:04 |
| Breath Volume | | 1.898 | 05:04 |
| Air Blank | | 0.000 | 05:05 |
| Air Blank | | 0.000 | 05:06 |
| Subject Sample | # 2 | 0.193 | 05:07 |
| Breath Volume | | 2.324 | 05:07 |
| Air Blank | | 0.000 | 05:08 |
| Control Test | | 0.078 | 05:08 |
| Air Blank | | 0.000 | 05:09 |
| Diagnostics Check | | OK | 05:09 |

Data Download Date/Time: 03/25/2020 15:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WINTERS

DATE/TIME OF TEST: 03/01/2020 05:24:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:18 |
| Air Blank | | 0.000 | 05:18 |
| Control Test | | 0.079 | 05:18 |
| Air Blank | | 0.000 | 05:19 |
| Subject Sample | # 1 | 0.133 | 05:19 |
| Breath Volume | | 1.160 | 05:19 |
| Air Blank | | 0.000 | 05:20 |
| Air Blank | | 0.000 | 05:22 |
| Subject Sample | # 2 | 0.128 | 05:22 |
| Breath Volume | | 1.593 | 05:22 |
| Air Blank | | 0.000 | 05:23 |
| Control Test | | 0.081 | 05:23 |
| Air Blank | | 0.000 | 05:24 |
| Diagnostics Check | | OK | 05:24 |

Data Download Date/Time: 03/25/2020 15:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYER,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 13:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: CROZIER

DATE/TIME OF TEST: 03/01/2020 14:13:00

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 14:09 | |
| Air Blank | | 0.000 | 14:09 | |
| Control Test | | 0.081 | 14:09 | |
| Air Blank | | 0.000 | 14:10 | |
| Subject Sample | # 1 | | 14:11 | Subject Test Refused |
| Breath Volume | | 0.000 | 14:11 | |
| Air Blank | | 0.000 | 14:12 | |
| Control Test | | 0.082 | 14:12 | |
| Air Blank | | 0.000 | 14:12 | |
| Diagnostics Check | | OK | 14:13 | |

Data Download Date/Time: 03/25/2020 15:22

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 03/02/2020 19:53:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:47 |
| Air Blank | | 0.000 | 19:47 |
| Control Test | | 0.080 | 19:47 |
| Air Blank | | 0.000 | 19:48 |
| Subject Sample | # 1 | 0.000 | 19:49 |
| Breath Volume | | 1.796 | 19:49 |
| Air Blank | | 0.000 | 19:49 |
| Air Blank | | 0.000 | 19:51 |
| Subject Sample | # 2 | 0.000 | 19:52 |
| Breath Volume | | 1.683 | 19:52 |
| Air Blank | | 0.000 | 19:52 |
| Control Test | | 0.081 | 19:53 |
| Air Blank | | 0.000 | 19:53 |
| Diagnostics Check | | OK | 19:53 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITMORE,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITMORE,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 03/02/2020 22:29:17

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:25 | |
| Air Blank | | 0.000 | 22:25 | |
| Control Test | | 0.081 | 22:26 | |
| Air Blank | | 0.000 | 22:26 | |
| Subject Sample | # 1 | | 22:28 | Sequence Aborted |
| Breath Volume | | 0.000 | 22:28 | |
| Air Blank | | 0.000 | 22:29 | |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTANEZ,STEVEN,
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: COLEMAN,QUINEY
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 03/03/2020 00:58:19

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:55 | |
| Air Blank | | 0.000 | 00:55 | |
| Control Test | | 0.081 | 00:55 | |
| Air Blank | | 0.000 | 00:56 | |
| Subject Sample | # 1 | | 00:56 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:56 | |
| Air Blank | | 0.000 | 00:57 | |
| Control Test | | 0.081 | 00:57 | |
| Air Blank | | 0.000 | 00:58 | |
| Diagnostics Check | | OK | 00:58 | |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTS,C,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BATTS,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HESS

DATE/TIME OF TEST: 03/04/2020 01:21:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:15 |
| Air Blank | | 0.000 | 01:15 |
| Control Test | | 0.081 | 01:16 |
| Air Blank | | 0.000 | 01:16 |
| Subject Sample | # 1 | 0.013 | 01:17 |
| Breath Volume | | 1.214 | 01:17 |
| Air Blank | | 0.000 | 01:17 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 2 | 0.011 | 01:19 |
| Breath Volume | | 1.328 | 01:19 |
| Air Blank | | 0.000 | 01:20 |
| Control Test | | 0.079 | 01:20 |
| Air Blank | | 0.000 | 01:21 |
| Diagnostics Check | | OK | 01:21 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 13:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 03/04/2020 15:14:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:08 |
| Air Blank | | 0.000 | 15:08 |
| Control Test | | 0.081 | 15:09 |
| Air Blank | | 0.000 | 15:09 |
| Subject Sample | # 1 | 0.294 | 15:10 |
| Breath Volume | | 1.351 | 15:10 |
| Air Blank | | 0.000 | 15:10 |
| Air Blank | | 0.000 | 15:12 |
| Subject Sample | # 2 | 0.309 | 15:13 |
| Breath Volume | | 1.769 | 15:13 |
| Air Blank | | 0.000 | 15:13 |
| Control Test | | 0.080 | 15:14 |
| Air Blank | | 0.000 | 15:14 |
| Diagnostics Check | | OK | 15:14 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,M,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MOORE,M
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/21/2021
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 03/06/2020 21:24:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:18 |
| Air Blank | | 0.000 | 21:18 |
| Control Test | | 0.082 | 21:19 |
| Air Blank | | 0.000 | 21:19 |
| Subject Sample | # 1 | 0.031 | 21:20 |
| Breath Volume | | 1.871 | 21:20 |
| Air Blank | | 0.000 | 21:21 |
| Air Blank | | 0.000 | 21:22 |
| Subject Sample | # 2 | 0.029 | 21:23 |
| Breath Volume | | 1.339 | 21:23 |
| Air Blank | | 0.000 | 21:23 |
| Control Test | | 0.082 | 21:24 |
| Air Blank | | 0.000 | 21:24 |
| Diagnostics Check | | OK | 21:24 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MARTON,TIMOTHY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ODUARDO RODRIGUEZ

DATE/TIME OF TEST: 03/07/2020 04:11:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:05 |
| Air Blank | | 0.000 | 04:05 |
| Control Test | | 0.082 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Subject Sample | # 1 | 0.195 | 04:07 |
| Breath Volume | | 2.445 | 04:07 |
| Air Blank | | 0.000 | 04:07 |
| Air Blank | | 0.000 | 04:09 |
| Subject Sample | # 2 | 0.191 | 04:10 |
| Breath Volume | | 1.519 | 04:10 |
| Air Blank | | 0.000 | 04:10 |
| Control Test | | 0.079 | 04:11 |
| Air Blank | | 0.000 | 04:11 |
| Diagnostics Check | | OK | 04:11 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALVET,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CALVET,CHRISTINA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LUIS

DATE/TIME OF TEST: 03/07/2020 23:45:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:38 |
| Air Blank | | 0.000 | 23:39 |
| Control Test | | 0.082 | 23:39 |
| Air Blank | | 0.000 | 23:40 |
| Subject Sample | # 1 | 0.181 | 23:40 |
| Breath Volume | | 2.066 | 23:40 |
| Air Blank | | 0.000 | 23:41 |
| Air Blank | | 0.000 | 23:43 |
| Subject Sample | # 2 | 0.174 | 23:43 |
| Breath Volume | | 2.316 | 23:43 |
| Air Blank | | 0.000 | 23:44 |
| Control Test | | 0.079 | 23:44 |
| Air Blank | | 0.000 | 23:45 |
| Diagnostics Check | | OK | 23:45 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITMORE,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITMORE,JAMES
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MOSER

DATE/TIME OF TEST: 03/08/2020 23:13:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:06 |
| Air Blank | | 0.000 | 23:06 |
| Control Test | | 0.081 | 23:06 |
| Air Blank | | 0.000 | 23:07 |
| Subject Sample | # 1 | 0.161 | 23:08 |
| Breath Volume | | 1.933 | 23:08 |
| Air Blank | | 0.000 | 23:08 |
| Air Blank | | 0.000 | 23:10 |
| Subject Sample | # 2 | 0.156 | 23:12 |
| Breath Volume | | 1.914 | 23:12 |
| Air Blank | | 0.000 | 23:12 |
| Control Test | | 0.080 | 23:13 |
| Air Blank | | 0.000 | 23:13 |
| Diagnostics Check | | OK | 23:13 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HOLIDAY

DATE/TIME OF TEST: 03/11/2020 03:04:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.081 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Subject Sample | # 1 | 0.171 | 02:58 |
| Breath Volume | | 1.605 | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Air Blank | | 0.000 | 03:01 |
| Subject Sample | # 2 | 0.168 | 03:02 |
| Breath Volume | | 1.503 | 03:02 |
| Air Blank | | 0.000 | 03:03 |
| Control Test | | 0.079 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Diagnostics Check | | OK | 03:04 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: YESS

DATE/TIME OF TEST: 03/11/2020 22:26:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:19 |
| Air Blank | | 0.000 | 22:20 |
| Control Test | | 0.081 | 22:20 |
| Air Blank | | 0.000 | 22:20 |
| Subject Sample | # 1 | 0.020 | 22:21 |
| Breath Volume | | 2.035 | 22:21 |
| Air Blank | | 0.000 | 22:22 |
| Air Blank | | 0.000 | 22:23 |
| Subject Sample | # 2 | 0.021 | 22:24 |
| Breath Volume | | 2.968 | 22:24 |
| Air Blank | | 0.000 | 22:25 |
| Control Test | | 0.081 | 22:25 |
| Air Blank | | 0.000 | 22:26 |
| Diagnostics Check | | OK | 22:26 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/12/2020 01:30:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:24 |
| Air Blank | | 0.000 | 01:24 |
| Control Test | | 0.080 | 01:24 |
| Air Blank | | 0.000 | 01:25 |
| Subject Sample | # 1 | 0.119 | 01:26 |
| Breath Volume | | 2.855 | 01:26 |
| Air Blank | | 0.000 | 01:26 |
| Air Blank | | 0.000 | 01:28 |
| Subject Sample | # 2 | 0.119 | 01:29 |
| Breath Volume | | 1.562 | 01:29 |
| Air Blank | | 0.000 | 01:29 |
| Control Test | | 0.080 | 01:30 |
| Air Blank | | 0.000 | 01:30 |
| Diagnostics Check | | OK | 01:30 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 03/12/2020 03:37:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:29 |
| Air Blank | | 0.000 | 03:30 |
| Control Test | | 0.081 | 03:30 |
| Air Blank | | 0.000 | 03:30 |
| Subject Sample | # 1 | 0.126 | 03:32 |
| Breath Volume | | 2.285 | 03:32 |
| Air Blank | | 0.000 | 03:32 |
| Air Blank | | 0.000 | 03:34 |
| Subject Sample | # 2 | 0.130 | 03:35 |
| Breath Volume | | 1.648 | 03:35 |
| Air Blank | | 0.000 | 03:36 |
| Control Test | | 0.080 | 03:36 |
| Air Blank | | 0.000 | 03:36 |
| Diagnostics Check | | OK | 03:37 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 03/12/2020 22:08:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:00 |
| Air Blank | | 0.000 | 22:00 |
| Control Test | | 0.081 | 22:01 |
| Air Blank | | 0.000 | 22:01 |
| Subject Sample | # 1 | 0.065 | 22:03 |
| Breath Volume | | 1.214 | 22:03 |
| Air Blank | | 0.000 | 22:04 |
| Air Blank | | 0.000 | 22:05 |
| Subject Sample | # 2 | 0.058 | 22:06 |
| Breath Volume | | 1.960 | 22:06 |
| Air Blank | | 0.000 | 22:07 |
| Control Test | | 0.079 | 22:07 |
| Air Blank | | 0.000 | 22:08 |
| Diagnostics Check | | OK | 22:08 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHEL,ASHLYN,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHEL,ASHLYN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/12/2020 23:57:04

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 23:55 | |
| Air Blank | 0.000 | 23:56 | |
| Control Test | 0.080 | 23:56 | |
| Air Blank | | 23:56 | Improper Sample |
| Air Blank | | 23:57 | Improper Sample |

Data Download Date/Time: 03/25/2020 15:23


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHEL,ASHLYN,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHEL,ASHLYN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 875088
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/13/2020 00:01:09

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 00:00 | |
| Air Blank | | 00:00 | Improper Sample |
| Air Blank | | 00:01 | Improper Sample |

Data Download Date/Time: 03/25/2020 15:23


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHEL,ASHLYN,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHEL,ASHLYN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/13/2020 00:08:36

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 00:07 | |
| Air Blank | | 00:08 | Ambient Fail |
| Air Blank | 0.000 | 00:08 | |

Data Download Date/Time: 03/25/2020 15:23


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHEL,A,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHEL,A
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 12/12/2020
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 03/13/2020 00:20:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:13 |
| Air Blank | | 0.000 | 00:13 |
| Control Test | | 0.077 | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Subject Sample | # 1 | 0.292 | 00:15 |
| Breath Volume | | 4.886 | 00:15 |
| Air Blank | | 0.000 | 00:16 |
| Air Blank | | 0.000 | 00:17 |
| Subject Sample | # 2 | 0.286 | 00:18 |
| Breath Volume | | 5.800 | 00:18 |
| Air Blank | | 0.000 | 00:19 |
| Control Test | | 0.079 | 00:19 |
| Air Blank | | 0.000 | 00:20 |
| Diagnostics Check | | OK | 00:20 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JENTSCH,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SCHELLINGER

DATE/TIME OF TEST: 03/14/2020 02:47:33

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:35 | |
| Air Blank | | 0.000 | 02:35 | |
| Control Test | | 0.081 | 02:36 | |
| Air Blank | | 0.000 | 02:36 | |
| Subject Sample | # 1 | 0.114 | 02:39 | Volume Not Met |
| Breath Volume | | 0.195 | 02:39 | |
| Air Blank | | 0.000 | 02:40 | |
| Air Blank | | 0.000 | 02:42 | |
| Subject Sample | # 2 | 0.116 | 02:43 | |
| Breath Volume | | 1.695 | 02:43 | |
| Air Blank | | 0.000 | 02:43 | |
| Air Blank | | 0.000 | 02:45 | |
| Subject Sample | # 3 | 0.114 | 02:45 | |
| Breath Volume | | 1.398 | 02:45 | |
| Air Blank | | 0.000 | 02:46 | |
| Control Test | | 0.080 | 02:46 | |
| Air Blank | | 0.000 | 02:47 | |
| Diagnostics Check | | OK | 02:47 | |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: SPILLER

DATE/TIME OF TEST: 03/14/2020 20:52:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:45 |
| Air Blank | | 0.000 | 20:46 |
| Control Test | | 0.081 | 20:46 |
| Air Blank | | 0.000 | 20:47 |
| Subject Sample | # 1 | 0.131 | 20:48 |
| Breath Volume | | 1.441 | 20:48 |
| Air Blank | | 0.000 | 20:48 |
| Air Blank | | 0.000 | 20:50 |
| Subject Sample | # 2 | 0.132 | 20:50 |
| Breath Volume | | 1.765 | 20:50 |
| Air Blank | | 0.000 | 20:51 |
| Control Test | | 0.081 | 20:51 |
| Air Blank | | 0.000 | 20:52 |
| Diagnostics Check | | OK | 20:52 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LEMKE-HERMAN

DATE/TIME OF TEST: 03/15/2020 03:17:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:11 |
| Air Blank | | 0.000 | 03:11 |
| Control Test | | 0.081 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Subject Sample | # 1 | 0.246 | 03:12 |
| Breath Volume | | 1.492 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Air Blank | | 0.000 | 03:15 |
| Subject Sample | # 2 | 0.253 | 03:15 |
| Breath Volume | | 1.792 | 03:15 |
| Air Blank | | 0.000 | 03:16 |
| Control Test | | 0.080 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Diagnostics Check | | OK | 03:17 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSACI,NEIL,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ROSACI,NEIL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: BUSSOW

DATE/TIME OF TEST: 03/17/2020 16:24:16

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:15 | |
| Air Blank | | 0.000 | 16:16 | |
| Control Test | | 0.081 | 16:16 | |
| Air Blank | | 0.000 | 16:16 | |
| Subject Sample | # 1 | | 16:20 | No Sample Provided |
| Breath Volume | | 0.000 | 16:20 | |
| Air Blank | | 0.000 | 16:20 | |
| Air Blank | | 0.000 | 16:22 | |
| Subject Sample | # 2 | | 16:22 | Subject Test Refused |
| Breath Volume | | 0.000 | 16:22 | |
| Air Blank | | 0.000 | 16:23 | |
| Control Test | | 0.081 | 16:23 | |
| Air Blank | | 0.000 | 16:24 | |
| Diagnostics Check | | OK | 16:24 | |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,ALYSSA,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WITT,ALYSSA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 101121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: FORTUNE

DATE/TIME OF TEST: 03/19/2020 01:11:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:04 |
| Air Blank | | 0.000 | 01:04 |
| Control Test | | 0.081 | 01:04 |
| Air Blank | | 0.000 | 01:05 |
| Subject Sample | # 1 | 0.257 | 01:06 |
| Breath Volume | | 1.824 | 01:06 |
| Air Blank | | 0.000 | 01:06 |
| Air Blank | | 0.000 | 01:08 |
| Subject Sample | # 2 | 0.251 | 01:09 |
| Breath Volume | | 1.855 | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Control Test | | 0.079 | 01:10 |
| Air Blank | | 0.000 | 01:11 |
| Diagnostics Check | | OK | 01:11 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELINSKI,H,T
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HELINSKI,H
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: KINSLER

DATE/TIME OF TEST: 03/21/2020 22:10:19

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:54 | |
| Air Blank | | 0.000 | 21:54 | |
| Control Test | | 0.081 | 21:55 | |
| Air Blank | | 0.000 | 21:55 | |
| Subject Sample | # 1 | 0.189 | 21:57 | |
| Breath Volume | | 1.191 | 21:57 | |
| Air Blank | | 0.000 | 21:58 | |
| Air Blank | | 0.000 | 22:00 | |
| Subject Sample | # 2 | 0.161 | 22:03 | Volume Not Met |
| Breath Volume | | 0.175 | 22:03 | |
| Air Blank | | 0.000 | 22:04 | |
| Air Blank | | 0.000 | 22:05 | |
| Subject Sample | # 3 | | 22:08 | No Sample Provided |
| Breath Volume | | 0.000 | 22:08 | |
| Air Blank | | 0.000 | 22:09 | |
| Control Test | | 0.079 | 22:09 | |
| Air Blank | | 0.000 | 22:10 | |
| Diagnostics Check | | OK | 22:10 | |

Data Download Date/Time: 03/25/2020 15:23


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GURRIERI,JOHN,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GURRIERI,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 03/22/2020 01:17:04

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:14 | |
| Air Blank | | 0.000 | 01:14 | |
| Control Test | | 0.080 | 01:14 | |
| Air Blank | | 0.000 | 01:15 | |
| Subject Sample | # 1 | | 01:15 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:15 | |
| Air Blank | | 0.000 | 01:16 | |
| Control Test | | 0.081 | 01:16 | |
| Air Blank | | 0.000 | 01:16 | |
| Diagnostics Check | | OK | 01:17 | |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/23/2020 16:25:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:19 |
| Air Blank | | 0.000 | 16:19 |
| Control Test | | 0.081 | 16:19 |
| Air Blank | | 0.000 | 16:20 |
| Subject Sample | # 1 | 0.000 | 16:20 |
| Breath Volume | | 2.035 | 16:20 |
| Air Blank | | 0.000 | 16:21 |
| Air Blank | | 0.000 | 16:23 |
| Subject Sample | # 2 | 0.000 | 16:23 |
| Breath Volume | | 1.878 | 16:23 |
| Air Blank | | 0.000 | 16:24 |
| Control Test | | 0.080 | 16:24 |
| Air Blank | | 0.000 | 16:25 |
| Diagnostics Check | | OK | 16:25 |

Data Download Date/Time: 03/25/2020 15:23

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/02/2020 00:42:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:35 |
| Air Blank | | 0.000 | 00:36 |
| Control Test | | 0.080 | 00:36 |
| Air Blank | | 0.000 | 00:36 |
| Subject Sample | # 1 | 0.185 | 00:37 |
| Breath Volume | | 1.613 | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Air Blank | | 0.000 | 00:40 |
| Subject Sample | # 2 | 0.170 | 00:40 |
| Breath Volume | | 2.234 | 00:40 |
| Air Blank | | 0.000 | 00:41 |
| Control Test | | 0.079 | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Diagnostics Check | | OK | 00:42 |

Data Download Date/Time: 03/12/2020 03:31

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY,WILLIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: METHVIN

DATE/TIME OF TEST: 03/10/2020 03:24:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:16 |
| Air Blank | | 0.000 | 03:17 |
| Control Test | | 0.079 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Subject Sample | # 1 | 0.197 | 03:18 |
| Breath Volume | | 2.000 | 03:18 |
| Air Blank | | 0.000 | 03:19 |
| Air Blank | | 0.000 | 03:21 |
| Subject Sample | # 2 | 0.198 | 03:22 |
| Breath Volume | | 2.347 | 03:22 |
| Air Blank | | 0.000 | 03:23 |
| Control Test | | 0.080 | 03:23 |
| Air Blank | | 0.000 | 03:24 |
| Diagnostics Check | | OK | 03:24 |

Data Download Date/Time: 03/12/2020 03:31

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: BUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 03/13/2020 17:45:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 17:33 | |
| Air Blank | | 0.000 | 17:34 | |
| Control Test | | 0.079 | 17:34 | |
| Air Blank | | 0.000 | 17:35 | |
| Subject Sample | # 1 | | 17:38 | No Sample Provided |
| Breath Volume | | 0.000 | 17:38 | |
| Air Blank | | 0.000 | 17:38 | |
| Air Blank | | 0.000 | 17:40 | |
| Subject Sample | # 2 | | 17:43 | No Sample Provided |
| Breath Volume | | 0.000 | 17:43 | |
| Air Blank | | 0.000 | 17:44 | |
| Control Test | | 0.080 | 17:44 | |
| Air Blank | | 0.000 | 17:45 | |
| Diagnostics Check | | OK | 17:45 | |

Data Download Date/Time: 04/27/2020 13:40


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREMRU,KIP,E
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BIGLIN,AMANDA
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 02/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: BATY

DATE/TIME OF TEST: 03/03/2020 19:52:39

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:46 | |
| Air Blank | | 0.000 | 19:47 | |
| Control Test | | 0.079 | 19:47 | |
| Air Blank | | 0.000 | 19:48 | |
| Subject Sample | # 1 | | 19:51 | Subject Test Refused |
| Breath Volume | | 0.000 | 19:51 | |
| Air Blank | | 0.000 | 19:51 | |
| Control Test | | 0.079 | 19:52 | |
| Air Blank | | 0.000 | 19:52 | |
| Diagnostics Check | | OK | 19:52 | |

Data Download Date/Time: 03/20/2020 13:06

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,KIMBERLY,L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: OSTROSKI,NICOLE
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 02/21/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: WINDHAM-MILTON

DATE/TIME OF TEST: 03/18/2020 21:51:12

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:38 | |
| Air Blank | | 0.000 | 21:39 | |
| Control Test | | 0.079 | 21:39 | |
| Air Blank | | 0.000 | 21:39 | |
| Subject Sample | # 1 | 0.165 | 21:40 | No .020 agreement |
| Breath Volume | | 2.511 | 21:40 | |
| Air Blank | | 0.000 | 21:41 | |
| Air Blank | | 0.000 | 21:43 | |
| Subject Sample | # 2 | 0.156 | 21:46 | Volume Not Met |
| Breath Volume | | 0.035 | 21:46 | |
| Air Blank | | 0.000 | 21:47 | |
| Air Blank | | 0.000 | 21:48 | |
| Subject Sample | # 3 | 0.141 | 21:49 | No .020 agreement |
| Breath Volume | | 2.140 | 21:49 | |
| Air Blank | | 0.000 | 21:50 | |
| Control Test | | 0.077 | 21:50 | |
| Air Blank | | 0.000 | 21:51 | |
| Diagnostics Check | | OK | 21:51 | |

Data Download Date/Time: 03/20/2020 13:06

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STYLES,LAUREN,N
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: STYLES,LAUREN
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 02/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14219080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BRACCI

DATE/TIME OF TEST: 03/19/2020 05:12:49

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:08 | |
| Air Blank | | 0.000 | 05:08 | |
| Control Test | | 0.079 | 05:09 | |
| Air Blank | | 0.000 | 05:09 | |
| Subject Sample | # 1 | | 05:11 | Subject Test Refused |
| Breath Volume | | 0.000 | 05:11 | |
| Air Blank | | 0.000 | 05:11 | |
| Control Test | | 0.080 | 05:12 | |
| Air Blank | | 0.000 | 05:12 | |
| Diagnostics Check | | OK | 05:12 | |

Data Download Date/Time: 03/20/2020 13:06

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,KYLE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HEARN,SHELDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/03/2020 17:02:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:54 |
| Air Blank | | 0.000 | 16:55 |
| Control Test | | 0.078 | 16:55 |
| Air Blank | | 0.000 | 16:55 |
| Subject Sample | # 1 | 0.230 | 16:57 |
| Breath Volume | | 2.250 | 16:57 |
| Air Blank | | 0.000 | 16:58 |
| Air Blank | | 0.000 | 17:00 |
| Subject Sample | # 2 | 0.231 | 17:00 |
| Breath Volume | | 2.328 | 17:00 |
| Air Blank | | 0.000 | 17:01 |
| Control Test | | 0.076 | 17:02 |
| Air Blank | | 0.000 | 17:02 |
| Diagnostics Check | | OK | 17:02 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MCMAHON,KENNETH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: CURRAN

DATE/TIME OF TEST: 03/05/2020 00:46:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:40 |
| Air Blank | | 0.000 | 00:40 |
| Control Test | | 0.078 | 00:40 |
| Air Blank | | 0.000 | 00:41 |
| Subject Sample | # 1 | 0.143 | 00:41 |
| Breath Volume | | 2.250 | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Air Blank | | 0.000 | 00:44 |
| Subject Sample | # 2 | 0.132 | 00:44 |
| Breath Volume | | 1.644 | 00:44 |
| Air Blank | | 0.000 | 00:45 |
| Control Test | | 0.078 | 00:45 |
| Air Blank | | 0.000 | 00:46 |
| Diagnostics Check | | OK | 00:46 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 03/06/2020 04:44:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:37 |
| Air Blank | | 0.000 | 04:37 |
| Control Test | | 0.079 | 04:38 |
| Air Blank | | 0.000 | 04:38 |
| Subject Sample | # 1 | 0.141 | 04:39 |
| Breath Volume | | 3.523 | 04:39 |
| Air Blank | | 0.000 | 04:40 |
| Air Blank | | 0.000 | 04:42 |
| Subject Sample | # 2 | 0.146 | 04:43 |
| Breath Volume | | 3.789 | 04:43 |
| Air Blank | | 0.000 | 04:43 |
| Control Test | | 0.078 | 04:44 |
| Air Blank | | 0.000 | 04:44 |
| Diagnostics Check | | OK | 04:44 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HARMON,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MIKOLACZYK

DATE/TIME OF TEST: 03/07/2020 04:01:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:47 | |
| Air Blank | | 0.000 | 03:47 | |
| Control Test | | 0.080 | 03:47 | |
| Air Blank | | 0.000 | 03:48 | |
| Subject Sample | # 1 | 0.185 | 03:49 | |
| Breath Volume | | 1.109 | 03:49 | |
| Air Blank | | 0.000 | 03:49 | |
| Air Blank | | 0.000 | 03:51 | |
| Subject Sample | # 2 | | 03:54 | No Sample Provided |
| Breath Volume | | 0.000 | 03:54 | |
| Air Blank | | 0.000 | 03:55 | |
| Air Blank | | 0.000 | 03:57 | |
| Subject Sample | # 3 | | 04:00 | No Sample Provided |
| Breath Volume | | 0.000 | 04:00 | |
| Air Blank | | 0.000 | 04:00 | |
| Control Test | | 0.080 | 04:01 | |
| Air Blank | | 0.000 | 04:01 | |
| Diagnostics Check | | OK | 04:01 | |

Data Download Date/Time: 03/29/2020 03:10


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/07/2020 04:13:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:07 |
| Air Blank | | 0.000 | 04:07 |
| Control Test | | 0.079 | 04:07 |
| Air Blank | | 0.000 | 04:08 |
| Subject Sample | # 1 | 0.128 | 04:08 |
| Breath Volume | | 1.140 | 04:08 |
| Air Blank | | 0.000 | 04:09 |
| Air Blank | | 0.000 | 04:11 |
| Subject Sample | # 2 | 0.123 | 04:12 |
| Breath Volume | | 1.128 | 04:12 |
| Air Blank | | 0.000 | 04:12 |
| Control Test | | 0.079 | 04:13 |
| Air Blank | | 0.000 | 04:13 |
| Diagnostics Check | | OK | 04:13 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HARMON,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MIKOLACZYK

DATE/TIME OF TEST: 03/07/2020 04:24:17

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 04:22 | |
| Air Blank | 0.000 | 04:23 | |
| Control Test | 0.073 | 04:23 | Control Outside Tolerance |
| Air Blank | 0.000 | 04:24 | |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HARMON,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MIKOLACZYK

DATE/TIME OF TEST: 03/07/2020 04:33:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:26 |
| Air Blank | | 0.000 | 04:27 |
| Control Test | | 0.076 | 04:27 |
| Air Blank | | 0.000 | 04:27 |
| Subject Sample | # 1 | 0.182 | 04:28 |
| Breath Volume | | 2.773 | 04:28 |
| Air Blank | | 0.000 | 04:29 |
| Air Blank | | 0.000 | 04:31 |
| Subject Sample | # 2 | 0.181 | 04:31 |
| Breath Volume | | 2.832 | 04:31 |
| Air Blank | | 0.000 | 04:32 |
| Control Test | | 0.076 | 04:32 |
| Air Blank | | 0.000 | 04:33 |
| Diagnostics Check | | OK | 04:33 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHASE,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 03/07/2020 15:33:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:26 |
| Air Blank | | 0.000 | 15:27 |
| Control Test | | 0.078 | 15:27 |
| Air Blank | | 0.000 | 15:27 |
| Subject Sample | # 1 | 0.158 | 15:28 |
| Breath Volume | | 1.757 | 15:28 |
| Air Blank | | 0.000 | 15:29 |
| Air Blank | | 0.000 | 15:31 |
| Subject Sample | # 2 | 0.160 | 15:31 |
| Breath Volume | | 1.480 | 15:31 |
| Air Blank | | 0.000 | 15:32 |
| Control Test | | 0.078 | 15:32 |
| Air Blank | | 0.000 | 15:32 |
| Diagnostics Check | | OK | 15:33 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRASER,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HEBEBRAND,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: DAVEY

DATE/TIME OF TEST: 03/10/2020 21:31:58

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:27 | |
| Air Blank | | 0.000 | 21:27 | |
| Control Test | | 0.078 | 21:28 | |
| Air Blank | | 0.000 | 21:28 | |
| Subject Sample | # 1 | | 21:30 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:30 | |
| Air Blank | | 0.000 | 21:31 | |
| Control Test | | 0.079 | 21:31 | |
| Air Blank | | 0.000 | 21:31 | |
| Diagnostics Check | | OK | 21:31 | |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MCMAHON,KENNETH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 03/12/2020 21:22:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:16 |
| Air Blank | | 0.000 | 21:16 |
| Control Test | | 0.079 | 21:17 |
| Air Blank | | 0.000 | 21:17 |
| Subject Sample | # 1 | 0.027 | 21:18 |
| Breath Volume | | 1.347 | 21:18 |
| Air Blank | | 0.000 | 21:18 |
| Air Blank | | 0.000 | 21:20 |
| Subject Sample | # 2 | 0.028 | 21:21 |
| Breath Volume | | 2.347 | 21:21 |
| Air Blank | | 0.000 | 21:22 |
| Control Test | | 0.079 | 21:22 |
| Air Blank | | 0.000 | 21:22 |
| Diagnostics Check | | OK | 21:22 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,KYLE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SKIPPER,KYLE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: HANNER

DATE/TIME OF TEST: 03/13/2020 02:48:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:36 | |
| Air Blank | | 0.000 | 02:37 | |
| Control Test | | 0.078 | 02:37 | |
| Air Blank | | 0.000 | 02:38 | |
| Subject Sample | # 1 | | 02:41 | No Sample Provided |
| Breath Volume | | 0.000 | 02:41 | |
| Air Blank | | 0.000 | 02:41 | |
| Air Blank | | 0.000 | 02:43 | |
| Subject Sample | # 2 | | 02:46 | No Sample Provided |
| Breath Volume | | 0.000 | 02:46 | |
| Air Blank | | 0.000 | 02:47 | |
| Control Test | | 0.078 | 02:47 | |
| Air Blank | | 0.000 | 02:48 | |
| Diagnostics Check | | OK | 02:48 | |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MCMAHON,KENNETH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: KRISHER

DATE/TIME OF TEST: 03/14/2020 03:13:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Control Test | | 0.079 | 03:08 |
| Air Blank | | 0.000 | 03:09 |
| Subject Sample | # 1 | 0.126 | 03:09 |
| Breath Volume | | 1.992 | 03:09 |
| Air Blank | | 0.000 | 03:10 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 2 | 0.117 | 03:12 |
| Breath Volume | | 3.386 | 03:12 |
| Air Blank | | 0.000 | 03:12 |
| Control Test | | 0.078 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Diagnostics Check | | OK | 03:13 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TASCHETTA,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TASCHETTA,JOSEPH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: SABOL

DATE/TIME OF TEST: 03/14/2020 04:16:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:10 |
| Air Blank | | 0.000 | 04:10 |
| Control Test | | 0.078 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Subject Sample | # 1 | 0.198 | 04:11 |
| Breath Volume | | 4.601 | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Air Blank | | 0.000 | 04:14 |
| Subject Sample | # 2 | 0.191 | 04:15 |
| Breath Volume | | 1.691 | 04:15 |
| Air Blank | | 0.000 | 04:15 |
| Control Test | | 0.077 | 04:16 |
| Air Blank | | 0.000 | 04:16 |
| Diagnostics Check | | OK | 04:16 |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACKEY,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/14/2020 15:01:46

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 14:59 | |
| Air Blank | | 0.000 | 14:59 | |
| Control Test | | 0.079 | 15:00 | |
| Air Blank | | 0.000 | 15:00 | |
| Subject Sample | # 1 | | 15:01 | RFI Detect |
| Breath Volume | | 1.042 | 15:01 | |
| Air Blank | | | 15:01 | RFI Detect |
| Air Blank | | 0.000 | 15:01 | |

Data Download Date/Time: 03/29/2020 03:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACKEY,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/14/2020 15:05:07

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 15:04 | |
| Air Blank | 0.000 | 15:04 | |
| | | 15:04 | RFI Detect |
| Air Blank | 0.000 | 15:05 | |

Data Download Date/Time: 03/29/2020 03:11

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACKEY,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/14/2020 15:08:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 15:06 | |
| Air Blank | | 0.000 | 15:07 | |
| Control Test | | 0.079 | 15:07 | |
| Air Blank | | 0.000 | 15:07 | |
| Subject Sample | # 1 | | 15:08 | RFI Detect |
| Breath Volume | | 0.000 | 15:08 | |
| Air Blank | | | 15:08 | RFI Detect |
| Air Blank | | 0.000 | 15:08 | |

Data Download Date/Time: 03/29/2020 03:11

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACKEY,NICHOLAS',C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/14/2020 15:16:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:10 |
| Air Blank | | 0.000 | 15:10 |
| Control Test | | 0.078 | 15:11 |
| Air Blank | | 0.000 | 15:11 |
| Subject Sample | # 1 | 0.198 | 15:11 |
| Breath Volume | | 2.761 | 15:11 |
| Air Blank | | 0.000 | 15:12 |
| Air Blank | | 0.000 | 15:14 |
| Subject Sample | # 2 | 0.199 | 15:14 |
| Breath Volume | | 2.316 | 15:14 |
| Air Blank | | 0.000 | 15:15 |
| Control Test | | 0.078 | 15:15 |
| Air Blank | | 0.000 | 15:16 |
| Diagnostics Check | | OK | 15:16 |

Data Download Date/Time: 03/29/2020 03:11

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEARN,SHELDON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HEARN,SHELDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 02/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: ATHANASIOS

DATE/TIME OF TEST: 03/20/2020 08:18:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:10 |
| Air Blank | | 0.000 | 08:11 |
| Control Test | | 0.079 | 08:11 |
| Air Blank | | 0.000 | 08:12 |
| Subject Sample | # 1 | 0.000 | 08:13 |
| Breath Volume | | 2.074 | 08:13 |
| Air Blank | | 0.000 | 08:14 |
| Air Blank | | 0.000 | 08:16 |
| Subject Sample | # 2 | 0.000 | 08:16 |
| Breath Volume | | 1.996 | 08:16 |
| Air Blank | | 0.000 | 08:17 |
| Control Test | | 0.078 | 08:17 |
| Air Blank | | 0.000 | 08:17 |
| Diagnostics Check | | OK | 08:18 |

Data Download Date/Time: 03/29/2020 03:11

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LEACH,ZACHARY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 03/29/2020 23:51:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:44 |
| Air Blank | | 0.000 | 23:45 |
| Control Test | | 0.079 | 23:45 |
| Air Blank | | 0.000 | 23:46 |
| Subject Sample | # 1 | 0.000 | 23:46 |
| Breath Volume | | 3.832 | 23:46 |
| Air Blank | | 0.000 | 23:47 |
| Air Blank | | 0.000 | 23:49 |
| Subject Sample | # 2 | 0.000 | 23:49 |
| Breath Volume | | 2.718 | 23:49 |
| Air Blank | | 0.000 | 23:50 |
| Control Test | | 0.078 | 23:50 |
| Air Blank | | 0.000 | 23:51 |
| Diagnostics Check | | OK | 23:51 |

Data Download Date/Time: 04/21/2020 00:48

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: PRESTON,MIKAYLA
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: KOPKA

DATE/TIME OF TEST: 03/01/2020 23:47:03

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:35 | |
| Air Blank | | 0.000 | 23:36 | |
| Control Test | | 0.081 | 23:36 | |
| Air Blank | | 0.000 | 23:36 | |
| Subject Sample | # 1 | | 23:40 | No Sample Provided |
| Breath Volume | | 0.000 | 23:40 | |
| Air Blank | | 0.000 | 23:40 | |
| Air Blank | | 0.000 | 23:42 | |
| Subject Sample | # 2 | | 23:45 | No Sample Provided |
| Breath Volume | | 0.000 | 23:45 | |
| Air Blank | | 0.000 | 23:46 | |
| Control Test | | 0.080 | 23:46 | |
| Air Blank | | 0.000 | 23:46 | |
| Diagnostics Check | | OK | 23:47 | |

Data Download Date/Time: 03/25/2020 14:24

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SHOWALTER

DATE/TIME OF TEST: 03/02/2020 21:01:27

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:54 | |
| Air Blank | | 0.000 | 20:54 | |
| Control Test | | 0.079 | 20:55 | |
| Air Blank | | 0.000 | 20:55 | |
| Subject Sample | # 1 | 0.212 | 20:56 | |
| Breath Volume | | 1.214 | 20:56 | |
| Air Blank | | 0.000 | 20:56 | |
| Air Blank | | 0.000 | 20:58 | |
| Subject Sample | # 2 | | 20:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:59 | |
| Air Blank | | 0.000 | 21:00 | |
| Control Test | | 0.077 | 21:00 | |
| Air Blank | | 0.000 | 21:01 | |
| Diagnostics Check | | OK | 21:01 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: MCNAMARA

DATE/TIME OF TEST: 03/03/2020 19:17:41

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:06 | |
| Air Blank | | 0.000 | 19:07 | |
| Control Test | | 0.077 | 19:07 | |
| Air Blank | | 0.000 | 19:08 | |
| Subject Sample | # 1 | 0.094 | 19:11 | Volume Not Met |
| Breath Volume | | 0.492 | 19:11 | |
| Air Blank | | 0.000 | 19:12 | |
| Air Blank | | 0.000 | 19:13 | |
| Subject Sample | # 2 | | 19:16 | Subject Test Refused |
| Breath Volume | | 0.535 | 19:16 | |
| Air Blank | | 0.000 | 19:16 | |
| Control Test | | 0.074 | 19:17 | Control Outside Tolerance |
| Air Blank | | 0.000 | 19:17 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: RICHERHAGEN

DATE/TIME OF TEST: 03/04/2020 03:31:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:23 |
| Air Blank | | 0.000 | 03:23 |
| Control Test | | 0.076 | 03:24 |
| Air Blank | | 0.000 | 03:24 |
| Subject Sample | # 1 | 0.127 | 03:25 |
| Breath Volume | | 1.511 | 03:25 |
| Air Blank | | 0.000 | 03:25 |
| Air Blank | | 0.000 | 03:27 |
| Subject Sample | # 2 | 0.123 | 03:29 |
| Breath Volume | | 1.601 | 03:29 |
| Air Blank | | 0.000 | 03:30 |
| Control Test | | 0.077 | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Diagnostics Check | | OK | 03:31 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORALUZZO,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 10:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: GABLE

DATE/TIME OF TEST: 03/05/2020 11:23:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:17 |
| Air Blank | | 0.000 | 11:18 |
| Control Test | | 0.079 | 11:18 |
| Air Blank | | 0.000 | 11:19 |
| Subject Sample | # 1 | 0.000 | 11:19 |
| Breath Volume | | 1.195 | 11:19 |
| Air Blank | | 0.000 | 11:19 |
| Air Blank | | 0.000 | 11:21 |
| Subject Sample | # 2 | 0.000 | 11:22 |
| Breath Volume | | 1.398 | 11:22 |
| Air Blank | | 0.000 | 11:22 |
| Control Test | | 0.077 | 11:23 |
| Air Blank | | 0.000 | 11:23 |
| Diagnostics Check | | OK | 11:23 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/06/2020 22:22:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:13 |
| Air Blank | | 0.000 | 22:14 |
| Control Test | | 0.080 | 22:14 |
| Air Blank | | 0.000 | 22:15 |
| Subject Sample | # 1 | 0.334 | 22:15 |
| Breath Volume | | 1.210 | 22:15 |
| Air Blank | | 0.000 | 22:16 |
| Air Blank | | 0.000 | 22:18 |
| Subject Sample | # 2 | 0.309 | 22:18 |
| Breath Volume | | 1.148 | 22:18 |
| Air Blank | | 0.000 | 22:19 |
| Air Blank | | 0.000 | 22:20 |
| Subject Sample | # 3 | 0.315 | 22:21 |
| Breath Volume | | 1.464 | 22:21 |
| Air Blank | | 0.000 | 22:22 |
| Control Test | | 0.077 | 22:22 |
| Air Blank | | 0.000 | 22:22 |
| Diagnostics Check | | OK | 22:22 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: STURGEON

DATE/TIME OF TEST: 03/07/2020 21:26:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:20 |
| Air Blank | | 0.000 | 21:20 |
| Control Test | | 0.080 | 21:21 |
| Air Blank | | 0.000 | 21:21 |
| Subject Sample | # 1 | 0.000 | 21:22 |
| Breath Volume | | 2.859 | 21:22 |
| Air Blank | | 0.000 | 21:22 |
| Air Blank | | 0.000 | 21:24 |
| Subject Sample | # 2 | 0.000 | 21:25 |
| Breath Volume | | 2.824 | 21:25 |
| Air Blank | | 0.000 | 21:25 |
| Control Test | | 0.077 | 21:25 |
| Air Blank | | 0.000 | 21:26 |
| Diagnostics Check | | OK | 21:26 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: YOUNGS

DATE/TIME OF TEST: 03/08/2020 04:08:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:01 |
| Air Blank | | 0.000 | 04:01 |
| Control Test | | 0.080 | 04:02 |
| Air Blank | | 0.000 | 04:02 |
| Subject Sample | # 1 | 0.193 | 04:03 |
| Breath Volume | | 1.417 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Air Blank | | 0.000 | 04:06 |
| Subject Sample | # 2 | 0.176 | 04:06 |
| Breath Volume | | 1.527 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Control Test | | 0.077 | 04:07 |
| Air Blank | | 0.000 | 04:08 |
| Diagnostics Check | | OK | 04:08 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 03/08/2020 22:13:16

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:06 | |
| Air Blank | | 0.000 | 22:06 | |
| Control Test | | 0.080 | 22:07 | |
| Air Blank | | 0.000 | 22:07 | |
| Subject Sample | # 1 | 0.106 | 22:08 | |
| Breath Volume | | 2.203 | 22:08 | |
| Air Blank | | 0.000 | 22:09 | |
| Air Blank | | 0.000 | 22:11 | |
| Subject Sample | # 2 | 0.096 | 22:11 | |
| Breath Volume | | 2.261 | 22:11 | |
| Air Blank | | 0.000 | 22:12 | |
| Control Test | | 0.071 | 22:12 | Control Outside Tolerance |
| Air Blank | | 0.000 | 22:13 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 03/08/2020 22:20:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:14 |
| Air Blank | | 0.000 | 22:15 |
| Control Test | | 0.077 | 22:15 |
| Air Blank | | 0.000 | 22:15 |
| Subject Sample | # 1 | 0.098 | 22:16 |
| Breath Volume | | 1.375 | 22:16 |
| Air Blank | | 0.000 | 22:17 |
| Air Blank | | 0.000 | 22:18 |
| Subject Sample | # 2 | 0.100 | 22:19 |
| Breath Volume | | 1.726 | 22:19 |
| Air Blank | | 0.000 | 22:19 |
| Control Test | | 0.077 | 22:20 |
| Air Blank | | 0.000 | 22:20 |
| Diagnostics Check | | OK | 22:20 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PASSOS

DATE/TIME OF TEST: 03/09/2020 03:31:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:24 |
| Air Blank | | 0.000 | 03:25 |
| Control Test | | 0.079 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Subject Sample | # 1 | 0.064 | 03:26 |
| Breath Volume | | 2.035 | 03:26 |
| Air Blank | | 0.000 | 03:27 |
| Air Blank | | 0.000 | 03:29 |
| Subject Sample | # 2 | 0.066 | 03:29 |
| Breath Volume | | 3.031 | 03:29 |
| Air Blank | | 0.000 | 03:30 |
| Control Test | | 0.077 | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Diagnostics Check | | OK | 03:31 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CHAMBLIN

DATE/TIME OF TEST: 03/09/2020 03:51:15

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:49 | |
| Air Blank | | 0.000 | 03:49 | |
| Control Test | | 0.078 | 03:49 | |
| Air Blank | | 0.000 | 03:50 | |
| Subject Sample | # 1 | | 03:50 | RFI Detect |
| Breath Volume | | 0.000 | 03:50 | |
| Air Blank | | | 03:50 | RFI Detect |
| Air Blank | | 0.000 | 03:51 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: CHAMBLIN

DATE/TIME OF TEST: 03/09/2020 03:59:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:52 |
| Air Blank | | 0.000 | 03:53 |
| Control Test | | 0.078 | 03:53 |
| Air Blank | | 0.000 | 03:53 |
| Subject Sample | # 1 | 0.226 | 03:54 |
| Breath Volume | | 1.421 | 03:54 |
| Air Blank | | 0.000 | 03:55 |
| Air Blank | | 0.000 | 03:57 |
| Subject Sample | # 2 | 0.225 | 03:57 |
| Breath Volume | | 1.167 | 03:57 |
| Air Blank | | 0.000 | 03:58 |
| Control Test | | 0.076 | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Diagnostics Check | | OK | 03:59 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,SCOTT,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BLOCH

DATE/TIME OF TEST: 03/10/2020 03:38:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:31 |
| Air Blank | | 0.000 | 03:31 |
| Control Test | | 0.078 | 03:32 |
| Air Blank | | 0.000 | 03:32 |
| Subject Sample | # 1 | 0.136 | 03:33 |
| Breath Volume | | 2.429 | 03:33 |
| Air Blank | | 0.000 | 03:34 |
| Air Blank | | 0.000 | 03:36 |
| Subject Sample | # 2 | 0.131 | 03:36 |
| Breath Volume | | 2.332 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Control Test | | 0.077 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Diagnostics Check | | OK | 03:38 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORTER,DENNIS,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: SCHWEINSBERG

DATE/TIME OF TEST: 03/11/2020 20:02:49

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:59 | |
| Air Blank | | 0.000 | 20:00 | |
| Control Test | | 0.080 | 20:00 | |
| Air Blank | | 0.000 | 20:00 | |
| Subject Sample | # 1 | | 20:01 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:01 | |
| Air Blank | | 0.000 | 20:01 | |
| Control Test | | 0.078 | 20:02 | |
| Air Blank | | 0.000 | 20:02 | |
| Diagnostics Check | | OK | 20:02 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: HIATT

DATE/TIME OF TEST: 03/11/2020 23:55:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:49 |
| Air Blank | | 0.000 | 23:49 |
| Control Test | | 0.079 | 23:49 |
| Air Blank | | 0.000 | 23:50 |
| Subject Sample | # 1 | 0.216 | 23:50 |
| Breath Volume | | 2.707 | 23:50 |
| Air Blank | | 0.000 | 23:51 |
| Air Blank | | 0.000 | 23:53 |
| Subject Sample | # 2 | 0.227 | 23:53 |
| Breath Volume | | 3.140 | 23:53 |
| Air Blank | | 0.000 | 23:54 |
| Control Test | | 0.077 | 23:54 |
| Air Blank | | 0.000 | 23:55 |
| Diagnostics Check | | OK | 23:55 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 10:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LESKO

DATE/TIME OF TEST: 03/12/2020 23:42:39

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:39 | |
| Air Blank | | 0.000 | 23:40 | |
| Control Test | | 0.079 | 23:40 | |
| Air Blank | | 0.000 | 23:40 | |
| Subject Sample | # 1 | | 23:41 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:41 | |
| Air Blank | | 0.000 | 23:41 | |
| Control Test | | 0.078 | 23:42 | |
| Air Blank | | 0.000 | 23:42 | |
| Diagnostics Check | | OK | 23:42 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MC GETRICK

DATE/TIME OF TEST: 03/13/2020 00:01:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:50 | |
| Air Blank | | 0.000 | 23:50 | |
| Control Test | | 0.077 | 23:51 | |
| Air Blank | | 0.000 | 23:51 | |
| Subject Sample | # 2 | | 23:54 | No Sample Provided |
| Breath Volume | | 0.000 | 23:54 | |
| Air Blank | | 0.000 | 23:55 | |
| Air Blank | | 0.000 | 23:57 | |
| Subject Sample | # 1 | | 00:00 | No Sample Provided |
| Breath Volume | | 0.000 | 00:00 | |
| Air Blank | | 0.000 | 00:00 | |
| Control Test | | 0.078 | 00:01 | |
| Air Blank | | 0.000 | 00:01 | |
| Diagnostics Check | | OK | 00:01 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 12/21/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: VEENSTRA

DATE/TIME OF TEST: 03/13/2020 04:37:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:29 |
| Air Blank | | 0.000 | 04:30 |
| Control Test | | 0.078 | 04:30 |
| Air Blank | | 0.000 | 04:31 |
| Subject Sample | # 1 | 0.155 | 04:32 |
| Breath Volume | | 1.757 | 04:32 |
| Air Blank | | 0.000 | 04:32 |
| Air Blank | | 0.000 | 04:34 |
| Subject Sample | # 2 | 0.160 | 04:35 |
| Breath Volume | | 1.855 | 04:35 |
| Air Blank | | 0.000 | 04:36 |
| Control Test | | 0.076 | 04:36 |
| Air Blank | | 0.000 | 04:37 |
| Diagnostics Check | | OK | 04:37 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIANNOTTA,DOMINIC,A
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: GIANNOTTA,DOMINIC
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SPANGLER

DATE/TIME OF TEST: 03/14/2020 00:17:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:10 |
| Air Blank | | 0.000 | 00:11 |
| Control Test | | 0.079 | 00:11 |
| Air Blank | | 0.000 | 00:12 |
| Subject Sample | # 1 | 0.112 | 00:13 |
| Breath Volume | | 2.644 | 00:13 |
| Air Blank | | 0.000 | 00:13 |
| Air Blank | | 0.000 | 00:15 |
| Subject Sample | # 2 | 0.098 | 00:16 |
| Breath Volume | | 3.917 | 00:16 |
| Air Blank | | 0.000 | 00:16 |
| Control Test | | 0.077 | 00:17 |
| Air Blank | | 0.000 | 00:17 |
| Diagnostics Check | | OK | 00:17 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 03/21/2020 00:19:27

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:10 | |
| Air Blank | | 0.000 | 00:11 | |
| Control Test | | 0.078 | 00:11 | |
| Air Blank | | 0.000 | 00:11 | |
| Subject Sample | # 1 | | 00:15 | No Sample Provided |
| Breath Volume | | 0.000 | 00:15 | |
| Air Blank | | 0.000 | 00:15 | |
| Air Blank | | 0.000 | 00:17 | |
| Subject Sample | # 2 | | 00:18 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:18 | |
| Air Blank | | 0.000 | 00:18 | |
| Control Test | | 0.078 | 00:18 | |
| Air Blank | | 0.000 | 00:19 | |
| Diagnostics Check | | OK | 00:19 | |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/22/2020 19:58:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:51 |
| Air Blank | | 0.000 | 19:52 |
| Control Test | | 0.079 | 19:52 |
| Air Blank | | 0.000 | 19:53 |
| Subject Sample | # 1 | 0.000 | 19:53 |
| Breath Volume | | 1.265 | 19:53 |
| Air Blank | | 0.000 | 19:54 |
| Air Blank | | 0.000 | 19:56 |
| Subject Sample | # 2 | 0.000 | 19:56 |
| Breath Volume | | 1.578 | 19:56 |
| Air Blank | | 0.000 | 19:57 |
| Control Test | | 0.077 | 19:57 |
| Air Blank | | 0.000 | 19:58 |
| Diagnostics Check | | OK | 19:58 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,BRIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ELDER,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: DIECKMAN

DATE/TIME OF TEST: 03/25/2020 11:13:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:06 |
| Air Blank | | 0.000 | 11:07 |
| Control Test | | 0.079 | 11:07 |
| Air Blank | | 0.000 | 11:07 |
| Subject Sample | # 1 | 0.000 | 11:08 |
| Breath Volume | | 1.968 | 11:08 |
| Air Blank | | 0.000 | 11:09 |
| Air Blank | | 0.000 | 11:11 |
| Subject Sample | # 2 | 0.000 | 11:12 |
| Breath Volume | | 3.667 | 11:12 |
| Air Blank | | 0.000 | 11:12 |
| Control Test | | 0.078 | 11:13 |
| Air Blank | | 0.000 | 11:13 |
| Diagnostics Check | | OK | 11:13 |

Data Download Date/Time: 03/25/2020 14:25

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRSHIKOVA,ALINA,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: TUSHA,JYLTON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1233
DRY GAS EXPIRY DATE: 02/23/2020
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: HAMANN

DATE/TIME OF TEST: 03/29/2020 18:48:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:41 |
| Air Blank | | 0.000 | 18:41 |
| Control Test | | 0.079 | 18:41 |
| Air Blank | | 0.000 | 18:42 |
| Subject Sample | # 1 | 0.198 | 18:43 |
| Breath Volume | | 2.253 | 18:43 |
| Air Blank | | 0.000 | 18:44 |
| Air Blank | | 0.000 | 18:46 |
| Subject Sample | # 2 | 0.184 | 18:46 |
| Breath Volume | | 1.613 | 18:46 |
| Air Blank | | 0.000 | 18:47 |
| Control Test | | 0.075 | 18:47 |
| Air Blank | | 0.000 | 18:48 |
| Diagnostics Check | | OK | 18:48 |

Data Download Date/Time: 04/28/2020 09:30

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LAMBERT,JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/01/2020 01:20:10

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:12 | |
| Air Blank | | 0.000 | 01:12 | |
| Control Test | | 0.078 | 01:13 | |
| Air Blank | | 0.000 | 01:13 | |
| Subject Sample | # 1 | 0.128 | 01:14 | |
| Breath Volume | | 1.976 | 01:14 | |
| Air Blank | | 0.000 | 01:15 | |
| Air Blank | | 0.000 | 01:16 | |
| Subject Sample | # 2 | 0.124 | 01:18 | |
| Breath Volume | | 1.417 | 01:18 | |
| Air Blank | | 0.000 | 01:19 | |
| Control Test | | 0.073 | 01:19 | Control Outside Tolerance |
| Air Blank | | 0.000 | 01:20 | |

Data Download Date/Time: 03/12/2020 06:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEOD,JOEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCLEOD,JOEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: SWITZER

DATE/TIME OF TEST: 03/02/2020 19:57:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:55 | |
| Air Blank | | 0.000 | 19:55 | |
| Control Test | | 0.077 | 19:55 | |
| Air Blank | | 0.000 | 19:56 | |
| Subject Sample | # 1 | | 19:56 | Subject Test Refused |
| Breath Volume | | 0.000 | 19:56 | |
| Air Blank | | 0.000 | 19:57 | |
| Control Test | | 0.076 | 19:57 | |
| Air Blank | | | 19:57 | RFI Detect |
| Air Blank | | 0.000 | 19:57 | |

Data Download Date/Time: 03/12/2020 06:55

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 03/06/2020 01:13:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:05 | |
| Air Blank | | 0.000 | 01:05 | |
| Control Test | | 0.078 | 01:06 | |
| Air Blank | | 0.000 | 01:06 | |
| Subject Sample | # 1 | 0.167 | 01:08 | |
| Breath Volume | | 2.265 | 01:08 | |
| Air Blank | | 0.000 | 01:09 | |
| Air Blank | | 0.000 | 01:10 | |
| Subject Sample | # 2 | 0.164 | 01:11 | |
| Breath Volume | | 1.503 | 01:11 | |
| Air Blank | | 0.000 | 01:12 | |
| Control Test | | 0.072 | 01:12 | Control Outside Tolerance |
| Air Blank | | 0.000 | 01:13 | |

Data Download Date/Time: 03/12/2020 06:56

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 03/06/2020 01:17:01

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 01:15 | |
| Air Blank | 0.000 | 01:16 | |
| Control Test | 0.072 | 01:16 | Control Outside Tolerance |
| Air Blank | 0.000 | 01:17 | |

Data Download Date/Time: 03/12/2020 06:56

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 03/06/2020 01:20:27

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 01:19 | |
| Air Blank | 0.000 | 01:19 | |
| Control Test | 0.074 | 01:19 | Control Outside Tolerance |
| Air Blank | 0.000 | 01:20 | |

Data Download Date/Time: 03/12/2020 06:56

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCALIA,ADAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HIGHFIELD

DATE/TIME OF TEST: 03/09/2020 23:48:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:37 | |
| Air Blank | | 0.000 | 23:37 | |
| Control Test | | 0.077 | 23:38 | |
| Air Blank | | 0.000 | 23:38 | |
| Subject Sample | # 1 | | 23:41 | No Sample Provided |
| Breath Volume | | 0.000 | 23:41 | |
| Air Blank | | 0.000 | 23:42 | |
| Air Blank | | 0.000 | 23:44 | |
| Subject Sample | # 2 | | 23:47 | No Sample Provided |
| Breath Volume | | 0.000 | 23:47 | |
| Air Blank | | 0.000 | 23:47 | |
| Control Test | | 0.077 | 23:48 | |
| Air Blank | | 0.000 | 23:48 | |
| Diagnostics Check | | OK | 23:48 | |

Data Download Date/Time: 03/12/2020 06:56

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,ROBERT,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KENNEDY,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: ORSBORN

DATE/TIME OF TEST: 03/31/2020 17:30:09

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 17:23 | |
| Air Blank | | 0.000 | 17:23 | |
| Control Test | | 0.076 | 17:24 | |
| Air Blank | | 0.000 | 17:24 | |
| Subject Sample | # 1 | 0.296 | 17:25 | |
| Breath Volume | | 2.292 | 17:25 | |
| Air Blank | | 0.000 | 17:26 | |
| Air Blank | | 0.000 | 17:27 | |
| Subject Sample | # 2 | 0.276 | 17:28 | |
| Breath Volume | | 1.773 | 17:28 | |
| Air Blank | | 0.000 | 17:29 | |
| Control Test | | 0.071 | 17:29 | Control Outside Tolerance |
| Air Blank | | 0.000 | 17:30 | |

Data Download Date/Time: 04/22/2020 16:19

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: YARBOROUGH,ERIC
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 03/06/2020 21:59:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:53 |
| Air Blank | | 0.000 | 21:53 |
| Control Test | | 0.079 | 21:53 |
| Air Blank | | 0.000 | 21:54 |
| Subject Sample | # 1 | 0.224 | 21:55 |
| Breath Volume | | 1.773 | 21:55 |
| Air Blank | | 0.000 | 21:55 |
| Air Blank | | 0.000 | 21:57 |
| Subject Sample | # 2 | 0.232 | 21:58 |
| Breath Volume | | 2.332 | 21:58 |
| Air Blank | | 0.000 | 21:58 |
| Control Test | | 0.078 | 21:59 |
| Air Blank | | 0.000 | 21:59 |
| Diagnostics Check | | OK | 21:59 |

Data Download Date/Time: 03/25/2020 14:09

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/08/2020 04:11:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:05 |
| Air Blank | | 0.000 | 04:05 |
| Control Test | | 0.079 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Subject Sample | # 1 | 0.119 | 04:06 |
| Breath Volume | | 3.085 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Air Blank | | 0.000 | 04:09 |
| Subject Sample | # 2 | 0.112 | 04:09 |
| Breath Volume | | 2.636 | 04:09 |
| Air Blank | | 0.000 | 04:10 |
| Control Test | | 0.079 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Diagnostics Check | | OK | 04:11 |

Data Download Date/Time: 03/25/2020 14:09

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: YARBOROUGH,ERIC
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: SCHENK

DATE/TIME OF TEST: 03/13/2020 03:48:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:44 | |
| Air Blank | | 0.000 | 03:44 | |
| Control Test | | 0.079 | 03:44 | |
| Air Blank | | 0.000 | 03:45 | |
| Subject Sample | # 1 | | 03:47 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:47 | |
| Air Blank | | 0.000 | 03:47 | |
| Control Test | | 0.078 | 03:48 | |
| Air Blank | | 0.000 | 03:48 | |
| Diagnostics Check | | OK | 03:48 | |

Data Download Date/Time: 03/25/2020 14:09

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: MCDONNELL

DATE/TIME OF TEST: 03/13/2020 21:29:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:23 |
| Air Blank | | 0.000 | 21:24 |
| Control Test | | 0.079 | 21:24 |
| Air Blank | | 0.000 | 21:25 |
| Subject Sample | # 1 | 0.097 | 21:25 |
| Breath Volume | | 1.847 | 21:25 |
| Air Blank | | 0.000 | 21:26 |
| Air Blank | | 0.000 | 21:27 |
| Subject Sample | # 2 | 0.084 | 21:28 |
| Breath Volume | | 1.109 | 21:28 |
| Air Blank | | 0.000 | 21:28 |
| Control Test | | 0.076 | 21:29 |
| Air Blank | | 0.000 | 21:29 |
| Diagnostics Check | | OK | 21:29 |

Data Download Date/Time: 03/25/2020 14:09

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 03/14/2020 03:38:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Control Test | | 0.079 | 03:31 |
| Air Blank | | 0.000 | 03:32 |
| Subject Sample | # 1 | 0.196 | 03:33 |
| Breath Volume | | 2.597 | 03:33 |
| Air Blank | | 0.000 | 03:34 |
| Air Blank | | 0.000 | 03:36 |
| Subject Sample | # 2 | 0.197 | 03:36 |
| Breath Volume | | 3.320 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Control Test | | 0.077 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Diagnostics Check | | OK | 03:38 |

Data Download Date/Time: 03/25/2020 14:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADO,RYAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RADO,RYAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 03/16/2020 02:09:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:03 |
| Air Blank | | 0.000 | 02:03 |
| Control Test | | 0.079 | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Subject Sample | # 1 | 0.155 | 02:05 |
| Breath Volume | | 2.500 | 02:05 |
| Air Blank | | 0.000 | 02:06 |
| Air Blank | | 0.000 | 02:07 |
| Subject Sample | # 2 | 0.152 | 02:08 |
| Breath Volume | | 1.574 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Control Test | | 0.076 | 02:09 |
| Air Blank | | 0.000 | 02:09 |
| Diagnostics Check | | OK | 02:09 |

Data Download Date/Time: 03/25/2020 14:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,S,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GARMON,B
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: KLINGLER

DATE/TIME OF TEST: 03/20/2020 22:51:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:42 |
| Air Blank | | 0.000 | 22:43 |
| Control Test | | 0.078 | 22:43 |
| Air Blank | | 0.000 | 22:43 |
| Subject Sample | # 1 | 0.211 | 22:44 |
| Breath Volume | | 2.703 | 22:44 |
| Air Blank | | 0.000 | 22:45 |
| Air Blank | | 0.000 | 22:46 |
| Subject Sample | # 2 | 0.190 | 22:47 |
| Breath Volume | | 1.746 | 22:47 |
| Air Blank | | 0.000 | 22:48 |
| Air Blank | | 0.000 | 22:49 |
| Subject Sample | # 3 | 0.195 | 22:50 |
| Breath Volume | | 1.613 | 22:50 |
| Air Blank | | 0.000 | 22:50 |
| Control Test | | 0.075 | 22:51 |
| Air Blank | | 0.000 | 22:51 |
| Diagnostics Check | | OK | 22:51 |

Data Download Date/Time: 03/25/2020 14:10

**Printed: 02 July 2020**

COBRA ©1997-2014 CMI, Inc.