Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS,GUILLERMO,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: TAMARGO,RAMON
ARRESTING OFFICER AGENCY: MIAMI SPRINGS PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 05/21/2020 08:36:33

Diagnostics Check		OK	08:33	
Air Blank		0.000	08:34	
Control Test		0.076	08:34	
Air Blank		0.000	08:35	
Subject Sample	# 1		08:35	Subject Test Refused
Breath Volume		0.000	08:35	
Air Blank		0.000	08:35	
Control Test		0.077	08:35	
Air Blank		0.000	08:36	
Diagnostics Check		OK	08:36	

Data Download Date/Time: 05/28/2020 15:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/03/2020 01:32:51

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1	0.136	01:27	
Breath Volume		3.074	01:27	
Air Blank		0.000	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 2		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:31	
Control Test		0.077	01:32	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 05/28/2020 15:31

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPLACELIERE,LEONARDO,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPLACELIERE,LEONARDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 05/06/2020 18:32:29

Diagnostics Check		OK	18:24
Air Blank		0.000	18:25
Control Test		0.081	18:25
Air Blank		0.000	18:26
Subject Sample	# 1	0.000	18:28
Breath Volume		1.828	18:28
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.000	18:31
Breath Volume		1.484	18:31
Air Blank		0.000	18:31
Control Test		0.079	18:31
Air Blank		0.000	18:32
Diagnostics Check		OK	18:32

Data Download Date/Time: 05/28/2020 15:31

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUBELA,CHANTEL,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: CUBELA,CHANTEL
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: BORJAS ROMERO

DATE/TIME OF TEST: 05/09/2020 08:19:26

Diagnostics Check		OK	08:10
Air Blank		0.000	08:10
Control Test		0.081	08:10
Air Blank		0.000	08:11
Subject Sample	# 1	0.201	08:13
Breath Volume		1.355	08:13
Air Blank		0.000	08:14
Air Blank		0.000	08:16
Subject Sample	# 2	0.207	08:17
Breath Volume		3.335	08:17
Air Blank		0.000	08:18
Control Test		0.077	08:18
Air Blank		0.000	08:19
Diagnostics Check		OK	08:19

Data Download Date/Time: 05/28/2020 15:31

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: A223440823750

DATE/TIME OF TEST: 05/18/2020 13:17:31

Diagnostics Check		OK	13:10
Air Blank		0.000	13:11
Control Test		0.080	13:11
Air Blank		0.000	13:11
Subject Sample	# 1	0.178	13:12
Breath Volume		2.265	13:12
Air Blank		0.000	13:13
Air Blank		0.000	13:15
Subject Sample	# 2	0.167	13:15
Breath Volume		1.648	13:15
Air Blank		0.000	13:16
Control Test		0.078	13:16
Air Blank		0.000	13:17
Diagnostics Check		OK	13:17

Data Download Date/Time: 05/28/2020 15:31

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSCULLUELA,BRYAN,M
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: BEAUTELUS,JOHNNY
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GUARISMA ACUA

DATE/TIME OF TEST: 05/22/2020 02:17:13

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.238	02:12
Breath Volume		2.378	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:14
Subject Sample	# 2	0.225	02:15
Breath Volume		1.800	02:15
Air Blank		0.000	02:16
Control Test		0.077	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 05/28/2020 15:32

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: OSORIA

DATE/TIME OF TEST: 05/29/2020 04:44:24

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:41	
Control Test		0.080	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:43	
Control Test		0.080	04:43	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 06/25/2020 11:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIVEIROS,KYLE,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEREYRA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 05/31/2020 03:39:28

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Control Test		0.081	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 06/25/2020 11:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 05/01/2020 00:57:12

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.091	00:52
Breath Volume		2.074	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.088	00:55
Breath Volume		1.492	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SILAS

DATE/TIME OF TEST: 05/01/2020 04:00:22

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.080	03:58	
Air Blank		0.000	03:58	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: APPLEBY

DATE/TIME OF TEST: 05/02/2020 01:34:49

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.206	01:26
Breath Volume		2.132	01:26
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.185	01:29
Breath Volume		1.410	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 3	0.197	01:33
Breath Volume		1.960	01:33
Air Blank		0.000	01:33
Control Test		0.078	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: SHULER

DATE/TIME OF TEST: 05/02/2020 20:40:26

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:38	
Control Test		0.078	20:38	
Air Blank		0.000	20:38	
Subject Sample	# 1		20:39	Subject Test Refused
Breath Volume		0.000	20:39	
Air Blank		0.000	20:39	
Control Test		0.079	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 05/03/2020 03:27:35

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.114	03:22
Breath Volume		1.351	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.112	03:26
Breath Volume		1.546	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HICKS,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: JASPER

DATE/TIME OF TEST: 05/03/2020 08:39:57

Diagnostics Check		OK	08:33
Air Blank		0.000	08:34
Control Test		0.078	08:34
Air Blank		0.000	08:34
Subject Sample	# 1	0.165	08:35
Breath Volume		2.761	08:35
Air Blank		0.000	08:36
Air Blank		0.000	08:37
Subject Sample	# 2	0.168	08:38
Breath Volume		2.652	08:38
Air Blank		0.000	08:39
Control Test		0.079	08:39
Air Blank		0.000	08:39
Diagnostics Check		OK	08:39

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONK,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: BATTISTA

DATE/TIME OF TEST: 05/03/2020 14:38:29

Diagnostics Check		OK	14:35	
Air Blank		0.000	14:35	
Control Test		0.079	14:36	
Air Blank		0.000	14:36	
Subject Sample	# 1		14:37	Subject Test Refused
Breath Volume		0.000	14:37	
Air Blank		0.000	14:37	
Control Test		0.079	14:37	
Air Blank		0.000	14:38	
Diagnostics Check		OK	14:38	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/04/2020 01:17:33

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.078	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.207	01:12
Breath Volume		1.601	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.196	01:16
Breath Volume		1.195	01:16
Air Blank		0.000	01:16
Control Test		0.078	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LUEDKE

DATE/TIME OF TEST: 05/04/2020 04:04:48

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.079	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KASKI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: GARCIA MUNOZ

DATE/TIME OF TEST: 05/05/2020 16:37:19

Diagnostics Check		OK	16:31
Air Blank		0.000	16:31
Control Test		0.079	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.170	16:32
Breath Volume		1.621	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.171	16:35
Breath Volume		1.968	16:35
Air Blank		0.000	16:36
Control Test		0.079	16:36
Air Blank		0.000	16:37
Diagnostics Check		OK	16:37

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MCBROOM,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:04

SUBJECT'S LAST NAME: PRIESTER

DATE/TIME OF TEST: 05/06/2020 10:31:17

Diagnostics Check		OK	10:24
Air Blank		0.000	10:25
Control Test		0.079	10:25
Air Blank		0.000	10:25
Subject Sample	# 1	0.000	10:27
Breath Volume		2.023	10:27
Air Blank		0.000	10:27
Air Blank		0.000	10:29
Subject Sample	# 2	0.000	10:29
Breath Volume		2.207	10:29
Air Blank		0.000	10:30
Control Test		0.079	10:30
Air Blank		0.000	10:31
Diagnostics Check		OK	10:31

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CHICO MARTINEZ

DATE/TIME OF TEST: 05/09/2020 02:36:47

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.178	02:32
Breath Volume		2.011	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.180	02:35
Breath Volume		3.324	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:51

SUBJECT'S LAST NAME: SALSBURG

DATE/TIME OF TEST: 05/09/2020 14:14:34

Diagnostics Check		OK	14:11	
Air Blank		0.000	14:12	
Control Test		0.079	14:12	
Air Blank		0.000	14:13	
Subject Sample	# 1		14:13	Subject Test Refused
Breath Volume		0.000	14:13	
Air Blank		0.000	14:13	
Control Test		0.080	14:13	
Air Blank		0.000	14:14	
Diagnostics Check		OK	14:14	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: JORANGER

DATE/TIME OF TEST: 05/09/2020 19:43:12

Diagnostics Check		OK	19:37
Air Blank		0.000	19:37
Control Test		0.080	19:38
Air Blank		0.000	19:38
Subject Sample	# 1	0.000	19:38
Breath Volume		3.363	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.000	19:41
Breath Volume		2.265	19:41
Air Blank		0.000	19:42
Control Test		0.080	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/09/2020 22:13:35

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.080	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.103	22:08
Breath Volume		2.039	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.103	22:12
Breath Volume		1.902	22:12
Air Blank		0.000	22:12
Control Test		0.079	22:12
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/10/2020 00:39:41

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:37	
Control Test		0.080	00:37	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:38	Subject Test Refused
Breath Volume		0.000	00:38	
Air Blank		0.000	00:38	
Control Test		0.080	00:39	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DIAMOND

DATE/TIME OF TEST: 05/10/2020 02:20:13

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.079	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.099	02:15
Breath Volume		2.652	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.099	02:18
Breath Volume		2.457	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GAFFNEY

DATE/TIME OF TEST: 05/10/2020 04:17:56

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.080	04:15	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:16	Subject Test Refused
Breath Volume		0.000	04:16	
Air Blank		0.000	04:17	
Control Test		0.080	04:17	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: HERNANDEZ LEON

DATE/TIME OF TEST: 05/10/2020 20:35:27

Diagnostics Check		OK	20:29
Air Blank		0.000	20:29
Control Test		0.080	20:30
Air Blank		0.000	20:30
Subject Sample	# 1	0.234	20:31
Breath Volume		1.734	20:31
Air Blank		0.000	20:31
Air Blank		0.000	20:33
Subject Sample	# 2	0.223	20:33
Breath Volume		2.820	20:33
Air Blank		0.000	20:34
Control Test		0.079	20:34
Air Blank		0.000	20:35
Diagnostics Check		OK	20:35

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KUUSELA,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 05/11/2020 02:30:21

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.263	02:25
Breath Volume		2.160	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.244	02:28
Breath Volume		1.835	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GOODE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:01

SUBJECT'S LAST NAME: WITTMAN

DATE/TIME OF TEST: 05/11/2020 13:28:15

Diagnostics Check		OK	13:21
Air Blank		0.000	13:22
Control Test		0.079	13:22
Air Blank		0.000	13:23
Subject Sample	# 1	0.000	13:23
Breath Volume		1.683	13:23
Air Blank		0.000	13:23
Air Blank		0.000	13:25
Subject Sample	# 2	0.000	13:26
Breath Volume		1.292	13:26
Air Blank		0.000	13:27
Control Test		0.080	13:27
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: SHRADER

DATE/TIME OF TEST: 05/12/2020 01:04:08

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.103	00:59
Breath Volume		1.628	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.100	01:02
Breath Volume		2.007	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: MAILEY

DATE/TIME OF TEST: 05/12/2020 04:27:49

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.212	04:23
Breath Volume		1.695	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.228	04:26
Breath Volume		1.765	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/13/2020 21:28:09

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:25	
Control Test		0.079	21:26	
Air Blank		0.000	21:26	
Subject Sample	# 1		21:26	Subject Test Refused
Breath Volume		0.000	21:26	
Air Blank		0.000	21:27	
Control Test		0.079	21:27	
Air Blank		0.000	21:28	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 05/14/2020 11:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SULLIVAN,C
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MONSEN

DATE/TIME OF TEST: 05/14/2020 03:09:38

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.080	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.080	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 05/14/2020 11:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CHIN

DATE/TIME OF TEST: 05/15/2020 02:16:36

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GRASSO

DATE/TIME OF TEST: 05/16/2020 00:35:24

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.000	00:31
Breath Volume		2.187	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.000	00:34
Breath Volume		2.285	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: STRUM

DATE/TIME OF TEST: 05/16/2020 02:00:25

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.180	01:55
Breath Volume		1.527	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.183	01:58
Breath Volume		1.851	01:58
Air Blank		0.000	01:59
Control Test		0.077	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAUSLER,T
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: SANCOMB

DATE/TIME OF TEST: 05/16/2020 23:41:58

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.085	23:37
Breath Volume		3.292	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.084	23:40
Breath Volume		3.332	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WALTER

DATE/TIME OF TEST: 05/17/2020 03:43:53

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.120	03:39
Breath Volume		1.289	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:41
Subject Sample	# 2	0.129	03:42
Breath Volume		1.929	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/18/2020 01:17:28

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.000	01:13
Breath Volume		3.546	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.000	01:16
Breath Volume		3.000	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/19/2020 01:38:16

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.166	01:33
Breath Volume		2.933	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.166	01:36
Breath Volume		2.453	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARTER,W
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: PETRIC

DATE/TIME OF TEST: 05/19/2020 16:09:09

Diagnostics Check		OK	16:01
Air Blank		0.000	16:01
Control Test		0.079	16:02
Air Blank		0.000	16:02
Subject Sample	# 1	0.230	16:04
Breath Volume		2.375	16:04
Air Blank		0.000	16:05
Air Blank		0.000	16:06
Subject Sample	# 2	0.237	16:07
Breath Volume		3.042	16:07
Air Blank		0.000	16:08
Control Test		0.079	16:08
Air Blank		0.000	16:09
Diagnostics Check		OK	16:09

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRAY,JORGEDREKA,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: BARTA

DATE/TIME OF TEST: 05/25/2020 20:14:08

Diagnostics Check		OK	20:06
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.190	20:09
Breath Volume		1.679	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:11
Subject Sample	# 2	0.195	20:12
Breath Volume		1.324	20:12
Air Blank		0.000	20:13
Control Test		0.079	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: GARBER

DATE/TIME OF TEST: 05/26/2020 20:18:59

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.080	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:18	
Control Test		0.079	20:18	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:18	

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: DRINAZI

DATE/TIME OF TEST: 05/27/2020 01:35:43

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.092	01:31
Breath Volume		2.191	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.093	01:34
Breath Volume		1.695	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRAY,JORGEDREKA,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPURRIER,JULIE
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: MADRIGAL

DATE/TIME OF TEST: 05/27/2020 05:38:03

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.080	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.176	05:33
Breath Volume		1.664	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:35
Subject Sample	# 2	0.177	05:36
Breath Volume		1.593	05:36
Air Blank		0.000	05:37
Control Test		0.079	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:38

Data Download Date/Time: 05/27/2020 09:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 05/29/2020 03:55:04

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.162	03:50
Breath Volume		4.609	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.166	03:53
Breath Volume		4.382	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:55

Data Download Date/Time: 06/17/2020 23:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPHERD,P,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MEDICA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/01/2020 17:09:05

Diagnostics Check		OK	17:02
Air Blank		0.000	17:03
Control Test		0.077	17:03
Air Blank		0.000	17:04
Subject Sample	# 1	0.271	17:04
Breath Volume		2.171	17:04
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.278	17:07
Breath Volume		1.867	17:07
Air Blank		0.000	17:08
Control Test		0.078	17:08
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VEREEN,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: COATS

DATE/TIME OF TEST: 05/02/2020 17:55:02

Diagnostics Check		OK	17:48
Air Blank		0.000	17:49
Control Test		0.078	17:49
Air Blank		0.000	17:50
Subject Sample	# 1	0.132	17:50
Breath Volume		1.308	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:52
Subject Sample	# 2	0.117	17:53
Breath Volume		1.152	17:53
Air Blank		0.000	17:54
Control Test		0.078	17:54
Air Blank		0.000	17:54
Diagnostics Check		OK	17:55

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IRRA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:59

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 05/09/2020 13:26:36

Diagnostics Check		OK	13:20
Air Blank		0.000	13:20
Control Test		0.079	13:21
Air Blank		0.000	13:21
Subject Sample	# 1	0.000	13:22
Breath Volume		2.183	13:22
Air Blank		0.000	13:22
Air Blank		0.000	13:24
Subject Sample	# 2	0.000	13:25
Breath Volume		2.695	13:25
Air Blank		0.000	13:25
Control Test		0.079	13:26
Air Blank		0.000	13:26
Diagnostics Check		OK	13:26

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETTY,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/09/2020 16:03:48

Diagnostics Check		OK	16:00	
Air Blank		0.000	16:01	
Control Test		0.079	16:01	
Air Blank		0.000	16:02	
Subject Sample	# 1		16:02	Subject Test Refused
Breath Volume		0.000	16:02	
Air Blank		0.000	16:02	
Control Test		0.079	16:03	
Air Blank		0.000	16:03	
Diagnostics Check		OK	16:03	

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,L,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: RICKETT

DATE/TIME OF TEST: 05/09/2020 18:33:24

Diagnostics Check		OK	18:30	
Air Blank		0.000	18:30	
Control Test		0.079	18:31	
Air Blank		0.000	18:31	
Subject Sample	# 1		18:32	Subject Test Refused
Breath Volume		0.000	18:32	
Air Blank		0.000	18:32	
Control Test		0.079	18:32	
Air Blank		0.000	18:33	
Diagnostics Check		OK	18:33	

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/09/2020 22:09:47

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:07	
Control Test		0.078	22:07	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:08	Subject Test Refused
Breath Volume		0.000	22:08	
Air Blank		0.000	22:08	
Control Test		0.078	22:09	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/10/2020 04:59:36

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.078	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.158	04:55
Breath Volume		2.238	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.157	04:58
Breath Volume		2.328	04:58
Air Blank		0.000	04:58
Control Test		0.079	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BLACKWELDER

DATE/TIME OF TEST: 05/11/2020 04:13:29

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.108	04:09
Breath Volume		3.171	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.105	04:12
Breath Volume		2.734	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CHIRINO

DATE/TIME OF TEST: 05/11/2020 21:39:19

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.079	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.000	21:35
Breath Volume		2.476	21:35
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.000	21:37
Breath Volume		3.386	21:37
Air Blank		0.000	21:38
Control Test		0.078	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: HERNANDEZ AGUILAR

DATE/TIME OF TEST: 05/12/2020 04:49:36

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.211	04:45
Breath Volume		1.394	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.210	04:48
Breath Volume		2.375	04:48
Air Blank		0.000	04:48
Control Test		0.078	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAPLEY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: ENGLE

DATE/TIME OF TEST: 05/12/2020 16:21:58

Diagnostics Check		OK	16:12
Air Blank		0.000	16:13
Control Test		0.079	16:13
Air Blank		0.000	16:13
Subject Sample	# 1	0.261	16:14
Breath Volume		2.488	16:14
Air Blank		0.000	16:15
Air Blank		0.000	16:16
Subject Sample	# 2	0.239	16:17
Breath Volume		1.265	16:17
Air Blank		0.000	16:18
Air Blank		0.000	16:20
Subject Sample	# 3	0.253	16:20
Breath Volume		1.851	16:20
Air Blank		0.000	16:21
Control Test		0.077	16:21
Air Blank		0.000	16:21
Diagnostics Check		OK	16:21

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHEELER,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ALLARD

DATE/TIME OF TEST: 05/13/2020 18:32:03

Diagnostics Check		OK	18:25
Air Blank		0.000	18:26
Control Test		0.077	18:26
Air Blank		0.000	18:27
Subject Sample	# 1	0.000	18:27
Breath Volume		3.148	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.000	18:30
Breath Volume		3.648	18:30
Air Blank		0.000	18:31
Control Test		0.078	18:31
Air Blank		0.000	18:31
Diagnostics Check		OK	18:32

Data Download Date/Time: 05/14/2020 10:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 05/14/2020 01:07:15

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.078	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.058	01:02
Breath Volume		1.226	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.060	01:05
Breath Volume		1.609	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 05/14/2020 10:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/14/2020 02:17:19

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.078	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Control Test		0.077	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 05/14/2020 10:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/04/2020 01:03:42

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.132	00:59
Breath Volume		2.121	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.134	01:02
Breath Volume		2.171	01:02
Air Blank		0.000	01:02
Control Test		0.078	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/14/2020 10:59

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/05/2020 19:44:36

Diagnostics Check		OK	19:38
Air Blank		0.000	19:38
Control Test		0.078	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.136	19:40
Breath Volume		1.687	19:40
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.136	19:43
Breath Volume		1.535	19:43
Air Blank		0.000	19:43
Control Test		0.079	19:44
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 05/14/2020 10:59

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BINKLEY,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 05/06/2020 04:04:22

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.250	03:56
Breath Volume		1.261	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.229	03:59
Breath Volume		1.480	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 3	0.245	04:02
Breath Volume		1.882	04:02
Air Blank		0.000	04:03
Control Test		0.078	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 05/14/2020 10:59

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: CANZANO-SMITH

DATE/TIME OF TEST: 05/07/2020 00:05:59

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.139	00:01
Breath Volume		2.183	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.140	00:04
Breath Volume		2.003	00:04
Air Blank		0.000	00:05
Control Test		0.078	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 05/14/2020 10:59

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DANIELS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: DAMIREDDY

DATE/TIME OF TEST: 05/07/2020 19:55:34

Diagnostics Check		OK	19:48
Air Blank		0.000	19:48
Control Test		0.079	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.266	19:50
Breath Volume		1.445	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.269	19:54
Breath Volume		1.488	19:54
Air Blank		0.000	19:54
Control Test		0.078	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 05/14/2020 10:59

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: PAGAN-CUELLO

DATE/TIME OF TEST: 05/09/2020 05:33:10

Diagnostics Check		OK	05:30	
Air Blank		0.000	05:30	
Control Test		0.080	05:31	
Air Blank		0.000	05:31	
Subject Sample	# 1		05:31	Subject Test Refused
Breath Volume		0.000	05:31	
Air Blank		0.000	05:32	
Control Test		0.080	05:32	
Air Blank		0.000	05:33	
Diagnostics Check		OK	05:33	

Data Download Date/Time: 05/14/2020 10:59

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAGWELL,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 05/17/2020 16:06:19

Diagnostics Check		OK	15:59
Air Blank		0.000	16:00
Control Test		0.079	16:00
Air Blank		0.000	16:01
Subject Sample	# 1	0.238	16:01
Breath Volume		1.234	16:01
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 2	0.232	16:04
Breath Volume		1.125	16:04
Air Blank		0.000	16:05
Control Test		0.079	16:05
Air Blank		0.000	16:06
Diagnostics Check		OK	16:06

Data Download Date/Time: 05/27/2020 11:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANDRES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: LIN

DATE/TIME OF TEST: 05/18/2020 10:32:42

Diagnostics Check		OK	10:25
Air Blank		0.000	10:26
Control Test		0.079	10:26
Air Blank		0.000	10:27
Subject Sample	# 1	0.000	10:27
Breath Volume		1.699	10:27
Air Blank		0.000	10:28
Air Blank		0.000	10:30
Subject Sample	# 2	0.000	10:31
Breath Volume		1.843	10:31
Air Blank		0.000	10:31
Control Test		0.079	10:32
Air Blank		0.000	10:32
Diagnostics Check		OK	10:32

Data Download Date/Time: 05/27/2020 11:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: DELIA

DATE/TIME OF TEST: 05/19/2020 15:53:09

Diagnostics Check		OK	15:46
Air Blank		0.000	15:47
Control Test		0.079	15:47
Air Blank		0.000	15:48
Subject Sample	# 1	0.155	15:48
Breath Volume		3.152	15:48
Air Blank		0.000	15:49
Air Blank		0.000	15:51
Subject Sample	# 2	0.153	15:51
Breath Volume		2.910	15:51
Air Blank		0.000	15:52
Control Test		0.078	15:52
Air Blank		0.000	15:53
Diagnostics Check		OK	15:53

Data Download Date/Time: 05/27/2020 11:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 05/19/2020 23:22:49

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.007	23:18
Breath Volume		2.398	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.000	23:21
Breath Volume		2.160	23:21
Air Blank		0.000	23:21
Control Test		0.079	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 05/27/2020 11:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 05/21/2020 23:48:16

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.151	23:43
Breath Volume		2.750	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.145	23:46
Breath Volume		1.738	23:46
Air Blank		0.000	23:47
Control Test		0.078	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/22/2020 22:59:49

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.079	22:54
Air Blank		0.000	22:54
Subject Sample	# 1	0.036	22:55
Breath Volume		2.253	22:55
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.036	22:58
Breath Volume		2.687	22:58
Air Blank		0.000	22:58
Control Test		0.079	22:59
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/23/2020 03:21:36

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.175	03:17
Breath Volume		1.746	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.176	03:20
Breath Volume		2.058	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SUTHERLAND

DATE/TIME OF TEST: 05/23/2020 22:38:35

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.079	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.000	22:34
Breath Volume		2.402	22:34
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.000	22:37
Breath Volume		1.597	22:37
Air Blank		0.000	22:37
Control Test		0.078	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: OSEGUERA

DATE/TIME OF TEST: 05/24/2020 01:15:30

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.200	01:11
Breath Volume		2.445	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.212	01:13
Breath Volume		2.394	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 05/24/2020 01:58:27

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.076	01:53
Breath Volume		1.781	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.074	01:56
Breath Volume		1.812	01:56
Air Blank		0.000	01:57
Control Test		0.078	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BOSTICK,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SAMUEL

DATE/TIME OF TEST: 05/24/2020 04:01:56

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.126	03:57
Breath Volume		2.250	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.121	04:00
Breath Volume		1.441	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: M630520654450

DATE/TIME OF TEST: 05/24/2020 22:49:02

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:44	
Control Test		0.079	22:45	
Air Blank		0.000	22:45	
Subject Sample	# 1		22:47	Subject Test Refused
Breath Volume		0.000	22:47	
Air Blank		0.000	22:48	
Control Test		0.079	22:48	
Air Blank		0.000	22:48	
Diagnostics Check		OK	22:49	

Data Download Date/Time: 05/27/2020 11:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 05/24/2020 23:37:34

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:36	
Control Test		0.079	23:36	
Air Blank		0.000	23:37	
Diagnostics Check		OK	23:37	

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAUSLER,T
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: BUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FUNEZ VILLANUEVA

DATE/TIME OF TEST: 05/25/2020 00:18:22

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:16	
Control Test		0.079	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:17	Subject Test Refused
Breath Volume		0.000	00:17	
Air Blank		0.000	00:17	
Control Test		0.079	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: VICKORIE

DATE/TIME OF TEST: 05/26/2020 01:23:56

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:23	
Control Test		0.080	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 05/27/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 05/28/2020 06:44:55

Diagnostics Check		OK	06:42	
Air Blank		0.000	06:42	
Control Test		0.079	06:42	
Air Blank		0.000	06:43	
Subject Sample	# 1		06:43	Subject Test Refused
Breath Volume		0.000	06:43	
Air Blank		0.000	06:44	
Control Test		0.079	06:44	
Air Blank		0.000	06:44	
Diagnostics Check		OK	06:44	

Data Download Date/Time: 06/17/2020 23:02

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: NASS

DATE/TIME OF TEST: 05/30/2020 02:44:26

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.118	02:39
Breath Volume		1.800	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.121	02:42
Breath Volume		2.457	02:42
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 06/17/2020 23:03

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: BIVENS

DATE/TIME OF TEST: 05/30/2020 06:04:58

Diagnostics Check		OK	05:58
Air Blank		0.000	05:58
Control Test		0.079	05:59
Air Blank		0.000	05:59
Subject Sample	# 1	0.155	06:00
Breath Volume		2.054	06:00
Air Blank		0.000	06:01
Air Blank		0.000	06:02
Subject Sample	# 2	0.157	06:03
Breath Volume		1.808	06:03
Air Blank		0.000	06:04
Control Test		0.078	06:04
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 06/17/2020 23:03

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: MATEO MORALES

DATE/TIME OF TEST: 05/31/2020 13:49:26

Diagnostics Check		OK	13:42
Air Blank		0.000	13:42
Control Test		0.079	13:42
Air Blank		0.000	13:43
Subject Sample	# 1	0.194	13:44
Breath Volume		3.203	13:44
Air Blank		0.000	13:45
Air Blank		0.000	13:47
Subject Sample	# 2	0.184	13:47
Breath Volume		2.007	13:47
Air Blank		0.000	13:48
Control Test		0.078	13:48
Air Blank		0.000	13:49
Diagnostics Check		OK	13:49

Data Download Date/Time: 06/17/2020 23:03

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BIGGS

DATE/TIME OF TEST: 05/09/2020 20:45:25

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:40	
Control Test		0.079	20:41	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:44	Subject Test Refused
Breath Volume		0.000	20:44	
Air Blank		0.000	20:44	
Control Test		0.079	20:44	
Air Blank		0.000	20:45	
Diagnostics Check		OK	20:45	

Data Download Date/Time: 07/03/2020 18:07

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRYDEN,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:07

SUBJECT'S LAST NAME: BLESSINGER

DATE/TIME OF TEST: 05/16/2020 09:34:35

Diagnostics Check		OK	09:28
Air Blank		0.000	09:28
Control Test		0.079	09:29
Air Blank		0.000	09:29
Subject Sample	# 1	0.055	09:30
Breath Volume		1.476	09:30
Air Blank		0.000	09:30
Air Blank		0.000	09:32
Subject Sample	# 2	0.054	09:33
Breath Volume		1.480	09:33
Air Blank		0.000	09:33
Control Test		0.078	09:33
Air Blank		0.000	09:34
Diagnostics Check		OK	09:34

Data Download Date/Time: 07/03/2020 18:07

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,SCOTT,R
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: SHEFFIELD,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: BUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HESS

DATE/TIME OF TEST: 05/24/2020 21:29:50

Diagnostics Check		OK	21:27	
Air Blank		0.000	21:27	
Control Test		0.079	21:27	
Air Blank		0.000	21:28	
Subject Sample	# 1		21:28	Subject Test Refused
Breath Volume		0.000	21:28	
Air Blank		0.000	21:28	
Control Test		0.079	21:29	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 07/03/2020 18:07

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,LUKE,O
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: DAVENPORT,LUKE
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 05/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: BUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/24/2020 22:39:05

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:36	
Control Test		0.080	22:37	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:37	Subject Test Refused
Breath Volume		0.000	22:37	
Air Blank		0.000	22:38	
Control Test		0.080	22:38	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 07/03/2020 18:07

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/11/2020 20:44:26

Diagnostics Check		OK	20:37
Air Blank		0.000	20:38
Control Test		0.080	20:38
Air Blank		0.000	20:39
Subject Sample	# 1	0.111	20:39
Breath Volume		1.812	20:39
Air Blank		0.000	20:40
Air Blank		0.000	20:42
Subject Sample	# 2	0.107	20:42
Breath Volume		1.667	20:42
Air Blank		0.000	20:43
Control Test		0.079	20:43
Air Blank		0.000	20:44
Diagnostics Check		OK	20:44

Data Download Date/Time: 05/31/2020 20:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON',A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: HARRIS JR

DATE/TIME OF TEST: 05/17/2020 21:24:40

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.080	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.233	21:20
Breath Volume		1.894	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.218	21:23
Breath Volume		1.976	21:23
Air Blank		0.000	21:23
Control Test		0.079	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 05/31/2020 20:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: RHODEN

DATE/TIME OF TEST: 05/30/2020 12:13:47

Diagnostics Check		OK	12:07
Air Blank		0.000	12:07
Control Test		0.081	12:07
Air Blank		0.000	12:08
Subject Sample	# 1	0.132	12:08
Breath Volume		1.812	12:08
Air Blank		0.000	12:09
Air Blank		0.000	12:11
Subject Sample	# 2	0.122	12:12
Breath Volume		1.753	12:12
Air Blank		0.000	12:12
Control Test		0.079	12:13
Air Blank		0.000	12:13
Diagnostics Check		OK	12:13

Data Download Date/Time: 05/31/2020 20:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/31/2020 18:13:17

Diagnostics Check		OK	18:07
Air Blank		0.000	18:07
Control Test		0.081	18:07
Air Blank		0.000	18:08
Subject Sample	# 1	0.000	18:08
Breath Volume		3.312	18:08
Air Blank		0.000	18:09
Air Blank		0.000	18:11
Subject Sample	# 2	0.000	18:11
Breath Volume		2.843	18:11
Air Blank		0.000	18:12
Control Test		0.079	18:12
Air Blank		0.000	18:13
Diagnostics Check		OK	18:13

Data Download Date/Time: 05/31/2020 20:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 05/03/2020 01:13:47

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.086	01:09
Breath Volume		1.839	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.080	01:12
Breath Volume		1.425	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 06/04/2020 11:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: NELOMS

DATE/TIME OF TEST: 05/04/2020 03:53:32

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 06/04/2020 11:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: BANACH,DONNA
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: ROSSIE

DATE/TIME OF TEST: 05/14/2020 19:22:58

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:17	
Control Test		0.075	19:18	
Air Blank		0.000	19:18	
Subject Sample	# 1		19:21	RFI Detect
Breath Volume		0.000	19:21	
Air Blank		0.000	19:22	
Control Test		0.073	19:22	Control Outside Tolerance
Air Blank		0.000	19:22	

Data Download Date/Time: 06/04/2020 11:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,DALTON,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: BANACH,DONNA
ARRESTING OFFICER AGENCY: Starke PD
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: ROSSIE

DATE/TIME OF TEST: 05/14/2020 19:38:42

Diagnostics Check		OK	19:30	
Air Blank		0.000	19:30	
Control Test		0.077	19:31	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:34	No Sample Provided
Breath Volume		0.000	19:34	
Air Blank		0.000	19:35	
Air Blank		0.000	19:37	
Subject Sample	# 2		19:37	Subject Test Refused
Breath Volume		0.000	19:37	
Air Blank		0.000	19:37	
Control Test		0.078	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:38	

Data Download Date/Time: 06/04/2020 11:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VISMAN,CLAYTON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VISMAN,CLAYTON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: THORSON

DATE/TIME OF TEST: 05/25/2020 21:59:40

Diagnostics Check		OK	21:52
Air Blank		0.000	21:53
Control Test		0.079	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.296	21:54
Breath Volume		1.890	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.284	21:58
Breath Volume		1.347	21:58
Air Blank		0.000	21:58
Control Test		0.079	21:59
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 06/04/2020 11:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: FLYNN,H
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 05/27/2020 15:18:46

Diagnostics Check		OK	15:04	
Air Blank		0.000	15:05	
Control Test		0.079	15:05	
Air Blank		0.000	15:05	
Subject Sample	# 1	0.026	15:06	
Breath Volume		1.832	15:06	
Air Blank		0.000	15:06	
Air Blank		0.000	15:08	
Subject Sample	# 2	0.023	15:11	Volume Not Met
Breath Volume		0.050	15:11	
Air Blank		0.000	15:12	
Air Blank		0.000	15:14	
Subject Sample	# 3	0.023	15:17	
Breath Volume		1.304	15:17	
Air Blank		0.000	15:17	
Control Test		0.078	15:18	
Air Blank		0.000	15:18	
Diagnostics Check		OK	15:18	

Data Download Date/Time: 06/04/2020 11:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HARDYMON

DATE/TIME OF TEST: 05/30/2020 21:39:25

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:37	
Control Test		0.078	21:37	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:38	Subject Test Refused
Breath Volume		0.000	21:38	
Air Blank		0.000	21:38	
Control Test		0.078	21:38	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 06/05/2020 08:38

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 05/29/2020 23:56:15

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:51
Breath Volume		2.667	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:54
Breath Volume		2.230	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 06/29/2020 18:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLEMENTS,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PELICO

DATE/TIME OF TEST: 05/30/2020 01:54:49

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.078	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:50	No Sample Provided
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	
Air Blank		0.000	01:53	
Subject Sample	# 2		01:53	Subject Test Refused
Breath Volume		0.000	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 06/29/2020 18:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLSOM,DAVID,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FOLSOM,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/30/2020 03:26:44

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.163	03:22
Breath Volume		1.882	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:24
Subject Sample	# 2	0.155	03:25
Breath Volume		1.441	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/29/2020 18:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERY,BAYLEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: CHANNELL

DATE/TIME OF TEST: 05/30/2020 22:31:55

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.079	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.262	22:27
Breath Volume		2.132	22:27
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.275	22:30
Breath Volume		2.625	22:30
Air Blank		0.000	22:31
Control Test		0.079	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 06/29/2020 18:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TOLBERT,JOEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 05/01/2020 02:14:57

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:04	
Control Test		0.081	02:04	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:07	No Sample Provided
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Air Blank		0.000	02:10	
Subject Sample	# 2		02:13	No Sample Provided
Breath Volume		0.000	02:13	
Air Blank		0.000	02:14	
Control Test		0.081	02:14	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,JAMISON,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 05/01/2020 05:39:57

Diagnostics Check		OK	05:33
Air Blank		0.000	05:34
Control Test		0.081	05:34
Air Blank		0.000	05:34
Subject Sample	# 1	0.238	05:35
Breath Volume		2.656	05:35
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.233	05:38
Breath Volume		2.945	05:38
Air Blank		0.000	05:39
Control Test		0.079	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TODD,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/02/2020 18:40:50

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:29	
Control Test		0.080	18:29	
Air Blank		0.000	18:29	
Subject Sample	# 1	0.114	18:30	
Breath Volume		1.875	18:30	
Air Blank		0.000	18:31	
Air Blank		0.000	18:33	
Subject Sample	# 2	0.103	18:36	Volume Not Met
Breath Volume		1.089	18:36	
Air Blank		0.000	18:37	
Air Blank		0.000	18:38	
Subject Sample	# 3	0.110	18:39	
Breath Volume		1.550	18:39	
Air Blank		0.000	18:39	
Control Test		0.079	18:40	
Air Blank		0.000	18:40	
Diagnostics Check		OK	18:40	

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 05/03/2020 21:05:17

Diagnostics Check		OK	20:59	
Air Blank		0.000	20:59	
Control Test		0.080	21:00	
Air Blank		0.000	21:00	
Subject Sample	# 1	0.099	21:01	
Breath Volume		2.410	21:01	
Air Blank		0.000	21:02	
Air Blank		0.000	21:03	
Subject Sample	# 2	0.108	21:04	
Breath Volume		5.343	21:04	
Air Blank			21:04	RFI Detect
Air Blank		0.000	21:05	

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN],MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 05/07/2020 18:27:11

Diagnostics Check		OK	18:20
Air Blank		0.000	18:21
Control Test		0.081	18:21
Air Blank		0.000	18:22
Subject Sample	# 1	0.089	18:22
Breath Volume		2.960	18:22
Air Blank		0.000	18:23
Air Blank		0.000	18:25
Subject Sample	# 2	0.085	18:25
Breath Volume		3.019	18:25
Air Blank		0.000	18:26
Control Test		0.079	18:26
Air Blank		0.000	18:27
Diagnostics Check		OK	18:27

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIRCH,JESSIE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 09:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:57

SUBJECT'S LAST NAME: HYMAN

DATE/TIME OF TEST: 05/08/2020 10:24:17

Diagnostics Check		OK	10:17
Air Blank		0.000	10:18
Control Test		0.081	10:18
Air Blank		0.000	10:19
Subject Sample	# 1	0.000	10:19
Breath Volume		1.304	10:19
Air Blank		0.000	10:20
Air Blank		0.000	10:22
Subject Sample	# 2	0.000	10:22
Breath Volume		1.535	10:22
Air Blank		0.000	10:23
Control Test		0.081	10:23
Air Blank		0.000	10:24
Diagnostics Check		OK	10:24

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 05/09/2020 19:26:33

Diagnostics Check		OK	19:20
Air Blank		0.000	19:20
Control Test		0.080	19:21
Air Blank		0.000	19:21
Subject Sample	# 1	0.000	19:22
Breath Volume		2.175	19:22
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 2	0.000	19:25
Breath Volume		2.058	19:25
Air Blank		0.000	19:25
Control Test		0.081	19:25
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BODI

DATE/TIME OF TEST: 05/10/2020 01:30:46

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.081	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Control Test		0.081	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARRIVEE-GRANT,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LARRIVEE-GRANT,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 05/12/2020 18:07:40

Diagnostics Check		OK	18:04	
Air Blank		0.000	18:04	
Control Test		0.080	18:05	
Air Blank		0.000	18:05	
Subject Sample	# 1		18:06	Subject Test Refused
Breath Volume		0.000	18:06	
Air Blank		0.000	18:06	
Control Test		0.080	18:07	
Air Blank		0.000	18:07	
Diagnostics Check		OK	18:07	

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: REAVES,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: KAHLER

DATE/TIME OF TEST: 05/14/2020 17:28:05

Diagnostics Check		OK	17:21
Air Blank		0.000	17:22
Control Test		0.081	17:22
Air Blank		0.000	17:23
Subject Sample	# 1	0.000	17:23
Breath Volume		3.265	17:23
Air Blank		0.000	17:24
Air Blank		0.000	17:26
Subject Sample	# 2	0.000	17:26
Breath Volume		3.445	17:26
Air Blank		0.000	17:27
Control Test		0.080	17:27
Air Blank		0.000	17:28
Diagnostics Check		OK	17:28

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,JOEL,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,JOEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 05/15/2020 06:13:51

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:07	
Control Test		0.081	06:07	
Air Blank		0.000	06:07	
Subject Sample	# 1	0.017	06:08	
Breath Volume		1.261	06:08	
Air Blank		0.000	06:09	
Air Blank		0.000	06:11	
Subject Sample	# 2		06:13	Sequence Aborted
Breath Volume		0.035	06:13	
Air Blank		0.000	06:13	

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,JOEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 05/15/2020 06:24:28

Diagnostics Check		OK	06:16
Air Blank		0.000	06:17
Control Test		0.079	06:17
Air Blank		0.000	06:17
Subject Sample	# 1	0.013	06:20
Breath Volume		1.523	06:20
Air Blank		0.000	06:20
Air Blank		0.000	06:22
Subject Sample	# 2	0.011	06:23
Breath Volume		1.695	06:23
Air Blank		0.000	06:23
Control Test		0.079	06:23
Air Blank		0.000	06:24
Diagnostics Check		OK	06:24

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,KEITH,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 08:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 05/15/2020 10:07:04

Diagnostics Check		OK	10:00
Air Blank		0.000	10:01
Control Test		0.080	10:01
Air Blank		0.000	10:02
Subject Sample	# 1	0.000	10:02
Breath Volume		2.800	10:02
Air Blank		0.000	10:03
Air Blank		0.000	10:05
Subject Sample	# 2	0.000	10:05
Breath Volume		1.781	10:05
Air Blank		0.000	10:06
Control Test		0.079	10:06
Air Blank		0.000	10:06
Diagnostics Check		OK	10:07

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MERRILL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: COONEY

DATE/TIME OF TEST: 05/15/2020 23:25:26

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:12	
Control Test		0.079	23:13	
Air Blank		0.000	23:13	
Subject Sample	# 1	0.225	23:16	Slope Not Level
Breath Volume		1.691	23:16	
Air Blank		0.000	23:17	
Air Blank		0.000	23:19	
Subject Sample	# 2	0.247	23:20	
Breath Volume		2.101	23:20	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 3	0.227	23:23	
Breath Volume		1.757	23:23	
Air Blank		0.000	23:24	
Control Test		0.078	23:24	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOGGETTE,MASON,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 05/15/2020 23:59:39

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.287	23:51
Breath Volume		1.214	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.258	23:54
Breath Volume		1.496	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 3	0.290	23:58
Breath Volume		2.273	23:58
Air Blank		0.000	23:58
Control Test		0.079	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MCALL-MOORE JR

DATE/TIME OF TEST: 05/16/2020 03:52:49

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.085	03:48
Breath Volume		4.691	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.082	03:51
Breath Volume		4.832	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 05/29/2020 19:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,J,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ST PIERRE

DATE/TIME OF TEST: 05/16/2020 21:54:55

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:41	
Control Test		0.080	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1	0.166	21:45	Volume Not Met
Breath Volume		0.113	21:45	
Air Blank		0.000	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 2	0.166	21:49	
Breath Volume		2.078	21:49	
Air Blank		0.000	21:50	
Air Blank		0.000	21:51	
Subject Sample	# 3	0.161	21:53	
Breath Volume		1.839	21:53	
Air Blank		0.000	21:54	
Control Test		0.078	21:54	
Air Blank		0.000	21:54	
Diagnostics Check		OK	21:54	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNSFORD,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DUNSFORD,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/16/2020 23:09:37

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.000	23:04
Breath Volume		1.328	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.000	23:08
Breath Volume		2.062	23:08
Air Blank		0.000	23:08
Control Test		0.079	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ASKEGARD

DATE/TIME OF TEST: 05/17/2020 02:17:40

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.184	02:11
Breath Volume		2.687	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.184	02:16
Breath Volume		3.070	02:16
Air Blank		0.000	02:16
Control Test		0.079	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POND,C
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/19/2020 21:41:54

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.080	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.315	21:37
Breath Volume		2.589	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.304	21:40
Breath Volume		2.703	21:40
Air Blank		0.000	21:41
Control Test		0.078	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JOSHUA,I
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 05/20/2020 05:47:34

Diagnostics Check	OK	05:47	
Air Blank		05:47	RFI Detect
Air Blank		05:47	RFI Detect

Data Download Date/Time: 05/29/2020 19:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JOSHUA,I
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 05/20/2020 05:52:04

Diagnostics Check	OK	05:51	
Air Blank		05:51	RFI Detect
Air Blank		05:52	RFI Detect

Data Download Date/Time: 05/29/2020 19:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JOSHUA,I
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SCHULTZ,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 05/20/2020 06:02:29

Diagnostics Check		OK	05:57	
Air Blank		0.000	05:57	
Control Test		0.081	05:57	
Air Blank		0.000	05:58	
Subject Sample	# 1	0.000	05:58	
Breath Volume		3.613	05:58	
Air Blank		0.000	05:59	
Air Blank			06:01	Improper Sample
Air Blank		0.000	06:01	
Control Test		0.080	06:01	
Air Blank		0.000	06:02	
Diagnostics Check		OK	06:02	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSMAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FORSMAN,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/22/2020 02:45:40

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.081	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.211	02:41
Breath Volume		2.414	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.200	02:44
Breath Volume		3.886	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CASIAS

DATE/TIME OF TEST: 05/23/2020 03:15:37

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.165	03:10
Breath Volume		3.074	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.168	03:14
Breath Volume		3.386	03:14
Air Blank		0.000	03:14
Control Test		0.081	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: MILSTEAD

DATE/TIME OF TEST: 05/23/2020 18:53:34

Diagnostics Check		OK	18:48	
Air Blank		0.000	18:48	
Control Test		0.081	18:48	
Air Blank		0.000	18:49	
Subject Sample	# 1	0.134	18:49	
Breath Volume		3.621	18:49	
Air Blank		0.000	18:50	
Air Blank		0.000	18:52	
Subject Sample	# 2	0.137	18:52	
Breath Volume		4.234	18:52	
Air Blank			18:53	RFI Detect
Air Blank		0.000	18:53	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RICE,HANAAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: T656453931470

DATE/TIME OF TEST: 05/23/2020 19:37:33

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:35	
Control Test		0.079	19:35	
Air Blank		0.000	19:36	
Subject Sample	# 1		19:36	Subject Test Refused
Breath Volume		0.000	19:36	
Air Blank		0.000	19:36	
Control Test		0.079	19:36	
Air Blank		0.000	19:37	
Diagnostics Check		OK	19:37	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLSOM,DAVID,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FOLSOM,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HAINES

DATE/TIME OF TEST: 05/24/2020 00:31:18

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.183	00:26
Breath Volume		2.886	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.168	00:29
Breath Volume		2.484	00:29
Air Blank		0.000	00:30
Control Test		0.077	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BROXSON

DATE/TIME OF TEST: 05/24/2020 01:31:03

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.115	01:26
Breath Volume		1.777	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:28
Subject Sample	# 2	0.118	01:29
Breath Volume		2.191	01:29
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:31

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: KNAUBER

DATE/TIME OF TEST: 05/25/2020 17:22:47

Diagnostics Check		OK	17:16
Air Blank		0.000	17:16
Control Test		0.080	17:17
Air Blank		0.000	17:17
Subject Sample	# 1	0.033	17:18
Breath Volume		2.730	17:18
Air Blank		0.000	17:18
Air Blank		0.000	17:20
Subject Sample	# 2	0.035	17:21
Breath Volume		2.222	17:21
Air Blank		0.000	17:21
Control Test		0.080	17:22
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAND,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAND,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: HOWINGTON

DATE/TIME OF TEST: 05/25/2020 19:52:50

Diagnostics Check		OK	19:48	
Air Blank		0.000	19:48	
Control Test		0.079	19:49	
Air Blank		0.000	19:49	
Subject Sample	# 1		19:52	Sequence Aborted
Breath Volume		0.000	19:52	
Air Blank		0.000	19:52	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAND,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAND,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: HOWINGTON

DATE/TIME OF TEST: 05/25/2020 21:00:40

Diagnostics Check		OK	20:57	
Air Blank		0.000	20:58	
Control Test		0.081	20:58	
Air Blank		0.000	20:59	
Subject Sample	# 1		20:59	Subject Test Refused
Breath Volume		0.000	20:59	
Air Blank		0.000	20:59	
Control Test		0.080	21:00	
Air Blank		0.000	21:00	
Diagnostics Check		OK	21:00	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/26/2020 01:12:41

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:09	
Control Test		0.079	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 1		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:11	
Control Test		0.080	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEVILLA,NOEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: BERGMANN

DATE/TIME OF TEST: 05/26/2020 19:10:25

Diagnostics Check		OK	19:01	
Air Blank		0.000	19:02	
Control Test		0.080	19:02	
Air Blank		0.000	19:03	
Subject Sample	# 1		19:06	No Sample Provided
Breath Volume		0.000	19:06	
Air Blank		0.000	19:06	
Air Blank		0.000	19:08	
Subject Sample	# 2		19:09	Subject Test Refused
Breath Volume		0.000	19:09	
Air Blank		0.000	19:09	
Control Test		0.080	19:09	
Air Blank		0.000	19:10	
Diagnostics Check		OK	19:10	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: FERRY

DATE/TIME OF TEST: 05/27/2020 00:26:53

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.080	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.164	00:22
Breath Volume		1.378	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.172	00:25
Breath Volume		2.519	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MERRILL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ANGELES

DATE/TIME OF TEST: 05/28/2020 00:01:01

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.080	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.080	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 05/29/2020 19:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,KEENAN,B
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: MOORE,BRANDON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 04/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 05/03/2020 01:28:19

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.194	01:23
Breath Volume		2.433	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.184	01:26
Breath Volume		1.945	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/05/2020 08:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JORDAN,WILFORD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/03/2020 17:27:10

Diagnostics Check		OK	17:20
Air Blank		0.000	17:21
Control Test		0.080	17:21
Air Blank		0.000	17:22
Subject Sample	# 1	0.227	17:22
Breath Volume		3.914	17:22
Air Blank		0.000	17:23
Air Blank		0.000	17:25
Subject Sample	# 2	0.218	17:25
Breath Volume		2.832	17:25
Air Blank		0.000	17:26
Control Test		0.078	17:26
Air Blank		0.000	17:27
Diagnostics Check		OK	17:27

Data Download Date/Time: 05/05/2020 08:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRISP,JESSICA,M
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: CRISP,JESSICA
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 04/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BAXLEY

DATE/TIME OF TEST: 05/04/2020 00:33:16

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.082	00:29	
Air Blank		0.000	00:29	
Subject Sample	# 1	0.178	00:30	
Breath Volume		3.484	00:30	
Air Blank		0.000	00:30	
Air Blank			00:32	RFI Detect
Air Blank		0.000	00:32	
			00:32	RFI Detect
Air Blank		0.000	00:33	

Data Download Date/Time: 05/05/2020 08:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRISP,JESSICA,M
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: CRISP,JESSICA
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 04/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BAXLEY

DATE/TIME OF TEST: 05/04/2020 01:29:31

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.161	01:25
Breath Volume		3.195	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.157	01:27
Breath Volume		3.191	01:27
Air Blank		0.000	01:28
Control Test		0.079	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 05/05/2020 08:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 9070670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/22/2020 14:18:37

Diagnostics Check	OK	14:17	
Air Blank		14:18	Ambient Fail
Air Blank	0.000	14:18	

Data Download Date/Time: 06/04/2020 12:28

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 9070670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: SYS CHECK
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/22/2020 14:24:00

Diagnostics Check		OK	14:20	
Air Blank		0.000	14:21	
Control Test		0.080	14:21	
Air Blank		0.000	14:22	
Subject Sample	# 1		14:22	Subject Test Refused
Breath Volume		0.000	14:22	
Air Blank		0.000	14:23	
Control Test		0.077	14:23	
Air Blank		0.000	14:23	
Diagnostics Check		OK	14:24	

Data Download Date/Time: 06/04/2020 12:28

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 9070670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: SYS CHECK
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/22/2020 19:24:16

Diagnostics Check	OK	19:23	
Air Blank		19:23	Ambient Fail
Air Blank	0.000	19:24	

Data Download Date/Time: 06/04/2020 12:28

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 9070670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: SYS CHECK
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/22/2020 19:28:14

Diagnostics Check	OK	19:27	
Air Blank		19:27	Ambient Fail
Air Blank		19:28	Ambient Fail

Data Download Date/Time: 06/04/2020 12:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 9070670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: SYS CHECK
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/22/2020 19:31:54

Diagnostics Check	OK	19:30	
Air Blank		19:31	Ambient Fail
Air Blank	0.000	19:31	

Data Download Date/Time: 06/04/2020 12:28

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFF,STEPHEN,
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NEFF,STEPHEN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: SYS CHECK
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2020 09:51:47

Diagnostics Check		OK	09:46	
Air Blank		0.000	09:46	
Control Test		0.078	09:46	
Air Blank		0.000	09:47	
Subject Sample	# 1	0.000	09:47	
Breath Volume		2.527	09:47	
Air Blank		0.000	09:48	
Air Blank		0.000	09:50	
Subject Sample	# 2		09:50	Subject Test Refused
Breath Volume		0.000	09:50	
Air Blank		0.000	09:50	
Control Test		0.078	09:51	
Air Blank		0.000	09:51	
Diagnostics Check		OK	09:51	

Data Download Date/Time: 06/04/2020 12:28

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: SYS CHECK
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/24/2020 18:08:01

Diagnostics Check		OK	18:02	
Air Blank		0.000	18:03	
Control Test		0.080	18:03	
Air Blank		0.000	18:04	
Subject Sample	# 1	0.061	18:04	
Breath Volume		3.320	18:04	
Air Blank		0.000	18:05	
Air Blank		0.000	18:06	
Subject Sample	# 2	0.057	18:07	
Breath Volume		2.593	18:07	
Air Blank			18:07	RFI Detect
Air Blank		0.000	18:08	

Data Download Date/Time: 06/04/2020 12:28

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: BUBAN

DATE/TIME OF TEST: 05/24/2020 19:47:46

Diagnostics Check		OK	19:41
Air Blank		0.000	19:41
Control Test		0.079	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.155	19:43
Breath Volume		4.140	19:43
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.150	19:46
Breath Volume		4.230	19:46
Air Blank		0.000	19:46
Control Test		0.079	19:47
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 06/04/2020 12:28

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALSOBROOKS,KEELIN,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALSOBROOKS,KEELIN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG816701
DRY GAS EXPIRY DATE: 06/16/2020
VIOLATION CODE: BUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 05/23/2020 17:24:59

Diagnostics Check		OK	17:18
Air Blank		0.000	17:18
Control Test		0.081	17:19
Air Blank		0.000	17:19
Subject Sample	# 1	0.143	17:20
Breath Volume		3.539	17:20
Air Blank		0.000	17:20
Air Blank		0.000	17:22
Subject Sample	# 2	0.142	17:23
Breath Volume		3.488	17:23
Air Blank		0.000	17:24
Control Test		0.080	17:24
Air Blank		0.000	17:24
Diagnostics Check		OK	17:24

Data Download Date/Time: 05/29/2020 12:51

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,JOHN,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: FOSTER,JOHN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 05/30/2020 03:08:59

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.193	03:04
Breath Volume		4.285	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.199	03:07
Breath Volume		3.726	03:07
Air Blank		0.000	03:08
Control Test		0.083	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 06/26/2020 16:20

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRITT,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SCHELL,GARRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: VISATHEP

DATE/TIME OF TEST: 05/04/2020 07:00:39

Diagnostics Check	OK	07:00	
Air Blank		07:00	RFI Detect
Air Blank	0.000	07:00	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRITT,DAKOTA,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SCHELL,GARY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: VISATHEP

DATE/TIME OF TEST: 05/04/2020 07:09:41

Diagnostics Check		OK	07:02
Air Blank		0.000	07:03
Control Test		0.080	07:03
Air Blank		0.000	07:04
Subject Sample	# 1	0.000	07:04
Breath Volume		1.605	07:04
Air Blank		0.000	07:05
Air Blank		0.000	07:07
Subject Sample	# 2	0.000	07:08
Breath Volume		1.738	07:08
Air Blank		0.000	07:08
Control Test		0.080	07:09
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: DEXTER

DATE/TIME OF TEST: 05/05/2020 19:58:32

Diagnostics Check		OK	19:52
Air Blank		0.000	19:52
Control Test		0.080	19:53
Air Blank		0.000	19:53
Subject Sample	# 1	0.134	19:53
Breath Volume		4.257	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.136	19:57
Breath Volume		4.621	19:57
Air Blank		0.000	19:57
Control Test		0.080	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CASTILLO-LABORDE

DATE/TIME OF TEST: 05/07/2020 02:37:36

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.081	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.000	02:33
Breath Volume		2.859	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.000	02:36
Breath Volume		2.019	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,RYAN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NELSON,RYAN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 05/08/2020 02:54:17

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.078	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.079	02:53	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHROW,TRACY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: STOPKA,COREY
ARRESTING OFFICER AGENCY: Parker PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 05/09/2020 01:10:47

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.177	01:06
Breath Volume		1.523	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.180	01:09
Breath Volume		1.632	01:09
Air Blank		0.000	01:09
Control Test		0.077	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,ALEX,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: GOBER,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RING

DATE/TIME OF TEST: 05/11/2020 01:23:22

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.080	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/12/2020 02:14:07

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:10	No Sample Provided
Breath Volume		0.000	02:10	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Control Test		0.079	02:13	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/15/2020 19:39:03

Diagnostics Check		OK	19:36	
Air Blank		0.000	19:36	
Control Test		0.080	19:36	
Air Blank		0.000	19:37	
Subject Sample	# 1		19:37	Subject Test Refused
Breath Volume		0.000	19:37	
Air Blank		0.000	19:38	
Control Test		0.080	19:38	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:39	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: POZADAS-PALMA

DATE/TIME OF TEST: 05/17/2020 04:23:50

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.080	04:13	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.068	04:16	Volume Not Met
Breath Volume		0.609	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 2	0.064	04:22	Volume Not Met
Breath Volume		0.496	04:22	
Air Blank		0.000	04:22	
Control Test		0.080	04:23	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: POZADAS-PALMA

DATE/TIME OF TEST: 05/17/2020 04:35:39

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.084	04:29
Breath Volume		1.414	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.081	04:34
Breath Volume		1.250	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 05/17/2020 14:15:31

Diagnostics Check		OK	14:12	
Air Blank		0.000	14:13	
Control Test		0.080	14:13	
Air Blank		0.000	14:13	
Subject Sample	# 1		14:14	Subject Test Refused
Breath Volume		0.000	14:14	
Air Blank		0.000	14:14	
Control Test		0.080	14:14	
Air Blank		0.000	14:15	
Diagnostics Check		OK	14:15	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHROW,TRACY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MATHIS,MARKEVIS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 05/19/2020 20:19:00

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.081	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:18	
Control Test		0.080	20:18	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:19	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHROW,TRACY,L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SULLIVAN,DAVID
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 05/22/2020 21:33:22

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:30	
Control Test		0.081	21:31	
Air Blank		0.000	21:31	
Subject Sample	# 1		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:32	
Control Test		0.080	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,JOHN,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: LAROCHE,THOMAS
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: 0080111720

DATE/TIME OF TEST: 05/23/2020 00:11:36

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.125	00:07
Breath Volume		1.683	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.124	00:10
Breath Volume		1.468	00:10
Air Blank		0.000	00:10
Control Test		0.081	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/23/2020 06:57:15

Diagnostics Check		OK	06:51
Air Blank		0.000	06:51
Control Test		0.080	06:51
Air Blank		0.000	06:52
Subject Sample	# 1	0.000	06:52
Breath Volume		2.312	06:52
Air Blank		0.000	06:53
Air Blank		0.000	06:55
Subject Sample	# 2	0.000	06:55
Breath Volume		2.453	06:55
Air Blank		0.000	06:56
Control Test		0.081	06:56
Air Blank		0.000	06:57
Diagnostics Check		OK	06:57

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: H655673709670

DATE/TIME OF TEST: 05/28/2020 01:46:16

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.214	01:41
Breath Volume		3.070	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.214	01:44
Breath Volume		2.941	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 05/29/2020 13:14

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EAKINS,SIERRA,M
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 05/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 869968
DRY GAS EXPIRY DATE: 12/06/2020
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 05/28/2020 15:00:24

Diagnostics Check		OK	14:57	
Air Blank		0.000	14:57	
Control Test		0.077	14:58	
Air Blank		0.000	14:58	
Subject Sample	# 1		14:59	Subject Test Refused
Breath Volume		0.000	14:59	
Air Blank		0.000	14:59	
Control Test		0.077	14:59	
Air Blank		0.000	15:00	
Diagnostics Check		OK	15:00	

Data Download Date/Time: 06/12/2020 14:27

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNING,LUKE,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PARSONS,C
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 05/08/2020 17:53:05

Diagnostics Check		OK	17:46
Air Blank		0.000	17:47
Control Test		0.080	17:47
Air Blank		0.000	17:48
Subject Sample	# 1	0.000	17:48
Breath Volume		1.609	17:48
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.000	17:51
Breath Volume		2.234	17:51
Air Blank		0.000	17:52
Control Test		0.080	17:52
Air Blank		0.000	17:53
Diagnostics Check		OK	17:53

Data Download Date/Time: 05/17/2020 15:25

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STONE,MICHAEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/10/2020 01:05:46

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:04	
Control Test		0.081	01:05	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 05/17/2020 15:25

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 05/10/2020 01:32:30

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.088	01:27
Breath Volume		2.140	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.081	01:30
Breath Volume		2.343	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 05/17/2020 15:25

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ALEXIS,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CREAMER,REMINGTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BOUGHNOU

DATE/TIME OF TEST: 05/10/2020 02:11:45

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.000	02:07
Breath Volume		1.968	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.000	02:10
Breath Volume		2.085	02:10
Air Blank		0.000	02:10
Control Test		0.081	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 05/17/2020 15:25

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ADKINSON,CODY
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 04/22/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 05/14/2020 07:26:38

Diagnostics Check		OK	07:19
Air Blank		0.000	07:20
Control Test		0.080	07:20
Air Blank		0.000	07:21
Subject Sample	# 1	0.216	07:21
Breath Volume		3.378	07:21
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 2	0.214	07:25
Breath Volume		3.261	07:25
Air Blank		0.000	07:25
Control Test		0.079	07:26
Air Blank		0.000	07:26
Diagnostics Check		OK	07:26

Data Download Date/Time: 05/17/2020 15:25

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGHEE,DARRYL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCGHEE,DARRYL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:56

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 05/21/2020 11:35:13

Diagnostics Check		OK	11:27
Air Blank		0.000	11:28
Control Test		0.079	11:28
Air Blank		0.000	11:29
Subject Sample	# 1	0.247	11:30
Breath Volume		2.265	11:30
Air Blank		0.000	11:30
Air Blank		0.000	11:32
Subject Sample	# 2	0.231	11:33
Breath Volume		2.109	11:33
Air Blank		0.000	11:34
Control Test		0.078	11:34
Air Blank		0.000	11:35
Diagnostics Check		OK	11:35

Data Download Date/Time: 06/28/2020 09:06

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WELLS,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 05/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: MAGOON

DATE/TIME OF TEST: 05/22/2020 03:34:43

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.080	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2		03:33	No Sample Provided
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.082	03:34	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 06/28/2020 09:06

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE,ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MAJERE,ARIC
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: ADMIN
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/01/2020 18:41:17

Diagnostics Check		OK	18:26	
Air Blank		0.000	18:27	
Control Test		0.080	18:27	
Air Blank		0.000	18:28	
Subject Sample	# 1	0.165	18:31	Volume Not Met
Breath Volume		0.761	18:31	
Air Blank		0.000	18:32	
Air Blank		0.000	18:33	
Subject Sample	# 2	0.178	18:34	
Breath Volume		1.277	18:34	
Air Blank		0.000	18:34	
Air Blank		0.000	18:36	
Subject Sample	# 3		18:39	No Sample Provided
Breath Volume		0.000	18:39	
Air Blank		0.000	18:40	
Control Test		0.079	18:40	
Air Blank		0.000	18:41	
Diagnostics Check		OK	18:41	

Data Download Date/Time: 05/11/2020 10:09

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO,MERCEDES,T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VARNADORE,RICHARD
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 04/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 05/02/2020 01:56:46

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.169	01:52
Breath Volume		2.289	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.174	01:55
Breath Volume		2.066	01:55
Air Blank		0.000	01:55
Control Test		0.078	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 05/11/2020 10:09

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: RENO,K
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/10/2020 03:34:38

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.081	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.000	03:29
Breath Volume		1.441	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.000	03:33
Breath Volume		1.671	03:33
Air Blank		0.000	03:33
Control Test		0.082	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 05/11/2020 10:09

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 05/10/2020 22:05:55

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.081	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:01
Breath Volume		3.640	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:04
Breath Volume		3.003	22:04
Air Blank		0.000	22:05
Control Test		0.081	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 05/11/2020 10:09

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: LOPEZ CHILEL

DATE/TIME OF TEST: 05/19/2020 02:20:41

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.081	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.135	02:16
Breath Volume		3.007	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.133	02:19
Breath Volume		2.589	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 06/09/2020 13:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MATAUSHEK

DATE/TIME OF TEST: 05/22/2020 23:33:10

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.081	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.145	23:28
Breath Volume		1.460	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.147	23:31
Breath Volume		1.789	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 06/09/2020 13:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: GREIS,B
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SHOCKLEY

DATE/TIME OF TEST: 05/23/2020 00:51:41

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.140	00:47
Breath Volume		2.171	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.130	00:50
Breath Volume		2.167	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 06/09/2020 13:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MOUTSCHKA,T
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 05/24/2020 00:51:11

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.148	00:46
Breath Volume		2.441	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.131	00:49
Breath Volume		1.773	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 06/09/2020 13:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,M
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: SNIPES

DATE/TIME OF TEST: 05/24/2020 21:29:49

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.081	21:24
Air Blank		0.000	21:25
Subject Sample	# 1	0.000	21:25
Breath Volume		2.437	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.000	21:28
Breath Volume		1.988	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 06/09/2020 13:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNA,BELEN,
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: GAVERN,JASON
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: LAFORCE

DATE/TIME OF TEST: 05/30/2020 16:54:46

Diagnostics Check		OK	16:43	
Air Blank		0.000	16:43	
Control Test		0.081	16:44	
Air Blank		0.000	16:44	
Subject Sample	# 1		16:47	No Sample Provided
Breath Volume		0.000	16:47	
Air Blank		0.000	16:48	
Air Blank		0.000	16:50	
Subject Sample	# 2		16:53	No Sample Provided
Breath Volume		0.000	16:53	
Air Blank		0.000	16:53	
Control Test		0.081	16:54	
Air Blank		0.000	16:54	
Diagnostics Check		OK	16:54	

Data Download Date/Time: 06/09/2020 13:11

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SALVADOR,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GARCIA,SALVADOR
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: DEL ROSARIO DE LAZAR

DATE/TIME OF TEST: 05/02/2020 18:34:59

Diagnostics Check		OK	18:28
Air Blank		0.000	18:29
Control Test		0.079	18:29
Air Blank		0.000	18:29
Subject Sample	# 1	0.202	18:30
Breath Volume		2.507	18:30
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.204	18:33
Breath Volume		2.578	18:33
Air Blank		0.000	18:34
Control Test		0.078	18:34
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHUNN,PETER,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LAWTON,BRYCE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/04/2020 09:27:07

Diagnostics Check		OK	09:20
Air Blank		0.000	09:21
Control Test		0.079	09:21
Air Blank		0.000	09:22
Subject Sample	# 1	0.000	09:22
Breath Volume		1.839	09:22
Air Blank		0.000	09:23
Air Blank		0.000	09:25
Subject Sample	# 2	0.000	09:25
Breath Volume		1.664	09:25
Air Blank		0.000	09:26
Control Test		0.079	09:26
Air Blank		0.000	09:27
Diagnostics Check		OK	09:27

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RICCARDI,A
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 05/08/2020 00:10:00

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.079	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:08	Subject Test Refused
Breath Volume		0.000	00:08	
Air Blank		0.000	00:09	
Control Test		0.079	00:09	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:10	

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ZAMFINO,ANDREW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: CARRIGAN

DATE/TIME OF TEST: 05/09/2020 20:40:28

Diagnostics Check		OK	20:31
Air Blank		0.000	20:31
Control Test		0.078	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.318	20:34
Breath Volume		2.406	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:36
Subject Sample	# 2	0.307	20:38
Breath Volume		1.687	20:38
Air Blank		0.000	20:39
Control Test		0.078	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 05/12/2020 01:54:58

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.235	01:50
Breath Volume		2.417	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.226	01:53
Breath Volume		1.500	01:53
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: ZAMFINO,ANDREW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 05/13/2020 20:37:07

Diagnostics Check		OK	20:31
Air Blank		0.000	20:31
Control Test		0.080	20:32
Air Blank		0.000	20:32
Subject Sample	# 1	0.000	20:32
Breath Volume		2.578	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.000	20:35
Breath Volume		2.187	20:35
Air Blank		0.000	20:36
Control Test		0.080	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,JAMIE,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: FAZIO,JOHN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 05/16/2020 07:51:35

Diagnostics Check		OK	07:45
Air Blank		0.000	07:45
Control Test		0.079	07:45
Air Blank		0.000	07:46
Subject Sample	# 1	0.084	07:47
Breath Volume		2.207	07:47
Air Blank		0.000	07:47
Air Blank		0.000	07:49
Subject Sample	# 2	0.077	07:50
Breath Volume		1.574	07:50
Air Blank		0.000	07:50
Control Test		0.079	07:50
Air Blank		0.000	07:51
Diagnostics Check		OK	07:51

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GHOBRIAL,MARTIN,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: YEBBA,RICKY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BEAN

DATE/TIME OF TEST: 05/18/2020 21:34:32

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.060	21:29
Breath Volume		3.269	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.053	21:33
Breath Volume		1.761	21:33
Air Blank		0.000	21:33
Control Test		0.079	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: REYNOLDS,MATTHEW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 05/23/2020 22:54:40

Diagnostics Check		OK	22:47
Air Blank		0.000	22:47
Control Test		0.080	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.246	22:49
Breath Volume		1.484	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:51
Subject Sample	# 2	0.260	22:53
Breath Volume		1.496	22:53
Air Blank		0.000	22:53
Control Test		0.078	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HELSETH,KEVIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PALLAIS GALEANO

DATE/TIME OF TEST: 05/25/2020 02:52:05

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.145	02:47
Breath Volume		3.261	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.147	02:50
Breath Volume		3.351	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RICCARDI,A
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 05/29/2020 04:35:47

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.079	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.000	04:31
Breath Volume		4.328	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.000	04:34
Breath Volume		4.296	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 05/29/2020 10:26

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SISTERELI,NICHOLAS
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 05/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: PELICO GONZALEZ

DATE/TIME OF TEST: 05/30/2020 06:21:28

Diagnostics Check		OK	06:05	
Air Blank		0.000	06:06	
Control Test		0.079	06:06	
Air Blank		0.000	06:07	
Subject Sample	# 1	0.129	06:08	
Breath Volume		1.796	06:08	
Air Blank		0.000	06:09	
Air Blank		0.000	06:11	
Subject Sample	# 2	0.115	06:14	Volume Not Met
Breath Volume		0.582	06:14	
Air Blank		0.000	06:15	
Air Blank		0.000	06:16	
Subject Sample	# 3	0.104	06:19	Volume Not Met
Breath Volume		0.406	06:19	
Air Blank		0.000	06:20	
Control Test		0.076	06:20	
Air Blank		0.000	06:21	
Diagnostics Check		OK	06:21	

Data Download Date/Time: 06/30/2020 13:06

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,SHAKAYLA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 04/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: STOREY

DATE/TIME OF TEST: 05/01/2020 03:06:56

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:04	
Control Test		0.080	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 05/01/2020 10:05

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLASSE,RAYMOND,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KAFER

DATE/TIME OF TEST: 05/02/2020 00:38:58

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.080	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.131	00:34
Breath Volume		1.582	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.139	00:37
Breath Volume		2.179	00:37
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLASSE,RAYMOND,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/03/2020 05:24:43

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:21	
Control Test		0.078	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:23	Subject Test Refused
Breath Volume		0.000	05:23	
Air Blank		0.000	05:23	
Control Test		0.078	05:24	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: A,DEAN
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 05/03/2020 19:52:37

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.079	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.177	19:48
Breath Volume		2.335	19:48
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.165	19:51
Breath Volume		2.460	19:51
Air Blank		0.000	19:51
Control Test		0.079	19:52
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BERMUDEZ,R
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: CENDEJAS

DATE/TIME OF TEST: 05/04/2020 20:39:49

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.080	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.000	20:35
Breath Volume		2.464	20:35
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.000	20:38
Breath Volume		1.468	20:38
Air Blank		0.000	20:38
Control Test		0.080	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OREHOSTKY,G
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 05/05/2020 17:58:59

Diagnostics Check		OK	17:56	
Air Blank		0.000	17:56	
Control Test		0.080	17:56	
Air Blank		0.000	17:57	
Subject Sample	# 1		17:57	Subject Test Refused
Breath Volume		0.000	17:57	
Air Blank		0.000	17:58	
Control Test		0.080	17:58	
Air Blank		0.000	17:58	
Diagnostics Check		OK	17:58	

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLASSE,RAYMOND,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CASTANO,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ODERDA

DATE/TIME OF TEST: 05/08/2020 01:27:14

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.192	01:22
Breath Volume		2.609	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.193	01:25
Breath Volume		3.316	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/08/2020 03:59:43

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.061	03:54
Breath Volume		3.210	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.057	03:58
Breath Volume		2.195	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GETCHELL,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 05/08/2020 13:15:55

Diagnostics Check		OK	13:13	
Air Blank		0.000	13:13	
Control Test		0.081	13:13	
Air Blank		0.000	13:14	
Subject Sample	# 1		13:14	Subject Test Refused
Breath Volume		0.000	13:14	
Air Blank		0.000	13:15	
Control Test		0.079	13:15	
Air Blank		0.000	13:15	
Diagnostics Check		OK	13:15	

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 05/08/2020 19:08:26

Diagnostics Check		OK	19:02
Air Blank		0.000	19:02
Control Test		0.081	19:03
Air Blank		0.000	19:03
Subject Sample	# 1	0.000	19:04
Breath Volume		2.218	19:04
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.000	19:07
Breath Volume		1.531	19:07
Air Blank		0.000	19:07
Control Test		0.080	19:07
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DAUGHTREY,MICHAEL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/09/2020 15:53:00

Diagnostics Check		OK	15:45
Air Blank		0.000	15:45
Control Test		0.079	15:46
Air Blank		0.000	15:46
Subject Sample	# 1	0.188	15:48
Breath Volume		1.257	15:48
Air Blank		0.000	15:48
Air Blank		0.000	15:50
Subject Sample	# 2	0.177	15:51
Breath Volume		1.906	15:51
Air Blank		0.000	15:52
Control Test		0.079	15:52
Air Blank		0.000	15:52
Diagnostics Check		OK	15:53

Data Download Date/Time: 06/05/2020 10:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WARD,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: DANSBURY

DATE/TIME OF TEST: 05/09/2020 19:19:49

Diagnostics Check		OK	19:13
Air Blank		0.000	19:13
Control Test		0.080	19:14
Air Blank		0.000	19:14
Subject Sample	# 1	0.183	19:15
Breath Volume		2.414	19:15
Air Blank		0.000	19:15
Air Blank		0.000	19:17
Subject Sample	# 2	0.179	19:18
Breath Volume		2.730	19:18
Air Blank		0.000	19:18
Control Test		0.079	19:19
Air Blank		0.000	19:19
Diagnostics Check		OK	19:19

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,MELINDA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CATALANO

DATE/TIME OF TEST: 05/10/2020 01:30:15

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:28	Subject Test Refused
Breath Volume		0.000	01:28	
Air Blank		0.000	01:29	
Control Test		0.079	01:29	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NESBITT,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/10/2020 22:39:21

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.346	22:34
Breath Volume		1.964	22:34
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.334	22:37
Breath Volume		1.648	22:37
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: MCCREA

DATE/TIME OF TEST: 05/11/2020 19:52:44

Diagnostics Check		OK	19:49	
Air Blank		0.000	19:50	
Control Test		0.080	19:50	
Air Blank		0.000	19:50	
Subject Sample	# 1		19:51	Subject Test Refused
Breath Volume		0.000	19:51	
Air Blank		0.000	19:51	
Control Test		0.080	19:52	
Air Blank		0.000	19:52	
Diagnostics Check		OK	19:52	

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GAST,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: COOLER

DATE/TIME OF TEST: 05/12/2020 19:58:20

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.080	19:51
Air Blank		0.000	19:52
Subject Sample	# 1	0.281	19:53
Breath Volume		1.535	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.287	19:56
Breath Volume		1.445	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 05/14/2020 01:30:27

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.242	01:22
Breath Volume		2.859	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.221	01:25
Breath Volume		1.542	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 3	0.213	01:28
Breath Volume		1.355	01:28
Air Blank		0.000	01:29
Control Test		0.078	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/14/2020 23:40:54

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.056	23:36
Breath Volume		1.128	23:36
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.057	23:39
Breath Volume		1.105	23:39
Air Blank		0.000	23:39
Control Test		0.080	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SCHIMENTI

DATE/TIME OF TEST: 05/15/2020 03:24:47

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:21	
Control Test		0.079	03:22	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:23	
Control Test		0.080	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: JOCELYN

DATE/TIME OF TEST: 05/15/2020 21:28:39

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:22
Subject Sample	# 1	0.225	21:22
Breath Volume		1.406	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.223	21:26
Breath Volume		1.597	21:26
Air Blank		0.000	21:27
Control Test		0.079	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: ESPANHOL

DATE/TIME OF TEST: 05/16/2020 22:27:40

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.181	22:21
Breath Volume		1.527	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.187	22:26
Breath Volume		2.707	22:26
Air Blank		0.000	22:26
Control Test		0.079	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 05/16/2020 23:38:49

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.080	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:37	Subject Test Refused
Breath Volume		0.000	23:37	
Air Blank		0.000	23:37	
Control Test		0.080	23:38	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/17/2020 02:38:49

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.080	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SEBRIS,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: FREDERICKS

DATE/TIME OF TEST: 05/17/2020 23:38:42

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.154	23:34
Breath Volume		1.847	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.159	23:37
Breath Volume		2.046	23:37
Air Blank		0.000	23:37
Control Test		0.080	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ISAN

DATE/TIME OF TEST: 05/20/2020 22:57:18

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.080	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.141	22:52
Breath Volume		3.125	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.132	22:55
Breath Volume		2.550	22:55
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DEROSA

DATE/TIME OF TEST: 05/23/2020 01:04:20

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.138	00:58
Breath Volume		1.640	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.131	01:02
Breath Volume		1.812	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MORENO BELLO

DATE/TIME OF TEST: 05/24/2020 02:57:33

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.080	02:55	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 05/24/2020 22:45:05

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.080	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.080	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 06/05/2020 10:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FARINACCI,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 05/24/2020 23:49:33

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.142	23:44
Breath Volume		2.644	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.148	23:48
Breath Volume		1.316	23:48
Air Blank		0.000	23:48
Control Test		0.082	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 05/25/2020 02:29:27

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.080	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 05/25/2020 03:48:37

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.000	03:48	RFI Detect
Air Blank		03:48	RFI Detect

Data Download Date/Time: 06/05/2020 10:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 05/25/2020 03:57:21

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.080	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:57	

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 05/25/2020 04:12:43

Diagnostics Check		OK	03:59	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.141	04:04	Volume Not Met
Breath Volume		0.101	04:04	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 2	0.192	04:07	
Breath Volume		2.410	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 3	0.176	04:11	
Breath Volume		1.171	04:11	
Air Blank		0.000	04:11	
Control Test		0.079	04:12	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGNORELLI,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 05/25/2020 22:25:44

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.080	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.072	22:21
Breath Volume		1.535	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.071	22:24
Breath Volume		1.433	22:24
Air Blank		0.000	22:24
Control Test		0.080	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BERMUDEZ,R
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: OLIVERI

DATE/TIME OF TEST: 05/30/2020 00:32:10

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.108	00:27
Breath Volume		2.078	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:29
Subject Sample	# 2	0.105	00:30
Breath Volume		2.781	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/30/2020 02:20:45

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.080	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:19	
Control Test		0.079	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AGUILAR,RODRIGO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MCPHEE

DATE/TIME OF TEST: 05/30/2020 03:46:55

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.230	03:42
Breath Volume		2.425	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.229	03:45
Breath Volume		2.125	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,ANGIE,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WARD,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/31/2020 00:32:38

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.080	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.165	00:27
Breath Volume		1.414	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.151	00:30
Breath Volume		1.308	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 05/31/2020 03:08:09

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.167	03:03
Breath Volume		2.039	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.160	03:06
Breath Volume		2.144	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1167545
DRY GAS EXPIRY DATE: 07/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MULLAN

DATE/TIME OF TEST: 05/31/2020 03:24:18

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.160	03:19
Breath Volume		2.226	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.171	03:22
Breath Volume		1.207	03:22
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 06/05/2020 10:42

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PAYNTER,JARED
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 05/02/2020 01:39:31

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.221	01:34
Breath Volume		1.488	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.223	01:37
Breath Volume		1.843	01:37
Air Blank		0.000	01:38
Control Test		0.081	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 05/28/2020 21:34

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PAEPLOW,DUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CORDLE

DATE/TIME OF TEST: 05/05/2020 03:54:27

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.082	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:53	Subject Test Refused
Breath Volume		0.000	03:53	
Air Blank		0.000	03:53	
Control Test		0.081	03:53	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 05/28/2020 21:34

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWELL,DECKITA,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BIRGE,ZACHARY
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 05/17/2020 14:45:37

Diagnostics Check		OK	14:39
Air Blank		0.000	14:39
Control Test		0.081	14:40
Air Blank		0.000	14:40
Subject Sample	# 1	0.280	14:41
Breath Volume		2.613	14:41
Air Blank		0.000	14:41
Air Blank		0.000	14:43
Subject Sample	# 2	0.277	14:44
Breath Volume		2.820	14:44
Air Blank		0.000	14:44
Control Test		0.081	14:45
Air Blank		0.000	14:45
Diagnostics Check		OK	14:45

Data Download Date/Time: 05/28/2020 21:34

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JOHN,P
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: VARGAS,MARIA
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/25/2020 02:01:02

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.081	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 05/28/2020 21:34

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BIRGE,ZACK
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 04/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 05/26/2020 21:34:18

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.081	21:32	
Air Blank		0.000	21:32	
Subject Sample	# 1		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:33	
Control Test		0.081	21:33	
Air Blank		0.000	21:34	
Diagnostics Check		OK	21:34	

Data Download Date/Time: 05/28/2020 21:34

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/02/2020 00:53:41

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.000	00:49
Breath Volume		1.152	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.000	00:52
Breath Volume		1.234	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: SOBEK

DATE/TIME OF TEST: 05/02/2020 16:12:16

Diagnostics Check		OK	16:05
Air Blank		0.000	16:06
Control Test		0.080	16:06
Air Blank		0.000	16:07
Subject Sample	# 1	0.000	16:07
Breath Volume		1.769	16:07
Air Blank		0.000	16:08
Air Blank		0.000	16:10
Subject Sample	# 2	0.000	16:10
Breath Volume		2.679	16:10
Air Blank		0.000	16:11
Control Test		0.080	16:11
Air Blank		0.000	16:12
Diagnostics Check		OK	16:12

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 05/02/2020 21:06:45

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.080	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.201	21:02
Breath Volume		2.949	21:02
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.204	21:05
Breath Volume		2.832	21:05
Air Blank		0.000	21:05
Control Test		0.078	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: YOUNG JR.

DATE/TIME OF TEST: 05/03/2020 03:14:33

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.278	03:09
Breath Volume		2.402	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.277	03:12
Breath Volume		2.453	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI,OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI,OFC.M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: SALO

DATE/TIME OF TEST: 05/03/2020 18:18:58

Diagnostics Check		OK	18:12
Air Blank		0.000	18:13
Control Test		0.077	18:13
Air Blank		0.000	18:13
Subject Sample	# 1	0.143	18:14
Breath Volume		3.070	18:14
Air Blank		0.000	18:15
Air Blank		0.000	18:16
Subject Sample	# 2	0.142	18:17
Breath Volume		3.609	18:17
Air Blank		0.000	18:18
Control Test		0.077	18:18
Air Blank		0.000	18:18
Diagnostics Check		OK	18:18

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: COUTO

DATE/TIME OF TEST: 05/03/2020 20:30:04

Diagnostics Check		OK	20:26	
Air Blank		0.000	20:26	
Control Test		0.080	20:26	
Air Blank		0.000	20:27	
Subject Sample	# 1		20:28	Subject Test Refused
Breath Volume		0.000	20:28	
Air Blank		0.000	20:29	
Control Test		0.080	20:29	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:30	

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 05/05/2020 02:26:32

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.172	02:21
Breath Volume		2.078	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.179	02:24
Breath Volume		2.097	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: WHITEHOUSE

DATE/TIME OF TEST: 05/07/2020 19:07:46

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.079	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.000	19:03
Breath Volume		5.000	19:03
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.000	19:06
Breath Volume		4.968	19:06
Air Blank		0.000	19:06
Control Test		0.079	19:07
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: SENTEL

DATE/TIME OF TEST: 05/08/2020 02:44:05

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.127	02:39
Breath Volume		2.246	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.126	02:42
Breath Volume		1.906	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:44

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PODHYSKI,PHILIP,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PODHYSKI,PHILIP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CALOMERIS

DATE/TIME OF TEST: 05/09/2020 00:39:19

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,OFC,W
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PERKINS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BRISSET

DATE/TIME OF TEST: 05/09/2020 01:55:23

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.080	01:40	
Air Blank		0.000	01:41	
Subject Sample	# 1	0.086	01:44	
Breath Volume		1.675	01:44	
Air Blank		0.000	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 2	0.085	01:50	Volume Not Met
Breath Volume		0.859	01:50	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 3	0.084	01:53	
Breath Volume		2.566	01:53	
Air Blank		0.000	01:54	
Control Test		0.078	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 05/12/2020 08:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,ANDREW
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/10/2020 06:18:57

Diagnostics Check		OK	06:07	
Air Blank		0.000	06:07	
Control Test		0.079	06:07	
Air Blank		0.000	06:08	
Subject Sample	# 1	0.164	06:11	Volume Not Met
Breath Volume		0.398	06:11	
Air Blank		0.000	06:12	
Air Blank		0.000	06:14	
Subject Sample	# 2	0.208	06:17	Volume Not Met
Breath Volume		0.050	06:17	
Air Blank		0.000	06:17	
Control Test		0.077	06:18	
Air Blank		0.000	06:18	
Diagnostics Check		OK	06:18	

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 05/11/2020 20:18:15

Diagnostics Check		OK	20:12
Air Blank		0.000	20:12
Control Test		0.079	20:12
Air Blank		0.000	20:13
Subject Sample	# 1	0.000	20:13
Breath Volume		3.277	20:13
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.000	20:16
Breath Volume		3.378	20:16
Air Blank		0.000	20:17
Control Test		0.078	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 05/12/2020 08:36

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 05/13/2020 01:45:51

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.106	01:41
Breath Volume		1.710	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.102	01:44
Breath Volume		1.636	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERNIER,N
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MERKT

DATE/TIME OF TEST: 05/15/2020 23:22:44

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.165	23:17
Breath Volume		1.441	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.165	23:21
Breath Volume		1.855	23:21
Air Blank		0.000	23:21
Control Test		0.078	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOW,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLOW,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 05/16/2020 19:13:40

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:01	
Control Test		0.079	19:01	
Air Blank		0.000	19:01	
Subject Sample	# 1	0.193	19:05	Volume Not Met
Breath Volume		0.714	19:05	
Air Blank		0.000	19:05	
Air Blank		0.000	19:07	
Subject Sample	# 2	0.197	19:08	
Breath Volume		1.753	19:08	
Air Blank		0.000	19:09	
Air Blank		0.000	19:10	
Subject Sample	# 3	0.183	19:12	
Breath Volume		1.328	19:12	
Air Blank		0.000	19:12	
Control Test		0.078	19:13	
Air Blank		0.000	19:13	
Diagnostics Check		OK	19:13	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 05/17/2020 06:13:19

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:12	
Control Test	0.079	06:12	
Air Blank		06:12	RFI Detect
Air Blank	0.000	06:13	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 05/17/2020 06:21:51

Diagnostics Check		OK	06:15
Air Blank		0.000	06:15
Control Test		0.078	06:15
Air Blank		0.000	06:16
Subject Sample	# 1	0.144	06:17
Breath Volume		1.808	06:17
Air Blank		0.000	06:18
Air Blank		0.000	06:19
Subject Sample	# 2	0.138	06:20
Breath Volume		1.796	06:20
Air Blank		0.000	06:20
Control Test		0.077	06:21
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/17/2020 23:10:27

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.080	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.193	23:05
Breath Volume		2.078	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.193	23:08
Breath Volume		2.980	23:08
Air Blank		0.000	23:09
Control Test		0.078	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: STINNETT

DATE/TIME OF TEST: 05/21/2020 00:42:55

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.241	00:38
Breath Volume		2.792	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:40
Subject Sample	# 2	0.247	00:41
Breath Volume		2.644	00:41
Air Blank		0.000	00:42
Control Test		0.078	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLLINS,OFC R
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: HOCHSTETLER

DATE/TIME OF TEST: 05/21/2020 20:41:54

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:38	
Control Test		0.079	20:39	
Air Blank		0.000	20:39	
Subject Sample	# 1		20:40	Slope Not Met
Breath Volume		1.007	20:40	
Air Blank		0.000	20:40	
Control Test		0.078	20:41	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLLINS,OFC R
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: HOCHSTETLER

DATE/TIME OF TEST: 05/21/2020 20:52:39

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.078	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.237	20:47
Breath Volume		1.406	20:47
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.249	20:51
Breath Volume		1.945	20:51
Air Blank		0.000	20:51
Control Test		0.076	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: GALEANA SALGADO

DATE/TIME OF TEST: 05/22/2020 05:01:15

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.080	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1	0.150	04:50	
Breath Volume		2.726	04:50	
Air Blank		0.000	04:50	
Air Blank		0.000	04:52	
Subject Sample	# 2	0.142	04:55	Volume Not Met
Breath Volume		0.371	04:55	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 3		04:59	Subject Test Refused
Breath Volume		0.167	04:59	
Air Blank		0.000	05:00	
Control Test		0.079	05:00	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: GALEANA SALGADO

DATE/TIME OF TEST: 05/22/2020 05:25:52

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:14	
Control Test		0.078	05:15	
Air Blank		0.000	05:15	
Subject Sample	# 1		05:18	No Sample Provided
Breath Volume		0.000	05:18	
Air Blank		0.000	05:19	
Air Blank		0.000	05:21	
Subject Sample	# 2		05:24	No Sample Provided
Breath Volume		0.000	05:24	
Air Blank		0.000	05:24	
Control Test		0.079	05:25	
Air Blank		0.000	05:25	
Diagnostics Check		OK	05:25	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 05/22/2020 23:44:13

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:39	
Control Test		0.080	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 1	0.147	23:40	
Breath Volume		3.246	23:40	
Air Blank		0.000	23:41	
Air Blank		0.000	23:43	
Subject Sample	# 2	0.147	23:43	
Breath Volume		3.062	23:43	
Air Blank			23:43	RFI Detect
Air Blank		0.000	23:44	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COKER,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 05/23/2020 22:22:55

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.080	22:17
Air Blank		0.000	22:18
Subject Sample	# 1	0.062	22:18
Breath Volume		2.113	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:20
Subject Sample	# 2	0.062	22:21
Breath Volume		2.800	22:21
Air Blank		0.000	22:22
Control Test		0.078	22:22
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHTAY,MEHER,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: CHTAY

DATE/TIME OF TEST: 05/24/2020 08:30:07

Diagnostics Check		OK	08:24	
Air Blank		0.000	08:25	
Control Test		0.079	08:25	
Air Blank		0.000	08:25	
Subject Sample	# 1	0.000	08:26	
Breath Volume		1.210	08:26	
Air Blank		0.000	08:27	
Air Blank			08:28	RFI Detect
Air Blank		0.000	08:29	
Control Test		0.079	08:29	
Air Blank			08:29	RFI Detect
Air Blank		0.000	08:30	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYRICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: CHTAY

DATE/TIME OF TEST: 05/24/2020 09:05:04

Diagnostics Check	OK	09:03	
Air Blank	0.000	09:03	
Control Test	0.080	09:04	
Air Blank		09:04	Sequence Aborted
Air Blank	0.000	09:05	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: CHTAY

DATE/TIME OF TEST: 05/24/2020 09:09:12

Diagnostics Check	OK	09:07	
Air Blank	0.000	09:08	
Control Test	0.079	09:08	
Air Blank		09:08	RFI Detect
Air Blank	0.000	09:09	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYRICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: CHTAY

DATE/TIME OF TEST: 05/24/2020 09:13:06

Diagnostics Check	OK	09:12	
Air Blank		09:12	RFI Detect
Air Blank	0.000	09:13	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: S,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: CHTAY

DATE/TIME OF TEST: 05/24/2020 09:19:51

Diagnostics Check		OK	09:15	
Air Blank		0.000	09:15	
Control Test		0.079	09:16	
Air Blank		0.000	09:16	
Subject Sample	# 1		09:18	Subject Test Refused
Breath Volume		0.953	09:18	
Air Blank		0.000	09:18	
Control Test		0.080	09:19	
Air Blank		0.000	09:19	
Diagnostics Check		OK	09:19	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,DEP,D
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHORT,STEVEN
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 05/25/2020 03:48:25

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.170	03:43
Breath Volume		3.769	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.163	03:46
Breath Volume		5.132	03:46
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: VUOTTO

DATE/TIME OF TEST: 05/26/2020 04:48:47

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.080	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.130	04:44
Breath Volume		4.285	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.130	04:47
Breath Volume		4.191	04:47
Air Blank		0.000	04:47
Control Test		0.080	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MCALEXANDER

DATE/TIME OF TEST: 05/26/2020 20:32:01

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:27
Subject Sample	# 1	0.000	20:27
Breath Volume		4.476	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.000	20:30
Breath Volume		4.562	20:30
Air Blank		0.000	20:31
Control Test		0.079	20:31
Air Blank		0.000	20:31
Diagnostics Check		OK	20:32

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CARDWELL,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 05/26/2020 23:44:00

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.080	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:43	
Control Test		0.080	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LAGARCE,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 05/28/2020 02:26:43

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.000	02:22
Breath Volume		3.070	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.000	02:25
Breath Volume		3.628	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 05/29/2020 17:43:08

Diagnostics Check		OK	17:36
Air Blank		0.000	17:37
Control Test		0.079	17:37
Air Blank		0.000	17:38
Subject Sample	# 1	0.000	17:38
Breath Volume		3.843	17:38
Air Blank		0.000	17:39
Air Blank		0.000	17:41
Subject Sample	# 2	0.000	17:41
Breath Volume		3.371	17:41
Air Blank		0.000	17:42
Control Test		0.079	17:42
Air Blank		0.000	17:43
Diagnostics Check		OK	17:43

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MENESES

DATE/TIME OF TEST: 05/30/2020 00:29:00

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.078	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.078	00:24
Breath Volume		3.617	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.075	00:27
Breath Volume		2.992	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:29

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: TREJOMARTINEZ

DATE/TIME OF TEST: 05/30/2020 05:19:27

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.080	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.102	05:15
Breath Volume		4.023	05:15
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.092	05:17
Breath Volume		2.089	05:17
Air Blank		0.000	05:18
Control Test		0.077	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: FITCHETT

DATE/TIME OF TEST: 05/31/2020 21:04:41

Diagnostics Check		OK	20:59	
Air Blank		0.000	20:59	
Control Test		0.078	20:59	
Air Blank		0.000	21:00	
Subject Sample	# 1	0.141	21:00	
Breath Volume		2.800	21:00	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 2	0.157	21:03	
Breath Volume		2.789	21:03	
Air Blank			21:04	RFI Detect
Air Blank		0.000	21:04	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: FITCHETT

DATE/TIME OF TEST: 05/31/2020 21:12:48

Diagnostics Check		OK	21:07	
Air Blank		0.000	21:07	
Control Test		0.079	21:08	
Air Blank		0.000	21:08	
Subject Sample	# 1	0.152	21:09	
Breath Volume		3.101	21:09	
Air Blank		0.000	21:09	
Air Blank		0.000	21:11	
Subject Sample	# 2	0.151	21:12	
Breath Volume		3.261	21:12	
Air Blank			21:12	RFI Detect
Air Blank		0.000	21:12	

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: FITCHETT

DATE/TIME OF TEST: 05/31/2020 21:21:24

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.078	21:15
Air Blank		0.000	21:16
Subject Sample	# 1	0.150	21:16
Breath Volume		3.582	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.141	21:19
Breath Volume		2.605	21:19
Air Blank		0.000	21:20
Control Test		0.078	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: DEACY

DATE/TIME OF TEST: 05/31/2020 23:31:34

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:26
Subject Sample	# 1	0.144	23:26
Breath Volume		2.707	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.137	23:30
Breath Volume		2.007	23:30
Air Blank		0.000	23:30
Control Test		0.079	23:30
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 06/30/2020 06:40

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,C,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STEELE,Z
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/10/2020 18:07:21

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:06	
Control Test	0.080	18:06	
Air Blank		18:06	RFI Detect
Air Blank	0.000	18:07	

Data Download Date/Time: 05/17/2020 09:44

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,C,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: STEELE,Z
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 04/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/10/2020 18:15:35

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.080	18:10
Air Blank		0.000	18:10
Subject Sample	# 1	0.092	18:11
Breath Volume		1.820	18:11
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.091	18:14
Breath Volume		2.179	18:14
Air Blank		0.000	18:14
Control Test		0.079	18:14
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 05/17/2020 09:44

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WADE,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/20/2020 01:04:03

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.080	01:03	
Air Blank		01:03	RFI Detect
Air Blank	0.000	01:04	

Data Download Date/Time: 06/22/2020 08:13

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WADE,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/20/2020 01:07:13

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank		01:06	RFI Detect
Air Blank	0.000	01:07	

Data Download Date/Time: 06/22/2020 08:13

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WADE,CODY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/20/2020 01:14:25

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.000	01:10
Breath Volume		4.730	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.000	01:13
Breath Volume		4.546	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 06/22/2020 08:13

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISKELL,ADAM,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: FRANQUIZ,JIM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1031121
DRY GAS EXPIRY DATE: 10/12/2021
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: PUCINO

DATE/TIME OF TEST: 05/20/2020 03:30:20

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.000	03:25
Breath Volume		1.488	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.000	03:28
Breath Volume		1.503	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/22/2020 08:13

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DEXTER

DATE/TIME OF TEST: 05/02/2020 21:16:43

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.080	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.235	21:12
Breath Volume		2.285	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.226	21:15
Breath Volume		1.886	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 05/12/2020 07:48

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: BARDEN

DATE/TIME OF TEST: 05/15/2020 17:38:07

Diagnostics Check		OK	17:34	
Air Blank		0.000	17:35	
Control Test		0.078	17:35	
Air Blank		0.000	17:35	
Subject Sample	# 1		17:36	Subject Test Refused
Breath Volume		0.000	17:36	
Air Blank		0.000	17:36	
Control Test		0.078	17:36	
Air Blank			17:37	Ambient Fail
Air Blank		0.000	17:38	

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BERGMANN

DATE/TIME OF TEST: 05/15/2020 21:21:54

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.078	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.247	21:17
Breath Volume		2.636	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.248	21:20
Breath Volume		2.824	21:20
Air Blank		0.000	21:21
Control Test		0.078	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BARDEN

DATE/TIME OF TEST: 05/15/2020 22:39:50

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.079	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.178	22:35
Breath Volume		3.046	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:37
Subject Sample	# 2	0.182	22:38
Breath Volume		3.136	22:38
Air Blank		0.000	22:38
Control Test		0.078	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: ERNST

DATE/TIME OF TEST: 05/16/2020 19:00:34

Diagnostics Check		OK	18:54
Air Blank		0.000	18:55
Control Test		0.079	18:55
Air Blank		0.000	18:55
Subject Sample	# 1	0.133	18:56
Breath Volume		1.875	18:56
Air Blank		0.000	18:56
Air Blank		0.000	18:58
Subject Sample	# 2	0.128	18:59
Breath Volume		2.074	18:59
Air Blank		0.000	18:59
Control Test		0.079	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEP J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: LAZINSKI

DATE/TIME OF TEST: 05/16/2020 20:08:29

Diagnostics Check	OK	20:07	
Air Blank		20:08	RFI Detect
Air Blank	0.000	20:08	

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEP J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: LAZINSKI

DATE/TIME OF TEST: 05/16/2020 20:17:13

Diagnostics Check		OK	20:11
Air Blank		0.000	20:11
Control Test		0.078	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.162	20:12
Breath Volume		4.042	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.157	20:15
Breath Volume		4.410	20:15
Air Blank		0.000	20:16
Control Test		0.077	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: TORREALBA

DATE/TIME OF TEST: 05/20/2020 21:59:44

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.079	21:54
Air Blank		0.000	21:54
Subject Sample	# 1	0.138	21:55
Breath Volume		4.261	21:55
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.137	21:58
Breath Volume		4.570	21:58
Air Blank		0.000	21:58
Control Test		0.078	21:59
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 05/21/2020 18:20:29

Diagnostics Check		OK	18:14
Air Blank		0.000	18:14
Control Test		0.079	18:14
Air Blank		0.000	18:15
Subject Sample	# 1	0.127	18:15
Breath Volume		2.035	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.121	18:19
Breath Volume		1.121	18:19
Air Blank		0.000	18:19
Control Test		0.079	18:19
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MCKOWN

DATE/TIME OF TEST: 05/22/2020 01:00:06

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.079	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1	0.232	00:58	Volume Not Met
Breath Volume		0.929	00:58	
Air Blank			00:58	RFI Detect
Air Blank		0.000	00:59	
Control Test		0.078	00:59	
Air Blank		0.000	01:00	
Diagnostics Check		OK	01:00	

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: MCKOWN

DATE/TIME OF TEST: 05/22/2020 01:07:23

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:03	
Control Test		0.078	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:06	RFI Detect
Breath Volume		0.367	01:06	
Air Blank			01:06	RFI Detect
Air Blank		0.000	01:07	

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MCKOWN

DATE/TIME OF TEST: 05/22/2020 01:16:28

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.234	01:11
Breath Volume		2.074	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.231	01:14
Breath Volume		1.550	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 05/22/2020 04:38:04

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.080	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.000	04:33
Breath Volume		2.734	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.000	04:36
Breath Volume		3.085	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:38

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 05/26/2020 23:14:09

Diagnostics Check		OK	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Subject Sample	# 1	0.225	23:09
Breath Volume		3.574	23:09
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.219	23:12
Breath Volume		3.156	23:12
Air Blank		0.000	23:13
Control Test		0.078	23:13
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: DAVALOS BERNAL

DATE/TIME OF TEST: 05/30/2020 02:38:09

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.079	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.226	02:33
Breath Volume		1.750	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.216	02:36
Breath Volume		2.203	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 05/31/2020 01:36:42

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:36	
Air Blank		01:36	RFI Detect
Air Blank	0.000	01:36	

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEP J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 05/31/2020 01:45:13

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.143	01:40
Breath Volume		2.441	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.129	01:43
Breath Volume		2.101	01:43
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 06/30/2020 06:45

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: GORDER

DATE/TIME OF TEST: 05/23/2020 22:45:00

Diagnostics Check		OK	22:38
Air Blank		0.000	22:39
Control Test		0.077	22:39
Air Blank		0.000	22:40
Subject Sample	# 1	0.194	22:40
Breath Volume		1.925	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.207	22:43
Breath Volume		1.484	22:43
Air Blank		0.000	22:44
Control Test		0.076	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:45

Data Download Date/Time: 06/30/2020 06:33

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEUHLER,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A4
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 05/25/2020 03:31:57

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.180	03:27
Breath Volume		3.929	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.180	03:30
Breath Volume		3.562	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 06/30/2020 06:33

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMPBELL,ADAM
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: MUSCENTE

DATE/TIME OF TEST: 05/12/2020 03:33:46

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.115	03:28
Breath Volume		1.226	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.122	03:32
Breath Volume		2.531	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 06/30/2020 07:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAYLOR,BOBBY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: DEMARCO

DATE/TIME OF TEST: 05/12/2020 22:17:16

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.210	22:12
Breath Volume		2.261	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.224	22:15
Breath Volume		1.914	22:15
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALLEN,MARTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: WHITIS

DATE/TIME OF TEST: 05/14/2020 01:16:03

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.172	01:11
Breath Volume		2.230	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.165	01:14
Breath Volume		1.953	01:14
Air Blank		0.000	01:15
Control Test		0.078	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:16

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ERICKSON,CHRISTINA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 05/16/2020 00:11:43

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.161	00:06
Breath Volume		1.726	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.165	00:10
Breath Volume		1.558	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KOVAL,KRISTEN
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CASTOR

DATE/TIME OF TEST: 05/16/2020 02:42:29

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.168	02:37
Breath Volume		3.257	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.159	02:41
Breath Volume		2.746	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: POORMON,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: LYSEK

DATE/TIME OF TEST: 05/19/2020 01:02:18

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.187	00:57
Breath Volume		2.242	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.172	01:00
Breath Volume		1.257	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMPBELL,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: OTHER
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CARBONE

DATE/TIME OF TEST: 05/21/2020 04:53:16

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:45	No Sample Provided
Breath Volume		0.000	04:45	
Air Blank		0.000	04:45	
Air Blank		0.000	04:47	
Subject Sample	# 2	0.158	04:48	
Breath Volume		2.503	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 3	0.165	04:51	
Breath Volume		2.832	04:51	
Air Blank		0.000	04:52	
Control Test		0.079	04:52	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORDSTROM,DEPUTY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NORDSTROM,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: CABRONE

DATE/TIME OF TEST: 05/21/2020 08:52:30

Diagnostics Check		OK	08:46
Air Blank		0.000	08:46
Control Test		0.079	08:47
Air Blank		0.000	08:47
Subject Sample	# 1	0.044	08:48
Breath Volume		2.171	08:48
Air Blank		0.000	08:48
Air Blank		0.000	08:50
Subject Sample	# 2	0.042	08:51
Breath Volume		1.984	08:51
Air Blank		0.000	08:51
Control Test		0.078	08:51
Air Blank		0.000	08:52
Diagnostics Check		OK	08:52

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORDSTROM,DEPUTY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NORDSTROM,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: CARBONE

DATE/TIME OF TEST: 05/21/2020 10:37:43

Diagnostics Check		OK	10:31
Air Blank		0.000	10:32
Control Test		0.079	10:32
Air Blank		0.000	10:33
Subject Sample	# 1	0.007	10:33
Breath Volume		2.433	10:33
Air Blank		0.000	10:33
Air Blank		0.000	10:36
Subject Sample	# 2	0.007	10:36
Breath Volume		2.378	10:36
Air Blank		0.000	10:36
Control Test		0.079	10:37
Air Blank		0.000	10:37
Diagnostics Check		OK	10:37

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELLISON,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: THARP

DATE/TIME OF TEST: 05/24/2020 03:46:30

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.124	03:41
Breath Volume		2.683	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.124	03:45
Breath Volume		2.292	03:45
Air Blank		0.000	03:45
Control Test		0.078	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CHANDLER,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LIBERA

DATE/TIME OF TEST: 05/25/2020 02:03:31

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.162	01:58
Breath Volume		2.261	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.164	02:02
Breath Volume		2.507	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/26/2020 17:18:03

Diagnostics Check		OK	17:06	
Air Blank		0.000	17:07	
Control Test		0.079	17:07	
Air Blank		0.000	17:07	
Subject Sample	# 1	0.000	17:11	Volume Not Met
Breath Volume		0.882	17:11	
Air Blank		0.000	17:11	
Air Blank		0.000	17:13	
Subject Sample	# 2	0.000	17:16	Volume Not Met
Breath Volume		0.929	17:16	
Air Blank		0.000	17:17	
Control Test		0.080	17:17	
Air Blank		0.000	17:17	
Diagnostics Check		OK	17:18	

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DUNN,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 05/26/2020 17:37:02

Diagnostics Check		OK	17:22	
Air Blank		0.000	17:22	
Control Test		0.079	17:23	
Air Blank		0.000	17:23	
Subject Sample	# 1	0.000	17:24	
Breath Volume		1.187	17:24	
Air Blank		0.000	17:24	
Air Blank		0.000	17:26	
Subject Sample	# 2	0.000	17:30	Volume Not Met
Breath Volume		0.062	17:30	
Air Blank		0.000	17:30	
Air Blank		0.000	17:32	
Subject Sample	# 3	0.000	17:35	Volume Not Met
Breath Volume		0.226	17:35	
Air Blank		0.000	17:36	
Control Test		0.079	17:36	
Air Blank		0.000	17:36	
Diagnostics Check		OK	17:37	

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARTER,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: ADMIN
ARREST TIME: 16:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: FILIPPI

DATE/TIME OF TEST: 05/26/2020 20:01:57

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:55	
Control Test		0.079	19:55	
Air Blank		0.000	19:56	
Subject Sample	# 1	0.000	19:59	Volume Not Met
Breath Volume		0.988	19:59	
Air Blank		0.000	20:00	
Air Blank		0.000	20:01	

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,B
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TRGO

DATE/TIME OF TEST: 05/27/2020 03:02:44

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.109	02:58
Breath Volume		4.265	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.106	03:01
Breath Volume		4.378	03:01
Air Blank		0.000	03:01
Control Test		0.079	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VETRI,THOMAS
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 05/27/2020 21:57:57

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.077	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.176	21:53
Breath Volume		1.648	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.181	21:56
Breath Volume		1.746	21:56
Air Blank		0.000	21:57
Control Test		0.078	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BARLEY,JAMES
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: OGARRO

DATE/TIME OF TEST: 05/28/2020 02:40:52

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.078	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.204	02:36
Breath Volume		1.390	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.214	02:39
Breath Volume		1.367	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WORLEY,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: FISCHER

DATE/TIME OF TEST: 05/28/2020 18:20:28

Diagnostics Check		OK	18:13
Air Blank		0.000	18:13
Control Test		0.079	18:14
Air Blank		0.000	18:14
Subject Sample	# 1	0.232	18:15
Breath Volume		2.359	18:15
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.225	18:18
Breath Volume		3.015	18:18
Air Blank		0.000	18:19
Control Test		0.077	18:19
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEUTHNER,N
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LENNON

DATE/TIME OF TEST: 05/30/2020 01:07:25

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:02
Breath Volume		2.242	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.000	01:06
Breath Volume		1.996	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,B
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10818080A5
DRY GAS EXPIRY DATE: 06/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DOYEN

DATE/TIME OF TEST: 05/30/2020 02:24:54

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.240	02:19
Breath Volume		2.351	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.242	02:23
Breath Volume		1.773	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 06/30/2020 07:57

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BIVINS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/10/2020 00:03:07

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:48	
Control Test		0.080	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 2		23:52	No Sample Provided
Breath Volume		0.000	23:52	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.227	23:56	
Breath Volume		3.394	23:56	
Air Blank		0.000	23:56	
Air Blank		0.000	23:58	
Subject Sample	# 1		00:01	No Sample Provided
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 05/27/2020 23:41

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 05/09/2020 22:01:57

Diagnostics Check		OK	21:57	
Air Blank		0.000	21:57	
Control Test		0.081	21:57	
Air Blank		0.000	21:58	
Subject Sample	# 1		22:00	Subject Test Refused
Breath Volume		0.000	22:00	
Air Blank		0.000	22:01	
Control Test		0.080	22:01	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 06/04/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKER,DEREK,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: LONG,KATHY
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 04/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: SERNA

DATE/TIME OF TEST: 05/29/2020 16:55:36

Diagnostics Check		OK	16:49
Air Blank		0.000	16:49
Control Test		0.081	16:50
Air Blank		0.000	16:50
Subject Sample	# 1	0.000	16:51
Breath Volume		4.160	16:51
Air Blank		0.000	16:51
Air Blank		0.000	16:53
Subject Sample	# 2	0.000	16:54
Breath Volume		4.296	16:54
Air Blank		0.000	16:54
Control Test		0.079	16:55
Air Blank		0.000	16:55
Diagnostics Check		OK	16:55

Data Download Date/Time: 06/04/2020 11:19

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAINS,HARRISON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/02/2020 00:02:45

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:59	
Control Test		0.080	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Control Test		0.081	00:02	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 05/06/2020 10:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCLEAN,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BARRETTODUARDO

DATE/TIME OF TEST: 05/03/2020 03:12:30

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.101	03:08
Breath Volume		3.320	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.098	03:10
Breath Volume		3.300	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 05/06/2020 10:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SOKOLOVA

DATE/TIME OF TEST: 05/06/2020 01:02:36

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.082	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.081	00:58
Breath Volume		2.308	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.081	01:00
Breath Volume		2.582	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 05/06/2020 10:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VANEK,TRAVIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: PAPANIKOLAOU

DATE/TIME OF TEST: 05/07/2020 04:48:03

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.082	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.000	04:43
Breath Volume		3.308	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.000	04:46
Breath Volume		1.664	04:46
Air Blank		0.000	04:47
Control Test		0.081	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:48

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GABALDON

DATE/TIME OF TEST: 05/08/2020 00:33:42

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.082	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.158	00:29
Breath Volume		3.839	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.155	00:32
Breath Volume		3.964	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SELPH,DANIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ALMON

DATE/TIME OF TEST: 05/09/2020 00:13:53

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:02	
Control Test		0.081	00:03	
Air Blank		0.000	00:03	
Subject Sample	# 1	0.099	00:06	Volume Not Met
Breath Volume		0.066	00:06	
Air Blank		0.000	00:07	
Air Blank		0.000	00:09	
Subject Sample	# 2	0.093	00:12	Volume Not Met
Breath Volume		0.417	00:12	
Air Blank		0.000	00:12	
Control Test		0.080	00:13	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,TAYLOR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROSARIO PAULINO

DATE/TIME OF TEST: 05/09/2020 02:48:35

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.102	02:37	
Breath Volume		1.378	02:37	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 2	0.104	02:42	Volume Not Met
Breath Volume		0.976	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.095	02:47	
Breath Volume		1.210	02:47	
Air Blank			02:48	Purge Fail
Air Blank		0.000	02:48	

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,TAYLOR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROSARIO PAULINO

DATE/TIME OF TEST: 05/09/2020 02:51:20

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.079	02:50	RFI Detect
Air Blank	0.000	02:51	

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,TAYLOR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROSARIO PAULINO

DATE/TIME OF TEST: 05/09/2020 02:54:12

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank		02:53	RFI Detect
Air Blank	0.000	02:54	

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,TAYLOR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROSARIO PAULINO

DATE/TIME OF TEST: 05/09/2020 03:02:35

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.101	02:57
Breath Volume		1.285	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.099	03:01
Breath Volume		1.632	03:01
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OSENKARSKI,MAX
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: LASHLEY

DATE/TIME OF TEST: 05/12/2020 22:56:54

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.186	22:52
Breath Volume		2.445	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:54
Subject Sample	# 2	0.178	22:55
Breath Volume		2.863	22:55
Air Blank		0.000	22:55
Control Test		0.080	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZ,JOSE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: SHETTY

DATE/TIME OF TEST: 05/13/2020 15:55:35

Diagnostics Check		OK	15:48
Air Blank		0.000	15:49
Control Test		0.081	15:49
Air Blank		0.000	15:50
Subject Sample	# 1	0.267	15:51
Breath Volume		1.839	15:51
Air Blank		0.000	15:51
Air Blank		0.000	15:53
Subject Sample	# 2	0.278	15:54
Breath Volume		2.042	15:54
Air Blank		0.000	15:54
Control Test		0.079	15:54
Air Blank		0.000	15:55
Diagnostics Check		OK	15:55

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 05/14/2020 00:04:47

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.058	23:59
Breath Volume		2.894	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.060	00:02
Breath Volume		3.492	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 05/14/2020 13:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARWOOD,ARMANDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/15/2020 01:19:52

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.081	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.000	01:15
Breath Volume		1.726	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.000	01:18
Breath Volume		2.785	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 05/21/2020 13:21

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARRERO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: FAY

DATE/TIME OF TEST: 05/16/2020 23:59:37

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.185	23:54
Breath Volume		1.582	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.180	23:58
Breath Volume		1.277	23:58
Air Blank		0.000	23:58
Control Test		0.080	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 05/21/2020 13:21

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILKES,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/20/2020 01:55:24

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.081	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.210	01:50
Breath Volume		2.097	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.198	01:53
Breath Volume		1.695	01:53
Air Blank		0.000	01:54
Control Test		0.081	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 05/21/2020 13:21

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ACHEY

DATE/TIME OF TEST: 05/20/2020 02:14:01

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.000	02:09
Breath Volume		2.824	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.000	02:12
Breath Volume		2.414	02:12
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:14

Data Download Date/Time: 05/21/2020 13:21

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WIEBER,DANIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/21/2020 19:35:06

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:29	
Control Test		0.081	19:30	
Air Blank		0.000	19:30	
Subject Sample	# 1	0.191	19:30	
Breath Volume		2.687	19:30	
Air Blank		0.000	19:31	
Air Blank		0.000	19:33	
Subject Sample	# 2	0.195	19:33	
Breath Volume		3.906	19:33	
Air Blank			19:34	Purge Fail
Air Blank		0.000	19:35	

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WIEBER,DANIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/21/2020 19:44:05

Diagnostics Check		OK	19:37
Air Blank		0.000	19:37
Control Test		0.081	19:38
Air Blank		0.000	19:38
Subject Sample	# 1	0.186	19:39
Breath Volume		2.277	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:41
Subject Sample	# 2	0.186	19:42
Breath Volume		2.621	19:42
Air Blank		0.000	19:43
Control Test		0.078	19:43
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HU SAM,JIMMY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: BIRCHELL

DATE/TIME OF TEST: 05/23/2020 04:19:17

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.193	04:14
Breath Volume		1.609	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.181	04:17
Breath Volume		1.222	04:17
Air Blank		0.000	04:18
Control Test		0.081	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HANSTINE,AUSTIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/23/2020 23:16:15

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.080	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.157	23:11
Breath Volume		2.804	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.165	23:14
Breath Volume		1.765	23:14
Air Blank		0.000	23:15
Control Test		0.080	23:15
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FITZGERALD,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: GAINA

DATE/TIME OF TEST: 05/24/2020 03:21:10

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.252	03:16
Breath Volume		1.382	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:18
Subject Sample	# 2	0.234	03:19
Breath Volume		1.312	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OSENKARSKI,MAX
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ALHARBI

DATE/TIME OF TEST: 05/24/2020 04:35:09

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.165	04:30
Breath Volume		1.355	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.171	04:33
Breath Volume		1.601	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORELLI,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/24/2020 07:35:02

Diagnostics Check		OK	07:23	
Air Blank		0.000	07:23	
Control Test		0.081	07:24	
Air Blank		0.000	07:24	
Subject Sample	# 1	0.196	07:27	Volume Not Met
Breath Volume		0.390	07:27	
Air Blank		0.000	07:28	
Air Blank		0.000	07:30	
Subject Sample	# 2	0.206	07:33	Volume Not Met
Breath Volume		0.253	07:33	
Air Blank		0.000	07:34	
Control Test		0.079	07:34	
Air Blank		0.000	07:34	
Diagnostics Check		OK	07:35	

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLARD,DALTON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: REKSTEN

DATE/TIME OF TEST: 05/25/2020 03:25:22

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.081	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2		03:24	No Sample Provided
Breath Volume		0.000	03:24	
Air Blank		0.000	03:24	
Control Test		0.081	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOOVER,ADAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:49

SUBJECT'S LAST NAME: LATTIMORE

DATE/TIME OF TEST: 05/25/2020 07:20:36

Diagnostics Check		OK	07:14
Air Blank		0.000	07:14
Control Test		0.081	07:14
Air Blank		0.000	07:15
Subject Sample	# 1	0.093	07:15
Breath Volume		1.347	07:15
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.088	07:18
Breath Volume		2.992	07:18
Air Blank		0.000	07:19
Control Test		0.081	07:19
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 05/28/2020 16:18

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BARBER,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: BUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 05/02/2020 22:33:53

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.134	22:29
Breath Volume		2.332	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.140	22:32
Breath Volume		1.710	22:32
Air Blank		0.000	22:32
Control Test		0.081	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 05/27/2020 01:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HANEY,M
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/03/2020 20:10:06

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:02	
Control Test		0.080	20:02	
Air Blank		0.000	20:02	
Subject Sample	# 1		20:06	No Sample Provided
Breath Volume		0.000	20:06	
Air Blank		0.000	20:06	
Air Blank		0.000	20:08	
Subject Sample	# 2		20:08	Subject Test Refused
Breath Volume		0.000	20:08	
Air Blank		0.000	20:09	
Control Test		0.080	20:09	
Air Blank		0.000	20:10	
Diagnostics Check		OK	20:10	

Data Download Date/Time: 05/27/2020 01:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ALEXANDER,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: CORNETT

DATE/TIME OF TEST: 05/16/2020 17:30:56

Diagnostics Check		OK	17:24
Air Blank		0.000	17:25
Control Test		0.080	17:25
Air Blank		0.000	17:25
Subject Sample	# 1	0.000	17:26
Breath Volume		1.457	17:26
Air Blank		0.000	17:27
Air Blank		0.000	17:29
Subject Sample	# 2	0.000	17:29
Breath Volume		1.371	17:29
Air Blank		0.000	17:30
Control Test		0.080	17:30
Air Blank		0.000	17:30
Diagnostics Check		OK	17:30

Data Download Date/Time: 05/27/2020 01:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T'NEISHA,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RAMOS,P
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 05/19/2020 09:24:09

Diagnostics Check		OK	09:18
Air Blank		0.000	09:18
Control Test		0.081	09:18
Air Blank		0.000	09:19
Subject Sample	# 1	0.000	09:19
Breath Volume		1.867	09:19
Air Blank		0.000	09:20
Air Blank		0.000	09:22
Subject Sample	# 2	0.000	09:22
Breath Volume		2.136	09:22
Air Blank		0.000	09:23
Control Test		0.080	09:23
Air Blank		0.000	09:24
Diagnostics Check		OK	09:24

Data Download Date/Time: 05/27/2020 01:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BARBER,E
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 04/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: BUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 05/23/2020 18:37:05

Diagnostics Check		OK	18:27
Air Blank		0.000	18:27
Control Test		0.080	18:27
Air Blank		0.000	18:28
Subject Sample	# 1	0.328	18:29
Breath Volume		1.738	18:29
Air Blank		0.000	18:29
Air Blank		0.000	18:31
Subject Sample	# 2	0.353	18:32
Breath Volume		3.195	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:35
Subject Sample	# 3	0.346	18:35
Breath Volume		2.734	18:35
Air Blank		0.000	18:36
Control Test		0.079	18:36
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 05/27/2020 01:01

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: RASMUS

DATE/TIME OF TEST: 05/30/2020 03:24:42

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.155	03:20
Breath Volume		1.156	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.161	03:23
Breath Volume		2.531	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 06/30/2020 05:44

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BAUMLIN

DATE/TIME OF TEST: 05/01/2020 00:09:38

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.081	00:04
Air Blank		0.000	00:04
Subject Sample	# 1	0.129	00:05
Breath Volume		2.414	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.125	00:08
Breath Volume		1.492	00:08
Air Blank		0.000	00:08
Control Test		0.081	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 05/06/2020 12:49

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: SWEIGART

DATE/TIME OF TEST: 05/01/2020 18:07:48

Diagnostics Check		OK	17:59	
Air Blank		0.000	17:59	
Control Test		0.081	18:00	
Air Blank		0.000	18:00	
Subject Sample	# 1		18:03	No Sample Provided
Breath Volume		0.000	18:03	
Air Blank		0.000	18:04	
Air Blank		0.000	18:06	
Subject Sample	# 2		18:06	Subject Test Refused
Breath Volume		0.000	18:06	
Air Blank		0.000	18:06	
Control Test		0.081	18:07	
Air Blank		0.000	18:07	
Diagnostics Check		OK	18:07	

Data Download Date/Time: 05/06/2020 12:49

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: PETROS

DATE/TIME OF TEST: 05/01/2020 20:46:33

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.081	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.127	20:41
Breath Volume		2.019	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.132	20:45
Breath Volume		2.296	20:45
Air Blank		0.000	20:45
Control Test		0.081	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 05/06/2020 12:49

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: AHO

DATE/TIME OF TEST: 05/01/2020 23:50:55

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:43	No Sample Provided
Breath Volume		0.000	23:43	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 2		23:49	No Sample Provided
Breath Volume		0.000	23:49	
Air Blank		0.000	23:50	
Control Test		0.081	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 05/06/2020 12:49

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: ABSHIRE

DATE/TIME OF TEST: 05/02/2020 19:14:46

Diagnostics Check		OK	19:08
Air Blank		0.000	19:09
Control Test		0.081	19:09
Air Blank		0.000	19:09
Subject Sample	# 1	0.161	19:10
Breath Volume		2.867	19:10
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.161	19:13
Breath Volume		2.816	19:13
Air Blank		0.000	19:13
Control Test		0.081	19:14
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 05/06/2020 12:49

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: MYRICK

DATE/TIME OF TEST: 05/05/2020 14:18:09

Diagnostics Check		OK	14:13	
Air Blank		0.000	14:14	
Control Test		0.082	14:14	
Air Blank		0.000	14:15	
Subject Sample	# 1		14:16	Subject Test Refused
Breath Volume		0.000	14:16	
Air Blank		0.000	14:17	
Control Test		0.081	14:17	
Air Blank		0.000	14:18	
Diagnostics Check		OK	14:18	

Data Download Date/Time: 05/06/2020 12:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BUNN

DATE/TIME OF TEST: 05/05/2020 23:21:17

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:10	
Control Test		0.082	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:14	No Sample Provided
Breath Volume		0.000	23:14	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 2		23:19	No Sample Provided
Breath Volume		0.000	23:19	
Air Blank		0.000	23:20	
Control Test		0.081	23:20	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 05/06/2020 12:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,P
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: VAZQUEZ-GONZALEZ

DATE/TIME OF TEST: 05/06/2020 00:44:12

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.082	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.200	00:39
Breath Volume		2.683	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.204	00:42
Breath Volume		2.761	00:42
Air Blank		0.000	00:43
Control Test		0.082	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 05/06/2020 12:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KOHLBECKER

DATE/TIME OF TEST: 05/06/2020 01:37:14

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.236	01:32
Breath Volume		2.285	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.233	01:35
Breath Volume		3.492	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 05/06/2020 12:50

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: TEIBERIS

DATE/TIME OF TEST: 05/06/2020 20:08:10

Diagnostics Check		OK	20:02
Air Blank		0.000	20:02
Control Test		0.081	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.000	20:03
Breath Volume		2.523	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.000	20:06
Breath Volume		2.730	20:06
Air Blank		0.000	20:07
Control Test		0.082	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 05/06/2020 23:03:42

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:52	
Control Test		0.082	22:53	
Air Blank		0.000	22:53	
Subject Sample	# 1	0.082	22:56	Volume Not Met
Breath Volume		0.621	22:56	
Air Blank		0.000	22:57	
Air Blank		0.000	22:59	
Subject Sample	# 2	0.087	23:02	Volume Not Met
Breath Volume		0.371	23:02	
Air Blank		0.000	23:02	
Control Test		0.081	23:03	
Air Blank		0.000	23:03	
Diagnostics Check		OK	23:03	

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 05/06/2020 23:18:21

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.081	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.090	23:13
Breath Volume		1.152	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.087	23:16
Breath Volume		1.449	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: REINKE

DATE/TIME OF TEST: 05/08/2020 00:38:01

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1	0.221	00:30	Volume Not Met
Breath Volume		0.707	00:30	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2	0.270	00:33	
Breath Volume		2.550	00:33	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 3	0.277	00:36	
Breath Volume		1.730	00:36	
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 05/08/2020 01:25:42

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.204	01:20
Breath Volume		2.433	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.191	01:24
Breath Volume		2.156	01:24
Air Blank		0.000	01:24
Control Test		0.081	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ESPINO

DATE/TIME OF TEST: 05/10/2020 00:06:28

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.379	00:02
Breath Volume		3.136	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.372	00:04
Breath Volume		3.074	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/12/2020 02:11:43

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.081	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.011	02:07
Breath Volume		1.710	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.011	02:10
Breath Volume		2.074	02:10
Air Blank		0.000	02:10
Control Test		0.081	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: GRESAK

DATE/TIME OF TEST: 05/12/2020 13:53:37

Diagnostics Check		OK	13:50	
Air Blank		0.000	13:51	
Control Test		0.081	13:51	
Air Blank		0.000	13:51	
Subject Sample	# 1		13:52	Subject Test Refused
Breath Volume		0.000	13:52	
Air Blank		0.000	13:52	
Control Test		0.082	13:53	
Air Blank		0.000	13:53	
Diagnostics Check		OK	13:53	

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: HAGLICH

DATE/TIME OF TEST: 05/12/2020 18:03:44

Diagnostics Check		OK	17:57
Air Blank		0.000	17:57
Control Test		0.081	17:58
Air Blank		0.000	17:58
Subject Sample	# 1	0.195	17:59
Breath Volume		3.179	17:59
Air Blank		0.000	17:59
Air Blank		0.000	18:01
Subject Sample	# 2	0.200	18:02
Breath Volume		3.128	18:02
Air Blank		0.000	18:02
Control Test		0.081	18:03
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: BRADISH

DATE/TIME OF TEST: 05/13/2020 22:50:10

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:38
Air Blank		0.000	22:39
Subject Sample	# 1	0.228	22:40
Breath Volume		1.343	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.255	22:45
Breath Volume		2.480	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:47
Subject Sample	# 3	0.245	22:48
Breath Volume		2.003	22:48
Air Blank		0.000	22:49
Control Test		0.080	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/14/2020 04:28:05

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.092	04:23
Breath Volume		2.480	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.086	04:26
Breath Volume		1.824	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 06/03/2020 08:53

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: BASCHUK

DATE/TIME OF TEST: 05/14/2020 21:55:34

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.082	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.124	21:51
Breath Volume		1.621	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.124	21:54
Breath Volume		3.140	21:54
Air Blank		0.000	21:54
Control Test		0.082	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: VOSE

DATE/TIME OF TEST: 05/14/2020 22:30:55

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.082	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.008	22:26
Breath Volume		1.687	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.000	22:29
Breath Volume		1.777	22:29
Air Blank		0.000	22:30
Control Test		0.081	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 05/14/2020 23:16:29

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:05	
Control Test		0.080	23:05	
Air Blank		0.000	23:06	
Subject Sample	# 1		23:09	No Sample Provided
Breath Volume		0.000	23:09	
Air Blank		0.000	23:10	
Air Blank		0.000	23:12	
Subject Sample	# 2		23:15	No Sample Provided
Breath Volume		0.000	23:15	
Air Blank		0.000	23:15	
Control Test		0.081	23:15	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:56

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 05/15/2020 11:19:59

Diagnostics Check		OK	11:16	
Air Blank		0.000	11:16	
Control Test		0.081	11:17	
Air Blank		0.000	11:17	
Subject Sample	# 1		11:18	Subject Test Refused
Breath Volume		0.000	11:18	
Air Blank		0.000	11:19	
Control Test		0.081	11:19	
Air Blank		0.000	11:19	
Diagnostics Check		OK	11:19	

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLAD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: XAYMONGKOL

DATE/TIME OF TEST: 05/15/2020 15:05:11

Diagnostics Check		OK	14:58
Air Blank		0.000	14:59
Control Test		0.082	14:59
Air Blank		0.000	14:59
Subject Sample	# 1	0.000	15:00
Breath Volume		2.488	15:00
Air Blank		0.000	15:00
Air Blank		0.000	15:02
Subject Sample	# 2	0.000	15:03
Breath Volume		1.562	15:03
Air Blank		0.000	15:04
Control Test		0.082	15:04
Air Blank		0.000	15:05
Diagnostics Check		OK	15:05

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/16/2020 01:12:46

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:58	
Control Test		0.080	00:58	
Air Blank		0.000	00:59	
Subject Sample	# 1		01:02	No Sample Provided
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 2	0.208	01:05	
Breath Volume		1.457	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 3	0.204	01:11	
Breath Volume		1.394	01:11	
Air Blank		0.000	01:11	
Control Test		0.081	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 05/16/2020 03:07:38

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.192	03:02
Breath Volume		1.144	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.201	03:06
Breath Volume		1.593	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: RENSHAW

DATE/TIME OF TEST: 05/16/2020 18:43:40

Diagnostics Check		OK	18:37
Air Blank		0.000	18:37
Control Test		0.081	18:38
Air Blank		0.000	18:38
Subject Sample	# 1	0.095	18:39
Breath Volume		1.339	18:39
Air Blank		0.000	18:39
Air Blank		0.000	18:41
Subject Sample	# 2	0.094	18:42
Breath Volume		1.199	18:42
Air Blank		0.000	18:42
Control Test		0.081	18:43
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VANDOREN

DATE/TIME OF TEST: 05/17/2020 00:59:17

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.081	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:55	No Sample Provided
Breath Volume		0.000	00:55	
Air Blank		0.000	00:55	
Air Blank		0.000	00:57	
Subject Sample	# 2		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Control Test		0.081	00:58	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: FLORES TORRES

DATE/TIME OF TEST: 05/17/2020 01:39:08

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.082	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.113	01:34
Breath Volume		1.585	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.124	01:37
Breath Volume		3.558	01:37
Air Blank		0.000	01:38
Control Test		0.081	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BUTTERFIELD,KELLEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: HELLEMS

DATE/TIME OF TEST: 05/19/2020 20:37:04

Diagnostics Check		OK	20:30
Air Blank		0.000	20:31
Control Test		0.081	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.000	20:32
Breath Volume		1.804	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.000	20:35
Breath Volume		1.886	20:35
Air Blank		0.000	20:36
Control Test		0.082	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:37

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/21/2020 00:52:49

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.082	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:45	No Sample Provided
Breath Volume		0.000	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2		00:51	No Sample Provided
Breath Volume		0.000	00:51	
Air Blank		0.000	00:51	
Control Test		0.082	00:52	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 05/21/2020 16:15:50

Diagnostics Check		OK	16:10
Air Blank		0.000	16:10
Control Test		0.082	16:10
Air Blank		0.000	16:11
Subject Sample	# 1	0.000	16:11
Breath Volume		2.109	16:11
Air Blank		0.000	16:12
Air Blank		0.000	16:14
Subject Sample	# 2	0.000	16:14
Breath Volume		2.117	16:14
Air Blank		0.000	16:14
Control Test		0.082	16:15
Air Blank		0.000	16:15
Diagnostics Check		OK	16:15

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 05/21/2020 17:40:29

Diagnostics Check		OK	17:28	
Air Blank		0.000	17:28	
Control Test		0.082	17:29	
Air Blank		0.000	17:29	
Subject Sample	# 1		17:32	No Sample Provided
Breath Volume		0.000	17:32	
Air Blank		0.000	17:33	
Air Blank		0.000	17:35	
Subject Sample	# 2	0.297	17:35	
Breath Volume		1.375	17:35	
Air Blank		0.000	17:36	
Air Blank		0.000	17:38	
Subject Sample	# 3	0.290	17:38	
Breath Volume		2.195	17:38	
Air Blank		0.000	17:39	
Control Test		0.081	17:39	
Air Blank		0.000	17:40	
Diagnostics Check		OK	17:40	

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: QUIAN

DATE/TIME OF TEST: 05/22/2020 00:23:39

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.254	00:15
Breath Volume		1.281	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:17
Subject Sample	# 2	0.278	00:18
Breath Volume		1.597	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 3	0.249	00:22
Breath Volume		1.292	00:22
Air Blank		0.000	00:22
Control Test		0.080	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MORADOS

DATE/TIME OF TEST: 05/22/2020 19:06:56

Diagnostics Check		OK	18:58
Air Blank		0.000	18:58
Control Test		0.082	18:59
Air Blank		0.000	18:59
Subject Sample	# 1	0.000	19:01
Breath Volume		1.109	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.000	19:05
Breath Volume		1.164	19:05
Air Blank		0.000	19:06
Control Test		0.082	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,P
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: FARROW

DATE/TIME OF TEST: 05/22/2020 20:44:31

Diagnostics Check		OK	20:34	
Air Blank		0.000	20:34	
Control Test		0.081	20:34	
Air Blank		0.000	20:35	
Subject Sample	# 1		20:38	No Sample Provided
Breath Volume		0.000	20:38	
Air Blank		0.000	20:38	
Air Blank		0.000	20:40	
Subject Sample	# 2		20:43	Subject Test Refused
Breath Volume		0.000	20:43	
Air Blank		0.000	20:43	
Control Test		0.080	20:43	
Air Blank		0.000	20:44	
Diagnostics Check		OK	20:44	

Data Download Date/Time: 06/03/2020 08:54

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: LOVING

DATE/TIME OF TEST: 05/22/2020 23:18:04

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.081	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.155	23:13
Breath Volume		3.250	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.156	23:16
Breath Volume		3.718	23:16
Air Blank		0.000	23:17
Control Test		0.081	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:18

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/23/2020 00:53:04

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.078	00:45	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:49	No Sample Provided
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2		00:51	Subject Test Refused
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Control Test		0.080	00:52	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: NEMBHARD

DATE/TIME OF TEST: 05/23/2020 02:15:47

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.000	02:11
Breath Volume		2.816	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.000	02:14
Breath Volume		2.910	02:14
Air Blank		0.000	02:14
Control Test		0.081	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,P
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: PAYEZ

DATE/TIME OF TEST: 05/24/2020 02:30:40

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.081	02:20	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:23	No Sample Provided
Breath Volume		0.000	02:23	
Air Blank		0.000	02:24	
Air Blank		0.000	02:26	
Subject Sample	# 2		02:29	No Sample Provided
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.082	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PAYEZ

DATE/TIME OF TEST: 05/24/2020 02:54:59

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.185	02:50
Breath Volume		2.222	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:52
Subject Sample	# 2	0.177	02:53
Breath Volume		1.550	02:53
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 05/24/2020 19:53:47

Diagnostics Check		OK	19:50	
Air Blank		0.000	19:51	
Control Test		0.081	19:51	
Air Blank		0.000	19:52	
Subject Sample	# 1		19:52	Subject Test Refused
Breath Volume		0.000	19:52	
Air Blank		0.000	19:52	
Control Test		0.081	19:53	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: SADDLER

DATE/TIME OF TEST: 05/24/2020 22:49:36

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.081	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.122	22:45
Breath Volume		4.292	22:45
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.107	22:48
Breath Volume		2.335	22:48
Air Blank		0.000	22:48
Control Test		0.080	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 05/25/2020 05:07:12

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.082	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.017	05:02
Breath Volume		2.972	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.015	05:05
Breath Volume		1.921	05:05
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WILLIAMS JR.

DATE/TIME OF TEST: 05/25/2020 22:58:19

Diagnostics Check		OK	22:51
Air Blank		0.000	22:52
Control Test		0.081	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.125	22:53
Breath Volume		3.437	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.123	22:56
Breath Volume		3.148	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:04

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 05/26/2020 10:30:49

Diagnostics Check		OK	10:24
Air Blank		0.000	10:25
Control Test		0.081	10:25
Air Blank		0.000	10:26
Subject Sample	# 1	0.176	10:26
Breath Volume		2.878	10:26
Air Blank		0.000	10:27
Air Blank		0.000	10:28
Subject Sample	# 2	0.180	10:29
Breath Volume		3.062	10:29
Air Blank		0.000	10:29
Control Test		0.080	10:30
Air Blank		0.000	10:30
Diagnostics Check		OK	10:30

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BUTTERFIELD,KELLY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/27/2020 02:26:10

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.082	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.000	02:21
Breath Volume		2.359	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.000	02:24
Breath Volume		2.484	02:24
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: SANDSTROM

DATE/TIME OF TEST: 05/28/2020 18:14:41

Diagnostics Check		OK	18:07
Air Blank		0.000	18:08
Control Test		0.081	18:08
Air Blank		0.000	18:08
Subject Sample	# 1	0.172	18:09
Breath Volume		2.492	18:09
Air Blank		0.000	18:10
Air Blank		0.000	18:12
Subject Sample	# 2	0.164	18:13
Breath Volume		2.738	18:13
Air Blank		0.000	18:13
Control Test		0.080	18:14
Air Blank		0.000	18:14
Diagnostics Check		OK	18:14

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 05/28/2020 21:14:27

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.081	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.000	21:10
Breath Volume		2.628	21:10
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.000	21:13
Breath Volume		2.703	21:13
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 05/28/2020 22:53:05

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.081	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.152	22:48
Breath Volume		2.417	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.149	22:51
Breath Volume		2.585	22:51
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BAEHMAN

DATE/TIME OF TEST: 05/29/2020 01:58:05

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:57	
Control Test		0.081	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DONTATE,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 05/29/2020 12:16:28

Diagnostics Check		OK	12:10
Air Blank		0.000	12:11
Control Test		0.081	12:11
Air Blank		0.000	12:11
Subject Sample	# 1	0.000	12:12
Breath Volume		3.031	12:12
Air Blank		0.000	12:12
Air Blank		0.000	12:14
Subject Sample	# 2	0.000	12:15
Breath Volume		2.507	12:15
Air Blank		0.000	12:15
Control Test		0.081	12:15
Air Blank		0.000	12:16
Diagnostics Check		OK	12:16

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: DRESCHER

DATE/TIME OF TEST: 05/29/2020 22:35:45

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.081	22:33	
Air Blank		0.000	22:33	
Subject Sample	# 1		22:34	Subject Test Refused
Breath Volume		0.000	22:34	
Air Blank		0.000	22:34	
Control Test		0.081	22:35	
Air Blank		0.000	22:35	
Diagnostics Check		OK	22:35	

Data Download Date/Time: 06/03/2020 08:55

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAFFERTY,NAOMI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 05/29/2020 23:51:12

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:44	No Sample Provided
Breath Volume		0.000	23:44	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 2		23:49	No Sample Provided
Breath Volume		0.000	23:49	
Air Blank		0.000	23:50	
Control Test		0.082	23:50	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:51	

Data Download Date/Time: 06/03/2020 08:56

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SUFFRIDGE

DATE/TIME OF TEST: 05/30/2020 00:30:44

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.146	00:25
Breath Volume		1.609	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.147	00:29
Breath Volume		2.371	00:29
Air Blank		0.000	00:29
Control Test		0.081	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 06/03/2020 08:56

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMOUR,ZACHARY,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 05/30/2020 02:54:31

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.128	02:50
Breath Volume		1.113	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.139	02:53
Breath Volume		1.210	02:53
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 06/03/2020 08:56

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/06/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 05/30/2020 03:44:38

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.229	03:40
Breath Volume		2.453	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.215	03:43
Breath Volume		1.753	03:43
Air Blank		0.000	03:43
Control Test		0.081	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 06/03/2020 08:56

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: NORTH MIAMI PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34418080A3
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LESPINASSE

DATE/TIME OF TEST: 05/26/2020 01:43:14

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.081	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.231	01:32	
Breath Volume		1.933	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 2		01:38	No Sample Provided
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 3		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Control Test		0.079	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 06/05/2020 14:44

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,COLLIN,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: DUNN,C
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ZUBER

DATE/TIME OF TEST: 05/17/2020 00:05:57

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.080	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.198	23:59
Breath Volume		1.816	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.196	00:04
Breath Volume		2.113	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 06/05/2020 19:02

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,EDWIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COLON,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: TURPIN

DATE/TIME OF TEST: 05/17/2020 07:07:05

Diagnostics Check		OK	06:54	
Air Blank		0.000	06:55	
Control Test		0.081	06:55	
Air Blank		0.000	06:56	
Subject Sample	# 1		06:59	No Sample Provided
Breath Volume		0.000	06:59	
Air Blank		0.000	06:59	
Air Blank		0.000	07:01	
Subject Sample	# 2	0.163	07:02	
Breath Volume		1.691	07:02	
Air Blank		0.000	07:03	
Air Blank		0.000	07:04	
Subject Sample	# 3	0.160	07:05	
Breath Volume		1.324	07:05	
Air Blank		0.000	07:06	
Control Test		0.079	07:06	
Air Blank		0.000	07:07	
Diagnostics Check		OK	07:07	

Data Download Date/Time: 06/05/2020 19:02

Printed: 17 August 2020

COBRA ©1997-2014 CMI, Inc.