Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BERRY,VICTORIA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/01/2020 00:17:51

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.000	00:13
Breath Volume		1.433	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.000	00:16
Breath Volume		1.656	00:16
Air Blank		0.000	00:16
Control Test		0.080	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MORGAN,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: MILLERET

DATE/TIME OF TEST: 08/01/2020 11:25:50

Diagnostics Check		OK	11:19
Air Blank		0.000	11:19
Control Test		0.081	11:20
Air Blank		0.000	11:20
Subject Sample	# 1	0.186	11:21
Breath Volume		3.882	11:21
Air Blank		0.000	11:22
Air Blank		0.000	11:23
Subject Sample	# 2	0.182	11:24
Breath Volume		3.929	11:24
Air Blank		0.000	11:24
Control Test		0.078	11:25
Air Blank		0.000	11:25
Diagnostics Check		OK	11:25

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SOREY

DATE/TIME OF TEST: 08/05/2020 14:59:51

Diagnostics Check		OK	14:53
Air Blank		0.000	14:53
Control Test		0.081	14:54
Air Blank		0.000	14:54
Subject Sample	# 1	0.000	14:55
Breath Volume		1.207	14:55
Air Blank		0.000	14:55
Air Blank		0.000	14:57
Subject Sample	# 2	0.000	14:58
Breath Volume		1.324	14:58
Air Blank		0.000	14:58
Control Test		0.081	14:59
Air Blank		0.000	14:59
Diagnostics Check		OK	14:59

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: WALSH,A
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 08/08/2020 09:31:45

Diagnostics Check		OK	09:26	
Air Blank		0.000	09:26	
Control Test		0.080	09:26	
Air Blank		0.000	09:27	
Subject Sample	# 1		09:30	Subject Test Refused
Breath Volume		0.000	09:30	
Air Blank		0.000	09:30	
Control Test		0.080	09:31	
Air Blank		0.000	09:31	
Diagnostics Check		OK	09:31	

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: WALSH,A
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 08/08/2020 09:59:31

Diagnostics Check		OK	09:50	
Air Blank		0.000	09:51	
Control Test		0.080	09:51	
Air Blank		0.000	09:52	
Subject Sample	# 1		09:55	No Sample Provided
Breath Volume		0.000	09:55	
Air Blank		0.000	09:55	
Air Blank		0.000	09:57	
Subject Sample	# 2		09:58	Subject Test Refused
Breath Volume		0.000	09:58	
Air Blank		0.000	09:58	
Control Test		0.080	09:58	
Air Blank		0.000	09:59	
Diagnostics Check		OK	09:59	

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 08/08/2020 19:11:47

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.081	19:02
Air Blank		0.000	19:03
Subject Sample	# 1	0.224	19:03
Breath Volume		3.394	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.199	19:06
Breath Volume		1.617	19:06
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 3	0.209	19:10
Breath Volume		2.144	19:10
Air Blank		0.000	19:10
Control Test		0.079	19:11
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 08/09/2020 01:41:00

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.314	01:36
Breath Volume		1.945	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:38
Subject Sample	# 2	0.318	01:39
Breath Volume		1.875	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:41

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: KOWALAK

DATE/TIME OF TEST: 08/09/2020 19:09:42

Diagnostics Check		OK	19:03
Air Blank		0.000	19:04
Control Test		0.080	19:04
Air Blank		0.000	19:04
Subject Sample	# 1	0.196	19:05
Breath Volume		1.628	19:05
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.199	19:08
Breath Volume		1.648	19:08
Air Blank		0.000	19:08
Control Test		0.080	19:09
Air Blank		0.000	19:09
Diagnostics Check		OK	19:09

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOPEZ,H
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: ASUNMAA

DATE/TIME OF TEST: 08/14/2020 17:06:12

Diagnostics Check		OK	16:59
Air Blank		0.000	17:00
Control Test		0.081	17:00
Air Blank		0.000	17:01
Subject Sample	# 1	0.000	17:01
Breath Volume		3.101	17:01
Air Blank		0.000	17:02
Air Blank		0.000	17:04
Subject Sample	# 2	0.000	17:04
Breath Volume		5.402	17:04
Air Blank		0.000	17:05
Control Test		0.079	17:05
Air Blank		0.000	17:06
Diagnostics Check		OK	17:06

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: CARNRIGHT

DATE/TIME OF TEST: 08/15/2020 03:19:02

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.098	03:14
Breath Volume		2.296	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.098	03:17
Breath Volume		2.652	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:19

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMUELS,ANDREW,G
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: REARDIGAN

DATE/TIME OF TEST: 08/26/2020 00:36:02

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.081	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.000	00:31
Breath Volume		2.429	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:34
Breath Volume		2.945	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:36

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: NORMANDIA

DATE/TIME OF TEST: 08/26/2020 19:55:20

Diagnostics Check		OK	19:49
Air Blank		0.000	19:49
Control Test		0.079	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.000	19:51
Breath Volume		1.324	19:51
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.000	19:53
Breath Volume		1.464	19:53
Air Blank		0.000	19:54
Control Test		0.079	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 08/28/2020 18:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BUENDIA,ERIK
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: BELFATTO

DATE/TIME OF TEST: 08/29/2020 15:19:26

Diagnostics Check		OK	15:13
Air Blank		0.000	15:13
Control Test		0.080	15:14
Air Blank		0.000	15:14
Subject Sample	# 1	0.198	15:14
Breath Volume		1.226	15:14
Air Blank		0.000	15:15
Air Blank		0.000	15:17
Subject Sample	# 2	0.196	15:17
Breath Volume		1.468	15:17
Air Blank		0.000	15:18
Control Test		0.079	15:18
Air Blank		0.000	15:19
Diagnostics Check		OK	15:19

Data Download Date/Time: 10/02/2020 18:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 05/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: AMBROLADZE

DATE/TIME OF TEST: 08/29/2020 23:43:50

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.078	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.153	23:38
Breath Volume		1.226	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.145	23:42
Breath Volume		1.683	23:42
Air Blank		0.000	23:42
Control Test		0.076	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 09/08/2020 20:05

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05218080A3
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: WIKE

DATE/TIME OF TEST: 08/30/2020 00:26:48

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:24	
Control Test		0.078	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.078	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 09/08/2020 20:05

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: SON

DATE/TIME OF TEST: 08/31/2020 01:03:49

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.249	00:59
Breath Volume		2.402	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.264	01:02
Breath Volume		2.363	01:02
Air Blank		0.000	01:02
Control Test		0.077	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 09/08/2020 20:05

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VETRI,N/A
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 08/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: VAILS

DATE/TIME OF TEST: 08/16/2020 03:07:37

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.249	03:02
Breath Volume		2.582	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:04
Subject Sample	# 2	0.246	03:06
Breath Volume		2.636	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 08/16/2020 22:09

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,E
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A2
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BERNHARD

DATE/TIME OF TEST: 08/16/2020 03:31:12

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.139	03:26
Breath Volume		1.195	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.141	03:29
Breath Volume		1.601	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 08/16/2020 22:09

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: FULKS

DATE/TIME OF TEST: 08/02/2020 01:32:52

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.000	01:28
Breath Volume		2.253	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:31
Breath Volume		2.613	01:31
Air Blank		0.000	01:31
Control Test		0.078	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,SETH,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,SETH
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/02/2020 21:20:16

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.078	21:09
Air Blank		0.000	21:10
Subject Sample	# 1	0.088	21:13
Breath Volume		2.761	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.081	21:18
Breath Volume		2.363	21:18
Air Blank		0.000	21:19
Control Test		0.079	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: LASSITER

DATE/TIME OF TEST: 08/07/2020 02:35:28

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.077	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.076	02:30
Breath Volume		1.957	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.071	02:33
Breath Volume		1.523	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MACLEOD

DATE/TIME OF TEST: 08/14/2020 19:31:25

Diagnostics Check		OK	19:25
Air Blank		0.000	19:25
Control Test		0.077	19:25
Air Blank		0.000	19:26
Subject Sample	# 1	0.000	19:27
Breath Volume		4.515	19:27
Air Blank		0.000	19:27
Air Blank		0.000	19:29
Subject Sample	# 2	0.000	19:30
Breath Volume		4.855	19:30
Air Blank		0.000	19:30
Control Test		0.077	19:30
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: PAES

DATE/TIME OF TEST: 08/15/2020 01:47:21

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.026	01:42
Breath Volume		4.285	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.023	01:45
Breath Volume		2.835	01:45
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 08/15/2020 03:47:44

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.121	03:42
Breath Volume		2.843	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.109	03:46
Breath Volume		1.746	03:46
Air Blank		0.000	03:46
Control Test		0.077	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MOGLE

DATE/TIME OF TEST: 08/19/2020 23:08:00

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.000	23:03
Breath Volume		2.976	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.000	23:06
Breath Volume		3.105	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:08

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 08/20/2020 02:23:33

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.078	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.130	02:18
Breath Volume		2.605	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.131	02:21
Breath Volume		3.320	02:21
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 08/20/2020 21:41:58

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:38	
Control Test		0.078	21:38	
Air Blank		0.000	21:38	
Subject Sample	# 1		21:40	Subject Test Refused
Breath Volume		0.000	21:40	
Air Blank		0.000	21:41	
Control Test		0.078	21:41	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 08/20/2020 21:51:23

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.077	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.339	21:46
Breath Volume		2.933	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.327	21:49
Breath Volume		2.582	21:49
Air Blank		0.000	21:50
Control Test		0.077	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASCKO,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LASCKO,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 08/24/2020 02:18:17

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.079	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1	0.176	02:13	
Breath Volume		2.992	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.174	02:16	
Breath Volume		3.011	02:16	
Air Blank		0.000	02:17	
Control Test		0.073	02:17	Control Outside Tolerance
Air Blank		0.000	02:18	

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: TREMMEL

DATE/TIME OF TEST: 08/24/2020 16:59:49

Diagnostics Check		OK	16:50
Air Blank		0.000	16:50
Control Test		0.078	16:50
Air Blank		0.000	16:51
Subject Sample	# 1	0.213	16:51
Breath Volume		2.476	16:51
Air Blank		0.000	16:52
Air Blank		0.000	16:54
Subject Sample	# 2	0.190	16:55
Breath Volume		2.300	16:55
Air Blank		0.000	16:55
Air Blank		0.000	16:57
Subject Sample	# 3	0.196	16:58
Breath Volume		2.281	16:58
Air Blank		0.000	16:58
Control Test		0.077	16:59
Air Blank		0.000	16:59
Diagnostics Check		OK	16:59

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARNES,B,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KARNES,B
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:59

SUBJECT'S LAST NAME: BRINSON

DATE/TIME OF TEST: 08/26/2020 09:32:53

Diagnostics Check		OK	09:24
Air Blank		0.000	09:24
Control Test		0.077	09:24
Air Blank		0.000	09:25
Subject Sample	# 1	0.207	09:26
Breath Volume		1.292	09:26
Air Blank		0.000	09:27
Air Blank		0.000	09:29
Subject Sample	# 2	0.194	09:31
Breath Volume		1.777	09:31
Air Blank		0.000	09:31
Control Test		0.076	09:32
Air Blank		0.000	09:32
Diagnostics Check		OK	09:32

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 08/29/2020 00:21:49

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.078	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.044	00:15
Breath Volume		3.128	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.044	00:20
Breath Volume		3.785	00:20
Air Blank		0.000	00:20
Control Test		0.078	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 08/30/2020 16:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WOODIE,L
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 08/29/2020 04:51:54

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.078	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.217	04:47
Breath Volume		3.121	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:49
Subject Sample	# 2	0.214	04:50
Breath Volume		2.269	04:50
Air Blank		0.000	04:50
Control Test		0.077	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 08/30/2020 16:39

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/30/2020 02:50:09

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.077	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:39	No Sample Provided
Breath Volume		0.000	02:39	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.072	02:43	
Breath Volume		1.726	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 3	0.070	02:48	
Breath Volume		1.339	02:48	
Air Blank		0.000	02:49	
Control Test		0.077	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 08/30/2020 16:39

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,R,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DAVIS,D
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 08/01/2020 04:30:23

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:22	
Control Test		0.079	04:22	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:26	No Sample Provided
Breath Volume		0.000	04:26	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:29	
Control Test		0.079	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: TOOLEY

DATE/TIME OF TEST: 08/01/2020 21:42:37

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:31	
Control Test		0.079	21:31	
Air Blank		0.000	21:32	
Subject Sample	# 1	0.175	21:35	Volume Not Met
Breath Volume		0.417	21:35	
Air Blank		0.000	21:36	
Air Blank		0.000	21:37	
Subject Sample	# 2	0.132	21:40	Volume Not Met
Breath Volume		0.035	21:40	
Air Blank		0.000	21:41	
Control Test		0.076	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: BUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: TOOLEY

DATE/TIME OF TEST: 08/01/2020 21:59:24

Diagnostics Check		OK	21:47	
Air Blank		0.000	21:48	
Control Test		0.077	21:48	
Air Blank		0.000	21:49	
Subject Sample	# 1	0.139	21:52	Volume Not Met
Breath Volume		0.074	21:52	
Air Blank		0.000	21:52	
Air Blank		0.000	21:54	
Subject Sample	# 2	0.134	21:57	Volume Not Met
Breath Volume		0.000	21:57	
Air Blank		0.000	21:58	
Control Test		0.076	21:58	
Air Blank		0.000	21:59	
Diagnostics Check		OK	21:59	

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: TANKS

DATE/TIME OF TEST: 08/03/2020 01:24:00

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.079	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:23	
Control Test		0.079	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 08/03/2020 19:20:38

Diagnostics Check		OK	19:13
Air Blank		0.000	19:14
Control Test		0.079	19:14
Air Blank		0.000	19:15
Subject Sample	# 1	0.166	19:16
Breath Volume		1.265	19:16
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.175	19:19
Breath Volume		1.503	19:19
Air Blank		0.000	19:19
Control Test		0.077	19:20
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MARQUETTE

DATE/TIME OF TEST: 08/04/2020 01:11:33

Diagnostics Check	OK	01:11	
Air Blank		01:11	RFI Detect
Air Blank		01:11	RFI Detect

Data Download Date/Time: 08/28/2020 08:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MARQUETTE

DATE/TIME OF TEST: 08/04/2020 01:20:37

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.214	01:15
Breath Volume		2.820	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.207	01:18
Breath Volume		2.054	01:18
Air Blank		0.000	01:19
Control Test		0.078	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: KELLMAN

DATE/TIME OF TEST: 08/05/2020 03:38:39

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.079	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:37	
Control Test		0.079	03:38	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 08/08/2020 01:01:07

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.084	00:54
Breath Volume		1.699	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.080	00:59
Breath Volume		1.640	00:59
Air Blank		0.000	01:00
Control Test		0.078	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: RODRIGUEZ-ROBLERO

DATE/TIME OF TEST: 08/09/2020 02:52:12

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.181	02:47
Breath Volume		2.562	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.181	02:50
Breath Volume		1.937	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SATURNA

DATE/TIME OF TEST: 08/13/2020 02:26:12

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.078	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.128	02:24	
Breath Volume		1.285	02:24	
Air Blank			02:24	RFI Detect
Air Blank		0.000	02:25	
Control Test		0.079	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SATURNA

DATE/TIME OF TEST: 08/13/2020 02:35:48

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.133	02:30
Breath Volume		1.890	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.127	02:34
Breath Volume		1.679	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAEUSSLER,BRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HAEUSSLER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: KENNEY

DATE/TIME OF TEST: 08/16/2020 21:29:54

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.079	21:24
Air Blank		0.000	21:24
Subject Sample	# 1	0.094	21:25
Breath Volume		1.769	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:27
Subject Sample	# 2	0.097	21:28
Breath Volume		2.183	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 08/28/2020 08:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BONILLA VILLATORO

DATE/TIME OF TEST: 08/01/2020 01:40:51

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:39	
Control Test		0.078	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PIERCY

DATE/TIME OF TEST: 08/02/2020 02:16:41

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.074	02:12
Breath Volume		2.832	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.074	02:15
Breath Volume		2.882	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 08/02/2020 15:35:20

Diagnostics Check		OK	15:28
Air Blank		0.000	15:28
Control Test		0.079	15:29
Air Blank		0.000	15:29
Subject Sample	# 1	0.163	15:30
Breath Volume		3.335	15:30
Air Blank		0.000	15:31
Air Blank		0.000	15:33
Subject Sample	# 2	0.170	15:33
Breath Volume		3.359	15:33
Air Blank		0.000	15:34
Control Test		0.078	15:34
Air Blank		0.000	15:35
Diagnostics Check		OK	15:35

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 08/03/2020 01:37:03

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.212	01:32
Breath Volume		1.613	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.213	01:35
Breath Volume		1.867	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 08/03/2020 04:10:30

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.000	04:06
Breath Volume		2.855	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.000	04:09
Breath Volume		2.691	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 08/06/2020 21:03:43

Diagnostics Check		OK	21:00	
Air Blank		0.000	21:01	
Control Test		0.078	21:01	
Air Blank		0.000	21:02	
Subject Sample	# 1		21:02	Subject Test Refused
Breath Volume		0.000	21:02	
Air Blank		0.000	21:02	
Control Test		0.078	21:03	
Air Blank		0.000	21:03	
Diagnostics Check		OK	21:03	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,T,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 08/07/2020 00:27:02

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.078	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:26	
Control Test		0.079	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 08/07/2020 02:40:13

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.137	02:34
Breath Volume		1.371	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.149	02:38
Breath Volume		1.472	02:38
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: QUINTANA,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: LYDE

DATE/TIME OF TEST: 08/07/2020 05:24:35

Diagnostics Check		OK	05:17
Air Blank		0.000	05:18
Control Test		0.079	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.279	05:19
Breath Volume		1.808	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.268	05:23
Breath Volume		1.710	05:23
Air Blank		0.000	05:23
Control Test		0.078	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BOODRAM

DATE/TIME OF TEST: 08/07/2020 20:36:48

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.078	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.369	20:32
Breath Volume		2.917	20:32
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.376	20:35
Breath Volume		1.937	20:35
Air Blank		0.000	20:35
Control Test		0.078	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 08/08/2020 01:50:09

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.079	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:49	
Control Test		0.079	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/08/2020 02:59:23

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.079	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.078	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HIMEBAUGH

DATE/TIME OF TEST: 08/08/2020 23:47:14

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.079	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:45	Subject Test Refused
Breath Volume		0.000	23:45	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOVACS,TYLER,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: CREED

DATE/TIME OF TEST: 08/09/2020 05:01:10

Diagnostics Check		OK	04:58	
Air Blank		0.000	04:58	
Control Test		0.079	04:59	
Air Blank		0.000	04:59	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	05:00	
Control Test		0.079	05:00	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 08/10/2020 01:11:29

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:08	
Control Test		0.079	01:09	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:10	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: STUBITS

DATE/TIME OF TEST: 08/12/2020 20:28:37

Diagnostics Check		OK	20:22
Air Blank		0.000	20:22
Control Test		0.079	20:23
Air Blank		0.000	20:23
Subject Sample	# 1	0.268	20:24
Breath Volume		2.976	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.268	20:27
Breath Volume		2.679	20:27
Air Blank		0.000	20:27
Control Test		0.077	20:27
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 08/14/2020 08:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: KOHN

DATE/TIME OF TEST: 08/14/2020 23:50:29

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.077	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.067	23:47	
Breath Volume		1.945	23:47	
Air Blank		0.000	23:47	
Air Blank		0.000	23:49	
Subject Sample	# 2		23:49	RFI Detect
Breath Volume		0.000	23:49	
Air Blank			23:50	RFI Detect
Air Blank		0.000	23:50	

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: KOHN

DATE/TIME OF TEST: 08/15/2020 00:20:04

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.078	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.057	00:15
Breath Volume		1.167	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:17
Subject Sample	# 2	0.056	00:18
Breath Volume		1.183	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:20

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RIGBY

DATE/TIME OF TEST: 08/15/2020 02:06:33

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.050	02:02
Breath Volume		2.226	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.046	02:05
Breath Volume		1.230	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NELSON,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: REBER

DATE/TIME OF TEST: 08/15/2020 04:44:10

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.080	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.188	04:39
Breath Volume		2.910	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.187	04:42
Breath Volume		2.714	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: TRADER

DATE/TIME OF TEST: 08/16/2020 21:05:59

Diagnostics Check		OK	20:59
Air Blank		0.000	20:59
Control Test		0.079	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.087	21:01
Breath Volume		1.421	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.092	21:04
Breath Volume		1.195	21:04
Air Blank		0.000	21:05
Control Test		0.079	21:05
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 08/17/2020 03:00:46

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.109	02:56
Breath Volume		1.531	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.110	02:59
Breath Volume		1.781	02:59
Air Blank		0.000	02:59
Control Test		0.079	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: LAUB

DATE/TIME OF TEST: 08/18/2020 22:29:21

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.079	22:24
Air Blank		0.000	22:24
Subject Sample	# 1	0.019	22:25
Breath Volume		2.695	22:25
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.017	22:27
Breath Volume		1.863	22:27
Air Blank		0.000	22:28
Control Test		0.079	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 08/19/2020 05:18:49

Diagnostics Check		OK	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.106	05:14
Breath Volume		1.605	05:14
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.102	05:17
Breath Volume		1.597	05:17
Air Blank		0.000	05:17
Control Test		0.079	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/20/2020 00:03:54

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.116	23:59
Breath Volume		1.914	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.113	00:02
Breath Volume		1.195	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GATCHELL

DATE/TIME OF TEST: 08/20/2020 01:31:21

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.161	01:26
Breath Volume		2.921	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.162	01:29
Breath Volume		2.539	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SANDIFORD,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 08/21/2020 06:17:09

Diagnostics Check		OK	06:10
Air Blank		0.000	06:11
Control Test		0.080	06:11
Air Blank		0.000	06:12
Subject Sample	# 1	0.121	06:12
Breath Volume		3.691	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.118	06:15
Breath Volume		4.113	06:15
Air Blank		0.000	06:16
Control Test		0.079	06:16
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLEBER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CAQUIAS LUGO

DATE/TIME OF TEST: 08/22/2020 02:39:16

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.012	02:34
Breath Volume		3.628	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.011	02:37
Breath Volume		3.632	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,CHRISTIAN,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 08/24/2020 23:30:17

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.078	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.217	23:25
Breath Volume		1.757	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.203	23:28
Breath Volume		1.824	23:28
Air Blank		0.000	23:29
Control Test		0.078	23:29
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,M,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DRAGOMIR,I
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/25/2020 02:13:59

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.079	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Control Test		0.078	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COREY,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: YANNI

DATE/TIME OF TEST: 08/25/2020 04:20:12

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.000	04:15
Breath Volume		1.515	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.000	04:18
Breath Volume		1.742	04:18
Air Blank		0.000	04:19
Control Test		0.078	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 08/26/2020 23:51:35

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.078	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.199	23:47
Breath Volume		2.031	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.199	23:50
Breath Volume		2.316	23:50
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 08/27/2020 03:49:09

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.000	03:44
Breath Volume		2.910	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.000	03:47
Breath Volume		3.519	03:47
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 08/27/2020 12:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMAS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HUTTO

DATE/TIME OF TEST: 08/27/2020 23:29:47

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:27	
Control Test		0.079	23:27	
Air Blank		0.000	23:28	
Subject Sample	# 1		23:28	Subject Test Refused
Breath Volume		0.000	23:28	
Air Blank		0.000	23:28	
Control Test		0.078	23:29	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 09/16/2020 15:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: DULLAS

DATE/TIME OF TEST: 08/28/2020 21:11:12

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:08	
Control Test		0.079	21:09	
Air Blank		0.000	21:09	
Subject Sample	# 1		21:09	Subject Test Refused
Breath Volume		0.000	21:09	
Air Blank		0.000	21:10	
Control Test		0.078	21:10	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 09/16/2020 15:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER',D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 08/29/2020 01:29:45

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.078	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.135	01:25
Breath Volume		1.691	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.140	01:28
Breath Volume		2.234	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 09/16/2020 15:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FREDERICK,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SALS

DATE/TIME OF TEST: 08/29/2020 05:21:38

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.078	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.205	05:17
Breath Volume		2.835	05:17
Air Blank		0.000	05:17
Air Blank		0.000	05:19
Subject Sample	# 2	0.205	05:20
Breath Volume		2.519	05:20
Air Blank		0.000	05:20
Control Test		0.079	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 09/16/2020 15:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SOSA,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/29/2020 16:19:58

Diagnostics Check		OK	16:14
Air Blank		0.000	16:14
Control Test		0.079	16:14
Air Blank		0.000	16:15
Subject Sample	# 1	0.000	16:15
Breath Volume		1.535	16:15
Air Blank		0.000	16:16
Air Blank		0.000	16:18
Subject Sample	# 2	0.000	16:18
Breath Volume		1.652	16:18
Air Blank		0.000	16:19
Control Test		0.078	16:19
Air Blank		0.000	16:19
Diagnostics Check		OK	16:19

Data Download Date/Time: 09/16/2020 15:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,CHRISTIAN,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 08/29/2020 21:03:31

Diagnostics Check		OK	21:00	
Air Blank		0.000	21:01	
Control Test		0.079	21:01	
Air Blank		0.000	21:01	
Subject Sample	# 1		21:02	Subject Test Refused
Breath Volume		0.000	21:02	
Air Blank		0.000	21:02	
Control Test		0.079	21:02	
Air Blank		0.000	21:03	
Diagnostics Check		OK	21:03	

Data Download Date/Time: 09/16/2020 15:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: WILLS

DATE/TIME OF TEST: 08/31/2020 16:55:41

Diagnostics Check		OK	16:52	
Air Blank		0.000	16:53	
Control Test		0.080	16:53	
Air Blank		0.000	16:54	
Subject Sample	# 1		16:54	Subject Test Refused
Breath Volume		0.000	16:54	
Air Blank		0.000	16:54	
Control Test		0.079	16:55	
Air Blank		0.000	16:55	
Diagnostics Check		OK	16:55	

Data Download Date/Time: 09/16/2020 15:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: BLANCO,ERICA
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 08/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/22/2020 01:42:50

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.161	01:38
Breath Volume		2.550	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.154	01:41
Breath Volume		3.269	01:41
Air Blank		0.000	01:41
Control Test		0.076	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 09/02/2020 14:41

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 08/14/2020 23:36:04

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.076	23:31	
Air Blank		0.000	23:31	
Subject Sample	# 1		23:34	Subject Test Refused
Breath Volume		0.843	23:34	
Air Blank		0.000	23:35	
Control Test		0.077	23:35	
Air Blank		0.000	23:35	
Diagnostics Check		OK	23:36	

Data Download Date/Time: 09/29/2020 13:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 08/16/2020 02:37:24

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.076	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.189	02:32
Breath Volume		2.167	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.197	02:35
Breath Volume		2.175	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 09/29/2020 13:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 08/26/2020 02:20:20

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.076	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.012	02:16
Breath Volume		3.082	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.010	02:18
Breath Volume		2.488	02:18
Air Blank		0.000	02:19
Control Test		0.077	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 09/29/2020 13:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: WEBB,MATT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/31/2022
VIOLATION CODE: BUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/15/2020 17:22:00

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.080	17:15
Air Blank		0.000	17:16
Subject Sample	# 1	0.154	17:17
Breath Volume		1.472	17:17
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.153	17:20
Breath Volume		1.226	17:20
Air Blank		0.000	17:21
Control Test		0.079	17:21
Air Blank		0.000	17:21
Diagnostics Check		OK	17:22

Data Download Date/Time: 08/28/2020 21:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: WEBB,ANDREW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/31/2022
VIOLATION CODE: BUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 08/15/2020 18:40:51

Diagnostics Check		OK	18:34
Air Blank		0.000	18:35
Control Test		0.079	18:35
Air Blank		0.000	18:35
Subject Sample	# 1	0.073	18:36
Breath Volume		3.550	18:36
Air Blank		0.000	18:36
Air Blank		0.000	18:38
Subject Sample	# 2	0.079	18:39
Breath Volume		4.648	18:39
Air Blank		0.000	18:39
Control Test		0.080	18:40
Air Blank		0.000	18:40
Diagnostics Check		OK	18:40

Data Download Date/Time: 08/28/2020 21:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MELENDEZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 08/07/2020 03:19:47

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.080	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.195	03:14
Breath Volume		2.890	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.197	03:17
Breath Volume		3.597	03:17
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 08/11/2020 15:42

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: VEGODA,JOSHUA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: PEREYRA

DATE/TIME OF TEST: 08/09/2020 22:05:26

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.080	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.210	22:00
Breath Volume		1.351	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.202	22:03
Breath Volume		1.464	22:03
Air Blank		0.000	22:04
Control Test		0.079	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 08/11/2020 15:42

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CHANLOPEZ

DATE/TIME OF TEST: 08/01/2020 03:33:06

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.091	03:28
Breath Volume		2.292	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.085	03:31
Breath Volume		2.136	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 08/05/2020 18:14

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 08/01/2020 23:18:29

Diagnostics Check		OK	23:12
Air Blank		0.000	23:12
Control Test		0.080	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.000	23:14
Breath Volume		3.222	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.000	23:17
Breath Volume		3.597	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 08/05/2020 18:14

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: PEREZ RUIZ

DATE/TIME OF TEST: 08/02/2020 21:11:40

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.081	21:05
Air Blank		0.000	21:06
Subject Sample	# 1	0.256	21:07
Breath Volume		2.140	21:07
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.258	21:10
Breath Volume		2.476	21:10
Air Blank		0.000	21:10
Control Test		0.080	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 08/05/2020 18:14

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28818080A1
DRY GAS EXPIRY DATE: 12/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/03/2020 03:40:14

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.132	03:35
Breath Volume		3.136	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.125	03:38
Breath Volume		1.210	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 08/05/2020 18:14

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BLANCO BONILLA

DATE/TIME OF TEST: 08/04/2020 00:27:40

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.079	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.131	00:23
Breath Volume		1.257	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.137	00:26
Breath Volume		1.667	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 08/05/2020 18:14

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 08/06/2020 02:06:09

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.000	02:01
Breath Volume		2.203	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.000	02:04
Breath Volume		1.859	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 09/02/2020 18:56

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 08/06/2020 23:23:27

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.101	23:18
Breath Volume		1.750	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.099	23:21
Breath Volume		2.125	23:21
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 09/02/2020 18:56

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: SPIKOWSKI

DATE/TIME OF TEST: 08/07/2020 02:37:47

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:32	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.080	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 09/02/2020 18:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: NESTOR SIXTOS

DATE/TIME OF TEST: 08/07/2020 18:35:41

Diagnostics Check		OK	18:29
Air Blank		0.000	18:29
Control Test		0.079	18:30
Air Blank		0.000	18:30
Subject Sample	# 1	0.196	18:31
Breath Volume		2.187	18:31
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.198	18:34
Breath Volume		2.468	18:34
Air Blank		0.000	18:34
Control Test		0.079	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 09/02/2020 18:56

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ESCAMILLA TREJO

DATE/TIME OF TEST: 08/08/2020 01:55:58

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.047	01:51
Breath Volume		3.707	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.043	01:54
Breath Volume		3.304	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 09/02/2020 18:56

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 08/08/2020 20:40:52

Diagnostics Check		OK	20:34
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:35
Subject Sample	# 1	0.229	20:36
Breath Volume		2.152	20:36
Air Blank		0.000	20:37
Air Blank		0.000	20:38
Subject Sample	# 2	0.224	20:39
Breath Volume		2.699	20:39
Air Blank		0.000	20:39
Control Test		0.078	20:40
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: COS-COS

DATE/TIME OF TEST: 08/09/2020 01:11:39

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.230	01:06
Breath Volume		1.519	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.224	01:10
Breath Volume		2.250	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CARRION NAZARIO

DATE/TIME OF TEST: 08/09/2020 04:09:02

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.120	04:03
Breath Volume		1.464	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.120	04:07
Breath Volume		1.980	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:09

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 08/09/2020 23:35:05

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.080	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.186	23:30
Breath Volume		1.250	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.176	23:33
Breath Volume		1.292	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: OSIASON

DATE/TIME OF TEST: 08/10/2020 20:45:56

Diagnostics Check		OK	20:38
Air Blank		0.000	20:39
Control Test		0.080	20:39
Air Blank		0.000	20:40
Subject Sample	# 1	0.190	20:40
Breath Volume		2.410	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.185	20:44
Breath Volume		2.167	20:44
Air Blank		0.000	20:45
Control Test		0.079	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: MONSEN-SCHMIDT

DATE/TIME OF TEST: 08/11/2020 23:12:40

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.081	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.164	23:08
Breath Volume		1.367	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.173	23:11
Breath Volume		2.097	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: PEREZ FONSECA

DATE/TIME OF TEST: 08/13/2020 01:04:38

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.000	01:00
Breath Volume		1.945	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.000	01:03
Breath Volume		1.359	01:03
Air Blank		0.000	01:03
Control Test		0.081	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MEJIA RIVERA

DATE/TIME OF TEST: 08/13/2020 23:06:20

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.081	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.120	23:01
Breath Volume		1.464	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.124	23:04
Breath Volume		1.683	23:04
Air Blank		0.000	23:05
Control Test		0.080	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 08/14/2020 04:43:39

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.085	04:38
Breath Volume		1.218	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.087	04:42
Breath Volume		1.703	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ANTHONIO

DATE/TIME OF TEST: 08/15/2020 01:26:45

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.076	01:22
Breath Volume		2.218	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.076	01:25
Breath Volume		2.417	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: SALAZAR GUZMAN

DATE/TIME OF TEST: 08/15/2020 04:33:34

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.211	04:29
Breath Volume		1.652	04:29
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.206	04:32
Breath Volume		1.832	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EVERETT,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: HAGAR

DATE/TIME OF TEST: 08/20/2020 18:32:10

Diagnostics Check		OK	18:25
Air Blank		0.000	18:25
Control Test		0.080	18:26
Air Blank		0.000	18:26
Subject Sample	# 1	0.169	18:27
Breath Volume		1.179	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:29
Subject Sample	# 2	0.188	18:30
Breath Volume		2.015	18:30
Air Blank		0.000	18:31
Control Test		0.079	18:31
Air Blank		0.000	18:32
Diagnostics Check		OK	18:32

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/23/2020 02:54:47

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.147	02:50
Breath Volume		2.019	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.148	02:53
Breath Volume		2.156	02:53
Air Blank		0.000	02:53
Control Test		0.079	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ALICEA MULERO

DATE/TIME OF TEST: 08/23/2020 03:24:34

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.079	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.125	03:20
Breath Volume		1.234	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.114	03:23
Breath Volume		1.144	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: ARNOLD JR

DATE/TIME OF TEST: 08/23/2020 05:54:34

Diagnostics Check		OK	05:48
Air Blank		0.000	05:49
Control Test		0.080	05:49
Air Blank		0.000	05:50
Subject Sample	# 1	0.000	05:50
Breath Volume		3.054	05:50
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.000	05:53
Breath Volume		2.871	05:53
Air Blank		0.000	05:53
Control Test		0.080	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BRISENO ALVAREZ

DATE/TIME OF TEST: 08/24/2020 23:30:37

Diagnostics Check		OK	23:24
Air Blank		0.000	23:24
Control Test		0.080	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.182	23:26
Breath Volume		1.433	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.176	23:29
Breath Volume		1.949	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SANCHEZ GODOY

DATE/TIME OF TEST: 08/29/2020 04:12:27

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.222	04:08
Breath Volume		1.152	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.227	04:10
Breath Volume		1.703	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 08/29/2020 05:20:57

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.179	05:15
Breath Volume		1.203	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.179	05:19
Breath Volume		2.683	05:19
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/30/2020 02:53:17

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.081	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.197	02:48
Breath Volume		3.886	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.199	02:51
Breath Volume		3.621	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 09/02/2020 18:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: XAVIER,STEEVE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: XAVIER,STEEVE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 08/04/2020 18:50:24

Diagnostics Check	OK	18:49	
Air Blank	0.000	18:49	
Control Test	0.080	18:49	
Air Blank		18:50	RFI Detect
Air Blank		18:50	RFI Detect

Data Download Date/Time: 10/01/2020 03:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: XAVIER,STEEVE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: XAVIER,STEEVE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 08/04/2020 19:05:03

Diagnostics Check		OK	18:56
Air Blank		0.000	18:56
Control Test		0.080	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.220	19:00
Breath Volume		1.882	19:00
Air Blank		0.000	19:01
Air Blank		0.000	19:02
Subject Sample	# 2	0.235	19:03
Breath Volume		1.972	19:03
Air Blank		0.000	19:04
Control Test		0.076	19:04
Air Blank		0.000	19:04
Diagnostics Check		OK	19:05

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ,DANIEL
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BECERRA BALCAZAR

DATE/TIME OF TEST: 08/13/2020 01:19:06

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.185	01:14
Breath Volume		2.105	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.174	01:17
Breath Volume		1.691	01:17
Air Blank		0.000	01:18
Control Test		0.077	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROJAS,GUILLERMO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SILVERA

DATE/TIME OF TEST: 08/15/2020 01:15:25

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.086	01:10
Breath Volume		2.878	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.082	01:13
Breath Volume		2.800	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,OMAR,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROMERO,OMAR
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/15/2020 04:22:16

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:19	
Control Test		0.080	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:20	Subject Test Refused
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROJAS,ROJAS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/16/2020 00:55:06

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1	0.236	00:50	
Breath Volume		2.136	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 2	0.249	00:54	
Breath Volume		1.375	00:54	
Air Blank			00:54	RFI Detect
Air Blank		0.000	00:55	

Data Download Date/Time: 10/01/2020 03:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CANO LOPEZ

DATE/TIME OF TEST: 08/19/2020 21:48:35

Diagnostics Check		OK	21:41
Air Blank		0.000	21:42
Control Test		0.081	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.182	21:44
Breath Volume		3.535	21:44
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.182	21:47
Breath Volume		1.851	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:47
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BRAND,ANDREW
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: BAIGORRIA

DATE/TIME OF TEST: 08/24/2020 06:22:36

Diagnostics Check		OK	06:19	
Air Blank		0.000	06:20	
Control Test		0.081	06:20	
Air Blank		0.000	06:21	
Subject Sample	# 1		06:21	Subject Test Refused
Breath Volume		0.000	06:21	
Air Blank		0.000	06:21	
Control Test		0.081	06:22	
Air Blank		0.000	06:22	
Diagnostics Check		OK	06:22	

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 08/29/2020 00:57:34

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.078	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.229	00:52
Breath Volume		2.746	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.230	00:55
Breath Volume		2.890	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DENNIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SEGUI,YOSVANY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 08/29/2020 07:21:11

Diagnostics Check		OK	07:15
Air Blank		0.000	07:15
Control Test		0.081	07:15
Air Blank		0.000	07:16
Subject Sample	# 1	0.093	07:16
Breath Volume		1.437	07:16
Air Blank		0.000	07:17
Air Blank		0.000	07:19
Subject Sample	# 2	0.088	07:19
Breath Volume		1.113	07:19
Air Blank		0.000	07:20
Control Test		0.080	07:20
Air Blank		0.000	07:21
Diagnostics Check		OK	07:21

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BERNABE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ABREU MARIN

DATE/TIME OF TEST: 08/30/2020 01:24:16

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.081	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:23	
Control Test		0.081	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 10/01/2020 03:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BOWERS,JOSEPH
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A5
DRY GAS EXPIRY DATE: 09/18/2020
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/23/2020 17:30:55

Diagnostics Check		OK	17:24
Air Blank		0.000	17:25
Control Test		0.078	17:25
Air Blank		0.000	17:26
Subject Sample	# 1	0.295	17:26
Breath Volume		2.511	17:26
Air Blank		0.000	17:27
Air Blank		0.000	17:28
Subject Sample	# 2	0.291	17:29
Breath Volume		2.480	17:29
Air Blank		0.000	17:30
Control Test		0.079	17:30
Air Blank		0.000	17:30
Diagnostics Check		OK	17:30

Data Download Date/Time: 09/18/2020 11:24

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DONOVAN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A5
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:51

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 08/31/2020 12:33:56

Diagnostics Check		OK	12:27	
Air Blank		0.000	12:27	
Control Test		0.078	12:28	
Air Blank		0.000	12:28	
Subject Sample	# 1	0.249	12:29	
Breath Volume		2.281	12:29	
Air Blank		0.000	12:29	
Air Blank		0.000	12:31	
Subject Sample	# 2	0.245	12:31	
Breath Volume		2.449	12:31	
Air Blank		0.000	12:32	
Control Test		0.077	12:32	
Air Blank		0.000	12:33	
Diagnostics Check		Fail	12:33	Temperature Regulation Test
Air Blank		0.000	12:33	

Data Download Date/Time: 09/18/2020 11:24

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORMLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 11:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: GOLDSMITH

DATE/TIME OF TEST: 08/09/2020 12:01:35

Diagnostics Check		OK	11:55
Air Blank		0.000	11:56
Control Test		0.080	11:56
Air Blank		0.000	11:56
Subject Sample	# 1	0.000	11:57
Breath Volume		1.476	11:57
Air Blank		0.000	11:57
Air Blank		0.000	11:59
Subject Sample	# 2	0.000	12:00
Breath Volume		1.695	12:00
Air Blank		0.000	12:00
Control Test		0.079	12:01
Air Blank		0.000	12:01
Diagnostics Check		OK	12:01

Data Download Date/Time: 08/22/2020 10:41

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: MAPHIS

DATE/TIME OF TEST: 08/26/2020 18:27:14

Diagnostics Check		OK	18:21
Air Blank		0.000	18:21
Control Test		0.079	18:22
Air Blank		0.000	18:22
Subject Sample	# 1	0.000	18:23
Breath Volume		2.539	18:23
Air Blank		0.000	18:23
Air Blank		0.000	18:25
Subject Sample	# 2	0.000	18:25
Breath Volume		3.195	18:25
Air Blank		0.000	18:26
Control Test		0.079	18:26
Air Blank		0.000	18:27
Diagnostics Check		OK	18:27

Data Download Date/Time: 09/19/2020 03:01

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HERNANDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/18/2020 22:55:15

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.000	22:50
Breath Volume		1.597	22:50
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.000	22:53
Breath Volume		2.261	22:53
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 09/01/2020 12:08

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CUEVAS,JOSE
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 08/25/2020 23:53:45

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:51	
Control Test		0.078	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 1		23:52	RFI Detect
Breath Volume		0.000	23:52	
Air Blank		0.000	23:52	
Control Test		0.078	23:53	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 09/01/2020 10:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CUEVAS,JOSE
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 08/25/2020 23:55:58

Diagnostics Check	Fail	23:55	Voltage/Current Test
Air Blank	0.000	23:55	

Data Download Date/Time: 09/01/2020 10:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,MAREIKE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DEWIT,HARALD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 08/01/2020 01:26:34

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.162	01:22
Breath Volume		4.179	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.163	01:25
Breath Volume		4.164	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,MAREIKE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SPEAKMAN

DATE/TIME OF TEST: 08/01/2020 02:09:15

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.222	02:04
Breath Volume		3.496	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.215	02:07
Breath Volume		3.410	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PASTOR-MIGUEL

DATE/TIME OF TEST: 08/01/2020 04:59:52

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.080	04:52	
Air Blank		0.000	04:52	
Subject Sample	# 1		04:55	No Sample Provided
Breath Volume		0.000	04:55	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 2		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:58	
Control Test		0.079	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/01/2020 22:32:48

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.080	22:30	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:31	Subject Test Refused
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Control Test		0.079	22:32	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL,SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 08/02/2020 03:36:52

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.080	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DOUCET

DATE/TIME OF TEST: 08/02/2020 19:43:31

Diagnostics Check		OK	19:40	
Air Blank		0.000	19:41	
Control Test		0.079	19:41	
Air Blank		0.000	19:41	
Subject Sample	# 1		19:42	Subject Test Refused
Breath Volume		0.000	19:42	
Air Blank		0.000	19:42	
Control Test		0.080	19:42	
Air Blank		0.000	19:43	
Diagnostics Check		OK	19:43	

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAMILTON,RAY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: COCKS

DATE/TIME OF TEST: 08/05/2020 04:27:39

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.212	04:23
Breath Volume		1.496	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.221	04:26
Breath Volume		1.644	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/06/2020 07:20:36

Diagnostics Check		OK	07:14
Air Blank		0.000	07:14
Control Test		0.080	07:15
Air Blank		0.000	07:15
Subject Sample	# 1	0.129	07:16
Breath Volume		1.628	07:16
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.123	07:19
Breath Volume		1.792	07:19
Air Blank		0.000	07:19
Control Test		0.079	07:20
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 08/07/2020 00:05:45

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.081	23:58
Air Blank		0.000	23:58
Subject Sample	# 1	0.000	00:01
Breath Volume		2.671	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.000	00:04
Breath Volume		2.289	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 08/08/2020 16:36:03

Diagnostics Check		OK	16:33	
Air Blank		0.000	16:33	
Control Test		0.079	16:33	
Air Blank		0.000	16:34	
Subject Sample	# 1		16:34	Subject Test Refused
Breath Volume		0.000	16:34	
Air Blank		0.000	16:35	
Control Test		0.080	16:35	
Air Blank		0.000	16:35	
Diagnostics Check		OK	16:36	

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREMBGEN,TIA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREMBGEN,TIA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: HAGEMEIER REYES

DATE/TIME OF TEST: 08/08/2020 21:46:04

Diagnostics Check		OK	21:42	
Air Blank		0.000	21:43	
Control Test		0.080	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 1		21:44	Subject Test Refused
Breath Volume		0.000	21:44	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BURNHAM,JD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/10/2020 02:53:48

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.070	02:49
Breath Volume		3.062	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.070	02:52
Breath Volume		3.386	02:52
Air Blank		0.000	02:52
Control Test		0.080	02:53
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NABULSI,TABITHA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 08/10/2020 20:43:39

Diagnostics Check		OK	20:37
Air Blank		0.000	20:37
Control Test		0.080	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.182	20:39
Breath Volume		1.625	20:39
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.183	20:42
Breath Volume		1.363	20:42
Air Blank		0.000	20:42
Control Test		0.079	20:43
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 08/14/2020 16:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/11/2020 20:39:11

Diagnostics Check		OK	20:29
Air Blank		0.000	20:30
Control Test		0.080	20:30
Air Blank		0.000	20:30
Subject Sample	# 1	0.319	20:31
Breath Volume		1.515	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.293	20:34
Breath Volume		1.367	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 3	0.300	20:37
Breath Volume		1.687	20:37
Air Blank		0.000	20:38
Control Test		0.079	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 08/14/2020 16:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/12/2020 17:45:31

Diagnostics Check		OK	17:35
Air Blank		0.000	17:36
Control Test		0.080	17:36
Air Blank		0.000	17:37
Subject Sample	# 1	0.204	17:37
Breath Volume		2.511	17:37
Air Blank		0.000	17:38
Air Blank		0.000	17:40
Subject Sample	# 2	0.172	17:40
Breath Volume		2.164	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 3	0.197	17:43
Breath Volume		3.183	17:43
Air Blank		0.000	17:44
Control Test		0.078	17:44
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 08/14/2020 16:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,ALLISON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 08/12/2020 22:23:43

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.200	22:19
Breath Volume		2.183	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.195	22:22
Breath Volume		1.507	22:22
Air Blank		0.000	22:22
Control Test		0.078	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 08/14/2020 16:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DEWIT,HARALD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GOODSON

DATE/TIME OF TEST: 08/14/2020 21:09:52

Diagnostics Check		OK	21:07	
Air Blank		0.000	21:07	
Control Test		0.079	21:07	
Air Blank		0.000	21:08	
Subject Sample	# 1		21:08	Subject Test Refused
Breath Volume		0.000	21:08	
Air Blank		0.000	21:08	
Control Test		0.080	21:09	
Air Blank		0.000	21:09	
Diagnostics Check		OK	21:09	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:28

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 08/15/2020 10:11:10

Diagnostics Check		OK	10:04
Air Blank		0.000	10:05
Control Test		0.080	10:05
Air Blank		0.000	10:06
Subject Sample	# 1	0.000	10:06
Breath Volume		2.890	10:06
Air Blank		0.000	10:07
Air Blank		0.000	10:09
Subject Sample	# 2	0.000	10:09
Breath Volume		1.335	10:09
Air Blank		0.000	10:10
Control Test		0.080	10:10
Air Blank		0.000	10:11
Diagnostics Check		OK	10:11

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SAUNDERS,COURTNEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: MIXON

DATE/TIME OF TEST: 08/15/2020 20:31:53

Diagnostics Check		OK	20:25
Air Blank		0.000	20:25
Control Test		0.079	20:25
Air Blank		0.000	20:26
Subject Sample	# 1	0.184	20:26
Breath Volume		2.281	20:26
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.183	20:30
Breath Volume		2.605	20:30
Air Blank		0.000	20:30
Control Test		0.079	20:31
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: HOLTON

DATE/TIME OF TEST: 08/17/2020 00:10:10

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.080	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:05
Breath Volume		1.984	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.000	00:08
Breath Volume		1.394	00:08
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAIN,ERIC,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SWAIN,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: BENNISON

DATE/TIME OF TEST: 08/18/2020 21:25:22

Diagnostics Check		OK	21:19
Air Blank		0.000	21:19
Control Test		0.081	21:20
Air Blank		0.000	21:20
Subject Sample	# 1	0.000	21:21
Breath Volume		2.269	21:21
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.000	21:24
Breath Volume		2.203	21:24
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GRUNDIN,GABREAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 09:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: CARUBA

DATE/TIME OF TEST: 08/19/2020 10:32:01

Diagnostics Check		OK	10:25
Air Blank		0.000	10:25
Control Test		0.081	10:25
Air Blank		0.000	10:26
Subject Sample	# 1	0.000	10:27
Breath Volume		2.378	10:27
Air Blank		0.000	10:28
Air Blank		0.000	10:30
Subject Sample	# 2	0.000	10:30
Breath Volume		1.984	10:30
Air Blank		0.000	10:31
Control Test		0.080	10:31
Air Blank		0.000	10:31
Diagnostics Check		OK	10:32

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GUEST

DATE/TIME OF TEST: 08/19/2020 20:37:08

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.081	20:31
Air Blank		0.000	20:31
Subject Sample	# 1	0.188	20:32
Breath Volume		2.714	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.188	20:35
Breath Volume		2.484	20:35
Air Blank		0.000	20:36
Control Test		0.079	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,KIRKLAND,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PAYNE,KIRKLAND
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 08/21/2020 04:46:52

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.000	04:42
Breath Volume		2.242	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.000	04:45
Breath Volume		1.585	04:45
Air Blank		0.000	04:45
Control Test		0.081	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: ABNEY

DATE/TIME OF TEST: 08/22/2020 05:28:00

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:25	
Control Test		0.079	05:25	
Air Blank		0.000	05:26	
Subject Sample	# 1		05:26	Subject Test Refused
Breath Volume		0.000	05:26	
Air Blank		0.000	05:27	
Control Test		0.080	05:27	
Air Blank		0.000	05:27	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,A,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PLANCK,A
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 11:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: BENTON JR

DATE/TIME OF TEST: 08/22/2020 12:32:05

Diagnostics Check		OK	12:20	
Air Blank		0.000	12:20	
Control Test		0.080	12:21	
Air Blank		0.000	12:21	
Subject Sample	# 1		12:24	No Sample Provided
Breath Volume		0.000	12:24	
Air Blank		0.000	12:25	
Air Blank		0.000	12:27	
Subject Sample	# 2	0.000	12:27	
Breath Volume		1.515	12:27	
Air Blank		0.000	12:28	
Air Blank		0.000	12:30	
Subject Sample	# 3	0.000	12:30	
Breath Volume		1.152	12:30	
Air Blank		0.000	12:31	
Control Test		0.080	12:31	
Air Blank		0.000	12:32	
Diagnostics Check		OK	12:32	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LIEB,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 08/23/2020 03:56:45

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.080	03:54	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.079	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,I,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARRINGTON,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 08/24/2020 01:45:51

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.157	01:41
Breath Volume		2.949	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.139	01:44
Breath Volume		1.539	01:44
Air Blank		0.000	01:44
Control Test		0.078	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 08/24/2020 02:11:15

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.079	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.079	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/25/2020 01:09:11

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.080	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:08	
Control Test		0.080	01:08	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUHLBACH,BRIAN,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HERRMANN,J
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: THOMPSON-RAYBON

DATE/TIME OF TEST: 08/27/2020 19:20:33

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.080	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.000	19:16
Breath Volume		2.328	19:16
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.000	19:19
Breath Volume		1.164	19:19
Air Blank		0.000	19:19
Control Test		0.079	19:19
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LAPEE,WYATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 08/28/2020 06:47:09

Diagnostics Check		OK	06:39
Air Blank		0.000	06:39
Control Test		0.080	06:40
Air Blank		0.000	06:40
Subject Sample	# 1	0.000	06:42
Breath Volume		2.554	06:42
Air Blank		0.000	06:43
Air Blank		0.000	06:45
Subject Sample	# 2	0.000	06:45
Breath Volume		1.898	06:45
Air Blank		0.000	06:46
Control Test		0.080	06:46
Air Blank		0.000	06:47
Diagnostics Check		OK	06:47

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: AFFLECK

DATE/TIME OF TEST: 08/28/2020 23:55:29

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.168	23:51
Breath Volume		2.792	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.162	23:53
Breath Volume		2.714	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAURICIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MAURICIO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SNEED

DATE/TIME OF TEST: 08/29/2020 01:48:14

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:45	
Control Test		0.080	01:46	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:47	
Control Test		0.080	01:47	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: FERRARIS REYES

DATE/TIME OF TEST: 08/30/2020 02:40:39

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Control Test		0.081	02:40	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/30/2020 03:54:53

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.198	03:49	
Breath Volume		1.550	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 3	0.203	03:53	
Breath Volume		1.300	03:53	
Air Blank		0.000	03:53	
Control Test		0.079	03:54	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CRAIG,ANYCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: LEBEL

DATE/TIME OF TEST: 08/30/2020 06:06:41

Diagnostics Check		OK	06:00
Air Blank		0.000	06:00
Control Test		0.079	06:01
Air Blank		0.000	06:01
Subject Sample	# 1	0.119	06:02
Breath Volume		3.210	06:02
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.119	06:05
Breath Volume		2.546	06:05
Air Blank		0.000	06:05
Control Test		0.079	06:06
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/30/2020 15:20:26

Diagnostics Check		OK	15:14
Air Blank		0.000	15:15
Control Test		0.081	15:15
Air Blank		0.000	15:15
Subject Sample	# 1	0.000	15:16
Breath Volume		1.398	15:16
Air Blank		0.000	15:16
Air Blank		0.000	15:18
Subject Sample	# 2	0.000	15:19
Breath Volume		2.000	15:19
Air Blank		0.000	15:19
Control Test		0.080	15:19
Air Blank		0.000	15:20
Diagnostics Check		OK	15:20

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/30/2020 22:54:52

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:52	
Control Test		0.080	22:52	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Control Test		0.079	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 09/17/2020 14:28

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 08/31/2020 02:32:50

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:30	
Control Test		0.080	02:30	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:31	Subject Test Refused
Breath Volume		0.000	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 09/17/2020 14:29

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: LOPEZ LUCAS

DATE/TIME OF TEST: 08/01/2020 18:03:39

Diagnostics Check		OK	17:57	
Air Blank		0.000	17:57	
Control Test		0.079	17:58	
Air Blank		0.000	17:58	
Subject Sample	# 1	0.223	17:59	
Breath Volume		1.761	17:59	
Air Blank		0.000	18:00	
Air Blank		0.000	18:01	
Subject Sample	# 2	0.214	18:02	
Breath Volume		2.640	18:02	
Air Blank			18:03	RFI Detect
Air Blank		0.000	18:03	

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: CASTELLON CASTILLO

DATE/TIME OF TEST: 08/02/2020 16:23:15

Diagnostics Check		OK	16:16
Air Blank		0.000	16:16
Control Test		0.079	16:17
Air Blank		0.000	16:17
Subject Sample	# 1	0.357	16:18
Breath Volume		2.644	16:18
Air Blank		0.000	16:19
Air Blank		0.000	16:21
Subject Sample	# 2	0.370	16:21
Breath Volume		2.578	16:21
Air Blank		0.000	16:22
Control Test		0.077	16:22
Air Blank		0.000	16:23
Diagnostics Check		OK	16:23

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: MADDENS

DATE/TIME OF TEST: 08/05/2020 20:29:18

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.078	20:23
Air Blank		0.000	20:24
Subject Sample	# 1	0.000	20:24
Breath Volume		1.128	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.000	20:27
Breath Volume		1.742	20:27
Air Blank		0.000	20:28
Control Test		0.077	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GUDGER

DATE/TIME OF TEST: 08/06/2020 21:32:04

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.077	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.059	21:27
Breath Volume		1.617	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.060	21:30
Breath Volume		1.785	21:30
Air Blank		0.000	21:31
Control Test		0.078	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:32

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STOWE

DATE/TIME OF TEST: 08/15/2020 03:29:22

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.155	03:24
Breath Volume		2.824	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.146	03:27
Breath Volume		2.445	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: KEKESKI

DATE/TIME OF TEST: 08/15/2020 18:15:56

Diagnostics Check		OK	18:09
Air Blank		0.000	18:10
Control Test		0.078	18:10
Air Blank		0.000	18:10
Subject Sample	# 1	0.262	18:11
Breath Volume		2.671	18:11
Air Blank		0.000	18:12
Air Blank		0.000	18:13
Subject Sample	# 2	0.262	18:14
Breath Volume		2.839	18:14
Air Blank		0.000	18:15
Control Test		0.077	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 08/15/2020 23:55:19

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.078	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.091	23:50
Breath Volume		2.449	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.094	23:53
Breath Volume		2.066	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: FORTE

DATE/TIME OF TEST: 08/20/2020 22:13:32

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.078	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.394	22:09
Breath Volume		5.128	22:09
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.386	22:11
Breath Volume		4.726	22:11
Air Blank		0.000	22:12
Control Test		0.076	22:12
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER:],DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: CASSAGNOL

DATE/TIME OF TEST: 08/21/2020 01:47:16

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.078	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.161	01:42
Breath Volume		1.476	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.157	01:45
Breath Volume		1.531	01:45
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 08/28/2020 12:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PEREZ RAMREZ

DATE/TIME OF TEST: 08/07/2020 01:54:44

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.232	01:50
Breath Volume		1.429	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.248	01:53
Breath Volume		1.324	01:53
Air Blank		0.000	01:53
Control Test		0.078	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARVER,J
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GUEVARA JR

DATE/TIME OF TEST: 08/07/2020 03:45:40

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.080	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:39	No Sample Provided
Breath Volume		0.000	03:39	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.081	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/07/2020 21:20:54

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.081	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.344	21:16
Breath Volume		2.851	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:18
Subject Sample	# 2	0.339	21:19
Breath Volume		2.160	21:19
Air Blank		0.000	21:19
Control Test		0.080	21:20
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/08/2020 01:47:21

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.081	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1	0.233	01:43	
Breath Volume		1.503	01:43	
Air Blank		0.000	01:44	
Air Blank		0.000	01:45	
			01:45	RFI Detect
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/08/2020 01:58:46

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.234	01:52
Breath Volume		1.417	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.223	01:57
Breath Volume		1.570	01:57
Air Blank		0.000	01:57
Control Test		0.080	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: SOBERON

DATE/TIME OF TEST: 08/08/2020 05:55:53

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.080	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.146	05:51
Breath Volume		2.781	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:53
Subject Sample	# 2	0.152	05:54
Breath Volume		2.996	05:54
Air Blank		0.000	05:54
Control Test		0.081	05:55
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 08/08/2020 23:29:20

Diagnostics Check		OK	23:26	
Air Blank		0.000	23:26	
Control Test		0.081	23:27	
Air Blank		0.000	23:27	
Subject Sample	# 1		23:27	Subject Test Refused
Breath Volume		0.000	23:27	
Air Blank		0.000	23:28	
Control Test		0.081	23:28	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: COVELLI

DATE/TIME OF TEST: 08/09/2020 00:32:22

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.081	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.085	00:28
Breath Volume		2.488	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.078	00:30
Breath Volume		2.167	00:30
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: FYE

DATE/TIME OF TEST: 08/09/2020 01:19:26

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.081	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1	0.239	01:12	Volume Not Met
Breath Volume		0.277	01:12	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 2	0.250	01:17	Volume Not Met
Breath Volume		0.191	01:17	
Air Blank		0.000	01:18	
Control Test		0.080	01:18	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: FYE

DATE/TIME OF TEST: 08/09/2020 01:29:51

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:25	No Sample Provided
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 2		01:28	Subject Test Refused
Breath Volume		0.000	01:28	
Air Blank		0.000	01:28	
Control Test		0.080	01:29	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/09/2020 02:46:48

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.054	02:42
Breath Volume		2.191	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.055	02:45
Breath Volume		2.277	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAMOS,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/09/2020 18:53:33

Diagnostics Check		OK	18:39	
Air Blank		0.000	18:40	
Control Test		0.081	18:40	
Air Blank		0.000	18:41	
Subject Sample	# 1	0.192	18:44	Volume Not Met
Breath Volume		0.945	18:44	
Air Blank		0.000	18:45	
Air Blank		0.000	18:46	
Subject Sample	# 2	0.201	18:47	
Breath Volume		1.949	18:47	
Air Blank		0.000	18:48	
Air Blank		0.000	18:50	
Subject Sample	# 3	0.190	18:51	
Breath Volume		1.222	18:51	
Air Blank		0.000	18:52	
Control Test		0.080	18:52	
Air Blank		0.000	18:53	
Diagnostics Check		OK	18:53	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 08/10/2020 02:30:52

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.131	02:26
Breath Volume		1.648	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.131	02:29
Breath Volume		1.800	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CROCKER

DATE/TIME OF TEST: 08/13/2020 02:39:22

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.055	02:33
Breath Volume		2.289	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.055	02:37
Breath Volume		1.472	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: TORRES MADERA

DATE/TIME OF TEST: 08/14/2020 03:45:14

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.081	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.132	03:37	Volume Not Met
Breath Volume		0.152	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.113	03:43	Volume Not Met
Breath Volume		0.367	03:43	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: TORRES MADERA

DATE/TIME OF TEST: 08/14/2020 04:02:38

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.141	03:49	
Breath Volume		1.351	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 2	0.112	03:55	Volume Not Met
Breath Volume		0.371	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 3	0.142	04:00	Slope Not Level
Breath Volume		1.316	04:00	
Air Blank		0.000	04:01	
Control Test		0.080	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: TORRES MADERA

DATE/TIME OF TEST: 08/14/2020 04:13:43

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.144	04:08
Breath Volume		1.289	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.155	04:12
Breath Volume		1.824	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/15/2020 02:56:47

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.072	02:52
Breath Volume		3.097	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:54
Subject Sample	# 2	0.071	02:55
Breath Volume		2.414	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 08/16/2020 04:42:35

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.146	04:37
Breath Volume		3.226	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.147	04:41
Breath Volume		3.335	04:41
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/17/2020 01:49:26

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.043	01:45
Breath Volume		2.996	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.036	01:47
Breath Volume		1.652	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 09/02/2020 18:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WHITFIELD JR

DATE/TIME OF TEST: 08/18/2020 03:17:14

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.171	03:12
Breath Volume		3.031	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.164	03:15
Breath Volume		2.636	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DELORENZO,D
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/19/2020 03:36:34

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.064	03:32
Breath Volume		1.929	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.064	03:35
Breath Volume		2.542	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SEREBRYANSKIY

DATE/TIME OF TEST: 08/20/2020 03:58:25

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.081	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.079	03:53
Breath Volume		2.289	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.084	03:56
Breath Volume		2.953	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SCOTT JR

DATE/TIME OF TEST: 08/22/2020 02:53:59

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.076	02:49
Breath Volume		2.109	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.072	02:52
Breath Volume		1.707	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARLOCK,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:47

SUBJECT'S LAST NAME: GUERRA VERA

DATE/TIME OF TEST: 08/22/2020 09:21:35

Diagnostics Check		OK	09:07	
Air Blank		0.000	09:08	
Control Test		0.079	09:08	
Air Blank		0.000	09:09	
Subject Sample	# 1	0.138	09:10	
Breath Volume		2.105	09:10	
Air Blank		0.000	09:10	
Air Blank		0.000	09:12	
Subject Sample	# 2	0.106	09:15	Volume Not Met
Breath Volume		0.093	09:15	
Air Blank		0.000	09:16	
Air Blank		0.000	09:18	
Subject Sample	# 3	0.118	09:20	
Breath Volume		1.285	09:20	
Air Blank		0.000	09:20	
Control Test		0.079	09:20	
Air Blank		0.000	09:21	
Diagnostics Check		OK	09:21	

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 08/23/2020 03:31:04

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.050	03:26
Breath Volume		1.839	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.046	03:29
Breath Volume		1.460	03:29
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:31

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 08/23/2020 04:57:36

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.077	04:53
Breath Volume		1.867	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.078	04:56
Breath Volume		1.539	04:56
Air Blank		0.000	04:56
Control Test		0.080	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BARTOLON CASTILLO

DATE/TIME OF TEST: 08/29/2020 00:11:36

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.080	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.113	00:07
Breath Volume		2.300	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.113	00:10
Breath Volume		2.367	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: REDDING JR

DATE/TIME OF TEST: 08/29/2020 01:54:58

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.144	01:50
Breath Volume		3.234	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.143	01:53
Breath Volume		3.433	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BLYDEN

DATE/TIME OF TEST: 08/29/2020 03:47:24

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.118	03:42
Breath Volume		3.394	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.118	03:45
Breath Volume		3.625	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/29/2020 04:32:11

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.121	04:27
Breath Volume		2.316	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.117	04:30
Breath Volume		2.546	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARRIS,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 08/30/2020 00:55:56

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.081	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.182	00:51
Breath Volume		2.929	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.169	00:54
Breath Volume		2.425	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURROUGHS,H
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 08/30/2020 03:16:12

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.080	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 1	0.039	03:08	Volume Not Met
Breath Volume		0.406	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2	0.047	03:12	
Breath Volume		1.519	03:12	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
			03:14	Improper Sample
Air Blank		0.000	03:15	
Control Test		0.080	03:15	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURROUGHS,H
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 08/30/2020 03:25:06

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.048	03:20
Breath Volume		2.781	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.047	03:23
Breath Volume		1.511	03:23
Air Blank		0.000	03:24
Control Test		0.081	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: GONZALEZ AGUIRRE

DATE/TIME OF TEST: 08/31/2020 05:41:33

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.081	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.122	05:37
Breath Volume		2.964	05:37
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.128	05:40
Breath Volume		2.937	05:40
Air Blank		0.000	05:40
Control Test		0.081	05:40
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 09/02/2020 18:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 08/21/2020 23:58:45

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.078	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.144	23:54
Breath Volume		2.292	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.145	23:57
Breath Volume		1.703	23:57
Air Blank		0.000	23:57
Control Test		0.077	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 09/02/2020 20:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 08/22/2020 01:20:44

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.078	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.143	01:16
Breath Volume		1.480	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.134	01:19
Breath Volume		1.921	01:19
Air Blank		0.000	01:19
Control Test		0.077	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 09/02/2020 20:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 08/22/2020 02:10:34

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.077	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.000	02:06
Breath Volume		1.230	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.000	02:09
Breath Volume		1.308	02:09
Air Blank		0.000	02:09
Control Test		0.077	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 09/02/2020 20:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ADAMES

DATE/TIME OF TEST: 08/22/2020 02:49:51

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:47	
Control Test		0.078	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.077	02:49	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 09/02/2020 20:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 08/22/2020 23:34:33

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:32	
Control Test		0.078	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:33	
Control Test		0.077	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 09/02/2020 20:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUM,ZACKARY,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: PRESSEY

DATE/TIME OF TEST: 08/02/2020 00:04:24

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:02	
Control Test		0.080	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	
Control Test		0.080	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 08/06/2020 08:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWEIGER,JONATHAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PARRILL

DATE/TIME OF TEST: 08/02/2020 01:54:05

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.079	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:52	Subject Test Refused
Breath Volume		0.000	01:52	
Air Blank		0.000	01:53	
Control Test		0.080	01:53	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 08/06/2020 08:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: DELUCA

DATE/TIME OF TEST: 08/03/2020 18:33:29

Diagnostics Check		OK	18:27
Air Blank		0.000	18:27
Control Test		0.079	18:28
Air Blank		0.000	18:28
Subject Sample	# 1	0.296	18:28
Breath Volume		2.566	18:28
Air Blank		0.000	18:29
Air Blank		0.000	18:31
Subject Sample	# 2	0.287	18:31
Breath Volume		2.468	18:31
Air Blank		0.000	18:32
Control Test		0.078	18:32
Air Blank		0.000	18:33
Diagnostics Check		OK	18:33

Data Download Date/Time: 08/06/2020 08:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFFERTY,NAOMI,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A3
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BAE

DATE/TIME OF TEST: 08/06/2020 01:29:43

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.168	01:25
Breath Volume		1.460	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.172	01:28
Breath Volume		1.710	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 08/06/2020 08:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,ZAKERIE,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: TUCKER,MARVILLE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 08/03/2020 05:26:25

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.190	05:21
Breath Volume		2.527	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.192	05:24
Breath Volume		1.390	05:24
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RETTE,CODY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2021
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DARVISH

DATE/TIME OF TEST: 08/09/2020 00:51:46

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.220	00:47
Breath Volume		2.246	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.219	00:50
Breath Volume		1.199	00:50
Air Blank		0.000	00:50
Control Test		0.079	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAVOTTA,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MORRIS,LANCE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 08/09/2020 22:39:38

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.269	22:35
Breath Volume		2.753	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.273	22:38
Breath Volume		3.097	22:38
Air Blank		0.000	22:38
Control Test		0.079	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,ZAKERIE,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALEVISATOS,JARRAD
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: MCKEEHAN

DATE/TIME OF TEST: 08/10/2020 22:33:14

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.099	22:28
Breath Volume		1.582	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.097	22:31
Breath Volume		2.640	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,ZAKERIE,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RAYNOR,JASON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 08/12/2020 02:31:01

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank		02:30	RFI Detect
Air Blank	0.000	02:31	

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,ZAKERIE,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RAYNOR,JASON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 08/12/2020 02:37:36

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:34	
Control Test		0.080	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.081	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 08/15/2020 02:13:41

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.081	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.151	02:09
Breath Volume		2.714	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.150	02:12
Breath Volume		2.773	02:12
Air Blank		0.000	02:12
Control Test		0.080	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 08/24/2020 03:51:57

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1	0.147	03:44	Volume Not Met
Breath Volume		0.542	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.148	03:50	Volume Not Met
Breath Volume		0.062	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 08/24/2020 04:00:50

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.159	03:55
Breath Volume		1.515	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.168	03:59
Breath Volume		2.507	03:59
Air Blank		0.000	03:59
Control Test		0.079	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: B525472013450

DATE/TIME OF TEST: 08/25/2020 00:00:51

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.000	23:56
Breath Volume		2.238	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.000	23:59
Breath Volume		2.691	23:59
Air Blank		0.000	23:59
Control Test		0.081	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,ZAKERIE,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: TURCHON,ANTHONY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GRODI

DATE/TIME OF TEST: 08/25/2020 00:34:00

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.208	00:28
Breath Volume		1.667	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.197	00:32
Breath Volume		1.390	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:34

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 08/29/2020 22:45:06

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.181	22:39
Breath Volume		1.140	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.165	22:43
Breath Volume		1.265	22:43
Air Blank		0.000	22:44
Control Test		0.078	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 08/31/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/01/2020 03:44:02

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.124	03:33	Volume Not Met
Breath Volume		0.582	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2	0.147	03:37	
Breath Volume		1.445	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 3	0.144	03:42	
Breath Volume		1.304	03:42	
Air Blank		0.000	03:43	
Control Test		0.078	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 08/13/2020 11:22

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VINCENT

DATE/TIME OF TEST: 08/03/2020 03:02:06

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:01	
Control Test		0.080	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 08/13/2020 11:22

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,J,
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: WEIGLE,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: MURDAUGH

DATE/TIME OF TEST: 08/16/2020 06:36:06

Diagnostics Check		OK	06:33	
Air Blank		0.000	06:33	
Control Test		0.081	06:33	
Air Blank		0.000	06:34	
Subject Sample	# 1		06:34	Subject Test Refused
Breath Volume		0.000	06:34	
Air Blank		0.000	06:35	
Control Test		0.081	06:35	
Air Blank		0.000	06:36	
Diagnostics Check		OK	06:36	

Data Download Date/Time: 09/09/2020 16:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: NORDLINGER

DATE/TIME OF TEST: 08/28/2020 22:44:14

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.081	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.101	22:39
Breath Volume		1.671	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:41
Subject Sample	# 2	0.094	22:42
Breath Volume		1.464	22:42
Air Blank		0.000	22:43
Control Test		0.079	22:43
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 09/09/2020 16:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: REHM

DATE/TIME OF TEST: 08/30/2020 03:20:32

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.081	03:18	
Air Blank		0.000	03:18	
Subject Sample	# 1		03:19	Subject Test Refused
Breath Volume		0.000	03:19	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 09/09/2020 16:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: REVERON

DATE/TIME OF TEST: 08/12/2020 00:34:40

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.184	00:30
Breath Volume		1.582	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.190	00:33
Breath Volume		1.472	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 08/19/2020 10:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ARIZA

DATE/TIME OF TEST: 08/14/2020 00:54:04

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.265	00:49
Breath Volume		1.539	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.252	00:52
Breath Volume		2.515	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:54

Data Download Date/Time: 08/19/2020 10:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: VILLEDA-GUERRERO

DATE/TIME OF TEST: 08/14/2020 01:57:12

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.260	01:51
Breath Volume		1.847	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.263	01:55
Breath Volume		1.851	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 08/19/2020 10:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 08/14/2020 21:27:06

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:24	
Control Test		0.080	21:25	
Air Blank		0.000	21:25	
Subject Sample	# 1		21:25	Subject Test Refused
Breath Volume		0.000	21:25	
Air Blank		0.000	21:26	
Control Test		0.078	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:27	

Data Download Date/Time: 08/19/2020 10:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LOJO

DATE/TIME OF TEST: 08/15/2020 01:04:19

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.151	00:59
Breath Volume		2.023	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.149	01:02
Breath Volume		1.640	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 08/19/2020 10:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CLAVERO-RODRIGUEZ

DATE/TIME OF TEST: 08/22/2020 02:47:00

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.081	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.139	02:41
Breath Volume		1.441	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.139	02:45
Breath Volume		1.511	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:47

Data Download Date/Time: 09/02/2020 13:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 08/22/2020 22:49:18

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.080	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.070	22:44
Breath Volume		1.742	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.073	22:47
Breath Volume		2.402	22:47
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 09/02/2020 13:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: DE LIMA

DATE/TIME OF TEST: 08/29/2020 01:57:34

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.080	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 09/02/2020 13:30

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:14

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 08/04/2020 10:41:15

Diagnostics Check		OK	10:35
Air Blank		0.000	10:35
Control Test		0.080	10:35
Air Blank		0.000	10:36
Subject Sample	# 1	0.000	10:36
Breath Volume		2.539	10:36
Air Blank		0.000	10:37
Air Blank		0.000	10:39
Subject Sample	# 2	0.000	10:39
Breath Volume		1.507	10:39
Air Blank		0.000	10:40
Control Test		0.079	10:40
Air Blank		0.000	10:41
Diagnostics Check		OK	10:41

Data Download Date/Time: 08/24/2020 10:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ROOP

DATE/TIME OF TEST: 08/08/2020 01:26:46

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.184	01:22
Breath Volume		2.671	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.182	01:25
Breath Volume		2.234	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 08/24/2020 10:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 08/08/2020 19:46:05

Diagnostics Check		OK	19:39
Air Blank		0.000	19:40
Control Test		0.080	19:40
Air Blank		0.000	19:40
Subject Sample	# 1	0.284	19:41
Breath Volume		2.175	19:41
Air Blank		0.000	19:42
Air Blank		0.000	19:43
Subject Sample	# 2	0.288	19:44
Breath Volume		1.250	19:44
Air Blank		0.000	19:45
Control Test		0.078	19:45
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 08/24/2020 10:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/24/2020 19:10:44

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.080	19:05
Air Blank		0.000	19:05
Subject Sample	# 1	0.313	19:06
Breath Volume		2.367	19:06
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.327	19:09
Breath Volume		3.656	19:09
Air Blank		0.000	19:09
Control Test		0.076	19:10
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 09/21/2020 10:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: PUTZ

DATE/TIME OF TEST: 08/29/2020 17:49:47

Diagnostics Check		OK	17:42
Air Blank		0.000	17:42
Control Test		0.079	17:43
Air Blank		0.000	17:43
Subject Sample	# 1	0.252	17:44
Breath Volume		2.062	17:44
Air Blank		0.000	17:45
Air Blank		0.000	17:47
Subject Sample	# 2	0.243	17:48
Breath Volume		1.902	17:48
Air Blank		0.000	17:48
Control Test		0.078	17:49
Air Blank		0.000	17:49
Diagnostics Check		OK	17:49

Data Download Date/Time: 09/21/2020 10:04

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 191483
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: RIVERA REYES

DATE/TIME OF TEST: 08/02/2020 22:58:42

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.078	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.176	22:53
Breath Volume		2.453	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.177	22:57
Breath Volume		2.332	22:57
Air Blank		0.000	22:57
Control Test		0.078	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 08/03/2020 16:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MCNUTT

DATE/TIME OF TEST: 08/09/2020 01:05:48

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:01
Breath Volume		2.996	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.000	01:04
Breath Volume		3.308	01:04
Air Blank		0.000	01:04
Control Test		0.079	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 08/28/2020 13:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: KEESEY

DATE/TIME OF TEST: 08/12/2020 20:49:57

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.078	20:41
Air Blank		0.000	20:42
Subject Sample	# 1	0.255	20:45
Breath Volume		1.550	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:47
Subject Sample	# 2	0.253	20:48
Breath Volume		1.449	20:48
Air Blank		0.000	20:49
Control Test		0.075	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 08/28/2020 13:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BOULERICE

DATE/TIME OF TEST: 08/08/2020 18:23:15

Diagnostics Check		OK	18:16
Air Blank		0.000	18:17
Control Test		0.078	18:17
Air Blank		0.000	18:18
Subject Sample	# 1	0.103	18:18
Breath Volume		3.074	18:18
Air Blank		0.000	18:19
Air Blank		0.000	18:21
Subject Sample	# 2	0.100	18:21
Breath Volume		1.992	18:21
Air Blank		0.000	18:22
Control Test		0.081	18:22
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 08/28/2020 17:39

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: PRESSER,JAMES
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 08/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 08/15/2020 20:49:14

Diagnostics Check		OK	20:46	
Air Blank		0.000	20:46	
Control Test		0.079	20:47	
Air Blank		0.000	20:47	
Subject Sample	# 1		20:47	Subject Test Refused
Breath Volume		0.000	20:47	
Air Blank		0.000	20:48	
Control Test		0.078	20:48	
Air Blank		0.000	20:49	
Diagnostics Check		OK	20:49	

Data Download Date/Time: 09/02/2020 14:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 08/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: MALPICA

DATE/TIME OF TEST: 08/16/2020 20:57:32

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:54	
Control Test		0.080	20:54	
Air Blank		0.000	20:55	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:56	
Control Test		0.079	20:56	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 09/02/2020 14:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: SAMPLE,CARSON
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 08/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: DORCILUS

DATE/TIME OF TEST: 08/16/2020 21:20:36

Diagnostics Check		OK	21:17	
Air Blank		0.000	21:17	
Control Test		0.079	21:18	
Air Blank		0.000	21:18	
Subject Sample	# 1		21:19	Subject Test Refused
Breath Volume		0.000	21:19	
Air Blank		0.000	21:19	
Control Test		0.079	21:19	
Air Blank		0.000	21:20	
Diagnostics Check		OK	21:20	

Data Download Date/Time: 09/02/2020 14:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: PRESSER,JAMES
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 08/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ESTRELLA

DATE/TIME OF TEST: 08/29/2020 20:49:36

Diagnostics Check		OK	20:43	
Air Blank		0.000	20:43	
Control Test		0.076	20:44	
Air Blank		0.000	20:44	
Subject Sample	# 1	0.179	20:45	
Breath Volume		3.234	20:45	
Air Blank		0.000	20:45	
Air Blank		0.000	20:47	
Subject Sample	# 2	0.179	20:48	
Breath Volume		1.859	20:48	
Air Blank		0.000	20:48	
Control Test		0.074	20:48	Control Outside Tolerance
Air Blank		0.000	20:49	

Data Download Date/Time: 09/02/2020 14:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: TZIMENATOS

DATE/TIME OF TEST: 08/16/2020 03:08:56

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.191	03:04
Breath Volume		2.429	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.192	03:07
Breath Volume		2.285	03:07
Air Blank		0.000	03:08
Control Test		0.077	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 09/16/2020 22:18

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 08/28/2020 22:31:01

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.075	22:25
Air Blank		0.000	22:25
Subject Sample	# 1	0.000	22:26
Breath Volume		1.730	22:26
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.000	22:29
Breath Volume		3.527	22:29
Air Blank		0.000	22:30
Control Test		0.081	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:31

Data Download Date/Time: 09/16/2020 22:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/30/2020 05:28:46

Diagnostics Check		OK	05:17	
Air Blank		0.000	05:17	
Control Test		0.080	05:18	
Air Blank		0.000	05:18	
Subject Sample	# 1	0.144	05:21	Volume Not Met
Breath Volume		0.152	05:21	
Air Blank		0.000	05:22	
Air Blank		0.000	05:24	
Subject Sample	# 2		05:27	No Sample Provided
Breath Volume		0.000	05:27	
Air Blank		0.000	05:27	
Control Test		0.077	05:28	
Air Blank		0.000	05:28	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 09/16/2020 22:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.