Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEVIN,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 08/03/2020 18:32:37

Diagnostics Check		OK	18:21	
Air Blank		0.000	18:21	
Control Test		0.076	18:21	
Air Blank		0.000	18:22	
Subject Sample	# 1	0.080	18:25	Volume Not Met
Breath Volume		0.335	18:25	
Air Blank		0.000	18:26	
Air Blank		0.000	18:27	
Subject Sample	# 2	0.097	18:31	Volume Not Met
Breath Volume		0.253	18:31	
Air Blank		0.000	18:31	
Control Test		0.079	18:32	
Air Blank		0.000	18:32	
Diagnostics Check		OK	18:32	

Data Download Date/Time: 08/04/2020 14:26

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DEVIN,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13518080A5
DRY GAS EXPIRY DATE: 08/05/2020
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 08/03/2020 18:42:16

Diagnostics Check		OK	18:35
Air Blank		0.000	18:36
Control Test		0.075	18:36
Air Blank		0.000	18:36
Subject Sample	# 1	0.103	18:37
Breath Volume		2.363	18:37
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.094	18:40
Breath Volume		1.765	18:40
Air Blank		0.000	18:41
Control Test		0.076	18:41
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 08/04/2020 14:26

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAZ,RICARDO,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: WANZENBERG,ALEX
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/03/2020 00:57:54

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:57	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: JONES,LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/04/2020 19:38:33

Diagnostics Check		OK	19:32
Air Blank		0.000	19:32
Control Test		0.079	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.114	19:34
Breath Volume		1.257	19:34
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.115	19:37
Breath Volume		1.589	19:37
Air Blank		0.000	19:37
Control Test		0.078	19:37
Air Blank		0.000	19:38
Diagnostics Check		OK	19:38

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ACOSTA,ALEX
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 08/04/2020 20:49:17

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.080	20:42
Air Blank		0.000	20:43
Subject Sample	# 1	0.180	20:44
Breath Volume		1.800	20:44
Air Blank		0.000	20:45
Air Blank		0.000	20:46
Subject Sample	# 2	0.178	20:47
Breath Volume		2.441	20:47
Air Blank		0.000	20:48
Control Test		0.077	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: RATCLIFF

DATE/TIME OF TEST: 08/05/2020 20:36:13

Diagnostics Check		OK	20:33	
Air Blank		0.000	20:33	
Control Test		0.079	20:34	
Air Blank		0.000	20:34	
Subject Sample	# 1		20:34	Subject Test Refused
Breath Volume		0.000	20:34	
Air Blank		0.000	20:35	
Control Test		0.079	20:35	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:36	

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALVAREZ,ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 08/08/2020 02:29:48

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.080	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILSON,MALCOLM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/16/2020 02:54:50

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.211	02:50
Breath Volume		1.730	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:52
Subject Sample	# 2	0.204	02:53
Breath Volume		2.335	02:53
Air Blank		0.000	02:53
Control Test		0.078	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: CHRISTIE,LUCAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 08/17/2020 00:51:25

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.101	00:47
Breath Volume		1.953	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.101	00:49
Breath Volume		2.316	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCREE,GARRETT
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: LOPEZ-PEREZ

DATE/TIME OF TEST: 08/18/2020 00:45:36

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.225	00:40
Breath Volume		1.976	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.230	00:43
Breath Volume		2.667	00:43
Air Blank		0.000	00:44
Control Test		0.077	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTSCHER,REBECCA,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PEKALA,TONY
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 08/19/2020 21:14:23

Diagnostics Check		OK	21:07
Air Blank		0.000	21:08
Control Test		0.079	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.140	21:09
Breath Volume		1.902	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.129	21:12
Breath Volume		1.484	21:12
Air Blank		0.000	21:13
Control Test		0.079	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,COURTNEY,M
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: MEYER,SAMUEL
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: LITTLES

DATE/TIME OF TEST: 08/20/2020 04:42:21

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.000	04:38
Breath Volume		3.261	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.000	04:40
Breath Volume		3.066	04:40
Air Blank		0.000	04:41
Control Test		0.077	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHUTTERLY,BROOKE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: WAGSTAFF

DATE/TIME OF TEST: 08/24/2020 00:41:08

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.079	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1	0.331	00:28	
Breath Volume		3.128	00:28	
Air Blank		0.000	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 2		00:34	No Sample Provided
Breath Volume		0.000	00:34	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 3		00:39	No Sample Provided
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHUTTERLY,BROOKE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: WAGSTAFF

DATE/TIME OF TEST: 08/24/2020 00:50:34

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.327	00:45
Breath Volume		3.816	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.326	00:49
Breath Volume		2.648	00:49
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 08/26/2020 14:33

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: WITTKE

DATE/TIME OF TEST: 08/29/2020 23:22:44

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.080	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.152	23:18
Breath Volume		1.906	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.148	23:21
Breath Volume		1.781	23:21
Air Blank		0.000	23:21
Control Test		0.079	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 09/23/2020 13:03

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: FEAGIN

DATE/TIME OF TEST: 08/02/2020 10:51:10

Diagnostics Check		OK	10:47	
Air Blank		0.000	10:48	
Control Test		0.081	10:48	
Air Blank		0.000	10:48	
Subject Sample	# 1		10:49	Subject Test Refused
Breath Volume		0.000	10:49	
Air Blank		0.000	10:50	
Control Test		0.080	10:50	
Air Blank		0.000	10:51	
Diagnostics Check		OK	10:51	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/08/2020 09:49:42

Diagnostics Check		OK	09:43
Air Blank		0.000	09:43
Control Test		0.080	09:44
Air Blank		0.000	09:44
Subject Sample	# 1	0.000	09:45
Breath Volume		2.792	09:45
Air Blank		0.000	09:45
Air Blank		0.000	09:47
Subject Sample	# 2	0.000	09:48
Breath Volume		1.347	09:48
Air Blank		0.000	09:48
Control Test		0.083	09:49
Air Blank		0.000	09:49
Diagnostics Check		OK	09:49

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,GABRIEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: DUNLAP,GABRIEL
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 08/08/2020 19:02:55

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:00	
Control Test		0.079	19:00	
Air Blank		0.000	19:01	
Subject Sample	# 1		19:01	Subject Test Refused
Breath Volume		0.000	19:01	
Air Blank		0.000	19:02	
Control Test		0.080	19:02	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,KEITH,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/08/2020 19:27:54

Diagnostics Check		OK	19:21
Air Blank		0.000	19:21
Control Test		0.079	19:21
Air Blank		0.000	19:22
Subject Sample	# 1	0.000	19:23
Breath Volume		4.179	19:23
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.000	19:26
Breath Volume		4.175	19:26
Air Blank		0.000	19:26
Control Test		0.083	19:27
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,JODY,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOWARD,JODY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 08/09/2020 02:52:49

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.077	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.214	02:48
Breath Volume		2.976	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.222	02:51
Breath Volume		2.714	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STALEY,KRIS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 08/11/2020 19:49:49

Diagnostics Check		OK	19:43
Air Blank		0.000	19:44
Control Test		0.080	19:44
Air Blank		0.000	19:44
Subject Sample	# 1	0.148	19:45
Breath Volume		2.410	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:47
Subject Sample	# 2	0.149	19:48
Breath Volume		1.554	19:48
Air Blank		0.000	19:48
Control Test		0.079	19:49
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,JON,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: GORDON,JON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 08/14/2020 01:51:15

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.181	01:46
Breath Volume		1.722	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.177	01:49
Breath Volume		1.503	01:49
Air Blank		0.000	01:50
Control Test		0.076	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: TELLES

DATE/TIME OF TEST: 08/15/2020 22:47:23

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.081	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.000	22:43
Breath Volume		4.414	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.000	22:46
Breath Volume		3.996	22:46
Air Blank		0.000	22:46
Control Test		0.084	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 08/17/2020 02:31:26

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.081	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Sequence Aborted
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 08/17/2020 02:57:11

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.101	02:52
Breath Volume		2.613	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.100	02:55
Breath Volume		2.195	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/17/2020 03:28:52

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:27	RFI Detect
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.080	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOBNOCK,PAULA,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BOBNOCK

DATE/TIME OF TEST: 08/18/2020 21:30:26

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.081	21:24
Air Blank		0.000	21:24
Subject Sample	# 1	0.049	21:25
Breath Volume		1.589	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.046	21:28
Breath Volume		1.726	21:28
Air Blank		0.000	21:29
Control Test		0.080	21:29
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENTINE,RONALD,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VALENTINE,RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 08/22/2020 00:16:05

Diagnostics Check	OK	00:15	
Air Blank		00:15	Ambient Fail
Air Blank	0.000	00:16	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENTINE,RONALD,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VALENTINE,RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SHERFFIELD

DATE/TIME OF TEST: 08/22/2020 00:26:56

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.220	00:22
Breath Volume		1.917	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.226	00:25
Breath Volume		1.949	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/22/2020 03:05:03

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.059	03:00
Breath Volume		1.351	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.068	03:03
Breath Volume		2.671	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:05

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 08/22/2020 16:23:02

Diagnostics Check		OK	16:16
Air Blank		0.000	16:17
Control Test		0.079	16:17
Air Blank		0.000	16:18
Subject Sample	# 1	0.160	16:18
Breath Volume		3.070	16:18
Air Blank		0.000	16:19
Air Blank		0.000	16:21
Subject Sample	# 2	0.163	16:21
Breath Volume		2.453	16:21
Air Blank		0.000	16:22
Control Test		0.079	16:22
Air Blank		0.000	16:22
Diagnostics Check		OK	16:23

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 08/22/2020 21:20:21

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:17	
Control Test		0.081	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1		21:18	Subject Test Refused
Breath Volume		0.000	21:18	
Air Blank		0.000	21:19	
Control Test		0.081	21:19	
Air Blank		0.000	21:20	
Diagnostics Check		OK	21:20	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 08/23/2020 04:06:03

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:05	
Control Test		0.081	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MIJARES

DATE/TIME OF TEST: 08/25/2020 02:27:07

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.000	02:22
Breath Volume		3.500	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.000	02:25
Breath Volume		4.066	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: STROPPEL

DATE/TIME OF TEST: 08/27/2020 22:43:03

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.081	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.000	22:38
Breath Volume		2.390	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.000	22:41
Breath Volume		1.933	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:43

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 08/29/2020 17:28:44

Diagnostics Check		OK	17:26	
Air Blank		0.000	17:26	
Control Test		0.081	17:27	
Air Blank		0.000	17:27	
Subject Sample	# 1		17:27	Subject Test Refused
Breath Volume		0.000	17:27	
Air Blank			17:28	RFI Detect
Air Blank		0.000	17:28	

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MOLLOY

DATE/TIME OF TEST: 08/30/2020 23:02:16

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.080	22:55
Air Blank		0.000	22:55
Subject Sample	# 1	0.203	22:57
Breath Volume		2.066	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.200	23:00
Breath Volume		2.105	23:00
Air Blank		0.000	23:01
Control Test		0.078	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 08/31/2020 12:31

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,ALLISON,T
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: LECUYER,JASON
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 07/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: JOYNER

DATE/TIME OF TEST: 08/15/2020 00:10:31

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.078	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.218	00:06
Breath Volume		2.589	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.212	00:09
Breath Volume		2.394	00:09
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 08/25/2020 00:35

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: GRINSTEINNER

DATE/TIME OF TEST: 08/01/2020 17:04:51

Diagnostics Check		OK	16:58
Air Blank		0.000	16:58
Control Test		0.081	16:59
Air Blank		0.000	16:59
Subject Sample	# 1	0.074	17:00
Breath Volume		1.601	17:00
Air Blank		0.000	17:00
Air Blank		0.000	17:02
Subject Sample	# 2	0.076	17:03
Breath Volume		1.960	17:03
Air Blank		0.000	17:03
Control Test		0.081	17:04
Air Blank		0.000	17:04
Diagnostics Check		OK	17:04

Data Download Date/Time: 08/11/2020 14:11

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: ROCKETTE

DATE/TIME OF TEST: 08/01/2020 18:04:53

Diagnostics Check		OK	18:00	
Air Blank		0.000	18:01	
Control Test		0.080	18:01	
Air Blank		0.000	18:02	
Subject Sample	# 1		18:03	Subject Test Refused
Breath Volume		0.500	18:03	
Air Blank		0.000	18:03	
Control Test		0.081	18:04	
Air Blank		0.000	18:04	
Diagnostics Check		OK	18:04	

Data Download Date/Time: 08/11/2020 14:11

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: TOBIASSEN

DATE/TIME OF TEST: 08/02/2020 21:44:08

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.081	21:38
Air Blank		0.000	21:38
Subject Sample	# 1	0.092	21:39
Breath Volume		4.613	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:41
Subject Sample	# 2	0.086	21:42
Breath Volume		4.226	21:42
Air Blank		0.000	21:43
Control Test		0.075	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 08/11/2020 14:11

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 08/12/2020 17:15:54

Diagnostics Check		OK	17:11	
Air Blank		0.000	17:11	
Control Test		0.082	17:12	
Air Blank		0.000	17:12	
Subject Sample	# 1		17:14	Subject Test Refused
Breath Volume		0.000	17:14	
Air Blank		0.000	17:14	
Control Test		0.082	17:15	
Air Blank		0.000	17:15	
Diagnostics Check		OK	17:15	

Data Download Date/Time: 09/15/2020 16:11

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMBERT,BRIAN,K
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAMBERT,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/16/2020 23:11:07

Diagnostics Check		OK	23:04
Air Blank		0.000	23:05
Control Test		0.082	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.168	23:06
Breath Volume		3.714	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:08
Subject Sample	# 2	0.168	23:09
Breath Volume		2.976	23:09
Air Blank		0.000	23:10
Control Test		0.081	23:10
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 09/15/2020 16:11

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JOHN,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAMBERT,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/22/2020 21:45:13

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:41	
Control Test		0.082	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:43	Subject Test Refused
Breath Volume		0.000	21:43	
Air Blank		0.000	21:44	
Control Test		0.083	21:44	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:45	

Data Download Date/Time: 09/15/2020 16:11

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT,COLTEN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT,COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: DYSON

DATE/TIME OF TEST: 08/30/2020 20:52:40

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.082	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.000	20:48
Breath Volume		2.835	20:48
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.000	20:51
Breath Volume		2.023	20:51
Air Blank		0.000	20:51
Control Test		0.081	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 09/15/2020 16:11

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLGOOD,JEFFREY,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAMBERT,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: BUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 08/01/2020 16:45:29

Diagnostics Check		OK	16:42	
Air Blank		0.000	16:43	
Control Test		0.083	16:43	
Air Blank		0.000	16:43	
Subject Sample	# 1		16:44	Subject Test Refused
Breath Volume		0.000	16:44	
Air Blank		0.000	16:44	
Control Test		0.083	16:44	
Air Blank		0.000	16:45	
Diagnostics Check		OK	16:45	

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 08/02/2020 22:14:22

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:09	
Control Test		0.084	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:12	Subject Test Refused
Breath Volume		0.761	22:12	
Air Blank		0.000	22:13	
Control Test		0.083	22:13	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: GAVARRETE

DATE/TIME OF TEST: 08/03/2020 19:40:51

Diagnostics Check		OK	19:34
Air Blank		0.000	19:35
Control Test		0.083	19:35
Air Blank		0.000	19:36
Subject Sample	# 1	0.078	19:36
Breath Volume		3.101	19:36
Air Blank		0.000	19:37
Air Blank		0.000	19:38
Subject Sample	# 2	0.076	19:39
Breath Volume		1.242	19:39
Air Blank		0.000	19:39
Control Test		0.083	19:40
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTZENBUEHLER,DOUGLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: HUTZENBUEHLER

DATE/TIME OF TEST: 08/06/2020 13:27:50

Diagnostics Check		OK	13:21
Air Blank		0.000	13:22
Control Test		0.083	13:22
Air Blank		0.000	13:23
Subject Sample	# 1	0.000	13:23
Breath Volume		1.203	13:23
Air Blank		0.000	13:23
Air Blank		0.000	13:25
Subject Sample	# 2	0.000	13:26
Breath Volume		1.265	13:26
Air Blank		0.000	13:26
Control Test		0.082	13:27
Air Blank		0.000	13:27
Diagnostics Check		OK	13:27

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCUM,JOSEPH,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARCUM,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: REARDON

DATE/TIME OF TEST: 08/07/2020 03:31:12

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.082	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.219	03:26
Breath Volume		1.628	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.210	03:29
Breath Volume		1.550	03:29
Air Blank		0.000	03:30
Control Test		0.082	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JOHN,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CLARK,JOHN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: BUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: CICERO

DATE/TIME OF TEST: 08/07/2020 18:48:10

Diagnostics Check		OK	18:45	
Air Blank		0.000	18:45	
Control Test		0.082	18:45	
Air Blank		0.000	18:46	
Subject Sample	# 1		18:46	Subject Test Refused
Breath Volume		0.000	18:46	
Air Blank		0.000	18:47	
Control Test		0.082	18:47	
Air Blank		0.000	18:48	
Diagnostics Check		OK	18:48	

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: COLSTON

DATE/TIME OF TEST: 08/07/2020 20:51:00

Diagnostics Check		OK	20:48	
Air Blank		0.000	20:48	
Control Test		0.082	20:48	
Air Blank		0.000	20:49	
Subject Sample	# 1		20:49	Subject Test Refused
Breath Volume		0.000	20:49	
Air Blank		0.000	20:50	
Control Test		0.082	20:50	
Air Blank		0.000	20:50	
Diagnostics Check		OK	20:51	

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JOHN,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CLARK,JOHN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: BUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 08/08/2020 16:50:26

Diagnostics Check		OK	16:47	
Air Blank		0.000	16:48	
Control Test		0.083	16:48	
Air Blank		0.000	16:48	
Subject Sample	# 1		16:49	Subject Test Refused
Breath Volume		0.000	16:49	
Air Blank		0.000	16:49	
Control Test		0.082	16:49	
Air Blank		0.000	16:50	
Diagnostics Check		OK	16:50	

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JOHN,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CLARK,JOHN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: HIMES

DATE/TIME OF TEST: 08/08/2020 18:19:23

Diagnostics Check		OK	18:12
Air Blank		0.000	18:12
Control Test		0.082	18:13
Air Blank		0.000	18:13
Subject Sample	# 1	0.217	18:14
Breath Volume		4.203	18:14
Air Blank		0.000	18:14
Air Blank		0.000	18:16
Subject Sample	# 2	0.197	18:17
Breath Volume		4.050	18:17
Air Blank		0.000	18:18
Control Test		0.079	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 08/11/2020 14:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SHORES

DATE/TIME OF TEST: 08/09/2020 05:15:19

Diagnostics Check	OK	05:14	
Air Blank		05:14	RFI Detect
Air Blank	0.000	05:15	

Data Download Date/Time: 09/02/2020 15:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: SHORES

DATE/TIME OF TEST: 08/09/2020 05:24:15

Diagnostics Check		OK	05:17
Air Blank		0.000	05:18
Control Test		0.082	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.202	05:19
Breath Volume		2.222	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.195	05:22
Breath Volume		1.714	05:22
Air Blank		0.000	05:23
Control Test		0.081	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 09/02/2020 15:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,ANTHONY,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: WINGROVE

DATE/TIME OF TEST: 08/18/2020 14:38:52

Diagnostics Check		OK	14:32
Air Blank		0.000	14:33
Control Test		0.082	14:33
Air Blank		0.000	14:33
Subject Sample	# 1	0.000	14:34
Breath Volume		1.953	14:34
Air Blank		0.000	14:34
Air Blank		0.000	14:36
Subject Sample	# 2	0.000	14:37
Breath Volume		4.128	14:37
Air Blank		0.000	14:37
Control Test		0.081	14:38
Air Blank		0.000	14:38
Diagnostics Check		OK	14:38

Data Download Date/Time: 09/02/2020 15:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PLEMMONS,L
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 08/02/2020 04:08:52

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.077	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Control Test		0.078	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 08/05/2020 19:08:22

Diagnostics Check		OK	19:05	
Air Blank		0.000	19:06	
Control Test		0.077	19:06	
Air Blank		0.000	19:06	
Subject Sample	# 1		19:07	Subject Test Refused
Breath Volume		0.000	19:07	
Air Blank		0.000	19:07	
Control Test		0.077	19:07	
Air Blank		0.000	19:08	
Diagnostics Check		OK	19:08	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: VARNER

DATE/TIME OF TEST: 08/07/2020 00:56:10

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.078	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.212	00:51
Breath Volume		2.371	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.194	00:54
Breath Volume		1.441	00:54
Air Blank		0.000	00:55
Control Test		0.077	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 08/07/2020 02:55:35

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.000	02:51
Breath Volume		2.035	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.000	02:54
Breath Volume		1.218	02:54
Air Blank		0.000	02:54
Control Test		0.077	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 08/08/2020 04:01:27

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.079	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:57	No Sample Provided
Breath Volume		0.000	03:57	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 2		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.078	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: LESCHANZ

DATE/TIME OF TEST: 08/08/2020 05:11:50

Diagnostics Check		OK	05:09	
Air Blank		0.000	05:09	
Control Test		0.079	05:09	
Air Blank		0.000	05:10	
Subject Sample	# 1		05:10	Subject Test Refused
Breath Volume		0.000	05:10	
Air Blank		0.000	05:10	
Control Test		0.079	05:11	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,BRETT,L
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: JOHNSON,ANDREW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 08/13/2020 21:25:16

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:16	
Control Test		0.078	21:17	
Air Blank		0.000	21:17	
Subject Sample	# 1	0.240	21:18	
Breath Volume		1.371	21:18	
Air Blank		0.000	21:19	
Air Blank		0.000	21:21	
Subject Sample	# 2		21:23	Subject Test Refused
Breath Volume		0.000	21:23	
Air Blank		0.000	21:24	
Control Test		0.076	21:24	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,BRETT,L
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: SULLIVAN,BRETT
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HALONSKI

DATE/TIME OF TEST: 08/16/2020 00:51:00

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.079	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.099	00:46
Breath Volume		2.367	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.095	00:49
Breath Volume		2.703	00:49
Air Blank		0.000	00:50
Control Test		0.077	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:51

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALDREE,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: POLVERE

DATE/TIME OF TEST: 08/16/2020 19:58:56

Diagnostics Check		OK	19:52
Air Blank		0.000	19:53
Control Test		0.078	19:53
Air Blank		0.000	19:54
Subject Sample	# 1	0.214	19:54
Breath Volume		2.078	19:54
Air Blank		0.000	19:55
Air Blank		0.000	19:56
Subject Sample	# 2	0.214	19:57
Breath Volume		2.066	19:57
Air Blank		0.000	19:58
Control Test		0.078	19:58
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,BRETT,L
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: JOHNSON,ANDREW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: EDENFIELD

DATE/TIME OF TEST: 08/19/2020 02:52:06

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.078	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.191	02:47
Breath Volume		2.269	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.183	02:50
Breath Volume		2.207	02:50
Air Blank		0.000	02:51
Control Test		0.077	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDREE,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALDREE,TASHAHNDA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 08/20/2020 05:33:43

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.000	05:29
Breath Volume		1.335	05:29
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.000	05:32
Breath Volume		2.320	05:32
Air Blank		0.000	05:32
Control Test		0.077	05:33
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ARGANBRIGHT

DATE/TIME OF TEST: 08/21/2020 00:54:13

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.078	00:46	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:49	No Sample Provided
Breath Volume		0.000	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 2		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Control Test		0.078	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATIVIDAD,DANIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NATIVIDAD,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 08/22/2020 03:25:07

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.188	03:19
Breath Volume		1.859	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.184	03:23
Breath Volume		2.781	03:23
Air Blank		0.000	03:24
Control Test		0.077	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: ELLIOTT,NIGEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/22/2020 23:13:43

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.078	23:06	
Air Blank		0.000	23:07	
Subject Sample	# 1	0.151	23:08	
Breath Volume		1.898	23:08	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2	0.146	23:12	
Breath Volume		1.539	23:12	
Air Blank		0.000	23:12	
Control Test		0.074	23:13	Control Outside Tolerance
Air Blank		0.000	23:13	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOUNDA

DATE/TIME OF TEST: 08/23/2020 02:05:25

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.079	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.048	02:04	
Breath Volume		2.234	02:04	
Air Blank			02:04	Ambient Fail
Air Blank			02:05	Ambient Fail

Data Download Date/Time: 08/31/2020 10:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOUNDA

DATE/TIME OF TEST: 08/23/2020 02:07:56

Diagnostics Check	OK	02:06	
Air Blank		02:07	Ambient Fail
Air Blank	0.000	02:07	

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOUNDA

DATE/TIME OF TEST: 08/23/2020 02:15:19

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.078	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.050	02:10
Breath Volume		1.660	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.051	02:13
Breath Volume		2.527	02:13
Air Blank		0.000	02:14
Control Test		0.077	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 08/31/2020 10:15

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/26/2020 22:42:24

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.076	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.062	22:37
Breath Volume		1.800	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.060	22:40
Breath Volume		2.859	22:40
Air Blank		0.000	22:41
Control Test		0.078	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,T,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ELLIOTT,N
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 08/28/2020 22:39:05

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:30	
Control Test		0.077	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 1	0.210	22:32	
Breath Volume		1.781	22:32	
Air Blank		0.000	22:32	
Air Blank		0.000	22:34	
Subject Sample	# 2	0.184	22:35	
Breath Volume		1.605	22:35	
Air Blank		0.000	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 3	0.194	22:38	
Breath Volume		1.437	22:38	
Air Blank			22:38	Improper Sample
Air Blank		0.000	22:39	

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,T,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ELLIOTT,N
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 08/28/2020 22:47:30

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.077	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.188	22:42
Breath Volume		2.750	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.185	22:45
Breath Volume		2.281	22:45
Air Blank		0.000	22:46
Control Test		0.076	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,TASHAHNDA,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GUERRY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MAGRINI

DATE/TIME OF TEST: 08/29/2020 03:36:02

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.076	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.156	03:31
Breath Volume		2.003	03:31
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.142	03:34
Breath Volume		1.718	03:34
Air Blank		0.000	03:35
Control Test		0.076	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:36

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,T,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: COODY,T
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SHIPP

DATE/TIME OF TEST: 08/29/2020 03:56:47

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.078	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.107	03:52
Breath Volume		2.265	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.111	03:55
Breath Volume		2.957	03:55
Air Blank		0.000	03:55
Control Test		0.077	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,T,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REGAL,T
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 08/29/2020 05:37:00

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:34	
Control Test		0.077	05:34	
Air Blank		0.000	05:35	
Subject Sample	# 1		05:35	Subject Test Refused
Breath Volume		0.000	05:35	
Air Blank		0.000	05:36	
Control Test		0.077	05:36	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COODY,T,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GUERRY,C
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: HERNANDEZ GARCIA

DATE/TIME OF TEST: 08/30/2020 03:18:51

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:16	
Control Test		0.078	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:17	
Control Test		0.078	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: TRUESDALE,J
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PARTIN

DATE/TIME OF TEST: 08/31/2020 02:22:09

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.151	02:17
Breath Volume		2.703	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.148	02:20
Breath Volume		2.699	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/31/2020 10:16

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ETHRIDGE,ALISA,B
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: DEMING,STORM
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/15/2021
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: BARROSO GUERRA

DATE/TIME OF TEST: 08/23/2020 16:16:10

Diagnostics Check		OK	16:08
Air Blank		0.000	16:09
Control Test		0.079	16:09
Air Blank		0.000	16:10
Subject Sample	# 1	0.256	16:11
Breath Volume		3.027	16:11
Air Blank		0.000	16:12
Air Blank		0.000	16:14
Subject Sample	# 2	0.250	16:14
Breath Volume		2.769	16:14
Air Blank		0.000	16:15
Control Test		0.075	16:15
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 09/09/2020 11:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WEDDLE

DATE/TIME OF TEST: 08/15/2020 04:01:13

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.083	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.143	03:56
Breath Volume		2.652	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.145	03:59
Breath Volume		2.343	03:59
Air Blank		0.000	04:00
Control Test		0.082	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 09/05/2020 20:10

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 08/20/2020 13:25:27

Diagnostics Check		OK	13:21	
Air Blank		0.000	13:22	
Control Test		0.079	13:22	
Air Blank		0.000	13:22	
Subject Sample	# 1	0.365	13:23	
Breath Volume		3.203	13:23	
Air Blank			13:23	Improper Sample
Air Blank		0.000	13:24	
Control Test		0.084	13:24	
Air Blank		0.000	13:25	
Diagnostics Check		OK	13:25	

Data Download Date/Time: 09/05/2020 20:10

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: STRATTON

DATE/TIME OF TEST: 08/20/2020 13:33:47

Diagnostics Check		OK	13:27
Air Blank		0.000	13:28
Control Test		0.081	13:28
Air Blank		0.000	13:28
Subject Sample	# 1	0.351	13:29
Breath Volume		2.851	13:29
Air Blank		0.000	13:30
Air Blank		0.000	13:31
Subject Sample	# 2	0.355	13:32
Breath Volume		3.257	13:32
Air Blank		0.000	13:32
Control Test		0.082	13:33
Air Blank		0.000	13:33
Diagnostics Check		OK	13:33

Data Download Date/Time: 09/05/2020 20:10

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADFORD,CHARLES,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WADFORD,CHARLES
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/31/2020 14:29:09

Diagnostics Check		OK	14:22
Air Blank		0.000	14:22
Control Test		0.082	14:23
Air Blank		0.000	14:23
Subject Sample	# 1	0.267	14:24
Breath Volume		1.558	14:24
Air Blank		0.000	14:25
Air Blank		0.000	14:27
Subject Sample	# 2	0.271	14:27
Breath Volume		1.519	14:27
Air Blank		0.000	14:28
Control Test		0.081	14:28
Air Blank		0.000	14:29
Diagnostics Check		OK	14:29

Data Download Date/Time: 09/05/2020 20:10

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SUMMERSET

DATE/TIME OF TEST: 08/31/2020 23:19:23

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:11	
Control Test		0.081	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1	0.311	23:12	
Breath Volume		2.320	23:12	
Air Blank		0.000	23:13	
Air Blank		0.000	23:15	
Subject Sample	# 2		23:17	Subject Test Refused
Breath Volume		0.742	23:17	
Air Blank			23:18	Purge Fail
Air Blank		0.000	23:19	

Data Download Date/Time: 09/05/2020 20:10

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SUMMERSET

DATE/TIME OF TEST: 08/31/2020 23:26:56

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.077	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1	0.245	23:25	Volume Not Met
Breath Volume		1.539	23:25	
Air Blank			23:26	Purge Fail
Air Blank		0.000	23:26	

Data Download Date/Time: 09/05/2020 20:10

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONEY,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MALONEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 906364
DRY GAS EXPIRY DATE: 02/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: POLSON

DATE/TIME OF TEST: 08/15/2020 21:47:39

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.079	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.183	21:43
Breath Volume		3.496	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.164	21:45
Breath Volume		1.507	21:45
Air Blank		0.000	21:46
Control Test		0.078	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 09/04/2020 13:39

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SOLOMON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 08:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 08/18/2020 08:56:46

Diagnostics Check		OK	08:50
Air Blank		0.000	08:50
Control Test		0.080	08:51
Air Blank		0.000	08:51
Subject Sample	# 1	0.216	08:52
Breath Volume		2.230	08:52
Air Blank		0.000	08:52
Air Blank		0.000	08:54
Subject Sample	# 2	0.225	08:55
Breath Volume		1.464	08:55
Air Blank		0.000	08:55
Control Test		0.078	08:56
Air Blank		0.000	08:56
Diagnostics Check		OK	08:56

Data Download Date/Time: 09/08/2020 08:46

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,JERRY,D
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: MAYO,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GRANDE

DATE/TIME OF TEST: 08/27/2020 20:26:39

Diagnostics Check		OK	20:16
Air Blank		0.000	20:17
Control Test		0.077	20:17
Air Blank		0.000	20:18
Subject Sample	# 1	0.277	20:18
Breath Volume		2.000	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.250	20:21
Breath Volume		2.382	20:21
Air Blank		0.000	20:22
Air Blank		0.000	20:24
Subject Sample	# 3	0.249	20:24
Breath Volume		2.445	20:24
Air Blank		0.000	20:25
Control Test		0.077	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 09/08/2020 08:46

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARDY,JOE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: BLAXTON

DATE/TIME OF TEST: 08/30/2020 10:48:47

Diagnostics Check		OK	10:45	
Air Blank		0.000	10:45	
Control Test		0.078	10:45	
Air Blank		0.000	10:46	
Subject Sample	# 1		10:47	Subject Test Refused
Breath Volume		0.000	10:47	
Air Blank		0.000	10:47	
Control Test		0.079	10:48	
Air Blank		0.000	10:48	
Diagnostics Check		OK	10:48	

Data Download Date/Time: 09/08/2020 08:46

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: FOSTER,LOGAN
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 08/30/2020 22:26:45

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.078	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.202	22:22
Breath Volume		3.894	22:22
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.195	22:25
Breath Volume		1.847	22:25
Air Blank		0.000	22:25
Control Test		0.078	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 09/08/2020 08:46

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 08/09/2020 03:39:07

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.082	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Control Test		0.082	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 09/04/2020 13:42

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 08/09/2020 04:52:14

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.236	04:46
Breath Volume		1.183	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.221	04:50
Breath Volume		1.121	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 09/04/2020 13:42

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: CALVILLO

DATE/TIME OF TEST: 08/09/2020 17:08:12

Diagnostics Check		OK	17:02	
Air Blank		0.000	17:03	
Control Test		0.081	17:03	
Air Blank		0.000	17:03	
Subject Sample	# 1	0.243	17:06	
Breath Volume		2.117	17:06	
Air Blank			17:07	Purge Fail
Air Blank		0.000	17:08	

Data Download Date/Time: 09/04/2020 13:42

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: CALVILLO

DATE/TIME OF TEST: 08/09/2020 17:13:06

Diagnostics Check	OK	17:11	
Air Blank	0.000	17:12	
Control Test	0.073	17:12	Control Outside Tolerance
Air Blank	0.000	17:13	

Data Download Date/Time: 09/04/2020 13:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: CALVILLO

DATE/TIME OF TEST: 08/09/2020 17:26:56

Diagnostics Check		OK	17:16	
Air Blank		0.000	17:16	
Control Test		0.078	17:16	
Air Blank		0.000	17:17	
Subject Sample	# 1	0.254	17:17	
Breath Volume		1.703	17:17	
Air Blank		0.000	17:18	
Air Blank		0.000	17:20	
Subject Sample	# 2	0.233	17:22	
Breath Volume		1.679	17:22	
Air Blank		0.000	17:23	
Air Blank		0.000	17:24	
Subject Sample	# 3	0.242	17:25	
Breath Volume		1.445	17:25	
Air Blank		0.000	17:25	
Control Test		0.073	17:26	Control Outside Tolerance
Air Blank		0.000	17:26	

Data Download Date/Time: 09/04/2020 13:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: CALVILLO

DATE/TIME OF TEST: 08/09/2020 17:36:15

Diagnostics Check		OK	17:29
Air Blank		0.000	17:30
Control Test		0.078	17:30
Air Blank		0.000	17:31
Subject Sample	# 1	0.241	17:31
Breath Volume		1.765	17:31
Air Blank		0.000	17:32
Air Blank		0.000	17:33
Subject Sample	# 2	0.240	17:34
Breath Volume		2.207	17:34
Air Blank		0.000	17:35
Control Test		0.077	17:35
Air Blank		0.000	17:36
Diagnostics Check		OK	17:36

Data Download Date/Time: 09/04/2020 13:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,BRETT,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WAGNER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LINDAY

DATE/TIME OF TEST: 08/13/2020 04:22:22

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.082	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.246	04:17
Breath Volume		1.375	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:19
Subject Sample	# 2	0.252	04:20
Breath Volume		1.925	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 09/04/2020 13:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: HINKLE

DATE/TIME OF TEST: 08/20/2020 14:38:19

Diagnostics Check		OK	14:31
Air Blank		0.000	14:32
Control Test		0.082	14:32
Air Blank		0.000	14:33
Subject Sample	# 1	0.184	14:33
Breath Volume		2.953	14:33
Air Blank		0.000	14:34
Air Blank		0.000	14:36
Subject Sample	# 2	0.185	14:36
Breath Volume		2.925	14:36
Air Blank		0.000	14:37
Control Test		0.079	14:37
Air Blank		0.000	14:38
Diagnostics Check		OK	14:38

Data Download Date/Time: 09/04/2020 13:47

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HENDRICKSON

DATE/TIME OF TEST: 08/20/2020 18:58:32

Diagnostics Check		OK	18:51
Air Blank		0.000	18:51
Control Test		0.081	18:51
Air Blank		0.000	18:52
Subject Sample	# 1	0.196	18:53
Breath Volume		1.222	18:53
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.210	18:56
Breath Volume		1.785	18:56
Air Blank		0.000	18:57
Control Test		0.078	18:57
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 09/04/2020 13:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 08/22/2020 00:19:15

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.082	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.000	00:14
Breath Volume		3.523	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.000	00:17
Breath Volume		4.367	00:17
Air Blank		0.000	00:18
Control Test		0.082	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 09/04/2020 13:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/24/2020 18:46:35

Diagnostics Check		OK	18:39
Air Blank		0.000	18:40
Control Test		0.082	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.000	18:41
Breath Volume		1.890	18:41
Air Blank		0.000	18:42
Air Blank		0.000	18:44
Subject Sample	# 2	0.000	18:45
Breath Volume		1.320	18:45
Air Blank		0.000	18:45
Control Test		0.082	18:45
Air Blank		0.000	18:46
Diagnostics Check		OK	18:46

Data Download Date/Time: 09/04/2020 13:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASSITER,STEVEN,P
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: BARTON,GREG
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: BROWNELL

DATE/TIME OF TEST: 08/17/2020 12:40:12

Diagnostics Check		OK	12:28	
Air Blank		0.000	12:29	
Control Test		0.081	12:29	
Air Blank		0.000	12:30	
Subject Sample	# 1		12:33	No Sample Provided
Breath Volume		0.000	12:33	
Air Blank		0.000	12:33	
Air Blank		0.000	12:35	
Subject Sample	# 2		12:38	No Sample Provided
Breath Volume		0.000	12:38	
Air Blank		0.000	12:39	
Control Test		0.081	12:39	
Air Blank		0.000	12:40	
Diagnostics Check		OK	12:40	

Data Download Date/Time: 09/02/2020 15:14

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CONNOR,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: HOSNY

DATE/TIME OF TEST: 08/21/2020 05:37:22

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.078	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.125	05:31
Breath Volume		1.218	05:31
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.122	05:35
Breath Volume		2.250	05:35
Air Blank		0.000	05:36
Control Test		0.077	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JUMPING EAGLE

DATE/TIME OF TEST: 08/21/2020 23:51:29

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.076	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1	0.230	23:47	
Breath Volume		2.730	23:47	
Air Blank		0.000	23:48	
Air Blank		0.000	23:50	
Subject Sample	# 2	0.219	23:50	
Breath Volume		2.097	23:50	
Air Blank			23:51	RFI Detect
Air Blank		0.000	23:51	

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JUMPING EAGLE

DATE/TIME OF TEST: 08/21/2020 23:59:52

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:54	
Control Test		0.077	23:54	
Air Blank		0.000	23:55	
Subject Sample	# 1	0.216	23:55	
Breath Volume		1.972	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:58	
			23:58	Improper Sample
Air Blank		0.000	23:58	
Control Test		0.077	23:59	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: JUMPING EAGLE

DATE/TIME OF TEST: 08/22/2020 00:08:10

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.076	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.220	00:02
Breath Volume		2.359	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.209	00:06
Breath Volume		1.945	00:06
Air Blank		0.000	00:07
Control Test		0.077	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WOODRUFF,MARK
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/22/2020 01:16:57

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.077	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.237	01:12
Breath Volume		1.640	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.240	01:15
Breath Volume		1.644	01:15
Air Blank		0.000	01:16
Control Test		0.077	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 08/22/2020 03:52:23

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.077	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.199	03:47
Breath Volume		2.390	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.194	03:50
Breath Volume		2.234	03:50
Air Blank		0.000	03:51
Control Test		0.077	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: WOODRUFF,MARK
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 08/22/2020 19:45:07

Diagnostics Check		OK	19:38
Air Blank		0.000	19:38
Control Test		0.078	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.193	19:40
Breath Volume		1.906	19:40
Air Blank		0.000	19:41
Air Blank		0.000	19:42
Subject Sample	# 2	0.198	19:43
Breath Volume		2.093	19:43
Air Blank		0.000	19:44
Control Test		0.076	19:44
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: COTTER

DATE/TIME OF TEST: 08/23/2020 00:37:42

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.119	00:33
Breath Volume		1.789	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.122	00:36
Breath Volume		1.457	00:36
Air Blank		0.000	00:36
Control Test		0.077	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,J,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ARNOLD,J
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: BUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: PAKARINEN

DATE/TIME OF TEST: 08/23/2020 20:30:24

Diagnostics Check		OK	20:23
Air Blank		0.000	20:24
Control Test		0.077	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.138	20:25
Breath Volume		2.433	20:25
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.142	20:28
Breath Volume		3.992	20:28
Air Blank		0.000	20:29
Control Test		0.076	20:29
Air Blank		0.000	20:30
Diagnostics Check		OK	20:30

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DODICH

DATE/TIME OF TEST: 08/24/2020 03:18:51

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.123	03:14
Breath Volume		2.855	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.115	03:17
Breath Volume		2.953	03:17
Air Blank		0.000	03:17
Control Test		0.076	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: ABDOR

DATE/TIME OF TEST: 08/25/2020 16:03:27

Diagnostics Check		OK	15:57
Air Blank		0.000	15:57
Control Test		0.078	15:58
Air Blank		0.000	15:58
Subject Sample	# 1	0.000	15:59
Breath Volume		2.500	15:59
Air Blank		0.000	15:59
Air Blank		0.000	16:01
Subject Sample	# 2	0.000	16:02
Breath Volume		2.589	16:02
Air Blank		0.000	16:02
Control Test		0.078	16:02
Air Blank		0.000	16:03
Diagnostics Check		OK	16:03

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 08/26/2020 16:52:33

Diagnostics Check		OK	16:46
Air Blank		0.000	16:46
Control Test		0.078	16:46
Air Blank		0.000	16:47
Subject Sample	# 1	0.147	16:48
Breath Volume		1.371	16:48
Air Blank		0.000	16:48
Air Blank		0.000	16:50
Subject Sample	# 2	0.133	16:50
Breath Volume		1.726	16:50
Air Blank		0.000	16:51
Control Test		0.077	16:51
Air Blank		0.000	16:52
Diagnostics Check		OK	16:52

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: FEENEY

DATE/TIME OF TEST: 08/29/2020 05:24:02

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.079	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.000	05:19
Breath Volume		3.675	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.000	05:22
Breath Volume		2.882	05:22
Air Blank		0.000	05:23
Control Test		0.076	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:24

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: SEARCHWELL

DATE/TIME OF TEST: 08/30/2020 20:38:00

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:25	
Control Test		0.077	20:25	
Air Blank		0.000	20:25	
Subject Sample	# 1		20:29	No Sample Provided
Breath Volume		0.000	20:29	
Air Blank		0.000	20:29	
Air Blank		0.000	20:31	
Subject Sample	# 2	0.130	20:33	
Breath Volume		2.617	20:33	
Air Blank		0.000	20:34	
Air Blank		0.000	20:36	
Subject Sample	# 3	0.117	20:36	
Breath Volume		2.339	20:36	
Air Blank		0.000	20:37	
Control Test		0.072	20:37	Control Outside Tolerance
Air Blank		0.000	20:38	

Data Download Date/Time: 09/15/2020 10:13

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERAVIL,JEAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: DERAVIL,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 08/01/2020 01:11:25

Diagnostics Check	OK	01:10	
Air Blank		01:10	RFI Detect
Air Blank	0.000	01:11	

Data Download Date/Time: 08/20/2020 09:50

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERAVIL,JEAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: DERAVIL,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 08/01/2020 01:20:45

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.175	01:16
Breath Volume		4.910	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.173	01:19
Breath Volume		4.062	01:19
Air Blank		0.000	01:19
Control Test		0.078	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 08/20/2020 09:50

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PINEDA HERNANDEZ

DATE/TIME OF TEST: 08/01/2020 04:14:15

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.118	04:09
Breath Volume		1.554	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.119	04:12
Breath Volume		2.175	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 08/20/2020 09:50

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 08/02/2020 03:27:00

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.000	03:22
Breath Volume		2.183	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.000	03:25
Breath Volume		1.292	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:27

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 08/03/2020 01:01:12

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.081	00:55	
Air Blank		0.000	00:55	
Subject Sample	# 1	0.314	00:56	
Breath Volume		1.722	00:56	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2		00:59	Slope Not Met
Breath Volume		1.308	00:59	
Air Blank		0.000	01:00	
Control Test		0.080	01:00	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 08/03/2020 01:10:01

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.303	01:04
Breath Volume		1.664	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.286	01:08
Breath Volume		1.296	01:08
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:10

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SEARS

DATE/TIME OF TEST: 08/04/2020 03:59:22

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.081	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.190	03:54
Breath Volume		4.156	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.193	03:57
Breath Volume		2.964	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 08/04/2020 20:07:54

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.081	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.175	20:02
Breath Volume		1.832	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.184	20:06
Breath Volume		2.039	20:06
Air Blank		0.000	20:06
Control Test		0.080	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/07/2020 19:14:02

Diagnostics Check		OK	19:05
Air Blank		0.000	19:06
Control Test		0.080	19:06
Air Blank		0.000	19:07
Subject Sample	# 1	0.060	19:09
Breath Volume		1.222	19:09
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.058	19:12
Breath Volume		1.296	19:12
Air Blank		0.000	19:13
Control Test		0.080	19:13
Air Blank		0.000	19:13
Diagnostics Check		OK	19:14

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 08/08/2020 01:29:09

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.189	01:24
Breath Volume		2.550	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.194	01:27
Breath Volume		3.296	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: COLANGELO,CODY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: WENGER

DATE/TIME OF TEST: 08/08/2020 03:21:12

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.172	03:16
Breath Volume		2.050	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.176	03:19
Breath Volume		2.769	03:19
Air Blank		0.000	03:20
Control Test		0.080	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: STEVESON

DATE/TIME OF TEST: 08/09/2020 18:44:27

Diagnostics Check		OK	18:38
Air Blank		0.000	18:38
Control Test		0.081	18:38
Air Blank		0.000	18:39
Subject Sample	# 1	0.166	18:40
Breath Volume		1.796	18:40
Air Blank		0.000	18:40
Air Blank		0.000	18:42
Subject Sample	# 2	0.164	18:42
Breath Volume		1.519	18:42
Air Blank		0.000	18:43
Control Test		0.080	18:43
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: THOROGOOD

DATE/TIME OF TEST: 08/11/2020 04:44:18

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.080	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.275	04:39
Breath Volume		3.496	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:41
Subject Sample	# 2	0.273	04:42
Breath Volume		2.921	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 08/12/2020 03:47:26

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.096	03:42
Breath Volume		2.250	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.094	03:45
Breath Volume		2.273	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGARD,RICKY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HAGGARD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NIETO GAYTAN

DATE/TIME OF TEST: 08/15/2020 04:34:35

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.166	04:29
Breath Volume		2.007	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.167	04:32
Breath Volume		2.632	04:32
Air Blank		0.000	04:33
Control Test		0.076	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: COLINDREZ MONTALVAN

DATE/TIME OF TEST: 08/17/2020 04:23:26

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.215	04:18
Breath Volume		1.957	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.209	04:21
Breath Volume		1.449	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 08/20/2020 09:51

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIPFORD,JASON,V
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GERBER,OLIVER
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2121
VIOLATION CODE: BUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: NUNERY

DATE/TIME OF TEST: 08/01/2020 02:21:06

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:10	
Control Test		0.080	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:14	No Sample Provided
Breath Volume		0.000	02:14	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 2		02:19	No Sample Provided
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.080	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 08/31/2020 15:26

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROSE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2121
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 08/23/2020 21:16:37

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.078	21:10
Air Blank		0.000	21:10
Subject Sample	# 1	0.149	21:11
Breath Volume		1.335	21:11
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.153	21:15
Breath Volume		1.714	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 08/31/2020 15:26

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: YOFFE

DATE/TIME OF TEST: 08/05/2020 23:45:29

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.080	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.171	23:41
Breath Volume		1.375	23:41
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.182	23:43
Breath Volume		1.640	23:43
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/06/2020 03:38:14

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.214	03:33
Breath Volume		1.644	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.222	03:36
Breath Volume		1.367	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: BEHNKE

DATE/TIME OF TEST: 08/06/2020 18:56:23

Diagnostics Check		OK	18:50
Air Blank		0.000	18:50
Control Test		0.079	18:50
Air Blank		0.000	18:51
Subject Sample	# 1	0.291	18:51
Breath Volume		1.628	18:51
Air Blank		0.000	18:52
Air Blank		0.000	18:54
Subject Sample	# 2	0.295	18:54
Breath Volume		1.617	18:54
Air Blank		0.000	18:55
Control Test		0.079	18:55
Air Blank		0.000	18:56
Diagnostics Check		OK	18:56

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 08/07/2020 03:51:13

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.081	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: ZERMENO ZERMENO

DATE/TIME OF TEST: 08/07/2020 19:22:10

Diagnostics Check		OK	19:15
Air Blank		0.000	19:15
Control Test		0.080	19:16
Air Blank		0.000	19:16
Subject Sample	# 1	0.135	19:17
Breath Volume		2.074	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.139	19:20
Breath Volume		2.343	19:20
Air Blank		0.000	19:21
Control Test		0.080	19:21
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: GABRIEL LOPEZ

DATE/TIME OF TEST: 08/07/2020 23:39:28

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.081	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.216	23:33
Breath Volume		1.679	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.224	23:37
Breath Volume		1.484	23:37
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: GRANTHAM

DATE/TIME OF TEST: 08/08/2020 02:11:20

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.130	02:06
Breath Volume		1.234	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.132	02:09
Breath Volume		2.355	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 08/08/2020 04:46:40

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.137	04:42
Breath Volume		2.640	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.131	04:45
Breath Volume		2.425	04:45
Air Blank		0.000	04:45
Control Test		0.079	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: FERMIN PENA

DATE/TIME OF TEST: 08/08/2020 21:29:21

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.081	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.210	21:24
Breath Volume		2.070	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.211	21:27
Breath Volume		1.234	21:27
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: VALDEZ JR

DATE/TIME OF TEST: 08/08/2020 23:06:35

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.081	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.000	23:02
Breath Volume		1.496	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.000	23:05
Breath Volume		1.332	23:05
Air Blank		0.000	23:05
Control Test		0.080	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 08/09/2020 04:06:31

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.251	04:01
Breath Volume		1.703	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.236	04:04
Breath Volume		2.226	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 08/09/2020 04:53:03

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.081	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.146	04:45
Breath Volume		1.648	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.178	04:48
Breath Volume		1.312	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:50
Subject Sample	# 3	0.167	04:51
Breath Volume		1.746	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:53

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HARVESTER,W
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: RODRIGUEZJUAREZ

DATE/TIME OF TEST: 08/09/2020 07:16:17

Diagnostics Check		OK	07:09
Air Blank		0.000	07:10
Control Test		0.081	07:10
Air Blank		0.000	07:10
Subject Sample	# 1	0.110	07:11
Breath Volume		3.585	07:11
Air Blank		0.000	07:12
Air Blank		0.000	07:14
Subject Sample	# 2	0.111	07:14
Breath Volume		4.019	07:14
Air Blank		0.000	07:15
Control Test		0.079	07:15
Air Blank		0.000	07:16
Diagnostics Check		OK	07:16

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: COSTON

DATE/TIME OF TEST: 08/09/2020 21:35:29

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.081	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.203	21:30
Breath Volume		2.261	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.197	21:33
Breath Volume		1.253	21:33
Air Blank		0.000	21:34
Control Test		0.080	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 09/02/2020 18:43

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MONTALVAN MACHADO

DATE/TIME OF TEST: 08/13/2020 00:51:44

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.000	00:46
Breath Volume		1.253	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.000	00:50
Breath Volume		2.156	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: DOMINGO MORALES

DATE/TIME OF TEST: 08/13/2020 04:32:11

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:19	
Control Test		0.081	04:19	
Air Blank		0.000	04:20	
Subject Sample	# 1	0.100	04:23	Volume Not Met
Breath Volume		0.792	04:23	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 2	0.130	04:26	
Breath Volume		3.003	04:26	
Air Blank		0.000	04:27	
Air Blank		0.000	04:29	
Subject Sample	# 3	0.123	04:30	
Breath Volume		1.851	04:30	
Air Blank		0.000	04:31	
Control Test		0.079	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: AGUIAR SOARES

DATE/TIME OF TEST: 08/15/2020 02:30:45

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.155	02:26
Breath Volume		2.367	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.150	02:29
Breath Volume		2.246	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 08/15/2020 04:35:10

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.081	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.214	04:30
Breath Volume		2.269	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.223	04:33
Breath Volume		1.394	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: ATCHISON

DATE/TIME OF TEST: 08/16/2020 00:31:19

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.081	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.198	00:26
Breath Volume		3.031	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.210	00:29
Breath Volume		2.207	00:29
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 08/17/2020 01:31:01

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.081	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 1	0.216	01:24	Volume Not Met
Breath Volume		0.214	01:24	
Air Blank		0.000	01:24	
Air Blank		0.000	01:26	
Subject Sample	# 2	0.176	01:29	Volume Not Met
Breath Volume		0.101	01:29	
Air Blank			01:30	Purge Fail
Air Blank		0.000	01:31	

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: DIAZ DIAZ

DATE/TIME OF TEST: 08/17/2020 22:02:14

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.080	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.200	21:57
Breath Volume		1.753	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.197	22:00
Breath Volume		1.707	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HARGROVE

DATE/TIME OF TEST: 08/18/2020 00:49:38

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.145	00:43
Breath Volume		1.156	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.154	00:48
Breath Volume		1.980	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MCNEIL

DATE/TIME OF TEST: 08/19/2020 23:54:06

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.080	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.000	23:49
Breath Volume		3.167	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.000	23:52
Breath Volume		3.406	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ALTAMIRANO GARCIA

DATE/TIME OF TEST: 08/20/2020 02:12:40

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.176	02:07
Breath Volume		1.199	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.166	02:11
Breath Volume		3.628	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DWYER,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 08/20/2020 19:39:35

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.081	19:33
Air Blank		0.000	19:34
Subject Sample	# 1	0.094	19:35
Breath Volume		1.714	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.093	19:38
Breath Volume		1.785	19:38
Air Blank		0.000	19:38
Control Test		0.080	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 08/21/2020 23:08:41

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.187	23:04
Breath Volume		1.511	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.181	23:07
Breath Volume		1.238	23:07
Air Blank		0.000	23:07
Control Test		0.079	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 08/22/2020 07:39:27

Diagnostics Check		OK	07:32
Air Blank		0.000	07:33
Control Test		0.080	07:33
Air Blank		0.000	07:34
Subject Sample	# 1	0.092	07:34
Breath Volume		2.281	07:34
Air Blank		0.000	07:35
Air Blank		0.000	07:37
Subject Sample	# 2	0.089	07:37
Breath Volume		2.214	07:37
Air Blank		0.000	07:38
Control Test		0.080	07:38
Air Blank		0.000	07:39
Diagnostics Check		OK	07:39

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BERG,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: WATTERS

DATE/TIME OF TEST: 08/22/2020 21:40:30

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.081	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.215	21:36
Breath Volume		3.226	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.219	21:38
Breath Volume		3.199	21:38
Air Blank		0.000	21:39
Control Test		0.079	21:39
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/23/2020 03:39:16

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.198	03:34
Breath Volume		1.984	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.212	03:37
Breath Volume		3.402	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KENO

DATE/TIME OF TEST: 08/23/2020 04:23:09

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.146	04:18
Breath Volume		1.402	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.150	04:21
Breath Volume		1.253	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOWARD,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/24/2020 05:37:14

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.081	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.171	05:32
Breath Volume		1.773	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.172	05:35
Breath Volume		1.726	05:35
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/25/2020 02:19:16

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.125	02:14
Breath Volume		1.636	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.119	02:17
Breath Volume		1.757	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: VELEZ SR

DATE/TIME OF TEST: 08/26/2020 01:15:19

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.250	01:10
Breath Volume		2.496	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.246	01:13
Breath Volume		2.347	01:13
Air Blank		0.000	01:14
Control Test		0.081	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SOUZA

DATE/TIME OF TEST: 08/28/2020 01:40:59

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.079	01:36
Breath Volume		2.570	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.081	01:39
Breath Volume		2.851	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KRIEGBAUM

DATE/TIME OF TEST: 08/28/2020 03:41:57

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.117	03:37
Breath Volume		1.152	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:39
Subject Sample	# 2	0.128	03:40
Breath Volume		1.562	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: PRATS RENDON

DATE/TIME OF TEST: 08/28/2020 04:20:36

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.181	04:15
Breath Volume		1.351	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.188	04:19
Breath Volume		1.574	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: SPITZER

DATE/TIME OF TEST: 08/28/2020 17:52:00

Diagnostics Check		OK	17:45
Air Blank		0.000	17:46
Control Test		0.080	17:46
Air Blank		0.000	17:46
Subject Sample	# 1	0.112	17:47
Breath Volume		1.152	17:47
Air Blank		0.000	17:48
Air Blank		0.000	17:49
Subject Sample	# 2	0.109	17:50
Breath Volume		1.101	17:50
Air Blank		0.000	17:51
Control Test		0.080	17:51
Air Blank		0.000	17:51
Diagnostics Check		OK	17:52

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 08/29/2020 00:25:11

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.121	00:20
Breath Volume		2.375	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.125	00:23
Breath Volume		1.671	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 08/29/2020 03:54:02

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.138	03:48
Breath Volume		1.101	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.148	03:52
Breath Volume		1.429	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:54

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: SANTIAGO RODRIGUEZ

DATE/TIME OF TEST: 08/29/2020 05:11:47

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:05	
Control Test		0.081	05:05	
Air Blank		0.000	05:06	
Subject Sample	# 1	0.198	05:07	
Breath Volume		1.988	05:07	
Air Blank		0.000	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 2		05:10	Slope Not Met
Breath Volume		0.707	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:11	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: SANTIAGO RODRIGUEZ

DATE/TIME OF TEST: 08/29/2020 05:40:11

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.080	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.180	05:34
Breath Volume		1.156	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.180	05:38
Breath Volume		1.941	05:38
Air Blank		0.000	05:39
Control Test		0.079	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/29/2020 21:04:26

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.081	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.148	20:59
Breath Volume		1.691	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.151	21:02
Breath Volume		1.593	21:02
Air Blank		0.000	21:03
Control Test		0.081	21:03
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: STALLWORTH

DATE/TIME OF TEST: 08/30/2020 06:28:32

Diagnostics Check		OK	06:22
Air Blank		0.000	06:22
Control Test		0.081	06:23
Air Blank		0.000	06:23
Subject Sample	# 1	0.000	06:24
Breath Volume		2.148	06:24
Air Blank		0.000	06:24
Air Blank		0.000	06:26
Subject Sample	# 2	0.000	06:27
Breath Volume		2.085	06:27
Air Blank		0.000	06:27
Control Test		0.080	06:27
Air Blank		0.000	06:28
Diagnostics Check		OK	06:28

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GREWE

DATE/TIME OF TEST: 08/31/2020 01:29:27

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.151	01:24
Breath Volume		1.386	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.154	01:27
Breath Volume		2.164	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 09/02/2020 18:44

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,ROBERT,B
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: ROBSON,WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 08/01/2020 23:01:24

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.078	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.192	22:57
Breath Volume		1.398	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.199	22:59
Breath Volume		1.339	22:59
Air Blank		0.000	23:00
Control Test		0.078	23:00
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 08/25/2020 10:32

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JAKE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS,JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: GLICKSMAN

DATE/TIME OF TEST: 08/14/2020 18:12:15

Diagnostics Check		OK	18:06
Air Blank		0.000	18:06
Control Test		0.078	18:06
Air Blank		0.000	18:07
Subject Sample	# 1	0.097	18:07
Breath Volume		2.886	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.103	18:10
Breath Volume		2.535	18:10
Air Blank		0.000	18:11
Control Test		0.078	18:11
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 08/25/2020 10:32

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PRATER

DATE/TIME OF TEST: 08/18/2020 23:15:02

Diagnostics Check		OK	23:08
Air Blank		0.000	23:08
Control Test		0.079	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.091	23:10
Breath Volume		1.781	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:12
Subject Sample	# 2	0.088	23:13
Breath Volume		1.937	23:13
Air Blank		0.000	23:14
Control Test		0.078	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:15

Data Download Date/Time: 08/25/2020 10:32

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATTHEW
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/07/2020 13:18:47

Diagnostics Check		OK	13:15	
Air Blank		0.000	13:15	
Control Test		0.079	13:16	
Air Blank		0.000	13:16	
Subject Sample	# 1		13:17	RFI Detect
Breath Volume		0.496	13:17	
Air Blank		0.000	13:17	
Control Test		0.080	13:18	
Air Blank		0.000	13:18	
Diagnostics Check		OK	13:18	

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATTHEW
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/07/2020 13:27:15

Diagnostics Check		OK	13:20
Air Blank		0.000	13:20
Control Test		0.079	13:21
Air Blank		0.000	13:21
Subject Sample	# 1	0.000	13:22
Breath Volume		1.425	13:22
Air Blank		0.000	13:22
Air Blank		0.000	13:24
Subject Sample	# 2	0.000	13:25
Breath Volume		2.203	13:25
Air Blank		0.000	13:26
Control Test		0.077	13:26
Air Blank		0.000	13:27
Diagnostics Check		OK	13:27

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 08/07/2020 17:16:55

Diagnostics Check		OK	17:10
Air Blank		0.000	17:11
Control Test		0.079	17:11
Air Blank		0.000	17:12
Subject Sample	# 1	0.000	17:12
Breath Volume		3.386	17:12
Air Blank		0.000	17:13
Air Blank		0.000	17:15
Subject Sample	# 2	0.000	17:15
Breath Volume		3.167	17:15
Air Blank		0.000	17:16
Control Test		0.080	17:16
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WIGGLESWORTH

DATE/TIME OF TEST: 08/10/2020 01:16:15

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.055	01:11
Breath Volume		2.339	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.050	01:14
Breath Volume		1.617	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAILLY,JOEL,N
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: MAILLY,JOEL
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: PRIDGEON

DATE/TIME OF TEST: 08/12/2020 21:51:26

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.079	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.139	21:46
Breath Volume		1.269	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.140	21:49
Breath Volume		2.480	21:49
Air Blank		0.000	21:50
Control Test		0.079	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDLE,CHARLES,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 08/13/2020 20:52:48

Diagnostics Check		OK	20:48	
Air Blank		0.000	20:48	
Control Test		0.079	20:49	
Air Blank		0.000	20:49	
Subject Sample	# 1		20:51	RFI Detect
Breath Volume		0.000	20:51	
Air Blank		0.000	20:51	
Control Test		0.079	20:52	
Air Blank		0.000	20:52	
Diagnostics Check		OK	20:52	

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,STEPHEN,A
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SCHRAM,AMANDA
ARRESTING OFFICER AGENCY: Taylor County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 08/16/2020 21:46:58

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.080	21:39
Air Blank		0.000	21:40
Subject Sample	# 1	0.140	21:42
Breath Volume		2.199	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:44
Subject Sample	# 2	0.141	21:45
Breath Volume		1.800	21:45
Air Blank		0.000	21:46
Control Test		0.078	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,LIN,
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: GRAY,LIN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: CEPHUS

DATE/TIME OF TEST: 08/19/2020 18:01:12

Diagnostics Check		OK	17:55
Air Blank		0.000	17:55
Control Test		0.080	17:56
Air Blank		0.000	17:56
Subject Sample	# 1	0.000	17:56
Breath Volume		3.152	17:56
Air Blank		0.000	17:57
Air Blank		0.000	17:59
Subject Sample	# 2	0.000	17:59
Breath Volume		2.117	17:59
Air Blank		0.000	18:00
Control Test		0.079	18:00
Air Blank		0.000	18:01
Diagnostics Check		OK	18:01

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,KOVEN,
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: PHELPS,KOVEN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/28/2020 23:58:35

Diagnostics Check	OK	23:57	
Air Blank		23:58	RFI Detect
Air Blank	0.000	23:58	

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,KOVEN,
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: PHELPS,KOVEN
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/29/2020 00:17:02

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:02	
Control Test		0.079	00:03	
Air Blank		0.000	00:03	
Subject Sample	# 1	0.073	00:04	
Breath Volume		2.054	00:04	
Air Blank		0.000	00:04	
Air Blank		0.000	00:06	
Subject Sample	# 2	0.064	00:09	Volume Not Met
Breath Volume		0.144	00:09	
Air Blank		0.000	00:10	
Air Blank		0.000	00:12	
Subject Sample	# 3	0.079	00:15	Volume Not Met
Breath Volume		0.089	00:15	
Air Blank		0.000	00:16	
Control Test		0.078	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 10/21/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,COLLEEN,E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 08/02/2020 00:42:07

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.000	00:37
Breath Volume		1.265	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:40
Subject Sample	# 2	0.000	00:40
Breath Volume		1.152	00:40
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 08/23/2020 23:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,COLLEEN,E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WHITLOCK,DANIEL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SANCHEZ-PEREZ

DATE/TIME OF TEST: 08/02/2020 01:19:54

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.081	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.205	01:14
Breath Volume		1.746	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.199	01:18
Breath Volume		1.343	01:18
Air Blank		0.000	01:18
Control Test		0.077	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 08/23/2020 23:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: QUINONEZ

DATE/TIME OF TEST: 08/06/2020 01:12:30

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.211	01:07
Breath Volume		1.457	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.214	01:10
Breath Volume		1.234	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 08/23/2020 23:57

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELM,DEREK,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HELM,DEREK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: WEBSTER

DATE/TIME OF TEST: 08/08/2020 21:28:41

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.080	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.125	21:24
Breath Volume		1.757	21:24
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.134	21:27
Breath Volume		1.925	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 08/09/2020 03:34:40

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.201	03:29
Breath Volume		2.597	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.200	03:32
Breath Volume		2.554	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: PRENTICE

DATE/TIME OF TEST: 08/09/2020 20:37:44

Diagnostics Check		OK	20:34	
Air Blank		0.000	20:35	
Control Test		0.080	20:35	
Air Blank		0.000	20:36	
Subject Sample	# 1		20:36	Subject Test Refused
Breath Volume		0.000	20:36	
Air Blank		0.000	20:36	
Control Test		0.080	20:37	
Air Blank		0.000	20:37	
Diagnostics Check		OK	20:37	

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BRAIDE

DATE/TIME OF TEST: 08/09/2020 23:55:11

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.195	23:47
Breath Volume		1.585	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.193	23:53
Breath Volume		1.414	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COURINGTON,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 08/10/2020 01:33:00

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.079	01:30	
Air Blank		0.000	01:31	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Control Test		0.080	01:32	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TOCCO

DATE/TIME OF TEST: 08/14/2020 23:08:05

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.081	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.183	23:03
Breath Volume		1.531	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:05
Subject Sample	# 2	0.189	23:06
Breath Volume		2.003	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:08

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 08/15/2020 22:39:44

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:37	
Control Test		0.080	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:38	Subject Test Refused
Breath Volume		0.000	22:38	
Air Blank		0.000	22:38	
Control Test		0.080	22:39	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 08/16/2020 00:54:51

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.185	00:50
Breath Volume		2.574	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.179	00:53
Breath Volume		2.460	00:53
Air Blank		0.000	00:53
Control Test		0.079	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SAVERINO,G
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 08/17/2020 23:42:10

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:41	
Control Test		0.080	23:41	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,D,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LEWIS,S
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LEVEROCK

DATE/TIME OF TEST: 08/19/2020 05:22:20

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.040	05:17
Breath Volume		2.406	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.039	05:20
Breath Volume		2.234	05:20
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 08/19/2020 16:30:41

Diagnostics Check		OK	16:23
Air Blank		0.000	16:23
Control Test		0.080	16:24
Air Blank		0.000	16:24
Subject Sample	# 1	0.205	16:25
Breath Volume		1.328	16:25
Air Blank		0.000	16:26
Air Blank		0.000	16:28
Subject Sample	# 2	0.207	16:28
Breath Volume		1.921	16:28
Air Blank		0.000	16:29
Control Test		0.079	16:30
Air Blank		0.000	16:30
Diagnostics Check		OK	16:30

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 08/23/2020 20:17:23

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.080	20:15	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:16	Subject Test Refused
Breath Volume		0.000	20:16	
Air Blank		0.000	20:16	
Control Test		0.081	20:16	
Air Blank		0.000	20:17	
Diagnostics Check		OK	20:17	

Data Download Date/Time: 08/23/2020 23:58

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,COLLEEN,E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BELFRY,NICOLAS
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: ARANDA

DATE/TIME OF TEST: 08/27/2020 22:06:00

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:49	
Control Test		0.081	21:50	
Air Blank		0.000	21:50	
Subject Sample	# 1	0.194	21:53	Volume Not Met
Breath Volume		0.988	21:53	
Air Blank		0.000	21:54	
Air Blank		0.000	21:56	
Subject Sample	# 2	0.197	21:58	
Breath Volume		1.406	21:58	
Air Blank		0.000	21:59	
Air Blank		0.000	22:01	
Subject Sample	# 3	0.192	22:04	Volume Not Met
Breath Volume		0.828	22:04	
Air Blank		0.000	22:05	
Control Test		0.077	22:05	
Air Blank		0.000	22:05	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 09/17/2020 13:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOAGLAN,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 08/28/2020 14:50:46

Diagnostics Check		OK	14:42
Air Blank		0.000	14:42
Control Test		0.079	14:43
Air Blank		0.000	14:43
Subject Sample	# 1	0.000	14:45
Breath Volume		1.187	14:45
Air Blank		0.000	14:46
Air Blank		0.000	14:48
Subject Sample	# 2	0.000	14:49
Breath Volume		1.488	14:49
Air Blank		0.000	14:49
Control Test		0.080	14:50
Air Blank		0.000	14:50
Diagnostics Check		OK	14:50

Data Download Date/Time: 09/17/2020 13:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOTCHKISS,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CHANEY

DATE/TIME OF TEST: 08/28/2020 18:43:13

Diagnostics Check		OK	18:40	
Air Blank		0.000	18:40	
Control Test		0.080	18:41	
Air Blank		0.000	18:41	
Subject Sample	# 1		18:41	Subject Test Refused
Breath Volume		0.000	18:41	
Air Blank		0.000	18:42	
Control Test		0.080	18:42	
Air Blank		0.000	18:43	
Diagnostics Check		OK	18:43	

Data Download Date/Time: 09/17/2020 13:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WESTON,EVAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 08/29/2020 21:38:07

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.081	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.196	21:33
Breath Volume		1.496	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:35
Subject Sample	# 2	0.186	21:36
Breath Volume		1.410	21:36
Air Blank		0.000	21:37
Control Test		0.078	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 09/17/2020 13:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 08/31/2020 23:49:40

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:47	
Control Test		0.081	23:47	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:48	Subject Test Refused
Breath Volume		0.000	23:48	
Air Blank		0.000	23:48	
Control Test		0.080	23:49	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 09/17/2020 13:23

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,C,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: RENAKER

DATE/TIME OF TEST: 08/02/2020 00:00:24

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Control Test		0.079	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,C,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/02/2020 05:32:55

Diagnostics Check	OK	05:32	
Air Blank		05:32	RFI Detect
Air Blank	0.000	05:32	

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFRTY,C,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/02/2020 05:41:21

Diagnostics Check		OK	05:34
Air Blank		0.000	05:35
Control Test		0.079	05:35
Air Blank		0.000	05:36
Subject Sample	# 1	0.136	05:36
Breath Volume		1.152	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.144	05:39
Breath Volume		1.894	05:39
Air Blank		0.000	05:40
Control Test		0.079	05:40
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: NARAISAN

DATE/TIME OF TEST: 08/15/2020 18:25:12

Diagnostics Check		OK	18:22	
Air Blank		0.000	18:22	
Control Test		0.078	18:23	
Air Blank		0.000	18:23	
Subject Sample	# 1		18:23	RFI Detect
Breath Volume		0.000	18:23	
Air Blank		0.000	18:24	
Control Test		0.079	18:24	
Air Blank		0.000	18:25	
Diagnostics Check		OK	18:25	

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: NARAISAN

DATE/TIME OF TEST: 08/15/2020 18:34:54

Diagnostics Check		OK	18:29
Air Blank		0.000	18:29
Control Test		0.079	18:29
Air Blank		0.000	18:30
Subject Sample	# 1	0.000	18:30
Breath Volume		1.695	18:30
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.000	18:33
Breath Volume		2.398	18:33
Air Blank		0.000	18:34
Control Test		0.079	18:34
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HORNE,S
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 08/15/2020 23:58:37

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:58	
		23:58	RFI Detect
Air Blank		23:58	RFI Detect

Data Download Date/Time: 08/27/2020 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HORNE,S
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 08/16/2020 00:07:13

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.160	00:02
Breath Volume		2.750	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.160	00:05
Breath Volume		2.550	00:05
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/16/2020 01:17:03

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.127	01:12
Breath Volume		2.898	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.126	01:15
Breath Volume		2.910	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:17

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HARRINGTON,A
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: PETRONE

DATE/TIME OF TEST: 08/17/2020 03:10:37

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.208	03:06
Breath Volume		1.722	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.201	03:09
Breath Volume		1.285	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 08/27/2020 12:21

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 08/30/2020 04:10:23

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.129	04:06
Breath Volume		2.496	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.121	04:08
Breath Volume		2.476	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 09/29/2020 10:22

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,ALAN,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: FRANZ

DATE/TIME OF TEST: 08/28/2020 19:10:11

Diagnostics Check		OK	19:03
Air Blank		0.000	19:04
Control Test		0.080	19:04
Air Blank		0.000	19:05
Subject Sample	# 1	0.274	19:05
Breath Volume		2.402	19:05
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.267	19:08
Breath Volume		2.214	19:08
Air Blank		0.000	19:09
Control Test		0.079	19:09
Air Blank		0.000	19:10
Diagnostics Check		OK	19:10

Data Download Date/Time: 09/25/2020 09:36

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,ALAN,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 08/02/2020 23:38:51

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.078	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.188	23:34
Breath Volume		2.285	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.187	23:37
Breath Volume		2.417	23:37
Air Blank		0.000	23:37
Control Test		0.078	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 08/26/2020 21:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 08/09/2020 13:42:01

Diagnostics Check		OK	13:35
Air Blank		0.000	13:36
Control Test		0.078	13:36
Air Blank		0.000	13:37
Subject Sample	# 1	0.000	13:37
Breath Volume		1.730	13:37
Air Blank		0.000	13:38
Air Blank		0.000	13:40
Subject Sample	# 2	0.000	13:40
Breath Volume		2.785	13:40
Air Blank		0.000	13:41
Control Test		0.078	13:41
Air Blank		0.000	13:41
Diagnostics Check		OK	13:42

Data Download Date/Time: 08/26/2020 21:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GILBERT,JASON
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/15/2020 05:02:06

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:50	
Control Test		0.078	04:51	
Air Blank		0.000	04:51	
Subject Sample	# 1	0.244	04:54	Volume Not Met
Breath Volume		0.113	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:57	
Subject Sample	# 2	0.249	05:00	Volume Not Met
Breath Volume		0.000	05:00	
Air Blank		0.000	05:01	
Control Test		0.077	05:01	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 08/26/2020 21:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GILBERT,JASON
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/15/2020 05:12:29

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:08	
Control Test		0.075	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:10	Subject Test Refused
Breath Volume		0.058	05:10	
Air Blank		0.000	05:11	
Control Test		0.077	05:11	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 08/26/2020 21:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GARZARO PAZ

DATE/TIME OF TEST: 08/17/2020 02:13:30

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.078	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.184	02:08
Breath Volume		2.074	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.187	02:11
Breath Volume		2.230	02:11
Air Blank		0.000	02:12
Control Test		0.078	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 08/26/2020 21:25

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: MAJERAN

DATE/TIME OF TEST: 08/05/2020 04:30:34

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.077	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.202	04:26
Breath Volume		1.449	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.198	04:29
Breath Volume		1.605	04:29
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 08/05/2020 23:35

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MAHANEY,JASON
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: FECHTENBURG

DATE/TIME OF TEST: 08/06/2020 20:04:42

Diagnostics Check		OK	19:57
Air Blank		0.000	19:58
Control Test		0.078	19:58
Air Blank		0.000	19:58
Subject Sample	# 1	0.160	20:00
Breath Volume		2.363	20:00
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.163	20:03
Breath Volume		1.710	20:03
Air Blank		0.000	20:03
Control Test		0.078	20:04
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUDAY,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BUDAY,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 08/06/2020 22:43:06

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:30	
Control Test		0.077	22:31	
Air Blank		0.000	22:31	
Subject Sample	# 1	0.218	22:34	Volume Not Met
Breath Volume		0.500	22:34	
Air Blank		0.000	22:35	
Air Blank		0.000	22:37	
Subject Sample	# 2	0.267	22:38	
Breath Volume		1.707	22:38	
Air Blank		0.000	22:38	
Air Blank		0.000	22:40	
Subject Sample	# 3	0.262	22:41	
Breath Volume		1.222	22:41	
Air Blank		0.000	22:42	
Control Test		0.077	22:42	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PEREZ,CONRAD
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 08/15/2020 03:40:29

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:38	
Control Test		0.078	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:39	Subject Test Refused
Breath Volume		0.000	03:39	
Air Blank		0.000	03:39	
Control Test		0.078	03:39	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WARDVALEZ

DATE/TIME OF TEST: 08/15/2020 04:35:38

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:32	
Control Test		0.078	04:32	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	
Control Test		0.078	04:35	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WILLERTON

DATE/TIME OF TEST: 08/16/2020 03:20:57

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.116	03:13
Breath Volume		1.132	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.147	03:16
Breath Volume		2.203	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:18
Subject Sample	# 3	0.150	03:19
Breath Volume		2.898	03:19
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 08/20/2020 21:27:34

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:25	
Control Test		0.077	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1		21:26	Subject Test Refused
Breath Volume		0.000	21:26	
Air Blank		0.000	21:26	
Control Test		0.077	21:26	
Air Blank		0.000	21:27	
Diagnostics Check		OK	21:27	

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BEAL,P
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 08/22/2020 03:46:03

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.077	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.099	03:41
Breath Volume		1.218	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.110	03:44
Breath Volume		3.039	03:44
Air Blank		0.000	03:45
Control Test		0.077	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:46

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LIPPINCOTT

DATE/TIME OF TEST: 08/23/2020 02:38:16

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.078	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.226	02:33
Breath Volume		2.621	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.234	02:36
Breath Volume		1.945	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: CRAWLEY

DATE/TIME OF TEST: 08/24/2020 14:24:58

Diagnostics Check		OK	14:18
Air Blank		0.000	14:19
Control Test		0.078	14:19
Air Blank		0.000	14:19
Subject Sample	# 1	0.013	14:20
Breath Volume		3.871	14:20
Air Blank		0.000	14:21
Air Blank		0.000	14:22
Subject Sample	# 2	0.011	14:23
Breath Volume		4.210	14:23
Air Blank		0.000	14:24
Control Test		0.078	14:24
Air Blank		0.000	14:24
Diagnostics Check		OK	14:24

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: GUY,KEVIN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 08/26/2020 02:25:36

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.078	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.000	02:21
Breath Volume		3.183	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.000	02:24
Breath Volume		1.507	02:24
Air Blank		0.000	02:24
Control Test		0.077	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: SMITH,ANGIE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/26/2020 20:41:31

Diagnostics Check		OK	20:34
Air Blank		0.000	20:34
Control Test		0.078	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.095	20:36
Breath Volume		1.570	20:36
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.101	20:40
Breath Volume		1.546	20:40
Air Blank		0.000	20:40
Control Test		0.078	20:40
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 08/27/2020 01:14:15

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.144	01:09
Breath Volume		2.039	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.150	01:12
Breath Volume		1.570	01:12
Air Blank		0.000	01:13
Control Test		0.078	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBI,UZIEL,Y
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: DUBI,UZIEL
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: ALLIE

DATE/TIME OF TEST: 08/29/2020 20:20:09

Diagnostics Check		OK	20:13
Air Blank		0.000	20:14
Control Test		0.078	20:14
Air Blank		0.000	20:15
Subject Sample	# 1	0.225	20:15
Breath Volume		1.972	20:15
Air Blank		0.000	20:16
Air Blank		0.000	20:18
Subject Sample	# 2	0.225	20:18
Breath Volume		2.128	20:18
Air Blank		0.000	20:19
Control Test		0.077	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 09/04/2020 18:00

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 08/07/2020 01:58:00

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.182	01:53
Breath Volume		1.546	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:55
Subject Sample	# 2	0.174	01:56
Breath Volume		1.886	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:58

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 08/07/2020 22:40:31

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.080	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.233	22:35
Breath Volume		3.296	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.228	22:38
Breath Volume		2.253	22:38
Air Blank		0.000	22:39
Control Test		0.079	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 08/15/2020 00:06:21

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.135	00:01
Breath Volume		2.855	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.137	00:04
Breath Volume		2.035	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/19/2020 22:57:03

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.063	22:52
Breath Volume		4.230	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:54
Subject Sample	# 2	0.054	22:55
Breath Volume		1.453	22:55
Air Blank		0.000	22:56
Control Test		0.081	22:56
Air Blank		0.000	22:56
Diagnostics Check		OK	22:57

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PADUANO,JOSEPH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 08/20/2020 22:52:44

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.081	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.149	22:48
Breath Volume		2.875	22:48
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.137	22:51
Breath Volume		1.484	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: KRAIN

DATE/TIME OF TEST: 08/21/2020 00:04:43

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.080	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.165	23:58
Breath Volume		1.648	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.174	00:03
Breath Volume		2.437	00:03
Air Blank		0.000	00:03
Control Test		0.079	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 08/28/2020 21:13:43

Diagnostics Check		OK	21:01	
Air Blank		0.000	21:02	
Control Test		0.079	21:02	
Air Blank		0.000	21:03	
Subject Sample	# 1	0.197	21:06	Volume Not Met
Breath Volume		0.957	21:06	
Air Blank		0.000	21:07	
Air Blank		0.000	21:08	
Subject Sample	# 2	0.230	21:12	Volume Not Met
Breath Volume		0.187	21:12	
Air Blank		0.000	21:12	
Control Test		0.077	21:13	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KRUSE,KRIS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 08/29/2020 02:00:10

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.230	01:54
Breath Volume		2.515	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.221	01:58
Breath Volume		1.871	01:58
Air Blank		0.000	01:59
Control Test		0.078	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 09/04/2020 17:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 08/01/2020 00:02:54

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.000	23:58
Breath Volume		1.964	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:01
Breath Volume		1.554	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 08/07/2020 17:36

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: COFFIE

DATE/TIME OF TEST: 08/01/2020 02:43:36

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.079	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.158	02:39
Breath Volume		1.550	02:39
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.148	02:42
Breath Volume		2.906	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 08/07/2020 17:36

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LENFANT,BENJAMINE
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/14/2020 01:21:13

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.080	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Control Test		0.080	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 09/28/2020 18:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LENFANT,BENJAMINE
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/14/2020 01:32:04

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.194	01:27
Breath Volume		2.167	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.192	01:30
Breath Volume		2.019	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:32

Data Download Date/Time: 09/28/2020 18:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SANVILLE

DATE/TIME OF TEST: 08/30/2020 03:36:05

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:35	
Control Test		0.081	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 09/28/2020 18:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: JURCZAK

DATE/TIME OF TEST: 08/08/2020 02:38:28

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.077	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1	0.185	02:32	
Breath Volume		4.085	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2	0.179	02:36	
Breath Volume		1.699	02:36	
Air Blank		0.000	02:37	
Control Test		0.074	02:37	Control Outside Tolerance
Air Blank		0.000	02:38	

Data Download Date/Time: 08/27/2020 12:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JURCZAK

DATE/TIME OF TEST: 08/08/2020 03:21:43

Diagnostics Check	OK	03:21	
Air Blank		03:21	RFI Detect
Air Blank		03:21	RFI Detect

Data Download Date/Time: 08/27/2020 12:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JURCZAK

DATE/TIME OF TEST: 08/08/2020 03:24:23

Diagnostics Check	OK	03:23	
Air Blank		03:23	RFI Detect
Air Blank	0.000	03:24	

Data Download Date/Time: 08/27/2020 12:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 07/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181692
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JURCZAK

DATE/TIME OF TEST: 08/08/2020 03:44:35

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:28	
Control Test		0.077	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1	0.166	03:32	
Breath Volume		1.898	03:32	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 2		03:37	No Sample Provided
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 3	0.153	03:42	
Breath Volume		1.832	03:42	
Air Blank		0.000	03:43	
Control Test		0.077	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 08/27/2020 12:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GURRIERI,JOHN,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GURRIERI,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/27/2020 21:44:31

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.079	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.217	21:40
Breath Volume		2.445	21:40
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.207	21:42
Breath Volume		1.281	21:42
Air Blank		0.000	21:43
Control Test		0.078	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDETTE,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FREDETTE,MICHAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PERK

DATE/TIME OF TEST: 08/28/2020 00:23:36

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.079	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:21	Slope Not Met
Breath Volume		1.488	00:21	
Air Blank		0.000	00:22	
Control Test		0.077	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDETTE,MICHAEL,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FREDETTE,MICHAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PERK

DATE/TIME OF TEST: 08/28/2020 00:32:46

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.378	00:27
Breath Volume		4.257	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.379	00:31
Breath Volume		4.105	00:31
Air Blank		0.000	00:31
Control Test		0.077	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,SHELBY,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HECKMAN,SHELBY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: WEFTFALL

DATE/TIME OF TEST: 08/29/2020 19:43:52

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.080	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.201	19:39
Breath Volume		3.644	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:41
Subject Sample	# 2	0.198	19:42
Breath Volume		2.882	19:42
Air Blank		0.000	19:42
Control Test		0.079	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 08/29/2020 20:10:52

Diagnostics Check		OK	19:59	
Air Blank		0.000	19:59	
Control Test		0.080	20:00	
Air Blank		0.000	20:00	
Subject Sample	# 1		20:03	No Sample Provided
Breath Volume		0.000	20:03	
Air Blank		0.000	20:04	
Air Blank		0.000	20:06	
Subject Sample	# 2		20:09	No Sample Provided
Breath Volume		0.000	20:09	
Air Blank		0.000	20:09	
Control Test		0.079	20:10	
Air Blank		0.000	20:10	
Diagnostics Check		OK	20:10	

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LYNCH,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: LIBBY

DATE/TIME OF TEST: 08/29/2020 20:29:58

Diagnostics Check		OK	20:23
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:24
Subject Sample	# 1	0.172	20:25
Breath Volume		1.652	20:25
Air Blank		0.000	20:26
Air Blank		0.000	20:27
Subject Sample	# 2	0.179	20:28
Breath Volume		1.675	20:28
Air Blank		0.000	20:29
Control Test		0.079	20:29
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANTON,CRYSTAL,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: KERN,KYLE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 08/30/2020 00:09:11

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.080	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.155	00:04
Breath Volume		2.316	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.157	00:07
Breath Volume		1.156	00:07
Air Blank		0.000	00:08
Control Test		0.076	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,A,
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,A
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: KARPOVICH

DATE/TIME OF TEST: 08/30/2020 00:25:03

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.148	00:20
Breath Volume		2.000	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.144	00:23
Breath Volume		1.738	00:23
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:25

Data Download Date/Time: 09/29/2020 10:19

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 08/05/2020 23:09:04

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.000	23:04
Breath Volume		2.863	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.000	23:07
Breath Volume		1.500	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:09

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: BRYSON

DATE/TIME OF TEST: 08/06/2020 06:39:08

Diagnostics Check		OK	06:32
Air Blank		0.000	06:33
Control Test		0.079	06:33
Air Blank		0.000	06:34
Subject Sample	# 1	0.088	06:34
Breath Volume		1.578	06:34
Air Blank		0.000	06:35
Air Blank		0.000	06:36
Subject Sample	# 2	0.089	06:37
Breath Volume		1.808	06:37
Air Blank		0.000	06:38
Control Test		0.079	06:38
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 08/13/2020 01:55:29

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.079	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1	0.206	01:44	
Breath Volume		1.359	01:44	
Air Blank		0.000	01:44	
Air Blank		0.000	01:46	
Subject Sample	# 2	0.182	01:49	Volume Not Met
Breath Volume		0.519	01:49	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 3	0.196	01:53	
Breath Volume		1.292	01:53	
Air Blank		0.000	01:54	
Control Test		0.076	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RUIZ,J
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: MAC GILVRAY

DATE/TIME OF TEST: 08/13/2020 20:34:01

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.079	20:28
Air Blank		0.000	20:29
Subject Sample	# 1	0.172	20:29
Breath Volume		3.382	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.172	20:32
Breath Volume		3.675	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:34

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 08/14/2020 03:02:41

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.145	02:58
Breath Volume		2.203	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.144	03:01
Breath Volume		2.320	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MAGALHAES VIEIRA

DATE/TIME OF TEST: 08/14/2020 23:12:01

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.079	23:05
Air Blank		0.000	23:06
Subject Sample	# 1	0.094	23:06
Breath Volume		1.828	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.088	23:10
Breath Volume		1.312	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:12

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: AMPEREZ

DATE/TIME OF TEST: 08/15/2020 00:21:26

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.079	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.105	00:16
Breath Volume		2.347	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.103	00:19
Breath Volume		3.953	00:19
Air Blank		0.000	00:20
Control Test		0.078	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PERUSKI

DATE/TIME OF TEST: 08/15/2020 23:08:15

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.130	23:03
Breath Volume		3.027	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.127	23:06
Breath Volume		1.683	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: MATOS BAEZ

DATE/TIME OF TEST: 08/16/2020 07:09:37

Diagnostics Check		OK	07:02
Air Blank		0.000	07:02
Control Test		0.079	07:03
Air Blank		0.000	07:03
Subject Sample	# 1	0.112	07:04
Breath Volume		1.324	07:04
Air Blank		0.000	07:05
Air Blank		0.000	07:07
Subject Sample	# 2	0.117	07:08
Breath Volume		2.238	07:08
Air Blank		0.000	07:08
Control Test		0.079	07:09
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ARAGON

DATE/TIME OF TEST: 08/17/2020 00:03:01

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:58
Breath Volume		2.933	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:01
Breath Volume		2.164	00:01
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:03

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: HERNANDEZ II

DATE/TIME OF TEST: 08/17/2020 23:46:44

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.222	23:38
Breath Volume		1.214	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.250	23:42
Breath Volume		2.898	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 3	0.227	23:45
Breath Volume		1.480	23:45
Air Blank		0.000	23:45
Control Test		0.078	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 08/18/2020 17:06:43

Diagnostics Check		OK	17:00
Air Blank		0.000	17:00
Control Test		0.080	17:01
Air Blank		0.000	17:01
Subject Sample	# 1	0.067	17:02
Breath Volume		1.351	17:02
Air Blank		0.000	17:02
Air Blank		0.000	17:04
Subject Sample	# 2	0.066	17:05
Breath Volume		1.625	17:05
Air Blank		0.000	17:05
Control Test		0.079	17:06
Air Blank		0.000	17:06
Diagnostics Check		OK	17:06

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: OCHOA-GALLEGOS

DATE/TIME OF TEST: 08/19/2020 02:12:24

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.229	02:07
Breath Volume		3.230	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.227	02:10
Breath Volume		2.855	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: VIENS

DATE/TIME OF TEST: 08/19/2020 23:35:55

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.202	23:28
Breath Volume		2.296	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.176	23:31
Breath Volume		1.703	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:33
Subject Sample	# 3	0.188	23:34
Breath Volume		1.683	23:34
Air Blank		0.000	23:35
Control Test		0.078	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: REID,W
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: BROUILLARD

DATE/TIME OF TEST: 08/20/2020 22:37:20

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.200	22:32
Breath Volume		1.316	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.202	22:35
Breath Volume		1.574	22:35
Air Blank		0.000	22:36
Control Test		0.079	22:36
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 08/21/2020 02:42:24

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.120	02:37
Breath Volume		1.488	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.128	02:40
Breath Volume		2.187	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PEELER

DATE/TIME OF TEST: 08/21/2020 22:39:04

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.078	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:37	Subject Test Refused
Breath Volume		0.000	22:37	
Air Blank		0.000	22:38	
Control Test		0.078	22:38	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 08/21/2020 23:20:08

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.078	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.125	23:15
Breath Volume		3.199	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.124	23:18
Breath Volume		2.457	23:18
Air Blank		0.000	23:19
Control Test		0.078	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 08/22/2020 01:31:00

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.150	01:26
Breath Volume		3.500	01:26
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.146	01:29
Breath Volume		3.820	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:31

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GOMEZ SANCHEZ

DATE/TIME OF TEST: 08/22/2020 03:30:47

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.078	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.079	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 09/02/2020 18:37

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RENDON,R
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 08/22/2020 21:27:41

Diagnostics Check		OK	21:21
Air Blank		0.000	21:21
Control Test		0.079	21:22
Air Blank		0.000	21:22
Subject Sample	# 1	0.198	21:23
Breath Volume		2.519	21:23
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.200	21:26
Breath Volume		2.679	21:26
Air Blank		0.000	21:26
Control Test		0.078	21:27
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 08/23/2020 22:24:51

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.170	22:20
Breath Volume		4.128	22:20
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.166	22:23
Breath Volume		3.187	22:23
Air Blank		0.000	22:23
Control Test		0.079	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: BAY

DATE/TIME OF TEST: 08/24/2020 00:52:33

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.000	00:48
Breath Volume		1.324	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:51
Breath Volume		1.527	00:51
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: ALMOND

DATE/TIME OF TEST: 08/24/2020 20:33:23

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.079	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.186	20:28
Breath Volume		2.242	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.174	20:31
Breath Volume		1.921	20:31
Air Blank		0.000	20:32
Control Test		0.079	20:32
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: FERGUSON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: STALTER

DATE/TIME OF TEST: 08/25/2020 00:09:43

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.193	00:04
Breath Volume		1.781	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.194	00:08
Breath Volume		2.074	00:08
Air Blank		0.000	00:08
Control Test		0.078	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MALOY

DATE/TIME OF TEST: 08/25/2020 02:47:29

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.281	02:43
Breath Volume		1.910	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.264	02:45
Breath Volume		1.867	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PEAN

DATE/TIME OF TEST: 08/26/2020 00:03:44

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.079	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.181	23:58
Breath Volume		1.242	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.162	00:02
Breath Volume		1.105	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FERRARA

DATE/TIME OF TEST: 08/27/2020 03:49:37

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.092	03:45
Breath Volume		1.175	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.091	03:48
Breath Volume		1.132	03:48
Air Blank		0.000	03:48
Control Test		0.080	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LEE JR

DATE/TIME OF TEST: 08/28/2020 00:14:29

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.080	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.192	00:09
Breath Volume		2.167	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.199	00:13
Breath Volume		2.437	00:13
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 08/28/2020 01:26:25

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:15	
Control Test		0.079	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1	0.112	01:19	Volume Not Met
Breath Volume		0.234	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2	0.110	01:24	Volume Not Met
Breath Volume		0.269	01:24	
Air Blank		0.000	01:25	
Control Test		0.079	01:25	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BLANCHARD,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 08/28/2020 03:27:37

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:25	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Control Test		0.080	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WASHINGTON,DESMAND
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: WAKIN

DATE/TIME OF TEST: 08/28/2020 08:21:44

Diagnostics Check		OK	08:14
Air Blank		0.000	08:14
Control Test		0.079	08:14
Air Blank		0.000	08:15
Subject Sample	# 1	0.249	08:16
Breath Volume		1.851	08:16
Air Blank		0.000	08:17
Air Blank		0.000	08:19
Subject Sample	# 2	0.235	08:20
Breath Volume		1.851	08:20
Air Blank		0.000	08:20
Control Test		0.079	08:21
Air Blank		0.000	08:21
Diagnostics Check		OK	08:21

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 08/28/2020 21:58:20

Diagnostics Check		OK	21:50
Air Blank		0.000	21:50
Control Test		0.079	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.190	21:53
Breath Volume		1.863	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.185	21:56
Breath Volume		1.578	21:56
Air Blank		0.000	21:57
Control Test		0.078	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: FAKESS

DATE/TIME OF TEST: 08/29/2020 03:30:09

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.125	03:25
Breath Volume		1.671	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.121	03:28
Breath Volume		1.507	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: LANDRY,L
ARRESTING OFFICER AGENCY: TEMPLE TERRACE PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: DORVILUS

DATE/TIME OF TEST: 08/29/2020 04:50:01

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.079	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.146	04:45
Breath Volume		3.453	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.145	04:48
Breath Volume		2.761	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:50

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ODONNELL,T
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:02

SUBJECT'S LAST NAME: WESTON

DATE/TIME OF TEST: 08/29/2020 08:28:46

Diagnostics Check		OK	08:22
Air Blank		0.000	08:23
Control Test		0.079	08:23
Air Blank		0.000	08:23
Subject Sample	# 1	0.201	08:24
Breath Volume		1.718	08:24
Air Blank		0.000	08:25
Air Blank		0.000	08:26
Subject Sample	# 2	0.205	08:27
Breath Volume		1.285	08:27
Air Blank		0.000	08:27
Control Test		0.079	08:28
Air Blank		0.000	08:28
Diagnostics Check		OK	08:28

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WILKINS,M
ARRESTING OFFICER AGENCY: FFWCC (SW REGION)
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/29/2020 19:55:35

Diagnostics Check		OK	19:49
Air Blank		0.000	19:49
Control Test		0.079	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.171	19:50
Breath Volume		3.378	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.172	19:54
Breath Volume		3.511	19:54
Air Blank		0.000	19:54
Control Test		0.080	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 08/30/2020 01:36:14

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.080	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.077	01:31
Breath Volume		2.109	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.076	01:34
Breath Volume		1.855	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 08/30/2020 02:43:05

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.000	02:38
Breath Volume		1.230	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.000	02:41
Breath Volume		1.156	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: BIVINS

DATE/TIME OF TEST: 08/30/2020 04:43:14

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.080	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.170	04:38
Breath Volume		3.566	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.167	04:41
Breath Volume		3.691	04:41
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: OQUENDO

DATE/TIME OF TEST: 08/30/2020 05:44:35

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.079	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.000	05:40
Breath Volume		2.875	05:40
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.000	05:43
Breath Volume		3.144	05:43
Air Blank		0.000	05:43
Control Test		0.080	05:44
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: STENBERG

DATE/TIME OF TEST: 08/30/2020 06:29:02

Diagnostics Check		OK	06:16	
Air Blank		0.000	06:16	
Control Test		0.080	06:17	
Air Blank		0.000	06:17	
Subject Sample	# 1	0.170	06:20	Volume Not Met
Breath Volume		0.718	06:20	
Air Blank		0.000	06:21	
Air Blank		0.000	06:23	
Subject Sample	# 2	0.175	06:24	
Breath Volume		1.261	06:24	
Air Blank		0.000	06:24	
Air Blank		0.000	06:26	
Subject Sample	# 3	0.173	06:27	
Breath Volume		1.171	06:27	
Air Blank		0.000	06:28	
Control Test		0.079	06:28	
Air Blank		0.000	06:28	
Diagnostics Check		OK	06:29	

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: FAYAD,J
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: CARVALHO

DATE/TIME OF TEST: 08/30/2020 16:45:55

Diagnostics Check		OK	16:28	
Air Blank		0.000	16:29	
Control Test		0.080	16:29	
Air Blank		0.000	16:30	
Subject Sample	# 1	0.000	16:33	Volume Not Met
Breath Volume		0.644	16:33	
Air Blank		0.000	16:33	
Air Blank		0.000	16:35	
Subject Sample	# 2	0.000	16:38	
Breath Volume		1.214	16:38	
Air Blank		0.000	16:39	
Air Blank		0.000	16:41	
Subject Sample	# 3	0.000	16:44	Volume Not Met
Breath Volume		0.863	16:44	
Air Blank		0.000	16:45	
Control Test		0.080	16:45	
Air Blank		0.000	16:45	
Diagnostics Check		OK	16:45	

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: FAYAD,J
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: CARVALHO

DATE/TIME OF TEST: 08/30/2020 16:59:58

Diagnostics Check		OK	16:48	
Air Blank		0.000	16:49	
Control Test		0.079	16:49	
Air Blank		0.000	16:49	
Subject Sample	# 1	0.000	16:53	Volume Not Met
Breath Volume		0.285	16:53	
Air Blank		0.000	16:53	
Air Blank		0.000	16:55	
Subject Sample	# 2	0.000	16:58	Volume Not Met
Breath Volume		0.488	16:58	
Air Blank		0.000	16:59	
Control Test		0.080	16:59	
Air Blank		0.000	16:59	
Diagnostics Check		OK	16:59	

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: FAYAD,J
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: CARAVLHO

DATE/TIME OF TEST: 08/30/2020 17:08:19

Diagnostics Check		OK	17:03	
Air Blank		0.000	17:03	
Control Test		0.079	17:04	
Air Blank		0.000	17:04	
Subject Sample	# 1		17:06	Subject Test Refused
Breath Volume		0.660	17:06	
Air Blank		0.000	17:07	
Control Test		0.080	17:07	
Air Blank		0.000	17:08	
Diagnostics Check		OK	17:08	

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: AYERS,M
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: LAY

DATE/TIME OF TEST: 08/30/2020 17:36:48

Diagnostics Check		OK	17:30
Air Blank		0.000	17:30
Control Test		0.079	17:31
Air Blank		0.000	17:31
Subject Sample	# 1	0.258	17:32
Breath Volume		2.253	17:32
Air Blank		0.000	17:32
Air Blank		0.000	17:34
Subject Sample	# 2	0.264	17:35
Breath Volume		2.175	17:35
Air Blank		0.000	17:35
Control Test		0.079	17:36
Air Blank		0.000	17:36
Diagnostics Check		OK	17:36

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HEIMER

DATE/TIME OF TEST: 08/31/2020 19:16:37

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.079	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.000	19:12
Breath Volume		2.589	19:12
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.000	19:15
Breath Volume		1.386	19:15
Air Blank		0.000	19:15
Control Test		0.079	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 09/02/2020 18:38

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GANO

DATE/TIME OF TEST: 08/01/2020 03:59:12

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.109	03:54
Breath Volume		2.199	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.108	03:57
Breath Volume		2.062	03:57
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 08/11/2020 13:54

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CONTI

DATE/TIME OF TEST: 08/02/2020 00:37:24

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.079	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.151	00:32
Breath Volume		3.714	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.151	00:35
Breath Volume		3.761	00:35
Air Blank		0.000	00:36
Control Test		0.077	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 08/11/2020 13:55

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DAVID,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 08/30/2020 02:56:46

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.082	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.067	02:52
Breath Volume		2.730	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.064	02:55
Breath Volume		2.359	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 09/04/2020 09:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DAVID,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 08/30/2020 03:15:12

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.078	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.180	03:10
Breath Volume		1.968	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:12
Subject Sample	# 2	0.174	03:13
Breath Volume		1.367	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 09/04/2020 09:48

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 08/19/2020 21:36:49

Diagnostics Check	OK	21:36	
Air Blank		21:36	RFI Detect
Air Blank		21:36	RFI Detect

Data Download Date/Time: 09/04/2020 18:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 08/19/2020 21:40:23

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:39	RFI Detect
Breath Volume		0.000	21:39	
Air Blank			21:39	RFI Detect
Air Blank		0.000	21:40	

Data Download Date/Time: 09/04/2020 18:02

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 08/19/2020 21:43:50

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.077	21:42	
Air Blank		0.000	21:43	
Subject Sample	# 1		21:43	Sequence Aborted
Breath Volume		0.000	21:43	
Air Blank		0.000	21:43	

Data Download Date/Time: 09/04/2020 18:02

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 08/01/2020 22:03:06

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.081	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.207	21:58
Breath Volume		1.757	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.210	22:01
Breath Volume		2.054	22:01
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 08/05/2020 08:14

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KOONCE

DATE/TIME OF TEST: 08/02/2020 02:06:49

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.080	02:04	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:05	Subject Test Refused
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 08/12/2020 10:52

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO,MARK,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: OCHS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: TERSIGNI

DATE/TIME OF TEST: 08/06/2020 10:58:50

Diagnostics Check		OK	10:52
Air Blank		0.000	10:53
Control Test		0.080	10:53
Air Blank		0.000	10:53
Subject Sample	# 1	0.000	10:54
Breath Volume		2.175	10:54
Air Blank		0.000	10:54
Air Blank		0.000	10:56
Subject Sample	# 2	0.000	10:57
Breath Volume		2.187	10:57
Air Blank		0.000	10:57
Control Test		0.080	10:58
Air Blank		0.000	10:58
Diagnostics Check		OK	10:58

Data Download Date/Time: 08/12/2020 10:52

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SAMANTHA,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BROWN,SAMANTHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 08/08/2020 00:16:08

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.104	00:11
Breath Volume		1.875	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.102	00:14
Breath Volume		1.820	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 08/12/2020 10:52

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO PEREZ,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: COLFER

DATE/TIME OF TEST: 08/08/2020 19:14:31

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.080	19:08
Air Blank		0.000	19:09
Subject Sample	# 1	0.230	19:10
Breath Volume		3.988	19:10
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.231	19:13
Breath Volume		4.058	19:13
Air Blank		0.000	19:13
Control Test		0.080	19:13
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 08/12/2020 10:52

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PERIAM,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CENATIS

DATE/TIME OF TEST: 08/11/2020 05:38:40

Diagnostics Check		OK	05:32
Air Blank		0.000	05:32
Control Test		0.079	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.000	05:34
Breath Volume		2.687	05:34
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.000	05:37
Breath Volume		3.339	05:37
Air Blank		0.000	05:37
Control Test		0.080	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 08/12/2020 10:52

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 08/13/2020 00:50:04

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.079	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.052	00:45
Breath Volume		1.597	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.050	00:48
Breath Volume		1.660	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:50

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,MARTINI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TARNOR

DATE/TIME OF TEST: 08/14/2020 22:42:41

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.000	22:38
Breath Volume		1.437	22:38
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.000	22:41
Breath Volume		1.195	22:41
Air Blank		0.000	22:41
Control Test		0.080	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARRISON,W
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 08/15/2020 00:57:10

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:54	
Control Test		0.080	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Control Test		0.079	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LUKOWSKI

DATE/TIME OF TEST: 08/15/2020 03:06:27

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.155	02:58
Breath Volume		1.703	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.178	03:02
Breath Volume		2.207	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 3	0.161	03:05
Breath Volume		1.398	03:05
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GARAY

DATE/TIME OF TEST: 08/16/2020 03:40:08

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.181	03:35
Breath Volume		2.218	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.185	03:38
Breath Volume		2.292	03:38
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: AUBRY,MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: RADKE

DATE/TIME OF TEST: 08/16/2020 14:43:59

Diagnostics Check	OK	14:43	
Air Blank		14:43	RFI Detect
Air Blank	0.000	14:43	

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: AUBRY,MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: RADKE

DATE/TIME OF TEST: 08/16/2020 14:51:40

Diagnostics Check		OK	14:45
Air Blank		0.000	14:46
Control Test		0.080	14:46
Air Blank		0.000	14:46
Subject Sample	# 1	0.046	14:47
Breath Volume		2.023	14:47
Air Blank		0.000	14:47
Air Blank		0.000	14:49
Subject Sample	# 2	0.045	14:50
Breath Volume		2.269	14:50
Air Blank		0.000	14:50
Control Test		0.079	14:51
Air Blank		0.000	14:51
Diagnostics Check		OK	14:51

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ENGH,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HINGSON-JOHNSON

DATE/TIME OF TEST: 08/18/2020 02:30:38

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.164	02:23
Breath Volume		1.144	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.162	02:29
Breath Volume		2.277	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SAMANTHA,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SCIMECA

DATE/TIME OF TEST: 08/18/2020 05:02:30

Diagnostics Check		OK	04:50	
Air Blank		0.000	04:51	
Control Test		0.081	04:51	
Air Blank		0.000	04:51	
Subject Sample	# 1	0.000	04:55	Volume Not Met
Breath Volume		0.234	04:55	
Air Blank		0.000	04:55	
Air Blank		0.000	04:57	
Subject Sample	# 2	0.000	04:58	
Breath Volume		2.027	04:58	
Air Blank		0.000	04:58	
Air Blank		0.000	05:00	
Subject Sample	# 3	0.000	05:01	
Breath Volume		2.000	05:01	
Air Blank		0.000	05:01	
Control Test		0.080	05:01	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/22/2020 02:55:32

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.221	02:50
Breath Volume		1.824	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.216	02:54
Breath Volume		1.605	02:54
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/22/2020 22:18:14

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.079	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.000	22:13
Breath Volume		3.660	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.000	22:16
Breath Volume		3.730	22:16
Air Blank		0.000	22:17
Control Test		0.080	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HOWES,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: SWARTZ

DATE/TIME OF TEST: 08/23/2020 23:13:12

Diagnostics Check		OK	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.170	23:08
Breath Volume		1.960	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.169	23:11
Breath Volume		1.718	23:11
Air Blank		0.000	23:12
Control Test		0.079	23:12
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYES,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: ANGULO

DATE/TIME OF TEST: 08/25/2020 01:29:40

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.212	01:25
Breath Volume		2.281	01:25
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.211	01:28
Breath Volume		1.839	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 08/28/2020 23:42:58

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.000	23:38
Breath Volume		1.402	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.000	23:41
Breath Volume		1.789	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 08/29/2020 02:57:35

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:45	
Control Test		0.081	02:45	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:49	No Sample Provided
Breath Volume		0.000	02:49	
Air Blank		0.000	02:49	
Air Blank		0.000	02:51	
Subject Sample	# 2	0.160	02:52	
Breath Volume		2.050	02:52	
Air Blank		0.000	02:53	
Air Blank		0.000	02:55	
Subject Sample	# 3	0.164	02:56	
Breath Volume		1.492	02:56	
Air Blank		0.000	02:56	
Control Test		0.079	02:57	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: BRAMLETT

DATE/TIME OF TEST: 08/29/2020 22:32:54

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:29	
Control Test		0.079	22:29	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:31	Subject Test Refused
Breath Volume		0.000	22:31	
Air Blank		0.000	22:32	
Control Test		0.079	22:32	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 09/25/2020 09:34

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: BUREN

DATE/TIME OF TEST: 08/25/2020 23:07:14

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.079	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1	0.142	22:59	Volume Not Met
Breath Volume		0.339	22:59	
Air Blank		0.000	23:00	
Air Blank		0.000	23:01	
Subject Sample	# 2	0.224	23:02	
Breath Volume		1.761	23:02	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 3	0.220	23:05	
Breath Volume		1.898	23:05	
Air Blank		0.000	23:06	
Control Test		0.079	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 08/28/2020 12:53

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LAGASSE JR

DATE/TIME OF TEST: 08/30/2020 01:21:13

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.079	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:19	RFI Detect
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Control Test		0.079	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 09/30/2020 08:09

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LAGASSE

DATE/TIME OF TEST: 08/30/2020 01:32:00

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:24	
Control Test		0.080	01:24	
Air Blank		0.000	01:24	
Subject Sample	# 1		01:27	No Sample Provided
Breath Volume		0.000	01:27	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 2		01:30	Subject Test Refused
Breath Volume		0.000	01:30	
Air Blank		0.000	01:31	
Control Test		0.079	01:31	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 09/30/2020 08:09

Printed: 18 November 2020

COBRA ©1997-2014 CMI, Inc.