Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BARTLETT

DATE/TIME OF TEST: 09/03/2020 00:50:45

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.178	00:46
Breath Volume		3.972	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.174	00:49
Breath Volume		3.511	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: TRIBBLE

DATE/TIME OF TEST: 09/04/2020 02:22:19

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.081	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.219	02:17
Breath Volume		2.191	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.218	02:20
Breath Volume		1.984	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: PIDGEON,A
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: MEJIAS

DATE/TIME OF TEST: 09/04/2020 10:48:53

Diagnostics Check		OK	10:42
Air Blank		0.000	10:43
Control Test		0.081	10:43
Air Blank		0.000	10:44
Subject Sample	# 1	0.000	10:44
Breath Volume		3.851	10:44
Air Blank		0.000	10:45
Air Blank		0.000	10:47
Subject Sample	# 2	0.000	10:47
Breath Volume		3.968	10:47
Air Blank		0.000	10:47
Control Test		0.080	10:48
Air Blank		0.000	10:48
Diagnostics Check		OK	10:48

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMUELS,ANDREW,G
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/04/2020 23:43:06

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.209	23:38
Breath Volume		1.730	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.226	23:41
Breath Volume		1.441	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGANJ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LOGAN,JONATHAN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: KLEIBER

DATE/TIME OF TEST: 09/12/2020 21:38:56

Diagnostics Check		OK	21:27	
Air Blank		0.000	21:27	
Control Test		0.081	21:27	
Air Blank		0.000	21:28	
Subject Sample	# 1	0.158	21:31	Volume Not Met
Breath Volume		0.062	21:31	
Air Blank		0.000	21:32	
Air Blank		0.000	21:34	
Subject Sample	# 2	0.149	21:37	Volume Not Met
Breath Volume		0.269	21:37	
Air Blank		0.000	21:38	
Control Test		0.078	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,COLE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MCDONALD,COLE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 09/18/2020 03:20:01

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.185	03:15
Breath Volume		2.261	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:17
Subject Sample	# 2	0.184	03:18
Breath Volume		2.343	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:20

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: KAPLAN,M
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: EDMUNDSON

DATE/TIME OF TEST: 09/26/2020 02:50:53

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.153	02:46
Breath Volume		1.816	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.152	02:49
Breath Volume		1.808	02:49
Air Blank		0.000	02:49
Control Test		0.079	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING,CODY,R
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: CRONIN,S
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: TORTORICE

DATE/TIME OF TEST: 09/28/2020 13:52:03

Diagnostics Check		OK	13:45
Air Blank		0.000	13:46
Control Test		0.080	13:46
Air Blank		0.000	13:46
Subject Sample	# 1	0.000	13:47
Breath Volume		2.375	13:47
Air Blank		0.000	13:48
Air Blank		0.000	13:50
Subject Sample	# 2	0.000	13:50
Breath Volume		1.492	13:50
Air Blank		0.000	13:51
Control Test		0.079	13:51
Air Blank		0.000	13:51
Diagnostics Check		OK	13:52

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMUELS,ANDREW,G
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SAMUELS,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 09/28/2020 23:56:13

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:51
Breath Volume		2.441	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:54
Breath Volume		2.785	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,D,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MORGAN,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: KENT

DATE/TIME OF TEST: 09/30/2020 17:46:14

Diagnostics Check		OK	17:40
Air Blank		0.000	17:40
Control Test		0.081	17:41
Air Blank		0.000	17:41
Subject Sample	# 1	0.000	17:41
Breath Volume		3.332	17:41
Air Blank		0.000	17:42
Air Blank		0.000	17:44
Subject Sample	# 2	0.000	17:44
Breath Volume		3.644	17:44
Air Blank		0.000	17:45
Control Test		0.080	17:45
Air Blank		0.000	17:46
Diagnostics Check		OK	17:46

Data Download Date/Time: 10/02/2020 18:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CAMERON,DAVID
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PINCKNEY

DATE/TIME OF TEST: 09/27/2020 02:25:15

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.079	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.185	02:19
Breath Volume		1.542	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.199	02:23
Breath Volume		1.292	02:23
Air Blank		0.000	02:24
Control Test		0.077	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 10/13/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: HOSFELD

DATE/TIME OF TEST: 09/03/2020 17:50:56

Diagnostics Check		OK	17:48	
Air Blank		0.000	17:48	
Control Test		0.079	17:48	
Air Blank		0.000	17:49	
Subject Sample	# 1		17:49	Subject Test Refused
Breath Volume		0.000	17:49	
Air Blank		0.000	17:49	
Control Test		0.078	17:50	
Air Blank		0.000	17:50	
Diagnostics Check		OK	17:50	

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 09/03/2020 23:23:44

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.213	23:17
Breath Volume		1.777	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.195	23:22
Breath Volume		1.601	23:22
Air Blank		0.000	23:22
Control Test		0.077	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: FISHER,CLIFFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SURRAT

DATE/TIME OF TEST: 09/04/2020 03:32:09

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.078	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.144	03:27
Breath Volume		3.535	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.136	03:30
Breath Volume		3.320	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BAILEM

DATE/TIME OF TEST: 09/04/2020 22:41:53

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:39	
Control Test		0.078	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.079	22:41	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CROCKETT

DATE/TIME OF TEST: 09/04/2020 23:42:00

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.051	23:37
Breath Volume		1.390	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.057	23:40
Breath Volume		3.464	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:42

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/05/2020 00:48:17

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.196	00:43
Breath Volume		1.117	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.192	00:46
Breath Volume		1.277	00:46
Air Blank		0.000	00:47
Control Test		0.078	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BOLOGNA

DATE/TIME OF TEST: 09/05/2020 01:53:26

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.078	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.012	01:49
Breath Volume		2.058	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.011	01:52
Breath Volume		1.894	01:52
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ALVAREZ CONDE

DATE/TIME OF TEST: 09/05/2020 02:32:05

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.158	02:27
Breath Volume		1.824	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:29
Subject Sample	# 2	0.157	02:30
Breath Volume		1.597	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:32

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PERRETTI

DATE/TIME OF TEST: 09/05/2020 03:32:05

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Subject Sample	# 1	0.000	03:27
Breath Volume		1.140	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.000	03:30
Breath Volume		2.203	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:32

Data Download Date/Time: 09/08/2020 20:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: STATON

DATE/TIME OF TEST: 09/10/2020 01:17:15

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.079	01:15	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:16	
Control Test		0.078	01:16	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 10/07/2020 15:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: EL MALIKI

DATE/TIME OF TEST: 09/11/2020 01:38:08

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.079	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.414	01:36	
Air Blank		0.000	01:37	
Control Test		0.079	01:37	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 10/07/2020 15:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 09/11/2020 03:46:52

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.000	03:42
Breath Volume		3.875	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.000	03:45
Breath Volume		4.218	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 10/07/2020 15:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/12/2020 04:03:01

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.154	03:58
Breath Volume		1.750	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.156	04:01
Breath Volume		1.535	04:01
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:03

Data Download Date/Time: 10/07/2020 15:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,STEWART,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CAMPBELL,JAMES
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: MCCULLEY

DATE/TIME OF TEST: 09/03/2020 08:31:26

Diagnostics Check		OK	08:24
Air Blank		0.000	08:24
Control Test		0.078	08:24
Air Blank		0.000	08:25
Subject Sample	# 1	0.241	08:26
Breath Volume		4.339	08:26
Air Blank		0.000	08:27
Air Blank		0.000	08:28
Subject Sample	# 2	0.251	08:29
Breath Volume		3.687	08:29
Air Blank		0.000	08:30
Control Test		0.076	08:30
Air Blank		0.000	08:31
Diagnostics Check		OK	08:31

Data Download Date/Time: 09/08/2020 17:00

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 09/05/2020 03:23:00

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.186	03:18
Breath Volume		1.527	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:20
Subject Sample	# 2	0.196	03:21
Breath Volume		1.953	03:21
Air Blank		0.000	03:22
Control Test		0.075	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:23

Data Download Date/Time: 09/08/2020 17:00

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: COLACCHIO

DATE/TIME OF TEST: 09/11/2020 20:08:15

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.078	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.188	20:03
Breath Volume		2.421	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:05
Subject Sample	# 2	0.183	20:06
Breath Volume		1.601	20:06
Air Blank		0.000	20:07
Control Test		0.076	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 10/31/2020 03:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,STEWART,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLS,ANTONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:08

SUBJECT'S LAST NAME: CASHWELL

DATE/TIME OF TEST: 09/18/2020 11:39:45

Diagnostics Check		OK	11:33
Air Blank		0.000	11:34
Control Test		0.078	11:34
Air Blank		0.000	11:34
Subject Sample	# 1	0.000	11:35
Breath Volume		3.308	11:35
Air Blank		0.000	11:35
Air Blank		0.000	11:37
Subject Sample	# 2	0.000	11:38
Breath Volume		2.085	11:38
Air Blank		0.000	11:38
Control Test		0.078	11:39
Air Blank		0.000	11:39
Diagnostics Check		OK	11:39

Data Download Date/Time: 10/31/2020 03:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU,OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 09/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A4
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PEASE

DATE/TIME OF TEST: 09/20/2020 00:59:02

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.077	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.072	00:54
Breath Volume		2.429	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.071	00:57
Breath Volume		2.101	00:57
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:59

Data Download Date/Time: 10/31/2020 03:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,ROBERT,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/22/2020 03:03:46

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.243	02:55
Breath Volume		2.003	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.222	02:58
Breath Volume		1.382	02:58
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 3	0.225	03:02
Breath Volume		1.859	03:02
Air Blank		0.000	03:02
Control Test		0.075	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FRASHIER

DATE/TIME OF TEST: 09/23/2020 02:18:54

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.079	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.079	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 09/24/2020 03:31:55

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.127	03:27
Breath Volume		1.902	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.128	03:30
Breath Volume		2.121	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DAVIDSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/24/2020 20:48:13

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:31	
Control Test		0.079	20:31	
Air Blank		0.000	20:32	
Subject Sample	# 1	0.240	20:35	Volume Not Met
Breath Volume		0.273	20:35	
Air Blank		0.000	20:36	
Air Blank		0.000	20:38	
Subject Sample	# 2	0.265	20:40	
Breath Volume		1.210	20:40	
Air Blank		0.000	20:41	
Air Blank		0.000	20:43	
Subject Sample	# 3		20:46	No Sample Provided
Breath Volume		0.000	20:46	
Air Blank		0.000	20:47	
Control Test		0.075	20:47	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,RONALD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: ROVIRA,ENZO
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 09/24/2020 23:50:58

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.078	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.051	23:46
Breath Volume		1.292	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.052	23:49
Breath Volume		1.925	23:49
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: REYES REYES

DATE/TIME OF TEST: 09/25/2020 03:12:18

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1	0.163	03:02	No .020 agreement
Breath Volume		1.960	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2	0.184	03:06	No .020 agreement
Breath Volume		1.878	03:06	
Air Blank		0.000	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 3	0.207	03:10	No .020 agreement
Breath Volume		2.140	03:10	
Air Blank		0.000	03:11	
Control Test		0.078	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATERMAN ORTIZ,JOHAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 09/25/2020 04:58:54

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.079	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.135	04:53
Breath Volume		2.890	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.132	04:57
Breath Volume		2.339	04:57
Air Blank		0.000	04:57
Control Test		0.078	04:58
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: SANDS,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MCCAULEY

DATE/TIME OF TEST: 09/25/2020 22:27:46

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.131	22:22
Breath Volume		2.589	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.125	22:25
Breath Volume		1.718	22:25
Air Blank		0.000	22:26
Control Test		0.078	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/29/2020 00:01:37

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.212	23:57
Breath Volume		1.582	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.203	00:00
Breath Volume		1.785	00:00
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MERINO VASQUEZ

DATE/TIME OF TEST: 09/30/2020 01:18:26

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.128	01:13
Breath Volume		2.609	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.124	01:16
Breath Volume		1.472	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: TOSE RIGELL

DATE/TIME OF TEST: 09/30/2020 04:58:06

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.043	04:53
Breath Volume		2.953	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.040	04:56
Breath Volume		2.390	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 10/20/2020 08:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/03/2020 00:36:12

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:31
Breath Volume		1.503	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:34
Breath Volume		3.437	00:34
Air Blank		0.000	00:35
Control Test		0.078	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 09/04/2020 00:20:45

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.103	00:15
Breath Volume		1.644	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.105	00:19
Breath Volume		1.828	00:19
Air Blank		0.000	00:19
Control Test		0.078	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TUCKER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: BRUNELLE

DATE/TIME OF TEST: 09/04/2020 00:57:56

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.079	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.000	00:53
Breath Volume		2.777	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.000	00:56
Breath Volume		2.707	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/04/2020 05:27:15

Diagnostics Check		OK	05:13	
Air Blank		0.000	05:14	
Control Test		0.079	05:14	
Air Blank		0.000	05:15	
Subject Sample	# 1	0.188	05:18	Volume Not Met
Breath Volume		0.792	05:18	
Air Blank		0.000	05:18	
Air Blank		0.000	05:20	
Subject Sample	# 2	0.193	05:21	
Breath Volume		1.718	05:21	
Air Blank		0.000	05:21	
Air Blank		0.000	05:23	
Subject Sample	# 3	0.201	05:25	
Breath Volume		1.804	05:25	
Air Blank		0.000	05:26	
Control Test		0.078	05:26	
Air Blank		0.000	05:27	
Diagnostics Check		OK	05:27	

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TUCKER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: DUNLAP

DATE/TIME OF TEST: 09/05/2020 03:20:53

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.079	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:19	Subject Test Refused
Breath Volume		0.000	03:19	
Air Blank		0.000	03:19	
Control Test		0.079	03:20	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,E,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPEICHER,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/08/2020 15:05:56

Diagnostics Check		OK	14:57
Air Blank		0.000	14:57
Control Test		0.079	14:58
Air Blank		0.000	14:58
Subject Sample	# 1	0.295	15:00
Breath Volume		2.804	15:00
Air Blank		0.000	15:01
Air Blank		0.000	15:02
Subject Sample	# 2	0.288	15:04
Breath Volume		1.378	15:04
Air Blank		0.000	15:05
Control Test		0.077	15:05
Air Blank		0.000	15:05
Diagnostics Check		OK	15:05

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KULBAKA,NAZARIY,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DALESSIO,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: KULBAKA

DATE/TIME OF TEST: 09/12/2020 07:37:04

Diagnostics Check		OK	07:34	
Air Blank		0.000	07:34	
Control Test		0.079	07:34	
Air Blank		0.000	07:35	
Subject Sample	# 1		07:35	Subject Test Refused
Breath Volume		0.000	07:35	
Air Blank		0.000	07:36	
Control Test		0.079	07:36	
Air Blank		0.000	07:36	
Diagnostics Check		OK	07:37	

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: SUMMERALL

DATE/TIME OF TEST: 09/12/2020 19:54:32

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:52	
Control Test		0.079	19:52	
Air Blank		0.000	19:52	
Subject Sample	# 1		19:53	Subject Test Refused
Breath Volume		0.000	19:53	
Air Blank		0.000	19:53	
Control Test		0.079	19:53	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SUMMERALL

DATE/TIME OF TEST: 09/12/2020 20:42:24

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.079	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.087	20:37
Breath Volume		1.726	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.086	20:40
Breath Volume		1.562	20:40
Air Blank		0.000	20:41
Control Test		0.080	20:41
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/12/2020 23:20:00

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.000	23:15
Breath Volume		2.738	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.000	23:18
Breath Volume		1.937	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:20

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SIERRA GONZALEZ

DATE/TIME OF TEST: 09/13/2020 02:27:16

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.177	02:19
Breath Volume		1.503	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.199	02:22
Breath Volume		3.089	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 3	0.188	02:25
Breath Volume		2.824	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROWELL,AUBREY,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SM WOLFORD,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: TROWELL

DATE/TIME OF TEST: 09/14/2020 16:25:06

Diagnostics Check		OK	16:22	
Air Blank		0.000	16:22	
Control Test		0.078	16:22	
Air Blank		0.000	16:23	
Subject Sample	# 1		16:23	Subject Test Refused
Breath Volume		0.000	16:23	
Air Blank		0.000	16:24	
Control Test		0.079	16:24	
Air Blank		0.000	16:25	
Diagnostics Check		OK	16:25	

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: OLEARY

DATE/TIME OF TEST: 09/14/2020 18:40:33

Diagnostics Check		OK	18:37	
Air Blank		0.000	18:38	
Control Test		0.078	18:38	
Air Blank		0.000	18:38	
Subject Sample	# 1		18:39	Subject Test Refused
Breath Volume		0.000	18:39	
Air Blank		0.000	18:39	
Control Test		0.078	18:39	
Air Blank		0.000	18:40	
Diagnostics Check		OK	18:40	

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,A,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: AIKEN

DATE/TIME OF TEST: 09/14/2020 22:40:46

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.079	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.000	22:36
Breath Volume		2.199	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.000	22:39
Breath Volume		1.886	22:39
Air Blank		0.000	22:39
Control Test		0.079	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 09/16/2020 01:14:55

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.078	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.078	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 09/16/2020 15:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 09/30/2020 21:52:01

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.079	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.160	21:45
Breath Volume		1.902	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.157	21:50
Breath Volume		1.582	21:50
Air Blank		0.000	21:51
Control Test		0.079	21:51
Air Blank		0.000	21:51
Diagnostics Check		OK	21:52

Data Download Date/Time: 10/21/2020 11:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNNEFELD,RALPH,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: NORMAN,BENJAMIN
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 09/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: OTHER
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/26/2020 06:14:26

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:12	
Control Test		0.080	06:12	
Air Blank		0.000	06:12	
Subject Sample	# 1		06:13	Subject Test Refused
Breath Volume		0.000	06:13	
Air Blank		0.000	06:13	
Control Test		0.080	06:13	
Air Blank		0.000	06:14	
Diagnostics Check		OK	06:14	

Data Download Date/Time: 10/21/2020 08:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

INSTRUMENT REGISTERED TO: SURFSIDE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: BLANCO,ERICA
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PERLA

DATE/TIME OF TEST: 09/05/2020 03:28:18

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.078	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 10/29/2020 12:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/11/2020 01:27:51

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.188	01:23
Breath Volume		2.250	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.173	01:26
Breath Volume		2.523	01:26
Air Blank		0.000	01:26
Control Test		0.077	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 09/29/2020 13:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: GEPHART

DATE/TIME OF TEST: 09/12/2020 19:49:35

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.076	19:44
Air Blank		0.000	19:44
Subject Sample	# 1	0.223	19:45
Breath Volume		1.941	19:45
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.216	19:47
Breath Volume		2.164	19:47
Air Blank		0.000	19:48
Control Test		0.075	19:48
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 09/29/2020 13:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: MCMILLON

DATE/TIME OF TEST: 09/25/2020 22:43:47

Diagnostics Check		OK	22:37
Air Blank		0.000	22:38
Control Test		0.077	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.000	22:39
Breath Volume		1.746	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.000	22:42
Breath Volume		1.222	22:42
Air Blank		0.000	22:42
Control Test		0.077	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 09/29/2020 13:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32518080A1
DRY GAS EXPIRY DATE: 02/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GONZALEZ-ORTEGA

DATE/TIME OF TEST: 09/27/2020 03:18:01

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.145	03:13
Breath Volume		2.054	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.142	03:16
Breath Volume		2.203	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 09/29/2020 13:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: JEDYNAK

DATE/TIME OF TEST: 09/07/2020 04:49:28

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.109	04:44
Breath Volume		4.742	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.107	04:48
Breath Volume		3.570	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 09/10/2020 14:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 09/24/2020 04:34:48

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.142	04:30
Breath Volume		4.273	04:30
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.140	04:33
Breath Volume		3.976	04:33
Air Blank		0.000	04:33
Control Test		0.080	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 10/22/2020 14:55

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: LOPEZ VELAZQUEZ

DATE/TIME OF TEST: 09/02/2020 23:53:44

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.079	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.176	23:48
Breath Volume		2.871	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.167	23:52
Breath Volume		2.074	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 10/20/2020 10:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: CRUZ SANTIAGO

DATE/TIME OF TEST: 09/03/2020 23:50:45

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.167	23:46
Breath Volume		2.425	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.171	23:49
Breath Volume		2.730	23:49
Air Blank		0.000	23:49
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 10/20/2020 10:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 09/04/2020 22:36:44

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.080	22:31
Air Blank		0.000	22:32
Subject Sample	# 1	0.101	22:32
Breath Volume		2.628	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:34
Subject Sample	# 2	0.113	22:35
Breath Volume		3.132	22:35
Air Blank		0.000	22:35
Control Test		0.079	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 10/20/2020 10:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: TAMIMI

DATE/TIME OF TEST: 09/05/2020 01:24:31

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.000	01:20
Breath Volume		1.667	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.000	01:23
Breath Volume		3.207	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 10/20/2020 10:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/05/2020 04:39:46

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.080	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1	0.152	04:30	Volume Not Met
Breath Volume		0.441	04:30	
Air Blank		0.000	04:30	
Air Blank		0.000	04:32	
Subject Sample	# 2	0.186	04:33	
Breath Volume		1.726	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 3	0.177	04:38	
Breath Volume		1.625	04:38	
Air Blank		0.000	04:38	
Control Test		0.078	04:39	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 10/20/2020 10:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: GATLIN

DATE/TIME OF TEST: 09/05/2020 06:50:40

Diagnostics Check		OK	06:44
Air Blank		0.000	06:45
Control Test		0.080	06:45
Air Blank		0.000	06:45
Subject Sample	# 1	0.181	06:46
Breath Volume		1.757	06:46
Air Blank		0.000	06:47
Air Blank		0.000	06:48
Subject Sample	# 2	0.168	06:49
Breath Volume		2.730	06:49
Air Blank		0.000	06:49
Control Test		0.079	06:50
Air Blank		0.000	06:50
Diagnostics Check		OK	06:50

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GILLOTTE,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: MADERA

DATE/TIME OF TEST: 09/05/2020 16:31:31

Diagnostics Check		OK	16:25
Air Blank		0.000	16:25
Control Test		0.079	16:25
Air Blank		0.000	16:26
Subject Sample	# 1	0.209	16:27
Breath Volume		2.523	16:27
Air Blank		0.000	16:27
Air Blank		0.000	16:29
Subject Sample	# 2	0.220	16:29
Breath Volume		2.546	16:29
Air Blank		0.000	16:30
Control Test		0.079	16:30
Air Blank		0.000	16:31
Diagnostics Check		OK	16:31

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: VALLEJO REYES

DATE/TIME OF TEST: 09/05/2020 22:21:02

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:09	
Control Test		0.078	22:10	
Air Blank		0.000	22:10	
Subject Sample	# 1	0.141	22:13	Volume Not Met
Breath Volume		0.136	22:13	
Air Blank		0.000	22:14	
Air Blank		0.000	22:16	
Subject Sample	# 2	0.189	22:19	Volume Not Met
Breath Volume		0.089	22:19	
Air Blank		0.000	22:20	
Control Test		0.078	22:20	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: VALLEJO REYES

DATE/TIME OF TEST: 09/05/2020 22:47:56

Diagnostics Check		OK	22:37
Air Blank		0.000	22:38
Control Test		0.078	22:38
Air Blank		0.000	22:39
Subject Sample	# 1	0.172	22:42
Breath Volume		1.527	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:44
Subject Sample	# 2	0.182	22:46
Breath Volume		2.601	22:46
Air Blank		0.000	22:47
Control Test		0.077	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: ABDULKARIM

DATE/TIME OF TEST: 09/06/2020 03:00:09

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.181	02:54
Breath Volume		1.972	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.189	02:58
Breath Volume		1.667	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LAGUERRE RODRIGUEZ

DATE/TIME OF TEST: 09/06/2020 03:16:02

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.140	03:11
Breath Volume		3.441	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.126	03:14
Breath Volume		1.281	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:16

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 09/06/2020 21:33:53

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:27
Air Blank		0.000	21:28
Subject Sample	# 1	0.179	21:29
Breath Volume		1.574	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.179	21:32
Breath Volume		1.648	21:32
Air Blank		0.000	21:32
Control Test		0.079	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARNA,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: COLEUS

DATE/TIME OF TEST: 09/07/2020 01:21:28

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.057	01:17
Breath Volume		2.535	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.046	01:19
Breath Volume		1.273	01:19
Air Blank		0.000	01:20
Control Test		0.078	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: AGRAMONTE MOREJON

DATE/TIME OF TEST: 09/07/2020 06:12:22

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.113	06:07
Breath Volume		2.691	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.108	06:10
Breath Volume		2.066	06:10
Air Blank		0.000	06:11
Control Test		0.080	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: LOPEZ VASQUEZ

DATE/TIME OF TEST: 09/07/2020 21:58:28

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.080	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.201	21:54
Breath Volume		2.492	21:54
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.199	21:56
Breath Volume		2.589	21:56
Air Blank		0.000	21:57
Control Test		0.080	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/09/2020 00:52:19

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.144	00:46
Breath Volume		2.761	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.152	00:50
Breath Volume		2.570	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DELORENZO,D
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: VERGARA RAMIREZ

DATE/TIME OF TEST: 09/10/2020 04:40:31

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.171	04:36
Breath Volume		1.738	04:36
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.169	04:38
Breath Volume		2.769	04:38
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 10/20/2020 10:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRUCE,KEATON
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 09/10/2020 17:34:09

Diagnostics Check		OK	17:27
Air Blank		0.000	17:28
Control Test		0.079	17:28
Air Blank		0.000	17:29
Subject Sample	# 1	0.287	17:29
Breath Volume		2.378	17:29
Air Blank		0.000	17:30
Air Blank		0.000	17:32
Subject Sample	# 2	0.290	17:32
Breath Volume		3.488	17:32
Air Blank		0.000	17:33
Control Test		0.083	17:33
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: HIEBNER

DATE/TIME OF TEST: 09/10/2020 23:17:48

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.078	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.000	23:13
Breath Volume		3.667	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.000	23:16
Breath Volume		3.609	23:16
Air Blank		0.000	23:16
Control Test		0.079	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CRESPO BORDON

DATE/TIME OF TEST: 09/11/2020 00:28:16

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.201	00:23
Breath Volume		2.378	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.214	00:26
Breath Volume		3.273	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/11/2020 21:31:13

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.080	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.050	21:25
Breath Volume		1.125	21:25
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.051	21:29
Breath Volume		1.683	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GONZALO

DATE/TIME OF TEST: 09/11/2020 22:16:53

Diagnostics Check		OK	22:09
Air Blank		0.000	22:09
Control Test		0.080	22:09
Air Blank		0.000	22:10
Subject Sample	# 1	0.183	22:12
Breath Volume		3.132	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:14
Subject Sample	# 2	0.186	22:15
Breath Volume		2.945	22:15
Air Blank		0.000	22:15
Control Test		0.078	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: ESPINOSA FELIZ

DATE/TIME OF TEST: 09/12/2020 05:02:39

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.182	04:58
Breath Volume		2.214	04:58
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.184	05:01
Breath Volume		2.667	05:01
Air Blank		0.000	05:01
Control Test		0.078	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: COSTELLO JR

DATE/TIME OF TEST: 09/12/2020 06:20:44

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.080	06:15
Air Blank		0.000	06:16
Subject Sample	# 1	0.086	06:16
Breath Volume		2.160	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:18
Subject Sample	# 2	0.085	06:19
Breath Volume		1.859	06:19
Air Blank		0.000	06:19
Control Test		0.080	06:20
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 09/12/2020 07:50:03

Diagnostics Check		OK	07:43
Air Blank		0.000	07:43
Control Test		0.079	07:44
Air Blank		0.000	07:44
Subject Sample	# 1	0.113	07:45
Breath Volume		1.460	07:45
Air Blank		0.000	07:46
Air Blank		0.000	07:48
Subject Sample	# 2	0.111	07:48
Breath Volume		1.480	07:48
Air Blank		0.000	07:49
Control Test		0.078	07:49
Air Blank		0.000	07:49
Diagnostics Check		OK	07:50

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/12/2020 21:41:51

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:30	
Control Test		0.080	21:31	
Air Blank		0.000	21:31	
Subject Sample	# 1	0.208	21:34	Volume Not Met
Breath Volume		0.070	21:34	
Air Blank		0.000	21:35	
Air Blank		0.000	21:37	
Subject Sample	# 2	0.168	21:40	Volume Not Met
Breath Volume		0.093	21:40	
Air Blank		0.000	21:40	
Control Test		0.077	21:41	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/12/2020 21:51:47

Diagnostics Check		OK	21:42	
Air Blank		0.000	21:43	
Control Test		0.079	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 1	0.207	21:47	Volume Not Met
Breath Volume		0.144	21:47	
Air Blank		0.000	21:47	
Air Blank		0.000	21:49	
Subject Sample	# 2		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.076	21:51	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 09/13/2020 04:04:51

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.138	04:00
Breath Volume		2.468	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.141	04:03
Breath Volume		2.597	04:03
Air Blank		0.000	04:03
Control Test		0.078	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: CASTRO LAZARO

DATE/TIME OF TEST: 09/13/2020 07:28:29

Diagnostics Check		OK	07:21
Air Blank		0.000	07:22
Control Test		0.080	07:22
Air Blank		0.000	07:22
Subject Sample	# 1	0.128	07:23
Breath Volume		1.316	07:23
Air Blank		0.000	07:24
Air Blank		0.000	07:26
Subject Sample	# 2	0.138	07:26
Breath Volume		2.812	07:26
Air Blank		0.000	07:27
Control Test		0.080	07:27
Air Blank		0.000	07:28
Diagnostics Check		OK	07:28

Data Download Date/Time: 10/20/2020 10:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 09/15/2020 04:02:53

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	RFI Detect
Breath Volume		0.000	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 09/15/2020 04:10:20

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.093	04:05
Breath Volume		2.636	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.094	04:08
Breath Volume		2.324	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: GONZALEZ ALFARO

DATE/TIME OF TEST: 09/15/2020 05:41:15

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.079	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.037	05:34
Breath Volume		1.140	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.035	05:39
Breath Volume		1.957	05:39
Air Blank		0.000	05:40
Control Test		0.081	05:40
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 09/15/2020 06:01:23

Diagnostics Check		OK	05:54
Air Blank		0.000	05:55
Control Test		0.079	05:55
Air Blank		0.000	05:55
Subject Sample	# 1	0.159	05:56
Breath Volume		1.988	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:59
Subject Sample	# 2	0.157	05:59
Breath Volume		1.367	05:59
Air Blank		0.000	06:00
Control Test		0.079	06:00
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: ARMATO

DATE/TIME OF TEST: 09/15/2020 22:56:26

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:40	
Control Test		0.080	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1	0.200	22:43	
Breath Volume		1.804	22:43	
Air Blank		0.000	22:44	
Air Blank		0.000	22:46	
Subject Sample	# 2	0.180	22:49	Volume Not Met
Breath Volume		0.066	22:49	
Air Blank		0.000	22:49	
Air Blank		0.000	22:51	
Subject Sample	# 3		22:54	No Sample Provided
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Control Test		0.078	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HERNSTROM

DATE/TIME OF TEST: 09/16/2020 00:47:49

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.079	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.149	00:41
Breath Volume		1.250	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.131	00:46
Breath Volume		1.246	00:46
Air Blank		0.000	00:46
Control Test		0.078	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LYRAS

DATE/TIME OF TEST: 09/17/2020 03:53:13

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.027	03:48
Breath Volume		2.070	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.027	03:51
Breath Volume		2.003	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GRANADOS

DATE/TIME OF TEST: 09/18/2020 01:20:12

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.116	01:15
Breath Volume		2.808	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.112	01:18
Breath Volume		2.828	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CONDE JR

DATE/TIME OF TEST: 09/18/2020 04:03:20

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.000	03:58
Breath Volume		3.117	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.000	04:01
Breath Volume		2.726	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: CIBOROWSKI

DATE/TIME OF TEST: 09/18/2020 17:59:43

Diagnostics Check		OK	17:53
Air Blank		0.000	17:53
Control Test		0.080	17:54
Air Blank		0.000	17:54
Subject Sample	# 1	0.189	17:55
Breath Volume		2.386	17:55
Air Blank		0.000	17:55
Air Blank		0.000	17:57
Subject Sample	# 2	0.196	17:58
Breath Volume		2.582	17:58
Air Blank		0.000	17:58
Control Test		0.079	17:59
Air Blank		0.000	17:59
Diagnostics Check		OK	17:59

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SANTANA QUINONES

DATE/TIME OF TEST: 09/19/2020 00:28:36

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.108	00:24
Breath Volume		2.671	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.108	00:27
Breath Volume		2.324	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JOSE GARCIA

DATE/TIME OF TEST: 09/19/2020 01:39:37

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.168	01:33
Breath Volume		1.589	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.156	01:38
Breath Volume		1.476	01:38
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/19/2020 03:07:52

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.054	03:03
Breath Volume		3.093	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.054	03:06
Breath Volume		2.921	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 10/20/2020 10:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MASCI

DATE/TIME OF TEST: 09/19/2020 05:37:23

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.080	05:31
Air Blank		0.000	05:32
Subject Sample	# 1	0.152	05:32
Breath Volume		3.261	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.145	05:35
Breath Volume		2.531	05:35
Air Blank		0.000	05:36
Control Test		0.079	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: KANCHE

DATE/TIME OF TEST: 09/19/2020 06:07:20

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.080	06:00
Air Blank		0.000	06:01
Subject Sample	# 1	0.194	06:02
Breath Volume		2.062	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.192	06:05
Breath Volume		1.585	06:05
Air Blank		0.000	06:06
Control Test		0.079	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KROUSE,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BUNGE

DATE/TIME OF TEST: 09/20/2020 03:46:24

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.027	03:41
Breath Volume		1.566	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.026	03:44
Breath Volume		1.656	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,KALEN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/20/2020 04:51:15

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.080	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.117	04:44	Volume Not Met
Breath Volume		0.195	04:44	
Air Blank		0.000	04:44	
Air Blank		0.000	04:46	
Subject Sample	# 2	0.120	04:49	Volume Not Met
Breath Volume		0.000	04:49	
Air Blank		0.000	04:50	
Control Test		0.079	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,KALEN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/20/2020 05:05:07

Diagnostics Check		OK	04:53	
Air Blank		0.000	04:54	
Control Test		0.078	04:54	
Air Blank		0.000	04:55	
Subject Sample	# 1	0.117	04:58	Volume Not Met
Breath Volume		0.250	04:58	
Air Blank		0.000	04:58	
Air Blank		0.000	05:00	
Subject Sample	# 2		05:03	No Sample Provided
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Control Test		0.079	05:04	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: LEONARDI

DATE/TIME OF TEST: 09/20/2020 22:08:09

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.080	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.000	22:03
Breath Volume		2.238	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.000	22:06
Breath Volume		4.023	22:06
Air Blank		0.000	22:07
Control Test		0.079	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 09/20/2020 23:27:44

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.250	23:23
Breath Volume		1.781	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.245	23:26
Breath Volume		1.273	23:26
Air Blank		0.000	23:26
Control Test		0.079	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 09/21/2020 00:40:53

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.080	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:35
Breath Volume		2.796	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.000	00:39
Breath Volume		1.667	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOSTER,E 5143
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 09/21/2020 14:09:24

Diagnostics Check		OK	14:01
Air Blank		0.000	14:01
Control Test		0.080	14:02
Air Blank		0.000	14:02
Subject Sample	# 1	0.000	14:04
Breath Volume		1.242	14:04
Air Blank		0.000	14:04
Air Blank		0.000	14:06
Subject Sample	# 2	0.000	14:08
Breath Volume		1.746	14:08
Air Blank		0.000	14:08
Control Test		0.080	14:08
Air Blank		0.000	14:09
Diagnostics Check		OK	14:09

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,L
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/22/2020 03:47:24

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.082	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.132	03:40	Volume Not Met
Breath Volume		0.136	03:40	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.119	03:45	Volume Not Met
Breath Volume		0.113	03:45	
Air Blank		0.000	03:46	
Control Test		0.078	03:46	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,L
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/22/2020 04:03:25

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.079	03:52	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.110	03:56	Volume Not Met
Breath Volume		0.152	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.114	04:01	Volume Not Met
Breath Volume		0.226	04:01	
Air Blank		0.000	04:02	
Control Test		0.077	04:02	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,SHELDON
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/22/2020 20:26:51

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.079	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.147	20:22
Breath Volume		1.804	20:22
Air Blank		0.000	20:22
Air Blank		0.000	20:24
Subject Sample	# 2	0.150	20:25
Breath Volume		1.339	20:25
Air Blank		0.000	20:25
Control Test		0.078	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:26

Data Download Date/Time: 10/20/2020 10:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RODRIGUEZ GUERRERO

DATE/TIME OF TEST: 09/24/2020 00:59:13

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.079	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.196	00:54
Breath Volume		1.628	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.198	00:57
Breath Volume		1.824	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 09/25/2020 00:59:51

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.205	00:55
Breath Volume		2.109	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:57
Subject Sample	# 2	0.213	00:58
Breath Volume		2.210	00:58
Air Blank		0.000	00:58
Control Test		0.078	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/25/2020 03:10:22

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.186	03:05
Breath Volume		1.386	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:07
Subject Sample	# 2	0.196	03:08
Breath Volume		1.707	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 09/25/2020 20:14:53

Diagnostics Check		OK	20:08
Air Blank		0.000	20:08
Control Test		0.080	20:09
Air Blank		0.000	20:09
Subject Sample	# 1	0.204	20:10
Breath Volume		2.218	20:10
Air Blank		0.000	20:11
Air Blank		0.000	20:12
Subject Sample	# 2	0.200	20:13
Breath Volume		2.371	20:13
Air Blank		0.000	20:13
Control Test		0.080	20:14
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: LIOTTA

DATE/TIME OF TEST: 09/26/2020 01:02:45

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.079	00:57
Breath Volume		2.832	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.077	01:01
Breath Volume		2.656	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: VASQUEZ RAMIREZ

DATE/TIME OF TEST: 09/26/2020 21:40:25

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.231	21:35
Breath Volume		1.339	21:35
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.220	21:38
Breath Volume		2.222	21:38
Air Blank		0.000	21:39
Control Test		0.078	21:39
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 09/26/2020 22:44:43

Diagnostics Check		OK	22:37
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Subject Sample	# 1	0.000	22:40
Breath Volume		3.109	22:40
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.000	22:43
Breath Volume		2.929	22:43
Air Blank		0.000	22:43
Control Test		0.080	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: QUILES

DATE/TIME OF TEST: 09/27/2020 04:10:23

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.198	04:05
Breath Volume		1.890	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.198	04:08
Breath Volume		2.269	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KIMMEL,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: SEITZ

DATE/TIME OF TEST: 09/27/2020 15:28:49

Diagnostics Check		OK	15:22
Air Blank		0.000	15:23
Control Test		0.080	15:23
Air Blank		0.000	15:24
Subject Sample	# 1	0.000	15:24
Breath Volume		2.265	15:24
Air Blank		0.000	15:24
Air Blank		0.000	15:26
Subject Sample	# 2	0.000	15:27
Breath Volume		2.449	15:27
Air Blank		0.000	15:27
Control Test		0.079	15:28
Air Blank		0.000	15:28
Diagnostics Check		OK	15:28

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 09/29/2020 00:59:42

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.104	00:55
Breath Volume		2.238	00:55
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.105	00:58
Breath Volume		1.847	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:59
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MCCOURT

DATE/TIME OF TEST: 09/30/2020 00:47:10

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.208	00:42
Breath Volume		1.757	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.217	00:45
Breath Volume		2.531	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: MAGNUSON

DATE/TIME OF TEST: 09/30/2020 20:13:42

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.138	20:09
Breath Volume		1.910	20:09
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.144	20:12
Breath Volume		1.890	20:12
Air Blank		0.000	20:12
Control Test		0.079	20:13
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 10/20/2020 10:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: AROCHA DIAZ

DATE/TIME OF TEST: 09/06/2020 02:19:27

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.079	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1	0.123	02:11	Volume Not Met
Breath Volume		0.441	02:11	
Air Blank		0.000	02:11	
Air Blank		0.000	02:13	
Subject Sample	# 2	0.133	02:14	
Breath Volume		1.390	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:17	
Subject Sample	# 3	0.139	02:17	
Breath Volume		1.464	02:17	
Air Blank		0.000	02:18	
Control Test		0.078	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SRIDHARAN

DATE/TIME OF TEST: 09/06/2020 03:37:31

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.146	03:32
Breath Volume		1.792	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.155	03:35
Breath Volume		1.941	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MURDOCK

DATE/TIME OF TEST: 09/06/2020 04:21:21

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:10	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1	0.152	04:14	Volume Not Met
Breath Volume		0.210	04:14	
Air Blank		0.000	04:14	
Air Blank		0.000	04:16	
Subject Sample	# 2	0.153	04:19	Volume Not Met
Breath Volume		0.046	04:19	
Air Blank		0.000	04:20	
Control Test		0.079	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: MURDOCK

DATE/TIME OF TEST: 09/06/2020 04:33:34

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.151	04:27
Breath Volume		1.683	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:29
Subject Sample	# 2	0.141	04:31
Breath Volume		1.648	04:31
Air Blank		0.000	04:32
Control Test		0.078	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LUGO,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: SKRZYSZOWSKI

DATE/TIME OF TEST: 09/09/2020 01:14:27

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.081	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.083	01:09
Breath Volume		2.000	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.077	01:12
Breath Volume		1.171	01:12
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 09/11/2020 03:00:24

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.081	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1	0.148	02:53	Volume Not Met
Breath Volume		0.261	02:53	
Air Blank		0.000	02:53	
Air Blank		0.000	02:55	
Subject Sample	# 2	0.147	02:58	Volume Not Met
Breath Volume		0.277	02:58	
Air Blank		0.000	02:59	
Control Test		0.079	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 09/11/2020 03:15:17

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:04	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.145	03:08	Volume Not Met
Breath Volume		0.265	03:08	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 2	0.122	03:13	Volume Not Met
Breath Volume		0.062	03:13	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: CHRISTIANOPOULOS

DATE/TIME OF TEST: 09/13/2020 04:55:49

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.195	04:51
Breath Volume		1.332	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.214	04:54
Breath Volume		2.121	04:54
Air Blank		0.000	04:54
Control Test		0.080	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 09/17/2020 21:44:07

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.081	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.177	21:39
Breath Volume		2.066	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.184	21:42
Breath Volume		1.558	21:42
Air Blank		0.000	21:43
Control Test		0.079	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KLENOFSKY,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: MANJA

DATE/TIME OF TEST: 09/17/2020 22:16:39

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.156	22:11
Breath Volume		3.675	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.157	22:15
Breath Volume		4.562	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 10/20/2020 10:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLASINI,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/18/2020 05:29:08

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.132	05:24
Breath Volume		2.695	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.134	05:27
Breath Volume		3.281	05:27
Air Blank		0.000	05:28
Control Test		0.080	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 10/20/2020 10:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GROSS,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/19/2020 06:18:41

Diagnostics Check		OK	06:12
Air Blank		0.000	06:13
Control Test		0.081	06:13
Air Blank		0.000	06:13
Subject Sample	# 1	0.080	06:14
Breath Volume		2.777	06:14
Air Blank		0.000	06:14
Air Blank		0.000	06:16
Subject Sample	# 2	0.075	06:17
Breath Volume		2.707	06:17
Air Blank		0.000	06:17
Control Test		0.080	06:18
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 10/20/2020 10:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 09/25/2020 00:09:42

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.081	23:57	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.175	00:01	Volume Not Met
Breath Volume		0.574	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.171	00:05	
Breath Volume		2.703	00:05	
Air Blank		0.000	00:05	
Air Blank		0.000	00:07	
Subject Sample	# 3	0.177	00:08	
Breath Volume		1.968	00:08	
Air Blank		0.000	00:08	
Control Test		0.078	00:09	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 10/20/2020 10:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: ROMO CASTRO

DATE/TIME OF TEST: 09/26/2020 20:55:12

Diagnostics Check		OK	20:48
Air Blank		0.000	20:49
Control Test		0.081	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.256	20:50
Breath Volume		2.464	20:50
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.248	20:53
Breath Volume		2.789	20:53
Air Blank		0.000	20:54
Control Test		0.080	20:54
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 10/20/2020 10:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: RONSKE

DATE/TIME OF TEST: 09/28/2020 04:12:37

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.188	04:08
Breath Volume		2.269	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.182	04:11
Breath Volume		1.800	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 10/20/2020 10:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,TREVOR,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROMERO,OMAR
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: FARRADA

DATE/TIME OF TEST: 09/02/2020 02:08:54

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.081	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Control Test		0.081	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/06/2020 01:43:04

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:32	
Control Test		0.080	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:36	No Sample Provided
Breath Volume		0.000	01:36	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 2		01:41	No Sample Provided
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/06/2020 23:43:52

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.081	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.088	23:39
Breath Volume		2.800	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:41
Subject Sample	# 2	0.076	23:42
Breath Volume		1.433	23:42
Air Blank		0.000	23:42
Control Test		0.079	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOSN,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CANTELAR,JOAN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 09/10/2020 02:38:49

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.242	02:33
Breath Volume		3.257	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.233	02:37
Breath Volume		3.113	02:37
Air Blank		0.000	02:37
Control Test		0.077	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FRANCESCHI,VANESSA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/13/2020 08:59:20

Diagnostics Check		OK	08:52	
Air Blank		0.000	08:53	
Control Test		0.082	08:53	
Air Blank		0.000	08:54	
Subject Sample	# 1	0.115	08:57	Volume Not Met
Breath Volume		0.000	08:57	
Air Blank			08:57	RFI Detect
Air Blank		0.000	08:58	
Control Test		0.082	08:58	
Air Blank		0.000	08:59	
Diagnostics Check		OK	08:59	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FRANCESCHI,VANESSA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/13/2020 09:06:50

Diagnostics Check		OK	09:04	
Air Blank		0.000	09:04	
Control Test		0.081	09:04	
Air Blank		0.000	09:05	
Subject Sample	# 1		09:05	Subject Test Refused
Breath Volume		0.000	09:05	
Air Blank		0.000	09:05	
Control Test		0.081	09:06	
Air Blank		0.000	09:06	
Diagnostics Check		OK	09:06	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FRANCESCHI,VANESSA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/13/2020 09:16:25

Diagnostics Check		OK	09:13	
Air Blank		0.000	09:14	
Control Test		0.081	09:14	
Air Blank		0.000	09:14	
Subject Sample	# 1		09:15	Subject Test Refused
Breath Volume		0.000	09:15	
Air Blank		0.000	09:15	
Control Test		0.082	09:15	
Air Blank		0.000	09:16	
Diagnostics Check		OK	09:16	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:51

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/13/2020 09:26:12

Diagnostics Check		OK	09:19
Air Blank		0.000	09:20
Control Test		0.081	09:20
Air Blank		0.000	09:20
Subject Sample	# 1	0.034	09:21
Breath Volume		1.156	09:21
Air Blank		0.000	09:22
Air Blank		0.000	09:24
Subject Sample	# 2	0.033	09:24
Breath Volume		1.726	09:24
Air Blank		0.000	09:25
Control Test		0.081	09:25
Air Blank		0.000	09:26
Diagnostics Check		OK	09:26

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SILVER

DATE/TIME OF TEST: 09/20/2020 03:00:33

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.081	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	02:59	
Control Test		0.081	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: LAFORTUNE

DATE/TIME OF TEST: 09/23/2020 18:01:24

Diagnostics Check		OK	17:55	
Air Blank		0.000	17:55	
Control Test		0.081	17:56	
Air Blank		0.000	17:56	
Subject Sample	# 1		17:59	Interferent Detect
Breath Volume		1.394	17:59	
Air Blank		0.000	18:00	
Control Test		0.082	18:00	
Air Blank		0.000	18:01	
Diagnostics Check		OK	18:01	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: L163540751010

DATE/TIME OF TEST: 09/23/2020 18:07:35

Diagnostics Check		OK	18:03	
Air Blank		0.000	18:03	
Control Test		0.080	18:04	
Air Blank		0.000	18:04	
Subject Sample	# 1		18:05	Interferent Detect
Breath Volume		1.367	18:05	
Air Blank		0.000	18:06	
Control Test		0.082	18:06	
Air Blank		0.000	18:07	
Diagnostics Check		OK	18:07	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: LAFORTUNE

DATE/TIME OF TEST: 09/23/2020 18:49:31

Diagnostics Check		OK	18:45	
Air Blank		0.000	18:46	
Control Test		0.080	18:46	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:47	Interferent Detect
Breath Volume		2.214	18:47	
Air Blank		0.000	18:48	
Control Test		0.080	18:48	
Air Blank		0.000	18:49	
Diagnostics Check		OK	18:49	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: LAFORTUNE

DATE/TIME OF TEST: 09/23/2020 19:02:09

Diagnostics Check		OK	18:58	
Air Blank		0.000	18:58	
Control Test		0.080	18:58	
Air Blank		0.000	18:59	
Subject Sample	# 1		19:00	Interferent Detect
Breath Volume		1.453	19:00	
Air Blank		0.000	19:01	
Control Test		0.081	19:01	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JUSTIN,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANCHEZ,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: NEWTON JR

DATE/TIME OF TEST: 09/25/2020 19:27:46

Diagnostics Check		OK	19:21
Air Blank		0.000	19:21
Control Test		0.081	19:22
Air Blank		0.000	19:22
Subject Sample	# 1	0.119	19:23
Breath Volume		2.105	19:23
Air Blank		0.000	19:24
Air Blank		0.000	19:25
Subject Sample	# 2	0.111	19:26
Breath Volume		1.527	19:26
Air Blank		0.000	19:26
Control Test		0.078	19:27
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FRANCESCHI,VANESSA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 09/27/2020 08:27:54

Diagnostics Check		OK	08:15	
Air Blank		0.000	08:16	
Control Test		0.081	08:16	
Air Blank		0.000	08:17	
Subject Sample	# 1	0.198	08:20	Volume Not Met
Breath Volume		0.339	08:20	
Air Blank		0.000	08:21	
Air Blank		0.000	08:22	
Subject Sample	# 2	0.189	08:25	Volume Not Met
Breath Volume		0.441	08:25	
Air Blank		0.000	08:26	
Control Test		0.078	08:27	
Air Blank		0.000	08:27	
Diagnostics Check		OK	08:27	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FRANCESCHI,VANESSA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 09/27/2020 08:42:37

Diagnostics Check		OK	08:31	
Air Blank		0.000	08:32	
Control Test		0.078	08:32	
Air Blank		0.000	08:32	
Subject Sample	# 1	0.179	08:36	Volume Not Met
Breath Volume		0.414	08:36	
Air Blank		0.000	08:37	
Air Blank		0.000	08:38	
Subject Sample	# 2	0.201	08:41	
Breath Volume		1.117	08:41	
Air Blank			08:42	Purge Fail
Air Blank		0.000	08:42	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FRANCESCHI,VANESSA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 09/27/2020 08:52:42

Diagnostics Check		OK	08:45	
Air Blank		0.000	08:45	
Control Test		0.077	08:46	
Air Blank		0.000	08:46	
Subject Sample	# 1	0.207	08:47	
Breath Volume		1.738	08:47	
Air Blank		0.000	08:48	
Air Blank		0.000	08:49	
Subject Sample	# 2	0.200	08:51	
Breath Volume		2.277	08:51	
Air Blank		0.000	08:51	
Control Test		0.073	08:52	Control Outside Tolerance
Air Blank		0.000	08:52	

Data Download Date/Time: 10/01/2020 03:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: KIERNAN

DATE/TIME OF TEST: 09/02/2020 02:27:30

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.079	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.245	02:20	Volume Not Met
Breath Volume		0.156	02:20	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.251	02:25	Volume Not Met
Breath Volume		0.210	02:25	
Air Blank		0.000	02:26	
Control Test		0.076	02:26	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 10/02/2020 13:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LAVALLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: BORECKY

DATE/TIME OF TEST: 09/02/2020 22:24:10

Diagnostics Check		OK	22:17
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.176	22:19
Breath Volume		1.625	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.169	22:22
Breath Volume		1.433	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 10/02/2020 13:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: DI GIANNI

DATE/TIME OF TEST: 09/03/2020 00:25:03

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.225	00:20
Breath Volume		2.238	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.232	00:23
Breath Volume		2.246	00:23
Air Blank		0.000	00:24
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:25

Data Download Date/Time: 10/02/2020 13:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: BALSER

DATE/TIME OF TEST: 09/03/2020 23:03:39

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.138	22:58
Breath Volume		2.023	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.135	23:02
Breath Volume		2.152	23:02
Air Blank		0.000	23:02
Control Test		0.078	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 10/02/2020 13:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAVALLE,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LAVALLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: KENNEY

DATE/TIME OF TEST: 09/16/2020 00:08:13

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.000	00:03
Breath Volume		3.671	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.000	00:06
Breath Volume		3.449	00:06
Air Blank		0.000	00:07
Control Test		0.080	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 10/02/2020 13:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TISDALE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/22/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 09/17/2020 21:46:56

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.081	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.000	21:42
Breath Volume		2.410	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.000	21:45
Breath Volume		2.441	21:45
Air Blank		0.000	21:46
Control Test		0.081	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 09/19/2020 03:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HERNANDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 190923
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: QUITIAN PENAGOS

DATE/TIME OF TEST: 09/07/2020 01:26:22

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:25	
Control Test		0.079	01:25	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 10/20/2020 12:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHOBREY,SPENCER,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHOBREY,SPENCER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 190923
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 09/18/2020 21:49:51

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.207	21:45
Breath Volume		2.250	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.206	21:48
Breath Volume		1.527	21:48
Air Blank		0.000	21:48
Control Test		0.079	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 10/20/2020 12:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 09/05/2020 00:06:05

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.142	00:01
Breath Volume		2.871	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.135	00:04
Breath Volume		2.910	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:06

Data Download Date/Time: 10/20/2020 12:13

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAYERS,MICHELLE,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SAYERS,MICHELLE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 09/05/2020 02:11:13

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.104	02:06
Breath Volume		2.816	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:08
Subject Sample	# 2	0.097	02:09
Breath Volume		1.375	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 10/20/2020 12:13

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITZGERALD,FRANKLIN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: FERNANDEZ,SHELSY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/19/2020 04:35:50

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.062	04:31
Breath Volume		1.140	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.063	04:34
Breath Volume		1.949	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 09/21/2020 09:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: ROVELTO

DATE/TIME OF TEST: 09/19/2020 18:19:09

Diagnostics Check		OK	18:13
Air Blank		0.000	18:13
Control Test		0.081	18:13
Air Blank		0.000	18:14
Subject Sample	# 1	0.000	18:14
Breath Volume		1.500	18:14
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.000	18:17
Breath Volume		1.347	18:17
Air Blank		0.000	18:18
Control Test		0.080	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 09/21/2020 09:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICKINGS,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BEHRENS,BRENT
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/30/2020 12:25:36

Diagnostics Check	OK	12:25	
Air Blank		12:25	RFI Detect
Air Blank		12:25	RFI Detect

Data Download Date/Time: 10/28/2020 11:05

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICKINGS,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BEHRENS,BRENT
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/30/2020 12:29:26

Diagnostics Check	OK	12:27	
Air Blank	0.000	12:28	
Control Test	0.081	12:28	
Air Blank		12:28	RFI Detect
Air Blank	0.000	12:29	

Data Download Date/Time: 10/28/2020 11:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICKINGS,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: BEHRENS,BRENT
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/30/2020 12:37:49

Diagnostics Check		OK	12:31
Air Blank		0.000	12:31
Control Test		0.080	12:31
Air Blank		0.000	12:32
Subject Sample	# 1	0.234	12:33
Breath Volume		2.488	12:33
Air Blank		0.000	12:34
Air Blank		0.000	12:35
Subject Sample	# 2	0.234	12:36
Breath Volume		2.039	12:36
Air Blank		0.000	12:36
Control Test		0.080	12:37
Air Blank		0.000	12:37
Diagnostics Check		OK	12:37

Data Download Date/Time: 10/28/2020 11:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MARTINEZHERNANDEZ

DATE/TIME OF TEST: 09/07/2020 00:27:48

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.041	00:23
Breath Volume		1.253	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.037	00:26
Breath Volume		2.621	00:26
Air Blank		0.000	00:26
Control Test		0.076	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 10/20/2020 12:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,DAVID,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 09/09/2020 02:41:15

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:38	
Control Test		0.078	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:39	RFI Detect
Breath Volume		0.089	02:39	
Air Blank		0.000	02:40	
Control Test		0.079	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 10/20/2020 12:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 09/28/2020 00:17:43

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
		00:17	RFI Detect
Air Blank	0.000	00:17	

Data Download Date/Time: 10/20/2020 12:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HANSELL,ROBERT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1190213
DRY GAS EXPIRY DATE: 08/28/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 09/28/2020 00:25:55

Diagnostics Check		OK	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.127	00:21
Breath Volume		1.750	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:23
Subject Sample	# 2	0.126	00:24
Breath Volume		2.289	00:24
Air Blank		0.000	00:24
Control Test		0.079	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 10/20/2020 12:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LIEB,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/02/2020 02:41:35

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: TROWBRIDGE,TIMOTHY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/03/2020 00:31:04

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.080	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 09/04/2020 02:06:10

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.009	02:01
Breath Volume		1.378	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.009	02:04
Breath Volume		1.191	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 09/05/2020 02:29:22

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.081	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1	0.194	02:19	Volume Not Met
Breath Volume		0.292	02:19	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.209	02:24	
Breath Volume		2.703	02:24	
Air Blank		0.000	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 3	0.224	02:27	
Breath Volume		2.039	02:27	
Air Blank		0.000	02:28	
Control Test		0.077	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,I,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SIMPKINS,BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/05/2020 05:06:39

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.205	05:01
Breath Volume		1.550	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.204	05:05
Breath Volume		2.187	05:05
Air Blank		0.000	05:05
Control Test		0.080	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MONTOYA MEJIA

DATE/TIME OF TEST: 09/06/2020 00:12:23

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.197	00:06
Breath Volume		2.734	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.200	00:10
Breath Volume		3.187	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,IAN,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: POWELL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 09/06/2020 01:58:30

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.079	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,IAN,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARRINGTON,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: S530330020080

DATE/TIME OF TEST: 09/06/2020 06:36:37

Diagnostics Check		OK	06:33	
Air Blank		0.000	06:33	
Control Test		0.080	06:34	
Air Blank		0.000	06:34	
Subject Sample	# 1		06:35	Subject Test Refused
Breath Volume		0.000	06:35	
Air Blank		0.000	06:35	
Control Test		0.079	06:36	
Air Blank		0.000	06:36	
Diagnostics Check		OK	06:36	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/06/2020 20:59:32

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.079	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.150	20:54
Breath Volume		3.546	20:54
Air Blank		0.000	20:55
Air Blank		0.000	20:56
Subject Sample	# 2	0.149	20:57
Breath Volume		4.199	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL,SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: KANTNER

DATE/TIME OF TEST: 09/07/2020 13:19:00

Diagnostics Check		OK	13:10
Air Blank		0.000	13:10
Control Test		0.080	13:11
Air Blank		0.000	13:11
Subject Sample	# 1	0.053	13:12
Breath Volume		1.945	13:12
Air Blank		0.000	13:12
Air Blank		0.000	13:14
Subject Sample	# 2	0.054	13:17
Breath Volume		1.558	13:17
Air Blank		0.000	13:18
Control Test		0.080	13:18
Air Blank		0.000	13:18
Diagnostics Check		OK	13:19

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 09/07/2020 14:38:56

Diagnostics Check		OK	14:32
Air Blank		0.000	14:32
Control Test		0.080	14:33
Air Blank		0.000	14:33
Subject Sample	# 1	0.019	14:34
Breath Volume		2.375	14:34
Air Blank		0.000	14:35
Air Blank		0.000	14:37
Subject Sample	# 2	0.018	14:37
Breath Volume		1.539	14:37
Air Blank		0.000	14:38
Control Test		0.080	14:38
Air Blank		0.000	14:38
Diagnostics Check		OK	14:38

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,IAN,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARRINGTON,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 09/10/2020 03:38:06

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.079	03:33
Breath Volume		3.847	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.078	03:36
Breath Volume		3.957	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,SHELBY,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 10:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:19

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 09/11/2020 12:50:21

Diagnostics Check		OK	12:43
Air Blank		0.000	12:44
Control Test		0.081	12:44
Air Blank		0.000	12:45
Subject Sample	# 1	0.000	12:45
Breath Volume		2.257	12:45
Air Blank		0.000	12:46
Air Blank		0.000	12:48
Subject Sample	# 2	0.000	12:48
Breath Volume		1.925	12:48
Air Blank		0.000	12:49
Control Test		0.080	12:49
Air Blank		0.000	12:50
Diagnostics Check		OK	12:50

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: BOLTON

DATE/TIME OF TEST: 09/11/2020 20:51:17

Diagnostics Check		OK	20:44
Air Blank		0.000	20:45
Control Test		0.080	20:45
Air Blank		0.000	20:46
Subject Sample	# 1	0.114	20:46
Breath Volume		2.316	20:46
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.113	20:49
Breath Volume		2.503	20:49
Air Blank		0.000	20:50
Control Test		0.079	20:50
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOHLIN,MARK,T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SWENSON,BRIAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: DEMARCO

DATE/TIME OF TEST: 09/12/2020 04:19:49

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.165	04:14
Breath Volume		1.402	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.170	04:18
Breath Volume		2.289	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SUQUINO GONZALEZ

DATE/TIME OF TEST: 09/13/2020 01:44:12

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.199	01:39
Breath Volume		2.503	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.192	01:42
Breath Volume		1.171	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: SUAZO MEJIA

DATE/TIME OF TEST: 09/13/2020 21:45:26

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.080	21:38
Air Blank		0.000	21:38
Subject Sample	# 1	0.191	21:40
Breath Volume		1.917	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.171	21:43
Breath Volume		1.652	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: THERIOT

DATE/TIME OF TEST: 09/13/2020 22:21:02

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.079	22:19	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank		0.000	22:20	
Control Test		0.080	22:20	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:21	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SIMPKINS,BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/15/2020 01:41:44

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.080	01:39	
Air Blank		0.000	01:40	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.080	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON,IAN,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ARRINGTON,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BERKLEY

DATE/TIME OF TEST: 09/16/2020 00:16:03

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.080	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1	0.200	00:08	Slope Not Level
Breath Volume		1.109	00:08	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.241	00:11	
Breath Volume		3.117	00:11	
Air Blank		0.000	00:12	
Air Blank		0.000	00:14	
Subject Sample	# 3	0.237	00:14	
Breath Volume		3.109	00:14	
Air Blank		0.000	00:15	
Control Test		0.078	00:15	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 09/17/2020 14:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: RUIZ OSORTO

DATE/TIME OF TEST: 09/18/2020 02:01:17

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.154	01:56
Breath Volume		2.433	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.149	01:59
Breath Volume		2.761	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 10/21/2020 16:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: POWELL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 09/19/2020 04:37:19

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.080	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:36	
Control Test		0.080	04:36	
Air Blank		0.000	04:37	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 10/21/2020 16:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,RAY,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: REES

DATE/TIME OF TEST: 09/19/2020 14:35:12

Diagnostics Check		OK	14:31	
Air Blank		0.000	14:31	
Control Test		0.077	14:32	
Air Blank		0.000	14:32	
Subject Sample	# 1		14:33	Subject Test Refused
Breath Volume		0.000	14:33	
Air Blank		0.000	14:34	
Control Test		0.078	14:34	
Air Blank		0.000	14:35	
Diagnostics Check		OK	14:35	

Data Download Date/Time: 10/21/2020 16:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONARCH,BRIAN,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MONARCH,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HOLL

DATE/TIME OF TEST: 09/20/2020 05:07:05

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.221	05:02
Breath Volume		3.035	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.223	05:05
Breath Volume		2.214	05:05
Air Blank		0.000	05:06
Control Test		0.078	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 10/21/2020 16:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGUSA,NICHOLE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK,JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/20/2020 08:16:15

Diagnostics Check		OK	08:13	
Air Blank		0.000	08:13	
Control Test		0.080	08:14	
Air Blank		0.000	08:14	
Subject Sample	# 1		08:14	Subject Test Refused
Breath Volume		0.000	08:14	
Air Blank		0.000	08:15	
Control Test		0.080	08:15	
Air Blank		0.000	08:16	
Diagnostics Check		OK	08:16	

Data Download Date/Time: 10/21/2020 16:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONARCH,BRIAN,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MONARCH,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CONLEY JR

DATE/TIME OF TEST: 09/21/2020 01:40:17

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.080	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Control Test		0.080	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 10/21/2020 16:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: \
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 09/22/2020 16:07:35

Diagnostics Check		OK	16:01
Air Blank		0.000	16:01
Control Test		0.080	16:02
Air Blank		0.000	16:02
Subject Sample	# 1	0.280	16:03
Breath Volume		2.195	16:03
Air Blank		0.000	16:03
Air Blank		0.000	16:05
Subject Sample	# 2	0.272	16:06
Breath Volume		2.250	16:06
Air Blank		0.000	16:06
Control Test		0.079	16:06
Air Blank		0.000	16:07
Diagnostics Check		OK	16:07

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 09/23/2020 05:23:24

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.080	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1	0.222	05:15	Volume Not Met
Breath Volume		0.937	05:15	
Air Blank		0.000	05:15	
Air Blank		0.000	05:17	
Subject Sample	# 2	0.250	05:17	
Breath Volume		1.488	05:17	
Air Blank		0.000	05:18	
Air Blank		0.000	05:20	
Subject Sample	# 3	0.253	05:21	
Breath Volume		1.960	05:21	
Air Blank		0.000	05:22	
Control Test		0.077	05:22	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 09/24/2020 00:09:18

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.081	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.117	00:04
Breath Volume		2.941	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.115	00:07
Breath Volume		2.890	00:07
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NABULSI,TABITHA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 09/24/2020 01:15:09

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.216	01:09
Breath Volume		2.171	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.207	01:13
Breath Volume		2.121	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRISTOL,JARRET,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BRISTOL,JARRET
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/24/2020 21:28:42

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:17	
Control Test		0.080	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1	0.395	21:18	
Breath Volume		1.285	21:18	
Air Blank		0.000	21:19	
Air Blank		0.000	21:21	
Subject Sample	# 2	0.364	21:23	
Breath Volume		2.730	21:23	
Air Blank		0.000	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 3	0.360	21:26	
Breath Volume		2.015	21:26	
Air Blank		0.000	21:27	
Control Test		0.074	21:28	Control Outside Tolerance
Air Blank		0.000	21:28	

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CHERAMIE

DATE/TIME OF TEST: 09/26/2020 20:04:32

Diagnostics Check		OK	19:56	
Air Blank		0.000	19:56	
Control Test		0.079	19:56	
Air Blank		0.000	19:57	
Subject Sample	# 1		20:00	No Sample Provided
Breath Volume		0.000	20:00	
Air Blank		0.000	20:00	
Air Blank		0.000	20:02	
Subject Sample	# 2		20:03	Subject Test Refused
Breath Volume		0.000	20:03	
Air Blank		0.000	20:03	
Control Test		0.079	20:03	
Air Blank		0.000	20:04	
Diagnostics Check		OK	20:04	

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 09/27/2020 07:15:26

Diagnostics Check		OK	07:06	
Air Blank		0.000	07:07	
Control Test		0.080	07:07	
Air Blank		0.000	07:08	
Subject Sample	# 1		07:11	No Sample Provided
Breath Volume		0.000	07:11	
Air Blank		0.000	07:11	
Air Blank		0.000	07:13	
Subject Sample	# 2		07:14	Subject Test Refused
Breath Volume		0.000	07:14	
Air Blank		0.000	07:14	
Control Test		0.080	07:14	
Air Blank		0.000	07:15	
Diagnostics Check		OK	07:15	

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: KALASHNIKOVA

DATE/TIME OF TEST: 09/30/2020 18:07:25

Diagnostics Check		OK	17:53	
Air Blank		0.000	17:53	
Control Test		0.080	17:53	
Air Blank		0.000	17:54	
Subject Sample	# 1	0.189	17:57	Volume Not Met
Breath Volume		0.367	17:57	
Air Blank		0.000	17:58	
Air Blank		0.000	17:59	
Subject Sample	# 2	0.207	18:00	
Breath Volume		1.601	18:00	
Air Blank		0.000	18:01	
Air Blank		0.000	18:03	
Subject Sample	# 3	0.200	18:05	
Breath Volume		1.464	18:05	
Air Blank		0.000	18:06	
Control Test		0.077	18:06	
Air Blank		0.000	18:07	
Diagnostics Check		OK	18:07	

Data Download Date/Time: 10/21/2020 16:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOKEN,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: LOKEN,JEFFREY
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 08/20/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/27/2020 06:53:05

Diagnostics Check		OK	06:46
Air Blank		0.000	06:46
Control Test		0.082	06:46
Air Blank		0.000	06:47
Subject Sample	# 1	0.162	06:48
Breath Volume		1.523	06:48
Air Blank		0.000	06:48
Air Blank		0.000	06:50
Subject Sample	# 2	0.160	06:51
Breath Volume		2.804	06:51
Air Blank		0.000	06:52
Control Test		0.080	06:52
Air Blank		0.000	06:53
Diagnostics Check		OK	06:53

Data Download Date/Time: 09/30/2020 15:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MENGUY

DATE/TIME OF TEST: 09/12/2020 02:19:12

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.000	02:15
Breath Volume		1.679	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.000	02:17
Breath Volume		1.203	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 10/10/2020 06:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: TOWNSON

DATE/TIME OF TEST: 09/17/2020 18:45:25

Diagnostics Check		OK	18:39
Air Blank		0.000	18:39
Control Test		0.076	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.251	18:41
Breath Volume		2.789	18:41
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.240	18:43
Breath Volume		2.078	18:43
Air Blank		0.000	18:44
Control Test		0.078	18:44
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 10/10/2020 06:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LAGARCE,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 09/18/2020 02:06:56

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.078	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.197	02:02
Breath Volume		2.621	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:04
Subject Sample	# 2	0.190	02:05
Breath Volume		1.726	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 10/10/2020 06:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SKAPYAK

DATE/TIME OF TEST: 09/18/2020 03:22:07

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.176	03:17
Breath Volume		4.246	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.168	03:20
Breath Volume		3.687	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 10/10/2020 06:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DACOSTAAGUIAR

DATE/TIME OF TEST: 09/25/2020 16:26:33

Diagnostics Check		OK	16:20
Air Blank		0.000	16:21
Control Test		0.079	16:21
Air Blank		0.000	16:21
Subject Sample	# 1	0.028	16:22
Breath Volume		2.832	16:22
Air Blank		0.000	16:22
Air Blank		0.000	16:24
Subject Sample	# 2	0.026	16:25
Breath Volume		2.332	16:25
Air Blank		0.000	16:25
Control Test		0.079	16:25
Air Blank		0.000	16:26
Diagnostics Check		OK	16:26

Data Download Date/Time: 10/10/2020 06:16

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SOWELL

DATE/TIME OF TEST: 09/04/2020 03:37:09

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.081	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.094	03:32
Breath Volume		2.691	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.090	03:35
Breath Volume		2.128	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 10/20/2020 09:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURROUGHS,H
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/05/2020 02:14:29

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.115	02:10
Breath Volume		2.144	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.116	02:12
Breath Volume		1.839	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 10/20/2020 09:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: NONTHE-CRUZ

DATE/TIME OF TEST: 09/05/2020 22:38:39

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.081	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.134	22:33
Breath Volume		1.402	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.142	22:37
Breath Volume		1.847	22:37
Air Blank		0.000	22:37
Control Test		0.080	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 10/20/2020 09:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/06/2020 02:15:27

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.257	02:10
Breath Volume		1.902	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.249	02:13
Breath Volume		1.691	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 10/20/2020 09:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CALVO PEREZ

DATE/TIME OF TEST: 09/06/2020 03:00:25

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.261	02:53
Breath Volume		1.421	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.240	02:55
Breath Volume		1.613	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 3	0.254	02:58
Breath Volume		2.398	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 10/20/2020 09:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GABARRETE ALVARENGA

DATE/TIME OF TEST: 09/07/2020 05:46:48

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.081	05:41
Air Blank		0.000	05:41
Subject Sample	# 1	0.219	05:42
Breath Volume		2.066	05:42
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.228	05:45
Breath Volume		2.023	05:45
Air Blank		0.000	05:45
Control Test		0.080	05:46
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 10/20/2020 09:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/07/2020 20:25:39

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.081	20:19
Air Blank		0.000	20:20
Subject Sample	# 1	0.000	20:21
Breath Volume		3.613	20:21
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.000	20:24
Breath Volume		2.539	20:24
Air Blank		0.000	20:24
Control Test		0.081	20:25
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 10/20/2020 09:29

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GARCIA DIAZ

DATE/TIME OF TEST: 09/11/2020 02:48:54

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.055	02:44
Breath Volume		2.871	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.051	02:47
Breath Volume		1.445	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BENCOSME

DATE/TIME OF TEST: 09/12/2020 04:33:21

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.170	04:27
Breath Volume		3.167	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.175	04:31
Breath Volume		3.082	04:31
Air Blank		0.000	04:32
Control Test		0.082	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ALFONSO II

DATE/TIME OF TEST: 09/18/2020 04:44:37

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.081	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.070	04:40
Breath Volume		3.535	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.074	04:43
Breath Volume		3.789	04:43
Air Blank		0.000	04:43
Control Test		0.081	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 09/19/2020 02:50:40

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.081	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.085	02:46
Breath Volume		1.152	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.081	02:49
Breath Volume		1.414	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LOPEZ CRUZ

DATE/TIME OF TEST: 09/20/2020 01:30:25

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:18	
Control Test		0.080	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1	0.092	01:19	
Breath Volume		1.324	01:19	
Air Blank		0.000	01:20	
Air Blank		0.000	01:22	
Subject Sample	# 2	0.076	01:25	Volume Not Met
Breath Volume		0.894	01:25	
Air Blank		0.000	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 3	0.088	01:28	
Breath Volume		1.535	01:28	
Air Blank		0.000	01:29	
Control Test		0.081	01:29	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 09/20/2020 05:15:26

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.228	05:10
Breath Volume		1.460	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.226	05:13
Breath Volume		2.046	05:13
Air Blank		0.000	05:14
Control Test		0.080	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: EASTLEY

DATE/TIME OF TEST: 09/24/2020 04:06:36

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.189	04:02
Breath Volume		2.820	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.172	04:05
Breath Volume		1.949	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BASSLER

DATE/TIME OF TEST: 09/25/2020 02:12:34

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.158	02:07
Breath Volume		2.679	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.161	02:11
Breath Volume		2.640	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 10/20/2020 09:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MERLO JR

DATE/TIME OF TEST: 09/25/2020 03:34:19

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.151	03:29
Breath Volume		3.808	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.155	03:32
Breath Volume		3.769	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 10/20/2020 09:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VASALLO,Y
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: APARICIO ROSALES

DATE/TIME OF TEST: 09/26/2020 21:29:36

Diagnostics Check		OK	21:23
Air Blank		0.000	21:23
Control Test		0.081	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.268	21:25
Breath Volume		2.035	21:25
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.275	21:28
Breath Volume		1.769	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 10/20/2020 09:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: AYZELMAN

DATE/TIME OF TEST: 09/27/2020 03:37:41

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.210	03:33
Breath Volume		1.828	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.209	03:36
Breath Volume		1.789	03:36
Air Blank		0.000	03:36
Control Test		0.081	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 10/20/2020 09:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 09/27/2020 05:03:47

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.081	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.194	04:58
Breath Volume		1.390	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.183	05:02
Breath Volume		1.546	05:02
Air Blank		0.000	05:02
Control Test		0.081	05:03
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 10/20/2020 09:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: OREARDON

DATE/TIME OF TEST: 09/29/2020 04:14:44

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.120	04:10
Breath Volume		3.906	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.110	04:13
Breath Volume		3.714	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 10/20/2020 09:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: PYLES

DATE/TIME OF TEST: 09/04/2020 23:50:26

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.077	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.180	23:45
Breath Volume		2.347	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.194	23:48
Breath Volume		2.890	23:48
Air Blank		0.000	23:49
Control Test		0.076	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: YOPP

DATE/TIME OF TEST: 09/05/2020 02:42:43

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.077	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.021	02:38
Breath Volume		2.210	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.022	02:41
Breath Volume		1.527	02:41
Air Blank		0.000	02:41
Control Test		0.077	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 09/05/2020 20:17:27

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:16	
Control Test	0.078	20:16	
Air Blank	0.000	20:16	
		20:16	RFI Detect
Air Blank	0.000	20:17	

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 09/05/2020 20:22:41

Diagnostics Check		OK	20:19	
Air Blank		0.000	20:20	
Control Test		0.077	20:20	
Air Blank		0.000	20:21	
Subject Sample	# 1		20:21	Subject Test Refused
Breath Volume		0.000	20:21	
Air Blank		0.000	20:21	
Control Test		0.077	20:22	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:22	

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: GARAPIC

DATE/TIME OF TEST: 09/06/2020 00:42:34

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.078	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.000	00:38
Breath Volume		3.707	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.000	00:41
Breath Volume		2.781	00:41
Air Blank		0.000	00:41
Control Test		0.077	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 09/06/2020 21:35:59

Diagnostics Check		OK	21:29
Air Blank		0.000	21:30
Control Test		0.078	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.102	21:31
Breath Volume		2.535	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:33
Subject Sample	# 2	0.099	21:34
Breath Volume		2.015	21:34
Air Blank		0.000	21:35
Control Test		0.078	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRILL,SHANNON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: RANGEL MEZQUITE

DATE/TIME OF TEST: 09/12/2020 03:36:53

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.079	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:35	Subject Test Refused
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.077	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ORTON

DATE/TIME OF TEST: 09/12/2020 23:59:38

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:55	
Control Test		0.077	23:55	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:58	Subject Test Refused
Breath Volume		0.000	23:58	
Air Blank		0.000	23:58	
Control Test		0.077	23:59	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HUSTER

DATE/TIME OF TEST: 09/13/2020 02:59:58

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.078	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.131	02:54
Breath Volume		1.328	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.123	02:58
Breath Volume		1.667	02:58
Air Blank		0.000	02:59
Control Test		0.077	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PONDS

DATE/TIME OF TEST: 09/17/2020 02:47:06

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.077	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.186	02:42
Breath Volume		1.355	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.185	02:45
Breath Volume		1.316	02:45
Air Blank		0.000	02:46
Control Test		0.076	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DEROSA

DATE/TIME OF TEST: 09/18/2020 01:44:52

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.077	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Control Test		0.077	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DOCHNAHL

DATE/TIME OF TEST: 09/19/2020 21:25:21

Diagnostics Check		OK	21:19
Air Blank		0.000	21:19
Control Test		0.078	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.231	21:20
Breath Volume		3.382	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.227	21:23
Breath Volume		3.300	21:23
Air Blank		0.000	21:24
Control Test		0.078	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: KUBASSEK

DATE/TIME OF TEST: 09/19/2020 23:08:44

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.078	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.229	23:03
Breath Volume		2.367	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.226	23:07
Breath Volume		2.457	23:07
Air Blank		0.000	23:07
Control Test		0.077	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 09/20/2020 00:50:39

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.078	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.181	00:46
Breath Volume		2.390	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.186	00:49
Breath Volume		2.144	00:49
Air Blank		0.000	00:49
Control Test		0.077	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: CHARIF

DATE/TIME OF TEST: 09/20/2020 02:09:37

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.078	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.077	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 10/16/2020 19:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEYER,CHARLES,L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: WIMBERLEY,KYLE
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 09/22/2020 22:01:19

Diagnostics Check		OK	21:53
Air Blank		0.000	21:53
Control Test		0.080	21:54
Air Blank		0.000	21:54
Subject Sample	# 1	0.133	21:56
Breath Volume		2.015	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:58
Subject Sample	# 2	0.128	21:59
Breath Volume		1.203	21:59
Air Blank		0.000	22:00
Control Test		0.077	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 10/21/2020 12:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLENN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 09/25/2020 23:47:12

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.077	23:42
Breath Volume		1.597	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.076	23:45
Breath Volume		1.527	23:45
Air Blank		0.000	23:46
Control Test		0.078	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 10/21/2020 12:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYMAN,THOMAS,C
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MAUTONE

DATE/TIME OF TEST: 09/01/2020 04:28:41

Diagnostics Check	OK	04:27	
Air Blank		04:28	RFI Detect
Air Blank	0.000	04:28	

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYMAN,THOMAS,C
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MAUTONE

DATE/TIME OF TEST: 09/01/2020 04:37:11

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.161	04:32
Breath Volume		1.929	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.148	04:35
Breath Volume		1.343	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: WHITE,JUSTIN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 09/05/2020 08:13:53

Diagnostics Check		OK	08:06
Air Blank		0.000	08:06
Control Test		0.080	08:07
Air Blank		0.000	08:07
Subject Sample	# 1	0.187	08:08
Breath Volume		2.183	08:08
Air Blank		0.000	08:09
Air Blank		0.000	08:11
Subject Sample	# 2	0.182	08:12
Breath Volume		2.226	08:12
Air Blank		0.000	08:12
Control Test		0.079	08:13
Air Blank		0.000	08:13
Diagnostics Check		OK	08:13

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALEVISATOS,JARRAD
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MASTERSON

DATE/TIME OF TEST: 09/06/2020 01:40:33

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.201	01:35
Breath Volume		2.640	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.198	01:39
Breath Volume		2.171	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RAINES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LOVELESS

DATE/TIME OF TEST: 09/07/2020 05:07:22

Diagnostics Check		OK	05:00
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.113	05:02
Breath Volume		1.148	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.116	05:05
Breath Volume		2.324	05:05
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SCHNIBBEN,BRIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/11/2020 02:27:50

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.090	02:23
Breath Volume		3.648	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.088	02:26
Breath Volume		3.230	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: PRYCE

DATE/TIME OF TEST: 09/11/2020 20:30:05

Diagnostics Check		OK	20:23
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:25
Subject Sample	# 1	0.070	20:25
Breath Volume		1.835	20:25
Air Blank		0.000	20:26
Air Blank		0.000	20:28
Subject Sample	# 2	0.066	20:28
Breath Volume		2.324	20:28
Air Blank		0.000	20:29
Control Test		0.079	20:29
Air Blank		0.000	20:30
Diagnostics Check		OK	20:30

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,DARREN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: PRESTON,WILLIAM
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 09/13/2020 03:31:04

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.336	03:26
Breath Volume		1.605	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:28
Subject Sample	# 2	0.331	03:29
Breath Volume		1.496	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:31

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RETTE,CODY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 09/19/2020 14:52:41

Diagnostics Check		OK	14:46
Air Blank		0.000	14:46
Control Test		0.081	14:47
Air Blank		0.000	14:47
Subject Sample	# 1	0.000	14:48
Breath Volume		1.140	14:48
Air Blank		0.000	14:48
Air Blank		0.000	14:50
Subject Sample	# 2	0.000	14:51
Breath Volume		1.695	14:51
Air Blank		0.000	14:51
Control Test		0.080	14:52
Air Blank		0.000	14:52
Diagnostics Check		OK	14:52

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RIZZO,BRIAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 08/13/2022
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: RUSCIANO

DATE/TIME OF TEST: 09/23/2020 05:48:48

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:37	
Control Test		0.081	05:37	
Air Blank		0.000	05:38	
Subject Sample	# 1	0.292	05:41	Volume Not Met
Breath Volume		0.105	05:41	
Air Blank		0.000	05:42	
Air Blank		0.000	05:44	
Subject Sample	# 2	0.314	05:47	Volume Not Met
Breath Volume		0.101	05:47	
Air Blank		0.000	05:47	
Control Test		0.079	05:48	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 09/29/2020 13:17

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HANCE

DATE/TIME OF TEST: 09/04/2020 01:00:34

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.078	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.139	00:55
Breath Volume		1.996	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.147	00:58
Breath Volume		4.500	00:58
Air Blank		0.000	00:59
Control Test		0.077	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 10/03/2020 00:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 09/05/2020 02:21:22

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.077	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.115	02:16
Breath Volume		3.277	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.120	02:19
Breath Volume		4.464	02:19
Air Blank		0.000	02:20
Control Test		0.078	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 10/03/2020 00:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 09/11/2020 02:30:41

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.078	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.130	02:26
Breath Volume		2.089	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.142	02:29
Breath Volume		1.375	02:29
Air Blank		0.000	02:29
Control Test		0.077	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 10/03/2020 00:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KOETTERS

DATE/TIME OF TEST: 09/12/2020 02:30:21

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:26	
Control Test		0.077	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Control Test		0.076	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 10/03/2020 00:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/18/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: GARNIER

DATE/TIME OF TEST: 09/15/2020 18:23:40

Diagnostics Check		OK	18:17
Air Blank		0.000	18:17
Control Test		0.077	18:17
Air Blank		0.000	18:18
Subject Sample	# 1	0.245	18:18
Breath Volume		1.351	18:18
Air Blank		0.000	18:19
Air Blank		0.000	18:21
Subject Sample	# 2	0.236	18:22
Breath Volume		1.281	18:22
Air Blank		0.000	18:22
Control Test		0.077	18:23
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 10/03/2020 00:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,D,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: WRIGHT,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/25/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 09/27/2020 00:35:55

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:31	
Control Test		0.079	00:31	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:34	Subject Test Refused
Breath Volume		0.175	00:34	
Air Blank		0.000	00:34	
Control Test		0.081	00:35	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 10/03/2020 00:27

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WYRICK

DATE/TIME OF TEST: 09/05/2020 02:40:32

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 09/09/2020 16:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 08/13/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GAMOLINS

DATE/TIME OF TEST: 09/06/2020 03:28:43

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.081	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.080	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 09/09/2020 16:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,J,
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: CAREY,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/09/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SABATTINI

DATE/TIME OF TEST: 09/11/2020 09:31:23

Diagnostics Check		OK	09:25
Air Blank		0.000	09:25
Control Test		0.081	09:25
Air Blank		0.000	09:26
Subject Sample	# 1	0.000	09:26
Breath Volume		3.429	09:26
Air Blank		0.000	09:27
Air Blank		0.000	09:29
Subject Sample	# 2	0.000	09:30
Breath Volume		3.316	09:30
Air Blank		0.000	09:30
Control Test		0.081	09:30
Air Blank		0.000	09:31
Diagnostics Check		OK	09:31

Data Download Date/Time: 10/20/2020 18:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,PAETSCH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: PRUETT

DATE/TIME OF TEST: 09/13/2020 21:30:28

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Subject Sample	# 1	0.156	21:25
Breath Volume		1.421	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.163	21:28
Breath Volume		1.257	21:28
Air Blank		0.000	21:29
Control Test		0.078	21:29
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 10/20/2020 18:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MAZURYK,MICHAEL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: HEDIN

DATE/TIME OF TEST: 09/16/2020 19:45:51

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:43	
Control Test		0.081	19:43	
Air Blank		0.000	19:44	
Subject Sample	# 1		19:44	Subject Test Refused
Breath Volume		0.000	19:44	
Air Blank		0.000	19:44	
Control Test		0.080	19:45	
Air Blank		0.000	19:45	
Diagnostics Check		OK	19:45	

Data Download Date/Time: 10/20/2020 18:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/27/2020 23:27:01

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.160	23:21
Breath Volume		1.402	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.152	23:25
Breath Volume		1.281	23:25
Air Blank		0.000	23:26
Control Test		0.078	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:27

Data Download Date/Time: 10/20/2020 18:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: OROZCO SANCHEZ

DATE/TIME OF TEST: 09/04/2020 01:37:28

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.226	01:32
Breath Volume		1.968	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.235	01:35
Breath Volume		2.953	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 10/28/2020 10:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: BUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: BLAY DE MALHERBE

DATE/TIME OF TEST: 09/05/2020 22:29:51

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.124	22:24
Breath Volume		1.953	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.129	22:28
Breath Volume		3.015	22:28
Air Blank		0.000	22:28
Control Test		0.079	22:29
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 10/28/2020 10:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 09/09/2020 02:26:07

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:23	
Control Test		0.079	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Control Test		0.082	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 10/28/2020 10:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: TARAJANO

DATE/TIME OF TEST: 09/19/2020 03:17:23

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.081	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:16	Subject Test Refused
Breath Volume		0.000	03:16	
Air Blank		0.000	03:16	
Control Test		0.079	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 10/28/2020 10:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ROJOP-PASTOR

DATE/TIME OF TEST: 09/20/2020 03:11:42

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.081	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.190	03:07
Breath Volume		3.613	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.194	03:10
Breath Volume		3.414	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 10/28/2020 10:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: VICTOR

DATE/TIME OF TEST: 09/24/2020 00:24:53

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.117	00:20
Breath Volume		2.250	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.114	00:23
Breath Volume		2.015	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 10/28/2020 10:25

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 09/06/2020 03:18:09

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.154	03:13
Breath Volume		1.609	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.158	03:16
Breath Volume		1.332	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 10/28/2020 10:21

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: HEWETT

DATE/TIME OF TEST: 09/11/2020 04:55:55

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.077	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.149	04:51
Breath Volume		1.593	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.136	04:54
Breath Volume		1.140	04:54
Air Blank		0.000	04:54
Control Test		0.077	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 10/28/2020 10:21

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,DAVID,L
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/01/2020 02:09:24

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.176	02:04
Breath Volume		1.398	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.177	02:07
Breath Volume		2.035	02:07
Air Blank		0.000	02:08
Control Test		0.075	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 09/21/2020 10:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 09/06/2020 04:46:50

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.166	04:42
Breath Volume		1.515	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.162	04:45
Breath Volume		1.312	04:45
Air Blank		0.000	04:45
Control Test		0.079	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 09/21/2020 10:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 08/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 09/19/2020 23:35:09

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.209	23:29
Breath Volume		1.332	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.204	23:33
Breath Volume		1.949	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 09/21/2020 10:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 09/26/2020 01:17:31

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.000	01:13
Breath Volume		1.988	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.000	01:16
Breath Volume		2.082	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 10/28/2020 11:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CONDE

DATE/TIME OF TEST: 09/28/2020 04:15:53

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.137	04:11
Breath Volume		2.281	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.132	04:14
Breath Volume		1.386	04:14
Air Blank		0.000	04:14
Control Test		0.078	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 10/28/2020 11:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/29/2020 04:30:23

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.094	04:25
Breath Volume		1.406	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.093	04:28
Breath Volume		1.707	04:28
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 10/28/2020 11:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: VILLAFANA CEBALLO

DATE/TIME OF TEST: 09/07/2020 18:30:38

Diagnostics Check		OK	18:23
Air Blank		0.000	18:23
Control Test		0.080	18:24
Air Blank		0.000	18:24
Subject Sample	# 1	0.144	18:25
Breath Volume		1.113	18:25
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.135	18:29
Breath Volume		1.781	18:29
Air Blank		0.000	18:29
Control Test		0.078	18:30
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 10/29/2020 13:07

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 09/04/2020 21:52:33

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.079	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.190	21:47
Breath Volume		3.378	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.189	21:50
Breath Volume		3.503	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 09/16/2020 22:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 09/05/2020 21:37:22

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.079	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.194	21:32
Breath Volume		1.867	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.204	21:35
Breath Volume		1.281	21:35
Air Blank		0.000	21:36
Control Test		0.079	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 09/16/2020 22:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 09/07/2020 23:59:02

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:49	
Control Test		0.078	23:49	
Air Blank		0.000	23:50	
Subject Sample	# 1	0.192	23:50	No .020 agreement
Breath Volume		1.699	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 2	0.171	23:53	No .020 agreement
Breath Volume		1.468	23:53	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.150	23:57	No .020 agreement
Breath Volume		1.625	23:57	
Air Blank		0.000	23:58	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 09/16/2020 22:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 09/08/2020 00:10:06

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.078	00:04
Air Blank		0.000	00:04
Subject Sample	# 1	0.167	00:05
Breath Volume		1.714	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:07
Subject Sample	# 2	0.173	00:08
Breath Volume		2.429	00:08
Air Blank		0.000	00:09
Control Test		0.077	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 09/16/2020 22:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HERNANDEZ-MIGUEL

DATE/TIME OF TEST: 09/10/2020 00:27:54

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.078	00:24	
Air Blank		0.000	00:25	
Subject Sample	# 1	0.162	00:26	
Breath Volume		2.941	00:26	
Air Blank			00:27	Purge Fail
Air Blank			00:27	Purge Fail

Data Download Date/Time: 09/16/2020 22:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HERNANDEZ-MIGUEL

DATE/TIME OF TEST: 09/10/2020 00:36:44

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.077	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.146	00:32
Breath Volume		2.304	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.148	00:35
Breath Volume		2.500	00:35
Air Blank		0.000	00:35
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 09/16/2020 22:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: WELLMAKER

DATE/TIME OF TEST: 09/12/2020 05:22:05

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:15	
Control Test		0.077	05:15	
Air Blank		0.000	05:16	
Subject Sample	# 1	0.202	05:16	
Breath Volume		2.253	05:16	
Air Blank		0.000	05:17	
Air Blank		0.000	05:19	
Subject Sample	# 2	0.195	05:19	
Breath Volume		3.308	05:19	
Air Blank		0.000	05:20	
Control Test		0.078	05:20	
Air Blank		0.000	05:21	
Diagnostics Check		Fail	05:21	Analytical Stability Test
Air Blank		0.000	05:22	

Data Download Date/Time: 09/16/2020 22:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: WELLMAKER

DATE/TIME OF TEST: 09/12/2020 05:30:18

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.078	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.205	05:25
Breath Volume		3.593	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.207	05:28
Breath Volume		2.964	05:28
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 09/16/2020 22:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 09/25/2020 21:51:07

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:44	
Control Test		0.078	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1	0.310	21:46	
Breath Volume		1.292	21:46	
Air Blank		0.000	21:47	
Air Blank		0.000	21:48	
Subject Sample	# 2	0.291	21:49	
Breath Volume		1.167	21:49	
Air Blank			21:50	Purge Fail
Air Blank		0.000	21:51	

Data Download Date/Time: 10/01/2020 02:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 09/25/2020 22:06:04

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.078	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.254	21:57	Volume Not Met
Breath Volume		0.757	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.290	22:00	
Breath Volume		1.621	22:00	
Air Blank		0.000	22:01	
Air Blank		0.000	22:03	
Subject Sample	# 3	0.304	22:04	
Breath Volume		1.203	22:04	
Air Blank		0.000	22:05	
Control Test		0.077	22:05	
Air Blank		0.000	22:05	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 10/01/2020 02:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SINKEVICH

DATE/TIME OF TEST: 09/27/2020 03:38:30

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.076	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.138	03:36	
Breath Volume		2.867	03:36	
Air Blank			03:37	Purge Fail
Air Blank			03:38	Purge Fail

Data Download Date/Time: 10/01/2020 02:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SINKEVICH

DATE/TIME OF TEST: 09/27/2020 03:48:10

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.130	03:42
Breath Volume		2.203	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.134	03:46
Breath Volume		2.824	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 10/01/2020 02:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.