Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIGER,CARL,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,-
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/22/2020 20:46:56

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:44	
Control Test		0.078	20:44	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:45	Subject Test Refused
Breath Volume		0.000	20:45	
Air Blank		0.000	20:46	
Control Test		0.077	20:46	
Air Blank		0.000	20:46	
Diagnostics Check		OK	20:46	

Data Download Date/Time: 12/11/2020 13:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CHASE,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/22/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: HALPIN

DATE/TIME OF TEST: 09/26/2020 19:49:31

Diagnostics Check		OK	19:46	
Air Blank		0.000	19:47	
Control Test		0.078	19:47	
Air Blank		0.000	19:47	
Subject Sample	# 1		19:48	Subject Test Refused
Breath Volume		0.000	19:48	
Air Blank		0.000	19:48	
Control Test		0.078	19:48	
Air Blank		0.000	19:49	
Diagnostics Check		OK	19:49	

Data Download Date/Time: 12/11/2020 13:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: COX,CLEVELAND
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PROPPER

DATE/TIME OF TEST: 09/01/2020 03:14:38

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.167	03:10
Breath Volume		1.148	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.173	03:13
Breath Volume		2.050	03:13
Air Blank		0.000	03:13
Control Test		0.078	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 09/23/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: YAKUBSIN,JAMES
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: STANONIS-KIMBERLING

DATE/TIME OF TEST: 09/02/2020 00:54:16

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.118	00:49
Breath Volume		2.292	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.114	00:52
Breath Volume		2.046	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 09/23/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MILLER,HAYLEY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BARBORAK

DATE/TIME OF TEST: 09/03/2020 21:10:51

Diagnostics Check		OK	20:59	
Air Blank		0.000	20:59	
Control Test		0.080	21:00	
Air Blank		0.000	21:00	
Subject Sample	# 1		21:03	No Sample Provided
Breath Volume		0.000	21:03	
Air Blank		0.000	21:04	
Air Blank		0.000	21:06	
Subject Sample	# 2		21:09	No Sample Provided
Breath Volume		0.000	21:09	
Air Blank		0.000	21:09	
Control Test		0.081	21:10	
Air Blank		0.000	21:10	
Diagnostics Check		OK	21:10	

Data Download Date/Time: 09/23/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,SARA,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ALLEN,SARA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: PADRON

DATE/TIME OF TEST: 09/05/2020 08:34:06

Diagnostics Check		OK	08:25
Air Blank		0.000	08:26
Control Test		0.081	08:26
Air Blank		0.000	08:26
Subject Sample	# 1	0.098	08:28
Breath Volume		2.250	08:28
Air Blank		0.000	08:29
Air Blank		0.000	08:30
Subject Sample	# 2	0.091	08:32
Breath Volume		1.890	08:32
Air Blank		0.000	08:33
Control Test		0.079	08:33
Air Blank		0.000	08:34
Diagnostics Check		OK	08:34

Data Download Date/Time: 09/23/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALVAREZ,ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: METTS

DATE/TIME OF TEST: 09/05/2020 22:22:31

Diagnostics Check	OK	22:21	
Air Blank		22:22	RFI Detect
Air Blank	0.000	22:22	

Data Download Date/Time: 09/23/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALVAREZ,ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: METTS

DATE/TIME OF TEST: 09/05/2020 22:25:16

Diagnostics Check	OK	22:24	
Air Blank		22:24	RFI Detect
Air Blank		22:25	RFI Detect

Data Download Date/Time: 09/23/2020 13:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALVAREZ,ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BOUTWELL

DATE/TIME OF TEST: 09/06/2020 03:13:29

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.279	03:07
Breath Volume		2.015	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.276	03:11
Breath Volume		2.253	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 09/23/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,DAVID,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RAILEY,BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: ROANE

DATE/TIME OF TEST: 09/06/2020 20:57:30

Diagnostics Check		OK	20:51
Air Blank		0.000	20:51
Control Test		0.079	20:51
Air Blank		0.000	20:52
Subject Sample	# 1	0.117	20:53
Breath Volume		1.671	20:53
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.116	20:55
Breath Volume		1.542	20:55
Air Blank		0.000	20:56
Control Test		0.078	20:56
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 09/23/2020 13:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MCVEY,TREVOR
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: TOMBLER

DATE/TIME OF TEST: 09/12/2020 04:02:40

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.215	03:55	Volume Not Met
Breath Volume		0.164	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 2	0.158	04:00	Volume Not Met
Breath Volume		0.054	04:00	
Air Blank		0.000	04:01	
Control Test		0.077	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 09/12/2020 04:42:32

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:39	
Control Test		0.079	04:40	
Air Blank		0.000	04:40	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: THAMES

DATE/TIME OF TEST: 09/13/2020 02:00:13

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.079	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Control Test		0.079	01:59	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: RICKS,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 09/13/2020 03:56:56

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.080	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1		03:49	No Sample Provided
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 2		03:55	No Sample Provided
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.081	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CABRAL,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 09/17/2020 02:13:05

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.079	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:06	No Sample Provided
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2		02:11	No Sample Provided
Breath Volume		0.000	02:11	
Air Blank		0.000	02:12	
Control Test		0.082	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 09/17/2020 18:14:22

Diagnostics Check		OK	18:09	
Air Blank		0.000	18:09	
Control Test		0.079	18:10	
Air Blank		0.000	18:10	
Subject Sample	# 1		18:12	RFI Detect
Breath Volume		0.000	18:12	
Air Blank		0.000	18:13	
Control Test		0.081	18:13	
Air Blank		0.000	18:14	
Diagnostics Check		OK	18:14	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 09/17/2020 18:31:26

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:28	
Control Test		0.080	18:29	
Air Blank		0.000	18:29	
Subject Sample	# 1		18:30	Subject Test Refused
Breath Volume		0.000	18:30	
Air Blank		0.000	18:30	
Control Test		0.082	18:30	
Air Blank		0.000	18:31	
Diagnostics Check		OK	18:31	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/19/2020 00:14:08

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.080	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1	0.224	00:08	
Breath Volume		1.140	00:08	
Air Blank		0.000	00:08	
Air Blank		0.000	00:10	
Subject Sample	# 2		00:12	RFI Detect
Breath Volume		0.281	00:12	
Air Blank		0.000	00:13	
Control Test		0.078	00:13	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:14	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/19/2020 00:24:41

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.077	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.219	00:18
Breath Volume		1.449	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:20
Subject Sample	# 2	0.208	00:23
Breath Volume		1.156	00:23
Air Blank		0.000	00:23
Control Test		0.078	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ROBERTO,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALVAREZ,ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/20/2020 03:59:57

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.138	03:55
Breath Volume		2.738	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.124	03:58
Breath Volume		2.300	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: STANFIELD,THOMAS
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/21/2020 01:44:08

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:28	
Control Test		0.080	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 1	0.149	01:32	Volume Not Met
Breath Volume		0.507	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.187	01:36	
Breath Volume		1.289	01:36	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 3	0.180	01:42	Volume Not Met
Breath Volume		0.136	01:42	
Air Blank		0.000	01:43	
Control Test		0.078	01:43	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 09/23/2020 13:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITMAN,AARON,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STALEY,K
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HAMBRICK

DATE/TIME OF TEST: 09/04/2020 01:47:45

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:31	
Control Test		0.080	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.187	01:34	
Breath Volume		3.437	01:34	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 2	0.200	01:39	Volume Not Met
Breath Volume		0.488	01:39	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 3	0.172	01:45	Volume Not Met
Breath Volume		0.347	01:45	
Air Blank		0.000	01:46	
Control Test		0.075	01:46	
Air Blank			01:47	Purge Fail
Air Blank		0.000	01:47	

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/05/2020 01:19:52

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Control Test		0.080	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: SLATE

DATE/TIME OF TEST: 09/05/2020 22:47:35

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:43	
Control Test		0.079	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:46	Subject Test Refused
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Control Test		0.079	22:46	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 09/06/2020 00:47:10

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.000	00:42
Breath Volume		2.468	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.000	00:45
Breath Volume		3.191	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LICHT

DATE/TIME OF TEST: 09/06/2020 03:18:46

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.157	03:10
Breath Volume		2.304	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.185	03:13
Breath Volume		4.421	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 3	0.196	03:17
Breath Volume		7.042	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ESQUIVEL TRANQUILINO

DATE/TIME OF TEST: 09/07/2020 02:30:40

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.000	02:26
Breath Volume		1.531	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.000	02:29
Breath Volume		2.984	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/11/2020 02:06:04

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.080	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Control Test		0.081	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: DELPOZO

DATE/TIME OF TEST: 09/11/2020 23:17:03

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.081	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.000	23:12
Breath Volume		1.785	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.000	23:15
Breath Volume		3.183	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:17

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: OKEEFE

DATE/TIME OF TEST: 09/14/2020 22:01:05

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.080	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.198	21:56
Breath Volume		1.437	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.195	21:59
Breath Volume		1.472	21:59
Air Blank		0.000	22:00
Control Test		0.080	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 09/19/2020 02:02:38

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.120	01:58
Breath Volume		2.218	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.115	02:01
Breath Volume		2.039	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/20/2020 23:17:26

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.000	23:12
Breath Volume		2.546	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.000	23:16
Breath Volume		3.671	23:16
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUS,CHRISTINA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CORLEY

DATE/TIME OF TEST: 09/22/2020 02:17:46

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.079	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 09/28/2020 12:45

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/23/2020 23:46:03

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.000	23:41
Breath Volume		2.445	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.000	23:44
Breath Volume		3.003	23:44
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:46

Data Download Date/Time: 09/28/2020 12:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUS,CHRISTINA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STALEY,KRIS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DESHIELDS

DATE/TIME OF TEST: 09/26/2020 03:39:54

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.077	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.140	03:34
Breath Volume		1.355	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.135	03:38
Breath Volume		1.707	03:38
Air Blank		0.000	03:38
Control Test		0.075	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 09/28/2020 12:46

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,CLEVE,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CORBIN,CLEVE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: STEDGE

DATE/TIME OF TEST: 09/28/2020 21:17:01

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:14	
Control Test		0.079	21:14	
Air Blank		0.000	21:15	
Subject Sample	# 1		21:15	Subject Test Refused
Breath Volume		0.000	21:15	
Air Blank		0.000	21:16	
Control Test		0.079	21:16	
Air Blank		0.000	21:16	
Diagnostics Check		OK	21:17	

Data Download Date/Time: 10/26/2020 11:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/30/2020 01:59:23

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.153	01:55
Breath Volume		3.437	01:55
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.156	01:57
Breath Volume		2.027	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 10/26/2020 11:30

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMBERT,BRIAN,K
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAMBERT,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: WICKER

DATE/TIME OF TEST: 09/05/2020 17:02:40

Diagnostics Check		OK	16:56
Air Blank		0.000	16:56
Control Test		0.081	16:56
Air Blank		0.000	16:57
Subject Sample	# 1	0.117	16:57
Breath Volume		1.257	16:57
Air Blank		0.000	16:58
Air Blank		0.000	17:00
Subject Sample	# 2	0.118	17:01
Breath Volume		1.566	17:01
Air Blank		0.000	17:01
Control Test		0.081	17:02
Air Blank		0.000	17:02
Diagnostics Check		OK	17:02

Data Download Date/Time: 09/15/2020 16:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLGOOD,JEFFREY,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLGOOD,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 09/05/2020 20:06:21

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.081	20:00
Air Blank		0.000	20:01
Subject Sample	# 1	0.058	20:01
Breath Volume		2.878	20:01
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.057	20:04
Breath Volume		3.531	20:04
Air Blank		0.000	20:05
Control Test		0.080	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 09/15/2020 16:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: YOUNG,TIMYRON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: HOLDT

DATE/TIME OF TEST: 09/09/2020 16:54:25

Diagnostics Check		OK	16:41	
Air Blank		0.000	16:42	
Control Test		0.081	16:42	
Air Blank		0.000	16:42	
Subject Sample	# 1	0.340	16:43	
Breath Volume		1.812	16:43	
Air Blank		0.000	16:44	
Air Blank		0.000	16:45	
Subject Sample	# 2	0.331	16:48	Volume Not Met
Breath Volume		0.390	16:48	
Air Blank		0.000	16:50	
Air Blank		0.000	16:51	
Subject Sample	# 3	0.323	16:52	
Breath Volume		1.218	16:52	
Air Blank		0.000	16:53	
Control Test		0.078	16:53	
Air Blank		0.000	16:54	
Diagnostics Check		OK	16:54	

Data Download Date/Time: 09/15/2020 16:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KITTRELL,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: ANDRONOVA

DATE/TIME OF TEST: 09/13/2020 22:44:48

Diagnostics Check		OK	22:37
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.171	22:39
Breath Volume		1.441	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.178	22:43
Breath Volume		1.625	22:43
Air Blank		0.000	22:43
Control Test		0.080	22:44
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 09/15/2020 16:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 09/14/2020 00:57:10

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.138	00:51
Breath Volume		1.406	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.135	00:55
Breath Volume		2.000	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 09/15/2020 16:12

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 09/18/2020 20:05:26

Diagnostics Check		OK	19:59
Air Blank		0.000	19:59
Control Test		0.082	20:00
Air Blank		0.000	20:00
Subject Sample	# 1	0.000	20:01
Breath Volume		2.355	20:01
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.000	20:04
Breath Volume		1.863	20:04
Air Blank		0.000	20:04
Control Test		0.081	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 10/13/2020 14:18

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MEDINA GUARDADO

DATE/TIME OF TEST: 09/19/2020 00:15:30

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.082	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.171	00:10
Breath Volume		4.367	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.166	00:13
Breath Volume		4.683	00:13
Air Blank		0.000	00:14
Control Test		0.081	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 10/13/2020 14:18

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: JACKMAN

DATE/TIME OF TEST: 09/25/2020 22:01:02

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:56	
Control Test		0.083	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.214	21:59	
Air Blank		0.000	22:00	
Control Test		0.083	22:00	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 10/13/2020 14:18

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/14/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1045940
DRY GAS EXPIRY DATE: 11/13/2021
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: JACKMAN

DATE/TIME OF TEST: 09/25/2020 22:28:50

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.083	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.000	22:24
Breath Volume		2.195	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.000	22:27
Breath Volume		1.996	22:27
Air Blank		0.000	22:27
Control Test		0.081	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 10/13/2020 14:18

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COX,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/06/2020 16:26:56

Diagnostics Check		OK	16:20
Air Blank		0.000	16:21
Control Test		0.081	16:21
Air Blank		0.000	16:22
Subject Sample	# 1	0.035	16:22
Breath Volume		2.761	16:22
Air Blank		0.000	16:23
Air Blank		0.000	16:25
Subject Sample	# 2	0.038	16:25
Breath Volume		2.964	16:25
Air Blank		0.000	16:26
Control Test		0.082	16:26
Air Blank		0.000	16:26
Diagnostics Check		OK	16:26

Data Download Date/Time: 10/15/2020 11:21

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,ANTHONY,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/07/2020 18:25:20

Diagnostics Check		OK	18:18	
Air Blank		0.000	18:19	
Control Test		0.081	18:19	
Air Blank		0.000	18:20	
Subject Sample	# 1	0.238	18:21	
Breath Volume		2.425	18:21	
Air Blank		0.000	18:22	
Air Blank			18:23	RFI Detect
Air Blank		0.000	18:24	
Control Test		0.082	18:24	
Air Blank		0.000	18:25	
Diagnostics Check		OK	18:25	

Data Download Date/Time: 10/15/2020 11:21

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,ANTHONY,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/07/2020 18:34:40

Diagnostics Check		OK	18:28
Air Blank		0.000	18:28
Control Test		0.081	18:29
Air Blank		0.000	18:29
Subject Sample	# 1	0.227	18:30
Breath Volume		3.328	18:30
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.224	18:33
Breath Volume		3.460	18:33
Air Blank		0.000	18:33
Control Test		0.080	18:34
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 10/15/2020 11:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JORDAN,WILFORD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: EXLEY

DATE/TIME OF TEST: 09/10/2020 19:33:52

Diagnostics Check		OK	19:28	
Air Blank		0.000	19:29	
Control Test		0.081	19:29	
Air Blank		0.000	19:30	
Subject Sample	# 1		19:32	Subject Test Refused
Breath Volume		0.000	19:32	
Air Blank		0.000	19:32	
Control Test		0.082	19:33	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 10/15/2020 11:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: DEMPS

DATE/TIME OF TEST: 09/17/2020 00:40:11

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.068	00:34
Breath Volume		1.894	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.075	00:38
Breath Volume		1.750	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 11/09/2020 10:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,JUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GARRISON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 09/17/2020 22:30:12

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.079	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.338	22:25
Breath Volume		2.214	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:27
Subject Sample	# 2	0.331	22:28
Breath Volume		2.832	22:28
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 11/09/2020 10:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADFORD,CHARLES,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WADFORD,CHARLES
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: TUMBLIN

DATE/TIME OF TEST: 09/24/2020 23:07:52

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.039	23:03
Breath Volume		2.949	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.037	23:06
Breath Volume		2.109	23:06
Air Blank		0.000	23:06
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 11/09/2020 10:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 09/26/2020 00:11:53

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.144	00:04
Breath Volume		4.320	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.123	00:07
Breath Volume		1.242	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 3	0.135	00:10
Breath Volume		4.476	00:10
Air Blank		0.000	00:10
Control Test		0.078	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 11/09/2020 10:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 09/26/2020 00:21:06

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.078	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.137	00:16
Breath Volume		4.441	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.134	00:19
Breath Volume		4.109	00:19
Air Blank		0.000	00:20
Control Test		0.077	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 11/09/2020 10:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 09/26/2020 23:01:57

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.147	22:57
Breath Volume		1.539	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.154	23:00
Breath Volume		2.964	23:00
Air Blank		0.000	23:01
Control Test		0.080	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 11/09/2020 10:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS,ALFRED,A
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: HERNANDEZ-GONZALEZ

DATE/TIME OF TEST: 09/27/2020 17:58:40

Diagnostics Check		OK	17:51
Air Blank		0.000	17:51
Control Test		0.080	17:51
Air Blank		0.000	17:52
Subject Sample	# 1	0.138	17:53
Breath Volume		1.136	17:53
Air Blank		0.000	17:54
Air Blank		0.000	17:55
Subject Sample	# 2	0.127	17:57
Breath Volume		1.617	17:57
Air Blank		0.000	17:57
Control Test		0.079	17:58
Air Blank		0.000	17:58
Diagnostics Check		OK	17:58

Data Download Date/Time: 11/09/2020 10:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 09/06/2020 20:07:05

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.085	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.177	20:02
Breath Volume		2.218	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:04
Subject Sample	# 2	0.163	20:05
Breath Volume		1.691	20:05
Air Blank		0.000	20:06
Control Test		0.083	20:06
Air Blank		0.000	20:06
Diagnostics Check		OK	20:07

Data Download Date/Time: 10/02/2020 12:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WILLCUTT

DATE/TIME OF TEST: 09/30/2020 02:02:58

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.084	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.084	01:58
Breath Volume		2.804	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.080	02:01
Breath Volume		2.664	02:01
Air Blank		0.000	02:02
Control Test		0.084	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 10/02/2020 12:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,BAYLEE,K
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERY,BAYLEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 09/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: HAMBRICK

DATE/TIME OF TEST: 09/30/2020 18:17:54

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.085	18:09
Air Blank		0.000	18:10
Subject Sample	# 1	0.000	18:13
Breath Volume		1.531	18:13
Air Blank		0.000	18:13
Air Blank		0.000	18:15
Subject Sample	# 2	0.000	18:16
Breath Volume		1.824	18:16
Air Blank		0.000	18:17
Control Test		0.083	18:17
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 10/02/2020 12:06

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,JASON,P
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BASS,JASON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: HOEHN

DATE/TIME OF TEST: 09/02/2020 14:17:34

Diagnostics Check		OK	14:10
Air Blank		0.000	14:10
Control Test		0.079	14:10
Air Blank		0.000	14:11
Subject Sample	# 1	0.227	14:13
Breath Volume		1.714	14:13
Air Blank		0.000	14:13
Air Blank		0.000	14:15
Subject Sample	# 2	0.222	14:16
Breath Volume		1.625	14:16
Air Blank		0.000	14:16
Control Test		0.078	14:16
Air Blank		0.000	14:17
Diagnostics Check		OK	14:17

Data Download Date/Time: 09/10/2020 12:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILSON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 09/06/2020 07:03:25

Diagnostics Check		OK	06:57
Air Blank		0.000	06:57
Control Test		0.079	06:57
Air Blank		0.000	06:58
Subject Sample	# 1	0.137	06:58
Breath Volume		2.214	06:58
Air Blank		0.000	06:59
Air Blank		0.000	07:01
Subject Sample	# 2	0.140	07:01
Breath Volume		2.898	07:01
Air Blank		0.000	07:02
Control Test		0.079	07:02
Air Blank		0.000	07:03
Diagnostics Check		OK	07:03

Data Download Date/Time: 09/10/2020 12:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ENGLISH,TRAVIS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/01/2020 22:41:51

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.078	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.229	22:37
Breath Volume		2.093	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:39
Subject Sample	# 2	0.224	22:40
Breath Volume		2.578	22:40
Air Blank		0.000	22:40
Control Test		0.076	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: AGUILA,S
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 09/02/2020 20:51:42

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.078	20:46
Air Blank		0.000	20:46
Subject Sample	# 1	0.000	20:47
Breath Volume		2.156	20:47
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.000	20:50
Breath Volume		2.468	20:50
Air Blank		0.000	20:50
Control Test		0.076	20:51
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PARISI,T
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 09/03/2020 04:02:44

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.078	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.141	03:58
Breath Volume		2.964	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.140	04:01
Breath Volume		1.468	04:01
Air Blank		0.000	04:01
Control Test		0.078	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: THEBO

DATE/TIME OF TEST: 09/05/2020 01:48:33

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.078	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.044	01:44
Breath Volume		3.031	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.047	01:47
Breath Volume		3.183	01:47
Air Blank		0.000	01:47
Control Test		0.077	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 09/05/2020 01:59:14

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.078	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.118	01:53
Breath Volume		1.812	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.122	01:57
Breath Volume		1.289	01:57
Air Blank		0.000	01:58
Control Test		0.078	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,R
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRASFIELD

DATE/TIME OF TEST: 09/05/2020 03:01:43

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.161	02:57
Breath Volume		2.945	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.153	03:00
Breath Volume		2.617	03:00
Air Blank		0.000	03:00
Control Test		0.076	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/05/2020 03:51:05

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.077	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.309	03:46
Breath Volume		2.847	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:48
Subject Sample	# 2	0.309	03:49
Breath Volume		1.398	03:49
Air Blank		0.000	03:50
Control Test		0.076	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,R
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: LAMPP

DATE/TIME OF TEST: 09/06/2020 00:24:22

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:23	
Control Test	0.078	00:23	
Air Blank		00:23	RFI Detect
Air Blank	0.000	00:24	

Data Download Date/Time: 10/06/2020 12:31

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,R
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: LAMPP

DATE/TIME OF TEST: 09/06/2020 00:32:19

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.078	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.092	00:27
Breath Volume		2.183	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.093	00:30
Breath Volume		3.937	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/06/2020 04:54:07

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.078	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.158	04:49
Breath Volume		2.414	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.162	04:52
Breath Volume		2.125	04:52
Air Blank		0.000	04:53
Control Test		0.078	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 09/06/2020 07:31:26

Diagnostics Check		OK	07:25
Air Blank		0.000	07:25
Control Test		0.078	07:26
Air Blank		0.000	07:26
Subject Sample	# 1	0.102	07:27
Breath Volume		1.648	07:27
Air Blank		0.000	07:27
Air Blank		0.000	07:29
Subject Sample	# 2	0.099	07:29
Breath Volume		1.351	07:29
Air Blank		0.000	07:30
Control Test		0.078	07:30
Air Blank		0.000	07:31
Diagnostics Check		OK	07:31

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: PERSON

DATE/TIME OF TEST: 09/09/2020 17:10:51

Diagnostics Check		OK	17:04
Air Blank		0.000	17:04
Control Test		0.078	17:05
Air Blank		0.000	17:05
Subject Sample	# 1	0.049	17:06
Breath Volume		2.132	17:06
Air Blank		0.000	17:06
Air Blank		0.000	17:08
Subject Sample	# 2	0.045	17:09
Breath Volume		1.671	17:09
Air Blank		0.000	17:09
Control Test		0.077	17:10
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/11/2020 17:07:11

Diagnostics Check	OK	17:06	
Air Blank	0.000	17:06	
		17:07	RFI Detect
Air Blank		17:07	RFI Detect

Data Download Date/Time: 10/06/2020 12:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: B650423652170

DATE/TIME OF TEST: 09/11/2020 17:15:54

Diagnostics Check		OK	17:09
Air Blank		0.000	17:09
Control Test		0.078	17:09
Air Blank		0.000	17:10
Subject Sample	# 1	0.000	17:11
Breath Volume		2.351	17:11
Air Blank		0.000	17:12
Air Blank		0.000	17:14
Subject Sample	# 2	0.000	17:14
Breath Volume		1.722	17:14
Air Blank		0.000	17:15
Control Test		0.076	17:15
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SKINNER,Z
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CALLAHAN

DATE/TIME OF TEST: 09/11/2020 22:56:17

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:51	
Control Test		0.078	22:51	
Air Blank		0.000	22:52	
Subject Sample	# 1	0.156	22:52	
Breath Volume		2.281	22:52	
Air Blank		0.000	22:53	
Air Blank		0.000	22:55	
Subject Sample	# 2	0.156	22:55	
Breath Volume		2.332	22:55	
Air Blank			22:55	RFI Detect
Air Blank		0.000	22:56	

Data Download Date/Time: 10/06/2020 12:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ENGLISH,T
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 09/19/2020 01:18:44

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.078	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.000	01:14
Breath Volume		2.109	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:17
Breath Volume		2.167	01:17
Air Blank		0.000	01:17
Control Test		0.077	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,JOSEPH,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOHNSTON,JOSEPH
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 09/19/2020 15:53:11

Diagnostics Check		OK	15:47
Air Blank		0.000	15:47
Control Test		0.079	15:47
Air Blank		0.000	15:48
Subject Sample	# 1	0.000	15:48
Breath Volume		3.542	15:48
Air Blank		0.000	15:49
Air Blank		0.000	15:51
Subject Sample	# 2	0.000	15:51
Breath Volume		2.218	15:51
Air Blank		0.000	15:52
Control Test		0.077	15:52
Air Blank		0.000	15:53
Diagnostics Check		OK	15:53

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PARISI,T
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCCUMBER

DATE/TIME OF TEST: 09/22/2020 21:36:29

Diagnostics Check		OK	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.135	21:32
Breath Volume		2.136	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.144	21:34
Breath Volume		2.453	21:34
Air Blank		0.000	21:35
Control Test		0.078	21:35
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROPER,SEAN,C
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROPER

DATE/TIME OF TEST: 09/23/2020 02:39:32

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.079	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 09/24/2020 23:53:10

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.079	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.106	23:48
Breath Volume		2.718	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.097	23:51
Breath Volume		1.925	23:51
Air Blank		0.000	23:52
Control Test		0.076	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 10/06/2020 12:32

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MUSGROVE,USTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DANCE

DATE/TIME OF TEST: 09/26/2020 23:25:01

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:21	
Control Test		0.078	23:22	
Air Blank		0.000	23:22	
Subject Sample	# 1		23:23	Subject Test Refused
Breath Volume		0.000	23:23	
Air Blank		0.000	23:24	
Control Test		0.078	23:24	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 10/06/2020 12:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: NASH,J
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DAMPIER

DATE/TIME OF TEST: 09/27/2020 01:16:28

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.078	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:12	No Sample Provided
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 2		01:15	Subject Test Refused
Breath Volume		0.000	01:15	
Air Blank		0.000	01:15	
Control Test		0.000	01:15	RFI Detect
Air Blank		0.000	01:16	

Data Download Date/Time: 10/06/2020 12:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: WYNNE,AUSTIN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 09/28/2020 00:08:26

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.078	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1	0.144	00:03	
Breath Volume		1.550	00:03	
Air Blank		0.000	00:04	
Air Blank		0.000	00:06	
Subject Sample	# 2	0.134	00:06	
Breath Volume		1.644	00:06	
Air Blank		0.000	00:07	
Control Test		0.074	00:07	Control Outside Tolerance
Air Blank		0.000	00:08	

Data Download Date/Time: 10/06/2020 12:33

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: MELIA

DATE/TIME OF TEST: 09/30/2020 17:14:55

Diagnostics Check		OK	17:09	
Air Blank		0.000	17:09	
Control Test		0.080	17:09	
Air Blank		0.000	17:10	
Subject Sample	# 1	0.204	17:10	
Breath Volume		1.613	17:10	
Air Blank		0.000	17:11	
Air Blank		0.000	17:13	
Subject Sample	# 2	0.186	17:13	
Breath Volume		1.480	17:13	
Air Blank			17:14	RFI Detect
Air Blank		0.000	17:14	

Data Download Date/Time: 10/06/2020 12:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,WESLEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA,COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 09/06/2020 05:44:56

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:36	
Control Test		0.079	05:36	
Air Blank		0.000	05:36	
Subject Sample	# 1	0.106	05:40	Volume Not Met
Breath Volume		0.589	05:40	
Air Blank		0.000	05:40	
Air Blank		0.000	05:42	
Subject Sample	# 2		05:43	Subject Test Refused
Breath Volume		0.000	05:43	
Air Blank		0.000	05:44	
Control Test		0.074	05:44	Control Outside Tolerance
Air Blank		0.000	05:44	

Data Download Date/Time: 09/09/2020 11:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,WESLEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA,COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A2
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 09/06/2020 06:02:40

Diagnostics Check		OK	05:47	
Air Blank		0.000	05:48	
Control Test		0.077	05:48	
Air Blank		0.000	05:49	
Subject Sample	# 1	0.094	05:52	Volume Not Met
Breath Volume		0.714	05:52	
Air Blank		0.000	05:53	
Air Blank		0.000	05:54	
Subject Sample	# 2	0.102	05:57	
Breath Volume		1.363	05:57	
Air Blank		0.000	05:58	
Air Blank		0.000	06:00	
Subject Sample	# 3	0.112	06:01	
Breath Volume		2.015	06:01	
Air Blank		0.000	06:01	
Control Test		0.075	06:02	
Air Blank		0.000	06:02	
Diagnostics Check		OK	06:02	

Data Download Date/Time: 09/09/2020 11:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,JUSTIN,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GARRISON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: SESSIONS

DATE/TIME OF TEST: 09/02/2020 12:48:54

Diagnostics Check		OK	12:44	
Air Blank		0.000	12:45	
Control Test		0.082	12:45	
Air Blank		0.000	12:46	
Subject Sample	# 1		12:47	Subject Test Refused
Breath Volume		0.000	12:47	
Air Blank		0.000	12:47	
Control Test		0.082	12:48	
Air Blank		0.000	12:48	
Diagnostics Check		OK	12:48	

Data Download Date/Time: 09/05/2020 20:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAIG,SHAWN,T
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: ESTEVEZ,MICHAEL
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 09/07/2020 03:54:11

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.083	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.234	03:48
Breath Volume		1.429	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.251	03:52
Breath Volume		3.097	03:52
Air Blank		0.000	03:53
Control Test		0.082	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 11/13/2020 16:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN,JASON,T
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: RHODEN,JASON
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/07/2020 20:26:17

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.183	20:25	Control Outside Tolerance
Air Blank	0.000	20:26	

Data Download Date/Time: 11/13/2020 16:02

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/02/2020 03:17:04

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.082	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.180	03:11
Breath Volume		2.179	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.186	03:15
Breath Volume		2.535	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:17

Data Download Date/Time: 09/04/2020 13:48

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YASUDA,SHOHEI,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YASUDA,SHOHEI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: HEWETT

DATE/TIME OF TEST: 09/05/2020 18:59:16

Diagnostics Check		OK	18:52
Air Blank		0.000	18:52
Control Test		0.082	18:53
Air Blank		0.000	18:53
Subject Sample	# 1	0.067	18:54
Breath Volume		1.128	18:54
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.064	18:57
Breath Volume		2.183	18:57
Air Blank		0.000	18:58
Control Test		0.081	18:58
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 09/06/2020 21:26:58

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.082	21:20
Air Blank		0.000	21:21
Subject Sample	# 1	0.171	21:22
Breath Volume		1.832	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:24
Subject Sample	# 2	0.168	21:25
Breath Volume		1.855	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 09/12/2020 00:56:16

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.081	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.303	00:51
Breath Volume		1.613	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.293	00:54
Breath Volume		1.433	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RODENBAUGH

DATE/TIME OF TEST: 09/12/2020 23:44:08

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.082	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.206	23:35
Breath Volume		1.867	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.185	23:38
Breath Volume		1.242	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 3	0.204	23:42
Breath Volume		1.152	23:42
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 09/16/2020 18:30:50

Diagnostics Check		OK	18:26	
Air Blank		0.000	18:26	
Control Test		0.081	18:26	
Air Blank		0.000	18:27	
Subject Sample	# 1	0.225	18:29	
Breath Volume		1.468	18:29	
Air Blank			18:30	Purge Fail
Air Blank		0.000	18:30	

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 09/16/2020 18:34:54

Diagnostics Check	OK	18:33	
Air Blank	0.000	18:33	
Control Test	0.073	18:34	Control Outside Tolerance
Air Blank	0.000	18:34	

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 09/16/2020 18:44:57

Diagnostics Check		OK	18:37
Air Blank		0.000	18:37
Control Test		0.079	18:37
Air Blank		0.000	18:38
Subject Sample	# 1	0.234	18:39
Breath Volume		1.710	18:39
Air Blank		0.000	18:40
Air Blank		0.000	18:41
Subject Sample	# 2	0.227	18:43
Breath Volume		1.382	18:43
Air Blank		0.000	18:43
Control Test		0.076	18:44
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/20/2020 01:38:28

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.082	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.177	01:33
Breath Volume		2.906	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.169	01:36
Breath Volume		2.480	01:36
Air Blank		0.000	01:37
Control Test		0.078	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 10/07/2020 07:41

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/24/2020 16:22:43

Diagnostics Check		OK	16:16
Air Blank		0.000	16:16
Control Test		0.080	16:16
Air Blank		0.000	16:17
Subject Sample	# 1	0.091	16:17
Breath Volume		1.121	16:17
Air Blank		0.000	16:18
Air Blank		0.000	16:20
Subject Sample	# 2	0.087	16:21
Breath Volume		1.421	16:21
Air Blank		0.000	16:21
Control Test		0.081	16:22
Air Blank		0.000	16:22
Diagnostics Check		OK	16:22

Data Download Date/Time: 10/16/2020 13:49

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SIMMS,KYLE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 09/06/2020 04:07:45

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.171	04:03
Breath Volume		2.628	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.180	04:06
Breath Volume		3.359	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 10/02/2020 15:14

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CARD

DATE/TIME OF TEST: 09/25/2020 22:08:09

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:57	
Control Test		0.078	21:57	
Air Blank		0.000	21:58	
Subject Sample	# 1		22:01	No Sample Provided
Breath Volume		0.000	22:01	
Air Blank		0.000	22:01	
Air Blank		0.000	22:03	
Subject Sample	# 2		22:06	No Sample Provided
Breath Volume		0.000	22:06	
Air Blank		0.000	22:07	
Control Test		0.079	22:07	
Air Blank		0.000	22:08	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 10/02/2020 15:14

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: CARD

DATE/TIME OF TEST: 09/25/2020 22:23:35

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.079	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.140	22:19
Breath Volume		2.226	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.149	22:22
Breath Volume		2.589	22:22
Air Blank		0.000	22:22
Control Test		0.079	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 10/02/2020 15:14

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 09/27/2020 01:20:26

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.200	01:15
Breath Volume		2.324	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.199	01:18
Breath Volume		1.820	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 10/02/2020 15:14

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEER,BRADLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: MEER,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/13/2020 06:23:43

Diagnostics Check		OK	06:16
Air Blank		0.000	06:16
Control Test		0.080	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.000	06:19
Breath Volume		2.281	06:19
Air Blank		0.000	06:20
Air Blank		0.000	06:22
Subject Sample	# 2	0.000	06:22
Breath Volume		2.117	06:22
Air Blank		0.000	06:22
Control Test		0.079	06:23
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 10/22/2020 15:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: MENDOZA ALAVA

DATE/TIME OF TEST: 09/15/2020 22:32:19

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:29	
Control Test		0.079	22:30	
Air Blank		0.000	22:30	
Subject Sample	# 1		22:30	Subject Test Refused
Breath Volume		0.000	22:30	
Air Blank		0.000	22:31	
Control Test		0.079	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 10/22/2020 15:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: GONZALEZ ORELLANA

DATE/TIME OF TEST: 09/17/2020 23:16:27

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.079	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.110	23:11
Breath Volume		2.707	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.110	23:14
Breath Volume		2.933	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 10/22/2020 15:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 09/18/2020 02:02:06

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.199	01:57
Breath Volume		3.203	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.197	02:00
Breath Volume		2.019	02:00
Air Blank		0.000	02:01
Control Test		0.077	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 10/22/2020 15:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LAROCCA

DATE/TIME OF TEST: 09/19/2020 19:57:00

Diagnostics Check		OK	19:45	
Air Blank		0.000	19:45	
Control Test		0.079	19:46	
Air Blank		0.000	19:46	
Subject Sample	# 1	0.021	19:49	Volume Not Met
Breath Volume		0.468	19:49	
Air Blank		0.000	19:50	
Air Blank		0.000	19:52	
Subject Sample	# 2	0.018	19:55	Volume Not Met
Breath Volume		0.335	19:55	
Air Blank		0.000	19:56	
Control Test		0.078	19:56	
Air Blank		0.000	19:56	
Diagnostics Check		OK	19:57	

Data Download Date/Time: 10/22/2020 15:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: LAROCCA

DATE/TIME OF TEST: 09/19/2020 20:05:45

Diagnostics Check		OK	19:59
Air Blank		0.000	19:59
Control Test		0.078	19:59
Air Blank		0.000	20:00
Subject Sample	# 1	0.018	20:00
Breath Volume		1.859	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.017	20:04
Breath Volume		1.953	20:04
Air Blank		0.000	20:04
Control Test		0.078	20:05
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 10/22/2020 15:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 09/20/2020 02:47:23

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.223	02:42
Breath Volume		2.152	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.228	02:45
Breath Volume		2.617	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRCHNER,MATTHEW,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: STICKNEY,RONALD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/20/2020 04:28:33

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.156	04:23
Breath Volume		2.386	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.149	04:26
Breath Volume		1.777	04:26
Air Blank		0.000	04:27
Control Test		0.077	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HATT,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HATT,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 09/20/2020 22:59:12

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:47	
Control Test		0.079	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 1	0.160	22:51	Volume Not Met
Breath Volume		0.156	22:51	
Air Blank		0.000	22:52	
Air Blank		0.000	22:53	
Subject Sample	# 2	0.210	22:54	
Breath Volume		2.734	22:54	
Air Blank		0.000	22:55	
Air Blank		0.000	22:57	
Subject Sample	# 3	0.206	22:57	
Breath Volume		3.230	22:57	
Air Blank		0.000	22:58	
Control Test		0.078	22:58	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CHIP,STEVE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: MAZZOCCO

DATE/TIME OF TEST: 09/22/2020 23:05:10

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.079	22:54	
Air Blank		0.000	22:54	
Subject Sample	# 1	0.270	22:57	Volume Not Met
Breath Volume		0.460	22:57	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 2	0.247	23:03	Volume Not Met
Breath Volume		0.449	23:03	
Air Blank		0.000	23:04	
Control Test		0.076	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CHIP,STEVE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03519080A5
DRY GAS EXPIRY DATE: 04/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: MAZZOCCO

DATE/TIME OF TEST: 09/22/2020 23:19:05

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:07	
Control Test		0.077	23:08	
Air Blank		0.000	23:08	
Subject Sample	# 1	0.253	23:11	Volume Not Met
Breath Volume		0.105	23:11	
Air Blank		0.000	23:12	
Air Blank		0.000	23:14	
Subject Sample	# 2	0.232	23:17	Volume Not Met
Breath Volume		0.105	23:17	
Air Blank		0.000	23:18	
Control Test		0.066	23:18	Control Outside Tolerance
Air Blank		0.000	23:19	

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERAVIL,JEAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: DERAVIL,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: F27413816804916

DATE/TIME OF TEST: 09/26/2020 20:06:27

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.080	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.000	20:02
Breath Volume		5.250	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.000	20:05
Breath Volume		5.425	20:05
Air Blank		0.000	20:05
Control Test		0.080	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: BOYLE

DATE/TIME OF TEST: 09/29/2020 16:11:16

Diagnostics Check		OK	16:04
Air Blank		0.000	16:05
Control Test		0.080	16:05
Air Blank		0.000	16:06
Subject Sample	# 1	0.000	16:06
Breath Volume		3.304	16:06
Air Blank		0.000	16:07
Air Blank		0.000	16:09
Subject Sample	# 2	0.000	16:09
Breath Volume		2.410	16:09
Air Blank		0.000	16:10
Control Test		0.080	16:10
Air Blank		0.000	16:11
Diagnostics Check		OK	16:11

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 09/10/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LARSEN

DATE/TIME OF TEST: 09/30/2020 00:22:18

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.080	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.212	00:17
Breath Volume		2.140	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.211	00:20
Breath Volume		1.937	00:20
Air Blank		0.000	00:21
Control Test		0.080	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 10/22/2020 15:39

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/03/2020 22:32:10

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.081	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.000	22:28
Breath Volume		3.222	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.000	22:30
Breath Volume		3.382	22:30
Air Blank		0.000	22:31
Control Test		0.079	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 09/11/2020 14:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: RHINIER

DATE/TIME OF TEST: 09/04/2020 22:54:11

Diagnostics Check		OK	22:47
Air Blank		0.000	22:48
Control Test		0.081	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.132	22:49
Breath Volume		2.660	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.135	22:52
Breath Volume		1.753	22:52
Air Blank		0.000	22:53
Control Test		0.081	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 09/11/2020 14:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/05/2020 00:41:39

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.214	00:36
Breath Volume		1.933	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.199	00:40
Breath Volume		1.871	00:40
Air Blank		0.000	00:40
Control Test		0.079	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 09/11/2020 14:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEER,BRADLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: MEER,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/19/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PADGITT

DATE/TIME OF TEST: 09/08/2020 01:07:24

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.187	01:02
Breath Volume		1.417	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.182	01:05
Breath Volume		1.296	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 09/11/2020 14:57

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: ALDAY,SETH
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2121
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/02/2020 19:32:41

Diagnostics Check		OK	19:26
Air Blank		0.000	19:27
Control Test		0.078	19:27
Air Blank		0.000	19:27
Subject Sample	# 1	0.131	19:28
Breath Volume		3.343	19:28
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.126	19:31
Breath Volume		2.843	19:31
Air Blank		0.000	19:31
Control Test		0.078	19:32
Air Blank		0.000	19:32
Diagnostics Check		OK	19:32

Data Download Date/Time: 09/28/2020 13:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY,WILLIAM,Q
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: COLMENARES

DATE/TIME OF TEST: 09/05/2020 23:33:05

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.171	23:28
Breath Volume		2.925	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.167	23:31
Breath Volume		2.593	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 09/28/2020 13:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: WEBB,ANDREW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2121
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BAGGETT

DATE/TIME OF TEST: 09/16/2020 18:28:53

Diagnostics Check		OK	18:26	
Air Blank		0.000	18:27	
Control Test		0.078	18:27	
Air Blank		0.000	18:28	
Subject Sample	# 1		18:28	Sequence Aborted
Breath Volume		0.000	18:28	
Air Blank		0.000	18:28	

Data Download Date/Time: 09/28/2020 13:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURCH,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: WEBB,ANDREW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2121
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BAGGETT

DATE/TIME OF TEST: 09/16/2020 18:56:33

Diagnostics Check		OK	18:53	
Air Blank		0.000	18:54	
Control Test		0.079	18:54	
Air Blank		0.000	18:54	
Subject Sample	# 1		18:55	Subject Test Refused
Breath Volume		0.000	18:55	
Air Blank		0.000	18:55	
Control Test		0.079	18:55	
Air Blank		0.000	18:56	
Diagnostics Check		OK	18:56	

Data Download Date/Time: 09/28/2020 13:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICKEY,WILLIAM,Q
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08819080A1
DRY GAS EXPIRY DATE: 06/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 09/19/2020 03:06:38

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:04	
Control Test		0.079	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:05	
Control Test		0.079	03:06	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 09/28/2020 13:56

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,BRADLEY,A
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: THOMAS,LARRY
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DOW

DATE/TIME OF TEST: 09/06/2020 03:27:50

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.077	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.000	03:23
Breath Volume		3.015	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.000	03:26
Breath Volume		2.886	03:26
Air Blank		0.000	03:26
Control Test		0.075	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 09/28/2020 18:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STOUT,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PETTIGREW

DATE/TIME OF TEST: 09/08/2020 01:19:38

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.071	01:19	Control Outside Tolerance
Air Blank	0.000	01:19	

Data Download Date/Time: 09/28/2020 18:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: STOUT,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PETTIGREW

DATE/TIME OF TEST: 09/08/2020 01:27:44

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.076	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.239	01:23
Breath Volume		2.886	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.227	01:26
Breath Volume		2.382	01:26
Air Blank		0.000	01:26
Control Test		0.075	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 09/28/2020 18:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ALMESTICA

DATE/TIME OF TEST: 09/13/2020 22:45:11

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.076	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.077	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 09/28/2020 18:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 09/01/2020 00:41:49

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.084	00:37
Breath Volume		1.281	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.087	00:40
Breath Volume		1.609	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 09/02/2020 18:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,DANIEL
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: VALDER

DATE/TIME OF TEST: 09/01/2020 18:26:32

Diagnostics Check		OK	18:19
Air Blank		0.000	18:20
Control Test		0.081	18:20
Air Blank		0.000	18:21
Subject Sample	# 1	0.034	18:22
Breath Volume		3.046	18:22
Air Blank		0.000	18:22
Air Blank		0.000	18:24
Subject Sample	# 2	0.033	18:25
Breath Volume		3.335	18:25
Air Blank		0.000	18:25
Control Test		0.080	18:25
Air Blank		0.000	18:26
Diagnostics Check		OK	18:26

Data Download Date/Time: 09/02/2020 18:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: COLLAZOS

DATE/TIME OF TEST: 09/02/2020 02:59:17

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.080	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:55	No Sample Provided
Breath Volume		0.000	02:55	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 09/02/2020 18:44

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GARTNER,ART
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/04/2020 02:53:31

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.081	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:52	RFI Detect
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.081	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 09/17/2020 13:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GARTNER,ART
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/04/2020 03:10:33

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.081	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1	0.150	03:00	Volume Not Met
Breath Volume		0.265	03:00	
Air Blank		0.000	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 2	0.141	03:05	
Breath Volume		1.320	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 3	0.145	03:08	
Breath Volume		1.140	03:08	
Air Blank		0.000	03:09	
Control Test		0.077	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 09/17/2020 13:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITLLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/05/2020 20:12:03

Diagnostics Check		OK	20:08	
Air Blank		0.000	20:08	
Control Test		0.081	20:09	
Air Blank		0.000	20:09	
Subject Sample	# 1		20:10	Subject Test Refused
Breath Volume		0.000	20:10	
Air Blank		0.000	20:11	
Control Test		0.081	20:11	
Air Blank		0.000	20:11	
Diagnostics Check		OK	20:12	

Data Download Date/Time: 09/17/2020 13:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: COUCH

DATE/TIME OF TEST: 09/06/2020 04:20:50

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.082	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.000	04:16
Breath Volume		2.710	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.000	04:19
Breath Volume		1.515	04:19
Air Blank		0.000	04:19
Control Test		0.081	04:20
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 09/17/2020 13:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIF,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: HALLOCK

DATE/TIME OF TEST: 09/08/2020 17:22:31

Diagnostics Check		OK	17:19	
Air Blank		0.000	17:20	
Control Test		0.080	17:20	
Air Blank		0.000	17:20	
Subject Sample	# 1		17:21	Subject Test Refused
Breath Volume		0.000	17:21	
Air Blank		0.000	17:21	
Control Test		0.080	17:21	
Air Blank		0.000	17:22	
Diagnostics Check		OK	17:22	

Data Download Date/Time: 09/17/2020 13:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CHOMIAK

DATE/TIME OF TEST: 09/13/2020 00:18:55

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:07	
Control Test		0.080	00:07	
Air Blank		0.000	00:08	
Subject Sample	# 1	0.190	00:11	Volume Not Met
Breath Volume		0.058	00:11	
Air Blank		0.000	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 2	0.174	00:17	Volume Not Met
Breath Volume		0.347	00:17	
Air Blank		0.000	00:17	
Control Test		0.077	00:18	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 09/17/2020 13:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,DANIEL,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CHAVEZ,T
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MANGRACINA

DATE/TIME OF TEST: 09/19/2020 04:48:53

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:44	
Control Test		0.081	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1		04:48	Sequence Aborted
Breath Volume		0.000	04:48	
Air Blank		0.000	04:48	

Data Download Date/Time: 10/23/2020 12:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: LAFONTAINE

DATE/TIME OF TEST: 09/19/2020 18:45:02

Diagnostics Check		OK	18:41	
Air Blank		0.000	18:42	
Control Test		0.080	18:42	
Air Blank		0.000	18:43	
Subject Sample	# 1		18:43	Subject Test Refused
Breath Volume		0.000	18:43	
Air Blank		0.000	18:44	
Control Test		0.080	18:44	
Air Blank		0.000	18:44	
Diagnostics Check		OK	18:45	

Data Download Date/Time: 10/23/2020 12:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,DANIEL,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GEORGE,S
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: IRENE

DATE/TIME OF TEST: 09/20/2020 05:47:49

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.081	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.286	05:42
Breath Volume		2.632	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.276	05:46
Breath Volume		1.804	05:46
Air Blank		0.000	05:46
Control Test		0.079	05:47
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 10/23/2020 12:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 09/21/2020 23:07:03

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:04	
Control Test		0.081	23:04	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:06	
Control Test		0.081	23:06	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 10/23/2020 12:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: RADECKI

DATE/TIME OF TEST: 09/24/2020 03:06:03

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.081	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.222	03:01
Breath Volume		2.722	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:03
Subject Sample	# 2	0.215	03:04
Breath Volume		2.226	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:06

Data Download Date/Time: 10/23/2020 12:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 09/24/2020 23:59:43

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.082	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.109	23:55
Breath Volume		1.347	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.101	23:58
Breath Volume		1.441	23:58
Air Blank		0.000	23:58
Control Test		0.080	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 10/23/2020 12:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 09/27/2020 00:46:50

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.000	00:42
Breath Volume		2.308	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.000	00:45
Breath Volume		2.062	00:45
Air Blank		0.000	00:45
Control Test		0.081	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 10/23/2020 12:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BARD,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: NEW

DATE/TIME OF TEST: 09/27/2020 02:59:52

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.081	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.239	02:54
Breath Volume		2.226	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.233	02:58
Breath Volume		2.007	02:58
Air Blank		0.000	02:58
Control Test		0.078	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 10/23/2020 12:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DRAKE,N
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/27/2020 05:26:22

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:22	
Control Test		0.080	05:23	
Air Blank		0.000	05:23	
Subject Sample	# 1		05:24	Subject Test Refused
Breath Volume		0.000	05:24	
Air Blank		0.000	05:25	
Control Test		0.081	05:25	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 10/23/2020 12:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: OCONNER

DATE/TIME OF TEST: 09/27/2020 23:17:21

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:05	
Control Test		0.081	23:06	
Air Blank		0.000	23:06	
Subject Sample	# 1		23:09	No Sample Provided
Breath Volume		0.000	23:09	
Air Blank		0.000	23:10	
Air Blank		0.000	23:12	
Subject Sample	# 2	0.164	23:15	Volume Not Met
Breath Volume		0.648	23:15	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:17	
Diagnostics Check		OK	23:17	

Data Download Date/Time: 10/23/2020 12:04

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/17/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: OCONNER

DATE/TIME OF TEST: 09/27/2020 23:23:44

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.079	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.080	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 10/23/2020 12:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/05/2020 02:54:25

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.000	02:50
Breath Volume		2.750	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.000	02:53
Breath Volume		3.109	02:53
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: OSBURN

DATE/TIME OF TEST: 09/05/2020 23:49:45

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.227	23:44
Breath Volume		1.492	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:46
Subject Sample	# 2	0.235	23:48
Breath Volume		1.402	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/06/2020 01:57:17

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.000	01:52
Breath Volume		2.882	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.000	01:55
Breath Volume		2.691	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: SEDNER

DATE/TIME OF TEST: 09/08/2020 20:03:03

Diagnostics Check		OK	20:00	
Air Blank		0.000	20:00	
Control Test		0.080	20:01	
Air Blank		0.000	20:01	
Subject Sample	# 1		20:02	RFI Detect
Breath Volume		0.128	20:02	
Air Blank			20:02	RFI Detect
Air Blank		0.000	20:03	

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: SEDNER

DATE/TIME OF TEST: 09/08/2020 20:16:09

Diagnostics Check		OK	20:04	
Air Blank		0.000	20:04	
Control Test		0.080	20:05	
Air Blank		0.000	20:05	
Subject Sample	# 1	0.229	20:08	Volume Not Met
Breath Volume		0.078	20:08	
Air Blank		0.000	20:09	
Air Blank		0.000	20:11	
Subject Sample	# 2	0.232	20:14	Volume Not Met
Breath Volume		0.015	20:14	
Air Blank		0.000	20:15	
Control Test		0.078	20:15	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,C,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: SEDNER

DATE/TIME OF TEST: 09/08/2020 20:19:01

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:18	
Control Test	0.074	20:18	Control Outside Tolerance
Air Blank	0.000	20:19	

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/13/2020 01:37:38

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.079	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.000	01:33
Breath Volume		1.359	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.000	01:36
Breath Volume		1.492	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: LEWANDOWSKI

DATE/TIME OF TEST: 09/26/2020 01:55:00

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:49	
Control Test		0.080	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1	0.161	01:50	
Breath Volume		1.464	01:50	
Air Blank		0.000	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 2		01:53	RFI Detect
Breath Volume		0.082	01:53	
Air Blank		0.000	01:54	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,C,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: LEWANDOWSKI

DATE/TIME OF TEST: 09/26/2020 01:59:25

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.078	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	RFI Detect
Breath Volume		0.000	01:58	
Air Blank		0.000	01:58	
Control Test		0.032	01:58	RFI Detect
Air Blank		0.000	01:59	

Data Download Date/Time: 09/29/2020 10:22

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CANNA,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 09/01/2020 19:33:42

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.079	19:27
Air Blank		0.000	19:27
Subject Sample	# 1	0.235	19:29
Breath Volume		1.445	19:29
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.248	19:32
Breath Volume		1.609	19:32
Air Blank		0.000	19:32
Control Test		0.078	19:33
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 09/05/2020 01:43:49

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.198	01:38
Breath Volume		1.683	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.193	01:42
Breath Volume		1.539	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 09/05/2020 22:36:32

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:33	
Control Test		0.079	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 1		22:35	Subject Test Refused
Breath Volume		0.000	22:35	
Air Blank		0.000	22:35	
Control Test		0.079	22:35	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:36	

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GRINON,REY
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: MOELLER

DATE/TIME OF TEST: 09/07/2020 19:38:06

Diagnostics Check		OK	19:35	
Air Blank		0.000	19:35	
Control Test		0.079	19:36	
Air Blank		0.000	19:36	
Subject Sample	# 1		19:36	Subject Test Refused
Breath Volume		0.000	19:36	
Air Blank		0.000	19:37	
Control Test		0.079	19:37	
Air Blank		0.000	19:38	
Diagnostics Check		OK	19:38	

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ZINN,BRYAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: BRONSON

DATE/TIME OF TEST: 09/07/2020 21:40:22

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:37	
Control Test		0.079	21:38	
Air Blank		0.000	21:38	
Subject Sample	# 1		21:38	Subject Test Refused
Breath Volume		0.000	21:38	
Air Blank		0.000	21:39	
Control Test		0.080	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CARNEY

DATE/TIME OF TEST: 09/08/2020 00:45:07

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.080	00:43	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.079	00:44	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LEE,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: OELKE

DATE/TIME OF TEST: 09/09/2020 21:52:36

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:41	
Control Test		0.079	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1	0.000	21:45	Volume Not Met
Breath Volume		0.019	21:45	
Air Blank		0.000	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 2	0.156	21:50	Volume Not Met
Breath Volume		0.000	21:50	
Air Blank		0.000	21:51	
Control Test		0.079	21:51	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 09/12/2020 01:13:33

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.192	01:08
Breath Volume		2.316	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.185	01:11
Breath Volume		2.207	01:11
Air Blank		0.000	01:12
Control Test		0.077	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: AMAR,GIOVANNI
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: CUMINGS

DATE/TIME OF TEST: 09/12/2020 22:59:03

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.079	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.246	22:54
Breath Volume		1.265	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:56
Subject Sample	# 2	0.247	22:57
Breath Volume		1.433	22:57
Air Blank		0.000	22:58
Control Test		0.078	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:59

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP MTZ,FELICIA,B
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: TEMPLIN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: KEESHAN

DATE/TIME OF TEST: 09/17/2020 09:04:36

Diagnostics Check		OK	08:57
Air Blank		0.000	08:58
Control Test		0.079	08:58
Air Blank		0.000	08:59
Subject Sample	# 1	0.194	09:00
Breath Volume		1.843	09:00
Air Blank		0.000	09:00
Air Blank		0.000	09:02
Subject Sample	# 2	0.188	09:02
Breath Volume		2.402	09:02
Air Blank		0.000	09:03
Control Test		0.079	09:03
Air Blank		0.000	09:04
Diagnostics Check		OK	09:04

Data Download Date/Time: 09/25/2020 09:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ORTIZ,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: DECKER

DATE/TIME OF TEST: 09/18/2020 18:47:43

Diagnostics Check		OK	18:40
Air Blank		0.000	18:40
Control Test		0.080	18:40
Air Blank		0.000	18:41
Subject Sample	# 1	0.000	18:42
Breath Volume		1.175	18:42
Air Blank		0.000	18:42
Air Blank		0.000	18:44
Subject Sample	# 2	0.000	18:46
Breath Volume		1.914	18:46
Air Blank		0.000	18:46
Control Test		0.078	18:47
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 09/25/2020 09:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: JACKSON,M
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: ETHRIDGE

DATE/TIME OF TEST: 09/18/2020 23:28:44

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.080	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.134	23:24
Breath Volume		2.031	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.124	23:27
Breath Volume		1.761	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 09/25/2020 09:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PETERSON,M
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 09/19/2020 01:35:36

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.079	01:34	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 09/25/2020 09:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WYATT,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARTMANN,HANS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 09/20/2020 07:33:31

Diagnostics Check		OK	07:26
Air Blank		0.000	07:27
Control Test		0.080	07:27
Air Blank		0.000	07:27
Subject Sample	# 1	0.179	07:28
Breath Volume		1.207	07:28
Air Blank		0.000	07:29
Air Blank		0.000	07:31
Subject Sample	# 2	0.191	07:31
Breath Volume		1.593	07:31
Air Blank		0.000	07:32
Control Test		0.077	07:32
Air Blank		0.000	07:33
Diagnostics Check		OK	07:33

Data Download Date/Time: 09/25/2020 09:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: KIBBEY,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CINDRIC

DATE/TIME OF TEST: 09/21/2020 01:12:29

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:10	
Control Test		0.079	01:10	
Air Blank		0.000	01:10	
Subject Sample	# 1		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:11	
Control Test		0.078	01:11	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 09/25/2020 09:37

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LEE,A
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: LANGLEY

DATE/TIME OF TEST: 09/24/2020 02:29:21

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.077	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.284	02:24
Breath Volume		2.300	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.272	02:27
Breath Volume		1.894	02:27
Air Blank		0.000	02:28
Control Test		0.077	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 10/27/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: OTERO,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 09/25/2020 01:35:58

Diagnostics Check		OK	01:29
Air Blank		0.000	01:30
Control Test		0.077	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.000	01:31
Breath Volume		3.070	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.000	01:34
Breath Volume		2.398	01:34
Air Blank		0.000	01:35
Control Test		0.077	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 10/27/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCCONNELL,AUSTIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 09/26/2020 02:11:53

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.077	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.189	02:06
Breath Volume		2.046	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.174	02:10
Breath Volume		1.816	02:10
Air Blank		0.000	02:10
Control Test		0.077	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 10/27/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COPENHAVER

DATE/TIME OF TEST: 09/26/2020 04:25:10

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:22	
Control Test		0.077	04:23	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:23	Subject Test Refused
Breath Volume		0.000	04:23	
Air Blank		0.000	04:24	
Control Test		0.077	04:24	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 10/27/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: UNBERG,JEREMY
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 09/21/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 09/26/2020 23:52:43

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.078	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.000	23:48
Breath Volume		1.550	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.000	23:51
Breath Volume		2.007	23:51
Air Blank		0.000	23:51
Control Test		0.078	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 10/27/2020 13:05

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BOTTONA,TRP
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: VARVEL

DATE/TIME OF TEST: 09/03/2020 00:53:57

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.080	00:41	
Air Blank		0.000	00:41	
Subject Sample	# 1	0.118	00:45	Volume Not Met
Breath Volume		0.562	00:45	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 2	0.168	00:48	
Breath Volume		1.468	00:48	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 3	0.171	00:52	
Breath Volume		2.289	00:52	
Air Blank		0.000	00:53	
Control Test		0.079	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 12/11/2020 13:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JOHNSON 3

DATE/TIME OF TEST: 09/03/2020 01:21:02

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.080	01:20	
Air Blank		01:20	RFI Detect
Air Blank	0.000	01:21	

Data Download Date/Time: 12/11/2020 13:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JOHNSON 3

DATE/TIME OF TEST: 09/03/2020 01:29:46

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.095	01:24
Breath Volume		1.484	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.110	01:28
Breath Volume		2.023	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 12/11/2020 13:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIGER JR,CARL,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ELIAS,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: KEARNS

DATE/TIME OF TEST: 09/03/2020 21:08:11

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.080	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.122	21:03
Breath Volume		2.746	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.125	21:06
Breath Volume		1.910	21:06
Air Blank		0.000	21:07
Control Test		0.080	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 12/11/2020 13:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 09/04/2020 03:30:42

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.223	03:25
Breath Volume		2.246	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.216	03:29
Breath Volume		1.308	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 12/11/2020 13:36

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 09/01/2020 01:19:36

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.078	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Control Test		0.078	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 09/04/2020 18:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: LINDAUER

DATE/TIME OF TEST: 09/03/2020 18:06:49

Diagnostics Check		OK	18:00
Air Blank		0.000	18:00
Control Test		0.078	18:01
Air Blank		0.000	18:01
Subject Sample	# 1	0.000	18:02
Breath Volume		3.441	18:02
Air Blank		0.000	18:02
Air Blank		0.000	18:04
Subject Sample	# 2	0.000	18:05
Breath Volume		3.070	18:05
Air Blank		0.000	18:05
Control Test		0.079	18:06
Air Blank		0.000	18:06
Diagnostics Check		OK	18:06

Data Download Date/Time: 09/04/2020 18:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRY,DEP
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 09/04/2020 00:07:37

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:52	
Control Test		0.078	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 3	0.182	23:56	Volume Not Met
Breath Volume		0.816	23:56	
Air Blank		0.000	23:56	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.181	00:00	
Breath Volume		1.339	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 2	0.170	00:06	Volume Not Met
Breath Volume		0.296	00:06	
Air Blank		0.000	00:06	
Control Test		0.078	00:07	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 09/04/2020 18:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRY,DEP
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 09/04/2020 00:20:43

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:09	
Control Test		0.077	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 1	0.184	00:13	Volume Not Met
Breath Volume		0.863	00:13	
Air Blank		0.000	00:14	
Air Blank		0.000	00:16	
Subject Sample	# 2	0.160	00:19	Volume Not Met
Breath Volume		0.765	00:19	
Air Blank		0.000	00:19	
Control Test		0.077	00:20	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 09/04/2020 18:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRY,DEP
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 09/04/2020 00:30:56

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.076	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.174	00:26	Volume Not Met
Breath Volume		0.054	00:26	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Control Test		0.076	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 09/04/2020 18:01

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LEGRAND

DATE/TIME OF TEST: 09/12/2020 00:14:37

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.077	00:08
Air Blank		0.000	00:09
Subject Sample	# 1	0.141	00:09
Breath Volume		2.558	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.135	00:13
Breath Volume		1.566	00:13
Air Blank		0.000	00:13
Control Test		0.076	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 10/07/2020 22:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KARNES

DATE/TIME OF TEST: 09/12/2020 01:36:31

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.076	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.245	01:32
Breath Volume		1.664	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.251	01:34
Breath Volume		1.937	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 10/07/2020 22:23

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMPTER,SCOTT,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: SUMPTER,SCOTT
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: BUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: MANNO

DATE/TIME OF TEST: 09/05/2020 19:39:54

Diagnostics Check		OK	19:32
Air Blank		0.000	19:32
Control Test		0.081	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.196	19:34
Breath Volume		2.238	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:36
Subject Sample	# 2	0.204	19:38
Breath Volume		1.515	19:38
Air Blank		0.000	19:38
Control Test		0.079	19:39
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 10/07/2020 21:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 09/05/2020 23:56:27

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.081	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.274	23:51
Breath Volume		3.214	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.285	23:54
Breath Volume		2.929	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 10/07/2020 21:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WILD,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: BUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: ATALSKI

DATE/TIME OF TEST: 09/06/2020 23:41:43

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:39	
Control Test		0.082	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:40	
Control Test		0.082	23:41	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 10/07/2020 21:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: WHEELER,BROOKE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 09/07/2020 21:07:05

Diagnostics Check		OK	21:04	
Air Blank		0.000	21:04	
Control Test		0.081	21:04	
Air Blank		0.000	21:05	
Subject Sample	# 1		21:05	Subject Test Refused
Breath Volume		0.000	21:05	
Air Blank		0.000	21:06	
Control Test		0.082	21:06	
Air Blank		0.000	21:06	
Diagnostics Check		OK	21:07	

Data Download Date/Time: 10/07/2020 21:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HOOBLALL

DATE/TIME OF TEST: 09/09/2020 03:13:15

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.081	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:12	
Control Test		0.081	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 10/07/2020 21:53

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: SAVANT

DATE/TIME OF TEST: 09/09/2020 15:38:19

Diagnostics Check		OK	15:30
Air Blank		0.000	15:31
Control Test		0.082	15:31
Air Blank		0.000	15:32
Subject Sample	# 1	0.283	15:33
Breath Volume		3.218	15:33
Air Blank		0.000	15:34
Air Blank		0.000	15:35
Subject Sample	# 2	0.284	15:36
Breath Volume		3.117	15:36
Air Blank		0.000	15:37
Control Test		0.081	15:37
Air Blank		0.000	15:38
Diagnostics Check		OK	15:38

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOOM,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LABRECQUE,COREY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: DOUGLASS

DATE/TIME OF TEST: 09/10/2020 00:00:43

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:58	
Control Test		0.081	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Control Test		0.081	00:00	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MATT,A
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: RICKETTS,JEREMIAH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: CHESKEY

DATE/TIME OF TEST: 09/12/2020 07:02:01

Diagnostics Check		OK	06:59	
Air Blank		0.000	06:59	
Control Test		0.081	06:59	
Air Blank		0.000	07:00	
Subject Sample	# 1		07:00	Subject Test Refused
Breath Volume		0.000	07:00	
Air Blank		0.000	07:01	
Control Test		0.082	07:01	
Air Blank		0.000	07:01	
Diagnostics Check		OK	07:02	

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,JASON,N
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: TAYLOR,OCTAVIOUS
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: OSORIO

DATE/TIME OF TEST: 09/13/2020 07:16:39

Diagnostics Check		OK	07:02	
Air Blank		0.000	07:03	
Control Test		0.081	07:03	
Air Blank		0.000	07:03	
Subject Sample	# 1		07:07	No Sample Provided
Breath Volume		0.000	07:07	
Air Blank		0.000	07:07	
Air Blank		0.000	07:09	
Subject Sample	# 2	0.186	07:10	
Breath Volume		1.539	07:10	
Air Blank		0.000	07:11	
Air Blank		0.000	07:13	
Subject Sample	# 3	0.175	07:14	
Breath Volume		1.378	07:14	
Air Blank		0.000	07:15	
Control Test		0.080	07:16	
Air Blank		0.000	07:16	
Diagnostics Check		OK	07:16	

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: CARROLL,F
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: LUKASZEWSKI

DATE/TIME OF TEST: 09/16/2020 21:32:07

Diagnostics Check		OK	21:24
Air Blank		0.000	21:25
Control Test		0.082	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.164	21:27
Breath Volume		2.156	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:29
Subject Sample	# 2	0.159	21:30
Breath Volume		2.468	21:30
Air Blank		0.000	21:31
Control Test		0.080	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LAMP

DATE/TIME OF TEST: 09/17/2020 22:22:10

Diagnostics Check		OK	22:15
Air Blank		0.000	22:16
Control Test		0.081	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.136	22:17
Breath Volume		1.441	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:19
Subject Sample	# 2	0.135	22:20
Breath Volume		1.472	22:20
Air Blank		0.000	22:21
Control Test		0.081	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DOUG,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BAUDERS,SEAN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BELDING

DATE/TIME OF TEST: 09/18/2020 02:29:32

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.081	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.106	02:24
Breath Volume		3.277	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.098	02:27
Breath Volume		1.804	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: SILVERIO,ALVIN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 09/18/2020 22:10:38

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.211	22:05
Breath Volume		1.769	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.210	22:09
Breath Volume		1.738	22:09
Air Blank		0.000	22:09
Control Test		0.081	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: DELAROSA VENTURA

DATE/TIME OF TEST: 09/19/2020 00:19:36

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.195	00:15
Breath Volume		2.777	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.189	00:18
Breath Volume		1.890	00:18
Air Blank		0.000	00:18
Control Test		0.081	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MAGEE,N
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/20/2020 00:24:16

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.081	00:16
Air Blank		0.000	00:16
Subject Sample	# 1	0.176	00:19
Breath Volume		2.117	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.174	00:22
Breath Volume		2.296	00:22
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 10/07/2020 21:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WINGO,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MCELWAIN

DATE/TIME OF TEST: 09/20/2020 01:35:09

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.082	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.134	01:30
Breath Volume		2.437	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:32
Subject Sample	# 2	0.130	01:33
Breath Volume		1.945	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RUST

DATE/TIME OF TEST: 09/24/2020 23:47:30

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.081	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.137	23:42
Breath Volume		2.550	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.136	23:45
Breath Volume		2.445	23:45
Air Blank		0.000	23:46
Control Test		0.082	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KESSINGER,CHRIS
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALVARENGA

DATE/TIME OF TEST: 09/25/2020 04:32:02

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.082	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.000	04:27
Breath Volume		2.566	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.000	04:30
Breath Volume		1.742	04:30
Air Blank		0.000	04:31
Control Test		0.081	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:32

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,WILLIAM,S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKENZIE,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BONNETTE

DATE/TIME OF TEST: 09/26/2020 03:53:26

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.082	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.250	03:43	
Breath Volume		3.273	03:43	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.007	03:48	Volume Not Met
Breath Volume		0.265	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 3	0.231	03:51	
Breath Volume		1.984	03:51	
Air Blank		0.000	03:52	
Control Test		0.081	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURCHINSON,RONNIE,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: GETFORD,CHEYENNE
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ISOM

DATE/TIME OF TEST: 09/26/2020 23:52:41

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:48	
Control Test		0.082	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.082	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURCHINSON,RONNIE,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 09/28/2020 02:04:20

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.274	01:59
Breath Volume		1.613	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.275	02:02
Breath Volume		1.250	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CASTEEL

DATE/TIME OF TEST: 09/29/2020 22:33:25

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.082	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.222	22:27
Breath Volume		4.015	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.222	22:31
Breath Volume		3.882	22:31
Air Blank		0.000	22:32
Control Test		0.081	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 10/07/2020 21:54

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILLIAMS,JONATHAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 09/15/2020 22:15:28

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:12	
Control Test		0.081	22:12	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:14	
Control Test		0.079	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 09/28/2020 18:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DAGOSTINO,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: VERES

DATE/TIME OF TEST: 09/16/2020 19:58:29

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:56	
Control Test		0.078	19:56	
Air Blank		0.000	19:56	
Subject Sample	# 1		19:57	Subject Test Refused
Breath Volume		0.000	19:57	
Air Blank		0.000	19:57	
Control Test		0.079	19:57	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:58	

Data Download Date/Time: 09/28/2020 18:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LEMFANT,BENJAMINE
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/07/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SEPULVEDA

DATE/TIME OF TEST: 09/24/2020 00:16:53

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:14	
Control Test		0.080	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 1		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:16	
Control Test		0.081	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 09/28/2020 18:19

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GURRIERI,JOHN,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GURRIERI,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: OIEN

DATE/TIME OF TEST: 09/02/2020 01:50:50

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.260	01:46
Breath Volume		3.273	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.255	01:49
Breath Volume		2.792	01:49
Air Blank		0.000	01:49
Control Test		0.078	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DUQUET

DATE/TIME OF TEST: 09/03/2020 00:05:40

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:51	
Control Test		0.079	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 3	0.123	23:55	Volume Not Met
Breath Volume		0.644	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1	0.125	00:00	
Breath Volume		2.636	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2	0.130	00:03	
Breath Volume		1.886	00:03	
Air Blank		0.000	00:04	
Control Test		0.079	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CABRAL,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: CORBITT JR

DATE/TIME OF TEST: 09/03/2020 11:34:41

Diagnostics Check		OK	11:28
Air Blank		0.000	11:28
Control Test		0.080	11:29
Air Blank		0.000	11:29
Subject Sample	# 1	0.000	11:30
Breath Volume		1.847	11:30
Air Blank		0.000	11:30
Air Blank		0.000	11:32
Subject Sample	# 2	0.000	11:33
Breath Volume		1.574	11:33
Air Blank		0.000	11:33
Control Test		0.080	11:34
Air Blank		0.000	11:34
Diagnostics Check		OK	11:34

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,SHELBY,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HECKMAN,SHELBY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 09/03/2020 23:17:02

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.151	23:12
Breath Volume		2.039	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.155	23:15
Breath Volume		2.660	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:17

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 09/05/2020 03:59:55

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.172	03:55
Breath Volume		1.546	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.178	03:58
Breath Volume		1.496	03:58
Air Blank		0.000	03:59
Control Test		0.077	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTLE,RYAN,F
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BATTLE,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 09/05/2020 04:54:20

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1	0.150	04:45	Volume Not Met
Breath Volume		1.011	04:45	
Air Blank		0.000	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 2	0.167	04:48	
Breath Volume		1.101	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 3	0.169	04:52	
Breath Volume		1.710	04:52	
Air Blank		0.000	04:53	
Control Test		0.078	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARR,LUCAS,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KARR,LUCAS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CYPRIAN

DATE/TIME OF TEST: 09/06/2020 04:31:37

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:20	
Control Test		0.077	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1		04:24	No Sample Provided
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2		04:30	No Sample Provided
Breath Volume		0.000	04:30	
Air Blank		0.000	04:30	
Control Test		0.078	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEST,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 09/07/2020 01:47:46

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:45	
Control Test		0.079	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:46	
Control Test		0.079	01:47	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 09/09/2020 04:05:53

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.078	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.114	04:01
Breath Volume		3.242	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.111	04:04
Breath Volume		2.257	04:04
Air Blank		0.000	04:04
Control Test		0.077	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JASON,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SANTOS,JASON
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: OZGERCIN

DATE/TIME OF TEST: 09/11/2020 01:24:49

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.143	01:20
Breath Volume		3.738	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.134	01:23
Breath Volume		4.046	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 09/14/2020 00:26:35

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.212	00:21
Breath Volume		2.945	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.206	00:25
Breath Volume		2.179	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON],KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/16/2020 14:21:18

Diagnostics Check		OK	14:15
Air Blank		0.000	14:15
Control Test		0.079	14:16
Air Blank		0.000	14:16
Subject Sample	# 1	0.000	14:17
Breath Volume		2.636	14:17
Air Blank		0.000	14:17
Air Blank		0.000	14:19
Subject Sample	# 2	0.000	14:19
Breath Volume		2.570	14:19
Air Blank		0.000	14:20
Control Test		0.080	14:20
Air Blank		0.000	14:21
Diagnostics Check		OK	14:21

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 09/19/2020 05:13:27

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:08	
Control Test		0.079	05:09	
Air Blank		0.000	05:09	
Subject Sample	# 1	0.159	05:09	
Breath Volume		2.488	05:09	
Air Blank		0.000	05:10	
Air Blank			05:12	RFI Detect
Air Blank		0.000	05:12	
Control Test		0.080	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 09/19/2020 05:21:57

Diagnostics Check		OK	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.152	05:17
Breath Volume		1.933	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.146	05:20
Breath Volume		2.121	05:20
Air Blank		0.000	05:21
Control Test		0.079	05:21
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,AUSTIN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COON,AUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CABAN

DATE/TIME OF TEST: 09/20/2020 03:27:19

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.145	03:22
Breath Volume		2.441	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.145	03:25
Breath Volume		2.593	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JASON,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SANTOS,JASON
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 09/20/2020 03:44:27

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:42	
Control Test		0.078	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:43	Sequence Aborted
Breath Volume		0.000	03:43	
Air Blank		0.000	03:44	

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: STRINGER,M
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HARDING

DATE/TIME OF TEST: 09/22/2020 04:20:15

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.102	04:15
Breath Volume		1.214	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.096	04:18
Breath Volume		1.304	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: KING III

DATE/TIME OF TEST: 09/24/2020 13:57:26

Diagnostics Check		OK	13:50
Air Blank		0.000	13:51
Control Test		0.081	13:51
Air Blank		0.000	13:52
Subject Sample	# 1	0.224	13:52
Breath Volume		3.328	13:52
Air Blank		0.000	13:53
Air Blank		0.000	13:55
Subject Sample	# 2	0.219	13:55
Breath Volume		3.191	13:55
Air Blank		0.000	13:56
Control Test		0.079	13:56
Air Blank		0.000	13:57
Diagnostics Check		OK	13:57

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTS,C,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BATTS,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: FORDE

DATE/TIME OF TEST: 09/24/2020 20:34:04

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.079	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.198	20:29
Breath Volume		2.699	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:31
Subject Sample	# 2	0.190	20:32
Breath Volume		2.148	20:32
Air Blank		0.000	20:33
Control Test		0.077	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:34

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,SHELBY,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HECKMAN,SHELBY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:06

SUBJECT'S LAST NAME: DRAGANAC

DATE/TIME OF TEST: 09/25/2020 12:33:20

Diagnostics Check		OK	12:26
Air Blank		0.000	12:27
Control Test		0.080	12:27
Air Blank		0.000	12:28
Subject Sample	# 1	0.247	12:28
Breath Volume		3.996	12:28
Air Blank		0.000	12:29
Air Blank		0.000	12:31
Subject Sample	# 2	0.249	12:31
Breath Volume		4.019	12:31
Air Blank		0.000	12:32
Control Test		0.079	12:32
Air Blank		0.000	12:33
Diagnostics Check		OK	12:33

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LEWANDOWSKI

DATE/TIME OF TEST: 09/26/2020 02:55:00

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.156	02:47
Breath Volume		1.871	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.157	02:53
Breath Volume		3.000	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:55

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BAKER,MATTHEW
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DIURLIN

DATE/TIME OF TEST: 09/27/2020 21:22:35

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.079	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:15	No Sample Provided
Breath Volume		0.000	21:15	
Air Blank		0.000	21:16	
Air Blank		0.000	21:18	
Subject Sample	# 2		21:21	No Sample Provided
Breath Volume		0.000	21:21	
Air Blank		0.000	21:21	
Control Test		0.080	21:21	
Air Blank		0.000	21:22	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 09/29/2020 10:20

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WALDRON

DATE/TIME OF TEST: 09/01/2020 00:44:10

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.200	00:38
Breath Volume		1.976	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.182	00:42
Breath Volume		1.820	00:42
Air Blank		0.000	00:43
Control Test		0.078	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 09/02/2020 18:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: GONZALEZ MARTINEZ

DATE/TIME OF TEST: 09/01/2020 01:57:32

Diagnostics Check		OK	01:51
Air Blank		0.000	01:51
Control Test		0.078	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.143	01:53
Breath Volume		1.585	01:53
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.146	01:56
Breath Volume		3.292	01:56
Air Blank		0.000	01:56
Control Test		0.078	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 09/02/2020 18:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TORRES DIAZ

DATE/TIME OF TEST: 09/01/2020 03:09:56

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:58	
Control Test		0.078	02:59	
Air Blank		0.000	02:59	
Subject Sample	# 1	0.110	03:02	Volume Not Met
Breath Volume		0.765	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2	0.082	03:08	Volume Not Met
Breath Volume		0.289	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 09/02/2020 18:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,D,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MALBRANCHE

DATE/TIME OF TEST: 09/02/2020 01:45:48

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.000	01:41
Breath Volume		1.371	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.000	01:44
Breath Volume		1.296	01:44
Air Blank		0.000	01:44
Control Test		0.077	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 09/02/2020 18:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 08/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: REYES CARASQUILLO

DATE/TIME OF TEST: 09/02/2020 04:12:56

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.212	04:08
Breath Volume		2.390	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.214	04:11
Breath Volume		1.703	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 09/02/2020 18:38

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: NEWTON,JENNA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GIAQUINTO

DATE/TIME OF TEST: 09/02/2020 20:40:01

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.080	20:34
Air Blank		0.000	20:34
Subject Sample	# 1	0.212	20:35
Breath Volume		2.042	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.216	20:38
Breath Volume		1.710	20:38
Air Blank		0.000	20:39
Control Test		0.079	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:40

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/04/2020 03:02:10

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.168	02:57
Breath Volume		3.546	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.163	03:00
Breath Volume		3.078	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ESPINO

DATE/TIME OF TEST: 09/04/2020 21:04:14

Diagnostics Check		OK	20:55
Air Blank		0.000	20:56
Control Test		0.080	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.209	20:57
Breath Volume		1.242	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.225	21:02
Breath Volume		1.531	21:02
Air Blank		0.000	21:03
Control Test		0.079	21:03
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK,MICHAEL,O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO-PEREZ,CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HARDWICK

DATE/TIME OF TEST: 09/05/2020 03:29:03

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.194	03:23
Breath Volume		1.945	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.194	03:27
Breath Volume		1.695	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:29

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,MATTHEW,K
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MILLS,MATTHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: DOBOS

DATE/TIME OF TEST: 09/05/2020 19:46:19

Diagnostics Check		OK	19:40
Air Blank		0.000	19:40
Control Test		0.080	19:41
Air Blank		0.000	19:41
Subject Sample	# 1	0.000	19:42
Breath Volume		3.750	19:42
Air Blank		0.000	19:42
Air Blank		0.000	19:44
Subject Sample	# 2	0.000	19:44
Breath Volume		3.531	19:44
Air Blank		0.000	19:45
Control Test		0.080	19:45
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: WEIMER

DATE/TIME OF TEST: 09/05/2020 22:47:22

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.041	22:42
Breath Volume		2.335	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.044	22:45
Breath Volume		3.101	22:45
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: JEAN CHARLES

DATE/TIME OF TEST: 09/06/2020 01:56:32

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.000	01:52
Breath Volume		1.726	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.000	01:55
Breath Volume		1.121	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTRELL,ROBERT,V
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHINKEL,MATTHEW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: HYDE

DATE/TIME OF TEST: 09/06/2020 10:28:57

Diagnostics Check		OK	10:22
Air Blank		0.000	10:22
Control Test		0.080	10:23
Air Blank		0.000	10:23
Subject Sample	# 1	0.000	10:24
Breath Volume		2.289	10:24
Air Blank		0.000	10:24
Air Blank		0.000	10:26
Subject Sample	# 2	0.000	10:27
Breath Volume		1.390	10:27
Air Blank		0.000	10:28
Control Test		0.080	10:28
Air Blank		0.000	10:28
Diagnostics Check		OK	10:28

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANDLEY,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: BROTHERS

DATE/TIME OF TEST: 09/06/2020 11:17:27

Diagnostics Check		OK	11:11
Air Blank		0.000	11:12
Control Test		0.080	11:12
Air Blank		0.000	11:12
Subject Sample	# 1	0.000	11:13
Breath Volume		2.480	11:13
Air Blank		0.000	11:13
Air Blank		0.000	11:15
Subject Sample	# 2	0.000	11:16
Breath Volume		2.765	11:16
Air Blank		0.000	11:16
Control Test		0.079	11:16
Air Blank		0.000	11:17
Diagnostics Check		OK	11:17

Data Download Date/Time: 09/25/2020 09:34

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ENGH,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 09/07/2020 05:05:04

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.000	04:59
Breath Volume		1.289	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.000	05:03
Breath Volume		1.433	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:05

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 09/08/2020 01:55:21

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.000	01:51
Breath Volume		2.335	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.000	01:54
Breath Volume		2.859	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 09/09/2020 02:06:11

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.000	02:01
Breath Volume		1.238	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.000	02:04
Breath Volume		1.843	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 09/09/2020 09:04:03

Diagnostics Check		OK	08:57
Air Blank		0.000	08:58
Control Test		0.080	08:58
Air Blank		0.000	08:59
Subject Sample	# 1	0.000	08:59
Breath Volume		1.980	08:59
Air Blank		0.000	09:00
Air Blank		0.000	09:02
Subject Sample	# 2	0.000	09:02
Breath Volume		2.578	09:02
Air Blank		0.000	09:03
Control Test		0.079	09:03
Air Blank		0.000	09:03
Diagnostics Check		OK	09:04

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 09/09/2020 22:54:23

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.000	22:49
Breath Volume		1.882	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.000	22:53
Breath Volume		1.574	22:53
Air Blank		0.000	22:53
Control Test		0.080	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACY,CODY,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: TIRICO

DATE/TIME OF TEST: 09/12/2020 10:31:01

Diagnostics Check		OK	10:24
Air Blank		0.000	10:24
Control Test		0.080	10:25
Air Blank		0.000	10:25
Subject Sample	# 1	0.000	10:26
Breath Volume		4.281	10:26
Air Blank		0.000	10:27
Air Blank		0.000	10:29
Subject Sample	# 2	0.000	10:29
Breath Volume		4.343	10:29
Air Blank		0.000	10:30
Control Test		0.080	10:30
Air Blank		0.000	10:30
Diagnostics Check		OK	10:31

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,PERRY,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: BLEY

DATE/TIME OF TEST: 09/20/2020 17:46:51

Diagnostics Check		OK	17:34
Air Blank		0.000	17:35
Control Test		0.079	17:35
Air Blank		0.000	17:36
Subject Sample	# 1	0.180	17:37
Breath Volume		2.648	17:37
Air Blank		0.000	17:38
Air Blank		0.000	17:39
Subject Sample	# 2	0.147	17:41
Breath Volume		2.113	17:41
Air Blank		0.000	17:42
Air Blank		0.000	17:44
Subject Sample	# 3	0.171	17:45
Breath Volume		2.410	17:45
Air Blank		0.000	17:45
Control Test		0.079	17:46
Air Blank		0.000	17:46
Diagnostics Check		OK	17:46

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACY,CODY,J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,SARAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: ZIEGLER

DATE/TIME OF TEST: 09/21/2020 11:12:10

Diagnostics Check		OK	11:04
Air Blank		0.000	11:04
Control Test		0.080	11:05
Air Blank		0.000	11:05
Subject Sample	# 1	0.000	11:07
Breath Volume		2.687	11:07
Air Blank		0.000	11:08
Air Blank		0.000	11:10
Subject Sample	# 2	0.000	11:10
Breath Volume		1.160	11:10
Air Blank		0.000	11:11
Control Test		0.079	11:11
Air Blank		0.000	11:12
Diagnostics Check		OK	11:12

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PERIAM,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: SOUVANNASING

DATE/TIME OF TEST: 09/22/2020 22:33:40

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.080	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.214	22:29
Breath Volume		1.648	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.211	22:32
Breath Volume		1.832	22:32
Air Blank		0.000	22:32
Control Test		0.078	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 09/25/2020 09:35

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 09/26/2020 20:08:45

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.079	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.195	20:03
Breath Volume		1.277	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.184	20:07
Breath Volume		1.296	20:07
Air Blank		0.000	20:07
Control Test		0.079	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 10/14/2020 14:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STRICKLAND,DANIELLE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 09/26/2020 22:19:09

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.078	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.277	22:14
Breath Volume		1.464	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:16
Subject Sample	# 2	0.260	22:17
Breath Volume		2.257	22:17
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 10/14/2020 14:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 09/27/2020 00:36:05

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.078	00:31
Breath Volume		2.093	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.078	00:34
Breath Volume		2.941	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:36

Data Download Date/Time: 10/14/2020 14:03

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCK,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/18/2020 02:24:52

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:24	Sequence Aborted
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	

Data Download Date/Time: 09/30/2020 08:09

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILHARDT,DEP,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLLINS,R
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WATEROUS

DATE/TIME OF TEST: 09/19/2020 02:44:16

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.080	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.179	02:38	
Breath Volume		3.687	02:38	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:43	
Control Test		0.080	02:43	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 09/30/2020 08:09

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MELICHAR

DATE/TIME OF TEST: 09/19/2020 02:58:39

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.239	02:54
Breath Volume		1.847	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.242	02:57
Breath Volume		2.746	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 09/30/2020 08:09

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SCAPPATURA

DATE/TIME OF TEST: 09/19/2020 04:08:38

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.079	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1	0.208	04:01	Volume Not Met
Breath Volume		0.726	04:01	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2	0.194	04:06	Volume Not Met
Breath Volume		0.531	04:06	
Air Blank		0.000	04:07	
Control Test		0.079	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SCAPPATURA

DATE/TIME OF TEST: 09/19/2020 04:17:52

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.078	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.203	04:13
Breath Volume		1.214	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.219	04:16
Breath Volume		1.515	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYRICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/19/2020 05:03:34

Diagnostics Check	OK	05:03	
Air Blank		05:03	RFI Detect
Air Blank		05:03	RFI Detect

Data Download Date/Time: 09/30/2020 08:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYRICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/19/2020 05:10:34

Diagnostics Check		OK	05:06	
Air Blank		0.000	05:07	
Control Test		0.080	05:07	
Air Blank		0.000	05:07	
Subject Sample	# 1	0.123	05:08	
Breath Volume		1.121	05:08	
Air Blank			05:09	RFI Detect
Air Blank		0.000	05:09	
Control Test		0.080	05:09	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUERO,JOYRICK,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SUERO,JOYRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/19/2020 05:21:12

Diagnostics Check		OK	05:14
Air Blank		0.000	05:14
Control Test		0.079	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.132	05:16
Breath Volume		2.304	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.128	05:19
Breath Volume		2.175	05:19
Air Blank		0.000	05:20
Control Test		0.079	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ZAWISLAK

DATE/TIME OF TEST: 09/19/2020 22:41:24

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.080	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.090	22:37
Breath Volume		3.093	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.089	22:39
Breath Volume		2.589	22:39
Air Blank		0.000	22:40
Control Test		0.081	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 09/20/2020 00:23:42

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.262	00:17
Breath Volume		1.906	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:19
Subject Sample	# 2	0.268	00:22
Breath Volume		2.195	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: DEDOMINICIS

DATE/TIME OF TEST: 09/20/2020 02:00:58

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.183	01:56
Breath Volume		3.750	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.180	01:59
Breath Volume		3.785	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/20/2020 18:35:43

Diagnostics Check		OK	18:29
Air Blank		0.000	18:30
Control Test		0.080	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.000	18:31
Breath Volume		1.835	18:31
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.000	18:34
Breath Volume		1.304	18:34
Air Blank		0.000	18:34
Control Test		0.080	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GREGOIRE JR

DATE/TIME OF TEST: 09/21/2020 03:19:18

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.000	03:15
Breath Volume		4.183	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.000	03:17
Breath Volume		4.125	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAEUSSLER,BRETT,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HAEUSSLER,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 09/22/2020 00:23:49

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.080	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.295	00:19
Breath Volume		1.957	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.286	00:22
Breath Volume		1.363	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 09/23/2020 23:30:04

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.081	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.000	23:25
Breath Volume		3.351	23:25
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.000	23:28
Breath Volume		2.851	23:28
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:30

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,DEP D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HEWITT

DATE/TIME OF TEST: 09/24/2020 00:48:20

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.081	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.184	00:42
Breath Volume		2.953	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.178	00:46
Breath Volume		3.117	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/24/2020 04:24:44

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.035	04:20
Breath Volume		2.289	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.029	04:23
Breath Volume		1.796	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEPUTY S
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MCAFOOSE

DATE/TIME OF TEST: 09/25/2020 20:00:29

Diagnostics Check		OK	19:51
Air Blank		0.000	19:51
Control Test		0.081	19:51
Air Blank		0.000	19:52
Subject Sample	# 1	0.374	19:53
Breath Volume		2.398	19:53
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.349	19:55
Breath Volume		2.890	19:55
Air Blank		0.000	19:56
Air Blank		0.000	19:58
Subject Sample	# 3	0.354	19:58
Breath Volume		3.867	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHAPIRO,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DUNLAP

DATE/TIME OF TEST: 09/26/2020 03:44:31

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.228	03:36
Breath Volume		1.179	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 2	0.198	03:39
Breath Volume		1.242	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 3	0.204	03:42
Breath Volume		1.609	03:42
Air Blank		0.000	03:43
Control Test		0.077	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PRUMMELL,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: DEEM

DATE/TIME OF TEST: 09/27/2020 03:08:41

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.207	03:03
Breath Volume		1.773	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.203	03:07
Breath Volume		2.992	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/28/2020 02:54:15

Diagnostics Check	OK	02:54	
Air Blank		02:54	RFI Detect
Air Blank		02:54	RFI Detect

Data Download Date/Time: 09/30/2020 08:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/28/2020 03:02:55

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.071	02:58
Breath Volume		1.261	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.069	03:01
Breath Volume		2.003	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BESKIN,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 08/28/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: RAFFAELE

DATE/TIME OF TEST: 09/29/2020 19:05:49

Diagnostics Check		OK	18:58
Air Blank		0.000	18:58
Control Test		0.080	18:58
Air Blank		0.000	18:59
Subject Sample	# 1	0.283	19:01
Breath Volume		2.621	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:03
Subject Sample	# 2	0.294	19:04
Breath Volume		2.878	19:04
Air Blank		0.000	19:04
Control Test		0.077	19:05
Air Blank		0.000	19:05
Diagnostics Check		OK	19:05

Data Download Date/Time: 09/30/2020 08:10

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 09/24/2020 22:28:19

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:16	
Control Test		0.080	22:17	
Air Blank		0.000	22:17	
Subject Sample	# 1	0.000	22:20	Volume Not Met
Breath Volume		0.390	22:20	
Air Blank		0.000	22:21	
Air Blank		0.000	22:23	
Subject Sample	# 2	0.000	22:24	
Breath Volume		2.289	22:24	
Air Blank		0.000	22:24	
Air Blank		0.000	22:26	
Subject Sample	# 3	0.000	22:27	
Breath Volume		1.687	22:27	
Air Blank		0.000	22:27	
Control Test		0.000	22:27	RFI Detect
Air Blank		0.000	22:28	

Data Download Date/Time: 10/15/2020 10:40

Printed: 07 January 2021

COBRA ©1997-2014 CMI, Inc.