Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 01/02/2021 23:47:32

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.142	23:42
Breath Volume		1.996	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.137	23:45
Breath Volume		1.480	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/27/2021 10:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: JACKSON,N
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FRY

DATE/TIME OF TEST: 01/09/2021 00:28:01

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.160	00:22
Breath Volume		1.945	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.160	00:26
Breath Volume		1.964	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:28

Data Download Date/Time: 01/27/2021 10:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: OTERO,H
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: OSTEEN

DATE/TIME OF TEST: 01/10/2021 21:48:32

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:46	
Control Test		0.081	21:46	
Air Blank		0.000	21:46	
Subject Sample	# 1		21:47	Subject Test Refused
Breath Volume		0.000	21:47	
Air Blank		0.000	21:47	
Control Test		0.080	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 01/27/2021 10:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HENSHAW,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 01/11/2021 00:03:32

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.081	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Control Test		0.081	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 01/27/2021 10:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,ERIC,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SCHMIDT,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 01/16/2021 03:30:05

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:27	
Control Test		0.081	03:27	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:29	
Control Test		0.081	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 01/27/2021 10:41

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,4REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:33

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 01/19/2021 15:01:42

Diagnostics Check		OK	14:55
Air Blank		0.000	14:56
Control Test		0.081	14:56
Air Blank		0.000	14:56
Subject Sample	# 1	0.000	14:57
Breath Volume		1.890	14:57
Air Blank		0.000	14:57
Air Blank		0.000	14:59
Subject Sample	# 2	0.000	15:00
Breath Volume		1.816	15:00
Air Blank		0.000	15:00
Control Test		0.081	15:01
Air Blank		0.000	15:01
Diagnostics Check		OK	15:01

Data Download Date/Time: 01/27/2021 10:41

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WELLS,D
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CANADA

DATE/TIME OF TEST: 01/21/2021 01:19:11

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.081	01:17	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.081	01:18	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 01/27/2021 10:41

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: JENNINGS,JOSHUA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/02/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: CLIFFORD

DATE/TIME OF TEST: 01/22/2021 01:13:12

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.000	01:08
Breath Volume		2.710	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.000	01:11
Breath Volume		2.527	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 01/27/2021 10:41

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBE,BRIAN,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: DUDLEY,STACY
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 12/24/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28420080A1
DRY GAS EXPIRY DATE: 12/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 01/30/2021 04:07:02

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.082	04:05	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:06	
Control Test		0.082	04:06	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 01/30/2021 18:47

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: High Springs PD
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 01/02/2021 16:27:22

Diagnostics Check		OK	16:21
Air Blank		0.000	16:21
Control Test		0.080	16:22
Air Blank		0.000	16:22
Subject Sample	# 1	0.000	16:23
Breath Volume		3.242	16:23
Air Blank		0.000	16:23
Air Blank		0.000	16:25
Subject Sample	# 2	0.000	16:25
Breath Volume		3.257	16:25
Air Blank		0.000	16:26
Control Test		0.079	16:26
Air Blank		0.000	16:27
Diagnostics Check		OK	16:27

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLE,MADDISON,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TELLE,MADDISON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/02/2021 22:49:32

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:41	
Control Test		0.080	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1		22:45	No Sample Provided
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Air Blank		0.000	22:47	
Subject Sample	# 2		22:48	Subject Test Refused
Breath Volume		0.000	22:48	
Air Blank		0.000	22:48	
Control Test		0.080	22:48	
Air Blank		0.000	22:49	
Diagnostics Check		OK	22:49	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/04/2021 04:02:38

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.081	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.109	03:55	Volume Not Met
Breath Volume		0.312	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 2	0.123	04:00	Volume Not Met
Breath Volume		0.132	04:00	
Air Blank		0.000	04:01	
Control Test		0.078	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: MCINNIS

DATE/TIME OF TEST: 01/07/2021 03:17:42

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.081	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:13	No Sample Provided
Breath Volume		0.000	03:13	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2		03:16	Subject Test Refused
Breath Volume		0.000	03:16	
Air Blank		0.000	03:16	
Control Test		0.082	03:17	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURRO,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KEYES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: PHIPPS

DATE/TIME OF TEST: 01/08/2021 10:54:58

Diagnostics Check		OK	10:51	
Air Blank		0.000	10:51	
Control Test		0.081	10:51	
Air Blank		0.000	10:52	
Subject Sample	# 1		10:54	Sequence Aborted
Breath Volume		0.000	10:54	
Air Blank		0.000	10:54	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURRO,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KEYES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: PHIPPS

DATE/TIME OF TEST: 01/08/2021 11:09:19

Diagnostics Check		OK	10:57	
Air Blank		0.000	10:58	
Control Test		0.080	10:58	
Air Blank		0.000	10:59	
Subject Sample	# 1		11:02	No Sample Provided
Breath Volume		0.000	11:02	
Air Blank		0.000	11:02	
Air Blank		0.000	11:04	
Subject Sample	# 2		11:07	No Sample Provided
Breath Volume		0.000	11:07	
Air Blank		0.000	11:08	
Control Test		0.081	11:08	
Air Blank		0.000	11:09	
Diagnostics Check		OK	11:09	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JONES,LEIGHA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 01/08/2021 21:07:06

Diagnostics Check		OK	20:59
Air Blank		0.000	20:59
Control Test		0.081	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.130	21:01
Breath Volume		1.230	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.149	21:05
Breath Volume		2.933	21:05
Air Blank		0.000	21:06
Control Test		0.080	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MOORE,GARY
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: STAMPER

DATE/TIME OF TEST: 01/09/2021 00:28:15

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:23
Breath Volume		2.144	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:26
Breath Volume		2.156	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLON,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DILLON,JUSTIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 01/11/2021 19:58:28

Diagnostics Check		OK	19:51
Air Blank		0.000	19:52
Control Test		0.079	19:52
Air Blank		0.000	19:53
Subject Sample	# 1	0.229	19:54
Breath Volume		1.242	19:54
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.221	19:56
Breath Volume		1.488	19:56
Air Blank		0.000	19:57
Control Test		0.079	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBSON,COATLAND,H
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: MCGILL,KALEN
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TAPIA-HERRERA

DATE/TIME OF TEST: 01/13/2021 03:11:43

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.073	03:07
Breath Volume		3.464	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:09
Subject Sample	# 2	0.071	03:10
Breath Volume		3.710	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 01/16/2021 02:23:08

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:10	
Control Test		0.081	02:10	
Air Blank		0.000	02:10	
Subject Sample	# 1		02:14	No Sample Provided
Breath Volume		0.000	02:14	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.157	02:18	
Breath Volume		1.683	02:18	
Air Blank		0.000	02:18	
Air Blank		0.000	02:20	
Subject Sample	# 3	0.164	02:21	
Breath Volume		2.566	02:21	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: VANAERNAM

DATE/TIME OF TEST: 01/16/2021 03:10:02

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.080	03:05
Breath Volume		4.355	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.081	03:08
Breath Volume		4.250	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:10

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FLORY,JOEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MALONEY

DATE/TIME OF TEST: 01/16/2021 17:51:04

Diagnostics Check		OK	17:37	
Air Blank		0.000	17:38	
Control Test		0.081	17:38	
Air Blank		0.000	17:39	
Subject Sample	# 1	0.000	17:41	
Breath Volume		1.289	17:41	
Air Blank		0.000	17:41	
Air Blank		0.000	17:43	
Subject Sample	# 2	0.000	17:46	Volume Not Met
Breath Volume		0.152	17:46	
Air Blank		0.000	17:47	
Air Blank		0.000	17:49	
Subject Sample	# 3	0.000	17:49	
Breath Volume		1.449	17:49	
Air Blank		0.000	17:50	
Control Test		0.077	17:50	
Air Blank		0.000	17:50	
Diagnostics Check		OK	17:51	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CAMPO

DATE/TIME OF TEST: 01/17/2021 03:52:20

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.081	03:50	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:50	Subject Test Refused
Breath Volume		0.000	03:50	
Air Blank		0.000	03:51	
Control Test		0.081	03:51	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/15/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BLEKE

DATE/TIME OF TEST: 01/17/2021 23:56:03

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.082	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	
Control Test		0.082	23:55	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 01/20/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,JON,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GORDON,JON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HULION

DATE/TIME OF TEST: 01/01/2021 22:32:41

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:21	
Control Test		0.081	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1		22:25	No Sample Provided
Breath Volume		0.000	22:25	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 2		22:31	No Sample Provided
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Control Test		0.082	22:32	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 01/28/2021 09:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,KEITH,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 01/05/2021 12:34:32

Diagnostics Check		OK	12:27	
Air Blank		0.000	12:27	
Control Test		0.082	12:27	
Air Blank		0.000	12:28	
Subject Sample	# 1	0.223	12:30	
Breath Volume		3.445	12:30	
Air Blank		0.000	12:30	
Air Blank		0.000	12:32	
Subject Sample	# 2		12:33	Subject Test Refused
Breath Volume		0.000	12:33	
Air Blank		0.000	12:33	
Control Test		0.079	12:33	
Air Blank		0.000	12:34	
Diagnostics Check		OK	12:34	

Data Download Date/Time: 01/28/2021 09:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: MIRAMBELL

DATE/TIME OF TEST: 01/19/2021 12:12:21

Diagnostics Check		OK	12:01	
Air Blank		0.000	12:01	
Control Test		0.082	12:02	
Air Blank		0.000	12:02	
Subject Sample	# 1		12:05	No Sample Provided
Breath Volume		0.000	12:05	
Air Blank		0.000	12:06	
Air Blank		0.000	12:08	
Subject Sample	# 2		12:10	Subject Test Refused
Breath Volume		0.000	12:10	
Air Blank		0.000	12:11	
Control Test		0.081	12:11	
Air Blank		0.000	12:12	
Diagnostics Check		OK	12:12	

Data Download Date/Time: 01/28/2021 09:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TISDALE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: LAPREAD

DATE/TIME OF TEST: 01/21/2021 02:44:23

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.081	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.145	02:39
Breath Volume		1.371	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.151	02:42
Breath Volume		1.734	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 01/28/2021 09:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VALENTINE,RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BERNARDO PADILLA

DATE/TIME OF TEST: 01/24/2021 04:33:02

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.081	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.166	04:28
Breath Volume		2.320	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:30
Subject Sample	# 2	0.168	04:31
Breath Volume		2.972	04:31
Air Blank		0.000	04:32
Control Test		0.078	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:33

Data Download Date/Time: 01/28/2021 09:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 01/25/2021 00:24:32

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.081	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.080	00:23	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 01/28/2021 09:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/29/2021 16:26:39

Diagnostics Check		OK	16:11	
Air Blank		0.000	16:12	
Control Test		0.081	16:12	
Air Blank		0.000	16:13	
Subject Sample	# 1	0.000	16:16	Volume Not Met
Breath Volume		0.468	16:16	
Air Blank		0.000	16:16	
Air Blank		0.000	16:18	
Subject Sample	# 2	0.000	16:19	
Breath Volume		1.144	16:19	
Air Blank		0.000	16:20	
Air Blank		0.000	16:22	
Subject Sample	# 3	0.000	16:25	Volume Not Met
Breath Volume		0.812	16:25	
Air Blank		0.000	16:25	
Control Test		0.082	16:26	
Air Blank		0.000	16:26	
Diagnostics Check		OK	16:26	

Data Download Date/Time: 02/25/2021 12:48

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: STURGES

DATE/TIME OF TEST: 01/01/2021 05:47:13

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.080	05:40
Air Blank		0.000	05:41
Subject Sample	# 1	0.127	05:42
Breath Volume		1.339	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.130	05:45
Breath Volume		1.832	05:45
Air Blank		0.000	05:46
Control Test		0.078	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 01/11/2021 17:01

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: PEEBLES

DATE/TIME OF TEST: 01/08/2021 21:13:51

Diagnostics Check		OK	21:06
Air Blank		0.000	21:07
Control Test		0.081	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.000	21:09
Breath Volume		1.320	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.000	21:12
Breath Volume		1.398	21:12
Air Blank		0.000	21:12
Control Test		0.081	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 01/11/2021 17:01

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PETTWAY

DATE/TIME OF TEST: 01/10/2021 02:18:03

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.082	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.108	02:13
Breath Volume		2.621	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.107	02:16
Breath Volume		2.257	02:16
Air Blank		0.000	02:17
Control Test		0.082	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:18

Data Download Date/Time: 01/11/2021 17:01

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: DEALE

DATE/TIME OF TEST: 01/11/2021 18:48:13

Diagnostics Check		OK	18:41
Air Blank		0.000	18:41
Control Test		0.081	18:41
Air Blank		0.000	18:42
Subject Sample	# 1	0.139	18:43
Breath Volume		1.929	18:43
Air Blank		0.000	18:44
Air Blank		0.000	18:46
Subject Sample	# 2	0.137	18:46
Breath Volume		2.457	18:46
Air Blank		0.000	18:47
Control Test		0.080	18:47
Air Blank		0.000	18:48
Diagnostics Check		OK	18:48

Data Download Date/Time: 02/04/2021 16:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA,RICARDO,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CHARLES,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: GRUNEWALD

DATE/TIME OF TEST: 01/12/2021 19:49:58

Diagnostics Check		OK	19:38	
Air Blank		0.000	19:39	
Control Test		0.081	19:39	
Air Blank		0.000	19:40	
Subject Sample	# 1	0.135	19:43	Volume Not Met
Breath Volume		0.367	19:43	
Air Blank		0.000	19:43	
Air Blank		0.000	19:45	
Subject Sample	# 2		19:48	Subject Test Refused
Breath Volume		0.128	19:48	
Air Blank		0.000	19:49	
Control Test		0.083	19:49	
Air Blank		0.000	19:49	
Diagnostics Check		OK	19:49	

Data Download Date/Time: 02/04/2021 16:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,C,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: GRUNEWALD

DATE/TIME OF TEST: 01/12/2021 20:13:58

Diagnostics Check		OK	20:02	
Air Blank		0.000	20:02	
Control Test		0.081	20:03	
Air Blank		0.000	20:03	
Subject Sample	# 1	0.119	20:06	Volume Not Met
Breath Volume		0.898	20:06	
Air Blank		0.000	20:07	
Air Blank		0.000	20:09	
Subject Sample	# 2	0.123	20:12	Volume Not Met
Breath Volume		1.027	20:12	
Air Blank		0.000	20:13	
Control Test		0.079	20:13	
Air Blank		0.000	20:13	
Diagnostics Check		OK	20:13	

Data Download Date/Time: 02/04/2021 16:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WOOD,DERRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DIAMOND

DATE/TIME OF TEST: 01/13/2021 03:42:42

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.198	03:38
Breath Volume		2.144	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.195	03:41
Breath Volume		2.136	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 02/04/2021 16:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,DERRICK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CULLIVER,TAVARIS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GAHIMER

DATE/TIME OF TEST: 01/16/2021 02:17:47

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.192	02:13
Breath Volume		1.851	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.184	02:16
Breath Volume		1.609	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 02/04/2021 16:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ARGENT,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 01/10/2021 03:26:53

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.083	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.220	03:19
Breath Volume		1.636	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.225	03:25
Breath Volume		1.539	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 02/04/2021 12:06

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LEACH,ZACHARY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1160200
DRY GAS EXPIRY DATE: 06/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/10/2021 21:37:05

Diagnostics Check		OK	21:30
Air Blank		0.000	21:31
Control Test		0.082	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.000	21:32
Breath Volume		4.332	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.000	21:35
Breath Volume		4.144	21:35
Air Blank		0.000	21:36
Control Test		0.080	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 02/04/2021 12:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA,ADALBERTO,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA,ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/01/2021 02:29:04

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.234	02:22
Breath Volume		1.886	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.242	02:27
Breath Volume		1.507	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:29

Data Download Date/Time: 01/11/2021 16:48

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING,CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 01/01/2021 08:01:02

Diagnostics Check		OK	07:54
Air Blank		0.000	07:55
Control Test		0.082	07:55
Air Blank		0.000	07:56
Subject Sample	# 1	0.173	07:56
Breath Volume		3.226	07:56
Air Blank		0.000	07:57
Air Blank		0.000	07:59
Subject Sample	# 2	0.164	07:59
Breath Volume		4.035	07:59
Air Blank		0.000	08:00
Control Test		0.081	08:00
Air Blank		0.000	08:00
Diagnostics Check		OK	08:01

Data Download Date/Time: 01/11/2021 16:48

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 01/01/2021 22:02:16

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.083	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.260	21:57
Breath Volume		1.496	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.259	22:00
Breath Volume		2.046	22:00
Air Blank		0.000	22:01
Control Test		0.082	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 01/11/2021 16:48

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,C,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 01/03/2021 15:06:14

Diagnostics Check		OK	15:00
Air Blank		0.000	15:00
Control Test		0.083	15:00
Air Blank		0.000	15:01
Subject Sample	# 1	0.000	15:01
Breath Volume		2.687	15:01
Air Blank		0.000	15:02
Air Blank		0.000	15:04
Subject Sample	# 2	0.000	15:04
Breath Volume		2.621	15:04
Air Blank		0.000	15:05
Control Test		0.083	15:05
Air Blank		0.000	15:06
Diagnostics Check		OK	15:06

Data Download Date/Time: 01/11/2021 16:48

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CHEADLE

DATE/TIME OF TEST: 01/04/2021 04:44:20

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.084	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.138	04:39
Breath Volume		1.890	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.144	04:42
Breath Volume		2.281	04:42
Air Blank		0.000	04:43
Control Test		0.083	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/11/2021 16:48

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BOOTHE,TIM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CUPP

DATE/TIME OF TEST: 01/12/2021 00:47:35

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.082	00:39	
Air Blank		0.000	00:40	
Subject Sample	# 1		00:43	No Sample Provided
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Air Blank		0.000	00:45	
Subject Sample	# 2		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:46	
Control Test		0.083	00:47	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 02/06/2021 14:28

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,C,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 01/15/2021 19:27:30

Diagnostics Check		OK	19:15	
Air Blank		0.000	19:16	
Control Test		0.083	19:16	
Air Blank		0.000	19:17	
Subject Sample	# 1	0.042	19:20	Volume Not Met
Breath Volume		0.156	19:20	
Air Blank		0.000	19:20	
Air Blank		0.000	19:22	
Subject Sample	# 2	0.040	19:26	Volume Not Met
Breath Volume		0.156	19:26	
Air Blank		0.000	19:26	
Control Test		0.081	19:26	
Air Blank		0.000	19:27	
Diagnostics Check		OK	19:27	

Data Download Date/Time: 02/06/2021 14:28

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,C,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALEXANDER,C
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 01/15/2021 19:39:21

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:30	
Control Test		0.083	19:30	
Air Blank		0.000	19:31	
Subject Sample	# 1	0.041	19:34	Volume Not Met
Breath Volume		0.441	19:34	
Air Blank		0.000	19:34	
Air Blank		0.000	19:36	
Subject Sample	# 2		19:37	Subject Test Refused
Breath Volume		0.507	19:37	
Air Blank		0.000	19:38	
Control Test		0.082	19:38	
Air Blank		0.000	19:39	
Diagnostics Check		OK	19:39	

Data Download Date/Time: 02/06/2021 14:28

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUPONT,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 01/18/2021 12:49:39

Diagnostics Check		OK	12:40
Air Blank		0.000	12:41
Control Test		0.084	12:41
Air Blank		0.000	12:42
Subject Sample	# 1	0.000	12:43
Breath Volume		1.507	12:43
Air Blank		0.000	12:44
Air Blank		0.000	12:46
Subject Sample	# 2	0.000	12:48
Breath Volume		1.183	12:48
Air Blank		0.000	12:48
Control Test		0.084	12:49
Air Blank		0.000	12:49
Diagnostics Check		OK	12:49

Data Download Date/Time: 02/06/2021 14:28

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCUM,JOSEPH,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARCUM,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 01/29/2021 00:25:54

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.084	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.224	00:20
Breath Volume		1.300	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.216	00:24
Breath Volume		1.578	00:24
Air Blank		0.000	00:24
Control Test		0.082	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 02/06/2021 14:28

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARCUM,JOSEPH,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARCUM,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: GULSBY

DATE/TIME OF TEST: 01/31/2021 05:26:28

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.082	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.161	05:20
Breath Volume		2.097	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.162	05:24
Breath Volume		3.027	05:24
Air Blank		0.000	05:25
Control Test		0.080	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 02/06/2021 14:28

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARTER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: JOHNSTON,TRENTON
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 12/05/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 01/17/2021 02:06:00

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.083	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Control Test		0.083	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 01/25/2021 09:17

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CABRAL,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BERRIER

DATE/TIME OF TEST: 01/22/2021 05:24:04

Diagnostics Check		OK	05:12	
Air Blank		0.000	05:13	
Control Test		0.080	05:13	
Air Blank		0.000	05:14	
Subject Sample	# 1		05:17	No Sample Provided
Breath Volume		0.000	05:17	
Air Blank		0.000	05:17	
Air Blank		0.000	05:19	
Subject Sample	# 2		05:22	No Sample Provided
Breath Volume		0.000	05:22	
Air Blank		0.000	05:23	
Control Test		0.081	05:23	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 02/18/2021 15:01

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURWELL,MARK,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MCVEIGH,TREVOR
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 01/29/2021 22:28:50

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.080	22:18	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:21	No Sample Provided
Breath Volume		0.000	22:21	
Air Blank		0.000	22:22	
Air Blank		0.000	22:24	
Subject Sample	# 2		22:27	No Sample Provided
Breath Volume		0.000	22:27	
Air Blank		0.000	22:27	
Control Test		0.082	22:28	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 02/18/2021 15:01

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: FORTNER

DATE/TIME OF TEST: 01/30/2021 19:55:34

Diagnostics Check		OK	19:48
Air Blank		0.000	19:48
Control Test		0.079	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.227	19:50
Breath Volume		2.316	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.243	19:53
Breath Volume		1.996	19:53
Air Blank		0.000	19:54
Control Test		0.077	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 02/18/2021 15:02

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DEEGAN

DATE/TIME OF TEST: 01/31/2021 01:47:38

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:44	
Control Test		0.079	01:45	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:46	
Control Test		0.078	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 02/18/2021 15:02

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 01/16/2021 02:09:00

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.081	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 01/16/2021 05:04:50

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.082	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.000	05:00
Breath Volume		3.460	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.000	05:03
Breath Volume		2.339	05:03
Air Blank		0.000	05:03
Control Test		0.081	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: HEWITT

DATE/TIME OF TEST: 01/16/2021 06:34:00

Diagnostics Check		OK	06:31	
Air Blank		0.000	06:31	
Control Test		0.082	06:31	
Air Blank		0.000	06:32	
Subject Sample	# 1		06:32	Subject Test Refused
Breath Volume		0.000	06:32	
Air Blank		0.000	06:33	
Control Test		0.081	06:33	
Air Blank		0.000	06:33	
Diagnostics Check		OK	06:34	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: GOMEZ GRACIA

DATE/TIME OF TEST: 01/16/2021 20:36:41

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.081	20:31
Air Blank		0.000	20:31
Subject Sample	# 1	0.133	20:32
Breath Volume		1.574	20:32
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.140	20:35
Breath Volume		1.503	20:35
Air Blank		0.000	20:35
Control Test		0.081	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 01/17/2021 00:50:23

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.081	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.219	00:46	Volume Not Met
Breath Volume		0.996	00:46	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 01/17/2021 03:01:54

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.081	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.211	02:57
Breath Volume		4.140	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.195	03:00
Breath Volume		2.777	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ARDOIN

DATE/TIME OF TEST: 01/17/2021 03:38:58

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.080	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1		03:37	Subject Test Refused
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Control Test		0.081	03:38	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EHRHART,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: VAN MATRE

DATE/TIME OF TEST: 01/17/2021 12:41:53

Diagnostics Check		OK	12:35
Air Blank		0.000	12:35
Control Test		0.082	12:36
Air Blank		0.000	12:36
Subject Sample	# 1	0.000	12:37
Breath Volume		1.695	12:37
Air Blank		0.000	12:37
Air Blank		0.000	12:39
Subject Sample	# 2	0.000	12:40
Breath Volume		1.457	12:40
Air Blank		0.000	12:40
Control Test		0.081	12:41
Air Blank		0.000	12:41
Diagnostics Check		OK	12:41

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: BEYER

DATE/TIME OF TEST: 01/17/2021 22:06:20

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:02
Breath Volume		1.425	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:04
Breath Volume		1.890	22:04
Air Blank		0.000	22:05
Control Test		0.081	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 01/18/2021 02:05:17

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:02	
Control Test		0.082	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:04	
Control Test		0.081	02:04	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EHRHART,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:24

SUBJECT'S LAST NAME: PAINTER

DATE/TIME OF TEST: 01/21/2021 14:51:15

Diagnostics Check		OK	14:44
Air Blank		0.000	14:45
Control Test		0.080	14:45
Air Blank		0.000	14:46
Subject Sample	# 1	0.196	14:46
Breath Volume		1.234	14:46
Air Blank		0.000	14:47
Air Blank		0.000	14:49
Subject Sample	# 2	0.199	14:49
Breath Volume		1.453	14:49
Air Blank		0.000	14:50
Control Test		0.079	14:50
Air Blank		0.000	14:51
Diagnostics Check		OK	14:51

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANCK,ARIEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PLANCK,ARIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 01/23/2021 01:59:49

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.081	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:58	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FEENEY,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 01/23/2021 20:31:02

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:28	
Control Test		0.081	20:28	
Air Blank		0.000	20:29	
Subject Sample	# 1		20:29	Subject Test Refused
Breath Volume		0.000	20:29	
Air Blank		0.000	20:30	
Control Test		0.081	20:30	
Air Blank		0.000	20:30	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEENEY,ERIC,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COPPER,T
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 01/23/2021 22:15:24

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.081	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.140	22:10
Breath Volume		2.542	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.139	22:13
Breath Volume		2.359	22:13
Air Blank		0.000	22:14
Control Test		0.081	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MILLER,WHITNEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 01/27/2021 01:44:52

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.081	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:37	No Sample Provided
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.137	01:40	
Breath Volume		2.562	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 3	0.132	01:43	
Breath Volume		1.417	01:43	
Air Blank		0.000	01:43	
Control Test		0.078	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANCK,ARIEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PAPPAS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 01/28/2021 05:21:50

Diagnostics Check		OK	05:18	
Air Blank		0.000	05:19	
Control Test		0.081	05:19	
Air Blank		0.000	05:20	
Subject Sample	# 1		05:20	Subject Test Refused
Breath Volume		0.000	05:20	
Air Blank		0.000	05:20	
Control Test		0.081	05:21	
Air Blank		0.000	05:21	
Diagnostics Check		OK	05:21	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAURICIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MAURICIO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 01/30/2021 02:05:41

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.082	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:04	Sequence Aborted
Breath Volume		1.042	02:04	
Air Blank		0.000	02:05	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAURICIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MAURICIO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 01/30/2021 02:15:02

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.110	02:10
Breath Volume		2.742	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.109	02:13
Breath Volume		2.523	02:13
Air Blank		0.000	02:14
Control Test		0.081	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:15

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 01/30/2021 02:28:40

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.082	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.203	02:23
Breath Volume		1.933	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.216	02:27
Breath Volume		1.687	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EHRHART,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ALTAMURA

DATE/TIME OF TEST: 01/30/2021 02:45:53

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.081	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.113	02:41
Breath Volume		2.011	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:43
Subject Sample	# 2	0.111	02:44
Breath Volume		1.964	02:44
Air Blank		0.000	02:44
Control Test		0.081	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATTS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WATTS,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 01/30/2021 04:21:19

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.082	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:14	No Sample Provided
Breath Volume		0.000	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 2		04:19	Subject Test Refused
Breath Volume		0.000	04:19	
Air Blank		0.000	04:20	
Control Test		0.082	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: TRILLIECI

DATE/TIME OF TEST: 01/30/2021 10:23:17

Diagnostics Check		OK	10:17
Air Blank		0.000	10:17
Control Test		0.082	10:18
Air Blank		0.000	10:18
Subject Sample	# 1	0.000	10:18
Breath Volume		2.515	10:18
Air Blank		0.000	10:19
Air Blank		0.000	10:21
Subject Sample	# 2	0.000	10:21
Breath Volume		3.402	10:21
Air Blank		0.000	10:22
Control Test		0.081	10:22
Air Blank		0.000	10:23
Diagnostics Check		OK	10:23

Data Download Date/Time: 02/04/2021 15:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 11:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: GURGANUS

DATE/TIME OF TEST: 01/05/2021 12:24:45

Diagnostics Check		OK	12:21	
Air Blank		0.000	12:22	
Control Test		0.076	12:22	
Air Blank		0.000	12:23	
Subject Sample	# 1		12:23	Subject Test Refused
Breath Volume		0.000	12:23	
Air Blank		0.000	12:23	
Control Test		0.076	12:24	
Air Blank		0.000	12:24	
Diagnostics Check		OK	12:24	

Data Download Date/Time: 02/02/2021 19:59

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00919080A3
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: LIEUPO

DATE/TIME OF TEST: 01/28/2021 15:23:15

Diagnostics Check		OK	15:20	
Air Blank		0.000	15:20	
Control Test		0.076	15:21	
Air Blank		0.000	15:21	
Subject Sample	# 1		15:21	Subject Test Refused
Breath Volume		0.000	15:21	
Air Blank		0.000	15:22	
Control Test		0.077	15:22	
Air Blank		0.000	15:23	
Diagnostics Check		OK	15:23	

Data Download Date/Time: 02/02/2021 19:59

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: KELEMEN

DATE/TIME OF TEST: 01/01/2021 13:43:48

Diagnostics Check		OK	13:40	
Air Blank		0.000	13:40	
Control Test		0.081	13:40	
Air Blank		0.000	13:41	
Subject Sample	# 1		13:42	Subject Test Refused
Breath Volume		0.000	13:42	
Air Blank		0.000	13:42	
Control Test		0.081	13:43	
Air Blank		0.000	13:43	
Diagnostics Check		OK	13:43	

Data Download Date/Time: 01/04/2021 14:27

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA,ADALBERTO,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA,ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 01/02/2021 06:58:34

Diagnostics Check	OK	06:57	
Air Blank	0.000	06:57	
Control Test	0.000	06:58	RFI Detect
Air Blank	0.000	06:58	

Data Download Date/Time: 01/04/2021 14:27

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA,ADALBERTO,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA,ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 01/02/2021 07:07:27

Diagnostics Check		OK	07:00
Air Blank		0.000	07:01
Control Test		0.080	07:01
Air Blank		0.000	07:02
Subject Sample	# 1	0.060	07:02
Breath Volume		2.589	07:02
Air Blank		0.000	07:03
Air Blank		0.000	07:05
Subject Sample	# 2	0.052	07:05
Breath Volume		2.085	07:05
Air Blank		0.000	07:06
Control Test		0.080	07:06
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

Data Download Date/Time: 01/04/2021 14:27

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SHINAFELT II

DATE/TIME OF TEST: 01/03/2021 02:04:38

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.081	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Control Test		0.081	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 01/04/2021 14:27

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,RYAN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HALFHILL,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 01/05/2021 04:06:33

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.000	04:02
Breath Volume		1.464	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:05
Breath Volume		1.417	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,ANN,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COPELAND,JUSTIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/06/2021 21:52:33

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:41	
Control Test		0.081	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:45	No Sample Provided
Breath Volume		0.000	21:45	
Air Blank		0.000	21:46	
Air Blank		0.000	21:48	
Subject Sample	# 2		21:51	No Sample Provided
Breath Volume		0.000	21:51	
Air Blank		0.000	21:51	
Control Test		0.081	21:51	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/09/2021 03:34:08

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.081	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2		03:32	No Sample Provided
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.081	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,JAMISON,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/09/2021 17:57:07

Diagnostics Check		OK	17:45	
Air Blank		0.000	17:46	
Control Test		0.081	17:46	
Air Blank		0.000	17:47	
Subject Sample	# 1		17:50	No Sample Provided
Breath Volume		0.000	17:50	
Air Blank		0.000	17:50	
Air Blank		0.000	17:52	
Subject Sample	# 2		17:55	No Sample Provided
Breath Volume		0.000	17:55	
Air Blank		0.000	17:56	
Control Test		0.082	17:56	
Air Blank		0.000	17:57	
Diagnostics Check		OK	17:57	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,DAKOTA,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JACKSON,DAKOTA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DEEN

DATE/TIME OF TEST: 01/10/2021 02:08:16

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.082	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Control Test		0.083	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSEY,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POSEY,BENJAMIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/10/2021 03:15:35

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.082	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:08	No Sample Provided
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2		03:14	No Sample Provided
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.082	03:14	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TISDALE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/13/2021 00:54:19

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.082	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.168	00:45	Volume Not Met
Breath Volume		0.855	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2	0.218	00:49	
Breath Volume		1.250	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 3	0.214	00:52	
Breath Volume		1.437	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DALGLEISH,KEVIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DALGLEISH,KEVIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 11:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 01/13/2021 12:57:13

Diagnostics Check		OK	12:50
Air Blank		0.000	12:51
Control Test		0.081	12:51
Air Blank		0.000	12:51
Subject Sample	# 1	0.000	12:52
Breath Volume		1.136	12:52
Air Blank		0.000	12:53
Air Blank		0.000	12:55
Subject Sample	# 2	0.000	12:55
Breath Volume		2.035	12:55
Air Blank		0.000	12:56
Control Test		0.081	12:56
Air Blank		0.000	12:57
Diagnostics Check		OK	12:57

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 01/14/2021 19:32:19

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.081	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.000	19:27
Breath Volume		2.433	19:27
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.000	19:30
Breath Volume		3.019	19:30
Air Blank		0.000	19:31
Control Test		0.081	19:31
Air Blank		0.000	19:32
Diagnostics Check		OK	19:32

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STEPHENSON,PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: OSZANIEC

DATE/TIME OF TEST: 01/17/2021 00:36:33

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.081	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.086	00:31
Breath Volume		1.960	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.085	00:34
Breath Volume		1.902	00:34
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PINTO,FELIPE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: BIGGER

DATE/TIME OF TEST: 01/17/2021 20:54:26

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.081	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.000	20:50
Breath Volume		2.972	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.000	20:53
Breath Volume		3.117	20:53
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,ANN,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: CUTTING

DATE/TIME OF TEST: 01/17/2021 22:33:37

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.081	22:23	
Air Blank		0.000	22:23	
Subject Sample	# 1		22:26	No Sample Provided
Breath Volume		0.000	22:26	
Air Blank		0.000	22:27	
Air Blank		0.000	22:29	
Subject Sample	# 2		22:32	No Sample Provided
Breath Volume		0.000	22:32	
Air Blank		0.000	22:32	
Control Test		0.081	22:33	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:33	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 01/18/2021 07:46:21

Diagnostics Check		OK	07:39
Air Blank		0.000	07:39
Control Test		0.082	07:40
Air Blank		0.000	07:40
Subject Sample	# 1	0.136	07:41
Breath Volume		1.519	07:41
Air Blank		0.000	07:42
Air Blank		0.000	07:44
Subject Sample	# 2	0.146	07:44
Breath Volume		2.132	07:44
Air Blank		0.000	07:45
Control Test		0.079	07:45
Air Blank		0.000	07:46
Diagnostics Check		OK	07:46

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,VIRGINIA,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBERSON,VIRGINIA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: MCGIRT

DATE/TIME OF TEST: 01/18/2021 16:59:05

Diagnostics Check		OK	16:52
Air Blank		0.000	16:53
Control Test		0.081	16:53
Air Blank		0.000	16:54
Subject Sample	# 1	0.000	16:54
Breath Volume		3.886	16:54
Air Blank		0.000	16:55
Air Blank		0.000	16:57
Subject Sample	# 2	0.000	16:57
Breath Volume		3.312	16:57
Air Blank		0.000	16:58
Control Test		0.081	16:58
Air Blank		0.000	16:59
Diagnostics Check		OK	16:59

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,J,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: O'DANIEL

DATE/TIME OF TEST: 01/22/2021 20:38:12

Diagnostics Check		OK	20:31
Air Blank		0.000	20:32
Control Test		0.080	20:32
Air Blank		0.000	20:32
Subject Sample	# 1	0.263	20:33
Breath Volume		2.859	20:33
Air Blank		0.000	20:34
Air Blank		0.000	20:35
Subject Sample	# 2	0.263	20:36
Breath Volume		3.859	20:36
Air Blank		0.000	20:37
Control Test		0.079	20:37
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 01/23/2021 01:25:51

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.201	01:15
Breath Volume		1.496	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.176	01:19
Breath Volume		2.234	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 3	0.183	01:24
Breath Volume		2.261	01:24
Air Blank		0.000	01:24
Control Test		0.076	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 01/23/2021 01:44:11

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:41	
Control Test		0.077	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BURKE,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 01/23/2021 05:53:28

Diagnostics Check		OK	05:40	
Air Blank		0.000	05:41	
Control Test		0.080	05:41	
Air Blank		0.000	05:42	
Subject Sample	# 1	0.017	05:45	Volume Not Met
Breath Volume		0.675	05:45	
Air Blank		0.000	05:45	
Air Blank		0.000	05:47	
Subject Sample	# 2	0.020	05:48	
Breath Volume		2.554	05:48	
Air Blank		0.000	05:49	
Air Blank		0.000	05:51	
Subject Sample	# 3	0.017	05:51	
Breath Volume		1.769	05:51	
Air Blank		0.000	05:52	
Control Test		0.080	05:52	
Air Blank		0.000	05:53	
Diagnostics Check		OK	05:53	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,JAMES,I
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MOORE,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/24/2021 05:04:57

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:02	
Control Test		0.081	05:02	
Air Blank		0.000	05:02	
Subject Sample	# 1		05:03	Subject Test Refused
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Control Test		0.080	05:04	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 01/24/2021 18:19:34

Diagnostics Check		OK	18:10
Air Blank		0.000	18:11
Control Test		0.080	18:11
Air Blank		0.000	18:12
Subject Sample	# 1	0.183	18:14
Breath Volume		3.125	18:14
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.190	18:17
Breath Volume		3.511	18:17
Air Blank		0.000	18:18
Control Test		0.080	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEICHNER,ADAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MARINO,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VINSON

DATE/TIME OF TEST: 01/29/2021 02:31:05

Diagnostics Check	OK	02:30	
Air Blank		02:30	RFI Detect
Air Blank	0.000	02:31	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEICHNER,ADAM,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MARINO,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VINSON

DATE/TIME OF TEST: 01/29/2021 02:35:33

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.081	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1		02:34	Subject Test Refused
Breath Volume		0.000	02:34	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,JAMES,I
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TIMMONS,ALLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LANDRUM

DATE/TIME OF TEST: 01/30/2021 03:34:33

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:32	
Control Test		0.081	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.082	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,JUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: TISDALE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/31/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: BAUCAS

DATE/TIME OF TEST: 01/30/2021 11:22:56

Diagnostics Check		OK	11:16
Air Blank		0.000	11:16
Control Test		0.081	11:16
Air Blank		0.000	11:17
Subject Sample	# 1	0.000	11:18
Breath Volume		2.953	11:18
Air Blank		0.000	11:19
Air Blank		0.000	11:21
Subject Sample	# 2	0.000	11:21
Breath Volume		2.664	11:21
Air Blank		0.000	11:22
Control Test		0.081	11:22
Air Blank		0.000	11:22
Diagnostics Check		OK	11:22

Data Download Date/Time: 02/01/2021 12:38

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LANTZ

DATE/TIME OF TEST: 01/01/2021 01:28:57

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.079	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:24	No Sample Provided
Breath Volume		0.000	01:24	
Air Blank		0.000	01:25	
Air Blank		0.000	01:27	
Subject Sample	# 2		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:28	
Control Test		0.078	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/02/2021 15:06

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DEPUTY
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 01/01/2021 02:03:50

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.078	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.075	01:58
Breath Volume		4.960	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.074	02:02
Breath Volume		2.906	02:02
Air Blank		0.000	02:02
Control Test		0.075	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/01/2021 02:43:45

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.077	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1	0.211	02:39	
Breath Volume		1.910	02:39	
Air Blank		0.000	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 2	0.213	02:42	
Breath Volume		2.128	02:42	
Air Blank		0.000	02:42	
Control Test		0.073	02:43	Control Outside Tolerance
Air Blank		0.000	02:43	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 01/01/2021 05:36:40

Diagnostics Check		OK	05:33	
Air Blank		0.000	05:34	
Control Test		0.077	05:34	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:35	Subject Test Refused
Breath Volume		0.000	05:35	
Air Blank		0.000	05:35	
Control Test		0.078	05:36	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GUERRY,CHAD
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/02/2021 00:34:38

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.078	00:33	
Air Blank		00:34	RFI Detect
Air Blank	0.000	00:34	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUSGROVE,DUSTIN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GUERRY,CHAD
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/02/2021 00:42:32

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.078	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.000	00:38
Breath Volume		3.011	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.000	00:41
Breath Volume		2.949	00:41
Air Blank		0.000	00:41
Control Test		0.078	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,C,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: KERNAN,M
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/03/2021 02:57:24

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.080	02:55	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,STUART,P
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: MONTEMURNO,MATTHEW
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 01/05/2021 15:27:52

Diagnostics Check		OK	15:24	
Air Blank		0.000	15:25	
Control Test		0.079	15:25	
Air Blank		0.000	15:26	
Subject Sample	# 1		15:26	Subject Test Refused
Breath Volume		0.000	15:26	
Air Blank		0.000	15:26	
Control Test		0.079	15:27	
Air Blank		0.000	15:27	
Diagnostics Check		OK	15:27	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: GOHDE,JAYME
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 01/07/2021 16:20:49

Diagnostics Check		OK	16:14
Air Blank		0.000	16:14
Control Test		0.078	16:15
Air Blank		0.000	16:15
Subject Sample	# 1	0.179	16:16
Breath Volume		1.613	16:16
Air Blank		0.000	16:16
Air Blank		0.000	16:18
Subject Sample	# 2	0.175	16:19
Breath Volume		1.238	16:19
Air Blank		0.000	16:19
Control Test		0.078	16:20
Air Blank		0.000	16:20
Diagnostics Check		OK	16:20

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PARISI,T
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 01/07/2021 21:21:05

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.079	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.280	21:15
Breath Volume		3.011	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.290	21:19
Breath Volume		3.125	21:19
Air Blank		0.000	21:20
Control Test		0.075	21:20
Air Blank		0.000	21:20
Diagnostics Check		OK	21:21

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BROWN,BRITTANY
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 01/10/2021 02:09:16

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.149	02:04
Breath Volume		1.292	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.143	02:07
Breath Volume		1.164	02:07
Air Blank		0.000	02:08
Control Test		0.076	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DANDRA,M
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: STEPHENS,MAURICE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VAZQUEZ MORALES

DATE/TIME OF TEST: 01/11/2021 03:13:02

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.078	03:02	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:11	Subject Test Refused
Breath Volume		0.039	03:11	
Air Blank		0.000	03:12	
Control Test		0.079	03:12	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SKUMATZ

DATE/TIME OF TEST: 01/14/2021 20:32:29

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.079	20:26
Air Blank		0.000	20:26
Subject Sample	# 1	0.068	20:27
Breath Volume		1.246	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.062	20:30
Breath Volume		1.410	20:30
Air Blank		0.000	20:31
Control Test		0.077	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASTOR,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KASTOR,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: CAULEY

DATE/TIME OF TEST: 01/15/2021 18:27:08

Diagnostics Check		OK	18:20	
Air Blank		0.000	18:21	
Control Test		0.077	18:21	
Air Blank		0.000	18:21	
Subject Sample	# 1	0.272	18:22	
Breath Volume		2.687	18:22	
Air Blank		0.000	18:23	
Air Blank		0.000	18:24	
Subject Sample	# 2	0.261	18:25	
Breath Volume		3.082	18:25	
Air Blank		0.000	18:26	
Control Test		0.073	18:26	Control Outside Tolerance
Air Blank		0.000	18:27	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PARISI,T
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 01/16/2021 23:54:25

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.242	23:49
Breath Volume		2.269	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.247	23:52
Breath Volume		2.328	23:52
Air Blank		0.000	23:53
Control Test		0.076	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: JEPKO,L
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 01/24/2021 03:45:47

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.122	03:41
Breath Volume		2.019	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.120	03:44
Breath Volume		2.101	03:44
Air Blank		0.000	03:44
Control Test		0.077	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: ROURKS

DATE/TIME OF TEST: 01/29/2021 03:29:20

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.213	03:21
Breath Volume		1.519	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.191	03:24
Breath Volume		1.613	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 3	0.211	03:27
Breath Volume		2.273	03:27
Air Blank		0.000	03:28
Control Test		0.075	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KAUS

DATE/TIME OF TEST: 01/30/2021 02:47:36

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.078	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.126	02:43
Breath Volume		3.390	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.134	02:45
Breath Volume		3.003	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: KOSKO,OFFICER
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: LITTLES

DATE/TIME OF TEST: 01/31/2021 03:05:33

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		03:00	No Sample Provided
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:04	
Control Test		0.080	03:04	
Air Blank		0.000	03:05	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/30/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 01/31/2021 21:26:38

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.078	21:24	
Air Blank		0.000	21:25	
Subject Sample	# 1		21:25	Subject Test Refused
Breath Volume		0.000	21:25	
Air Blank		0.000	21:25	
Control Test		0.078	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 02/02/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 01/04/2021 04:40:05

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:33	
Control Test		0.081	04:33	
Air Blank		0.000	04:33	
Subject Sample	# 1	0.230	04:35	
Breath Volume		2.082	04:35	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 2	0.221	04:38	
Breath Volume		1.640	04:38	
Air Blank		0.000	04:39	
Control Test		0.086	04:39	Control Outside Tolerance
Air Blank		0.000	04:40	

Data Download Date/Time: 01/19/2021 11:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BARGE,SUSAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 01/04/2021 04:49:01

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.084	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.237	04:43
Breath Volume		2.382	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.246	04:47
Breath Volume		2.695	04:47
Air Blank		0.000	04:48
Control Test		0.081	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:49

Data Download Date/Time: 01/19/2021 11:21

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,DAVID,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COX,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: CRIPES

DATE/TIME OF TEST: 01/02/2021 07:44:53

Diagnostics Check		OK	07:38
Air Blank		0.000	07:39
Control Test		0.079	07:39
Air Blank		0.000	07:39
Subject Sample	# 1	0.111	07:40
Breath Volume		1.914	07:40
Air Blank		0.000	07:41
Air Blank		0.000	07:42
Subject Sample	# 2	0.112	07:43
Breath Volume		2.261	07:43
Air Blank		0.000	07:43
Control Test		0.078	07:44
Air Blank		0.000	07:44
Diagnostics Check		OK	07:44

Data Download Date/Time: 01/19/2021 13:24

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYLE,HARRY,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PYLE,HARRY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 01/03/2021 17:58:05

Diagnostics Check		OK	17:51
Air Blank		0.000	17:51
Control Test		0.079	17:51
Air Blank		0.000	17:52
Subject Sample	# 1	0.135	17:53
Breath Volume		1.679	17:53
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.141	17:56
Breath Volume		2.710	17:56
Air Blank		0.000	17:57
Control Test		0.078	17:57
Air Blank		0.000	17:58
Diagnostics Check		OK	17:58

Data Download Date/Time: 01/19/2021 13:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LERNER,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: LERNER,MATTHEW
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MCMILLAN

DATE/TIME OF TEST: 01/04/2021 02:00:30

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.101	01:56
Breath Volume		3.851	01:56
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.101	01:58
Breath Volume		4.050	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 01/19/2021 13:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FHP TROOP A
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,WILFREDO,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ARROYO,WILFREDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HEWETT

DATE/TIME OF TEST: 01/05/2021 21:59:12

Diagnostics Check		OK	21:52
Air Blank		0.000	21:53
Control Test		0.078	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.099	21:54
Breath Volume		1.464	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.101	21:57
Breath Volume		2.015	21:57
Air Blank		0.000	21:58
Control Test		0.079	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 01/19/2021 13:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAWYER,DARRYL,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAWYER,DARRYL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/03/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 01/10/2021 05:06:55

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.083	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1		05:05	Subject Test Refused
Breath Volume		0.000	05:05	
Air Blank		0.000	05:05	
Control Test		0.083	05:06	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 01/20/2021 14:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 01/18/2021 23:59:34

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.083	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.304	23:54
Breath Volume		2.203	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.307	23:57
Breath Volume		2.320	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 01/20/2021 14:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROSE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: KESLER

DATE/TIME OF TEST: 01/25/2021 20:56:24

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:53	
Control Test		0.082	20:54	
Air Blank		0.000	20:54	
Subject Sample	# 1		20:55	Subject Test Refused
Breath Volume		0.000	20:55	
Air Blank		0.000	20:55	
Control Test		0.082	20:55	
Air Blank		0.000	20:56	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 02/08/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/16/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: WENIG

DATE/TIME OF TEST: 01/30/2021 17:25:38

Diagnostics Check		OK	17:19
Air Blank		0.000	17:20
Control Test		0.082	17:20
Air Blank		0.000	17:20
Subject Sample	# 1	0.000	17:21
Breath Volume		1.898	17:21
Air Blank		0.000	17:21
Air Blank		0.000	17:23
Subject Sample	# 2	0.000	17:24
Breath Volume		1.789	17:24
Air Blank		0.000	17:24
Control Test		0.081	17:25
Air Blank		0.000	17:25
Diagnostics Check		OK	17:25

Data Download Date/Time: 02/08/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/31/2021 04:32:47

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.082	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.086	04:27
Breath Volume		2.984	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.086	04:30
Breath Volume		4.476	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 02/08/2021 15:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JESSICA,L
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: MAYO,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 01/10/2021 02:52:39

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.189	02:46
Breath Volume		1.515	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.189	02:50
Breath Volume		1.414	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 01/25/2021 09:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,JOEY,C
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: BARTON,GREG
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 12/08/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/21/2021 23:41:31

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:39	
Control Test		0.081	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 01/25/2021 09:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:11

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/31/2021 12:50:55

Diagnostics Check		OK	12:44
Air Blank		0.000	12:44
Control Test		0.080	12:45
Air Blank		0.000	12:45
Subject Sample	# 1	0.283	12:46
Breath Volume		2.402	12:46
Air Blank		0.000	12:47
Air Blank		0.000	12:48
Subject Sample	# 2	0.282	12:49
Breath Volume		2.117	12:49
Air Blank		0.000	12:50
Control Test		0.078	12:50
Air Blank		0.000	12:50
Diagnostics Check		OK	12:50

Data Download Date/Time: 02/12/2021 12:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: STOKOVICH,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: FELL

DATE/TIME OF TEST: 01/01/2021 01:18:00

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.078	01:15	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.078	01:17	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 01/04/2021 13:32

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/03/2021 01:36:25

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.080	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.023	01:32
Breath Volume		3.050	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.021	01:34
Breath Volume		3.328	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 01/04/2021 13:32

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/11/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BRITSCH

DATE/TIME OF TEST: 01/04/2021 00:10:17

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.141	00:05
Breath Volume		1.480	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.137	00:08
Breath Volume		2.324	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 01/04/2021 13:32

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 01/06/2021 23:29:58

Diagnostics Check		OK	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:25
Subject Sample	# 1	0.196	23:25
Breath Volume		1.871	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.177	23:28
Breath Volume		1.125	23:28
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 02/03/2021 23:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: LUNDAY

DATE/TIME OF TEST: 01/08/2021 23:46:53

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.000	23:42
Breath Volume		4.042	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.000	23:45
Breath Volume		4.125	23:45
Air Blank		0.000	23:45
Control Test		0.082	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 02/03/2021 23:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1094494
DRY GAS EXPIRY DATE: 02/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 01/16/2021 23:58:15

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.080	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.042	23:53
Breath Volume		1.664	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.040	23:56
Breath Volume		1.699	23:56
Air Blank		0.000	23:57
Control Test		0.081	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 02/03/2021 23:40

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGLIONI,DAVID,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CHRISTY

DATE/TIME OF TEST: 01/01/2021 20:33:04

Diagnostics Check	Fail	20:32	Internal Printer Test
Air Blank	0.000	20:33	

Data Download Date/Time: 01/28/2021 07:57

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGLIONI,DAVID,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CHRISTY

DATE/TIME OF TEST: 01/01/2021 20:41:55

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.081	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.176	20:37
Breath Volume		2.816	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.184	20:40
Breath Volume		3.234	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 01/28/2021 07:57

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGLIONI,DAVID,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CASTANOS

DATE/TIME OF TEST: 01/02/2021 01:58:21

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.127	01:53
Breath Volume		1.785	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.132	01:56
Breath Volume		2.128	01:56
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 01/28/2021 07:57

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: POPOVICH,K
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/02/2021 05:43:58

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.081	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.199	05:38
Breath Volume		2.531	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.187	05:42
Breath Volume		2.601	05:42
Air Blank		0.000	05:43
Control Test		0.077	05:43
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINERNEY,JESSE,L
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MASTRANGIOLA,NATALIE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MOORHEAD

DATE/TIME OF TEST: 01/04/2021 02:11:53

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.301	02:05
Breath Volume		2.566	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.294	02:10
Breath Volume		2.601	02:10
Air Blank		0.000	02:10
Control Test		0.078	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRAWDY,LAINE,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DRAWDY,LAINE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: ATTARDI

DATE/TIME OF TEST: 01/08/2021 18:11:33

Diagnostics Check		OK	18:05
Air Blank		0.000	18:05
Control Test		0.081	18:05
Air Blank		0.000	18:06
Subject Sample	# 1	0.137	18:07
Breath Volume		2.871	18:07
Air Blank		0.000	18:07
Air Blank		0.000	18:09
Subject Sample	# 2	0.127	18:10
Breath Volume		1.359	18:10
Air Blank		0.000	18:10
Control Test		0.081	18:10
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 01/10/2021 02:35:31

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.082	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.225	02:31
Breath Volume		2.445	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.224	02:33
Breath Volume		2.972	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEFLANG,DANIEL,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SEMBOOKA,KUNJAROO
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: FORLENZA

DATE/TIME OF TEST: 01/10/2021 23:10:34

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.081	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.205	23:06
Breath Volume		1.175	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.205	23:08
Breath Volume		1.566	23:08
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGARD,RICKY,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HAGGARD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DEVITO

DATE/TIME OF TEST: 01/12/2021 04:17:05

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.000	04:12
Breath Volume		3.484	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.000	04:15
Breath Volume		4.195	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: FIELD

DATE/TIME OF TEST: 01/14/2021 00:50:29

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:48	
Control Test		0.080	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.080	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: FIELD

DATE/TIME OF TEST: 01/14/2021 01:01:06

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.149	00:56
Breath Volume		1.777	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:58
Subject Sample	# 2	0.165	00:59
Breath Volume		2.667	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 01/14/2021 20:41:53

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.081	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.000	20:37
Breath Volume		3.257	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.000	20:40
Breath Volume		1.843	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/18/2021 04:54:13

Diagnostics Check		OK	04:39	
Air Blank		0.000	04:40	
Control Test		0.080	04:40	
Air Blank		0.000	04:41	
Subject Sample	# 1	0.210	04:42	No .020 agreement
Breath Volume		1.253	04:42	
Air Blank		0.000	04:43	
Air Blank		0.000	04:45	
Subject Sample	# 2	0.178	04:46	No .020 agreement
Breath Volume		1.750	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:49	
Subject Sample	# 3	0.197	04:52	Volume Not Met
Breath Volume		0.289	04:52	
Air Blank		0.000	04:53	
Control Test		0.079	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/18/2021 05:03:01

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.175	04:58
Breath Volume		2.234	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:00
Subject Sample	# 2	0.187	05:01
Breath Volume		3.585	05:01
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:03

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: OBREGON,JULIO
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: HENKELS

DATE/TIME OF TEST: 01/20/2021 00:19:07

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.081	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.126	00:14
Breath Volume		3.136	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:16
Subject Sample	# 2	0.121	00:17
Breath Volume		3.187	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: THINEL

DATE/TIME OF TEST: 01/22/2021 00:00:32

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.395	23:53
Breath Volume		1.210	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.379	23:58
Breath Volume		2.062	23:58
Air Blank		0.000	23:59
Control Test		0.079	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: ANTON,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 01/24/2021 03:06:28

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.130	03:02
Breath Volume		1.406	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.121	03:04
Breath Volume		1.273	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 01/28/2021 07:52

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCUOPPO,ANTHONY,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: CHAPARRO,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: KOONTZ

DATE/TIME OF TEST: 01/24/2021 23:32:51

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.081	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.009	23:28
Breath Volume		1.503	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.007	23:31
Breath Volume		1.589	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 01/28/2021 07:53

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CORTLAND,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: JONES,CORTLAND
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/23/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: SAN GEORGE

DATE/TIME OF TEST: 01/26/2021 17:55:35

Diagnostics Check		OK	17:49
Air Blank		0.000	17:49
Control Test		0.080	17:50
Air Blank		0.000	17:50
Subject Sample	# 1	0.037	17:51
Breath Volume		2.382	17:51
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.039	17:54
Breath Volume		2.183	17:54
Air Blank		0.000	17:54
Control Test		0.080	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 01/28/2021 07:53

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ESTANISLAU,DANIEL
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PEREZ RODRIGUEZ

DATE/TIME OF TEST: 01/30/2021 00:29:42

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.081	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.258	00:25
Breath Volume		2.242	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.259	00:28
Breath Volume		1.609	00:28
Air Blank		0.000	00:28
Control Test		0.080	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/25/2021 14:36

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: VIAUD

DATE/TIME OF TEST: 01/12/2021 23:47:30

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.082	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.184	23:43
Breath Volume		3.191	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.173	23:45
Breath Volume		3.085	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/14/2021 11:30

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/13/2021 02:52:02

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.072	02:47
Breath Volume		2.593	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:49
Subject Sample	# 2	0.067	02:50
Breath Volume		1.582	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:52

Data Download Date/Time: 01/14/2021 11:30

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 01/15/2021 03:19:17

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.081	03:18	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 02/16/2021 13:41

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 01/15/2021 03:35:17

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.081	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.137	03:27	Volume Not Met
Breath Volume		0.691	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 2	0.166	03:30	
Breath Volume		1.515	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 3	0.161	03:33	
Breath Volume		1.281	03:33	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 02/16/2021 13:41

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 01/15/2021 05:02:13

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.081	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.147	04:57
Breath Volume		1.558	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.149	05:00
Breath Volume		1.214	05:00
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 02/16/2021 13:41

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: STEPHEN

DATE/TIME OF TEST: 01/15/2021 21:39:59

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.081	21:32
Air Blank		0.000	21:32
Subject Sample	# 1	0.147	21:35
Breath Volume		1.542	21:35
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.150	21:38
Breath Volume		1.164	21:38
Air Blank		0.000	21:39
Control Test		0.079	21:39
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 01/16/2021 04:30:08

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.187	04:24
Breath Volume		1.625	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.183	04:28
Breath Volume		1.535	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SOTO,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 01/16/2021 05:54:39

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.081	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.166	05:50
Breath Volume		2.164	05:50
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.158	05:53
Breath Volume		2.390	05:53
Air Blank		0.000	05:53
Control Test		0.081	05:54
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CABRERA-PORTILLO

DATE/TIME OF TEST: 01/17/2021 02:24:54

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.082	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.153	02:20
Breath Volume		1.914	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.158	02:23
Breath Volume		1.617	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 01/17/2021 04:04:35

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.081	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.152	04:00
Breath Volume		1.656	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.144	04:03
Breath Volume		1.183	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SCANNELL JR

DATE/TIME OF TEST: 01/18/2021 01:13:49

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.150	01:08
Breath Volume		1.542	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.152	01:12
Breath Volume		1.566	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: EVERSLEIGH

DATE/TIME OF TEST: 01/19/2021 06:13:12

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.081	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.127	06:08
Breath Volume		1.898	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.132	06:11
Breath Volume		1.562	06:11
Air Blank		0.000	06:12
Control Test		0.082	06:12
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CASTLE

DATE/TIME OF TEST: 01/20/2021 02:10:06

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.000	02:05
Breath Volume		2.972	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.000	02:08
Breath Volume		3.214	02:08
Air Blank		0.000	02:09
Control Test		0.081	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 01/20/2021 03:11:17

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.082	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.140	03:06
Breath Volume		2.691	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.143	03:09
Breath Volume		2.531	03:09
Air Blank		0.000	03:10
Control Test		0.081	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/20/2021 06:13:17

Diagnostics Check		OK	06:05
Air Blank		0.000	06:05
Control Test		0.081	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.169	06:08
Breath Volume		1.167	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.163	06:11
Breath Volume		1.906	06:11
Air Blank		0.000	06:12
Control Test		0.080	06:12
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I 253212
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: YOUNG JR

DATE/TIME OF TEST: 01/20/2021 08:57:16

Diagnostics Check		OK	08:48
Air Blank		0.000	08:49
Control Test		0.081	08:49
Air Blank		0.000	08:50
Subject Sample	# 1	0.000	08:52
Breath Volume		2.140	08:52
Air Blank		0.000	08:53
Air Blank		0.000	08:55
Subject Sample	# 2	0.000	08:55
Breath Volume		2.714	08:55
Air Blank		0.000	08:56
Control Test		0.081	08:56
Air Blank		0.000	08:57
Diagnostics Check		OK	08:57

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 01/22/2021 00:57:08

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.194	00:52
Breath Volume		3.460	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.191	00:55
Breath Volume		1.527	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 01/22/2021 02:02:29

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.202	01:57
Breath Volume		2.144	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.210	02:00
Breath Volume		2.097	02:00
Air Blank		0.000	02:01
Control Test		0.081	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/22/2021 05:09:12

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.081	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.131	05:04
Breath Volume		2.257	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.132	05:07
Breath Volume		1.738	05:07
Air Blank		0.000	05:08
Control Test		0.079	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 01/22/2021 21:18:55

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.080	21:03	
Air Blank		0.000	21:04	
Subject Sample	# 1	0.114	21:07	Volume Not Met
Breath Volume		0.531	21:07	
Air Blank		0.000	21:08	
Air Blank		0.000	21:10	
Subject Sample	# 2	0.132	21:12	
Breath Volume		1.414	21:12	
Air Blank		0.000	21:13	
Air Blank		0.000	21:15	
Subject Sample	# 3	0.135	21:17	
Breath Volume		1.261	21:17	
Air Blank		0.000	21:17	
Control Test		0.079	21:18	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MC KEE

DATE/TIME OF TEST: 01/23/2021 00:09:56

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:04
Air Blank		0.000	00:04
Subject Sample	# 1	0.283	00:05
Breath Volume		2.574	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:07
Subject Sample	# 2	0.294	00:08
Breath Volume		2.890	00:08
Air Blank		0.000	00:09
Control Test		0.081	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 02/16/2021 13:42

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: VALENTIN DIAZ

DATE/TIME OF TEST: 01/23/2021 03:29:44

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.129	03:22	Volume Not Met
Breath Volume		0.078	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.162	03:28	Volume Not Met
Breath Volume		0.636	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: VALENTIN DIAZ

DATE/TIME OF TEST: 01/23/2021 03:43:16

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.079	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.159	03:35	Volume Not Met
Breath Volume		0.421	03:35	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.145	03:41	Volume Not Met
Breath Volume		0.789	03:41	
Air Blank		0.000	03:42	
Control Test		0.079	03:42	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: DRIGGERS

DATE/TIME OF TEST: 01/23/2021 04:59:21

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.080	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.000	04:55
Breath Volume		2.375	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.000	04:57
Breath Volume		2.199	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 01/23/2021 07:54:12

Diagnostics Check		OK	07:47	
Air Blank		0.000	07:48	
Control Test		0.080	07:48	
Air Blank		0.000	07:48	
Subject Sample	# 1	0.214	07:49	
Breath Volume		3.031	07:49	
Air Blank		0.000	07:50	
Air Blank		0.000	07:52	
Subject Sample	# 2		07:52	Slope Not Met
Breath Volume		1.148	07:52	
Air Blank		0.000	07:53	
Control Test		0.080	07:53	
Air Blank		0.000	07:54	
Diagnostics Check		OK	07:54	

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 01/23/2021 08:21:43

Diagnostics Check		OK	08:15
Air Blank		0.000	08:16
Control Test		0.080	08:16
Air Blank		0.000	08:16
Subject Sample	# 1	0.199	08:17
Breath Volume		3.867	08:17
Air Blank		0.000	08:17
Air Blank		0.000	08:19
Subject Sample	# 2	0.192	08:20
Breath Volume		2.417	08:20
Air Blank		0.000	08:20
Control Test		0.079	08:21
Air Blank		0.000	08:21
Diagnostics Check		OK	08:21

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 01/24/2021 00:18:19

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.104	00:13
Breath Volume		1.664	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.102	00:16
Breath Volume		2.054	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/24/2021 02:48:32

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.095	02:43
Breath Volume		1.644	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.095	02:46
Breath Volume		1.382	02:46
Air Blank		0.000	02:47
Control Test		0.081	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RODRIGUEZ-AGUILAR

DATE/TIME OF TEST: 01/24/2021 05:01:51

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Subject Sample	# 1	0.072	04:57
Breath Volume		2.332	04:57
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.073	05:00
Breath Volume		1.980	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:01
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILKINS,MALACHI
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: BUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: BIFULCO

DATE/TIME OF TEST: 01/24/2021 18:08:58

Diagnostics Check		OK	18:02
Air Blank		0.000	18:02
Control Test		0.081	18:03
Air Blank		0.000	18:03
Subject Sample	# 1	0.000	18:04
Breath Volume		1.375	18:04
Air Blank		0.000	18:04
Air Blank		0.000	18:06
Subject Sample	# 2	0.000	18:07
Breath Volume		4.382	18:07
Air Blank		0.000	18:08
Control Test		0.081	18:08
Air Blank		0.000	18:08
Diagnostics Check		OK	18:08

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: LASSALLE

DATE/TIME OF TEST: 01/24/2021 22:25:16

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.081	22:15
Air Blank		0.000	22:16
Subject Sample	# 1	0.290	22:17
Breath Volume		1.535	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:19
Subject Sample	# 2	0.269	22:20
Breath Volume		1.917	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 3	0.283	22:23
Breath Volume		1.507	22:23
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: COATE

DATE/TIME OF TEST: 01/24/2021 23:45:00

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.081	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.179	23:40
Breath Volume		2.402	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.173	23:43
Breath Volume		1.882	23:43
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:45

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,DESMAND
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 01/25/2021 05:22:44

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:20	
Control Test		0.080	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 1		05:21	Subject Test Refused
Breath Volume		0.000	05:21	
Air Blank		0.000	05:21	
Control Test		0.079	05:22	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MOYE

DATE/TIME OF TEST: 01/26/2021 01:18:13

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.187	01:13
Breath Volume		3.000	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.197	01:16
Breath Volume		2.730	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 01/26/2021 05:10:19

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.081	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.191	05:05
Breath Volume		2.640	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.185	05:08
Breath Volume		2.519	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAVER,A 242967
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: SHOATS

DATE/TIME OF TEST: 01/26/2021 16:46:40

Diagnostics Check		OK	16:38
Air Blank		0.000	16:39
Control Test		0.081	16:39
Air Blank		0.000	16:40
Subject Sample	# 1	0.339	16:41
Breath Volume		1.886	16:41
Air Blank		0.000	16:42
Air Blank		0.000	16:44
Subject Sample	# 2	0.356	16:45
Breath Volume		2.476	16:45
Air Blank		0.000	16:45
Control Test		0.079	16:46
Air Blank		0.000	16:46
Diagnostics Check		OK	16:46

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SORIANO

DATE/TIME OF TEST: 01/26/2021 23:14:29

Diagnostics Check		OK	23:07
Air Blank		0.000	23:08
Control Test		0.081	23:08
Air Blank		0.000	23:09
Subject Sample	# 1	0.157	23:09
Breath Volume		2.480	23:09
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.153	23:12
Breath Volume		1.980	23:12
Air Blank		0.000	23:13
Control Test		0.080	23:13
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/27/2021 01:52:17

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.081	01:41	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:45	No Sample Provided
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 2	0.105	01:50	Volume Not Met
Breath Volume		0.035	01:50	
Air Blank		0.000	01:51	
Control Test		0.081	01:51	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 02/16/2021 13:43

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CSONTOS

DATE/TIME OF TEST: 01/28/2021 01:23:35

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.000	01:19
Breath Volume		1.472	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.000	01:22
Breath Volume		1.593	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 01/28/2021 03:00:19

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.103	02:55
Breath Volume		2.902	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.098	02:58
Breath Volume		2.230	02:58
Air Blank		0.000	02:59
Control Test		0.081	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:56

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 01/28/2021 08:29:16

Diagnostics Check		OK	08:22
Air Blank		0.000	08:22
Control Test		0.082	08:22
Air Blank		0.000	08:23
Subject Sample	# 1	0.062	08:24
Breath Volume		1.289	08:24
Air Blank		0.000	08:25
Air Blank		0.000	08:27
Subject Sample	# 2	0.058	08:27
Breath Volume		1.203	08:27
Air Blank		0.000	08:28
Control Test		0.081	08:28
Air Blank		0.000	08:29
Diagnostics Check		OK	08:29

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULLIVAN,T 4034
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: SEHRES

DATE/TIME OF TEST: 01/28/2021 14:29:34

Diagnostics Check		OK	14:23
Air Blank		0.000	14:23
Control Test		0.082	14:23
Air Blank		0.000	14:24
Subject Sample	# 1	0.000	14:24
Breath Volume		3.433	14:24
Air Blank		0.000	14:25
Air Blank		0.000	14:27
Subject Sample	# 2	0.000	14:28
Breath Volume		3.398	14:28
Air Blank		0.000	14:28
Control Test		0.081	14:28
Air Blank		0.000	14:29
Diagnostics Check		OK	14:29

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 01/29/2021 01:44:06

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.171	01:39
Breath Volume		2.199	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.178	01:42
Breath Volume		2.527	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 01/29/2021 03:04:07

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.082	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.195	02:59
Breath Volume		1.355	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.199	03:02
Breath Volume		1.671	03:02
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: DURSO

DATE/TIME OF TEST: 01/30/2021 02:16:36

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.083	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.082	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: MARTINEZ GARCIA

DATE/TIME OF TEST: 01/30/2021 21:44:08

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.082	21:38
Air Blank		0.000	21:38
Subject Sample	# 1	0.166	21:39
Breath Volume		1.906	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.168	21:42
Breath Volume		1.449	21:42
Air Blank		0.000	21:43
Control Test		0.081	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LOPEZ ROJAS

DATE/TIME OF TEST: 01/31/2021 00:35:00

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.082	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.175	00:30
Breath Volume		2.324	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.166	00:33
Breath Volume		3.082	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:35

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: LAM

DATE/TIME OF TEST: 01/31/2021 05:32:53

Diagnostics Check		OK	05:27	
Air Blank		0.000	05:28	
Control Test		0.081	05:28	
Air Blank		0.000	05:29	
Subject Sample	# 1	0.169	05:29	
Breath Volume		1.679	05:29	
Air Blank		0.000	05:30	
Air Blank		0.000	05:32	
Subject Sample	# 2		05:32	Sequence Aborted
Breath Volume		0.000	05:32	
Air Blank		0.000	05:32	

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: LAM

DATE/TIME OF TEST: 01/31/2021 06:06:03

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.081	06:00
Air Blank		0.000	06:01
Subject Sample	# 1	0.168	06:01
Breath Volume		1.441	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:03
Subject Sample	# 2	0.167	06:04
Breath Volume		2.300	06:04
Air Blank		0.000	06:05
Control Test		0.081	06:05
Air Blank		0.000	06:05
Diagnostics Check		OK	06:06

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: JUICA

DATE/TIME OF TEST: 01/31/2021 19:14:48

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.081	19:08
Air Blank		0.000	19:09
Subject Sample	# 1	0.192	19:10
Breath Volume		2.222	19:10
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.182	19:13
Breath Volume		1.148	19:13
Air Blank		0.000	19:13
Control Test		0.080	19:14
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: DIAZ NEGRON

DATE/TIME OF TEST: 01/31/2021 20:24:33

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.082	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.083	20:20
Breath Volume		2.445	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.082	20:23
Breath Volume		2.742	20:23
Air Blank		0.000	20:23
Control Test		0.081	20:23
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 02/16/2021 13:44

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROFT,BRETT,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROFT,BRETT
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: OTOOLE

DATE/TIME OF TEST: 01/01/2021 07:39:04

Diagnostics Check		OK	07:29
Air Blank		0.000	07:30
Control Test		0.081	07:30
Air Blank		0.000	07:31
Subject Sample	# 1	0.123	07:33
Breath Volume		1.914	07:33
Air Blank		0.000	07:34
Air Blank		0.000	07:36
Subject Sample	# 2	0.119	07:37
Breath Volume		1.578	07:37
Air Blank		0.000	07:38
Control Test		0.078	07:38
Air Blank		0.000	07:38
Diagnostics Check		OK	07:39

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: KILLOREN

DATE/TIME OF TEST: 01/02/2021 02:12:22

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.081	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.166	02:06	Volume Not Met
Breath Volume		0.933	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 2		02:10	Subject Test Refused
Breath Volume		0.062	02:10	
Air Blank		0.000	02:11	
Control Test		0.077	02:11	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CANDELA,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HERWARTH

DATE/TIME OF TEST: 01/02/2021 03:38:01

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.235	03:33
Breath Volume		3.531	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.224	03:36
Breath Volume		2.000	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:38

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: DEHESA

DATE/TIME OF TEST: 01/06/2021 01:54:40

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.081	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.217	01:50
Breath Volume		1.164	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.211	01:52
Breath Volume		1.597	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: BRUNETTI

DATE/TIME OF TEST: 01/06/2021 21:16:27

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.082	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.000	21:12
Breath Volume		1.664	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.000	21:15
Breath Volume		1.164	21:15
Air Blank		0.000	21:15
Control Test		0.081	21:15
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/09/2021 22:45:07

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.081	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.194	22:40
Breath Volume		1.859	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:42
Subject Sample	# 2	0.192	22:43
Breath Volume		1.648	22:43
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBEY,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: PERYEA

DATE/TIME OF TEST: 01/10/2021 00:11:12

Diagnostics Check		OK	00:02	
Air Blank		0.000	00:03	
Control Test		0.081	00:03	
Air Blank		0.000	00:04	
Subject Sample	# 1	0.000	00:07	Volume Not Met
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Air Blank		0.000	00:09	
Subject Sample	# 2		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:10	
Control Test		0.078	00:10	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,TRCY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 01/10/2021 16:44:48

Diagnostics Check		OK	16:38
Air Blank		0.000	16:38
Control Test		0.080	16:38
Air Blank		0.000	16:39
Subject Sample	# 1	0.305	16:39
Breath Volume		1.992	16:39
Air Blank		0.000	16:40
Air Blank		0.000	16:42
Subject Sample	# 2	0.292	16:43
Breath Volume		2.218	16:43
Air Blank		0.000	16:43
Control Test		0.078	16:44
Air Blank		0.000	16:44
Diagnostics Check		OK	16:44

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,IVAN,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DOLAN,N
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/11/2021 19:34:23

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:21	
Control Test		0.080	19:21	
Air Blank		0.000	19:22	
Subject Sample	# 1	0.339	19:23	No .020 agreement
Breath Volume		1.402	19:23	
Air Blank		0.000	19:24	
Air Blank		0.000	19:25	
Subject Sample	# 2	0.313	19:26	No .020 agreement
Breath Volume		1.789	19:26	
Air Blank		0.000	19:27	
Air Blank		0.000	19:29	
Subject Sample	# 3	0.310	19:32	Volume Not Met
Breath Volume		0.000	19:32	
Air Blank		0.000	19:33	
Control Test		0.076	19:33	
Air Blank		0.000	19:34	
Diagnostics Check		OK	19:34	

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/12/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DE ALMEIDA

DATE/TIME OF TEST: 01/12/2021 00:08:47

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.079	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.064	00:03
Breath Volume		1.308	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.064	00:07
Breath Volume		1.710	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 01/20/2021 11:12

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,TERENCE,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MCDONALD,A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 01/27/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GRAMMAN

DATE/TIME OF TEST: 01/29/2021 00:36:44

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.077	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.149	00:31
Breath Volume		1.605	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.147	00:35
Breath Volume		2.093	00:35
Air Blank		0.000	00:35
Control Test		0.076	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 02/17/2021 11:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 01/01/2021 00:45:01

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.170	00:40
Breath Volume		1.750	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.166	00:43
Breath Volume		1.640	00:43
Air Blank		0.000	00:44
Control Test		0.078	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:45

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEVORAK,MATT,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RIZZO,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/29/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: COLDSNOW

DATE/TIME OF TEST: 01/02/2021 02:46:01

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.078	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.170	02:41
Breath Volume		2.203	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:43
Subject Sample	# 2	0.166	02:44
Breath Volume		3.058	02:44
Air Blank		0.000	02:45
Control Test		0.076	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:46

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,BRANDON,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 01/03/2021 00:37:36

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:35	
Control Test		0.078	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:36	Subject Test Refused
Breath Volume		0.000	00:36	
Air Blank		0.000	00:36	
Control Test		0.077	00:36	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,BRANDON,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RIZZO,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: PAYNE-CUPETO

DATE/TIME OF TEST: 01/03/2021 05:01:32

Diagnostics Check		OK	04:58	
Air Blank		0.000	04:59	
Control Test		0.078	04:59	
Air Blank		0.000	04:59	
Subject Sample	# 1		05:00	Subject Test Refused
Breath Volume		0.000	05:00	
Air Blank		0.000	05:00	
Control Test		0.078	05:00	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: MCCULLOUGH,X
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 01/09/2021 01:08:13

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.144	01:03
Breath Volume		2.257	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.139	01:06
Breath Volume		1.683	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,BRANDON,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RIZZO,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HOLDEN

DATE/TIME OF TEST: 01/15/2021 22:58:53

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:46	
Control Test		0.078	22:46	
Air Blank		0.000	22:47	
Subject Sample	# 1		22:50	No Sample Provided
Breath Volume		0.000	22:50	
Air Blank		0.000	22:50	
Air Blank		0.000	22:52	
Subject Sample	# 2	0.174	22:53	
Breath Volume		6.421	22:53	
Air Blank		0.000	22:54	
Air Blank		0.000	22:56	
Subject Sample	# 3	0.176	22:57	
Breath Volume		5.765	22:57	
Air Blank		0.000	22:57	
Control Test		0.077	22:58	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,BRANDON,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BOSWELL,A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/16/2021 00:46:06

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:41	
Control Test		0.077	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:45	
Control Test		0.077	00:45	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,BRANDON,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HANEY,M
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: ANDERSON MARHEINE

DATE/TIME OF TEST: 01/16/2021 23:33:02

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:30	
Control Test		0.078	23:30	
Air Blank		0.000	23:31	
Subject Sample	# 1		23:31	Subject Test Refused
Breath Volume		0.000	23:31	
Air Blank		0.000	23:32	
Control Test		0.078	23:32	
Air Blank		0.000	23:32	
Diagnostics Check		OK	23:33	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HORTON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: YOUNCE

DATE/TIME OF TEST: 01/18/2021 15:07:51

Diagnostics Check		OK	15:00
Air Blank		0.000	15:01
Control Test		0.079	15:01
Air Blank		0.000	15:02
Subject Sample	# 1	0.000	15:03
Breath Volume		1.417	15:03
Air Blank		0.000	15:03
Air Blank		0.000	15:05
Subject Sample	# 2	0.000	15:06
Breath Volume		1.984	15:06
Air Blank		0.000	15:06
Control Test		0.078	15:07
Air Blank		0.000	15:07
Diagnostics Check		OK	15:07

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/19/2021 21:06:46

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.078	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.175	21:02
Breath Volume		1.464	21:02
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.174	21:05
Breath Volume		2.203	21:05
Air Blank		0.000	21:05
Control Test		0.078	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LUX

DATE/TIME OF TEST: 01/20/2021 22:59:51

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:57	
Control Test		0.077	22:57	
Air Blank		0.000	22:58	
Subject Sample	# 1		22:58	Subject Test Refused
Breath Volume		0.000	22:58	
Air Blank		0.000	22:58	
Control Test		0.077	22:59	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,BRANDON,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FAGOUSEH,A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/21/2021 02:59:04

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.078	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.078	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PAOLI

DATE/TIME OF TEST: 01/23/2021 00:47:01

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.078	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1	0.178	00:34	Volume Not Met
Breath Volume		0.726	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 2	0.191	00:39	
Breath Volume		1.484	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 3	0.190	00:45	
Breath Volume		1.792	00:45	
Air Blank		0.000	00:46	
Control Test		0.076	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BOX

DATE/TIME OF TEST: 01/23/2021 02:20:13

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.076	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:19	
Control Test		0.077	02:19	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: MCCULLOUGH,X
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 01/23/2021 04:57:31

Diagnostics Check		OK	04:54	
Air Blank		0.000	04:54	
Control Test		0.078	04:54	
Air Blank		0.000	04:55	
Subject Sample	# 1		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Control Test		0.078	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASON,CHRISTOPHER,N
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: RAMJIT,E
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 01/24/2021 03:54:14

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.078	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:47	No Sample Provided
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2		03:52	No Sample Provided
Breath Volume		0.000	03:52	
Air Blank		0.000	03:53	
Control Test		0.078	03:53	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARCE,TROY,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VIERA,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PEARCE

DATE/TIME OF TEST: 01/29/2021 03:44:47

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.078	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.000	03:40
Breath Volume		2.492	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.000	03:43
Breath Volume		2.132	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MELCHIONNE

DATE/TIME OF TEST: 01/29/2021 22:57:05

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.077	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.249	22:52
Breath Volume		1.675	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.250	22:55
Breath Volume		1.371	22:55
Air Blank		0.000	22:56
Control Test		0.077	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 02/17/2021 09:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: BUEHLER

DATE/TIME OF TEST: 01/09/2021 16:30:02

Diagnostics Check		OK	16:23
Air Blank		0.000	16:24
Control Test		0.077	16:24
Air Blank		0.000	16:25
Subject Sample	# 1	0.342	16:25
Breath Volume		2.445	16:25
Air Blank		0.000	16:26
Air Blank		0.000	16:28
Subject Sample	# 2	0.348	16:28
Breath Volume		3.113	16:28
Air Blank		0.000	16:29
Control Test		0.078	16:29
Air Blank		0.000	16:29
Diagnostics Check		OK	16:30

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DE MOLA

DATE/TIME OF TEST: 01/10/2021 02:20:08

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.077	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.000	02:15
Breath Volume		2.652	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.000	02:18
Breath Volume		1.503	02:18
Air Blank		0.000	02:19
Control Test		0.078	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 01/10/2021 15:06:46

Diagnostics Check		OK	15:00
Air Blank		0.000	15:00
Control Test		0.077	15:01
Air Blank		0.000	15:01
Subject Sample	# 1	0.258	15:02
Breath Volume		2.753	15:02
Air Blank		0.000	15:03
Air Blank		0.000	15:04
Subject Sample	# 2	0.258	15:05
Breath Volume		2.257	15:05
Air Blank		0.000	15:05
Control Test		0.077	15:06
Air Blank		0.000	15:06
Diagnostics Check		OK	15:06

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADE,CODY,L
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RAPP,ERIC
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/12/2021 00:43:32

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.077	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.226	00:39
Breath Volume		2.464	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.207	00:41
Breath Volume		2.417	00:41
Air Blank		0.000	00:42
Control Test		0.075	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURCHINSON,RONNIE,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RICKETTS,JEREMIAH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 01/13/2021 04:14:08

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.077	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.183	04:09
Breath Volume		2.566	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.177	04:12
Breath Volume		1.296	04:12
Air Blank		0.000	04:13
Control Test		0.077	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RONALD,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: RODRIGUEZ,RONALD
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: PEDRAZA-MEZA

DATE/TIME OF TEST: 01/17/2021 04:35:53

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.077	04:33	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	
Control Test		0.077	04:35	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: COLSTON BARNES,MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: RESENDIZ-HERNANDEZ

DATE/TIME OF TEST: 01/17/2021 16:12:20

Diagnostics Check		OK	16:09	
Air Blank		0.000	16:09	
Control Test		0.077	16:09	
Air Blank		0.000	16:10	
Subject Sample	# 1		16:10	Subject Test Refused
Breath Volume		0.000	16:10	
Air Blank		0.000	16:11	
Control Test		0.078	16:11	
Air Blank		0.000	16:12	
Diagnostics Check		OK	16:12	

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: WILLIAMS,AUSTIN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: ANDERSEN

DATE/TIME OF TEST: 01/19/2021 14:02:11

Diagnostics Check		OK	13:55
Air Blank		0.000	13:55
Control Test		0.077	13:56
Air Blank		0.000	13:56
Subject Sample	# 1	0.179	13:57
Breath Volume		3.164	13:57
Air Blank		0.000	13:58
Air Blank		0.000	14:00
Subject Sample	# 2	0.173	14:00
Breath Volume		2.421	14:00
Air Blank		0.000	14:01
Control Test		0.077	14:01
Air Blank		0.000	14:02
Diagnostics Check		OK	14:02

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RONALD,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: RODRIGUEZ,RONALD
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: WOODRUFF

DATE/TIME OF TEST: 01/21/2021 17:25:29

Diagnostics Check		OK	17:19
Air Blank		0.000	17:19
Control Test		0.077	17:20
Air Blank		0.000	17:20
Subject Sample	# 1	0.000	17:21
Breath Volume		2.523	17:21
Air Blank		0.000	17:21
Air Blank		0.000	17:23
Subject Sample	# 2	0.000	17:24
Breath Volume		3.023	17:24
Air Blank		0.000	17:24
Control Test		0.077	17:24
Air Blank		0.000	17:25
Diagnostics Check		OK	17:25

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,DAVID,E
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RATLIFF,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/22/2021 20:51:02

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:38	
Control Test		0.076	20:38	
Air Blank		0.000	20:39	
Subject Sample	# 1	0.200	20:42	Volume Not Met
Breath Volume		0.980	20:42	
Air Blank		0.000	20:43	
Air Blank		0.000	20:44	
Subject Sample	# 2	0.203	20:46	
Breath Volume		1.484	20:46	
Air Blank		0.000	20:46	
Air Blank		0.000	20:48	
Subject Sample	# 3	0.203	20:49	
Breath Volume		1.843	20:49	
Air Blank		0.000	20:50	
Control Test		0.076	20:50	
Air Blank		0.000	20:50	
Diagnostics Check		OK	20:51	

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: COMBEE

DATE/TIME OF TEST: 01/23/2021 00:24:38

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.076	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.076	00:24	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRY,ROBERT
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: FABRICO

DATE/TIME OF TEST: 01/23/2021 04:31:52

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.076	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.192	04:27
Breath Volume		1.468	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.186	04:30
Breath Volume		1.105	04:30
Air Blank		0.000	04:30
Control Test		0.076	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,EDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BENENATI

DATE/TIME OF TEST: 01/24/2021 02:52:12

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.148	02:47
Breath Volume		2.519	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.134	02:50
Breath Volume		1.707	02:50
Air Blank		0.000	02:51
Control Test		0.077	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: TOLE

DATE/TIME OF TEST: 01/25/2021 21:44:36

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.076	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.188	21:39
Breath Volume		2.769	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.184	21:43
Breath Volume		2.843	21:43
Air Blank		0.000	21:43
Control Test		0.076	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: CALKINS

DATE/TIME OF TEST: 01/30/2021 00:02:34

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.077	23:55
Air Blank		0.000	23:55
Subject Sample	# 2	0.213	23:57
Breath Volume		4.292	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.206	00:01
Breath Volume		4.164	00:01
Air Blank		0.000	00:01
Control Test		0.077	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,JOSHUA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: LUPIAN

DATE/TIME OF TEST: 01/30/2021 03:11:02

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.077	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.172	03:06
Breath Volume		3.269	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:08
Subject Sample	# 2	0.166	03:09
Breath Volume		3.839	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:11

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,BROOKE,M
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: WHEELER,BROOKE
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/30/2021 03:43:45

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.077	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Control Test		0.077	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: SYLDOR,JEAN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 01/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: HERNANDEZ-DIAZ

DATE/TIME OF TEST: 01/31/2021 15:39:13

Diagnostics Check		OK	15:31
Air Blank		0.000	15:31
Control Test		0.076	15:31
Air Blank		0.000	15:32
Subject Sample	# 1	0.228	15:34
Breath Volume		1.972	15:34
Air Blank		0.000	15:35
Air Blank		0.000	15:37
Subject Sample	# 2	0.237	15:37
Breath Volume		1.214	15:37
Air Blank		0.000	15:38
Control Test		0.075	15:38
Air Blank		0.000	15:39
Diagnostics Check		OK	15:39

Data Download Date/Time: 02/03/2021 23:58

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UNION,LYDIA,
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: TAYLOR,OCTAVIOUS
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: UNION

DATE/TIME OF TEST: 01/01/2021 02:06:35

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.081	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:05	Subject Test Refused
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Control Test		0.082	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 01/13/2021 16:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BAYLARK

DATE/TIME OF TEST: 01/01/2021 04:01:20

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.082	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.134	03:56
Breath Volume		2.300	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.129	03:59
Breath Volume		1.792	03:59
Air Blank		0.000	04:00
Control Test		0.082	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 01/13/2021 16:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CUTLIFF

DATE/TIME OF TEST: 01/01/2021 21:07:07

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.082	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.166	21:02
Breath Volume		2.902	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:04
Subject Sample	# 2	0.166	21:05
Breath Volume		2.453	21:05
Air Blank		0.000	21:06
Control Test		0.082	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 01/13/2021 16:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ARCE NAVARRETE

DATE/TIME OF TEST: 01/02/2021 00:44:08

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.083	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.000	00:39
Breath Volume		1.535	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.000	00:42
Breath Volume		2.117	00:42
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/02/2021 06:53:01

Diagnostics Check		OK	06:45
Air Blank		0.000	06:46
Control Test		0.081	06:46
Air Blank		0.000	06:47
Subject Sample	# 1	0.259	06:48
Breath Volume		3.617	06:48
Air Blank		0.000	06:48
Air Blank		0.000	06:50
Subject Sample	# 2	0.252	06:51
Breath Volume		2.949	06:51
Air Blank		0.000	06:52
Control Test		0.081	06:52
Air Blank		0.000	06:52
Diagnostics Check		OK	06:53

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WILLIAMS,Z
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: FREE

DATE/TIME OF TEST: 01/04/2021 00:18:13

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.083	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:16	Subject Test Refused
Breath Volume		0.000	00:16	
Air Blank		0.000	00:17	
Control Test		0.083	00:17	
Air Blank		0.000	00:18	
Diagnostics Check		OK	00:18	

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: SCOTT,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: LAVA-SABLAN

DATE/TIME OF TEST: 01/04/2021 00:42:09

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.083	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	
Control Test		0.083	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LANDA,JAVIER
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: JEWELL

DATE/TIME OF TEST: 01/04/2021 22:09:36

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.083	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.161	22:05
Breath Volume		3.296	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.150	22:08
Breath Volume		3.683	22:08
Air Blank		0.000	22:08
Control Test		0.080	22:09
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: JENNEY

DATE/TIME OF TEST: 01/05/2021 03:45:31

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.083	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.084	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: JOHNSON,DEREK
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 01/08/2021 02:00:16

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.083	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Control Test		0.082	01:59	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,C,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RATLIFF,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 01/08/2021 10:51:35

Diagnostics Check		OK	10:46	
Air Blank		0.000	10:47	
Control Test		0.082	10:47	
Air Blank		0.000	10:47	
Subject Sample	# 1	0.221	10:49	
Breath Volume		1.554	10:49	
Air Blank			10:49	Improper Sample
Air Blank		0.000	10:50	
Control Test		0.083	10:50	
Air Blank		0.000	10:51	
Diagnostics Check		OK	10:51	

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,CHRIS,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RATLIFF,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 01/08/2021 11:01:15

Diagnostics Check		OK	10:53
Air Blank		0.000	10:53
Control Test		0.080	10:54
Air Blank		0.000	10:54
Subject Sample	# 1	0.223	10:56
Breath Volume		1.773	10:56
Air Blank		0.000	10:57
Air Blank		0.000	10:58
Subject Sample	# 2	0.225	10:59
Breath Volume		2.636	10:59
Air Blank		0.000	11:00
Control Test		0.078	11:00
Air Blank		0.000	11:01
Diagnostics Check		OK	11:01

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1099238
DRY GAS EXPIRY DATE: 03/01/2022
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: MONTERO

DATE/TIME OF TEST: 01/08/2021 11:10:11

Diagnostics Check		OK	11:03
Air Blank		0.000	11:04
Control Test		0.081	11:04
Air Blank		0.000	11:05
Subject Sample	# 1	0.000	11:05
Breath Volume		2.312	11:05
Air Blank		0.000	11:06
Air Blank		0.000	11:08
Subject Sample	# 2	0.000	11:08
Breath Volume		2.464	11:08
Air Blank		0.000	11:09
Control Test		0.082	11:09
Air Blank		0.000	11:10
Diagnostics Check		OK	11:10

Data Download Date/Time: 01/13/2021 16:05

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WALSH,SHANNON
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/02/2021 03:33:08

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.081	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.163	03:28
Breath Volume		1.195	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.165	03:31
Breath Volume		1.289	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 01/20/2021 20:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBBS,K,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 01/05/2021 14:50:53

Diagnostics Check		OK	14:43
Air Blank		0.000	14:43
Control Test		0.081	14:44
Air Blank		0.000	14:44
Subject Sample	# 1	0.274	14:46
Breath Volume		2.304	14:46
Air Blank		0.000	14:47
Air Blank		0.000	14:48
Subject Sample	# 2	0.268	14:49
Breath Volume		1.660	14:49
Air Blank		0.000	14:49
Control Test		0.079	14:50
Air Blank		0.000	14:50
Diagnostics Check		OK	14:50

Data Download Date/Time: 01/20/2021 20:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LENFANT,BENJAMIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 01/08/2021 22:41:26

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:39	
Control Test		0.081	22:39	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.080	22:40	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 01/20/2021 20:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LENFANT,BENJAMIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 01/10/2021 22:25:11

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.082	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.000	22:20
Breath Volume		3.472	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.000	22:23
Breath Volume		3.531	22:23
Air Blank		0.000	22:24
Control Test		0.080	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 01/20/2021 20:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: PEPPERS,BENJAMIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/04/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ROSENMUND

DATE/TIME OF TEST: 01/14/2021 03:16:22

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.080	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.080	03:15	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 01/20/2021 20:04

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DAGOSTINO,STEFANO
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 01/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 01/21/2021 02:51:38

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.081	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Control Test		0.083	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 03/02/2021 13:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DISMORE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BONAMY

DATE/TIME OF TEST: 01/01/2021 01:10:39

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.080	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:03	No Sample Provided
Breath Volume		0.000	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2		01:09	No Sample Provided
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.080	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 01/11/2021 18:24

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HERNBROTT

DATE/TIME OF TEST: 01/01/2021 02:27:15

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.000	02:22
Breath Volume		3.339	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.000	02:25
Breath Volume		1.847	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 01/11/2021 18:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 01/01/2021 02:52:20

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.109	02:47
Breath Volume		1.984	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.106	02:50
Breath Volume		1.808	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 01/11/2021 18:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/09/2021 02:02:41

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.172	01:58
Breath Volume		1.847	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.176	02:01
Breath Volume		1.464	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 01/11/2021 18:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SANCHEZ ZETINO

DATE/TIME OF TEST: 01/10/2021 02:11:00

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.131	02:06
Breath Volume		1.808	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.139	02:09
Breath Volume		2.925	02:09
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/11/2021 18:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 12/09/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/10/2021 21:00:16

Diagnostics Check		OK	20:53
Air Blank		0.000	20:54
Control Test		0.080	20:54
Air Blank		0.000	20:54
Subject Sample	# 1	0.283	20:55
Breath Volume		3.710	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.287	20:58
Breath Volume		3.351	20:58
Air Blank		0.000	20:59
Control Test		0.080	20:59
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 01/11/2021 18:25

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/13/2021 04:45:49

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.081	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.177	04:41
Breath Volume		1.343	04:41
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.167	04:44
Breath Volume		1.164	04:44
Air Blank		0.000	04:44
Control Test		0.079	04:45
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 03/01/2021 11:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: COTELO BADIA

DATE/TIME OF TEST: 01/14/2021 00:51:13

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.206	00:46
Breath Volume		1.417	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.204	00:49
Breath Volume		1.859	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 03/01/2021 11:07

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WEBB-HERNANDEZ

DATE/TIME OF TEST: 01/14/2021 05:39:40

Diagnostics Check		OK	05:32
Air Blank		0.000	05:32
Control Test		0.080	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.081	05:35
Breath Volume		1.585	05:35
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.082	05:38
Breath Volume		1.230	05:38
Air Blank		0.000	05:38
Control Test		0.079	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 03/01/2021 11:08

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: STEPHEN JR

DATE/TIME OF TEST: 01/15/2021 21:47:28

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.079	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.170	21:42
Breath Volume		3.359	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.171	21:45
Breath Volume		3.226	21:45
Air Blank		0.000	21:46
Control Test		0.079	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 03/01/2021 11:08

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/16/2021 03:01:41

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.168	02:57
Breath Volume		2.503	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.172	03:00
Breath Volume		2.378	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 03/01/2021 11:08

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MONTES CARABALLO

DATE/TIME OF TEST: 01/16/2021 04:12:50

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.080	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:12	Sequence Aborted
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MONTES CARABALLO

DATE/TIME OF TEST: 01/16/2021 04:40:48

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.156	04:35
Breath Volume		1.199	04:35
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.150	04:39
Breath Volume		1.207	04:39
Air Blank		0.000	04:39
Control Test		0.079	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: INMAN

DATE/TIME OF TEST: 01/16/2021 05:08:22

Diagnostics Check		OK	05:01
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.096	05:03
Breath Volume		3.789	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.096	05:06
Breath Volume		4.128	05:06
Air Blank		0.000	05:07
Control Test		0.078	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: HANDY

DATE/TIME OF TEST: 01/16/2021 23:03:28

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.080	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.184	22:58
Breath Volume		2.351	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.176	23:01
Breath Volume		1.742	23:01
Air Blank		0.000	23:02
Control Test		0.079	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: ZANDER

DATE/TIME OF TEST: 01/17/2021 06:39:31

Diagnostics Check		OK	06:33
Air Blank		0.000	06:33
Control Test		0.080	06:33
Air Blank		0.000	06:34
Subject Sample	# 1	0.165	06:34
Breath Volume		1.492	06:34
Air Blank		0.000	06:35
Air Blank		0.000	06:37
Subject Sample	# 2	0.170	06:38
Breath Volume		3.171	06:38
Air Blank		0.000	06:38
Control Test		0.080	06:38
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 01/18/2021 01:24:07

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.118	01:19
Breath Volume		1.972	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.126	01:22
Breath Volume		1.730	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/22/2021 04:52:09

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.080	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:51	Sequence Aborted
Breath Volume		0.000	04:51	
Air Blank		0.000	04:52	

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/22/2021 05:23:23

Diagnostics Check		OK	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.075	05:18
Breath Volume		1.574	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:20
Subject Sample	# 2	0.077	05:21
Breath Volume		2.136	05:21
Air Blank		0.000	05:22
Control Test		0.080	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: REVERE III

DATE/TIME OF TEST: 01/23/2021 02:01:32

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.116	01:56
Breath Volume		1.726	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.114	02:00
Breath Volume		1.406	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/23/2021 02:18:49

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.112	02:14
Breath Volume		3.023	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.116	02:17
Breath Volume		3.050	02:17
Air Blank		0.000	02:17
Control Test		0.079	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/23/2021 04:16:57

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.179	04:12
Breath Volume		2.992	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.176	04:15
Breath Volume		2.343	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: LEON-TORRES

DATE/TIME OF TEST: 01/23/2021 07:45:59

Diagnostics Check		OK	07:34	
Air Blank		0.000	07:34	
Control Test		0.079	07:35	
Air Blank		0.000	07:35	
Subject Sample	# 1		07:38	No Sample Provided
Breath Volume		0.000	07:38	
Air Blank		0.000	07:39	
Air Blank		0.000	07:41	
Subject Sample	# 2	0.139	07:44	Volume Not Met
Breath Volume		0.308	07:44	
Air Blank		0.000	07:45	
Control Test		0.079	07:45	
Air Blank		0.000	07:45	
Diagnostics Check		OK	07:45	

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: LEON-TORRES

DATE/TIME OF TEST: 01/23/2021 07:56:06

Diagnostics Check		OK	07:49
Air Blank		0.000	07:49
Control Test		0.078	07:49
Air Blank		0.000	07:50
Subject Sample	# 1	0.156	07:51
Breath Volume		3.070	07:51
Air Blank		0.000	07:51
Air Blank		0.000	07:53
Subject Sample	# 2	0.145	07:54
Breath Volume		2.394	07:54
Air Blank		0.000	07:55
Control Test		0.078	07:55
Air Blank		0.000	07:56
Diagnostics Check		OK	07:56

Data Download Date/Time: 03/01/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 01/24/2021 02:03:00

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.118	01:58
Breath Volume		2.429	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.116	02:01
Breath Volume		2.460	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:03

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MATA

DATE/TIME OF TEST: 01/24/2021 05:07:18

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:56	
Control Test		0.080	04:56	
Air Blank		0.000	04:57	
Subject Sample	# 1	0.187	05:00	Volume Not Met
Breath Volume		0.960	05:00	
Air Blank		0.000	05:00	
Air Blank		0.000	05:02	
Subject Sample	# 2	0.168	05:05	Volume Not Met
Breath Volume		0.199	05:05	
Air Blank		0.000	05:06	
Control Test		0.079	05:06	
Air Blank		0.000	05:07	
Diagnostics Check		OK	05:07	

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RODIGUEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: ELVIR GALLARDO

DATE/TIME OF TEST: 01/24/2021 05:37:11

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.079	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.154	05:32
Breath Volume		2.789	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.154	05:35
Breath Volume		1.996	05:35
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: EGIDO

DATE/TIME OF TEST: 01/24/2021 06:01:22

Diagnostics Check		OK	05:54
Air Blank		0.000	05:55
Control Test		0.079	05:55
Air Blank		0.000	05:56
Subject Sample	# 1	0.133	05:56
Breath Volume		2.101	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:59
Subject Sample	# 2	0.131	05:59
Breath Volume		1.101	05:59
Air Blank		0.000	06:00
Control Test		0.079	06:00
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: KONRAD

DATE/TIME OF TEST: 01/25/2021 00:45:55

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.080	00:38	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:41	No Sample Provided
Breath Volume		0.000	00:41	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:45	
Control Test		0.080	00:45	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: LANZA

DATE/TIME OF TEST: 01/25/2021 03:03:22

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.079	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 1	0.026	02:56	Volume Not Met
Breath Volume		0.523	02:56	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 2	0.021	03:01	Volume Not Met
Breath Volume		0.492	03:01	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: LANZA

DATE/TIME OF TEST: 01/25/2021 03:21:59

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.079	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1	0.025	03:11	
Breath Volume		1.230	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2	0.022	03:17	Volume Not Met
Breath Volume		0.613	03:17	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 3	0.022	03:20	
Breath Volume		1.929	03:20	
Air Blank		0.000	03:21	
Control Test		0.079	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RODRIGUEZ,D
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: SARDJA

DATE/TIME OF TEST: 01/25/2021 05:12:49

Diagnostics Check		OK	05:06
Air Blank		0.000	05:07
Control Test		0.079	05:07
Air Blank		0.000	05:07
Subject Sample	# 1	0.000	05:08
Breath Volume		1.414	05:08
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.000	05:11
Breath Volume		2.042	05:11
Air Blank		0.000	05:11
Control Test		0.079	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HERNANDEZ,A 3455
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:41

SUBJECT'S LAST NAME: ROSENCRANS

DATE/TIME OF TEST: 01/25/2021 14:15:06

Diagnostics Check		OK	14:05
Air Blank		0.000	14:05
Control Test		0.080	14:06
Air Blank		0.000	14:06
Subject Sample	# 1	0.303	14:07
Breath Volume		1.617	14:07
Air Blank		0.000	14:07
Air Blank		0.000	14:09
Subject Sample	# 2	0.326	14:10
Breath Volume		2.425	14:10
Air Blank		0.000	14:11
Air Blank		0.000	14:12
Subject Sample	# 3	0.324	14:13
Breath Volume		2.832	14:13
Air Blank		0.000	14:14
Control Test		0.079	14:14
Air Blank		0.000	14:15
Diagnostics Check		OK	14:15

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: TILLIS

DATE/TIME OF TEST: 01/26/2021 20:56:14

Diagnostics Check		OK	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.000	20:50
Breath Volume		1.398	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.000	20:54
Breath Volume		1.714	20:54
Air Blank		0.000	20:55
Control Test		0.079	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/27/2021 04:23:59

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.080	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:22	Subject Test Refused
Breath Volume		0.000	04:22	
Air Blank		0.000	04:23	
Control Test		0.079	04:23	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: RODRIGUEZ,D 241685
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 01/28/2021 07:44:50

Diagnostics Check		OK	07:38
Air Blank		0.000	07:38
Control Test		0.080	07:39
Air Blank		0.000	07:39
Subject Sample	# 1	0.000	07:40
Breath Volume		4.738	07:40
Air Blank		0.000	07:40
Air Blank		0.000	07:42
Subject Sample	# 2	0.000	07:43
Breath Volume		3.667	07:43
Air Blank		0.000	07:43
Control Test		0.080	07:44
Air Blank		0.000	07:44
Diagnostics Check		OK	07:44

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: DRISCOLL,B
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: BAUGHN

DATE/TIME OF TEST: 01/28/2021 23:04:28

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.081	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.020	22:59
Breath Volume		1.843	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.021	23:03
Breath Volume		3.000	23:03
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 03/01/2021 11:10

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ARREDONDO

DATE/TIME OF TEST: 01/30/2021 02:14:27

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.159	02:09
Breath Volume		2.199	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.160	02:12
Breath Volume		3.261	02:12
Air Blank		0.000	02:13
Control Test		0.081	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/30/2021 03:16:33

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.099	03:11
Breath Volume		2.066	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.102	03:15
Breath Volume		2.617	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 01/30/2021 04:20:18

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.095	04:15
Breath Volume		1.367	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.092	04:18
Breath Volume		1.992	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: TEMPLE TERRACE PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DONAGHY II

DATE/TIME OF TEST: 01/31/2021 01:10:01

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.108	01:05
Breath Volume		3.304	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.104	01:08
Breath Volume		2.820	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:10

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: FRIDRIKSSON

DATE/TIME OF TEST: 01/31/2021 02:50:51

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.080	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:49	
Control Test		0.080	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: HILLSBOROUGH CO SO
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ERB

DATE/TIME OF TEST: 01/31/2021 04:35:41

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.080	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.191	04:31
Breath Volume		2.484	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.175	04:34
Breath Volume		2.074	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: TAMPA PD
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BOWE

DATE/TIME OF TEST: 01/31/2021 05:06:37

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.130	05:00
Breath Volume		1.699	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.137	05:05
Breath Volume		1.425	05:05
Air Blank		0.000	05:05
Control Test		0.080	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: HILLSBOROUGH CO SO
ARRESTING OFFICER: HALE,W
ARRESTING OFFICER AGENCY: FHP TROOP C
LAST AGENCY INSPECTION: 01/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 01/31/2021 08:37:48

Diagnostics Check		OK	08:30
Air Blank		0.000	08:31
Control Test		0.080	08:31
Air Blank		0.000	08:32
Subject Sample	# 1	0.178	08:33
Breath Volume		2.230	08:33
Air Blank		0.000	08:33
Air Blank		0.000	08:35
Subject Sample	# 2	0.179	08:36
Breath Volume		1.675	08:36
Air Blank		0.000	08:36
Control Test		0.079	08:37
Air Blank		0.000	08:37
Diagnostics Check		OK	08:37

Data Download Date/Time: 03/01/2021 11:11

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/01/2021 17:43:23

Diagnostics Check	OK	17:42	
Air Blank		17:42	RFI Detect
Air Blank	0.000	17:43	

Data Download Date/Time: 01/08/2021 21:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/01/2021 17:50:54

Diagnostics Check		OK	17:44
Air Blank		0.000	17:45
Control Test		0.079	17:45
Air Blank		0.000	17:46
Subject Sample	# 1	0.000	17:46
Breath Volume		2.992	17:46
Air Blank		0.000	17:46
Air Blank		0.000	17:48
Subject Sample	# 2	0.000	17:49
Breath Volume		2.886	17:49
Air Blank		0.000	17:49
Control Test		0.077	17:50
Air Blank		0.000	17:50
Diagnostics Check		OK	17:50

Data Download Date/Time: 01/08/2021 21:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: WALSH,S
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/16/2020

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 01/03/2021 01:51:42

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.000	01:47
Breath Volume		1.566	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.000	01:50
Breath Volume		1.488	01:50
Air Blank		0.000	01:50
Control Test		0.078	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 01/08/2021 21:55

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 01/17/2021 05:52:53

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.081	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.189	05:48
Breath Volume		2.175	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.191	05:51
Breath Volume		1.914	05:51
Air Blank		0.000	05:51
Control Test		0.081	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 01/19/2021 13:20

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER,EDWARD,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER,EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MUIR

DATE/TIME OF TEST: 01/15/2021 22:19:04

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:15	
Control Test		0.078	22:16	
Air Blank		0.000	22:16	
Subject Sample	# 1	0.190	22:18	
Breath Volume		1.933	22:18	
Air Blank			22:18	RFI Detect
Air Blank			22:18	RFI Detect
Air Blank		0.000	22:19	

Data Download Date/Time: 01/19/2021 10:02

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER,EDWARD,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SWOGGER,EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MUIR

DATE/TIME OF TEST: 01/15/2021 22:27:05

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.190	22:22
Breath Volume		1.789	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:24
Subject Sample	# 2	0.180	22:25
Breath Volume		2.246	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:27

Data Download Date/Time: 01/19/2021 10:02

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DRENNING

DATE/TIME OF TEST: 01/17/2021 00:22:03

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.188	00:17
Breath Volume		1.109	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:19
Subject Sample	# 2	0.188	00:20
Breath Volume		1.746	00:20
Air Blank		0.000	00:21
Control Test		0.078	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:22

Data Download Date/Time: 01/19/2021 10:02

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: LEBEAU

DATE/TIME OF TEST: 01/17/2021 22:15:03

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:12	
Control Test		0.079	22:13	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.000	22:13	
Air Blank		0.000	22:14	
Control Test		0.079	22:14	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 01/19/2021 10:02

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/19/2021 17:58:02

Diagnostics Check		OK	17:51	
Air Blank		0.000	17:51	
Control Test		0.079	17:52	
Air Blank		0.000	17:52	
Subject Sample	# 1	0.000	17:53	
Breath Volume		1.781	17:53	
Air Blank		0.000	17:54	
Air Blank		0.000	17:56	
Subject Sample	# 2	0.000	17:56	
Breath Volume		1.687	17:56	
Air Blank		0.000	17:57	
Control Test		0.080	17:57	
Air Blank			17:57	RFI Detect
Air Blank		0.000	17:58	

Data Download Date/Time: 01/20/2021 11:09

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: LEDER

DATE/TIME OF TEST: 01/20/2021 14:38:48

Diagnostics Check		OK	14:32
Air Blank		0.000	14:32
Control Test		0.080	14:33
Air Blank		0.000	14:33
Subject Sample	# 1	0.415	14:34
Breath Volume		2.003	14:34
Air Blank		0.000	14:34
Air Blank		0.000	14:36
Subject Sample	# 2	0.419	14:37
Breath Volume		1.890	14:37
Air Blank		0.000	14:37
Control Test		0.079	14:38
Air Blank		0.000	14:38
Diagnostics Check		OK	14:38

Data Download Date/Time: 02/03/2021 08:32

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIDAL,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 01/22/2021 22:41:27

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:38	
Control Test		0.079	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.079	22:40	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 02/03/2021 08:32

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: STICKNEY

DATE/TIME OF TEST: 01/27/2021 04:22:27

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.080	04:19	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:21	Subject Test Refused
Breath Volume		0.000	04:21	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 02/03/2021 08:32

Printed: 19 May 2021

COBRA ©1997-2014 CMI, Inc.