Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 10/08/2021 01:03:55

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:58	
Control Test		0.077	00:58	
Air Blank		0.000	00:59	
Subject Sample	# 1	0.133	00:59	
Breath Volume		1.386	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 2		01:02	Interferent Detect
Breath Volume		1.316	01:02	
Air Blank		0.000	01:02	
Control Test		0.078	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: KAYE

DATE/TIME OF TEST: 10/08/2021 22:35:00

Diagnostics Check	OK	22:34	
Air Blank		22:34	Ambient Fail
Air Blank	0.000	22:35	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: KAYE

DATE/TIME OF TEST: 10/08/2021 22:43:00

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.159	22:38
Breath Volume		2.281	22:38
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.154	22:41
Breath Volume		3.132	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:43

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/14/2021 02:17:34

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.078	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:16	RFI Detect
Breath Volume		0.015	02:16	
Air Blank		0.000	02:16	
Control Test		0.077	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/14/2021 02:29:59

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.073	02:22
Breath Volume		1.414	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.074	02:28
Breath Volume		1.539	02:28
Air Blank		0.000	02:29
Control Test		0.077	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SIMS,WALKER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 10/14/2021 03:57:59

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.075	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.081	03:57	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 10/16/2021 02:40:22

Diagnostics Check	OK	02:39	
Air Blank		02:39	Ambient Fail
Air Blank	0.000	02:40	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 10/16/2021 02:42:37

Diagnostics Check	OK	02:41	
Air Blank		02:42	Ambient Fail
Air Blank	0.000	02:42	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 10/16/2021 02:50:08

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.210	02:45
Breath Volume		1.394	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.205	02:48
Breath Volume		2.121	02:48
Air Blank		0.000	02:49
Control Test		0.076	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 10/16/2021 23:05:10

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.121	23:00
Breath Volume		3.503	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.114	23:03
Breath Volume		2.175	23:03
Air Blank		0.000	23:04
Control Test		0.076	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: OROURKE

DATE/TIME OF TEST: 10/17/2021 02:25:07

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.097	02:20
Breath Volume		2.304	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.094	02:23
Breath Volume		1.699	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 10/20/2021 23:31:10

Diagnostics Check	OK	23:30	
Air Blank		23:30	Ambient Fail
Air Blank	0.000	23:31	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 10/20/2021 23:33:27

Diagnostics Check	OK	23:32	
Air Blank		23:32	RFI Detect
Air Blank	0.000	23:33	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 10/20/2021 23:36:19

Diagnostics Check	OK	23:35	
Air Blank		23:35	Ambient Fail
Air Blank	0.000	23:36	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 10/20/2021 23:43:50

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:38	
Control Test		0.079	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1	0.213	23:39	
Breath Volume		1.449	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 2	0.205	23:42	
Breath Volume		1.542	23:42	
Air Blank		0.000	23:42	
Control Test		0.074	23:43	Interferent Detect
Air Blank		0.000	23:43	

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: CARTOLANO

DATE/TIME OF TEST: 10/22/2021 01:25:42

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.000	01:21
Breath Volume		2.312	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.000	01:24
Breath Volume		2.777	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/22/2021 02:34:00

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.136	02:29
Breath Volume		2.648	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.134	02:32
Breath Volume		2.472	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 10/23/2021 03:16:08

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.082	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.144	03:11
Breath Volume		2.445	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.142	03:14
Breath Volume		1.605	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ALLADATIN

DATE/TIME OF TEST: 10/24/2021 01:26:49

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.081	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.102	01:22
Breath Volume		1.730	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.102	01:25
Breath Volume		1.847	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PORRAS

DATE/TIME OF TEST: 10/28/2021 02:47:43

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.119	02:43
Breath Volume		1.292	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.121	02:46
Breath Volume		1.734	02:46
Air Blank		0.000	02:46
Control Test		0.078	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/29/2021 11:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WHEELESS

DATE/TIME OF TEST: 10/02/2021 23:18:07

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.080	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.000	23:13
Breath Volume		2.707	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.000	23:16
Breath Volume		2.640	23:16
Air Blank		0.000	23:17
Control Test		0.081	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SIMS,WALKER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 10/09/2021 00:58:35

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.081	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.111	00:54
Breath Volume		2.843	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.094	00:57
Breath Volume		2.437	00:57
Air Blank		0.000	00:57
Control Test		0.081	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HITCH,ANDREW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: HANCOCK

DATE/TIME OF TEST: 10/10/2021 20:28:40

Diagnostics Check		OK	20:22
Air Blank		0.000	20:23
Control Test		0.081	20:23
Air Blank		0.000	20:23
Subject Sample	# 1	0.000	20:24
Breath Volume		2.906	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.000	20:27
Breath Volume		2.007	20:27
Air Blank		0.000	20:27
Control Test		0.082	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: COMBS,BRADLEY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/13/2021 18:57:48

Diagnostics Check		OK	18:51
Air Blank		0.000	18:52
Control Test		0.081	18:52
Air Blank		0.000	18:52
Subject Sample	# 1	0.000	18:53
Breath Volume		1.593	18:53
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.000	18:56
Breath Volume		4.777	18:56
Air Blank		0.000	18:56
Control Test		0.081	18:57
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: N,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DANIELS,ANDREW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 10/13/2021 21:01:55

Diagnostics Check		OK	20:55
Air Blank		0.000	20:55
Control Test		0.081	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.259	20:57
Breath Volume		1.921	20:57
Air Blank		0.000	20:58
Air Blank		0.000	20:59
Subject Sample	# 2	0.258	21:00
Breath Volume		1.722	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BROWN,BENJAMIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: LANGLEY

DATE/TIME OF TEST: 10/17/2021 21:45:18

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:41	
Control Test		0.081	21:42	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:43	Subject Test Refused
Breath Volume		0.000	21:43	
Air Blank		0.000	21:44	
Control Test		0.081	21:44	
Air Blank		0.000	21:45	
Diagnostics Check		OK	21:45	

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BAKER,JERRY
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 10/18/2021 02:16:43

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.081	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.081	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: NELSON,ROBERT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 10/19/2021 12:04:52

Diagnostics Check		OK	11:58
Air Blank		0.000	11:58
Control Test		0.081	11:59
Air Blank		0.000	11:59
Subject Sample	# 1	0.000	12:00
Breath Volume		2.878	12:00
Air Blank		0.000	12:00
Air Blank		0.000	12:02
Subject Sample	# 2	0.000	12:03
Breath Volume		2.660	12:03
Air Blank		0.000	12:03
Control Test		0.082	12:04
Air Blank		0.000	12:04
Diagnostics Check		OK	12:04

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TOWNSEND,THOMAS
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PERRELLA

DATE/TIME OF TEST: 10/22/2021 19:29:34

Diagnostics Check		OK	19:26	
Air Blank		0.000	19:26	
Control Test		0.081	19:27	
Air Blank		0.000	19:27	
Subject Sample	# 1		19:28	Subject Test Refused
Breath Volume		0.000	19:28	
Air Blank		0.000	19:28	
Control Test		0.081	19:28	
Air Blank		0.000	19:29	
Diagnostics Check		OK	19:29	

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SIMS,WALKER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WEICHT

DATE/TIME OF TEST: 10/23/2021 02:38:43

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.109	02:34
Breath Volume		2.734	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.109	02:37
Breath Volume		2.808	02:37
Air Blank		0.000	02:37
Control Test		0.081	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 10/29/2021 11:22

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DEMNER,AUTUMN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: WIGTON

DATE/TIME OF TEST: 10/21/2021 21:32:25

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.081	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.266	21:27
Breath Volume		1.453	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.275	21:30
Breath Volume		1.519	21:30
Air Blank		0.000	21:31
Control Test		0.081	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 11/29/2021 15:20

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GRABILL,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MUNION

DATE/TIME OF TEST: 10/22/2021 01:06:29

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.082	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:01
Breath Volume		3.355	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.000	01:05
Breath Volume		3.265	01:05
Air Blank		0.000	01:05
Control Test		0.082	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 11/29/2021 15:20

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,R,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ALI,S
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: PEARCE

DATE/TIME OF TEST: 10/24/2021 20:07:54

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:55	
Control Test		0.082	19:55	
Air Blank		0.000	19:56	
Subject Sample	# 1	0.108	19:59	Slope Not Level
Breath Volume		1.476	19:59	
Air Blank		0.000	20:00	
Air Blank		0.000	20:01	
Subject Sample	# 2	0.114	20:03	
Breath Volume		1.542	20:03	
Air Blank		0.000	20:04	
Air Blank		0.000	20:05	
Subject Sample	# 3	0.114	20:06	
Breath Volume		1.429	20:06	
Air Blank		0.000	20:06	
Control Test		0.080	20:07	
Air Blank		0.000	20:07	
Diagnostics Check		OK	20:07	

Data Download Date/Time: 11/29/2021 15:20

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATTS,SHANNON,N
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: STEINER,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:07

SUBJECT'S LAST NAME: LAGUNAS

DATE/TIME OF TEST: 10/26/2021 12:49:55

Diagnostics Check		OK	12:42	
Air Blank		0.000	12:42	
Control Test		0.082	12:43	
Air Blank		0.000	12:43	
Subject Sample	# 1	0.000	12:44	
Breath Volume		1.406	12:44	
Air Blank		0.000	12:45	
Air Blank		0.000	12:47	
Subject Sample	# 2		12:48	Subject Test Refused
Breath Volume		0.382	12:48	
Air Blank		0.000	12:49	
Control Test		0.080	12:49	
Air Blank		0.000	12:49	
Diagnostics Check		OK	12:49	

Data Download Date/Time: 10/29/2021 13:41

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,CRAIG,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WINLAND,BOBBY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: STIRES

DATE/TIME OF TEST: 10/26/2021 14:07:28

Diagnostics Check		OK	14:00	
Air Blank		0.000	14:01	
Control Test		0.082	14:01	
Air Blank		0.000	14:02	
Subject Sample	# 1	0.000	14:02	
Breath Volume		2.160	14:02	
Air Blank		0.000	14:03	
Air Blank		0.000	14:05	
Subject Sample	# 2	0.000	14:05	
Breath Volume		1.480	14:05	
Air Blank		0.000	14:06	
Control Test		0.080	14:06	
Air Blank		0.000	14:06	
Diagnostics Check		Fail	14:06	Temperature Regulation Test
Air Blank		0.000	14:07	

Data Download Date/Time: 10/29/2021 13:41

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOUGERE,MARCEL,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: REIF-JUNGWIRTH

DATE/TIME OF TEST: 10/03/2021 21:15:08

Diagnostics Check		OK	21:02	
Air Blank		0.000	21:03	
Control Test		0.080	21:03	
Air Blank		0.000	21:04	
Subject Sample	# 1	0.146	21:07	Volume Not Met
Breath Volume		0.613	21:07	
Air Blank		0.000	21:07	
Air Blank		0.000	21:09	
Subject Sample	# 2	0.165	21:10	
Breath Volume		2.093	21:10	
Air Blank		0.000	21:10	
Air Blank		0.000	21:12	
Subject Sample	# 3	0.157	21:13	
Breath Volume		1.175	21:13	
Air Blank		0.000	21:14	
Control Test		0.079	21:14	
Air Blank		0.000	21:15	
Diagnostics Check		OK	21:15	

Data Download Date/Time: 11/10/2021 12:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/10/2021 01:45:58

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.322	01:41
Breath Volume		1.496	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.315	01:44
Breath Volume		1.378	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 10/18/2021 14:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 10/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: RESENDEZ

DATE/TIME OF TEST: 10/20/2021 00:42:33

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.153	00:36
Breath Volume		1.304	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.149	00:40
Breath Volume		1.257	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 11/15/2021 14:50

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: GALLOWAY,PARKER
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: WOMBLE

DATE/TIME OF TEST: 10/06/2021 01:34:40

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.082	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.363	01:30
Breath Volume		2.472	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.356	01:33
Breath Volume		2.562	01:33
Air Blank		0.000	01:33
Control Test		0.081	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 10/07/2021 10:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: EVANS,MAURINE
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 10/09/2021 20:00:15

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:58	
Control Test		0.080	19:58	
Air Blank		0.000	19:59	
Subject Sample	# 1		19:59	Subject Test Refused
Breath Volume		0.000	19:59	
Air Blank			19:59	RFI Detect
Air Blank		0.000	20:00	

Data Download Date/Time: 10/22/2021 11:50

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: EVANS,MAURINE
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BRUNS

DATE/TIME OF TEST: 10/09/2021 20:05:53

Diagnostics Check		OK	20:03	
Air Blank		0.000	20:03	
Control Test		0.081	20:03	
Air Blank		0.000	20:04	
Subject Sample	# 1		20:04	Subject Test Refused
Breath Volume		0.000	20:04	
Air Blank		0.000	20:04	
Control Test		0.081	20:05	
Air Blank		0.000	20:05	
Diagnostics Check		OK	20:05	

Data Download Date/Time: 10/22/2021 11:50

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GRABILL,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: F43644556164

DATE/TIME OF TEST: 10/11/2021 22:15:39

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.079	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.245	22:09
Breath Volume		1.332	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.242	22:13
Breath Volume		1.750	22:13
Air Blank		0.000	22:14
Control Test		0.080	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 10/22/2021 11:50

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: NANGANO

DATE/TIME OF TEST: 10/15/2021 20:54:00

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.080	20:48
Air Blank		0.000	20:48
Subject Sample	# 1	0.114	20:49
Breath Volume		1.777	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:51
Subject Sample	# 2	0.111	20:52
Breath Volume		2.640	20:52
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:53
Diagnostics Check		OK	20:54

Data Download Date/Time: 10/22/2021 11:50

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SANTANA DEL MONTE

DATE/TIME OF TEST: 10/18/2021 01:53:34

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.143	01:48
Breath Volume		2.906	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.137	01:52
Breath Volume		2.886	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 10/22/2021 11:50

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SPELLER,BENJAMIN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/22/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/24/2021 01:17:06

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.185	01:11
Breath Volume		1.351	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.182	01:15
Breath Volume		1.730	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 11/15/2021 12:44

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 10/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SCARCELLA

DATE/TIME OF TEST: 10/30/2021 05:19:01

Diagnostics Check		OK	05:16	
Air Blank		0.000	05:16	
Control Test		0.081	05:16	
Air Blank		0.000	05:17	
Subject Sample	# 1		05:17	Subject Test Refused
Breath Volume		0.000	05:17	
Air Blank		0.000	05:18	
Control Test		0.080	05:18	
Air Blank		0.000	05:18	
Diagnostics Check		OK	05:19	

Data Download Date/Time: 11/15/2021 12:44

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CARLUCCI

DATE/TIME OF TEST: 10/01/2021 19:57:15

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.081	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.183	19:52
Breath Volume		1.632	19:52
Air Blank		0.000	19:53
Air Blank		0.000	19:55
Subject Sample	# 2	0.186	19:55
Breath Volume		1.164	19:55
Air Blank		0.000	19:56
Control Test		0.080	19:56
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 10/30/2021 18:27

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,JIMMIE,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: MORALES,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/03/2021 19:34:18

Diagnostics Check		OK	19:27
Air Blank		0.000	19:28
Control Test		0.080	19:28
Air Blank		0.000	19:29
Subject Sample	# 1	0.258	19:29
Breath Volume		2.484	19:29
Air Blank		0.000	19:30
Air Blank		0.000	19:32
Subject Sample	# 2	0.243	19:32
Breath Volume		1.429	19:32
Air Blank		0.000	19:33
Control Test		0.080	19:33
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 10/30/2021 18:27

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: DU,WESLEY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 10/06/2021 02:09:14

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.043	02:04
Breath Volume		2.062	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.043	02:07
Breath Volume		1.714	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 10/30/2021 18:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 10/21/2021 21:00:33

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:53	
Control Test		0.080	20:54	
Air Blank		0.000	20:54	
Subject Sample	# 1	0.169	20:56	
Breath Volume		2.582	20:56	
Air Blank		0.000	20:57	
Air Blank		0.000	20:58	
Subject Sample	# 2	0.184	20:59	
Breath Volume		1.445	20:59	
Air Blank			21:00	RFI Detect
Air Blank		0.000	21:00	

Data Download Date/Time: 10/30/2021 18:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 10/21/2021 21:10:52

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.177	21:06
Breath Volume		2.089	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.180	21:09
Breath Volume		2.226	21:09
Air Blank		0.000	21:09
Control Test		0.081	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 10/30/2021 18:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 10/26/2021 02:34:40

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.081	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.000	02:30
Breath Volume		2.136	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.000	02:33
Breath Volume		2.699	02:33
Air Blank		0.000	02:33
Control Test		0.081	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 10/30/2021 18:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LABRECQUE,COREY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LE BLANC JR

DATE/TIME OF TEST: 10/28/2021 22:34:01

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:31	
Control Test		0.080	22:31	
Air Blank		0.000	22:32	
Subject Sample	# 1		22:32	Subject Test Refused
Breath Volume		0.000	22:32	
Air Blank		0.000	22:33	
Control Test		0.080	22:33	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 10/30/2021 18:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 10/29/2021 19:03:04

Diagnostics Check		OK	18:54
Air Blank		0.000	18:55
Control Test		0.080	18:55
Air Blank		0.000	18:56
Subject Sample	# 1	0.176	18:58
Breath Volume		2.214	18:58
Air Blank		0.000	18:58
Air Blank		0.000	19:00
Subject Sample	# 2	0.184	19:01
Breath Volume		2.445	19:01
Air Blank		0.000	19:02
Control Test		0.081	19:02
Air Blank		0.000	19:02
Diagnostics Check		OK	19:03

Data Download Date/Time: 10/30/2021 18:28

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,JIMMIE,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 10/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: HUSAIN

DATE/TIME OF TEST: 10/30/2021 22:09:14

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.081	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.189	22:04
Breath Volume		2.238	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.189	22:07
Breath Volume		2.257	22:07
Air Blank		0.000	22:08
Control Test		0.081	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 11/17/2021 10:40

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DORR

DATE/TIME OF TEST: 10/09/2021 03:55:12

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.143	03:50
Breath Volume		2.160	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.151	03:53
Breath Volume		1.625	03:53
Air Blank		0.000	03:54
Control Test		0.078	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: SQUIER

DATE/TIME OF TEST: 10/09/2021 05:33:32

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.081	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.095	05:28
Breath Volume		2.207	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.096	05:32
Breath Volume		1.507	05:32
Air Blank		0.000	05:32
Control Test		0.081	05:32
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GEISS

DATE/TIME OF TEST: 10/15/2021 04:26:21

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.081	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1	0.123	04:14	Volume Not Met
Breath Volume		1.007	04:14	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 2	0.118	04:19	
Breath Volume		1.394	04:19	
Air Blank		0.000	04:20	
Air Blank		0.000	04:22	
Subject Sample	# 3	0.112	04:24	
Breath Volume		1.300	04:24	
Air Blank		0.000	04:25	
Control Test		0.081	04:25	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,HUNTER,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HARMON,JONTE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/17/2021 18:45:40

Diagnostics Check		OK	18:41	
Air Blank		0.000	18:42	
Control Test		0.081	18:42	
Air Blank		0.000	18:43	
Subject Sample	# 1		18:44	Subject Test Refused
Breath Volume		0.000	18:44	
Air Blank		0.000	18:44	
Control Test		0.082	18:45	
Air Blank		0.000	18:45	
Diagnostics Check		OK	18:45	

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 10:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:07

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 10/18/2021 11:33:58

Diagnostics Check		OK	11:27
Air Blank		0.000	11:28
Control Test		0.081	11:28
Air Blank		0.000	11:29
Subject Sample	# 1	0.098	11:29
Breath Volume		2.558	11:29
Air Blank		0.000	11:30
Air Blank		0.000	11:31
Subject Sample	# 2	0.093	11:32
Breath Volume		2.691	11:32
Air Blank		0.000	11:33
Control Test		0.081	11:33
Air Blank		0.000	11:33
Diagnostics Check		OK	11:33

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BOTZOC PANA

DATE/TIME OF TEST: 10/20/2021 01:00:13

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.183	00:55
Breath Volume		1.734	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:57
Subject Sample	# 2	0.169	00:58
Breath Volume		1.328	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: TURNER,WILLIAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ARCOS

DATE/TIME OF TEST: 10/20/2021 01:32:48

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1	0.178	01:25	Volume Not Met
Breath Volume		0.644	01:25	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 2	0.172	01:31	Volume Not Met
Breath Volume		0.117	01:31	
Air Blank		0.000	01:31	
Control Test		0.080	01:32	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: TURNER,WILLIAM
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: ARCOS

DATE/TIME OF TEST: 10/20/2021 01:41:01

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.192	01:36
Breath Volume		1.582	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.190	01:39
Breath Volume		1.449	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:41

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COUPE,JESSICA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 10/23/2021 09:02:40

Diagnostics Check		OK	08:56
Air Blank		0.000	08:56
Control Test		0.081	08:56
Air Blank		0.000	08:57
Subject Sample	# 1	0.181	08:58
Breath Volume		1.464	08:58
Air Blank		0.000	08:58
Air Blank		0.000	09:00
Subject Sample	# 2	0.189	09:01
Breath Volume		1.132	09:01
Air Blank		0.000	09:01
Control Test		0.082	09:02
Air Blank		0.000	09:02
Diagnostics Check		OK	09:02

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: RENTERIA

DATE/TIME OF TEST: 10/24/2021 07:34:55

Diagnostics Check		OK	07:28
Air Blank		0.000	07:28
Control Test		0.081	07:29
Air Blank		0.000	07:29
Subject Sample	# 1	0.152	07:30
Breath Volume		1.363	07:30
Air Blank		0.000	07:31
Air Blank		0.000	07:32
Subject Sample	# 2	0.153	07:33
Breath Volume		1.597	07:33
Air Blank		0.000	07:34
Control Test		0.081	07:34
Air Blank		0.000	07:34
Diagnostics Check		OK	07:34

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 10/30/2021 03:53:27

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.081	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.207	03:44	
Breath Volume		2.015	03:44	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 2	0.177	03:49	Volume Not Met
Breath Volume		0.097	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 3		03:52	Sequence Aborted
Breath Volume		0.000	03:52	
Air Blank		0.000	03:53	

Data Download Date/Time: 11/01/2021 11:02

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,HUNTER,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 10/30/2021 21:05:00

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.081	20:59
Air Blank		0.000	20:59
Subject Sample	# 1	0.296	21:00
Breath Volume		3.402	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:02
Subject Sample	# 2	0.300	21:03
Breath Volume		3.542	21:03
Air Blank		0.000	21:04
Control Test		0.081	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:05

Data Download Date/Time: 11/01/2021 11:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RODRIGUEZ PITRE,V
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: KNOLL

DATE/TIME OF TEST: 10/02/2021 03:49:28

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.156	03:45
Breath Volume		3.023	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.151	03:47
Breath Volume		2.531	03:47
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 10/13/2021 09:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PULIZZOTTO,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: GRIJALVA

DATE/TIME OF TEST: 10/03/2021 22:19:12

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.080	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.256	22:14
Breath Volume		2.105	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.261	22:17
Breath Volume		2.343	22:17
Air Blank		0.000	22:18
Control Test		0.079	22:18
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 10/13/2021 09:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RODRIGUEZ-PITRE,V
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/03/2021 23:48:57

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.196	23:44
Breath Volume		3.371	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:46
Subject Sample	# 2	0.190	23:47
Breath Volume		2.546	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 10/13/2021 09:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENRRIQUEZ,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WARD,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MENDEZDIAZ

DATE/TIME OF TEST: 10/06/2021 01:48:19

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.244	01:43
Breath Volume		2.480	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.242	01:46
Breath Volume		2.089	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 10/13/2021 09:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: VITERITTI

DATE/TIME OF TEST: 10/06/2021 22:25:34

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.080	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.000	22:21
Breath Volume		3.476	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.000	22:24
Breath Volume		3.039	22:24
Air Blank		0.000	22:24
Control Test		0.080	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 10/13/2021 09:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: COONFIELD

DATE/TIME OF TEST: 10/07/2021 00:15:11

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.208	00:10
Breath Volume		2.183	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.209	00:13
Breath Volume		2.156	00:13
Air Blank		0.000	00:14
Control Test		0.080	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 10/13/2021 09:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: ABRAHAMGONZALEZ

DATE/TIME OF TEST: 10/07/2021 21:18:15

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.129	21:13
Breath Volume		2.058	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.113	21:16
Breath Volume		1.421	21:16
Air Blank		0.000	21:17
Control Test		0.079	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 10/13/2021 09:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARVIN,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 10/08/2021 03:29:02

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.197	03:24
Breath Volume		3.046	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:26
Subject Sample	# 2	0.189	03:27
Breath Volume		2.019	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:29

Data Download Date/Time: 10/13/2021 09:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: EMANS,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WIMSATT

DATE/TIME OF TEST: 10/09/2021 04:15:11

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1	0.000	04:11	
Breath Volume		2.085	04:11	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 2		04:13	RFI Detect
Breath Volume		0.000	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 10/13/2021 09:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: EMANS,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WIMSATT

DATE/TIME OF TEST: 10/09/2021 04:23:59

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.000	04:19
Breath Volume		1.839	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.000	04:22
Breath Volume		2.140	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 10/13/2021 09:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BRENNAN,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: O'HARE

DATE/TIME OF TEST: 10/09/2021 18:00:44

Diagnostics Check	OK	17:59	
Air Blank		18:00	RFI Detect
Air Blank	0.000	18:00	

Data Download Date/Time: 10/13/2021 09:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BRENNAN,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: O'HARE

DATE/TIME OF TEST: 10/09/2021 18:03:52

Diagnostics Check	OK	18:03	
Air Blank		18:03	RFI Detect
Air Blank	0.000	18:03	

Data Download Date/Time: 10/13/2021 09:14

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BRENNAN,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: O'HARE

DATE/TIME OF TEST: 10/09/2021 18:06:58

Diagnostics Check	OK	18:06	
Air Blank		18:06	RFI Detect
Air Blank		18:06	Improper Sample

Data Download Date/Time: 10/13/2021 09:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BRENNAN,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: O'HARE

DATE/TIME OF TEST: 10/09/2021 18:10:07

Diagnostics Check	OK	18:09	
Air Blank		18:09	RFI Detect
Air Blank	0.000	18:10	

Data Download Date/Time: 10/13/2021 09:15

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRERO,C,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BRENNAN,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: O'HARE

DATE/TIME OF TEST: 10/09/2021 18:24:31

Diagnostics Check		OK	18:18
Air Blank		0.000	18:18
Control Test		0.079	18:19
Air Blank		0.000	18:19
Subject Sample	# 1	0.178	18:20
Breath Volume		4.242	18:20
Air Blank		0.000	18:20
Air Blank		0.000	18:22
Subject Sample	# 2	0.169	18:22
Breath Volume		3.718	18:22
Air Blank		0.000	18:23
Control Test		0.077	18:23
Air Blank		0.000	18:24
Diagnostics Check		OK	18:24

Data Download Date/Time: 10/13/2021 09:15

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: VIECHEC

DATE/TIME OF TEST: 10/09/2021 20:25:28

Diagnostics Check		OK	20:18
Air Blank		0.000	20:19
Control Test		0.080	20:19
Air Blank		0.000	20:20
Subject Sample	# 1	0.096	20:20
Breath Volume		1.539	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.092	20:23
Breath Volume		1.660	20:23
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 10/13/2021 09:15

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: REED,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: COMERIATO

DATE/TIME OF TEST: 10/11/2021 11:53:39

Diagnostics Check		OK	11:47
Air Blank		0.000	11:47
Control Test		0.080	11:48
Air Blank		0.000	11:48
Subject Sample	# 1	0.045	11:49
Breath Volume		2.250	11:49
Air Blank		0.000	11:49
Air Blank		0.000	11:51
Subject Sample	# 2	0.046	11:52
Breath Volume		2.410	11:52
Air Blank		0.000	11:52
Control Test		0.080	11:53
Air Blank		0.000	11:53
Diagnostics Check		OK	11:53

Data Download Date/Time: 10/13/2021 09:15

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASALA,J,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZMORALES,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GAVINA LARA

DATE/TIME OF TEST: 10/11/2021 21:19:05

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.083	21:14
Breath Volume		2.300	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.081	21:17
Breath Volume		2.144	21:17
Air Blank		0.000	21:18
Control Test		0.080	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 10/13/2021 09:15

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,MAXWELL,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JOHNSON,TONY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ALEKSANROV

DATE/TIME OF TEST: 10/12/2021 02:15:35

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:04	
Control Test		0.081	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:08	No Sample Provided
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:10	
Subject Sample	# 2		02:14	No Sample Provided
Breath Volume		0.000	02:14	
Air Blank		0.000	02:14	
Control Test		0.080	02:15	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 10/13/2021 09:15

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: GODINEZ PEREZ

DATE/TIME OF TEST: 10/14/2021 00:35:25

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.185	00:30
Breath Volume		3.023	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.183	00:33
Breath Volume		3.007	00:33
Air Blank		0.000	00:34
Control Test		0.079	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 11/19/2021 11:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: MELECO

DATE/TIME OF TEST: 10/14/2021 20:35:57

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.080	20:29
Air Blank		0.000	20:29
Subject Sample	# 1	0.291	20:30
Breath Volume		1.210	20:30
Air Blank		0.000	20:31
Air Blank		0.000	20:32
Subject Sample	# 2	0.295	20:34
Breath Volume		2.292	20:34
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:35
Diagnostics Check		OK	20:35

Data Download Date/Time: 11/19/2021 11:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,M,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PIERSTORFF-MCLAIN,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/15/2021 03:56:21

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.082	03:51
Breath Volume		2.761	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.075	03:54
Breath Volume		1.773	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 11/19/2021 11:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 10/23/2021 03:29:34

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.223	03:25
Breath Volume		3.894	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.206	03:28
Breath Volume		3.335	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 11/19/2021 11:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARIMON,KARLHEINZ,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BATSON,MITCHEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SILVACASIANO

DATE/TIME OF TEST: 10/24/2021 04:16:49

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.228	04:12
Breath Volume		2.007	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.231	04:15
Breath Volume		2.332	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 11/19/2021 11:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MEJIA PEREZ

DATE/TIME OF TEST: 10/25/2021 03:13:15

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.080	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.191	03:08
Breath Volume		2.054	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.186	03:11
Breath Volume		2.207	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 11/19/2021 11:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BAXTER,JUSTIN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 10/30/2021 22:46:36

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.173	22:42
Breath Volume		2.296	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.156	22:45
Breath Volume		1.121	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 11/19/2021 11:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RAMIREZ PEREZ

DATE/TIME OF TEST: 10/01/2021 02:47:29

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.222	02:42
Breath Volume		2.250	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.235	02:45
Breath Volume		1.343	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/14/2021 09:21

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUONPANE,V
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 10/02/2021 02:08:19

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.166	02:03
Breath Volume		1.976	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.166	02:06
Breath Volume		1.253	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 10/14/2021 09:21

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORE,RICHARD,A
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BIDWELL,P
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/21/2021 12:04:56

Diagnostics Check		OK	11:58
Air Blank		0.000	11:58
Control Test		0.080	11:59
Air Blank		0.000	11:59
Subject Sample	# 1	0.000	12:00
Breath Volume		1.683	12:00
Air Blank		0.000	12:00
Air Blank		0.000	12:02
Subject Sample	# 2	0.000	12:03
Breath Volume		1.136	12:03
Air Blank		0.000	12:04
Control Test		0.080	12:04
Air Blank		0.000	12:04
Diagnostics Check		OK	12:04

Data Download Date/Time: 12/03/2021 12:31

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PRICE,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 10/20/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 08:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 10/24/2021 08:59:48

Diagnostics Check		OK	08:53
Air Blank		0.000	08:53
Control Test		0.080	08:54
Air Blank		0.000	08:54
Subject Sample	# 1	0.150	08:55
Breath Volume		1.355	08:55
Air Blank		0.000	08:55
Air Blank		0.000	08:57
Subject Sample	# 2	0.163	08:58
Breath Volume		1.714	08:58
Air Blank		0.000	08:58
Control Test		0.080	08:59
Air Blank		0.000	08:59
Diagnostics Check		OK	08:59

Data Download Date/Time: 12/03/2021 12:31

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: GAUSE

DATE/TIME OF TEST: 10/01/2021 22:32:05

Diagnostics Check		OK	22:27	
Air Blank		0.000	22:27	
Control Test		0.082	22:28	
Air Blank		0.000	22:28	
Subject Sample	# 1		22:30	Subject Test Refused
Breath Volume		0.000	22:30	
Air Blank		0.000	22:31	
Control Test		0.081	22:31	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 10/07/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEYMOUR,KEVIN,R
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: SEYMOUR,KEVIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 09/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 10/16/2021 01:39:02

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.078	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.176	01:34
Breath Volume		2.687	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.168	01:37
Breath Volume		2.585	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:39

Data Download Date/Time: 10/21/2021 17:14

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEYMOUR,KEVIN,R
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: SEYMOUR,KEVIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 10/25/2021 23:50:01

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.078	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.144	23:45
Breath Volume		1.101	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:47
Subject Sample	# 2	0.149	23:48
Breath Volume		1.339	23:48
Air Blank		0.000	23:49
Control Test		0.076	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:50

Data Download Date/Time: 11/29/2021 10:14

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHARETTE,JONATHAN
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 10/02/2021 19:51:13

Diagnostics Check		OK	19:39	
Air Blank		0.000	19:40	
Control Test		0.079	19:40	
Air Blank		0.000	19:41	
Subject Sample	# 1		19:44	No Sample Provided
Breath Volume		0.000	19:44	
Air Blank		0.000	19:44	
Air Blank		0.000	19:46	
Subject Sample	# 2		19:49	No Sample Provided
Breath Volume		0.000	19:49	
Air Blank		0.000	19:50	
Control Test		0.080	19:50	
Air Blank		0.000	19:51	
Diagnostics Check		OK	19:51	

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRELL,AUSTIN,G
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILE,JESSICA
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 10/07/2021 05:04:27

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.144	05:00
Breath Volume		1.609	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.143	05:02
Breath Volume		2.070	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TZUL CAAL

DATE/TIME OF TEST: 10/10/2021 02:56:27

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.165	02:49
Breath Volume		1.378	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.169	02:54
Breath Volume		1.230	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: STEARNS

DATE/TIME OF TEST: 10/10/2021 04:34:25

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.081	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.116	04:28
Breath Volume		1.441	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.110	04:32
Breath Volume		1.867	04:32
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: MCQUEEN

DATE/TIME OF TEST: 10/14/2021 19:53:24

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:42	
Control Test		0.080	19:42	
Air Blank		0.000	19:43	
Subject Sample	# 1		19:46	No Sample Provided
Breath Volume		0.000	19:46	
Air Blank		0.000	19:46	
Air Blank		0.000	19:48	
Subject Sample	# 2		19:52	No Sample Provided
Breath Volume		0.000	19:52	
Air Blank		0.000	19:52	
Control Test		0.080	19:52	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 10/17/2021 23:44:04

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:33	
Control Test		0.079	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:37	No Sample Provided
Breath Volume		0.000	23:37	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:42	No Sample Provided
Breath Volume		0.000	23:42	
Air Blank		0.000	23:43	
Control Test		0.080	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: REWIS

DATE/TIME OF TEST: 10/17/2021 23:58:37

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.080	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.200	23:54
Breath Volume		1.960	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.204	23:57
Breath Volume		2.308	23:57
Air Blank		0.000	23:57
Control Test		0.080	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MEDEROS

DATE/TIME OF TEST: 10/18/2021 04:13:24

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.193	04:08
Breath Volume		1.937	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.193	04:11
Breath Volume		1.875	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 10/18/2021 09:26

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 10/18/2021 11:45:19

Diagnostics Check		OK	11:38
Air Blank		0.000	11:39
Control Test		0.080	11:39
Air Blank		0.000	11:40
Subject Sample	# 1	0.000	11:40
Breath Volume		1.730	11:40
Air Blank		0.000	11:41
Air Blank		0.000	11:43
Subject Sample	# 2	0.000	11:43
Breath Volume		1.539	11:43
Air Blank		0.000	11:44
Control Test		0.080	11:44
Air Blank		0.000	11:45
Diagnostics Check		OK	11:45

Data Download Date/Time: 11/23/2021 14:59

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BEHRENS,BRENT
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: LYNN

DATE/TIME OF TEST: 10/21/2021 21:30:27

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:24
Subject Sample	# 1	0.094	21:25
Breath Volume		1.171	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.094	21:28
Breath Volume		1.308	21:28
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:30
Diagnostics Check		OK	21:30

Data Download Date/Time: 11/23/2021 14:59

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRELL,AUSTIN,G
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 10/22/2021 02:13:00

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.000	02:08
Breath Volume		1.292	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.000	02:11
Breath Volume		1.261	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:13

Data Download Date/Time: 11/23/2021 14:59

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/22/2021 03:04:47

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:53	
Control Test		0.080	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:57	No Sample Provided
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Air Blank		0.000	02:59	
Subject Sample	# 2	0.174	03:00	
Breath Volume		2.085	03:00	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 3	0.174	03:03	
Breath Volume		1.648	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 11/23/2021 14:59

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,JOLNETH,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOLNETH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: OTERO-HERNANDEZ

DATE/TIME OF TEST: 10/23/2021 03:00:44

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.103	02:56
Breath Volume		1.910	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.104	02:59
Breath Volume		1.847	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 11/23/2021 14:59

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 10/23/2021 04:13:32

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.125	04:08
Breath Volume		2.355	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.125	04:12
Breath Volume		1.273	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 11/23/2021 14:59

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,SHERLY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAFONTANT,SEDONA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MCCOLLEY

DATE/TIME OF TEST: 10/23/2021 05:57:04

Diagnostics Check		OK	05:54	
Air Blank		0.000	05:54	
Control Test		0.079	05:54	
Air Blank		0.000	05:55	
Subject Sample	# 1		05:55	Subject Test Refused
Breath Volume		0.000	05:55	
Air Blank		0.000	05:56	
Control Test		0.080	05:56	
Air Blank		0.000	05:56	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTALBANO,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MONTALBANO,JONATHAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:59

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 10/23/2021 08:31:38

Diagnostics Check		OK	08:20	
Air Blank		0.000	08:20	
Control Test		0.081	08:20	
Air Blank		0.000	08:21	
Subject Sample	# 1	0.158	08:24	Volume Not Met
Breath Volume		0.558	08:24	
Air Blank		0.000	08:25	
Air Blank		0.000	08:26	
Subject Sample	# 2	0.173	08:30	Volume Not Met
Breath Volume		0.066	08:30	
Air Blank		0.000	08:30	
Control Test		0.080	08:31	
Air Blank		0.000	08:31	
Diagnostics Check		OK	08:31	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 10/23/2021 20:46:22

Diagnostics Check		OK	20:35	
Air Blank		0.000	20:35	
Control Test		0.081	20:35	
Air Blank		0.000	20:36	
Subject Sample	# 1		20:39	No Sample Provided
Breath Volume		0.000	20:39	
Air Blank		0.000	20:39	
Air Blank		0.000	20:41	
Subject Sample	# 2		20:45	No Sample Provided
Breath Volume		0.000	20:45	
Air Blank		0.000	20:45	
Control Test		0.081	20:45	
Air Blank		0.000	20:46	
Diagnostics Check		OK	20:46	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 10/24/2021 02:40:27

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.081	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.102	02:35
Breath Volume		2.937	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.095	02:38
Breath Volume		2.539	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: STANFIELD

DATE/TIME OF TEST: 10/24/2021 02:58:14

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:50	No Sample Provided
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.065	02:53	
Breath Volume		2.742	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 3	0.061	02:56	
Breath Volume		1.566	02:56	
Air Blank		0.000	02:57	
Control Test		0.079	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/24/2021 03:58:04

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.080	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:51	No Sample Provided
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:57	
Control Test		0.081	03:57	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:58	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 10/24/2021 06:00:29

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.081	05:54
Air Blank		0.000	05:55
Subject Sample	# 1	0.142	05:55
Breath Volume		2.187	05:55
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 2	0.141	05:58
Breath Volume		1.929	05:58
Air Blank		0.000	05:59
Control Test		0.080	05:59
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 10/24/2021 20:08:14

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.080	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.227	20:03
Breath Volume		1.660	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.211	20:06
Breath Volume		2.089	20:06
Air Blank		0.000	20:07
Control Test		0.079	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ZELAYA LAINEZ

DATE/TIME OF TEST: 10/26/2021 03:11:11

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.077	03:06
Breath Volume		2.062	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:08
Subject Sample	# 2	0.076	03:09
Breath Volume		1.304	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELCHER,ALEX,B
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 10/27/2021 04:46:54

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:35	
Control Test		0.081	04:36	
Air Blank		0.000	04:36	
Subject Sample	# 1		04:39	No Sample Provided
Breath Volume		0.000	04:39	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 2		04:45	No Sample Provided
Breath Volume		0.000	04:45	
Air Blank		0.000	04:46	
Control Test		0.081	04:46	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRON,DAVID,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SITLNICK,SARAH
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:06

SUBJECT'S LAST NAME: BECKERMAN

DATE/TIME OF TEST: 10/27/2021 14:40:48

Diagnostics Check		OK	14:31
Air Blank		0.000	14:32
Control Test		0.080	14:32
Air Blank		0.000	14:33
Subject Sample	# 1	0.144	14:36
Breath Volume		1.742	14:36
Air Blank		0.000	14:37
Air Blank		0.000	14:38
Subject Sample	# 2	0.141	14:39
Breath Volume		1.195	14:39
Air Blank		0.000	14:39
Control Test		0.078	14:40
Air Blank		0.000	14:40
Diagnostics Check		OK	14:40

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GASKIN

DATE/TIME OF TEST: 10/28/2021 02:12:12

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:01	
Control Test		0.080	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:05	No Sample Provided
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 2		02:10	No Sample Provided
Breath Volume		0.000	02:10	
Air Blank		0.000	02:11	
Control Test		0.080	02:11	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: STINSON

DATE/TIME OF TEST: 10/28/2021 02:23:23

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.200	02:18
Breath Volume		1.996	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.198	02:21
Breath Volume		1.574	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,RICKY,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BINET,RICKY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 10/29/2021 03:37:57

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:35	
Control Test		0.081	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,BRIANA,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLETCHER,BRIANA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: KOSIER

DATE/TIME OF TEST: 10/29/2021 04:33:02

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.149	04:28
Breath Volume		2.277	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.146	04:31
Breath Volume		2.171	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:33

Data Download Date/Time: 11/23/2021 15:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/29/2021 20:05:46

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.081	19:59
Air Blank		0.000	19:59
Subject Sample	# 1	0.117	20:01
Breath Volume		1.113	20:01
Air Blank		0.000	20:02
Air Blank		0.000	20:03
Subject Sample	# 2	0.122	20:04
Breath Volume		1.507	20:04
Air Blank		0.000	20:04
Control Test		0.079	20:05
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: EMBRY

DATE/TIME OF TEST: 10/29/2021 22:45:10

Diagnostics Check		OK	22:36
Air Blank		0.000	22:36
Control Test		0.079	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.154	22:40
Breath Volume		1.757	22:40
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.161	22:43
Breath Volume		1.691	22:43
Air Blank		0.000	22:44
Control Test		0.078	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGAN,LAKEBRA,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/30/2021 02:52:22

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.080	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:45	No Sample Provided
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 2		02:51	No Sample Provided
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Control Test		0.081	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: JURADO

DATE/TIME OF TEST: 10/30/2021 03:38:04

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.127	03:33
Breath Volume		1.449	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:35
Subject Sample	# 2	0.135	03:36
Breath Volume		2.250	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:38

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 10/30/2021 03:59:14

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.000	03:54
Breath Volume		1.617	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.000	03:57
Breath Volume		1.578	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HODGES,MARCO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 10/30/2021 23:08:58

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.129	23:04
Breath Volume		1.714	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.128	23:07
Breath Volume		2.582	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ALBERT,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/31/2021 01:12:14

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2		01:10	No Sample Provided
Breath Volume		0.000	01:10	
Air Blank		0.000	01:11	
Control Test		0.080	01:11	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 10/31/2021 01:24:22

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.183	01:19
Breath Volume		1.535	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.175	01:22
Breath Volume		1.800	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTS,JAMES,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RYAN,DEBORAH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: OTHER
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: BUURMA

DATE/TIME OF TEST: 10/31/2021 07:40:46

Diagnostics Check		OK	07:33
Air Blank		0.000	07:34
Control Test		0.081	07:34
Air Blank		0.000	07:35
Subject Sample	# 1	0.108	07:36
Breath Volume		1.656	07:36
Air Blank		0.000	07:36
Air Blank		0.000	07:38
Subject Sample	# 2	0.111	07:39
Breath Volume		1.707	07:39
Air Blank		0.000	07:39
Control Test		0.080	07:40
Air Blank		0.000	07:40
Diagnostics Check		OK	07:40

Data Download Date/Time: 11/23/2021 15:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RICKETTS,J
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/17/2021 01:57:11

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.081	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 10/27/2021 11:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181622
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 10/22/2021 03:28:36

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.233	03:23
Breath Volume		2.101	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.229	03:27
Breath Volume		2.078	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 10/27/2021 11:01

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: MENDOZA,JONATHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 10/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 10/08/2021 19:52:27

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.080	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.103	19:47
Breath Volume		2.097	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.099	19:50
Breath Volume		2.027	19:50
Air Blank		0.000	19:51
Control Test		0.078	19:51
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 11/16/2021 00:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: MARTIN,BAILEY
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 10/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 10/13/2021 01:57:18

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.203	01:52
Breath Volume		1.785	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:54
Subject Sample	# 2	0.215	01:55
Breath Volume		1.628	01:55
Air Blank		0.000	01:56
Control Test		0.077	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 11/16/2021 00:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 10/03/2021 22:22:54

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.194	22:18
Breath Volume		2.507	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:20
Subject Sample	# 2	0.192	22:21
Breath Volume		2.066	22:21
Air Blank		0.000	22:21
Control Test		0.081	22:22
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 10/04/2021 11:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MANDESE,ANGELO
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 10/16/2021 03:51:18

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.083	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.000	03:46
Breath Volume		3.027	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.000	03:49
Breath Volume		2.578	03:49
Air Blank		0.000	03:50
Control Test		0.083	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 11/03/2021 12:27

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMBSH,CARLYTON,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: HAMBSH,CARLYTON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 10/19/2021 17:11:43

Diagnostics Check		OK	17:05
Air Blank		0.000	17:05
Control Test		0.082	17:05
Air Blank		0.000	17:06
Subject Sample	# 1	0.038	17:06
Breath Volume		1.289	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:09
Subject Sample	# 2	0.037	17:10
Breath Volume		1.578	17:10
Air Blank		0.000	17:10
Control Test		0.082	17:11
Air Blank		0.000	17:11
Diagnostics Check		OK	17:11

Data Download Date/Time: 11/03/2021 12:27

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MEANEY,TREVOR
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: NASTI

DATE/TIME OF TEST: 10/22/2021 00:48:51

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.082	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.107	00:43
Breath Volume		1.328	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.108	00:47
Breath Volume		1.261	00:47
Air Blank		0.000	00:47
Control Test		0.081	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 11/03/2021 12:27

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: VELEZ,EDWARD
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/23/2021 22:22:21

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.082	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.251	22:17
Breath Volume		1.929	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.258	22:20
Breath Volume		1.722	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 11/03/2021 12:27

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,NICOLE,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: ALEVISATOS,JARRAD
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 10/26/2021 03:16:30

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.083	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.137	03:07
Breath Volume		1.339	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.164	03:12
Breath Volume		1.742	03:12
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 3	0.140	03:14
Breath Volume		1.406	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 11/03/2021 12:27

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,RYAN,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: CHIAPPA,JOSEPH
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 10/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: LEBRON

DATE/TIME OF TEST: 10/21/2021 03:45:33

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.000	03:40
Breath Volume		1.484	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.000	03:44
Breath Volume		1.132	03:44
Air Blank		0.000	03:44
Control Test		0.083	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 11/13/2021 19:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 10/05/2021 01:06:59

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:02
Breath Volume		3.968	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.000	01:05
Breath Volume		3.937	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 10/28/2021 09:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MCNALLY

DATE/TIME OF TEST: 10/06/2021 00:36:07

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.080	00:25	
Air Blank		0.000	00:26	
Subject Sample	# 1		00:29	No Sample Provided
Breath Volume		0.000	00:29	
Air Blank		0.000	00:29	
Air Blank		0.000	00:31	
Subject Sample	# 2		00:34	No Sample Provided
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Control Test		0.080	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 10/28/2021 09:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BILLER

DATE/TIME OF TEST: 10/01/2021 00:48:30

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.081	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.139	00:44
Breath Volume		1.957	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.147	00:46
Breath Volume		1.777	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 10/14/2021 09:25

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RIVERATORRES

DATE/TIME OF TEST: 10/01/2021 03:49:08

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.081	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.136	03:44
Breath Volume		3.476	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.133	03:47
Breath Volume		3.519	03:47
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 10/14/2021 09:25

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARBONE,ANDRES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: TIDWELL

DATE/TIME OF TEST: 10/01/2021 04:36:20

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.171	04:31
Breath Volume		2.050	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.163	04:34
Breath Volume		2.160	04:34
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 10/14/2021 09:25

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BIVENS

DATE/TIME OF TEST: 10/06/2021 03:45:24

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.000	03:41
Breath Volume		3.242	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.000	03:44
Breath Volume		2.539	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 10/14/2021 09:25

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 10/15/2021 01:47:28

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.102	01:42
Breath Volume		1.558	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.091	01:45
Breath Volume		1.175	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUTCHINSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/15/2021 04:16:16

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.081	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.225	04:08
Breath Volume		4.808	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.163	04:11
Breath Volume		1.429	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 3	0.215	04:14
Breath Volume		5.136	04:14
Air Blank		0.000	04:15
Control Test		0.078	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: YATGAAL

DATE/TIME OF TEST: 10/16/2021 01:37:33

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.081	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.155	01:33
Breath Volume		1.593	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.147	01:35
Breath Volume		1.675	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: COLON SILVA

DATE/TIME OF TEST: 10/16/2021 20:53:19

Diagnostics Check		OK	20:46
Air Blank		0.000	20:47
Control Test		0.081	20:47
Air Blank		0.000	20:48
Subject Sample	# 1	0.139	20:48
Breath Volume		1.371	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.135	20:51
Breath Volume		2.109	20:51
Air Blank		0.000	20:52
Control Test		0.081	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATKINS,ROMAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/17/2021 07:36:51

Diagnostics Check		OK	07:27	
Air Blank		0.000	07:28	
Control Test		0.081	07:28	
Air Blank		0.000	07:28	
Subject Sample	# 1		07:32	No Sample Provided
Breath Volume		0.000	07:32	
Air Blank		0.000	07:32	
Air Blank		0.000	07:34	
Subject Sample	# 2		07:35	Subject Test Refused
Breath Volume		0.000	07:35	
Air Blank		0.000	07:35	
Control Test		0.081	07:36	
Air Blank		0.000	07:36	
Diagnostics Check		OK	07:36	

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MITCHELL,CEDRIC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: RODRIGUEZSUAZO

DATE/TIME OF TEST: 10/17/2021 08:18:46

Diagnostics Check		OK	08:12
Air Blank		0.000	08:12
Control Test		0.081	08:12
Air Blank		0.000	08:13
Subject Sample	# 1	0.109	08:14
Breath Volume		1.867	08:14
Air Blank		0.000	08:14
Air Blank		0.000	08:16
Subject Sample	# 2	0.108	08:17
Breath Volume		2.335	08:17
Air Blank		0.000	08:17
Control Test		0.080	08:18
Air Blank		0.000	08:18
Diagnostics Check		OK	08:18

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 10/18/2021 20:32:11

Diagnostics Check		OK	20:26
Air Blank		0.000	20:26
Control Test		0.082	20:26
Air Blank		0.000	20:27
Subject Sample	# 1	0.000	20:27
Breath Volume		3.871	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.000	20:30
Breath Volume		3.093	20:30
Air Blank		0.000	20:31
Control Test		0.079	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KILLON,JONATHAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 10/20/2021 01:35:30

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.081	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.000	01:30
Breath Volume		2.085	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.000	01:34
Breath Volume		2.375	01:34
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONE,MAURICE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CODE

DATE/TIME OF TEST: 10/21/2021 23:42:42

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.080	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1	0.174	23:35	Volume Not Met
Breath Volume		0.019	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 2	0.158	23:40	Volume Not Met
Breath Volume		0.218	23:40	
Air Blank		0.000	23:41	
Control Test		0.077	23:42	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 11/03/2021 10:33

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ECHEVERRIA VIDAL,XAVIER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 10/22/2021 04:52:14

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.107	04:47
Breath Volume		2.183	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:49
Subject Sample	# 2	0.105	04:50
Breath Volume		2.136	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 10/22/2021 05:08:41

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.081	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.169	05:04
Breath Volume		2.496	05:04
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.159	05:07
Breath Volume		2.925	05:07
Air Blank		0.000	05:07
Control Test		0.079	05:08
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 10/22/2021 05:27:16

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.080	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.190	05:22
Breath Volume		4.296	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.170	05:25
Breath Volume		4.355	05:25
Air Blank		0.000	05:26
Control Test		0.077	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAMB,ASHLEY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: VELASQUEZ RAMOS

DATE/TIME OF TEST: 10/23/2021 21:02:44

Diagnostics Check		OK	20:56
Air Blank		0.000	20:57
Control Test		0.081	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.222	20:58
Breath Volume		1.597	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.235	21:01
Breath Volume		2.875	21:01
Air Blank		0.000	21:01
Control Test		0.080	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRYANT,ALFONZIA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: WIELAND

DATE/TIME OF TEST: 10/26/2021 01:37:45

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.081	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.097	01:33
Breath Volume		3.214	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.096	01:36
Breath Volume		2.632	01:36
Air Blank		0.000	01:36
Control Test		0.080	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: DOOBRAJ

DATE/TIME OF TEST: 10/27/2021 03:03:54

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.303	02:56
Breath Volume		1.523	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.312	03:01
Breath Volume		2.101	03:01
Air Blank		0.000	03:02
Control Test		0.076	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRELAND,DAVID
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ADAMS TORRES

DATE/TIME OF TEST: 10/27/2021 03:40:31

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.077	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.155	03:35
Breath Volume		3.218	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.151	03:38
Breath Volume		2.835	03:38
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALVAREZ,PEDRO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 10/28/2021 00:04:12

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.081	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:03	
Control Test		0.080	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 10/28/2021 04:06:00

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.084	04:01
Breath Volume		1.535	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.083	04:04
Breath Volume		2.851	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:06

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWNS,NATALIE
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 10/28/2021 20:55:43

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:44	
Control Test		0.080	20:45	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:48	No Sample Provided
Breath Volume		0.000	20:48	
Air Blank		0.000	20:49	
Air Blank		0.000	20:51	
Subject Sample	# 2		20:54	No Sample Provided
Breath Volume		0.000	20:54	
Air Blank		0.000	20:54	
Control Test		0.082	20:55	
Air Blank		0.000	20:55	
Diagnostics Check		OK	20:55	

Data Download Date/Time: 11/03/2021 10:34

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZ,RICHARD
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SANTOSCASTRO

DATE/TIME OF TEST: 10/14/2021 01:41:59

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.082	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.139	01:37
Breath Volume		1.410	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.140	01:40
Breath Volume		1.492	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 10/14/2021 08:42

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,JEFFREY
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 10/20/2021 01:21:57

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.082	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.154	01:17
Breath Volume		2.117	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:19
Subject Sample	# 2	0.158	01:20
Breath Volume		1.351	01:20
Air Blank		0.000	01:21
Control Test		0.082	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHAPLIN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 10/20/2021 22:32:41

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.082	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.238	22:28
Breath Volume		1.500	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.245	22:31
Breath Volume		1.359	22:31
Air Blank		0.000	22:31
Control Test		0.081	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: PHILLIP

DATE/TIME OF TEST: 10/22/2021 01:08:00

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.083	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.169	01:03
Breath Volume		2.882	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.166	01:06
Breath Volume		2.917	01:06
Air Blank		0.000	01:07
Control Test		0.082	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:08

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BAHENAENCISO

DATE/TIME OF TEST: 10/23/2021 00:02:29

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.082	23:54
Air Blank		0.000	23:54
Subject Sample	# 2	0.130	23:57
Breath Volume		1.375	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.122	00:00
Breath Volume		1.664	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KING,ARTHUR
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: CARVALHAES

DATE/TIME OF TEST: 10/23/2021 02:08:45

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.082	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.153	02:04
Breath Volume		1.316	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.151	02:07
Breath Volume		2.296	02:07
Air Blank		0.000	02:07
Control Test		0.082	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWNS,NATALIE
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 10/28/2021 20:57:57

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:51
Subject Sample	# 1	0.390	20:52
Breath Volume		1.371	20:52
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.396	20:56
Breath Volume		1.195	20:56
Air Blank		0.000	20:56
Control Test		0.080	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FAGUNDES,FELIPE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MORENOJAIMES

DATE/TIME OF TEST: 10/29/2021 01:49:18

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.228	01:44
Breath Volume		3.605	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.232	01:47
Breath Volume		3.277	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUCIANO MONTES,NICOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: GUZMANHEREDIA

DATE/TIME OF TEST: 10/29/2021 04:58:53

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.082	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.222	04:51
Breath Volume		1.535	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.246	04:54
Breath Volume		2.730	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 3	0.238	04:57
Breath Volume		2.214	04:57
Air Blank		0.000	04:57
Control Test		0.081	04:58
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD,MIGUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: JANASIAK

DATE/TIME OF TEST: 10/29/2021 05:11:50

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.081	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.138	05:07
Breath Volume		1.500	05:07
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.145	05:10
Breath Volume		1.437	05:10
Air Blank		0.000	05:10
Control Test		0.081	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DISLA,EMILIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: ABRIOLA

DATE/TIME OF TEST: 10/29/2021 07:46:22

Diagnostics Check		OK	07:34	
Air Blank		0.000	07:35	
Control Test		0.082	07:35	
Air Blank		0.000	07:36	
Subject Sample	# 1	0.000	07:39	Volume Not Met
Breath Volume		0.359	07:39	
Air Blank		0.000	07:39	
Air Blank		0.000	07:41	
Subject Sample	# 2	0.000	07:44	Volume Not Met
Breath Volume		0.097	07:44	
Air Blank		0.000	07:45	
Control Test		0.082	07:45	
Air Blank		0.000	07:46	
Diagnostics Check		OK	07:46	

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BROWN,DARRELL
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: KEPPEN

DATE/TIME OF TEST: 10/30/2021 02:58:30

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:47	
Control Test		0.083	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:51	No Sample Provided
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Air Blank		0.000	02:54	
Subject Sample	# 2		02:57	No Sample Provided
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.081	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MILLNER

DATE/TIME OF TEST: 10/30/2021 04:13:03

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.082	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.172	04:08
Breath Volume		1.226	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.176	04:11
Breath Volume		1.429	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:13

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 10/30/2021 09:47:09

Diagnostics Check		OK	09:40
Air Blank		0.000	09:41
Control Test		0.082	09:41
Air Blank		0.000	09:42
Subject Sample	# 1	0.161	09:42
Breath Volume		2.785	09:42
Air Blank		0.000	09:43
Air Blank		0.000	09:45
Subject Sample	# 2	0.163	09:45
Breath Volume		2.515	09:45
Air Blank		0.000	09:46
Control Test		0.082	09:46
Air Blank		0.000	09:47
Diagnostics Check		OK	09:47

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 10/31/2021 01:27:42

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.082	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.120	01:23
Breath Volume		2.625	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.122	01:26
Breath Volume		3.164	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WALLACE,TIMOTHY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: BRACY

DATE/TIME OF TEST: 10/31/2021 20:12:09

Diagnostics Check		OK	20:05
Air Blank		0.000	20:06
Control Test		0.083	20:06
Air Blank		0.000	20:07
Subject Sample	# 1	0.186	20:07
Breath Volume		1.953	20:07
Air Blank		0.000	20:08
Air Blank		0.000	20:10
Subject Sample	# 2	0.192	20:10
Breath Volume		3.265	20:10
Air Blank		0.000	20:11
Control Test		0.082	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 11/03/2021 13:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: KINDLE

DATE/TIME OF TEST: 10/02/2021 03:56:33

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.187	03:51
Breath Volume		2.640	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.195	03:55
Breath Volume		3.070	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 10/14/2021 09:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CELLI

DATE/TIME OF TEST: 10/02/2021 05:02:32

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.080	04:51	
Air Blank		0.000	04:52	
Subject Sample	# 1		04:55	No Sample Provided
Breath Volume		0.000	04:55	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 2		05:01	No Sample Provided
Breath Volume		0.000	05:01	
Air Blank		0.000	05:01	
Control Test		0.080	05:01	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 10/14/2021 09:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANCEL,KAREN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SWANIGAN

DATE/TIME OF TEST: 10/02/2021 05:18:16

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.080	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.000	05:12
Breath Volume		1.722	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.000	05:16
Breath Volume		2.515	05:16
Air Blank		0.000	05:17
Control Test		0.081	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 10/14/2021 09:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PAGAN,KEVIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: CALVILLOLOPEZ

DATE/TIME OF TEST: 10/02/2021 14:13:12

Diagnostics Check		OK	14:07
Air Blank		0.000	14:07
Control Test		0.080	14:07
Air Blank		0.000	14:08
Subject Sample	# 1	0.216	14:08
Breath Volume		1.308	14:08
Air Blank		0.000	14:09
Air Blank		0.000	14:11
Subject Sample	# 2	0.222	14:11
Breath Volume		1.718	14:11
Air Blank		0.000	14:12
Control Test		0.079	14:12
Air Blank		0.000	14:13
Diagnostics Check		OK	14:13

Data Download Date/Time: 10/14/2021 09:04

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HERSHOCK

DATE/TIME OF TEST: 10/03/2021 01:14:08

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.205	01:09
Breath Volume		2.812	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.207	01:12
Breath Volume		3.687	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCFERREN,JASON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: DUCKWORTH

DATE/TIME OF TEST: 10/03/2021 02:37:36

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.098	02:32
Breath Volume		1.171	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.106	02:36
Breath Volume		2.707	02:36
Air Blank		0.000	02:36
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: JIMENEZMIRANDA

DATE/TIME OF TEST: 10/06/2021 23:50:30

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.184	23:45
Breath Volume		1.511	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.187	23:48
Breath Volume		2.074	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUBIC,JUDE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 10/07/2021 03:24:05

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.080	03:13	
Air Blank		0.000	03:13	
Subject Sample	# 1	0.214	03:16	Volume Not Met
Breath Volume		0.132	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 2	0.172	03:22	Volume Not Met
Breath Volume		0.437	03:22	
Air Blank		0.000	03:23	
Control Test		0.078	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ADLER,ERIK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SHETTY

DATE/TIME OF TEST: 10/07/2021 23:35:29

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.241	23:31
Breath Volume		2.875	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.243	23:33
Breath Volume		2.968	23:33
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NDOUMBE,FREDERIC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: ALBORNOZQUINTANA

DATE/TIME OF TEST: 10/08/2021 10:18:08

Diagnostics Check		OK	10:12
Air Blank		0.000	10:12
Control Test		0.080	10:12
Air Blank		0.000	10:13
Subject Sample	# 1	0.146	10:13
Breath Volume		2.812	10:13
Air Blank		0.000	10:14
Air Blank		0.000	10:16
Subject Sample	# 2	0.150	10:16
Breath Volume		3.097	10:16
Air Blank		0.000	10:17
Control Test		0.079	10:17
Air Blank		0.000	10:18
Diagnostics Check		OK	10:18

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAUNDERS,HUNTER
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BUETTNER JR

DATE/TIME OF TEST: 10/09/2021 02:56:02

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.097	02:51
Breath Volume		3.132	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.096	02:54
Breath Volume		3.101	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:56

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GOHMAN

DATE/TIME OF TEST: 10/09/2021 03:18:58

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.082	03:14
Breath Volume		1.882	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.082	03:17
Breath Volume		1.941	03:17
Air Blank		0.000	03:18
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: DEVINE

DATE/TIME OF TEST: 10/09/2021 03:49:07

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.170	03:44
Breath Volume		1.570	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:46
Subject Sample	# 2	0.160	03:47
Breath Volume		2.351	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LOPEZ TORREGROSA

DATE/TIME OF TEST: 10/09/2021 04:25:44

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.069	04:21
Breath Volume		2.335	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.070	04:24
Breath Volume		1.722	04:24
Air Blank		0.000	04:24
Control Test		0.079	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HENRIQUEZ,ENRIQUEZ
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 10/09/2021 07:29:01

Diagnostics Check		OK	07:22
Air Blank		0.000	07:22
Control Test		0.080	07:22
Air Blank		0.000	07:23
Subject Sample	# 1	0.182	07:24
Breath Volume		1.390	07:24
Air Blank		0.000	07:25
Air Blank		0.000	07:26
Subject Sample	# 2	0.187	07:27
Breath Volume		1.878	07:27
Air Blank		0.000	07:28
Control Test		0.078	07:28
Air Blank		0.000	07:28
Diagnostics Check		OK	07:29

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORDON,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: NOLASCO

DATE/TIME OF TEST: 10/09/2021 10:06:15

Diagnostics Check		OK	10:00
Air Blank		0.000	10:00
Control Test		0.080	10:00
Air Blank		0.000	10:01
Subject Sample	# 1	0.197	10:01
Breath Volume		2.839	10:01
Air Blank		0.000	10:02
Air Blank		0.000	10:04
Subject Sample	# 2	0.191	10:04
Breath Volume		2.640	10:04
Air Blank		0.000	10:05
Control Test		0.079	10:05
Air Blank		0.000	10:06
Diagnostics Check		OK	10:06

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FLEURIMOND,JEANMARC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: NANTON

DATE/TIME OF TEST: 10/10/2021 02:24:50

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.121	02:20
Breath Volume		1.292	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.124	02:23
Breath Volume		1.464	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 10/10/2021 23:26:11

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.243	23:21
Breath Volume		3.453	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.243	23:24
Breath Volume		3.242	23:24
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ADLER,ERIK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/13/2021 03:06:47

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.083	03:02
Breath Volume		3.078	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.082	03:05
Breath Volume		2.902	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCFERREN,JASON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/13/2021 04:16:31

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.130	04:12
Breath Volume		3.039	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.119	04:14
Breath Volume		1.710	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 10/14/2021 09:05

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: NOLAND,P
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 10/02/2021 22:16:27

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:14	
Control Test		0.080	22:14	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:15	Subject Test Refused
Breath Volume		0.000	22:15	
Air Blank		0.000	22:15	
Control Test		0.079	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 10/13/2021 18:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/09/2021 00:57:45

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.193	00:52
Breath Volume		1.222	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:54
Subject Sample	# 2	0.187	00:56
Breath Volume		1.636	00:56
Air Blank		0.000	00:56
Control Test		0.076	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 10/13/2021 18:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: VILLACRES,NICHOLAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CARRINGTON

DATE/TIME OF TEST: 10/09/2021 03:25:07

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.080	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.146	03:20
Breath Volume		1.371	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.144	03:23
Breath Volume		1.546	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 10/13/2021 18:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICHECO,JOHN,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: PROVENCHER,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: NESBITT

DATE/TIME OF TEST: 10/10/2021 00:58:07

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.075	00:53
Breath Volume		4.406	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.071	00:56
Breath Volume		2.500	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 10/13/2021 18:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: COLLINS,J
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 10/14/2021 19:49:34

Diagnostics Check		OK	19:46	
Air Blank		0.000	19:47	
Control Test		0.079	19:47	
Air Blank		0.000	19:48	
Subject Sample	# 1		19:48	Subject Test Refused
Breath Volume		0.000	19:48	
Air Blank		0.000	19:48	
Control Test		0.080	19:48	
Air Blank		0.000	19:49	
Diagnostics Check		OK	19:49	

Data Download Date/Time: 10/22/2021 20:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICHECO,JOHN,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HARTZELL,EDWARD
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: KNOBEN

DATE/TIME OF TEST: 10/17/2021 02:13:03

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.078	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.240	02:08
Breath Volume		2.628	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:10
Subject Sample	# 2	0.243	02:11
Breath Volume		4.136	02:11
Air Blank		0.000	02:12
Control Test		0.078	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:13

Data Download Date/Time: 10/22/2021 20:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICHECO,JOHN,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: CHRISTOFFEL,JAMES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BARNHART

DATE/TIME OF TEST: 10/17/2021 04:49:13

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.119	04:44
Breath Volume		1.296	04:44
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.133	04:47
Breath Volume		1.515	04:47
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 10/22/2021 20:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: NOLAND,P
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 10/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: AMES

DATE/TIME OF TEST: 10/20/2021 20:37:18

Diagnostics Check		OK	20:31
Air Blank		0.000	20:31
Control Test		0.080	20:31
Air Blank		0.000	20:32
Subject Sample	# 1	0.113	20:32
Breath Volume		1.664	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:35
Subject Sample	# 2	0.116	20:35
Breath Volume		2.472	20:35
Air Blank		0.000	20:36
Control Test		0.079	20:36
Air Blank		0.000	20:37
Diagnostics Check		OK	20:37

Data Download Date/Time: 10/22/2021 20:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: HOLBROOK,NICHOLAS
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: KUITKOWSKI

DATE/TIME OF TEST: 10/02/2021 22:46:43

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.077	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.169	22:42
Breath Volume		1.800	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:44
Subject Sample	# 2	0.155	22:45
Breath Volume		1.687	22:45
Air Blank		0.000	22:45
Control Test		0.078	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 10/06/2021 14:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BAUERNFEIND,BRANDON
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 10/13/2021 03:43:53

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.078	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Control Test		0.080	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 10/22/2021 11:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JAMES,P
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SCHALLON,CHRISTIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 10/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LEHMANN

DATE/TIME OF TEST: 10/25/2021 01:34:46

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.268	01:30
Breath Volume		3.468	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.268	01:33
Breath Volume		2.574	01:33
Air Blank		0.000	01:33
Control Test		0.078	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 10/27/2021 09:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,OLIVIA,C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DOLCI,P
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ARRIAGA-STEVENS

DATE/TIME OF TEST: 10/03/2021 06:39:50

Diagnostics Check		OK	06:32
Air Blank		0.000	06:33
Control Test		0.082	06:33
Air Blank		0.000	06:33
Subject Sample	# 1	0.000	06:35
Breath Volume		1.582	06:35
Air Blank		0.000	06:36
Air Blank		0.000	06:38
Subject Sample	# 2	0.000	06:38
Breath Volume		1.964	06:38
Air Blank		0.000	06:38
Control Test		0.081	06:39
Air Blank		0.000	06:39
Diagnostics Check		OK	06:39

Data Download Date/Time: 10/11/2021 14:39

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,FELICIA,N
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: GOULD,WILLIAM
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: MILONE

DATE/TIME OF TEST: 10/03/2021 17:52:16

Diagnostics Check		OK	17:44
Air Blank		0.000	17:45
Control Test		0.081	17:45
Air Blank		0.000	17:45
Subject Sample	# 1	0.253	17:46
Breath Volume		1.371	17:46
Air Blank		0.000	17:47
Air Blank		0.000	17:48
Subject Sample	# 2	0.237	17:50
Breath Volume		1.593	17:50
Air Blank		0.000	17:51
Control Test		0.079	17:51
Air Blank		0.000	17:52
Diagnostics Check		OK	17:52

Data Download Date/Time: 10/11/2021 14:39

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZEWICZ,ALISON,E
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAZEWICZ,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MELLEN

DATE/TIME OF TEST: 10/13/2021 22:32:26

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.000	22:26
Breath Volume		1.234	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.000	22:31
Breath Volume		1.109	22:31
Air Blank		0.000	22:31
Control Test		0.081	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 10/25/2021 13:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZEWICZ,ALISON,E
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LOPEZ,AMANDA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/15/2021 00:04:21

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.082	23:57
Air Blank		0.000	23:57
Subject Sample	# 1	0.000	00:00
Breath Volume		1.730	00:00
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 2	0.000	00:02
Breath Volume		1.621	00:02
Air Blank		0.000	00:03
Control Test		0.082	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 10/25/2021 15:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BADIKYAN,DAVID
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/18/2021 16:37:58

Diagnostics Check		OK	16:31
Air Blank		0.000	16:32
Control Test		0.081	16:32
Air Blank		0.000	16:33
Subject Sample	# 1	0.000	16:33
Breath Volume		2.371	16:33
Air Blank		0.000	16:34
Air Blank		0.000	16:36
Subject Sample	# 2	0.000	16:36
Breath Volume		3.925	16:36
Air Blank		0.000	16:37
Control Test		0.082	16:37
Air Blank		0.000	16:37
Diagnostics Check		OK	16:37

Data Download Date/Time: 10/25/2021 15:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,A,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MOFFATT,D
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 10/22/2021 04:23:35

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.000	04:19
Breath Volume		2.500	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.000	04:22
Breath Volume		2.125	04:22
Air Blank		0.000	04:22
Control Test		0.082	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 10/25/2021 15:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,A,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LOPEZ,A
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 10/23/2021 02:13:53

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.176	02:05
Breath Volume		1.320	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.143	02:08
Breath Volume		2.078	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 3	0.149	02:12
Breath Volume		1.792	02:12
Air Blank		0.000	02:12
Control Test		0.080	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 10/25/2021 15:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,A,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LOPEZ,A
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: KORNOW

DATE/TIME OF TEST: 10/23/2021 22:22:25

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.081	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.155	22:17
Breath Volume		1.992	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.153	22:20
Breath Volume		2.031	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 10/25/2021 15:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,A,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: GUIDONE,P
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/24/2021 02:57:09

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.082	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.000	02:52
Breath Volume		1.308	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.000	02:55
Breath Volume		1.390	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 10/25/2021 15:03

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,A,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: GUIDONE,P
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/26/2021 18:58:52

Diagnostics Check		OK	18:52
Air Blank		0.000	18:53
Control Test		0.082	18:53
Air Blank		0.000	18:54
Subject Sample	# 1	0.000	18:54
Breath Volume		4.410	18:54
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.000	18:57
Breath Volume		4.226	18:57
Air Blank		0.000	18:57
Control Test		0.081	18:58
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 11/30/2021 11:36

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZEWICZ,ALISON,E
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAZEWICZ,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 10/29/2021 04:06:35

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.082	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.000	04:01
Breath Volume		1.109	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:05
Breath Volume		1.433	04:05
Air Blank		0.000	04:05
Control Test		0.082	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 11/30/2021 11:37

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 10/06/2021 19:18:14

Diagnostics Check		OK	19:11	
Air Blank		0.000	19:12	
Control Test		0.078	19:12	
Air Blank		0.000	19:12	
Subject Sample	# 1	0.208	19:13	
Breath Volume		2.359	19:13	
Air Blank		0.000	19:13	
Air Blank		0.000	19:15	
Subject Sample	# 2	0.212	19:16	
Breath Volume		2.144	19:16	
Air Blank		0.000	19:16	
Control Test		0.079	19:17	
Air Blank		0.000	19:17	
Diagnostics Check		Fail	19:17	Analytical Stability Test
Air Blank		0.000	19:18	

Data Download Date/Time: 10/27/2021 16:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 10/06/2021 19:29:54

Diagnostics Check		OK	19:23
Air Blank		0.000	19:23
Control Test		0.078	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.206	19:25
Breath Volume		2.355	19:25
Air Blank		0.000	19:26
Air Blank		0.000	19:27
Subject Sample	# 2	0.206	19:28
Breath Volume		2.019	19:28
Air Blank		0.000	19:28
Control Test		0.078	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 10/27/2021 16:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CABREA

DATE/TIME OF TEST: 10/12/2021 01:39:24

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.082	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.000	01:35
Breath Volume		1.570	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.000	01:38
Breath Volume		1.683	01:38
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 10/27/2021 16:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 10/14/2021 23:47:25

Diagnostics Check		OK	23:41
Air Blank		0.000	23:41
Control Test		0.078	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.108	23:43
Breath Volume		2.386	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.108	23:45
Breath Volume		1.972	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 10/27/2021 16:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 10/24/2021 00:39:43

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:35
Breath Volume		1.863	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.000	00:38
Breath Volume		2.042	00:38
Air Blank		0.000	00:38
Control Test		0.079	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 10/27/2021 16:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SANTARROSA LUCAS

DATE/TIME OF TEST: 10/24/2021 03:52:49

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.081	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1	0.207	03:41	Volume Not Met
Breath Volume		0.132	03:41	
Air Blank		0.000	03:42	
Air Blank		0.000	03:44	
Subject Sample	# 2	0.234	03:46	
Breath Volume		1.367	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 3	0.247	03:51	
Breath Volume		1.937	03:51	
Air Blank		0.000	03:51	
Control Test		0.078	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 10/27/2021 16:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/25/2021 02:13:38

Diagnostics Check		OK	01:59	
Air Blank		0.000	01:59	
Control Test		0.078	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 1	0.147	02:03	Volume Not Met
Breath Volume		0.265	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 2	0.136	02:08	
Breath Volume		2.027	02:08	
Air Blank		0.000	02:08	
Air Blank		0.000	02:10	
Subject Sample	# 3	0.133	02:12	
Breath Volume		1.476	02:12	
Air Blank		0.000	02:12	
Control Test		0.077	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 10/27/2021 16:52

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ELMAKOUDI

DATE/TIME OF TEST: 10/01/2021 01:20:54

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.058	01:15
Breath Volume		2.527	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.052	01:19
Breath Volume		1.125	01:19
Air Blank		0.000	01:19
Control Test		0.079	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 10/03/2021 22:29

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 10/01/2021 21:34:56

Diagnostics Check		OK	21:28
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:30
Subject Sample	# 1	0.112	21:30
Breath Volume		2.585	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.110	21:33
Breath Volume		2.593	21:33
Air Blank		0.000	21:34
Control Test		0.079	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 10/03/2021 22:29

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BELT,WILLIAM
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: CRONIN

DATE/TIME OF TEST: 10/02/2021 03:42:35

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.175	03:37
Breath Volume		1.531	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.191	03:41
Breath Volume		2.988	03:41
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 10/03/2021 22:30

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 10/02/2021 23:19:40

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.078	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.240	23:15
Breath Volume		2.468	23:15
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.235	23:18
Breath Volume		2.390	23:18
Air Blank		0.000	23:18
Control Test		0.079	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 10/03/2021 22:30

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: LIGAC

DATE/TIME OF TEST: 10/04/2021 18:46:31

Diagnostics Check		OK	18:40
Air Blank		0.000	18:40
Control Test		0.080	18:41
Air Blank		0.000	18:41
Subject Sample	# 1	0.056	18:42
Breath Volume		1.851	18:42
Air Blank		0.000	18:42
Air Blank		0.000	18:44
Subject Sample	# 2	0.057	18:45
Breath Volume		1.933	18:45
Air Blank		0.000	18:45
Control Test		0.080	18:45
Air Blank		0.000	18:46
Diagnostics Check		OK	18:46

Data Download Date/Time: 10/31/2021 04:47

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: KEE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: FROMANG

DATE/TIME OF TEST: 10/04/2021 20:31:16

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:18	
Control Test		0.080	20:19	
Air Blank		0.000	20:19	
Subject Sample	# 1	0.042	20:22	Volume Not Met
Breath Volume		1.429	20:22	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 2	0.038	20:25	
Breath Volume		1.183	20:25	
Air Blank		0.000	20:26	
Air Blank		0.000	20:28	
Subject Sample	# 3	0.035	20:29	
Breath Volume		1.339	20:29	
Air Blank		0.000	20:30	
Control Test		0.078	20:30	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 10/31/2021 04:47

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PFANNERSTILL,JACOB
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: COPPOLO

DATE/TIME OF TEST: 10/04/2021 22:42:01

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.080	22:36
Air Blank		0.000	22:37
Subject Sample	# 1	0.000	22:37
Breath Volume		4.089	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.000	22:40
Breath Volume		1.906	22:40
Air Blank		0.000	22:41
Control Test		0.079	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:42

Data Download Date/Time: 10/31/2021 04:47

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENCH,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: GONGALEZ,J
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: PAULIN

DATE/TIME OF TEST: 10/06/2021 09:49:55

Diagnostics Check		OK	09:43
Air Blank		0.000	09:43
Control Test		0.080	09:44
Air Blank		0.000	09:44
Subject Sample	# 1	0.224	09:45
Breath Volume		2.250	09:45
Air Blank		0.000	09:45
Air Blank		0.000	09:47
Subject Sample	# 2	0.228	09:48
Breath Volume		2.042	09:48
Air Blank		0.000	09:49
Control Test		0.079	09:49
Air Blank		0.000	09:49
Diagnostics Check		OK	09:49

Data Download Date/Time: 10/31/2021 04:47

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JONES,C
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: BLEVINS

DATE/TIME OF TEST: 10/08/2021 16:48:54

Diagnostics Check		OK	16:46	
Air Blank		0.000	16:46	
Control Test		0.080	16:46	
Air Blank		0.000	16:47	
Subject Sample	# 1		16:47	RFI Detect
Breath Volume		0.000	16:47	
Air Blank		0.000	16:47	
Control Test		0.080	16:48	
Air Blank		0.000	16:48	
Diagnostics Check		OK	16:48	

Data Download Date/Time: 10/31/2021 04:47

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JONES,C
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: BLEVINS

DATE/TIME OF TEST: 10/08/2021 16:54:38

Diagnostics Check		OK	16:51	
Air Blank		0.000	16:51	
Control Test		0.080	16:52	
Air Blank		0.000	16:52	
Subject Sample	# 1		16:53	Subject Test Refused
Breath Volume		0.000	16:53	
Air Blank		0.000	16:53	
Control Test		0.080	16:54	
Air Blank		0.000	16:54	
Diagnostics Check		OK	16:54	

Data Download Date/Time: 10/31/2021 04:47

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HELTON,RYAN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RIZZO

DATE/TIME OF TEST: 10/19/2021 02:03:18

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.079	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.106	01:58
Breath Volume		1.730	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.117	02:01
Breath Volume		1.289	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 10/31/2021 04:47

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KILLON,J
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 10/21/2021 04:50:09

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.080	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.000	04:45
Breath Volume		3.105	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.000	04:48
Breath Volume		2.468	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 10/31/2021 04:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 10/22/2021 01:06:59

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.145	01:01
Breath Volume		2.187	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.156	01:05
Breath Volume		2.320	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 10/31/2021 04:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SAAVEDRA,JUAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 10/22/2021 22:04:26

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.079	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.279	21:59
Breath Volume		1.480	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.281	22:02
Breath Volume		2.386	22:02
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 10/31/2021 04:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NSOULI,C
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 09/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: BRAXTON

DATE/TIME OF TEST: 10/23/2021 07:23:11

Diagnostics Check		OK	07:13	
Air Blank		0.000	07:13	
Control Test		0.078	07:13	
Air Blank		0.000	07:14	
Subject Sample	# 1	0.215	07:17	Volume Not Met
Breath Volume		0.171	07:17	
Air Blank		0.000	07:18	
Air Blank		0.000	07:20	
Subject Sample	# 2		07:21	Subject Test Refused
Breath Volume		0.125	07:21	
Air Blank		0.000	07:22	
Control Test		0.078	07:22	
Air Blank		0.000	07:23	
Diagnostics Check		OK	07:23	

Data Download Date/Time: 10/31/2021 04:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HANNAH,CHANDLER
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 10/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GRIMM

DATE/TIME OF TEST: 10/30/2021 03:40:01

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.158	03:35
Breath Volume		3.476	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.163	03:38
Breath Volume		3.203	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:40

Data Download Date/Time: 10/31/2021 04:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVICI,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/01/2021 00:09:45

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.080	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.088	00:04
Breath Volume		1.582	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.087	00:08
Breath Volume		1.464	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 10/29/2021 19:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: SALVAGGIO

DATE/TIME OF TEST: 10/01/2021 10:54:40

Diagnostics Check		OK	10:47
Air Blank		0.000	10:48
Control Test		0.080	10:48
Air Blank		0.000	10:48
Subject Sample	# 1	0.000	10:50
Breath Volume		2.066	10:50
Air Blank		0.000	10:50
Air Blank		0.000	10:52
Subject Sample	# 2	0.000	10:53
Breath Volume		1.929	10:53
Air Blank		0.000	10:53
Control Test		0.079	10:54
Air Blank		0.000	10:54
Diagnostics Check		OK	10:54

Data Download Date/Time: 10/29/2021 19:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BALCH

DATE/TIME OF TEST: 10/02/2021 02:11:50

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.000	02:07
Breath Volume		1.515	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.000	02:10
Breath Volume		2.195	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 10/29/2021 19:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/02/2021 05:13:51

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:02	
Control Test		0.079	05:03	
Air Blank		0.000	05:03	
Subject Sample	# 1	0.110	05:06	Volume Not Met
Breath Volume		0.882	05:06	
Air Blank		0.000	05:07	
Air Blank		0.000	05:09	
Subject Sample	# 2		05:12	No Sample Provided
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.078	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 10/29/2021 19:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/02/2021 05:40:51

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:26	
Control Test		0.079	05:27	
Air Blank		0.000	05:27	
Subject Sample	# 1	0.106	05:28	
Breath Volume		1.562	05:28	
Air Blank		0.000	05:28	
Air Blank		0.000	05:30	
Subject Sample	# 2	0.096	05:33	Slope Not Level
Breath Volume		1.101	05:33	
Air Blank		0.000	05:34	
Air Blank		0.000	05:36	
Subject Sample	# 3		05:39	No Sample Provided
Breath Volume		0.000	05:39	
Air Blank		0.000	05:39	
Control Test		0.078	05:40	
Air Blank		0.000	05:40	
Diagnostics Check		OK	05:40	

Data Download Date/Time: 10/29/2021 19:09

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/02/2021 05:56:30

Diagnostics Check		OK	05:50
Air Blank		0.000	05:50
Control Test		0.079	05:50
Air Blank		0.000	05:51
Subject Sample	# 1	0.095	05:51
Breath Volume		1.402	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.095	05:54
Breath Volume		1.394	05:54
Air Blank		0.000	05:55
Control Test		0.079	05:55
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 10/03/2021 01:02:26

Diagnostics Check		OK	00:59	
Air Blank		0.000	01:00	
Control Test		0.079	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.079	01:01	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARLINGTON,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/03/2021 01:13:16

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: HAENGEL

DATE/TIME OF TEST: 10/03/2021 02:10:21

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.184	02:05
Breath Volume		2.042	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.186	02:08
Breath Volume		2.121	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCGARRY

DATE/TIME OF TEST: 10/03/2021 02:52:08

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.000	02:47
Breath Volume		1.820	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.000	02:50
Breath Volume		2.011	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SINGHAVONG

DATE/TIME OF TEST: 10/03/2021 03:27:26

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.118	03:22
Breath Volume		2.074	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.113	03:25
Breath Volume		1.664	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: KUHLES

DATE/TIME OF TEST: 10/03/2021 06:07:47

Diagnostics Check		OK	06:04	
Air Blank		0.000	06:05	
Control Test		0.080	06:05	
Air Blank		0.000	06:06	
Subject Sample	# 1		06:06	Subject Test Refused
Breath Volume		0.000	06:06	
Air Blank		0.000	06:06	
Control Test		0.079	06:07	
Air Blank		0.000	06:07	
Diagnostics Check		OK	06:07	

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 10/03/2021 22:57:06

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.080	22:50
Air Blank		0.000	22:50
Subject Sample	# 1	0.000	22:51
Breath Volume		1.183	22:51
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.000	22:55
Breath Volume		1.109	22:55
Air Blank		0.000	22:56
Control Test		0.080	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 10/05/2021 21:34:43

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:29
Air Blank		0.000	21:29
Subject Sample	# 1	0.000	21:30
Breath Volume		1.328	21:30
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.000	21:33
Breath Volume		2.164	21:33
Air Blank		0.000	21:33
Control Test		0.080	21:34
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JAGODA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/05/2021 23:33:30

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.080	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.187	23:28
Breath Volume		2.222	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.183	23:31
Breath Volume		2.089	23:31
Air Blank		0.000	23:32
Control Test		0.079	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG012501
DRY GAS EXPIRY DATE: 05/04/2022
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: BOTTA

DATE/TIME OF TEST: 10/06/2021 21:52:10

Diagnostics Check		OK	21:45
Air Blank		0.000	21:45
Control Test		0.079	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.185	21:47
Breath Volume		1.582	21:47
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.182	21:50
Breath Volume		3.300	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: RIDER

DATE/TIME OF TEST: 10/08/2021 19:59:33

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.079	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.072	19:55
Breath Volume		2.445	19:55
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.071	19:57
Breath Volume		2.070	19:57
Air Blank		0.000	19:58
Control Test		0.080	19:58
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: LONABAUGH

DATE/TIME OF TEST: 10/09/2021 00:33:23

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.000	00:29
Breath Volume		3.085	00:29
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.000	00:31
Breath Volume		2.789	00:31
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DICKMAN

DATE/TIME OF TEST: 10/09/2021 02:51:07

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:50	
Control Test		0.079	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GRADWELL

DATE/TIME OF TEST: 10/10/2021 00:37:15

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:32
Breath Volume		1.609	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.000	00:35
Breath Volume		2.226	00:35
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,V
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 10/11/2021 01:13:35

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.099	01:08
Breath Volume		1.636	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.101	01:11
Breath Volume		1.609	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DONLEY

DATE/TIME OF TEST: 10/11/2021 04:50:03

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.212	04:45
Breath Volume		1.496	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:47
Subject Sample	# 2	0.215	04:48
Breath Volume		1.335	04:48
Air Blank		0.000	04:49
Control Test		0.078	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:50

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUTH,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: HAGES

DATE/TIME OF TEST: 10/11/2021 06:55:53

Diagnostics Check		OK	06:49
Air Blank		0.000	06:50
Control Test		0.079	06:50
Air Blank		0.000	06:51
Subject Sample	# 1	0.000	06:51
Breath Volume		2.062	06:51
Air Blank		0.000	06:51
Air Blank		0.000	06:53
Subject Sample	# 2	0.000	06:54
Breath Volume		2.007	06:54
Air Blank		0.000	06:54
Control Test		0.080	06:55
Air Blank		0.000	06:55
Diagnostics Check		OK	06:55

Data Download Date/Time: 10/29/2021 19:10

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZIMMERMAN,AUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/12/2021 16:52:23

Diagnostics Check		OK	16:45
Air Blank		0.000	16:45
Control Test		0.079	16:46
Air Blank		0.000	16:46
Subject Sample	# 1	0.333	16:47
Breath Volume		1.417	16:47
Air Blank		0.000	16:48
Air Blank		0.000	16:50
Subject Sample	# 2	0.327	16:50
Breath Volume		1.441	16:50
Air Blank		0.000	16:51
Control Test		0.078	16:51
Air Blank		0.000	16:52
Diagnostics Check		OK	16:52

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOUGHERTY,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 10/12/2021 23:56:08

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.078	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.137	23:50
Breath Volume		1.312	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:52
Subject Sample	# 2	0.144	23:54
Breath Volume		1.949	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ROOT II

DATE/TIME OF TEST: 10/15/2021 02:14:22

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.080	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:13	
Control Test		0.079	02:13	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WYMAN

DATE/TIME OF TEST: 10/15/2021 03:43:43

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.080	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Control Test		0.079	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRILL,SHANNON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MORRIS JR

DATE/TIME OF TEST: 10/15/2021 05:11:52

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:07	
Control Test		0.080	05:08	
Air Blank		0.000	05:08	
Subject Sample	# 1		05:10	Subject Test Refused
Breath Volume		0.000	05:10	
Air Blank		0.000	05:10	
Control Test		0.079	05:11	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: IVANCVICI,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/15/2021 23:27:35

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:19	
Control Test		0.079	23:19	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:23	No Sample Provided
Breath Volume		0.000	23:23	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 2		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.080	23:26	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SMALLEY

DATE/TIME OF TEST: 10/16/2021 03:44:07

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.079	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/16/2021 22:11:06

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.141	22:06
Breath Volume		3.132	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.139	22:09
Breath Volume		3.242	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/17/2021 23:40:37

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.221	23:36
Breath Volume		2.156	23:36
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.217	23:38
Breath Volume		2.031	23:38
Air Blank		0.000	23:39
Control Test		0.078	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,ZOE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: RODAS

DATE/TIME OF TEST: 10/18/2021 00:16:54

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.263	00:12
Breath Volume		2.289	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.265	00:15
Breath Volume		1.800	00:15
Air Blank		0.000	00:15
Control Test		0.078	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,ZOE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: OCSASIO

DATE/TIME OF TEST: 10/18/2021 02:42:11

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.260	02:37
Breath Volume		1.621	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.272	02:40
Breath Volume		2.195	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SINGSON,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: HORD

DATE/TIME OF TEST: 10/18/2021 23:52:44

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:48	
Control Test		0.079	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:51	Subject Test Refused
Breath Volume		0.000	23:51	
Air Blank		0.000	23:51	
Control Test		0.079	23:52	
Air Blank		0.000	23:52	
Diagnostics Check		OK	23:52	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KOWALSKI,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MICKUS

DATE/TIME OF TEST: 10/19/2021 01:11:40

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.078	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:11	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 10/19/2021 01:48:52

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.111	01:44
Breath Volume		1.839	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:46
Subject Sample	# 2	0.110	01:47
Breath Volume		1.660	01:47
Air Blank		0.000	01:47
Control Test		0.078	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEWIS,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RABBERS

DATE/TIME OF TEST: 10/19/2021 04:05:18

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.078	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:04	
Control Test		0.078	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,K,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SWISHER,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 10:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 10/19/2021 11:21:02

Diagnostics Check		OK	11:14
Air Blank		0.000	11:15
Control Test		0.079	11:15
Air Blank		0.000	11:16
Subject Sample	# 1	0.000	11:16
Breath Volume		2.398	11:16
Air Blank		0.000	11:17
Air Blank		0.000	11:19
Subject Sample	# 2	0.000	11:19
Breath Volume		2.593	11:19
Air Blank		0.000	11:20
Control Test		0.079	11:20
Air Blank		0.000	11:20
Diagnostics Check		OK	11:21

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,M,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KELLY,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 10/20/2021 03:36:50

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.158	03:32
Breath Volume		1.953	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.163	03:35
Breath Volume		2.152	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 10/20/2021 06:05:32

Diagnostics Check		OK	06:02	
Air Blank		0.000	06:03	
Control Test		0.080	06:03	
Air Blank		0.000	06:03	
Subject Sample	# 1		06:04	Subject Test Refused
Breath Volume		0.000	06:04	
Air Blank		0.000	06:04	
Control Test		0.079	06:04	
Air Blank		0.000	06:05	
Diagnostics Check		OK	06:05	

Data Download Date/Time: 10/29/2021 19:11

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: QUIMBY

DATE/TIME OF TEST: 10/20/2021 23:31:01

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.080	23:28	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:29	Subject Test Refused
Breath Volume		0.000	23:29	
Air Blank		0.000	23:30	
Control Test		0.079	23:30	
Air Blank		0.000	23:30	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ROBB,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: SZAST

DATE/TIME OF TEST: 10/21/2021 22:11:09

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.079	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.167	22:06
Breath Volume		2.417	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.156	22:09
Breath Volume		1.554	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DUCKERS,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SKOLE

DATE/TIME OF TEST: 10/22/2021 00:33:13

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.079	00:31	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:31	Subject Test Refused
Breath Volume		0.000	00:31	
Air Blank		0.000	00:32	
Control Test		0.079	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,ZOE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/22/2021 02:40:00

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.079	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.174	02:35
Breath Volume		1.281	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:37
Subject Sample	# 2	0.174	02:38
Breath Volume		1.425	02:38
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,STEVEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TURNER,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/22/2021 18:41:12

Diagnostics Check		OK	18:35
Air Blank		0.000	18:35
Control Test		0.078	18:35
Air Blank		0.000	18:36
Subject Sample	# 1	0.294	18:36
Breath Volume		1.968	18:36
Air Blank		0.000	18:37
Air Blank		0.000	18:39
Subject Sample	# 2	0.286	18:39
Breath Volume		1.824	18:39
Air Blank		0.000	18:40
Control Test		0.079	18:40
Air Blank		0.000	18:41
Diagnostics Check		OK	18:41

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: GRANDSTROM

DATE/TIME OF TEST: 10/22/2021 23:39:36

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.079	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.080	23:35
Breath Volume		3.808	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.080	23:37
Breath Volume		3.847	23:37
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 10/23/2021 00:29:56

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:27	
Control Test		0.078	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Control Test		0.079	00:29	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/23/2021 02:46:57

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.177	02:42
Breath Volume		2.167	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.171	02:45
Breath Volume		2.453	02:45
Air Blank		0.000	02:45
Control Test		0.079	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: WARRANT

DATE/TIME OF TEST: 10/23/2021 03:19:47

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.186	03:14
Breath Volume		2.375	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.169	03:18
Breath Volume		2.195	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/23/2021 05:22:00

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:19	
Control Test		0.080	05:19	
Air Blank		0.000	05:20	
Subject Sample	# 1		05:20	Subject Test Refused
Breath Volume		0.000	05:20	
Air Blank		0.000	05:21	
Control Test		0.080	05:21	
Air Blank		0.000	05:21	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERRETTI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: HORNAK

DATE/TIME OF TEST: 10/23/2021 19:25:11

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.079	19:17
Air Blank		0.000	19:17
Subject Sample	# 1	0.279	19:20
Breath Volume		2.714	19:20
Air Blank		0.000	19:21
Air Blank		0.000	19:22
Subject Sample	# 2	0.279	19:23
Breath Volume		2.109	19:23
Air Blank		0.000	19:24
Control Test		0.078	19:24
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/23/2021 23:08:05

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.166	23:03
Breath Volume		1.632	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.165	23:06
Breath Volume		2.062	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:08

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 10/23/2021 23:22:22

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.078	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.145	23:17
Breath Volume		1.796	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.144	23:20
Breath Volume		2.023	23:20
Air Blank		0.000	23:21
Control Test		0.077	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEAVER,N
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: SHIMER

DATE/TIME OF TEST: 10/24/2021 00:34:00

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.079	00:31	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:32	Subject Test Refused
Breath Volume		0.000	00:32	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERRETTI,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SIDLOSKAS

DATE/TIME OF TEST: 10/24/2021 01:07:00

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.078	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.278	01:02
Breath Volume		2.289	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.287	01:05
Breath Volume		1.746	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:07

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COLMAN,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 10/24/2021 03:55:08

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.079	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:53	Subject Test Refused
Breath Volume		0.000	03:53	
Air Blank		0.000	03:54	
Control Test		0.080	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICOTTILLI,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MIELE,ROBERT
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: STEADMIRE

DATE/TIME OF TEST: 10/25/2021 01:43:29

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.159	01:38
Breath Volume		1.250	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.171	01:41
Breath Volume		1.742	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTWICK,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRAME,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BEAM

DATE/TIME OF TEST: 10/25/2021 05:17:19

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.080	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.182	05:12
Breath Volume		1.265	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:14
Subject Sample	# 2	0.175	05:15
Breath Volume		1.355	05:15
Air Blank		0.000	05:16
Control Test		0.079	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 10/29/2021 19:12

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFFORD,BRIAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PULIZZANO,ALYSSA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 10/25/2021 15:11:16

Diagnostics Check		OK	15:04
Air Blank		0.000	15:05
Control Test		0.079	15:05
Air Blank		0.000	15:06
Subject Sample	# 1	0.000	15:06
Breath Volume		1.117	15:06
Air Blank		0.000	15:07
Air Blank		0.000	15:09
Subject Sample	# 2	0.000	15:09
Breath Volume		1.488	15:09
Air Blank		0.000	15:10
Control Test		0.079	15:10
Air Blank		0.000	15:11
Diagnostics Check		OK	15:11

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: NAYLOR

DATE/TIME OF TEST: 10/25/2021 19:17:30

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.080	19:11
Air Blank		0.000	19:12
Subject Sample	# 1	0.115	19:12
Breath Volume		1.984	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.110	19:15
Breath Volume		1.488	19:15
Air Blank		0.000	19:16
Control Test		0.079	19:16
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IFFT,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 10/26/2021 01:59:29

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.079	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:55	No Sample Provided
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:58	
Control Test		0.080	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RILL,AUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/27/2021 04:19:02

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.079	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1		04:11	No Sample Provided
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 2	0.315	04:16	Volume Not Met
Breath Volume		0.058	04:16	
Air Blank		0.000	04:18	
Control Test		0.078	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,KAYLA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: MURROW

DATE/TIME OF TEST: 10/27/2021 18:25:32

Diagnostics Check		OK	18:22	
Air Blank		0.000	18:22	
Control Test		0.079	18:23	
Air Blank		0.000	18:23	
Subject Sample	# 1		18:24	Subject Test Refused
Breath Volume		0.000	18:24	
Air Blank		0.000	18:24	
Control Test		0.080	18:24	
Air Blank		0.000	18:25	
Diagnostics Check		OK	18:25	

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: ROLLENHAGEN

DATE/TIME OF TEST: 10/27/2021 19:44:44

Diagnostics Check		OK	19:38
Air Blank		0.000	19:39
Control Test		0.079	19:39
Air Blank		0.000	19:40
Subject Sample	# 1	0.000	19:40
Breath Volume		1.921	19:40
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.000	19:43
Breath Volume		1.457	19:43
Air Blank		0.000	19:43
Control Test		0.079	19:44
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: HINCHEY

DATE/TIME OF TEST: 10/27/2021 22:53:05

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.079	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.039	22:48
Breath Volume		2.734	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.039	22:51
Breath Volume		2.097	22:51
Air Blank		0.000	22:52
Control Test		0.078	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 10/28/2021 01:59:32

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.192	01:52
Breath Volume		1.511	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.186	01:57
Breath Volume		1.230	01:57
Air Blank		0.000	01:58
Control Test		0.077	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ G,YOEL-ALEXIZ,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ALVAREZ G,YOEL-ALEXIZ
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 10/28/2021 23:44:56

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.252	23:37
Breath Volume		3.875	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:39
Subject Sample	# 2	0.279	23:40
Breath Volume		3.703	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:42
Subject Sample	# 3	0.271	23:43
Breath Volume		3.800	23:43
Air Blank		0.000	23:43
Control Test		0.078	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARUSO,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 10/29/2021 01:19:55

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.078	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Control Test		0.078	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONAHUE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HANRATTY

DATE/TIME OF TEST: 10/29/2021 01:42:48

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.186	01:37
Breath Volume		2.425	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.189	01:41
Breath Volume		2.152	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JAGODA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MEDLOCK

DATE/TIME OF TEST: 10/29/2021 02:36:30

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.099	02:31
Breath Volume		2.140	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.095	02:34
Breath Volume		1.703	02:34
Air Blank		0.000	02:35
Control Test		0.078	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 10/29/2021 19:13

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DANIEL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: BONAWITZ

DATE/TIME OF TEST: 10/29/2021 19:54:28

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:52	
Control Test		0.079	19:52	
Air Blank		0.000	19:52	
Subject Sample	# 1		19:53	Subject Test Refused
Breath Volume		0.000	19:53	
Air Blank		0.000	19:53	
Control Test		0.079	19:53	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 11/24/2021 19:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IFFT,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/30/2021 02:48:11

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.222	02:43
Breath Volume		2.726	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.224	02:46
Breath Volume		2.949	02:46
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 11/24/2021 19:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KELLY,BRENDAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: CHESLEY

DATE/TIME OF TEST: 10/30/2021 18:07:31

Diagnostics Check		OK	18:00
Air Blank		0.000	18:01
Control Test		0.079	18:01
Air Blank		0.000	18:02
Subject Sample	# 1	0.256	18:02
Breath Volume		1.808	18:02
Air Blank		0.000	18:03
Air Blank		0.000	18:05
Subject Sample	# 2	0.243	18:05
Breath Volume		1.199	18:05
Air Blank		0.000	18:06
Control Test		0.078	18:06
Air Blank		0.000	18:07
Diagnostics Check		OK	18:07

Data Download Date/Time: 11/24/2021 19:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 10/30/2021 20:24:58

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.078	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.236	20:20
Breath Volume		1.691	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:22
Subject Sample	# 2	0.235	20:23
Breath Volume		1.890	20:23
Air Blank		0.000	20:24
Control Test		0.078	20:24
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 11/24/2021 19:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 10/31/2021 01:19:03

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.080	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 11/24/2021 19:49

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REIDY,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SLAGLE

DATE/TIME OF TEST: 10/31/2021 04:29:48

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.080	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.079	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 11/24/2021 19:49

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKEE,KATHLEEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HUTCHERSON

DATE/TIME OF TEST: 10/31/2021 22:31:53

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:20	
Control Test		0.080	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:24	No Sample Provided
Breath Volume		0.000	22:24	
Air Blank		0.000	22:25	
Air Blank		0.000	22:27	
Subject Sample	# 2		22:30	No Sample Provided
Breath Volume		0.000	22:30	
Air Blank		0.000	22:30	
Control Test		0.079	22:31	
Air Blank		0.000	22:31	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 11/24/2021 19:49

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKEE,KATHLEEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HUTCHERSON

DATE/TIME OF TEST: 10/31/2021 22:39:36

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.114	22:35
Breath Volume		2.160	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.116	22:37
Breath Volume		1.589	22:37
Air Blank		0.000	22:38
Control Test		0.078	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 11/24/2021 19:49

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/01/2021 01:34:27

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.229	01:27
Breath Volume		1.441	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.255	01:30
Breath Volume		1.507	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 3	0.239	01:32
Breath Volume		1.519	01:32
Air Blank		0.000	01:33
Control Test		0.078	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 10/30/2021 00:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 10/01/2021 02:52:38

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:50	
Control Test		0.079	02:50	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Control Test		0.079	02:52	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 10/30/2021 00:48

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: LONABAUGH

DATE/TIME OF TEST: 10/01/2021 04:54:13

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.079	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.000	04:49
Breath Volume		1.753	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.000	04:52
Breath Volume		1.398	04:52
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 10/30/2021 02:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,ZOE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: MAKINSON

DATE/TIME OF TEST: 10/02/2021 06:11:58

Diagnostics Check		OK	06:05
Air Blank		0.000	06:06
Control Test		0.078	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.127	06:07
Breath Volume		1.484	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:09
Subject Sample	# 2	0.128	06:10
Breath Volume		1.992	06:10
Air Blank		0.000	06:11
Control Test		0.078	06:11
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 10/30/2021 02:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DEG,KENISHA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 10/03/2021 17:12:07

Diagnostics Check		OK	17:08	
Air Blank		0.000	17:09	
Control Test		0.080	17:09	
Air Blank		0.000	17:10	
Subject Sample	# 1		17:10	Subject Test Refused
Breath Volume		0.000	17:10	
Air Blank		0.000	17:11	
Control Test		0.079	17:11	
Air Blank		0.000	17:12	
Diagnostics Check		OK	17:12	

Data Download Date/Time: 10/30/2021 02:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SAPP,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 10/03/2021 22:13:00

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:05	
Control Test		0.079	22:05	
Air Blank		0.000	22:05	
Subject Sample	# 1		22:09	No Sample Provided
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Air Blank		0.000	22:11	
Subject Sample	# 2		22:11	Subject Test Refused
Breath Volume		0.000	22:11	
Air Blank		0.000	22:12	
Control Test		0.079	22:12	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 10/30/2021 02:16

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,HOWARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 10/04/2021 07:00:27

Diagnostics Check		OK	06:54
Air Blank		0.000	06:54
Control Test		0.080	06:54
Air Blank		0.000	06:55
Subject Sample	# 1	0.170	06:55
Breath Volume		1.875	06:55
Air Blank		0.000	06:56
Air Blank		0.000	06:58
Subject Sample	# 2	0.159	06:58
Breath Volume		1.812	06:58
Air Blank		0.000	06:59
Control Test		0.079	06:59
Air Blank		0.000	07:00
Diagnostics Check		OK	07:00

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:36

SUBJECT'S LAST NAME: ALTAJAR

DATE/TIME OF TEST: 10/04/2021 11:04:08

Diagnostics Check		OK	10:57
Air Blank		0.000	10:58
Control Test		0.079	10:58
Air Blank		0.000	10:59
Subject Sample	# 1	0.000	10:59
Breath Volume		3.398	10:59
Air Blank		0.000	11:00
Air Blank		0.000	11:02
Subject Sample	# 2	0.000	11:02
Breath Volume		3.679	11:02
Air Blank		0.000	11:03
Control Test		0.079	11:03
Air Blank		0.000	11:04
Diagnostics Check		OK	11:04

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PETRACCO,JORDAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 10/06/2021 01:18:48

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.153	01:13
Breath Volume		2.101	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.159	01:17
Breath Volume		1.929	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 10/06/2021 03:05:36

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.078	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 1	0.218	02:57	Volume Not Met
Breath Volume		0.015	02:57	
Air Blank		0.000	02:58	
Air Blank		0.000	03:00	
Subject Sample	# 2	0.207	03:00	
Breath Volume		1.601	03:00	
Air Blank		0.000	03:01	
Air Blank		0.000	03:03	
Subject Sample	# 3	0.209	03:04	
Breath Volume		1.355	03:04	
Air Blank		0.000	03:04	
Control Test		0.079	03:05	
Air Blank		0.000	03:05	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,STEVEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TURNER,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 10/08/2021 22:24:55

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.080	22:22	
Air Blank		0.000	22:23	
Subject Sample	# 1		22:23	Subject Test Refused
Breath Volume		0.000	22:23	
Air Blank		0.000	22:24	
Control Test		0.079	22:24	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BYRD,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/09/2021 05:12:21

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:09	
Control Test		0.079	05:09	
Air Blank		0.000	05:10	
Subject Sample	# 1		05:10	Subject Test Refused
Breath Volume		0.000	05:10	
Air Blank		0.000	05:11	
Control Test		0.079	05:11	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PIXLER

DATE/TIME OF TEST: 10/09/2021 22:44:30

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.151	22:39
Breath Volume		2.070	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.138	22:42
Breath Volume		1.343	22:42
Air Blank		0.000	22:43
Control Test		0.079	22:43
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 10/12/2021 21:55:18

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:51	
Control Test		0.079	21:51	
Air Blank		0.000	21:51	
Subject Sample	# 1		21:53	Subject Test Refused
Breath Volume		0.000	21:53	
Air Blank		0.000	21:54	
Control Test		0.079	21:54	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BRATHWAITE

DATE/TIME OF TEST: 10/14/2021 02:22:56

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.034	02:18
Breath Volume		1.652	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.033	02:21
Breath Volume		1.128	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: O242805735190

DATE/TIME OF TEST: 10/14/2021 20:48:46

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.079	20:43
Air Blank		0.000	20:43
Subject Sample	# 1	0.000	20:44
Breath Volume		2.355	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.000	20:47
Breath Volume		1.882	20:47
Air Blank		0.000	20:47
Control Test		0.079	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZANANDRIE,KELLEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: DORRITIE

DATE/TIME OF TEST: 10/15/2021 04:14:53

Diagnostics Check		OK	04:12	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:13	Subject Test Refused
Breath Volume		0.000	04:13	
Air Blank		0.000	04:13	
Control Test		0.079	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONGMIRE,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LONGMIRE,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 10/15/2021 18:33:23

Diagnostics Check		OK	18:30	
Air Blank		0.000	18:31	
Control Test		0.078	18:31	
Air Blank		0.000	18:31	
Subject Sample	# 1		18:32	Subject Test Refused
Breath Volume		0.000	18:32	
Air Blank		0.000	18:32	
Control Test		0.079	18:32	
Air Blank		0.000	18:33	
Diagnostics Check		OK	18:33	

Data Download Date/Time: 10/30/2021 02:17

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 10/16/2021 03:57:33

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.079	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.126	03:50	Volume Not Met
Breath Volume		0.015	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 2	0.136	03:56	Volume Not Met
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Control Test		0.078	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 10/16/2021 04:03:17

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.078	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Control Test		0.079	04:02	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: TRUBECK

DATE/TIME OF TEST: 10/16/2021 20:22:46

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:16
Subject Sample	# 1	0.254	20:18
Breath Volume		1.250	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:20
Subject Sample	# 2	0.261	20:21
Breath Volume		1.644	20:21
Air Blank		0.000	20:21
Control Test		0.079	20:22
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,RANDY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: MACIAS TERREROS

DATE/TIME OF TEST: 10/17/2021 06:30:30

Diagnostics Check		OK	06:27	
Air Blank		0.000	06:27	
Control Test		0.079	06:28	
Air Blank		0.000	06:28	
Subject Sample	# 1		06:29	Subject Test Refused
Breath Volume		0.000	06:29	
Air Blank		0.000	06:29	
Control Test		0.079	06:29	
Air Blank		0.000	06:30	
Diagnostics Check		OK	06:30	

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MARION

DATE/TIME OF TEST: 10/19/2021 00:05:24

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.078	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.048	00:00
Breath Volume		2.609	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.049	00:03
Breath Volume		2.222	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,THOMAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: BALDRIDGE

DATE/TIME OF TEST: 10/19/2021 22:41:01

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.079	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.137	22:36
Breath Volume		1.183	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:38
Subject Sample	# 2	0.140	22:39
Breath Volume		1.832	22:39
Air Blank		0.000	22:40
Control Test		0.079	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:41

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORGAN,STEPHEN,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BONETTI

DATE/TIME OF TEST: 10/20/2021 03:27:54

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.321	03:23
Breath Volume		2.128	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:25
Subject Sample	# 2	0.316	03:26
Breath Volume		1.835	03:26
Air Blank		0.000	03:26
Control Test		0.078	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 10/23/2021 02:17:36

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.079	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,C,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: TAHO

DATE/TIME OF TEST: 10/27/2021 21:13:10

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.079	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.147	21:08
Breath Volume		1.273	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:10
Subject Sample	# 2	0.147	21:11
Breath Volume		1.695	21:11
Air Blank		0.000	21:12
Control Test		0.078	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BIBB

DATE/TIME OF TEST: 10/28/2021 02:18:38

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.078	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.079	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CHEK

DATE/TIME OF TEST: 10/29/2021 04:45:13

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.080	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:44	
Control Test		0.080	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: KOUBOS

DATE/TIME OF TEST: 10/29/2021 19:01:42

Diagnostics Check		OK	18:55
Air Blank		0.000	18:56
Control Test		0.080	18:56
Air Blank		0.000	18:56
Subject Sample	# 1	0.000	18:57
Breath Volume		1.171	18:57
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.000	19:00
Breath Volume		1.363	19:00
Air Blank		0.000	19:00
Control Test		0.079	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 10/30/2021 02:18

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 10/31/2021 01:46:22

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.182	01:42
Breath Volume		2.453	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.180	01:44
Breath Volume		2.253	01:44
Air Blank		0.000	01:45
Control Test		0.078	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 11/24/2021 21:51

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.