Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ATTARDI

DATE/TIME OF TEST: 03/13/2021 23:30:07

Diagnostics Check		OK	23:22	
Air Blank		0.000	23:22	
Control Test		0.083	23:23	
Air Blank		0.000	23:23	
Subject Sample	# 1	0.153	23:25	
Breath Volume		2.078	23:25	
Air Blank		0.000	23:25	
Air Blank		0.000	23:27	
Subject Sample	# 2		23:28	Interferent Detect
Breath Volume		2.894	23:28	
Air Blank		0.000	23:29	
Control Test		0.082	23:29	
Air Blank		0.000	23:30	
Diagnostics Check		OK	23:30	

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ATTARDI

DATE/TIME OF TEST: 03/13/2021 23:38:15

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.163	23:33
Breath Volume		2.593	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.161	23:36
Breath Volume		2.695	23:36
Air Blank		0.000	23:37
Control Test		0.078	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/18/2021 01:11:26

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.076	01:05	
Air Blank		0.000	01:05	
Subject Sample	# 1	0.222	01:06	
Breath Volume		1.777	01:06	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.227	01:09	
Breath Volume		2.074	01:09	
Air Blank		0.000	01:10	
Control Test		0.086	01:10	Control Outside Tolerance
Air Blank		0.000	01:11	

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/18/2021 01:14:43

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:14	Interferent Detect
Air Blank		01:14	Ambient Fail

Data Download Date/Time: 03/23/2021 14:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/18/2021 01:23:10

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:17
Air Blank		0.000	01:17
Subject Sample	# 1	0.222	01:18
Breath Volume		1.617	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:20
Subject Sample	# 2	0.227	01:21
Breath Volume		2.539	01:21
Air Blank		0.000	01:22
Control Test		0.076	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SPAGNOLO

DATE/TIME OF TEST: 03/18/2021 03:02:28

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.097	03:01	Control Outside Tolerance
Air Blank	0.000	03:02	

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SPAGNOLO

DATE/TIME OF TEST: 03/18/2021 03:05:13

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:04	
Control Test	0.085	03:04	Interferent Detect
Air Blank	0.000	03:05	

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SPAGNOLO

DATE/TIME OF TEST: 03/18/2021 03:13:50

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.081	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1	0.183	03:08	
Breath Volume		2.886	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2	0.193	03:11	
Breath Volume		2.000	03:11	
Air Blank		0.000	03:12	
Control Test		0.094	03:12	Control Outside Tolerance
Air Blank			03:13	Purge Fail

Data Download Date/Time: 03/23/2021 14:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SPAGNOLO

DATE/TIME OF TEST: 03/18/2021 03:17:21

Diagnostics Check	OK	03:16	
Air Blank		03:16	Ambient Fail
Air Blank	0.000	03:17	

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SPAGNOLO

DATE/TIME OF TEST: 03/18/2021 03:21:42

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.075	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:20	Slope Not Met
Breath Volume		1.539	03:20	
Air Blank			03:21	Purge Fail
Air Blank		0.000	03:21	

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DOE,JOHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/18/2021 13:36:00

Diagnostics Check	OK	13:34	
Air Blank	0.000	13:35	
Control Test	0.073	13:35	Control Outside Tolerance
Air Blank	0.000	13:36	

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/22/2021 17:12:58

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.082	17:06
Air Blank		0.000	17:07
Subject Sample	# 1	0.000	17:08
Breath Volume		1.125	17:08
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 2	0.000	17:11
Breath Volume		1.742	17:11
Air Blank		0.000	17:12
Control Test		0.082	17:12
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 03/23/2021 14:32

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RODRIGUEZ PAZ

DATE/TIME OF TEST: 03/28/2021 01:32:55

Diagnostics Check		OK	01:26	
Air Blank		0.000	01:26	
Control Test		0.078	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1	0.200	01:28	
Breath Volume		1.207	01:28	
Air Blank		0.000	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 2	0.219	01:31	
Breath Volume		2.773	01:31	
Air Blank			01:32	RFI Detect
Air Blank		0.000	01:32	

Data Download Date/Time: 04/16/2021 13:56

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RODRIGUEZ PAZ

DATE/TIME OF TEST: 03/28/2021 01:41:42

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.207	01:36
Breath Volume		1.789	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.201	01:39
Breath Volume		2.617	01:39
Air Blank		0.000	01:40
Control Test		0.082	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 04/16/2021 13:56

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: DOLAND

DATE/TIME OF TEST: 03/11/2021 19:00:16

Diagnostics Check		OK	18:53
Air Blank		0.000	18:53
Control Test		0.082	18:53
Air Blank		0.000	18:54
Subject Sample	# 1	0.120	18:55
Breath Volume		1.531	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:57
Subject Sample	# 2	0.116	18:58
Breath Volume		1.582	18:58
Air Blank		0.000	18:59
Control Test		0.078	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 03/24/2021 13:02

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER III,RICHARD,F
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER III,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SOSA-CASTELLON

DATE/TIME OF TEST: 03/12/2021 00:38:18

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.082	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.164	00:33
Breath Volume		2.632	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.163	00:36
Breath Volume		2.292	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 03/24/2021 13:03

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 03/15/2021 02:58:32

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.082	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.000	02:54
Breath Volume		3.441	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.000	02:57
Breath Volume		3.277	02:57
Air Blank		0.000	02:57
Control Test		0.081	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 03/24/2021 13:03

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SPAGNOLO

DATE/TIME OF TEST: 03/18/2021 03:54:31

Diagnostics Check	OK	03:53	
Air Blank		03:54	RFI Detect
Air Blank		03:54	RFI Detect

Data Download Date/Time: 03/24/2021 13:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SPAGNOLO

DATE/TIME OF TEST: 03/18/2021 04:02:36

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.076	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.165	03:57
Breath Volume		1.632	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.168	04:00
Breath Volume		2.105	04:00
Air Blank		0.000	04:01
Control Test		0.075	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 03/24/2021 13:03

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: CUTLER

DATE/TIME OF TEST: 03/18/2021 23:09:39

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.078	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.162	23:05
Breath Volume		2.152	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.159	23:08
Breath Volume		1.386	23:08
Air Blank		0.000	23:08
Control Test		0.076	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 03/24/2021 13:03

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SPITLER

DATE/TIME OF TEST: 03/01/2021 00:00:57

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.080	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.000	23:56
Breath Volume		3.785	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		4.535	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 03/10/2021 13:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SCARBOROUGH

DATE/TIME OF TEST: 03/04/2021 04:15:50

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.081	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.190	04:10
Breath Volume		2.125	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.172	04:14
Breath Volume		1.386	04:14
Air Blank		0.000	04:14
Control Test		0.079	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 03/10/2021 13:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONCION,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: DE MENESES

DATE/TIME OF TEST: 03/06/2021 04:11:54

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.130	04:06
Breath Volume		2.296	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.119	04:10
Breath Volume		1.128	04:10
Air Blank		0.000	04:10
Control Test		0.081	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 03/10/2021 13:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 03/07/2021 22:10:33

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:04
Subject Sample	# 1	0.000	22:06
Breath Volume		2.375	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.000	22:09
Breath Volume		2.214	22:09
Air Blank		0.000	22:09
Control Test		0.081	22:09
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 03/10/2021 13:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: EASLER

DATE/TIME OF TEST: 03/07/2021 23:54:35

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.174	23:49
Breath Volume		2.339	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.169	23:52
Breath Volume		2.433	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 03/10/2021 13:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MORALES RODRIGUEZ

DATE/TIME OF TEST: 03/08/2021 00:44:42

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.153	00:40
Breath Volume		2.468	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.156	00:43
Breath Volume		2.085	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 03/10/2021 13:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GARCIA VALLEJOS

DATE/TIME OF TEST: 03/08/2021 01:16:09

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.189	01:11
Breath Volume		3.011	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.188	01:14
Breath Volume		1.542	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 03/10/2021 13:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/13/2021 02:17:06

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.087	02:10
Breath Volume		1.457	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.084	02:15
Breath Volume		1.167	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 03/14/2021 01:25:34

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.142	01:20
Breath Volume		1.425	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.133	01:23
Breath Volume		2.359	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: TAJAH

DATE/TIME OF TEST: 03/18/2021 03:36:02

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.129	03:31
Breath Volume		2.468	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.128	03:34
Breath Volume		1.597	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:36

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: SALGADO RAMIREZ

DATE/TIME OF TEST: 03/19/2021 00:15:37

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.081	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.216	00:11
Breath Volume		1.550	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.211	00:13
Breath Volume		1.859	00:13
Air Blank		0.000	00:14
Control Test		0.077	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 03/20/2021 03:59:22

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.225	03:54
Breath Volume		2.726	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.219	03:57
Breath Volume		2.800	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: TIBBS

DATE/TIME OF TEST: 03/20/2021 04:19:48

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.123	04:15
Breath Volume		1.265	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.120	04:18
Breath Volume		1.585	04:18
Air Blank		0.000	04:18
Control Test		0.079	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SALTOS ARAUJO

DATE/TIME OF TEST: 03/21/2021 04:24:57

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.081	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.181	04:20
Breath Volume		2.828	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.191	04:23
Breath Volume		2.007	04:23
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 03/22/2021 04:26:38

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:13	
Control Test		0.081	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1	0.121	04:17	Volume Not Met
Breath Volume		0.226	04:17	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2	0.130	04:22	
Breath Volume		1.878	04:22	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 3	0.124	04:25	
Breath Volume		1.167	04:25	
Air Blank		0.000	04:25	
Control Test		0.081	04:26	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/24/2021 03:45:45

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.163	03:40
Breath Volume		1.726	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.176	03:44
Breath Volume		2.882	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 03/27/2021 04:44:30

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.081	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.166	04:39
Breath Volume		1.699	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.165	04:42
Breath Volume		1.914	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 04/06/2021 16:19

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 03/29/2021 03:31:22

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.197	03:26
Breath Volume		1.445	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:28
Subject Sample	# 2	0.190	03:29
Breath Volume		1.109	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 04/06/2021 16:20

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DENKER,B
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: QUESTELL

DATE/TIME OF TEST: 03/05/2021 10:26:53

Diagnostics Check		OK	10:20
Air Blank		0.000	10:21
Control Test		0.080	10:21
Air Blank		0.000	10:22
Subject Sample	# 1	0.172	10:22
Breath Volume		2.636	10:22
Air Blank		0.000	10:23
Air Blank		0.000	10:24
Subject Sample	# 2	0.186	10:25
Breath Volume		3.054	10:25
Air Blank		0.000	10:25
Control Test		0.080	10:26
Air Blank		0.000	10:26
Diagnostics Check		OK	10:26

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RONEMOUS

DATE/TIME OF TEST: 03/06/2021 00:08:18

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:56
Subject Sample	# 3	0.185	23:58
Breath Volume		1.863	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.163	00:02
Breath Volume		1.355	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.158	00:06
Breath Volume		1.718	00:06
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: YATSKO

DATE/TIME OF TEST: 03/11/2021 01:04:25

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.127	00:58
Breath Volume		1.140	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.125	01:02
Breath Volume		1.656	01:02
Air Blank		0.000	01:03
Control Test		0.076	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BRAUN

DATE/TIME OF TEST: 03/11/2021 02:15:15

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.081	02:10
Breath Volume		2.839	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.081	02:13
Breath Volume		2.890	02:13
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 03/12/2021 01:47:55

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.164	01:43
Breath Volume		1.886	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.154	01:46
Breath Volume		2.035	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,JEFFREY,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SOTO VALDES

DATE/TIME OF TEST: 03/13/2021 00:58:46

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.211	00:54
Breath Volume		2.085	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.214	00:57
Breath Volume		2.828	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MARTINEZ HURTADO

DATE/TIME OF TEST: 03/14/2021 01:47:59

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.146	01:43
Breath Volume		1.835	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.145	01:46
Breath Volume		2.027	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CLAY,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 03/14/2021 21:03:10

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.079	20:56
Air Blank		0.000	20:57
Subject Sample	# 1	0.211	20:58
Breath Volume		1.546	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.198	21:01
Breath Volume		1.679	21:01
Air Blank		0.000	21:02
Control Test		0.077	21:02
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GUERRERO,M
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 03/15/2021 02:16:30

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.080	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 04/02/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 03/17/2021 02:23:50

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.096	02:19
Breath Volume		2.355	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.087	02:22
Breath Volume		1.597	02:22
Air Blank		0.000	02:22
Control Test		0.078	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 04/02/2021 09:38

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/21/2021 03:13:32

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.080	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 04/02/2021 09:38

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: STRACK,ROBERT
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 03/24/2021 10:26:39

Diagnostics Check		OK	10:20
Air Blank		0.000	10:20
Control Test		0.081	10:20
Air Blank		0.000	10:21
Subject Sample	# 1	0.000	10:22
Breath Volume		1.394	10:22
Air Blank		0.000	10:22
Air Blank		0.000	10:24
Subject Sample	# 2	0.000	10:25
Breath Volume		1.363	10:25
Air Blank		0.000	10:25
Control Test		0.078	10:26
Air Blank		0.000	10:26
Diagnostics Check		OK	10:26

Data Download Date/Time: 04/02/2021 09:38

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 03/25/2021 01:59:14

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.081	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:58	
Control Test		0.081	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 04/02/2021 09:38

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CHATTERTON

DATE/TIME OF TEST: 03/25/2021 02:53:15

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.081	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.132	02:48
Breath Volume		1.773	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:50
Subject Sample	# 2	0.132	02:51
Breath Volume		1.488	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 04/02/2021 09:38

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: SENTEIO

DATE/TIME OF TEST: 03/26/2021 07:03:49

Diagnostics Check		OK	06:50	
Air Blank		0.000	06:51	
Control Test		0.080	06:51	
Air Blank		0.000	06:51	
Subject Sample	# 1	0.124	06:55	Volume Not Met
Breath Volume		0.636	06:55	
Air Blank		0.000	06:55	
Air Blank		0.000	06:57	
Subject Sample	# 2	0.124	06:58	
Breath Volume		1.164	06:58	
Air Blank		0.000	06:59	
Air Blank		0.000	07:01	
Subject Sample	# 3	0.116	07:02	
Breath Volume		1.125	07:02	
Air Blank		0.000	07:02	
Control Test		0.079	07:03	
Air Blank		0.000	07:03	
Diagnostics Check		OK	07:03	

Data Download Date/Time: 04/02/2021 09:38

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CAPELA,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/30/2021 00:15:41

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.153	00:10
Breath Volume		2.214	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.142	00:13
Breath Volume		2.496	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 04/02/2021 09:38

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFREY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:42

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 03/20/2021 09:17:47

Diagnostics Check		OK	09:06	
Air Blank		0.000	09:07	
Control Test		0.079	09:07	
Air Blank		0.000	09:08	
Subject Sample	# 1	0.162	09:11	Volume Not Met
Breath Volume		0.195	09:11	
Air Blank		0.000	09:12	
Air Blank		0.000	09:13	
Subject Sample	# 2		09:16	Subject Test Refused
Breath Volume		0.000	09:16	
Air Blank		0.000	09:16	
Control Test		0.079	09:17	
Air Blank		0.000	09:17	
Diagnostics Check		OK	09:17	

Data Download Date/Time: 03/31/2021 17:25

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: THOMPSON,C
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 03/06/2021 01:18:19

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.082	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.166	01:13
Breath Volume		2.953	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.168	01:16
Breath Volume		2.046	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ALVAREZ,NORGE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: FAJARDO HERNANDEZ

DATE/TIME OF TEST: 03/13/2021 19:50:37

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.081	19:45
Air Blank		0.000	19:45
Subject Sample	# 1	0.187	19:46
Breath Volume		2.761	19:46
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.194	19:49
Breath Volume		2.664	19:49
Air Blank		0.000	19:49
Control Test		0.079	19:50
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CASTILLO,CARLOS
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: NUNEZ PION

DATE/TIME OF TEST: 03/15/2021 01:50:11

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.082	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.085	01:45
Breath Volume		1.921	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.085	01:48
Breath Volume		2.976	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,JUAN,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JOSEPH,JEAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 03/15/2021 09:32:06

Diagnostics Check		OK	09:29	
Air Blank		0.000	09:29	
Control Test		0.081	09:30	
Air Blank		0.000	09:30	
Subject Sample	# 1		09:30	Subject Test Refused
Breath Volume		0.000	09:30	
Air Blank		0.000	09:31	
Control Test		0.081	09:31	
Air Blank		0.000	09:32	
Diagnostics Check		OK	09:32	

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPOTE,DUVAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CAPOTE,DUVAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 03/17/2021 01:24:50

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.000	01:20
Breath Volume		3.605	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.000	01:23
Breath Volume		2.574	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILNES,K,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MILNES,K
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 03/21/2021 03:47:25

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:45	
Control Test		0.081	03:45	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:46	
Control Test		0.082	03:46	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: WROY,W
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/21/2021 23:40:06

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.081	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.181	23:35
Breath Volume		2.609	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:37
Subject Sample	# 2	0.179	23:38
Breath Volume		2.214	23:38
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,CHARLES,E
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GIZ,JOAQUIN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: VARONA MONZON

DATE/TIME OF TEST: 03/25/2021 20:19:02

Diagnostics Check		OK	20:12
Air Blank		0.000	20:13
Control Test		0.081	20:13
Air Blank		0.000	20:13
Subject Sample	# 1	0.140	20:14
Breath Volume		2.332	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.137	20:17
Breath Volume		1.804	20:17
Air Blank		0.000	20:18
Control Test		0.081	20:18
Air Blank		0.000	20:18
Diagnostics Check		OK	20:19

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BASOA,J
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RIOS TELLEZ

DATE/TIME OF TEST: 03/26/2021 03:40:03

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.146	03:34
Breath Volume		2.089	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.138	03:38
Breath Volume		1.105	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:40

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,JOSE,M
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: RODRIGUEZ,NICHOLAS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: DETRINIDAD

DATE/TIME OF TEST: 03/27/2021 02:04:14

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.077	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.342	01:59
Breath Volume		2.328	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.327	02:02
Breath Volume		2.160	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBO,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SANCHEZ,CELESTE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: OLVERA

DATE/TIME OF TEST: 03/27/2021 18:56:08

Diagnostics Check		OK	18:48
Air Blank		0.000	18:49
Control Test		0.081	18:49
Air Blank		0.000	18:50
Subject Sample	# 1	0.243	18:51
Breath Volume		1.953	18:51
Air Blank		0.000	18:52
Air Blank		0.000	18:54
Subject Sample	# 2	0.243	18:54
Breath Volume		2.082	18:54
Air Blank		0.000	18:55
Control Test		0.080	18:55
Air Blank		0.000	18:56
Diagnostics Check		OK	18:56

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MARRERO,MICHAEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: COLOMBO

DATE/TIME OF TEST: 03/27/2021 19:20:37

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:18	
Control Test		0.081	19:18	
Air Blank		0.000	19:19	
Subject Sample	# 1		19:19	Subject Test Refused
Breath Volume		0.000	19:19	
Air Blank		0.000	19:19	
Control Test		0.081	19:20	
Air Blank		0.000	19:20	
Diagnostics Check		OK	19:20	

Data Download Date/Time: 04/05/2021 00:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: JARQUIN

DATE/TIME OF TEST: 03/26/2021 05:07:29

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:06	
Control Test	0.077	05:06	Interferent Detect
Air Blank	0.000	05:07	

Data Download Date/Time: 03/30/2021 15:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: JARQUIN

DATE/TIME OF TEST: 03/26/2021 05:15:45

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.078	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.153	05:11
Breath Volume		3.972	05:11
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.146	05:14
Breath Volume		3.132	05:14
Air Blank		0.000	05:14
Control Test		0.077	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 03/30/2021 15:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: CHRISTIE

DATE/TIME OF TEST: 03/03/2021 01:38:05

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.078	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.180	01:33
Breath Volume		3.054	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.177	01:36
Breath Volume		2.750	01:36
Air Blank		0.000	01:37
Control Test		0.076	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 03/16/2021 10:46

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEX,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEX
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: THURMAN

DATE/TIME OF TEST: 03/06/2021 03:51:03

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.078	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:50	
Control Test		0.078	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 03/16/2021 10:46

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 03/12/2021 00:52:35

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.078	00:47
Breath Volume		3.023	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.078	00:51
Breath Volume		3.097	00:51
Air Blank		0.000	00:51
Control Test		0.077	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 03/16/2021 10:46

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MCCUTCHEON

DATE/TIME OF TEST: 03/21/2021 01:22:51

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:17
Subject Sample	# 1	0.170	01:18
Breath Volume		3.375	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:20
Subject Sample	# 2	0.173	01:21
Breath Volume		2.339	01:21
Air Blank		0.000	01:21
Control Test		0.076	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 04/18/2021 07:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANHORN,MICHAEL,
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: VANHORN,MICHAEL
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 03/31/2021 21:19:16

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.078	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.287	21:14
Breath Volume		2.402	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:16
Subject Sample	# 2	0.271	21:17
Breath Volume		2.679	21:17
Air Blank		0.000	21:18
Control Test		0.077	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 04/18/2021 07:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SCHWARTZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: ALTAMIRANO

DATE/TIME OF TEST: 03/13/2021 05:51:18

Diagnostics Check		OK	05:44
Air Blank		0.000	05:44
Control Test		0.080	05:45
Air Blank		0.000	05:45
Subject Sample	# 1	0.138	05:46
Breath Volume		2.484	05:46
Air Blank		0.000	05:47
Air Blank		0.000	05:48
Subject Sample	# 2	0.131	05:49
Breath Volume		2.335	05:49
Air Blank		0.000	05:50
Control Test		0.079	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 04/07/2021 15:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLIER,JEREMEY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: TELLIER,JEREMEY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: JANNEY

DATE/TIME OF TEST: 03/13/2021 22:34:40

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:32	
Control Test		0.079	22:32	
Air Blank		0.000	22:33	
Subject Sample	# 1		22:33	Subject Test Refused
Breath Volume		0.000	22:33	
Air Blank		0.000	22:33	
Control Test		0.079	22:34	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 04/07/2021 15:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GUILLEMETTE

DATE/TIME OF TEST: 03/15/2021 04:33:14

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:30	
Control Test		0.079	04:31	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:32	
Control Test		0.079	04:32	
Air Blank		0.000	04:33	
Diagnostics Check		OK	04:33	

Data Download Date/Time: 04/07/2021 15:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATKINS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: KUNTZMANN

DATE/TIME OF TEST: 03/16/2021 03:48:36

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:46	
Control Test		0.080	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Control Test		0.079	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 04/07/2021 15:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: GARCIA RODRIGUEZ

DATE/TIME OF TEST: 03/18/2021 04:40:55

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:38	
Control Test		0.079	04:38	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:39	Subject Test Refused
Breath Volume		0.000	04:39	
Air Blank		0.000	04:40	
Control Test		0.078	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 04/07/2021 15:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SCHWARTZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: TOMCZUK

DATE/TIME OF TEST: 03/18/2021 07:22:23

Diagnostics Check		OK	07:18	
Air Blank		0.000	07:18	
Control Test		0.079	07:19	
Air Blank		0.000	07:19	
Subject Sample	# 1		07:21	Subject Test Refused
Breath Volume		0.000	07:21	
Air Blank		0.000	07:21	
Control Test		0.079	07:21	
Air Blank		0.000	07:22	
Diagnostics Check		OK	07:22	

Data Download Date/Time: 04/07/2021 16:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CALERO

DATE/TIME OF TEST: 03/19/2021 00:48:19

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.117	00:43
Breath Volume		2.777	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.120	00:46
Breath Volume		3.882	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 04/07/2021 16:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PEREZ VICENTE

DATE/TIME OF TEST: 03/20/2021 02:21:46

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:19	
Control Test		0.080	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Control Test		0.080	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 04/07/2021 16:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: BAR

DATE/TIME OF TEST: 03/22/2021 07:27:41

Diagnostics Check		OK	07:20
Air Blank		0.000	07:21
Control Test		0.081	07:21
Air Blank		0.000	07:21
Subject Sample	# 1	0.167	07:23
Breath Volume		2.050	07:23
Air Blank		0.000	07:23
Air Blank		0.000	07:25
Subject Sample	# 2	0.160	07:26
Breath Volume		1.437	07:26
Air Blank		0.000	07:26
Control Test		0.078	07:27
Air Blank		0.000	07:27
Diagnostics Check		OK	07:27

Data Download Date/Time: 04/07/2021 16:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 03/25/2021 03:11:30

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.101	03:06
Breath Volume		1.863	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.103	03:09
Breath Volume		2.828	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 04/07/2021 16:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KIEFER,RUSSELL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: SCHOONOVER

DATE/TIME OF TEST: 03/26/2021 00:03:48

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Control Test		0.080	00:03	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 04/07/2021 16:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEXANDER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 10:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MCAULEY

DATE/TIME OF TEST: 03/30/2021 00:43:33

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:39	No Sample Provided
Breath Volume		0.000	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Control Test		0.080	00:42	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 04/07/2021 16:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MERIDA

DATE/TIME OF TEST: 03/20/2021 01:09:48

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.218	01:05
Breath Volume		2.929	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:07
Subject Sample	# 2	0.209	01:08
Breath Volume		3.015	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 03/23/2021 12:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 03/25/2021 19:37:40

Diagnostics Check		OK	19:25
Air Blank		0.000	19:26
Control Test		0.079	19:26
Air Blank		0.000	19:26
Subject Sample	# 1	0.311	19:28
Breath Volume		1.925	19:28
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.284	19:33
Breath Volume		1.519	19:33
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 3	0.289	19:36
Breath Volume		1.464	19:36
Air Blank		0.000	19:36
Control Test		0.076	19:37
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 04/16/2021 11:49

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ULIN

DATE/TIME OF TEST: 03/03/2021 03:14:58

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.249	03:09
Breath Volume		1.804	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.259	03:13
Breath Volume		2.132	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 03/23/2021 12:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: BLOODSWORTH

DATE/TIME OF TEST: 03/29/2021 21:39:42

Diagnostics Check		OK	21:32
Air Blank		0.000	21:32
Control Test		0.081	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.117	21:34
Breath Volume		1.640	21:34
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.111	21:38
Breath Volume		1.941	21:38
Air Blank		0.000	21:38
Control Test		0.081	21:39
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 04/19/2021 08:33

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/04/2021 21:46:37

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.080	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.142	21:42
Breath Volume		1.207	21:42
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.145	21:45
Breath Volume		1.601	21:45
Air Blank		0.000	21:45
Control Test		0.079	21:45
Air Blank		0.000	21:46
Diagnostics Check		OK	21:46

Data Download Date/Time: 04/09/2021 12:02

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: THRONE

DATE/TIME OF TEST: 03/03/2021 01:41:49

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.078	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.108	01:37
Breath Volume		1.687	01:37
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.118	01:40
Breath Volume		2.507	01:40
Air Blank		0.000	01:40
Control Test		0.077	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 04/07/2021 19:35

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 03/06/2021 01:38:40

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.080	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.227	01:34
Breath Volume		2.503	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.234	01:36
Breath Volume		1.957	01:36
Air Blank		0.000	01:37
Control Test		0.077	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 04/07/2021 19:35

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAMDEN,D
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MCDANIELS

DATE/TIME OF TEST: 03/06/2021 02:56:36

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.078	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.065	02:52
Breath Volume		1.878	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.070	02:55
Breath Volume		1.968	02:55
Air Blank		0.000	02:55
Control Test		0.078	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 04/07/2021 19:35

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LEROUX,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: LUMAJ

DATE/TIME OF TEST: 03/10/2021 19:15:36

Diagnostics Check		OK	19:09
Air Blank		0.000	19:09
Control Test		0.079	19:10
Air Blank		0.000	19:10
Subject Sample	# 1	0.000	19:11
Breath Volume		2.664	19:11
Air Blank		0.000	19:11
Air Blank		0.000	19:13
Subject Sample	# 2	0.000	19:14
Breath Volume		1.433	19:14
Air Blank		0.000	19:14
Control Test		0.078	19:15
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 04/07/2021 19:35

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PELAGGI

DATE/TIME OF TEST: 03/11/2021 02:58:56

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.078	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.103	02:54
Breath Volume		2.722	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:56
Subject Sample	# 2	0.099	02:57
Breath Volume		2.550	02:57
Air Blank		0.000	02:58
Control Test		0.079	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 03/16/2021 03:12:11

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.023	03:07
Breath Volume		2.273	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.022	03:10
Breath Volume		1.640	03:10
Air Blank		0.000	03:11
Control Test		0.076	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DALLAIRE

DATE/TIME OF TEST: 03/20/2021 01:32:27

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.066	01:27
Breath Volume		3.308	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.068	01:30
Breath Volume		4.261	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 03/21/2021 02:22:07

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.110	02:17
Breath Volume		2.828	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:19
Subject Sample	# 2	0.110	02:20
Breath Volume		2.847	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: CADEK

DATE/TIME OF TEST: 03/21/2021 21:37:23

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.077	21:32
Air Blank		0.000	21:32
Subject Sample	# 1	0.316	21:32
Breath Volume		1.656	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.306	21:35
Breath Volume		1.476	21:35
Air Blank		0.000	21:36
Control Test		0.077	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MANSHIP

DATE/TIME OF TEST: 03/21/2021 22:17:01

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.152	22:12
Breath Volume		2.214	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:14
Subject Sample	# 2	0.140	22:15
Breath Volume		1.605	22:15
Air Blank		0.000	22:16
Control Test		0.077	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:17

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/22/2021 00:27:20

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.078	00:22	
Air Blank		0.000	00:22	
Subject Sample	# 1	0.142	00:22	
Breath Volume		2.378	00:22	
Air Blank		0.000	00:23	
Air Blank		0.000	00:25	
Subject Sample	# 2	0.170	00:26	
Breath Volume		1.496	00:26	
Air Blank			00:26	Purge Fail
Air Blank		0.000	00:27	

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 03/25/2021 23:11:43

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.077	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.014	23:06
Breath Volume		1.886	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.013	23:10
Breath Volume		3.640	23:10
Air Blank		0.000	23:10
Control Test		0.075	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PRAMAGIOULIS

DATE/TIME OF TEST: 03/26/2021 03:21:46

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.098	03:17
Breath Volume		2.406	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.094	03:20
Breath Volume		2.109	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 04/07/2021 19:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 03/10/2021 21:52:46

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.081	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.000	21:48
Breath Volume		1.421	21:48
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.000	21:51
Breath Volume		1.355	21:51
Air Blank		0.000	21:51
Control Test		0.080	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 04/15/2021 19:29

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 03/14/2021 03:41:22

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.114	03:36
Breath Volume		2.757	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.105	03:39
Breath Volume		1.710	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 04/15/2021 19:29

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/19/2021 01:11:27

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.081	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.193	01:06
Breath Volume		2.027	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:08
Subject Sample	# 2	0.210	01:09
Breath Volume		2.121	01:09
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 04/15/2021 19:29

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 03/21/2021 02:02:57

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.093	01:57
Breath Volume		1.703	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.097	02:01
Breath Volume		2.082	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 04/15/2021 19:29

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: GRIFFIS

DATE/TIME OF TEST: 03/21/2021 04:46:37

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.127	04:41
Breath Volume		1.937	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.121	04:44
Breath Volume		1.800	04:44
Air Blank		0.000	04:45
Control Test		0.079	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 04/15/2021 19:29

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,G
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 03/05/2021 02:23:03

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.078	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1	0.247	02:21	Volume Not Met
Breath Volume		0.167	02:21	
Air Blank			02:22	Purge Fail
Air Blank		0.000	02:23	

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FRANCO,G
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 03/05/2021 02:33:59

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.076	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.269	02:29
Breath Volume		1.574	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:31
Subject Sample	# 2	0.262	02:32
Breath Volume		1.421	02:32
Air Blank		0.000	02:33
Control Test		0.077	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/06/2021 04:57:06

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.068	04:52
Breath Volume		1.488	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:54
Subject Sample	# 2	0.070	04:55
Breath Volume		2.953	04:55
Air Blank		0.000	04:56
Control Test		0.078	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ESPINOSA MARIN

DATE/TIME OF TEST: 03/07/2021 02:28:31

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.173	02:21	Volume Not Met
Breath Volume		0.792	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2	0.183	02:26	Volume Not Met
Breath Volume		0.746	02:26	
Air Blank		0.000	02:27	
Control Test		0.077	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ESPINOZA MARIN

DATE/TIME OF TEST: 03/07/2021 02:41:37

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:33	
Control Test		0.078	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.168	02:37	Volume Not Met
Breath Volume		0.230	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.078	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/07/2021 06:24:18

Diagnostics Check		OK	06:17
Air Blank		0.000	06:17
Control Test		0.079	06:18
Air Blank		0.000	06:18
Subject Sample	# 1	0.100	06:19
Breath Volume		2.746	06:19
Air Blank		0.000	06:20
Air Blank		0.000	06:21
Subject Sample	# 2	0.096	06:22
Breath Volume		2.878	06:22
Air Blank		0.000	06:23
Control Test		0.079	06:23
Air Blank		0.000	06:24
Diagnostics Check		OK	06:24

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PEREZ,Y
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 03/07/2021 06:52:13

Diagnostics Check		OK	06:46	
Air Blank		0.000	06:47	
Control Test		0.079	06:47	
Air Blank		0.000	06:48	
Subject Sample	# 1		06:50	Subject Test Refused
Breath Volume		0.000	06:50	
Air Blank		0.000	06:51	
Control Test		0.077	06:51	
Air Blank		0.000	06:52	
Diagnostics Check		OK	06:52	

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: GUZMAN CACERES

DATE/TIME OF TEST: 03/20/2021 05:25:27

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.179	05:20
Breath Volume		1.531	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.181	05:23
Breath Volume		3.406	05:23
Air Blank		0.000	05:24
Control Test		0.079	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 03/27/2021 03:30

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/28/2021 02:48:37

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.078	02:46	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:47	
Control Test		0.077	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 04/26/2021 17:09

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROBERTS JR.

DATE/TIME OF TEST: 03/03/2021 01:42:14

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.129	01:37
Breath Volume		3.195	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.130	01:40
Breath Volume		2.847	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 03/03/2021 11:06

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: LANUM

DATE/TIME OF TEST: 03/04/2021 18:01:24

Diagnostics Check		OK	17:58	
Air Blank		0.000	17:58	
Control Test		0.079	17:58	
Air Blank		0.000	17:59	
Subject Sample	# 1	0.000	17:59	
Breath Volume		2.976	17:59	
Air Blank			18:00	RFI Detect
Air Blank		0.000	18:00	
Control Test		0.079	18:00	
Air Blank		0.000	18:01	
Diagnostics Check		OK	18:01	

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: LANUM

DATE/TIME OF TEST: 03/04/2021 18:06:22

Diagnostics Check	OK	18:04	
Air Blank	0.000	18:05	
Control Test	0.078	18:05	
Air Blank		18:05	RFI Detect
Air Blank	0.000	18:06	

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: LANUM

DATE/TIME OF TEST: 03/04/2021 18:15:55

Diagnostics Check		OK	18:09
Air Blank		0.000	18:10
Control Test		0.078	18:10
Air Blank		0.000	18:10
Subject Sample	# 1	0.000	18:11
Breath Volume		2.078	18:11
Air Blank		0.000	18:12
Air Blank		0.000	18:14
Subject Sample	# 2	0.000	18:14
Breath Volume		1.429	18:14
Air Blank		0.000	18:14
Control Test		0.078	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANTOINE,KAHLYL
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/05/2021 02:34:34

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:32	
Control Test		0.079	02:32	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:33	RFI Detect
Breath Volume		0.000	02:33	
Air Blank		0.000	02:33	
Control Test		0.078	02:33	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANTOINE,KAHLYL
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/05/2021 02:50:09

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:37	
Control Test		0.078	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1	0.126	02:39	
Breath Volume		1.628	02:39	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 2		02:44	No Sample Provided
Breath Volume		0.000	02:44	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 3	0.116	02:48	
Breath Volume		1.496	02:48	
Air Blank		0.000	02:49	
Control Test		0.079	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: AUGUSMAR

DATE/TIME OF TEST: 03/05/2021 03:42:00

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.078	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.000	03:37
Breath Volume		1.761	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.000	03:40
Breath Volume		1.878	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:42

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BITTLER

DATE/TIME OF TEST: 03/05/2021 05:04:01

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:03	
Control Test	0.079	05:03	
Air Blank		05:03	RFI Detect
Air Blank		05:04	RFI Detect

Data Download Date/Time: 04/07/2021 11:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BITTLER

DATE/TIME OF TEST: 03/05/2021 05:12:56

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.077	05:06
Air Blank		0.000	05:07
Subject Sample	# 1	0.010	05:08
Breath Volume		1.601	05:08
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.010	05:11
Breath Volume		1.203	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: DEVIEUX

DATE/TIME OF TEST: 03/07/2021 05:09:14

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.092	05:04
Breath Volume		2.160	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.090	05:07
Breath Volume		2.496	05:07
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,N
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: SIMENTAL

DATE/TIME OF TEST: 03/13/2021 07:32:23

Diagnostics Check		OK	07:25
Air Blank		0.000	07:26
Control Test		0.079	07:26
Air Blank		0.000	07:27
Subject Sample	# 1	0.139	07:27
Breath Volume		2.761	07:27
Air Blank		0.000	07:28
Air Blank		0.000	07:30
Subject Sample	# 2	0.138	07:30
Breath Volume		2.832	07:30
Air Blank		0.000	07:31
Control Test		0.079	07:31
Air Blank		0.000	07:32
Diagnostics Check		OK	07:32

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 03/14/2021 00:40:16

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.132	00:35
Breath Volume		1.480	00:35
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.132	00:38
Breath Volume		1.714	00:38
Air Blank		0.000	00:39
Control Test		0.077	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/14/2021 05:53:12

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.078	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.183	05:48
Breath Volume		1.609	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.178	05:51
Breath Volume		1.183	05:51
Air Blank		0.000	05:52
Control Test		0.079	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 04/07/2021 11:16

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPIC,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: COX,G
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: SHERROD

DATE/TIME OF TEST: 03/14/2021 06:17:49

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.077	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.238	06:13
Breath Volume		1.476	06:13
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.244	06:16
Breath Volume		1.667	06:16
Air Blank		0.000	06:16
Control Test		0.076	06:17
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: CALITO VASQUEZ

DATE/TIME OF TEST: 03/14/2021 21:57:46

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.078	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.211	21:53
Breath Volume		1.281	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.207	21:56
Breath Volume		2.121	21:56
Air Blank		0.000	21:56
Control Test		0.078	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,JASIEL
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: CRUZ ORELLANA

DATE/TIME OF TEST: 03/15/2021 05:07:24

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.078	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.189	05:02
Breath Volume		1.734	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:04
Subject Sample	# 2	0.191	05:05
Breath Volume		2.070	05:05
Air Blank		0.000	05:06
Control Test		0.077	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL,T'KEYAH
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BROZIC

DATE/TIME OF TEST: 03/20/2021 00:18:19

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.079	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.252	00:13
Breath Volume		2.167	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.254	00:16
Breath Volume		2.003	00:16
Air Blank		0.000	00:17
Control Test		0.077	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: PEREZ PAZ

DATE/TIME OF TEST: 03/20/2021 01:14:54

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.079	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.078	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SOTO CARDONA

DATE/TIME OF TEST: 03/20/2021 05:08:36

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.078	05:02
Air Blank		0.000	05:02
Subject Sample	# 1	0.176	05:03
Breath Volume		2.472	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:05
Subject Sample	# 2	0.175	05:06
Breath Volume		2.312	05:06
Air Blank		0.000	05:07
Control Test		0.078	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: CANALES SALGADO

DATE/TIME OF TEST: 03/20/2021 05:59:08

Diagnostics Check		OK	05:52
Air Blank		0.000	05:52
Control Test		0.079	05:53
Air Blank		0.000	05:53
Subject Sample	# 1	0.184	05:54
Breath Volume		1.640	05:54
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.182	05:57
Breath Volume		1.113	05:57
Air Blank		0.000	05:58
Control Test		0.077	05:58
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SANTAMARIA,DOMINIQUE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/28/2021 20:14:43

Diagnostics Check		OK	20:02	
Air Blank		0.000	20:03	
Control Test		0.078	20:03	
Air Blank		0.000	20:04	
Subject Sample	# 1	0.172	20:07	Volume Not Met
Breath Volume		0.183	20:07	
Air Blank		0.000	20:08	
Air Blank		0.000	20:09	
Subject Sample	# 2	0.159	20:12	Volume Not Met
Breath Volume		0.171	20:12	
Air Blank		0.000	20:13	
Control Test		0.076	20:14	
Air Blank		0.000	20:14	
Diagnostics Check		OK	20:14	

Data Download Date/Time: 04/07/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SWENSON,BRIAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 03/06/2021 06:31:06

Diagnostics Check		OK	06:25	
Air Blank		0.000	06:26	
Control Test		0.081	06:26	
Air Blank		0.000	06:27	
Subject Sample	# 1		06:29	Subject Test Refused
Breath Volume		0.000	06:29	
Air Blank		0.000	06:30	
Control Test		0.081	06:30	
Air Blank		0.000	06:31	
Diagnostics Check		OK	06:31	

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILSON,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/08/2021 23:10:30

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.083	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.172	23:05
Breath Volume		1.738	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:07
Subject Sample	# 2	0.177	23:08
Breath Volume		2.695	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/10/2021 19:24:25

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:22	
Control Test		0.081	19:22	
Air Blank		0.000	19:22	
Subject Sample	# 1		19:23	Subject Test Refused
Breath Volume		0.000	19:23	
Air Blank		0.000	19:23	
Control Test		0.081	19:23	
Air Blank		0.000	19:24	
Diagnostics Check		OK	19:24	

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CARTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 11:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: MERAZ NAJERA

DATE/TIME OF TEST: 03/13/2021 12:13:00

Diagnostics Check		OK	12:05
Air Blank		0.000	12:06
Control Test		0.081	12:06
Air Blank		0.000	12:07
Subject Sample	# 1	0.319	12:08
Breath Volume		1.914	12:08
Air Blank		0.000	12:09
Air Blank		0.000	12:10
Subject Sample	# 2	0.321	12:11
Breath Volume		1.671	12:11
Air Blank		0.000	12:12
Control Test		0.080	12:12
Air Blank		0.000	12:12
Diagnostics Check		OK	12:13

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CALLOWAY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/13/2021 23:45:47

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.082	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:44	Subject Test Refused
Breath Volume		0.000	23:44	
Air Blank		0.000	23:44	
Control Test		0.082	23:45	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 03/14/2021 00:38:00

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.082	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:31	No Sample Provided
Breath Volume		0.000	00:31	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 2		00:36	No Sample Provided
Breath Volume		0.000	00:36	
Air Blank		0.000	00:37	
Control Test		0.078	00:37	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 03/14/2021 01:08:10

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.199	01:02
Breath Volume		2.027	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.192	01:06
Breath Volume		1.597	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TWIDDY,RYAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: LAROCCA

DATE/TIME OF TEST: 03/15/2021 14:28:50

Diagnostics Check		OK	14:22	
Air Blank		0.000	14:22	
Control Test		0.081	14:23	
Air Blank		0.000	14:23	
Subject Sample	# 1	0.222	14:24	
Breath Volume		1.609	14:24	
Air Blank		0.000	14:25	
Air Blank		0.000	14:26	
Subject Sample	# 2	0.227	14:28	
Breath Volume		1.761	14:28	
Air Blank			14:28	RFI Detect
Air Blank		0.000	14:28	

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENBURG,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TWIDDY,RYAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: LAROCCA

DATE/TIME OF TEST: 03/15/2021 14:38:18

Diagnostics Check		OK	14:31
Air Blank		0.000	14:32
Control Test		0.079	14:32
Air Blank		0.000	14:32
Subject Sample	# 1	0.225	14:33
Breath Volume		1.546	14:33
Air Blank		0.000	14:34
Air Blank		0.000	14:35
Subject Sample	# 2	0.221	14:36
Breath Volume		1.636	14:36
Air Blank		0.000	14:37
Control Test		0.080	14:37
Air Blank		0.000	14:38
Diagnostics Check		OK	14:38

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: W325117912411

DATE/TIME OF TEST: 03/21/2021 03:45:03

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.149	03:40
Breath Volume		4.730	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.150	03:43
Breath Volume		4.832	03:43
Air Blank		0.000	03:44
Control Test		0.082	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:45

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CALLOWAY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/23/2021 00:50:24

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.081	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:49	Subject Test Refused
Breath Volume		0.000	00:49	
Air Blank		0.000	00:49	
Control Test		0.081	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOHLIN,MARK,T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CALLOWAY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: KEARNEY

DATE/TIME OF TEST: 03/25/2021 01:49:32

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.082	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.041	01:44
Breath Volume		2.152	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.039	01:47
Breath Volume		3.179	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 04/19/2021 04:26

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 03/11/2021 00:14:56

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:12	
Control Test		0.078	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 1		00:13	Subject Test Refused
Breath Volume		0.000	00:13	
Air Blank		0.000	00:14	
Control Test		0.079	00:14	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:14	

Data Download Date/Time: 04/08/2021 01:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ENOS

DATE/TIME OF TEST: 03/18/2021 00:57:12

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.152	00:52
Breath Volume		2.339	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.153	00:55
Breath Volume		2.234	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 04/08/2021 01:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/18/2021 03:49:05

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.261	03:44
Breath Volume		3.796	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.245	03:47
Breath Volume		3.570	03:47
Air Blank		0.000	03:48
Control Test		0.077	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 04/08/2021 01:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHM,JONATHAN,H
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DESCARTES

DATE/TIME OF TEST: 03/20/2021 00:41:54

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.079	00:38	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:40	
Control Test		0.080	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 04/08/2021 01:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ICENHOWER

DATE/TIME OF TEST: 03/24/2021 01:50:47

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.078	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.111	01:45
Breath Volume		4.007	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.107	01:49
Breath Volume		4.718	01:49
Air Blank		0.000	01:49
Control Test		0.078	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 04/08/2021 01:00

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/28/2021 18:37:56

Diagnostics Check		OK	18:34	
Air Blank		0.000	18:35	
Control Test		0.078	18:35	
Air Blank		0.000	18:36	
Subject Sample	# 1		18:36	Subject Test Refused
Breath Volume		0.000	18:36	
Air Blank		0.000	18:37	
Control Test		0.077	18:37	
Air Blank		0.000	18:37	
Diagnostics Check		OK	18:37	

Data Download Date/Time: 04/08/2021 00:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,C A,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WHITE,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 03/15/2021 01:40:14

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.078	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.145	01:35
Breath Volume		1.710	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.156	01:38
Breath Volume		1.875	01:38
Air Blank		0.000	01:39
Control Test		0.077	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 03/19/2021 16:53

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/02/2021 23:43:35

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.191	23:38
Breath Volume		2.199	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.187	23:41
Breath Volume		1.445	23:41
Air Blank		0.000	23:42
Control Test		0.077	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ERICKSON,CHRISTINA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/03/2021 04:16:03

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.165	04:11
Breath Volume		2.234	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.161	04:14
Breath Volume		2.074	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:16

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 03/06/2021 01:44:32

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.174	01:39
Breath Volume		1.988	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.180	01:42
Breath Volume		2.519	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEE,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MULDOON

DATE/TIME OF TEST: 03/07/2021 22:42:17

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.080	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.178	22:37
Breath Volume		2.976	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:39
Subject Sample	# 2	0.178	22:40
Breath Volume		3.257	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DE LEON,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 03/08/2021 22:10:46

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.080	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.000	22:06
Breath Volume		4.585	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.000	22:09
Breath Volume		4.542	22:09
Air Blank		0.000	22:09
Control Test		0.080	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DOPPLERBAUER,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 03/10/2021 23:23:08

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.110	23:18
Breath Volume		3.371	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.113	23:21
Breath Volume		2.949	23:21
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ERICKSON,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: HENNESSY

DATE/TIME OF TEST: 03/11/2021 02:13:47

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.134	02:08
Breath Volume		2.136	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.131	02:12
Breath Volume		1.906	02:12
Air Blank		0.000	02:12
Control Test		0.078	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BECKNER

DATE/TIME OF TEST: 03/11/2021 04:42:12

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.079	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.064	04:37
Breath Volume		1.601	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:39
Subject Sample	# 2	0.061	04:40
Breath Volume		1.503	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EARNER,JACOB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: FERRONE

DATE/TIME OF TEST: 03/12/2021 01:11:19

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.081	01:06
Breath Volume		1.335	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.082	01:09
Breath Volume		1.398	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,BRANDON
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 03/15/2021 00:24:46

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.079	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.366	00:19
Breath Volume		1.968	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.348	00:22
Breath Volume		1.832	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMPBELL,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TUSET

DATE/TIME OF TEST: 03/15/2021 01:17:28

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.128	01:12
Breath Volume		4.808	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.127	01:15
Breath Volume		3.375	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KLEIN,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: TELFORT

DATE/TIME OF TEST: 03/16/2021 02:48:59

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.023	02:44
Breath Volume		2.171	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.023	02:47
Breath Volume		2.644	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STEYSKAL,T
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 03/16/2021 13:52:07

Diagnostics Check		OK	13:49	
Air Blank		0.000	13:49	
Control Test		0.079	13:49	
Air Blank		0.000	13:50	
Subject Sample	# 1		13:50	Subject Test Refused
Breath Volume		0.000	13:50	
Air Blank		0.000	13:51	
Control Test		0.079	13:51	
Air Blank		0.000	13:52	
Diagnostics Check		OK	13:52	

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DELEON,ALEXANDRIA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/17/2021 19:17:17

Diagnostics Check		OK	19:11
Air Blank		0.000	19:11
Control Test		0.079	19:12
Air Blank		0.000	19:12
Subject Sample	# 1	0.089	19:12
Breath Volume		3.269	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.087	19:15
Breath Volume		3.152	19:15
Air Blank		0.000	19:16
Control Test		0.078	19:16
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,JACOB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: D'AGOSTINO

DATE/TIME OF TEST: 03/17/2021 22:31:29

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.079	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.102	22:27
Breath Volume		2.046	22:27
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.100	22:29
Breath Volume		1.585	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DE LEON,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/18/2021 03:13:51

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.000	03:09
Breath Volume		4.750	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.000	03:12
Breath Volume		4.683	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RUSSELL,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/18/2021 12:52:11

Diagnostics Check		OK	12:45
Air Blank		0.000	12:46
Control Test		0.079	12:46
Air Blank		0.000	12:46
Subject Sample	# 1	0.204	12:47
Breath Volume		5.042	12:47
Air Blank		0.000	12:48
Air Blank		0.000	12:49
Subject Sample	# 2	0.207	12:50
Breath Volume		5.285	12:50
Air Blank		0.000	12:51
Control Test		0.078	12:51
Air Blank		0.000	12:52
Diagnostics Check		OK	12:52

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RUSSELL,OFFICER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/18/2021 16:10:09

Diagnostics Check		OK	16:02
Air Blank		0.000	16:03
Control Test		0.079	16:03
Air Blank		0.000	16:04
Subject Sample	# 1	0.116	16:05
Breath Volume		5.199	16:05
Air Blank		0.000	16:05
Air Blank		0.000	16:07
Subject Sample	# 2	0.113	16:08
Breath Volume		5.175	16:08
Air Blank		0.000	16:09
Control Test		0.079	16:09
Air Blank		0.000	16:10
Diagnostics Check		OK	16:10

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CRADER

DATE/TIME OF TEST: 03/18/2021 17:05:31

Diagnostics Check		OK	16:58
Air Blank		0.000	16:59
Control Test		0.077	16:59
Air Blank		0.000	16:59
Subject Sample	# 1	0.217	17:00
Breath Volume		1.160	17:00
Air Blank		0.000	17:01
Air Blank		0.000	17:03
Subject Sample	# 2	0.229	17:03
Breath Volume		1.437	17:03
Air Blank		0.000	17:04
Control Test		0.079	17:04
Air Blank		0.000	17:05
Diagnostics Check		OK	17:05

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KLEIVER,N/A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 12:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/18/2021 19:54:27

Diagnostics Check		OK	19:48
Air Blank		0.000	19:48
Control Test		0.079	19:48
Air Blank		0.000	19:49
Subject Sample	# 1	0.020	19:49
Breath Volume		4.953	19:49
Air Blank		0.000	19:50
Air Blank		0.000	19:52
Subject Sample	# 2	0.019	19:52
Breath Volume		4.210	19:52
Air Blank		0.000	19:53
Control Test		0.079	19:53
Air Blank		0.000	19:54
Diagnostics Check		OK	19:54

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRAY,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: CRADER

DATE/TIME OF TEST: 03/19/2021 01:45:59

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.068	01:40
Breath Volume		1.109	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.068	01:44
Breath Volume		1.304	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK',R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRAY,AMBER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: CRADER

DATE/TIME OF TEST: 03/19/2021 03:12:54

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.043	03:08
Breath Volume		1.871	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.044	03:11
Breath Volume		1.878	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRAY,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: CRADER

DATE/TIME OF TEST: 03/19/2021 04:13:14

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.028	04:08
Breath Volume		1.332	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.028	04:11
Breath Volume		1.757	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CIOLINO,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 03/25/2021 21:38:37

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:32
Subject Sample	# 1	0.209	21:33
Breath Volume		2.574	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.210	21:37
Breath Volume		2.738	21:37
Air Blank		0.000	21:37
Control Test		0.078	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 03/25/2021 22:54:35

Diagnostics Check		OK	22:47
Air Blank		0.000	22:47
Control Test		0.079	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.193	22:49
Breath Volume		1.515	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.193	22:52
Breath Volume		1.984	22:52
Air Blank		0.000	22:53
Control Test		0.078	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/28/2021 03:50:38

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.230	03:46
Breath Volume		1.691	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.229	03:49
Breath Volume		2.292	03:49
Air Blank		0.000	03:49
Control Test		0.079	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LOPEZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 03/28/2021 16:51:14

Diagnostics Check		OK	16:36	
Air Blank		0.000	16:36	
Control Test		0.079	16:37	
Air Blank		0.000	16:37	
Subject Sample	# 1	0.085	16:38	
Breath Volume		1.742	16:38	
Air Blank		0.000	16:39	
Air Blank		0.000	16:40	
Subject Sample	# 2	0.075	16:44	Volume Not Met
Breath Volume		0.468	16:44	
Air Blank		0.000	16:44	
Air Blank		0.000	16:46	
Subject Sample	# 3	0.064	16:49	Volume Not Met
Breath Volume		0.050	16:49	
Air Blank		0.000	16:50	
Control Test		0.078	16:50	
Air Blank		0.000	16:51	
Diagnostics Check		OK	16:51	

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OKEEFE,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: NUCERO

DATE/TIME OF TEST: 03/29/2021 15:38:23

Diagnostics Check		OK	15:26
Air Blank		0.000	15:26
Control Test		0.079	15:26
Air Blank		0.000	15:27
Subject Sample	# 1	0.282	15:29
Breath Volume		2.390	15:29
Air Blank		0.000	15:29
Air Blank		0.000	15:31
Subject Sample	# 2	0.260	15:33
Breath Volume		1.167	15:33
Air Blank		0.000	15:34
Air Blank		0.000	15:35
Subject Sample	# 3	0.274	15:36
Breath Volume		1.972	15:36
Air Blank		0.000	15:37
Control Test		0.076	15:37
Air Blank		0.000	15:38
Diagnostics Check		OK	15:38

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SKOLNICK,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/30/2021 22:40:27

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.080	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.162	22:35
Breath Volume		2.761	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.166	22:38
Breath Volume		2.785	22:38
Air Blank		0.000	22:39
Control Test		0.079	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/31/2021 04:02:55

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.159	03:58
Breath Volume		5.468	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.156	04:01
Breath Volume		4.593	04:01
Air Blank		0.000	04:01
Control Test		0.079	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RICE,LACEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 03/31/2021 06:03:46

Diagnostics Check		OK	05:56
Air Blank		0.000	05:57
Control Test		0.079	05:57
Air Blank		0.000	05:58
Subject Sample	# 1	0.186	05:59
Breath Volume		1.878	05:59
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.184	06:02
Breath Volume		1.687	06:02
Air Blank		0.000	06:02
Control Test		0.078	06:03
Air Blank		0.000	06:03
Diagnostics Check		OK	06:03

Data Download Date/Time: 03/31/2021 06:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: SOTOLONGO

DATE/TIME OF TEST: 03/02/2021 11:12:24

Diagnostics Check	OK	11:11	
Air Blank		11:11	RFI Detect
Air Blank	0.000	11:12	

Data Download Date/Time: 03/27/2021 04:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUTIERREZ,K
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: SOTOLONGO

DATE/TIME OF TEST: 03/02/2021 11:17:21

Diagnostics Check		OK	11:13	
Air Blank		0.000	11:14	
Control Test		0.080	11:14	
Air Blank		0.000	11:15	
Subject Sample	# 1		11:15	Slope Not Met
Breath Volume		1.746	11:15	
Air Blank		0.000	11:16	
Control Test		0.079	11:16	
Air Blank		0.000	11:17	
Diagnostics Check		OK	11:17	

Data Download Date/Time: 03/27/2021 04:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: SOTOLONGO

DATE/TIME OF TEST: 03/02/2021 11:25:17

Diagnostics Check		OK	11:18
Air Blank		0.000	11:19
Control Test		0.080	11:19
Air Blank		0.000	11:20
Subject Sample	# 1	0.341	11:20
Breath Volume		2.554	11:20
Air Blank		0.000	11:21
Air Blank		0.000	11:23
Subject Sample	# 2	0.332	11:23
Breath Volume		2.261	11:23
Air Blank		0.000	11:24
Control Test		0.078	11:24
Air Blank		0.000	11:25
Diagnostics Check		OK	11:25

Data Download Date/Time: 03/27/2021 04:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 03/12/2021 03:15:14

Diagnostics Check	OK	03:14	
Air Blank		03:15	RFI Detect
Air Blank		03:15	RFI Detect

Data Download Date/Time: 03/27/2021 04:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 03/12/2021 03:19:57

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.101	03:19	Control Outside Tolerance
Air Blank	0.000	03:19	

Data Download Date/Time: 03/27/2021 04:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 03/12/2021 03:27:32

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.116	03:26	Control Outside Tolerance
Air Blank	0.000	03:27	

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 03/12/2021 03:35:11

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.103	03:34	Interferent Detect
Air Blank	0.000	03:35	

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 03/12/2021 03:40:18

Diagnostics Check	OK	03:39	
Air Blank		03:39	Ambient Fail
Air Blank	0.000	03:40	

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: GUADAMUS

DATE/TIME OF TEST: 03/12/2021 03:45:19

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.099	03:44	Control Outside Tolerance
Air Blank	0.000	03:45	

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GUADAMUZ

DATE/TIME OF TEST: 03/12/2021 04:55:20

Diagnostics Check		OK	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:49
Subject Sample	# 1	0.087	04:50
Breath Volume		2.375	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.091	04:53
Breath Volume		2.000	04:53
Air Blank		0.000	04:54
Control Test		0.076	04:54
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OCNER,W
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 03/13/2021 20:12:30

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.080	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.155	20:07
Breath Volume		1.617	20:07
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.153	20:10
Breath Volume		2.082	20:10
Air Blank		0.000	20:11
Control Test		0.080	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCEGUERA,YASMANI,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OCEGUERA,YASMANI
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: LINK

DATE/TIME OF TEST: 03/15/2021 19:25:24

Diagnostics Check		OK	19:22	
Air Blank		0.000	19:22	
Control Test		0.078	19:23	
Air Blank		0.000	19:23	
Subject Sample	# 1		19:24	Subject Test Refused
Breath Volume		0.000	19:24	
Air Blank		0.000	19:24	
Control Test		0.078	19:24	
Air Blank		0.000	19:25	
Diagnostics Check		OK	19:25	

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCEGUERA,YASMANI,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CHAVARRIA,JOANNA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 03/21/2021 18:57:57

Diagnostics Check		OK	18:52	
Air Blank		0.000	18:52	
Control Test		0.079	18:52	
Air Blank		0.000	18:53	
Subject Sample	# 1		18:56	Subject Test Refused
Breath Volume		0.000	18:56	
Air Blank		0.000	18:57	
Control Test		0.080	18:57	
Air Blank		0.000	18:57	
Diagnostics Check		OK	18:57	

Data Download Date/Time: 03/27/2021 04:08

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDA,ROBERT,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LAYPORT,JOHN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/08/2021 14:45:36

Diagnostics Check		OK	14:37
Air Blank		0.000	14:37
Control Test		0.078	14:37
Air Blank		0.000	14:38
Subject Sample	# 1	0.000	14:40
Breath Volume		1.363	14:40
Air Blank		0.000	14:41
Air Blank		0.000	14:43
Subject Sample	# 2	0.000	14:44
Breath Volume		1.863	14:44
Air Blank		0.000	14:44
Control Test		0.077	14:44
Air Blank		0.000	14:45
Diagnostics Check		OK	14:45

Data Download Date/Time: 04/02/2021 10:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JESSICA,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ADLER,JAMES
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 181632
DRY GAS EXPIRY DATE: 03/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 03/20/2021 07:23:58

Diagnostics Check		OK	07:17
Air Blank		0.000	07:17
Control Test		0.078	07:18
Air Blank		0.000	07:18
Subject Sample	# 1	0.069	07:19
Breath Volume		2.699	07:19
Air Blank		0.000	07:19
Air Blank		0.000	07:21
Subject Sample	# 2	0.073	07:22
Breath Volume		3.382	07:22
Air Blank		0.000	07:23
Control Test		0.078	07:23
Air Blank		0.000	07:23
Diagnostics Check		OK	07:23

Data Download Date/Time: 04/02/2021 10:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PEREZ BETANCOURT

DATE/TIME OF TEST: 03/01/2021 02:32:53

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:26	
Control Test		0.078	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1	0.172	02:28	
Breath Volume		1.800	02:28	
Air Blank		0.000	02:28	
Air Blank		0.000	02:30	
Subject Sample	# 2	0.175	02:31	
Breath Volume		1.789	02:31	
Air Blank		0.000	02:32	
Control Test		0.000	02:32	Control Outside Tolerance
Air Blank		0.000	02:32	

Data Download Date/Time: 03/09/2021 15:06

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CAMILO

DATE/TIME OF TEST: 03/04/2021 01:24:41

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.080	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.080	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 03/09/2021 15:06

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: GROSSO

DATE/TIME OF TEST: 03/04/2021 04:57:14

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.157	04:52
Breath Volume		1.421	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.157	04:55
Breath Volume		1.855	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 03/09/2021 15:06

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHINSON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/08/2021 22:09:30

Diagnostics Check		OK	22:02
Air Blank		0.000	22:03
Control Test		0.081	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.218	22:04
Breath Volume		3.171	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.221	22:07
Breath Volume		2.777	22:07
Air Blank		0.000	22:08
Control Test		0.081	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 03/09/2021 15:06

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: CRENSHAW

DATE/TIME OF TEST: 03/08/2021 23:53:23

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.209	23:48
Breath Volume		2.296	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.210	23:51
Breath Volume		2.019	23:51
Air Blank		0.000	23:52
Control Test		0.080	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 03/09/2021 15:06

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ECKERT

DATE/TIME OF TEST: 03/09/2021 22:38:03

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.081	22:30
Air Blank		0.000	22:31
Subject Sample	# 1	0.255	22:32
Breath Volume		1.128	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:34
Subject Sample	# 2	0.242	22:36
Breath Volume		1.917	22:36
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:37
Diagnostics Check		OK	22:38

Data Download Date/Time: 04/06/2021 15:39

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 03/09/2021 23:07:47

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.079	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.000	23:03
Breath Volume		4.117	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.000	23:06
Breath Volume		2.300	23:06
Air Blank		0.000	23:06
Control Test		0.081	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 03/12/2021 20:53:38

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.079	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.200	20:49
Breath Volume		1.757	20:49
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.186	20:51
Breath Volume		1.265	20:51
Air Blank		0.000	20:52
Control Test		0.078	20:53
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: STIDUM

DATE/TIME OF TEST: 03/13/2021 03:19:40

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.062	03:15
Breath Volume		1.488	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.058	03:18
Breath Volume		1.769	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: CATARINO MATEO

DATE/TIME OF TEST: 03/13/2021 18:03:29

Diagnostics Check		OK	17:57
Air Blank		0.000	17:57
Control Test		0.080	17:57
Air Blank		0.000	17:58
Subject Sample	# 1	0.107	17:58
Breath Volume		2.152	17:58
Air Blank		0.000	17:59
Air Blank		0.000	18:01
Subject Sample	# 2	0.108	18:01
Breath Volume		3.011	18:01
Air Blank		0.000	18:02
Control Test		0.080	18:02
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HARVEY JR

DATE/TIME OF TEST: 03/14/2021 01:13:34

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.094	01:08
Breath Volume		1.886	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.092	01:11
Breath Volume		1.875	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MONDRAGON POZOS

DATE/TIME OF TEST: 03/14/2021 04:04:49

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.159	03:59
Breath Volume		4.132	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.156	04:03
Breath Volume		4.343	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MESA REVUELTA

DATE/TIME OF TEST: 03/15/2021 03:17:41

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.203	03:13
Breath Volume		2.007	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.201	03:15
Breath Volume		1.761	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: PENINGER

DATE/TIME OF TEST: 03/17/2021 23:37:36

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.078	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.142	23:33
Breath Volume		2.558	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.145	23:35
Breath Volume		2.164	23:35
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 03/18/2021 03:52:00

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.172	03:47
Breath Volume		1.226	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.175	03:50
Breath Volume		1.722	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:52

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/20/2021 01:59:30

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.203	01:54
Breath Volume		2.488	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.198	01:57
Breath Volume		2.593	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 04/06/2021 15:40

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GAZZOLA,JULIEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/07/2021 21:34:09

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.081	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.143	21:29
Breath Volume		3.562	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:31
Subject Sample	# 2	0.144	21:32
Breath Volume		1.503	21:32
Air Blank		0.000	21:33
Control Test		0.080	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WHITT,JESSE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: LEAL

DATE/TIME OF TEST: 03/12/2021 22:08:55

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:54	
Control Test		0.080	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.046	21:58	Volume Not Met
Breath Volume		0.652	21:58	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.057	22:03	
Breath Volume		2.613	22:03	
Air Blank		0.000	22:04	
Air Blank		0.000	22:06	
Subject Sample	# 3		22:07	RFI Detect
Breath Volume		1.136	22:07	
Air Blank		0.000	22:08	
Control Test		0.080	22:08	
Air Blank		0.000	22:08	
Diagnostics Check		OK	22:08	

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BRUDER,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80-007168
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: FOLGOSO

DATE/TIME OF TEST: 03/13/2021 17:50:13

Diagnostics Check		OK	17:45	
Air Blank		0.000	17:46	
Control Test		0.080	17:46	
Air Blank		0.000	17:47	
Subject Sample	# 1		17:49	Sequence Aborted
Breath Volume		0.000	17:49	
Air Blank		0.000	17:50	

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BRUDER,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: FOLGOSO

DATE/TIME OF TEST: 03/13/2021 17:59:35

Diagnostics Check		OK	17:53
Air Blank		0.000	17:53
Control Test		0.081	17:54
Air Blank		0.000	17:54
Subject Sample	# 1	0.169	17:54
Breath Volume		2.773	17:54
Air Blank		0.000	17:55
Air Blank		0.000	17:57
Subject Sample	# 2	0.166	17:57
Breath Volume		2.328	17:57
Air Blank		0.000	17:58
Control Test		0.079	17:58
Air Blank		0.000	17:59
Diagnostics Check		OK	17:59

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINLEY,STEVEN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: STANTON,TARYN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: VALLADARES ALAYA

DATE/TIME OF TEST: 03/13/2021 20:26:27

Diagnostics Check		OK	20:22	
Air Blank		0.000	20:22	
Control Test		0.080	20:22	
Air Blank		0.000	20:23	
Subject Sample	# 1	0.138	20:25	
Breath Volume		1.347	20:25	
Air Blank			20:25	Purge Fail
Air Blank		0.000	20:26	

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: DESPIAN,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BOFFIL-ROMAN

DATE/TIME OF TEST: 03/15/2021 23:15:51

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.080	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.113	23:11
Breath Volume		3.640	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:13
Subject Sample	# 2	0.116	23:14
Breath Volume		3.746	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GAZZOLA,JULIEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:49

SUBJECT'S LAST NAME: CHAVIANO

DATE/TIME OF TEST: 03/20/2021 17:35:06

Diagnostics Check		OK	17:32	
Air Blank		0.000	17:32	
Control Test		0.080	17:33	
Air Blank		0.000	17:33	
Subject Sample	# 1		17:33	Subject Test Refused
Breath Volume		0.000	17:33	
Air Blank		0.000	17:34	
Control Test		0.081	17:34	
Air Blank		0.000	17:35	
Diagnostics Check		OK	17:35	

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GAZZOLA,JULIEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 03/20/2021 22:43:17

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:40	
Control Test		0.079	22:41	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:42	
Control Test		0.080	22:42	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDGEPETH,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GEORGEVICH,THOMAS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: VIGIL

DATE/TIME OF TEST: 03/26/2021 19:51:11

Diagnostics Check		OK	19:48	
Air Blank		0.000	19:48	
Control Test		0.081	19:49	
Air Blank		0.000	19:49	
Subject Sample	# 1		19:49	Subject Test Refused
Breath Volume		0.000	19:49	
Air Blank		0.000	19:50	
Control Test		0.081	19:50	
Air Blank		0.000	19:51	
Diagnostics Check		OK	19:51	

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCK,NATHAN,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CONROY,JESSICA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 02/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: SADLER VEIGA

DATE/TIME OF TEST: 03/27/2021 21:42:08

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.081	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.032	21:37
Breath Volume		2.613	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.028	21:40
Breath Volume		1.988	21:40
Air Blank		0.000	21:41
Control Test		0.082	21:41
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 04/02/2021 14:07

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TEJADA PELAEZ

DATE/TIME OF TEST: 03/01/2021 03:56:23

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.220	03:52	
Breath Volume		2.406	03:52	
Air Blank		0.000	03:53	
Air Blank			03:54	Improper Sample
Air Blank		0.000	03:55	
Control Test		0.080	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 03/02/2021 11:56

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TEJADA PELAEZ

DATE/TIME OF TEST: 03/01/2021 04:10:45

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.201	04:03
Breath Volume		2.492	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.228	04:06
Breath Volume		2.292	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 3	0.208	04:09
Breath Volume		2.468	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 03/02/2021 15:59

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 03/07/2021 05:44:45

Diagnostics Check		OK	05:41	
Air Blank		0.000	05:42	
Control Test		0.080	05:42	
Air Blank		0.000	05:43	
Subject Sample	# 1		05:43	Subject Test Refused
Breath Volume		0.000	05:43	
Air Blank		0.000	05:43	
Control Test		0.079	05:44	
Air Blank		0.000	05:44	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 04/08/2021 10:48

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 03/13/2021 02:59:37

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.163	02:54
Breath Volume		3.398	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.162	02:57
Breath Volume		3.113	02:57
Air Blank		0.000	02:58
Control Test		0.080	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 04/08/2021 10:48

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ITURREY

DATE/TIME OF TEST: 03/18/2021 00:54:12

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:51	
Control Test		0.080	00:52	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 04/08/2021 10:48

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: QUENZER

DATE/TIME OF TEST: 03/19/2021 01:21:49

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.120	01:15
Breath Volume		1.406	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.134	01:20
Breath Volume		2.007	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 04/08/2021 10:48

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURCIER,LEON,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: PEAKE

DATE/TIME OF TEST: 03/20/2021 02:30:05

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.145	02:25
Breath Volume		1.609	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.148	02:28
Breath Volume		1.320	02:28
Air Blank		0.000	02:29
Control Test		0.077	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 04/08/2021 10:48

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 03/24/2021 02:47:51

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.079	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.242	02:42
Breath Volume		1.847	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.240	02:46
Breath Volume		1.742	02:46
Air Blank		0.000	02:46
Control Test		0.078	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 04/08/2021 10:48

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CRISTEA,VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: AUMILLER

DATE/TIME OF TEST: 03/25/2021 21:53:20

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.080	21:48
Air Blank		0.000	21:48
Subject Sample	# 1	0.000	21:49
Breath Volume		3.183	21:49
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.000	21:51
Breath Volume		2.019	21:51
Air Blank		0.000	21:52
Control Test		0.080	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 04/08/2021 10:48

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ONGUARD,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 03/28/2021 23:53:23

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:42	
Control Test		0.080	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:46	No Sample Provided
Breath Volume		0.000	23:46	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 2		23:52	No Sample Provided
Breath Volume		0.000	23:52	
Air Blank		0.000	23:52	
Control Test		0.081	23:52	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 04/08/2021 10:49

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN-FRANCOIS,LEEF,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JEAN-FRANCOIS,LEEF
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 11:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HEYD

DATE/TIME OF TEST: 03/24/2021 00:37:50

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:37	Sequence Aborted
Breath Volume		0.000	00:37	
Air Blank		0.000	00:37	

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,J,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: STRECIWILK

DATE/TIME OF TEST: 03/24/2021 21:29:20

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.080	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.152	21:24
Breath Volume		2.328	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.149	21:27
Breath Volume		2.292	21:27
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEPHART,HUNTER,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KEPHART,HUNTER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: CZAPLICK

DATE/TIME OF TEST: 03/24/2021 22:01:54

Diagnostics Check		OK	21:55
Air Blank		0.000	21:55
Control Test		0.079	21:56
Air Blank		0.000	21:56
Subject Sample	# 1	0.130	21:57
Breath Volume		1.320	21:57
Air Blank		0.000	21:58
Air Blank		0.000	21:59
Subject Sample	# 2	0.126	22:00
Breath Volume		1.242	22:00
Air Blank		0.000	22:00
Control Test		0.078	22:01
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ORCHARD,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 03/24/2021 23:10:04

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.077	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.190	23:05
Breath Volume		1.898	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.189	23:08
Breath Volume		1.910	23:08
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:10

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCOY,BRANDEN,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SMITH,MICAH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GOLDING

DATE/TIME OF TEST: 03/27/2021 03:14:11

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.086	03:08
Breath Volume		2.156	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.081	03:12
Breath Volume		2.273	03:12
Air Blank		0.000	03:13
Control Test		0.078	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COLEMAN,JONATHAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 03/27/2021 20:17:05

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.079	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.199	20:12
Breath Volume		2.183	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:14
Subject Sample	# 2	0.199	20:15
Breath Volume		2.238	20:15
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,ALAN,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: TAPANES,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 03/28/2021 04:25:19

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.106	04:20
Breath Volume		1.527	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.111	04:23
Breath Volume		1.949	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: ROHR

DATE/TIME OF TEST: 03/28/2021 04:54:06

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.073	04:49
Breath Volume		1.968	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.070	04:52
Breath Volume		1.765	04:52
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 04/27/2021 14:23

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBB,B,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GRUBB,B
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PHILON

DATE/TIME OF TEST: 03/04/2021 04:37:05

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.000	04:32
Breath Volume		2.464	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.000	04:35
Breath Volume		3.546	04:35
Air Blank		0.000	04:36
Control Test		0.081	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,JOEDICKE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 03/05/2021 01:17:41

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.152	01:12
Breath Volume		2.550	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.154	01:16
Breath Volume		3.183	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ORCHARD,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 03/05/2021 18:35:05

Diagnostics Check		OK	18:31	
Air Blank		0.000	18:32	
Control Test		0.081	18:32	
Air Blank		0.000	18:32	
Subject Sample	# 1		18:33	Subject Test Refused
Breath Volume		0.000	18:33	
Air Blank		0.000	18:34	
Control Test		0.081	18:34	
Air Blank		0.000	18:34	
Diagnostics Check		OK	18:35	

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/12/2021 10:54:04

Diagnostics Check		OK	10:47
Air Blank		0.000	10:48
Control Test		0.081	10:48
Air Blank		0.000	10:49
Subject Sample	# 1	0.158	10:49
Breath Volume		3.156	10:49
Air Blank		0.000	10:50
Air Blank		0.000	10:51
Subject Sample	# 2	0.162	10:52
Breath Volume		3.019	10:52
Air Blank		0.000	10:53
Control Test		0.080	10:53
Air Blank		0.000	10:53
Diagnostics Check		OK	10:54

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON,CODY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BUI,LEON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: DUSKA

DATE/TIME OF TEST: 03/12/2021 15:58:16

Diagnostics Check		OK	15:51
Air Blank		0.000	15:52
Control Test		0.080	15:52
Air Blank		0.000	15:53
Subject Sample	# 1	0.232	15:53
Breath Volume		2.378	15:53
Air Blank		0.000	15:54
Air Blank		0.000	15:56
Subject Sample	# 2	0.226	15:56
Breath Volume		2.457	15:56
Air Blank		0.000	15:57
Control Test		0.079	15:57
Air Blank		0.000	15:58
Diagnostics Check		OK	15:58

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAKER,MARK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 03/12/2021 17:32:16

Diagnostics Check		OK	17:26
Air Blank		0.000	17:26
Control Test		0.081	17:26
Air Blank		0.000	17:27
Subject Sample	# 1	0.009	17:27
Breath Volume		2.578	17:27
Air Blank		0.000	17:28
Air Blank		0.000	17:30
Subject Sample	# 2	0.008	17:30
Breath Volume		2.757	17:30
Air Blank		0.000	17:31
Control Test		0.080	17:31
Air Blank		0.000	17:32
Diagnostics Check		OK	17:32

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JAMES,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: MARZI

DATE/TIME OF TEST: 03/12/2021 18:39:49

Diagnostics Check		OK	18:33
Air Blank		0.000	18:34
Control Test		0.080	18:34
Air Blank		0.000	18:35
Subject Sample	# 1	0.031	18:35
Breath Volume		2.875	18:35
Air Blank		0.000	18:36
Air Blank		0.000	18:37
Subject Sample	# 2	0.029	18:38
Breath Volume		2.890	18:38
Air Blank		0.000	18:38
Control Test		0.079	18:39
Air Blank		0.000	18:39
Diagnostics Check		OK	18:39

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYER,ZACHARY,
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FELICIEN,PHILLIP
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:23

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 03/14/2021 12:59:59

Diagnostics Check		OK	12:50
Air Blank		0.000	12:50
Control Test		0.080	12:50
Air Blank		0.000	12:51
Subject Sample	# 1	0.397	12:51
Breath Volume		2.078	12:51
Air Blank		0.000	12:52
Air Blank		0.000	12:54
Subject Sample	# 2	0.370	12:54
Breath Volume		1.582	12:54
Air Blank		0.000	12:55
Air Blank		0.000	12:57
Subject Sample	# 3	0.366	12:57
Breath Volume		2.140	12:57
Air Blank		0.000	12:59
Control Test		0.078	12:59
Air Blank		0.000	12:59
Diagnostics Check		OK	12:59

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 03/16/2021 23:30:34

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.080	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.165	23:24
Breath Volume		1.128	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.176	23:28
Breath Volume		1.945	23:28
Air Blank		0.000	23:29
Control Test		0.079	23:29
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEPHART,HUNTER,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KEPHART,HUNTER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ORDUNO SANCHEZ

DATE/TIME OF TEST: 03/17/2021 00:00:33

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.126	23:55
Breath Volume		2.433	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.122	23:58
Breath Volume		2.472	23:58
Air Blank		0.000	23:59
Control Test		0.078	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 03/20/2021 03:26:06

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.081	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.135	03:21
Breath Volume		3.003	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:23
Subject Sample	# 2	0.126	03:24
Breath Volume		3.007	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 03/20/2021 23:12:04

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.080	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.162	23:07
Breath Volume		1.964	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:09
Subject Sample	# 2	0.163	23:10
Breath Volume		2.187	23:10
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:12

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JAMES,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GADSON

DATE/TIME OF TEST: 03/21/2021 04:14:21

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:05	
Control Test		0.080	04:06	
Air Blank		0.000	04:06	
Subject Sample	# 1	0.071	04:07	
Breath Volume		1.578	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 2	0.062	04:12	
Breath Volume		2.261	04:12	
Air Blank		0.000	04:13	
Control Test		0.072	04:13	Control Outside Tolerance
Air Blank		0.000	04:14	

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JAMES,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GADSON

DATE/TIME OF TEST: 03/21/2021 04:24:12

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.063	04:19
Breath Volume		2.234	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:21
Subject Sample	# 2	0.060	04:22
Breath Volume		1.484	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 03/21/2021 04:38:18

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.158	04:33
Breath Volume		2.292	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.157	04:36
Breath Volume		1.390	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECCUE,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BECCUE,JOSHUA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: STANFORD

DATE/TIME OF TEST: 03/21/2021 04:59:42

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.179	04:52
Breath Volume		2.273	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.171	04:57
Breath Volume		1.582	04:57
Air Blank		0.000	04:58
Control Test		0.079	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ORCHARD,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: MATOUSEK

DATE/TIME OF TEST: 03/22/2021 07:19:25

Diagnostics Check		OK	07:16	
Air Blank		0.000	07:16	
Control Test		0.079	07:17	
Air Blank		0.000	07:17	
Subject Sample	# 1		07:18	Subject Test Refused
Breath Volume		0.000	07:18	
Air Blank		0.000	07:18	
Control Test		0.080	07:18	
Air Blank		0.000	07:19	
Diagnostics Check		OK	07:19	

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COLEMAN,JONATHAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ONEIL

DATE/TIME OF TEST: 03/22/2021 22:12:40

Diagnostics Check		OK	22:05
Air Blank		0.000	22:06
Control Test		0.080	22:06
Air Blank		0.000	22:07
Subject Sample	# 1	0.206	22:07
Breath Volume		1.640	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.203	22:10
Breath Volume		1.437	22:10
Air Blank		0.000	22:11
Control Test		0.079	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 03/23/2021 01:57:03

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.142	01:52
Breath Volume		3.121	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:54
Subject Sample	# 2	0.145	01:55
Breath Volume		3.605	01:55
Air Blank		0.000	01:56
Control Test		0.078	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:57

Data Download Date/Time: 03/23/2021 15:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/20/2021 05:17:20

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.209	05:12
Breath Volume		1.230	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.213	05:15
Breath Volume		1.546	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 04/17/2021 13:43

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,JON,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GORDON,JON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/21/2021 22:12:28

Diagnostics Check		OK	22:05
Air Blank		0.000	22:06
Control Test		0.080	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.163	22:07
Breath Volume		1.847	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.163	22:10
Breath Volume		1.742	22:10
Air Blank		0.000	22:11
Control Test		0.076	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 04/17/2021 13:43

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 03/26/2021 01:40:06

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:29	
Control Test		0.079	01:29	
Air Blank		0.000	01:29	
Subject Sample	# 1		01:33	No Sample Provided
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 2		01:38	No Sample Provided
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Control Test		0.080	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 04/17/2021 13:43

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,DAVID,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BUDUEN,PABLO
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CRAYTON

DATE/TIME OF TEST: 03/28/2021 19:21:16

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.079	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.126	19:16
Breath Volume		1.468	19:16
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.114	19:19
Breath Volume		1.148	19:19
Air Blank		0.000	19:20
Control Test		0.082	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 04/17/2021 13:43

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/01/2021 01:07:52

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.079	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Control Test		0.078	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 03/18/2021 09:44

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DARBY,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: CALES

DATE/TIME OF TEST: 03/02/2021 07:09:19

Diagnostics Check		OK	07:03
Air Blank		0.000	07:03
Control Test		0.079	07:04
Air Blank		0.000	07:04
Subject Sample	# 1	0.000	07:05
Breath Volume		1.832	07:05
Air Blank		0.000	07:05
Air Blank		0.000	07:07
Subject Sample	# 2	0.000	07:07
Breath Volume		3.503	07:07
Air Blank		0.000	07:08
Control Test		0.078	07:08
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

Data Download Date/Time: 03/18/2021 09:44

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: PIETRZAKOWSKI

DATE/TIME OF TEST: 03/02/2021 20:58:18

Diagnostics Check		OK	20:55	
Air Blank		0.000	20:55	
Control Test		0.079	20:56	
Air Blank		0.000	20:56	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:57	
Control Test		0.078	20:57	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 03/18/2021 09:44

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDLE,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORDLE,ABIGAIL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 03/03/2021 04:12:24

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.078	04:10	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Control Test		0.078	04:11	
Air Blank		0.000	04:12	
Diagnostics Check		OK	04:12	

Data Download Date/Time: 03/18/2021 09:44

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 03/05/2021 03:07:02

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:55	
Control Test		0.079	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:59	No Sample Provided
Breath Volume		0.000	02:59	
Air Blank		0.000	03:00	
Air Blank		0.000	03:02	
Subject Sample	# 2		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 03/18/2021 09:44

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDLE,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: WALLACE,RYAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 03/06/2021 03:56:17

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.174	03:51
Breath Volume		2.023	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.173	03:54
Breath Volume		2.046	03:54
Air Blank		0.000	03:55
Control Test		0.077	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 03/18/2021 09:44

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 03/09/2021 15:19:06

Diagnostics Check		OK	15:13
Air Blank		0.000	15:13
Control Test		0.078	15:13
Air Blank		0.000	15:14
Subject Sample	# 1	0.000	15:14
Breath Volume		3.421	15:14
Air Blank		0.000	15:15
Air Blank		0.000	15:17
Subject Sample	# 2	0.000	15:17
Breath Volume		3.101	15:17
Air Blank		0.000	15:18
Control Test		0.078	15:18
Air Blank		0.000	15:19
Diagnostics Check		OK	15:19

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: THOMPSON,MASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 03/10/2021 20:21:11

Diagnostics Check		OK	20:15
Air Blank		0.000	20:15
Control Test		0.079	20:15
Air Blank		0.000	20:16
Subject Sample	# 1	0.237	20:16
Breath Volume		1.761	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.251	20:19
Breath Volume		2.359	20:19
Air Blank		0.000	20:20
Control Test		0.077	20:20
Air Blank		0.000	20:21
Diagnostics Check		OK	20:21

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KAMMER,ROBERT
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 03/10/2021 21:14:57

Diagnostics Check		OK	21:07
Air Blank		0.000	21:08
Control Test		0.078	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.171	21:10
Breath Volume		1.699	21:10
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.171	21:13
Breath Volume		1.335	21:13
Air Blank		0.000	21:14
Control Test		0.076	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDLE,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORDLE,ABIGAIL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/11/2021 03:18:16

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.078	03:13
Breath Volume		2.085	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.082	03:16
Breath Volume		2.496	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYERS,SPENCER,O
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: EVANS,IAN
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: ROMINE

DATE/TIME OF TEST: 03/13/2021 20:39:07

Diagnostics Check		OK	20:32
Air Blank		0.000	20:32
Control Test		0.079	20:33
Air Blank		0.000	20:33
Subject Sample	# 1	0.155	20:34
Breath Volume		2.027	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:36
Subject Sample	# 2	0.152	20:37
Breath Volume		3.265	20:37
Air Blank		0.000	20:38
Control Test		0.079	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WHITEHEAD,JAMES
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: CANELOS

DATE/TIME OF TEST: 03/13/2021 21:10:04

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.078	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.159	21:05
Breath Volume		1.343	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.172	21:08
Breath Volume		1.367	21:08
Air Blank		0.000	21:09
Control Test		0.077	21:09
Air Blank		0.000	21:09
Diagnostics Check		OK	21:10

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PLANT

DATE/TIME OF TEST: 03/14/2021 03:36:22

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.068	03:35	RFI Detect
Air Blank	0.000	03:36	

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MUNCH

DATE/TIME OF TEST: 03/17/2021 02:06:43

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.000	02:02
Breath Volume		2.843	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.000	02:05
Breath Volume		2.671	02:05
Air Blank		0.000	02:05
Control Test		0.078	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 03/17/2021 23:17:59

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.077	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.084	23:12
Breath Volume		1.214	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.082	23:16
Breath Volume		1.167	23:16
Air Blank		0.000	23:17
Control Test		0.077	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,M
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JACALONE

DATE/TIME OF TEST: 03/18/2021 03:33:33

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.135	03:29
Breath Volume		1.621	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.139	03:32
Breath Volume		1.671	03:32
Air Blank		0.000	03:32
Control Test		0.077	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 03/18/2021 09:45

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONIE,JARED,P
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MONIE,JARED
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: KIMBALL

DATE/TIME OF TEST: 03/21/2021 22:39:16

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.079	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:37	Subject Test Refused
Breath Volume		0.000	22:37	
Air Blank		0.000	22:38	
Control Test		0.078	22:38	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 04/16/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/23/2021 17:45:32

Diagnostics Check	OK	17:44	
Air Blank	0.000	17:44	
Control Test	0.072	17:44	Control Outside Tolerance
Air Blank	0.000	17:45	

Data Download Date/Time: 04/16/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HETZLER,BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 03/25/2021 02:42:01

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.142	02:37
Breath Volume		1.933	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.136	02:40
Breath Volume		1.500	02:40
Air Blank		0.000	02:41
Control Test		0.077	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:42

Data Download Date/Time: 04/16/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DARBY,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WINTERS,ERIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: GROGG

DATE/TIME OF TEST: 03/25/2021 05:48:16

Diagnostics Check		OK	05:42
Air Blank		0.000	05:42
Control Test		0.078	05:43
Air Blank		0.000	05:43
Subject Sample	# 1	0.000	05:44
Breath Volume		5.636	05:44
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.000	05:46
Breath Volume		5.507	05:46
Air Blank		0.000	05:47
Control Test		0.077	05:47
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 04/16/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: MADEJ

DATE/TIME OF TEST: 03/25/2021 17:46:32

Diagnostics Check		OK	17:40
Air Blank		0.000	17:40
Control Test		0.079	17:41
Air Blank		0.000	17:41
Subject Sample	# 1	0.000	17:42
Breath Volume		1.324	17:42
Air Blank		0.000	17:42
Air Blank		0.000	17:44
Subject Sample	# 2	0.000	17:45
Breath Volume		1.324	17:45
Air Blank		0.000	17:45
Control Test		0.078	17:45
Air Blank		0.000	17:46
Diagnostics Check		OK	17:46

Data Download Date/Time: 04/16/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,ALAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,M
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MCCAWLEY

DATE/TIME OF TEST: 03/26/2021 21:52:21

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.078	21:42
Air Blank		0.000	21:43
Subject Sample	# 1	0.235	21:44
Breath Volume		3.105	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.213	21:47
Breath Volume		1.101	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 3	0.228	21:50
Breath Volume		1.511	21:50
Air Blank		0.000	21:51
Control Test		0.078	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 04/16/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BROWN,JOHNATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 03/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 03/31/2021 21:10:46

Diagnostics Check		OK	21:07	
Air Blank		0.000	21:07	
Control Test		0.078	21:08	
Air Blank		0.000	21:08	
Subject Sample	# 1		21:09	Subject Test Refused
Breath Volume		0.000	21:09	
Air Blank		0.000	21:09	
Control Test		0.078	21:10	
Air Blank		0.000	21:10	
Diagnostics Check		OK	21:10	

Data Download Date/Time: 04/16/2021 11:17

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGHER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CREAMER,REMINGTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 03/04/2021 20:42:39

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:31	
Control Test		0.078	20:32	
Air Blank		0.000	20:32	
Subject Sample	# 1		20:35	No Sample Provided
Breath Volume		0.000	20:35	
Air Blank		0.000	20:36	
Air Blank		0.000	20:38	
Subject Sample	# 2		20:41	No Sample Provided
Breath Volume		0.000	20:41	
Air Blank		0.000	20:41	
Control Test		0.079	20:42	
Air Blank		0.000	20:42	
Diagnostics Check		OK	20:42	

Data Download Date/Time: 03/15/2021 18:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,KEITH,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: REINLIE

DATE/TIME OF TEST: 03/06/2021 06:49:10

Diagnostics Check		OK	06:42	
Air Blank		0.000	06:43	
Control Test		0.077	06:43	
Air Blank		0.000	06:44	
Subject Sample	# 1	0.069	06:45	
Breath Volume		2.121	06:45	
Air Blank		0.000	06:46	
Air Blank		0.000	06:47	
Subject Sample	# 2	0.067	06:48	
Breath Volume		3.191	06:48	
Air Blank			06:48	RFI Detect
Air Blank		0.000	06:49	

Data Download Date/Time: 03/15/2021 18:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,KEITH,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,KEITH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: REINLIE

DATE/TIME OF TEST: 03/06/2021 06:58:05

Diagnostics Check		OK	06:51
Air Blank		0.000	06:51
Control Test		0.077	06:52
Air Blank		0.000	06:52
Subject Sample	# 1	0.065	06:53
Breath Volume		1.507	06:53
Air Blank		0.000	06:54
Air Blank		0.000	06:56
Subject Sample	# 2	0.063	06:56
Breath Volume		2.359	06:56
Air Blank		0.000	06:57
Control Test		0.077	06:57
Air Blank		0.000	06:57
Diagnostics Check		OK	06:58

Data Download Date/Time: 03/15/2021 18:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGHER,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 03/31/2021 23:44:01

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.078	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.000	23:39
Breath Volume		3.773	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.000	23:42
Breath Volume		2.273	23:42
Air Blank		0.000	23:43
Control Test		0.077	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:44

Data Download Date/Time: 04/17/2021 09:50

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GORGONIO GOMEZ

DATE/TIME OF TEST: 03/01/2021 02:01:18

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.161	01:56
Breath Volume		3.234	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.166	01:59
Breath Volume		3.285	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 03/03/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RIVAS,HUMBERTO
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 02/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KAHLA

DATE/TIME OF TEST: 03/03/2021 01:16:22

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.081	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.169	01:11
Breath Volume		1.101	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.173	01:14
Breath Volume		1.382	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 03/03/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CHURLY

DATE/TIME OF TEST: 03/03/2021 02:15:06

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.111	02:10
Breath Volume		1.550	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.109	02:13
Breath Volume		1.472	02:13
Air Blank		0.000	02:14
Control Test		0.078	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 03/03/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 03/04/2021 03:07:13

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.117	03:02
Breath Volume		2.925	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.128	03:05
Breath Volume		2.726	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 04/13/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ABOUD

DATE/TIME OF TEST: 03/04/2021 03:57:05

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:46	
Control Test		0.080	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.075	03:50	Volume Not Met
Breath Volume		0.027	03:50	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 2		03:55	No Sample Provided
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 04/13/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALONSO,BERNIE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/05/2021 21:05:02

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.081	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.201	20:59
Breath Volume		1.339	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.203	21:03
Breath Volume		1.644	21:03
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:05

Data Download Date/Time: 04/13/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PERSAD

DATE/TIME OF TEST: 03/05/2021 23:50:41

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.080	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.251	23:45
Breath Volume		1.453	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.256	23:48
Breath Volume		1.660	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 04/13/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: VARGHESE

DATE/TIME OF TEST: 03/06/2021 02:32:43

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.128	02:28
Breath Volume		1.371	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.126	02:31
Breath Volume		1.449	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 04/13/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANDRADE,DAVID
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: JOHANSEN

DATE/TIME OF TEST: 03/10/2021 20:06:34

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.081	20:00
Air Blank		0.000	20:01
Subject Sample	# 1	0.000	20:02
Breath Volume		2.226	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.000	20:05
Breath Volume		2.800	20:05
Air Blank		0.000	20:05
Control Test		0.081	20:05
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 04/13/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 03/10/2021 23:56:47

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:45	
Control Test		0.081	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.186	23:49	Volume Not Met
Breath Volume		0.054	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 2	0.219	23:55	Volume Not Met
Breath Volume		0.058	23:55	
Air Blank		0.000	23:55	
Control Test		0.078	23:56	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 04/13/2021 11:12

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 03/13/2021 05:31:45

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.081	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1	0.000	05:27	Volume Not Met
Breath Volume		0.000	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:30	
Subject Sample	# 2		05:30	Subject Test Refused
Breath Volume		0.000	05:30	
Air Blank		0.000	05:30	
Control Test		0.080	05:31	
Air Blank		0.000	05:31	
Diagnostics Check		OK	05:31	

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: SAXBERG

DATE/TIME OF TEST: 03/13/2021 22:05:20

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.081	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.227	22:00
Breath Volume		2.750	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:02
Subject Sample	# 2	0.229	22:03
Breath Volume		3.160	22:03
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BLUM

DATE/TIME OF TEST: 03/14/2021 00:32:40

Diagnostics Check		OK	00:18	
Air Blank		0.000	00:19	
Control Test		0.081	00:19	
Air Blank		0.000	00:19	
Subject Sample	# 1	0.159	00:23	Volume Not Met
Breath Volume		0.332	00:23	
Air Blank		0.000	00:23	
Air Blank		0.000	00:25	
Subject Sample	# 2	0.178	00:26	
Breath Volume		1.359	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 3	0.188	00:30	
Breath Volume		1.515	00:30	
Air Blank		0.000	00:31	
Control Test		0.079	00:32	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUBREUIL,ALAIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: TRONSEN

DATE/TIME OF TEST: 03/14/2021 23:41:42

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.136	23:36
Breath Volume		2.652	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.129	23:40
Breath Volume		2.601	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,NORBENS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: PHILIPPE

DATE/TIME OF TEST: 03/17/2021 01:10:29

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.177	01:05
Breath Volume		1.472	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.175	01:08
Breath Volume		1.230	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JENNINGS,ARTHUR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 03/17/2021 09:49:45

Diagnostics Check		OK	09:43
Air Blank		0.000	09:43
Control Test		0.081	09:43
Air Blank		0.000	09:44
Subject Sample	# 1	0.000	09:45
Breath Volume		2.605	09:45
Air Blank		0.000	09:45
Air Blank		0.000	09:47
Subject Sample	# 2	0.000	09:48
Breath Volume		2.394	09:48
Air Blank		0.000	09:48
Control Test		0.081	09:49
Air Blank		0.000	09:49
Diagnostics Check		OK	09:49

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIS,JASHAUNIE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: BRODERICK

DATE/TIME OF TEST: 03/17/2021 19:26:19

Diagnostics Check		OK	19:20
Air Blank		0.000	19:20
Control Test		0.081	19:20
Air Blank		0.000	19:21
Subject Sample	# 1	0.000	19:21
Breath Volume		2.640	19:21
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 2	0.000	19:24
Breath Volume		2.589	19:24
Air Blank		0.000	19:25
Control Test		0.081	19:25
Air Blank		0.000	19:26
Diagnostics Check		OK	19:26

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 03/18/2021 05:04:57

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.081	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.201	04:59
Breath Volume		1.980	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.205	05:03
Breath Volume		1.992	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 04/13/2021 11:13

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LAMPIASI,JAKE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: YARWOOD

DATE/TIME OF TEST: 03/20/2021 21:35:41

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.081	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.190	21:30
Breath Volume		2.304	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.200	21:33
Breath Volume		2.253	21:33
Air Blank		0.000	21:34
Control Test		0.080	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/21/2021 03:03:28

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.081	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.133	02:58
Breath Volume		2.007	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.137	03:01
Breath Volume		2.710	03:01
Air Blank		0.000	03:02
Control Test		0.078	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: AMAZAN

DATE/TIME OF TEST: 03/21/2021 04:53:37

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.081	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.080	04:48
Breath Volume		1.722	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.078	04:51
Breath Volume		1.652	04:51
Air Blank		0.000	04:52
Control Test		0.081	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ-MURILLO,REYNA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: CARDONA LUCAS

DATE/TIME OF TEST: 03/21/2021 19:47:04

Diagnostics Check		OK	19:39
Air Blank		0.000	19:40
Control Test		0.081	19:40
Air Blank		0.000	19:41
Subject Sample	# 1	0.214	19:42
Breath Volume		1.273	19:42
Air Blank		0.000	19:43
Air Blank		0.000	19:44
Subject Sample	# 2	0.216	19:45
Breath Volume		1.644	19:45
Air Blank		0.000	19:46
Control Test		0.077	19:46
Air Blank		0.000	19:46
Diagnostics Check		OK	19:47

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HEPBURN,ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: LAS MARIAS

DATE/TIME OF TEST: 03/21/2021 20:22:59

Diagnostics Check		OK	20:16
Air Blank		0.000	20:16
Control Test		0.080	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.163	20:18
Breath Volume		1.179	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:20
Subject Sample	# 2	0.164	20:21
Breath Volume		1.406	20:21
Air Blank		0.000	20:22
Control Test		0.080	20:22
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: PINA ROMERO

DATE/TIME OF TEST: 03/24/2021 02:40:19

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.156	02:35
Breath Volume		1.539	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.155	02:38
Breath Volume		1.339	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WALLEN

DATE/TIME OF TEST: 03/26/2021 01:16:32

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:03	
Control Test		0.081	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 1	0.192	01:07	Volume Not Met
Breath Volume		0.820	01:07	
Air Blank		0.000	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 2	0.196	01:11	
Breath Volume		1.386	01:11	
Air Blank		0.000	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 3	0.204	01:14	
Breath Volume		1.281	01:14	
Air Blank		0.000	01:15	
Control Test		0.078	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:39

SUBJECT'S LAST NAME: CIDRAK

DATE/TIME OF TEST: 03/27/2021 07:27:26

Diagnostics Check		OK	07:19
Air Blank		0.000	07:20
Control Test		0.082	07:20
Air Blank		0.000	07:20
Subject Sample	# 1	0.109	07:21
Breath Volume		1.355	07:21
Air Blank		0.000	07:22
Air Blank		0.000	07:24
Subject Sample	# 2	0.109	07:25
Breath Volume		1.675	07:25
Air Blank		0.000	07:26
Control Test		0.081	07:26
Air Blank		0.000	07:27
Diagnostics Check		OK	07:27

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VIERA,A
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:31

SUBJECT'S LAST NAME: DASILVA

DATE/TIME OF TEST: 03/27/2021 08:05:52

Diagnostics Check		OK	07:58
Air Blank		0.000	07:59
Control Test		0.080	07:59
Air Blank		0.000	08:00
Subject Sample	# 1	0.000	08:01
Breath Volume		1.535	08:01
Air Blank		0.000	08:01
Air Blank		0.000	08:03
Subject Sample	# 2	0.000	08:04
Breath Volume		1.625	08:04
Air Blank		0.000	08:04
Control Test		0.080	08:05
Air Blank		0.000	08:05
Diagnostics Check		OK	08:05

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,JASIEL
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 03/27/2021 22:29:31

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.082	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.164	22:24
Breath Volume		1.281	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.159	22:27
Breath Volume		1.332	22:27
Air Blank		0.000	22:28
Control Test		0.078	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: DE LEON RAYMUNDO

DATE/TIME OF TEST: 03/28/2021 22:47:23

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.198	22:42
Breath Volume		2.546	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.186	22:45
Breath Volume		1.316	22:45
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 04/13/2021 11:14

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: FUNEZ VASQUEZ

DATE/TIME OF TEST: 03/30/2021 00:59:34

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.082	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.147	00:54
Breath Volume		1.785	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.144	00:57
Breath Volume		2.078	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 04/13/2021 11:15

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JACKS

DATE/TIME OF TEST: 03/02/2021 01:40:55

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.086	01:36
Breath Volume		1.500	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:38
Subject Sample	# 2	0.088	01:39
Breath Volume		2.753	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 03/02/2021 10:31

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 03/04/2021 23:17:06

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.042	23:12
Breath Volume		3.277	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.043	23:15
Breath Volume		3.207	23:15
Air Blank		0.000	23:16
Control Test		0.080	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 04/05/2021 09:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRYDEN,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: MCCAHILL

DATE/TIME OF TEST: 03/05/2021 17:22:08

Diagnostics Check		OK	17:19	
Air Blank		0.000	17:19	
Control Test		0.080	17:19	
Air Blank		0.000	17:20	
Subject Sample	# 1		17:20	Subject Test Refused
Breath Volume		0.000	17:20	
Air Blank		0.000	17:21	
Control Test		0.081	17:21	
Air Blank		0.000	17:22	
Diagnostics Check		OK	17:22	

Data Download Date/Time: 04/05/2021 09:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRS,RUSSELL,B
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HOWARD,ROGER
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 03/07/2021 15:37:23

Diagnostics Check		OK	15:28
Air Blank		0.000	15:29
Control Test		0.081	15:29
Air Blank		0.000	15:30
Subject Sample	# 1	0.043	15:31
Breath Volume		1.121	15:31
Air Blank		0.000	15:32
Air Blank		0.000	15:33
Subject Sample	# 2	0.040	15:35
Breath Volume		1.859	15:35
Air Blank		0.000	15:36
Control Test		0.079	15:36
Air Blank		0.000	15:37
Diagnostics Check		OK	15:37

Data Download Date/Time: 04/05/2021 09:36

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAUNCEY,DAYLENE,L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: STEWART,TYREN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 03/10/2021 03:02:50

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.080	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.075	02:55	Volume Not Met
Breath Volume		0.000	02:55	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 2		03:01	No Sample Provided
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.079	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BAGLEY

DATE/TIME OF TEST: 03/11/2021 01:04:46

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:03	Subject Test Refused
Breath Volume		0.000	01:03	
Air Blank		0.000	01:03	
Control Test		0.080	01:04	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 03/13/2021 19:21:51

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:18	
Control Test		0.081	19:19	
Air Blank		0.000	19:19	
Subject Sample	# 1	0.229	19:20	
Breath Volume		1.585	19:20	
Air Blank			19:21	Purge Fail
Air Blank		0.000	19:21	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 03/13/2021 19:30:07

Diagnostics Check		OK	19:23
Air Blank		0.000	19:24
Control Test		0.076	19:24
Air Blank		0.000	19:24
Subject Sample	# 1	0.220	19:25
Breath Volume		1.890	19:25
Air Blank		0.000	19:26
Air Blank		0.000	19:27
Subject Sample	# 2	0.225	19:28
Breath Volume		2.031	19:28
Air Blank		0.000	19:29
Control Test		0.075	19:29
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAUNCEY,DAYLENE,L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: CHAUNCEY,DAYLENE
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LEON FIGEROA

DATE/TIME OF TEST: 03/15/2021 01:25:10

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.198	01:20
Breath Volume		1.460	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.214	01:23
Breath Volume		1.488	01:23
Air Blank		0.000	01:24
Control Test		0.078	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: NICHOLSON,MICHAEL
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 03/15/2021 21:08:15

Diagnostics Check		OK	20:59
Air Blank		0.000	21:00
Control Test		0.080	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.192	21:03
Breath Volume		1.503	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.194	21:06
Breath Volume		1.796	21:06
Air Blank		0.000	21:07
Control Test		0.076	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN,JASON,T
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: RHODEN,JASON
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/18/2021 00:14:18

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.167	00:03	
Breath Volume		1.289	00:03	
Air Blank		0.000	00:04	
Air Blank		0.000	00:05	
Subject Sample	# 2	0.148	00:09	Volume Not Met
Breath Volume		0.855	00:09	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 3	0.162	00:12	
Breath Volume		1.652	00:12	
Air Blank		0.000	00:13	
Control Test		0.075	00:13	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:14	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: NICHOLSON,MICHAEL
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GARCIA MARTIN

DATE/TIME OF TEST: 03/21/2021 00:51:57

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.079	00:47	
Air Blank		0.000	00:48	
Subject Sample	# 1	0.286	00:48	
Breath Volume		3.125	00:48	
Air Blank		0.000	00:49	
Air Blank			00:51	Ambient Fail
Air Blank		0.000	00:51	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: NICHOLSON,MICHAEL
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GARCIA MARTIN

DATE/TIME OF TEST: 03/21/2021 01:00:29

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.077	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.281	00:55
Breath Volume		2.375	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.273	00:58
Breath Volume		2.171	00:58
Air Blank		0.000	00:59
Control Test		0.076	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DURAN-PITTMAN

DATE/TIME OF TEST: 03/22/2021 01:27:35

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.129	01:22
Breath Volume		2.117	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.127	01:25
Breath Volume		2.832	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACFADDEN,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MACFADDEN,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: EARNEST

DATE/TIME OF TEST: 03/23/2021 19:42:22

Diagnostics Check		OK	19:35	
Air Blank		0.000	19:36	
Control Test		0.079	19:36	
Air Blank		0.000	19:37	
Subject Sample	# 1	0.109	19:37	
Breath Volume		3.742	19:37	
Air Blank		0.000	19:38	
Air Blank		0.000	19:40	
Subject Sample	# 2	0.098	19:40	
Breath Volume		2.562	19:40	
Air Blank		0.000	19:41	
Control Test		0.072	19:41	Control Outside Tolerance
Air Blank		0.000	19:42	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,M,
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: LUCKETTE,J
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/26/2021 03:16:04

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.080	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.119	03:11
Breath Volume		1.835	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.123	03:14
Breath Volume		3.031	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:16

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DANIEL,C
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: CARTER,DANIEL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 03/26/2021 20:14:29

Diagnostics Check		OK	20:10	
Air Blank		0.000	20:11	
Control Test		0.080	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1	0.193	20:13	
Breath Volume		1.945	20:13	
Air Blank			20:13	Purge Fail
Air Blank		0.000	20:14	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DANIEL,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: CARTER,DANIEL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 03/26/2021 20:23:45

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.078	20:18
Air Blank		0.000	20:18
Subject Sample	# 1	0.187	20:19
Breath Volume		2.144	20:19
Air Blank		0.000	20:20
Air Blank		0.000	20:21
Subject Sample	# 2	0.185	20:22
Breath Volume		1.789	20:22
Air Blank		0.000	20:22
Control Test		0.077	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DANIEL,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: ROBERTS,TOMMY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: BOWLES

DATE/TIME OF TEST: 03/28/2021 22:06:25

Diagnostics Check		OK	22:02	
Air Blank		0.000	22:02	
Control Test		0.080	22:03	
Air Blank		0.000	22:03	
Subject Sample	# 1	0.264	22:05	
Breath Volume		1.421	22:05	
Air Blank			22:05	Purge Fail
Air Blank		0.000	22:06	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DANIEL,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: ROBERTS,TOMMY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: BOWLES

DATE/TIME OF TEST: 03/28/2021 22:10:05

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:09	
Control Test	0.073	22:09	Control Outside Tolerance
Air Blank	0.000	22:10	

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DANIEL,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: ROBERTS,TOMMY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: BOWLES

DATE/TIME OF TEST: 03/28/2021 22:19:45

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.079	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.261	22:14
Breath Volume		3.156	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.250	22:17
Breath Volume		3.148	22:17
Air Blank		0.000	22:18
Control Test		0.078	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,M,
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: LUCKETTE,J
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/30/2021 05:14:16

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.011	05:10
Breath Volume		1.707	05:10
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.009	05:12
Breath Volume		2.644	05:12
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 03/31/2021 00:09:59

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.205	00:05
Breath Volume		1.730	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.209	00:08
Breath Volume		1.960	00:08
Air Blank		0.000	00:09
Control Test		0.077	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER,WILLIAM,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BONTRAGER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DELONEY

DATE/TIME OF TEST: 03/31/2021 01:57:03

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.000	01:52
Breath Volume		2.296	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.000	01:55
Breath Volume		2.390	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:56
Diagnostics Check		OK	01:57

Data Download Date/Time: 04/05/2021 09:37

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/23/2021 13:21:53

Diagnostics Check		OK	13:15
Air Blank		0.000	13:15
Control Test		0.082	13:16
Air Blank		0.000	13:16
Subject Sample	# 1	0.000	13:17
Breath Volume		1.839	13:17
Air Blank		0.000	13:17
Air Blank		0.000	13:19
Subject Sample	# 2	0.000	13:20
Breath Volume		1.613	13:20
Air Blank		0.000	13:20
Control Test		0.081	13:21
Air Blank		0.000	13:21
Diagnostics Check		OK	13:21

Data Download Date/Time: 04/05/2021 12:41

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/23/2021 13:33:02

Diagnostics Check		OK	13:26	
Air Blank		0.000	13:26	
Control Test		0.081	13:26	
Air Blank		0.000	13:27	
Subject Sample	# 1	0.000	13:28	
Breath Volume		1.558	13:28	
Air Blank		0.000	13:28	
Air Blank		0.000	13:30	
Subject Sample	# 2	0.000	13:31	
Breath Volume		1.292	13:31	
Air Blank		0.000	13:31	
Control Test		0.081	13:31	
Air Blank		0.000	13:32	
Diagnostics Check		Fail	13:32	Internal Printer Test
Air Blank		0.000	13:33	

Data Download Date/Time: 04/05/2021 12:41

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/23/2021 14:23:38

Diagnostics Check		OK	14:16
Air Blank		0.000	14:17
Control Test		0.080	14:17
Air Blank		0.000	14:18
Subject Sample	# 1	0.000	14:18
Breath Volume		1.351	14:18
Air Blank		0.000	14:19
Air Blank		0.000	14:21
Subject Sample	# 2	0.000	14:22
Breath Volume		1.203	14:22
Air Blank		0.000	14:22
Control Test		0.081	14:22
Air Blank		0.000	14:23
Diagnostics Check		OK	14:23

Data Download Date/Time: 04/05/2021 12:41

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/24/2021 13:57:50

Diagnostics Check	Fail	13:57	Voltage/Current Test
Air Blank	0.000	13:57	

Data Download Date/Time: 04/05/2021 12:41

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/24/2021 14:10:57

Diagnostics Check		OK	13:59	
Air Blank		0.000	13:59	
Control Test		0.081	14:00	
Air Blank		0.000	14:00	
Subject Sample	# 1		14:03	No Sample Provided
Breath Volume		0.000	14:03	
Air Blank		0.000	14:04	
Air Blank		0.000	14:06	
Subject Sample	# 2	0.000	14:06	
Breath Volume		1.152	14:06	
Air Blank		0.000	14:07	
Air Blank		0.000	14:09	
Subject Sample	# 3	0.000	14:09	
Breath Volume		1.492	14:09	
Air Blank		0.000	14:10	
Control Test		0.081	14:10	
Air Blank		0.000	14:10	
Diagnostics Check		OK	14:10	

Data Download Date/Time: 04/05/2021 12:41

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/24/2021 15:36:45

Diagnostics Check		OK	15:30
Air Blank		0.000	15:31
Control Test		0.077	15:31
Air Blank		0.000	15:32
Subject Sample	# 1	0.000	15:32
Breath Volume		1.261	15:32
Air Blank		0.000	15:32
Air Blank		0.000	15:34
Subject Sample	# 2	0.000	15:35
Breath Volume		1.117	15:35
Air Blank		0.000	15:35
Control Test		0.081	15:36
Air Blank		0.000	15:36
Diagnostics Check		OK	15:36

Data Download Date/Time: 04/05/2021 17:11

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/24/2021 15:58:07

Diagnostics Check		OK	15:51
Air Blank		0.000	15:52
Control Test		0.081	15:52
Air Blank		0.000	15:53
Subject Sample	# 1	0.000	15:53
Breath Volume		1.414	15:53
Air Blank		0.000	15:54
Air Blank		0.000	15:56
Subject Sample	# 2	0.000	15:56
Breath Volume		1.140	15:56
Air Blank		0.000	15:57
Control Test		0.081	15:57
Air Blank		0.000	15:58
Diagnostics Check		OK	15:58

Data Download Date/Time: 04/05/2021 17:11

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/18/2021 09:52:08

Diagnostics Check		OK	09:43
Air Blank		0.000	09:44
Control Test		0.079	09:44
Air Blank		0.000	09:45
Subject Sample	# 1	0.000	09:46
Breath Volume		1.199	09:46
Air Blank		0.000	09:46
Air Blank		0.000	09:48
Subject Sample	# 2	0.000	09:50
Breath Volume		1.507	09:50
Air Blank		0.000	09:51
Control Test		0.081	09:51
Air Blank		0.000	09:52
Diagnostics Check		OK	09:52

Data Download Date/Time: 04/02/2021 12:46

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/18/2021 13:30:43

Diagnostics Check		OK	13:24
Air Blank		0.000	13:24
Control Test		0.081	13:25
Air Blank		0.000	13:25
Subject Sample	# 1	0.000	13:26
Breath Volume		1.574	13:26
Air Blank		0.000	13:26
Air Blank		0.000	13:28
Subject Sample	# 2	0.000	13:29
Breath Volume		1.250	13:29
Air Blank		0.000	13:29
Control Test		0.082	13:30
Air Blank		0.000	13:30
Diagnostics Check		OK	13:30

Data Download Date/Time: 04/02/2021 12:46

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/18/2021 13:52:16

Diagnostics Check		OK	13:45
Air Blank		0.000	13:46
Control Test		0.081	13:46
Air Blank		0.000	13:47
Subject Sample	# 1	0.000	13:47
Breath Volume		1.218	13:47
Air Blank		0.000	13:48
Air Blank		0.000	13:50
Subject Sample	# 2	0.000	13:50
Breath Volume		1.187	13:50
Air Blank		0.000	13:51
Control Test		0.081	13:51
Air Blank		0.000	13:52
Diagnostics Check		OK	13:52

Data Download Date/Time: 04/02/2021 12:46

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/26/2021 10:02:37

Diagnostics Check		OK	09:56
Air Blank		0.000	09:56
Control Test		0.081	09:57
Air Blank		0.000	09:57
Subject Sample	# 1	0.000	09:58
Breath Volume		1.351	09:58
Air Blank		0.000	09:58
Air Blank		0.000	10:00
Subject Sample	# 2	0.000	10:01
Breath Volume		1.515	10:01
Air Blank		0.000	10:01
Control Test		0.081	10:02
Air Blank		0.000	10:02
Diagnostics Check		OK	10:02

Data Download Date/Time: 04/05/2021 16:56

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/26/2021 10:25:10

Diagnostics Check		OK	10:19
Air Blank		0.000	10:19
Control Test		0.081	10:19
Air Blank		0.000	10:20
Subject Sample	# 1	0.000	10:20
Breath Volume		1.335	10:20
Air Blank		0.000	10:21
Air Blank		0.000	10:23
Subject Sample	# 2	0.000	10:23
Breath Volume		1.265	10:23
Air Blank		0.000	10:24
Control Test		0.081	10:24
Air Blank		0.000	10:25
Diagnostics Check		OK	10:25

Data Download Date/Time: 04/05/2021 16:56

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 08:44:39

Diagnostics Check		OK	08:38
Air Blank		0.000	08:38
Control Test		0.082	08:39
Air Blank		0.000	08:39
Subject Sample	# 1	0.000	08:40
Breath Volume		1.335	08:40
Air Blank		0.000	08:40
Air Blank		0.000	08:42
Subject Sample	# 2	0.000	08:43
Breath Volume		1.429	08:43
Air Blank		0.000	08:43
Control Test		0.081	08:43
Air Blank		0.000	08:44
Diagnostics Check		OK	08:44

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 09:03:12

Diagnostics Check		OK	08:56	
Air Blank		0.000	08:57	
Control Test		0.081	08:57	
Air Blank		0.000	08:58	
Subject Sample	# 1	0.000	08:58	
Breath Volume		1.214	08:58	
Air Blank		0.000	08:58	
Air Blank		0.000	09:00	
Subject Sample	# 2	0.000	09:01	
Breath Volume		1.289	09:01	
Air Blank		0.000	09:01	
Control Test		0.080	09:02	
Air Blank		0.000	09:02	
Diagnostics Check		Fail	09:02	Internal Printer Test
Air Blank		0.000	09:03	

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 09:41:21

Diagnostics Check		OK	09:34
Air Blank		0.000	09:34
Control Test		0.081	09:35
Air Blank		0.000	09:35
Subject Sample	# 1	0.000	09:37
Breath Volume		1.406	09:37
Air Blank		0.000	09:37
Air Blank		0.000	09:39
Subject Sample	# 2	0.000	09:39
Breath Volume		1.343	09:39
Air Blank		0.000	09:40
Control Test		0.081	09:40
Air Blank		0.000	09:41
Diagnostics Check		OK	09:41

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 09:56:43

Diagnostics Check		OK	09:49
Air Blank		0.000	09:50
Control Test		0.081	09:50
Air Blank		0.000	09:51
Subject Sample	# 1	0.000	09:51
Breath Volume		1.128	09:51
Air Blank		0.000	09:51
Air Blank		0.000	09:53
Subject Sample	# 2	0.000	09:55
Breath Volume		1.347	09:55
Air Blank		0.000	09:55
Control Test		0.081	09:56
Air Blank		0.000	09:56
Diagnostics Check		OK	09:56

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 11:28:30

Diagnostics Check	OK	11:27	
Air Blank	0.000	11:27	
Control Test	0.097	11:27	Control Outside Tolerance
Air Blank	0.000	11:28	

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 11:31:22

Diagnostics Check	Fail	11:30	Internal Printer Test
Air Blank	0.000	11:31	

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 11:40:36

Diagnostics Check		OK	11:32	
Air Blank		0.000	11:33	
Control Test		0.081	11:33	
Air Blank		0.000	11:34	
Subject Sample	# 1	0.000	11:35	
Breath Volume		1.250	11:35	
Air Blank		0.000	11:36	
Air Blank		0.000	11:38	
Subject Sample	# 2	0.000	11:38	
Breath Volume		1.656	11:38	
Air Blank		0.000	11:39	
Control Test		0.081	11:39	
Air Blank		0.000	11:39	
Diagnostics Check		Fail	11:40	Internal Printer Test
Air Blank		0.000	11:40	

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

INSTRUMENT REGISTERED TO:
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 03/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 03/31/2021 11:48:24

Diagnostics Check		OK	11:42
Air Blank		0.000	11:42
Control Test		0.082	11:42
Air Blank		0.000	11:43
Subject Sample	# 1	0.000	11:44
Breath Volume		1.328	11:44
Air Blank		0.000	11:44
Air Blank		0.000	11:46
Subject Sample	# 2	0.000	11:46
Breath Volume		1.265	11:46
Air Blank		0.000	11:47
Control Test		0.081	11:47
Air Blank		0.000	11:48
Diagnostics Check		OK	11:48

Data Download Date/Time: 04/05/2021 17:24

Printed: 24 June 2021

COBRA ©1997-2014 CMI, Inc.