

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2021 00:20:43

Diagnostics Check	OK	00:19	
Air Blank		00:20	RFI Detect
Air Blank		00:20	RFI Detect

Data Download Date/Time: 06/14/2021 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2021 00:23:19

Diagnostics Check	OK	00:22	
Air Blank		00:22	Ambient Fail
Air Blank	0.000	00:23	

Data Download Date/Time: 06/14/2021 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2021 00:33:40

Diagnostics Check	OK	00:32	
Air Blank		00:33	Ambient Fail
Air Blank	0.000	00:33	

Data Download Date/Time: 06/14/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2021 15:24:40

Diagnostics Check	OK	15:23	
Air Blank		15:23	Ambient Fail
Air Blank		15:24	Purge Fail

Data Download Date/Time: 06/14/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2021 15:50:12

Diagnostics Check	OK	15:44
Air Blank	0.000	15:44
Control Test	0.080	15:45
Air Blank	0.000	15:45
Subject Sample # 1	0.000	15:45
Breath Volume	3.019	15:45
Air Blank	0.000	15:46
Air Blank	0.000	15:48
Subject Sample # 2	0.000	15:48
Breath Volume	2.636	15:48
Air Blank	0.000	15:49
Control Test	0.081	15:49
Air Blank	0.000	15:50
Diagnostics Check	OK	15:50

Data Download Date/Time: 06/14/2021 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 05/23/2021 16:43:43

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.078	16:38
Air Blank	0.000	16:38
Subject Sample # 1	0.007	16:39
Breath Volume	1.542	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:41
Subject Sample # 2	0.000	16:42
Breath Volume	1.878	16:42
Air Blank	0.000	16:42
Control Test	0.079	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 06/14/2021 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ORTIZ PEREZ

DATE/TIME OF TEST: 05/24/2021 01:48:35

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.339	01:40
Breath Volume	2.835	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:42
Subject Sample # 2	0.314	01:43
Breath Volume	1.660	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 3	0.310	01:46
Breath Volume	3.605	01:46
Air Blank	0.000	01:47
Control Test	0.078	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 06/14/2021 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: CLEMENT

DATE/TIME OF TEST: 05/24/2021 04:39:37

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.200	04:34
Breath Volume	1.609	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.212	04:37
Breath Volume	1.835	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 06/14/2021 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER III,RICHARD,F
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER III,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MARTIN MARTIN

DATE/TIME OF TEST: 05/28/2021 02:23:21

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.298	02:18
Breath Volume	2.210	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.286	02:21
Breath Volume	2.335	02:21
Air Blank	0.000	02:22
Control Test	0.076	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/14/2021 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 05/30/2021 03:25:41

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.085	03:25	Interferent Detect
Air Blank	0.000	03:25	

Data Download Date/Time: 06/14/2021 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 05/30/2021 03:34:18

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.177	03:29
Breath Volume	1.554	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.167	03:32
Breath Volume	1.445	03:32
Air Blank	0.000	03:33
Control Test	0.085	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/14/2021 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: COBON PALACIOS

DATE/TIME OF TEST: 05/01/2021 04:08:05

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.206	04:03
Breath Volume	1.472	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.217	04:06
Breath Volume	3.136	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/13/2021 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMRICH,IAN,W
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: SANDS,BRIAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: SORRELL

DATE/TIME OF TEST: 05/01/2021 18:09:35

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.080	18:07	
Air Blank	0.000	18:07	
Subject Sample # 1	0.000	18:08	
Breath Volume	4.820	18:08	
Air Blank		18:08	Improper Sample
Air Blank	0.000	18:08	
Control Test	0.080	18:08	
Air Blank	0.000	18:09	
Diagnostics Check	OK	18:09	

Data Download Date/Time: 05/13/2021 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMRICH,IAN,W
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: SANDS,BRIAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: SORRELL

DATE/TIME OF TEST: 05/01/2021 18:18:11

Diagnostics Check	OK	18:11
Air Blank	0.000	18:12
Control Test	0.080	18:12
Air Blank	0.000	18:12
Subject Sample # 1	0.000	18:13
Breath Volume	1.902	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.000	18:16
Breath Volume	1.824	18:16
Air Blank	0.000	18:17
Control Test	0.080	18:17
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 05/13/2021 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/05/2021 05:47:11

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:35	
Control Test	0.080	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:39	No Sample Provided
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Air Blank	0.000	05:41	
Subject Sample	# 2	0.132	05:42
Breath Volume	4.632	05:42	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample	# 3	0.130	05:45
Breath Volume	3.722	05:45	
Air Blank	0.000	05:46	
Control Test	0.076	05:46	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 05/13/2021 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RAUDA CHAVEZ

DATE/TIME OF TEST: 05/06/2021 00:03:49

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.156	23:58
Breath Volume	1.855	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.150	00:02
Breath Volume	2.222	00:02
Air Blank	0.000	00:02
Control Test	0.078	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 05/13/2021 15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRAO, MICHAEL, W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BRUTUS

DATE/TIME OF TEST: 05/11/2021 21:44:58

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.081	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.008	21:40
Breath Volume	1.476	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.000	21:43
Breath Volume	1.996	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 05/13/2021 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WORTMAN

DATE/TIME OF TEST: 05/12/2021 02:42:40

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.055	02:37
Breath Volume	5.210	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.054	02:41
Breath Volume	5.121	02:41
Air Blank	0.000	02:41
Control Test	0.081	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/13/2021 15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: DROHAN

DATE/TIME OF TEST: 05/17/2021 16:50:17

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.080	16:44
Air Blank	0.000	16:44
Subject Sample # 1	0.194	16:45
Breath Volume	1.660	16:45
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 2	0.191	16:48
Breath Volume	1.460	16:48
Air Blank	0.000	16:49
Control Test	0.079	16:49
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 06/10/2021 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: HERNDEN

DATE/TIME OF TEST: 05/23/2021 02:20:45

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.136	02:15
Breath Volume	2.222	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.145	02:19
Breath Volume	3.324	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 06/10/2021 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: CORSIGLIA

DATE/TIME OF TEST: 05/31/2021 17:44:54

Diagnostics Check	OK	17:38
Air Blank	0.000	17:39
Control Test	0.080	17:39
Air Blank	0.000	17:40
Subject Sample # 1	0.165	17:40
Breath Volume	3.675	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:42
Subject Sample # 2	0.172	17:43
Breath Volume	4.050	17:43
Air Blank	0.000	17:43
Control Test	0.080	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 06/10/2021 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 05/01/2021 03:09:52

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.162	03:05
Breath Volume	2.152	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.153	03:08
Breath Volume	1.714	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/04/2021 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 05/01/2021 04:21:03

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1 0.172	04:14	Volume Not Met
Breath Volume	0.210	04:14	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	04:19	No Sample Provided
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.079	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 05/04/2021 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 05/01/2021 22:48:24

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.141	22:43
Breath Volume	1.230	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.151	22:46
Breath Volume	1.554	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 05/04/2021 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/01/2021 23:29:44

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.081	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.351	23:24
Breath Volume	1.414	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.346	23:28
Breath Volume	1.515	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 05/04/2021 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: DAVALOS BERNAL

DATE/TIME OF TEST: 05/02/2021 05:55:47

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.080	05:42	
Air Blank	0.000	05:42	
Subject Sample	# 1 0.118	05:45	Volume Not Met
Breath Volume	0.023	05:45	
Air Blank	0.000	05:46	
Air Blank	0.000	05:48	
Subject Sample	# 2 0.115	05:51	
Breath Volume	1.972	05:51	
Air Blank	0.000	05:51	
Air Blank	0.000	05:53	
Subject Sample	# 3 0.109	05:54	
Breath Volume	1.468	05:54	
Air Blank	0.000	05:54	
Control Test	0.077	05:55	
Air Blank	0.000	05:55	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 05/04/2021 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/06/2021 00:59:53

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.150	00:55
Breath Volume	2.296	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.147	00:58
Breath Volume	2.523	00:58
Air Blank	0.000	00:58
Control Test	0.078	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 06/01/2021 21:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 05/06/2021 03:00:26

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.185	02:55
Breath Volume	3.031	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.182	02:58
Breath Volume	2.683	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 06/01/2021 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILL,JOHN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 05/06/2021 03:54:59

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.131	03:50
Breath Volume	1.289	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.139	03:53
Breath Volume	1.570	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BRYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SOMERS,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 05/07/2021 19:34:31

Diagnostics Check	OK	19:33	
Air Blank		19:34	RFI Detect
Air Blank		19:34	RFI Detect

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BRYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SOMERS,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 05/07/2021 19:44:51

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.081	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.146	19:39
Breath Volume	3.242	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.144	19:43
Breath Volume	2.910	19:43
Air Blank	0.000	19:43
Control Test	0.080	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BRYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CLAY,AARON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: DEAZEVEDO

DATE/TIME OF TEST: 05/13/2021 20:31:02

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.081	20:29	
Air Blank	0.000	20:29	
Subject Sample # 1	0.000	20:30	
Breath Volume	3.902	20:30	
Air Blank		20:30	RFI Detect
Air Blank		20:30	RFI Detect
Air Blank	0.000	20:31	

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BRYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CLAY,BRYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: DEAZEVEDO

DATE/TIME OF TEST: 05/13/2021 20:39:51

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:35
Breath Volume	3.574	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:38
Breath Volume	3.796	20:38
Air Blank	0.000	20:38
Control Test	0.081	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WICHMAN, MICHELLE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: SHAMBURG

DATE/TIME OF TEST: 05/18/2021 07:12:54

Diagnostics Check	OK	07:06
Air Blank	0.000	07:06
Control Test	0.080	07:06
Air Blank	0.000	07:07
Subject Sample # 1	0.000	07:08
Breath Volume	1.121	07:08
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample # 2	0.000	07:11
Breath Volume	1.910	07:11
Air Blank	0.000	07:11
Control Test	0.080	07:12
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHUM,JASON,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BERGMAN,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: CHUM

DATE/TIME OF TEST: 05/19/2021 08:41:43

Diagnostics Check	OK	08:34
Air Blank	0.000	08:34
Control Test	0.080	08:34
Air Blank	0.000	08:35
Subject Sample # 1	0.090	08:36
Breath Volume	1.894	08:36
Air Blank	0.000	08:37
Air Blank	0.000	08:39
Subject Sample # 2	0.084	08:40
Breath Volume	1.265	08:40
Air Blank	0.000	08:40
Control Test	0.077	08:41
Air Blank	0.000	08:41
Diagnostics Check	OK	08:41

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,J,G
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JOHNSON,DAVID
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BOSTIC

DATE/TIME OF TEST: 05/20/2021 16:14:15

Diagnostics Check	OK	16:08
Air Blank	0.000	16:08
Control Test	0.081	16:09
Air Blank	0.000	16:09
Subject Sample # 1	0.000	16:10
Breath Volume	2.390	16:10
Air Blank	0.000	16:10
Air Blank	0.000	16:12
Subject Sample # 2	0.000	16:12
Breath Volume	2.308	16:12
Air Blank	0.000	16:13
Control Test	0.078	16:13
Air Blank	0.000	16:14
Diagnostics Check	OK	16:14

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/20/2021 23:40:20

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.149	23:35
Breath Volume	1.160	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.144	23:38
Breath Volume	2.468	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 06/01/2021 22:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 05/21/2021 03:03:42

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.194	02:59
Breath Volume	1.640	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.202	03:02
Breath Volume	1.679	03:02
Air Blank	0.000	03:02
Control Test	0.079	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 06/01/2021 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: LASKOWSKI

DATE/TIME OF TEST: 05/22/2021 18:30:11

Diagnostics Check	OK	18:28	
Air Blank	0.000	18:29	
Control Test	0.080	18:29	
Air Blank		18:29	RFI Detect
Air Blank	0.000	18:30	

Data Download Date/Time: 06/01/2021 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI, PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: LASKOWSKI

DATE/TIME OF TEST: 05/22/2021 18:37:34

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.000	18:33
Breath Volume	1.957	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.000	18:36
Breath Volume	2.535	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:36
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 06/01/2021 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI, PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TIZZIO

DATE/TIME OF TEST: 05/29/2021 01:58:51

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.000	01:52
Breath Volume	1.847	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.000	01:57
Breath Volume	1.687	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 06/01/2021 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,E,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DUNKLE

DATE/TIME OF TEST: 05/30/2021 01:35:55

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 06/01/2021 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CANTELAR,JOAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/01/2021 00:31:32

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.229	00:26
Breath Volume	1.707	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.232	00:29
Breath Volume	1.597	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 06/02/2021 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: CUADROS

DATE/TIME OF TEST: 05/01/2021 04:23:15

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.082	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.018	04:18
Breath Volume	1.933	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.015	04:21
Breath Volume	2.003	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 06/02/2021 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JUSTIN,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MADRIGAL,ADRIAN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MAURIZ

DATE/TIME OF TEST: 05/02/2021 01:29:20

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.081	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1 0.148	01:19	
Breath Volume	1.796	01:19	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2 0.131	01:24	Volume Not Met
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.146	01:27	
Breath Volume	2.289	01:27	
Air Blank	0.000	01:28	
Control Test	0.079	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PEREZ CARDONA

DATE/TIME OF TEST: 05/02/2021 02:06:59

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.081	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.170	02:02
Breath Volume	2.000	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.164	02:05
Breath Volume	2.101	02:05
Air Blank	0.000	02:06
Control Test	0.076	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO,OMAR,E
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROMERO,OMAR
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/04/2021 22:38:06

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.081	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.201	22:33
Breath Volume	2.355	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.194	22:36
Breath Volume	2.152	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SANCHEZ,JUSTIN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ROCA VARGAS

DATE/TIME OF TEST: 05/06/2021 00:11:30

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.227	00:07
Breath Volume	2.511	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.220	00:09
Breath Volume	2.066	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JUSTIN,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: VAZQUEZ
GONZALEZ,WILLIAM
ARRESTING OFFICER AGENCY: CORAL GABLES PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female

OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: NEUWEG REYNA

DATE/TIME OF TEST: 05/06/2021 05:13:52

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.081	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:12	Subject Test Refused
Breath Volume	0.000	05:12	
Air Blank	0.000	05:12	
Control Test	0.081	05:13	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO VALDES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ACEVEDO VALDES,CRISTIAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: ROVIRA

DATE/TIME OF TEST: 05/07/2021 16:54:11

Diagnostics Check	OK	16:47
Air Blank	0.000	16:47
Control Test	0.081	16:48
Air Blank	0.000	16:48
Subject Sample # 1	0.307	16:49
Breath Volume	1.824	16:49
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 2	0.297	16:52
Breath Volume	1.761	16:52
Air Blank	0.000	16:53
Control Test	0.080	16:53
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: FALCON,IVAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SANCHEZ SANTOS

DATE/TIME OF TEST: 05/09/2021 00:32:14

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.203	00:27
Breath Volume	3.562	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.189	00:30
Breath Volume	1.273	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,RICHARD,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARIAS,RICHARD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CALA

DATE/TIME OF TEST: 05/15/2021 03:39:17

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.082	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.175	03:34
Breath Volume	2.195	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.172	03:37
Breath Volume	2.039	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANIO,PIERRE,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: LLANIO,PIERRE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: GARCIA JIMENEZ

DATE/TIME OF TEST: 05/15/2021 07:55:55

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.081	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.100	07:48
Breath Volume	1.390	07:48
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.100	07:54
Breath Volume	1.578	07:54
Air Blank	0.000	07:54
Control Test	0.081	07:55
Air Blank	0.000	07:55
Diagnostics Check	OK	07:55

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LEON,ANTONIO,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DE LEON,ANTONIO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/16/2021 20:25:44

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:23	
Control Test	0.081	20:23	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:24	
Control Test	0.081	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LALANE,LYNN,G
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: LALANE,LYNN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 05/17/2021 02:44:00

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.175	02:39
Breath Volume	2.261	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.175	02:42
Breath Volume	2.738	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,RICHARD,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARIAS,RICHARD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LABANINO

DATE/TIME OF TEST: 05/19/2021 01:17:46

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:14	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.082	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,RICHARD,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ARIAS,RICHARD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LABANINO

DATE/TIME OF TEST: 05/19/2021 01:30:07

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.163	01:25
Breath Volume	1.675	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.155	01:28
Breath Volume	1.101	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,RICHARD,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: THOMPSON,CHARLOTTE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PADOVANO

DATE/TIME OF TEST: 05/22/2021 04:51:37

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:47	
Control Test	0.082	04:48	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:50	
Control Test	0.082	04:51	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,LEON,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JOHNSON,LEON
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 05/23/2021 09:02:57

Diagnostics Check	OK	08:56
Air Blank	0.000	08:57
Control Test	0.081	08:57
Air Blank	0.000	08:57
Subject Sample # 1	0.153	08:58
Breath Volume	2.398	08:58
Air Blank	0.000	08:59
Air Blank	0.000	09:00
Subject Sample # 2	0.155	09:01
Breath Volume	1.355	09:01
Air Blank	0.000	09:02
Control Test	0.080	09:02
Air Blank	0.000	09:02
Diagnostics Check	OK	09:02

Data Download Date/Time: 06/02/2021 21:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERO ORTIZ,JOHAN,S
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BARRERO ORTIZ,JOHAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 05/29/2021 05:38:04

Diagnostics Check	OK	05:24	
Air Blank	0.000	05:24	
Control Test	0.083	05:24	
Air Blank	0.000	05:25	
Subject Sample	# 1 0.195	05:28	
Breath Volume	2.199	05:28	
Air Blank	0.000	05:29	
Air Blank	0.000	05:30	
Subject Sample	# 2 0.174	05:31	
Breath Volume	1.332	05:31	
Air Blank	0.000	05:32	
Air Blank	0.000	05:34	
Subject Sample	# 3 0.170	05:36	
Breath Volume	2.039	05:36	
Air Blank		05:37	Purge Fail
Air Blank	0.000	05:38	

Data Download Date/Time: 06/02/2021 21:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERO ORTIZ,JOHAN,S
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BARRERO ORTIZ,JOHAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 05/29/2021 05:44:45

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.077	05:42	
Air Blank	0.000	05:42	
Subject Sample	# 1 0.178	05:43	
Breath Volume	2.156	05:43	
Air Blank		05:44	Purge Fail
Air Blank	0.000	05:44	

Data Download Date/Time: 06/02/2021 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERO ORTIZ,JOHAN,S
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BARRERO ORTIZ,JOHAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 05/29/2021 06:12:44

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.079	06:07
Air Blank	0.000	06:07
Subject Sample # 1	0.176	06:08
Breath Volume	2.355	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 2	0.157	06:10
Breath Volume	1.320	06:10
Air Blank	0.000	06:11
Control Test	0.079	06:12
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 06/02/2021 21:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANDON,ERICK,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: TORUNO,KARYN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: DOMAS BORGES

DATE/TIME OF TEST: 05/30/2021 23:27:21

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.082	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Control Test	0.081	23:26	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 06/02/2021 21:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,JOSEPH,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CHAPMAN,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BLACKFORD

DATE/TIME OF TEST: 05/27/2021 22:02:40

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.173	21:58
Breath Volume	2.476	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.161	22:00
Breath Volume	2.507	22:00
Air Blank	0.000	22:01
Control Test	0.077	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 06/11/2021 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HERNANDEZ GARCIA

DATE/TIME OF TEST: 05/02/2021 01:22:51

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:18
Breath Volume	1.566	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:21
Breath Volume	2.347	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 05/17/2021 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: TAITE

DATE/TIME OF TEST: 05/06/2021 02:15:58

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.174	02:09
Breath Volume	1.101	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.160	02:14
Breath Volume	1.132	02:14
Air Blank	0.000	02:15
Control Test	0.077	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 05/17/2021 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/30/2021 00:35:21

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.214	00:28
Breath Volume	1.968	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.197	00:33
Breath Volume	1.601	00:33
Air Blank	0.000	00:34
Control Test	0.076	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 06/14/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: BUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/30/2021 18:16:33

Diagnostics Check	OK	18:13	
Air Blank	0.000	18:13	
Control Test	0.078	18:14	
Air Blank	0.000	18:14	
Subject Sample	# 1	18:15	RFI Detect
Breath Volume	0.000	18:15	
Air Blank	0.000	18:15	
Control Test	0.078	18:15	
Air Blank	0.000	18:16	
Diagnostics Check	OK	18:16	

Data Download Date/Time: 06/14/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: BUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/30/2021 18:25:17

Diagnostics Check	OK	18:18
Air Blank	0.000	18:18
Control Test	0.079	18:19
Air Blank	0.000	18:19
Subject Sample # 1	0.220	18:20
Breath Volume	3.484	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:22
Subject Sample # 2	0.207	18:23
Breath Volume	2.609	18:23
Air Blank	0.000	18:24
Control Test	0.078	18:24
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 06/14/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIDSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DAVIDSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: BUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: POOL

DATE/TIME OF TEST: 05/30/2021 18:51:48

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:40	
Control Test	0.078	18:41	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:44	No Sample Provided
Breath Volume	0.000	18:44	
Air Blank	0.000	18:45	
Air Blank	0.000	18:47	
Subject Sample	# 2	18:50	No Sample Provided
Breath Volume	0.000	18:50	
Air Blank	0.000	18:50	
Control Test	0.078	18:51	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 06/14/2021 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: BUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 05/30/2021 19:16:10

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:13	
Control Test	0.079	19:14	
Air Blank	0.000	19:14	
Subject Sample	# 1	19:14	Subject Test Refused
Breath Volume	0.000	19:14	
Air Blank	0.000	19:15	
Control Test	0.079	19:15	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 06/14/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 05/03/2021 22:43:11

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.081	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1 0.119	22:38	
Breath Volume	2.781	22:38	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample	# 2	22:41	Interferent Detect
Breath Volume	2.636	22:41	
Air Blank	0.000	22:42	
Control Test	0.080	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 05/17/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 05/03/2021 22:51:08

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.077	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.117	22:46
Breath Volume	2.457	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.121	22:49
Breath Volume	1.191	22:49
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 05/17/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LEZAMA-MONTALVAN

DATE/TIME OF TEST: 05/08/2021 02:18:49

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.206	02:14
Breath Volume	1.378	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.202	02:17
Breath Volume	2.886	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 05/17/2021 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BLAIR, TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: WEATHERS

DATE/TIME OF TEST: 05/28/2021 02:26:33

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:24	Interferent Detect
Breath Volume	1.488	02:24	
Air Blank	0.000	02:25	
Control Test	0.075	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 06/25/2021 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BLAIR, TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: WEATHERS

DATE/TIME OF TEST: 05/28/2021 02:32:53

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:30	Interferent Detect
Breath Volume	1.574	02:30	
Air Blank	0.000	02:31	
Control Test	0.076	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	Fail	02:32	Voltage/Current Test
Air Blank	0.000	02:32	

Data Download Date/Time: 06/25/2021 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HARLEYJR

DATE/TIME OF TEST: 05/29/2021 23:25:02

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.000	23:20
Breath Volume	2.222	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.000	23:23
Breath Volume	3.226	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:25

Data Download Date/Time: 06/25/2021 04:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/01/2021 19:55:07

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.078	19:46
Air Blank	0.000	19:46
Subject Sample # 1	0.349	19:47
Breath Volume	2.328	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.290	19:49
Breath Volume	1.117	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 3	0.334	19:53
Breath Volume	1.937	19:53
Air Blank	0.000	19:54
Control Test	0.076	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 05/28/2021 18:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: O BRIEN

DATE/TIME OF TEST: 05/02/2021 04:12:17

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.078	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.078	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 05/28/2021 18:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 05/16/2021 01:06:56

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.151	01:02
Breath Volume	2.031	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.146	01:05
Breath Volume	1.617	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 05/28/2021 18:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SLOVER

DATE/TIME OF TEST: 05/26/2021 01:30:40

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.135	01:25
Breath Volume	1.515	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.154	01:28
Breath Volume	4.183	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 05/28/2021 18:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULAIMAN, RAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER, ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MOFFATT

DATE/TIME OF TEST: 05/29/2021 01:35:58

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.151	01:31
Breath Volume	1.402	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.151	01:34
Breath Volume	1.289	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/01/2021 16:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: DOWD-ALUSH

DATE/TIME OF TEST: 05/01/2021 20:43:20

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.203	20:38
Breath Volume	3.167	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.192	20:41
Breath Volume	2.703	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 05/11/2021 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: THORNBRUGH,ELAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/02/2021 00:29:27

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.174	00:24
Breath Volume	2.425	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.172	00:27
Breath Volume	2.921	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 05/11/2021 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: SCHWARTZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 05/08/2021 20:27:25

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.133	20:22
Breath Volume	3.386	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.119	20:25
Breath Volume	1.332	20:25
Air Blank	0.000	20:26
Control Test	0.079	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 05/11/2021 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: OGGERI,TIMOTHY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: ZUTANT

DATE/TIME OF TEST: 05/11/2021 22:02:18

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:48	
Control Test	0.080	21:48	
Air Blank	0.000	21:49	
Subject Sample	# 1 0.000	21:50	
Breath Volume	1.269	21:50	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 2 0.000	21:56	Volume Not Met
Breath Volume	0.898	21:56	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample	# 3 0.000	22:00	
Breath Volume	1.238	22:00	
Air Blank	0.000	22:01	
Control Test	0.079	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEXANDER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/15/2021 03:04:26

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:59	No Sample Provided
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 2	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATKINS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WAITE,K
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: PARAMO OLIVAS

DATE/TIME OF TEST: 05/20/2021 05:00:07

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.168	04:55
Breath Volume	2.820	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.164	04:58
Breath Volume	1.960	04:58
Air Blank	0.000	04:59
Control Test	0.078	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYSE,EVELYN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BRISCO

DATE/TIME OF TEST: 05/21/2021 20:19:01

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.079	20:13
Air Blank	0.000	20:13
Subject Sample # 1	0.197	20:14
Breath Volume	2.265	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:16
Subject Sample # 2	0.194	20:17
Breath Volume	2.609	20:17
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:19

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: SCHWARTZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CLIFTON

DATE/TIME OF TEST: 05/24/2021 02:48:30

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: JUSTICE

DATE/TIME OF TEST: 05/24/2021 03:05:05

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.085	03:00
Breath Volume	1.953	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.087	03:03
Breath Volume	2.546	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: OLAH

DATE/TIME OF TEST: 05/27/2021 23:19:39

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:15
Breath Volume	1.902	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:18
Breath Volume	1.390	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: THORNBRUGH,ELAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/29/2021 06:43:15

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.080	06:37
Air Blank	0.000	06:37
Subject Sample # 1	0.142	06:38
Breath Volume	2.320	06:38
Air Blank	0.000	06:39
Air Blank	0.000	06:40
Subject Sample # 2	0.130	06:41
Breath Volume	1.253	06:41
Air Blank	0.000	06:42
Control Test	0.080	06:42
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: MARIN BUITRAGO

DATE/TIME OF TEST: 05/29/2021 18:50:10

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.059	18:45
Breath Volume	1.582	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.057	18:48
Breath Volume	1.390	18:48
Air Blank	0.000	18:49
Control Test	0.079	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: HEFTY

DATE/TIME OF TEST: 05/30/2021 00:22:07

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.149	00:17
Breath Volume	3.386	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.156	00:20
Breath Volume	3.183	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 06/09/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/03/2021 21:54:11

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.080	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1 0.168	21:45	Slope Not Level
Breath Volume	1.304	21:45	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample	# 2 0.172	21:49	
Breath Volume	2.097	21:49	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 3 0.180	21:52	
Breath Volume	1.980	21:52	
Air Blank	0.000	21:53	
Control Test	0.080	21:53	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 05/17/2021 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 05/15/2021 02:59:48

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.070	02:55
Breath Volume	2.390	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.060	02:58
Breath Volume	1.781	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 05/17/2021 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1076482
BREATH TEST OPERATOR: RUBEL,NEIL,F	DRY GAS EXPIRY DATE: 01/15/2022
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: MAYO,RICHARD	ARREST TIME: 11:53
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Female
LAST AGENCY INSPECTION: 05/12/2021	OBSERVATION PERIOD BEGAN: 11:53

SUBJECT'S LAST NAME: KEEN	DATE/TIME OF TEST: 05/21/2021	12:19:50
---------------------------	-------------------------------	----------

Diagnostics Check	OK	12:13	
Air Blank	0.000	12:13	
Control Test	0.080	12:13	
Air Blank	0.000	12:14	
Subject Sample	# 1 0.184	12:15	
Breath Volume	1.453	12:15	
Air Blank	0.000	12:16	
Air Blank	0.000	12:18	
Subject Sample	# 2 0.170	12:19	
Breath Volume	1.648	12:19	
Air Blank		12:19	RFI Detect
Air Blank	0.000	12:19	

Data Download Date/Time: 06/09/2021 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1076482
BREATH TEST OPERATOR: RUBEL,NEIL,F	DRY GAS EXPIRY DATE: 01/15/2022
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: MAYO,RICHARD	ARREST TIME: 11:53
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Female
LAST AGENCY INSPECTION: 05/12/2021	OBSERVATION PERIOD BEGAN: 11:53

SUBJECT'S LAST NAME: KEEN	DATE/TIME OF TEST: 05/21/2021 12:30:55
---------------------------	--

Diagnostics Check	OK	12:22
Air Blank	0.000	12:23
Control Test	0.079	12:23
Air Blank	0.000	12:23
Subject Sample # 1	0.186	12:24
Breath Volume	1.113	12:24
Air Blank	0.000	12:25
Air Blank	0.000	12:27
Subject Sample # 2	0.168	12:29
Breath Volume	2.738	12:29
Air Blank	0.000	12:29
Control Test	0.080	12:30
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 06/09/2021 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BLANTON, T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 05/01/2021 23:46:41

Diagnostics Check	OK	23:45	
Air Blank		23:46	RFI Detect
Air Blank	0.000	23:46	

Data Download Date/Time: 05/06/2021 19:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BLANTON, T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 05/01/2021 23:55:31

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:50
Breath Volume	1.289	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.000	23:54
Breath Volume	1.183	23:54
Air Blank	0.000	23:54
Control Test	0.078	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 05/06/2021 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GUERNON

DATE/TIME OF TEST: 05/03/2021 00:56:30

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.195	00:51
Breath Volume	2.070	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.205	00:54
Breath Volume	1.742	00:54
Air Blank	0.000	00:55
Control Test	0.078	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 05/06/2021 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/11/2021 21:07:09

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.077	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.000	21:02
Breath Volume	2.132	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.000	21:05
Breath Volume	2.695	21:05
Air Blank	0.000	21:06
Control Test	0.077	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 06/22/2021 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 05/12/2021 00:55:40

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.077	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.166	00:50
Breath Volume	2.410	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.166	00:53
Breath Volume	1.234	00:53
Air Blank	0.000	00:54
Control Test	0.077	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/22/2021 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: SWEAT

DATE/TIME OF TEST: 05/15/2021 02:38:53

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.116	02:34
Breath Volume	2.128	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.116	02:37
Breath Volume	2.246	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 06/22/2021 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: PESHEK

DATE/TIME OF TEST: 05/20/2021 01:29:53

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.147	01:25
Breath Volume	2.339	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.149	01:28
Breath Volume	2.941	01:28
Air Blank	0.000	01:28
Control Test	0.077	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 06/22/2021 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MCGLOCKING

DATE/TIME OF TEST: 05/26/2021 01:42:51

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.077	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.109	01:37
Breath Volume	2.332	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.109	01:41
Breath Volume	1.464	01:41
Air Blank	0.000	01:41
Control Test	0.078	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 06/22/2021 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 05/29/2021 21:58:59

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.078	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:56	Interferent Detect
Breath Volume	2.640	21:56	
Air Blank	0.000	21:58	
Control Test	0.084	21:58	Interferent Detect
Air Blank	0.000	21:58	

Data Download Date/Time: 06/22/2021 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: HARRIOTT

DATE/TIME OF TEST: 05/02/2021 01:53:54

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.080	01:49
Breath Volume	2.441	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.080	01:52
Breath Volume	2.792	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 05/06/2021 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 04/11/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MUNOZ JARAMILLO

DATE/TIME OF TEST: 05/06/2021 00:24:14

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.079	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1 0.077	00:13	
Breath Volume	1.152	00:13	
Air Blank	0.000	00:14	
Air Blank	0.000	00:16	
Subject Sample	# 2 0.066	00:19	Volume Not Met
Breath Volume	0.488	00:19	
Air Blank	0.000	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 3 0.074	00:22	
Breath Volume	1.105	00:22	
Air Blank	0.000	00:23	
Control Test	0.078	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 05/06/2021 19:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PEREZ-MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: CHILDERS

DATE/TIME OF TEST: 05/15/2021 18:29:15

Diagnostics Check	OK	18:19
Air Blank	0.000	18:19
Control Test	0.081	18:19
Air Blank	0.000	18:20
Subject Sample # 1	0.202	18:21
Breath Volume	1.925	18:21
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.179	18:24
Breath Volume	1.773	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 3	0.181	18:27
Breath Volume	1.992	18:27
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 05/17/2021 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TERUZ,GABRIEL,S

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: PEREZ-MORALES,JAVIER

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1111579

DRY GAS EXPIRY DATE: 03/19/2022

VIOLATION CODE: DUI

ARREST TIME: 02:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: NARANJO SANCHEZ

DATE/TIME OF TEST: 05/30/2021 03:35:26

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.144	03:25	Volume Not Met
Breath Volume	0.406	03:25	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2 0.168	03:28	
Breath Volume	1.714	03:28	
Air Blank	0.000	03:28	
Air Blank	0.000	03:30	
Subject Sample	# 3 0.151	03:33	Volume Not Met
Breath Volume	0.074	03:33	
Air Blank	0.000	03:34	
Control Test	0.081	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 06/14/2021 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TERUZ,GABRIEL,S

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: PEREZ-MORALES,JAVIER

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1111579

DRY GAS EXPIRY DATE: 03/19/2022

VIOLATION CODE: DUI

ARREST TIME: 04:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: LARA VALLE

DATE/TIME OF TEST: 05/31/2021 05:33:02

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.082	05:20	
Air Blank	0.000	05:20	
Subject Sample	# 1 0.200	05:21	
Breath Volume	1.613	05:21	
Air Blank	0.000	05:22	
Air Blank	0.000	05:24	
Subject Sample	# 2 0.208	05:27	Volume Not Met
Breath Volume	0.062	05:27	
Air Blank	0.000	05:28	
Air Blank	0.000	05:30	
Subject Sample	# 3 0.190	05:31	
Breath Volume	1.484	05:31	
Air Blank	0.000	05:32	
Control Test	0.079	05:32	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 06/14/2021 12:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: VIVEROS

DATE/TIME OF TEST: 05/01/2021 06:31:44

Diagnostics Check	OK	06:28	
Air Blank	0.000	06:29	
Control Test	0.078	06:29	
Air Blank	0.000	06:29	
Subject Sample	# 1	06:30	Subject Test Refused
Breath Volume	0.000	06:30	
Air Blank	0.000	06:30	
Control Test	0.077	06:31	
Air Blank	0.000	06:31	
Diagnostics Check	OK	06:31	

Data Download Date/Time: 05/18/2021 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOFF,SCOTT,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RUDOFF,SCOTT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: POSE

DATE/TIME OF TEST: 05/08/2021 00:57:27

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:50	
Subject Sample # 1	0.161	00:53	
Breath Volume	2.453	00:53	
Air Blank	0.000	00:54	
Air Blank	0.000	00:55	
		00:55	Improper Sample
Air Blank	0.000	00:56	
Control Test	0.078	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 05/18/2021 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOFF,SCOTT,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RUDOFF,SCOTT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: POSE

DATE/TIME OF TEST: 05/08/2021 01:06:01

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.077	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.165	01:01
Breath Volume	3.097	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.161	01:04
Breath Volume	2.558	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:06

Data Download Date/Time: 05/18/2021 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 04/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: STROZIER

DATE/TIME OF TEST: 05/08/2021 03:04:06

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.143	02:59
Breath Volume	1.648	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.143	03:02
Breath Volume	1.640	03:02
Air Blank	0.000	03:03
Control Test	0.077	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/18/2021 05:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BEAUPLAN

DATE/TIME OF TEST: 05/20/2021 05:32:21

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.200	05:27
Breath Volume	1.777	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.208	05:30
Breath Volume	1.589	05:30
Air Blank	0.000	05:31
Control Test	0.078	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 06/03/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/22/2021 02:48:41

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.077	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 06/03/2021 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUEZ,LEWIS,P
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LARSH,MATTHEW
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 05/22/2021 23:38:34

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 06/03/2021 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARSH, MATTHEW,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LARSH, MATTHEW
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 05/22/2021 23:44:11

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.079	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.077	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 06/03/2021 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SANDIANES CAMINERO

DATE/TIME OF TEST: 05/29/2021 23:14:40

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.077	23:12	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:13	
Control Test	0.076	23:14	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 06/03/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: IBALMEA

DATE/TIME OF TEST: 05/30/2021 02:14:44

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.077	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.073	02:09
Breath Volume	3.109	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.072	02:13
Breath Volume	1.847	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 06/03/2021 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHERSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: MARTINIANO DA SILVA

DATE/TIME OF TEST: 05/01/2021 21:02:17

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.127	20:57
Breath Volume	2.140	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.124	21:00
Breath Volume	2.132	21:00
Air Blank	0.000	21:01
Control Test	0.079	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DOREUS

DATE/TIME OF TEST: 05/02/2021 00:53:20

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.078	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.226	00:48
Breath Volume	1.589	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.226	00:51
Breath Volume	1.328	00:51
Air Blank	0.000	00:52
Control Test	0.075	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BUFF

DATE/TIME OF TEST: 05/02/2021 01:35:17

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.115	01:30
Breath Volume	1.742	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.116	01:33
Breath Volume	2.011	01:33
Air Blank	0.000	01:34
Control Test	0.076	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BROWN,EDDIE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 05/02/2021 05:12:01

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:57	
Control Test	0.078	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1 0.114	05:01	Volume Not Met
Breath Volume	0.199	05:01	
Air Blank	0.000	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 2 0.145	05:05	
Breath Volume	1.242	05:05	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample	# 3 0.140	05:10	
Breath Volume	1.312	05:10	
Air Blank	0.000	05:11	
Control Test	0.076	05:11	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: PALMA AGUILAR

DATE/TIME OF TEST: 05/02/2021 06:18:27

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.077	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.123	06:13
Breath Volume	1.523	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.121	06:16
Breath Volume	1.949	06:16
Air Blank	0.000	06:17
Control Test	0.078	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 05/04/2021 02:55:38

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.193	02:50
Breath Volume	2.070	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.189	02:54
Breath Volume	1.398	02:54
Air Blank	0.000	02:54
Control Test	0.075	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: AUDAIN

DATE/TIME OF TEST: 05/06/2021 23:22:31

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.000	23:18
Breath Volume	1.382	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.000	23:21
Breath Volume	1.503	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SURINACH

DATE/TIME OF TEST: 05/08/2021 04:04:18

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.130	03:59
Breath Volume	2.644	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.127	04:02
Breath Volume	2.937	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CARILLO RAMIREZ

DATE/TIME OF TEST: 05/10/2021 00:14:05

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.153	00:09
Breath Volume	1.968	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.155	00:12
Breath Volume	2.093	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 05/12/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TWIDDY,RYAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: FELDER

DATE/TIME OF TEST: 05/01/2021 04:13:20

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.083	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.153	04:05	Volume Not Met
Breath Volume	0.085	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 2 0.170	04:08	
Breath Volume	2.210	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample	# 3 0.170	04:11	
Breath Volume	2.605	04:11	
Air Blank	0.000	04:12	
Control Test	0.080	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 06/07/2021 19:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 05/01/2021 22:33:52

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.082	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.200	22:28
Breath Volume	1.277	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.202	22:32
Breath Volume	1.300	22:32
Air Blank	0.000	22:32
Control Test	0.081	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MYRICK,JASMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: HOMAN

DATE/TIME OF TEST: 05/04/2021 02:35:12

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Control Test	0.081	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRICKS,JOHN,M
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HENDRICKS,JOHN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: FRYER

DATE/TIME OF TEST: 05/05/2021 17:11:51

Diagnostics Check	OK	17:04
Air Blank	0.000	17:04
Control Test	0.082	17:05
Air Blank	0.000	17:05
Subject Sample # 1	0.284	17:06
Breath Volume	2.218	17:06
Air Blank	0.000	17:07
Air Blank	0.000	17:08
Subject Sample # 2	0.293	17:10
Breath Volume	1.777	17:10
Air Blank	0.000	17:10
Control Test	0.081	17:11
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOHLIN,MARK,T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOHLIN,MARK
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GUIDRY

DATE/TIME OF TEST: 05/09/2021 02:44:08

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.153	02:39
Breath Volume	1.750	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.149	02:42
Breath Volume	1.183	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 05/09/2021 20:49:51

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:41	
Control Test	0.081	20:41	
Air Blank	0.000	20:41	
Subject Sample	# 1	20:45	No Sample Provided
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Air Blank	0.000	20:47	
Subject Sample	# 2	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Control Test	0.081	20:49	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HAIL,WILLIAM
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BAYLESS JR

DATE/TIME OF TEST: 05/11/2021 04:45:02

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.082	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.082	04:44	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS, ABNER, K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: FRIES, NICHOLAS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MONGE

DATE/TIME OF TEST: 05/15/2021 04:49:22

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.081	04:38	
Air Blank	0.000	04:38	
Subject Sample	# 1 0.195	04:42	Volume Not Met
Breath Volume	0.585	04:42	
Air Blank	0.000	04:42	
Air Blank	0.000	04:44	
Subject Sample	# 2 0.218	04:47	Volume Not Met
Breath Volume	0.656	04:47	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS, ABNER, K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: FRIES, NICHOLAS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MONGE

DATE/TIME OF TEST: 05/15/2021 04:59:25

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.234	04:53
Breath Volume	2.191	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.219	04:57
Breath Volume	1.324	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOHLIN,MARK,T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: DICK

DATE/TIME OF TEST: 05/29/2021 04:55:08

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.082	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:53	Subject Test Refused
Breath Volume	0.000	04:53	
Air Blank	0.000	04:54	
Control Test	0.082	04:54	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOHLIN,MARK,T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOHLIN,MARK
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 04/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/29/2021 21:51:35

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.082	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.219	21:45
Breath Volume	2.667	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.218	21:49
Breath Volume	1.925	21:49
Air Blank	0.000	21:50
Control Test	0.081	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 06/07/2021 19:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: PENNAY GONDUSKY

DATE/TIME OF TEST: 05/01/2021 18:03:03

Diagnostics Check	OK	17:56
Air Blank	0.000	17:56
Control Test	0.077	17:57
Air Blank	0.000	17:57
Subject Sample # 1	0.000	17:58
Breath Volume	2.195	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.000	18:01
Breath Volume	2.820	18:01
Air Blank	0.000	18:02
Control Test	0.078	18:02
Air Blank	0.000	18:02
Diagnostics Check	OK	18:03

Data Download Date/Time: 05/12/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: FORRE

DATE/TIME OF TEST: 05/05/2021 03:23:13

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.077	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.078	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 05/12/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: IUCULANO

DATE/TIME OF TEST: 05/05/2021 23:35:26

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.077	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.071	23:29
Breath Volume	1.703	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.072	23:33
Breath Volume	1.937	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 05/12/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 05/06/2021 03:33:25

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.078	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.110	03:26	Volume Not Met
Breath Volume	0.687	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2 0.105	03:31	Slope Not Level
Breath Volume	1.503	03:31	
Air Blank	0.000	03:32	
Control Test	0.077	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 05/12/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: KLAR

DATE/TIME OF TEST: 05/07/2021 00:42:42

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.077	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.077	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 05/12/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WERNERT

DATE/TIME OF TEST: 05/12/2021 03:05:33

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.077	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 05/12/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER, ANAYA, S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HEPBURN, ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 970670
DRY GAS EXPIRY DATE: 06/07/2021
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/28/2021 19:43:03

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.079	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.000	19:38
Breath Volume	1.601	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.000	19:41
Breath Volume	1.507	19:41
Air Blank	0.000	19:42
Control Test	0.079	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:43

Data Download Date/Time: 06/02/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/09/2021 04:14:24

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.190	04:09
Breath Volume	1.148	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.189	04:12
Breath Volume	2.882	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 05/12/2021 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: RAU

DATE/TIME OF TEST: 05/31/2021 17:37:56

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.077	17:32
Air Blank	0.000	17:32
Subject Sample # 1	0.088	17:33
Breath Volume	4.667	17:33
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.087	17:36
Breath Volume	3.890	17:36
Air Blank	0.000	17:37
Control Test	0.078	17:37
Air Blank	0.000	17:37
Diagnostics Check	OK	17:37

Data Download Date/Time: 06/02/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/31/2021 23:49:42

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.078	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:48	
Control Test	0.079	23:49	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 06/02/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: FERACOTTA

DATE/TIME OF TEST: 05/15/2021 00:14:18

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.080	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1 0.154	00:10	
Breath Volume	2.171	00:10	
Air Blank	0.000	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 2	00:12	RFI Detect
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 06/02/2021 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: FERACOTTA

DATE/TIME OF TEST: 05/15/2021 00:22:54

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.159	00:18
Breath Volume	2.164	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.153	00:21
Breath Volume	1.984	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 06/02/2021 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 05/20/2021 03:22:40

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.148	03:17
Breath Volume	3.378	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.159	03:21
Breath Volume	2.496	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 06/02/2021 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: GEBHARDT

DATE/TIME OF TEST: 05/30/2021 16:05:38

Diagnostics Check	OK	16:02	
Air Blank	0.000	16:02	
Control Test	0.079	16:03	
Air Blank	0.000	16:03	
Subject Sample	# 1	16:04	Subject Test Refused
Breath Volume	0.000	16:04	
Air Blank	0.000	16:04	
Control Test	0.079	16:05	
Air Blank	0.000	16:05	
Diagnostics Check	OK	16:05	

Data Download Date/Time: 06/02/2021 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 05/01/2021 00:31:23

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.146	00:26
Breath Volume	1.480	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.143	00:29
Breath Volume	1.792	00:29
Air Blank	0.000	00:30
Control Test	0.077	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 05/11/2021 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LIMA,J
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FEE

DATE/TIME OF TEST: 05/01/2021 03:55:26

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.000	03:51
Breath Volume	2.761	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.000	03:54
Breath Volume	2.398	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 05/11/2021 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,STEPHEN,J
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MORALES,G
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 05/01/2021 11:49:33

Diagnostics Check	OK	11:43
Air Blank	0.000	11:43
Control Test	0.078	11:44
Air Blank	0.000	11:44
Subject Sample # 1	0.019	11:45
Breath Volume	2.410	11:45
Air Blank	0.000	11:45
Air Blank	0.000	11:47
Subject Sample # 2	0.018	11:48
Breath Volume	2.562	11:48
Air Blank	0.000	11:48
Control Test	0.078	11:48
Air Blank	0.000	11:49
Diagnostics Check	OK	11:49

Data Download Date/Time: 05/11/2021 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: JOHNSON,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 05/01/2021 21:15:51

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.078	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.268	21:10
Breath Volume	3.554	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.279	21:14
Breath Volume	3.648	21:14
Air Blank	0.000	21:14
Control Test	0.078	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 05/11/2021 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (SW REGION)
ARRESTING OFFICER: PLUSSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (SW REGION)
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: BUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: RUIZ PEREZ

DATE/TIME OF TEST: 05/01/2021 21:26:36

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.077	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.147	21:21
Breath Volume	2.296	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.153	21:24
Breath Volume	2.203	21:24
Air Blank	0.000	21:25
Control Test	0.077	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASH,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MIRELES

DATE/TIME OF TEST: 05/02/2021 04:25:32

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.078	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.145	04:20
Breath Volume	3.804	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.138	04:23
Breath Volume	1.523	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (SW REGION)
ARRESTING OFFICER: PLUSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (SW REGION)
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: BUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/02/2021 17:19:13

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.078	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.093	17:14
Breath Volume	3.519	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.091	17:17
Breath Volume	1.851	17:17
Air Blank	0.000	17:18
Control Test	0.078	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: PLUSSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 05/02/2021 17:29:18

Diagnostics Check	OK	17:23
Air Blank	0.000	17:23
Control Test	0.078	17:23
Air Blank	0.000	17:24
Subject Sample # 1	0.093	17:24
Breath Volume	3.707	17:24
Air Blank	0.000	17:25
Air Blank	0.000	17:27
Subject Sample # 2	0.089	17:27
Breath Volume	2.593	17:27
Air Blank	0.000	17:28
Control Test	0.077	17:28
Air Blank	0.000	17:29
Diagnostics Check	OK	17:29

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,MAXWELL,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: JOHNSON,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SAVELL

DATE/TIME OF TEST: 05/03/2021 02:38:05

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.078	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.648	02:36	
Air Blank	0.000	02:37	
Control Test	0.077	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECLERC,A,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MAHOLTZ,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:21

SUBJECT'S LAST NAME: ANTOINE

DATE/TIME OF TEST: 05/04/2021 09:56:32

Diagnostics Check	OK	09:48
Air Blank	0.000	09:48
Control Test	0.078	09:49
Air Blank	0.000	09:49
Subject Sample # 1	0.000	09:51
Breath Volume	2.316	09:51
Air Blank	0.000	09:51
Air Blank	0.000	09:53
Subject Sample # 2	0.000	09:55
Breath Volume	1.976	09:55
Air Blank	0.000	09:55
Control Test	0.078	09:55
Air Blank	0.000	09:56
Diagnostics Check	OK	09:56

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLINGS,EMANUEL,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: HORNBECK,DANNY
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: HAUSIN

DATE/TIME OF TEST: 05/07/2021 12:18:38

Diagnostics Check	OK	12:11
Air Blank	0.000	12:11
Control Test	0.078	12:12
Air Blank	0.000	12:12
Subject Sample # 1	0.250	12:13
Breath Volume	2.242	12:13
Air Blank	0.000	12:14
Air Blank	0.000	12:16
Subject Sample # 2	0.256	12:17
Breath Volume	2.628	12:17
Air Blank	0.000	12:17
Control Test	0.077	12:18
Air Blank	0.000	12:18
Diagnostics Check	OK	12:18

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT-FORT,PAUL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: AUSTIN,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: CABELLO

DATE/TIME OF TEST: 05/07/2021 16:38:34

Diagnostics Check	OK	16:31
Air Blank	0.000	16:32
Control Test	0.078	16:32
Air Blank	0.000	16:33
Subject Sample # 1	0.275	16:33
Breath Volume	2.625	16:33
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.277	16:36
Breath Volume	2.773	16:36
Air Blank	0.000	16:37
Control Test	0.075	16:37
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: DIKEMAN,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119080A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/08/2021 03:44:36

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.130	03:40
Breath Volume	2.902	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.127	03:43
Breath Volume	3.074	03:43
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CARBAJAL,T
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/10/2021 03:31:59

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.155	03:26
Breath Volume	2.890	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.160	03:30
Breath Volume	4.281	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,D,S
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24119070A1
DRY GAS EXPIRY DATE: 11/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 05/10/2021 10:44:42

Diagnostics Check	OK	10:38
Air Blank	0.000	10:38
Control Test	0.078	10:39
Air Blank	0.000	10:39
Subject Sample # 1	0.029	10:40
Breath Volume	3.121	10:40
Air Blank	0.000	10:40
Air Blank	0.000	10:42
Subject Sample # 2	0.028	10:43
Breath Volume	2.718	10:43
Air Blank	0.000	10:43
Control Test	0.078	10:44
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 05/11/2021 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZERPA,EDWIN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: RAMOSGARCIA

DATE/TIME OF TEST: 05/29/2021 19:50:11

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.079	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.240	19:45
Breath Volume	3.046	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:47
Subject Sample # 2	0.236	19:48
Breath Volume	3.289	19:48
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 06/09/2021 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,BRIAN,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PUTMAN,BRIAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 05/02/2021 04:32:52

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.000	04:28
Breath Volume	2.746	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.000	04:31
Breath Volume	2.960	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STEALEY,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MCCARRON

DATE/TIME OF TEST: 05/03/2021 03:46:00

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.079	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:41	No Sample Provided
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 2	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.077	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HAUSE,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 05/08/2021 19:29:23

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.140	19:24
Breath Volume	4.292	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:26
Subject Sample # 2	0.131	19:27
Breath Volume	3.535	19:27
Air Blank	0.000	19:28
Control Test	0.079	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BARLEY,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HOETING

DATE/TIME OF TEST: 05/10/2021 02:33:38

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1 0.177	02:23	
Breath Volume	2.062	02:23	
Air Blank	0.000	02:24	
Air Blank	0.000	02:26	
Subject Sample	# 2 0.149	02:29	Volume Not Met
Breath Volume	0.226	02:29	
Air Blank	0.000	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 3	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.078	02:33	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HEGWOOD,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 05/14/2021 17:39:27

Diagnostics Check	OK	17:32
Air Blank	0.000	17:32
Control Test	0.079	17:33
Air Blank	0.000	17:33
Subject Sample # 1	0.219	17:34
Breath Volume	1.386	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.234	17:37
Breath Volume	1.988	17:37
Air Blank	0.000	17:38
Control Test	0.079	17:38
Air Blank	0.000	17:39
Diagnostics Check	OK	17:39

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 05/15/2021 20:02:05

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.080	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.000	19:57
Breath Volume	2.460	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.000	20:00
Breath Volume	2.101	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:01
Diagnostics Check	OK	20:02

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HOLLINGSWORTH,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/16/2021 03:28:26

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.233	03:23
Breath Volume	2.492	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.238	03:26
Breath Volume	2.648	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REILLY,S.,M

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: LEMEN,L.

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A3

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: DOWNER

DATE/TIME OF TEST: 05/16/2021 05:14:11

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.163	05:09
Breath Volume	3.554	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.147	05:12
Breath Volume	1.507	05:12
Air Blank	0.000	05:13
Control Test	0.079	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEPUTY,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEPUTY,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 05/16/2021 17:28:04

Diagnostics Check	OK	17:21
Air Blank	0.000	17:21
Control Test	0.080	17:21
Air Blank	0.000	17:22
Subject Sample # 1	0.000	17:23
Breath Volume	2.292	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.000	17:26
Breath Volume	2.152	17:26
Air Blank	0.000	17:27
Control Test	0.080	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:28

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DZUGAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DZUGAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MCAVOY

DATE/TIME OF TEST: 05/18/2021 01:22:54

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:18
Breath Volume	3.062	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:21
Breath Volume	3.253	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STEALEY,LANE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: KELLUM

DATE/TIME OF TEST: 05/18/2021 01:57:15

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.041	01:52
Breath Volume	1.527	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.038	01:55
Breath Volume	1.355	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 05/22/2021 23:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DIPIAZZA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: ROCHEK

DATE/TIME OF TEST: 05/18/2021 10:36:59

Diagnostics Check	OK	10:30
Air Blank	0.000	10:30
Control Test	0.079	10:30
Air Blank	0.000	10:31
Subject Sample # 1	0.000	10:32
Breath Volume	1.105	10:32
Air Blank	0.000	10:32
Air Blank	0.000	10:34
Subject Sample # 2	0.000	10:35
Breath Volume	1.261	10:35
Air Blank	0.000	10:36
Control Test	0.080	10:36
Air Blank	0.000	10:36
Diagnostics Check	OK	10:36

Data Download Date/Time: 05/22/2021 23:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TORRES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/18/2021 15:59:01

Diagnostics Check	OK	15:52
Air Blank	0.000	15:52
Control Test	0.079	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.223	15:54
Breath Volume	2.921	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.218	15:57
Breath Volume	3.117	15:57
Air Blank	0.000	15:58
Control Test	0.078	15:58
Air Blank	0.000	15:58
Diagnostics Check	OK	15:59

Data Download Date/Time: 05/22/2021 23:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORDSTROM,JAMES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PARO,JOSIAH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/19/2021 06:25:26

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:17	
Control Test	0.080	06:17	
Air Blank	0.000	06:18	
Subject Sample	# 1 0.190	06:20	
Breath Volume	1.117	06:20	
Air Blank	0.000	06:21	
Air Blank	0.000	06:23	
Subject Sample	# 2 0.186	06:23	
Breath Volume	1.453	06:23	
Air Blank	0.000	06:24	
Control Test	0.068	06:24	Control Outside Tolerance
Air Blank	0.000	06:25	

Data Download Date/Time: 05/22/2021 23:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WYMER,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CHAREUNSAB

DATE/TIME OF TEST: 05/20/2021 00:26:55

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.162	00:22
Breath Volume	2.628	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.163	00:25
Breath Volume	2.812	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 05/22/2021 23:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,T
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:33

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/21/2021 10:59:32

Diagnostics Check	OK	10:53
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Subject Sample # 1	0.251	10:54
Breath Volume	2.164	10:54
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.244	10:57
Breath Volume	1.894	10:57
Air Blank	0.000	10:58
Control Test	0.077	10:58
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 05/22/2021 23:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,T
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCSHANE

DATE/TIME OF TEST: 05/21/2021 21:31:13

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.080	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.000	21:26
Breath Volume	1.304	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.000	21:29
Breath Volume	1.757	21:29
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 05/22/2021 23:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: URSOMANNO,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 05/21/2021 23:31:00

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.241	23:26
Breath Volume	1.625	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.245	23:29
Breath Volume	2.238	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:31

Data Download Date/Time: 05/22/2021 23:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JOHNSON,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/23/2021 22:18:09

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.190	22:13
Breath Volume	1.511	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.197	22:16
Breath Volume	1.292	22:16
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 06/22/2021 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEE,ELIJAH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/28/2021 22:21:01

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.000	22:16
Breath Volume	2.398	22:16
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.000	22:19
Breath Volume	2.031	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:21

Data Download Date/Time: 06/22/2021 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KIMBERLIN,DEVIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/29/2021 21:24:41

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.181	21:19
Breath Volume	3.191	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.178	21:23
Breath Volume	3.039	21:23
Air Blank	0.000	21:23
Control Test	0.080	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 06/22/2021 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VETRI,T
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: CRIDER

DATE/TIME OF TEST: 05/29/2021 22:24:06

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.103	22:19
Breath Volume	1.246	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.103	22:22
Breath Volume	1.421	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 06/22/2021 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 05/30/2021 04:32:01

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.171	04:27
Breath Volume	2.242	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.183	04:30
Breath Volume	2.734	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:32

Data Download Date/Time: 06/22/2021 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OWENS,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 05/31/2021 00:00:06

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.190	23:55
Breath Volume	3.011	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.190	23:58
Breath Volume	3.210	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 06/22/2021 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 05/31/2021 01:58:18

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:55	
Control Test	0.079	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Control Test	0.080	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 06/22/2021 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 05/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CHIRINO MARTINEZ

DATE/TIME OF TEST: 05/30/2021 03:09:28

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Subject Sample # 1	0.231	03:03	
Breath Volume	1.917	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample # 2		03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 06/08/2021 05:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 05/06/2021 22:23:09

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.078	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.000	22:18
Breath Volume	1.425	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:21
Subject Sample # 2	0.000	22:21
Breath Volume	1.535	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 05/12/2021 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 05/13/2021 03:41:44

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.139	03:36
Breath Volume	1.750	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.144	03:40
Breath Volume	1.339	03:40
Air Blank	0.000	03:40
Control Test	0.077	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 06/02/2021 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 05/15/2021 01:31:20

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.077	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.237	01:26
Breath Volume	1.593	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.218	01:29
Breath Volume	1.488	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 06/02/2021 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPIC,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: ORJUELA TAYLOR

DATE/TIME OF TEST: 05/15/2021 06:24:14

Diagnostics Check	OK	06:17
Air Blank	0.000	06:17
Control Test	0.078	06:17
Air Blank	0.000	06:18
Subject Sample # 1	0.100	06:19
Breath Volume	1.675	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.090	06:22
Breath Volume	1.773	06:22
Air Blank	0.000	06:23
Control Test	0.078	06:23
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 06/02/2021 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 05/23/2021 04:57:39

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.154	04:52
Breath Volume	3.140	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.144	04:56
Breath Volume	2.941	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 06/02/2021 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 05/30/2021 00:28:08

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.078	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.068	00:23
Breath Volume	2.304	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.067	00:26
Breath Volume	2.085	00:26
Air Blank	0.000	00:27
Control Test	0.078	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 06/02/2021 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: GREENSWORD

DATE/TIME OF TEST: 05/31/2021 00:17:27

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.216	00:11
Breath Volume	1.609	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.224	00:15
Breath Volume	1.843	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 06/02/2021 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/28/2021 02:13:01

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.081	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.126	02:08
Breath Volume	4.167	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.121	02:11
Breath Volume	3.828	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/09/2021 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1349789
BREATH TEST OPERATOR: HUDSON,JOSHUA,K	DRY GAS EXPIRY DATE: 07/15/2023
BREATH TEST OPERATOR AGENCY: Pensacola PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HUDSON,JOSHUA	ARREST TIME: 02:00
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Female
LAST AGENCY INSPECTION: 05/26/2021	OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PROVOST	DATE/TIME OF TEST: 05/29/2021 02:39:02
------------------------------	--

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.000	02:34
Breath Volume	2.203	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:37
Breath Volume	2.371	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 06/09/2021 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY, MIA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 05/30/2021 05:22:45

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.231	05:17
Breath Volume	4.003	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.233	05:21
Breath Volume	4.058	05:21
Air Blank	0.000	05:21
Control Test	0.079	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 06/09/2021 00:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILLIAMS,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/30/2021 22:59:21

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.081	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.304	22:54
Breath Volume	1.882	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.300	22:57
Breath Volume	1.957	22:57
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 06/09/2021 00:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 05/20/2021 18:58:59

Diagnostics Check	OK	18:51
Air Blank	0.000	18:52
Control Test	0.079	18:52
Air Blank	0.000	18:53
Subject Sample # 1	0.182	18:54
Breath Volume	4.238	18:54
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.182	18:57
Breath Volume	4.664	18:57
Air Blank	0.000	18:58
Control Test	0.077	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 07/01/2021 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: PINERO

DATE/TIME OF TEST: 05/04/2021 22:56:35

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.079	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1 0.289	22:46	
Breath Volume	1.816	22:46	
Air Blank	0.000	22:47	
Air Blank	0.000	22:48	
Subject Sample	# 2 0.024	22:51	Volume Not Met
Breath Volume	0.539	22:51	
Air Blank	0.000	22:52	
Air Blank	0.000	22:54	
Subject Sample	# 3 0.289	22:54	
Breath Volume	3.257	22:54	
Air Blank	0.000	22:55	
Control Test	0.077	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 06/02/2021 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: AMEZCUA MORENO

DATE/TIME OF TEST: 05/08/2021 00:26:07

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.183	00:20
Breath Volume	1.996	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.179	00:24
Breath Volume	1.230	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 06/02/2021 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 05/08/2021 02:13:34

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.078	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.000	02:09
Breath Volume	3.882	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.000	02:12
Breath Volume	3.187	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/02/2021 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PEREZ ZAPATA

DATE/TIME OF TEST: 05/09/2021 02:16:56

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.219	02:11
Breath Volume	3.117	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.238	02:15
Breath Volume	3.238	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 06/02/2021 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ROBLEDO

DATE/TIME OF TEST: 05/13/2021 20:29:09

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.078	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.000	20:24
Breath Volume	2.078	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.000	20:27
Breath Volume	1.945	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 06/02/2021 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CRUZ MILAN

DATE/TIME OF TEST: 05/14/2021 01:10:02

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.178	01:05
Breath Volume	1.671	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.183	01:08
Breath Volume	1.261	01:08
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:10

Data Download Date/Time: 06/02/2021 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 05/14/2021 23:30:31

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.081	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1 0.153	23:23	Volume Not Met
Breath Volume	0.070	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 2 0.161	23:28	Volume Not Met
Breath Volume	0.136	23:28	
Air Blank	0.000	23:29	
Control Test	0.079	23:29	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 06/02/2021 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: ELMEGREHI

DATE/TIME OF TEST: 05/15/2021 06:12:54

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.080	06:05
Air Blank	0.000	06:06
Subject Sample # 1	0.166	06:07
Breath Volume	1.484	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:09
Subject Sample # 2	0.166	06:11
Breath Volume	1.605	06:11
Air Blank	0.000	06:11
Control Test	0.079	06:12
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 06/02/2021 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BUIE

DATE/TIME OF TEST: 05/19/2021 00:23:51

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.009	00:19
Breath Volume	2.652	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.009	00:22
Breath Volume	2.566	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 06/02/2021 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: TORRES JR

DATE/TIME OF TEST: 05/22/2021 19:06:38

Diagnostics Check	OK	18:59
Air Blank	0.000	19:00
Control Test	0.080	19:00
Air Blank	0.000	19:01
Subject Sample # 1	0.323	19:01
Breath Volume	1.722	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.305	19:04
Breath Volume	2.050	19:04
Air Blank	0.000	19:05
Control Test	0.078	19:06
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 06/02/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/23/2021 01:12:07

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.119	01:07
Breath Volume	1.726	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.112	01:10
Breath Volume	1.761	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 06/02/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 05/24/2021 23:54:53

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.082	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.267	23:50
Breath Volume	2.398	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.282	23:53
Breath Volume	2.113	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 06/02/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: HOUCK

DATE/TIME OF TEST: 05/26/2021 02:43:08

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1 0.204	02:35	Volume Not Met
Breath Volume	0.800	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample	# 2 0.184	02:41	Volume Not Met
Breath Volume	0.710	02:41	
Air Blank	0.000	02:42	
Control Test	0.078	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 06/02/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: HOUCK

DATE/TIME OF TEST: 05/26/2021 02:54:13

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.219	02:49
Breath Volume	1.671	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:51
Subject Sample # 2	0.214	02:52
Breath Volume	1.691	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 06/02/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 05/29/2021 01:40:33

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.226	01:35
Breath Volume	2.546	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.232	01:38
Breath Volume	2.105	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 06/02/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 05/29/2021 23:48:36

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:47	
Control Test	0.079	23:47	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 06/02/2021 10:08

Subject Test Electronic Data

Subject Test Electronic Data

DATE/TIME OF TEST: 05/31/2021 21:29:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS,MALACHI

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 05/01/2021 18:55:55

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.078	18:49
Air Blank	0.000	18:49
Subject Sample # 1	0.114	18:51
Breath Volume	2.851	18:51
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.110	18:54
Breath Volume	2.082	18:54
Air Blank	0.000	18:54
Control Test	0.077	18:55
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 05/21/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILLIAMS,CLINT

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 20:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 05/01/2021 21:04:47

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:53	
Control Test	0.078	20:54	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:57	No Sample Provided
Breath Volume	0.000	20:57	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample	# 2	21:03	No Sample Provided
Breath Volume	0.000	21:03	
Air Blank	0.000	21:03	
Control Test	0.079	21:04	
Air Blank	0.000	21:04	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 05/21/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILKINS,MALACHI

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 23:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: STAMM

DATE/TIME OF TEST: 05/01/2021 23:50:24

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.079	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:43	No Sample Provided
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 2	23:49	No Sample Provided
Breath Volume	0.000	23:49	
Air Blank	0.000	23:49	
Control Test	0.079	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 05/21/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACH,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,ZACH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: CREGO

DATE/TIME OF TEST: 05/02/2021 17:11:35

Diagnostics Check	OK	17:04
Air Blank	0.000	17:04
Control Test	0.079	17:04
Air Blank	0.000	17:05
Subject Sample # 1	0.107	17:05
Breath Volume	1.714	17:05
Air Blank	0.000	17:06
Air Blank	0.000	17:08
Subject Sample # 2	0.110	17:09
Breath Volume	1.160	17:09
Air Blank	0.000	17:10
Control Test	0.079	17:10
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 05/21/2021 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOPPE,ZACH,D

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: HOPPE,ZACH

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 21:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: DIUSMEL

DATE/TIME OF TEST: 05/02/2021 21:25:54

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:22	
Control Test	0.078	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.075	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 05/21/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACH,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BERG,TIM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: RODRIGUEZ CAMARILLO

DATE/TIME OF TEST: 05/08/2021 17:56:03

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:53	
Control Test	0.080	17:53	
Air Blank	0.000	17:54	
Subject Sample	# 1	17:54	Subject Test Refused
Breath Volume	0.000	17:54	
Air Blank	0.000	17:55	
Control Test	0.080	17:55	
Air Blank	0.000	17:55	
Diagnostics Check	OK	17:56	

Data Download Date/Time: 05/21/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACH,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,ZACH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 05/08/2021 23:11:59

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 05/21/2021 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: PETRU,MIROS

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 18:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: OBERGH

DATE/TIME OF TEST: 05/09/2021 19:25:54

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.078	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.000	19:21
Breath Volume	2.046	19:21
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.000	19:24
Breath Volume	2.765	19:24
Air Blank	0.000	19:24
Control Test	0.078	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 05/21/2021 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACH,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 05/09/2021 20:06:26

Diagnostics Check	OK	19:59
Air Blank	0.000	20:00
Control Test	0.079	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.086	20:01
Breath Volume	1.703	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.071	20:04
Breath Volume	1.125	20:04
Air Blank	0.000	20:05
Control Test	0.078	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 05/21/2021 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILKINS,MALACHI,C

BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)

ARRESTING OFFICER: WILLIAMS,CLINT

ARRESTING OFFICER AGENCY: FFWCC (Lakeland)

LAST AGENCY INSPECTION: 04/19/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 31420080A2

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: BUI

ARREST TIME: 20:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/09/2021 20:48:50

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.079	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:41	No Sample Provided
Breath Volume	0.000	20:41	
Air Blank	0.000	20:42	
Air Blank	0.000	20:44	
Subject Sample	# 2	20:47	No Sample Provided
Breath Volume	0.000	20:47	
Air Blank	0.000	20:47	
Control Test	0.079	20:48	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 05/21/2021 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURCIER, LEON, M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO, JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: COMAS HERRADA

DATE/TIME OF TEST: 05/02/2021 05:11:24

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.080	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:10	
Control Test	0.081	05:10	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY, CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: BENAGLIO

DATE/TIME OF TEST: 05/04/2021 21:25:25

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.080	21:19	
Air Blank	0.000	21:20	
Subject Sample # 1	0.154	21:21	
Breath Volume	1.980	21:21	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample # 2		21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.080	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURCIER, LEON, M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO, JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: RATER

DATE/TIME OF TEST: 05/06/2021 23:12:00

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.232	23:07
Breath Volume	4.535	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:09
Subject Sample # 2	0.228	23:10
Breath Volume	3.289	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:12

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DZIURZYNSKI

DATE/TIME OF TEST: 05/07/2021 22:44:05

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.279	22:39
Breath Volume	3.015	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:41
Subject Sample # 2	0.286	22:42
Breath Volume	3.347	22:42
Air Blank	0.000	22:43
Control Test	0.078	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FOSTER,MEGHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 05/08/2021 01:40:42

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.037	01:36
Breath Volume	2.984	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.037	01:39
Breath Volume	2.347	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY, CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: DENNISON

DATE/TIME OF TEST: 05/08/2021 20:41:08

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.114	20:36
Breath Volume	1.250	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.113	20:39
Breath Volume	1.347	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PURYEAR

DATE/TIME OF TEST: 05/08/2021 23:59:49

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.129	23:55
Breath Volume	3.632	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.127	23:58
Breath Volume	3.156	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,CORBIN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY,CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/09/2021 02:46:06

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.079	02:45	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY, CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 05/09/2021 22:50:30

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.218	22:45
Breath Volume	2.074	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.236	22:48
Breath Volume	2.046	22:48
Air Blank	0.000	22:49
Control Test	0.078	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHANDLER,KYLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: WIEDEMANN

DATE/TIME OF TEST: 05/10/2021 19:04:31

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:02	
Control Test	0.080	19:02	
Air Blank	0.000	19:02	
Subject Sample	# 1	19:03	Subject Test Refused
Breath Volume	0.000	19:03	
Air Blank	0.000	19:03	
Control Test	0.080	19:03	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:04	

Data Download Date/Time: 05/12/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HRADECKY, CORBIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 05/14/2021 00:03:57

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.164	23:59
Breath Volume	2.597	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.161	00:02
Breath Volume	2.667	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/15/2021 01:25:27

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:24	Subject Test Refused
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Control Test	0.079	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: NORTH

DATE/TIME OF TEST: 05/16/2021 00:28:35

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.192	00:23
Breath Volume	2.968	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.195	00:26
Breath Volume	2.632	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTERBAND,FRANK,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WESTERBAND,FRANK
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:08

SUBJECT'S LAST NAME: RICKETTS

DATE/TIME OF TEST: 05/20/2021 09:36:14

Diagnostics Check	OK	09:29
Air Blank	0.000	09:30
Control Test	0.080	09:30
Air Blank	0.000	09:31
Subject Sample # 1	0.000	09:31
Breath Volume	3.183	09:31
Air Blank	0.000	09:32
Air Blank	0.000	09:34
Subject Sample # 2	0.000	09:34
Breath Volume	1.679	09:34
Air Blank	0.000	09:35
Control Test	0.080	09:35
Air Blank	0.000	09:36
Diagnostics Check	OK	09:36

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTERBAND,FRANK,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SZYMANSKI,AMIE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: PRIBYL

DATE/TIME OF TEST: 05/20/2021 19:28:56

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:17	
Control Test	0.080	19:18	
Air Blank	0.000	19:18	
Subject Sample	# 1	19:21	No Sample Provided
Breath Volume	0.000	19:21	
Air Blank	0.000	19:22	
Air Blank	0.000	19:24	
Subject Sample	# 2	19:27	No Sample Provided
Breath Volume	0.000	19:27	
Air Blank	0.000	19:28	
Control Test	0.080	19:28	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: BRACKIN

DATE/TIME OF TEST: 05/23/2021 01:22:25

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.180	01:17
Breath Volume	2.480	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.180	01:20
Breath Volume	2.363	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTERBAND,FRANK,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WESTERBAND,FRANK
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: BLALOCK

DATE/TIME OF TEST: 05/30/2021 17:14:38

Diagnostics Check	OK	17:07
Air Blank	0.000	17:08
Control Test	0.080	17:08
Air Blank	0.000	17:09
Subject Sample # 1	0.189	17:09
Breath Volume	2.027	17:09
Air Blank	0.000	17:10
Air Blank	0.000	17:12
Subject Sample # 2	0.187	17:12
Breath Volume	2.269	17:12
Air Blank	0.000	17:13
Control Test	0.079	17:14
Air Blank	0.000	17:14
Diagnostics Check	OK	17:14

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 05/30/2021 19:11:42

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.079	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.192	19:06
Breath Volume	2.015	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:08
Subject Sample # 2	0.193	19:09
Breath Volume	2.566	19:09
Air Blank	0.000	19:10
Control Test	0.079	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RUFFIN, CHRIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/31/2021 20:22:22

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.207	20:17
Breath Volume	3.019	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:19
Subject Sample # 2	0.203	20:20
Breath Volume	2.570	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 06/07/2021 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONELY,JARRICK,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DONELY,JARRICK
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: CONKLIN

DATE/TIME OF TEST: 05/27/2021 00:26:08

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.000	00:21
Breath Volume	1.492	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:24
Breath Volume	2.164	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN, JONATHAN, M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WILDEMAN-STUART, JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 05/28/2021 03:35:53

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.100	03:31
Breath Volume	1.535	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.104	03:34
Breath Volume	1.371	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEPHART,HUNTER,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KEPHART,HUNTER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: URAM

DATE/TIME OF TEST: 05/28/2021 22:47:49

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:45	
Control Test	0.081	22:45	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Control Test	0.081	22:47	
Air Blank	0.000	22:47	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/28/2021 23:59:05

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.000	23:54
Breath Volume	1.894	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:57
Breath Volume	1.199	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,KIP,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PETERSON,KIP
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/29/2021 02:02:49

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.189	01:57
Breath Volume	2.035	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.185	02:01
Breath Volume	1.234	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAS,ASA,E
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: CHRISTMAS,FOLIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: BUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 05/29/2021 19:11:05

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:09	
Control Test	0.081	19:09	
Air Blank		19:10	Purge Fail
Air Blank	0.000	19:11	

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAS,ASA,E
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: CHRISTMAS,FOLIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: BUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 05/29/2021 19:16:08

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:14	
Control Test	0.078	19:14	
Air Blank		19:15	Purge Fail
Air Blank	0.000	19:16	

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAS,ASA,E
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: CHRISTMAS,FOLIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: BUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 05/29/2021 19:30:48

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.081	19:24
Air Blank	0.000	19:24
Subject Sample # 1	0.125	19:25
Breath Volume	3.460	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.124	19:29
Breath Volume	3.773	19:29
Air Blank	0.000	19:29
Control Test	0.080	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIDORI,CASEY,A
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: POLIDORI,CASEY
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HALVERSON

DATE/TIME OF TEST: 05/29/2021 23:50:51

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.081	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	1.046	23:49	
Air Blank	0.000	23:49	
Control Test	0.081	23:50	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MALESKE,RACHEL
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: HUERTA RUGGIERO

DATE/TIME OF TEST: 05/31/2021 05:00:23

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.129	04:55
Breath Volume	3.742	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.135	04:58
Breath Volume	3.503	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECCUE,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BECCUE,JOSHUA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 05/31/2021 20:14:22

Diagnostics Check	OK	20:06
Air Blank	0.000	20:07
Control Test	0.081	20:07
Air Blank	0.000	20:08
Subject Sample # 1	0.199	20:09
Breath Volume	1.144	20:09
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.190	20:12
Breath Volume	1.453	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWDER,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 05/31/2021 21:23:50

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	
Control Test	0.081	21:23	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 06/23/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ORCHARD,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/01/2021 16:45:59

Diagnostics Check	OK	16:39
Air Blank	0.000	16:40
Control Test	0.080	16:40
Air Blank	0.000	16:41
Subject Sample # 1	0.000	16:41
Breath Volume	2.816	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.000	16:44
Breath Volume	1.886	16:44
Air Blank	0.000	16:45
Control Test	0.080	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIPP,JORDAN,E
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: TRIPP,JORDAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KAVANAUGH

DATE/TIME OF TEST: 05/02/2021 22:10:32

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:00	
Control Test	0.080	22:01	
Air Blank	0.000	22:01	
Subject Sample	# 1 0.238	22:04	
Breath Volume	1.785	22:04	
Air Blank	0.000	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 2 0.237	22:09	
Breath Volume	1.828	22:09	
Air Blank		22:10	Purge Fail
Air Blank	0.000	22:10	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BUSICK

DATE/TIME OF TEST: 05/05/2021 03:27:03

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.000	03:22
Breath Volume	1.460	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.000	03:25
Breath Volume	2.429	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETWEILER,BRETT,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,DILLON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: ROLFS

DATE/TIME OF TEST: 05/05/2021 09:15:19

Diagnostics Check	OK	09:08
Air Blank	0.000	09:09
Control Test	0.080	09:09
Air Blank	0.000	09:10
Subject Sample # 1	0.179	09:10
Breath Volume	1.886	09:10
Air Blank	0.000	09:11
Air Blank	0.000	09:13
Subject Sample # 2	0.168	09:13
Breath Volume	2.078	09:13
Air Blank	0.000	09:14
Control Test	0.079	09:14
Air Blank	0.000	09:15
Diagnostics Check	OK	09:15

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HATTAWAY

DATE/TIME OF TEST: 05/07/2021 02:56:35

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.092	02:51
Breath Volume	2.621	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.087	02:54
Breath Volume	2.593	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 05/08/2021 03:22:39

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.166	03:18
Breath Volume	2.890	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.161	03:20
Breath Volume	1.640	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MCMURRAY,HUNTER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: GRASSMYER

DATE/TIME OF TEST: 05/08/2021 13:59:43

Diagnostics Check	OK	13:48	
Air Blank	0.000	13:48	
Control Test	0.081	13:49	
Air Blank	0.000	13:49	
Subject Sample	# 1	13:52	No Sample Provided
Breath Volume	0.000	13:52	
Air Blank	0.000	13:53	
Air Blank	0.000	13:55	
Subject Sample	# 2	13:58	Volume Not Met
Breath Volume	0.421	13:58	
Air Blank	0.000	13:58	
Control Test	0.082	13:59	
Air Blank	0.000	13:59	
Diagnostics Check	OK	13:59	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MCMURRAY,HUNTER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: GRASSMYER

DATE/TIME OF TEST: 05/08/2021 14:39:13

Diagnostics Check	OK	14:24	
Air Blank	0.000	14:24	
Control Test	0.081	14:24	
Air Blank	0.000	14:25	
Subject Sample	# 1 0.000	14:26	
Breath Volume	1.121	14:26	
Air Blank	0.000	14:26	
Air Blank	0.000	14:28	
Subject Sample	# 2 0.000	14:32	Volume Not Met
Breath Volume	0.269	14:32	
Air Blank	0.000	14:32	
Air Blank	0.000	14:34	
Subject Sample	# 3 0.000	14:37	Volume Not Met
Breath Volume	0.375	14:37	
Air Blank	0.000	14:38	
Control Test	0.079	14:38	
Air Blank	0.000	14:39	
Diagnostics Check	OK	14:39	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONELY,JARRICK,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DONELY,JARRICK
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: ALCIME

DATE/TIME OF TEST: 05/09/2021 04:45:10

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.201	04:40
Breath Volume	1.828	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.196	04:43
Breath Volume	1.636	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/09/2021 04:59:15

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.114	04:51
Breath Volume	1.320	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.113	04:57
Breath Volume	1.339	04:57
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 05/09/2021 21:10:40

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.080	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.196	21:05
Breath Volume	1.734	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.198	21:08
Breath Volume	1.609	21:08
Air Blank	0.000	21:09
Control Test	0.080	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: FRITTS

DATE/TIME OF TEST: 05/15/2021 04:52:31

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.082	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.000	04:48
Breath Volume	1.871	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.000	04:51
Breath Volume	2.492	04:51
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINTON, BLAKE, E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MINTON, BLAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/15/2021 23:09:44

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.080	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1 0.194	23:05	Volume Not Met
Breath Volume	0.214	23:05	
Air Blank	0.000	23:06	
Air Blank	0.000	23:08	
Subject Sample	# 2	23:08	RFI Detect
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.078	23:09	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINTON, BLAKE, E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MINTON, BLAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/15/2021 23:25:12

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.080	23:12	
Air Blank	0.000	23:13	
Subject Sample	# 1 0.182	23:16	Volume Not Met
Breath Volume	0.312	23:16	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 2 0.176	23:19	No .020 agreement
Breath Volume	1.316	23:19	
Air Blank	0.000	23:20	
Air Blank	0.000	23:22	
Subject Sample	# 3 0.198	23:23	No .020 agreement
Breath Volume	1.152	23:23	
Air Blank	0.000	23:24	
Control Test	0.079	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHER,ASHLYN,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHER,ASHLYN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: SHERRILL

DATE/TIME OF TEST: 05/16/2021 00:53:37

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:52	
Control Test	0.081	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: VAZQUEZ-ROGERO

DATE/TIME OF TEST: 05/16/2021 05:28:08

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.150	05:22
Breath Volume	1.101	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.148	05:26
Breath Volume	2.457	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMPTER,SCOTT,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: SUMPTER,SCOTT
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: MCGREAL

DATE/TIME OF TEST: 05/16/2021 17:07:15

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.081	17:01
Air Blank	0.000	17:01
Subject Sample # 1	0.008	17:02
Breath Volume	2.980	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.000	17:05
Breath Volume	2.824	17:05
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIMMEL,AUTUM,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HIMMEL,AUTUM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: SAMUELS

DATE/TIME OF TEST: 05/16/2021 19:06:24

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.081	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.059	19:01
Breath Volume	1.597	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:03
Subject Sample # 2	0.054	19:04
Breath Volume	1.542	19:04
Air Blank	0.000	19:05
Control Test	0.079	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HOLLINGSWORTH,DALE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/16/2021 22:40:26

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Sequence Aborted
Breath Volume	0.000	22:39	
Air Blank	0.000	22:40	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD, DAVID, L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HOLLINGSWORTH, DALE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/16/2021 23:12:00

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:04	No Sample Provided
Breath Volume	0.000	23:04	
Air Blank	0.000	23:04	
Air Blank	0.000	23:06	
Subject Sample	# 2	0.264 23:07	
Breath Volume	1.625	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 3	0.270 23:10	
Breath Volume	3.683	23:10	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,YOANDY,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MIRANDA,YOANDY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 05/20/2021 08:31:08

Diagnostics Check	OK	08:19	
Air Blank	0.000	08:19	
Control Test	0.081	08:19	
Air Blank	0.000	08:20	
Subject Sample	# 1	08:23	No Sample Provided
Breath Volume	0.000	08:23	
Air Blank	0.000	08:23	
Air Blank	0.000	08:25	
Subject Sample	# 2	0.170	08:26
Breath Volume	2.027	08:26	
Air Blank	0.000	08:27	
Air Blank	0.000	08:28	
Subject Sample	# 3	0.171	08:29
Breath Volume	1.894	08:29	
Air Blank	0.000	08:30	
Control Test	0.081	08:30	
Air Blank	0.000	08:31	
Diagnostics Check	OK	08:31	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HERRMANN

DATE/TIME OF TEST: 05/20/2021 23:07:06

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.000	23:02
Breath Volume	2.511	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.000	23:05
Breath Volume	2.703	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 05/22/2021 02:25:11

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.251	02:17
Breath Volume	1.214	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.247	02:23
Breath Volume	1.648	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHER,ASHLYN,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHER,ASHLYN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CHAMPION

DATE/TIME OF TEST: 05/22/2021 02:49:32

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.078	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:48	
Control Test	0.078	02:48	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 05/22/2021 14:24:49

Diagnostics Check	OK	14:10	
Air Blank	0.000	14:10	
Control Test	0.080	14:10	
Air Blank	0.000	14:11	
Subject Sample # 1	0.000	14:12	
Breath Volume	1.578	14:12	
Air Blank	0.000	14:13	
Air Blank	0.000	14:15	
Subject Sample # 2	0.000	14:18	Volume Not Met
Breath Volume	0.218	14:18	
Air Blank	0.000	14:18	
Air Blank	0.000	14:20	
Subject Sample # 3	0.000	14:23	
Breath Volume	1.246	14:23	
Air Blank	0.000	14:23	
Control Test	0.077	14:24	
Air Blank	0.000	14:24	
Diagnostics Check	OK	14:24	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ALCIME

DATE/TIME OF TEST: 05/23/2021 04:34:27

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.000	04:29
Breath Volume	1.597	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.000	04:33
Breath Volume	2.574	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BARNES,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: FEDERICO KENNEDY

DATE/TIME OF TEST: 05/23/2021 06:16:34

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:02	
Control Test	0.079	06:03	
Air Blank	0.000	06:03	
Subject Sample	# 1 0.159	06:06	Volume Not Met
Breath Volume	0.382	06:06	
Air Blank	0.000	06:07	
Air Blank	0.000	06:09	
Subject Sample	# 2 0.177	06:10	
Breath Volume	1.101	06:10	
Air Blank	0.000	06:11	
Air Blank	0.000	06:12	
Subject Sample	# 3 0.165	06:14	
Breath Volume	1.949	06:14	
Air Blank	0.000	06:15	
Control Test	0.079	06:15	
Air Blank	0.000	06:16	
Diagnostics Check	OK	06:16	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 05/23/2021 17:08:33

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.080	17:03
Air Blank	0.000	17:03
Subject Sample # 1	0.000	17:04
Breath Volume	2.218	17:04
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample # 2	0.000	17:07
Breath Volume	1.402	17:07
Air Blank	0.000	17:07
Control Test	0.080	17:07
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ELLIOTT,ZACHARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 05/24/2021 20:03:17

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.223	19:57
Breath Volume	1.113	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.226	20:01
Breath Volume	1.382	20:01
Air Blank	0.000	20:02
Control Test	0.077	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DILLON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,DILLON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/25/2021 17:47:42

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:36	
Control Test	0.080	17:36	
Air Blank	0.000	17:37	
Subject Sample	# 1 0.129	17:40	Volume Not Met
Breath Volume	0.277	17:40	
Air Blank	0.000	17:41	
Air Blank	0.000	17:42	
Subject Sample	# 2 0.123	17:45	Volume Not Met
Breath Volume	0.792	17:45	
Air Blank	0.000	17:46	
Control Test	0.078	17:47	
Air Blank	0.000	17:47	
Diagnostics Check	OK	17:47	

Data Download Date/Time: 05/26/2021 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: HERNANDEZ,RAFAEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WITHERSPOON

DATE/TIME OF TEST: 05/25/2021 21:58:59

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.080	21:56	
Air Blank	0.000	21:57	
Subject Sample	# 1	21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Control Test	0.080	21:58	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 05/26/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GURRIERI,JOHN,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GURRIERI,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MOSELER

DATE/TIME OF TEST: 05/26/2021 03:32:54

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.190	03:28
Breath Volume	2.835	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.192	03:31
Breath Volume	1.570	03:31
Air Blank	0.000	03:31
Control Test	0.077	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 05/26/2021 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SYVERSON,PAUL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1012059
DRY GAS EXPIRY DATE: 09/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/02/2021 21:21:10

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.078	21:18	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:20	
Control Test	0.078	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 05/15/2021 01:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: BESS

DATE/TIME OF TEST: 05/01/2021 21:27:42

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.078	21:15	
Air Blank	0.000	21:16	
Subject Sample	# 1 0.118	21:17	
Breath Volume	1.460	21:17	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2 0.095	21:23	Volume Not Met
Breath Volume	0.691	21:23	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 3 0.122	21:26	
Breath Volume	2.683	21:26	
Air Blank	0.000	21:26	
Control Test	0.077	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: KOSTREVSKI

DATE/TIME OF TEST: 05/02/2021 22:04:07

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:52	
Control Test	0.078	21:53	
Air Blank	0.000	21:53	
Subject Sample	# 1 0.204	21:56	Volume Not Met
Breath Volume	0.308	21:56	
Air Blank	0.000	21:57	
Air Blank	0.000	21:59	
Subject Sample	# 2 0.213	22:02	Volume Not Met
Breath Volume	0.210	22:02	
Air Blank	0.000	22:03	
Control Test	0.078	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DOWNING,TYLER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: GAY-WERTS

DATE/TIME OF TEST: 05/03/2021 08:12:05

Diagnostics Check	OK	08:05
Air Blank	0.000	08:05
Control Test	0.078	08:05
Air Blank	0.000	08:06
Subject Sample # 1	0.127	08:07
Breath Volume	2.160	08:07
Air Blank	0.000	08:07
Air Blank	0.000	08:09
Subject Sample # 2	0.126	08:10
Breath Volume	2.183	08:10
Air Blank	0.000	08:11
Control Test	0.077	08:11
Air Blank	0.000	08:11
Diagnostics Check	OK	08:12

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MILLS,SCOTT,T

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MILLS,SCOTT

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:03

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 05/04/2021 01:01:42

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.000	00:57
Breath Volume	1.964	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	01:00
Breath Volume	2.691	01:00
Air Blank	0.000	01:00
Control Test	0.078	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GREEN,J,A

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: GREEN,J

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: VANDERLINDEN

DATE/TIME OF TEST: 05/04/2021 03:07:59

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.069	03:03
Breath Volume	2.242	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.069	03:06
Breath Volume	3.070	03:06
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/06/2021 03:18:10

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.087	03:13
Breath Volume	1.527	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.088	03:16
Breath Volume	1.511	03:16
Air Blank	0.000	03:17
Control Test	0.077	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FIGUEROA,K
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/09/2021 04:48:29

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.078	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.183	04:44
Breath Volume	1.878	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.178	04:46
Breath Volume	2.250	04:46
Air Blank	0.000	04:47
Control Test	0.077	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GIANNOTTA,D
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: LAVOSKY

DATE/TIME OF TEST: 05/09/2021 05:15:28

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.214	05:10
Breath Volume	2.945	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.225	05:13
Breath Volume	2.039	05:13
Air Blank	0.000	05:14
Control Test	0.077	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WOFFORD,JOSEPH
ARRESTING OFFICER AGENCY: OTHER County
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: V340660741020

DATE/TIME OF TEST: 05/10/2021 14:03:38

Diagnostics Check	OK	13:57
Air Blank	0.000	13:57
Control Test	0.078	13:57
Air Blank	0.000	13:58
Subject Sample # 1	0.000	13:59
Breath Volume	3.332	13:59
Air Blank	0.000	13:59
Air Blank	0.000	14:01
Subject Sample # 2	0.000	14:02
Breath Volume	2.367	14:02
Air Blank	0.000	14:02
Control Test	0.079	14:03
Air Blank	0.000	14:03
Diagnostics Check	OK	14:03

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PARKER,DARBY,J

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: BURKE,LYLE

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 17:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: FITZBURGH

DATE/TIME OF TEST: 05/11/2021 18:42:23

Diagnostics Check	OK	18:35
Air Blank	0.000	18:36
Control Test	0.078	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.306	18:37
Breath Volume	3.480	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.291	18:40
Breath Volume	3.386	18:40
Air Blank	0.000	18:41
Control Test	0.076	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/13/2021 02:40:10

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.195	02:35
Breath Volume	1.539	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.202	02:38
Breath Volume	1.582	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,SCOTT,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: AKIN,BARRY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/13/2021 15:53:56

Diagnostics Check	OK	15:47
Air Blank	0.000	15:47
Control Test	0.078	15:48
Air Blank	0.000	15:48
Subject Sample # 1	0.373	15:49
Breath Volume	2.730	15:49
Air Blank	0.000	15:49
Air Blank	0.000	15:51
Subject Sample # 2	0.367	15:52
Breath Volume	2.800	15:52
Air Blank	0.000	15:53
Control Test	0.075	15:53
Air Blank	0.000	15:53
Diagnostics Check	OK	15:53

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAYWARD,KYLE,M

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: HAYWARD,KYLE

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14020080A2

DRY GAS EXPIRY DATE: 07/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/14/2021 02:10:48

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.027	02:09	
Air Blank	0.000	02:09	
Control Test	0.078	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: FLOYD,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KINION

DATE/TIME OF TEST: 05/15/2021 21:32:39

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.221	21:24
Breath Volume	3.089	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.194	21:27
Breath Volume	2.093	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 3	0.203	21:31
Breath Volume	3.500	21:31
Air Blank	0.000	21:31
Control Test	0.075	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ERAZO

DATE/TIME OF TEST: 05/15/2021 23:50:38

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.223	23:46
Breath Volume	2.371	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.227	23:49
Breath Volume	2.484	23:49
Air Blank	0.000	23:49
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HAND,B
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 05/18/2021 20:33:10

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.079	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:31	Subject Test Refused
Breath Volume	0.000	20:31	
Air Blank	0.000	20:32	
Control Test	0.079	20:32	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SIMS,WALKER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PINNELLA

DATE/TIME OF TEST: 05/22/2021 01:11:53

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.070	01:07
Breath Volume	2.707	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample # 2	0.069	01:10
Breath Volume	2.347	01:10
Air Blank	0.000	01:10
Control Test	0.077	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/26/2021 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: COMBS,BRADLEY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 05/22/2021 01:21:01

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.078	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.175	01:16
Breath Volume	3.843	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.168	01:19
Breath Volume	3.730	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:21

Data Download Date/Time: 05/26/2021 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,SEAN,C
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BAKER,JERRY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BIRO

DATE/TIME OF TEST: 05/22/2021 19:07:11

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.078	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.275	19:02
Breath Volume	2.019	19:02
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.277	19:05
Breath Volume	1.886	19:05
Air Blank	0.000	19:06
Control Test	0.078	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 05/26/2021 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: WIESNER

DATE/TIME OF TEST: 05/22/2021 20:22:10

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.079	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.156	20:17
Breath Volume	1.781	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.165	20:20
Breath Volume	2.207	20:20
Air Blank	0.000	20:21
Control Test	0.077	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 05/26/2021 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HAYWARD,KYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 05/23/2021 01:55:20

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.134	01:50
Breath Volume	1.632	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.133	01:53
Breath Volume	1.496	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 05/26/2021 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORALUZZO,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/28/2021 17:57:07

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.079	17:51
Air Blank	0.000	17:52
Subject Sample # 1	0.263	17:52
Breath Volume	2.664	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:54
Subject Sample # 2	0.244	17:55
Breath Volume	2.535	17:55
Air Blank	0.000	17:56
Control Test	0.078	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 06/10/2021 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: NELSON,ROBERT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 05/31/2021 11:54:33

Diagnostics Check	OK	11:48
Air Blank	0.000	11:48
Control Test	0.079	11:49
Air Blank	0.000	11:49
Subject Sample # 1	0.000	11:50
Breath Volume	2.074	11:50
Air Blank	0.000	11:50
Air Blank	0.000	11:52
Subject Sample # 2	0.000	11:53
Breath Volume	2.406	11:53
Air Blank	0.000	11:53
Control Test	0.079	11:53
Air Blank	0.000	11:54
Diagnostics Check	OK	11:54

Data Download Date/Time: 06/10/2021 19:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BICKHART,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: O'BLENIS

DATE/TIME OF TEST: 05/31/2021 22:45:11

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.000	22:40
Breath Volume	1.964	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.000	22:43
Breath Volume	1.820	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 06/10/2021 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANCLEVE,TRENTON,D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VANCLEVE,TRENTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 04/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/14/2021 23:12:30

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.380	23:07
Breath Volume	2.628	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.371	23:10
Breath Volume	1.890	23:10
Air Blank	0.000	23:11
Control Test	0.075	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 05/15/2021 01:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: D'HAITI,DWAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: FACEY

DATE/TIME OF TEST: 05/01/2021 00:15:09

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.081	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.124	00:07	Volume Not Met
Breath Volume	0.019	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:10	
Subject Sample	# 2 0.129	00:13	Volume Not Met
Breath Volume	0.109	00:13	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 05/03/2021 20:26:33

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:15	
Control Test	0.081	20:15	
Air Blank	0.000	20:16	
Subject Sample	# 1 0.103	20:19	Volume Not Met
Breath Volume	0.078	20:19	
Air Blank	0.000	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 2 0.093	20:24	Volume Not Met
Breath Volume	0.121	20:24	
Air Blank	0.000	20:25	
Control Test	0.079	20:25	
Air Blank	0.000	20:26	
Diagnostics Check	OK	20:26	

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MALAVE,WILLIAM
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ESCALANTE JR

DATE/TIME OF TEST: 05/04/2021 01:57:47

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.081	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 2	0.237	01:50
Breath Volume	1.328	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 3	0.195	01:56
Breath Volume	0.000	01:56	Volume Not Met
Air Blank	0.000	01:56	
Control Test	0.079	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,DILLON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MOROZENKO

DATE/TIME OF TEST: 05/05/2021 19:16:25

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:12	
Control Test	0.082	19:13	
Air Blank	0.000	19:13	
Subject Sample	# 1	19:15	RFI Detect
Breath Volume	0.000	19:15	
Air Blank		19:15	RFI Detect
Air Blank	0.000	19:16	

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RUSSELL,DILLON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MOROZENKO

DATE/TIME OF TEST: 05/05/2021 19:31:04

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.082	19:20	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:24	No Sample Provided
Breath Volume	0.000	19:24	
Air Blank	0.000	19:24	
Air Blank	0.000	19:26	
Subject Sample	# 2	19:29	No Sample Provided
Breath Volume	0.000	19:29	
Air Blank	0.000	19:30	
Control Test	0.083	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:31	

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 05/05/2021 22:10:12

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.082	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.000	22:05
Breath Volume	1.781	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.000	22:08
Breath Volume	1.796	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SWAN

DATE/TIME OF TEST: 05/06/2021 01:27:23

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.082	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.051	01:22
Breath Volume	3.062	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.050	01:25
Breath Volume	2.800	01:25
Air Blank	0.000	01:26
Control Test	0.082	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LIPTAK,LAUREN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/09/2021 02:30:33

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.172	02:25
Breath Volume	2.699	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.173	02:28
Breath Volume	2.394	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: CASTRO-RAMIREZ

DATE/TIME OF TEST: 05/09/2021 05:30:34

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.082	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.199	05:25
Breath Volume	2.761	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.195	05:28
Breath Volume	2.750	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 05/12/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SLAYTON,GARY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:56

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 05/16/2021 09:34:50

Diagnostics Check	OK	09:28
Air Blank	0.000	09:28
Control Test	0.083	09:29
Air Blank	0.000	09:29
Subject Sample # 1	0.000	09:30
Breath Volume	3.656	09:30
Air Blank	0.000	09:30
Air Blank	0.000	09:32
Subject Sample # 2	0.000	09:33
Breath Volume	2.773	09:33
Air Blank	0.000	09:33
Control Test	0.081	09:34
Air Blank	0.000	09:34
Diagnostics Check	OK	09:34

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 05/19/2021 02:54:35

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.163	02:49
Breath Volume	2.468	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.161	02:52
Breath Volume	1.562	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DENARDO,RYAN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RAMLOCHAN

DATE/TIME OF TEST: 05/20/2021 22:44:42

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.082	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.185	22:39
Breath Volume	1.382	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.186	22:42
Breath Volume	1.402	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: COURT ORD
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: DUTRA

DATE/TIME OF TEST: 05/21/2021 22:03:55

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.082	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.000	21:59
Breath Volume	1.515	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:02
Breath Volume	1.300	22:02
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PAIGE,EMMETT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 05/22/2021 01:47:29

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.083	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.222	01:42
Breath Volume	2.003	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.226	01:45
Breath Volume	2.742	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ULLOA,ADRIAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: MEDRANO ALVAREZ

DATE/TIME OF TEST: 05/23/2021 06:01:02

Diagnostics Check	OK	05:54
Air Blank	0.000	05:55
Control Test	0.082	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.149	05:56
Breath Volume	2.507	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:58
Subject Sample # 2	0.146	05:59
Breath Volume	2.617	05:59
Air Blank	0.000	06:00
Control Test	0.081	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:01

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DAVIS,MATTHEW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 05/26/2021 01:51:54

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.083	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.044	01:46
Breath Volume	1.855	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.045	01:50
Breath Volume	2.679	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SIUGZDINIENE

DATE/TIME OF TEST: 05/28/2021 00:39:44

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.082	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.226	00:34
Breath Volume	1.453	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.224	00:37
Breath Volume	1.523	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL,TKEYAH
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MANUEL

DATE/TIME OF TEST: 05/28/2021 03:04:29

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.179	02:59
Breath Volume	3.109	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.180	03:02
Breath Volume	2.757	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL,TKEYAH
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: JELONEK

DATE/TIME OF TEST: 05/28/2021 19:24:26

Diagnostics Check	OK	19:18
Air Blank	0.000	19:18
Control Test	0.082	19:18
Air Blank	0.000	19:19
Subject Sample # 1	0.000	19:20
Breath Volume	3.125	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.000	19:23
Breath Volume	2.738	19:23
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JURADO

DATE/TIME OF TEST: 05/31/2021 03:37:29

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.083	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.094	03:32
Breath Volume	1.238	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.097	03:35
Breath Volume	2.453	03:35
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 05/31/2021 04:38:44

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.082	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.168	04:34
Breath Volume	1.496	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.170	04:37
Breath Volume	3.242	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 06/02/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BLANEY

DATE/TIME OF TEST: 05/12/2021 16:46:55

Diagnostics Check	OK	16:40
Air Blank	0.000	16:40
Control Test	0.080	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.119	16:42
Breath Volume	2.152	16:42
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.126	16:45
Breath Volume	2.250	16:45
Air Blank	0.000	16:45
Control Test	0.081	16:46
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARTLETT,AMY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: FANSLOW

DATE/TIME OF TEST: 05/13/2021 19:42:25

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.081	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:37
Breath Volume	3.015	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.000	19:41
Breath Volume	2.175	19:41
Air Blank	0.000	19:41
Control Test	0.081	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DAVIS,MATTHEW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 05/14/2021 02:30:42

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.000	02:26
Breath Volume	1.121	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.000	02:29
Breath Volume	1.796	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALFORD,MARK
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 09:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:33

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/14/2021 11:01:54

Diagnostics Check	OK	10:55
Air Blank	0.000	10:56
Control Test	0.081	10:56
Air Blank	0.000	10:57
Subject Sample # 1	0.200	10:57
Breath Volume	1.796	10:57
Air Blank	0.000	10:58
Air Blank	0.000	10:59
Subject Sample # 2	0.197	11:00
Breath Volume	1.765	11:00
Air Blank	0.000	11:00
Control Test	0.080	11:01
Air Blank	0.000	11:01
Diagnostics Check	OK	11:01

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ-
MURILLO,REYNA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male

OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/22/2021 19:16:53

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.081	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.176	19:10
Breath Volume	1.199	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.187	19:15
Breath Volume	1.105	19:15
Air Blank	0.000	19:15
Control Test	0.079	19:16
Air Blank	0.000	19:16
Diagnostics Check	OK	19:16

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/22/2021 23:04:45

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.270	23:00
Breath Volume	2.179	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.273	23:03
Breath Volume	2.414	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 05/24/2021 00:40:20

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.186	00:35
Breath Volume	1.136	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.190	00:38
Breath Volume	1.828	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MC CORKLE

DATE/TIME OF TEST: 05/24/2021 04:06:56

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.079	04:02
Breath Volume	2.734	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.078	04:05
Breath Volume	3.445	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MARIN HERNANDEZ

DATE/TIME OF TEST: 05/26/2021 02:32:30

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.115	02:27
Breath Volume	1.867	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.116	02:30
Breath Volume	2.089	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DENARDO,RYAN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: COLWELL

DATE/TIME OF TEST: 05/27/2021 02:24:23

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.162	02:11	
Breath Volume	1.703	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2 0.160	02:16	Volume Not Met
Breath Volume	0.417	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 3 0.178	02:22	Volume Not Met
Breath Volume	0.066	02:22	
Air Blank	0.000	02:23	
Control Test	0.079	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOHM,JONATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/27/2021 04:38:46

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1 0.189	04:31	Volume Not Met
Breath Volume	0.414	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 2 0.180	04:36	Volume Not Met
Breath Volume	0.183	04:36	
Air Blank	0.000	04:37	
Control Test	0.078	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 05/27/2021 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER, WILLIAM, L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BONTRAGER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: DAVIS-BOLTON

DATE/TIME OF TEST: 05/02/2021 08:11:36

Diagnostics Check	OK	08:05	
Air Blank	0.000	08:06	
Control Test	0.081	08:06	
Air Blank	0.000	08:07	
Subject Sample	# 1	08:09	Subject Test Refused
Breath Volume	0.968	08:09	
Air Blank	0.000	08:10	
Control Test	0.080	08:10	
Air Blank	0.000	08:11	
Diagnostics Check	OK	08:11	

Data Download Date/Time: 05/06/2021 11:54