

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A4

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BELOM

DATE/TIME OF TEST: 06/11/2021 02:59:08

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.082	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.000	02:54
Breath Volume	1.863	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.000	02:57
Breath Volume	2.281	02:57
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 06/28/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLCROFT,AMANDA,L
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HOLLCROFT,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/12/2021 16:58:24

Diagnostics Check	OK	16:51
Air Blank	0.000	16:52
Control Test	0.081	16:52
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:53
Breath Volume	1.117	16:53
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.000	16:57
Breath Volume	1.394	16:57
Air Blank	0.000	16:57
Control Test	0.081	16:57
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 06/28/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: STEWART-REID

DATE/TIME OF TEST: 06/14/2021 00:37:06

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.191	00:31
Breath Volume	1.386	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.187	00:35
Breath Volume	1.953	00:35
Air Blank	0.000	00:36
Control Test	0.078	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 06/28/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: PULSIFER,MARK
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/12/2021 04:47:44

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:45	
Control Test	0.079	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:46	Subject Test Refused
Breath Volume	0.000	04:46	
Air Blank	0.000	04:46	
Control Test	0.079	04:47	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SCHUBART

DATE/TIME OF TEST: 06/12/2021 21:40:31

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.080	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.206	21:36
Breath Volume	2.546	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.197	21:38
Breath Volume	1.531	21:38
Air Blank	0.000	21:39
Control Test	0.079	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: RAYNOR

DATE/TIME OF TEST: 06/15/2021 10:56:39

Diagnostics Check	OK	10:50
Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:52
Subject Sample # 1	0.000	10:52
Breath Volume	1.816	10:52
Air Blank	0.000	10:52
Air Blank	0.000	10:54
Subject Sample # 2	0.000	10:55
Breath Volume	1.890	10:55
Air Blank	0.000	10:55
Control Test	0.079	10:56
Air Blank	0.000	10:56
Diagnostics Check	OK	10:56

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 06/20/2021 04:00:06

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.078	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KOLBICKA,LAWRENCE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 06/21/2021 09:44:36

Diagnostics Check	OK	09:38
Air Blank	0.000	09:39
Control Test	0.080	09:39
Air Blank	0.000	09:40
Subject Sample # 1	0.000	09:40
Breath Volume	2.152	09:40
Air Blank	0.000	09:40
Air Blank	0.000	09:42
Subject Sample # 2	0.000	09:43
Breath Volume	2.488	09:43
Air Blank	0.000	09:43
Control Test	0.080	09:44
Air Blank	0.000	09:44
Diagnostics Check	OK	09:44

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUFF,ZACARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: STOPA

DATE/TIME OF TEST: 06/23/2021 10:28:24

Diagnostics Check	OK	10:21
Air Blank	0.000	10:22
Control Test	0.080	10:22
Air Blank	0.000	10:23
Subject Sample # 1	0.136	10:23
Breath Volume	2.414	10:23
Air Blank	0.000	10:24
Air Blank	0.000	10:26
Subject Sample # 2	0.136	10:26
Breath Volume	2.089	10:26
Air Blank	0.000	10:27
Control Test	0.079	10:27
Air Blank	0.000	10:28
Diagnostics Check	OK	10:28

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: HANNUM

DATE/TIME OF TEST: 06/25/2021 03:04:17

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAWARD,MORGANA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: EDWARDS,AUSTIN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: DUONG

DATE/TIME OF TEST: 06/26/2021 14:24:56

Diagnostics Check	OK	14:18
Air Blank	0.000	14:19
Control Test	0.080	14:19
Air Blank	0.000	14:19
Subject Sample # 1	0.000	14:20
Breath Volume	2.015	14:20
Air Blank	0.000	14:20
Air Blank	0.000	14:22
Subject Sample # 2	0.000	14:23
Breath Volume	2.136	14:23
Air Blank	0.000	14:24
Control Test	0.080	14:24
Air Blank	0.000	14:24
Diagnostics Check	OK	14:24

Data Download Date/Time: 07/28/2021 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: RADOVANIC

DATE/TIME OF TEST: 06/01/2021 05:12:54

Diagnostics Check	OK	04:59	
Air Blank	0.000	05:00	
Control Test	0.078	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:04	No Sample Provided
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Air Blank	0.000	05:06	
Subject Sample	# 2	05:07	
Breath Volume	1.140	05:07	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 3	05:11	
Breath Volume	1.164	05:11	
Air Blank	0.000	05:11	
Control Test	0.078	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 06/01/2021 18:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARNES,B,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KARNES,B
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: VIERRA

DATE/TIME OF TEST: 06/01/2021 22:10:52

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.115	22:06
Breath Volume	2.933	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.115	22:09
Breath Volume	1.992	22:09
Air Blank	0.000	22:09
Control Test	0.078	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU, OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU, OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/01/2021 23:22:54

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.000	23:18
Breath Volume	1.441	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:21
Breath Volume	1.605	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RODRIGUEZ,XAVIER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/03/2021 02:02:58

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.064	01:58
Breath Volume	1.304	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.058	02:01
Breath Volume	1.265	02:01
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: ALBAUGH

DATE/TIME OF TEST: 06/03/2021 16:55:22

Diagnostics Check	OK	16:45
Air Blank	0.000	16:46
Control Test	0.077	16:46
Air Blank	0.000	16:47
Subject Sample # 1	0.257	16:48
Breath Volume	1.128	16:48
Air Blank	0.000	16:48
Air Blank	0.000	16:50
Subject Sample # 2	0.262	16:53
Breath Volume	1.351	16:53
Air Blank	0.000	16:54
Control Test	0.076	16:54
Air Blank	0.000	16:55
Diagnostics Check	OK	16:55

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENEZIA,STEVEN,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 06/04/2021 04:38:25

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.078	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:34	No Sample Provided
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample	# 2	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:37	
Control Test	0.078	04:37	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GONZALES,PAUL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: OLMSTEAD

DATE/TIME OF TEST: 06/04/2021 23:12:56

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.078	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:05	No Sample Provided
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Air Blank	0.000	23:07	
Subject Sample	# 2	0.000	23:08
Breath Volume	1.699	23:08	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 3	0.000	23:11
Breath Volume	1.589	23:11	
Air Blank	0.000	23:11	
Control Test	0.078	23:12	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: VALDIVIA PINTO

DATE/TIME OF TEST: 06/05/2021 05:09:41

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.077	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.178	05:04
Breath Volume	3.023	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.170	05:07
Breath Volume	2.671	05:07
Air Blank	0.000	05:08
Control Test	0.077	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBRIEN,DANIEL,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,A
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MONAST

DATE/TIME OF TEST: 06/06/2021 03:30:24

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.077	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.066	03:25
Breath Volume	2.058	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.063	03:28
Breath Volume	1.925	03:28
Air Blank	0.000	03:29
Control Test	0.076	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SILIS

DATE/TIME OF TEST: 06/06/2021 03:50:06

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.145	03:45
Breath Volume	2.425	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.139	03:48
Breath Volume	1.468	03:48
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHEO,FRANCISCO,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILILLO,NICHOLAS
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 06/07/2021 18:51:30

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:48	
Control Test	0.078	18:48	
Air Blank	0.000	18:49	
Subject Sample	# 1 0.197	18:49	
Breath Volume	1.597	18:49	
Air Blank		18:50	Purge Fail
Air Blank		18:51	Purge Fail

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHEO,FRANCISCO,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILILLO,NICHOLAS
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 06/07/2021 18:54:32

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:53	
Control Test	0.050	18:53	Control Outside Tolerance
Air Blank	0.000	18:54	

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNEAU, OFFICER,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU, OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/12/2021 16:27:32

Diagnostics Check	OK	16:20
Air Blank	0.000	16:21
Control Test	0.077	16:21
Air Blank	0.000	16:22
Subject Sample # 1	0.097	16:22
Breath Volume	2.562	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.096	16:25
Breath Volume	2.433	16:25
Air Blank	0.000	16:26
Control Test	0.077	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: ORTEGA BOLANOS

DATE/TIME OF TEST: 06/12/2021 22:31:37

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.078	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.079	22:31	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 06/16/2021 02:43:18

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.234	02:38
Breath Volume	1.578	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.250	02:41
Breath Volume	2.300	02:41
Air Blank	0.000	02:42
Control Test	0.076	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: QUINN SCRUGGS

DATE/TIME OF TEST: 06/17/2021 01:11:40

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.223	01:02
Breath Volume	1.855	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.259	01:06
Breath Volume	3.207	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 3	0.231	01:09
Breath Volume	1.523	01:09
Air Blank	0.000	01:10
Control Test	0.076	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HAAGSMA

DATE/TIME OF TEST: 06/18/2021 02:10:21

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.078	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Control Test	0.077	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRUNEAU,DAVE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/19/2021 03:07:04

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.219	03:02
Breath Volume	2.949	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.213	03:05
Breath Volume	2.457	03:05
Air Blank	0.000	03:06
Control Test	0.077	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:07

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MORFIN

DATE/TIME OF TEST: 06/19/2021 04:16:31

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.091	04:10
Breath Volume	1.468	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.093	04:14
Breath Volume	1.222	04:14
Air Blank	0.000	04:15
Control Test	0.077	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 07/02/2021 20:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARNES,B,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 06/19/2021 19:45:42

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.078	19:39
Air Blank	0.000	19:39
Subject Sample # 1	0.154	19:40
Breath Volume	1.582	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.155	19:44
Breath Volume	1.687	19:44
Air Blank	0.000	19:44
Control Test	0.077	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LAPENNA

DATE/TIME OF TEST: 06/19/2021 23:38:49

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.077	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.082	23:34
Breath Volume	2.988	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.082	23:37
Breath Volume	3.527	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MARQUEZ RODRIGUEZ

DATE/TIME OF TEST: 06/20/2021 01:18:52

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.000	01:14
Breath Volume	3.027	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.000	01:17
Breath Volume	3.195	01:17
Air Blank	0.000	01:17
Control Test	0.078	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: ROBINSON 2

DATE/TIME OF TEST: 06/20/2021 02:26:34

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.077	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.124	02:21
Breath Volume	2.609	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.136	02:24
Breath Volume	2.964	02:24
Air Blank	0.000	02:25
Control Test	0.077	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 06/20/2021 03:33:27

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.131	03:28
Breath Volume	2.019	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.128	03:31
Breath Volume	1.878	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PETRI

DATE/TIME OF TEST: 06/24/2021 02:08:07

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.057	02:03
Breath Volume	1.539	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.053	02:06
Breath Volume	2.757	02:06
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 06/24/2021 23:19:33

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.079	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:12	No Sample Provided
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample	# 2	23:18	No Sample Provided
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/25/2021 01:28:36

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.117	01:23
Breath Volume	2.890	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.119	01:26
Breath Volume	2.218	01:26
Air Blank	0.000	01:27
Control Test	0.076	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/25/2021 23:05:26

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.078	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.066	23:00
Breath Volume	2.292	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.064	23:03
Breath Volume	1.851	23:03
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: WINNING

DATE/TIME OF TEST: 06/26/2021 02:07:54

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.069	02:03
Breath Volume	4.109	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.070	02:06
Breath Volume	4.179	02:06
Air Blank	0.000	02:06
Control Test	0.077	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,S,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,S
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KOGAN

DATE/TIME OF TEST: 06/26/2021 03:45:37

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.164	03:39
Breath Volume	3.476	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.157	03:43
Breath Volume	2.832	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,HENRY,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,HENRY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: MAYE PENA

DATE/TIME OF TEST: 06/26/2021 23:12:01

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.077	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:07
Breath Volume	2.250	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:10
Breath Volume	2.300	23:10
Air Blank	0.000	23:11
Control Test	0.077	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:12

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REID,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: GRIMALDI

DATE/TIME OF TEST: 06/27/2021 05:14:59

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.077	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.187	05:09
Breath Volume	1.628	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.176	05:13
Breath Volume	1.410	05:13
Air Blank	0.000	05:14
Control Test	0.078	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 07/02/2021 20:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FLORES HERNANDEZ

DATE/TIME OF TEST: 06/03/2021 03:41:17

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.000	03:36
Breath Volume	2.316	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.000	03:39
Breath Volume	2.660	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 06/10/2021 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BLENMAN JR

DATE/TIME OF TEST: 06/03/2021 04:54:49

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.215	04:50
Breath Volume	2.019	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.221	04:53
Breath Volume	2.238	04:53
Air Blank	0.000	04:53
Control Test	0.079	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 06/10/2021 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 06/09/2021 04:42:08

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.079	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.170	04:37
Breath Volume	2.968	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:39
Subject Sample # 2	0.156	04:40
Breath Volume	1.671	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 06/10/2021 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,RYAN,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: OSTERTAG,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: KRAWCZAK

DATE/TIME OF TEST: 06/12/2021 01:37:06

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.079	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 07/12/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 06/12/2021 04:25:57

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.199	04:20
Breath Volume	1.449	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.204	04:24
Breath Volume	2.289	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 07/12/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: REINIG

DATE/TIME OF TEST: 06/12/2021 16:02:23

Diagnostics Check	OK	15:58	
Air Blank	0.000	15:59	
Control Test	0.079	15:59	
Air Blank	0.000	16:00	
Subject Sample	# 1	16:00	Subject Test Refused
Breath Volume	0.000	16:00	
Air Blank	0.000	16:01	
Control Test	0.078	16:01	
Air Blank	0.000	16:02	
Diagnostics Check	OK	16:02	

Data Download Date/Time: 07/12/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: STAMM

DATE/TIME OF TEST: 06/17/2021 01:53:33

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.218	01:47
Breath Volume	1.503	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample # 2	0.237	01:51
Breath Volume	2.457	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 07/12/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/17/2021 23:17:28

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.078	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.285	23:12
Breath Volume	2.414	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.287	23:15
Breath Volume	2.382	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 07/12/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/18/2021 23:03:31

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.079	23:02	
Air Blank		23:03	RFI Detect
Air Blank	0.000	23:03	

Data Download Date/Time: 07/12/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/18/2021 23:08:54

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.079	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1 0.288	23:07	
Breath Volume	1.218	23:07	
Air Blank		23:08	Purge Fail
Air Blank	0.000	23:08	

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALLUCCA,MARIO,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/18/2021 23:18:20

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.076	23:11	
Air Blank	0.000	23:12	
Subject Sample # 1	0.285	23:13	
Breath Volume	2.167	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:15	
Subject Sample # 2	0.284	23:16	
Breath Volume	1.988	23:16	
Air Blank	0.000	23:17	
Control Test	0.073	23:17	Control Outside Tolerance
Air Blank	0.000	23:18	

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,ADAM,M
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MOORE,ADAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 06/19/2021 09:07:46

Diagnostics Check	OK	09:01
Air Blank	0.000	09:01
Control Test	0.078	09:02
Air Blank	0.000	09:02
Subject Sample # 1	0.103	09:03
Breath Volume	2.175	09:03
Air Blank	0.000	09:03
Air Blank	0.000	09:05
Subject Sample # 2	0.103	09:06
Breath Volume	1.757	09:06
Air Blank	0.000	09:06
Control Test	0.078	09:07
Air Blank	0.000	09:07
Diagnostics Check	OK	09:07

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:02

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/20/2021 12:26:43

Diagnostics Check	OK	12:23	
Air Blank	0.000	12:23	
Control Test	0.078	12:23	
Air Blank	0.000	12:24	
Subject Sample	# 1 0.209	12:24	
Breath Volume	2.476	12:24	
Air Blank		12:25	RFI Detect
Air Blank	0.000	12:25	
Control Test	0.079	12:26	
Air Blank	0.000	12:26	
Diagnostics Check	OK	12:26	

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:02

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/20/2021 12:34:18

Diagnostics Check	OK	12:27
Air Blank	0.000	12:28
Control Test	0.078	12:28
Air Blank	0.000	12:29
Subject Sample # 1	0.212	12:29
Breath Volume	2.437	12:29
Air Blank	0.000	12:30
Air Blank	0.000	12:32
Subject Sample # 2	0.195	12:32
Breath Volume	1.851	12:32
Air Blank	0.000	12:33
Control Test	0.078	12:33
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMBERLIN,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TOMBERLIN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 12:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/22/2021 16:12:23

Diagnostics Check	OK	16:09	
Air Blank	0.000	16:09	
Control Test	0.079	16:10	
Air Blank	0.000	16:10	
Subject Sample	# 1	16:11	Subject Test Refused
Breath Volume	0.000	16:11	
Air Blank	0.000	16:11	
Control Test	0.079	16:11	
Air Blank	0.000	16:12	
Diagnostics Check	OK	16:12	

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MAURICIO AVIAS

DATE/TIME OF TEST: 06/26/2021 02:09:29

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.208	02:04
Breath Volume	1.253	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.220	02:07
Breath Volume	1.410	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ALBANEZ DIAZ

DATE/TIME OF TEST: 06/28/2021 02:32:11

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 07/12/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 06/16/2021 23:03:27

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.078	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 06/17/2021 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/17/2021 04:52:48

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.043	04:48
Breath Volume	2.042	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.043	04:51
Breath Volume	2.109	04:51
Air Blank	0.000	04:51
Control Test	0.078	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 06/17/2021 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BRAYER

DATE/TIME OF TEST: 06/18/2021 03:33:00

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.204	03:28
Breath Volume	3.503	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.216	03:31
Breath Volume	3.457	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 06/18/2021 22:45:17

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.078	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.187	22:40
Breath Volume	2.464	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.190	22:43
Breath Volume	2.007	22:43
Air Blank	0.000	22:44
Control Test	0.078	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 06/19/2021 00:12:16

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:09	
Control Test	0.079	00:10	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Control Test	0.079	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS, A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 06/19/2021 02:54:30

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.058	02:50
Breath Volume	2.632	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.055	02:53
Breath Volume	2.753	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: RITCHIE

DATE/TIME OF TEST: 06/19/2021 05:59:37

Diagnostics Check	OK	05:53
Air Blank	0.000	05:53
Control Test	0.078	05:54
Air Blank	0.000	05:54
Subject Sample # 1	0.067	05:55
Breath Volume	3.832	05:55
Air Blank	0.000	05:55
Air Blank	0.000	05:57
Subject Sample # 2	0.069	05:58
Breath Volume	3.417	05:58
Air Blank	0.000	05:58
Control Test	0.078	05:59
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: ANDRADE GALVEZ

DATE/TIME OF TEST: 06/19/2021 19:54:31

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.078	19:52	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:53	Subject Test Refused
Breath Volume	0.000	19:53	
Air Blank	0.000	19:53	
Control Test	0.078	19:53	
Air Blank	0.000	19:54	
Diagnostics Check	OK	19:54	

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BUSTOS GUZMAN

DATE/TIME OF TEST: 06/20/2021 01:13:01

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.078	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.145	01:08
Breath Volume	2.808	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.139	01:11
Breath Volume	3.273	01:11
Air Blank	0.000	01:12
Control Test	0.077	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ARCA-SANTANA

DATE/TIME OF TEST: 06/20/2021 03:05:05

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.078	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.078	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: QUINTANA,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/20/2021 04:48:37

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.078	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.072	04:44
Breath Volume	3.558	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.075	04:47
Breath Volume	3.859	04:47
Air Blank	0.000	04:47
Control Test	0.078	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 06/21/2021 00:42:11

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.078	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.135	00:37
Breath Volume	1.281	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.137	00:40
Breath Volume	1.546	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 06/21/2021 22:59:35

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.247	22:55
Breath Volume	2.812	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.246	22:58
Breath Volume	2.582	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/01/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: KRAMP

DATE/TIME OF TEST: 06/22/2021 21:54:03

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.079	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:53	
Control Test	0.077	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGGINS,R,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOODY,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: MONTOYA BIASINO

DATE/TIME OF TEST: 06/23/2021 20:52:57

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.079	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:52	
Control Test	0.078	20:52	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: EASTERLING

DATE/TIME OF TEST: 06/23/2021 23:29:29

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.151	23:24
Breath Volume	2.222	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.150	23:27
Breath Volume	2.484	23:27
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 06/24/2021 01:57:28

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.134	01:52
Breath Volume	3.058	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.141	01:55
Breath Volume	2.394	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORIS,MATUS,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MATIS

DATE/TIME OF TEST: 06/24/2021 03:21:45

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.153	03:17
Breath Volume	1.410	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.151	03:20
Breath Volume	1.718	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KRAMER,S
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/25/2021 02:42:22

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.131	02:37
Breath Volume	1.570	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.136	02:40
Breath Volume	1.738	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 06/25/2021 23:34:47

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.271	23:30
Breath Volume	1.902	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.266	23:33
Breath Volume	1.988	23:33
Air Blank	0.000	23:33
Control Test	0.078	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: OLEJNICZAK

DATE/TIME OF TEST: 06/26/2021 03:01:48

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.156	02:57
Breath Volume	3.210	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.146	03:00
Breath Volume	2.441	03:00
Air Blank	0.000	03:00
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART, A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: JEWETT

DATE/TIME OF TEST: 06/26/2021 04:34:50

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.161	04:30
Breath Volume	1.554	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.161	04:33
Breath Volume	1.710	04:33
Air Blank	0.000	04:33
Control Test	0.078	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DRY

DATE/TIME OF TEST: 06/26/2021 23:46:53

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.065	23:42
Breath Volume	2.472	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.065	23:45
Breath Volume	2.781	23:45
Air Blank	0.000	23:45
Control Test	0.078	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MCCOLLEY

DATE/TIME OF TEST: 06/27/2021 02:07:22

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.078	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS, CHRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS, A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/27/2021 03:06:57

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.000	03:02
Breath Volume	2.312	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.000	03:05
Breath Volume	1.585	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: QUATTLEBAUM

DATE/TIME OF TEST: 06/27/2021 21:49:27

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.079	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.079	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 06/27/2021 22:49:24

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.078	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.094	22:44
Breath Volume	2.144	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:46
Subject Sample # 2	0.093	22:47
Breath Volume	2.421	22:47
Air Blank	0.000	22:48
Control Test	0.078	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 07/01/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MENDOZA GARCIA

DATE/TIME OF TEST: 06/27/2021 23:57:50

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.166	23:53
Breath Volume	2.128	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.161	23:56
Breath Volume	2.031	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 07/01/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 06/28/2021 01:24:50

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.079	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 07/01/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/29/2021 01:08:59

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.078	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 07/01/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: BUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GICALONE

DATE/TIME OF TEST: 06/30/2021 00:11:23

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.102	00:06
Breath Volume	1.917	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.101	00:09
Breath Volume	1.953	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 07/01/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CASH

DATE/TIME OF TEST: 06/30/2021 01:59:42

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:57	
Control Test	0.078	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.079	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 07/01/2021 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOODY,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/30/2021 03:52:54

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.026	03:48
Breath Volume	1.277	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.025	03:51
Breath Volume	1.378	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/01/2021 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: YOUNG,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/30/2021 19:50:38

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:48	
Control Test	0.078	19:48	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:49	
Control Test	0.078	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 07/01/2021 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSINO,LUIS,H
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: JUSINO,LUIS
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DUNNOM

DATE/TIME OF TEST: 06/20/2021 01:20:11

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:15	
Air Blank	0.000	01:15	
Subject Sample # 1	0.305	01:16	
Breath Volume	3.503	01:16	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample # 2	0.308	01:19	
Breath Volume	2.031	01:19	
Air Blank		01:19	RFI Detect
Air Blank	0.000	01:20	

Data Download Date/Time: 07/30/2021 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNNEFELD,RALPH,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HILL,CAMRYN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:08

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 06/05/2021 09:34:15

Diagnostics Check	OK	09:28
Air Blank	0.000	09:28
Control Test	0.080	09:28
Air Blank	0.000	09:29
Subject Sample # 1	0.000	09:29
Breath Volume	1.835	09:29
Air Blank	0.000	09:30
Air Blank	0.000	09:32
Subject Sample # 2	0.000	09:32
Breath Volume	3.480	09:32
Air Blank	0.000	09:33
Control Test	0.080	09:33
Air Blank	0.000	09:34
Diagnostics Check	OK	09:34

Data Download Date/Time: 06/30/2021 07:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: VELLADARES,CARLOS
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 980558
DRY GAS EXPIRY DATE: 07/09/2021
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DAOSTA

DATE/TIME OF TEST: 06/09/2021 05:29:37

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:27	
Control Test	0.079	05:27	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:28	
Control Test	0.079	05:29	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

Data Download Date/Time: 06/23/2021 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CRUZ,EVI,

BREATH TEST OPERATOR AGENCY: HIALEAH PD

ARRESTING OFFICER: NUNEZ,J

ARRESTING OFFICER AGENCY: HIALEAH PD

LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1224982

DRY GAS EXPIRY DATE: 10/31/2022

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: DOLZ

DATE/TIME OF TEST: 06/03/2021 03:11:18

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.078	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.077	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 06/10/2021 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LAWLER,ADDISON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/01/2021 01:46:55

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.196	01:42
Breath Volume	1.621	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.180	01:45
Breath Volume	2.789	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 06/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 06/10/2021 23:45:50

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.203	23:41
Breath Volume	2.144	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.201	23:44
Breath Volume	2.628	23:44
Air Blank	0.000	23:44
Control Test	0.077	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 06/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 06/16/2021 20:26:42

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.078	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.000	20:22
Breath Volume	1.773	20:22
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.000	20:25
Breath Volume	1.867	20:25
Air Blank	0.000	20:25
Control Test	0.078	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 07/08/2021 07:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/26/2021 21:59:54

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.078	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.082	21:55
Breath Volume	2.671	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.086	21:58
Breath Volume	3.074	21:58
Air Blank	0.000	21:58
Control Test	0.079	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 07/08/2021 07:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARROQUIN,RICHARD,J

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: PROVENCHER,JOSH

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG009801

DRY GAS EXPIRY DATE: 04/07/2022

VIOLATION CODE: DUI

ARREST TIME: 16:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 06/05/2021 17:21:31

Diagnostics Check	OK	17:21	
Air Blank		17:21	RFI Detect
Air Blank		17:21	RFI Detect

Data Download Date/Time: 06/30/2021 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARROQUIN,RICHARD,J

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: PROVENCHER,JOSH

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 05/20/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG009801

DRY GAS EXPIRY DATE: 04/07/2022

VIOLATION CODE: DUI

ARREST TIME: 16:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 06/05/2021 17:30:40

Diagnostics Check	OK	17:24
Air Blank	0.000	17:24
Control Test	0.077	17:25
Air Blank	0.000	17:25
Subject Sample # 1	0.133	17:26
Breath Volume	3.488	17:26
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.129	17:29
Breath Volume	3.359	17:29
Air Blank	0.000	17:29
Control Test	0.076	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 06/30/2021 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: GORE,MATTHEW,A	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: GORE,MATTHEW	ARREST TIME: 18:38
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/25/2021	OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BAUGHMAN DATE/TIME OF TEST: 06/06/2021 19:33:47

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.079	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.107	19:29
Breath Volume	3.542	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.092	19:32
Breath Volume	1.359	19:32
Air Blank	0.000	19:32
Control Test	0.079	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 06/28/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD,NICOLE	ARREST TIME: 16:38
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/25/2021	OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: ALLEN DATE/TIME OF TEST: 06/11/2021 17:09:16

Diagnostics Check	OK	17:08	
Air Blank		17:09	RFI Detect
Air Blank		17:09	RFI Detect

Data Download Date/Time: 06/28/2021 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1249794
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 12/13/2022
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BASFORD,NICOLE	ARREST TIME: 16:38
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/25/2021	OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: ALLEN	DATE/TIME OF TEST: 06/11/2021 17:22:39
----------------------------	--

Diagnostics Check	OK	17:19	
Air Blank	0.000	17:20	
Control Test	0.079	17:20	
Air Blank	0.000	17:21	
Subject Sample	# 1	17:21	Subject Test Refused
Breath Volume	0.000	17:21	
Air Blank	0.000	17:21	
Control Test	0.079	17:22	
Air Blank	0.000	17:22	
Diagnostics Check	OK	17:22	

Data Download Date/Time: 06/28/2021 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHANSEN,LLOYD,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HAMMELMAN,TYLER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/01/2021 03:13:37

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.000	03:09
Breath Volume	1.343	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.007	03:12
Breath Volume	1.628	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 06/16/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMARCO,AUSTIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON,WILLIAM
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: XOLO SOSME

DATE/TIME OF TEST: 06/04/2021 22:32:46

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.195	22:28
Breath Volume	2.988	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.197	22:31
Breath Volume	1.910	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 06/16/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/05/2021 02:46:55

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.232	02:42
Breath Volume	2.242	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.245	02:45
Breath Volume	2.753	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 06/16/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SAUER

DATE/TIME OF TEST: 06/06/2021 04:09:15

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.081	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.000	04:00	Volume Not Met
Breath Volume	0.195	04:00	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2 0.000	04:04	
Breath Volume	1.613	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 3 0.000	04:07	
Breath Volume	1.511	04:07	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 06/16/2021 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY, RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/07/2021 00:55:23

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.229	00:47
Breath Volume	1.171	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.258	00:50
Breath Volume	2.308	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 3	0.255	00:53
Breath Volume	1.917	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/16/2021 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: KEELING

DATE/TIME OF TEST: 06/07/2021 23:40:04

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.080	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1 0.017	23:33	Volume Not Met
Breath Volume	0.203	23:33	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2 0.015	23:38	Volume Not Met
Breath Volume	0.105	23:38	
Air Blank	0.000	23:39	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/08/2021 14:55:24

Diagnostics Check	OK	14:50	
Air Blank	0.000	14:50	
Control Test	0.079	14:51	
Air Blank	0.000	14:51	
Subject Sample	# 1	14:54	Subject Test Refused
Breath Volume	0.000	14:54	
Air Blank	0.000	14:54	
Control Test	0.080	14:54	
Air Blank	0.000	14:55	
Diagnostics Check	OK	14:55	

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMAR,GIOVANNI,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MERCIER FEHR

DATE/TIME OF TEST: 06/09/2021 01:54:36

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.009	01:50
Breath Volume	2.691	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.010	01:53
Breath Volume	2.179	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMARCO,AUSTIN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DIMARCO,AUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HULSEN

DATE/TIME OF TEST: 06/10/2021 18:59:09

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.080	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.056	18:54
Breath Volume	3.812	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.057	18:57
Breath Volume	3.902	18:57
Air Blank	0.000	18:58
Control Test	0.080	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NARINE,NANDKISHORE,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/10/2021 22:19:11

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:09	
Control Test	0.080	22:09	
Air Blank	0.000	22:10	
Subject Sample	# 1 0.076	22:13	Volume Not Met
Breath Volume	0.191	22:13	
Air Blank	0.000	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 2	22:17	Subject Test Refused
Breath Volume	0.136	22:17	
Air Blank	0.000	22:18	
Control Test	0.080	22:18	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NARINE,NANDKISHORE,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/10/2021 22:32:03

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.080	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1 0.070	22:26	Volume Not Met
Breath Volume	0.300	22:26	
Air Blank	0.000	22:27	
Air Blank	0.000	22:28	
Subject Sample	# 2	22:30	Subject Test Refused
Breath Volume	0.359	22:30	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: VERGARA

DATE/TIME OF TEST: 06/11/2021 07:35:13

Diagnostics Check	OK	07:28
Air Blank	0.000	07:29
Control Test	0.079	07:29
Air Blank	0.000	07:30
Subject Sample # 1	0.133	07:30
Breath Volume	4.347	07:30
Air Blank	0.000	07:31
Air Blank	0.000	07:33
Subject Sample # 2	0.131	07:33
Breath Volume	4.019	07:33
Air Blank	0.000	07:34
Control Test	0.080	07:34
Air Blank	0.000	07:35
Diagnostics Check	OK	07:35

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: A411253914130

DATE/TIME OF TEST: 06/13/2021 05:23:18

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.079	05:21	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 06/14/2021 04:59:29

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.040	04:54
Breath Volume	3.562	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.040	04:58
Breath Volume	3.585	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 06/16/2021 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RUIZCHAVEZ

DATE/TIME OF TEST: 06/05/2021 03:38:25

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.187	03:33
Breath Volume	3.445	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.180	03:36
Breath Volume	3.195	03:36
Air Blank	0.000	03:37
Control Test	0.077	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 08/02/2021 23:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ORDONEZ

DATE/TIME OF TEST: 06/06/2021 04:29:20

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.133	04:24
Breath Volume	1.316	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.137	04:27
Breath Volume	2.722	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/02/2021 23:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: ENCARNACION,HECTOR
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BENAVIDES

DATE/TIME OF TEST: 06/10/2021 02:51:54

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.078	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1 0.190	02:40	Volume Not Met
Breath Volume	0.066	02:40	
Air Blank	0.000	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 2 0.221	02:46	
Breath Volume	1.664	02:46	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 3 0.214	02:50	
Breath Volume	1.539	02:50	
Air Blank	0.000	02:50	
Control Test	0.078	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 08/02/2021 23:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: ENCARNACION,HECTOR
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BENAVIDES

DATE/TIME OF TEST: 06/10/2021 03:03:58

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.078	02:56	
Air Blank	0.000	02:57	
Subject Sample # 1	0.222	02:59	
Breath Volume	1.398	02:59	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample # 2	0.204	03:02	
Breath Volume	1.863	03:02	
Air Blank	0.000	03:03	
Control Test	0.074	03:03	Control Outside Tolerance
Air Blank	0.000	03:03	

Data Download Date/Time: 08/02/2021 23:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: ENCARNACION,HECTOR
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BENAVIDES

DATE/TIME OF TEST: 06/10/2021 03:18:09

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.077	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.207	03:13
Breath Volume	1.468	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.197	03:16
Breath Volume	1.375	03:16
Air Blank	0.000	03:17
Control Test	0.076	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 08/02/2021 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: PAVON

DATE/TIME OF TEST: 06/26/2021 04:15:12

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank		04:14	RFI Detect
Air Blank	0.000	04:15	

Data Download Date/Time: 08/02/2021 23:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYERS,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 06/01/2021 03:15:23

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.205	03:10
Breath Volume	1.527	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.209	03:13
Breath Volume	2.144	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 06/02/2021 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: TRUCKS

DATE/TIME OF TEST: 06/02/2021 00:36:54

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.197	00:30
Breath Volume	1.359	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.212	00:35
Breath Volume	2.667	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 06/02/2021 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRIAN,A
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: SMALL,BRIAN
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/22/2021 14:45:16

Diagnostics Check	OK	14:43	
Air Blank	0.000	14:44	
Control Test	0.201	14:44	Control Outside Tolerance
Air Blank	0.000	14:45	

Data Download Date/Time: 07/08/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRIAN,A
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: SMALL,BRIAN
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/22/2021 15:00:30

Diagnostics Check	OK	14:54
Air Blank	0.000	14:54
Control Test	0.077	14:55
Air Blank	0.000	14:55
Subject Sample # 1	0.203	14:56
Breath Volume	1.671	14:56
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.210	14:59
Breath Volume	1.523	14:59
Air Blank	0.000	14:59
Control Test	0.077	14:59
Air Blank	0.000	15:00
Diagnostics Check	OK	15:00

Data Download Date/Time: 07/08/2021 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRIAN,A
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: SMALL,BRIAN
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/22/2021 16:08:15

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.078	16:02
Air Blank	0.000	16:03
Subject Sample # 1	0.000	16:04
Breath Volume	1.953	16:04
Air Blank	0.000	16:04
Air Blank	0.000	16:06
Subject Sample # 2	0.000	16:06
Breath Volume	2.171	16:06
Air Blank	0.000	16:07
Control Test	0.078	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 07/08/2021 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUI,KHANG,M
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: BUI,KHANG
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/30/2021 11:32:59

Diagnostics Check	OK	11:27
Air Blank	0.000	11:27
Control Test	0.078	11:27
Air Blank	0.000	11:28
Subject Sample # 1	0.000	11:28
Breath Volume	1.769	11:28
Air Blank	0.000	11:29
Air Blank	0.000	11:31
Subject Sample # 2	0.000	11:31
Breath Volume	1.800	11:31
Air Blank	0.000	11:32
Control Test	0.078	11:32
Air Blank	0.000	11:32
Diagnostics Check	OK	11:32

Data Download Date/Time: 07/08/2021 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 06/11/2021 13:39:02

Diagnostics Check	OK	13:32	
Air Blank	0.000	13:33	
Control Test	0.079	13:33	
Air Blank	0.000	13:33	
Subject Sample	# 1 0.238	13:34	
Breath Volume	1.433	13:34	
Air Blank	0.000	13:34	
Air Blank	0.000	13:36	
Subject Sample	# 2 0.257	13:37	
Breath Volume	2.011	13:37	
Air Blank	0.000	13:37	
Control Test	0.080	13:38	
Air Blank		13:38	RFI Detect
Air Blank	0.000	13:39	

Data Download Date/Time: 06/25/2021 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 01:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ALDERMAN

DATE/TIME OF TEST: 06/19/2021 02:36:10

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.149	02:31
Breath Volume	1.945	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.150	02:34
Breath Volume	1.953	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 07/09/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 01:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/19/2021 02:40:24

Diagnostics Check	OK	02:39	
Air Blank		02:39	RFI Detect
Air Blank	0.000	02:40	

Data Download Date/Time: 07/09/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 01:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/19/2021 02:49:01

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.120	02:44
Breath Volume	1.250	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.116	02:47
Breath Volume	2.472	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:49

Data Download Date/Time: 07/09/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: LAFONTANT,SEDONA

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 03:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HUTTO

DATE/TIME OF TEST: 06/19/2021 04:09:14

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2	0.274	04:04
Breath Volume	1.640	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 3	0.261	04:07
Breath Volume	1.425	04:07	
Air Blank	0.000	04:08	
Control Test	0.080	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 07/09/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICKINGS,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: JAMES,JACKIE
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/24/2021 13:19:40

Diagnostics Check	OK	13:13
Air Blank	0.000	13:13
Control Test	0.079	13:13
Air Blank	0.000	13:14
Subject Sample # 1	0.232	13:15
Breath Volume	2.023	13:15
Air Blank	0.000	13:15
Air Blank	0.000	13:17
Subject Sample # 2	0.217	13:18
Breath Volume	1.273	13:18
Air Blank	0.000	13:18
Control Test	0.080	13:19
Air Blank	0.000	13:19
Diagnostics Check	OK	13:19

Data Download Date/Time: 07/09/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 00:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WATTERS

DATE/TIME OF TEST: 06/26/2021 01:27:00

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.081	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.150	01:22
Breath Volume	2.148	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.144	01:25
Breath Volume	1.750	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:27

Data Download Date/Time: 07/09/2021 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 18:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HUTCHINS

DATE/TIME OF TEST: 06/27/2021 19:26:21

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.000	19:22
Breath Volume	3.078	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.000	19:24
Breath Volume	2.984	19:24
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 07/09/2021 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CURRIE,MATTHEW,B

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: CURRIE,MATTHEW

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 05/30/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: DUI

ARREST TIME: 22:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: RIVERA-RODRIGUEZ

DATE/TIME OF TEST: 06/26/2021 23:07:17

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.080	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:06	
Control Test	0.081	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 07/06/2021 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 06/01/2021 22:28:03

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.188	22:23
Breath Volume	2.683	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.196	22:26
Breath Volume	2.269	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:28

Data Download Date/Time: 06/24/2021 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: BOURCIER

DATE/TIME OF TEST: 06/04/2021 20:51:09

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.000	20:46
Breath Volume	2.812	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.000	20:49
Breath Volume	2.890	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 06/24/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 06/05/2021 21:39:19

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.086	21:34
Breath Volume	3.058	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.087	21:37
Breath Volume	3.078	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 06/24/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MORALES MORALES

DATE/TIME OF TEST: 06/07/2021 02:58:31

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.122	02:54
Breath Volume	2.031	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.122	02:57
Breath Volume	1.882	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 06/24/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/10/2021 22:12:42

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.256	22:08
Breath Volume	1.117	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.270	22:11
Breath Volume	1.523	22:11
Air Blank	0.000	22:11
Control Test	0.077	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 06/24/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 06/14/2021 21:04:17

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.079	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.000	21:00
Breath Volume	2.265	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:02
Breath Volume	2.015	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 06/24/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: LANDIS

DATE/TIME OF TEST: 06/19/2021 03:04:56

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.080	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:59	No Sample Provided
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 2	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.080	03:04	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 06/24/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: BREEDEN

DATE/TIME OF TEST: 06/23/2021 16:46:27

Diagnostics Check	OK	16:39
Air Blank	0.000	16:39
Control Test	0.080	16:40
Air Blank	0.000	16:40
Subject Sample # 1	0.183	16:41
Breath Volume	1.183	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:43
Subject Sample # 2	0.178	16:44
Breath Volume	1.570	16:44
Air Blank	0.000	16:45
Control Test	0.079	16:45
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 06/24/2021 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 06/25/2021 18:59:11

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:54
Subject Sample # 1	0.056	18:54
Breath Volume	2.972	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.054	18:57
Breath Volume	3.265	18:57
Air Blank	0.000	18:58
Control Test	0.078	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 08/05/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,S,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERNIER,N
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 06/26/2021 00:50:39

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.179	00:46
Breath Volume	1.832	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.178	00:49
Breath Volume	1.722	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 08/05/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,C
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MYSLINSKI

DATE/TIME OF TEST: 06/26/2021 02:18:59

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.091	02:14
Breath Volume	2.679	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.089	02:17
Breath Volume	2.828	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 08/05/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SKINNER,OFC
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CABREJA

DATE/TIME OF TEST: 06/26/2021 03:53:54

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.194	03:46
Breath Volume	2.167	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.167	03:49
Breath Volume	1.269	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 3	0.182	03:52
Breath Volume	1.886	03:52
Air Blank	0.000	03:52
Control Test	0.076	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 08/05/2021 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A2
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 06/27/2021 05:44:20

Diagnostics Check	OK	05:37
Air Blank	0.000	05:37
Control Test	0.079	05:38
Air Blank	0.000	05:38
Subject Sample # 1	0.101	05:39
Breath Volume	1.164	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.102	05:42
Breath Volume	1.507	05:42
Air Blank	0.000	05:43
Control Test	0.080	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 08/05/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 31420080A2
BREATH TEST OPERATOR: DOSTALY, STEVENS,	DRY GAS EXPIRY DATE: 01/05/2023
BREATH TEST OPERATOR AGENCY: NORTH MIAMI BEACH PD	VIOLATION CODE: DUI
ARRESTING OFFICER: QUINTERO, JUSTIN	ARREST TIME: 23:58
ARRESTING OFFICER AGENCY: NORTH MIAMI PD	GENDER: Male
LAST AGENCY INSPECTION: 06/01/2021	OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: GUTIERREZ	DATE/TIME OF TEST: 06/29/2021 00:40:29
--------------------------------	--

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.081	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 07/22/2021 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: UZONWANNE

DATE/TIME OF TEST: 06/04/2021 00:35:14

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.179	00:30
Breath Volume	2.429	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.173	00:33
Breath Volume	1.914	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/04/2021 02:14:34

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.113	02:09
Breath Volume	2.328	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.115	02:13
Breath Volume	2.250	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: CLEMENT

DATE/TIME OF TEST: 06/04/2021 04:21:10

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.076	04:16
Breath Volume	3.042	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.068	04:19
Breath Volume	2.039	04:19
Air Blank	0.000	04:20
Control Test	0.077	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: MAHECHA-SEPULVEDA

DATE/TIME OF TEST: 06/04/2021 04:51:01

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.173	04:44
Breath Volume	1.550	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.168	04:49
Breath Volume	1.707	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:51

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: MASSARO

DATE/TIME OF TEST: 06/04/2021 22:18:55

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.079	22:07	
Air Blank	0.000	22:07	
Subject Sample # 1	0.000	22:08	
Breath Volume	1.113	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample # 2	0.000	22:14	Volume Not Met
Breath Volume	0.605	22:14	
Air Blank	0.000	22:14	
Air Blank	0.000	22:16	
Subject Sample # 3	0.000	22:17	
Breath Volume	1.628	22:17	
Air Blank	0.000	22:18	
Control Test	0.079	22:18	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 06/04/2021 23:59:58

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.088	23:55
Breath Volume	2.714	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.086	23:58
Breath Volume	3.273	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 06/05/2021 03:50:17

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.208	03:45
Breath Volume	1.722	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.204	03:48
Breath Volume	1.371	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAIRSTON,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WORTH

DATE/TIME OF TEST: 06/06/2021 01:54:01

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.151	01:49
Breath Volume	2.011	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.153	01:52
Breath Volume	2.066	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: HERNANDEZ REVOLORIO

DATE/TIME OF TEST: 06/06/2021 02:35:55

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.219	02:31
Breath Volume	3.453	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.234	02:34
Breath Volume	3.152	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 06/06/2021 06:00:02

Diagnostics Check	OK	05:53
Air Blank	0.000	05:53
Control Test	0.079	05:53
Air Blank	0.000	05:54
Subject Sample # 1	0.125	05:55
Breath Volume	2.117	05:55
Air Blank	0.000	05:55
Air Blank	0.000	05:57
Subject Sample # 2	0.141	05:58
Breath Volume	2.105	05:58
Air Blank	0.000	05:59
Control Test	0.079	05:59
Air Blank	0.000	05:59
Diagnostics Check	OK	06:00

Data Download Date/Time: 07/16/2021 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEWART,J 257781
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: SCHAEFFLER

DATE/TIME OF TEST: 06/07/2021 15:13:59

Diagnostics Check	OK	15:03
Air Blank	0.000	15:04
Control Test	0.080	15:04
Air Blank	0.000	15:04
Subject Sample # 1	0.208	15:05
Breath Volume	1.449	15:05
Air Blank	0.000	15:06
Air Blank	0.000	15:07
Subject Sample # 2	0.235	15:09
Breath Volume	2.425	15:09
Air Blank	0.000	15:09
Air Blank	0.000	15:11
Subject Sample # 3	0.215	15:12
Breath Volume	1.828	15:12
Air Blank	0.000	15:13
Control Test	0.078	15:13
Air Blank	0.000	15:13
Diagnostics Check	OK	15:13

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: SGOBBO

DATE/TIME OF TEST: 06/08/2021 16:50:52

Diagnostics Check	OK	16:45
Air Blank	0.000	16:45
Control Test	0.080	16:45
Air Blank	0.000	16:46
Subject Sample # 1	0.000	16:46
Breath Volume	2.199	16:46
Air Blank	0.000	16:47
Air Blank	0.000	16:49
Subject Sample # 2	0.000	16:49
Breath Volume	2.300	16:49
Air Blank	0.000	16:49
Control Test	0.080	16:50
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 06/10/2021 01:32:33

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.133	01:28
Breath Volume	3.433	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.127	01:31
Breath Volume	2.074	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: P
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 06/10/2021 03:16:40

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.175	03:12
Breath Volume	2.324	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.171	03:15
Breath Volume	2.507	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 06/11/2021 01:02:48

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.160	00:58
Breath Volume	2.382	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.155	01:01
Breath Volume	1.761	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 06/11/2021 01:42:14

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.167	01:37
Breath Volume	1.339	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.167	01:40
Breath Volume	1.281	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 06/12/2021 03:13:19

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.164	03:08
Breath Volume	3.132	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.167	03:11
Breath Volume	2.945	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: PORTOCARRERA

DATE/TIME OF TEST: 06/13/2021 00:34:22

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.221	00:26
Breath Volume	3.050	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.198	00:29
Breath Volume	1.554	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 3	0.226	00:32
Breath Volume	2.949	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SANCHEZ CIPRIANO

DATE/TIME OF TEST: 06/13/2021 00:57:27

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.247	00:52
Breath Volume	1.281	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.253	00:55
Breath Volume	1.632	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CORMIER

DATE/TIME OF TEST: 06/13/2021 03:41:07

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.150	03:36
Breath Volume	3.265	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.141	03:39
Breath Volume	3.113	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: VASQUEZ BAUTISTA

DATE/TIME OF TEST: 06/13/2021 04:16:09

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.216	04:10
Breath Volume	1.804	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:12
Subject Sample # 2	0.230	04:14
Breath Volume	1.703	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MURDOCK

DATE/TIME OF TEST: 06/13/2021 04:36:28

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.211	04:31
Breath Volume	2.800	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.212	04:34
Breath Volume	2.699	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARSON,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:46

SUBJECT'S LAST NAME: RAINEY

DATE/TIME OF TEST: 06/13/2021 11:13:17

Diagnostics Check	OK	11:07
Air Blank	0.000	11:07
Control Test	0.079	11:08
Air Blank	0.000	11:08
Subject Sample # 1	0.000	11:09
Breath Volume	1.453	11:09
Air Blank	0.000	11:09
Air Blank	0.000	11:11
Subject Sample # 2	0.000	11:11
Breath Volume	1.542	11:11
Air Blank	0.000	11:12
Control Test	0.079	11:12
Air Blank	0.000	11:13
Diagnostics Check	OK	11:13

Data Download Date/Time: 07/16/2021 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: AMBROSIO

DATE/TIME OF TEST: 06/14/2021 00:01:08

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.079	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1 0.150	23:51	Volume Not Met
Breath Volume	0.550	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 2 0.177	23:56	
Breath Volume	1.574	23:56	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample	# 3 0.162	23:59	
Breath Volume	1.292	23:59	
Air Blank	0.000	00:00	
Control Test	0.079	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: EMERSON

DATE/TIME OF TEST: 06/14/2021 01:14:21

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.079	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1 0.201	01:03	Slope Not Level
Breath Volume	1.460	01:03	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 2 0.204	01:07	
Breath Volume	1.343	01:07	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample	# 3 0.201	01:12	Volume Not Met
Breath Volume	0.804	01:12	
Air Blank	0.000	01:13	
Control Test	0.079	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 06/15/2021 04:05:24

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.198	04:00
Breath Volume	3.273	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.213	04:03
Breath Volume	2.054	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PARKER,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: GRIFFIN III

DATE/TIME OF TEST: 06/17/2021 03:00:01

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.081	02:55
Breath Volume	1.675	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.079	02:58
Breath Volume	1.582	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	03:00

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/18/2021 04:11:39

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.148	04:05
Breath Volume	1.199	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.147	04:10
Breath Volume	1.425	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 06/18/2021 07:34:14

Diagnostics Check	OK	07:27
Air Blank	0.000	07:28
Control Test	0.080	07:28
Air Blank	0.000	07:28
Subject Sample # 1	0.142	07:29
Breath Volume	1.550	07:29
Air Blank	0.000	07:30
Air Blank	0.000	07:31
Subject Sample # 2	0.151	07:32
Breath Volume	2.152	07:32
Air Blank	0.000	07:33
Control Test	0.079	07:33
Air Blank	0.000	07:34
Diagnostics Check	OK	07:34

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 06/18/2021 22:22:30

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:15	No Sample Provided
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample	# 2	22:21	No Sample Provided
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.080	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 06/18/2021 23:25:16

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.105	23:17	Volume Not Met
Breath Volume	0.253	23:17	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 2 0.124	23:20	
Breath Volume	1.660	23:20	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample	# 3 0.131	23:23	
Breath Volume	1.121	23:23	
Air Blank	0.000	23:24	
Control Test	0.079	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: CASH

DATE/TIME OF TEST: 06/19/2021 04:49:15

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:45	
Control Test	0.079	04:45	
Air Blank	0.000	04:46	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 06/20/2021 03:22:43

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.200	03:18
Breath Volume	1.585	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.197	03:21
Breath Volume	1.355	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: EADS

DATE/TIME OF TEST: 06/20/2021 03:38:12

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:27	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: VANDERSCHUUR

DATE/TIME OF TEST: 06/21/2021 03:24:46

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.242	03:20
Breath Volume	2.636	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.250	03:23
Breath Volume	3.175	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRIGHTWELL,R
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 06/22/2021 02:26:08

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:21
Breath Volume	2.113	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:24
Breath Volume	2.582	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/16/2021 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 06/22/2021 06:05:20

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.079	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.170	06:00
Breath Volume	2.273	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.162	06:03
Breath Volume	1.820	06:03
Air Blank	0.000	06:04
Control Test	0.079	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BEECHE

DATE/TIME OF TEST: 06/23/2021 01:30:49

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.188	01:25
Breath Volume	1.187	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.197	01:29
Breath Volume	1.628	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 06/25/2021 00:49:52

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.174	00:45
Breath Volume	2.035	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.186	00:48
Breath Volume	2.796	00:48
Air Blank	0.000	00:48
Control Test	0.079	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 06/25/2021 04:06:30

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.084	04:01
Breath Volume	2.105	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.080	04:05
Breath Volume	1.261	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 06/25/2021 19:34:14

Diagnostics Check	OK	19:28
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.182	19:29
Breath Volume	2.660	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.189	19:32
Breath Volume	2.871	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELAZQUEZ,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 06/26/2021 05:01:47

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.000	04:57
Breath Volume	2.531	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.000	05:00
Breath Volume	1.527	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUIZ,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/26/2021 23:34:19

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.210	23:29
Breath Volume	2.023	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.196	23:32
Breath Volume	1.605	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: PEREZ LUCAS

DATE/TIME OF TEST: 06/26/2021 23:53:22

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.130	23:48
Breath Volume	1.148	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.133	23:51
Breath Volume	1.292	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CURBELO,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: ODELL JR

DATE/TIME OF TEST: 06/27/2021 06:26:26

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.079	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.126	06:22
Breath Volume	2.679	06:22
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.126	06:24
Breath Volume	3.238	06:24
Air Blank	0.000	06:25
Control Test	0.079	06:25
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOTHEM,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: FLANNING

DATE/TIME OF TEST: 06/27/2021 07:35:49

Diagnostics Check	OK	07:23	
Air Blank	0.000	07:23	
Control Test	0.080	07:24	
Air Blank	0.000	07:24	
Subject Sample	# 1 0.086	07:27	Volume Not Met
Breath Volume	0.890	07:27	
Air Blank	0.000	07:28	
Air Blank	0.000	07:30	
Subject Sample	# 2 0.094	07:31	
Breath Volume	1.429	07:31	
Air Blank	0.000	07:31	
Air Blank	0.000	07:33	
Subject Sample	# 3 0.094	07:34	
Breath Volume	1.566	07:34	
Air Blank	0.000	07:34	
Control Test	0.079	07:35	
Air Blank	0.000	07:35	
Diagnostics Check	OK	07:35	

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BOGART

DATE/TIME OF TEST: 06/28/2021 02:52:00

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.228	02:47
Breath Volume	1.839	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.230	02:50
Breath Volume	2.023	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/16/2021 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/28/2021 17:51:57

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.080	17:46
Air Blank	0.000	17:46
Subject Sample # 1	0.304	17:47
Breath Volume	1.742	17:47
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.303	17:50
Breath Volume	2.136	17:50
Air Blank	0.000	17:51
Control Test	0.079	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 07/16/2021 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/29/2021 00:57:46

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.080	00:55	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:56	Subject Test Refused
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:57	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 07/16/2021 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: LOATMAN

DATE/TIME OF TEST: 06/29/2021 05:28:13

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:11	
Control Test	0.080	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1 0.171	05:15	Volume Not Met
Breath Volume	0.054	05:15	
Air Blank	0.000	05:16	
Air Blank	0.000	05:17	
Subject Sample	# 2 0.166	05:20	
Breath Volume	1.312	05:20	
Air Blank	0.000	05:21	
Air Blank	0.000	05:23	
Subject Sample	# 3 0.175	05:26	Volume Not Met
Breath Volume	0.019	05:26	
Air Blank	0.000	05:27	
Control Test	0.078	05:27	
Air Blank	0.000	05:28	
Diagnostics Check	OK	05:28	

Data Download Date/Time: 07/16/2021 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: LOATMAN

DATE/TIME OF TEST: 06/29/2021 06:02:23

Diagnostics Check	OK	05:54
Air Blank	0.000	05:55
Control Test	0.079	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.164	05:57
Breath Volume	1.953	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.156	06:00
Breath Volume	2.156	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 07/16/2021 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAIRSTON,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: HARWOOD

DATE/TIME OF TEST: 06/30/2021 04:53:34

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.104	04:48
Breath Volume	2.718	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.109	04:52
Breath Volume	2.992	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 07/16/2021 10:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 06/30/2021 05:32:58

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.228	05:27
Breath Volume	1.589	05:27
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.214	05:31
Breath Volume	1.355	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 07/16/2021 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: NODARSE

DATE/TIME OF TEST: 06/11/2021 19:17:36

Diagnostics Check	OK	19:11
Air Blank	0.000	19:12
Control Test	0.082	19:12
Air Blank	0.000	19:12
Subject Sample # 1	0.115	19:13
Breath Volume	2.011	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.115	19:16
Breath Volume	1.855	19:16
Air Blank	0.000	19:16
Control Test	0.082	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 06/30/2021 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,RUBEN,G
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: GOMEZ,RUBEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: NODARSE

DATE/TIME OF TEST: 06/11/2021 20:17:38

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.083	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.092	20:13
Breath Volume	1.687	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.091	20:16
Breath Volume	1.304	20:16
Air Blank	0.000	20:16
Control Test	0.082	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 06/30/2021 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDGEPEETH,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HERNANDEZ,KEITH
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: BUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/13/2021 18:13:00

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:10	
Control Test	0.083	18:10	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:11	Subject Test Refused
Breath Volume	0.000	18:11	
Air Blank	0.000	18:12	
Control Test	0.081	18:12	
Air Blank	0.000	18:12	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 06/30/2021 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAMOS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MUNOZ-CORTEZ

DATE/TIME OF TEST: 06/01/2021 01:12:52

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.078	01:06	
Air Blank	0.000	01:07	
Subject Sample # 1	0.200	01:08	
Breath Volume	3.242	01:08	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample # 2	0.198	01:11	
Breath Volume	2.464	01:11	
Air Blank		01:12	Purge Fail
Air Blank	0.000	01:12	

Data Download Date/Time: 06/14/2021 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAMOS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MUNOZ-CORTEZ

DATE/TIME OF TEST: 06/01/2021 01:20:56

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.077	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Interferent Detect
Breath Volume	2.082	01:19	
Air Blank	0.000	01:19	
Control Test	0.078	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 06/14/2021 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAMOS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MUNOZ-CORTEZ

DATE/TIME OF TEST: 06/01/2021 01:28:37

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.077	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.182	01:24
Breath Volume	1.839	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.175	01:27
Breath Volume	1.550	01:27
Air Blank	0.000	01:27
Control Test	0.077	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 06/14/2021 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 06/04/2021 00:17:31

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.162	00:12
Breath Volume	2.777	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.161	00:15
Breath Volume	3.062	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 06/14/2021 17:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LONGO,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CORDI

DATE/TIME OF TEST: 06/06/2021 03:15:46

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.239	03:10
Breath Volume	1.648	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.233	03:14
Breath Volume	2.492	03:14
Air Blank	0.000	03:14
Control Test	0.078	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 06/14/2021 17:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 06/07/2021 00:01:40

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	Interferent Detect
Air Blank	0.000	00:01	

Data Download Date/Time: 06/14/2021 17:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 06/07/2021 00:11:20

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.241	00:06	
Breath Volume	1.859	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2 0.241	00:09	
Breath Volume	1.480	00:09	
Air Blank	0.000	00:10	
Control Test	0.078	00:10	Interferent Detect
Air Blank	0.000	00:11	

Data Download Date/Time: 06/14/2021 17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 06/06/2021 00:05:17

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.000	00:00
Breath Volume	1.468	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:03
Breath Volume	2.054	00:03
Air Blank	0.000	00:04
Control Test	0.081	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 06/16/2021 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 06/11/2021 03:57:30

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.081	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 06/16/2021 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HERNANDEZ CERVANTES

DATE/TIME OF TEST: 06/11/2021 22:34:29

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.081	22:30
Breath Volume	1.140	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.081	22:32
Breath Volume	1.152	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 06/16/2021 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/13/2021 04:11:50

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 06/16/2021 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: OBIDOWICZ-URBANEK

DATE/TIME OF TEST: 06/16/2021 20:00:16

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.081	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.235	19:55
Breath Volume	2.445	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.238	19:58
Breath Volume	2.187	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 06/16/2021 23:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG011102
DRY GAS EXPIRY DATE: 04/20/2022
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: LYGNOS

DATE/TIME OF TEST: 06/24/2021 18:18:39

Diagnostics Check	OK	18:15	
Air Blank	0.000	18:16	
Control Test	0.080	18:16	
Air Blank	0.000	18:17	
Subject Sample	# 1	18:17	Subject Test Refused
Breath Volume	0.000	18:17	
Air Blank	0.000	18:17	
Control Test	0.080	18:18	
Air Blank	0.000	18:18	
Diagnostics Check	OK	18:18	

Data Download Date/Time: 07/09/2021 04:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH,THOMAS,E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: YOVANOVICH,THOMAS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/20/2021 09:47:16

Diagnostics Check	OK	09:40
Air Blank	0.000	09:41
Control Test	0.081	09:41
Air Blank	0.000	09:41
Subject Sample # 1	0.000	09:43
Breath Volume	1.863	09:43
Air Blank	0.000	09:43
Air Blank	0.000	09:45
Subject Sample # 2	0.000	09:45
Breath Volume	1.753	09:45
Air Blank	0.000	09:46
Control Test	0.081	09:46
Air Blank	0.000	09:47
Diagnostics Check	OK	09:47

Data Download Date/Time: 07/09/2021 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: FISCHETTI,JAMES
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/25/2021 13:50:26

Diagnostics Check	OK	13:44
Air Blank	0.000	13:44
Control Test	0.081	13:45
Air Blank	0.000	13:45
Subject Sample # 1	0.000	13:46
Breath Volume	2.816	13:46
Air Blank	0.000	13:46
Air Blank	0.000	13:48
Subject Sample # 2	0.000	13:49
Breath Volume	2.464	13:49
Air Blank	0.000	13:49
Control Test	0.081	13:49
Air Blank	0.000	13:50
Diagnostics Check	OK	13:50

Data Download Date/Time: 07/09/2021 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MORRIS, LANCE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HAMMERSLEY

DATE/TIME OF TEST: 06/27/2021 03:32:49

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.172	03:28
Breath Volume	2.019	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.166	03:31
Breath Volume	1.398	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/09/2021 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GULYA

DATE/TIME OF TEST: 06/19/2021 00:54:13

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.135	00:49
Breath Volume	2.835	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 2	0.130	00:52
Breath Volume	1.957	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 07/06/2021 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: KOSKI,BRIAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: PENTICOFF

DATE/TIME OF TEST: 06/19/2021 17:49:56

Diagnostics Check	OK	17:43
Air Blank	0.000	17:43
Control Test	0.080	17:43
Air Blank	0.000	17:44
Subject Sample # 1	0.000	17:45
Breath Volume	3.968	17:45
Air Blank	0.000	17:45
Air Blank	0.000	17:47
Subject Sample # 2	0.000	17:48
Breath Volume	3.648	17:48
Air Blank	0.000	17:48
Control Test	0.079	17:49
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

Data Download Date/Time: 07/06/2021 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/19/2021 21:56:27

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:44	
Control Test	0.081	21:44	
Air Blank	0.000	21:45	
Subject Sample	# 1 0.188	21:48	Volume Not Met
Breath Volume	0.175	21:48	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2 0.196	21:54	Volume Not Met
Breath Volume	0.019	21:54	
Air Blank	0.000	21:55	
Control Test	0.076	21:55	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 07/06/2021 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/19/2021 22:11:14

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:58	
Control Test	0.075	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1 0.204	22:00	No .020 agreement
Breath Volume	1.183	22:00	
Air Blank	0.000	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 2 0.190	22:05	Volume Not Met
Breath Volume	0.218	22:05	
Air Blank	0.000	22:07	
Air Blank	0.000	22:08	
Subject Sample	# 3 0.183	22:09	No .020 agreement
Breath Volume	2.144	22:09	
Air Blank	0.000	22:10	
Control Test	0.076	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 07/06/2021 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WAY,JOSEPH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/23/2021 16:07:28

Diagnostics Check	OK	16:00
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:01
Subject Sample # 1	0.000	16:02
Breath Volume	3.207	16:02
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.000	16:06
Breath Volume	3.511	16:06
Air Blank	0.000	16:06
Control Test	0.079	16:06
Air Blank	0.000	16:07
Diagnostics Check	OK	16:07

Data Download Date/Time: 07/06/2021 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: VICIOSO,Y
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:12

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/30/2021 13:42:21

Diagnostics Check	OK	13:35
Air Blank	0.000	13:36
Control Test	0.081	13:36
Air Blank	0.000	13:37
Subject Sample # 1	0.000	13:37
Breath Volume	1.183	13:37
Air Blank	0.000	13:38
Air Blank	0.000	13:40
Subject Sample # 2	0.000	13:40
Breath Volume	1.312	13:40
Air Blank	0.000	13:41
Control Test	0.080	13:41
Air Blank	0.000	13:42
Diagnostics Check	OK	13:42

Data Download Date/Time: 07/06/2021 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBERTS,JORDAN,L

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: WATSON,WILLIAM

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 24119080A1

DRY GAS EXPIRY DATE: 11/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/12/2021 23:38:51

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.079	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1 0.325	23:34	
Breath Volume	1.480	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 2	23:37	Slope Not Met
Breath Volume	0.832	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 06/29/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBERTS,JORDAN,L

BREATH TEST OPERATOR AGENCY: FHP Troop C

ARRESTING OFFICER: WATSON,WILLIAM

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 24119080A1

DRY GAS EXPIRY DATE: 11/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/12/2021 23:48:31

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.287	23:43
Breath Volume	1.847	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.294	23:46
Breath Volume	1.421	23:46
Air Blank	0.000	23:47
Control Test	0.077	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 06/29/2021 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,DEBBIE,L
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON,WILLIAM
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 05/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 06/17/2021 20:59:12

Diagnostics Check	OK	20:52
Air Blank	0.000	20:52
Control Test	0.081	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:54
Breath Volume	1.781	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.000	20:57
Breath Volume	2.687	20:57
Air Blank	0.000	20:58
Control Test	0.076	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 06/29/2021 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MARTIN,STERLING
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: FELTS

DATE/TIME OF TEST: 06/03/2021 00:30:23

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.218	00:25
Breath Volume	2.460	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.218	00:28
Breath Volume	2.398	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 06/30/2021 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: DE STEFANO

DATE/TIME OF TEST: 06/03/2021 10:44:46

Diagnostics Check	OK	10:38
Air Blank	0.000	10:39
Control Test	0.081	10:39
Air Blank	0.000	10:39
Subject Sample # 1	0.000	10:40
Breath Volume	1.421	10:40
Air Blank	0.000	10:40
Air Blank	0.000	10:42
Subject Sample # 2	0.000	10:43
Breath Volume	1.714	10:43
Air Blank	0.000	10:43
Control Test	0.080	10:44
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 06/30/2021 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROWLAND,KIERSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: STRUNK

DATE/TIME OF TEST: 06/03/2021 18:57:21

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.080	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.000	18:52
Breath Volume	2.726	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.000	18:55
Breath Volume	2.460	18:55
Air Blank	0.000	18:56
Control Test	0.080	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 06/30/2021 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODKIN,MICHEAL,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROWLAND,KIERSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/07/2021 03:03:50

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.000	02:59
Breath Volume	1.949	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.000	03:02
Breath Volume	2.519	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 06/30/2021 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODKIN,MICHEAL,W
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TUCKER,MARVILLE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/07/2021 03:13:26

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.063	03:08
Breath Volume	2.265	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.059	03:11
Breath Volume	2.250	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 06/30/2021 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MARTIN,STERLING
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SALVATO

DATE/TIME OF TEST: 06/08/2021 20:48:07

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.000	20:43
Breath Volume	1.414	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.000	20:46
Breath Volume	1.144	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 06/30/2021 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,MARVILLE,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: TUCKER,MARVILLE
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/10/2021 02:31:25

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.184	02:26
Breath Volume	2.105	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.188	02:29
Breath Volume	2.371	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 06/30/2021 09:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: OTERI,MALLORY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 06/13/2021 02:20:41

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.198	02:15
Breath Volume	2.015	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.205	02:19
Breath Volume	2.902	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 06/30/2021 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JAIME,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LAFASO,TERRY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/13/2021 07:50:24

Diagnostics Check	OK	07:43
Air Blank	0.000	07:44
Control Test	0.080	07:44
Air Blank	0.000	07:45
Subject Sample # 1	0.000	07:46
Breath Volume	1.941	07:46
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.000	07:49
Breath Volume	1.585	07:49
Air Blank	0.000	07:49
Control Test	0.079	07:49
Air Blank	0.000	07:50
Diagnostics Check	OK	07:50

Data Download Date/Time: 06/30/2021 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,MARVILLE,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MOFFATT,DANIEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: THOMAS IV

DATE/TIME OF TEST: 06/14/2021 05:05:17

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.135	05:00
Breath Volume	1.781	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.136	05:03
Breath Volume	4.664	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 06/30/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,R,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GLASS,A
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: RAMOS III

DATE/TIME OF TEST: 06/20/2021 14:13:26

Diagnostics Check	OK	14:08	
Air Blank	0.000	14:09	
Control Test	0.081	14:09	
Air Blank	0.000	14:09	
Subject Sample	# 1	14:12	Sequence Aborted
Breath Volume	0.019	14:12	
Air Blank	0.000	14:13	

Data Download Date/Time: 06/30/2021 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,R,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GLASS,A
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:46

SUBJECT'S LAST NAME: NELSON III

DATE/TIME OF TEST: 06/20/2021 14:22:15

Diagnostics Check	OK	14:15
Air Blank	0.000	14:16
Control Test	0.078	14:16
Air Blank	0.000	14:16
Subject Sample # 1	0.174	14:17
Breath Volume	1.890	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:19
Subject Sample # 2	0.159	14:20
Breath Volume	1.339	14:20
Air Blank	0.000	14:21
Control Test	0.080	14:21
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

Data Download Date/Time: 06/30/2021 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CAMPOVERDE,LUIS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/21/2021 16:30:38

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.081	16:24
Air Blank	0.000	16:24
Subject Sample # 1	0.000	16:26
Breath Volume	2.710	16:26
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.000	16:29
Breath Volume	2.796	16:29
Air Blank	0.000	16:29
Control Test	0.079	16:30
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 06/30/2021 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BREMER

DATE/TIME OF TEST: 06/11/2021 01:56:03

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.082	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.196	01:51
Breath Volume	1.371	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.191	01:54
Breath Volume	1.507	01:54
Air Blank	0.000	01:55
Control Test	0.084	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:56

Data Download Date/Time: 06/28/2021 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PINEAULT

DATE/TIME OF TEST: 06/12/2021 00:38:50

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.183	00:34
Breath Volume	1.593	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.183	00:37
Breath Volume	1.253	00:37
Air Blank	0.000	00:37
Control Test	0.078	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 06/28/2021 16:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TOSKEY

DATE/TIME OF TEST: 06/13/2021 03:33:10

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.088	03:32	Control Outside Tolerance
Air Blank	0.000	03:33	

Data Download Date/Time: 06/28/2021 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULKEARNS,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: SIEH,Z
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MURRETT

DATE/TIME OF TEST: 06/17/2021 01:13:19

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.091	01:12	Control Outside Tolerance
Air Blank	0.000	01:13	

Data Download Date/Time: 06/28/2021 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUEVARA,ERICK,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GUEVARA,E
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 06/07/2021 13:59:31

Diagnostics Check	OK	13:52	
Air Blank	0.000	13:52	
Control Test	0.079	13:52	
Air Blank	0.000	13:53	
Subject Sample	# 1 0.283	13:54	
Breath Volume	2.328	13:54	
Air Blank	0.000	13:55	
Air Blank	0.000	13:57	
Subject Sample	# 2 0.301	13:57	
Breath Volume	2.089	13:57	
Air Blank	0.000	13:58	
Control Test	0.074	13:58	Control Outside Tolerance
Air Blank	0.000	13:59	

Data Download Date/Time: 07/02/2021 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: PROPHETE

DATE/TIME OF TEST: 06/04/2021 06:41:17

Diagnostics Check	OK	06:28	
Air Blank	0.000	06:29	
Control Test	0.078	06:29	
Air Blank	0.000	06:29	
Subject Sample	# 1 0.169	06:33	Volume Not Met
Breath Volume	0.441	06:33	
Air Blank	0.000	06:33	
Air Blank	0.000	06:35	
Subject Sample	# 2 0.178	06:36	
Breath Volume	2.996	06:36	
Air Blank	0.000	06:36	
Air Blank	0.000	06:38	
Subject Sample	# 3 0.168	06:39	
Breath Volume	1.847	06:39	
Air Blank	0.000	06:40	
Control Test	0.077	06:40	
Air Blank	0.000	06:41	
Diagnostics Check	OK	06:41	

Data Download Date/Time: 07/02/2021 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: CABALLERO

DATE/TIME OF TEST: 06/04/2021 22:00:16

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.080	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.033	21:55
Breath Volume	1.242	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.032	21:58
Breath Volume	1.324	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 07/02/2021 09:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/05/2021 23:23:41

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.078	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:22	Subject Test Refused
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Control Test	0.079	23:23	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 07/02/2021 09:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CARRANZA-VARGAS

DATE/TIME OF TEST: 06/12/2021 02:51:42

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.135	02:47
Breath Volume	2.757	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.133	02:50
Breath Volume	2.566	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/02/2021 09:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VALCIN,JEFFREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: FRANZEN

DATE/TIME OF TEST: 06/04/2021 00:39:19

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.207	00:34
Breath Volume	3.464	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 2	0.206	00:37
Breath Volume	3.320	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOPEZ,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 06/05/2021 02:50:40

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.214	02:45
Breath Volume	3.160	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.206	02:48
Breath Volume	2.535	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WITT,KELLY,

BREATH TEST OPERATOR AGENCY: Fort Myers PD

ARRESTING OFFICER: BRADFORD,HARLEE

ARRESTING OFFICER AGENCY: Fort Myers PD

LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: ROMEO

DATE/TIME OF TEST: 06/05/2021 20:24:19

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.000	20:19
Breath Volume	4.457	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.000	20:22
Breath Volume	4.222	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WITT,KELLY,

BREATH TEST OPERATOR AGENCY: Fort Myers PD

ARRESTING OFFICER: RODRIGUEZ,LAYNOR

ARRESTING OFFICER AGENCY: Fort Myers PD

LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: DUI

ARREST TIME: 20:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/05/2021 22:07:44

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.261	22:02
Breath Volume	2.046	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.260	22:05
Breath Volume	2.339	22:05
Air Blank	0.000	22:06
Control Test	0.078	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOPEZ,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/06/2021 00:57:07

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.239	00:52
Breath Volume	2.355	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.230	00:55
Breath Volume	1.863	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODAKOSKI,SARAH,R
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HODAKOSKI,SARAH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: TODDY

DATE/TIME OF TEST: 06/06/2021 18:58:33

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.079	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.000	18:54
Breath Volume	4.949	18:54
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.000	18:57
Breath Volume	4.085	18:57
Air Blank	0.000	18:57
Control Test	0.079	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD 18096,AB,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CANFIELD 18096,AB
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: GONZALEZ FRADES

DATE/TIME OF TEST: 06/07/2021 21:21:24

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.080	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.210	21:16
Breath Volume	2.597	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.200	21:19
Breath Volume	2.246	21:19
Air Blank	0.000	21:20
Control Test	0.078	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROBEL,R,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STROBEL,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: CUFF

DATE/TIME OF TEST: 06/09/2021 10:42:57

Diagnostics Check	OK	10:36
Air Blank	0.000	10:36
Control Test	0.080	10:36
Air Blank	0.000	10:37
Subject Sample # 1	0.137	10:38
Breath Volume	1.535	10:38
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.138	10:41
Breath Volume	1.269	10:41
Air Blank	0.000	10:41
Control Test	0.080	10:42
Air Blank	0.000	10:42
Diagnostics Check	OK	10:42

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,KELLY,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SIN,PHILLIP
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 06/10/2021 19:07:56

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.225	19:03
Breath Volume	1.929	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.212	19:06
Breath Volume	2.171	19:06
Air Blank	0.000	19:06
Control Test	0.078	19:07
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WITT,KELLY,

BREATH TEST OPERATOR AGENCY: Fort Myers PD

ARRESTING OFFICER: RODRIGUEZ,LAYNOR

ARRESTING OFFICER AGENCY: Fort Myers PD

LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: UTRILLA SOLIS

DATE/TIME OF TEST: 06/11/2021 02:36:36

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.166	02:31
Breath Volume	1.367	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.173	02:34
Breath Volume	2.390	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WITT,KELLY,

BREATH TEST OPERATOR AGENCY: Fort Myers PD

ARRESTING OFFICER: BRADFORD,HARLEE

ARRESTING OFFICER AGENCY: Fort Myers PD

LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 202303

DRY GAS EXPIRY DATE: 01/05/2023

VIOLATION CODE: DUI

ARREST TIME: 22:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/11/2021 02:56:38

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:52
Breath Volume	2.710	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:55
Breath Volume	2.121	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACH,BRANDON,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SINGLETON,CORY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/13/2021 08:32:39

Diagnostics Check	OK	08:26
Air Blank	0.000	08:27
Control Test	0.081	08:27
Air Blank	0.000	08:28
Subject Sample # 1	0.000	08:28
Breath Volume	2.898	08:28
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.000	08:31
Breath Volume	2.398	08:31
Air Blank	0.000	08:31
Control Test	0.078	08:32
Air Blank	0.000	08:32
Diagnostics Check	OK	08:32

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WAY,JOSEPH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/15/2021 17:01:59

Diagnostics Check	OK	16:53
Air Blank	0.000	16:54
Control Test	0.079	16:54
Air Blank	0.000	16:54
Subject Sample # 1	0.000	16:57
Breath Volume	1.957	16:57
Air Blank	0.000	16:57
Air Blank	0.000	16:59
Subject Sample # 2	0.000	17:00
Breath Volume	4.152	17:00
Air Blank	0.000	17:01
Control Test	0.076	17:01
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 06/16/2021 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17919080A1
DRY GAS EXPIRY DATE: 08/05/2021
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME:]

DATE/TIME OF TEST: 06/04/2021 11:02:08

Diagnostics Check	OK	10:50	
Air Blank	0.000	10:50	
Control Test	0.079	10:51	
Air Blank	0.000	10:51	
Subject Sample	# 1 0.000	10:54	Volume Not Met
Breath Volume	0.402	10:54	
Air Blank	0.000	10:55	
Air Blank	0.000	10:57	
Subject Sample	# 2 0.000	10:57	
Breath Volume	1.253	10:57	
Air Blank	0.000	10:58	
Air Blank	0.000	11:00	
Subject Sample	# 3 0.000	11:00	
Breath Volume	1.886	11:00	
Air Blank	0.000	11:01	
Control Test	0.079	11:01	
Air Blank	0.000	11:02	
Diagnostics Check	OK	11:02	

Data Download Date/Time: 06/17/2021 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WICKINGS,JUSTIN,R

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: HENDERSON,EMMA

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 14:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: MCKENNA

DATE/TIME OF TEST: 06/10/2021 15:24:39

Diagnostics Check	OK	15:15
Air Blank	0.000	15:15
Control Test	0.081	15:15
Air Blank	0.000	15:16
Subject Sample # 1	0.178	15:18
Breath Volume	1.984	15:18
Air Blank	0.000	15:19
Air Blank	0.000	15:20
Subject Sample # 2	0.186	15:22
Breath Volume	1.148	15:22
Air Blank	0.000	15:23
Control Test	0.076	15:24
Air Blank	0.000	15:24
Diagnostics Check	OK	15:24

Data Download Date/Time: 06/17/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: COBY

DATE/TIME OF TEST: 06/11/2021 23:13:35

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:11	
Subject Sample	# 1 0.059	23:12	
Breath Volume	3.285	23:12	
Air Blank		23:12	Purge Fail
Air Blank		23:13	Purge Fail

Data Download Date/Time: 06/17/2021 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 22:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: COBY

DATE/TIME OF TEST: 06/11/2021 23:23:30

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.051	23:18
Breath Volume	3.296	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.051	23:21
Breath Volume	3.195	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 06/17/2021 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: JONES,KEN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 14:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: BETTARD

DATE/TIME OF TEST: 06/12/2021 16:09:08

Diagnostics Check	OK	16:02
Air Blank	0.000	16:02
Control Test	0.079	16:03
Air Blank	0.000	16:03
Subject Sample # 1	0.000	16:04
Breath Volume	1.890	16:04
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 2	0.000	16:07
Breath Volume	1.835	16:07
Air Blank	0.000	16:08
Control Test	0.079	16:08
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 06/17/2021 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 05/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 17919080A1

DRY GAS EXPIRY DATE: 08/05/2021

VIOLATION CODE: DUI

ARREST TIME: 01:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: OLAY

DATE/TIME OF TEST: 06/13/2021 02:20:04

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.215	02:15
Breath Volume	1.824	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.218	02:18
Breath Volume	2.398	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:20

Data Download Date/Time: 06/17/2021 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 06/01/2021 03:47:53

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.162	03:43
Breath Volume	1.308	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.173	03:46
Breath Volume	1.164	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 06/11/2021 17:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGI,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DADE

DATE/TIME OF TEST: 06/02/2021 01:34:09

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.000	01:29
Breath Volume	1.328	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.000	01:32
Breath Volume	2.941	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 06/11/2021 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: SOTO MENDEZ

DATE/TIME OF TEST: 06/02/2021 03:38:47

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.209	03:33
Breath Volume	1.964	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.204	03:37
Breath Volume	1.894	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRINIDAD,OSCAR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/03/2021 02:47:49

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.160	02:42
Breath Volume	2.347	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.155	02:46
Breath Volume	1.902	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SMITH,MARK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: STORMS

DATE/TIME OF TEST: 06/03/2021 14:57:32

Diagnostics Check	OK	14:51
Air Blank	0.000	14:51
Control Test	0.080	14:52
Air Blank	0.000	14:52
Subject Sample # 1	0.000	14:53
Breath Volume	3.820	14:53
Air Blank	0.000	14:53
Air Blank	0.000	14:55
Subject Sample # 2	0.000	14:56
Breath Volume	3.566	14:56
Air Blank	0.000	14:56
Control Test	0.080	14:56
Air Blank	0.000	14:57
Diagnostics Check	OK	14:57

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 06/03/2021 20:59:30

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.199	20:54
Breath Volume	1.781	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.187	20:57
Breath Volume	1.562	20:57
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: CHAN

DATE/TIME OF TEST: 06/04/2021 23:08:16

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.257	23:02
Breath Volume	1.527	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.246	23:06
Breath Volume	1.605	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TINIDAD,OSCAR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GODOY

DATE/TIME OF TEST: 06/05/2021 03:02:19

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.233	02:57
Breath Volume	1.519	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.228	03:00
Breath Volume	1.828	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WASLIN

DATE/TIME OF TEST: 06/05/2021 03:48:14

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.104	03:43
Breath Volume	3.558	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.105	03:46
Breath Volume	3.582	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: GUZMAN-COBO

DATE/TIME OF TEST: 06/05/2021 19:42:41

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.081	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.040	19:38
Breath Volume	1.242	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:40
Subject Sample # 2	0.039	19:41
Breath Volume	1.359	19:41
Air Blank	0.000	19:41
Control Test	0.080	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 06/05/2021 23:42:01

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.250	23:37
Breath Volume	2.417	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.245	23:40
Breath Volume	1.648	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:42

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ESTEBAN

DATE/TIME OF TEST: 06/06/2021 03:24:46

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.235	03:19
Breath Volume	2.222	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.233	03:23
Breath Volume	1.253	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 06/11/2021 17:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PELLEREY

DATE/TIME OF TEST: 06/06/2021 23:22:11

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.258	23:17
Breath Volume	1.320	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.269	23:20
Breath Volume	1.164	23:20
Air Blank	0.000	23:21
Control Test	0.077	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 06/07/2021 01:41:56

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.220	01:37
Breath Volume	1.570	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.230	01:40
Breath Volume	1.976	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MITCHELL,COLE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: TURCO

DATE/TIME OF TEST: 06/07/2021 03:56:50

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.204	03:52
Breath Volume	2.218	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.187	03:55
Breath Volume	1.539	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS,OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: BARTEE

DATE/TIME OF TEST: 06/08/2021 00:13:46

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample # 1	0.126	00:09	
Breath Volume	1.804	00:09	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample # 2	0.129	00:11	
Breath Volume	1.921	00:11	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	Interferent Detect
Air Blank		00:13	Purge Fail

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS,OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BARTEE

DATE/TIME OF TEST: 06/08/2021 00:45:11

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Interferent Detect
Breath Volume	1.812	00:43	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	Interferent Detect
Air Blank	0.000	00:45	

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS,OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BARTEE

DATE/TIME OF TEST: 06/08/2021 01:53:06

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.076	01:48
Breath Volume	1.753	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.073	01:51
Breath Volume	1.675	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MITCHELL,COLE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/08/2021 08:03:52

Diagnostics Check	OK	07:55
Air Blank	0.000	07:56
Control Test	0.080	07:56
Air Blank	0.000	07:57
Subject Sample # 1	0.199	07:58
Breath Volume	1.625	07:58
Air Blank	0.000	07:59
Air Blank	0.000	08:00
Subject Sample # 2	0.196	08:02
Breath Volume	1.453	08:02
Air Blank	0.000	08:02
Control Test	0.080	08:03
Air Blank	0.000	08:03
Diagnostics Check	OK	08:03

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHARELUS,OBED
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 06/09/2021 00:27:11

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.097	00:21
Breath Volume	1.445	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.096	00:25
Breath Volume	1.894	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MUSSO CALDERON

DATE/TIME OF TEST: 06/09/2021 02:55:49

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.080	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.221	02:47	Volume Not Met
Breath Volume	0.214	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2 0.226	02:50	
Breath Volume	1.964	02:50	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 3 0.221	02:54	
Breath Volume	1.367	02:54	
Air Blank	0.000	02:54	
Control Test	0.078	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LAURINCE

DATE/TIME OF TEST: 06/10/2021 03:13:39

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.108	03:08
Breath Volume	2.226	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.099	03:11
Breath Volume	1.148	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CALVILLO

DATE/TIME OF TEST: 06/11/2021 00:42:04

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.138	00:36
Breath Volume	1.308	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.129	00:40
Breath Volume	1.113	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:42

Data Download Date/Time: 06/11/2021 17:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA, GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA, GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 14:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:51

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/05/2021 15:42:38

Diagnostics Check	OK	15:36
Air Blank	0.000	15:36
Control Test	0.080	15:36
Air Blank	0.000	15:37
Subject Sample # 1	0.055	15:38
Breath Volume	1.468	15:38
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.049	15:41
Breath Volume	1.230	15:41
Air Blank	0.000	15:41
Control Test	0.080	15:41
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 06/23/2021 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,S
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CATALANO,BRANDON
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 06/05/2021 22:43:44

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.080	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:36	No Sample Provided
Breath Volume	0.000	22:36	
Air Blank	0.000	22:37	
Air Blank	0.000	22:39	
Subject Sample	# 2	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:42	
Control Test	0.081	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 06/23/2021 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA,GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA,GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: DIERESTI

DATE/TIME OF TEST: 06/06/2021 15:54:42

Diagnostics Check	OK	15:53	
Air Blank	0.000	15:53	
Control Test	0.065	15:54	Control Outside Tolerance
Air Blank	0.000	15:54	

Data Download Date/Time: 06/23/2021 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTAYA, GUILLERMO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARTAYA, GUILLERMO
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: DIERESTIL

DATE/TIME OF TEST: 06/06/2021 16:04:04

Diagnostics Check	OK	15:57
Air Blank	0.000	15:57
Control Test	0.077	15:58
Air Blank	0.000	15:58
Subject Sample # 1	0.096	15:59
Breath Volume	1.425	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:01
Subject Sample # 2	0.091	16:02
Breath Volume	1.558	16:02
Air Blank	0.000	16:03
Control Test	0.076	16:03
Air Blank	0.000	16:03
Diagnostics Check	OK	16:04

Data Download Date/Time: 06/23/2021 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDGEPEETH,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: WHITT,JESSE
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/20/2021 23:06:22

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.081	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1 0.098	22:58	Volume Not Met
Breath Volume	0.070	22:58	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2 0.085	23:04	Volume Not Met
Breath Volume	0.449	23:04	
Air Blank	0.000	23:05	
Control Test	0.077	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 06/23/2021 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDGEPEETH,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: WHITT,JESSE
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: BUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/20/2021 23:22:23

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.079	23:10	
Air Blank	0.000	23:11	
Subject Sample	# 1 0.084	23:14	Volume Not Met
Breath Volume	0.132	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 2 0.079	23:17	
Breath Volume	2.152	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 3 0.085	23:20	
Breath Volume	1.648	23:20	
Air Blank	0.000	23:21	
Control Test	0.076	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 06/23/2021 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 06/19/2021 02:51:27

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.208	02:46
Breath Volume	2.328	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.205	02:49
Breath Volume	2.218	02:49
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 07/23/2021 04:28