Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ASH

DATE/TIME OF TEST: 06/05/2021 01:58:10

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.082	01:46	
Air Blank		0.000	01:46	
Subject Sample	# 1	0.107	01:50	Volume Not Met
Breath Volume		0.207	01:50	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 2	0.116	01:53	
Breath Volume		2.046	01:53	
Air Blank		0.000	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 3	0.116	01:56	
Breath Volume		2.039	01:56	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 06/14/2021 11:12

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 06/15/2021 01:51:50

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.086	01:51	Interferent Detect
Air Blank	0.000	01:51	

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 06/15/2021 02:00:15

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.148	01:55
Breath Volume		2.109	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:57
Subject Sample	# 2	0.138	01:58
Breath Volume		1.667	01:58
Air Blank		0.000	01:59
Control Test		0.077	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MORALES MIGUEL

DATE/TIME OF TEST: 06/19/2021 03:21:44

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.082	03:21	Interferent Detect
Air Blank	0.000	03:21	

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MORALES MIGUEL

DATE/TIME OF TEST: 06/19/2021 03:32:46

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.077	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.141	03:27
Breath Volume		1.410	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.144	03:31
Breath Volume		1.695	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: VERDUGO ROBLERO

DATE/TIME OF TEST: 06/20/2021 02:48:28

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.079	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.163	02:38	
Breath Volume		1.582	02:38	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.146	02:43	Volume Not Met
Breath Volume		0.949	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 3	0.159	02:46	
Breath Volume		1.289	02:46	
Air Blank		0.000	02:47	
Control Test		0.078	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: QUAGLIERI

DATE/TIME OF TEST: 06/20/2021 23:18:11

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1	0.007	23:13	
Breath Volume		2.644	23:13	
Air Blank		0.000	23:13	
Air Blank		0.000	23:15	
Subject Sample	# 2		23:16	Interferent Detect
Breath Volume		3.328	23:16	
Air Blank		0.000	23:17	
Control Test		0.085	23:17	Interferent Detect
Air Blank			23:18	Purge Fail

Data Download Date/Time: 07/12/2021 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: QUAGLIERI

DATE/TIME OF TEST: 06/20/2021 23:21:46

Diagnostics Check	OK	23:20	
Air Blank		23:20	Ambient Fail
Air Blank	0.000	23:21	

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: LARUE

DATE/TIME OF TEST: 06/25/2021 00:00:45

Diagnostics Check	OK	23:59	
Air Blank		00:00	Ambient Fail
Air Blank	0.000	00:00	

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: LARUE

DATE/TIME OF TEST: 06/25/2021 00:08:50

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.007	00:04
Breath Volume		3.085	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.000	00:07
Breath Volume		2.761	00:07
Air Blank		0.000	00:07
Control Test		0.079	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 07/12/2021 09:14

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: GARZON FORERO

DATE/TIME OF TEST: 06/04/2021 17:21:59

Diagnostics Check		OK	17:13
Air Blank		0.000	17:13
Control Test		0.079	17:14
Air Blank		0.000	17:14
Subject Sample	# 1	0.280	17:17
Breath Volume		1.371	17:17
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.287	17:20
Breath Volume		1.476	17:20
Air Blank		0.000	17:21
Control Test		0.078	17:21
Air Blank		0.000	17:21
Diagnostics Check		OK	17:21

Data Download Date/Time: 06/10/2021 09:19

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: MESSANO

DATE/TIME OF TEST: 06/04/2021 23:22:37

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.080	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.227	23:18
Breath Volume		2.480	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.214	23:20
Breath Volume		2.535	23:20
Air Blank		0.000	23:21
Control Test		0.079	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 06/10/2021 09:19

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BERDUO PEREZ

DATE/TIME OF TEST: 06/05/2021 02:03:26

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:50	
Control Test		0.080	01:50	
Air Blank		0.000	01:51	
Subject Sample	# 1	0.238	01:54	Volume Not Met
Breath Volume		0.027	01:54	
Air Blank		0.000	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 2	0.250	01:58	
Breath Volume		2.191	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 3	0.255	02:01	
Breath Volume		1.453	02:01	
Air Blank		0.000	02:02	
Control Test		0.075	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 06/10/2021 09:19

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: MCDONALD JR

DATE/TIME OF TEST: 06/02/2021 23:27:27

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.081	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.180	23:22
Breath Volume		1.542	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:24
Subject Sample	# 2	0.197	23:25
Breath Volume		2.082	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 07/15/2021 15:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: DAUPHIN

DATE/TIME OF TEST: 06/03/2021 04:12:40

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.074	04:08
Breath Volume		3.148	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.074	04:11
Breath Volume		3.535	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 07/15/2021 15:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CARSWELL

DATE/TIME OF TEST: 06/04/2021 00:33:18

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.000	00:28
Breath Volume		1.910	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.000	00:31
Breath Volume		1.660	00:31
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/15/2021 15:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MODUGNO

DATE/TIME OF TEST: 06/04/2021 01:03:35

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.000	00:59
Breath Volume		1.101	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.000	01:02
Breath Volume		2.066	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/15/2021 15:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: WUSSOW

DATE/TIME OF TEST: 06/04/2021 20:53:46

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.080	20:47
Air Blank		0.000	20:47
Subject Sample	# 1	0.151	20:48
Breath Volume		1.207	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:50
Subject Sample	# 2	0.158	20:52
Breath Volume		1.308	20:52
Air Blank		0.000	20:52
Control Test		0.080	20:53
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 07/15/2021 15:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOPPE,Z
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: BUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: LANESE

DATE/TIME OF TEST: 06/04/2021 22:55:14

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.080	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.184	22:48
Breath Volume		2.347	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:50
Subject Sample	# 2	0.179	22:53
Breath Volume		2.160	22:53
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 07/15/2021 15:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: GRIMSLEY BURDOIN

DATE/TIME OF TEST: 06/05/2021 00:09:53

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.081	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.200	00:04
Breath Volume		2.335	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.195	00:08
Breath Volume		2.335	00:08
Air Blank		0.000	00:08
Control Test		0.079	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 07/15/2021 15:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BALTAZAR TIERRABLANC

DATE/TIME OF TEST: 06/05/2021 01:52:55

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:51	
Control Test		0.081	01:52	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 07/15/2021 15:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SPILLANE

DATE/TIME OF TEST: 06/05/2021 02:51:03

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.025	02:46
Breath Volume		1.300	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.029	02:49
Breath Volume		2.261	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:51

Data Download Date/Time: 07/15/2021 15:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONCION,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: ODUM

DATE/TIME OF TEST: 06/05/2021 17:37:20

Diagnostics Check		OK	17:25
Air Blank		0.000	17:26
Control Test		0.080	17:26
Air Blank		0.000	17:27
Subject Sample	# 1	0.071	17:30
Breath Volume		1.527	17:30
Air Blank		0.000	17:31
Air Blank		0.000	17:32
Subject Sample	# 2	0.070	17:35
Breath Volume		1.695	17:35
Air Blank		0.000	17:36
Control Test		0.079	17:36
Air Blank		0.000	17:37
Diagnostics Check		OK	17:37

Data Download Date/Time: 07/15/2021 15:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: DIAZ TORRES

DATE/TIME OF TEST: 06/05/2021 17:53:35

Diagnostics Check		OK	17:46
Air Blank		0.000	17:46
Control Test		0.079	17:46
Air Blank		0.000	17:47
Subject Sample	# 1	0.036	17:48
Breath Volume		1.714	17:48
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.031	17:52
Breath Volume		1.160	17:52
Air Blank		0.000	17:52
Control Test		0.080	17:52
Air Blank		0.000	17:53
Diagnostics Check		OK	17:53

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SAILOR

DATE/TIME OF TEST: 06/06/2021 00:46:36

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.138	00:42
Breath Volume		1.664	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.132	00:45
Breath Volume		1.375	00:45
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BOWLS

DATE/TIME OF TEST: 06/06/2021 01:44:31

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.190	01:39
Breath Volume		3.113	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.191	01:42
Breath Volume		3.402	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SCANNELL

DATE/TIME OF TEST: 06/06/2021 03:13:58

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.106	03:09
Breath Volume		2.683	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:11
Subject Sample	# 2	0.105	03:12
Breath Volume		3.289	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: JODAR VELAZQUEZ

DATE/TIME OF TEST: 06/06/2021 04:33:48

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.231	04:29
Breath Volume		2.832	04:29
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.232	04:32
Breath Volume		2.710	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BERG,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: BUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: FARAJ

DATE/TIME OF TEST: 06/06/2021 16:58:28

Diagnostics Check		OK	16:51
Air Blank		0.000	16:52
Control Test		0.081	16:52
Air Blank		0.000	16:53
Subject Sample	# 1	0.031	16:54
Breath Volume		2.289	16:54
Air Blank		0.000	16:54
Air Blank		0.000	16:56
Subject Sample	# 2	0.030	16:56
Breath Volume		2.441	16:56
Air Blank		0.000	16:57
Control Test		0.081	16:57
Air Blank		0.000	16:58
Diagnostics Check		OK	16:58

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: TOLER

DATE/TIME OF TEST: 06/06/2021 18:54:46

Diagnostics Check		OK	18:48
Air Blank		0.000	18:48
Control Test		0.080	18:49
Air Blank		0.000	18:49
Subject Sample	# 1	0.030	18:50
Breath Volume		1.289	18:50
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 2	0.029	18:53
Breath Volume		1.128	18:53
Air Blank		0.000	18:53
Control Test		0.081	18:54
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REID,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 06/07/2021 01:05:00

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.162	01:00
Breath Volume		2.070	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.155	01:03
Breath Volume		1.546	01:03
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 06/07/2021 02:54:38

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.137	02:49
Breath Volume		2.871	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.140	02:52
Breath Volume		4.628	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: MCGRIFF

DATE/TIME OF TEST: 06/07/2021 21:39:07

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:26	
Control Test		0.081	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 1	0.088	21:30	Volume Not Met
Breath Volume		0.531	21:30	
Air Blank		0.000	21:31	
Air Blank		0.000	21:33	
Subject Sample	# 2	0.099	21:34	
Breath Volume		2.042	21:34	
Air Blank		0.000	21:34	
Air Blank		0.000	21:36	
Subject Sample	# 3	0.099	21:37	
Breath Volume		2.027	21:37	
Air Blank		0.000	21:38	
Control Test		0.080	21:38	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 07/15/2021 15:44

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DOBOSIEWIUZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: OBERMAN

DATE/TIME OF TEST: 06/08/2021 05:53:15

Diagnostics Check		OK	05:44	
Air Blank		0.000	05:45	
Control Test		0.081	05:45	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:49	No Sample Provided
Breath Volume		0.000	05:49	
Air Blank		0.000	05:49	
Air Blank		0.000	05:51	
Subject Sample	# 2		05:51	Subject Test Refused
Breath Volume		0.000	05:51	
Air Blank		0.000	05:52	
Control Test		0.081	05:52	
Air Blank		0.000	05:53	
Diagnostics Check		OK	05:53	

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: PALU

DATE/TIME OF TEST: 06/10/2021 00:28:53

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.081	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.143	00:24
Breath Volume		2.691	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.137	00:27
Breath Volume		2.480	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 06/10/2021 23:52:41

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.145	23:48
Breath Volume		2.148	23:48
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.148	23:51
Breath Volume		2.113	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CRUZ RAMOS

DATE/TIME OF TEST: 06/11/2021 00:34:51

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.081	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.000	00:30
Breath Volume		1.414	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.000	00:33
Breath Volume		1.843	00:33
Air Blank		0.000	00:33
Control Test		0.081	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: LADIKA

DATE/TIME OF TEST: 06/12/2021 00:42:45

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.081	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.145	00:38
Breath Volume		2.742	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.139	00:41
Breath Volume		2.308	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/12/2021 00:57:54

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.150	00:53
Breath Volume		3.050	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.142	00:56
Breath Volume		1.816	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MCGETTRICK

DATE/TIME OF TEST: 06/12/2021 02:13:33

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:10	
Control Test		0.080	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:12	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MORALES MAZARIEGOS

DATE/TIME OF TEST: 06/12/2021 03:07:31

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.081	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.210	03:02
Breath Volume		1.113	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.222	03:05
Breath Volume		2.593	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: ANTONIO DOMINGUEZ

DATE/TIME OF TEST: 06/12/2021 21:10:22

Diagnostics Check		OK	20:58	
Air Blank		0.000	20:58	
Control Test		0.081	20:58	
Air Blank		0.000	20:59	
Subject Sample	# 1	0.096	21:02	Volume Not Met
Breath Volume		0.156	21:02	
Air Blank		0.000	21:03	
Air Blank		0.000	21:05	
Subject Sample	# 2	0.136	21:05	
Breath Volume		3.148	21:05	
Air Blank		0.000	21:06	
Air Blank		0.000	21:08	
Subject Sample	# 3	0.117	21:08	
Breath Volume		1.621	21:08	
Air Blank		0.000	21:09	
Control Test		0.080	21:09	
Air Blank		0.000	21:10	
Diagnostics Check		OK	21:10	

Data Download Date/Time: 07/15/2021 15:45

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MAYS

DATE/TIME OF TEST: 06/12/2021 21:43:28

Diagnostics Check		OK	21:36
Air Blank		0.000	21:36
Control Test		0.080	21:36
Air Blank		0.000	21:37
Subject Sample	# 1	0.205	21:38
Breath Volume		3.671	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.220	21:41
Breath Volume		3.062	21:41
Air Blank		0.000	21:42
Control Test		0.080	21:42
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: ROWSE

DATE/TIME OF TEST: 06/13/2021 19:31:18

Diagnostics Check		OK	19:24
Air Blank		0.000	19:24
Control Test		0.081	19:24
Air Blank		0.000	19:25
Subject Sample	# 1	0.130	19:26
Breath Volume		2.804	19:26
Air Blank		0.000	19:27
Air Blank		0.000	19:28
Subject Sample	# 2	0.127	19:29
Breath Volume		1.394	19:29
Air Blank		0.000	19:30
Control Test		0.081	19:30
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: URIAS DE PAZ

DATE/TIME OF TEST: 06/14/2021 05:04:09

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.082	04:59
Breath Volume		1.582	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.092	05:02
Breath Volume		2.445	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/15/2021 04:00:05

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.174	03:55
Breath Volume		1.457	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.179	03:58
Breath Volume		2.734	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EVERETT,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: RAHAW

DATE/TIME OF TEST: 06/15/2021 19:00:19

Diagnostics Check		OK	18:53
Air Blank		0.000	18:54
Control Test		0.081	18:54
Air Blank		0.000	18:54
Subject Sample	# 1	0.217	18:55
Breath Volume		2.214	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:58
Subject Sample	# 2	0.218	18:58
Breath Volume		2.082	18:58
Air Blank		0.000	18:59
Control Test		0.080	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DEROCHE

DATE/TIME OF TEST: 06/16/2021 01:57:10

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.227	01:52
Breath Volume		1.730	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:54
Subject Sample	# 2	0.213	01:55
Breath Volume		1.460	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: KALWARY

DATE/TIME OF TEST: 06/16/2021 02:32:52

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.081	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1	0.240	02:22	
Breath Volume		2.160	02:22	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 2		02:27	No Sample Provided
Breath Volume		0.000	02:27	
Air Blank		0.000	02:28	
Air Blank		0.000	02:30	
Subject Sample	# 3	0.235	02:31	
Breath Volume		2.390	02:31	
Air Blank		0.000	02:31	
Control Test		0.079	02:32	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEWART,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:39

SUBJECT'S LAST NAME: STUMP

DATE/TIME OF TEST: 06/16/2021 14:10:18

Diagnostics Check		OK	14:00
Air Blank		0.000	14:01
Control Test		0.081	14:01
Air Blank		0.000	14:02
Subject Sample	# 1	0.211	14:04
Breath Volume		1.472	14:04
Air Blank		0.000	14:05
Air Blank		0.000	14:07
Subject Sample	# 2	0.211	14:08
Breath Volume		1.394	14:08
Air Blank		0.000	14:09
Control Test		0.077	14:09
Air Blank		0.000	14:10
Diagnostics Check		OK	14:10

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 06/18/2021 02:42:57

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.221	02:38
Breath Volume		2.539	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.225	02:41
Breath Volume		1.445	02:41
Air Blank		0.000	02:42
Control Test		0.078	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 06/18/2021 06:30:25

Diagnostics Check		OK	06:23
Air Blank		0.000	06:23
Control Test		0.081	06:24
Air Blank		0.000	06:24
Subject Sample	# 1	0.122	06:25
Breath Volume		2.292	06:25
Air Blank		0.000	06:26
Air Blank		0.000	06:28
Subject Sample	# 2	0.116	06:28
Breath Volume		1.156	06:28
Air Blank		0.000	06:29
Control Test		0.080	06:29
Air Blank		0.000	06:30
Diagnostics Check		OK	06:30

Data Download Date/Time: 07/15/2021 15:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 06/18/2021 21:12:16

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.081	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.174	21:07
Breath Volume		3.210	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.176	21:10
Breath Volume		3.433	21:10
Air Blank		0.000	21:11
Control Test		0.079	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: VAZQUEZ-DE JESUS

DATE/TIME OF TEST: 06/19/2021 02:49:04

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.081	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.197	02:43
Breath Volume		1.910	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.183	02:47
Breath Volume		1.500	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:49

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: VEGA-COLON

DATE/TIME OF TEST: 06/19/2021 04:30:19

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.080	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.081	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WINTER,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ALGARIN NIEVES

DATE/TIME OF TEST: 06/20/2021 00:39:14

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.081	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.155	00:34
Breath Volume		1.550	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.151	00:37
Breath Volume		1.363	00:37
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: COTES CEDENO

DATE/TIME OF TEST: 06/20/2021 02:23:32

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.270	02:17
Breath Volume		1.460	02:17
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.265	02:21
Breath Volume		1.363	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RODRIGUEZ RAMOS

DATE/TIME OF TEST: 06/20/2021 03:00:42

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.000	02:56
Breath Volume		1.988	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.000	02:59
Breath Volume		2.531	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURANCEAU,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: SANTOS RUIZ

DATE/TIME OF TEST: 06/20/2021 05:06:53

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.081	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.191	05:01
Breath Volume		1.761	05:01
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.197	05:05
Breath Volume		2.742	05:05
Air Blank		0.000	05:05
Control Test		0.081	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDREWS,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: VAZQUEZ SOLIS

DATE/TIME OF TEST: 06/20/2021 19:41:00

Diagnostics Check		OK	19:34
Air Blank		0.000	19:35
Control Test		0.081	19:35
Air Blank		0.000	19:35
Subject Sample	# 1	0.175	19:36
Breath Volume		2.425	19:36
Air Blank		0.000	19:37
Air Blank		0.000	19:38
Subject Sample	# 2	0.164	19:39
Breath Volume		2.601	19:39
Air Blank		0.000	19:40
Control Test		0.080	19:40
Air Blank		0.000	19:40
Diagnostics Check		OK	19:41

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/21/2021 00:19:12

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.073	00:13
Breath Volume		1.417	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.068	00:17
Breath Volume		1.367	00:17
Air Blank		0.000	00:18
Control Test		0.078	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MOLINA CORREA

DATE/TIME OF TEST: 06/21/2021 00:57:43

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.150	00:51
Breath Volume		1.468	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.147	00:56
Breath Volume		1.164	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:54

SUBJECT'S LAST NAME: MUNOZ GONZALEZ

DATE/TIME OF TEST: 06/21/2021 09:24:36

Diagnostics Check		OK	09:15
Air Blank		0.000	09:16
Control Test		0.080	09:16
Air Blank		0.000	09:17
Subject Sample	# 1	0.164	09:19
Breath Volume		2.738	09:19
Air Blank		0.000	09:20
Air Blank		0.000	09:22
Subject Sample	# 2	0.148	09:22
Breath Volume		1.945	09:22
Air Blank		0.000	09:23
Control Test		0.078	09:23
Air Blank		0.000	09:24
Diagnostics Check		OK	09:24

Data Download Date/Time: 07/15/2021 15:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: PARTON II

DATE/TIME OF TEST: 06/21/2021 21:30:57

Diagnostics Check		OK	21:19	
Air Blank		0.000	21:19	
Control Test		0.080	21:20	
Air Blank		0.000	21:20	
Subject Sample	# 1		21:23	No Sample Provided
Breath Volume		0.000	21:23	
Air Blank		0.000	21:24	
Air Blank		0.000	21:26	
Subject Sample	# 2		21:29	No Sample Provided
Breath Volume		0.000	21:29	
Air Blank		0.000	21:30	
Control Test		0.081	21:30	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 07/15/2021 15:48

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SNYDERS JR

DATE/TIME OF TEST: 06/22/2021 00:51:41

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.191	00:47
Breath Volume		1.191	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.192	00:50
Breath Volume		1.246	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 07/15/2021 15:48

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: FINDLEY

DATE/TIME OF TEST: 06/22/2021 03:06:32

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.081	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.155	03:01
Breath Volume		1.289	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.158	03:04
Breath Volume		1.238	03:04
Air Blank		0.000	03:05
Control Test		0.081	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 07/15/2021 15:48

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/22/2021 22:09:58

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.081	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.000	22:05
Breath Volume		3.359	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.000	22:08
Breath Volume		3.503	22:08
Air Blank		0.000	22:09
Control Test		0.081	22:09
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 07/15/2021 15:49

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 06/23/2021 03:01:58

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.000	02:57
Breath Volume		2.476	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.000	03:00
Breath Volume		1.390	03:00
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 07/15/2021 15:49

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: DELANGEL-CRUZ

DATE/TIME OF TEST: 06/23/2021 06:14:33

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.081	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.115	06:08
Breath Volume		1.917	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.114	06:12
Breath Volume		1.574	06:12
Air Blank		0.000	06:13
Control Test		0.080	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 07/15/2021 15:57

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAIRSTON,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 06/24/2021 02:35:55

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.165	02:31
Breath Volume		2.222	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.170	02:34
Breath Volume		2.593	02:34
Air Blank		0.000	02:34
Control Test		0.081	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 07/15/2021 15:57

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: FADIGAN JR

DATE/TIME OF TEST: 06/24/2021 20:39:59

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.081	20:34
Air Blank		0.000	20:34
Subject Sample	# 1	0.328	20:35
Breath Volume		2.613	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:37
Subject Sample	# 2	0.315	20:38
Breath Volume		1.687	20:38
Air Blank		0.000	20:39
Control Test		0.079	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 06/25/2021 19:21:27

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.081	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.206	19:16
Breath Volume		1.988	19:16
Air Blank		0.000	19:17
Air Blank		0.000	19:18
Subject Sample	# 2	0.211	19:19
Breath Volume		1.390	19:19
Air Blank		0.000	19:20
Control Test		0.080	19:20
Air Blank		0.000	19:21
Diagnostics Check		OK	19:21

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DAMEWOOD

DATE/TIME OF TEST: 06/26/2021 03:47:29

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.081	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.148	03:40	Volume Not Met
Breath Volume		0.859	03:40	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.160	03:45	Volume Not Met
Breath Volume		0.203	03:45	
Air Blank		0.000	03:46	
Control Test		0.079	03:46	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CURBELO,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/26/2021 04:19:01

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.145	04:14
Breath Volume		1.894	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:16
Subject Sample	# 2	0.145	04:17
Breath Volume		1.949	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:19

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 06/26/2021 23:14:42

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:03	
Control Test		0.080	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:07	No Sample Provided
Breath Volume		0.000	23:07	
Air Blank		0.000	23:08	
Air Blank		0.000	23:10	
Subject Sample	# 2		23:13	No Sample Provided
Breath Volume		0.000	23:13	
Air Blank		0.000	23:13	
Control Test		0.081	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:14	

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: WILLIAMES JR

DATE/TIME OF TEST: 06/27/2021 00:27:32

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.081	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.281	00:19
Breath Volume		2.664	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.258	00:22
Breath Volume		1.769	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 3	0.267	00:25
Breath Volume		3.070	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SALES

DATE/TIME OF TEST: 06/27/2021 00:48:23

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.206	00:39
Breath Volume		1.449	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:41
Subject Sample	# 2	0.184	00:42
Breath Volume		1.429	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 3	0.208	00:46
Breath Volume		2.929	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/27/2021 05:07:53

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.080	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.197	05:03
Breath Volume		2.621	05:03
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.195	05:06
Breath Volume		1.617	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:07
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: HETRICK

DATE/TIME OF TEST: 06/27/2021 05:31:16

Diagnostics Check		OK	05:24
Air Blank		0.000	05:25
Control Test		0.081	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.143	05:26
Breath Volume		1.324	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.147	05:29
Breath Volume		1.507	05:29
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: TOJUOLA

DATE/TIME OF TEST: 06/27/2021 05:56:43

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.081	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.126	05:51
Breath Volume		1.144	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.119	05:55
Breath Volume		1.304	05:55
Air Blank		0.000	05:55
Control Test		0.080	05:56
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/27/2021 06:13:49

Diagnostics Check		OK	06:06
Air Blank		0.000	06:06
Control Test		0.080	06:06
Air Blank		0.000	06:07
Subject Sample	# 1	0.192	06:08
Breath Volume		1.207	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.201	06:12
Breath Volume		1.726	06:12
Air Blank		0.000	06:12
Control Test		0.080	06:13
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 06/28/2021 00:41:00

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.081	00:30	
Air Blank		0.000	00:30	
Subject Sample	# 1	0.194	00:32	
Breath Volume		1.492	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 2	0.177	00:37	Volume Not Met
Breath Volume		0.945	00:37	
Air Blank		0.000	00:38	
Air Blank		0.000	00:40	
Subject Sample	# 3		00:40	Sequence Aborted
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 06/28/2021 00:57:23

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.195	00:52
Breath Volume		1.460	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.200	00:55
Breath Volume		1.335	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/15/2021 15:58

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LUGO,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 06/29/2021 01:04:23

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.081	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1	0.019	00:57	Volume Not Met
Breath Volume		0.246	00:57	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2	0.019	01:02	Volume Not Met
Breath Volume		0.230	01:02	
Air Blank		0.000	01:03	
Control Test		0.081	01:03	
Air Blank		0.000	01:04	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 07/15/2021 15:59

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: VIRES

DATE/TIME OF TEST: 06/29/2021 05:27:58

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.082	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.305	05:22
Breath Volume		1.816	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.303	05:26
Breath Volume		2.410	05:26
Air Blank		0.000	05:27
Control Test		0.080	05:27
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 07/15/2021 15:59

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 06/29/2021 20:13:50

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.152	20:08
Breath Volume		1.644	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.155	20:12
Breath Volume		2.152	20:12
Air Blank		0.000	20:12
Control Test		0.080	20:13
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 07/15/2021 15:59

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GRIDER

DATE/TIME OF TEST: 06/30/2021 00:34:46

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:23	
Control Test		0.081	00:23	
Air Blank		0.000	00:24	
Subject Sample	# 1		00:27	No Sample Provided
Breath Volume		0.000	00:27	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2	0.000	00:30	
Breath Volume		3.601	00:30	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 3	0.000	00:33	
Breath Volume		3.015	00:33	
Air Blank		0.000	00:33	
Control Test		0.081	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 07/15/2021 15:59

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ZAPATA,A
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: H325546775280

DATE/TIME OF TEST: 06/01/2021 17:08:10

Diagnostics Check		OK	17:05	
Air Blank		0.000	17:05	
Control Test		0.081	17:06	
Air Blank		0.000	17:06	
Subject Sample	# 1		17:06	Subject Test Refused
Breath Volume		0.000	17:06	
Air Blank		0.000	17:07	
Control Test		0.080	17:07	
Air Blank		0.000	17:08	
Diagnostics Check		OK	17:08	

Data Download Date/Time: 06/01/2021 22:08

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: ROGATENKO

DATE/TIME OF TEST: 06/04/2021 02:55:58

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.097	02:51
Breath Volume		3.089	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.102	02:54
Breath Volume		2.792	02:54
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/29/2021 12:05

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WEAVER,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: MCVEY

DATE/TIME OF TEST: 06/05/2021 17:19:17

Diagnostics Check		OK	17:12
Air Blank		0.000	17:12
Control Test		0.080	17:13
Air Blank		0.000	17:13
Subject Sample	# 1	0.316	17:14
Breath Volume		1.371	17:14
Air Blank		0.000	17:15
Air Blank		0.000	17:17
Subject Sample	# 2	0.330	17:17
Breath Volume		1.148	17:17
Air Blank		0.000	17:18
Control Test		0.079	17:18
Air Blank		0.000	17:19
Diagnostics Check		OK	17:19

Data Download Date/Time: 06/29/2021 12:05

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SLAGA,JASMINE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 06/05/2021 21:01:58

Diagnostics Check		OK	20:54
Air Blank		0.000	20:54
Control Test		0.080	20:55
Air Blank		0.000	20:55
Subject Sample	# 1	0.215	20:57
Breath Volume		1.375	20:57
Air Blank		0.000	20:58
Air Blank		0.000	20:59
Subject Sample	# 2	0.225	21:00
Breath Volume		1.511	21:00
Air Blank		0.000	21:01
Control Test		0.078	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 06/29/2021 12:05

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCASTRO,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CUTHBERT,ANTHONY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 06/06/2021 08:39:40

Diagnostics Check		OK	08:32
Air Blank		0.000	08:33
Control Test		0.080	08:33
Air Blank		0.000	08:34
Subject Sample	# 1	0.200	08:34
Breath Volume		2.937	08:34
Air Blank		0.000	08:35
Air Blank		0.000	08:37
Subject Sample	# 2	0.197	08:37
Breath Volume		2.933	08:37
Air Blank		0.000	08:38
Control Test		0.079	08:39
Air Blank		0.000	08:39
Diagnostics Check		OK	08:39

Data Download Date/Time: 06/29/2021 12:05

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BRYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CUTHBERT,ANTHONY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:56

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 06/06/2021 09:23:54

Diagnostics Check		OK	09:16
Air Blank		0.000	09:17
Control Test		0.080	09:17
Air Blank		0.000	09:18
Subject Sample	# 1	0.197	09:19
Breath Volume		2.660	09:19
Air Blank		0.000	09:19
Air Blank		0.000	09:21
Subject Sample	# 2	0.190	09:22
Breath Volume		2.953	09:22
Air Blank		0.000	09:22
Control Test		0.079	09:23
Air Blank		0.000	09:23
Diagnostics Check		OK	09:23

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LANE,M
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MARTIR VEGA

DATE/TIME OF TEST: 06/08/2021 04:20:09

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.022	04:13
Breath Volume		2.121	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.021	04:18
Breath Volume		1.179	04:18
Air Blank		0.000	04:19
Control Test		0.082	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGMAN,D/S,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BERGMAN,D/S
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/12/2021 01:05:30

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.080	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:04	
Control Test		0.079	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BURRELL

DATE/TIME OF TEST: 06/13/2021 02:14:39

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.080	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:13	Sequence Aborted
Breath Volume		0.191	02:13	
Air Blank		0.000	02:14	

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BURRELL

DATE/TIME OF TEST: 06/13/2021 02:22:44

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.077	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.200	02:18	
Breath Volume		1.312	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 2	0.199	02:21	
Breath Volume		2.730	02:21	
Air Blank		0.000	02:21	
Control Test		0.071	02:22	Control Outside Tolerance
Air Blank		0.000	02:22	

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MAGANA

DATE/TIME OF TEST: 06/14/2021 00:01:48

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.086	23:57
Breath Volume		2.945	23:57
Air Blank		0.000	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.083	00:00
Breath Volume		2.957	00:00
Air Blank		0.000	00:00
Control Test		0.079	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ALDERSON

DATE/TIME OF TEST: 06/14/2021 03:19:05

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: BOSTIC

DATE/TIME OF TEST: 06/15/2021 01:55:54

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.000	01:51
Breath Volume		1.457	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.000	01:54
Breath Volume		1.382	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GREEN,S
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/18/2021 03:25:24

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.233	03:20
Breath Volume		2.078	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.214	03:23
Breath Volume		1.710	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 06/29/2021 12:06

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WICHMAN,MICHELLE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: COTTON

DATE/TIME OF TEST: 06/19/2021 03:53:16

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.217	03:48
Breath Volume		3.566	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.219	03:51
Breath Volume		3.445	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 06/29/2021 12:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: DIRKS

DATE/TIME OF TEST: 06/19/2021 09:44:52

Diagnostics Check		OK	09:38
Air Blank		0.000	09:38
Control Test		0.081	09:38
Air Blank		0.000	09:39
Subject Sample	# 1	0.130	09:40
Breath Volume		2.585	09:40
Air Blank		0.000	09:41
Air Blank		0.000	09:42
Subject Sample	# 2	0.138	09:43
Breath Volume		4.269	09:43
Air Blank		0.000	09:43
Control Test		0.078	09:44
Air Blank		0.000	09:44
Diagnostics Check		OK	09:44

Data Download Date/Time: 06/29/2021 12:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: LOPACKI

DATE/TIME OF TEST: 06/21/2021 22:38:08

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.081	22:32
Air Blank		0.000	22:32
Subject Sample	# 1	0.122	22:33
Breath Volume		2.261	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.120	22:36
Breath Volume		2.597	22:36
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 06/29/2021 12:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: FLORENTINO

DATE/TIME OF TEST: 06/24/2021 23:21:59

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.209	23:17
Breath Volume		2.511	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:19
Subject Sample	# 2	0.203	23:20
Breath Volume		2.246	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 06/29/2021 12:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: CAULEY

DATE/TIME OF TEST: 06/25/2021 21:05:11

Diagnostics Check		OK	20:57
Air Blank		0.000	20:57
Control Test		0.081	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.000	21:00
Breath Volume		1.414	21:00
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.000	21:03
Breath Volume		1.824	21:03
Air Blank		0.000	21:04
Control Test		0.079	21:04
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 06/29/2021 12:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LOGSDON,DEREK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: CABRAL

DATE/TIME OF TEST: 06/26/2021 18:50:01

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.080	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.316	18:45
Breath Volume		2.417	18:45
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.317	18:48
Breath Volume		2.214	18:48
Air Blank		0.000	18:49
Control Test		0.079	18:49
Air Blank		0.000	18:49
Diagnostics Check		OK	18:50

Data Download Date/Time: 06/29/2021 12:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: STRACK,ROBERT
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LEBRON AYALA

DATE/TIME OF TEST: 06/27/2021 04:26:59

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.058	04:22
Breath Volume		1.625	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:24
Subject Sample	# 2	0.061	04:25
Breath Volume		1.789	04:25
Air Blank		0.000	04:26
Control Test		0.078	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 06/29/2021 12:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GOLDEN,LISA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: UURANNIEMI

DATE/TIME OF TEST: 06/03/2021 19:12:53

Diagnostics Check		OK	19:04
Air Blank		0.000	19:05
Control Test		0.079	19:05
Air Blank		0.000	19:06
Subject Sample	# 1	0.367	19:07
Breath Volume		1.226	19:07
Air Blank		0.000	19:08
Air Blank		0.000	19:09
Subject Sample	# 2	0.355	19:10
Breath Volume		1.363	19:10
Air Blank		0.000	19:11
Control Test		0.077	19:12
Air Blank		0.000	19:12
Diagnostics Check		OK	19:12

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TWYFORD

DATE/TIME OF TEST: 06/16/2021 02:41:02

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.074	02:36
Breath Volume		2.511	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:38
Subject Sample	# 2	0.071	02:39
Breath Volume		3.332	02:39
Air Blank		0.000	02:40
Control Test		0.077	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:41

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BENARROCHE

DATE/TIME OF TEST: 06/17/2021 01:52:12

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:49	
Control Test		0.078	01:50	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:50	Subject Test Refused
Breath Volume		0.000	01:50	
Air Blank		0.000	01:51	
Control Test		0.080	01:51	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFREY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 06/17/2021 02:22:12

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.079	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:22	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WALKER,KIANA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: SARNER

DATE/TIME OF TEST: 06/21/2021 21:57:17

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:54	
Control Test		0.079	21:55	
Air Blank		0.000	21:55	
Subject Sample	# 1		21:55	Subject Test Refused
Breath Volume		0.000	21:55	
Air Blank		0.000	21:56	
Control Test		0.079	21:56	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: BERMAN

DATE/TIME OF TEST: 06/21/2021 22:36:12

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:26	
Control Test		0.078	22:26	
Air Blank		0.000	22:26	
Subject Sample	# 1	0.142	22:30	
Breath Volume		1.968	22:30	
Air Blank		0.000	22:31	
Air Blank		0.000	22:32	
Subject Sample	# 2	0.134	22:34	
Breath Volume		2.195	22:34	
Air Blank			22:35	Purge Fail
Air Blank			22:36	Purge Fail

Data Download Date/Time: 06/30/2021 13:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,5
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 13:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/23/2021 15:08:01

Diagnostics Check	OK	15:06	
Air Blank	0.000	15:07	
Control Test	0.059	15:07	Control Outside Tolerance
Air Blank	0.000	15:08	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 13:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/23/2021 15:10:43

Diagnostics Check	OK	15:09	
Air Blank	0.000	15:09	
Control Test	0.067	15:10	Control Outside Tolerance
Air Blank	0.000	15:10	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 14:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: VAN CAMP

DATE/TIME OF TEST: 06/23/2021 15:26:16

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:25	
Control Test	0.069	15:25	Control Outside Tolerance
Air Blank	0.000	15:26	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,A,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 06/23/2021 22:44:49

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.072	22:44	Control Outside Tolerance
Air Blank	0.000	22:44	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASAS,J,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 06/23/2021 22:47:44

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.074	22:47	Control Outside Tolerance
Air Blank	0.000	22:47	

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VC,JOE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: ADMIN
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/24/2021 15:26:38

Diagnostics Check		OK	15:20
Air Blank		0.000	15:21
Control Test		0.079	15:21
Air Blank		0.000	15:22
Subject Sample	# 1	0.000	15:22
Breath Volume		1.480	15:22
Air Blank		0.000	15:22
Air Blank		0.000	15:24
Subject Sample	# 2	0.000	15:25
Breath Volume		1.644	15:25
Air Blank		0.000	15:25
Control Test		0.079	15:26
Air Blank		0.000	15:26
Diagnostics Check		OK	15:26

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 06/24/2021 03:34:09

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.181	03:29
Breath Volume		2.042	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.180	03:32
Breath Volume		2.968	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 06/30/2021 13:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXIMILIEN,RONALD,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MAXIMILIEN,RONALD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BAUGH

DATE/TIME OF TEST: 06/03/2021 02:28:41

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.082	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 1	0.208	02:27	Volume Not Met
Breath Volume		0.429	02:27	
Air Blank			02:28	Purge Fail
Air Blank		0.000	02:28	

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAXIMILIEN,RONALD,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MAXIMILIEN,RONALD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BAUGH

DATE/TIME OF TEST: 06/03/2021 02:40:03

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.078	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.235	02:34
Breath Volume		1.378	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.227	02:38
Breath Volume		1.351	02:38
Air Blank		0.000	02:39
Control Test		0.076	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANDON,ERICK,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BLANDON,ERICK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/06/2021 01:44:59

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:42	
Control Test		0.081	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:43	Subject Test Refused
Breath Volume		0.000	01:43	
Air Blank		0.000	01:44	
Control Test		0.082	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,RUSSELL,W
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: OCHOA,GUILLERMO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CENDAN

DATE/TIME OF TEST: 06/06/2021 07:38:41

Diagnostics Check		OK	07:35	
Air Blank		0.000	07:36	
Control Test		0.081	07:36	
Air Blank		0.000	07:37	
Subject Sample	# 1		07:37	Subject Test Refused
Breath Volume		0.000	07:37	
Air Blank		0.000	07:37	
Control Test		0.081	07:38	
Air Blank		0.000	07:38	
Diagnostics Check		OK	07:38	

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,ALBERT,M
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: COLLINS,ALBERT
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/10/2021 07:06:13

Diagnostics Check		OK	07:03	
Air Blank		0.000	07:03	
Control Test		0.081	07:04	
Air Blank		0.000	07:04	
Subject Sample	# 1		07:04	Subject Test Refused
Breath Volume		0.000	07:04	
Air Blank		0.000	07:05	
Control Test		0.081	07:05	
Air Blank		0.000	07:06	
Diagnostics Check		OK	07:06	

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 06/12/2021 03:56:22

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.082	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.166	03:51
Breath Volume		1.906	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.163	03:54
Breath Volume		1.714	03:54
Air Blank		0.000	03:55
Control Test		0.082	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,RUSSELL,W
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CRUZ,RUSSELL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: FLORES MALDONADO

DATE/TIME OF TEST: 06/20/2021 03:15:14

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.224	03:10
Breath Volume		2.769	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.227	03:13
Breath Volume		3.472	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,RUSSELL],W
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/20/2021 04:25:55

Diagnostics Check		OK	04:23	
Air Blank		0.000	04:23	
Control Test		0.082	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:24	Subject Test Refused
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Control Test		0.081	04:25	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,RUSSEL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/20/2021 04:42:57

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:38	
Control Test		0.081	04:39	
Air Blank		0.000	04:39	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Control Test		0.081	04:42	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ELLISON,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: TERCERO TALAVERA

DATE/TIME OF TEST: 06/20/2021 06:38:47

Diagnostics Check		OK	06:29	
Air Blank		0.000	06:30	
Control Test		0.080	06:30	
Air Blank		0.000	06:30	
Subject Sample	# 1	0.146	06:31	
Breath Volume		3.046	06:31	
Air Blank		0.000	06:32	
Air Blank		0.000	06:34	
Subject Sample	# 2	0.125	06:34	
Breath Volume		1.468	06:34	
Air Blank		0.000	06:35	
Air Blank			06:37	Improper Sample
Air Blank		0.000	06:37	
Control Test		0.079	06:38	
Air Blank		0.000	06:38	
Diagnostics Check		OK	06:38	

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: THOMPSON,THOMPSON
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: ROQUE VASQUEZ

DATE/TIME OF TEST: 06/20/2021 07:20:34

Diagnostics Check		OK	07:11
Air Blank		0.000	07:11
Control Test		0.081	07:12
Air Blank		0.000	07:12
Subject Sample	# 1	0.246	07:15
Breath Volume		2.187	07:15
Air Blank		0.000	07:16
Air Blank		0.000	07:17
Subject Sample	# 2	0.266	07:18
Breath Volume		2.171	07:18
Air Blank		0.000	07:19
Control Test		0.076	07:19
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,ALBERT,M
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: COLLINS,ALBERT
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 06/25/2021 00:34:57

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.082	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.169	00:30
Breath Volume		1.414	00:30
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.175	00:33
Breath Volume		1.328	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 08/02/2021 23:27

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: LARA,ALEXANDER
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 06/26/2021 00:24:17

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.081	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.146	00:19
Breath Volume		2.296	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.145	00:22
Breath Volume		2.257	00:22
Air Blank		0.000	00:23
Control Test		0.078	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 08/02/2021 23:28

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN,EMMANUEL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JEAN,EMMANUEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/29/2021 19:49:17

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.082	19:43
Air Blank		0.000	19:44
Subject Sample	# 1	0.250	19:44
Breath Volume		1.820	19:44
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.249	19:47
Breath Volume		1.312	19:47
Air Blank		0.000	19:48
Control Test		0.079	19:48
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 08/02/2021 23:28

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAIOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MAHANEY

DATE/TIME OF TEST: 06/03/2021 01:51:03

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.078	01:39	
Air Blank		0.000	01:40	
Subject Sample	# 1	0.129	01:43	Volume Not Met
Breath Volume		0.656	01:43	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2	0.165	01:47	
Breath Volume		1.566	01:47	
Air Blank		0.000	01:48	
Air Blank			01:49	RFI Detect
Air Blank		0.000	01:50	
Control Test		0.079	01:50	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 06/14/2021 11:04

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MAHANEY

DATE/TIME OF TEST: 06/03/2021 02:02:27

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.077	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.134	01:57
Breath Volume		1.273	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.121	02:00
Breath Volume		1.312	02:00
Air Blank		0.000	02:01
Control Test		0.076	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 06/14/2021 11:04

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 06/16/2021 18:20:11

Diagnostics Check	OK	18:19	
Air Blank		18:19	RFI Detect
Air Blank		18:20	RFI Detect

Data Download Date/Time: 07/12/2021 09:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 06/16/2021 18:22:49

Diagnostics Check	OK	18:22	
Air Blank		18:22	RFI Detect
Air Blank		18:22	RFI Detect

Data Download Date/Time: 07/12/2021 09:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULAIMAN,RAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULAIMAN,RAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 972301
DRY GAS EXPIRY DATE: 06/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SLEDGE

DATE/TIME OF TEST: 06/03/2021 01:05:01

Diagnostics Check	OK	01:04	
Air Blank		01:04	RFI Detect
Air Blank	0.000	01:05	

Data Download Date/Time: 07/01/2021 16:16

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULAIMAN,RAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULAIMAN,RAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 972301
DRY GAS EXPIRY DATE: 06/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SLEDGE

DATE/TIME OF TEST: 06/03/2021 01:14:28

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:07	
Control Test		0.078	01:07	
Air Blank		0.000	01:07	
Subject Sample	# 1	0.187	01:08	
Breath Volume		1.558	01:08	
Air Blank		0.000	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 2	0.187	01:12	
Breath Volume		1.726	01:12	
Air Blank		0.000	01:13	
Control Test		0.077	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		Fail	01:14	Analytical Stability Test
Air Blank		0.000	01:14	

Data Download Date/Time: 07/01/2021 16:16

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULAIMAN,RAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULAIMAN,RAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 972301
DRY GAS EXPIRY DATE: 06/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SLEDGE

DATE/TIME OF TEST: 06/03/2021 01:24:27

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.077	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.195	01:19
Breath Volume		1.855	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.182	01:22
Breath Volume		1.859	01:22
Air Blank		0.000	01:23
Control Test		0.076	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 07/01/2021 16:16

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 972301
DRY GAS EXPIRY DATE: 06/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/12/2021 01:04:35

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.078	00:58	
Air Blank		0.000	00:58	
Subject Sample	# 1	0.250	00:59	
Breath Volume		2.480	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2	0.256	01:02	
Breath Volume		2.093	01:02	
Air Blank		0.000	01:03	
Control Test		0.077	01:04	Interferent Detect
Air Blank		0.000	01:04	

Data Download Date/Time: 07/01/2021 16:16

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 972301
DRY GAS EXPIRY DATE: 06/15/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/12/2021 01:15:53

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.253	01:11
Breath Volume		1.878	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.252	01:14
Breath Volume		1.785	01:14
Air Blank		0.000	01:14
Control Test		0.077	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 07/01/2021 16:16

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PAYNTER,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 972301
DRY GAS EXPIRY DATE: 06/15/2021
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SILVEIRA

DATE/TIME OF TEST: 06/12/2021 05:02:35

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.195	04:57
Breath Volume		2.546	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.181	05:00
Breath Volume		1.847	05:00
Air Blank		0.000	05:01
Control Test		0.080	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 07/01/2021 16:16

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,ROBERT,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,ROBERT
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PERSON

DATE/TIME OF TEST: 06/10/2021 04:16:31

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.169	04:11
Breath Volume		2.296	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.162	04:14
Breath Volume		2.089	04:14
Air Blank		0.000	04:15
Control Test		0.077	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHT,JAMES,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VAUGHT,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PEKRUL

DATE/TIME OF TEST: 06/11/2021 04:07:54

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.124	04:03
Breath Volume		3.398	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:05
Subject Sample	# 2	0.111	04:06
Breath Volume		2.343	04:06
Air Blank		0.000	04:06
Control Test		0.077	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/11/2021 19:20:40

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.080	19:14
Air Blank		0.000	19:15
Subject Sample	# 1	0.172	19:15
Breath Volume		2.414	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.177	19:18
Breath Volume		3.554	19:18
Air Blank		0.000	19:19
Control Test		0.077	19:20
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/12/2021 02:03:51

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.080	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.209	01:54
Breath Volume		3.023	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.184	01:58
Breath Volume		1.117	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 3	0.192	02:02
Breath Volume		2.769	02:02
Air Blank		0.000	02:02
Control Test		0.075	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MIKLOS

DATE/TIME OF TEST: 06/12/2021 20:43:11

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.080	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.000	20:38
Breath Volume		1.906	20:38
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.000	20:41
Breath Volume		2.000	20:41
Air Blank		0.000	20:42
Control Test		0.078	20:42
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 06/15/2021 03:47:03

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.193	03:42
Breath Volume		1.468	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.194	03:45
Breath Volume		2.199	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 06/16/2021 19:16:36

Diagnostics Check		OK	19:10
Air Blank		0.000	19:10
Control Test		0.080	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.000	19:12
Breath Volume		3.847	19:12
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.000	19:15
Breath Volume		4.203	19:15
Air Blank		0.000	19:15
Control Test		0.080	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GIRGIS

DATE/TIME OF TEST: 06/17/2021 04:50:53

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.081	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:49	Subject Test Refused
Breath Volume		0.000	04:49	
Air Blank		0.000	04:49	
Control Test		0.080	04:50	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: MENDEZ CARRRILLO

DATE/TIME OF TEST: 06/21/2021 23:26:55

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:20	
Control Test		0.080	23:20	
Air Blank		0.000	23:21	
Subject Sample	# 1	0.197	23:22	
Breath Volume		1.726	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 2		23:25	RFI Detect
Breath Volume		0.000	23:25	
Air Blank		0.000	23:26	
Control Test		0.079	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 07/12/2021 16:51

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: MENDEZ CARRRILLO

DATE/TIME OF TEST: 06/21/2021 23:37:45

Diagnostics Check		OK	23:29
Air Blank		0.000	23:30
Control Test		0.079	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.190	23:32
Breath Volume		1.406	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.185	23:36
Breath Volume		2.171	23:36
Air Blank		0.000	23:36
Control Test		0.078	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 07/12/2021 16:52

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/23/2021 01:54:44

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.180	01:49
Breath Volume		3.312	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.173	01:53
Breath Volume		3.371	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 07/12/2021 16:52

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: LEVASSEUR

DATE/TIME OF TEST: 06/26/2021 20:56:03

Diagnostics Check		OK	20:49
Air Blank		0.000	20:49
Control Test		0.080	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.272	20:51
Breath Volume		2.523	20:51
Air Blank		0.000	20:52
Air Blank		0.000	20:53
Subject Sample	# 2	0.270	20:54
Breath Volume		2.761	20:54
Air Blank		0.000	20:55
Control Test		0.077	20:55
Air Blank		0.000	20:55
Diagnostics Check		OK	20:56

Data Download Date/Time: 07/12/2021 16:52

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CHACON OLIVEROS

DATE/TIME OF TEST: 06/27/2021 01:05:03

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.124	01:00
Breath Volume		1.464	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.114	01:03
Breath Volume		1.863	01:03
Air Blank		0.000	01:04
Control Test		0.078	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

Data Download Date/Time: 07/12/2021 16:52

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/27/2021 22:18:49

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.079	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.212	22:14
Breath Volume		1.558	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.218	22:17
Breath Volume		1.507	22:17
Air Blank		0.000	22:17
Control Test		0.079	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 07/12/2021 16:52

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROSA,CARLY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LECHNER

DATE/TIME OF TEST: 06/03/2021 19:06:18

Diagnostics Check		OK	19:00
Air Blank		0.000	19:00
Control Test		0.078	19:01
Air Blank		0.000	19:01
Subject Sample	# 1	0.000	19:02
Breath Volume		3.320	19:02
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.000	19:04
Breath Volume		3.066	19:04
Air Blank		0.000	19:05
Control Test		0.078	19:05
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 06/25/2021 04:00

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FAULK

DATE/TIME OF TEST: 06/05/2021 03:45:45

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.075	03:41
Breath Volume		2.347	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.069	03:44
Breath Volume		1.734	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/25/2021 04:00

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROSA,CARLY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 06/12/2021 01:40:57

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.172	01:36
Breath Volume		2.957	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.166	01:39
Breath Volume		1.882	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 06/25/2021 04:00

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROSAS,CARLY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/17/2021 00:17:16

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.077	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.000	00:13
Breath Volume		1.421	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.000	00:15
Breath Volume		1.226	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 06/25/2021 04:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MACK,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 06/17/2021 04:02:40

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.192	03:58
Breath Volume		2.632	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:00
Subject Sample	# 2	0.194	04:01
Breath Volume		2.039	04:01
Air Blank		0.000	04:01
Control Test		0.079	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 06/25/2021 04:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROSAS,CARLY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 06/26/2021 02:23:27

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.078	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.103	02:19
Breath Volume		3.753	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.097	02:21
Breath Volume		3.179	02:21
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 08/04/2021 08:11

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROSAS,CARLY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 06/27/2021 00:12:20

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.196	00:07
Breath Volume		2.816	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.191	00:10
Breath Volume		1.894	00:10
Air Blank		0.000	00:11
Control Test		0.075	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 08/04/2021 08:11

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 06/02/2021 00:20:11

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.078	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:18	Subject Test Refused
Breath Volume		0.000	00:18	
Air Blank		0.000	00:19	
Control Test		0.078	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 07/01/2021 16:21

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ESPRONCEDA OSORIA

DATE/TIME OF TEST: 06/05/2021 04:04:07

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.077	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.021	03:59
Breath Volume		1.144	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.019	04:02
Breath Volume		1.226	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 07/01/2021 16:21

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/05/2021 05:06:27

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.077	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.037	05:01
Breath Volume		2.328	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.035	05:04
Breath Volume		1.757	05:04
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 07/01/2021 16:22

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLAWAY,GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/06/2021 22:50:21

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:46	
Control Test		0.078	22:47	
Air Blank		0.000	22:47	
Subject Sample	# 1		22:48	Subject Test Refused
Breath Volume		0.000	22:48	
Air Blank		0.000	22:49	
Control Test		0.078	22:49	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 07/01/2021 16:22

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LAFON

DATE/TIME OF TEST: 06/08/2021 03:56:43

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.205	03:48
Breath Volume		1.460	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.237	03:51
Breath Volume		1.343	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 3	0.230	03:54
Breath Volume		1.359	03:54
Air Blank		0.000	03:55
Control Test		0.077	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 07/01/2021 16:22

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 06/14/2021 05:34:21

Diagnostics Check		OK	05:27
Air Blank		0.000	05:27
Control Test		0.078	05:28
Air Blank		0.000	05:28
Subject Sample	# 1	0.184	05:29
Breath Volume		1.156	05:29
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.198	05:32
Breath Volume		1.445	05:32
Air Blank		0.000	05:33
Control Test		0.078	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 07/01/2021 16:22

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DARLING,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BEIDINGER

DATE/TIME OF TEST: 06/01/2021 23:52:32

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.247	23:47
Breath Volume		2.835	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.237	23:51
Breath Volume		2.773	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 07/01/2021 16:12

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ROGERS-WALSKY

DATE/TIME OF TEST: 06/12/2021 04:21:46

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.130	04:17
Breath Volume		1.375	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.130	04:20
Breath Volume		1.703	04:20
Air Blank		0.000	04:20
Control Test		0.079	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 07/01/2021 16:12

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMAR,GIOVANNI,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HERNAN

DATE/TIME OF TEST: 06/13/2021 02:23:16

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1	0.193	02:11	
Breath Volume		1.496	02:11	
Air Blank		0.000	02:11	
Air Blank		0.000	02:13	
Subject Sample	# 2	0.160	02:17	Volume Not Met
Breath Volume		0.046	02:17	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 3	0.204	02:21	
Breath Volume		1.425	02:21	
Air Blank		0.000	02:22	
Control Test		0.078	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 07/01/2021 16:12

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/14/2021 01:22:15

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.144	01:17
Breath Volume		2.726	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.136	01:20
Breath Volume		1.878	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 07/01/2021 16:12

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARTNER,ARTHUR,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GARTNER,ARTHUR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LAUREANO TORRES

DATE/TIME OF TEST: 06/23/2021 01:42:38

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.195	01:37
Breath Volume		2.152	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.188	01:41
Breath Volume		2.136	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/01/2021 16:12

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 06/01/2021 05:52:35

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.080	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.139	05:47
Breath Volume		1.273	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.138	05:50
Breath Volume		1.414	05:50
Air Blank		0.000	05:51
Control Test		0.078	05:51
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 06/09/2021 13:38

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: JAHN

DATE/TIME OF TEST: 06/05/2021 05:32:25

Diagnostics Check		OK	05:28	
Air Blank		0.000	05:29	
Control Test		0.080	05:29	
Air Blank		0.000	05:29	
Subject Sample	# 1		05:31	Subject Test Refused
Breath Volume		0.000	05:31	
Air Blank		0.000	05:31	
Control Test		0.079	05:31	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 06/09/2021 13:38

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: MUIR

DATE/TIME OF TEST: 06/09/2021 21:49:04

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.080	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.110	21:44
Breath Volume		2.652	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:46
Subject Sample	# 2	0.109	21:47
Breath Volume		2.886	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:49

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 06/11/2021 18:32:39

Diagnostics Check		OK	18:29	
Air Blank		0.000	18:30	
Control Test		0.080	18:30	
Air Blank		0.000	18:31	
Subject Sample	# 1		18:31	Subject Test Refused
Breath Volume		0.000	18:31	
Air Blank		0.000	18:31	
Control Test		0.079	18:32	
Air Blank		0.000	18:32	
Diagnostics Check		OK	18:32	

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: COYLE

DATE/TIME OF TEST: 06/12/2021 23:51:05

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.213	23:46
Breath Volume		2.324	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.222	23:49
Breath Volume		2.226	23:49
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:51

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: FOGARTY

DATE/TIME OF TEST: 06/13/2021 19:19:11

Diagnostics Check		OK	19:06	
Air Blank		0.000	19:07	
Control Test		0.080	19:07	
Air Blank		0.000	19:08	
Subject Sample	# 1	0.206	19:08	
Breath Volume		1.824	19:08	
Air Blank		0.000	19:09	
Air Blank		0.000	19:11	
Subject Sample	# 2	0.236	19:13	
Breath Volume		4.285	19:13	
Air Blank		0.000	19:14	
Air Blank		0.000	19:16	
Subject Sample	# 3	0.225	19:17	
Breath Volume		2.476	19:17	
Air Blank		0.000	19:18	
Control Test		0.074	19:18	Control Outside Tolerance
Air Blank		0.000	19:19	

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: FOGARTY

DATE/TIME OF TEST: 06/13/2021 19:34:28

Diagnostics Check		OK	19:27
Air Blank		0.000	19:27
Control Test		0.077	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.195	19:29
Breath Volume		1.140	19:29
Air Blank		0.000	19:30
Air Blank		0.000	19:32
Subject Sample	# 2	0.192	19:32
Breath Volume		2.320	19:32
Air Blank		0.000	19:33
Control Test		0.078	19:33
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: CAMBRIDGE

DATE/TIME OF TEST: 06/14/2021 03:26:21

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.116	03:21
Breath Volume		3.816	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.108	03:24
Breath Volume		3.523	03:24
Air Blank		0.000	03:25
Control Test		0.076	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: ARAGON

DATE/TIME OF TEST: 06/15/2021 23:15:19

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:06	
Control Test		0.080	23:06	
Air Blank		0.000	23:06	
Subject Sample	# 1	0.100	23:10	Volume Not Met
Breath Volume		0.070	23:10	
Air Blank		0.000	23:10	
Air Blank		0.000	23:12	
Subject Sample	# 2		23:13	Subject Test Refused
Breath Volume		0.339	23:13	
Air Blank		0.000	23:14	
Control Test		0.076	23:14	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATKINS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WATKINS,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 06/17/2021 02:43:33

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.079	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: BRANHAM

DATE/TIME OF TEST: 06/18/2021 18:54:20

Diagnostics Check		OK	18:47
Air Blank		0.000	18:48
Control Test		0.079	18:48
Air Blank		0.000	18:49
Subject Sample	# 1	0.177	18:49
Breath Volume		1.980	18:49
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 2	0.175	18:52
Breath Volume		2.304	18:52
Air Blank		0.000	18:53
Control Test		0.079	18:53
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 07/01/2021 10:17

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BUSCEMI,ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/18/2021 19:22:50

Diagnostics Check		OK	19:11
Air Blank		0.000	19:12
Control Test		0.079	19:12
Air Blank		0.000	19:13
Subject Sample	# 1	0.286	19:13
Breath Volume		2.441	19:13
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.262	19:18
Breath Volume		1.773	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:20
Subject Sample	# 3	0.267	19:21
Breath Volume		2.527	19:21
Air Blank		0.000	19:21
Control Test		0.077	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/20/2021 03:41:32

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.079	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:40	
Control Test		0.078	03:40	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 06/25/2021 01:46:41

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.123	01:41
Breath Volume		2.250	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.130	01:45
Breath Volume		2.640	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 06/25/2021 03:17:25

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.146	03:12
Breath Volume		2.257	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.151	03:15
Breath Volume		2.289	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER,EDWARD,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER,EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/25/2021 04:27:40

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.163	04:22
Breath Volume		2.062	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.157	04:26
Breath Volume		1.792	04:26
Air Blank		0.000	04:26
Control Test		0.079	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JOHN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALLEN,JOHN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 06/25/2021 17:09:39

Diagnostics Check		OK	17:03
Air Blank		0.000	17:03
Control Test		0.080	17:03
Air Blank		0.000	17:04
Subject Sample	# 1	0.089	17:05
Breath Volume		4.238	17:05
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.087	17:08
Breath Volume		4.179	17:08
Air Blank		0.000	17:08
Control Test		0.079	17:09
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,BRANDON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 06/26/2021 01:37:16

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:35	Subject Test Refused
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Control Test		0.080	01:36	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,BRANDON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 06/26/2021 01:43:00

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.080	01:40	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWOGGER,EDWARD,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SWOGGER,EDWARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 06/26/2021 19:55:57

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Subject Sample	# 1	0.000	19:51
Breath Volume		3.441	19:51
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.000	19:54
Breath Volume		3.453	19:54
Air Blank		0.000	19:55
Control Test		0.080	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKY,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 06/26/2021 22:30:55

Diagnostics Check		OK	22:23
Air Blank		0.000	22:24
Control Test		0.080	22:24
Air Blank		0.000	22:25
Subject Sample	# 1	0.237	22:26
Breath Volume		1.203	22:26
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.244	22:29
Breath Volume		1.335	22:29
Air Blank		0.000	22:29
Control Test		0.079	22:30
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,BRANDON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CHAMZINSKAYA

DATE/TIME OF TEST: 06/27/2021 01:10:31

Diagnostics Check		OK	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.106	01:05
Breath Volume		1.984	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.115	01:08
Breath Volume		2.742	01:08
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: GRESHAM

DATE/TIME OF TEST: 06/29/2021 23:56:04

Diagnostics Check		OK	23:53	
Air Blank		0.000	23:53	
Control Test		0.081	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:55	
Control Test		0.081	23:55	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/30/2021 01:14:47

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.088	01:10
Breath Volume		2.718	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.084	01:13
Breath Volume		2.593	01:13
Air Blank		0.000	01:13
Control Test		0.080	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 07/01/2021 10:18

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ANTHONY,A
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: GROVE,DAVID
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 05/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: DIOS

DATE/TIME OF TEST: 06/09/2021 06:50:57

Diagnostics Check		OK	06:39	
Air Blank		0.000	06:39	
Control Test		0.078	06:39	
Air Blank		0.000	06:40	
Subject Sample	# 1		06:43	No Sample Provided
Breath Volume		0.000	06:43	
Air Blank		0.000	06:44	
Air Blank		0.000	06:46	
Subject Sample	# 2	0.023	06:46	
Breath Volume		2.500	06:46	
Air Blank		0.000	06:47	
Air Blank		0.000	06:49	
Subject Sample	# 3	0.020	06:49	
Breath Volume		2.253	06:49	
Air Blank		0.000	06:50	
Control Test		0.077	06:50	
Air Blank		0.000	06:50	
Diagnostics Check		OK	06:50	

Data Download Date/Time: 06/14/2021 13:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,WILFORD,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JORDAN,WILFORD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: DOOLEY

DATE/TIME OF TEST: 06/09/2021 14:45:41

Diagnostics Check		OK	14:32	
Air Blank		0.000	14:32	
Control Test		0.078	14:33	
Air Blank		0.000	14:33	
Subject Sample	# 1	0.198	14:36	Volume Not Met
Breath Volume		0.410	14:36	
Air Blank		0.000	14:37	
Air Blank		0.000	14:39	
Subject Sample	# 2	0.248	14:40	
Breath Volume		1.246	14:40	
Air Blank		0.000	14:41	
Air Blank		0.000	14:43	
Subject Sample	# 3	0.236	14:44	
Breath Volume		1.375	14:44	
Air Blank		0.000	14:44	
Control Test		0.076	14:45	
Air Blank		0.000	14:45	
Diagnostics Check		OK	14:45	

Data Download Date/Time: 07/07/2021 10:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 06/03/2021 03:26:24

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.077	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.135	03:21
Breath Volume		1.417	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.149	03:24
Breath Volume		2.367	03:24
Air Blank		0.000	03:25
Control Test		0.076	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BECK,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 06/04/2021 17:56:43

Diagnostics Check	OK	17:55	
Air Blank	0.000	17:55	
Control Test	0.081	17:56	Interferent Detect
Air Blank	0.000	17:56	

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BECK,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 06/04/2021 18:00:15

Diagnostics Check	OK	17:58	
Air Blank		17:59	Ambient Fail
Air Blank	0.000	18:00	

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BECK,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: BUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 06/04/2021 18:03:41

Diagnostics Check	OK	18:02	
Air Blank		18:02	Ambient Fail
Air Blank	0.000	18:03	

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: YOUNG,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: BUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 06/04/2021 20:09:43

Diagnostics Check	OK	20:09	
Air Blank		20:09	RFI Detect
Air Blank		20:09	RFI Detect

Data Download Date/Time: 06/22/2021 14:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SATTERFIELD

DATE/TIME OF TEST: 06/08/2021 03:12:18

Diagnostics Check	OK	03:11	
Air Blank		03:11	Ambient Fail
Air Blank		03:12	Purge Fail

Data Download Date/Time: 06/22/2021 14:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/12/2021 02:02:05

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.077	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.156	01:56	
Breath Volume		1.167	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2	0.145	01:59	
Breath Volume		2.390	01:59	
Air Blank		0.000	02:00	
Control Test		0.077	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		Fail	02:01	Analytical Stability Test
Air Blank		0.000	02:02	

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 06/12/2021 02:09:09

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.159	02:04
Breath Volume		1.507	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.154	02:07
Breath Volume		1.597	02:07
Air Blank		0.000	02:08
Control Test		0.077	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 06/13/2021 01:57:17

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.077	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.009	01:52
Breath Volume		1.421	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.008	01:55
Breath Volume		1.164	01:55
Air Blank		0.000	01:56
Control Test		0.075	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: EKLUND

DATE/TIME OF TEST: 06/14/2021 00:48:39

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.078	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.143	00:44
Breath Volume		3.007	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.129	00:47
Breath Volume		1.488	00:47
Air Blank		0.000	00:47
Control Test		0.077	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 06/22/2021 14:01

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MCDONNELL

DATE/TIME OF TEST: 06/26/2021 22:36:51

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.076	22:31
Air Blank		0.000	22:32
Subject Sample	# 1	0.000	22:32
Breath Volume		4.304	22:32
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.000	22:35
Breath Volume		4.035	22:35
Air Blank		0.000	22:35
Control Test		0.076	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 07/27/2021 17:11

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: AWE

DATE/TIME OF TEST: 06/27/2021 01:18:38

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.076	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.127	01:13
Breath Volume		1.386	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.134	01:16
Breath Volume		1.496	01:16
Air Blank		0.000	01:17
Control Test		0.076	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 07/27/2021 17:12

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: JAKEY

DATE/TIME OF TEST: 06/12/2021 23:56:07

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.081	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1	0.186	23:43	
Breath Volume		1.546	23:43	
Air Blank		0.000	23:43	
Air Blank		0.000	23:45	
Subject Sample	# 2	0.124	23:48	Volume Not Met
Breath Volume		0.414	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:51	
Subject Sample	# 3	0.129	23:54	Volume Not Met
Breath Volume		0.121	23:54	
Air Blank		0.000	23:55	
Control Test		0.079	23:55	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 06/14/2021 12:36

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 190362
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 06/12/2021 04:10:44

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.041	04:10	Control Outside Tolerance
Air Blank	0.000	04:10	

Data Download Date/Time: 06/14/2021 11:21

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 05/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MELLION

DATE/TIME OF TEST: 06/02/2021 03:41:15

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.078	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.104	03:36
Breath Volume		1.128	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 2	0.112	03:39
Breath Volume		1.945	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 06/03/2021 09:09

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ACCINELLI

DATE/TIME OF TEST: 06/05/2021 01:03:51

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.077	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.149	00:59
Breath Volume		1.992	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.155	01:02
Breath Volume		1.894	01:02
Air Blank		0.000	01:02
Control Test		0.077	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SANCHEZ,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: OSPINA

DATE/TIME OF TEST: 06/10/2021 03:11:13

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.078	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.077	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/12/2021 01:21:02

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.076	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.134	01:15
Breath Volume		1.507	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.130	01:19
Breath Volume		1.640	01:19
Air Blank		0.000	01:20
Control Test		0.077	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:21

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 06/16/2021 04:29:28

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.078	04:27	
Air Blank		0.000	04:27	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.076	04:28	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MENDEZ REYES

DATE/TIME OF TEST: 06/19/2021 01:29:27

Diagnostics Check		OK	01:26	
Air Blank		0.000	01:26	
Control Test		0.077	01:27	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:28	Subject Test Refused
Breath Volume		0.000	01:28	
Air Blank		0.000	01:28	
Control Test		0.075	01:28	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: SOSSA

DATE/TIME OF TEST: 06/24/2021 01:03:59

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.078	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.149	00:59
Breath Volume		1.621	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.147	01:02
Breath Volume		1.777	01:02
Air Blank		0.000	01:03
Control Test		0.078	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINLAND,J,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SERRANO ROSA

DATE/TIME OF TEST: 06/25/2021 02:12:43

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.183	02:07
Breath Volume		1.148	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.189	02:11
Breath Volume		2.011	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINALND,J,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: WHITTLE

DATE/TIME OF TEST: 06/26/2021 04:42:42

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.173	04:38
Breath Volume		3.070	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.170	04:41
Breath Volume		2.468	04:41
Air Blank		0.000	04:41
Control Test		0.076	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINLAND,J,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PETITFRERE,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MADRID-FERRERA

DATE/TIME OF TEST: 06/27/2021 00:43:22

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.077	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.208	00:38
Breath Volume		2.609	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.210	00:41
Breath Volume		1.777	00:41
Air Blank		0.000	00:42
Control Test		0.076	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 07/14/2021 09:53

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRICKS,JOHN,M
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HENDRICKS,JOHN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: LABORDE

DATE/TIME OF TEST: 06/01/2021 18:11:21

Diagnostics Check		OK	18:04
Air Blank		0.000	18:04
Control Test		0.082	18:04
Air Blank		0.000	18:05
Subject Sample	# 1	0.214	18:06
Breath Volume		2.074	18:06
Air Blank		0.000	18:07
Air Blank		0.000	18:08
Subject Sample	# 2	0.212	18:09
Breath Volume		2.148	18:09
Air Blank		0.000	18:10
Control Test		0.078	18:10
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 06/07/2021 19:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SIRMANS,ABNER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 06/02/2021 02:57:53

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:55	
Control Test		0.082	02:55	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.082	02:57	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 06/07/2021 19:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MYRICK,JASMINE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: RIOS PENALOZA

DATE/TIME OF TEST: 06/13/2021 01:17:24

Diagnostics Check		OK	01:14	
Air Blank		0.000	01:14	
Control Test		0.082	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:16	
Control Test		0.082	01:16	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 06/30/2021 20:10

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HAIL,WILLIAM
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 05/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GUITY

DATE/TIME OF TEST: 06/27/2021 04:04:18

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.082	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1	0.113	03:54	Volume Not Met
Breath Volume		0.113	03:54	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2	0.134	03:59	
Breath Volume		1.816	03:59	
Air Blank		0.000	03:59	
Air Blank		0.000	04:01	
Subject Sample	# 3	0.125	04:02	
Breath Volume		1.421	04:02	
Air Blank		0.000	04:03	
Control Test		0.000	04:03	RFI Detect
Air Blank		0.000	04:04	

Data Download Date/Time: 06/30/2021 20:10

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES',W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/05/2021 03:34:16

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.078	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.165	03:29
Breath Volume		2.371	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.175	03:32
Breath Volume		2.933	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 07/14/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/05/2021 06:14:31

Diagnostics Check		OK	06:07
Air Blank		0.000	06:08
Control Test		0.078	06:08
Air Blank		0.000	06:09
Subject Sample	# 1	0.000	06:09
Breath Volume		1.457	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.000	06:13
Breath Volume		1.476	06:13
Air Blank		0.000	06:13
Control Test		0.078	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 07/14/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 06/10/2021 22:44:08

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.077	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.228	22:39
Breath Volume		3.777	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.231	22:42
Breath Volume		3.421	22:42
Air Blank		0.000	22:43
Control Test		0.077	22:43
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 07/14/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MAHIN

DATE/TIME OF TEST: 06/14/2021 22:15:40

Diagnostics Check		OK	22:08
Air Blank		0.000	22:09
Control Test		0.077	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.164	22:10
Breath Volume		1.355	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:12
Subject Sample	# 2	0.154	22:14
Breath Volume		2.476	22:14
Air Blank		0.000	22:14
Control Test		0.078	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 07/14/2021 10:41

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: CROWELL

DATE/TIME OF TEST: 06/01/2021 22:53:12

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.079	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.000	22:48
Breath Volume		2.011	22:48
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.000	22:51
Breath Volume		1.714	22:51
Air Blank		0.000	22:52
Control Test		0.079	22:52
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 06/02/2021 10:03

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: FOLEY,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 06/09/2021 22:33:54

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.078	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.081	22:29
Breath Volume		1.117	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.091	22:32
Breath Volume		1.906	22:32
Air Blank		0.000	22:32
Control Test		0.078	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 07/14/2021 15:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SELLER,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HOVART

DATE/TIME OF TEST: 06/10/2021 02:15:16

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.220	02:10
Breath Volume		2.265	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.218	02:13
Breath Volume		2.867	02:13
Air Blank		0.000	02:14
Control Test		0.078	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 07/14/2021 15:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONGORA,ANTONIO,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GONGORA,ANTONIO
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: HEMINGWAY

DATE/TIME OF TEST: 06/10/2021 16:06:37

Diagnostics Check		OK	16:00
Air Blank		0.000	16:00
Control Test		0.078	16:01
Air Blank		0.000	16:01
Subject Sample	# 1	0.242	16:02
Breath Volume		2.488	16:02
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 2	0.223	16:05
Breath Volume		1.496	16:05
Air Blank		0.000	16:05
Control Test		0.077	16:06
Air Blank		0.000	16:06
Diagnostics Check		OK	16:06

Data Download Date/Time: 07/14/2021 15:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: NAPLES PD
ARRESTING OFFICER: TOBLER,J
ARRESTING OFFICER AGENCY: NAPLES PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GROENINGS

DATE/TIME OF TEST: 06/12/2021 03:43:53

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.077	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.153	03:39
Breath Volume		3.140	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.151	03:42
Breath Volume		2.898	03:42
Air Blank		0.000	03:42
Control Test		0.077	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 07/14/2021 15:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: KISTLER

DATE/TIME OF TEST: 06/13/2021 04:15:35

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.193	04:11
Breath Volume		2.929	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.176	04:13
Breath Volume		2.125	04:13
Air Blank		0.000	04:14
Control Test		0.078	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: REED,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/13/2021 08:01:43

Diagnostics Check		OK	07:54
Air Blank		0.000	07:54
Control Test		0.077	07:54
Air Blank		0.000	07:55
Subject Sample	# 1	0.000	07:56
Breath Volume		1.164	07:56
Air Blank		0.000	07:56
Air Blank		0.000	07:58
Subject Sample	# 2	0.000	08:00
Breath Volume		1.261	08:00
Air Blank		0.000	08:00
Control Test		0.077	08:01
Air Blank		0.000	08:01
Diagnostics Check		OK	08:01

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: RIVERA OLGUIN

DATE/TIME OF TEST: 06/13/2021 22:20:44

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.077	22:14
Air Blank		0.000	22:15
Subject Sample	# 1	0.238	22:16
Breath Volume		2.687	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.244	22:19
Breath Volume		2.738	22:19
Air Blank		0.000	22:19
Control Test		0.075	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROA,ALEXANDRA,S
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: JEAN,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: RIVERASOTO

DATE/TIME OF TEST: 06/16/2021 15:35:40

Diagnostics Check		OK	15:29
Air Blank		0.000	15:29
Control Test		0.077	15:29
Air Blank		0.000	15:30
Subject Sample	# 1	0.258	15:30
Breath Volume		1.765	15:30
Air Blank		0.000	15:31
Air Blank		0.000	15:33
Subject Sample	# 2	0.257	15:33
Breath Volume		2.621	15:33
Air Blank		0.000	15:34
Control Test		0.076	15:35
Air Blank		0.000	15:35
Diagnostics Check		OK	15:35

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,MAXWELL,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: RODRIGUEZ-PITRE,VALERIE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 06/18/2021 05:20:02

Diagnostics Check		OK	05:12
Air Blank		0.000	05:13
Control Test		0.078	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.160	05:14
Breath Volume		1.742	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.154	05:18
Breath Volume		2.753	05:18
Air Blank		0.000	05:19
Control Test		0.075	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:20

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CONROY,JESSICA
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: WYNESS

DATE/TIME OF TEST: 06/19/2021 20:44:52

Diagnostics Check		OK	20:38
Air Blank		0.000	20:38
Control Test		0.078	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.096	20:40
Breath Volume		1.910	20:40
Air Blank		0.000	20:40
Air Blank		0.000	20:42
Subject Sample	# 2	0.088	20:43
Breath Volume		2.035	20:43
Air Blank		0.000	20:43
Control Test		0.078	20:44
Air Blank		0.000	20:44
Diagnostics Check		OK	20:44

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,THOMAS,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ENDELMANN,K
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ALTHOUSE

DATE/TIME OF TEST: 06/19/2021 23:59:38

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.221	23:54
Breath Volume		3.515	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.216	23:57
Breath Volume		3.347	23:57
Air Blank		0.000	23:58
Control Test		0.077	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELIAS,D,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ELIAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/21/2021 04:38:04

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.078	04:32
Air Blank		0.000	04:32
Subject Sample	# 1	0.159	04:33
Breath Volume		2.101	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:35
Subject Sample	# 2	0.170	04:36
Breath Volume		3.453	04:36
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:38

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: LITKA,A
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 06/22/2021 17:59:11

Diagnostics Check		OK	17:53
Air Blank		0.000	17:53
Control Test		0.078	17:53
Air Blank		0.000	17:54
Subject Sample	# 1	0.000	17:54
Breath Volume		1.921	17:54
Air Blank		0.000	17:55
Air Blank		0.000	17:57
Subject Sample	# 2	0.000	17:57
Breath Volume		1.562	17:57
Air Blank		0.000	17:58
Control Test		0.079	17:58
Air Blank		0.000	17:59
Diagnostics Check		OK	17:59

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WALKER,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LUCIO MORALES

DATE/TIME OF TEST: 06/23/2021 05:12:16

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.078	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.218	05:04
Breath Volume		1.609	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.195	05:07
Breath Volume		1.660	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 3	0.197	05:10
Breath Volume		1.906	05:10
Air Blank		0.000	05:11
Control Test		0.076	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBIN,L,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BORCHERS,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:04

SUBJECT'S LAST NAME: MCGEARY

DATE/TIME OF TEST: 06/23/2021 11:34:02

Diagnostics Check		OK	11:27
Air Blank		0.000	11:28
Control Test		0.078	11:28
Air Blank		0.000	11:28
Subject Sample	# 1	0.000	11:29
Breath Volume		2.488	11:29
Air Blank		0.000	11:30
Air Blank		0.000	11:32
Subject Sample	# 2	0.000	11:32
Breath Volume		2.527	11:32
Air Blank		0.000	11:33
Control Test		0.079	11:33
Air Blank		0.000	11:33
Diagnostics Check		OK	11:34

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GARCIA,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MACE

DATE/TIME OF TEST: 06/24/2021 19:58:41

Diagnostics Check		OK	19:52
Air Blank		0.000	19:52
Control Test		0.078	19:53
Air Blank		0.000	19:53
Subject Sample	# 1	0.214	19:54
Breath Volume		1.679	19:54
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 2	0.212	19:57
Breath Volume		1.683	19:57
Air Blank		0.000	19:57
Control Test		0.077	19:58
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GOODMAN,R
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: SANJUAN

DATE/TIME OF TEST: 06/26/2021 05:29:58

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.078	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.050	05:25
Breath Volume		4.484	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:27
Subject Sample	# 2	0.049	05:28
Breath Volume		4.648	05:28
Air Blank		0.000	05:29
Control Test		0.077	05:29
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GROSSKLAS,D
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: OLDS

DATE/TIME OF TEST: 06/27/2021 02:34:32

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.078	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.164	02:29
Breath Volume		2.367	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:31
Subject Sample	# 2	0.173	02:32
Breath Volume		3.656	02:32
Air Blank		0.000	02:33
Control Test		0.077	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,MAXWELL,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: JOHNSON,TONY
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ALVARADO ALVARADO

DATE/TIME OF TEST: 06/28/2021 00:31:48

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.078	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.211	00:27
Breath Volume		3.781	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:29
Subject Sample	# 2	0.207	00:30
Breath Volume		4.417	00:30
Air Blank		0.000	00:30
Control Test		0.077	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 07/14/2021 15:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,S,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: SHERAN,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 06/28/2021 01:38:27

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.078	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.000	01:33
Breath Volume		3.285	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.000	01:37
Breath Volume		3.281	01:37
Air Blank		0.000	01:37
Control Test		0.078	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 07/14/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,N/A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: OLAH

DATE/TIME OF TEST: 06/01/2021 02:54:22

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.000	02:49
Breath Volume		4.023	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.000	02:52
Breath Volume		3.324	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 06/22/2021 10:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BRASCH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/01/2021 05:59:28

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.079	05:53
Air Blank		0.000	05:54
Subject Sample	# 1	0.000	05:55
Breath Volume		2.160	05:55
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.000	05:58
Breath Volume		1.562	05:58
Air Blank		0.000	05:58
Control Test		0.079	05:58
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 06/22/2021 10:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REILLY,S.,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REILLY,S.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 06/03/2021 02:50:30

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.080	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.030	02:45
Breath Volume		2.941	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.026	02:48
Breath Volume		2.582	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 06/22/2021 10:42

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PERO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: AGNANT

DATE/TIME OF TEST: 06/05/2021 02:32:21

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.209	02:26
Breath Volume		1.851	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.192	02:30
Breath Volume		1.355	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WELCH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: FURNISS

DATE/TIME OF TEST: 06/05/2021 04:01:10

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.000	03:56
Breath Volume		4.027	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.000	03:59
Breath Volume		4.339	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: C. SMITH,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: CHICHESTER

DATE/TIME OF TEST: 06/06/2021 02:13:27

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.108	02:08
Breath Volume		2.054	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.107	02:11
Breath Volume		2.136	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OKEEFE,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/08/2021 17:39:05

Diagnostics Check		OK	17:32
Air Blank		0.000	17:32
Control Test		0.080	17:32
Air Blank		0.000	17:33
Subject Sample	# 1	0.130	17:34
Breath Volume		2.308	17:34
Air Blank		0.000	17:34
Air Blank		0.000	17:36
Subject Sample	# 2	0.119	17:37
Breath Volume		1.156	17:37
Air Blank		0.000	17:38
Control Test		0.080	17:38
Air Blank		0.000	17:39
Diagnostics Check		OK	17:39

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,BRIAN,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,JOHN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 06/09/2021 01:42:11

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.292	01:36
Breath Volume		3.785	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.292	01:40
Breath Volume		3.949	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 06/13/2021 03:44:16

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.125	03:39
Breath Volume		1.273	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.132	03:42
Breath Volume		1.695	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEE,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 06/14/2021 04:36:49

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.244	04:31
Breath Volume		2.562	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.242	04:35
Breath Volume		2.925	04:35
Air Blank		0.000	04:35
Control Test		0.079	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GUEVARA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 06/14/2021 21:13:36

Diagnostics Check		OK	21:02	
Air Blank		0.000	21:02	
Control Test		0.079	21:02	
Air Blank		0.000	21:03	
Subject Sample	# 1	0.154	21:06	Volume Not Met
Breath Volume		0.312	21:06	
Air Blank		0.000	21:07	
Air Blank		0.000	21:09	
Subject Sample	# 2		21:12	No Sample Provided
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Control Test		0.078	21:12	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GEARHART,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 06/15/2021 04:51:33

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.080	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.000	04:47
Breath Volume		4.031	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.000	04:50
Breath Volume		3.421	04:50
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HILE

DATE/TIME OF TEST: 06/17/2021 04:23:10

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.112	04:18
Breath Volume		3.964	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:20
Subject Sample	# 2	0.110	04:21
Breath Volume		3.496	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CICCARONE

DATE/TIME OF TEST: 06/17/2021 22:39:16

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.079	22:33
Air Blank		0.000	22:33
Subject Sample	# 1	0.165	22:34
Breath Volume		1.679	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:36
Subject Sample	# 2	0.166	22:37
Breath Volume		1.359	22:37
Air Blank		0.000	22:38
Control Test		0.079	22:38
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BRASCH,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: NATAL

DATE/TIME OF TEST: 06/20/2021 03:03:37

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.084	02:58
Breath Volume		2.488	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.083	03:02
Breath Volume		4.339	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 06/22/2021 10:43

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/23/2021 04:15:48

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.237	04:11
Breath Volume		2.183	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.223	04:14
Breath Volume		2.480	04:14
Air Blank		0.000	04:14
Control Test		0.079	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: POOLE,T
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/25/2021 18:46:05

Diagnostics Check		OK	18:39
Air Blank		0.000	18:39
Control Test		0.079	18:39
Air Blank		0.000	18:40
Subject Sample	# 1	0.160	18:41
Breath Volume		4.132	18:41
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.165	18:44
Breath Volume		4.234	18:44
Air Blank		0.000	18:45
Control Test		0.079	18:45
Air Blank		0.000	18:46
Diagnostics Check		OK	18:46

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HEGWOOD,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 06/26/2021 19:08:26

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.079	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.194	19:03
Breath Volume		1.539	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:05
Subject Sample	# 2	0.203	19:06
Breath Volume		2.375	19:06
Air Blank		0.000	19:07
Control Test		0.079	19:07
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: ADSIDE

DATE/TIME OF TEST: 06/27/2021 04:07:11

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.080	04:02
Breath Volume		2.257	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.077	04:05
Breath Volume		2.957	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HOLLINGSWORTH,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 06/28/2021 01:00:18

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.216	00:55
Breath Volume		2.183	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:57
Subject Sample	# 2	0.211	00:58
Breath Volume		2.753	00:58
Air Blank		0.000	00:59
Control Test		0.079	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PARRA-HERMIDA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WEINRICH

DATE/TIME OF TEST: 06/28/2021 22:02:45

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.097	21:57
Breath Volume		1.343	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.089	22:01
Breath Volume		2.328	22:01
Air Blank		0.000	22:01
Control Test		0.079	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PARRA-HERMIDA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WEINRICH

DATE/TIME OF TEST: 06/29/2021 04:02:30

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.000	03:58
Breath Volume		2.339	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.000	04:01
Breath Volume		2.996	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,ALAN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON J,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 06/30/2021 02:10:22

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.000	02:05
Breath Volume		2.765	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.000	02:08
Breath Volume		1.640	02:08
Air Blank		0.000	02:09
Control Test		0.081	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 08/02/2021 11:31

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 04/22/2022
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/30/2021 23:11:21

Diagnostics Check		OK	23:07	
Air Blank		0.000	23:07	
Control Test		0.080	23:07	
Air Blank		0.000	23:08	
Subject Sample	# 1	0.242	23:10	
Breath Volume		1.828	23:10	
Air Blank			23:10	RFI Detect
Air Blank			23:10	RFI Detect
Air Blank		0.000	23:11	

Data Download Date/Time: 07/14/2021 10:10

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06419080A2
DRY GAS EXPIRY DATE: 04/22/2022
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 06/30/2021 23:19:46

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.078	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.249	23:14
Breath Volume		2.253	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.232	23:17
Breath Volume		2.500	23:17
Air Blank		0.000	23:18
Control Test		0.077	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 07/14/2021 10:10

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,B
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 06/05/2021 02:39:05

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.078	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.154	02:34
Breath Volume		1.714	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:36
Subject Sample	# 2	0.165	02:37
Breath Volume		2.074	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 07/14/2021 10:48

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CANTAVE

DATE/TIME OF TEST: 06/06/2021 02:05:17

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.078	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		02:01	No Sample Provided
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2		02:04	Sequence Aborted
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	

Data Download Date/Time: 07/14/2021 10:48

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CANTAVE

DATE/TIME OF TEST: 06/06/2021 02:23:20

Diagnostics Check	Fail	02:22	Internal Printer Test
Air Blank	0.000	02:23	

Data Download Date/Time: 07/14/2021 10:48

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CANTAVE

DATE/TIME OF TEST: 06/06/2021 02:35:03

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.237	02:28
Breath Volume		1.656	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.228	02:33
Breath Volume		2.062	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

Data Download Date/Time: 07/14/2021 10:48

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ROSSMAN

DATE/TIME OF TEST: 06/11/2021 01:45:04

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.161	01:40
Breath Volume		1.660	01:40
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.154	01:43
Breath Volume		1.605	01:43
Air Blank		0.000	01:44
Control Test		0.077	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:45

Data Download Date/Time: 07/14/2021 10:49

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: RODRIGUEZ HERRERA

DATE/TIME OF TEST: 06/11/2021 22:41:10

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.163	22:36
Breath Volume		1.308	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.177	22:39
Breath Volume		2.199	22:39
Air Blank		0.000	22:40
Control Test		0.078	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 07/14/2021 10:49

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: STEFFEY

DATE/TIME OF TEST: 06/27/2021 01:23:55

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:23	
Control Test		0.077	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 07/14/2021 10:49

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GLASS,DREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 06/19/2021 03:44:24

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.082	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.188	03:39
Breath Volume		1.843	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.187	03:42
Breath Volume		1.593	03:42
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MARTINEZ,HENDRICK
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 06/19/2021 19:15:08

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.081	19:02
Air Blank		0.000	19:03
Subject Sample	# 1	0.287	19:04
Breath Volume		2.785	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.238	19:07
Breath Volume		1.183	19:07
Air Blank		0.000	19:08
Air Blank		0.000	19:10
Subject Sample	# 3	0.234	19:13
Breath Volume		3.339	19:13
Air Blank		0.000	19:14
Control Test		0.079	19:14
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GLASS,DREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ELSING

DATE/TIME OF TEST: 06/19/2021 23:51:19

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.081	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.057	23:46
Breath Volume		1.707	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.062	23:49
Breath Volume		3.992	23:49
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MAY,MIYA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 06/21/2021 21:37:05

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.081	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.131	21:32
Breath Volume		2.886	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.133	21:35
Breath Volume		2.835	21:35
Air Blank		0.000	21:36
Control Test		0.080	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUZBEE,HEATHER,N
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MLINAC,MADISON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BOC-CHEX

DATE/TIME OF TEST: 06/22/2021 20:15:40

Diagnostics Check		OK	20:08
Air Blank		0.000	20:09
Control Test		0.080	20:09
Air Blank		0.000	20:10
Subject Sample	# 1	0.194	20:10
Breath Volume		2.066	20:10
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.196	20:13
Breath Volume		2.546	20:13
Air Blank		0.000	20:14
Control Test		0.079	20:15
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MCMAHON,KENNETH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 06/22/2021 20:45:15

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.080	20:39
Air Blank		0.000	20:40
Subject Sample	# 1	0.000	20:40
Breath Volume		3.261	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.000	20:43
Breath Volume		3.734	20:43
Air Blank		0.000	20:44
Control Test		0.080	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LEACH,ZACHARY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 06/25/2021 21:42:59

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.081	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.191	21:37
Breath Volume		1.550	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.184	21:41
Breath Volume		2.277	21:41
Air Blank		0.000	21:42
Control Test		0.076	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MCMAHON,KENNETH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 06/30/2021 22:17:34

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.082	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.000	22:13
Breath Volume		1.894	22:13
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.000	22:16
Breath Volume		2.441	22:16
Air Blank		0.000	22:16
Control Test		0.081	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 08/12/2021 10:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JESSICA,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/08/2021 15:51:35

Diagnostics Check		OK	15:48	
Air Blank		0.000	15:49	
Control Test		0.079	15:49	
Air Blank		0.000	15:50	
Subject Sample	# 1		15:50	Subject Test Refused
Breath Volume		0.000	15:50	
Air Blank		0.000	15:50	
Control Test		0.080	15:50	
Air Blank		0.000	15:51	
Diagnostics Check		OK	15:51	

Data Download Date/Time: 07/01/2021 14:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: EVERSOLE,DONALD
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/17/2021 23:52:07

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.084	23:47
Breath Volume		3.597	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:49
Subject Sample	# 2	0.088	23:50
Breath Volume		1.613	23:50
Air Blank		0.000	23:51
Control Test		0.077	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 07/01/2021 14:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LOPEZ-RUIZ

DATE/TIME OF TEST: 06/27/2021 03:29:16

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.126	03:23
Breath Volume		1.742	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.118	03:27
Breath Volume		1.890	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 07/01/2021 14:47

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SKEETE

DATE/TIME OF TEST: 06/05/2021 02:15:25

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.078	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.192	02:10
Breath Volume		3.050	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.195	02:13
Breath Volume		2.468	02:13
Air Blank		0.000	02:14
Control Test		0.078	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 07/23/2021 09:56

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 06/05/2021 03:09:11

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.151	03:04
Breath Volume		1.714	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.155	03:07
Breath Volume		2.945	03:07
Air Blank		0.000	03:08
Control Test		0.077	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 07/23/2021 09:56

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: GALINDO CACEREZ

DATE/TIME OF TEST: 06/20/2021 21:37:03

Diagnostics Check		OK	21:29
Air Blank		0.000	21:30
Control Test		0.079	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.178	21:31
Breath Volume		1.816	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.175	21:35
Breath Volume		1.882	21:35
Air Blank		0.000	21:36
Control Test		0.077	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:37

Data Download Date/Time: 07/23/2021 09:56

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: RAMIREZ ORTEGA

DATE/TIME OF TEST: 06/20/2021 22:20:17

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.078	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.125	22:15
Breath Volume		2.214	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:17
Subject Sample	# 2	0.125	22:18
Breath Volume		1.898	22:18
Air Blank		0.000	22:19
Control Test		0.077	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 07/23/2021 09:56

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 06/21/2021 00:35:09

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.257	00:30
Breath Volume		3.074	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.250	00:33
Breath Volume		2.917	00:33
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 07/23/2021 09:56

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: GRANVILLE

DATE/TIME OF TEST: 06/30/2021 00:37:47

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.000	00:33
Breath Volume		1.972	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.000	00:36
Breath Volume		1.101	00:36
Air Blank		0.000	00:36
Control Test		0.078	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 07/23/2021 09:56

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: BERGNES

DATE/TIME OF TEST: 06/01/2021 19:25:37

Diagnostics Check		OK	19:18
Air Blank		0.000	19:19
Control Test		0.079	19:19
Air Blank		0.000	19:20
Subject Sample	# 1	0.185	19:20
Breath Volume		2.792	19:20
Air Blank		0.000	19:21
Air Blank		0.000	19:23
Subject Sample	# 2	0.183	19:23
Breath Volume		1.535	19:23
Air Blank		0.000	19:24
Control Test		0.079	19:25
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 06/07/2021 12:50

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: VANHOOSE,CHRIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PICKEN

DATE/TIME OF TEST: 06/06/2021 00:54:36

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.369	00:50
Breath Volume		1.972	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.381	00:52
Breath Volume		1.894	00:52
Air Blank		0.000	00:53
Control Test		0.077	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 06/07/2021 12:50

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRA,WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 06/07/2021 02:10:04

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.140	02:05
Breath Volume		1.484	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.139	02:08
Breath Volume		1.894	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:10

Data Download Date/Time: 06/07/2021 12:50

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURCIER,LEON,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SALAZAR,JUSTIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: BUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: HENNESSEY

DATE/TIME OF TEST: 06/12/2021 00:09:17

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.080	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Control Test		0.080	00:08	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/13/2021 01:28:40

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,KYLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: NUNLEY

DATE/TIME OF TEST: 06/13/2021 23:36:18

Diagnostics Check		OK	23:27	
Air Blank		0.000	23:28	
Control Test		0.080	23:28	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:32	No Sample Provided
Breath Volume		0.000	23:32	
Air Blank		0.000	23:32	
Air Blank		0.000	23:34	
Subject Sample	# 2		23:34	Subject Test Refused
Breath Volume		0.000	23:34	
Air Blank		0.000	23:35	
Control Test		0.079	23:35	
Air Blank		0.000	23:36	
Diagnostics Check		OK	23:36	

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOMPOINT,DESLILES,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MOMPOINT,DESLILES
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 06/20/2021 16:14:31

Diagnostics Check		OK	16:07
Air Blank		0.000	16:08
Control Test		0.080	16:08
Air Blank		0.000	16:09
Subject Sample	# 1	0.176	16:09
Breath Volume		1.699	16:09
Air Blank		0.000	16:10
Air Blank		0.000	16:12
Subject Sample	# 2	0.189	16:12
Breath Volume		1.734	16:12
Air Blank		0.000	16:13
Control Test		0.079	16:13
Air Blank		0.000	16:14
Diagnostics Check		OK	16:14

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: VANHOOSE,CHRIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 06/20/2021 23:59:07

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:56	
Control Test		0.080	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:57	Subject Test Refused
Breath Volume		0.000	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HEMPHILL

DATE/TIME OF TEST: 06/23/2021 00:54:34

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.084	00:49
Breath Volume		1.746	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.081	00:52
Breath Volume		1.972	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: VANHOOSE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: STORK

DATE/TIME OF TEST: 06/24/2021 20:37:00

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:33	
Control Test		0.080	20:33	
Air Blank		0.000	20:34	
Subject Sample	# 1		20:35	Subject Test Refused
Breath Volume		0.000	20:35	
Air Blank		0.000	20:36	
Control Test		0.080	20:36	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:37	

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MANSO

DATE/TIME OF TEST: 06/26/2021 02:09:41

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.103	02:04
Breath Volume		1.226	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.097	02:08
Breath Volume		1.761	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEIRD,ANDREW,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LEIRD,ANDREW
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: ZERO TOL
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: BRULAND

DATE/TIME OF TEST: 06/26/2021 12:41:27

Diagnostics Check		OK	12:35
Air Blank		0.000	12:35
Control Test		0.080	12:35
Air Blank		0.000	12:36
Subject Sample	# 1	0.011	12:37
Breath Volume		2.824	12:37
Air Blank		0.000	12:37
Air Blank		0.000	12:39
Subject Sample	# 2	0.008	12:40
Breath Volume		1.652	12:40
Air Blank		0.000	12:40
Control Test		0.080	12:40
Air Blank		0.000	12:41
Diagnostics Check		OK	12:41

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: GREIDER

DATE/TIME OF TEST: 06/30/2021 21:23:56

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.080	21:21	
Air Blank		0.000	21:21	
Subject Sample	# 1		21:22	Subject Test Refused
Breath Volume		0.000	21:22	
Air Blank		0.000	21:22	
Control Test		0.080	21:23	
Air Blank		0.000	21:23	
Diagnostics Check		OK	21:23	

Data Download Date/Time: 07/01/2021 14:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: HOLDERFIELD

DATE/TIME OF TEST: 06/01/2021 17:55:37

Diagnostics Check		OK	17:47
Air Blank		0.000	17:48
Control Test		0.081	17:48
Air Blank		0.000	17:49
Subject Sample	# 1	0.000	17:50
Breath Volume		2.000	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.000	17:54
Breath Volume		2.054	17:54
Air Blank		0.000	17:54
Control Test		0.080	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEPHART,HUNTER,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KEPHART,HUNTER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CHAVIS

DATE/TIME OF TEST: 06/03/2021 03:56:26

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.080	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.281	03:51
Breath Volume		1.742	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.278	03:54
Breath Volume		1.921	03:54
Air Blank		0.000	03:55
Control Test		0.079	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: BAUERSCHMIDT

DATE/TIME OF TEST: 06/03/2021 15:00:42

Diagnostics Check		OK	14:54
Air Blank		0.000	14:54
Control Test		0.080	14:54
Air Blank		0.000	14:55
Subject Sample	# 1	0.000	14:56
Breath Volume		3.187	14:56
Air Blank		0.000	14:56
Air Blank		0.000	14:58
Subject Sample	# 2	0.000	14:59
Breath Volume		1.906	14:59
Air Blank		0.000	14:59
Control Test		0.080	15:00
Air Blank		0.000	15:00
Diagnostics Check		OK	15:00

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGANO,JORDAN,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BRAUN,DANIELLE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/05/2021 19:09:30

Diagnostics Check		OK	19:02
Air Blank		0.000	19:02
Control Test		0.080	19:03
Air Blank		0.000	19:03
Subject Sample	# 1	0.146	19:04
Breath Volume		2.074	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.135	19:07
Breath Volume		1.769	19:07
Air Blank		0.000	19:08
Control Test		0.079	19:08
Air Blank		0.000	19:09
Diagnostics Check		OK	19:09

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: SITZER

DATE/TIME OF TEST: 06/06/2021 01:15:01

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.179	01:09
Breath Volume		1.675	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:11
Subject Sample	# 2	0.170	01:13
Breath Volume		1.757	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:15

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/11/2021 00:59:43

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.080	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1	0.174	00:46	
Breath Volume		1.136	00:46	
Air Blank		0.000	00:47	
Air Blank		0.000	00:49	
Subject Sample	# 2	0.110	00:52	Volume Not Met
Breath Volume		0.128	00:52	
Air Blank		0.000	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 3	0.159	00:58	Volume Not Met
Breath Volume		0.859	00:58	
Air Blank		0.000	00:58	
Control Test		0.078	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RUSSO',CHRISTOPHER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/11/2021 01:15:52

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.079	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.159	01:04	
Breath Volume		1.589	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.109	01:10	Volume Not Met
Breath Volume		0.667	01:10	
Air Blank		0.000	01:10	
Air Blank		0.000	01:12	
Subject Sample	# 3	0.149	01:13	
Breath Volume		1.285	01:13	
Air Blank		0.000	01:14	
Control Test		0.077	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,CODY,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LISS,CODY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 06/12/2021 23:11:24

Diagnostics Check		OK	23:04
Air Blank		0.000	23:05
Control Test		0.081	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.209	23:06
Breath Volume		1.425	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.220	23:09
Breath Volume		1.929	23:09
Air Blank		0.000	23:10
Control Test		0.079	23:10
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODNOW,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: GOODNOW,TIMOTHY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: BOJORGE LOPEZ

DATE/TIME OF TEST: 06/13/2021 23:14:18

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.081	23:04
Air Blank		0.000	23:05
Subject Sample	# 1	0.081	23:07
Breath Volume		1.945	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:09
Subject Sample	# 2	0.073	23:12
Breath Volume		2.796	23:12
Air Blank		0.000	23:13
Control Test		0.078	23:13
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,HOLDEN,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: POWELL,HOLDEN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ARGAST

DATE/TIME OF TEST: 06/14/2021 02:32:26

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.101	02:27
Breath Volume		1.574	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.106	02:30
Breath Volume		1.636	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYES,JARED,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: HAYES,JARED
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: BUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: MALHEIRO

DATE/TIME OF TEST: 06/14/2021 19:53:36

Diagnostics Check		OK	19:46
Air Blank		0.000	19:47
Control Test		0.081	19:47
Air Blank		0.000	19:48
Subject Sample	# 1	0.157	19:49
Breath Volume		1.445	19:49
Air Blank		0.000	19:50
Air Blank		0.000	19:51
Subject Sample	# 2	0.158	19:51
Breath Volume		1.226	19:51
Air Blank		0.000	19:52
Control Test		0.078	19:52
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: THORSEN

DATE/TIME OF TEST: 06/15/2021 01:20:41

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.081	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.217	01:15
Breath Volume		2.214	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.204	01:18
Breath Volume		1.863	01:18
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORCHARD,DUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ORCHARD,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: OBANDO

DATE/TIME OF TEST: 06/16/2021 23:08:56

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.081	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.099	23:04
Breath Volume		1.289	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 2	0.098	23:07
Breath Volume		1.296	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,AUSTIN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COON,AUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:21

SUBJECT'S LAST NAME: BLEVINS

DATE/TIME OF TEST: 06/18/2021 10:47:30

Diagnostics Check		OK	10:41
Air Blank		0.000	10:41
Control Test		0.080	10:41
Air Blank		0.000	10:42
Subject Sample	# 1	0.293	10:42
Breath Volume		1.464	10:42
Air Blank		0.000	10:43
Air Blank		0.000	10:45
Subject Sample	# 2	0.301	10:45
Breath Volume		1.171	10:45
Air Blank		0.000	10:46
Control Test		0.079	10:46
Air Blank		0.000	10:47
Diagnostics Check		OK	10:47

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PARSON

DATE/TIME OF TEST: 06/20/2021 02:18:44

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.081	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 06/23/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,LANE,K
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: JACOBS,LANE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/24/2021 16:40:53

Diagnostics Check		OK	16:33
Air Blank		0.000	16:33
Control Test		0.080	16:34
Air Blank		0.000	16:34
Subject Sample	# 1	0.210	16:35
Breath Volume		3.035	16:35
Air Blank		0.000	16:36
Air Blank		0.000	16:38
Subject Sample	# 2	0.215	16:39
Breath Volume		2.082	16:39
Air Blank		0.000	16:39
Control Test		0.077	16:40
Air Blank		0.000	16:40
Diagnostics Check		OK	16:40

Data Download Date/Time: 07/12/2021 11:08

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECCUE,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BECCUE,JOSHUA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MUNROE

DATE/TIME OF TEST: 06/24/2021 20:09:10

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.078	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.134	20:04
Breath Volume		1.269	20:04
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.147	20:07
Breath Volume		1.105	20:07
Air Blank		0.000	20:08
Control Test		0.076	20:08
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 07/12/2021 11:08

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GARAY ABREGO

DATE/TIME OF TEST: 06/27/2021 01:01:34

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.204	00:55
Breath Volume		2.421	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.201	00:59
Breath Volume		3.140	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 07/12/2021 11:08

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: BALDEO

DATE/TIME OF TEST: 06/28/2021 17:29:15

Diagnostics Check		OK	17:22
Air Blank		0.000	17:23
Control Test		0.078	17:23
Air Blank		0.000	17:24
Subject Sample	# 1	0.000	17:24
Breath Volume		1.363	17:24
Air Blank		0.000	17:25
Air Blank		0.000	17:27
Subject Sample	# 2	0.000	17:27
Breath Volume		1.238	17:27
Air Blank		0.000	17:28
Control Test		0.077	17:28
Air Blank		0.000	17:29
Diagnostics Check		OK	17:29

Data Download Date/Time: 07/12/2021 11:08

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,KRISTAL,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: OSHIELDS,TY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 13:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 06/30/2021 13:56:05

Diagnostics Check		OK	13:50
Air Blank		0.000	13:50
Control Test		0.080	13:50
Air Blank		0.000	13:51
Subject Sample	# 1	0.000	13:51
Breath Volume		1.707	13:51
Air Blank		0.000	13:52
Air Blank		0.000	13:54
Subject Sample	# 2	0.000	13:54
Breath Volume		2.460	13:54
Air Blank		0.000	13:55
Control Test		0.079	13:55
Air Blank		0.000	13:55
Diagnostics Check		OK	13:56

Data Download Date/Time: 07/12/2021 11:08

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HAYWARD,KYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/01/2021 01:14:40

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.079	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: YOUNGBLOOD,MALLORY
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 06/01/2021 16:19:51

Diagnostics Check		OK	16:04	
Air Blank		0.000	16:05	
Control Test		0.078	16:05	
Air Blank		0.000	16:06	
Subject Sample	# 1	0.000	16:07	
Breath Volume		2.015	16:07	
Air Blank		0.000	16:07	
Air Blank		0.000	16:09	
Subject Sample	# 2	0.000	16:12	Volume Not Met
Breath Volume		0.472	16:12	
Air Blank		0.000	16:13	
Air Blank		0.000	16:15	
Subject Sample	# 3		16:18	No Sample Provided
Breath Volume		0.000	16:18	
Air Blank		0.000	16:18	
Control Test		0.078	16:19	
Air Blank		0.000	16:19	
Diagnostics Check		OK	16:19	

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 06/02/2021 03:02:06

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.136	02:57
Breath Volume		3.035	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.135	03:00
Breath Volume		3.078	03:00
Air Blank		0.000	03:01
Control Test		0.077	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WINTERS,E
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/02/2021 19:29:04

Diagnostics Check		OK	19:22
Air Blank		0.000	19:23
Control Test		0.079	19:23
Air Blank		0.000	19:24
Subject Sample	# 1	0.227	19:24
Breath Volume		2.710	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.225	19:27
Breath Volume		2.621	19:27
Air Blank		0.000	19:28
Control Test		0.077	19:28
Air Blank		0.000	19:28
Diagnostics Check		OK	19:29

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/03/2021 06:57:14

Diagnostics Check		OK	06:51
Air Blank		0.000	06:51
Control Test		0.078	06:51
Air Blank		0.000	06:52
Subject Sample	# 1	0.211	06:52
Breath Volume		2.503	06:52
Air Blank		0.000	06:53
Air Blank		0.000	06:55
Subject Sample	# 2	0.213	06:55
Breath Volume		2.722	06:55
Air Blank		0.000	06:56
Control Test		0.078	06:56
Air Blank		0.000	06:57
Diagnostics Check		OK	06:57

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: PARRETTE

DATE/TIME OF TEST: 06/05/2021 18:09:49

Diagnostics Check		OK	18:03
Air Blank		0.000	18:04
Control Test		0.078	18:04
Air Blank		0.000	18:05
Subject Sample	# 1	0.209	18:05
Breath Volume		2.917	18:05
Air Blank		0.000	18:06
Air Blank		0.000	18:07
Subject Sample	# 2	0.196	18:08
Breath Volume		1.953	18:08
Air Blank		0.000	18:08
Control Test		0.076	18:09
Air Blank		0.000	18:09
Diagnostics Check		OK	18:09

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: ACKISS

DATE/TIME OF TEST: 06/05/2021 22:09:41

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:07	
Control Test		0.078	22:07	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:08	Subject Test Refused
Breath Volume		0.000	22:08	
Air Blank		0.000	22:08	
Control Test		0.079	22:09	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: COLOGNE,B
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: DRAWDY

DATE/TIME OF TEST: 06/06/2021 00:34:11

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.078	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.000	00:29
Breath Volume		2.109	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.000	00:32
Breath Volume		2.074	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SATTERFIELD

DATE/TIME OF TEST: 06/08/2021 04:04:24

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.071	03:59
Breath Volume		2.113	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.075	04:02
Breath Volume		1.460	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 06/10/2021 19:46

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,JOHN,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 06/11/2021 00:15:01

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.078	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.194	00:10
Breath Volume		1.285	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:12
Subject Sample	# 2	0.209	00:13
Breath Volume		2.531	00:13
Air Blank		0.000	00:14
Control Test		0.077	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:15

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 06/11/2021 03:06:11

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.084	03:01
Breath Volume		1.468	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.082	03:04
Breath Volume		2.457	03:04
Air Blank		0.000	03:05
Control Test		0.077	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: MCHALE

DATE/TIME OF TEST: 06/16/2021 01:10:02

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.078	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.291	01:05
Breath Volume		2.101	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.279	01:08
Breath Volume		2.644	01:08
Air Blank		0.000	01:09
Control Test		0.077	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:10

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/16/2021 01:37:40

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.078	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.141	01:33
Breath Volume		1.667	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.141	01:36
Breath Volume		1.656	01:36
Air Blank		0.000	01:36
Control Test		0.078	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDLE,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: HARVEY,CALEB
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 06/17/2021 23:50:50

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.078	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.192	23:46
Breath Volume		1.511	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.184	23:49
Breath Volume		1.519	23:49
Air Blank		0.000	23:49
Control Test		0.076	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ROMEIN,C
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: THALER

DATE/TIME OF TEST: 06/21/2021 01:28:53

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.078	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.066	01:24
Breath Volume		1.773	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.072	01:27
Breath Volume		2.531	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KUKAR,ROBERT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FORREST JR

DATE/TIME OF TEST: 06/22/2021 00:41:42

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.076	00:31	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:34	No Sample Provided
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:37	
Subject Sample	# 2		00:40	No Sample Provided
Breath Volume		0.000	00:40	
Air Blank		0.000	00:40	
Control Test		0.079	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORALUZZO,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 06/22/2021 01:31:49

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.000	01:27
Breath Volume		2.699	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:30
Breath Volume		3.746	01:30
Air Blank		0.000	01:30
Control Test		0.078	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,INDIA,N
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GREEN,INDIA
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: SYS CHECK
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 06/24/2021 23:57:42

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.079	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.163	23:53
Breath Volume		1.683	23:53
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.168	23:56
Breath Volume		1.761	23:56
Air Blank		0.000	23:56
Control Test		0.078	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHROCK,L,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRIPLETT,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: BANTA

DATE/TIME OF TEST: 06/26/2021 21:23:44

Diagnostics Check		OK	21:17
Air Blank		0.000	21:17
Control Test		0.079	21:17
Air Blank		0.000	21:18
Subject Sample	# 1	0.165	21:19
Breath Volume		1.214	21:19
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.169	21:22
Breath Volume		1.882	21:22
Air Blank		0.000	21:22
Control Test		0.077	21:23
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 06/27/2021 04:08:05

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.144	04:03
Breath Volume		3.550	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.156	04:06
Breath Volume		2.707	04:06
Air Blank		0.000	04:07
Control Test		0.077	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 06/28/2021 21:35:50

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.079	21:29
Air Blank		0.000	21:30
Subject Sample	# 1	0.143	21:31
Breath Volume		3.925	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.162	21:34
Breath Volume		3.710	21:34
Air Blank		0.000	21:34
Control Test		0.078	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DELANEY,DANIEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GODIN

DATE/TIME OF TEST: 06/30/2021 03:06:24

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.161	03:01
Breath Volume		2.527	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.163	03:04
Breath Volume		2.562	03:04
Air Blank		0.000	03:05
Control Test		0.078	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 08/04/2021 15:39

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,JON,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GORDON,JON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ECHTERNACHT

DATE/TIME OF TEST: 06/18/2021 00:31:22

Diagnostics Check		OK	00:16	
Air Blank		0.000	00:16	
Control Test		0.077	00:17	
Air Blank		0.000	00:17	
Subject Sample	# 1	0.227	00:18	
Breath Volume		1.515	00:18	
Air Blank		0.000	00:19	
Air Blank		0.000	00:20	
Subject Sample	# 2	0.119	00:24	Volume Not Met
Breath Volume		0.132	00:24	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 3	0.195	00:29	Volume Not Met
Breath Volume		0.066	00:29	
Air Blank		0.000	00:30	
Control Test		0.070	00:30	Control Outside Tolerance
Air Blank		0.000	00:31	

Data Download Date/Time: 07/19/2021 22:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/10/2021 01:28:01

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.079	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.000	01:23
Breath Volume		1.636	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.000	01:26
Breath Volume		1.578	01:26
Air Blank		0.000	01:27
Control Test		0.078	01:27
Air Blank		0.000	01:27
Diagnostics Check		OK	01:28

Data Download Date/Time: 07/01/2021 16:50

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: CHESHIRE

DATE/TIME OF TEST: 06/10/2021 02:04:11

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.079	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1	0.362	01:59	
Breath Volume		2.207	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 2		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Control Test		0.078	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 07/01/2021 16:50

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DAVIS,MATTHEW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: ST FELIX

DATE/TIME OF TEST: 06/01/2021 06:16:21

Diagnostics Check		OK	05:59	
Air Blank		0.000	06:00	
Control Test		0.082	06:00	
Air Blank		0.000	06:01	
Subject Sample	# 1	0.202	06:04	Volume Not Met
Breath Volume		0.554	06:04	
Air Blank		0.000	06:05	
Air Blank		0.000	06:06	
Subject Sample	# 2	0.209	06:08	
Breath Volume		1.382	06:08	
Air Blank		0.000	06:09	
Air Blank		0.000	06:11	
Subject Sample	# 3	0.199	06:14	Volume Not Met
Breath Volume		0.585	06:14	
Air Blank		0.000	06:15	
Control Test		0.079	06:15	
Air Blank		0.000	06:16	
Diagnostics Check		OK	06:16	

Data Download Date/Time: 06/02/2021 13:07

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 06/07/2021 02:14:32

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.081	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.155	02:09
Breath Volume		1.269	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.163	02:12
Breath Volume		1.367	02:12
Air Blank		0.000	02:13
Control Test		0.077	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 07/14/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 06/08/2021 01:03:40

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.091	00:58
Breath Volume		2.335	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.092	01:01
Breath Volume		2.152	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/14/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRICENO,OSVALDO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 06/13/2021 07:24:49

Diagnostics Check		OK	07:17
Air Blank		0.000	07:18
Control Test		0.081	07:18
Air Blank		0.000	07:19
Subject Sample	# 1	0.110	07:19
Breath Volume		3.273	07:19
Air Blank		0.000	07:20
Air Blank		0.000	07:22
Subject Sample	# 2	0.105	07:23
Breath Volume		2.382	07:23
Air Blank		0.000	07:23
Control Test		0.081	07:24
Air Blank		0.000	07:24
Diagnostics Check		OK	07:24

Data Download Date/Time: 07/14/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 06/13/2021 23:46:19

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.121	23:41
Breath Volume		1.835	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.127	23:44
Breath Volume		2.945	23:44
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 07/14/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUERRA,ALVARO
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: GORDILLO

DATE/TIME OF TEST: 06/14/2021 00:41:57

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.082	00:29	
Air Blank		0.000	00:29	
Subject Sample	# 1	0.133	00:32	Volume Not Met
Breath Volume		0.195	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 2	0.144	00:36	
Breath Volume		1.851	00:36	
Air Blank		0.000	00:37	
Air Blank		0.000	00:39	
Subject Sample	# 3	0.137	00:40	
Breath Volume		1.164	00:40	
Air Blank		0.000	00:41	
Control Test		0.079	00:41	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 07/14/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 06/14/2021 01:24:49

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.080	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.176	01:20
Breath Volume		1.542	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.173	01:23
Breath Volume		1.503	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 07/14/2021 10:23

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HIGUITA,WILLIAM
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/15/2021 23:09:01

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.105	23:04
Breath Volume		2.742	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 2	0.105	23:07
Breath Volume		1.812	23:07
Air Blank		0.000	23:08
Control Test		0.081	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:09

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 06/15/2021 23:54:12

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.080	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.170	23:49
Breath Volume		3.378	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.159	23:52
Breath Volume		3.457	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: HERRERA GIRALDO

DATE/TIME OF TEST: 06/16/2021 04:19:30

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.176	04:14
Breath Volume		2.339	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.166	04:17
Breath Volume		2.289	04:17
Air Blank		0.000	04:18
Control Test		0.080	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/18/2021 20:11:45

Diagnostics Check		OK	20:05
Air Blank		0.000	20:06
Control Test		0.081	20:06
Air Blank		0.000	20:06
Subject Sample	# 1	0.000	20:07
Breath Volume		1.867	20:07
Air Blank		0.000	20:07
Air Blank		0.000	20:09
Subject Sample	# 2	0.000	20:10
Breath Volume		2.308	20:10
Air Blank		0.000	20:10
Control Test		0.081	20:11
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HARRIS JR

DATE/TIME OF TEST: 06/18/2021 23:02:46

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.080	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.189	22:58
Breath Volume		1.761	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.179	23:01
Breath Volume		1.519	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MENDOZA GUILLEN

DATE/TIME OF TEST: 06/19/2021 03:02:11

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.219	02:57
Breath Volume		2.738	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.214	03:00
Breath Volume		2.359	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/19/2021 03:49:54

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.081	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.089	03:45
Breath Volume		1.851	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.087	03:48
Breath Volume		1.656	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: YASKO,JOSEPH
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 06/19/2021 09:24:49

Diagnostics Check		OK	09:17
Air Blank		0.000	09:18
Control Test		0.082	09:18
Air Blank		0.000	09:19
Subject Sample	# 1	0.067	09:20
Breath Volume		3.101	09:20
Air Blank		0.000	09:20
Air Blank		0.000	09:22
Subject Sample	# 2	0.067	09:23
Breath Volume		2.769	09:23
Air Blank		0.000	09:23
Control Test		0.081	09:24
Air Blank		0.000	09:24
Diagnostics Check		OK	09:24

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROOTS,DIANTE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 08:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: DELANEY

DATE/TIME OF TEST: 06/19/2021 10:59:47

Diagnostics Check		OK	10:53
Air Blank		0.000	10:53
Control Test		0.081	10:53
Air Blank		0.000	10:54
Subject Sample	# 1	0.173	10:54
Breath Volume		1.996	10:54
Air Blank		0.000	10:55
Air Blank		0.000	10:57
Subject Sample	# 2	0.174	10:58
Breath Volume		1.835	10:58
Air Blank		0.000	10:58
Control Test		0.080	10:59
Air Blank		0.000	10:59
Diagnostics Check		OK	10:59

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SCHOENBORN

DATE/TIME OF TEST: 06/20/2021 02:55:05

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.081	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.000	02:50
Breath Volume		1.488	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.000	02:53
Breath Volume		1.289	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CHECCHI

DATE/TIME OF TEST: 06/20/2021 23:33:23

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.165	23:28
Breath Volume		1.210	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:30
Subject Sample	# 2	0.166	23:31
Breath Volume		2.261	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,DWAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 06/21/2021 01:37:57

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.081	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.150	01:33
Breath Volume		1.378	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.152	01:36
Breath Volume		2.515	01:36
Air Blank		0.000	01:37
Control Test		0.081	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMERO,ELISA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: JR. LEE

DATE/TIME OF TEST: 06/23/2021 09:04:43

Diagnostics Check		OK	08:57
Air Blank		0.000	08:58
Control Test		0.081	08:58
Air Blank		0.000	08:58
Subject Sample	# 1	0.102	09:00
Breath Volume		1.316	09:00
Air Blank		0.000	09:00
Air Blank		0.000	09:02
Subject Sample	# 2	0.101	09:03
Breath Volume		1.210	09:03
Air Blank		0.000	09:03
Control Test		0.080	09:04
Air Blank		0.000	09:04
Diagnostics Check		OK	09:04

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: COROMINAS,BENNY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: REY

DATE/TIME OF TEST: 06/24/2021 03:55:34

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.203	03:50
Breath Volume		1.652	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:52
Subject Sample	# 2	0.204	03:53
Breath Volume		2.078	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARDRADE,DAVID
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: ARCINIEGAS-VALDES

DATE/TIME OF TEST: 06/26/2021 07:24:42

Diagnostics Check		OK	07:18
Air Blank		0.000	07:18
Control Test		0.081	07:18
Air Blank		0.000	07:19
Subject Sample	# 1	0.168	07:19
Breath Volume		1.964	07:19
Air Blank		0.000	07:20
Air Blank		0.000	07:22
Subject Sample	# 2	0.169	07:23
Breath Volume		1.421	07:23
Air Blank		0.000	07:23
Control Test		0.080	07:24
Air Blank		0.000	07:24
Diagnostics Check		OK	07:24

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KRUMENACKER,MARK
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: HARNARRINE

DATE/TIME OF TEST: 06/26/2021 15:02:12

Diagnostics Check		OK	14:52
Air Blank		0.000	14:52
Control Test		0.081	14:52
Air Blank		0.000	14:53
Subject Sample	# 1	0.224	14:54
Breath Volume		1.519	14:54
Air Blank		0.000	14:55
Air Blank		0.000	14:57
Subject Sample	# 2	0.245	14:57
Breath Volume		1.367	14:57
Air Blank		0.000	14:58
Air Blank		0.000	15:00
Subject Sample	# 3	0.243	15:00
Breath Volume		1.515	15:00
Air Blank		0.000	15:01
Control Test		0.079	15:01
Air Blank		0.000	15:02
Diagnostics Check		OK	15:02

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,NORBENS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: HENRIQUE SANDOVAL

DATE/TIME OF TEST: 06/27/2021 06:14:51

Diagnostics Check		OK	06:08
Air Blank		0.000	06:08
Control Test		0.081	06:08
Air Blank		0.000	06:09
Subject Sample	# 1	0.240	06:10
Breath Volume		1.242	06:10
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.231	06:13
Breath Volume		1.304	06:13
Air Blank		0.000	06:13
Control Test		0.080	06:14
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 07/14/2021 10:24

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ACOSTA PENA

DATE/TIME OF TEST: 06/28/2021 02:30:42

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.140	02:25
Breath Volume		1.664	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.139	02:29
Breath Volume		2.121	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 07/14/2021 10:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRIGNONI,CRYSTAL
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: VON DEMBOWSKI

DATE/TIME OF TEST: 06/28/2021 10:20:37

Diagnostics Check		OK	10:13
Air Blank		0.000	10:14
Control Test		0.081	10:14
Air Blank		0.000	10:15
Subject Sample	# 1	0.342	10:15
Breath Volume		2.925	10:15
Air Blank		0.000	10:16
Air Blank		0.000	10:18
Subject Sample	# 2	0.334	10:18
Breath Volume		1.546	10:18
Air Blank		0.000	10:19
Control Test		0.079	10:19
Air Blank		0.000	10:20
Diagnostics Check		OK	10:20

Data Download Date/Time: 07/14/2021 10:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:19

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 06/28/2021 11:25:38

Diagnostics Check		OK	11:19
Air Blank		0.000	11:19
Control Test		0.081	11:19
Air Blank		0.000	11:20
Subject Sample	# 1	0.000	11:21
Breath Volume		2.035	11:21
Air Blank		0.000	11:21
Air Blank		0.000	11:23
Subject Sample	# 2	0.000	11:24
Breath Volume		2.570	11:24
Air Blank		0.000	11:24
Control Test		0.080	11:25
Air Blank		0.000	11:25
Diagnostics Check		OK	11:25

Data Download Date/Time: 07/14/2021 10:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,DANIEL,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HARMON,TYLER
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 06/01/2021 17:20:39

Diagnostics Check		OK	17:14
Air Blank		0.000	17:14
Control Test		0.080	17:15
Air Blank		0.000	17:15
Subject Sample	# 1	0.169	17:16
Breath Volume		1.160	17:16
Air Blank		0.000	17:16
Air Blank		0.000	17:18
Subject Sample	# 2	0.170	17:18
Breath Volume		1.437	17:18
Air Blank		0.000	17:19
Control Test		0.079	17:20
Air Blank		0.000	17:20
Diagnostics Check		OK	17:20

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,M,
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: LUCKETTE,J
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/08/2021 23:56:14

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.131	23:51
Breath Volume		2.503	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.128	23:54
Breath Volume		2.253	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEGEL,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIEGEL,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: NOAH

DATE/TIME OF TEST: 06/10/2021 18:17:16

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.080	18:11
Air Blank		0.000	18:11
Subject Sample	# 1	0.000	18:12
Breath Volume		4.355	18:12
Air Blank		0.000	18:12
Air Blank		0.000	18:14
Subject Sample	# 2	0.000	18:15
Breath Volume		4.367	18:15
Air Blank		0.000	18:16
Control Test		0.078	18:16
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 06/12/2021 00:57:22

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.075	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.090	00:52
Breath Volume		1.476	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.092	00:55
Breath Volume		1.437	00:55
Air Blank		0.000	00:56
Control Test		0.076	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIMMERMANN,JOHN,C
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: STOFEL,WESLEY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: JUSTICE

DATE/TIME OF TEST: 06/12/2021 18:55:43

Diagnostics Check		OK	18:48
Air Blank		0.000	18:49
Control Test		0.080	18:49
Air Blank		0.000	18:50
Subject Sample	# 1	0.117	18:50
Breath Volume		3.484	18:50
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.111	18:53
Breath Volume		2.273	18:53
Air Blank		0.000	18:54
Control Test		0.077	18:55
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAIG,SHAWN,T
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: CRAIG,SHAWN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 06/15/2021 12:24:55

Diagnostics Check		OK	12:17
Air Blank		0.000	12:18
Control Test		0.080	12:18
Air Blank		0.000	12:19
Subject Sample	# 1	0.265	12:19
Breath Volume		2.269	12:19
Air Blank		0.000	12:21
Air Blank		0.000	12:22
Subject Sample	# 2	0.265	12:23
Breath Volume		1.289	12:23
Air Blank		0.000	12:23
Control Test		0.077	12:24
Air Blank		0.000	12:24
Diagnostics Check		OK	12:24

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLSON,M,
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: HARMON,T
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/23/2021 00:46:08

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.090	00:41
Breath Volume		3.144	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.089	00:44
Breath Volume		3.164	00:44
Air Blank		0.000	00:45
Control Test		0.078	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAIG,SHAWN,T
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: CRAIG,SHAWN
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: VELASGUES

DATE/TIME OF TEST: 06/26/2021 13:32:41

Diagnostics Check		OK	13:24
Air Blank		0.000	13:25
Control Test		0.079	13:25
Air Blank		0.000	13:26
Subject Sample	# 1	0.255	13:27
Breath Volume		2.222	13:27
Air Blank		0.000	13:28
Air Blank		0.000	13:29
Subject Sample	# 2	0.253	13:30
Breath Volume		2.937	13:30
Air Blank		0.000	13:31
Control Test		0.077	13:32
Air Blank		0.000	13:32
Diagnostics Check		OK	13:32

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADFORD,CHARLES,T
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: STOFEL,WESLEY
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 06/27/2021 00:23:54

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.080	00:21	
Air Blank		0.000	00:22	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:22	
Control Test		0.080	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 07/01/2021 15:25

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/06/2021 01:38:38

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.097	01:34
Breath Volume		1.171	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.092	01:37
Breath Volume		1.320	01:37
Air Blank		0.000	01:37
Control Test		0.077	01:38
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 07/14/2021 11:22

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/06/2021 02:09:12

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.171	02:04
Breath Volume		1.281	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.178	02:07
Breath Volume		1.578	02:07
Air Blank		0.000	02:08
Control Test		0.078	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 07/14/2021 11:22

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VALENTIN,ENRIQUE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: RIVERA DIAZ

DATE/TIME OF TEST: 06/17/2021 17:14:26

Diagnostics Check		OK	17:08
Air Blank		0.000	17:08
Control Test		0.080	17:09
Air Blank		0.000	17:09
Subject Sample	# 1	0.000	17:10
Breath Volume		1.855	17:10
Air Blank		0.000	17:10
Air Blank		0.000	17:12
Subject Sample	# 2	0.000	17:13
Breath Volume		2.105	17:13
Air Blank		0.000	17:13
Control Test		0.081	17:13
Air Blank		0.000	17:14
Diagnostics Check		OK	17:14

Data Download Date/Time: 07/14/2021 11:22

Printed: 22 September 2021

COBRA ©1997-2014 CMI, Inc.