

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: RAKOVSKY

DATE/TIME OF TEST: 07/01/2021 01:39:52

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.078	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2	01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Control Test	0.078	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 07/14/2021 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: XAVIER,STEEVE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: XAVIER,STEEVE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GANTT JR

DATE/TIME OF TEST: 07/03/2021 00:53:37

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.115	00:48
Breath Volume	2.082	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.110	00:52
Breath Volume	2.515	00:52
Air Blank	0.000	00:52
Control Test	0.076	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 07/14/2021 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 07/03/2021 08:07:09

Diagnostics Check	Fail	08:06	Voltage/Current Test
Air Blank	0.000	08:07	

Data Download Date/Time: 07/14/2021 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORIAN,ROLANDO,I
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: FLORIAN,ROLANDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SEQUEIRA

DATE/TIME OF TEST: 07/26/2021 03:29:01

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.193	03:23
Breath Volume	1.218	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.210	03:27
Breath Volume	1.328	03:27
Air Blank	0.000	03:28
Control Test	0.077	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:29

Data Download Date/Time: 08/06/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,G
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/29/2021 18:48:38

Diagnostics Check	OK	18:39	
Air Blank	0.000	18:39	
Control Test	0.075	18:39	
Air Blank	0.000	18:40	
Subject Sample	# 1 0.229	18:43	Volume Not Met
Breath Volume	0.250	18:43	
Air Blank	0.000	18:44	
Air Blank	0.000	18:46	
Subject Sample	# 2	18:47	Sequence Aborted
Breath Volume	0.160	18:47	
Air Blank	0.000	18:48	

Data Download Date/Time: 08/06/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,G
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/29/2021 18:53:03

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:52	
Control Test	0.074	18:52	Control Outside Tolerance
Air Blank	0.000	18:53	

Data Download Date/Time: 08/06/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,G
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/29/2021 19:01:47

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:58	
Control Test	0.076	18:58	
Air Blank	0.000	18:59	
Subject Sample	# 1	19:00	Sequence Aborted
Breath Volume	0.015	19:00	
Air Blank	0.000	19:01	

Data Download Date/Time: 08/06/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,G
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/29/2021 19:07:48

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:06	
Control Test	0.074	19:07	Control Outside Tolerance
Air Blank	0.000	19:07	

Data Download Date/Time: 08/06/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,G
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/29/2021 19:16:55

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:12	
Control Test	0.076	19:13	
Air Blank	0.000	19:13	
Subject Sample	# 1	19:15	Subject Test Refused
Breath Volume	0.480	19:15	
Air Blank	0.000	19:16	
Control Test	0.077	19:16	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:16	

Data Download Date/Time: 08/06/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA, TERESITA, D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DIEUDONNE, K
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: CIRREA JR

DATE/TIME OF TEST: 07/03/2021 08:00:03

Diagnostics Check	OK	07:44	
Air Blank	0.000	07:45	
Control Test	0.080	07:45	
Air Blank	0.000	07:46	
Subject Sample # 1	0.152	07:46	
Breath Volume	2.230	07:46	
Air Blank	0.000	07:47	
Air Blank	0.000	07:49	
Subject Sample # 2		07:52	No Sample Provided
Breath Volume	0.000	07:52	
Air Blank	0.000	07:52	
Air Blank	0.000	07:54	
Subject Sample # 3	0.126	07:58	Volume Not Met
Breath Volume	0.417	07:58	
Air Blank	0.000	07:59	
Control Test	0.077	07:59	
Air Blank	0.000	07:59	
Diagnostics Check	OK	08:00	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA, TERESITA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DIEUDONNE, K
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: CORREA JR

DATE/TIME OF TEST: 07/03/2021 08:23:52

Diagnostics Check	OK	08:09	
Air Blank	0.000	08:10	
Control Test	0.075	08:10	
Air Blank	0.000	08:11	
Subject Sample # 1	0.147	08:13	
Breath Volume	1.414	08:13	
Air Blank	0.000	08:14	
Air Blank	0.000	08:15	
Subject Sample # 2	0.126	08:18	Volume Not Met
Breath Volume	0.382	08:18	
Air Blank	0.000	08:19	
Air Blank	0.000	08:21	
Subject Sample # 3	0.138	08:22	
Breath Volume	1.437	08:22	
Air Blank	0.000	08:22	
Control Test	0.073	08:23	Control Outside Tolerance
Air Blank	0.000	08:23	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 07/03/2021 09:07:52

Diagnostics Check		OK	09:01
Air Blank		0.000	09:01
Control Test		0.079	09:02
Air Blank		0.000	09:02
Subject Sample	# 1	0.109	09:03
Breath Volume		1.292	09:03
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.110	09:06
Breath Volume		1.703	09:06
Air Blank		0.000	09:06
Control Test		0.078	09:07
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MORENO,FRANCISCO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: BONWIT

DATE/TIME OF TEST: 07/03/2021 18:10:07

Diagnostics Check	OK	18:07	
Air Blank	0.000	18:07	
Control Test	0.081	18:07	
Air Blank	0.000	18:08	
Subject Sample	# 1	18:08	Subject Test Refused
Breath Volume	0.000	18:08	
Air Blank	0.000	18:09	
Control Test	0.080	18:09	
Air Blank	0.000	18:10	
Diagnostics Check	OK	18:10	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDERO,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PLANAS,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 07/03/2021 22:16:19

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.165	22:11
Breath Volume	2.621	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 2	0.162	22:14
Breath Volume	2.316	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS,WALEX,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOUIS,WALEX
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/04/2021 00:13:20

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.080	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1 0.254	00:12	
Breath Volume	1.867	00:12	
Air Blank		00:12	Purge Fail
Air Blank	0.000	00:13	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/04/2021 05:34:39

Diagnostics Check		OK	05:24
Air Blank		0.000	05:25
Control Test		0.080	05:25
Air Blank		0.000	05:26
Subject Sample	# 1	0.340	05:26
Breath Volume		2.019	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.370	05:29
Breath Volume		3.265	05:29
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 3	0.360	05:32
Breath Volume		2.355	05:32
Air Blank		0.000	05:33
Control Test		0.079	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CRUZ,DANIEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/05/2021 09:52:17

Diagnostics Check	OK	09:45
Air Blank	0.000	09:46
Control Test	0.079	09:46
Air Blank	0.000	09:47
Subject Sample # 1	0.072	09:47
Breath Volume	1.707	09:47
Air Blank	0.000	09:48
Air Blank	0.000	09:50
Subject Sample # 2	0.074	09:50
Breath Volume	1.785	09:50
Air Blank	0.000	09:51
Control Test	0.079	09:51
Air Blank	0.000	09:52
Diagnostics Check	OK	09:52

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: EGUSQUIZA

DATE/TIME OF TEST: 07/09/2021 00:22:30

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:18	No Sample Provided
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample	# 2	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 07/11/2021 00:56:39

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.080	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1	0.245	00:49	Volume Not Met
Breath Volume		0.664	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:51	
Subject Sample	# 2	0.237	00:54	Volume Not Met
Breath Volume		0.019	00:54	
Air Blank		0.000	00:55	
Control Test		0.078	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 07/11/2021 01:03:54

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.079	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 1 0.191	01:02	Volume Not Met
Breath Volume	0.070	01:02	
Air Blank		01:03	Purge Fail
Air Blank	0.000	01:03	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 07/11/2021 01:21:38

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:10	
Control Test		0.077	01:10	
Air Blank		0.000	01:10	
Subject Sample	# 1	0.138	01:14	Volume Not Met
Breath Volume		0.148	01:14	
Air Blank		0.000	01:14	
Air Blank		0.000	01:16	
Subject Sample	# 2	0.179	01:19	Volume Not Met
Breath Volume		0.066	01:19	
Air Blank		0.000	01:20	
Control Test		0.078	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: XAVIER,STEEVE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: XAVIER,STEEVE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/11/2021 01:43:02

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.079	01:42	
Air Blank		01:42	RFI Detect
Air Blank	0.000	01:43	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILIAN,WANDA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LLOSA,JOSEPH
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: BACA

DATE/TIME OF TEST: 07/11/2021 19:08:28

Diagnostics Check		OK	18:58
Air Blank		0.000	18:58
Control Test		0.080	18:59
Air Blank		0.000	18:59
Subject Sample	# 1	0.310	19:00
Breath Volume		1.105	19:00
Air Blank		0.000	19:01
Air Blank		0.000	19:02
Subject Sample	# 2	0.288	19:03
Breath Volume		1.160	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:05
Subject Sample	# 3	0.295	19:06
Breath Volume		1.437	19:06
Air Blank		0.000	19:07
Control Test		0.076	19:07
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILIAN,WANDA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DORCELY,SAMUEL
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ROSALES HUECK

DATE/TIME OF TEST: 07/11/2021 19:32:07

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.281	19:27
Breath Volume	2.406	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:29
Subject Sample # 2	0.272	19:30
Breath Volume	2.464	19:30
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 07/24/2021 03:21:48

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.080	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,NATALIE,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: RIVERO,NATALIE
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/27/2021 03:02:52

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEY, LAZARO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: COURT ORD
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/28/2021 11:11:23

Diagnostics Check	OK	11:05
Air Blank	0.000	11:05
Control Test	0.081	11:06
Air Blank	0.000	11:06
Subject Sample # 1	0.000	11:06
Breath Volume	1.765	11:06
Air Blank	0.000	11:07
Air Blank	0.000	11:09
Subject Sample # 2	0.000	11:09
Breath Volume	1.492	11:09
Air Blank	0.000	11:10
Control Test	0.080	11:10
Air Blank	0.000	11:11
Diagnostics Check	OK	11:11

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LEY, LAZARO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: COURT ORD
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/29/2021 11:08:53

Diagnostics Check	OK	11:02
Air Blank	0.000	11:03
Control Test	0.080	11:03
Air Blank	0.000	11:03
Subject Sample # 1	0.000	11:04
Breath Volume	1.878	11:04
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.000	11:07
Breath Volume	1.480	11:07
Air Blank	0.000	11:07
Control Test	0.081	11:08
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 08/19/2021 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 07/01/2021 20:10:14

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:07	
Control Test	0.079	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.078	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 07/02/2021 01:53:23

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.079	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ZIVALJEVIC

DATE/TIME OF TEST: 07/02/2021 23:56:07

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: TAANK

DATE/TIME OF TEST: 07/03/2021 02:05:01

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.049	02:00
Breath Volume	1.843	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.050	02:03
Breath Volume	2.023	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:05

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO, E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 07/03/2021 23:30:56

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:28	
Control Test	0.079	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:30	
Control Test	0.078	23:30	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO, E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MALDONADO MANCIA

DATE/TIME OF TEST: 07/04/2021 03:12:12

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPURRIER, J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: COSAR

DATE/TIME OF TEST: 07/04/2021 06:10:34

Diagnostics Check	OK	06:07	
Air Blank	0.000	06:08	
Control Test	0.080	06:08	
Air Blank	0.000	06:09	
Subject Sample	# 1	06:09	Subject Test Refused
Breath Volume	0.000	06:09	
Air Blank	0.000	06:09	
Control Test	0.079	06:09	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:10	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPLATT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: PHELPS JR

DATE/TIME OF TEST: 07/04/2021 16:06:19

Diagnostics Check	OK	16:03	
Air Blank	0.000	16:03	
Control Test	0.078	16:04	
Air Blank	0.000	16:04	
Subject Sample	# 1	16:04	Subject Test Refused
Breath Volume	0.000	16:04	
Air Blank	0.000	16:05	
Control Test	0.079	16:05	
Air Blank	0.000	16:06	
Diagnostics Check	OK	16:06	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 07/04/2021 21:33:00

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:30	
Control Test	0.079	21:30	
Air Blank	0.000	21:31	
Subject Sample	# 1	21:31	Subject Test Refused
Breath Volume	0.000	21:31	
Air Blank	0.000	21:32	
Control Test	0.079	21:32	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: WDL61N4D823B

DATE/TIME OF TEST: 07/05/2021 00:30:06

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.156	00:25
Breath Volume	1.800	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.155	00:28
Breath Volume	1.847	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 07/15/2021 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SLIGH

DATE/TIME OF TEST: 07/05/2021 03:11:35

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPURRIER,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/05/2021 05:32:05

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:29	
Control Test	0.079	05:29	
Air Blank	0.000	05:30	
Subject Sample	# 1	05:30	Subject Test Refused
Breath Volume	0.000	05:30	
Air Blank	0.000	05:31	
Control Test	0.079	05:31	
Air Blank	0.000	05:31	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: KELLEHER

DATE/TIME OF TEST: 07/06/2021 00:46:35

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.210	00:41
Breath Volume		1.882	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.200	00:44
Breath Volume		2.753	00:44
Air Blank		0.000	00:45
Control Test		0.078	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DADAMO

DATE/TIME OF TEST: 07/06/2021 01:44:04

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.000	01:39
Breath Volume	4.023	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.000	01:42
Breath Volume	3.402	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:44

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 07/06/2021 04:28:11

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.230	04:23
Breath Volume	2.644	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.228	04:26
Breath Volume	3.593	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:59

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/06/2021 14:26:05

Diagnostics Check	OK	14:19
Air Blank	0.000	14:20
Control Test	0.078	14:20
Air Blank	0.000	14:21
Subject Sample # 1	0.182	14:21
Breath Volume	4.113	14:21
Air Blank	0.000	14:22
Air Blank	0.000	14:24
Subject Sample # 2	0.196	14:24
Breath Volume	4.019	14:24
Air Blank	0.000	14:25
Control Test	0.078	14:25
Air Blank	0.000	14:25
Diagnostics Check	OK	14:26

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 07/06/2021 21:28:03

Diagnostics Check	OK	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.000	21:23
Breath Volume	1.613	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.000	21:26
Breath Volume	1.507	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:28

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/07/2021 04:58:57

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.079	04:56	
Air Blank	0.000	04:57	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:58	
Control Test	0.079	04:58	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 07/09/2021 05:55:16

Diagnostics Check		OK	05:48
Air Blank		0.000	05:49
Control Test		0.079	05:49
Air Blank		0.000	05:50
Subject Sample	# 1	0.162	05:50
Breath Volume		1.832	05:50
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 2	0.166	05:53
Breath Volume		1.921	05:53
Air Blank		0.000	05:54
Control Test		0.079	05:54
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROBES,S,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BARBOUR

DATE/TIME OF TEST: 07/15/2021 01:17:27

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.180	01:13
Breath Volume		2.019	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.181	01:15
Breath Volume		2.320	01:15
Air Blank		0.000	01:16
Control Test		0.078	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: MIRNA

DATE/TIME OF TEST: 07/15/2021 02:42:00

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROBES,S,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/15/2021 05:16:20

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.171	05:11
Breath Volume	1.492	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.173	05:14
Breath Volume	1.429	05:14
Air Blank	0.000	05:15
Control Test	0.078	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 07/15/2021 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROBES,S,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSON,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MATHEWS

DATE/TIME OF TEST: 07/16/2021 00:58:34

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.079	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTHRAN,D,A
BREATH TEST OPERATOR AGENCY: Neptune Beach PD
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 07/16/2021 02:40:25

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.194	02:34
Breath Volume	1.636	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.189	02:38
Breath Volume	1.246	02:38
Air Blank	0.000	02:39
Control Test	0.077	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROBES,S,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 07/16/2021 03:25:26

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.083	03:20
Breath Volume	1.191	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.085	03:23
Breath Volume	1.261	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTHRAN,D,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 07/17/2021 02:42:20

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.221	02:36
Breath Volume	2.386	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.212	02:40
Breath Volume	1.667	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HARTJE

DATE/TIME OF TEST: 07/18/2021 01:01:36

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.079	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: ESPERANZA

DATE/TIME OF TEST: 07/19/2021 02:29:22

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: CARLTON

DATE/TIME OF TEST: 07/20/2021 03:13:44

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.179	03:09
Breath Volume	2.894	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.185	03:12
Breath Volume	2.277	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/22/2021 20:21:05

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.079	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.000	20:16
Breath Volume	2.996	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.000	20:19
Breath Volume	2.933	20:19
Air Blank	0.000	20:20
Control Test	0.079	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: GUTIEREZ

DATE/TIME OF TEST: 07/23/2021 03:05:09

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.080	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1	0.178	02:55	
Breath Volume		1.171	02:55	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2	0.169	03:00	Volume Not Met
Breath Volume		0.082	03:00	
Air Blank		0.000	03:01	
Air Blank		0.000	03:03	
Subject Sample	# 3	0.183	03:03	
Breath Volume		1.515	03:03	
Air Blank		0.000	03:04	
Control Test		0.077	03:04	
Air Blank		0.000	03:05	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: DRINNON

DATE/TIME OF TEST: 07/25/2021 01:43:19

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.032	01:38
Breath Volume	3.011	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.032	01:41
Breath Volume	2.976	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/25/2021 04:29:31

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.128	04:25
Breath Volume	2.878	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.123	04:28
Breath Volume	1.519	04:28
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMP,W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: GEREN

DATE/TIME OF TEST: 07/25/2021 23:49:53

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:48	
Control Test	0.080	23:49	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MALDONADO ACOSTA

DATE/TIME OF TEST: 07/26/2021 01:30:02

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.173	01:25
Breath Volume	1.863	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.167	01:28
Breath Volume	2.015	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/26/2021 04:10:04

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:08	Subject Test Refused
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: ORELLANA ENAMORADO

DATE/TIME OF TEST: 07/26/2021 20:57:56

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.170	20:53
Breath Volume	1.546	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.170	20:56
Breath Volume	1.730	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 07/29/2021 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: PLACE

DATE/TIME OF TEST: 07/27/2021 02:37:22

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.240	02:31
Breath Volume		1.210	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.220	02:35
Breath Volume		3.050	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 07/29/2021 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BOATENG

DATE/TIME OF TEST: 07/28/2021 01:40:27

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.232	01:35
Breath Volume	1.523	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.220	01:38
Breath Volume	1.636	01:38
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 07/29/2021 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/28/2021 05:15:54

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.079	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.079	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 07/29/2021 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/04/2021 18:18:30

Diagnostics Check	OK	18:15	
Air Blank	0.000	18:16	
Control Test	0.080	18:16	
Air Blank	0.000	18:16	
Subject Sample	# 1	18:17	Subject Test Refused
Breath Volume	0.000	18:17	
Air Blank	0.000	18:17	
Control Test	0.081	18:17	
Air Blank	0.000	18:18	
Diagnostics Check	OK	18:18	

Data Download Date/Time: 07/16/2021 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CANTY,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/11/2021 04:04:39

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.000	04:00
Breath Volume	3.714	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.000	04:03
Breath Volume	3.480	04:03
Air Blank	0.000	04:03
Control Test	0.084	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 07/16/2021 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ABRANTE

DATE/TIME OF TEST: 07/12/2021 03:45:13

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 07/16/2021 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: ASKEW

DATE/TIME OF TEST: 07/14/2021 20:09:14

Diagnostics Check		OK	20:03
Air Blank		0.000	20:03
Control Test		0.081	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.281	20:04
Breath Volume		2.722	20:04
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 2	0.272	20:07
Breath Volume		2.609	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 07/16/2021 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 07/15/2021 01:21:08

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.230	01:16
Breath Volume		2.949	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.230	01:19
Breath Volume		2.304	01:19
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 07/16/2021 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: LOFFER

DATE/TIME OF TEST: 07/15/2021 02:50:50

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 07/16/2021 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 07/15/2021 21:33:23

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.082	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.184	21:28
Breath Volume	2.394	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.168	21:31
Breath Volume	2.304	21:31
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 07/16/2021 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESCOTT,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GIBSON,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 07/16/2021 01:27:10

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.211	01:22
Breath Volume		1.246	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.226	01:25
Breath Volume		2.191	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 07/16/2021 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 07/17/2021 02:32:04

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.163	02:27
Breath Volume		4.292	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.157	02:30
Breath Volume		4.222	02:30
Air Blank		0.000	02:31
Control Test		0.083	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:32

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JENSEN,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: COPTI GASPARATO

DATE/TIME OF TEST: 07/17/2021 04:31:18

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.234	04:26
Breath Volume	2.312	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.237	04:29
Breath Volume	2.144	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HEATHCO

DATE/TIME OF TEST: 07/17/2021 05:13:02

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:10	
Control Test	0.081	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Control Test	0.081	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 07/18/2021 03:19:56

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.082	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.000	03:15
Breath Volume	2.316	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.000	03:18
Breath Volume	2.406	03:18
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,R,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 07/22/2021 03:28:38

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Control Test	0.080	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 07/24/2021 04:08:03

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.068	04:02
Breath Volume		1.230	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.062	04:06
Breath Volume		1.222	04:06
Air Blank		0.000	04:07
Control Test		0.081	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:08

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPHERD,P,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IRRA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/24/2021 09:18:30

Diagnostics Check	OK	09:15	
Air Blank	0.000	09:16	
Control Test	0.080	09:16	
Air Blank	0.000	09:16	
Subject Sample	# 1	09:17	Subject Test Refused
Breath Volume	0.000	09:17	
Air Blank	0.000	09:17	
Control Test	0.081	09:17	
Air Blank	0.000	09:18	
Diagnostics Check	OK	09:18	

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MORALES-ROSARIO

DATE/TIME OF TEST: 07/25/2021 02:32:35

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.105	02:28
Breath Volume	2.351	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.105	02:31
Breath Volume	1.632	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PARRISH,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: TUOT

DATE/TIME OF TEST: 07/25/2021 04:44:11

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.151	04:39
Breath Volume	1.261	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.148	04:42
Breath Volume	1.304	04:42
Air Blank	0.000	04:43
Control Test	0.081	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBY,C,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: HOOKS

DATE/TIME OF TEST: 07/25/2021 05:28:19

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.170	05:23
Breath Volume	1.503	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.171	05:26
Breath Volume	1.425	05:26
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: O'CONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: VALLIERE

DATE/TIME OF TEST: 07/27/2021 17:06:46

Diagnostics Check	OK	17:00
Air Blank	0.000	17:00
Control Test	0.080	17:00
Air Blank	0.000	17:01
Subject Sample # 1	0.149	17:02
Breath Volume	1.960	17:02
Air Blank	0.000	17:02
Air Blank	0.000	17:04
Subject Sample # 2	0.150	17:05
Breath Volume	1.343	17:05
Air Blank	0.000	17:05
Control Test	0.080	17:06
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,R,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAWSON,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 07/27/2021 18:27:37

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.080	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.136	18:23
Breath Volume	1.613	18:23
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.135	18:26
Breath Volume	1.589	18:26
Air Blank	0.000	18:26
Control Test	0.080	18:27
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 07/29/2021 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JAKE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS,JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: TORRE ROJAS

DATE/TIME OF TEST: 07/02/2021 19:23:53

Diagnostics Check	OK	19:23	
Air Blank		19:23	RFI Detect
Air Blank	0.000	19:23	

Data Download Date/Time: 08/06/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS, JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: TORRE ROJAS

DATE/TIME OF TEST: 07/02/2021 19:31:11

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.116	19:26
Breath Volume	1.625	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.110	19:29
Breath Volume	1.289	19:29
Air Blank	0.000	19:30
Control Test	0.079	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 08/06/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JAIRUS,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALLEN,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DESSERT

DATE/TIME OF TEST: 07/06/2021 15:34:48

Diagnostics Check	OK	15:32	
Air Blank	0.000	15:32	
Control Test	0.080	15:32	
Air Blank	0.000	15:33	
Subject Sample	# 1	15:33	Subject Test Refused
Breath Volume	0.000	15:33	
Air Blank	0.000	15:33	
Control Test	0.079	15:34	
Air Blank	0.000	15:34	
Diagnostics Check	OK	15:34	

Data Download Date/Time: 08/06/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 07/30/2021 22:13:39

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.076	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.181	22:05
Breath Volume		1.707	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.157	22:08
Breath Volume		1.351	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 3	0.150	22:12
Breath Volume		1.800	22:12
Air Blank		0.000	22:12
Control Test		0.079	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 08/06/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: LEWIS,L
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: OLIVER RAWLS

DATE/TIME OF TEST: 07/10/2021 02:06:38

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.282	02:02
Breath Volume		1.582	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.285	02:05
Breath Volume		1.726	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 09/26/2021 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSHAUGHNESSY,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MCNEAL,M
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DIAZ GARCIA

DATE/TIME OF TEST: 07/02/2021 01:06:06

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.081	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:04	Subject Test Refused
Breath Volume	0.058	01:04	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 09/26/2021 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALLEN,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 07/09/2021 17:31:36

Diagnostics Check	OK	17:25
Air Blank	0.000	17:25
Control Test	0.075	17:26
Air Blank	0.000	17:26
Subject Sample # 1	0.000	17:27
Breath Volume	2.933	17:27
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 2	0.000	17:30
Breath Volume	2.687	17:30
Air Blank	0.000	17:30
Control Test	0.075	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 09/26/2021 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN, JOSHUA, B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALLEN, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/14/2021 16:11:55

Diagnostics Check	OK	16:05
Air Blank	0.000	16:06
Control Test	0.076	16:06
Air Blank	0.000	16:07
Subject Sample # 1	0.000	16:07
Breath Volume	1.921	16:07
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.000	16:10
Breath Volume	2.117	16:10
Air Blank	0.000	16:11
Control Test	0.076	16:11
Air Blank	0.000	16:11
Diagnostics Check	OK	16:11

Data Download Date/Time: 09/26/2021 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWM,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG000606
DRY GAS EXPIRY DATE: 01/06/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SECUNDINO

DATE/TIME OF TEST: 07/16/2021 00:16:47

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.077	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.134	00:12
Breath Volume		2.136	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.139	00:15
Breath Volume		1.453	00:15
Air Blank		0.000	00:15
Control Test		0.078	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 09/26/2021 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/30/2021 06:46:09

Diagnostics Check	OK	06:39
Air Blank	0.000	06:39
Control Test	0.080	06:40
Air Blank	0.000	06:40
Subject Sample # 1	0.173	06:41
Breath Volume	3.238	06:41
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.179	06:44
Breath Volume	3.468	06:44
Air Blank	0.000	06:45
Control Test	0.080	06:45
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 08/02/2021 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MCGLONE

DATE/TIME OF TEST: 07/31/2021 00:27:01

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.000	00:22
Breath Volume	1.734	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.000	00:25
Breath Volume	3.441	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:27

Data Download Date/Time: 08/02/2021 17:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: DRIVER

DATE/TIME OF TEST: 07/31/2021 04:20:32

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.081	04:16
Breath Volume	2.683	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.080	04:19
Breath Volume	4.031	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 08/02/2021 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BLACK,KURT
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LEAL-FERNANDEZ

DATE/TIME OF TEST: 07/01/2021 20:11:55

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:01	
Control Test	0.082	20:01	
Air Blank	0.000	20:01	
Subject Sample	# 1	20:04	No Sample Provided
Breath Volume	0.000	20:04	
Air Blank	0.000	20:05	
Air Blank	0.000	20:07	
Subject Sample	# 2	20:10	No Sample Provided
Breath Volume	0.000	20:10	
Air Blank	0.000	20:11	
Control Test	0.081	20:11	
Air Blank	0.000	20:11	
Diagnostics Check	OK	20:11	

Data Download Date/Time: 08/02/2021 17:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LEVIER,MATTHEW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 07/03/2021 02:26:46

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.082	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.077	02:22
Breath Volume	3.566	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.074	02:25
Breath Volume	3.648	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/02/2021 17:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: BROCKWELL

DATE/TIME OF TEST: 07/04/2021 01:15:26

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.172	01:10
Breath Volume	1.167	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.165	01:13
Breath Volume	1.968	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: EXLINE

DATE/TIME OF TEST: 07/04/2021 20:26:24

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:24	
Control Test	0.084	20:25	
Air Blank		20:25	Purge Fail
Air Blank	0.000	20:26	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: BAILLARGEON

DATE/TIME OF TEST: 07/04/2021 21:27:05

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.081	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.101	21:22
Breath Volume	1.785	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.099	21:25
Breath Volume	1.585	21:25
Air Blank	0.000	21:26
Control Test	0.083	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 07/04/2021 22:23:25

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.083	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.081	22:22	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAHL,RANDY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DAHL,RANDY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/05/2021 02:31:55

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.244	02:26
Breath Volume		2.031	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.247	02:30
Breath Volume		1.191	02:30
Air Blank		0.000	02:30
Control Test		0.081	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 07/08/2021 04:00:01

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.214	03:55
Breath Volume	2.316	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.214	03:58
Breath Volume	1.789	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERA,DANTE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BURROUGHS

DATE/TIME OF TEST: 07/09/2021 00:52:58

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.082	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,VIRGINIA,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBERSON,VIRGINIA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ICAL

DATE/TIME OF TEST: 07/09/2021 01:40:48

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.082	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample	# 2	0.109	01:36
Breath Volume	1.308	01:36	
Air Blank	0.000	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 3	0.107	01:39
Breath Volume	2.062	01:39	
Air Blank	0.000	01:39	
Control Test	0.078	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: NORWOOD

DATE/TIME OF TEST: 07/09/2021 22:45:03

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.081	22:34	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:38	No Sample Provided
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Air Blank	0.000	22:40	
Subject Sample	# 2	22:43	No Sample Provided
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Control Test	0.080	22:44	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ROBINS II,JERRY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LUKER

DATE/TIME OF TEST: 07/11/2021 02:18:10

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.081	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: FRATE

DATE/TIME OF TEST: 07/13/2021 22:52:09

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.082	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Control Test	0.081	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HALFHILL,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 07/15/2021 23:26:32

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:25	Subject Test Refused
Breath Volume	0.000	23:25	
Air Blank	0.000	23:25	
Control Test	0.081	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 08/02/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/16/2021 01:29:20

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.081	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:25
Breath Volume	1.921	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:27
Breath Volume	2.281	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/16/2021 23:16:58

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.082	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:12	No Sample Provided
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample	# 2	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Control Test	0.081	23:16	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: BIGELOW

DATE/TIME OF TEST: 07/17/2021 19:22:03

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:19	
Control Test	0.082	19:19	
Air Blank	0.000	19:20	
Subject Sample	# 1	19:20	Subject Test Refused
Breath Volume	0.000	19:20	
Air Blank	0.000	19:21	
Control Test	0.081	19:21	
Air Blank	0.000	19:21	
Diagnostics Check	OK	19:22	

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JAMES,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: REAVES,CHRISTINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: FORE

DATE/TIME OF TEST: 07/18/2021 02:30:23

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.082	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 07/18/2021 22:52:11

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:45	
Air Blank	0.000	22:46	
Subject Sample	# 1 0.190	22:47	
Breath Volume	2.718	22:47	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 2 0.179	22:50	
Breath Volume	1.714	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	Fail	22:51	Analytical Stability Test
Air Blank	0.000	22:52	

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSMAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FORSMAN,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/19/2021 02:29:34

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.081	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 07/21/2021 00:43:10

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.082	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.082	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/22/2021 02:18:55

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:14
Breath Volume	3.851	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:17
Breath Volume	3.964	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURST,CARY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HURST,CARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: BECKLUND

DATE/TIME OF TEST: 07/23/2021 19:11:35

Diagnostics Check	OK	19:05
Air Blank	0.000	19:05
Control Test	0.082	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.215	19:07
Breath Volume	1.960	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.222	19:10
Breath Volume	2.000	19:10
Air Blank	0.000	19:10
Control Test	0.084	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: BODDA

DATE/TIME OF TEST: 07/24/2021 00:23:17

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.083	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:18
Breath Volume	1.273	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:21
Breath Volume	3.113	00:21
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MALAMOS

DATE/TIME OF TEST: 07/25/2021 01:58:42

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.084	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.083	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 08/02/2021 17:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,J,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: MEADE

DATE/TIME OF TEST: 07/25/2021 21:58:23

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.083	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1 0.225	21:51	Volume Not Met
Breath Volume	0.066	21:51	
Air Blank	0.000	21:51	
Air Blank	0.000	21:53	
Subject Sample	# 2	21:56	No Sample Provided
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.079	21:57	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 07/28/2021 00:38:34

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.082	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.000	00:34
Breath Volume	1.242	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.000	00:37
Breath Volume	2.230	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MARTONICK

DATE/TIME OF TEST: 07/28/2021 03:30:52

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.082	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.085	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: FLANIJEN

DATE/TIME OF TEST: 07/28/2021 15:32:49

Diagnostics Check	OK	15:21	
Air Blank	0.000	15:21	
Control Test	0.082	15:22	
Air Blank	0.000	15:22	
Subject Sample	# 1	15:25	No Sample Provided
Breath Volume	0.000	15:25	
Air Blank	0.000	15:26	
Air Blank	0.000	15:28	
Subject Sample	# 2	15:31	No Sample Provided
Breath Volume	0.000	15:31	
Air Blank	0.000	15:31	
Control Test	0.082	15:32	
Air Blank	0.000	15:32	
Diagnostics Check	OK	15:32	

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: FRALEY

DATE/TIME OF TEST: 07/28/2021 22:24:08

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.082	22:22	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.082	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PIENIAZEK

DATE/TIME OF TEST: 07/29/2021 02:46:51

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.083	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.016	02:42
Breath Volume	3.953	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.014	02:45
Breath Volume	3.648	02:45
Air Blank	0.000	02:45
Control Test	0.083	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,JOHN,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: CORLEY

DATE/TIME OF TEST: 07/29/2021 05:34:25

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.082	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.000	05:29
Breath Volume	4.933	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.000	05:33
Breath Volume	5.082	05:33
Air Blank	0.000	05:33
Control Test	0.082	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,JAMISON,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: PAQUETTE

DATE/TIME OF TEST: 07/29/2021 22:22:46

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.082	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1 0.160	22:10	
Breath Volume	1.542	22:10	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2	22:15	No Sample Provided
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:18	
Subject Sample	# 3	22:21	No Sample Provided
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.082	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 08/02/2021 17:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,DAVID,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COX,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/13/2021 16:01:47

Diagnostics Check	OK	15:55
Air Blank	0.000	15:56
Control Test	0.082	15:56
Air Blank	0.000	15:56
Subject Sample # 1	0.153	15:57
Breath Volume	2.988	15:57
Air Blank	0.000	15:58
Air Blank	0.000	15:59
Subject Sample # 2	0.153	16:00
Breath Volume	3.093	16:00
Air Blank	0.000	16:00
Control Test	0.080	16:01
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 09/05/2021 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/13/2021 18:39:50

Diagnostics Check	OK	18:33
Air Blank	0.000	18:34
Control Test	0.081	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.000	18:35
Breath Volume	2.761	18:35
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.000	18:38
Breath Volume	2.601	18:38
Air Blank	0.000	18:38
Control Test	0.082	18:39
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time: 09/05/2021 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: HOWELL,KEENAN
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BISS

DATE/TIME OF TEST: 07/23/2021 23:58:37

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.240	23:54
Breath Volume	1.562	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.233	23:57
Breath Volume	1.195	23:57
Air Blank	0.000	23:57
Control Test	0.082	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 09/05/2021 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HUNTER,JOHN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: BARGE

DATE/TIME OF TEST: 07/03/2021 18:51:35

Diagnostics Check	OK	18:45
Air Blank	0.000	18:45
Control Test	0.081	18:45
Air Blank	0.000	18:46
Subject Sample # 1	0.195	18:46
Breath Volume	2.535	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:49
Subject Sample # 2	0.190	18:50
Breath Volume	2.800	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 07/12/2021 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKINS,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: NASH,JARRED
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WYCHE

DATE/TIME OF TEST: 07/02/2021 22:35:18

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.026	22:31
Breath Volume	2.519	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.023	22:33
Breath Volume	1.292	22:33
Air Blank	0.000	22:34
Control Test	0.077	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 07/14/2021 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA, COLBY, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 07/11/2021 11:45:46

Diagnostics Check	OK	11:36
Air Blank	0.000	11:36
Control Test	0.078	11:36
Air Blank	0.000	11:37
Subject Sample # 1	0.151	11:40
Breath Volume	1.640	11:40
Air Blank	0.000	11:41
Air Blank	0.000	11:42
Subject Sample # 2	0.146	11:44
Breath Volume	1.601	11:44
Air Blank	0.000	11:44
Control Test	0.076	11:45
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time: 07/14/2021 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARGROVE,ERIK,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BIEBERITZ,CHRIS
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BOULETTE

DATE/TIME OF TEST: 07/06/2021 00:12:00

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.457	00:04
Breath Volume		2.296	00:04
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.435	00:07
Breath Volume		3.152	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 3	0.448	00:10
Breath Volume		2.140	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:12

Data Download Date/Time: 07/16/2021 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RODGERS,HALEY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CROW

DATE/TIME OF TEST: 07/10/2021 03:35:24

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1	0.131	03:22	
Breath Volume		1.257	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2	0.116	03:28	Slope Not Level
Breath Volume		1.335	03:28	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 3	0.113	03:33	Slope Not Level
Breath Volume		1.480	03:33	
Air Blank		0.000	03:34	
Control Test		0.079	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 07/16/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PROCTER,JASON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 07/10/2021 22:28:27

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.211	22:23
Breath Volume	1.492	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.207	22:26
Breath Volume	2.914	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 07/16/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/12/2021 02:41:56

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.169	02:37
Breath Volume	1.324	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.183	02:40
Breath Volume	2.035	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 07/16/2021 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: DAYMONT

DATE/TIME OF TEST: 07/14/2021 20:15:16

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.080	20:13	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:13	Subject Test Refused
Breath Volume	0.000	20:13	
Air Blank	0.000	20:14	
Control Test	0.081	20:14	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 07/16/2021 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOLMES,KYLE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/16/2021 23:15:29

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:14	
Control Test	0.081	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 08/04/2021 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAGUADOG,SHAUN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER,DILLON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/18/2021 16:03:05

Diagnostics Check	OK	15:55
Air Blank	0.000	15:56
Control Test	0.080	15:56
Air Blank	0.000	15:57
Subject Sample # 1	0.000	15:57
Breath Volume	1.695	15:57
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.000	16:01
Breath Volume	1.871	16:01
Air Blank	0.000	16:02
Control Test	0.080	16:02
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

Data Download Date/Time: 08/04/2021 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HIRES

DATE/TIME OF TEST: 07/20/2021 19:06:35

Diagnostics Check	OK	19:00
Air Blank	0.000	19:00
Control Test	0.081	19:01
Air Blank	0.000	19:01
Subject Sample # 1	0.104	19:02
Breath Volume	2.746	19:02
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.098	19:05
Breath Volume	1.187	19:05
Air Blank	0.000	19:05
Control Test	0.080	19:06
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 08/04/2021 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JONES, DANIEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: GARCIA ABAD

DATE/TIME OF TEST: 07/22/2021 08:02:50

Diagnostics Check	OK	07:55
Air Blank	0.000	07:56
Control Test	0.080	07:56
Air Blank	0.000	07:57
Subject Sample # 1	0.110	07:58
Breath Volume	1.726	07:58
Air Blank	0.000	07:58
Air Blank	0.000	08:00
Subject Sample # 2	0.113	08:01
Breath Volume	1.695	08:01
Air Blank	0.000	08:01
Control Test	0.080	08:02
Air Blank	0.000	08:02
Diagnostics Check	OK	08:02

Data Download Date/Time: 08/04/2021 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 07/24/2021 01:03:37

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.122	00:59
Breath Volume	1.652	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.120	01:02
Breath Volume	1.386	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 08/04/2021 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BYRD,BILLY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: NEYLAND

DATE/TIME OF TEST: 07/24/2021 01:37:25

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.225	01:32
Breath Volume		2.593	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.235	01:35
Breath Volume		2.933	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 08/04/2021 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GAVIN

DATE/TIME OF TEST: 07/24/2021 03:59:06

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.153	03:54
Breath Volume	1.898	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.156	03:57
Breath Volume	1.894	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 08/04/2021 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ALEX,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ROACHE,DUSTIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RISH

DATE/TIME OF TEST: 07/27/2021 03:57:42

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.080	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 08/04/2021 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 07/29/2021 02:07:22

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.080	02:04	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.078	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 08/04/2021 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,KENNETH,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ALLEN,F
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG923908
DRY GAS EXPIRY DATE: 08/27/2021
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 07/29/2021 07:51:49

Diagnostics Check	OK	07:40	
Air Blank	0.000	07:40	
Control Test	0.079	07:41	
Air Blank	0.000	07:41	
Subject Sample	# 1	07:44	No Sample Provided
Breath Volume	0.000	07:44	
Air Blank	0.000	07:45	
Air Blank	0.000	07:47	
Subject Sample	# 2	07:50	No Sample Provided
Breath Volume	0.000	07:50	
Air Blank	0.000	07:50	
Control Test	0.079	07:51	
Air Blank	0.000	07:51	
Diagnostics Check	OK	07:51	

Data Download Date/Time: 08/04/2021 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POVEDA,CARLOS,M
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: WENDT

DATE/TIME OF TEST: 07/21/2021 14:05:08

Diagnostics Check	OK	14:02	
Air Blank	0.000	14:02	
Control Test	0.079	14:03	
Air Blank	0.000	14:03	
Subject Sample	# 1	14:03	Subject Test Refused
Breath Volume	0.000	14:03	
Air Blank	0.000	14:04	
Control Test	0.079	14:04	
Air Blank	0.000	14:05	
Diagnostics Check	OK	14:05	

Data Download Date/Time: 09/03/2021 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DYKES,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: HALL,DANIEL
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: O'SHEA

DATE/TIME OF TEST: 07/25/2021 02:49:19

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Sequence Aborted
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	

Data Download Date/Time: 09/03/2021 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DYKES,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: HALL,DANIEL
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: O'SHEA

DATE/TIME OF TEST: 07/25/2021 03:00:23

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.167	02:54
Breath Volume	2.144	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.165	02:58
Breath Volume	2.542	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 09/03/2021 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNGE,JACOB,L
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: DYKES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 07/29/2021 03:21:25

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2	03:20	No Sample Provided
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 09/03/2021 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNGE,JACOB,L
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: DYKES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 07/29/2021 03:34:44

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.191	03:29
Breath Volume	2.976	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.171	03:33
Breath Volume	1.218	03:33
Air Blank	0.000	03:33
Control Test	0.076	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/03/2021 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGH,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346773
DRY GAS EXPIRY DATE: 07/08/2023
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/31/2021 01:38:51

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.000	01:34
Breath Volume	3.175	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.000	01:37
Breath Volume	2.425	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 08/21/2021 21:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: FOLTZ

DATE/TIME OF TEST: 07/03/2021 04:51:17

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.110	04:46
Breath Volume	1.437	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:48
Subject Sample # 2	0.107	04:49
Breath Volume	1.894	04:49
Air Blank	0.000	04:50
Control Test	0.081	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 07/03/2021 05:02:47

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.081	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.219	04:58
Breath Volume		4.136	04:58
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.217	05:01
Breath Volume		3.910	05:01
Air Blank		0.000	05:01
Control Test		0.082	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MICHUA CASTANEDA

DATE/TIME OF TEST: 07/04/2021 05:06:26

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.081	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.138	05:02
Breath Volume		2.203	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.135	05:04
Breath Volume		2.152	05:04
Air Blank		0.000	05:05
Control Test		0.082	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,ERIC,C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: FOWLER,ERIC
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: BUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 07/04/2021 18:58:46

Diagnostics Check	OK	18:57	
Air Blank	0.000	18:57	
Control Test	0.081	18:57	
Air Blank		18:58	RFI Detect
Air Blank	0.000	18:58	

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,ERIC,C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: FOWLER,ERIC
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: BUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 07/04/2021 19:04:15

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:01	
Control Test	0.080	19:01	
Air Blank	0.000	19:02	
Subject Sample	# 1	19:02	Subject Test Refused
Breath Volume	0.000	19:02	
Air Blank	0.000	19:03	
Control Test	0.082	19:03	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:04	

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: AJIATAZ TAJIBOY

DATE/TIME OF TEST: 07/05/2021 05:11:58

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.096	05:07
Breath Volume	1.554	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.096	05:10
Breath Volume	1.433	05:10
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KOUCHNIR,EUGENE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/11/2021 00:23:46

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.234	00:18
Breath Volume	1.953	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.215	00:22
Breath Volume	1.496	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUSTER,CORY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GUSTER,CORY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/12/2021 22:24:01

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.081	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.085	22:18
Breath Volume	1.191	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.088	22:22
Breath Volume	1.902	22:22
Air Blank	0.000	22:23
Control Test	0.082	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:24

Data Download Date/Time: 07/15/2021 21:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUNN,CONNOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 07/17/2021 03:00:32

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.109	02:56
Breath Volume	1.949	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.111	02:59
Breath Volume	3.019	02:59
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 08/21/2021 21:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/23/2021 04:10:30

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.167	04:06
Breath Volume	2.191	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.160	04:08
Breath Volume	2.140	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/21/2021 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,CHRIS,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: KAISER,ANTHONY
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/28/2021 22:05:52

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.081	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.164	22:01
Breath Volume		1.750	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:03
Subject Sample	# 2	0.157	22:04
Breath Volume		3.121	22:04
Air Blank		0.000	22:04
Control Test		0.079	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 08/21/2021 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUNN,CONNOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 07/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 07/31/2021 01:31:36

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 08/21/2021 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO, MERCEDES, T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VARNDORE, RICHARD
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SCRUTON

DATE/TIME OF TEST: 07/02/2021 01:02:57

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.080	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 07/02/2021 04:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, MARK, A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: DOVAL, R
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 07/03/2021 00:56:37

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.140	00:52
Breath Volume	3.476	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.137	00:55
Breath Volume	3.937	00:55
Air Blank	0.000	00:55
Control Test	0.080	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/03/2021 02:56:56

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.209	02:51
Breath Volume		1.285	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.196	02:55
Breath Volume		1.738	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/03/2021 22:19:06

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.252	22:14
Breath Volume	3.578	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.249	22:17
Breath Volume	3.695	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/04/2021 02:13:16

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.151	02:08
Breath Volume	2.484	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.152	02:11
Breath Volume	2.617	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/04/2021 04:25:44

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.080	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: COX,TYLER
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 07/04/2021 15:55:10

Diagnostics Check	OK	15:52	
Air Blank	0.000	15:52	
Control Test	0.080	15:53	
Air Blank	0.000	15:53	
Subject Sample	# 1	15:53	Subject Test Refused
Breath Volume	0.000	15:53	
Air Blank	0.000	15:54	
Control Test	0.080	15:54	
Air Blank	0.000	15:55	
Diagnostics Check	OK	15:55	

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 07/04/2021 19:27:49

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.000	19:23
Breath Volume	1.660	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.000	19:26
Breath Volume	1.445	19:26
Air Blank	0.000	19:26
Control Test	0.081	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: SANCHEZ RODRIGUEZ

DATE/TIME OF TEST: 07/04/2021 23:31:22

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.268	23:26
Breath Volume	1.597	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.255	23:29
Breath Volume	1.253	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: LANNING,MARK
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 07/07/2021 17:04:43

Diagnostics Check	OK	16:58
Air Blank	0.000	16:59
Control Test	0.081	16:59
Air Blank	0.000	17:00
Subject Sample # 1	0.000	17:00
Breath Volume	1.937	17:00
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.000	17:03
Breath Volume	2.820	17:03
Air Blank	0.000	17:03
Control Test	0.081	17:04
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 07/08/2021 03:26:44

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.000	03:22
Breath Volume	2.148	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.000	03:25
Breath Volume	1.765	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LGUERRE

DATE/TIME OF TEST: 07/09/2021 02:56:53

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:52
Breath Volume	1.371	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:55
Breath Volume	1.460	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/19/2021 21:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: VANDIVER

DATE/TIME OF TEST: 07/16/2021 21:45:08

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Subject Sample # 1	0.000	21:40
Breath Volume	2.105	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.000	21:43
Breath Volume	2.429	21:43
Air Blank	0.000	21:44
Control Test	0.081	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: DOVAL,RAMIRO
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 07/17/2021 00:07:34

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.079	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 07/17/2021 20:26:14

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.081	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.191	20:20
Breath Volume	1.578	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.197	20:24
Breath Volume	1.308	20:24
Air Blank	0.000	20:25
Control Test	0.080	20:25
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/18/2021 01:45:35

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.057	01:41
Breath Volume	2.339	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.056	01:44
Breath Volume	2.437	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CASIAN,SKYLER
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 07/21/2021 05:35:11

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.147	05:30
Breath Volume	2.632	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.155	05:33
Breath Volume	2.535	05:33
Air Blank	0.000	05:34
Control Test	0.080	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: MARTIN CRAWFORD

DATE/TIME OF TEST: 07/22/2021 22:35:16

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:20
Subject Sample # 1	0.246	22:23
Breath Volume	1.273	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.220	22:27
Breath Volume	1.605	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 3	0.220	22:33
Breath Volume	1.363	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/22/2021 23:32:31

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:31	Subject Test Refused
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 07/23/2021 02:30:20

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.094	02:25
Breath Volume	3.468	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.093	02:28
Breath Volume	2.976	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO, MERCEDES, T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: JONES, JONES
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: ALDRIDE

DATE/TIME OF TEST: 07/25/2021 05:11:27

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.080	05:09	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:10	
Control Test	0.080	05:10	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WILLETTE

DATE/TIME OF TEST: 07/26/2021 22:37:02

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Subject Sample # 1	0.176	22:32
Breath Volume	3.132	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.177	22:35
Breath Volume	2.687	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:37

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CARL,C
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: BROUGHTON,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MEDEROS

DATE/TIME OF TEST: 07/29/2021 04:17:22

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.217	04:09
Breath Volume	1.878	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.238	04:12
Breath Volume	3.996	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 3	0.235	04:15
Breath Volume	2.718	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CARL,C
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: BROUGHTON,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: LERMA

DATE/TIME OF TEST: 07/30/2021 00:29:19

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.162	00:24
Breath Volume	1.218	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.159	00:27
Breath Volume	1.953	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: DE OLIVEIRA

DATE/TIME OF TEST: 07/30/2021 17:13:35

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.079	17:08
Air Blank	0.000	17:08
Subject Sample # 1	0.000	17:09
Breath Volume	2.750	17:09
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.000	17:12
Breath Volume	2.500	17:12
Air Blank	0.000	17:12
Control Test	0.080	17:12
Air Blank	0.000	17:13
Diagnostics Check	OK	17:13

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/31/2021 03:13:29

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: DOVAL,RAMIRO
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: KIGER

DATE/TIME OF TEST: 07/31/2021 03:45:28

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.132	03:40
Breath Volume	1.503	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.151	03:43
Breath Volume	2.886	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 08/19/2021 21:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,K
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 07/02/2021 11:19:16

Diagnostics Check	OK	11:13
Air Blank	0.000	11:13
Control Test	0.080	11:13
Air Blank	0.000	11:14
Subject Sample # 1	0.060	11:14
Breath Volume	2.214	11:14
Air Blank	0.000	11:15
Air Blank	0.000	11:17
Subject Sample # 2	0.064	11:17
Breath Volume	2.320	11:17
Air Blank	0.000	11:18
Control Test	0.079	11:18
Air Blank	0.000	11:19
Diagnostics Check	OK	11:19

Data Download Date/Time: 07/23/2021 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,K
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: DOBSON-RUIZ

DATE/TIME OF TEST: 07/08/2021 10:33:22

Diagnostics Check	OK	10:27
Air Blank	0.000	10:28
Control Test	0.081	10:28
Air Blank	0.000	10:28
Subject Sample # 1	0.000	10:29
Breath Volume	2.222	10:29
Air Blank	0.000	10:29
Air Blank	0.000	10:31
Subject Sample # 2	0.000	10:32
Breath Volume	2.257	10:32
Air Blank	0.000	10:32
Control Test	0.079	10:32
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 07/23/2021 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SCARINGE,M
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: KEANE

DATE/TIME OF TEST: 07/10/2021 04:37:20

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:36	
Control Test	0.081	04:36	
Air Blank	0.000	04:37	
		04:37	RFI Detect
Air Blank		04:37	RFI Detect

Data Download Date/Time: 07/23/2021 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: SCARINGE,M
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: KEANE

DATE/TIME OF TEST: 07/10/2021 04:48:12

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.197	04:43
Breath Volume	1.109	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample # 2	0.203	04:46
Breath Volume	1.886	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 07/23/2021 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHUNN,PETER,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DOMINGUEZ,IVAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HOWSE

DATE/TIME OF TEST: 07/18/2021 22:29:02

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.229	22:24
Breath Volume	2.851	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.216	22:27
Breath Volume	2.710	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:29

Data Download Date/Time: 07/23/2021 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUMBLIN,WILLIS,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: VEGA,B
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MARIAN

DATE/TIME OF TEST: 07/21/2021 23:29:47

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.081	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1	0.184	23:20	Volume Not Met
Breath Volume		0.355	23:20	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 2	0.195	23:22	
Breath Volume		1.312	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 3	0.203	23:28	
Breath Volume		2.390	23:28	
Air Blank		0.000	23:28	
Control Test		0.078	23:29	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 07/23/2021 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONCALVES,JOAO,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPLEBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: QUINTANILLA-VALENCIA

DATE/TIME OF TEST: 07/23/2021 20:57:48

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.000	20:53
Breath Volume	2.976	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:56
Breath Volume	3.039	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 08/31/2021 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPLBAUM,K
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: NOGUERA

DATE/TIME OF TEST: 07/24/2021 01:57:10

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.097	01:52
Breath Volume	3.800	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.099	01:55
Breath Volume	3.804	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 08/31/2021 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: TERAPAK

DATE/TIME OF TEST: 07/29/2021 05:30:06

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.088	05:25
Breath Volume	2.843	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.085	05:28
Breath Volume	2.628	05:28
Air Blank	0.000	05:29
Control Test	0.081	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 08/31/2021 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IANNONE,JOSEPH,V
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: IANNONE,JOSEPH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/31/2021 01:09:12

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:04
Breath Volume	1.890	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:07
Breath Volume	2.097	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 08/31/2021 17:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE, BRANDON
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: JARRIEL

DATE/TIME OF TEST: 07/17/2021 01:30:59

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.152	01:26
Breath Volume	1.964	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.152	01:29
Breath Volume	1.792	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 07/19/2021 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SWENSEN

DATE/TIME OF TEST: 07/17/2021 02:20:10

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.079	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:19	
Control Test	0.079	02:19	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 07/19/2021 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: INESTROZA

DATE/TIME OF TEST: 07/18/2021 19:23:13

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.079	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.052	19:18
Breath Volume	2.066	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.051	19:21
Breath Volume	2.265	19:21
Air Blank	0.000	19:22
Control Test	0.078	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 07/19/2021 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOFFMAN,DANIEL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BROUGH

DATE/TIME OF TEST: 07/20/2021 14:11:31

Diagnostics Check	OK	14:02	
Air Blank	0.000	14:03	
Control Test	0.079	14:03	
Air Blank	0.000	14:03	
Subject Sample	# 1 0.000	14:06	Volume Not Met
Breath Volume	0.250	14:06	
Air Blank	0.000	14:07	
Air Blank	0.000	14:09	
Subject Sample	# 2	14:10	Subject Test Refused
Breath Volume	0.484	14:10	
Air Blank	0.000	14:10	
Control Test	0.077	14:10	
Air Blank	0.000	14:11	
Diagnostics Check	OK	14:11	

Data Download Date/Time: 09/04/2021 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MEDINA ZAVALA

DATE/TIME OF TEST: 07/20/2021 18:50:07

Diagnostics Check	OK	18:49	
Air Blank		18:49	RFI Detect
Air Blank		18:50	RFI Detect

Data Download Date/Time: 09/04/2021 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MEDINA ZAVALA

DATE/TIME OF TEST: 07/20/2021 18:54:57

Diagnostics Check	OK	18:52	
Air Blank	0.000	18:52	
Control Test	0.079	18:52	
Air Blank	0.000	18:53	
Subject Sample	# 1	18:53	Subject Test Refused
Breath Volume	0.000	18:53	
Air Blank	0.000	18:54	
Control Test	0.079	18:54	
Air Blank	0.000	18:54	
Diagnostics Check	OK	18:54	

Data Download Date/Time: 09/04/2021 23:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEILL,DEPUTY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 07/21/2021 00:57:38

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:52
Breath Volume	1.941	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:56
Breath Volume	2.234	00:56
Air Blank	0.000	00:56
Control Test	0.079	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/04/2021 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT, D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: CORBO

DATE/TIME OF TEST: 07/21/2021 20:06:39

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:04	
Control Test	0.079	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:05	Subject Test Refused
Breath Volume	0.000	20:05	
Air Blank	0.000	20:05	
Control Test	0.079	20:06	
Air Blank	0.000	20:06	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 09/04/2021 23:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ, M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SPARKER

DATE/TIME OF TEST: 07/22/2021 01:44:03

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.079	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 09/04/2021 23:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE, B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 07/22/2021 03:36:26

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.079	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.078	03:35	RFI Detect
Air Blank	0.000	03:36	

Data Download Date/Time: 09/04/2021 23:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 07/22/2021 03:42:34

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.079	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 09/04/2021 23:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: GRIMSEY

DATE/TIME OF TEST: 07/27/2021 20:36:24

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.079	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.218	20:29
Breath Volume	1.531	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.227	20:34
Breath Volume	1.378	20:34
Air Blank	0.000	20:35
Control Test	0.078	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 09/04/2021 23:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZE,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FINCIK

DATE/TIME OF TEST: 07/28/2021 01:48:54

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.069	01:43
Breath Volume	2.855	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.058	01:47
Breath Volume	1.175	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 09/04/2021 23:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/29/2021 01:51:56

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 09/04/2021 23:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/31/2021 21:28:02

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.080	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.079	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 09/04/2021 23:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PATERNO

DATE/TIME OF TEST: 07/01/2021 00:19:48

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.182	00:15
Breath Volume	2.031	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.178	00:18
Breath Volume	1.585	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 07/16/2021 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTMAN,KEVIN,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/01/2021 02:04:14

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 07/16/2021 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE, D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: GLOETZNER

DATE/TIME OF TEST: 07/02/2021 19:57:33

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:55	
Control Test	0.082	19:55	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:56	Subject Test Refused
Breath Volume	0.000	19:56	
Air Blank	0.000	19:56	
Control Test	0.081	19:56	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:57	

Data Download Date/Time: 07/16/2021 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 07/03/2021 01:51:02

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.082	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 07/16/2021 17:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SHAW,NICHOLAS
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 07/03/2021 21:34:17

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:31	
Control Test	0.080	21:32	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:33	
Control Test	0.081	21:33	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 07/16/2021 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEAN,A
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MONTO

DATE/TIME OF TEST: 07/03/2021 23:08:22

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.081	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.147	23:03
Breath Volume	1.207	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.149	23:06
Breath Volume	1.191	23:06
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE, B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SANGBUSH

DATE/TIME OF TEST: 07/04/2021 02:00:53

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.080	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SHAW,NICHOLAS
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: FETTEROLF

DATE/TIME OF TEST: 07/04/2021 23:08:12

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.082	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1 0.107	23:03	Volume Not Met
Breath Volume	0.335	23:03	
Air Blank	0.000	23:03	
Air Blank	0.000	23:05	
Subject Sample	# 2	23:06	Subject Test Refused
Breath Volume	0.273	23:06	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 07/04/2021 23:45:34

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.080	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:41	No Sample Provided
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RICHARD, M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ROMERO HERNANDEZ

DATE/TIME OF TEST: 07/05/2021 00:32:48

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.132	00:28
Breath Volume	1.773	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.132	00:31
Breath Volume	1.589	00:31
Air Blank	0.000	00:31
Control Test	0.080	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOFFMAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: LETTON

DATE/TIME OF TEST: 07/05/2021 19:01:10

Diagnostics Check	OK	18:54
Air Blank	0.000	18:55
Control Test	0.081	18:55
Air Blank	0.000	18:56
Subject Sample # 1	0.044	18:56
Breath Volume	2.339	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.043	18:59
Breath Volume	2.222	18:59
Air Blank	0.000	19:00
Control Test	0.082	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARCARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: PINEO

DATE/TIME OF TEST: 07/07/2021 01:12:49

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.081	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:11	
Control Test	0.081	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	Fail	01:12	DSP Test
Air Blank	0.000	01:12	

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIGMORELLIA,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRADY,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 13:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: DEVORE

DATE/TIME OF TEST: 07/07/2021 14:52:53

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.081	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.117	14:48
Breath Volume	1.464	14:48
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.117	14:51
Breath Volume	1.363	14:51
Air Blank	0.000	14:51
Control Test	0.081	14:52
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DELEON,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 07/10/2021 01:37:21

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:30	No Sample Provided
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:35	Subject Test Refused
Breath Volume	0.062	01:35	
Air Blank	0.000	01:35	
Control Test	0.082	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	Fail	01:36	DSP Test
Air Blank	0.000	01:37	

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 07/10/2021 02:23:50

Diagnostics Check	Fail	02:23	DSP Test
Air Blank	0.000	02:23	

Data Download Date/Time: 07/16/2021 17:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 07/10/2021 02:53:34

Diagnostics Check	Fail	02:53	DSP Test
Air Blank	0.000	02:53	

Data Download Date/Time: 07/16/2021 17:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOFFMAN,DANIEL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: GITELMAN

DATE/TIME OF TEST: 07/10/2021 11:31:46

Diagnostics Check	OK	11:24	
Air Blank	0.000	11:25	
Control Test	0.081	11:25	
Air Blank	0.000	11:25	
Subject Sample # 1	0.000	11:26	
Breath Volume	2.472	11:26	
Air Blank	0.000	11:27	
Air Blank	0.000	11:29	
Subject Sample # 2	0.000	11:29	
Breath Volume	2.050	11:29	
Air Blank	0.000	11:30	
Control Test	0.081	11:30	
Air Blank	0.000	11:31	
Diagnostics Check	Fail	11:31	DSP Test
Air Blank	0.000	11:31	

Data Download Date/Time: 07/16/2021 17:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HOILETT,CPL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/10/2021 19:56:24

Diagnostics Check	Fail	19:55	DSP Test
Air Blank	0.000	19:56	

Data Download Date/Time: 07/16/2021 17:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILHARDT,DEP,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/01/2021 03:19:57

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.192	03:15
Breath Volume	3.195	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.186	03:18
Breath Volume	2.753	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/30/2021 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,S,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOPE,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 07/01/2021 04:48:59

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.078	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.158	04:43
Breath Volume	2.546	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.159	04:47
Breath Volume	2.445	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILHARDT,DEP,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 07/02/2021 00:00:56

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.159	23:55
Breath Volume	1.769	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.166	23:59
Breath Volume	1.796	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 07/02/2021 00:52:30

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.222	00:46
Breath Volume	1.496	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.207	00:50
Breath Volume	2.335	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TINFINA,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 07/03/2021 23:49:52

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:45
Breath Volume	3.417	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:48
Breath Volume	3.402	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: WAITERS

DATE/TIME OF TEST: 07/04/2021 21:41:18

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.168	21:35
Breath Volume	2.085	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.168	21:39
Breath Volume	2.164	21:39
Air Blank	0.000	21:40
Control Test	0.078	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,A
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CAMPION

DATE/TIME OF TEST: 07/05/2021 00:55:41

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.212	00:46
Breath Volume	1.363	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.185	00:50
Breath Volume	1.238	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 3	0.198	00:54
Breath Volume	1.187	00:54
Air Blank	0.000	00:54
Control Test	0.078	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHAPIRO,J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 07/05/2021 06:12:57

Diagnostics Check	OK	06:01	
Air Blank	0.000	06:01	
Control Test	0.079	06:01	
Air Blank	0.000	06:02	
Subject Sample	# 1 0.227	06:02	
Breath Volume	1.531	06:02	
Air Blank	0.000	06:03	
Air Blank	0.000	06:05	
Subject Sample	# 2 0.247	06:08	Volume Not Met
Breath Volume	0.078	06:08	
Air Blank	0.000	06:09	
Air Blank	0.000	06:10	
Subject Sample	# 3 0.219	06:11	
Breath Volume	2.167	06:11	
Air Blank	0.000	06:12	
Control Test	0.071	06:12	Control Outside Tolerance
Air Blank	0.000	06:12	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHAPIRO,J
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 07/05/2021 06:33:06

Diagnostics Check	OK	06:25	
Air Blank	0.000	06:26	
Control Test	0.077	06:26	
Air Blank	0.000	06:27	
Subject Sample	# 1 0.214	06:27	
Breath Volume	1.816	06:27	
Air Blank	0.000	06:28	
Air Blank	0.000	06:29	
Subject Sample	# 2	06:31	Subject Test Refused
Breath Volume	0.066	06:31	
Air Blank	0.000	06:32	
Control Test	0.077	06:32	
Air Blank	0.000	06:33	
Diagnostics Check	OK	06:33	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KIEFER,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: SIZEMORE

DATE/TIME OF TEST: 07/05/2021 12:17:03

Diagnostics Check		OK	12:05	
Air Blank		0.000	12:05	
Control Test		0.079	12:06	
Air Blank		0.000	12:06	
Subject Sample	# 1	0.000	12:09	Volume Not Met
Breath Volume		1.023	12:09	
Air Blank		0.000	12:10	
Air Blank		0.000	12:12	
Subject Sample	# 2	0.000	12:12	
Breath Volume		1.691	12:12	
Air Blank		0.000	12:13	
Air Blank		0.000	12:15	
Subject Sample	# 3	0.000	12:15	
Breath Volume		2.292	12:15	
Air Blank		0.000	12:16	
Control Test		0.079	12:16	
Air Blank		0.000	12:16	
Diagnostics Check		OK	12:17	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HARBAUGH

DATE/TIME OF TEST: 07/07/2021 03:00:54

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.133	02:52
Breath Volume	1.699	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:54
Subject Sample # 2	0.162	02:55
Breath Volume	1.644	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 3	0.145	02:59
Breath Volume	2.242	02:59
Air Blank	0.000	02:59
Control Test	0.077	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: AMSDEN

DATE/TIME OF TEST: 07/07/2021 19:48:21

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.080	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.031	19:43
Breath Volume	2.417	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.030	19:46
Breath Volume	2.453	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SKINNER,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 07/11/2021 04:08:28

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.215	03:58	
Breath Volume	1.492	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 2 0.213	04:04	Volume Not Met
Breath Volume	0.035	04:04	
Air Blank	0.000	04:04	
Air Blank	0.000	04:06	
Subject Sample	# 3 0.195	04:06	
Breath Volume	1.230	04:06	
Air Blank	0.000	04:07	
Control Test	0.074	04:07	Control Outside Tolerance
Air Blank	0.000	04:08	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLLINS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 07/12/2021 02:04:20

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.202	01:59
Breath Volume	3.523	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.194	02:02
Breath Volume	3.277	02:02
Air Blank	0.000	02:03
Control Test	0.077	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILHARDT,DEP,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLEMAN,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 07/15/2021 01:31:43

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.000	01:27
Breath Volume	2.121	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.000	01:30
Breath Volume	1.972	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/15/2021 03:33:26

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.115	03:28
Breath Volume	2.355	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.113	03:31
Breath Volume	4.082	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JABLONSKI,OFC
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CARRIGAN

DATE/TIME OF TEST: 07/16/2021 06:32:13

Diagnostics Check		OK	06:20	
Air Blank		0.000	06:21	
Control Test		0.079	06:21	
Air Blank		0.000	06:21	
Subject Sample	# 1	0.088	06:25	Volume Not Met
Breath Volume		0.441	06:25	
Air Blank		0.000	06:25	
Air Blank		0.000	06:27	
Subject Sample	# 2	0.093	06:30	Volume Not Met
Breath Volume		0.339	06:30	
Air Blank		0.000	06:31	
Control Test		0.079	06:31	
Air Blank		0.000	06:32	
Diagnostics Check		OK	06:32	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 07/17/2021 06:07:42

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.079	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.107	06:03
Breath Volume	3.074	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.103	06:06
Breath Volume	2.457	06:06
Air Blank	0.000	06:06
Control Test	0.078	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,OFC
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: GULLIFA

DATE/TIME OF TEST: 07/17/2021 22:32:25

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.064	22:27
Breath Volume	1.937	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.064	22:30
Breath Volume	2.101	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MENDEZ SALES

DATE/TIME OF TEST: 07/18/2021 05:00:01

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank		04:59	RFI Detect
Air Blank	0.000	05:00	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MENDEZ SALES

DATE/TIME OF TEST: 07/18/2021 05:09:18

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.131	05:04
Breath Volume		2.203	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.135	05:07
Breath Volume		2.195	05:07
Air Blank		0.000	05:08
Control Test		0.079	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MCCORKLE

DATE/TIME OF TEST: 07/18/2021 20:46:21

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.080	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.154	20:38
Breath Volume		2.203	20:38
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.133	20:41
Breath Volume		1.753	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:43
Subject Sample	# 3	0.141	20:44
Breath Volume		2.816	20:44
Air Blank		0.000	20:45
Control Test		0.077	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 07/19/2021 02:25:07

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.230	02:19
Breath Volume	1.328	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.223	02:23
Breath Volume	2.300	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILHARDT,DEP,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/20/2021 04:43:13

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.128	04:38
Breath Volume	1.699	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.135	04:41
Breath Volume	1.339	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PANDELOGLOU

DATE/TIME OF TEST: 07/21/2021 02:28:04

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.133	02:22
Breath Volume	2.332	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.129	02:26
Breath Volume	2.058	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,J,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/21/2021 14:42:37

Diagnostics Check	OK	14:35
Air Blank	0.000	14:36
Control Test	0.079	14:36
Air Blank	0.000	14:37
Subject Sample # 1	0.173	14:37
Breath Volume	1.812	14:37
Air Blank	0.000	14:38
Air Blank	0.000	14:40
Subject Sample # 2	0.173	14:41
Breath Volume	1.863	14:41
Air Blank	0.000	14:41
Control Test	0.079	14:42
Air Blank	0.000	14:42
Diagnostics Check	OK	14:42

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: BENVENISTI

DATE/TIME OF TEST: 07/22/2021 17:51:44

Diagnostics Check	OK	17:44
Air Blank	0.000	17:45
Control Test	0.080	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.000	17:47
Breath Volume	2.214	17:47
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.000	17:50
Breath Volume	1.394	17:50
Air Blank	0.000	17:50
Control Test	0.079	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: TREHARNE

DATE/TIME OF TEST: 07/22/2021 22:59:18

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.183	22:54
Breath Volume	1.152	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.183	22:57
Breath Volume	1.398	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BRIONES-VEGA

DATE/TIME OF TEST: 07/23/2021 02:16:26

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.160	02:08
Breath Volume	1.328	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.184	02:11
Breath Volume	2.054	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 3	0.170	02:14
Breath Volume	1.574	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/23/2021 22:52:45

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.079	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 07/24/2021 00:41:13

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.204	00:36
Breath Volume		2.308	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.213	00:39
Breath Volume		2.582	00:39
Air Blank		0.000	00:40
Control Test		0.079	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DEP,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MILLER,DEP R
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MACIAS HERNANDEZ

DATE/TIME OF TEST: 07/24/2021 03:40:57

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.158	03:35
Breath Volume		1.558	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.161	03:39
Breath Volume		1.980	03:39
Air Blank		0.000	03:40
Control Test		0.077	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 07/24/2021 22:52:05

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.167	22:47
Breath Volume	2.125	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.161	22:50
Breath Volume	1.968	22:50
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:52

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: WINKLER

DATE/TIME OF TEST: 07/24/2021 23:04:55

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.155	23:00
Breath Volume	1.308	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.149	23:03
Breath Volume	1.410	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 07/26/2021 04:10:27

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.195	04:06
Breath Volume		2.062	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.183	04:08
Breath Volume		1.800	04:08
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: VESSELLE

DATE/TIME OF TEST: 07/27/2021 02:16:22

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.154	02:11
Breath Volume		1.320	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:13
Subject Sample	# 2	0.150	02:14
Breath Volume		1.140	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: LUSCAVAGE

DATE/TIME OF TEST: 07/27/2021 19:26:04

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.079	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.248	19:21
Breath Volume	2.761	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.248	19:24
Breath Volume	2.523	19:24
Air Blank	0.000	19:25
Control Test	0.078	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:26

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: GOMEZGUZMAN

DATE/TIME OF TEST: 07/28/2021 02:26:18

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STANALAND,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 07/28/2021 03:37:19

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.187	03:32
Breath Volume	2.300	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.184	03:35
Breath Volume	2.871	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CAVALIERE

DATE/TIME OF TEST: 07/30/2021 00:38:51

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.080	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.191	00:34
Breath Volume	2.347	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 2	0.175	00:37
Breath Volume	1.425	00:37
Air Blank	0.000	00:37
Control Test	0.077	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 07/30/2021 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 07/31/2021 00:09:28

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.080	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.150	00:04
Breath Volume		2.730	00:04
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.153	00:07
Breath Volume		2.488	00:07
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 09/06/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV, OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON, OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: MYKYTYN

DATE/TIME OF TEST: 07/31/2021 23:11:22

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.224	23:06
Breath Volume	3.425	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.221	23:09
Breath Volume	3.191	23:09
Air Blank	0.000	23:10
Control Test	0.078	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/06/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,CHANNING,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WHITFIELD,CIARA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: YANCEY

DATE/TIME OF TEST: 07/03/2021 02:45:41

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.143	02:41
Breath Volume	2.671	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.142	02:44
Breath Volume	2.406	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/04/2021 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WHITFIELD,CIARA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 07/07/2021 00:21:41

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:10	
Control Test		0.081	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 1	0.156	00:14	Volume Not Met
Breath Volume		0.636	00:14	
Air Blank		0.000	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 2	0.160	00:20	Volume Not Met
Breath Volume		0.519	00:20	
Air Blank		0.000	00:20	
Control Test		0.080	00:21	
Air Blank		0.000	00:21	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 08/04/2021 08:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MIRABAL,GIANCARLO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 07/23/2021 06:01:39

Diagnostics Check		OK	05:51
Air Blank		0.000	05:52
Control Test		0.080	05:52
Air Blank		0.000	05:53
Subject Sample	# 1	0.163	05:55
Breath Volume		1.582	05:55
Air Blank		0.000	05:56
Air Blank		0.000	05:57
Subject Sample	# 2	0.174	06:00
Breath Volume		1.691	06:00
Air Blank		0.000	06:00
Control Test		0.077	06:01
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 08/05/2021 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAUDRA,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/24/2021 00:48:39

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.000	00:43
Breath Volume	1.109	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.000	00:47
Breath Volume	2.218	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 08/05/2021 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PHARES

DATE/TIME OF TEST: 07/25/2021 00:05:50

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.077	00:01
Breath Volume	1.171	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.080	00:04
Breath Volume	2.058	00:04
Air Blank	0.000	00:04
Control Test	0.079	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 08/05/2021 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORITZ,DFC,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 07/25/2021 01:58:27

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.080	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1	0.138	01:51	Volume Not Met
Breath Volume		0.156	01:51	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 2	0.127	01:56	Volume Not Met
Breath Volume		0.000	01:56	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 08/05/2021 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LARKINS,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MARCHMANN

DATE/TIME OF TEST: 07/26/2021 02:47:19

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.183	02:42
Breath Volume		2.253	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.179	02:45
Breath Volume		1.371	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/05/2021 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SEVOR,ETHAN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LOPEZ MALDONADO

DATE/TIME OF TEST: 07/05/2021 03:54:16

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.231	03:49
Breath Volume		1.363	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.230	03:52
Breath Volume		1.582	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: TIRADO,FREDERID
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: LIGHTNER

DATE/TIME OF TEST: 07/09/2021 03:39:18

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.255	03:34
Breath Volume		1.687	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.247	03:37
Breath Volume		1.707	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SUTTON,ISAAC
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/09/2021 23:42:20

Diagnostics Check	OK	23:41	
Air Blank		23:41	RFI Detect
Air Blank	0.000	23:42	

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SUTTON,ISAAC
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/09/2021 23:50:27

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.165	23:46
Breath Volume		1.105	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.171	23:48
Breath Volume		1.394	23:48
Air Blank		0.000	23:49
Control Test		0.078	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: CHRISTIE,LUCAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ESPINOSA BRAVO

DATE/TIME OF TEST: 07/16/2021 03:51:01

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:49	Slope Not Met
Breath Volume	1.003	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: CHRISTIE,LUCAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: ESPINOSA BRAVO

DATE/TIME OF TEST: 07/16/2021 03:59:34

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.078	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.192	03:54
Breath Volume	2.324	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.172	03:58
Breath Volume	1.812	03:58
Air Blank	0.000	03:58
Control Test	0.076	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: CHRISTIE,LUCAS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: MORI

DATE/TIME OF TEST: 07/24/2021 04:39:49

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.176	04:35
Breath Volume	3.093	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.174	04:38
Breath Volume	3.429	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: RYAN,DUSTIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/27/2021 23:18:57

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:14	No Sample Provided
Breath Volume	0.000	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 2	23:17	Subject Test Refused
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:18	

Data Download Date/Time: 07/30/2021 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART,ROBERT,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: OSBORNE,O
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A4
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: HARDING

DATE/TIME OF TEST: 07/31/2021 13:43:36

Diagnostics Check	OK	13:37
Air Blank	0.000	13:38
Control Test	0.080	13:38
Air Blank	0.000	13:38
Subject Sample # 1	0.000	13:39
Breath Volume	2.128	13:39
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.000	13:42
Breath Volume	1.171	13:42
Air Blank	0.000	13:42
Control Test	0.079	13:43
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 08/26/2021 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CALATCHI,RALPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/03/2021 01:03:24

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:52	
Control Test		0.084	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1	0.190	00:56	Volume Not Met
Breath Volume		0.433	00:56	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2	0.174	01:01	Volume Not Met
Breath Volume		0.019	01:01	
Air Blank		0.000	01:02	
Control Test		0.081	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 07/06/2021 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZ,RICHARD
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: STARCK

DATE/TIME OF TEST: 07/05/2021 01:30:42

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:19	
Control Test		0.084	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1	0.264	01:23	Volume Not Met
Breath Volume		0.242	01:23	
Air Blank		0.000	01:24	
Air Blank		0.000	01:26	
Subject Sample	# 2	0.224	01:29	Volume Not Met
Breath Volume		0.406	01:29	
Air Blank		0.000	01:29	
Control Test		0.081	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 07/06/2021 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/06/2021 00:10:38

Diagnostics Check		OK	00:04
Air Blank		0.000	00:04
Control Test		0.084	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:06
Breath Volume		1.296	00:06
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.000	00:09
Breath Volume		1.464	00:09
Air Blank		0.000	00:09
Control Test		0.083	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 07/06/2021 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NELSON,GREGG
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ROBILLARD

DATE/TIME OF TEST: 07/07/2021 00:45:33

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.084	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.000	00:41
Breath Volume	2.734	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:44
Breath Volume	2.683	00:44
Air Blank	0.000	00:44
Control Test	0.084	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 08/12/2021 07:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAMMAD,JIHAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: GOODNIGHT

DATE/TIME OF TEST: 07/07/2021 05:20:45

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:17	
Control Test	0.083	05:18	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:19	Subject Test Refused
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Control Test	0.083	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 08/12/2021 07:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BOYD,RENEE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: AMES

DATE/TIME OF TEST: 07/07/2021 21:50:18

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.084	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.083	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 08/12/2021 07:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: BOYD JR

DATE/TIME OF TEST: 07/09/2021 04:32:07

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.084	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.000	04:27
Breath Volume	3.867	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.000	04:30
Breath Volume	4.253	04:30
Air Blank	0.000	04:31
Control Test	0.083	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 08/12/2021 07:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/10/2021 02:10:27

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.084	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.243	02:05
Breath Volume	2.250	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.246	02:08
Breath Volume	1.304	02:08
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 08/12/2021 07:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRULE,JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MITTWEDE JR

DATE/TIME OF TEST: 07/10/2021 03:03:22

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.083	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.186	02:59
Breath Volume		1.750	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.184	03:01
Breath Volume		1.660	03:01
Air Blank		0.000	03:02
Control Test		0.082	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 08/12/2021 07:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: AMBRASHKA

DATE/TIME OF TEST: 07/10/2021 03:46:46

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.083	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.077	03:42
Breath Volume		1.960	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.075	03:45
Breath Volume		1.449	03:45
Air Blank		0.000	03:45
Control Test		0.083	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 08/12/2021 07:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUBIC,JUDE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 07/10/2021 23:37:31

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.084	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.206	23:33
Breath Volume	2.136	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.210	23:36
Breath Volume	1.988	23:36
Air Blank	0.000	23:36
Control Test	0.082	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 08/12/2021 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PROFFER,KATIE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 07/11/2021 04:15:32

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.083	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.085	04:11
Breath Volume		1.609	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.084	04:14
Breath Volume		2.156	04:14
Air Blank		0.000	04:14
Control Test		0.082	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 08/12/2021 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS, MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/11/2021 07:30:22

Diagnostics Check	OK	07:22
Air Blank	0.000	07:23
Control Test	0.084	07:23
Air Blank	0.000	07:24
Subject Sample # 1	0.198	07:25
Breath Volume	1.621	07:25
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.203	07:28
Breath Volume	1.792	07:28
Air Blank	0.000	07:29
Control Test	0.081	07:29
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 08/12/2021 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO, OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MERRYMAN

DATE/TIME OF TEST: 07/12/2021 02:28:12

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.083	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.147	02:23
Breath Volume	3.105	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.147	02:26
Breath Volume	2.898	02:26
Air Blank	0.000	02:27
Control Test	0.083	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 08/12/2021 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: QUIMBAYO, MONICA
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CABELLO

DATE/TIME OF TEST: 07/12/2021 03:15:01

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.084	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.171	03:10
Breath Volume	3.183	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.173	03:13
Breath Volume	3.042	03:13
Air Blank	0.000	03:14
Control Test	0.082	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 08/12/2021 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATTHEWS,GREG
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: BUTTERFIELD

DATE/TIME OF TEST: 07/13/2021 16:34:18

Diagnostics Check	OK	16:28
Air Blank	0.000	16:28
Control Test	0.084	16:29
Air Blank	0.000	16:29
Subject Sample # 1	0.225	16:30
Breath Volume	2.226	16:30
Air Blank	0.000	16:30
Air Blank	0.000	16:32
Subject Sample # 2	0.211	16:32
Breath Volume	2.046	16:32
Air Blank	0.000	16:33
Control Test	0.083	16:33
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 08/12/2021 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/13/2021 23:06:29

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.083	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.140	23:02
Breath Volume		2.285	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.139	23:05
Breath Volume		2.125	23:05
Air Blank		0.000	23:05
Control Test		0.083	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 08/12/2021 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUBIC,JUDE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: PANTON

DATE/TIME OF TEST: 07/14/2021 19:52:16

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.085	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.062	19:47
Breath Volume	1.449	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.064	19:50
Breath Volume	1.121	19:50
Air Blank	0.000	19:51
Control Test	0.083	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 08/12/2021 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STANLEY,JONATHAN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: KATO

DATE/TIME OF TEST: 07/15/2021 03:30:52

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.083	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.206	03:26
Breath Volume	3.906	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.210	03:29
Breath Volume	3.191	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 08/12/2021 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOFFETT,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: GUTIERREZCORTEZ

DATE/TIME OF TEST: 07/15/2021 21:46:14

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.084	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.188	21:41
Breath Volume	2.808	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.195	21:44
Breath Volume	2.476	21:44
Air Blank	0.000	21:45
Control Test	0.083	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 08/12/2021 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MIRABAL FLORES

DATE/TIME OF TEST: 07/16/2021 02:56:40

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.084	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.077	02:52
Breath Volume	1.234	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.077	02:55
Breath Volume	1.363	02:55
Air Blank	0.000	02:55
Control Test	0.084	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDERSON,JENNIFER
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: GIRALDI

DATE/TIME OF TEST: 07/16/2021 09:07:25

Diagnostics Check		OK	08:59
Air Blank		0.000	08:59
Control Test		0.083	08:59
Air Blank		0.000	09:00
Subject Sample	# 1	0.229	09:02
Breath Volume		2.523	09:02
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.212	09:05
Breath Volume		2.175	09:05
Air Blank		0.000	09:06
Control Test		0.082	09:06
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAYDEN,BENJAMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: DESILVA

DATE/TIME OF TEST: 07/16/2021 23:40:44

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.084	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.171	23:36
Breath Volume	1.898	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.164	23:39
Breath Volume	1.500	23:39
Air Blank	0.000	23:39
Control Test	0.083	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BERRIOS,MARK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ESCOBARCABRERA

DATE/TIME OF TEST: 07/17/2021 03:46:32

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.084	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.110	03:41
Breath Volume	1.148	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.102	03:45
Breath Volume	1.171	03:45
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: LOPEZVARGAS

DATE/TIME OF TEST: 07/17/2021 08:48:00

Diagnostics Check	OK	08:41
Air Blank	0.000	08:42
Control Test	0.083	08:42
Air Blank	0.000	08:43
Subject Sample # 1	0.071	08:43
Breath Volume	2.011	08:43
Air Blank	0.000	08:44
Air Blank	0.000	08:45
Subject Sample # 2	0.065	08:46
Breath Volume	1.398	08:46
Air Blank	0.000	08:47
Control Test	0.083	08:47
Air Blank	0.000	08:47
Diagnostics Check	OK	08:48

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS, TABITHA, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BERRIOS, MARK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: STREET

DATE/TIME OF TEST: 07/18/2021 02:32:57

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.084	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.122	02:28
Breath Volume	3.847	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:30
Subject Sample # 2	0.130	02:31
Breath Volume	3.765	02:31
Air Blank	0.000	02:32
Control Test	0.083	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 07/18/2021 05:17:28

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.083	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1 0.151	05:07	
Breath Volume	1.738	05:07	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2 0.144	05:13	Volume Not Met
Breath Volume	0.484	05:13	
Air Blank	0.000	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 3	05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.082	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 07/19/2021 02:14:05

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.084	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.206	02:09
Breath Volume	1.144	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.204	02:12
Breath Volume	1.320	02:12
Air Blank	0.000	02:13
Control Test	0.083	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCFERREN,JASON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 07/19/2021 15:35:59

Diagnostics Check	OK	15:29
Air Blank	0.000	15:30
Control Test	0.084	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.000	15:31
Breath Volume	3.597	15:31
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.000	15:34
Breath Volume	3.308	15:34
Air Blank	0.000	15:35
Control Test	0.083	15:35
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MELLA,MELISSA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 07/20/2021 22:00:38

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.085	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.000	21:56
Breath Volume	1.636	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.000	21:59
Breath Volume	1.281	21:59
Air Blank	0.000	21:59
Control Test	0.084	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PROFFER,KATIE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: NAUREDDINE

DATE/TIME OF TEST: 07/21/2021 03:04:58

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.084	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.069	03:00
Breath Volume	1.648	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.065	03:03
Breath Volume	1.714	03:03
Air Blank	0.000	03:04
Control Test	0.083	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESCANDON,TIMOTHY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 07/24/2021 00:41:43

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.084	00:35
Air Blank		0.000	00:35
Subject Sample	# 1	0.234	00:36
Breath Volume		1.800	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.235	00:40
Breath Volume		1.378	00:40
Air Blank		0.000	00:40
Control Test		0.083	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TUERO, ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SAMS

DATE/TIME OF TEST: 07/27/2021 04:02:05

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.084	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.201	03:57
Breath Volume	1.679	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.204	04:00
Breath Volume	1.898	04:00
Air Blank	0.000	04:01
Control Test	0.084	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BOYOL, RENEE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 07/27/2021 05:15:58

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.083	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.165	05:11
Breath Volume	1.187	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.165	05:14
Breath Volume	1.257	05:14
Air Blank	0.000	05:15
Control Test	0.083	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 08/23/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/28/2021 02:45:54

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.084	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.204	02:41
Breath Volume	3.312	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.205	02:44
Breath Volume	3.246	02:44
Air Blank	0.000	02:45
Control Test	0.083	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/23/2021 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINS,CONNOR,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PARRISH,KAYLA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 07/29/2021 03:00:44

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.084	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.217	02:51
Breath Volume		1.410	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.195	02:56
Breath Volume		1.617	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:58
Subject Sample	# 3	0.208	02:59
Breath Volume		1.222	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 08/23/2021 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDERSON,NEAL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: RODRIGUEZVELEZ

DATE/TIME OF TEST: 07/30/2021 03:20:36

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.084	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.137	03:15
Breath Volume	1.835	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.134	03:19
Breath Volume	1.960	03:19
Air Blank	0.000	03:19
Control Test	0.083	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 08/23/2021 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATSON,ERIC
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 07/30/2021 03:53:20

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.083	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.210	03:48
Breath Volume	1.539	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.212	03:51
Breath Volume	2.031	03:51
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 08/23/2021 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PROFFER,KATIE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/30/2021 05:50:22

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.083	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.210	05:43
Breath Volume	3.382	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.190	05:48
Breath Volume	2.105	05:48
Air Blank	0.000	05:49
Control Test	0.083	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 08/23/2021 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: WETHERHOLT

DATE/TIME OF TEST: 07/31/2021 04:43:38

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.084	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.187	04:39
Breath Volume	3.816	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.190	04:42
Breath Volume	3.546	04:42
Air Blank	0.000	04:42
Control Test	0.083	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 08/23/2021 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OGLETREE,JERRELL
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: DIGGINS

DATE/TIME OF TEST: 07/31/2021 06:30:20

Diagnostics Check	OK	06:21
Air Blank	0.000	06:21
Control Test	0.084	06:22
Air Blank	0.000	06:22
Subject Sample # 1	0.079	06:24
Breath Volume	1.140	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.073	06:28
Breath Volume	1.285	06:28
Air Blank	0.000	06:29
Control Test	0.082	06:29
Air Blank	0.000	06:30
Diagnostics Check	OK	06:30

Data Download Date/Time: 08/23/2021 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 07/07/2021 02:17:55

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:13
Breath Volume	1.820	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:16
Breath Volume	1.554	02:16
Air Blank	0.000	02:17
Control Test	0.082	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 08/12/2021 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 07/07/2021 05:38:25

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.082	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.061	05:33
Breath Volume	1.183	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.056	05:36
Breath Volume	1.164	05:36
Air Blank	0.000	05:37
Control Test	0.081	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 08/12/2021 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: FRAIFELD

DATE/TIME OF TEST: 07/10/2021 05:04:16

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.000	05:00
Breath Volume	1.429	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.000	05:02
Breath Volume	1.777	05:02
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 08/12/2021 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: PEDIGO

DATE/TIME OF TEST: 07/13/2021 04:36:04

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.127	04:29
Breath Volume	1.269	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.130	04:34
Breath Volume	1.476	04:34
Air Blank	0.000	04:35
Control Test	0.082	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:36

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VIDAL,XAVIER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RAMASSAR

DATE/TIME OF TEST: 07/16/2021 01:00:49

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.000	00:56
Breath Volume	1.214	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	00:59
Breath Volume	1.226	00:59
Air Blank	0.000	00:59
Control Test	0.081	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PROFFER,KATIE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: EASTERLING

DATE/TIME OF TEST: 07/16/2021 03:45:03

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.082	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.275	03:40
Breath Volume		2.238	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.257	03:43
Breath Volume		2.757	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:45

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DENNY

DATE/TIME OF TEST: 07/18/2021 03:56:53

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.082	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.114	03:52
Breath Volume	3.492	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.107	03:55
Breath Volume	2.878	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FAGUNDES, FELIPE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: SHARA

DATE/TIME OF TEST: 07/19/2021 00:10:08

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.081	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:03	No Sample Provided
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2	00:08	No Sample Provided
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONGSHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/19/2021 23:05:26

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.167	23:00
Breath Volume		4.261	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.169	23:03
Breath Volume		4.828	23:03
Air Blank		0.000	23:04
Control Test		0.081	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/24/2021 06:28:20

Diagnostics Check	OK	06:22
Air Blank	0.000	06:22
Control Test	0.082	06:22
Air Blank	0.000	06:23
Subject Sample # 1	0.217	06:23
Breath Volume	1.753	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample # 2	0.203	06:26
Breath Volume	2.042	06:26
Air Blank	0.000	06:27
Control Test	0.082	06:27
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA, ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: DIAZROMERO

DATE/TIME OF TEST: 07/24/2021 20:59:22

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.082	20:50	
Air Blank	0.000	20:50	
Subject Sample	# 1 0.269	20:54	Volume Not Met
Breath Volume	0.171	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:56	
Subject Sample	# 2	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:58	
Control Test	0.078	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BERRIOS, MARK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 07/25/2021 04:01:06

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.081	03:56
Breath Volume	1.183	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.085	03:59
Breath Volume	1.105	03:59
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS, TABITHA, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO, OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: CALDERONCALDERON

DATE/TIME OF TEST: 07/25/2021 04:47:53

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.082	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.131	04:43
Breath Volume	1.628	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.115	04:46
Breath Volume	1.566	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: APPLEBY,JACE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: AGUILARFUENTES

DATE/TIME OF TEST: 07/25/2021 06:50:53

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.082	06:43
Air Blank	0.000	06:43
Subject Sample # 1	0.177	06:46
Breath Volume	1.625	06:46
Air Blank	0.000	06:47
Air Blank	0.000	06:48
Subject Sample # 2	0.177	06:49
Breath Volume	2.253	06:49
Air Blank	0.000	06:49
Control Test	0.081	06:50
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVIA, NICHOLAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WHITE JR

DATE/TIME OF TEST: 07/26/2021 01:15:04

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.082	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.000	01:10
Breath Volume	1.582	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.000	01:13
Breath Volume	2.519	01:13
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:15

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BERRIOS,MARK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 07/26/2021 04:04:09

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.081	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.159	03:59
Breath Volume	1.750	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.168	04:02
Breath Volume	2.281	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERVAIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ERICKSON,JORDAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/27/2021 05:06:57

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:03	
Control Test	0.082	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:05	Subject Test Refused
Breath Volume	0.000	05:05	
Air Blank	0.000	05:06	
Control Test	0.082	05:06	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 08/12/2021 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEWAN,JANICE,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HALL JR

DATE/TIME OF TEST: 07/29/2021 03:12:01

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.082	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.000	03:07
Breath Volume	1.417	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:10
Breath Volume	1.859	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

Data Download Date/Time: 08/12/2021 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: CONVERY

DATE/TIME OF TEST: 07/01/2021 00:26:48

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.106	00:21
Breath Volume	3.054	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.114	00:25
Breath Volume	3.093	00:25
Air Blank	0.000	00:25
Control Test	0.077	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 07/23/2021 01:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 07/01/2021 23:05:06

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.266	22:57
Breath Volume		1.351	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.293	23:00
Breath Volume		2.121	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:02
Subject Sample	# 3	0.284	23:03
Breath Volume		1.925	23:03
Air Blank		0.000	23:04
Control Test		0.077	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 07/23/2021 01:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASLINE,T,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WESTER,A
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: BUI
ARREST TIME: 13:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 07/03/2021 15:08:41

Diagnostics Check	OK	15:02
Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:03
Subject Sample # 1	0.112	15:03
Breath Volume	1.832	15:03
Air Blank	0.000	15:04
Air Blank	0.000	15:06
Subject Sample # 2	0.108	15:07
Breath Volume	1.187	15:07
Air Blank	0.000	15:07
Control Test	0.079	15:08
Air Blank	0.000	15:08
Diagnostics Check	OK	15:08

Data Download Date/Time: 07/28/2021 08:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: CORREIA

DATE/TIME OF TEST: 07/10/2021 23:35:12

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.082	23:27	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:31	No Sample Provided
Breath Volume	0.000	23:31	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 2	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:34	
Control Test	0.080	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 07/28/2021 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARLAND,TIA,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: MCVANNEL,TREAVOR
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 07/03/2021 02:13:38

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.082	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.081	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 07/12/2021 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANG,JOHN,G
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: CHANG,JOHN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BRISTER

DATE/TIME OF TEST: 07/05/2021 02:34:11

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1 0.147	02:30	
Breath Volume	1.898	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2 0.144	02:33	
Breath Volume	1.742	02:33	
Air Blank		02:34	RFI Detect
Air Blank		02:34	RFI Detect

Data Download Date/Time: 07/12/2021 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANG,JOHN,G
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: CHANG,JOHN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BRISTER

DATE/TIME OF TEST: 07/05/2021 02:46:16

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.146	02:40
Breath Volume	2.203	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.143	02:44
Breath Volume	1.664	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 07/12/2021 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VIRNOCHE,JOSIAH,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: VIRNOCHE,JOSIAH
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 07/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/18/2021 04:45:46

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.000	04:41
Breath Volume	1.570	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.000	04:44
Breath Volume	1.925	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 08/24/2021 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JARED,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: MCCRANEY,CHARLES
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 07/04/2021 02:23:31

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.079	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.078	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 08/11/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JARED,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: MCCRANEY,C
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MURROW

DATE/TIME OF TEST: 07/05/2021 02:46:28

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.077	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.062	02:42
Breath Volume	3.132	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.060	02:44
Breath Volume	2.312	02:44
Air Blank	0.000	02:45
Control Test	0.076	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/11/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JARED,R
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: MCCRANEY,CHARLES
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/07/2021 22:30:42

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.078	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.085	22:26
Breath Volume	2.539	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.084	22:29
Breath Volume	2.792	22:29
Air Blank	0.000	22:29
Control Test	0.076	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 08/11/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,KENNETH,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ROBERTS,KENNETH
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: KEENEY

DATE/TIME OF TEST: 07/09/2021 23:05:02

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.138	23:00
Breath Volume		2.960	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.134	23:03
Breath Volume		3.218	23:03
Air Blank		0.000	23:04
Control Test		0.078	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:05

Data Download Date/Time: 08/11/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCASTRO,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: KOSSEY,NICHOLAS
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A2
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: BUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: MCTAGGART

DATE/TIME OF TEST: 07/17/2021 18:03:29

Diagnostics Check	OK	17:56
Air Blank	0.000	17:57
Control Test	0.079	17:57
Air Blank	0.000	17:58
Subject Sample # 1	0.040	17:58
Breath Volume	2.941	17:58
Air Blank	0.000	17:59
Air Blank	0.000	18:01
Subject Sample # 2	0.039	18:01
Breath Volume	3.199	18:01
Air Blank	0.000	18:02
Control Test	0.077	18:02
Air Blank	0.000	18:03
Diagnostics Check	OK	18:03

Data Download Date/Time: 08/11/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RADDER

DATE/TIME OF TEST: 07/22/2021 00:42:52

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.078	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.215	00:37
Breath Volume	2.164	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.212	00:41
Breath Volume	2.042	00:41
Air Blank	0.000	00:41
Control Test	0.077	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 08/11/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYANT, BRANDEN, J
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: MAYNARD, J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:16

SUBJECT'S LAST NAME: BURKES

DATE/TIME OF TEST: 07/30/2021 15:42:47

Diagnostics Check	OK	15:36
Air Blank	0.000	15:36
Control Test	0.076	15:37
Air Blank	0.000	15:37
Subject Sample # 1	0.214	15:38
Breath Volume	1.292	15:38
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.209	15:41
Breath Volume	1.117	15:41
Air Blank	0.000	15:41
Control Test	0.075	15:42
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 08/11/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: STESLICKI

DATE/TIME OF TEST: 07/04/2021 01:20:46

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.082	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.272	01:12
Breath Volume		2.750	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.235	01:15
Breath Volume		1.113	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 3	0.258	01:19
Breath Volume		1.902	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 07/28/2021 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: TEXTOR

DATE/TIME OF TEST: 07/10/2021 00:34:03

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:23	
Control Test		0.081	00:23	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.219	00:27	Volume Not Met
Breath Volume		0.136	00:27	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:33	
Control Test		0.080	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 07/28/2021 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 07/22/2021 02:04:31

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.253	02:00
Breath Volume	1.851	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.253	02:03
Breath Volume	1.472	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 07/28/2021 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL, DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ANELLO

DATE/TIME OF TEST: 07/29/2021 04:34:21

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.082	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.000	04:29
Breath Volume	1.960	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.000	04:33
Breath Volume	2.730	04:33
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 08/31/2021 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: ZINN

DATE/TIME OF TEST: 07/30/2021 23:20:45

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.081	23:10	
Air Blank		0.000	23:10	
Subject Sample	# 1	0.072	23:13	Volume Not Met
Breath Volume		0.031	23:13	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 2		23:19	No Sample Provided
Breath Volume		0.000	23:19	
Air Blank		0.000	23:19	
Control Test		0.080	23:20	
Air Blank		0.000	23:20	
Diagnostics Check		OK	23:20	

Data Download Date/Time: 08/31/2021 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/31/2021 00:12:26

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.081	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.199	00:08
Breath Volume		2.097	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.203	00:10
Breath Volume		2.378	00:10
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 08/31/2021 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SHUMWAY

DATE/TIME OF TEST: 07/31/2021 02:08:48

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.117	02:04
Breath Volume		2.140	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.122	02:07
Breath Volume		2.261	02:07
Air Blank		0.000	02:07
Control Test		0.081	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 08/31/2021 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,JULIO,R
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,FINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 07/11/2021 06:08:01

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:06	
Control Test	0.081	06:06	
Air Blank	0.000	06:07	
Subject Sample	# 1	06:07	RFI Detect
Breath Volume	0.000	06:07	
Air Blank		06:07	RFI Detect
Air Blank	0.000	06:08	

Data Download Date/Time: 07/29/2021 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,JULIO,R
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,FINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 07/11/2021 06:13:55

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:11	
Control Test	0.080	06:11	
Air Blank	0.000	06:12	
Subject Sample	# 1	06:12	Subject Test Refused
Breath Volume	0.000	06:12	
Air Blank	0.000	06:13	
Control Test	0.079	06:13	
Air Blank	0.000	06:13	
Diagnostics Check	OK	06:13	

Data Download Date/Time: 07/29/2021 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORA,MANUEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTILLO,ISMAELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: SERAFIN

DATE/TIME OF TEST: 07/16/2021 16:08:06

Diagnostics Check	OK	16:01
Air Blank	0.000	16:01
Control Test	0.081	16:02
Air Blank	0.000	16:02
Subject Sample # 1	0.182	16:03
Breath Volume	1.785	16:03
Air Blank	0.000	16:04
Air Blank	0.000	16:05
Subject Sample # 2	0.177	16:06
Breath Volume	2.617	16:06
Air Blank	0.000	16:07
Control Test	0.080	16:07
Air Blank	0.000	16:08
Diagnostics Check	OK	16:08

Data Download Date/Time: 07/29/2021 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORA,MANUEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTILLO,ISMAELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: SERAFIN

DATE/TIME OF TEST: 07/16/2021 17:11:54

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:10	
Control Test	0.081	17:11	
Air Blank		17:11	RFI Detect
Air Blank	0.000	17:11	

Data Download Date/Time: 07/29/2021 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORA,MANUEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTILLO,I
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: SERAFIN

DATE/TIME OF TEST: 07/16/2021 17:14:35

Diagnostics Check	OK	17:13	
Air Blank		17:14	RFI Detect
Air Blank	0.000	17:14	

Data Download Date/Time: 07/29/2021 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,ROBERTO,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,HILDA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BRITO

DATE/TIME OF TEST: 07/17/2021 01:02:27

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 07/29/2021 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DE PRADO,DANIEL,L

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: DE PRADO,DANIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 989917

DRY GAS EXPIRY DATE: 07/25/2021

VIOLATION CODE: DUI

ARREST TIME: 07:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 07/17/2021 08:17:41

Diagnostics Check	OK	08:16	
Air Blank	0.000	08:17	
Control Test	0.080	08:17	
Air Blank		08:17	RFI Detect
Air Blank		08:17	RFI Detect

Data Download Date/Time: 07/29/2021 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE PRADO,DANIEL,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DE PRADO,DANIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: TRAN
DATE/TIME OF TEST: 07/17/2021 08:42:37

Diagnostics Check	OK	08:34	
Air Blank	0.000	08:34	
Control Test	0.078	08:34	
Air Blank	0.000	08:35	
Subject Sample # 1	0.106	08:36	
Breath Volume	1.449	08:36	
Air Blank	0.000	08:37	
Air Blank	0.000	08:38	
Subject Sample # 2	0.104	08:41	
Breath Volume	1.750	08:41	
Air Blank		08:42	Purge Fail
Air Blank	0.000	08:42	

Data Download Date/Time: 07/29/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PEREZ,VICTOR
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: MORENO CRUZ

DATE/TIME OF TEST: 07/24/2021 20:09:15

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:05	
Control Test	0.082	20:06	
Air Blank	0.000	20:06	
Subject Sample	# 1	20:07	Subject Test Refused
Breath Volume	0.000	20:07	
Air Blank	0.000	20:08	
Control Test	0.081	20:08	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 07/29/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,ALEXIA,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BAUZA,TANIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/25/2021 04:44:56

Diagnostics Check	OK	04:44	
Air Blank		04:44	RFI Detect
Air Blank		04:44	RFI Detect

Data Download Date/Time: 07/29/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,ALEXIA,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BAUZA,TANIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/25/2021 04:49:24

Diagnostics Check	OK	04:48	
Air Blank		04:48	RFI Detect
Air Blank	0.000	04:49	

Data Download Date/Time: 07/29/2021 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,ALEXIA,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BAUZA,TANIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/25/2021 04:59:08

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.082	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.231	04:54
Breath Volume		2.121	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.225	04:57
Breath Volume		1.414	04:57
Air Blank		0.000	04:58
Control Test		0.081	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 07/29/2021 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISI,BERNARDO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GRISI,BERNARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/25/2021 07:13:03

Diagnostics Check	OK	07:10	
Air Blank	0.000	07:10	
Control Test	0.081	07:10	
Air Blank	0.000	07:11	
Subject Sample	# 1	07:11	Subject Test Refused
Breath Volume	0.000	07:11	
Air Blank	0.000	07:12	
Control Test	0.081	07:12	
Air Blank	0.000	07:12	
Diagnostics Check	OK	07:13	

Data Download Date/Time: 07/29/2021 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORALUZZO,MICHAEL,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CORALUZZO,CORALUZZO
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: FURLONG

DATE/TIME OF TEST: 07/02/2021 00:03:43

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.081	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.085	23:59
Breath Volume	3.218	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.078	00:02
Breath Volume	1.617	00:02
Air Blank	0.000	00:02
Control Test	0.081	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMDEN,DENNIS,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: LIU

DATE/TIME OF TEST: 07/02/2021 05:37:47

Diagnostics Check		OK	05:30
Air Blank		0.000	05:31
Control Test		0.081	05:31
Air Blank		0.000	05:32
Subject Sample	# 1	0.068	05:33
Breath Volume		2.406	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.070	05:36
Breath Volume		2.367	05:36
Air Blank		0.000	05:36
Control Test		0.082	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EVANS,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 07/05/2021 20:52:11

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.082	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.200	20:47
Breath Volume	3.140	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.205	20:50
Breath Volume	1.683	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CRAWFORD,JUDSON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 07/06/2021 00:12:38

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.371	00:05
Breath Volume	2.230	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.338	00:08
Breath Volume	1.988	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 3	0.371	00:11
Breath Volume	2.308	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: CAVICCHIO

DATE/TIME OF TEST: 07/07/2021 18:38:25

Diagnostics Check	OK	18:32
Air Blank	0.000	18:32
Control Test	0.082	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.000	18:34
Breath Volume	2.464	18:34
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.000	18:37
Breath Volume	2.531	18:37
Air Blank	0.000	18:37
Control Test	0.081	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMDEN,DENNIS,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: KISHEK

DATE/TIME OF TEST: 07/10/2021 02:53:42

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.062	02:49
Breath Volume	4.015	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.061	02:52
Breath Volume	3.460	02:52
Air Blank	0.000	02:52
Control Test	0.082	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EVANS,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 07/11/2021 05:51:32

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.081	05:45
Air Blank		0.000	05:46
Subject Sample	# 1	0.182	05:46
Breath Volume		4.523	05:46
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.184	05:50
Breath Volume		4.085	05:50
Air Blank		0.000	05:50
Control Test		0.080	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,CALEB,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HARVEY,CALEB
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: TRAVALEE

DATE/TIME OF TEST: 07/11/2021 06:47:49

Diagnostics Check	OK	06:46	
Air Blank	0.000	06:47	
		06:47	RFI Detect
Air Blank	0.000	06:47	

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,CALEB,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HARVEY,CALEB
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: TRAVALEE

DATE/TIME OF TEST: 07/11/2021 06:59:03

Diagnostics Check	OK	06:51
Air Blank	0.000	06:52
Control Test	0.081	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.121	06:54
Breath Volume	1.371	06:54
Air Blank	0.000	06:54
Air Blank	0.000	06:56
Subject Sample # 2	0.129	06:57
Breath Volume	2.273	06:57
Air Blank	0.000	06:58
Control Test	0.079	06:58
Air Blank	0.000	06:58
Diagnostics Check	OK	06:59

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BURKE,LYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 07/11/2021 21:02:47

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.081	20:58
Breath Volume	2.273	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:00
Subject Sample # 2	0.077	21:01
Breath Volume	2.234	21:01
Air Blank	0.000	21:01
Control Test	0.081	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/12/2021 18:34:59

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.082	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.139	18:29
Breath Volume	1.566	18:29
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.129	18:33
Breath Volume	2.457	18:33
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: KOKARAS

DATE/TIME OF TEST: 07/12/2021 22:31:19

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.059	22:26
Breath Volume	1.460	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.065	22:29
Breath Volume	1.386	22:29
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: GUILLOTTE

DATE/TIME OF TEST: 07/14/2021 02:13:45

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.000	02:09
Breath Volume	2.660	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.000	02:12
Breath Volume	1.886	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CRAWFORD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ERIKSON

DATE/TIME OF TEST: 07/15/2021 23:41:13

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.082	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.176	23:36
Breath Volume	1.656	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.172	23:39
Breath Volume	1.792	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMDEN,DENNIS,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: JONES-MORENO

DATE/TIME OF TEST: 07/16/2021 03:16:12

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.081	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Control Test	0.082	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,C
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:14

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 07/16/2021 11:40:57

Diagnostics Check	OK	11:34
Air Blank	0.000	11:35
Control Test	0.081	11:35
Air Blank	0.000	11:36
Subject Sample # 1	0.000	11:36
Breath Volume	2.421	11:36
Air Blank	0.000	11:37
Air Blank	0.000	11:39
Subject Sample # 2	0.000	11:39
Breath Volume	2.210	11:39
Air Blank	0.000	11:40
Control Test	0.082	11:40
Air Blank	0.000	11:40
Diagnostics Check	OK	11:40

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LEAR,T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:33

SUBJECT'S LAST NAME: BEVERLY

DATE/TIME OF TEST: 07/17/2021 10:59:25

Diagnostics Check	OK	10:53
Air Blank	0.000	10:53
Control Test	0.081	10:54
Air Blank	0.000	10:54
Subject Sample # 1	0.000	10:55
Breath Volume	2.171	10:55
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.000	10:58
Breath Volume	1.738	10:58
Air Blank	0.000	10:58
Control Test	0.081	10:58
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: LUONGO

DATE/TIME OF TEST: 07/20/2021 20:06:31

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.076	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.140	20:02
Breath Volume	2.718	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.137	20:04
Breath Volume	3.832	20:04
Air Blank	0.000	20:05
Control Test	0.079	20:05
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 07/23/2021 00:25:35

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.081	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1	0.169	00:21	Volume Not Met
Breath Volume		0.656	00:21	
Air Blank		0.000	00:22	
Air Blank		0.000	00:24	
Subject Sample	# 2		00:24	Subject Test Refused
Breath Volume		0.000	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:24	
Air Blank		0.000	00:25	
Diagnostics Check		OK	00:25	

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WINTERS,ERIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 07/23/2021 18:00:24

Diagnostics Check	OK	17:49	
Air Blank	0.000	17:49	
Control Test	0.082	17:49	
Air Blank	0.000	17:50	
Subject Sample	# 1	17:53	No Sample Provided
Breath Volume	0.000	17:53	
Air Blank	0.000	17:53	
Air Blank	0.000	17:55	
Subject Sample	# 2	17:59	No Sample Provided
Breath Volume	0.000	17:59	
Air Blank	0.000	17:59	
Control Test	0.081	17:59	
Air Blank	0.000	18:00	
Diagnostics Check	OK	18:00	

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMDEN,DENNIS,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 07/24/2021 00:05:43

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.082	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:01
Breath Volume	3.664	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:04
Breath Volume	2.429	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EVANS,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BEAL

DATE/TIME OF TEST: 07/24/2021 03:01:48

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.082	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:54	No Sample Provided
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 2	03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EVANS,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 07/25/2021 03:13:06

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.081	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:11	No Sample Provided
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.080	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 07/29/2021 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WILLIAMS,SHELTON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 07/30/2021 03:39:13

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.081	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1	0.201	03:28	Volume Not Met
Breath Volume		0.125	03:28	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.190	03:33	
Breath Volume		1.656	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 3	0.187	03:37	
Breath Volume		1.320	03:37	
Air Blank		0.000	03:38	
Control Test		0.076	03:38	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 08/27/2021 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 07/30/2021 19:47:01

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.081	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.247	19:42
Breath Volume	2.812	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:44
Subject Sample # 2	0.236	19:45
Breath Volume	2.957	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:47

Data Download Date/Time: 08/27/2021 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/31/2021 01:27:46

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.141	01:23
Breath Volume	1.988	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.143	01:26
Breath Volume	2.925	01:26
Air Blank	0.000	01:26
Control Test	0.081	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 08/27/2021 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 07/31/2021 18:58:46

Diagnostics Check	OK	18:55	
Air Blank	0.000	18:55	
Control Test	0.081	18:56	
Air Blank	0.000	18:56	
Subject Sample	# 1	18:57	Subject Test Refused
Breath Volume	0.000	18:57	
Air Blank	0.000	18:57	
Control Test	0.082	18:58	
Air Blank	0.000	18:58	
Diagnostics Check	OK	18:58	

Data Download Date/Time: 08/27/2021 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LANTZ

DATE/TIME OF TEST: 07/31/2021 21:57:31

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.082	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.112	21:53
Breath Volume	3.371	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.107	21:56
Breath Volume	4.378	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 08/27/2021 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BROWN,J
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: BAYER

DATE/TIME OF TEST: 07/31/2021 22:16:16

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.232	22:11
Breath Volume	1.984	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.242	22:14
Breath Volume	3.097	22:14
Air Blank	0.000	22:15
Control Test	0.081	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 08/27/2021 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,EDWIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: COLON,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ALMONTE

DATE/TIME OF TEST: 07/10/2021 03:44:53

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.081	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	RFI Detect
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 07/26/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,EDWIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: COLON,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ALMONTE

DATE/TIME OF TEST: 07/10/2021 03:53:31

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.082	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.070	03:49
Breath Volume		2.800	03:49
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.068	03:51
Breath Volume		2.847	03:51
Air Blank		0.000	03:52
Control Test		0.081	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 07/26/2021 12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,EDWIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: COLON,EDWIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 07/14/2021 05:19:20

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.082	05:13	
Air Blank	0.000	05:13	
Subject Sample # 1	0.161	05:15	
Breath Volume	1.371	05:15	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample # 2	0.174	05:17	
Breath Volume	1.136	05:17	
Air Blank		05:18	Ambient Fail
Air Blank	0.000	05:19	

Data Download Date/Time: 07/26/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,EDWIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BEHRENSHOUSER,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/25/2021 03:50:25

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.082	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.093	03:45
Breath Volume	4.179	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.093	03:48
Breath Volume	3.773	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 07/26/2021 12:13