Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MULL

DATE/TIME OF TEST: 07/03/2021 00:12:27

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.077	00:06
Air Blank		0.000	00:07
Subject Sample	# 1	0.181	00:07
Breath Volume		2.746	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.179	00:10
Breath Volume		2.558	00:10
Air Blank		0.000	00:11
Control Test		0.076	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 07/12/2021 09:14

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VAN DE ROER

DATE/TIME OF TEST: 07/13/2021 03:44:42

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.081	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.000	03:39
Breath Volume		2.128	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.000	03:43
Breath Volume		1.492	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 08/09/2021 15:31

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: MALOY

DATE/TIME OF TEST: 07/14/2021 19:00:15

Diagnostics Check		OK	18:54
Air Blank		0.000	18:54
Control Test		0.079	18:54
Air Blank		0.000	18:55
Subject Sample	# 1	0.035	18:55
Breath Volume		2.378	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:58
Subject Sample	# 2	0.033	18:58
Breath Volume		2.660	18:58
Air Blank		0.000	18:59
Control Test		0.080	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 08/09/2021 15:31

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BOLDEN

DATE/TIME OF TEST: 07/16/2021 00:11:20

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:08	
Control Test		0.080	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1		00:09	Subject Test Refused
Breath Volume		0.000	00:09	
Air Blank		0.000	00:10	
Control Test		0.079	00:10	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 08/09/2021 15:31

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACK,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CAROLLO

DATE/TIME OF TEST: 07/17/2021 02:23:37

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.227	02:18
Breath Volume		1.570	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.239	02:21
Breath Volume		2.851	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 08/09/2021 15:31

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACK,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: YACK,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LAYETT VALECILLOS

DATE/TIME OF TEST: 07/18/2021 03:32:31

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.131	03:26
Breath Volume		1.910	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.117	03:30
Breath Volume		1.367	03:30
Air Blank		0.000	03:31
Control Test		0.076	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 08/09/2021 15:31

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACK,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: YACK,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: GARCIA VAZQUEZ

DATE/TIME OF TEST: 07/19/2021 05:07:31

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.080	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1		05:06	Subject Test Refused
Breath Volume		0.000	05:06	
Air Blank		0.000	05:06	
Control Test		0.080	05:06	
Air Blank		0.000	05:07	
Diagnostics Check		OK	05:07	

Data Download Date/Time: 08/09/2021 15:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: MEDRZYCKI

DATE/TIME OF TEST: 07/19/2021 16:12:45

Diagnostics Check		OK	16:00	
Air Blank		0.000	16:01	
Control Test		0.081	16:01	
Air Blank		0.000	16:02	
Subject Sample	# 1	0.112	16:05	Volume Not Met
Breath Volume		0.757	16:05	
Air Blank		0.000	16:05	
Air Blank		0.000	16:07	
Subject Sample	# 2	0.132	16:08	
Breath Volume		2.292	16:08	
Air Blank		0.000	16:08	
Air Blank		0.000	16:10	
Subject Sample	# 3	0.133	16:11	
Breath Volume		2.285	16:11	
Air Blank		0.000	16:11	
Control Test		0.080	16:12	
Air Blank		0.000	16:12	
Diagnostics Check		OK	16:12	

Data Download Date/Time: 08/09/2021 15:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PONTBRIAND

DATE/TIME OF TEST: 07/20/2021 22:57:37

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.080	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.159	22:53
Breath Volume		1.808	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.164	22:56
Breath Volume		2.222	22:56
Air Blank		0.000	22:56
Control Test		0.080	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 08/09/2021 15:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/25/2021 18:40:51

Diagnostics Check		OK	18:34
Air Blank		0.000	18:34
Control Test		0.080	18:34
Air Blank		0.000	18:35
Subject Sample	# 1	0.363	18:35
Breath Volume		2.351	18:35
Air Blank		0.000	18:36
Air Blank		0.000	18:38
Subject Sample	# 2	0.357	18:39
Breath Volume		2.261	18:39
Air Blank		0.000	18:39
Control Test		0.079	18:40
Air Blank		0.000	18:40
Diagnostics Check		OK	18:40

Data Download Date/Time: 08/09/2021 15:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: VANCE

DATE/TIME OF TEST: 07/29/2021 00:38:50

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.123	00:33
Breath Volume		1.972	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.118	00:37
Breath Volume		1.507	00:37
Air Blank		0.000	00:37
Control Test		0.080	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 08/09/2021 15:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 07/29/2021 18:25:02

Diagnostics Check		OK	18:19
Air Blank		0.000	18:19
Control Test		0.080	18:19
Air Blank		0.000	18:20
Subject Sample	# 1	0.000	18:20
Breath Volume		1.589	18:20
Air Blank		0.000	18:21
Air Blank		0.000	18:23
Subject Sample	# 2	0.000	18:23
Breath Volume		1.640	18:23
Air Blank		0.000	18:24
Control Test		0.081	18:24
Air Blank		0.000	18:24
Diagnostics Check		OK	18:25

Data Download Date/Time: 08/09/2021 15:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER III,RICHARD,F
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER III,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: HAYDEN

DATE/TIME OF TEST: 07/29/2021 19:03:05

Diagnostics Check		OK	18:56
Air Blank		0.000	18:56
Control Test		0.080	18:57
Air Blank		0.000	18:57
Subject Sample	# 1	0.124	18:58
Breath Volume		1.101	18:58
Air Blank		0.000	18:58
Air Blank		0.000	19:00
Subject Sample	# 2	0.124	19:01
Breath Volume		1.464	19:01
Air Blank		0.000	19:02
Control Test		0.080	19:02
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 08/09/2021 15:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PAZ DEL SOL

DATE/TIME OF TEST: 07/01/2021 04:38:36

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.234	04:34
Breath Volume		1.226	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.248	04:36
Breath Volume		1.726	04:36
Air Blank		0.000	04:37
Control Test		0.079	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: NUNN

DATE/TIME OF TEST: 07/02/2021 01:19:23

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.081	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.143	01:14
Breath Volume		1.792	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:16
Subject Sample	# 2	0.145	01:17
Breath Volume		2.679	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: DORVILUS

DATE/TIME OF TEST: 07/02/2021 05:04:07

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.081	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.159	04:54
Breath Volume		1.359	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:56
Subject Sample	# 2	0.137	04:58
Breath Volume		1.558	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 3	0.127	05:02
Breath Volume		1.707	05:02
Air Blank		0.000	05:03
Control Test		0.078	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 07/03/2021 00:24:14

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.081	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.153	00:19
Breath Volume		1.164	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:21
Subject Sample	# 2	0.167	00:22
Breath Volume		1.148	00:22
Air Blank		0.000	00:23
Control Test		0.080	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAIRSTON,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 07/03/2021 01:40:11

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.099	01:35
Breath Volume		3.496	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.096	01:38
Breath Volume		3.492	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: OVERBY

DATE/TIME OF TEST: 07/03/2021 04:04:12

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:51	
Control Test		0.080	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.075	03:55	Volume Not Met
Breath Volume		0.441	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.084	03:59	
Breath Volume		1.156	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 3	0.080	04:02	
Breath Volume		1.410	04:02	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PLAISIR

DATE/TIME OF TEST: 07/03/2021 05:09:44

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.081	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.205	05:05
Breath Volume		1.757	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.212	05:08
Breath Volume		2.566	05:08
Air Blank		0.000	05:08
Control Test		0.079	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: GARCIA TORRES

DATE/TIME OF TEST: 07/03/2021 06:10:16

Diagnostics Check		OK	06:02
Air Blank		0.000	06:03
Control Test		0.080	06:03
Air Blank		0.000	06:04
Subject Sample	# 1	0.184	06:05
Breath Volume		1.445	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 2	0.184	06:08
Breath Volume		1.195	06:08
Air Blank		0.000	06:09
Control Test		0.079	06:09
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: UTRERA VIERA

DATE/TIME OF TEST: 07/03/2021 21:34:28

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.188	21:29
Breath Volume		1.851	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.191	21:32
Breath Volume		2.210	21:32
Air Blank		0.000	21:33
Control Test		0.080	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 07/15/2021 15:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/04/2021 00:44:32

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.175	00:39
Breath Volume		2.449	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.185	00:42
Breath Volume		2.109	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 07/04/2021 01:54:03

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.195	01:49
Breath Volume		1.820	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.196	01:52
Breath Volume		1.621	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:54

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: COLIS

DATE/TIME OF TEST: 07/04/2021 04:36:37

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.078	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.136	04:31
Breath Volume		1.500	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.137	04:35
Breath Volume		1.988	04:35
Air Blank		0.000	04:35
Control Test		0.078	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: WONG ITURRIAGA

DATE/TIME OF TEST: 07/04/2021 05:03:16

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.174	04:58
Breath Volume		1.582	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:00
Subject Sample	# 2	0.185	05:01
Breath Volume		2.058	05:01
Air Blank		0.000	05:02
Control Test		0.077	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: ORELLANA REGALADO

DATE/TIME OF TEST: 07/04/2021 05:28:32

Diagnostics Check		OK	05:21
Air Blank		0.000	05:22
Control Test		0.077	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.177	05:23
Breath Volume		1.398	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.195	05:26
Breath Volume		1.742	05:26
Air Blank		0.000	05:27
Control Test		0.077	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 07/04/2021 07:04:05

Diagnostics Check		OK	06:56
Air Blank		0.000	06:57
Control Test		0.077	06:57
Air Blank		0.000	06:58
Subject Sample	# 1	0.000	06:59
Breath Volume		2.218	06:59
Air Blank		0.000	07:00
Air Blank		0.000	07:02
Subject Sample	# 2	0.000	07:02
Breath Volume		1.558	07:02
Air Blank		0.000	07:03
Control Test		0.077	07:03
Air Blank		0.000	07:04
Diagnostics Check		OK	07:04

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/05/2021 04:21:14

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:15	
Control Test		0.077	04:15	
Air Blank		0.000	04:16	
Subject Sample	# 1	0.149	04:17	
Breath Volume		1.410	04:17	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2		04:20	Sequence Aborted
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/05/2021 04:52:32

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.077	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.147	04:47
Breath Volume		1.792	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.147	04:50
Breath Volume		1.464	04:50
Air Blank		0.000	04:51
Control Test		0.076	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: ULMER

DATE/TIME OF TEST: 07/05/2021 07:41:03

Diagnostics Check		OK	07:31
Air Blank		0.000	07:31
Control Test		0.076	07:32
Air Blank		0.000	07:32
Subject Sample	# 1	0.228	07:33
Breath Volume		1.722	07:33
Air Blank		0.000	07:34
Air Blank		0.000	07:35
Subject Sample	# 2	0.201	07:36
Breath Volume		1.855	07:36
Air Blank		0.000	07:37
Air Blank		0.000	07:38
Subject Sample	# 3	0.227	07:39
Breath Volume		2.964	07:39
Air Blank		0.000	07:40
Control Test		0.076	07:40
Air Blank		0.000	07:40
Diagnostics Check		OK	07:41

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 07/05/2021 23:15:37

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.079	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.000	23:11
Breath Volume		2.695	23:11
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.000	23:14
Breath Volume		1.503	23:14
Air Blank		0.000	23:14
Control Test		0.078	23:15
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 07/06/2021 02:47:19

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.077	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.109	02:42
Breath Volume		4.156	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.104	02:45
Breath Volume		3.449	02:45
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 07/08/2021 02:42:40

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.110	02:37
Breath Volume		1.281	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.112	02:41
Breath Volume		1.328	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 07/15/2021 16:00

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 07/08/2021 04:31:10

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.150	04:26
Breath Volume		1.710	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.149	04:29
Breath Volume		1.714	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: BURGOS

DATE/TIME OF TEST: 07/08/2021 14:02:10

Diagnostics Check		OK	13:54
Air Blank		0.000	13:55
Control Test		0.081	13:55
Air Blank		0.000	13:56
Subject Sample	# 1	0.108	13:57
Breath Volume		1.753	13:57
Air Blank		0.000	13:58
Air Blank		0.000	14:00
Subject Sample	# 2	0.108	14:00
Breath Volume		1.351	14:00
Air Blank		0.000	14:01
Control Test		0.080	14:01
Air Blank		0.000	14:02
Diagnostics Check		OK	14:02

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 07/08/2021 23:01:12

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:47	
Control Test		0.081	22:48	
Air Blank		0.000	22:48	
Subject Sample	# 1	0.120	22:50	
Breath Volume		1.589	22:50	
Air Blank		0.000	22:51	
Air Blank		0.000	22:53	
Subject Sample	# 2	0.122	22:56	Volume Not Met
Breath Volume		0.238	22:56	
Air Blank		0.000	22:57	
Air Blank		0.000	22:58	
Subject Sample	# 3	0.115	22:59	
Breath Volume		1.726	22:59	
Air Blank		0.000	23:00	
Control Test		0.078	23:00	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:01	

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: HERNANDEZ II

DATE/TIME OF TEST: 07/09/2021 03:56:12

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.185	03:51
Breath Volume		1.843	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.198	03:54
Breath Volume		1.429	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/09/2021 20:14:26

Diagnostics Check		OK	20:06
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:07
Subject Sample	# 1	0.215	20:08
Breath Volume		1.320	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:10
Subject Sample	# 2	0.204	20:12
Breath Volume		1.695	20:12
Air Blank		0.000	20:13
Control Test		0.080	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: RUSCH

DATE/TIME OF TEST: 07/09/2021 23:07:58

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.150	23:03
Breath Volume		3.261	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:05
Subject Sample	# 2	0.141	23:06
Breath Volume		3.222	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 07/10/2021 00:51:29

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.080	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.170	00:38	
Breath Volume		1.964	00:38	
Air Blank		0.000	00:39	
Air Blank		0.000	00:41	
Subject Sample	# 2	0.135	00:44	Volume Not Met
Breath Volume		0.082	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 3	0.167	00:49	
Breath Volume		1.371	00:49	
Air Blank		0.000	00:50	
Control Test		0.078	00:50	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NAVARRO III

DATE/TIME OF TEST: 07/10/2021 04:46:55

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.172	04:42
Breath Volume		1.500	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.173	04:45
Breath Volume		2.078	04:45
Air Blank		0.000	04:45
Control Test		0.079	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: O'NEILL

DATE/TIME OF TEST: 07/10/2021 22:26:37

Diagnostics Check		OK	22:17
Air Blank		0.000	22:17
Control Test		0.081	22:18
Air Blank		0.000	22:18
Subject Sample	# 1	0.259	22:20
Breath Volume		1.578	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.256	22:24
Breath Volume		1.656	22:24
Air Blank		0.000	22:25
Control Test		0.079	22:25
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HURTADO-SANCHEZ

DATE/TIME OF TEST: 07/11/2021 23:47:20

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.132	23:42
Breath Volume		3.554	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.133	23:45
Breath Volume		3.328	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOREJON,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 07/12/2021 22:28:33

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.081	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.126	22:23
Breath Volume		1.410	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.117	22:26
Breath Volume		1.640	22:26
Air Blank		0.000	22:27
Control Test		0.081	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CARRASCO-AGUILAR

DATE/TIME OF TEST: 07/14/2021 02:43:16

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.211	02:36
Breath Volume		1.867	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.221	02:41
Breath Volume		2.074	02:41
Air Blank		0.000	02:42
Control Test		0.080	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 07/15/2021 16:01

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 07/16/2021 23:25:11

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:12	
Control Test		0.081	23:12	
Air Blank		0.000	23:12	
Subject Sample	# 1		23:16	No Sample Provided
Breath Volume		0.000	23:16	
Air Blank		0.000	23:16	
Air Blank		0.000	23:18	
Subject Sample	# 2	0.297	23:19	
Breath Volume		2.367	23:19	
Air Blank		0.000	23:20	
Air Blank		0.000	23:22	
Subject Sample	# 3	0.295	23:23	
Breath Volume		1.972	23:23	
Air Blank		0.000	23:24	
Control Test		0.079	23:24	
Air Blank		0.000	23:25	
Diagnostics Check		OK	23:25	

Data Download Date/Time: 08/11/2021 13:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MONDRAGON VELASQUEZ

DATE/TIME OF TEST: 07/17/2021 03:00:53

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.163	02:56
Breath Volume		2.160	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.166	02:59
Breath Volume		2.269	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 08/11/2021 13:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: PANTALEON RAMIREZ

DATE/TIME OF TEST: 07/17/2021 04:31:45

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.190	04:27
Breath Volume		2.496	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.187	04:30
Breath Volume		2.878	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ROSADO JR

DATE/TIME OF TEST: 07/18/2021 03:08:32

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.145	03:04
Breath Volume		2.363	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.143	03:06
Breath Volume		2.359	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: TORRES MARCHIANI

DATE/TIME OF TEST: 07/23/2021 01:02:44

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.081	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.227	00:58
Breath Volume		2.960	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:00
Subject Sample	# 2	0.214	01:01
Breath Volume		1.558	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CARNEY

DATE/TIME OF TEST: 07/25/2021 00:41:54

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:36
Breath Volume		1.257	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.000	00:40
Breath Volume		1.855	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PARKER,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/25/2021 03:15:59

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.080	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.000	03:11
Breath Volume		2.878	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.000	03:14
Breath Volume		3.167	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 07/25/2021 05:24:49

Diagnostics Check		OK	05:13	
Air Blank		0.000	05:13	
Control Test		0.079	05:13	
Air Blank		0.000	05:14	
Subject Sample	# 1		05:17	No Sample Provided
Breath Volume		0.000	05:17	
Air Blank		0.000	05:18	
Air Blank		0.000	05:20	
Subject Sample	# 2		05:23	No Sample Provided
Breath Volume		0.000	05:23	
Air Blank		0.000	05:23	
Control Test		0.077	05:24	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: JIMENEZ PEREZ

DATE/TIME OF TEST: 07/27/2021 00:11:38

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.160	00:05
Breath Volume		2.398	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.156	00:09
Breath Volume		2.042	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CONNELL,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/29/2021 04:00:06

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.025	03:55
Breath Volume		1.886	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.021	03:58
Breath Volume		1.179	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CROOKS

DATE/TIME OF TEST: 07/31/2021 00:52:07

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Subject Sample	# 1	0.161	00:46
Breath Volume		1.394	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.151	00:50
Breath Volume		1.527	00:50
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CAMARENA-MARTINEZ

DATE/TIME OF TEST: 07/31/2021 02:25:14

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.186	02:20
Breath Volume		2.593	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.182	02:23
Breath Volume		1.945	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 08/11/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 07/31/2021 02:51:34

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:37	
Control Test		0.080	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1	0.204	02:41	Volume Not Met
Breath Volume		0.175	02:41	
Air Blank		0.000	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 2	0.198	02:46	
Breath Volume		1.277	02:46	
Air Blank		0.000	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 3	0.204	02:49	
Breath Volume		1.367	02:49	
Air Blank		0.000	02:50	
Control Test		0.077	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 08/11/2021 13:41

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 07/31/2021 04:09:35

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.131	04:04
Breath Volume		1.339	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.132	04:07
Breath Volume		2.539	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 08/11/2021 13:41

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 07/31/2021 04:43:47

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:31	
Control Test		0.080	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.106	04:35	Volume Not Met
Breath Volume		0.132	04:35	
Air Blank		0.000	04:35	
Air Blank		0.000	04:37	
Subject Sample	# 2	0.121	04:38	
Breath Volume		1.101	04:38	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 3	0.115	04:42	
Breath Volume		1.457	04:42	
Air Blank		0.000	04:42	
Control Test		0.079	04:43	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 08/11/2021 13:41

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: HUEWITT

DATE/TIME OF TEST: 07/31/2021 05:07:24

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:03	
Control Test		0.079	05:04	
Air Blank		0.000	05:04	
Subject Sample	# 1		05:06	Subject Test Refused
Breath Volume		0.000	05:06	
Air Blank		0.000	05:06	
Control Test		0.080	05:06	
Air Blank		0.000	05:07	
Diagnostics Check		OK	05:07	

Data Download Date/Time: 08/11/2021 13:41

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SLAGA,JASMINE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 07/03/2021 04:57:13

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.081	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.143	04:52
Breath Volume		1.625	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.143	04:55
Breath Volume		1.226	04:55
Air Blank		0.000	04:56
Control Test		0.080	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 07/08/2021 09:19

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MARKWELL

DATE/TIME OF TEST: 07/04/2021 00:09:30

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.080	00:07	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:08	Subject Test Refused
Breath Volume		0.000	00:08	
Air Blank		0.000	00:08	
Control Test		0.080	00:08	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 07/08/2021 09:23

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ABERCROMBIE

DATE/TIME OF TEST: 07/04/2021 02:17:28

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.217	02:09
Breath Volume		1.750	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:11
Subject Sample	# 2	0.238	02:12
Breath Volume		1.222	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 3	0.221	02:15
Breath Volume		2.187	02:15
Air Blank		0.000	02:16
Control Test		0.076	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 07/08/2021 09:23

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISS,ADAM,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GREEN,SETH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 07/05/2021 01:45:00

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.256	01:40
Breath Volume		5.292	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:42
Subject Sample	# 2	0.251	01:43
Breath Volume		5.285	01:43
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:45

Data Download Date/Time: 07/08/2021 09:23

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISS,ADAM,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WICHMAN,MICHELLE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 07/05/2021 04:08:12

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.000	04:03
Breath Volume		4.820	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.000	04:06
Breath Volume		3.312	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 07/08/2021 09:23

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ZALAK,JONATHAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CUSICK

DATE/TIME OF TEST: 07/07/2021 23:43:19

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.081	23:41	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.080	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 07/08/2021 09:24

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GREEN,SETH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 07/10/2021 00:46:39

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:44	
Control Test		0.081	00:44	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:45	Subject Test Refused
Breath Volume		0.000	00:45	
Air Blank		0.000	00:45	
Control Test		0.081	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 08/19/2021 14:58

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/10/2021 02:38:45

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: CHELLAND

DATE/TIME OF TEST: 07/13/2021 18:39:09

Diagnostics Check		OK	18:35	
Air Blank		0.000	18:36	
Control Test		0.081	18:36	
Air Blank		0.000	18:37	
Subject Sample	# 1		18:37	Subject Test Refused
Breath Volume		0.000	18:37	
Air Blank		0.000	18:38	
Control Test		0.080	18:38	
Air Blank		0.000	18:39	
Diagnostics Check		OK	18:39	

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: FAIRCLOTH,EDWARD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SARDEN

DATE/TIME OF TEST: 07/25/2021 01:39:28

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.104	01:34
Breath Volume		2.003	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.103	01:37
Breath Volume		1.808	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PALMADESSO

DATE/TIME OF TEST: 07/25/2021 23:57:52

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.081	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1	0.200	23:50	Volume Not Met
Breath Volume		0.660	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.220	23:56	Volume Not Met
Breath Volume		1.011	23:56	
Air Blank		0.000	23:56	
Control Test		0.077	23:57	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: '
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: PALMADESSO

DATE/TIME OF TEST: 07/26/2021 00:07:23

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.271	00:02
Breath Volume		2.976	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.265	00:05
Breath Volume		2.765	00:05
Air Blank		0.000	00:06
Control Test		0.076	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR,EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CLAY

DATE/TIME OF TEST: 07/30/2021 01:23:29

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.081	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.080	01:22	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PIERRE,STEVE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 07/30/2021 21:11:22

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:08	
Control Test		0.080	21:09	
Air Blank		0.000	21:09	
Subject Sample	# 1		21:09	Subject Test Refused
Breath Volume		0.000	21:09	
Air Blank		0.000	21:10	
Control Test		0.080	21:10	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PIERRE,STEVE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 07/30/2021 21:34:22

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.286	21:29
Breath Volume		1.699	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.289	21:32
Breath Volume		1.777	21:32
Air Blank		0.000	21:33
Control Test		0.078	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WICHMAN,MICHELLE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: VERZI

DATE/TIME OF TEST: 07/31/2021 00:44:11

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:39	
Control Test		0.080	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1	0.255	00:40	
Breath Volume		3.046	00:40	
Air Blank		0.000	00:41	
Air Blank			00:42	Improper Sample
Air Blank		0.000	00:43	
Control Test		0.078	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WICHMAN,MICHELLE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: VERZI

DATE/TIME OF TEST: 07/31/2021 00:52:35

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.078	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.245	00:48
Breath Volume		1.929	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.259	00:50
Breath Volume		2.351	00:50
Air Blank		0.000	00:51
Control Test		0.077	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 08/19/2021 14:59

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,HALEY,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 07/01/2021 23:31:57

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.079	23:29	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:30	Subject Test Refused
Breath Volume		0.000	23:30	
Air Blank		0.000	23:31	
Control Test		0.078	23:31	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 07/12/2021 13:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GOLDEN,LISA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BERKLEY

DATE/TIME OF TEST: 07/02/2021 02:40:56

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.131	02:33
Breath Volume		1.480	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.124	02:39
Breath Volume		1.339	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 07/12/2021 13:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,HALEY,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HEDRICK

DATE/TIME OF TEST: 07/02/2021 03:44:10

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:38	
Control Test		0.078	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1	0.211	03:39	
Breath Volume		2.640	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 2	0.192	03:42	
Breath Volume		1.183	03:42	
Air Blank		0.000	03:43	
Control Test		0.074	03:43	Control Outside Tolerance
Air Blank		0.000	03:44	

Data Download Date/Time: 07/12/2021 13:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,HALEY,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/24/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HEDRICK

DATE/TIME OF TEST: 07/02/2021 03:54:32

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.076	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.204	03:49
Breath Volume		2.296	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.195	03:52
Breath Volume		1.265	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/12/2021 13:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIDRON,JULIO,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: TAYLOR,SANDRA
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DOBARGANES

DATE/TIME OF TEST: 07/05/2021 02:01:16

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.173	01:51
Breath Volume		1.687	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.206	01:55
Breath Volume		3.648	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 3	0.208	01:59
Breath Volume		3.582	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: ALMECIGA

DATE/TIME OF TEST: 07/06/2021 17:47:01

Diagnostics Check		OK	17:40
Air Blank		0.000	17:41
Control Test		0.079	17:41
Air Blank		0.000	17:42
Subject Sample	# 1	0.186	17:42
Breath Volume		1.699	17:42
Air Blank		0.000	17:43
Air Blank		0.000	17:44
Subject Sample	# 2	0.177	17:45
Breath Volume		1.222	17:45
Air Blank		0.000	17:46
Control Test		0.078	17:46
Air Blank		0.000	17:46
Diagnostics Check		OK	17:47

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CANTELAR,JOAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: HALLIVIS

DATE/TIME OF TEST: 07/08/2021 17:30:25

Diagnostics Check		OK	17:24
Air Blank		0.000	17:24
Control Test		0.079	17:24
Air Blank		0.000	17:25
Subject Sample	# 1	0.073	17:25
Breath Volume		1.585	17:25
Air Blank		0.000	17:26
Air Blank		0.000	17:28
Subject Sample	# 2	0.073	17:28
Breath Volume		1.511	17:28
Air Blank		0.000	17:29
Control Test		0.079	17:29
Air Blank		0.000	17:30
Diagnostics Check		OK	17:30

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROJAS,GUILLERMO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/11/2021 21:53:19

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.081	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.000	21:48
Breath Volume		2.949	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.000	21:51
Breath Volume		3.179	21:51
Air Blank		0.000	21:52
Control Test		0.079	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ELLER

DATE/TIME OF TEST: 07/17/2021 02:46:47

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.082	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.000	02:42
Breath Volume		1.734	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.000	02:45
Breath Volume		1.484	02:45
Air Blank		0.000	02:45
Control Test		0.082	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO,ANGELA,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROSARIO,JUAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TZEP COJ

DATE/TIME OF TEST: 07/17/2021 04:20:07

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.081	04:19	
Air Blank		04:19	Improper Sample
Air Blank		04:20	Improper Sample

Data Download Date/Time: 08/02/2021 23:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO,ANGELA,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROSARIO,JUAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TZEP COJ

DATE/TIME OF TEST: 07/17/2021 04:22:34

Diagnostics Check	OK	04:22	
Air Blank		04:22	Improper Sample
Air Blank		04:22	Improper Sample

Data Download Date/Time: 08/02/2021 23:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO,ANGELA,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROSARIO,JUAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TZEP COJ

DATE/TIME OF TEST: 07/17/2021 04:27:07

Diagnostics Check	OK	04:26	
Air Blank		04:26	Improper Sample
Air Blank	0.000	04:27	

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO,ANGELA,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROSARIO,JUAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TZEP COJ

DATE/TIME OF TEST: 07/17/2021 04:36:31

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.078	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.213	04:31
Breath Volume		1.882	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:33
Subject Sample	# 2	0.196	04:34
Breath Volume		1.710	04:34
Air Blank		0.000	04:35
Control Test		0.078	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: BRAVO ABREU

DATE/TIME OF TEST: 07/17/2021 19:40:52

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.081	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.163	19:35
Breath Volume		1.429	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.176	19:39
Breath Volume		2.246	19:39
Air Blank		0.000	19:39
Control Test		0.078	19:40
Air Blank		0.000	19:40
Diagnostics Check		OK	19:40

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BRUNO,FRANCISCO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: MILTZ

DATE/TIME OF TEST: 07/18/2021 14:59:08

Diagnostics Check		OK	14:52
Air Blank		0.000	14:53
Control Test		0.082	14:53
Air Blank		0.000	14:54
Subject Sample	# 1	0.056	14:54
Breath Volume		2.691	14:54
Air Blank		0.000	14:55
Air Blank		0.000	14:57
Subject Sample	# 2	0.055	14:57
Breath Volume		2.925	14:57
Air Blank		0.000	14:58
Control Test		0.081	14:58
Air Blank		0.000	14:59
Diagnostics Check		OK	14:59

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROJAS,GUILLERMO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/18/2021 22:07:02

Diagnostics Check		OK	21:54
Air Blank		0.000	21:54
Control Test		0.080	21:54
Air Blank		0.000	21:55
Subject Sample	# 1	0.221	21:58
Breath Volume		3.816	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.183	22:01
Breath Volume		2.406	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:03
Subject Sample	# 3	0.191	22:05
Breath Volume		3.386	22:05
Air Blank		0.000	22:06
Control Test		0.076	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:07

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCCURN

DATE/TIME OF TEST: 07/20/2021 03:10:16

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.210	03:04
Breath Volume		1.101	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.200	03:08
Breath Volume		2.007	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,IVAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: FALCON,IVAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: CABRERA GARCIA

DATE/TIME OF TEST: 07/25/2021 22:37:53

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.082	22:35	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.082	22:37	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 08/02/2021 23:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RAWLS,GAVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/15/2021 20:28:42

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.077	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.218	20:24
Breath Volume		1.390	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.220	20:27
Breath Volume		1.890	20:27
Air Blank		0.000	20:27
Control Test		0.077	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 08/09/2021 14:27

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KELLS,BRANDON
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: TREVINO

DATE/TIME OF TEST: 07/30/2021 02:02:57

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.077	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.115	01:58
Breath Volume		1.523	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.121	02:01
Breath Volume		2.074	02:01
Air Blank		0.000	02:02
Control Test		0.077	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 08/09/2021 14:27

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SANDS,BRIAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/30/2021 02:41:16

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.077	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.203	02:35
Breath Volume		3.078	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.196	02:39
Breath Volume		1.703	02:39
Air Blank		0.000	02:40
Control Test		0.076	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 08/09/2021 14:27

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CASIQUE

DATE/TIME OF TEST: 07/04/2021 03:26:44

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.182	03:21
Breath Volume		2.914	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.174	03:24
Breath Volume		1.898	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 07/12/2021 16:52

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: NAVA

DATE/TIME OF TEST: 07/04/2021 03:55:12

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.173	03:50
Breath Volume		2.011	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.182	03:53
Breath Volume		1.570	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 07/12/2021 16:52

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHT,JAMES,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VAUGHT,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: YESCAS

DATE/TIME OF TEST: 07/04/2021 20:14:18

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:09
Subject Sample	# 1	0.269	20:09
Breath Volume		3.464	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.252	20:12
Breath Volume		3.613	20:12
Air Blank		0.000	20:13
Control Test		0.077	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 07/12/2021 16:52

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHT,JAMES,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VAUGHT,JAMES
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MERIZALDE

DATE/TIME OF TEST: 07/08/2021 02:54:15

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.311	02:49
Breath Volume		1.390	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.304	02:52
Breath Volume		1.718	02:52
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 07/12/2021 16:52

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/08/2021 03:31:49

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.161	03:27
Breath Volume		3.355	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.157	03:30
Breath Volume		3.097	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 07/12/2021 16:52

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 07/03/2021 03:15:04

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.097	03:10
Breath Volume		2.550	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.098	03:13
Breath Volume		2.109	03:13
Air Blank		0.000	03:14
Control Test		0.075	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:15

Data Download Date/Time: 08/04/2021 08:11

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: DUPRE

DATE/TIME OF TEST: 07/03/2021 20:50:45

Diagnostics Check		OK	20:47	
Air Blank		0.000	20:48	
Control Test		0.079	20:48	
Air Blank		0.000	20:49	
Subject Sample	# 1		20:49	Subject Test Refused
Breath Volume		0.000	20:49	
Air Blank		0.000	20:49	
Control Test		0.079	20:50	
Air Blank		0.000	20:50	
Diagnostics Check		OK	20:50	

Data Download Date/Time: 08/02/2021 13:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: ESCALON

DATE/TIME OF TEST: 07/10/2021 01:38:16

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.235	01:33
Breath Volume		1.789	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.233	01:36
Breath Volume		1.281	01:36
Air Blank		0.000	01:37
Control Test		0.078	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 08/02/2021 13:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/10/2021 03:28:36

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.190	03:23
Breath Volume		1.136	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.192	03:27
Breath Volume		1.597	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 08/02/2021 13:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: CAPPA

DATE/TIME OF TEST: 07/11/2021 07:17:01

Diagnostics Check		OK	07:10
Air Blank		0.000	07:10
Control Test		0.078	07:10
Air Blank		0.000	07:11
Subject Sample	# 1	0.099	07:12
Breath Volume		1.968	07:12
Air Blank		0.000	07:12
Air Blank		0.000	07:14
Subject Sample	# 2	0.100	07:15
Breath Volume		1.453	07:15
Air Blank		0.000	07:16
Control Test		0.079	07:16
Air Blank		0.000	07:16
Diagnostics Check		OK	07:17

Data Download Date/Time: 08/02/2021 13:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 07/16/2021 04:41:16

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.080	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.198	04:36
Breath Volume		1.437	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:38
Subject Sample	# 2	0.189	04:39
Breath Volume		1.632	04:39
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 08/02/2021 13:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SAPP,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ISSAEV

DATE/TIME OF TEST: 07/17/2021 01:49:49

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.079	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.079	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 08/02/2021 13:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1101850
DRY GAS EXPIRY DATE: 03/02/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: DIPIETRO

DATE/TIME OF TEST: 07/17/2021 03:52:41

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.078	03:49	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Control Test		0.080	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 08/02/2021 13:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDS,BRIAN,V
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: SANDS,BRIAN
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/02/2021 22:55:05

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.077	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.183	22:50
Breath Volume		1.910	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:52
Subject Sample	# 2	0.183	22:53
Breath Volume		1.800	22:53
Air Blank		0.000	22:54
Control Test		0.077	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 07/29/2021 13:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: HURT,ALEXANDER
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 08/02/2022
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 07/05/2021 21:58:11

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.077	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.205	21:53
Breath Volume		1.218	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:55
Subject Sample	# 2	0.205	21:56
Breath Volume		1.230	21:56
Air Blank		0.000	21:57
Control Test		0.075	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 07/29/2021 13:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESANTIS,ADAM,B
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: DAVIS,DONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175970
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/24/2021 04:57:22

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.077	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.098	04:52
Breath Volume		1.781	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.095	04:55
Breath Volume		1.625	04:55
Air Blank		0.000	04:56
Control Test		0.076	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 07/29/2021 13:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 07/01/2021 00:59:12

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.000	00:54
Breath Volume		2.656	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.000	00:57
Breath Volume		1.277	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 07/01/2021 10:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 07/29/2021 03:26:20

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.253	03:17
Breath Volume		1.273	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.228	03:21
Breath Volume		1.750	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 3	0.254	03:24
Breath Volume		3.617	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 08/09/2021 10:30

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: TAYLOR MITCHELL

DATE/TIME OF TEST: 07/02/2021 02:02:01

Diagnostics Check		OK	01:59	
Air Blank		0.000	01:59	
Control Test		0.076	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 1		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:01	
Control Test		0.075	02:01	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CRON

DATE/TIME OF TEST: 07/04/2021 02:02:00

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.076	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.170	01:57
Breath Volume		3.500	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.159	02:00
Breath Volume		4.050	02:00
Air Blank		0.000	02:01
Control Test		0.075	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:02

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMDEN,DENNIS,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: KISHEK

DATE/TIME OF TEST: 07/10/2021 02:18:14

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.076	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1	0.069	02:14	
Breath Volume		3.535	02:14	
Air Blank		0.000	02:15	
Air Blank			02:16	RFI Detect
Air Blank		0.000	02:17	
Control Test		0.076	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ERIKSEN

DATE/TIME OF TEST: 07/19/2021 22:07:28

Diagnostics Check		OK	22:01
Air Blank		0.000	22:01
Control Test		0.076	22:01
Air Blank		0.000	22:02
Subject Sample	# 1	0.174	22:02
Breath Volume		1.496	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:05
Subject Sample	# 2	0.173	22:05
Breath Volume		1.703	22:05
Air Blank		0.000	22:06
Control Test		0.076	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CASSIN

DATE/TIME OF TEST: 07/20/2021 02:24:17

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:18	
Control Test		0.077	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 1	0.217	02:19	
Breath Volume		1.367	02:19	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.200	02:22	
Breath Volume		1.976	02:22	
Air Blank		0.000	02:23	
Control Test		0.074	02:23	Control Outside Tolerance
Air Blank		0.000	02:24	

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CASSIN

DATE/TIME OF TEST: 07/20/2021 02:32:48

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.076	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.203	02:28
Breath Volume		1.828	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.202	02:31
Breath Volume		1.664	02:31
Air Blank		0.000	02:31
Control Test		0.077	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/20/2021 22:02:55

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.076	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.142	21:58
Breath Volume		1.398	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:00
Subject Sample	# 2	0.150	22:01
Breath Volume		1.933	22:01
Air Blank		0.000	22:01
Control Test		0.076	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BLEEKER

DATE/TIME OF TEST: 07/21/2021 02:46:23

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.077	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.120	02:41
Breath Volume		1.699	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.123	02:44
Breath Volume		1.328	02:44
Air Blank		0.000	02:45
Control Test		0.078	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 07/24/2021 00:46:19

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.076	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.225	00:41
Breath Volume		1.375	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.245	00:44
Breath Volume		1.984	00:44
Air Blank		0.000	00:45
Control Test		0.076	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: TRIPP

DATE/TIME OF TEST: 07/24/2021 22:45:11

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.076	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.000	22:43
Breath Volume		2.984	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: WOODRING

DATE/TIME OF TEST: 07/25/2021 00:38:53

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.076	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.202	00:34
Breath Volume		1.882	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:36
Subject Sample	# 2	0.200	00:37
Breath Volume		1.253	00:37
Air Blank		0.000	00:37
Control Test		0.077	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: PURDIN

DATE/TIME OF TEST: 07/25/2021 03:12:03

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.076	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.041	03:06
Breath Volume		1.769	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.039	03:10
Breath Volume		1.238	03:10
Air Blank		0.000	03:11
Control Test		0.077	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:12

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35019080A3
DRY GAS EXPIRY DATE: 02/25/2022
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: OLI

DATE/TIME OF TEST: 07/26/2021 00:32:36

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.077	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.140	00:27
Breath Volume		2.300	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:29
Subject Sample	# 2	0.131	00:31
Breath Volume		3.078	00:31
Air Blank		0.000	00:31
Control Test		0.077	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 07/27/2021 17:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: GARCIA MACEDO

DATE/TIME OF TEST: 07/23/2021 20:23:17

Diagnostics Check		OK	20:16
Air Blank		0.000	20:16
Control Test		0.081	20:17
Air Blank		0.000	20:17
Subject Sample	# 1	0.216	20:18
Breath Volume		2.156	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.210	20:21
Breath Volume		2.054	20:21
Air Blank		0.000	20:22
Control Test		0.079	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 08/09/2021 11:06

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 07/29/2021 22:37:14

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:31
Subject Sample	# 1	0.088	22:32
Breath Volume		1.429	22:32
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.083	22:35
Breath Volume		1.953	22:35
Air Blank		0.000	22:36
Control Test		0.080	22:36
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 08/09/2021 11:06

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINLAND,J,D
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MOYER,A
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 07/11/2021 07:11:56

Diagnostics Check		OK	07:05
Air Blank		0.000	07:06
Control Test		0.077	07:06
Air Blank		0.000	07:06
Subject Sample	# 1	0.176	07:07
Breath Volume		2.078	07:07
Air Blank		0.000	07:08
Air Blank		0.000	07:09
Subject Sample	# 2	0.161	07:10
Breath Volume		2.304	07:10
Air Blank		0.000	07:11
Control Test		0.076	07:11
Air Blank		0.000	07:11
Diagnostics Check		OK	07:11

Data Download Date/Time: 07/14/2021 09:53

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOFF,SCOTT,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RUDOFF,SCOTT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: FIALLOS

DATE/TIME OF TEST: 07/11/2021 08:25:44

Diagnostics Check		OK	08:18
Air Blank		0.000	08:19
Control Test		0.077	08:19
Air Blank		0.000	08:19
Subject Sample	# 1	0.125	08:21
Breath Volume		2.355	08:21
Air Blank		0.000	08:21
Air Blank		0.000	08:23
Subject Sample	# 2	0.124	08:24
Breath Volume		2.769	08:24
Air Blank		0.000	08:24
Control Test		0.076	08:25
Air Blank		0.000	08:25
Diagnostics Check		OK	08:25

Data Download Date/Time: 07/14/2021 09:54

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GONCALVES

DATE/TIME OF TEST: 07/17/2021 02:36:04

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.018	02:31
Breath Volume		2.101	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 2	0.017	02:34
Breath Volume		2.312	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:36

Data Download Date/Time: 08/19/2021 01:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOUCHELLE,ROBERT
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: DAO

DATE/TIME OF TEST: 07/24/2021 20:00:59

Diagnostics Check		OK	19:54
Air Blank		0.000	19:55
Control Test		0.081	19:55
Air Blank		0.000	19:55
Subject Sample	# 1	0.150	19:56
Breath Volume		2.167	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:58
Subject Sample	# 2	0.156	19:59
Breath Volume		2.425	19:59
Air Blank		0.000	20:00
Control Test		0.082	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:00

Data Download Date/Time: 08/19/2021 01:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: JEANTY

DATE/TIME OF TEST: 07/25/2021 03:48:13

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.121	03:43
Breath Volume		1.863	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.124	03:46
Breath Volume		1.312	03:46
Air Blank		0.000	03:47
Control Test		0.081	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 08/19/2021 01:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSH
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/31/2021 08:31:05

Diagnostics Check		OK	08:24
Air Blank		0.000	08:25
Control Test		0.079	08:25
Air Blank		0.000	08:26
Subject Sample	# 1	0.000	08:26
Breath Volume		1.980	08:26
Air Blank		0.000	08:27
Air Blank		0.000	08:29
Subject Sample	# 2	0.000	08:29
Breath Volume		1.683	08:29
Air Blank		0.000	08:30
Control Test		0.080	08:30
Air Blank		0.000	08:31
Diagnostics Check		OK	08:31

Data Download Date/Time: 08/19/2021 01:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAYES,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: GAYLOR

DATE/TIME OF TEST: 07/31/2021 09:29:38

Diagnostics Check		OK	09:22
Air Blank		0.000	09:22
Control Test		0.080	09:22
Air Blank		0.000	09:23
Subject Sample	# 1	0.141	09:24
Breath Volume		2.128	09:24
Air Blank		0.000	09:24
Air Blank		0.000	09:26
Subject Sample	# 2	0.136	09:28
Breath Volume		2.359	09:28
Air Blank		0.000	09:28
Control Test		0.079	09:29
Air Blank		0.000	09:29
Diagnostics Check		OK	09:29

Data Download Date/Time: 08/19/2021 01:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MAJEAU

DATE/TIME OF TEST: 07/25/2021 23:25:00

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.081	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.172	23:20
Breath Volume		1.804	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:22
Subject Sample	# 2	0.183	23:23
Breath Volume		1.570	23:23
Air Blank		0.000	23:24
Control Test		0.080	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:25

Data Download Date/Time: 08/04/2021 05:41

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 07/26/2021 06:02:36

Diagnostics Check		OK	05:56
Air Blank		0.000	05:56
Control Test		0.081	05:57
Air Blank		0.000	05:57
Subject Sample	# 1	0.138	05:57
Breath Volume		1.886	05:57
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.141	06:00
Breath Volume		1.875	06:00
Air Blank		0.000	06:01
Control Test		0.081	06:01
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 08/04/2021 05:41

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: HEILMAN

DATE/TIME OF TEST: 07/30/2021 17:18:40

Diagnostics Check		OK	17:12
Air Blank		0.000	17:12
Control Test		0.081	17:12
Air Blank		0.000	17:13
Subject Sample	# 1	0.195	17:14
Breath Volume		1.363	17:14
Air Blank		0.000	17:14
Air Blank		0.000	17:16
Subject Sample	# 2	0.196	17:17
Breath Volume		2.660	17:17
Air Blank		0.000	17:17
Control Test		0.081	17:18
Air Blank		0.000	17:18
Diagnostics Check		OK	17:18

Data Download Date/Time: 08/04/2021 05:42

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SIRMANS,ABNER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LABORDE

DATE/TIME OF TEST: 07/30/2021 23:58:16

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.082	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:57	
Control Test		0.082	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 08/04/2021 05:42

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 07/02/2021 02:54:08

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.077	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.270	02:49
Breath Volume		4.035	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.280	02:52
Breath Volume		4.023	02:52
Air Blank		0.000	02:53
Control Test		0.077	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 07/14/2021 10:44

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 08:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/05/2021 08:52:32

Diagnostics Check		OK	08:44	
Air Blank		0.000	08:45	
Control Test		0.075	08:45	
Air Blank		0.000	08:46	
Subject Sample	# 1	0.185	08:47	
Breath Volume		2.503	08:47	
Air Blank		0.000	08:48	
Air Blank		0.000	08:50	
Subject Sample	# 2	0.177	08:51	
Breath Volume		1.386	08:51	
Air Blank		0.000	08:51	
Control Test		0.000	08:52	Control Outside Tolerance
Air Blank		0.000	08:52	

Data Download Date/Time: 07/14/2021 10:44

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 08:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/05/2021 09:02:00

Diagnostics Check		OK	08:54
Air Blank		0.000	08:55
Control Test		0.077	08:55
Air Blank		0.000	08:56
Subject Sample	# 1	0.179	08:57
Breath Volume		1.863	08:57
Air Blank		0.000	08:57
Air Blank		0.000	08:59
Subject Sample	# 2	0.183	09:00
Breath Volume		2.488	09:00
Air Blank		0.000	09:01
Control Test		0.077	09:01
Air Blank		0.000	09:01
Diagnostics Check		OK	09:02

Data Download Date/Time: 07/14/2021 10:44

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: NINO LAGUNA

DATE/TIME OF TEST: 07/09/2021 23:41:09

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.077	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.219	23:32
Breath Volume		2.207	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.196	23:36
Breath Volume		1.156	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:38
Subject Sample	# 3	0.192	23:39
Breath Volume		1.324	23:39
Air Blank		0.000	23:40
Control Test		0.076	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 07/14/2021 10:44

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HEUSNER

DATE/TIME OF TEST: 07/16/2021 03:32:49

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.077	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.201	03:27
Breath Volume		1.441	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.201	03:31
Breath Volume		1.460	03:31
Air Blank		0.000	03:31
Control Test		0.078	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 08/18/2021 10:44

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 07/17/2021 05:02:05

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.076	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.178	04:57
Breath Volume		1.859	04:57
Air Blank		0.000	04:58
Air Blank		0.000	04:59
Subject Sample	# 2	0.182	05:00
Breath Volume		3.085	05:00
Air Blank		0.000	05:01
Control Test		0.077	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 08/18/2021 10:44

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: FOX JR

DATE/TIME OF TEST: 07/21/2021 02:21:00

Diagnostics Check		OK	02:13
Air Blank		0.000	02:14
Control Test		0.078	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.173	02:16
Breath Volume		1.382	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.171	02:19
Breath Volume		1.269	02:19
Air Blank		0.000	02:20
Control Test		0.078	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:21

Data Download Date/Time: 08/18/2021 10:44

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 07/21/2021 23:25:27

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.077	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.082	23:20
Breath Volume		4.042	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.078	23:23
Breath Volume		2.492	23:23
Air Blank		0.000	23:24
Control Test		0.078	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 08/18/2021 10:45

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BIEN AIME

DATE/TIME OF TEST: 07/29/2021 01:20:24

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.077	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1	0.179	01:14	
Breath Volume		1.683	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.173	01:18	
Breath Volume		1.167	01:18	
Air Blank		0.000	01:19	
Control Test		0.000	01:19	Control Outside Tolerance
Air Blank		0.000	01:20	

Data Download Date/Time: 08/18/2021 10:53

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BIEN AIME

DATE/TIME OF TEST: 07/29/2021 01:30:01

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.076	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.159	01:24
Breath Volume		1.367	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.171	01:28
Breath Volume		1.343	01:28
Air Blank		0.000	01:29
Control Test		0.077	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:30

Data Download Date/Time: 08/18/2021 10:54

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 07/30/2021 01:15:46

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.000	01:11
Breath Volume		2.722	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.000	01:14
Breath Volume		2.687	01:14
Air Blank		0.000	01:14
Control Test		0.080	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 08/18/2021 10:54

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BUSTILLO

DATE/TIME OF TEST: 07/31/2021 21:47:33

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.077	21:40
Air Blank		0.000	21:41
Subject Sample	# 1	0.183	21:42
Breath Volume		1.476	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.178	21:45
Breath Volume		1.394	21:45
Air Blank		0.000	21:46
Control Test		0.076	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 08/18/2021 10:04

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALUSMA,JOE,S
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ALUSMA,JOE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/01/2021 22:58:32

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:56	
Control Test		0.078	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 1		22:57	RFI Detect
Breath Volume		0.000	22:57	
Air Blank			22:58	RFI Detect
Air Blank		0.000	22:58	

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALUSMA,JOE,S
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ALUSMA,JOE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/01/2021 23:00:54

Diagnostics Check	OK	23:00	
Air Blank		23:00	RFI Detect
Air Blank	0.000	23:00	

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALUSMA,JOE,S
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ALUSMA,JOE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/01/2021 23:08:22

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.077	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.173	23:03
Breath Volume		1.890	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.160	23:06
Breath Volume		1.433	23:06
Air Blank		0.000	23:07
Control Test		0.078	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,D,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CHAPMAN,M
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 07/03/2021 00:12:03

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.233	00:04
Breath Volume		2.058	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.212	00:07
Breath Volume		1.480	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:09
Subject Sample	# 3	0.230	00:10
Breath Volume		2.203	00:10
Air Blank		0.000	00:11
Control Test		0.077	00:11
Air Blank		0.000	00:11
Diagnostics Check		OK	00:12

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,CHASE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CARIAS CHACON

DATE/TIME OF TEST: 07/07/2021 03:12:26

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.078	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Control Test		0.078	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: GASH,CHASE
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CARIAS CHACON

DATE/TIME OF TEST: 07/07/2021 03:22:54

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.091	03:17
Breath Volume		1.347	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.086	03:21
Breath Volume		1.203	03:21
Air Blank		0.000	03:21
Control Test		0.076	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: JOHNSON,J
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 07/09/2021 23:05:43

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.205	23:00
Breath Volume		2.117	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.204	23:04
Breath Volume		1.359	23:04
Air Blank		0.000	23:04
Control Test		0.078	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAFFEE,C,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: CHAFFEE,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: MEZA

DATE/TIME OF TEST: 07/11/2021 11:56:46

Diagnostics Check		OK	11:50
Air Blank		0.000	11:51
Control Test		0.078	11:51
Air Blank		0.000	11:51
Subject Sample	# 1	0.000	11:52
Breath Volume		3.324	11:52
Air Blank		0.000	11:52
Air Blank		0.000	11:54
Subject Sample	# 2	0.000	11:55
Breath Volume		2.929	11:55
Air Blank		0.000	11:55
Control Test		0.079	11:56
Air Blank		0.000	11:56
Diagnostics Check		OK	11:56

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: APONTE,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/13/2021 01:50:04

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.211	01:45
Breath Volume		1.539	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.208	01:48
Breath Volume		1.511	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:50

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,L,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TARR,L
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: PRASAD

DATE/TIME OF TEST: 07/13/2021 03:24:14

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.218	03:19
Breath Volume		2.652	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.212	03:22
Breath Volume		1.777	03:22
Air Blank		0.000	03:23
Control Test		0.077	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 07/16/2021 15:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARMENTER,JACOB
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 07/14/2021 22:05:04

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.080	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.227	21:57
Breath Volume		2.582	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.224	22:03
Breath Volume		1.925	22:03
Air Blank		0.000	22:04
Control Test		0.079	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:05

Data Download Date/Time: 08/03/2021 15:36

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: ROMERO CAMACHO

DATE/TIME OF TEST: 07/04/2021 00:03:48

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.153	23:59
Breath Volume		3.921	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.153	00:02
Breath Volume		3.820	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 08/02/2021 11:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/04/2021 01:54:45

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.172	01:48
Breath Volume		1.394	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.173	01:53
Breath Volume		1.265	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 08/02/2021 11:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,TROY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 07/04/2021 22:34:16

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.099	22:29
Breath Volume		3.726	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.097	22:32
Breath Volume		3.554	22:32
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 08/02/2021 11:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GUEVARA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CULVER

DATE/TIME OF TEST: 07/05/2021 01:34:54

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.134	01:29
Breath Volume		2.097	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:31
Subject Sample	# 2	0.134	01:33
Breath Volume		1.644	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 08/02/2021 11:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORITZ,DFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 07/06/2021 01:03:46

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.080	00:52
Air Blank		0.000	00:52
Subject Sample	# 1	0.221	00:54
Breath Volume		1.429	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.254	00:58
Breath Volume		3.117	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 3	0.213	01:02
Breath Volume		1.558	01:02
Air Blank		0.000	01:02
Control Test		0.076	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 08/02/2021 11:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: BELAND

DATE/TIME OF TEST: 07/06/2021 04:29:11

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.080	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.128	04:24
Breath Volume		1.710	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.133	04:27
Breath Volume		3.574	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 08/02/2021 11:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: POOLE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: BONILLA SANTIAGO

DATE/TIME OF TEST: 07/06/2021 06:57:31

Diagnostics Check		OK	06:52	
Air Blank		0.000	06:52	
Control Test		0.079	06:52	
Air Blank		0.000	06:53	
Subject Sample	# 1		06:56	Subject Test Refused
Breath Volume		0.000	06:56	
Air Blank		0.000	06:56	
Control Test		0.080	06:56	
Air Blank		0.000	06:57	
Diagnostics Check		OK	06:57	

Data Download Date/Time: 08/02/2021 11:32

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JACQUINTO

DATE/TIME OF TEST: 07/07/2021 03:43:00

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:34	
Control Test		0.080	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1	0.000	03:38	Volume Not Met
Breath Volume		0.113	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2		03:41	Subject Test Refused
Breath Volume		0.000	03:41	
Air Blank		0.000	03:42	
Control Test		0.080	03:42	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 08/02/2021 11:33

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 07/09/2021 23:29:44

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.195	23:23
Breath Volume		1.371	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.190	23:28
Breath Volume		1.207	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 08/02/2021 11:33

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: SUTHERLAND

DATE/TIME OF TEST: 07/10/2021 01:04:33

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.101	00:59
Breath Volume		1.980	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.098	01:02
Breath Volume		1.921	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 08/02/2021 11:33

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MATTEINI,V
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: BECKHAM

DATE/TIME OF TEST: 07/10/2021 03:46:53

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.226	03:42
Breath Volume		1.769	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.218	03:45
Breath Volume		1.226	03:45
Air Blank		0.000	03:45
Control Test		0.079	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 08/02/2021 11:33

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GOSSETT

DATE/TIME OF TEST: 07/10/2021 04:47:26

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.148	04:42
Breath Volume		3.167	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.150	04:45
Breath Volume		3.136	04:45
Air Blank		0.000	04:46
Control Test		0.077	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 08/02/2021 11:33

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MEDDAUGH,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: KESSINGER

DATE/TIME OF TEST: 07/10/2021 11:09:28

Diagnostics Check		OK	11:02
Air Blank		0.000	11:03
Control Test		0.080	11:03
Air Blank		0.000	11:03
Subject Sample	# 1	0.362	11:04
Breath Volume		1.796	11:04
Air Blank		0.000	11:05
Air Blank		0.000	11:07
Subject Sample	# 2	0.370	11:07
Breath Volume		1.988	11:07
Air Blank		0.000	11:08
Control Test		0.080	11:08
Air Blank		0.000	11:09
Diagnostics Check		OK	11:09

Data Download Date/Time: 08/02/2021 11:33

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BERMUDEZ,G
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: FARNSWORTH

DATE/TIME OF TEST: 07/10/2021 22:03:30

Diagnostics Check		OK	21:56
Air Blank		0.000	21:57
Control Test		0.079	21:57
Air Blank		0.000	21:58
Subject Sample	# 1	0.131	21:58
Breath Volume		2.929	21:58
Air Blank		0.000	21:59
Air Blank		0.000	22:01
Subject Sample	# 2	0.127	22:01
Breath Volume		3.156	22:01
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:03
Diagnostics Check		OK	22:03

Data Download Date/Time: 08/02/2021 11:33

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REILLY,S.,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MCHLAUGLIN,T.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: LOVERING

DATE/TIME OF TEST: 07/10/2021 22:39:58

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.000	22:35
Breath Volume		3.960	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.000	22:38
Breath Volume		2.183	22:38
Air Blank		0.000	22:39
Control Test		0.080	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 08/02/2021 11:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARKINS,CLINTON,T
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/12/2021 04:51:53

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.174	04:47
Breath Volume		2.199	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.181	04:50
Breath Volume		2.617	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 08/02/2021 11:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ETIENNE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: CAISSE

DATE/TIME OF TEST: 07/12/2021 18:50:20

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.080	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.000	18:46
Breath Volume		1.808	18:46
Air Blank		0.000	18:46
Air Blank		0.000	18:48
Subject Sample	# 2	0.000	18:48
Breath Volume		1.992	18:48
Air Blank		0.000	18:49
Control Test		0.080	18:49
Air Blank		0.000	18:50
Diagnostics Check		OK	18:50

Data Download Date/Time: 08/02/2021 11:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWAARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: NICKLES

DATE/TIME OF TEST: 07/15/2021 03:12:46

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.111	03:08
Breath Volume		1.800	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.110	03:11
Breath Volume		2.683	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 08/02/2021 11:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ROUTON

DATE/TIME OF TEST: 07/15/2021 23:46:31

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.080	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:44	Subject Test Refused
Breath Volume		0.140	23:44	
Air Blank			23:45	Purge Fail
Air Blank			23:46	Purge Fail

Data Download Date/Time: 08/02/2021 11:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD P,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,TAYLOR
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 07/18/2021 03:48:17

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.070	03:47	Control Outside Tolerance
Air Blank	0.000	03:48	

Data Download Date/Time: 08/02/2021 11:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,TAYLOR
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 07/18/2021 03:52:45

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.072	03:52	Control Outside Tolerance
Air Blank	0.000	03:52	

Data Download Date/Time: 08/02/2021 11:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 07/18/2021 08:26:50

Diagnostics Check	OK	08:25	
Air Blank	0.000	08:25	
Control Test	0.070	08:26	Control Outside Tolerance
Air Blank	0.000	08:26	

Data Download Date/Time: 08/02/2021 11:34

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAMORA,DEP,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 07/18/2021 08:31:20

Diagnostics Check	OK	08:29	
Air Blank	0.000	08:30	
Control Test	0.070	08:30	Control Outside Tolerance
Air Blank	0.000	08:31	

Data Download Date/Time: 08/02/2021 11:35

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MORENO CARRANZA

DATE/TIME OF TEST: 07/03/2021 03:25:46

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.080	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1	0.196	03:12	Volume Not Met
Breath Volume		0.324	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2	0.243	03:18	
Breath Volume		2.285	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 3	0.182	03:23	Volume Not Met
Breath Volume		0.324	03:23	
Air Blank		0.000	03:24	
Control Test		0.076	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 07/14/2021 10:10

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MORENO CARRANZA

DATE/TIME OF TEST: 07/03/2021 03:45:27

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.077	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1	0.238	03:32	No .020 agreement
Breath Volume		1.375	03:32	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 2	0.213	03:37	No .020 agreement
Breath Volume		1.316	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 3	0.197	03:43	Volume Not Met
Breath Volume		0.503	03:43	
Air Blank		0.000	03:44	
Control Test		0.075	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 07/14/2021 10:10

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: RAMRAKHIANI

DATE/TIME OF TEST: 07/07/2021 05:08:42

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:06	
Control Test		0.081	05:06	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:07	Subject Test Refused
Breath Volume		0.000	05:07	
Air Blank		0.000	05:07	
Control Test		0.081	05:08	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 07/14/2021 10:10

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/08/2021 03:10:05

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.082	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 07/14/2021 10:10

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 07/09/2021 02:23:56

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.082	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.171	02:18
Breath Volume		3.066	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.154	02:22
Breath Volume		1.136	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 07/14/2021 10:10

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 07/11/2021 05:50:06

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.080	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.202	05:45
Breath Volume		2.117	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:47
Subject Sample	# 2	0.212	05:48
Breath Volume		2.828	05:48
Air Blank		0.000	05:49
Control Test		0.079	05:49
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 07/14/2021 10:10

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DALEY

DATE/TIME OF TEST: 07/03/2021 05:10:00

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.077	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.157	05:05
Breath Volume		1.226	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.167	05:08
Breath Volume		1.136	05:08
Air Blank		0.000	05:09
Control Test		0.078	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:10

Data Download Date/Time: 07/14/2021 10:49

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/04/2021 03:20:42

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.079	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.018	03:16
Breath Volume		1.867	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.016	03:19
Breath Volume		1.343	03:19
Air Blank		0.000	03:19
Control Test		0.078	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 07/14/2021 10:49

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TRUJILLO GARCIA

DATE/TIME OF TEST: 07/11/2021 02:21:50

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.078	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.157	02:16
Breath Volume		2.125	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.152	02:20
Breath Volume		2.335	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 07/14/2021 10:49

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 07/17/2021 01:51:15

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.198	01:46
Breath Volume		1.906	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.203	01:49
Breath Volume		2.839	01:49
Air Blank		0.000	01:50
Control Test		0.078	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 08/19/2021 01:19

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 07/18/2021 02:47:35

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.000	02:43
Breath Volume		2.500	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:46
Breath Volume		2.792	02:46
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/19/2021 01:19

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: SEMISTA

DATE/TIME OF TEST: 07/21/2021 23:36:52

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.078	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.205	23:32
Breath Volume		2.132	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.212	23:35
Breath Volume		1.460	23:35
Air Blank		0.000	23:35
Control Test		0.077	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 08/19/2021 01:19

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 07/23/2021 04:58:16

Diagnostics Check	Fail	04:57	Internal Printer Test
Air Blank	0.000	04:58	

Data Download Date/Time: 08/19/2021 01:20

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 07/23/2021 05:06:59

Diagnostics Check		OK	04:59	
Air Blank		0.000	05:00	
Control Test		0.076	05:00	
Air Blank		0.000	05:01	
Subject Sample	# 1	0.167	05:01	
Breath Volume		1.613	05:01	
Air Blank		0.000	05:02	
Air Blank		0.000	05:04	
Subject Sample	# 2	0.164	05:05	
Breath Volume		1.343	05:05	
Air Blank		0.000	05:06	
Control Test		0.000	05:06	Control Outside Tolerance
Air Blank		0.000	05:06	

Data Download Date/Time: 08/19/2021 01:20

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 07/25/2021 04:09:08

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.173	04:04
Breath Volume		1.238	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.175	04:07
Breath Volume		1.644	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 08/19/2021 01:20

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MAZUR

DATE/TIME OF TEST: 07/29/2021 04:08:03

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.212	04:03
Breath Volume		1.878	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.205	04:06
Breath Volume		1.582	04:06
Air Blank		0.000	04:07
Control Test		0.077	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:08

Data Download Date/Time: 08/19/2021 01:20

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/16/2021 02:41:34

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.076	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.218	02:37
Breath Volume		1.941	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.223	02:39
Breath Volume		1.902	02:39
Air Blank		0.000	02:40
Control Test		0.077	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 08/11/2021 11:15

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 07/16/2021 04:13:24

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:10	
Control Test		0.078	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 08/11/2021 11:15

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: STARK

DATE/TIME OF TEST: 07/16/2021 05:36:32

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.078	05:30
Air Blank		0.000	05:30
Subject Sample	# 1	0.123	05:31
Breath Volume		2.960	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.113	05:34
Breath Volume		2.632	05:34
Air Blank		0.000	05:35
Control Test		0.077	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 08/11/2021 11:15

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/20/2021 21:49:22

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.078	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.000	21:44
Breath Volume		2.703	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.000	21:48
Breath Volume		2.714	21:48
Air Blank		0.000	21:48
Control Test		0.078	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 08/11/2021 11:15

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PARADA RODRIGUEZ

DATE/TIME OF TEST: 07/22/2021 02:35:34

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.078	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.146	02:31
Breath Volume		2.042	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.147	02:33
Breath Volume		2.113	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 08/11/2021 11:15

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: TEBALAN CHIC

DATE/TIME OF TEST: 07/23/2021 21:11:44

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.078	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.201	21:07
Breath Volume		2.992	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:09
Subject Sample	# 2	0.182	21:10
Breath Volume		1.812	21:10
Air Blank		0.000	21:10
Control Test		0.077	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 08/11/2021 11:15

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: MCNEAL JR

DATE/TIME OF TEST: 07/25/2021 20:10:05

Diagnostics Check		OK	20:03
Air Blank		0.000	20:04
Control Test		0.079	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.195	20:05
Breath Volume		1.632	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.197	20:08
Breath Volume		2.007	20:08
Air Blank		0.000	20:09
Control Test		0.077	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 08/11/2021 11:16

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GROOKETT

DATE/TIME OF TEST: 07/01/2021 01:51:31

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.081	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1	0.000	01:48	
Breath Volume		2.726	01:48	
Air Blank		0.000	01:48	
Air Blank		0.000	01:50	
Subject Sample	# 2		01:50	RFI Detect
Breath Volume		0.000	01:50	
Air Blank			01:51	RFI Detect
Air Blank		0.000	01:51	

Data Download Date/Time: 08/12/2021 10:47

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GROOKETT

DATE/TIME OF TEST: 07/01/2021 02:02:45

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.000	01:58
Breath Volume		4.476	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.000	02:01
Breath Volume		4.105	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 08/12/2021 10:47

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BROWN,BRENDAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MCVOY

DATE/TIME OF TEST: 07/02/2021 23:08:30

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.082	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.000	23:04
Breath Volume		1.351	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.000	23:07
Breath Volume		1.761	23:07
Air Blank		0.000	23:07
Control Test		0.081	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JAMIE,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BAUTISTA,JAQUELYN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 07/03/2021 00:28:22

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.082	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.000	00:24
Breath Volume		1.128	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.000	00:27
Breath Volume		1.148	00:27
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MUNDT,JADA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: SANBORN

DATE/TIME OF TEST: 07/03/2021 19:43:42

Diagnostics Check		OK	19:36
Air Blank		0.000	19:37
Control Test		0.081	19:37
Air Blank		0.000	19:38
Subject Sample	# 1	0.204	19:38
Breath Volume		2.144	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.195	19:42
Breath Volume		1.703	19:42
Air Blank		0.000	19:42
Control Test		0.080	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TOMASEGOVICH,KODY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: BERCAW

DATE/TIME OF TEST: 07/04/2021 03:33:51

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.274	03:29
Breath Volume		2.175	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.273	03:32
Breath Volume		4.046	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMAHON,KENNETH,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MCMAHON,KENNETH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 07/04/2021 04:35:44

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.137	04:31
Breath Volume		1.996	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:33
Subject Sample	# 2	0.136	04:34
Breath Volume		1.277	04:34
Air Blank		0.000	04:34
Control Test		0.081	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK',J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ROPER,W
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FATU

DATE/TIME OF TEST: 07/05/2021 23:47:33

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:36	
Control Test		0.079	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.191	23:40	Volume Not Met
Breath Volume		0.855	23:40	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2	0.202	23:42	
Breath Volume		1.667	23:42	
Air Blank		0.000	23:43	
Air Blank		0.000	23:45	
Subject Sample	# 3	0.205	23:45	
Breath Volume		2.476	23:45	
Air Blank		0.000	23:46	
Control Test		0.078	23:46	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GLASS,DREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/09/2021 02:09:25

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.193	02:04
Breath Volume		1.339	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.203	02:07
Breath Volume		3.164	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ARGENT,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/10/2021 03:45:52

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.246	03:41
Breath Volume		2.914	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.239	03:44
Breath Volume		2.859	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MARIN ORONZO

DATE/TIME OF TEST: 07/11/2021 01:21:21

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.081	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.087	01:16
Breath Volume		1.757	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.091	01:19
Breath Volume		2.328	01:19
Air Blank		0.000	01:20
Control Test		0.077	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MLINAC,MADISON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/11/2021 03:38:52

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.261	03:30
Breath Volume		1.660	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.231	03:34
Breath Volume		1.410	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 3	0.245	03:37
Breath Volume		2.109	03:37
Air Blank		0.000	03:37
Control Test		0.076	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 07/15/2021 01:57:20

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.109	01:52
Breath Volume		2.246	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.107	01:55
Breath Volume		2.015	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 07/17/2021 00:29:55

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.203	00:23
Breath Volume		1.246	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.212	00:28
Breath Volume		1.371	00:28
Air Blank		0.000	00:28
Control Test		0.079	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GLASS,DREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 07/17/2021 03:23:42

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.213	03:18
Breath Volume		1.718	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.204	03:22
Breath Volume		1.296	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 08/12/2021 10:48

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: VASQUEZ-DIMOS

DATE/TIME OF TEST: 07/09/2021 22:34:58

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.080	22:32	
Air Blank		0.000	22:33	
Subject Sample	# 1		22:33	Subject Test Refused
Breath Volume		0.000	22:33	
Air Blank		0.000	22:34	
Control Test		0.079	22:34	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 09/04/2021 11:46

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ABBOTT,RYAN
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: GONSALEZ

DATE/TIME OF TEST: 07/13/2021 07:08:27

Diagnostics Check		OK	07:01
Air Blank		0.000	07:01
Control Test		0.079	07:01
Air Blank		0.000	07:02
Subject Sample	# 1	0.173	07:03
Breath Volume		3.792	07:03
Air Blank		0.000	07:04
Air Blank		0.000	07:05
Subject Sample	# 2	0.159	07:06
Breath Volume		3.679	07:06
Air Blank		0.000	07:07
Control Test		0.077	07:07
Air Blank		0.000	07:08
Diagnostics Check		OK	07:08

Data Download Date/Time: 09/04/2021 11:46

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: LEON BETANCOURT

DATE/TIME OF TEST: 07/15/2021 23:05:49

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.079	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.169	23:00
Breath Volume		1.769	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.179	23:04
Breath Volume		2.710	23:04
Air Blank		0.000	23:04
Control Test		0.076	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 09/04/2021 11:46

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: EVERSOLE,DONNIE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MORALES BESANILLA

DATE/TIME OF TEST: 07/16/2021 00:37:42

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.171	00:32
Breath Volume		1.445	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:34
Subject Sample	# 2	0.167	00:35
Breath Volume		2.082	00:35
Air Blank		0.000	00:36
Control Test		0.076	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 09/04/2021 11:46

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BROWN,JAKEVIS
ARRESTING OFFICER AGENCY: Bowling Green PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BENITEZ SANCHEZ

DATE/TIME OF TEST: 07/24/2021 01:52:07

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.159	01:47
Breath Volume		2.941	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.161	01:50
Breath Volume		2.433	01:50
Air Blank		0.000	01:51
Control Test		0.078	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 09/04/2021 11:46

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JESSICA,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: TYLER,RHETT
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: SOLES

DATE/TIME OF TEST: 07/24/2021 06:53:46

Diagnostics Check		OK	06:47
Air Blank		0.000	06:47
Control Test		0.080	06:48
Air Blank		0.000	06:48
Subject Sample	# 1	0.205	06:49
Breath Volume		2.125	06:49
Air Blank		0.000	06:49
Air Blank		0.000	06:51
Subject Sample	# 2	0.204	06:52
Breath Volume		2.867	06:52
Air Blank		0.000	06:52
Control Test		0.078	06:53
Air Blank		0.000	06:53
Diagnostics Check		OK	06:53

Data Download Date/Time: 09/04/2021 11:46

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/01/2021 18:54:01

Diagnostics Check		OK	18:47
Air Blank		0.000	18:48
Control Test		0.078	18:48
Air Blank		0.000	18:49
Subject Sample	# 1	0.205	18:49
Breath Volume		3.234	18:49
Air Blank		0.000	18:50
Air Blank		0.000	18:51
Subject Sample	# 2	0.203	18:52
Breath Volume		3.511	18:52
Air Blank		0.000	18:53
Control Test		0.078	18:53
Air Blank		0.000	18:53
Diagnostics Check		OK	18:54

Data Download Date/Time: 07/23/2021 09:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/08/2021 01:55:32

Diagnostics Check	Fail	01:55	Internal Printer Test
Air Blank	0.000	01:55	

Data Download Date/Time: 07/23/2021 09:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/08/2021 02:04:30

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.215	01:59
Breath Volume		1.726	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.223	02:02
Breath Volume		2.609	02:02
Air Blank		0.000	02:03
Control Test		0.077	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 07/23/2021 09:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PURMORT

DATE/TIME OF TEST: 07/10/2021 01:09:48

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.078	01:04
Air Blank		0.000	01:04
Subject Sample	# 1	0.000	01:05
Breath Volume		1.527	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.000	01:08
Breath Volume		2.132	01:08
Air Blank		0.000	01:08
Control Test		0.078	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 07/23/2021 09:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: HASSEY JR

DATE/TIME OF TEST: 07/11/2021 06:02:05

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.080	05:55
Air Blank		0.000	05:55
Subject Sample	# 1	0.147	05:56
Breath Volume		1.339	05:56
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 2	0.142	06:00
Breath Volume		1.214	06:00
Air Blank		0.000	06:01
Control Test		0.078	06:01
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 07/23/2021 09:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CHAVES

DATE/TIME OF TEST: 07/12/2021 04:42:24

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.147	04:37
Breath Volume		1.574	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.150	04:40
Breath Volume		1.152	04:40
Air Blank		0.000	04:41
Control Test		0.077	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 07/23/2021 09:56

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARROLL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: BUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: AMBROSINI

DATE/TIME OF TEST: 07/04/2021 19:06:32

Diagnostics Check	Fail	19:06	Internal Printer Test
Air Blank	0.000	19:06	

Data Download Date/Time: 08/06/2021 11:04

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: CARROLL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1380119
DRY GAS EXPIRY DATE: 09/22/2023
VIOLATION CODE: BUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: AMBROSINI

DATE/TIME OF TEST: 07/04/2021 19:12:12

Diagnostics Check		OK	19:08	
Air Blank		0.000	19:08	
Control Test		0.079	19:09	
Air Blank		0.000	19:09	
Subject Sample	# 1		19:10	Subject Test Refused
Breath Volume		0.000	19:10	
Air Blank		0.000	19:11	
Control Test		0.079	19:11	
Air Blank		0.000	19:12	
Diagnostics Check		OK	19:12	

Data Download Date/Time: 08/06/2021 11:04

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 07/03/2021 05:27:28

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.180	05:22
Breath Volume		2.613	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.172	05:25
Breath Volume		1.300	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 08/17/2021 13:11

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: MCNEAL

DATE/TIME OF TEST: 07/03/2021 21:09:16

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:06	
Control Test		0.080	21:06	
Air Blank		0.000	21:06	
Subject Sample	# 1		21:07	Subject Test Refused
Breath Volume		0.000	21:07	
Air Blank		0.000	21:08	
Control Test		0.079	21:08	
Air Blank		0.000	21:09	
Diagnostics Check		OK	21:09	

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ELIZABETH,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: YETTER,JAMES
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: BUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 07/03/2021 21:25:31

Diagnostics Check		OK	21:21	
Air Blank		0.000	21:21	
Control Test		0.081	21:21	
Air Blank		0.000	21:22	
Subject Sample	# 1		21:24	Subject Test Refused
Breath Volume		0.000	21:24	
Air Blank		0.000	21:24	
Control Test		0.081	21:24	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BUSCEMI,ANTHONY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: JACK

DATE/TIME OF TEST: 07/03/2021 22:26:12

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:23	
Control Test		0.079	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1		22:24	Subject Test Refused
Breath Volume		0.000	22:24	
Air Blank		0.000	22:25	
Control Test		0.080	22:25	
Air Blank		0.000	22:26	
Diagnostics Check		OK	22:26	

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: RIVERA ALVARADO

DATE/TIME OF TEST: 07/04/2021 00:48:24

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.054	00:43
Breath Volume		2.078	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.057	00:46
Breath Volume		2.738	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: GRECH

DATE/TIME OF TEST: 07/06/2021 20:44:40

Diagnostics Check		OK	20:38
Air Blank		0.000	20:38
Control Test		0.080	20:38
Air Blank		0.000	20:39
Subject Sample	# 1	0.000	20:39
Breath Volume		1.328	20:39
Air Blank		0.000	20:40
Air Blank		0.000	20:42
Subject Sample	# 2	0.000	20:43
Breath Volume		1.964	20:43
Air Blank		0.000	20:43
Control Test		0.080	20:44
Air Blank		0.000	20:44
Diagnostics Check		OK	20:44

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: STANT-ROJAS

DATE/TIME OF TEST: 07/09/2021 22:09:18

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:06	
Control Test		0.080	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 1		22:07	Subject Test Refused
Breath Volume		0.000	22:07	
Air Blank		0.000	22:08	
Control Test		0.080	22:08	
Air Blank		0.000	22:09	
Diagnostics Check		OK	22:09	

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CURRY,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/10/2021 04:25:50

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.043	04:21
Breath Volume		2.222	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.042	04:24
Breath Volume		1.851	04:24
Air Blank		0.000	04:24
Control Test		0.080	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASON,CONNER,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HERNANDEZ,LEONARDO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: NASON

DATE/TIME OF TEST: 07/11/2021 03:31:53

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.195	03:27
Breath Volume		3.460	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.193	03:30
Breath Volume		3.222	03:30
Air Blank		0.000	03:30
Control Test		0.079	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 08/17/2021 13:12

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/11/2021 22:47:58

Diagnostics Check		OK	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Subject Sample	# 1	0.160	22:43
Breath Volume		1.804	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.158	22:46
Breath Volume		1.906	22:46
Air Blank		0.000	22:47
Control Test		0.080	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 08/17/2021 13:13

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SCHMITT

DATE/TIME OF TEST: 07/12/2021 20:26:29

Diagnostics Check		OK	20:23	
Air Blank		0.000	20:23	
Control Test		0.080	20:24	
Air Blank		0.000	20:24	
Subject Sample	# 1		20:25	Subject Test Refused
Breath Volume		0.000	20:25	
Air Blank		0.000	20:25	
Control Test		0.080	20:25	
Air Blank		0.000	20:26	
Diagnostics Check		OK	20:26	

Data Download Date/Time: 08/17/2021 13:17

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN,NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SHORTER

DATE/TIME OF TEST: 07/13/2021 23:35:16

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.081	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.132	23:30
Breath Volume		1.691	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.120	23:33
Breath Volume		1.750	23:33
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 08/17/2021 13:17

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,SAMUEL,F
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 07/14/2021 19:02:42

Diagnostics Check		OK	18:49	
Air Blank		0.000	18:49	
Control Test		0.080	18:49	
Air Blank		0.000	18:50	
Subject Sample	# 1	0.222	18:51	
Breath Volume		1.285	18:51	
Air Blank		0.000	18:52	
Air Blank		0.000	18:53	
Subject Sample	# 2	0.195	18:56	Volume Not Met
Breath Volume		0.929	18:56	
Air Blank		0.000	18:57	
Air Blank		0.000	18:59	
Subject Sample	# 3	0.214	19:00	
Breath Volume		1.351	19:00	
Air Blank		0.000	19:01	
Control Test		0.076	19:02	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 08/17/2021 13:17

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,BRANDON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: LORCA

DATE/TIME OF TEST: 07/15/2021 02:11:21

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.080	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,BRIAN,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SAPP,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: BUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CANN

DATE/TIME OF TEST: 07/15/2021 22:18:34

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:15	
Control Test		0.080	22:16	
Air Blank		0.000	22:16	
Subject Sample	# 1		22:17	Subject Test Refused
Breath Volume		0.000	22:17	
Air Blank		0.000	22:17	
Control Test		0.080	22:17	
Air Blank		0.000	22:18	
Diagnostics Check		OK	22:18	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: HAHNER

DATE/TIME OF TEST: 07/16/2021 01:40:57

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.080	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:40	
Control Test		0.080	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIERRO,ALEXANDER,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FIERRO,ALEXANDER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 07/16/2021 02:24:14

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.079	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2	0.190	02:19	
Breath Volume		3.945	02:19	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 3	0.186	02:22	
Breath Volume		1.363	02:22	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTIN-REYES,JUAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: MAZZUCA

DATE/TIME OF TEST: 07/18/2021 19:09:51

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:00	
Control Test		0.079	19:01	
Air Blank		0.000	19:01	
Subject Sample	# 1		19:04	No Sample Provided
Breath Volume		0.000	19:04	
Air Blank		0.000	19:05	
Air Blank		0.000	19:07	
Subject Sample	# 2		19:08	Subject Test Refused
Breath Volume		0.000	19:08	
Air Blank		0.000	19:08	
Control Test		0.079	19:08	
Air Blank			19:09	RFI Detect
Air Blank		0.000	19:09	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNN',NOLAN,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WYNN,NOLAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KORZEN

DATE/TIME OF TEST: 07/19/2021 00:01:20

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.081	23:59	
Air Blank		0.000	23:59	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Control Test		0.081	00:00	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,BRANDON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: FRANCO DE FRANCISCO

DATE/TIME OF TEST: 07/20/2021 03:31:09

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.081	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Control Test		0.080	03:30	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,JENNIFER,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUGHES,JENNIFER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: MOLERO

DATE/TIME OF TEST: 07/20/2021 17:14:10

Diagnostics Check		OK	17:11	
Air Blank		0.000	17:11	
Control Test		0.080	17:12	
Air Blank		0.000	17:12	
Subject Sample	# 1		17:12	Subject Test Refused
Breath Volume		0.000	17:12	
Air Blank		0.000	17:13	
Control Test		0.080	17:13	
Air Blank		0.000	17:14	
Diagnostics Check		OK	17:14	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 07/24/2021 04:14:29

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.221	04:09
Breath Volume		3.046	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.215	04:12
Breath Volume		2.804	04:12
Air Blank		0.000	04:13
Control Test		0.078	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: MALBROUGH

DATE/TIME OF TEST: 07/24/2021 20:02:12

Diagnostics Check		OK	19:55
Air Blank		0.000	19:55
Control Test		0.080	19:56
Air Blank		0.000	19:56
Subject Sample	# 1	0.246	19:57
Breath Volume		2.246	19:57
Air Blank		0.000	19:58
Air Blank		0.000	19:59
Subject Sample	# 2	0.252	20:00
Breath Volume		1.113	20:00
Air Blank		0.000	20:01
Control Test		0.079	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,BRIAN,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BOLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LIZ

DATE/TIME OF TEST: 07/25/2021 01:18:25

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.080	01:07	
Air Blank		0.000	01:07	
Subject Sample	# 1	0.126	01:10	Volume Not Met
Breath Volume		1.214	01:10	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 2	0.131	01:16	Volume Not Met
Breath Volume		0.906	01:16	
Air Blank		0.000	01:17	
Control Test		0.077	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,BRIAN,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BOLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LIZ

DATE/TIME OF TEST: 07/25/2021 01:28:38

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.078	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.157	01:23
Breath Volume		2.640	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.153	01:26
Breath Volume		2.187	01:26
Air Blank		0.000	01:27
Control Test		0.077	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,BRANDON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HERRADA

DATE/TIME OF TEST: 07/28/2021 05:14:33

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:13	Subject Test Refused
Breath Volume		0.000	05:13	
Air Blank		0.000	05:13	
Control Test		0.080	05:13	
Air Blank		0.000	05:14	
Diagnostics Check		OK	05:14	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LEVASSEUR

DATE/TIME OF TEST: 07/28/2021 21:07:05

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.080	21:03	
Air Blank		0.000	21:04	
Subject Sample	# 1		21:05	Subject Test Refused
Breath Volume		0.000	21:05	
Air Blank		0.000	21:06	
Control Test		0.081	21:06	
Air Blank		0.000	21:07	
Diagnostics Check		OK	21:07	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LEVASSEUR

DATE/TIME OF TEST: 07/28/2021 21:18:02

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.080	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.185	21:12
Breath Volume		1.246	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.183	21:16
Breath Volume		1.324	21:16
Air Blank		0.000	21:17
Control Test		0.079	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:18

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 12:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:53

SUBJECT'S LAST NAME: PHU

DATE/TIME OF TEST: 07/28/2021 13:22:05

Diagnostics Check		OK	13:15
Air Blank		0.000	13:15
Control Test		0.080	13:16
Air Blank		0.000	13:16
Subject Sample	# 1	0.097	13:17
Breath Volume		2.097	13:17
Air Blank		0.000	13:18
Air Blank		0.000	13:19
Subject Sample	# 2	0.096	13:20
Breath Volume		2.613	13:20
Air Blank		0.000	13:21
Control Test		0.077	13:21
Air Blank		0.000	13:21
Diagnostics Check		OK	13:22

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 07/28/2021 15:30:29

Diagnostics Check		OK	15:27	
Air Blank		0.000	15:27	
Control Test		0.079	15:28	
Air Blank		0.000	15:28	
Subject Sample	# 1		15:29	Subject Test Refused
Breath Volume		0.000	15:29	
Air Blank		0.000	15:29	
Control Test		0.078	15:29	
Air Blank		0.000	15:30	
Diagnostics Check		OK	15:30	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: DORIA

DATE/TIME OF TEST: 07/29/2021 12:23:26

Diagnostics Check		OK	12:16
Air Blank		0.000	12:16
Control Test		0.080	12:16
Air Blank		0.000	12:17
Subject Sample	# 1	0.178	12:18
Breath Volume		1.539	12:18
Air Blank		0.000	12:19
Air Blank		0.000	12:20
Subject Sample	# 2	0.164	12:21
Breath Volume		1.578	12:21
Air Blank		0.000	12:22
Control Test		0.078	12:22
Air Blank		0.000	12:23
Diagnostics Check		OK	12:23

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/31/2021 20:22:55

Diagnostics Check		OK	20:20	
Air Blank		0.000	20:20	
Control Test		0.080	20:20	
Air Blank		0.000	20:21	
Subject Sample	# 1		20:21	Subject Test Refused
Breath Volume		0.000	20:21	
Air Blank		0.000	20:22	
Control Test		0.080	20:22	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:22	

Data Download Date/Time: 08/17/2021 13:18

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RORER,JESSICA,L
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PASS,JUSTIN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/04/2021 19:27:32

Diagnostics Check		OK	19:20
Air Blank		0.000	19:21
Control Test		0.079	19:21
Air Blank		0.000	19:22
Subject Sample	# 1	0.256	19:22
Breath Volume		4.156	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.255	19:25
Breath Volume		3.031	19:25
Air Blank		0.000	19:26
Control Test		0.077	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 07/14/2021 10:11

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

INSTRUMENT REGISTERED TO: FWC TAMPA
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPPE,ZACH,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: WILKINS,MALACHI
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: BUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: PIPER

DATE/TIME OF TEST: 07/24/2021 20:19:21

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.080	20:17	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:18	
Control Test		0.080	20:18	
Air Blank		0.000	20:19	
Diagnostics Check		OK	20:19	

Data Download Date/Time: 08/30/2021 13:15

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRA,WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/03/2021 01:14:55

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.080	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.080	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 08/16/2021 13:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ELMORE,BENJAMIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BERNINI

DATE/TIME OF TEST: 07/03/2021 04:35:57

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:33	
Control Test		0.079	04:33	
Air Blank		0.000	04:33	
Subject Sample	# 1		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:35	
Control Test		0.079	04:35	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 08/16/2021 13:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRAGG,GARETT
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: POPCOCHE

DATE/TIME OF TEST: 07/03/2021 22:47:41

Diagnostics Check		OK	22:44	
Air Blank		0.000	22:45	
Control Test		0.080	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:46	Subject Test Refused
Breath Volume		0.000	22:46	
Air Blank		0.000	22:46	
Control Test		0.079	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 08/16/2021 13:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRAGG,GARETT
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: OREA-IGARZA

DATE/TIME OF TEST: 07/05/2021 02:02:04

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.139	01:57
Breath Volume		1.656	01:57
Air Blank		0.000	01:58
Air Blank		0.000	01:59
Subject Sample	# 2	0.143	02:00
Breath Volume		1.468	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:02

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ELMORE,BENJAMIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 07/07/2021 20:15:48

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:13	
Control Test		0.080	20:13	
Air Blank		0.000	20:13	
Subject Sample	# 1		20:14	Subject Test Refused
Breath Volume		0.000	20:14	
Air Blank		0.000	20:14	
Control Test		0.080	20:15	
Air Blank		0.000	20:15	
Diagnostics Check		OK	20:15	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: THUNE

DATE/TIME OF TEST: 07/10/2021 22:30:31

Diagnostics Check		OK	22:27	
Air Blank		0.000	22:28	
Control Test		0.080	22:28	
Air Blank		0.000	22:28	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.080	22:29	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTERBAND,FRANK,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS,WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: LA CORTE

DATE/TIME OF TEST: 07/13/2021 17:42:41

Diagnostics Check		OK	17:35
Air Blank		0.000	17:36
Control Test		0.079	17:36
Air Blank		0.000	17:36
Subject Sample	# 1	0.313	17:38
Breath Volume		2.328	17:38
Air Blank		0.000	17:38
Air Blank		0.000	17:40
Subject Sample	# 2	0.303	17:40
Breath Volume		2.648	17:40
Air Blank		0.000	17:41
Control Test		0.077	17:42
Air Blank		0.000	17:42
Diagnostics Check		OK	17:42

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/14/2021 22:29:26

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:26	
Control Test		0.080	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Control Test		0.080	22:28	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WESTERBAND,FRANK,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RUFIN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 08:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: BUESO

DATE/TIME OF TEST: 07/15/2021 09:34:57

Diagnostics Check		OK	09:23	
Air Blank		0.000	09:23	
Control Test		0.080	09:24	
Air Blank		0.000	09:24	
Subject Sample	# 1		09:27	No Sample Provided
Breath Volume		0.000	09:27	
Air Blank		0.000	09:28	
Air Blank		0.000	09:30	
Subject Sample	# 2		09:33	No Sample Provided
Breath Volume		0.000	09:33	
Air Blank		0.000	09:34	
Control Test		0.079	09:34	
Air Blank		0.000	09:34	
Diagnostics Check		OK	09:34	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: ANTROBUS

DATE/TIME OF TEST: 07/17/2021 15:50:16

Diagnostics Check		OK	15:40
Air Blank		0.000	15:41
Control Test		0.080	15:41
Air Blank		0.000	15:41
Subject Sample	# 1	0.192	15:42
Breath Volume		1.777	15:42
Air Blank		0.000	15:43
Air Blank		0.000	15:45
Subject Sample	# 2	0.168	15:45
Breath Volume		1.273	15:45
Air Blank		0.000	15:46
Air Blank		0.000	15:48
Subject Sample	# 3	0.188	15:48
Breath Volume		1.820	15:48
Air Blank		0.000	15:49
Control Test		0.078	15:49
Air Blank		0.000	15:50
Diagnostics Check		OK	15:50

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: NIEMIEC,NICHOLAS
ARRESTING OFFICER AGENCY: Key Colony Beach PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 07/18/2021 21:29:19

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.079	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:28	
Control Test		0.080	21:28	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: NIEMIEC,NICHOLAS
ARRESTING OFFICER AGENCY: Key Colony Beach PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 07/18/2021 22:10:21

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:56	
Control Test		0.078	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1	0.174	22:00	Volume Not Met
Breath Volume		0.066	22:00	
Air Blank		0.000	22:01	
Air Blank		0.000	22:03	
Subject Sample	# 2	0.166	22:04	
Breath Volume		2.808	22:04	
Air Blank		0.000	22:05	
Air Blank		0.000	22:07	
Subject Sample	# 3	0.155	22:08	
Breath Volume		2.144	22:08	
Air Blank		0.000	22:09	
Control Test		0.078	22:09	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: YAQUINTO

DATE/TIME OF TEST: 07/22/2021 19:35:16

Diagnostics Check		OK	19:32	
Air Blank		0.000	19:32	
Control Test		0.080	19:32	
Air Blank		0.000	19:33	
Subject Sample	# 1		19:33	Subject Test Refused
Breath Volume		0.000	19:33	
Air Blank		0.000	19:34	
Control Test		0.079	19:34	
Air Blank		0.000	19:35	
Diagnostics Check		OK	19:35	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GIODIGIOVANNI,KRIS
ARRESTING OFFICER AGENCY: Key Colony Beach PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: PULLEY

DATE/TIME OF TEST: 07/23/2021 06:25:05

Diagnostics Check		OK	06:22	
Air Blank		0.000	06:22	
Control Test		0.080	06:22	
Air Blank		0.000	06:23	
Subject Sample	# 1		06:23	Subject Test Refused
Breath Volume		0.000	06:23	
Air Blank		0.000	06:24	
Control Test		0.081	06:24	
Air Blank		0.000	06:25	
Diagnostics Check		OK	06:25	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CRISTEA,VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/23/2021 23:16:31

Diagnostics Check		OK	23:07
Air Blank		0.000	23:08
Control Test		0.080	23:08
Air Blank		0.000	23:09
Subject Sample	# 1	0.078	23:11
Breath Volume		1.738	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:13
Subject Sample	# 2	0.069	23:14
Breath Volume		1.343	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 07/25/2021 04:48:53

Diagnostics Check		OK	04:46	
Air Blank		0.000	04:46	
Control Test		0.080	04:46	
Air Blank		0.000	04:47	
Subject Sample	# 1		04:47	Subject Test Refused
Breath Volume		0.000	04:47	
Air Blank		0.000	04:47	
Control Test		0.080	04:48	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 08/16/2021 13:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARIAS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ARIAS,WILLIAM
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 07/01/2021 02:02:24

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1	0.111	01:57	
Breath Volume		2.398	01:57	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2	0.099	02:00	
Breath Volume		2.207	02:00	
Air Blank		0.000	02:01	
Control Test		0.072	02:01	Control Outside Tolerance
Air Blank		0.000	02:02	

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,JOSIE,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WELLS,JOSIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 10:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: ABNEY

DATE/TIME OF TEST: 07/01/2021 11:37:15

Diagnostics Check		OK	11:31
Air Blank		0.000	11:31
Control Test		0.079	11:31
Air Blank		0.000	11:32
Subject Sample	# 1	0.000	11:32
Breath Volume		3.152	11:32
Air Blank		0.000	11:33
Air Blank		0.000	11:35
Subject Sample	# 2	0.000	11:35
Breath Volume		2.800	11:35
Air Blank		0.000	11:36
Control Test		0.079	11:36
Air Blank		0.000	11:37
Diagnostics Check		OK	11:37

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,CODY,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LISS,CODY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 07/02/2021 03:55:36

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.079	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2		03:54	No Sample Provided
Breath Volume		0.000	03:54	
Air Blank		0.000	03:54	
Control Test		0.079	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/02/2021 21:21:51

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.080	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.032	21:16
Breath Volume		1.214	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:18
Subject Sample	# 2	0.029	21:20
Breath Volume		1.250	21:20
Air Blank		0.000	21:20
Control Test		0.080	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLASENOR,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: VILLASENOR,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ALLEY

DATE/TIME OF TEST: 07/03/2021 01:10:58

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.000	01:06
Breath Volume		3.796	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.000	01:09
Breath Volume		3.867	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CROFOOT

DATE/TIME OF TEST: 07/03/2021 01:41:20

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.144	01:36
Breath Volume		2.835	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.135	01:39
Breath Volume		3.167	01:39
Air Blank		0.000	01:40
Control Test		0.079	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALL,CODY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BALL,CODY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 07/03/2021 23:46:53

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.177	23:39
Breath Volume		1.925	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.194	23:44
Breath Volume		3.046	23:44
Air Blank		0.000	23:45
Control Test		0.078	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VANMATRE

DATE/TIME OF TEST: 07/04/2021 01:42:17

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.000	01:37
Breath Volume		4.074	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.000	01:40
Breath Volume		1.738	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: REDDY

DATE/TIME OF TEST: 07/04/2021 20:59:58

Diagnostics Check		OK	20:54
Air Blank		0.000	20:54
Control Test		0.080	20:54
Air Blank		0.000	20:55
Subject Sample	# 1	0.000	20:55
Breath Volume		2.382	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.000	20:58
Breath Volume		1.824	20:58
Air Blank		0.000	20:59
Control Test		0.078	20:59
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEPHART,HUNTER,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KEPHART,H
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PETH

DATE/TIME OF TEST: 07/06/2021 00:44:34

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.080	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Control Test		0.080	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: PHILPOT

DATE/TIME OF TEST: 07/06/2021 22:34:11

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:22	
Control Test		0.079	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1	0.113	22:26	Volume Not Met
Breath Volume		0.441	22:26	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 2	0.117	22:29	
Breath Volume		2.167	22:29	
Air Blank		0.000	22:29	
Air Blank		0.000	22:31	
Subject Sample	# 3	0.117	22:32	
Breath Volume		1.281	22:32	
Air Blank		0.000	22:33	
Control Test		0.078	22:33	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCQUAIG,JASMINE,T
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MCQUAIG,JASMINE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: AGUDELO

DATE/TIME OF TEST: 07/07/2021 01:31:44

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.076	01:19	
Breath Volume		1.777	01:19	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 2		01:24	No Sample Provided
Breath Volume		0.000	01:24	
Air Blank		0.000	01:25	
Air Blank		0.000	01:27	
Subject Sample	# 3		01:30	No Sample Provided
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Control Test		0.079	01:31	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BAKER,MATTHEW
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: KOERNSCHILD

DATE/TIME OF TEST: 07/07/2021 17:55:18

Diagnostics Check		OK	17:49
Air Blank		0.000	17:49
Control Test		0.079	17:49
Air Blank		0.000	17:50
Subject Sample	# 1	0.255	17:50
Breath Volume		3.542	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.239	17:53
Breath Volume		3.097	17:53
Air Blank		0.000	17:54
Control Test		0.079	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,JORDAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CESPEDES,ROXANA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 07/08/2021 03:21:59

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.078	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.208	03:16
Breath Volume		1.621	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.223	03:20
Breath Volume		1.714	03:20
Air Blank		0.000	03:21
Control Test		0.077	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: TAPANES,CODY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 07/08/2021 18:44:57

Diagnostics Check		OK	18:38
Air Blank		0.000	18:38
Control Test		0.079	18:39
Air Blank		0.000	18:39
Subject Sample	# 1	0.172	18:40
Breath Volume		2.679	18:40
Air Blank		0.000	18:41
Air Blank		0.000	18:42
Subject Sample	# 2	0.160	18:43
Breath Volume		2.277	18:43
Air Blank		0.000	18:44
Control Test		0.078	18:44
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CUTINO

DATE/TIME OF TEST: 07/09/2021 00:33:26

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.000	00:28
Breath Volume		2.320	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.000	00:32
Breath Volume		1.683	00:32
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,JOSIE,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WELLS,JOSIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ABREGO VELIZ

DATE/TIME OF TEST: 07/10/2021 00:36:27

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.079	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1	0.148	00:34	
Breath Volume		1.195	00:34	
Air Blank			00:34	RFI Detect
Air Blank		0.000	00:35	
Control Test		0.080	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,JOSIE,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WELLS,JOSIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ABREGO VELIZ

DATE/TIME OF TEST: 07/10/2021 00:49:14

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.077	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.164	00:44
Breath Volume		2.105	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.176	00:47
Breath Volume		2.429	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALVET,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HOLLINGSWORTH,ROGER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/10/2021 03:02:28

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.159	02:57
Breath Volume		3.292	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.158	03:00
Breath Volume		3.917	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEST,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 07/10/2021 05:42:18

Diagnostics Check		OK	05:30	
Air Blank		0.000	05:31	
Control Test		0.079	05:31	
Air Blank		0.000	05:32	
Subject Sample	# 1	0.177	05:35	Volume Not Met
Breath Volume		0.855	05:35	
Air Blank		0.000	05:36	
Air Blank		0.000	05:37	
Subject Sample	# 2	0.201	05:40	Volume Not Met
Breath Volume		0.292	05:40	
Air Blank			05:41	Purge Fail
Air Blank		0.000	05:42	

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEST,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 06/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 07/10/2021 06:11:02

Diagnostics Check		OK	06:03
Air Blank		0.000	06:04
Control Test		0.078	06:04
Air Blank		0.000	06:05
Subject Sample	# 1	0.192	06:05
Breath Volume		1.332	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 2	0.194	06:09
Breath Volume		2.042	06:09
Air Blank		0.000	06:10
Control Test		0.079	06:10
Air Blank		0.000	06:10
Diagnostics Check		OK	06:11

Data Download Date/Time: 07/12/2021 11:08

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: CRAFFORD

DATE/TIME OF TEST: 07/16/2021 06:43:09

Diagnostics Check		OK	06:36
Air Blank		0.000	06:37
Control Test		0.076	06:37
Air Blank		0.000	06:37
Subject Sample	# 1	0.129	06:38
Breath Volume		1.484	06:38
Air Blank		0.000	06:39
Air Blank		0.000	06:40
Subject Sample	# 2	0.122	06:41
Breath Volume		1.707	06:41
Air Blank		0.000	06:42
Control Test		0.077	06:42
Air Blank		0.000	06:43
Diagnostics Check		OK	06:43

Data Download Date/Time: 08/04/2021 15:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HEYMANS,M
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A2
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:22

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 07/27/2021 09:52:31

Diagnostics Check		OK	09:46
Air Blank		0.000	09:46
Control Test		0.078	09:47
Air Blank		0.000	09:47
Subject Sample	# 1	0.000	09:48
Breath Volume		2.226	09:48
Air Blank		0.000	09:48
Air Blank		0.000	09:50
Subject Sample	# 2	0.000	09:51
Breath Volume		1.765	09:51
Air Blank		0.000	09:51
Control Test		0.078	09:51
Air Blank		0.000	09:52
Diagnostics Check		OK	09:52

Data Download Date/Time: 08/04/2021 15:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDLE,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORDLE,ABIGAIL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: TATRU

DATE/TIME OF TEST: 07/01/2021 23:48:04

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:46	Subject Test Refused
Breath Volume		0.000	23:46	
Air Blank		0.000	23:47	
Control Test		0.078	23:47	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BROWN,J
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: CULBERTH

DATE/TIME OF TEST: 07/02/2021 16:11:52

Diagnostics Check		OK	16:05
Air Blank		0.000	16:05
Control Test		0.079	16:06
Air Blank		0.000	16:06
Subject Sample	# 1	0.131	16:07
Breath Volume		2.542	16:07
Air Blank		0.000	16:07
Air Blank		0.000	16:09
Subject Sample	# 2	0.129	16:10
Breath Volume		2.335	16:10
Air Blank		0.000	16:10
Control Test		0.078	16:11
Air Blank		0.000	16:11
Diagnostics Check		OK	16:11

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SISSON

DATE/TIME OF TEST: 07/03/2021 03:34:25

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:20	
Control Test		0.079	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1	0.118	03:21	
Breath Volume		1.750	03:21	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 3		03:33	No Sample Provided
Breath Volume		0.000	03:33	
Air Blank		0.000	03:33	
Control Test		0.078	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LEWIS,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: VAN FLEET

DATE/TIME OF TEST: 07/04/2021 05:23:13

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.079	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1	0.148	05:13	
Breath Volume		1.503	05:13	
Air Blank		0.000	05:13	
Air Blank		0.000	05:15	
Subject Sample	# 2		05:18	No Sample Provided
Breath Volume		0.000	05:18	
Air Blank		0.000	05:19	
Air Blank		0.000	05:21	
Subject Sample	# 3		05:21	Subject Test Refused
Breath Volume		0.000	05:21	
Air Blank		0.000	05:22	
Control Test		0.079	05:22	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: PRESTON,MIKAYLA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: RAVAN

DATE/TIME OF TEST: 07/05/2021 23:46:14

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.158	23:41
Breath Volume		2.000	23:41
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.167	23:44
Breath Volume		1.585	23:44
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HERITAGE

DATE/TIME OF TEST: 07/08/2021 03:30:21

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:29	RFI Detect
Breath Volume		0.000	03:29	
Air Blank			03:29	RFI Detect
Air Blank		0.000	03:30	

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HERITAGE

DATE/TIME OF TEST: 07/08/2021 03:38:31

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.275	03:34
Breath Volume		3.738	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.266	03:36
Breath Volume		1.906	03:36
Air Blank		0.000	03:37
Control Test		0.077	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: HULL

DATE/TIME OF TEST: 07/08/2021 21:50:47

Diagnostics Check		OK	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:45
Subject Sample	# 1	0.319	21:45
Breath Volume		4.019	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.329	21:48
Breath Volume		3.671	21:48
Air Blank		0.000	21:49
Control Test		0.077	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSORIA,JAROLIN,X
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: LINSKY,MICHAEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: OSORIA

DATE/TIME OF TEST: 07/18/2021 02:34:25

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.172	02:29
Breath Volume		1.171	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.173	02:32
Breath Volume		1.421	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,WALKER,L
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SIMS,WALKER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: QUILAL-LAN

DATE/TIME OF TEST: 07/18/2021 06:03:31

Diagnostics Check		OK	05:57
Air Blank		0.000	05:57
Control Test		0.079	05:58
Air Blank		0.000	05:58
Subject Sample	# 1	0.065	05:59
Breath Volume		2.316	05:59
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.065	06:02
Breath Volume		3.371	06:02
Air Blank		0.000	06:02
Control Test		0.079	06:02
Air Blank		0.000	06:03
Diagnostics Check		OK	06:03

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JOHNATHAN,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BROWN,JOHNATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/21/2021 18:57:50

Diagnostics Check		OK	18:50
Air Blank		0.000	18:50
Control Test		0.079	18:51
Air Blank		0.000	18:51
Subject Sample	# 1	0.237	18:53
Breath Volume		1.156	18:53
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.256	18:56
Breath Volume		1.734	18:56
Air Blank		0.000	18:56
Control Test		0.078	18:57
Air Blank		0.000	18:57
Diagnostics Check		OK	18:57

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESTON,MIKAYLA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WINTERS,ERIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 07/23/2021 17:36:54

Diagnostics Check		OK	17:25	
Air Blank		0.000	17:25	
Control Test		0.079	17:26	
Air Blank		0.000	17:26	
Subject Sample	# 1		17:29	No Sample Provided
Breath Volume		0.000	17:29	
Air Blank		0.000	17:30	
Air Blank		0.000	17:32	
Subject Sample	# 2		17:35	No Sample Provided
Breath Volume		0.000	17:35	
Air Blank		0.000	17:35	
Control Test		0.079	17:36	
Air Blank		0.000	17:36	
Diagnostics Check		OK	17:36	

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DABROWSKI

DATE/TIME OF TEST: 07/25/2021 01:59:15

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:58	
Control Test		0.079	01:58	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: PADGETTE,E
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: ZANNINO

DATE/TIME OF TEST: 07/27/2021 15:41:13

Diagnostics Check		OK	15:31
Air Blank		0.000	15:31
Control Test		0.079	15:32
Air Blank		0.000	15:32
Subject Sample	# 1	0.190	15:33
Breath Volume		1.261	15:33
Air Blank		0.000	15:34
Air Blank		0.000	15:36
Subject Sample	# 2	0.216	15:36
Breath Volume		3.355	15:36
Air Blank		0.000	15:37
Air Blank		0.000	15:39
Subject Sample	# 3	0.197	15:39
Breath Volume		2.800	15:39
Air Blank		0.000	15:40
Control Test		0.078	15:40
Air Blank		0.000	15:41
Diagnostics Check		OK	15:41

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 06/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/27/2021 23:08:00

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.116	23:03
Breath Volume		2.503	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.097	23:06
Breath Volume		1.285	23:06
Air Blank		0.000	23:07
Control Test		0.077	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:08

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 07/28/2021 23:10:51

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:07	
Control Test		0.077	23:07	
Air Blank		0.000	23:08	
Subject Sample	# 1		23:09	Subject Test Refused
Breath Volume		0.000	23:09	
Air Blank		0.000	23:09	
Control Test		0.077	23:10	
Air Blank		0.000	23:10	
Diagnostics Check		OK	23:10	

Data Download Date/Time: 07/29/2021 14:07

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/15/2021 02:36:48

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.082	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.000	02:32
Breath Volume		1.121	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.000	02:35
Breath Volume		1.445	02:35
Air Blank		0.000	02:35
Control Test		0.080	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 08/18/2021 12:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ISAKH,FAREISHA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 07/16/2021 03:28:50

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.082	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.000	03:22
Breath Volume		1.449	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.000	03:27
Breath Volume		1.355	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 08/18/2021 12:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 07/17/2021 02:06:02

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.082	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.191	02:01
Breath Volume		3.078	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.198	02:04
Breath Volume		3.218	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:06

Data Download Date/Time: 08/18/2021 12:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: RIVIERA MURILLO

DATE/TIME OF TEST: 07/17/2021 03:52:48

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.082	03:45	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.312	03:48	Volume Not Met
Breath Volume		0.636	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 2		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Control Test		0.077	03:52	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 08/18/2021 12:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SLAYTON,GARY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 07/17/2021 06:40:17

Diagnostics Check		OK	06:31
Air Blank		0.000	06:32
Control Test		0.082	06:32
Air Blank		0.000	06:33
Subject Sample	# 1	0.201	06:35
Breath Volume		1.871	06:35
Air Blank		0.000	06:36
Air Blank		0.000	06:38
Subject Sample	# 2	0.211	06:38
Breath Volume		1.449	06:38
Air Blank		0.000	06:39
Control Test		0.078	06:39
Air Blank		0.000	06:40
Diagnostics Check		OK	06:40

Data Download Date/Time: 08/18/2021 12:38

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: KORSMAN

DATE/TIME OF TEST: 07/18/2021 23:44:12

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.081	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.162	23:39
Breath Volume		2.203	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.158	23:42
Breath Volume		2.144	23:42
Air Blank		0.000	23:43
Control Test		0.081	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 08/18/2021 12:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVAARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PENSO

DATE/TIME OF TEST: 07/22/2021 01:54:27

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.082	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.200	01:49
Breath Volume		1.890	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.201	01:52
Breath Volume		1.601	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 08/18/2021 12:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 07/23/2021 05:52:45

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.082	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.145	05:47
Breath Volume		1.527	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.143	05:51
Breath Volume		1.363	05:51
Air Blank		0.000	05:51
Control Test		0.081	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 08/18/2021 12:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MCFARLANE

DATE/TIME OF TEST: 07/25/2021 23:38:39

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.081	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.218	23:33
Breath Volume		1.496	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.215	23:36
Breath Volume		1.355	23:36
Air Blank		0.000	23:37
Control Test		0.079	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 08/18/2021 12:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: ESCOBOZA

DATE/TIME OF TEST: 07/26/2021 02:12:42

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.082	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.037	02:08
Breath Volume		3.421	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.036	02:11
Breath Volume		2.898	02:11
Air Blank		0.000	02:11
Control Test		0.082	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 08/18/2021 12:39

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ISAKH,FAREISHA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALPHONSE

DATE/TIME OF TEST: 07/02/2021 04:40:49

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.030	04:35
Breath Volume		1.261	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.030	04:39
Breath Volume		2.265	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: QUIGG

DATE/TIME OF TEST: 07/03/2021 17:58:03

Diagnostics Check		OK	17:51
Air Blank		0.000	17:52
Control Test		0.081	17:52
Air Blank		0.000	17:53
Subject Sample	# 1	0.000	17:53
Breath Volume		2.183	17:53
Air Blank		0.000	17:54
Air Blank		0.000	17:56
Subject Sample	# 2	0.000	17:56
Breath Volume		1.890	17:56
Air Blank		0.000	17:57
Control Test		0.081	17:57
Air Blank		0.000	17:57
Diagnostics Check		OK	17:58

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOLLOWAY,DEVARIOUS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 07/04/2021 03:50:58

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.000	03:46
Breath Volume		2.445	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.000	03:49
Breath Volume		2.953	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SZEGEDY

DATE/TIME OF TEST: 07/05/2021 03:02:03

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.081	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.240	02:57
Breath Volume		2.351	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.240	03:00
Breath Volume		2.121	03:00
Air Blank		0.000	03:01
Control Test		0.076	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RAMIREZ-RUIZ

DATE/TIME OF TEST: 07/07/2021 01:09:48

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:55	
Control Test		0.081	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1	0.216	00:59	
Breath Volume		1.613	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 2	0.214	01:04	Volume Not Met
Breath Volume		0.882	01:04	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 3	0.219	01:08	
Breath Volume		2.355	01:08	
Air Blank		0.000	01:08	
Control Test		0.075	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 07/07/2021 04:30:00

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.302	04:25
Breath Volume		2.140	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:27
Subject Sample	# 2	0.301	04:28
Breath Volume		2.253	04:28
Air Blank		0.000	04:29
Control Test		0.079	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:30

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHINCHILLA,SERGIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/07/2021 22:25:04

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.081	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.158	22:20
Breath Volume		1.621	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:22
Subject Sample	# 2	0.159	22:23
Breath Volume		1.199	22:23
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:25

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MORGAN,DAVID
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/08/2021 20:49:44

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.081	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.172	20:45
Breath Volume		2.312	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.166	20:48
Breath Volume		2.292	20:48
Air Blank		0.000	20:48
Control Test		0.080	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ISAKH,FAREISHA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MASTROGIANNI

DATE/TIME OF TEST: 07/09/2021 04:02:06

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.081	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.111	03:57
Breath Volume		2.320	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.112	04:00
Breath Volume		1.953	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 07/10/2021 03:50:14

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.081	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.103	03:45
Breath Volume		1.640	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.112	03:48
Breath Volume		1.632	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WOODARD,DONTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 07/10/2021 18:03:21

Diagnostics Check		OK	17:56
Air Blank		0.000	17:57
Control Test		0.080	17:57
Air Blank		0.000	17:58
Subject Sample	# 1	0.091	17:58
Breath Volume		2.542	17:58
Air Blank		0.000	17:59
Air Blank		0.000	18:01
Subject Sample	# 2	0.093	18:01
Breath Volume		2.632	18:01
Air Blank		0.000	18:02
Control Test		0.081	18:02
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 07/14/2021 10:25

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEE,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: GARCIAZ

DATE/TIME OF TEST: 07/04/2021 23:03:20

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.080	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.231	22:57
Breath Volume		1.121	22:57
Air Blank		0.000	22:58
Air Blank		0.000	22:59
Subject Sample	# 2	0.236	23:01
Breath Volume		2.597	23:01
Air Blank		0.000	23:02
Control Test		0.080	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 07/14/2021 11:22

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TRABUE,THOMAS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LOVINGSON

DATE/TIME OF TEST: 07/05/2021 02:10:36

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.168	02:04
Breath Volume		2.613	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.173	02:08
Breath Volume		1.378	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 07/14/2021 11:22

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ABREU,LESTER
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 07/08/2021 10:54:52

Diagnostics Check		OK	10:46
Air Blank		0.000	10:46
Control Test		0.080	10:47
Air Blank		0.000	10:47
Subject Sample	# 1	0.049	10:48
Breath Volume		1.945	10:48
Air Blank		0.000	10:48
Air Blank		0.000	10:50
Subject Sample	# 2	0.048	10:53
Breath Volume		2.898	10:53
Air Blank		0.000	10:53
Control Test		0.081	10:54
Air Blank		0.000	10:54
Diagnostics Check		OK	10:54

Data Download Date/Time: 07/14/2021 11:22

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FANFAN,KUSHNER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: COLQUITT

DATE/TIME OF TEST: 07/11/2021 01:42:57

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.080	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 1	0.158	01:35	Volume Not Met
Breath Volume		0.472	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 2	0.165	01:41	Volume Not Met
Breath Volume		0.417	01:41	
Air Blank		0.000	01:42	
Control Test		0.080	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 07/14/2021 11:22

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ISAKH,FARIESHA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SMELLIE

DATE/TIME OF TEST: 07/13/2021 02:12:32

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.081	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.000	02:08
Breath Volume		1.457	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.000	02:11
Breath Volume		2.125	02:11
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 07/14/2021 11:22

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUE,RUDY
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CORRALES

DATE/TIME OF TEST: 07/27/2021 20:25:11

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.080	20:15
Air Blank		0.000	20:15
Subject Sample	# 1	0.225	20:17
Breath Volume		1.285	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.212	20:23
Breath Volume		1.132	20:23
Air Blank		0.000	20:24
Control Test		0.079	20:24
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 08/19/2021 01:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: O'HERRON

DATE/TIME OF TEST: 07/28/2021 02:54:25

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.162	02:49
Breath Volume		2.867	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.156	02:52
Breath Volume		2.683	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 08/19/2021 01:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMERO,ELISA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GRANIZO

DATE/TIME OF TEST: 07/29/2021 00:31:54

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.130	00:27
Breath Volume		1.703	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.130	00:30
Breath Volume		1.878	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 08/19/2021 01:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ISAKH,FAREISHA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 07/29/2021 04:20:15

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.194	04:15
Breath Volume		1.457	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.192	04:18
Breath Volume		2.136	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 08/19/2021 01:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 07/29/2021 16:14:29

Diagnostics Check		OK	16:06
Air Blank		0.000	16:06
Control Test		0.080	16:06
Air Blank		0.000	16:07
Subject Sample	# 1	0.262	16:09
Breath Volume		1.457	16:09
Air Blank		0.000	16:10
Air Blank		0.000	16:12
Subject Sample	# 2	0.257	16:12
Breath Volume		1.644	16:12
Air Blank		0.000	16:13
Control Test		0.079	16:13
Air Blank		0.000	16:14
Diagnostics Check		OK	16:14

Data Download Date/Time: 08/19/2021 01:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/30/2021 22:15:00

Diagnostics Check		OK	22:07
Air Blank		0.000	22:07
Control Test		0.080	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.269	22:09
Breath Volume		1.687	22:09
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.272	22:13
Breath Volume		1.613	22:13
Air Blank		0.000	22:14
Control Test		0.081	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:15

Data Download Date/Time: 08/19/2021 01:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/31/2021 01:52:43

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.080	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 1	0.146	01:40	
Breath Volume		1.351	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 2		01:45	No Sample Provided
Breath Volume		0.000	01:45	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 3		01:51	No Sample Provided
Breath Volume		0.000	01:51	
Air Blank		0.000	01:51	
Control Test		0.081	01:52	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 08/19/2021 01:28

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/31/2021 02:18:00

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.081	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.140	02:13
Breath Volume		1.898	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.137	02:16
Breath Volume		1.484	02:16
Air Blank		0.000	02:17
Control Test		0.078	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:18

Data Download Date/Time: 08/19/2021 01:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOLDBERG,JOSH
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/31/2021 16:05:21

Diagnostics Check		OK	15:58
Air Blank		0.000	15:58
Control Test		0.080	15:58
Air Blank		0.000	15:59
Subject Sample	# 1	0.201	16:00
Breath Volume		1.144	16:00
Air Blank		0.000	16:01
Air Blank		0.000	16:03
Subject Sample	# 2	0.219	16:03
Breath Volume		1.539	16:03
Air Blank		0.000	16:04
Control Test		0.079	16:04
Air Blank		0.000	16:05
Diagnostics Check		OK	16:05

Data Download Date/Time: 08/19/2021 01:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 07/31/2021 19:36:59

Diagnostics Check		OK	19:30
Air Blank		0.000	19:31
Control Test		0.081	19:31
Air Blank		0.000	19:32
Subject Sample	# 1	0.085	19:32
Breath Volume		3.390	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:34
Subject Sample	# 2	0.087	19:35
Breath Volume		3.894	19:35
Air Blank		0.000	19:36
Control Test		0.081	19:36
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 08/19/2021 01:29

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,Q
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03521080A3
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LOVELACE

DATE/TIME OF TEST: 07/23/2021 23:34:42

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.204	23:30
Breath Volume		3.164	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.212	23:33
Breath Volume		2.960	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 08/25/2021 20:40

Printed: 19 October 2021

COBRA ©1997-2014 CMI, Inc.