

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GAMBY

DATE/TIME OF TEST: 08/11/2021 23:38:42

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.077	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:34
Breath Volume	1.292	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.000	23:37
Breath Volume	1.375	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON,CASEY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: KEEL

DATE/TIME OF TEST: 08/13/2021 00:21:41

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.077	00:11	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:14	No Sample Provided
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Air Blank	0.000	00:17	
Subject Sample	# 2	00:20	No Sample Provided
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.077	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH, SEAN, J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON, CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LEVINSKY

DATE/TIME OF TEST: 08/13/2021 01:51:39

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.077	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:44	No Sample Provided
Breath Volume	0.000	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 2	01:50	No Sample Provided
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.077	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCULLOUGH, SEAN, J

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: ROBISON, CASEY

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 21:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 08/13/2021 21:50:56

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.077	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.000	21:46
Breath Volume	3.941	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.000	21:49
Breath Volume	3.925	21:49
Air Blank	0.000	21:50
Control Test	0.077	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: KAPO

DATE/TIME OF TEST: 08/14/2021 01:56:39

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.178	01:52
Breath Volume	3.105	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.174	01:55
Breath Volume	2.835	01:55
Air Blank	0.000	01:55
Control Test	0.077	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: MCCULLOUGH,SEAN

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: POTOCZEK

DATE/TIME OF TEST: 08/14/2021 03:36:27

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.077	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.249	03:29
Breath Volume	2.277	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.223	03:31
Breath Volume	1.296	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 3	0.244	03:34
Breath Volume	1.324	03:34
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ,SAMANTHA,E

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: ROBISON,CASEY

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 08/15/2021 00:14:30

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.076	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.000	00:10
Breath Volume	1.339	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.000	00:13
Breath Volume	1.679	00:13
Air Blank	0.000	00:13
Control Test	0.077	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 08/15/2021 03:39:46

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.000	03:35
Breath Volume	1.167	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.000	03:38
Breath Volume	1.539	03:38
Air Blank	0.000	03:38
Control Test	0.077	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/17/2021 18:30:37

Diagnostics Check	OK	18:19	
Air Blank	0.000	18:19	
Control Test	0.077	18:20	
Air Blank	0.000	18:20	
Subject Sample	# 1	18:23	No Sample Provided
Breath Volume	0.000	18:23	
Air Blank	0.000	18:24	
Air Blank	0.000	18:26	
Subject Sample	# 2	18:29	No Sample Provided
Breath Volume	0.000	18:29	
Air Blank	0.000	18:29	
Control Test	0.078	18:30	
Air Blank	0.000	18:30	
Diagnostics Check	OK	18:30	

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PALOMINO

DATE/TIME OF TEST: 08/17/2021 19:44:30

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:31	
Control Test	0.077	19:31	
Air Blank	0.000	19:32	
Subject Sample	# 1	19:35	No Sample Provided
Breath Volume	0.000	19:35	
Air Blank	0.000	19:35	
Air Blank	0.000	19:37	
Subject Sample	# 2	0.164	19:39
Breath Volume	3.207	19:39	
Air Blank	0.000	19:40	
Air Blank	0.000	19:42	
Subject Sample	# 3	0.161	19:43
Breath Volume	1.246	19:43	
Air Blank	0.000	19:43	
Control Test	0.077	19:43	
Air Blank	0.000	19:44	
Diagnostics Check	OK	19:44	

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VAN HOUTEN, JONATHAN, J

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ, SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 08/18/2021 01:39:55

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.077	01:39	
Air Blank		01:39	RFI Detect
Air Blank	0.000	01:39	

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VAN HOUTEN, JONATHAN, J

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ, SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 08/18/2021 01:54:04

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.077	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.077	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VAN HOUTEN, JONATHAN, J

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ, SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:52

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LESHKO

DATE/TIME OF TEST: 08/19/2021 01:06:05

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.077	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.222	01:01
Breath Volume	1.429	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.237	01:04
Breath Volume	1.664	01:04
Air Blank	0.000	01:05
Control Test	0.077	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:10

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: LEPICARD

DATE/TIME OF TEST: 08/19/2021 19:29:54

Diagnostics Check	OK	19:23
Air Blank	0.000	19:24
Control Test	0.077	19:24
Air Blank	0.000	19:25
Subject Sample # 1	0.000	19:25
Breath Volume	1.863	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.000	19:28
Breath Volume	1.878	19:28
Air Blank	0.000	19:29
Control Test	0.078	19:29
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 08/31/2021 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: TABB

DATE/TIME OF TEST: 08/20/2021 02:45:56

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.177	02:39
Breath Volume	1.410	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.160	02:44
Breath Volume	2.308	02:44
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/31/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LONARDO

DATE/TIME OF TEST: 08/21/2021 01:34:16

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.078	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2	0.199	01:28
Breath Volume	1.289	01:28	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample	# 3	0.201	01:32
Breath Volume	1.679	01:32	
Air Blank	0.000	01:33	
Control Test	0.077	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 08/31/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ,SAMANTHA,E

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: ROBISON,CASEY

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: IM

DATE/TIME OF TEST: 08/21/2021 02:59:07

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.077	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.147	02:54
Breath Volume	1.312	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.143	02:57
Breath Volume	1.390	02:57
Air Blank	0.000	02:58
Control Test	0.077	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 08/31/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/21/2021 03:54:55

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.213	03:50
Breath Volume	2.132	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.225	03:53
Breath Volume	2.433	03:53
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 08/31/2021 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,HENRY,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,HENRY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 00521080A2

DRY GAS EXPIRY DATE: 02/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CORIOLAN

DATE/TIME OF TEST: 08/07/2021 19:30:29

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.000	19:25
Breath Volume	2.175	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.000	19:29
Breath Volume	1.937	19:29
Air Blank	0.000	19:29
Control Test	0.081	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,HENRY,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,HENRY

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 00521080A2

DRY GAS EXPIRY DATE: 02/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/08/2021 03:10:14

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.063	03:05
Breath Volume	3.414	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.061	03:08
Breath Volume	1.671	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KURINSKY,DAVID,L

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: WOODIE,LANGSTON

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 00521080A2

DRY GAS EXPIRY DATE: 02/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/09/2021 22:37:29

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.000	22:32
Breath Volume	3.750	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.000	22:36
Breath Volume	3.304	22:36
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REED,CHAD,R

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: REID,JONATHAN

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 00521080A2

DRY GAS EXPIRY DATE: 02/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/10/2021 05:04:42

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.081	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.028	04:59
Breath Volume	1.707	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.028	05:03
Breath Volume	1.636	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HURT,JONATHAN,R

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HURT,JONATHAN

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 00521080A2

DRY GAS EXPIRY DATE: 02/05/2023

VIOLATION CODE: DUI

ARREST TIME: 08:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: CZERNIAWSKI

DATE/TIME OF TEST: 08/12/2021 09:01:33

Diagnostics Check	OK	08:54
Air Blank	0.000	08:54
Control Test	0.082	08:55
Air Blank	0.000	08:55
Subject Sample # 1	0.243	08:56
Breath Volume	3.273	08:56
Air Blank	0.000	08:57
Air Blank	0.000	08:59
Subject Sample # 2	0.238	08:59
Breath Volume	3.867	08:59
Air Blank	0.000	09:00
Control Test	0.081	09:00
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,J,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/13/2021 00:52:15

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.081	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	RFI Detect
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUIS,J,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: BUIS,J

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13921080A2

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 08/13/2021 00:59:40

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.043	00:55
Breath Volume	2.117	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.044	00:58
Breath Volume	2.089	00:58
Air Blank	0.000	00:58
Control Test	0.082	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13921080A2

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 08/13/2021 02:08:59

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.169	02:04
Breath Volume	2.144	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:06
Subject Sample # 2	0.170	02:07
Breath Volume	2.121	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13921080A2

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 06:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: LOPEZ-MORA

DATE/TIME OF TEST: 08/15/2021 06:57:31

Diagnostics Check	OK	06:50
Air Blank	0.000	06:51
Control Test	0.081	06:51
Air Blank	0.000	06:52
Subject Sample # 1	0.199	06:52
Breath Volume	1.593	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.205	06:55
Breath Volume	2.503	06:55
Air Blank	0.000	06:56
Control Test	0.081	06:56
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KROLICK,ALEX,M

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: KROLICK,ALEX

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13921080A2

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 20:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RODRIGUEZ-SANCHES

DATE/TIME OF TEST: 08/15/2021 21:38:01

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.000	21:33
Breath Volume	1.156	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.000	21:36
Breath Volume	1.558	21:36
Air Blank	0.000	21:37
Control Test	0.081	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:38

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HURT,JONATHAN,R

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: HURT,JONATHAN

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13921080A2

DRY GAS EXPIRY DATE: 07/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GUITIERREZ

DATE/TIME OF TEST: 08/18/2021 21:42:40

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.083	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.220	21:37
Breath Volume	1.582	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.240	21:41
Breath Volume	2.316	21:41
Air Blank	0.000	21:41
Control Test	0.081	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 10/27/2021 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MCCLOSKEY

DATE/TIME OF TEST: 08/07/2021 02:25:56

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.078	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.173	02:20
Breath Volume	1.367	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.165	02:24
Breath Volume	1.957	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 08/08/2021 01:58:40

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.077	01:54
Breath Volume	1.945	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.072	01:57
Breath Volume	2.222	01:57
Air Blank	0.000	01:57
Control Test	0.076	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: LARUE

DATE/TIME OF TEST: 08/22/2021 02:25:01

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.183	02:19
Breath Volume	1.328	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.179	02:23
Breath Volume	1.433	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:25

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:33

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ECHIKSON

DATE/TIME OF TEST: 08/25/2021 22:50:34

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.230	22:46
Breath Volume	2.617	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.219	22:49
Breath Volume	1.843	22:49
Air Blank	0.000	22:49
Control Test	0.078	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 08/26/2021 02:01:35

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.113	01:57
Breath Volume	2.441	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.113	02:00
Breath Volume	2.855	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: FONGER

DATE/TIME OF TEST: 08/27/2021 00:32:44

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.125	00:28
Breath Volume	1.316	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.140	00:31
Breath Volume	4.218	00:31
Air Blank	0.000	00:31
Control Test	0.077	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 08/27/2021 22:28:30

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.079	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.215	22:23
Breath Volume	1.683	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.205	22:27
Breath Volume	2.519	22:27
Air Blank	0.000	22:27
Control Test	0.078	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VILLAMOR

DATE/TIME OF TEST: 08/28/2021 01:16:39

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.078	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.109	01:12
Breath Volume	2.082	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.107	01:15
Breath Volume	1.304	01:15
Air Blank	0.000	01:15
Control Test	0.077	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: MORGAN,JONATHAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/28/2021 03:42:16

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.171	03:37
Breath Volume	1.234	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.175	03:40
Breath Volume	1.253	03:40
Air Blank	0.000	03:41
Control Test	0.077	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARLEY,JOSEPH,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: FARLEY,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 35019080A3

DRY GAS EXPIRY DATE: 02/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/29/2021 02:21:47

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.187	02:17
Breath Volume	5.183	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.184	02:20
Breath Volume	4.675	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 08/30/2021 17:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 08/07/2021 21:18:38

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
Control Test	0.079	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:17	
Control Test	0.079	21:18	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:18	

Data Download Date/Time: 08/31/2021 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN, TYLER, J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: DEAN, TYLER
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 08/11/2021 00:18:56

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.080	00:12	
Air Blank	0.000	00:12	
Subject Sample	# 1 0.277	00:13	
Breath Volume	3.347	00:13	
Air Blank	0.000	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 2	00:17	Subject Test Refused
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Control Test	0.079	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENTINE, RONALD, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VALENTINE, RONALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 08/14/2021 21:54:26

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.080	21:52	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:53	Subject Test Refused
Breath Volume	0.000	21:53	
Air Blank	0.000	21:53	
Control Test	0.080	21:53	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:54	

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REED,SAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DICARLO

DATE/TIME OF TEST: 08/17/2021 02:46:09

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.029	02:41
Breath Volume	2.406	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.027	02:44
Breath Volume	1.921	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMALL,BRANNON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: CULVER

DATE/TIME OF TEST: 08/19/2021 04:52:17

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.153	04:47
Breath Volume	2.730	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.151	04:50
Breath Volume	2.421	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RUBLEY

DATE/TIME OF TEST: 08/26/2021 23:07:16

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:06	Sequence Aborted
Breath Volume	0.000	23:06	
Air Blank	0.000	23:07	

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANGER,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RUBLEY

DATE/TIME OF TEST: 08/26/2021 23:26:26

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.000	23:22
Breath Volume	2.679	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.000	23:25
Breath Volume	2.867	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 08/28/2021 04:22:18

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.000	04:17
Breath Volume	2.929	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.000	04:20
Breath Volume	3.679	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 08/28/2021 20:51:58

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.152	20:47
Breath Volume	4.070	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.146	20:50
Breath Volume	5.093	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 08/28/2021 22:20:19

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.122	22:15
Breath Volume	2.582	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.116	22:18
Breath Volume	2.339	22:18
Air Blank	0.000	22:19
Control Test	0.077	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALEY,KRISTOPHER,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STALEY,KRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: CINA

DATE/TIME OF TEST: 08/28/2021 22:50:59

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:46	
Control Test	0.079	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:50	Sequence Aborted
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALEY,KRISTOPHER,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STALEY,KRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: CINA

DATE/TIME OF TEST: 08/28/2021 22:57:51

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:55	
Control Test	0.079	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.078	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 08/31/2021 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 11:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 08/19/2021 01:42:07

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.130	01:37
Breath Volume	1.296	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.129	01:40
Breath Volume	1.390	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 09/09/2021 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTE,JACOB,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LLOYD,GEORGE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: PEREA

DATE/TIME OF TEST: 08/22/2021 23:25:07

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.184	23:17
Breath Volume	1.777	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.230	23:20
Breath Volume	4.285	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 3	0.227	23:23
Breath Volume	3.429	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 09/09/2021 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: RHODEN,DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 08/24/2021 19:11:17

Diagnostics Check	OK	19:03
Air Blank	0.000	19:04
Control Test	0.080	19:04
Air Blank	0.000	19:04
Subject Sample # 1	0.347	19:05
Breath Volume	1.789	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.338	19:09
Breath Volume	1.148	19:09
Air Blank	0.000	19:10
Control Test	0.080	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 09/09/2021 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLINS,JACKSON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: BUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WILLMANN

DATE/TIME OF TEST: 08/26/2021 00:03:41

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.081	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 2	23:59	No Sample Provided
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 09/09/2021 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORMLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GORMLEY, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LOPICCOLO

DATE/TIME OF TEST: 08/26/2021 22:28:17

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.080	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 09/09/2021 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT, COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: HILSON

DATE/TIME OF TEST: 08/27/2021 01:28:39

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.185	01:24
Breath Volume	3.300	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.189	01:27
Breath Volume	3.242	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/09/2021 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLANDA, RICARDO, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HOLANDA, RICARDO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/29/2021 03:29:03

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.179	03:24
Breath Volume	2.535	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.178	03:27
Breath Volume	1.613	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:29

Data Download Date/Time: 09/09/2021 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 08/16/2021 22:58:49

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.045	22:50
Breath Volume	1.554	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.007	22:54
Breath Volume	2.136	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 3	0.046	22:57
Breath Volume	4.273	22:57
Air Blank	0.000	22:57
Control Test	0.077	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 09/09/2021 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT, COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 08/22/2021 03:58:32

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.170	03:53
Breath Volume	1.292	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.162	03:56
Breath Volume	1.296	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: PASCHAL

DATE/TIME OF TEST: 08/01/2021 00:37:30

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.147	00:32
Breath Volume	2.324	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.152	00:35
Breath Volume	1.609	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 08/12/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 08/06/2021 20:10:25

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.081	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.180	20:06
Breath Volume	2.042	20:06
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.177	20:08
Breath Volume	1.972	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 08/12/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON, EMILY, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MELTON, EMILY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 08/07/2021 21:02:52

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.082	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.000	20:57
Breath Volume	2.527	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.000	21:01
Breath Volume	1.585	21:01
Air Blank	0.000	21:01
Control Test	0.082	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 08/12/2021 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 08/07/2021 21:32:49

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.082	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.241	21:28
Breath Volume	2.437	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.243	21:31
Breath Volume	1.898	21:31
Air Blank	0.000	21:31
Control Test	0.081	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 08/12/2021 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARDY,JOE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 08/08/2021 22:23:04

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.166	22:18
Breath Volume	1.558	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.165	22:21
Breath Volume	1.425	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:23

Data Download Date/Time: 09/05/2021 05:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ALMANZA

DATE/TIME OF TEST: 08/01/2021 03:01:17

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.165	02:56
Breath Volume	1.414	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.163	02:59
Breath Volume	1.402	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 08/25/2021 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAEZ,JORDI
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: PASCUAL MORALES

DATE/TIME OF TEST: 08/01/2021 23:53:41

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 08/25/2021 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 08/03/2021 00:34:58

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.079	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 08/25/2021 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CLAYTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 08/06/2021 02:13:05

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.079	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1 0.240	02:00	
Breath Volume	1.785	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 2	02:05	No Sample Provided
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 3 0.202	02:11	Volume Not Met
Breath Volume	1.007	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 08/25/2021 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CABRAL,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CALLAWAY

DATE/TIME OF TEST: 08/20/2021 01:07:27

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.000	01:02
Breath Volume	2.250	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.000	01:06
Breath Volume	2.578	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 08/25/2021 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: RILEY,BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 07/13/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: GARCIA-MENDEZ

DATE/TIME OF TEST: 08/22/2021 06:04:53

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:54	
Subject Sample	# 1 0.135	05:57	Volume Not Met
Breath Volume	0.015	05:57	
Air Blank	0.000	05:58	
Air Blank	0.000	06:00	
Subject Sample	# 2	06:03	No Sample Provided
Breath Volume	0.000	06:03	
Air Blank	0.000	06:03	
Control Test	0.079	06:04	
Air Blank	0.000	06:04	
Diagnostics Check	OK	06:04	

Data Download Date/Time: 08/25/2021 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 08/25/2021 23:21:20

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.150	23:16
Breath Volume	2.269	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.148	23:19
Breath Volume	1.796	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 09/21/2021 15:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/26/2021 01:56:37

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.151	01:52
Breath Volume	1.640	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.149	01:55
Breath Volume	1.226	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 09/21/2021 15:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRAL,JONATHAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FLAHERTY,ZACHARY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DELL

DATE/TIME OF TEST: 08/29/2021 02:16:55

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:12
Breath Volume	2.511	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.000	02:15
Breath Volume	1.957	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY, JAKE, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOWERY, BRITTANY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 08/02/2021 05:21:14

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:07	
Control Test	0.081	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1 0.249	05:09	
Breath Volume	1.113	05:09	
Air Blank	0.000	05:10	
Air Blank	0.000	05:12	
Subject Sample	# 2	05:15	No Sample Provided
Breath Volume	0.000	05:15	
Air Blank	0.000	05:16	
Air Blank	0.000	05:18	
Subject Sample	# 3 0.248	05:19	
Breath Volume	2.281	05:19	
Air Blank	0.000	05:20	
Control Test	0.081	05:20	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LONDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MASON, N
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 08/02/2021 22:33:29

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Control Test	0.080	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 08/06/2021 01:49:50

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.081	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON, IAN, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE, JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: RAINWATER

DATE/TIME OF TEST: 08/07/2021 04:11:26

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.235	04:06
Breath Volume	4.667	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.236	04:09
Breath Volume	3.644	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARRINGTON, IAN, M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE, JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/08/2021 03:01:19

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GUNDLACH

DATE/TIME OF TEST: 08/08/2021 03:41:34

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.075	03:37
Breath Volume	3.792	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.076	03:40
Breath Volume	4.097	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: LINDSAY

DATE/TIME OF TEST: 08/08/2021 04:32:43

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.076	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:32	Sequence Aborted
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: LINDSAY

DATE/TIME OF TEST: 08/08/2021 05:00:34

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.212	04:56
Breath Volume	1.542	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.215	04:58
Breath Volume	1.492	04:58
Air Blank	0.000	04:59
Control Test	0.076	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/13/2021 22:11:15

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:08	
Control Test	0.081	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:09	Subject Test Refused
Breath Volume	0.000	22:09	
Air Blank	0.000	22:10	
Control Test	0.081	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EHRHART,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ROBISON

DATE/TIME OF TEST: 08/14/2021 03:37:18

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.082	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BULLOCK,ISIAH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CORMIER

DATE/TIME OF TEST: 08/14/2021 06:49:19

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.081	06:42
Air Blank	0.000	06:43
Subject Sample # 1	0.189	06:44
Breath Volume	2.011	06:44
Air Blank	0.000	06:45
Air Blank	0.000	06:47
Subject Sample # 2	0.190	06:47
Breath Volume	2.226	06:47
Air Blank	0.000	06:48
Control Test	0.080	06:48
Air Blank	0.000	06:49
Diagnostics Check	OK	06:49

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/15/2021 05:53:50

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.082	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.000	05:49
Breath Volume	3.707	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.000	05:52
Breath Volume	4.332	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EHRHART,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 08/19/2021 03:33:19

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.081	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAIN,ERIC,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 08/21/2021 04:12:11

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.081	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LANDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STILL, ASHLEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 08/21/2021 22:04:03

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:01	
Control Test	0.081	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LONDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE, JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: AMMANN

DATE/TIME OF TEST: 08/22/2021 02:46:37

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 08/23/2021 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MARSHALL,ALEXANDRIA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/23/2021 23:50:46

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.348	23:45
Breath Volume	1.121	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.367	23:48
Breath Volume	1.542	23:48
Air Blank	0.000	23:49
Control Test	0.077	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: DOOLY

DATE/TIME OF TEST: 08/24/2021 07:45:35

Diagnostics Check	OK	07:39
Air Blank	0.000	07:39
Control Test	0.081	07:39
Air Blank	0.000	07:40
Subject Sample # 1	0.234	07:40
Breath Volume	3.101	07:40
Air Blank	0.000	07:41
Air Blank	0.000	07:43
Subject Sample # 2	0.219	07:43
Breath Volume	3.589	07:43
Air Blank	0.000	07:44
Control Test	0.080	07:44
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MLYNARSKI, NIKLAS, E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MLYNARSKI, NIKLAS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: ESKALD

DATE/TIME OF TEST: 08/24/2021 17:08:08

Diagnostics Check	OK	17:01
Air Blank	0.000	17:01
Control Test	0.080	17:01
Air Blank	0.000	17:02
Subject Sample # 1	0.183	17:03
Breath Volume	2.765	17:03
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.179	17:06
Breath Volume	2.742	17:06
Air Blank	0.000	17:07
Control Test	0.080	17:07
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: KROUSE

DATE/TIME OF TEST: 08/24/2021 22:49:59

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.082	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.000	22:45
Breath Volume	2.714	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.000	22:48
Breath Volume	2.210	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAINES,BENJAMIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: BANG

DATE/TIME OF TEST: 08/25/2021 06:11:05

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.081	06:05
Air Blank	0.000	06:06
Subject Sample # 1	0.000	06:06
Breath Volume	3.792	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.000	06:09
Breath Volume	3.117	06:09
Air Blank	0.000	06:10
Control Test	0.080	06:10
Air Blank	0.000	06:10
Diagnostics Check	OK	06:11

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARD,JONATHAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 08/28/2021 21:05:10

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1 0.180	20:54	
Breath Volume	1.300	20:54	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 2 0.170	20:59	Volume Not Met
Breath Volume	0.414	20:59	
Air Blank	0.000	21:00	
Air Blank	0.000	21:02	
Subject Sample	# 3 0.184	21:03	
Breath Volume	1.503	21:03	
Air Blank	0.000	21:04	
Control Test	0.073	21:04	Control Outside Tolerance
Air Blank	0.000	21:05	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 08/29/2021 02:30:28

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.080	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:23	No Sample Provided
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 2	02:29	No Sample Provided
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 08/29/2021 06:44:56

Diagnostics Check	OK	06:38
Air Blank	0.000	06:38
Control Test	0.081	06:38
Air Blank	0.000	06:39
Subject Sample # 1	0.119	06:40
Breath Volume	1.617	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:42
Subject Sample # 2	0.118	06:43
Breath Volume	1.570	06:43
Air Blank	0.000	06:44
Control Test	0.080	06:44
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAUNCEY, DAYLENE, L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: DELSASSO, MICHAEL
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 08/07/2021 01:11:26

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.251	01:06
Breath Volume	2.898	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.237	01:09
Breath Volume	2.523	01:09
Air Blank	0.000	01:10
Control Test	0.076	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 09/04/2021 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAUNCEY, DAYLENE, L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: DELSASSO, MICHAEL
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/09/2021 01:10:52

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.078	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.122	01:06
Breath Volume	2.503	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.130	01:09
Breath Volume	2.972	01:09
Air Blank	0.000	01:09
Control Test	0.077	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 09/04/2021 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LATUNSKI

DATE/TIME OF TEST: 08/26/2021 22:33:40

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.078	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.000	22:29
Breath Volume	2.113	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.000	22:32
Breath Volume	2.210	22:32
Air Blank	0.000	22:32
Control Test	0.077	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 09/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 08/27/2021 02:45:06

Diagnostics Check	OK	02:44	
Air Blank		02:44	RFI Detect
Air Blank	0.000	02:45	

Data Download Date/Time: 09/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 08/27/2021 02:53:52

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.202	02:48
Breath Volume	2.085	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.185	02:52
Breath Volume	1.835	02:52
Air Blank	0.000	02:52
Control Test	0.077	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 09/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JOSHUA,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 08/31/2021 01:34:59

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample # 1	0.081	01:33	
Breath Volume	1.890	01:33	
Air Blank		01:34	RFI Detect
Air Blank		01:34	RFI Detect
Air Blank	0.000	01:34	

Data Download Date/Time: 09/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JOSHUA,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WRIGHT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 08/31/2021 01:44:22

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.078	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.079	01:39
Breath Volume	2.511	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.077	01:42
Breath Volume	1.457	01:42
Air Blank	0.000	01:43
Control Test	0.078	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 09/04/2021 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNTER, REGEN, B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES, ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 08/03/2021 00:37:56

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.000	00:33
Breath Volume	2.953	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.000	00:36
Breath Volume	3.261	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: TSIKURIS

DATE/TIME OF TEST: 08/03/2021 03:21:14

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BOURGEOIS

DATE/TIME OF TEST: 08/04/2021 03:29:57

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.260	03:25
Breath Volume	2.191	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.259	03:28
Breath Volume	2.074	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:09

SUBJECT'S LAST NAME: HARDAWAY

DATE/TIME OF TEST: 08/06/2021 08:37:17

Diagnostics Check	OK	08:30
Air Blank	0.000	08:30
Control Test	0.080	08:30
Air Blank	0.000	08:31
Subject Sample # 1	0.094	08:32
Breath Volume	2.656	08:32
Air Blank	0.000	08:33
Air Blank	0.000	08:35
Subject Sample # 2	0.093	08:35
Breath Volume	2.843	08:35
Air Blank	0.000	08:36
Control Test	0.077	08:36
Air Blank	0.000	08:37
Diagnostics Check	OK	08:37

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: TOUCHTON,STUART
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: BUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: DUPLECHAIN

DATE/TIME OF TEST: 08/06/2021 19:40:25

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.081	19:35
Air Blank	0.000	19:35
Subject Sample # 1	0.000	19:36
Breath Volume	3.289	19:36
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.000	19:39
Breath Volume	3.359	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MONZINGO

DATE/TIME OF TEST: 08/06/2021 22:24:36

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.080	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.080	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGAN,MARLON,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 08/07/2021 00:31:32

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.073	00:27
Breath Volume	1.738	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.073	00:30
Breath Volume	1.683	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FERRERIS

DATE/TIME OF TEST: 08/07/2021 03:02:27

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.000	02:58
Breath Volume	1.816	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.000	03:01
Breath Volume	2.597	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ADKINS, JAKE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520090A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 08/16/2021 01:22:30

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.120	01:18
Breath Volume	1.121	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.117	01:20
Breath Volume	1.339	01:20
Air Blank	0.000	01:21
Control Test	0.077	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 09/21/2021 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 08/03/2021 10:38:45

Diagnostics Check	OK	10:35	
Air Blank	0.000	10:36	
Control Test	0.079	10:36	
Air Blank	0.000	10:37	
Subject Sample	# 1	10:37	Subject Test Refused
Breath Volume	0.000	10:37	
Air Blank	0.000	10:37	
Control Test	0.079	10:38	
Air Blank	0.000	10:38	
Diagnostics Check	OK	10:38	

Data Download Date/Time: 08/17/2021 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,BRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PITTMAN,BRETT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TREVERROW

DATE/TIME OF TEST: 08/08/2021 23:54:07

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.078	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.079	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 08/17/2021 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 08/15/2021 00:04:01

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 08/17/2021 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER,WILLIAM,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BONTRAGER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/15/2021 00:24:50

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.079	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 08/17/2021 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 08/21/2021 03:42:04

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.160	03:37
Breath Volume	2.675	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.174	03:40
Breath Volume	1.804	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:42

Data Download Date/Time: 10/15/2021 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEAT,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SIRMANS

DATE/TIME OF TEST: 08/22/2021 02:47:38

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.048	02:43
Breath Volume	2.687	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.045	02:46
Breath Volume	1.949	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 10/15/2021 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GANUS,GLEN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 08/26/2021 01:55:30

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.000	01:50
Breath Volume	1.316	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.000	01:54
Breath Volume	2.355	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 10/15/2021 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TURNSPLENTY

DATE/TIME OF TEST: 08/28/2021 03:51:00

Diagnostics Check	OK	03:50	
Air Blank		03:50	RFI Detect
Air Blank	0.000	03:51	

Data Download Date/Time: 10/15/2021 08:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TURNSPENTY

DATE/TIME OF TEST: 08/28/2021 03:58:12

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.078	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank		03:57	RFI Detect
Air Blank	0.000	03:58	

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROSE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/05/2021 19:40:26

Diagnostics Check	OK	19:36	
Air Blank	0.000	19:37	
Control Test	0.080	19:37	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:39	Subject Test Refused
Breath Volume	0.000	19:39	
Air Blank	0.000	19:39	
Control Test	0.081	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 08/08/2021 05:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 09:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: WHITTAKER

DATE/TIME OF TEST: 08/10/2021 09:57:29

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.081	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.000	09:53
Breath Volume	3.464	09:53
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample # 2	0.000	09:56
Breath Volume	3.472	09:56
Air Blank	0.000	09:56
Control Test	0.081	09:56
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 09/10/2021 04:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINCOLN,TREY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LINCOLN,TREY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: RIDGEWAY

DATE/TIME OF TEST: 08/14/2021 07:58:52

Diagnostics Check	OK	07:50
Air Blank	0.000	07:50
Control Test	0.080	07:51
Air Blank	0.000	07:51
Subject Sample # 1	0.136	07:53
Breath Volume	1.492	07:53
Air Blank	0.000	07:53
Air Blank	0.000	07:55
Subject Sample # 2	0.135	07:56
Breath Volume	1.687	07:56
Air Blank	0.000	07:57
Control Test	0.078	07:58
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 09/10/2021 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: DEROSIA

DATE/TIME OF TEST: 08/15/2021 14:16:44

Diagnostics Check	OK	14:08
Air Blank	0.000	14:09
Control Test	0.080	14:09
Air Blank	0.000	14:10
Subject Sample # 1	0.184	14:10
Breath Volume	1.125	14:10
Air Blank	0.000	14:11
Air Blank	0.000	14:13
Subject Sample # 2	0.186	14:14
Breath Volume	1.671	14:14
Air Blank	0.000	14:15
Control Test	0.078	14:16
Air Blank	0.000	14:16
Diagnostics Check	OK	14:16

Data Download Date/Time: 09/10/2021 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 13:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 08/29/2021 15:28:42

Diagnostics Check	OK	15:21
Air Blank	0.000	15:22
Control Test	0.080	15:22
Air Blank	0.000	15:23
Subject Sample # 1	0.000	15:23
Breath Volume	4.421	15:23
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.000	15:27
Breath Volume	5.031	15:27
Air Blank	0.000	15:27
Control Test	0.080	15:28
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 09/10/2021 04:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ARROYO,WILFREDO,

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: ARROYO,WILFREDO

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 08/07/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1206004

DRY GAS EXPIRY DATE: 09/20/2022

VIOLATION CODE: DUI

ARREST TIME: 19:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SCHOTT

DATE/TIME OF TEST: 08/30/2021 21:07:02

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.000	21:02
Breath Volume	4.816	21:02
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.000	21:05
Breath Volume	4.558	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:07

Data Download Date/Time: 09/10/2021 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GLASS,DREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: LINDLEY

DATE/TIME OF TEST: 08/18/2021 20:49:19

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.143	20:44
Breath Volume	1.492	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.148	20:47
Breath Volume	1.906	20:47
Air Blank	0.000	20:48
Control Test	0.081	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 09/02/2021 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: HALL,RICHARD
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: NORWOOD

DATE/TIME OF TEST: 08/01/2021 00:46:28

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.081	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	0.192	00:42
Breath Volume	1.195	00:42	
Air Blank	0.000	00:42	
Air Blank	0.000	00:44	
Subject Sample	# 3	0.179	00:44
Breath Volume	1.554	00:44	
Air Blank	0.000	00:45	
Control Test	0.078	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ZAPATA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PEREZ MEJIA

DATE/TIME OF TEST: 08/08/2021 04:16:20

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.178	04:11
Breath Volume	2.269	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.168	04:14
Breath Volume	1.640	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LARA INCER

DATE/TIME OF TEST: 08/08/2021 04:38:11

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.191	04:33
Breath Volume	1.894	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.194	04:36
Breath Volume	2.351	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA,NELSON,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: COLANGELO,NICHOLAS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 08/15/2021 04:16:34

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.176	04:12
Breath Volume	1.503	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.189	04:14
Breath Volume	2.460	04:14
Air Blank	0.000	04:15
Control Test	0.075	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 08/20/2021 00:00:20

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.000	23:55
Breath Volume	2.656	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:58
Breath Volume	2.820	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PONDER

DATE/TIME OF TEST: 08/21/2021 02:42:00

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.070	02:37
Breath Volume	2.316	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.068	02:40
Breath Volume	2.394	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: WICKLE

DATE/TIME OF TEST: 08/21/2021 23:03:20

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.137	22:58
Breath Volume	3.105	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.142	23:01
Breath Volume	3.277	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYFORD,GREGG,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: DOZIER,TERRELL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: TORREZ GUTIERREZ

DATE/TIME OF TEST: 08/22/2021 14:55:41

Diagnostics Check	OK	14:49
Air Blank	0.000	14:50
Control Test	0.080	14:50
Air Blank	0.000	14:50
Subject Sample # 1	0.167	14:51
Breath Volume	2.035	14:51
Air Blank	0.000	14:52
Air Blank	0.000	14:53
Subject Sample # 2	0.171	14:54
Breath Volume	2.625	14:54
Air Blank	0.000	14:54
Control Test	0.079	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 07/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: STROUD

DATE/TIME OF TEST: 08/23/2021 19:47:10

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.080	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 1 0.000	19:40	Volume Not Met
Breath Volume	0.203	19:40	
Air Blank	0.000	19:40	
Air Blank	0.000	19:42	
Subject Sample	# 2 0.000	19:45	Volume Not Met
Breath Volume	0.257	19:45	
Air Blank	0.000	19:46	
Control Test	0.079	19:46	
Air Blank	0.000	19:47	
Diagnostics Check	OK	19:47	

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RICHARD,D
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIONEL,CONYERS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 08/30/2021 01:55:17

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample	# 2	0.280	01:50
Breath Volume	1.605	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 3	0.284	01:53
Breath Volume	2.378	01:53	
Air Blank	0.000	01:54	
Control Test	0.078	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: CARTY

DATE/TIME OF TEST: 08/31/2021 21:25:32

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.081	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.097	21:21
Breath Volume	2.320	21:21
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.095	21:24
Breath Volume	1.863	21:24
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 09/01/2021 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN, TY, C
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: VAN HEERDEN, TY
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BUTTS

DATE/TIME OF TEST: 08/04/2021 01:47:24

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.158	01:43
Breath Volume	2.460	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.151	01:45
Breath Volume	1.988	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/03/2021 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN, TY, C
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: VAN HEERDEN, TY
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 08/23/2021 00:33:21

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:24	
Air Blank	0.000	00:24	
Subject Sample	# 1 0.195	00:27	Volume Not Met
Breath Volume	0.085	00:27	
Air Blank	0.000	00:28	
Air Blank	0.000	00:30	
Subject Sample	# 2	00:31	Subject Test Refused
Breath Volume	0.320	00:31	
Air Blank	0.000	00:32	
Control Test	0.078	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 09/03/2021 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HEERDEN, TY, C
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: POLAND, ADRIAN
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 08/26/2021 01:10:31

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.079	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 09/03/2021 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERBER,OLIVER,B
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GERBER,OLIVER
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 08/28/2021 18:59:40

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:54
Subject Sample # 1	0.127	18:55
Breath Volume	1.589	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.128	18:58
Breath Volume	3.582	18:58
Air Blank	0.000	18:58
Control Test	0.078	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 09/03/2021 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARSONS,JOHN,M
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: TEDJESKE,JAMES
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 08/03/2021 22:38:32

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.389	22:34
Breath Volume	3.308	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.371	22:36
Breath Volume	2.363	22:36
Air Blank	0.000	22:37
Control Test	0.079	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 10/03/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,BRADLEY,A
BREATH TEST OPERATOR AGENCY: Hamilton County SO
ARRESTING OFFICER: HERRERA,JASON
ARRESTING OFFICER AGENCY: Hamilton County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/09/2021 04:19:37

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.199	04:15
Breath Volume	2.042	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.188	04:18
Breath Volume	1.765	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 10/03/2021 17:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASTOR,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 08/18/2021 16:46:12

Diagnostics Check	OK	16:40
Air Blank	0.000	16:40
Control Test	0.080	16:40
Air Blank	0.000	16:41
Subject Sample # 1	0.159	16:41
Breath Volume	2.605	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.154	16:44
Breath Volume	1.277	16:44
Air Blank	0.000	16:45
Control Test	0.079	16:45
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 10/03/2021 17:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BRENNEMAN

DATE/TIME OF TEST: 08/13/2021 03:52:10

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.187	03:47
Breath Volume	2.921	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.194	03:50
Breath Volume	3.726	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/10/2021 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: LACZA

DATE/TIME OF TEST: 08/13/2021 04:53:18

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.082	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.213	04:45
Breath Volume	1.468	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.186	04:48
Breath Volume	1.914	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 3	0.215	04:51
Breath Volume	3.839	04:51
Air Blank	0.000	04:52
Control Test	0.081	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 09/10/2021 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ANTOLINO

DATE/TIME OF TEST: 08/14/2021 02:46:48

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.227	02:42
Breath Volume	2.898	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.222	02:45
Breath Volume	1.925	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 09/10/2021 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: CASAUBON

DATE/TIME OF TEST: 08/14/2021 03:00:59

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.127	02:56
Breath Volume	1.546	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.128	02:59
Breath Volume	1.578	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 09/10/2021 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 08/14/2021 04:04:10

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.215	03:59
Breath Volume	3.453	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.201	04:02
Breath Volume	1.421	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 09/10/2021 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: TAGGART

DATE/TIME OF TEST: 08/15/2021 00:18:54

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1 0.112	00:11	Volume Not Met
Breath Volume	1.023	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2 0.109	00:17	Volume Not Met
Breath Volume	0.800	00:17	
Air Blank	0.000	00:17	
Control Test	0.080	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 09/10/2021 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ORIHUELA

DATE/TIME OF TEST: 08/15/2021 03:49:33

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.081	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.075	03:40	Volume Not Met
Breath Volume	0.281	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 2 0.096	03:43	
Breath Volume	2.195	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 3 0.089	03:47	
Breath Volume	1.394	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 09/10/2021 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: BORGES JR

DATE/TIME OF TEST: 08/15/2021 04:46:54

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.000	04:42
Breath Volume	3.000	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.000	04:45
Breath Volume	1.515	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: GADDIS

DATE/TIME OF TEST: 08/15/2021 05:23:28

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.263	05:15
Breath Volume	2.105	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.232	05:18
Breath Volume	1.300	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 3	0.253	05:21
Breath Volume	2.179	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: WALLNER

DATE/TIME OF TEST: 08/18/2021 21:07:52

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.082	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.109	21:03
Breath Volume	3.230	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.106	21:06
Breath Volume	2.468	21:06
Air Blank	0.000	21:06
Control Test	0.081	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATRON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PEREZ FERMIN

DATE/TIME OF TEST: 08/21/2021 00:53:29

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.118	00:49
Breath Volume	2.757	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.122	00:51
Breath Volume	1.906	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BRIM

DATE/TIME OF TEST: 08/21/2021 04:14:57

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.189	04:10
Breath Volume	1.324	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:12
Subject Sample # 2	0.187	04:13
Breath Volume	1.359	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 08/21/2021 07:21:40

Diagnostics Check	OK	07:14
Air Blank	0.000	07:15
Control Test	0.081	07:15
Air Blank	0.000	07:15
Subject Sample # 1	0.096	07:16
Breath Volume	3.136	07:16
Air Blank	0.000	07:17
Air Blank	0.000	07:19
Subject Sample # 2	0.096	07:20
Breath Volume	3.468	07:20
Air Blank	0.000	07:20
Control Test	0.080	07:21
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: TOSADO

DATE/TIME OF TEST: 08/22/2021 01:06:29

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.166	01:01
Breath Volume	3.351	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.171	01:04
Breath Volume	3.222	01:04
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RAINEY JR.

DATE/TIME OF TEST: 08/22/2021 02:41:58

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.057	02:37
Breath Volume	1.183	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.052	02:40
Breath Volume	1.152	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: SALAZAR-MORALES

DATE/TIME OF TEST: 08/22/2021 05:31:54

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.081	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.144	05:27
Breath Volume	1.125	05:27
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.155	05:30
Breath Volume	3.000	05:30
Air Blank	0.000	05:30
Control Test	0.081	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: CARTER JR

DATE/TIME OF TEST: 08/23/2021 00:10:45

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.082	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.199	00:05
Breath Volume	2.257	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.194	00:09
Breath Volume	1.562	00:09
Air Blank	0.000	00:09
Control Test	0.081	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULAIMAN,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CHAPPELL

DATE/TIME OF TEST: 08/23/2021 02:34:17

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.262	02:29
Breath Volume	1.601	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:31
Subject Sample # 2	0.273	02:32
Breath Volume	2.167	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: SEL-MAQUIN

DATE/TIME OF TEST: 08/23/2021 05:26:15

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.214	05:21
Breath Volume	1.742	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.204	05:24
Breath Volume	1.382	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 09/10/2021 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HEATHERLEY

DATE/TIME OF TEST: 08/24/2021 00:56:25

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.082	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.000	00:51
Breath Volume	3.589	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.000	00:55
Breath Volume	3.421	00:55
Air Blank	0.000	00:55
Control Test	0.082	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND, M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 08/24/2021 02:33:52

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.197	02:28
Breath Volume	1.878	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.207	02:32
Breath Volume	2.328	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/24/2021 22:59:22

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.145	22:53
Breath Volume	1.585	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.155	22:57
Breath Volume	1.960	22:57
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LOPER

DATE/TIME OF TEST: 08/25/2021 03:56:33

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.157	03:51
Breath Volume	3.285	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.151	03:54
Breath Volume	3.578	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GREENWELL

DATE/TIME OF TEST: 08/26/2021 01:50:34

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.082	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.287	01:45
Breath Volume	3.453	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.286	01:48
Breath Volume	4.820	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/26/2021 02:23:38

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.162	02:18
Breath Volume	2.230	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.163	02:22
Breath Volume	3.042	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: BORGES JR.

DATE/TIME OF TEST: 08/26/2021 19:04:56

Diagnostics Check	OK	18:58
Air Blank	0.000	18:58
Control Test	0.081	18:58
Air Blank	0.000	18:59
Subject Sample # 1	0.000	19:00
Breath Volume	3.070	19:00
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.000	19:03
Breath Volume	3.121	19:03
Air Blank	0.000	19:04
Control Test	0.081	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PAREJO

DATE/TIME OF TEST: 08/27/2021 01:35:25

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.218	01:30
Breath Volume	1.820	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.221	01:33
Breath Volume	1.351	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: ABREU-CAMILO

DATE/TIME OF TEST: 08/27/2021 03:26:57

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.158	03:22
Breath Volume	2.554	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.153	03:25
Breath Volume	2.496	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: AGNETA

DATE/TIME OF TEST: 08/28/2021 00:25:57

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.203	00:20
Breath Volume	1.953	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.207	00:24
Breath Volume	2.667	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BERG,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: BUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: MARTINEZ DUPREY

DATE/TIME OF TEST: 08/28/2021 18:34:03

Diagnostics Check	OK	18:27
Air Blank	0.000	18:28
Control Test	0.081	18:28
Air Blank	0.000	18:28
Subject Sample # 1	0.011	18:29
Breath Volume	3.343	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.011	18:32
Breath Volume	3.421	18:32
Air Blank	0.000	18:33
Control Test	0.081	18:33
Air Blank	0.000	18:33
Diagnostics Check	OK	18:34

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 08/29/2021 01:58:28

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.140	01:53
Breath Volume	1.753	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.146	01:56
Breath Volume	2.335	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 08/29/2021 02:28:51

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.202	02:21
Breath Volume	1.824	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.223	02:24
Breath Volume	1.675	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 3	0.222	02:27
Breath Volume	1.761	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 09/10/2021 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: WARREN JR

DATE/TIME OF TEST: 08/29/2021 04:39:33

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.000	04:35
Breath Volume	2.457	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.000	04:38
Breath Volume	1.777	04:38
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 08/29/2021 23:57:36

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1 0.194	23:45	Slope Not Level
Breath Volume	2.003	23:45	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 2 0.207	23:51	
Breath Volume	1.570	23:51	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 3	23:55	Slope Not Met
Breath Volume	1.445	23:55	
Air Blank	0.000	23:56	
Control Test	0.080	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: PINGOR

DATE/TIME OF TEST: 08/30/2021 00:15:31

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1 0.181	00:07	Volume Not Met
Breath Volume	0.718	00:07	
Air Blank	0.000	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 2 0.175	00:10	
Breath Volume	1.402	00:10	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample	# 3 0.186	00:13	
Breath Volume	1.308	00:13	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DRAKE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: HAWKES JR

DATE/TIME OF TEST: 08/30/2021 14:54:49

Diagnostics Check	OK	14:47
Air Blank	0.000	14:48
Control Test	0.082	14:48
Air Blank	0.000	14:49
Subject Sample # 1	0.265	14:50
Breath Volume	1.531	14:50
Air Blank	0.000	14:50
Air Blank	0.000	14:52
Subject Sample # 2	0.270	14:53
Breath Volume	1.730	14:53
Air Blank	0.000	14:53
Control Test	0.081	14:54
Air Blank	0.000	14:54
Diagnostics Check	OK	14:54

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: HUMERICKHOUSE

DATE/TIME OF TEST: 08/23/2021 12:28:23

Diagnostics Check	OK	12:21
Air Blank	0.000	12:21
Control Test	0.080	12:22
Air Blank	0.000	12:22
Subject Sample # 1	0.000	12:23
Breath Volume	1.261	12:23
Air Blank	0.000	12:24
Air Blank	0.000	12:26
Subject Sample # 2	0.000	12:27
Breath Volume	1.207	12:27
Air Blank	0.000	12:27
Control Test	0.077	12:27
Air Blank	0.000	12:28
Diagnostics Check	OK	12:28

Data Download Date/Time: 09/22/2021 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALZA, PAOLOGENE, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BALZA, PAOLOGENE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 08/01/2021 08:10:29

Diagnostics Check	OK	08:04
Air Blank	0.000	08:04
Control Test	0.078	08:05
Air Blank	0.000	08:05
Subject Sample # 1	0.009	08:06
Breath Volume	1.839	08:06
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.008	08:09
Breath Volume	2.792	08:09
Air Blank	0.000	08:09
Control Test	0.078	08:09
Air Blank	0.000	08:10
Diagnostics Check	OK	08:10

Data Download Date/Time: 08/12/2021 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: NICOLINI

DATE/TIME OF TEST: 08/04/2021 01:08:03

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.034	01:03
Breath Volume	1.363	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.034	01:06
Breath Volume	1.464	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:08

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 08/08/2021 01:48:35

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.081	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.080	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIBBEY,BRADLEY,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SCHWARZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SPOONER

DATE/TIME OF TEST: 08/08/2021 18:01:02

Diagnostics Check	OK	17:54
Air Blank	0.000	17:54
Control Test	0.081	17:54
Air Blank	0.000	17:55
Subject Sample # 1	0.125	17:56
Breath Volume	1.675	17:56
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample # 2	0.126	17:59
Breath Volume	1.667	17:59
Air Blank	0.000	18:00
Control Test	0.079	18:00
Air Blank	0.000	18:00
Diagnostics Check	OK	18:01

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/09/2021 02:05:55

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:01
Breath Volume	2.273	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.000	02:04
Breath Volume	2.207	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYATT,JESSICA,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 08/10/2021 19:20:14

Diagnostics Check	OK	19:08	
Air Blank	0.000	19:08	
Control Test	0.079	19:09	
Air Blank	0.000	19:09	
Subject Sample	# 1 0.075	19:12	Volume Not Met
Breath Volume	0.328	19:12	
Air Blank	0.000	19:13	
Air Blank	0.000	19:15	
Subject Sample	# 2 0.062	19:18	Volume Not Met
Breath Volume	0.472	19:18	
Air Blank	0.000	19:19	
Control Test	0.078	19:19	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAMJIT,E
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: THOGODE

DATE/TIME OF TEST: 08/13/2021 01:29:40

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.080	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HAUENSTEIN,C
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 08/13/2021 22:40:52

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.273	22:35
Breath Volume	1.667	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.281	22:38
Breath Volume	1.652	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: AMAR,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 08/14/2021 16:05:55

Diagnostics Check	OK	15:59
Air Blank	0.000	15:59
Control Test	0.079	15:59
Air Blank	0.000	16:00
Subject Sample # 1	0.202	16:01
Breath Volume	3.035	16:01
Air Blank	0.000	16:01
Air Blank	0.000	16:03
Subject Sample # 2	0.202	16:04
Breath Volume	2.765	16:04
Air Blank	0.000	16:04
Control Test	0.079	16:05
Air Blank	0.000	16:05
Diagnostics Check	OK	16:05

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYATT,JESSICA,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,T
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: BUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 08/14/2021 18:03:06

Diagnostics Check	OK	18:00	
Air Blank	0.000	18:00	
Control Test	0.080	18:00	
Air Blank	0.000	18:01	
Subject Sample	# 1	18:01	Subject Test Refused
Breath Volume	0.000	18:01	
Air Blank	0.000	18:02	
Control Test	0.080	18:02	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 08/15/2021 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: BERENYI

DATE/TIME OF TEST: 08/20/2021 21:56:20

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.080	21:54	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Control Test	0.081	21:55	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 09/16/2021 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASCITELLI,V,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:34

SUBJECT'S LAST NAME: SCHULTZ

DATE/TIME OF TEST: 08/22/2021 12:01:26

Diagnostics Check	OK	11:54
Air Blank	0.000	11:55
Control Test	0.081	11:55
Air Blank	0.000	11:56
Subject Sample # 1	0.174	11:56
Breath Volume	1.906	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.167	11:59
Breath Volume	2.066	11:59
Air Blank	0.000	12:00
Control Test	0.079	12:00
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 09/16/2021 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PODRAZA,BRIAN,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WHITAKER,GABRIEL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 08/28/2021 03:25:13

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.081	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: PORRO

DATE/TIME OF TEST: 08/29/2021 19:59:48

Diagnostics Check	OK	19:53
Air Blank	0.000	19:53
Control Test	0.080	19:54
Air Blank	0.000	19:54
Subject Sample # 1	0.224	19:55
Breath Volume	1.875	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:57
Subject Sample # 2	0.224	19:58
Breath Volume	1.980	19:58
Air Blank	0.000	19:58
Control Test	0.077	19:59
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: REYNOLDS,DONOVAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 08/06/2021 23:41:17

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.078	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Control Test	0.080	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 08/10/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: JACKSON,NICHOLAS
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PROVENCHER

DATE/TIME OF TEST: 08/07/2021 01:59:40

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.074	01:55	Volume Not Met
Breath Volume	0.222	01:55	
Air Blank	0.000	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 2	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.073	01:59	Control Outside Tolerance
Air Blank	0.000	01:59	

Data Download Date/Time: 08/10/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: KUFAHL

DATE/TIME OF TEST: 08/09/2021 00:51:35

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.078	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.201	00:46
Breath Volume	1.457	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.216	00:49
Breath Volume	2.152	00:49
Air Blank	0.000	00:50
Control Test	0.077	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/10/2021 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHALL,JALEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: DECOTEAU,TIMOTHY
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: NARVAEZ MALDONADO

DATE/TIME OF TEST: 08/09/2021 02:38:20

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.079	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 08/10/2021 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,CODY,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: LEONARDO

DATE/TIME OF TEST: 08/11/2021 04:07:24

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.078	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1 0.140	04:00	Volume Not Met
Breath Volume	0.246	04:00	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2 0.137	04:05	Volume Not Met
Breath Volume	0.242	04:05	
Air Blank	0.000	04:06	
Control Test	0.077	04:06	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,CODY,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CRIPE

DATE/TIME OF TEST: 08/14/2021 22:33:10

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.079	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:32	
Control Test	0.079	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FLORENCE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 08/16/2021 22:19:37

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.080	22:17	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.079	22:18	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GUINANE,JOHN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 08/16/2021 23:00:11

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:57	
Control Test	0.079	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:59	
Control Test	0.079	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DRAWDY

DATE/TIME OF TEST: 08/20/2021 04:35:52

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.000	04:31
Breath Volume	3.050	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.000	04:34
Breath Volume	2.894	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/21/2021 21:44:26

Diagnostics Check	OK	21:30	
Air Blank	0.000	21:31	
Control Test	0.078	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1 0.177	21:35	Volume Not Met
Breath Volume	0.390	21:35	
Air Blank	0.000	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 2 0.183	21:38	
Breath Volume	1.546	21:38	
Air Blank	0.000	21:39	
Air Blank	0.000	21:41	
Subject Sample	# 3 0.182	21:42	
Breath Volume	2.050	21:42	
Air Blank	0.000	21:43	
Control Test	0.078	21:43	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 08/24/2021 03:02:30

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.077	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.149	02:57
Breath Volume	1.367	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.143	03:00
Breath Volume	1.347	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CANNA,DERRICK
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 08/27/2021 01:41:55

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.078	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1 0.169	01:36	Volume Not Met
Breath Volume	0.207	01:36	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 2	01:40	Subject Test Refused
Breath Volume	0.031	01:40	
Air Blank	0.000	01:40	
Control Test	0.077	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,CODY,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNOR
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 08/27/2021 22:19:17

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.194	22:14
Breath Volume	2.644	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:16
Subject Sample # 2	0.186	22:17
Breath Volume	1.664	22:17
Air Blank	0.000	22:18
Control Test	0.078	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: EVERS

DATE/TIME OF TEST: 08/28/2021 21:57:20

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
		21:56	Sequence Aborted
Air Blank	0.000	21:57	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NELSON,GARRETT
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: EVERS

DATE/TIME OF TEST: 08/28/2021 22:22:31

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.104	22:17
Breath Volume	2.324	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:19
Subject Sample # 2	0.102	22:20
Breath Volume	2.429	22:20
Air Blank	0.000	22:21
Control Test	0.078	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDLE,PAUL,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SANCHEZ AGUIRE

DATE/TIME OF TEST: 08/29/2021 01:47:54

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.212	01:43
Breath Volume	2.332	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.212	01:46
Breath Volume	2.253	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: VELEZ,A
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CALKA

DATE/TIME OF TEST: 08/29/2021 23:34:31

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.078	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DADDARIO,ASHLEY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TITUS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 08/01/2021 20:14:32

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:12	
Control Test	0.078	20:12	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:13	Subject Test Refused
Breath Volume	0.000	20:13	
Air Blank	0.000	20:13	
Control Test	0.077	20:13	
Air Blank	0.000	20:14	
Diagnostics Check	OK	20:14	

Data Download Date/Time: 09/24/2021 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,NICHOLAS,L
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ,CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DELGADO II

DATE/TIME OF TEST: 08/01/2021 03:50:17

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 09/01/2021 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BULTRON,CARLOS
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROBISON

DATE/TIME OF TEST: 08/02/2021 01:48:39

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.216	01:43
Breath Volume	2.023	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.207	01:47
Breath Volume	1.316	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEUSER,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: PEREZ,CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: FARADAY

DATE/TIME OF TEST: 08/07/2021 03:49:07

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEUSER,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: PARISH,CLINTON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MELENDEZ ZAYAS

DATE/TIME OF TEST: 08/07/2021 04:12:41

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.181	04:07
Breath Volume	3.750	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.184	04:11
Breath Volume	2.445	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODEN,ANDREW,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 08/07/2021 05:56:24

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:48	
Control Test	0.079	05:48	
Air Blank	0.000	05:49	
Subject Sample	# 1	05:52	No Sample Provided
Breath Volume	0.000	05:52	
Air Blank	0.000	05:52	
Air Blank	0.000	05:54	
Subject Sample	# 2	05:55	Subject Test Refused
Breath Volume	0.000	05:55	
Air Blank	0.000	05:55	
Control Test	0.079	05:55	
Air Blank	0.000	05:56	
Diagnostics Check	OK	05:56	

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,JOSHUA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: GONZALEZ,L
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JAVIER

DATE/TIME OF TEST: 08/08/2021 02:07:13

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1	02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 2	02:05	No Sample Provided
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.079	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: GONZALEZ,L
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JAVIER CONTRERAS

DATE/TIME OF TEST: 08/08/2021 02:13:27

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: ROMAN,ESTRELLA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 08/10/2021 02:54:08

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.199	02:49
Breath Volume	2.992	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.203	02:52
Breath Volume	2.324	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODEN,ANDREW,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: DINGMAN

DATE/TIME OF TEST: 08/11/2021 01:32:43

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.168	01:28
Breath Volume	2.128	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.171	01:31
Breath Volume	3.320	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODEN,ANDREW,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: RODEN,ANDREW
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: BONILLA PADILLA

DATE/TIME OF TEST: 08/13/2021 03:11:54

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 09/01/2021 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:22

SUBJECT'S LAST NAME: REEDER

DATE/TIME OF TEST: 08/13/2021 11:49:03

Diagnostics Check	OK	11:42
Air Blank	0.000	11:43
Control Test	0.079	11:43
Air Blank	0.000	11:44
Subject Sample # 1	0.000	11:44
Breath Volume	2.003	11:44
Air Blank	0.000	11:45
Air Blank	0.000	11:47
Subject Sample # 2	0.000	11:47
Breath Volume	1.570	11:47
Air Blank	0.000	11:48
Control Test	0.080	11:48
Air Blank	0.000	11:48
Diagnostics Check	OK	11:49

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNICK,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BANNICK,COURTNEY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LEMON

DATE/TIME OF TEST: 08/15/2021 00:11:14

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.079	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.079	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUDERS,S,A
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: BAUDERS,S
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GARCES

DATE/TIME OF TEST: 08/15/2021 02:20:49

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.080	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.105	02:13	Volume Not Met
Breath Volume	0.386	02:13	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2 0.115	02:19	Volume Not Met
Breath Volume	0.984	02:19	
Air Blank	0.000	02:19	
Control Test	0.080	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUDERS,S,A
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: BAUDERS,S
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GARCES

DATE/TIME OF TEST: 08/15/2021 03:57:36

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 2	03:56	No Sample Provided
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, RONALD,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: RODRIGUEZ, RONALD
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 08/15/2021 16:18:35

Diagnostics Check	OK	16:11
Air Blank	0.000	16:11
Control Test	0.080	16:12
Air Blank	0.000	16:12
Subject Sample # 1	0.279	16:13
Breath Volume	1.472	16:13
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.268	16:17
Breath Volume	1.113	16:17
Air Blank	0.000	16:17
Control Test	0.080	16:17
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY, WAYNE, G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: ABDELMESSIH, ADAM
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: HALLNAN

DATE/TIME OF TEST: 08/16/2021 06:08:59

Diagnostics Check	OK	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:03
Subject Sample # 1	0.247	06:04
Breath Volume	2.617	06:04
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.248	06:07
Breath Volume	1.519	06:07
Air Blank	0.000	06:08
Control Test	0.079	06:08
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUEGER, ELIZABETH, A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: PEREZ, CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 08/16/2021 23:58:14

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.118	23:53
Breath Volume	1.601	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.114	23:56
Breath Volume	1.628	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 08/17/2021 02:17:31

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.037	02:13
Breath Volume	1.816	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.036	02:16
Breath Volume	2.191	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBI,UZIEL,Y
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BERRIDGE,KATIE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: ENOS

DATE/TIME OF TEST: 08/18/2021 20:53:22

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.079	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:52	
Control Test	0.079	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: DANIELS,DEVIN
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: CITERA

DATE/TIME OF TEST: 08/20/2021 16:39:59

Diagnostics Check	OK	16:31
Air Blank	0.000	16:32
Control Test	0.079	16:32
Air Blank	0.000	16:33
Subject Sample # 1	0.210	16:34
Breath Volume	2.464	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.200	16:38
Breath Volume	1.449	16:38
Air Blank	0.000	16:39
Control Test	0.080	16:39
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE, KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CONTI VIANA

DATE/TIME OF TEST: 08/21/2021 04:37:18

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Air Blank	0.000	04:32	
Subject Sample	# 2	04:35	No Sample Provided
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PARDO,JUAN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: HERNANDEZ-SERVIN

DATE/TIME OF TEST: 08/21/2021 07:58:06

Diagnostics Check	OK	07:49
Air Blank	0.000	07:50
Control Test	0.080	07:50
Air Blank	0.000	07:51
Subject Sample # 1	0.164	07:53
Breath Volume	2.230	07:53
Air Blank	0.000	07:54
Air Blank	0.000	07:56
Subject Sample # 2	0.149	07:56
Breath Volume	1.320	07:56
Air Blank	0.000	07:57
Control Test	0.079	07:57
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 09/01/2021 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUP,STEVE,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LOPEZ,RICARDO
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 08/21/2021 08:34:04

Diagnostics Check	OK	08:25
Air Blank	0.000	08:25
Control Test	0.079	08:26
Air Blank	0.000	08:26
Subject Sample # 1	0.208	08:29
Breath Volume	1.687	08:29
Air Blank	0.000	08:29
Air Blank	0.000	08:31
Subject Sample # 2	0.206	08:32
Breath Volume	2.722	08:32
Air Blank	0.000	08:33
Control Test	0.079	08:33
Air Blank	0.000	08:33
Diagnostics Check	OK	08:34

Data Download Date/Time: 09/01/2021 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE, KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BODDIE

DATE/TIME OF TEST: 08/22/2021 04:02:00

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.079	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 09/01/2021 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: BUEHLER,ANTHONY
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 07/26/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BATTERBEE

DATE/TIME OF TEST: 08/25/2021 03:36:03

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 09/01/2021 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLDHAM,TAYLOR,D
BREATH TEST OPERATOR AGENCY: Mascotte PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: ZOSS

DATE/TIME OF TEST: 08/28/2021 16:05:26

Diagnostics Check	OK	16:03	
Air Blank	0.000	16:03	
Control Test	0.079	16:04	
Air Blank	0.000	16:04	
Subject Sample	# 1	16:04	Sequence Aborted
Breath Volume	0.000	16:04	
Air Blank	0.000	16:05	

Data Download Date/Time: 09/01/2021 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/29/2021 18:45:12

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.079	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.130	18:39
Breath Volume	1.312	18:39
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.137	18:43
Breath Volume	1.843	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 09/01/2021 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCLAIN, DAVID, R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: DAGASTINO, STEFANO

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 20:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/01/2021 21:16:16

Diagnostics Check	OK	21:11	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:14	Subject Test Refused
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 08/30/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LENFANT,BENJAMIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: DELRIO

DATE/TIME OF TEST: 08/08/2021 20:54:51

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.081	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:53	
Control Test	0.081	20:54	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 08/30/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: OSWALT

DATE/TIME OF TEST: 08/11/2021 00:51:52

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 08/30/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCLAIN,DAVID,R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: MCCLAIN,DAVID

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:48

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 08/15/2021 04:52:47

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:39	
Control Test	0.080	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1 0.146	04:43	Volume Not Met
Breath Volume	0.421	04:43	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample	# 2 0.175	04:47	
Breath Volume	1.125	04:47	
Air Blank	0.000	04:48	
Air Blank	0.000	04:50	
Subject Sample	# 3 0.162	04:51	
Breath Volume	1.171	04:51	
Air Blank	0.000	04:51	
Control Test	0.078	04:52	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 08/30/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DAGASTINO, STEFANO
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: WILLEY

DATE/TIME OF TEST: 08/24/2021 00:01:09

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.300	23:56
Breath Volume	1.691	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.290	23:59
Breath Volume	1.699	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 08/30/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DAGASTINO, STEFANO
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 30219080A3
DRY GAS EXPIRY DATE: 01/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/25/2021 03:19:32

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.150	03:14
Breath Volume	1.203	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:16
Subject Sample # 2	0.164	03:18
Breath Volume	1.527	03:18
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 08/30/2021 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCLAIN, DAVID, R

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: DAGASTINO, STEFANO

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 30219080A3

DRY GAS EXPIRY DATE: 01/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: TEATE

DATE/TIME OF TEST: 08/28/2021 20:30:45

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.081	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:29	Subject Test Refused
Breath Volume	0.000	20:29	
Air Blank	0.000	20:29	
Control Test	0.080	20:30	
Air Blank	0.000	20:30	
Diagnostics Check	OK	20:30	

Data Download Date/Time: 08/30/2021 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,LAYNE,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: CESPEDES,ROXANA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 08/01/2021 03:15:07

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.083	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.080	03:10
Breath Volume	1.605	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.073	03:13
Breath Volume	1.730	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 08/03/2021 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,TRAVON,X
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BOYD,TRAVON
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 08/02/2021 03:49:58

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.082	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample	# 2	03:48	Volume Not Met
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.082	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 08/03/2021 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MOORE,M
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 08/07/2021 01:56:11

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.000	01:52
Breath Volume	1.906	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.000	01:54
Breath Volume	2.066	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 09/23/2021 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRIS
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 08/08/2021 02:20:09

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.190	02:15
Breath Volume	1.718	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.187	02:18
Breath Volume	1.246	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 09/23/2021 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRIS,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 08/10/2021 18:09:31

Diagnostics Check	OK	18:03
Air Blank	0.000	18:04
Control Test	0.081	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.000	18:05
Breath Volume	2.046	18:05
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.000	18:08
Breath Volume	2.121	18:08
Air Blank	0.000	18:08
Control Test	0.080	18:08
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 09/23/2021 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: SHAFRANEK

DATE/TIME OF TEST: 08/11/2021 19:34:52

Diagnostics Check	OK	19:28
Air Blank	0.000	19:28
Control Test	0.079	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.126	19:30
Breath Volume	1.390	19:30
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.129	19:33
Breath Volume	2.000	19:33
Air Blank	0.000	19:33
Control Test	0.078	19:34
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 08/20/2021 00:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/12/2021 18:23:18

Diagnostics Check	OK	18:16
Air Blank	0.000	18:16
Control Test	0.079	18:17
Air Blank	0.000	18:17
Subject Sample # 1	0.200	18:18
Breath Volume	1.484	18:18
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample # 2	0.220	18:21
Breath Volume	2.355	18:21
Air Blank	0.000	18:22
Control Test	0.078	18:22
Air Blank	0.000	18:23
Diagnostics Check	OK	18:23

Data Download Date/Time: 08/20/2021 00:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ISHAM,MATTHEW,H
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ISHAM,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 08/05/2021 00:48:38

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.203	00:44
Breath Volume	2.269	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.198	00:46
Breath Volume	2.253	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 08/19/2021 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: WHITSON

DATE/TIME OF TEST: 08/07/2021 08:14:31

Diagnostics Check	OK	08:07
Air Blank	0.000	08:07
Control Test	0.080	08:08
Air Blank	0.000	08:08
Subject Sample # 1	0.075	08:09
Breath Volume	1.839	08:09
Air Blank	0.000	08:10
Air Blank	0.000	08:12
Subject Sample # 2	0.068	08:12
Breath Volume	1.414	08:12
Air Blank	0.000	08:13
Control Test	0.080	08:13
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time: 08/19/2021 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ELAINE

DATE/TIME OF TEST: 08/13/2021 02:14:25

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.169	02:09
Breath Volume	1.667	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.160	02:12
Breath Volume	1.625	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 08/19/2021 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BIERMAN

DATE/TIME OF TEST: 08/13/2021 04:40:34

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.312	04:35
Breath Volume	3.222	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.315	04:38
Breath Volume	2.597	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 08/19/2021 22:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: RAMIREZ JACINTO

DATE/TIME OF TEST: 08/14/2021 17:57:04

Diagnostics Check	OK	17:50
Air Blank	0.000	17:50
Control Test	0.080	17:50
Air Blank	0.000	17:51
Subject Sample # 1	0.074	17:52
Breath Volume	2.757	17:52
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.073	17:55
Breath Volume	2.507	17:55
Air Blank	0.000	17:56
Control Test	0.079	17:56
Air Blank	0.000	17:56
Diagnostics Check	OK	17:57

Data Download Date/Time: 08/19/2021 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 08/14/2021 23:36:13

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.079	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 08/19/2021 22:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MORRIS,KEVIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: PEREZ VELASQUEZ

DATE/TIME OF TEST: 08/21/2021 17:48:41

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.079	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.117	17:44
Breath Volume	1.570	17:44
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.122	17:47
Breath Volume	1.597	17:47
Air Blank	0.000	17:47
Control Test	0.078	17:48
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

Data Download Date/Time: 09/08/2021 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,DAVID,E
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: JACOBSON,DAVID
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: WINSKI

DATE/TIME OF TEST: 08/22/2021 16:43:17

Diagnostics Check	OK	16:40	
Air Blank	0.000	16:40	
Control Test	0.079	16:41	
Air Blank	0.000	16:41	
Subject Sample	# 1	16:41	Subject Test Refused
Breath Volume	0.000	16:41	
Air Blank	0.000	16:42	
Control Test	0.079	16:42	
Air Blank	0.000	16:43	
Diagnostics Check	OK	16:43	

Data Download Date/Time: 09/09/2021 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL, MICHAEL, J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO, GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: VIAMONTES

DATE/TIME OF TEST: 08/25/2021 21:36:11

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.000	21:31
Breath Volume	3.847	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:34
Breath Volume	3.324	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 09/09/2021 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: LUTTON

DATE/TIME OF TEST: 08/28/2021 16:30:11

Diagnostics Check	OK	16:24
Air Blank	0.000	16:24
Control Test	0.080	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.000	16:25
Breath Volume	1.136	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.000	16:28
Breath Volume	1.792	16:28
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 09/09/2021 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,T
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 08/29/2021 04:36:04

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.206	04:31
Breath Volume	1.992	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.188	04:34
Breath Volume	2.304	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:36

Data Download Date/Time: 09/09/2021 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,CHARLES,S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DURHAM,KARI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: ZIGLAR

DATE/TIME OF TEST: 08/01/2021 15:32:15

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.081	15:27
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:28
Breath Volume	1.195	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.000	15:30
Breath Volume	3.382	15:30
Air Blank	0.000	15:31
Control Test	0.081	15:31
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 08/06/2021 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/03/2021 22:39:30

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.000	22:35
Breath Volume	1.851	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.000	22:38
Breath Volume	1.511	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 08/06/2021 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 08/05/2021 02:36:07

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.035	02:31
Breath Volume	3.199	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.035	02:34
Breath Volume	1.796	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 08/06/2021 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOWE,ALEXANDER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 08/05/2021 19:21:52

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.081	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.000	19:17
Breath Volume	1.476	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.000	19:20
Breath Volume	1.656	19:20
Air Blank	0.000	19:20
Control Test	0.081	19:21
Air Blank	0.000	19:21
Diagnostics Check	OK	19:21

Data Download Date/Time: 08/06/2021 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: LECREUX

DATE/TIME OF TEST: 08/06/2021 22:37:24

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.364	22:29
Breath Volume	3.066	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.342	22:32
Breath Volume	1.402	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 3	0.327	22:35
Breath Volume	2.144	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOETHEN, K
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GARCIA FLORES

DATE/TIME OF TEST: 08/07/2021 04:26:38

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.107	04:22
Breath Volume	3.523	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.108	04:25
Breath Volume	3.664	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: LABRADA LEIVA

DATE/TIME OF TEST: 08/10/2021 04:36:11

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.000	04:31
Breath Volume	2.429	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.000	04:34
Breath Volume	2.933	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHRABOT,NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: SATURNE

DATE/TIME OF TEST: 08/10/2021 06:36:12

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.081	06:29
Air Blank	0.000	06:30
Subject Sample # 1	0.171	06:31
Breath Volume	1.148	06:31
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample # 2	0.159	06:34
Breath Volume	1.285	06:34
Air Blank	0.000	06:35
Control Test	0.080	06:35
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,KELVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: JOLLY

DATE/TIME OF TEST: 08/12/2021 05:34:27

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.123	05:29
Breath Volume	1.621	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.126	05:33
Breath Volume	1.652	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WINBURN

DATE/TIME OF TEST: 08/15/2021 05:07:00

Diagnostics Check	OK	05:01	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:02	
Subject Sample # 1	0.158	05:03	
Breath Volume	1.789	05:03	
Air Blank	0.000	05:03	
Air Blank		05:05	RFI Detect
Air Blank	0.000	05:06	
Control Test	0.081	05:06	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WINBURN

DATE/TIME OF TEST: 08/15/2021 05:16:19

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.156	05:11
Breath Volume	2.667	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.141	05:14
Breath Volume	1.839	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTSON, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DOTSON, WILLIAM
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 08/16/2021 18:19:48

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:11	
Control Test	0.081	18:12	
Air Blank	0.000	18:12	
Subject Sample	# 1	18:15	No Sample Provided
Breath Volume	0.000	18:15	
Air Blank	0.000	18:16	
Air Blank	0.000	18:18	
Subject Sample	# 2	18:18	Subject Test Refused
Breath Volume	0.000	18:18	
Air Blank	0.000	18:18	
Control Test	0.080	18:19	
Air Blank	0.000	18:19	
Diagnostics Check	OK	18:19	

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CANNOVA,CASSANDRA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: LAFOND

DATE/TIME OF TEST: 08/18/2021 04:11:28

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.198	04:07
Breath Volume	2.769	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.196	04:10
Breath Volume	2.289	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LEIVA

DATE/TIME OF TEST: 08/19/2021 02:48:35

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.000	02:44
Breath Volume	4.296	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:47
Breath Volume	4.527	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SWAZUK

DATE/TIME OF TEST: 08/21/2021 01:00:11

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.209	00:55
Breath Volume	1.910	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.210	00:58
Breath Volume	2.382	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNOVA,CASSANDRA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FUGIEL,ADRIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 08/21/2021 05:11:08

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.000	05:06
Breath Volume	1.218	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.000	05:09
Breath Volume	1.632	05:09
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: STUBENVOLL

DATE/TIME OF TEST: 08/21/2021 23:55:53

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.024	23:51
Breath Volume	1.613	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.023	23:54
Breath Volume	1.800	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SAMANTHA,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BROWN,SAMANTHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RICKETTI

DATE/TIME OF TEST: 08/22/2021 23:11:39

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.000	23:07
Breath Volume	3.125	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:10
Breath Volume	2.957	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIOS,DAVIS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RIOS,DAVIS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ATHANS

DATE/TIME OF TEST: 08/28/2021 03:30:50

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.000	03:26
Breath Volume	2.035	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.000	03:29
Breath Volume	2.070	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 08/29/2021 03:36:14

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.248	03:31
Breath Volume	2.308	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.250	03:34
Breath Volume	3.066	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 08/29/2021 03:42:34

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.079	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.080	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP,RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: AGUILAR RAMIREZ

DATE/TIME OF TEST: 08/29/2021 04:07:04

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.202	04:02
Breath Volume	3.886	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.212	04:05
Breath Volume	3.441	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:07

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JENNA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: NEWTON,JENNA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SOMMER

DATE/TIME OF TEST: 08/29/2021 04:38:28

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.000	04:33
Breath Volume	3.238	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.000	04:37
Breath Volume	3.109	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: FRETTE

DATE/TIME OF TEST: 08/29/2021 20:07:20

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.110	20:02
Breath Volume	3.582	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.105	20:05
Breath Volume	1.957	20:05
Air Blank	0.000	20:06
Control Test	0.080	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/29/2021 21:32:17

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.000	21:27
Breath Volume	2.714	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.000	21:30
Breath Volume	2.507	21:30
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 09/03/2021 11:19