

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACK, REAGHN, S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER, RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PENA-VILLEDA

DATE/TIME OF TEST: 08/02/2021 02:37:23

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1 0.186	02:29	Volume Not Met
Breath Volume	0.558	02:29	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 2 0.184	02:35	Volume Not Met
Breath Volume	1.078	02:35	
Air Blank	0.000	02:36	
Control Test	0.076	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 08/09/2021 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: KIEFER

DATE/TIME OF TEST: 08/05/2021 15:43:36

Diagnostics Check	OK	15:37
Air Blank	0.000	15:37
Control Test	0.080	15:37
Air Blank	0.000	15:38
Subject Sample # 1	0.336	15:39
Breath Volume	2.320	15:39
Air Blank	0.000	15:39
Air Blank	0.000	15:41
Subject Sample # 2	0.331	15:41
Breath Volume	2.171	15:41
Air Blank	0.000	15:42
Control Test	0.081	15:42
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 08/09/2021 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 08/08/2021 03:12:20

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.080	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1 0.070	03:05	
Breath Volume	1.390	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:10	Subject Test Refused
Breath Volume	0.476	03:10	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 08/09/2021 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GONZALEZ JARQUIN

DATE/TIME OF TEST: 08/12/2021 01:55:53

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.102	01:50
Breath Volume	2.023	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.088	01:54
Breath Volume	1.253	01:54
Air Blank	0.000	01:54
Control Test	0.076	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CULLIFER

DATE/TIME OF TEST: 08/13/2021 00:00:57

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.000	23:56
Breath Volume	2.035	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:59
Subject Sample # 2	0.000	23:59
Breath Volume	2.085	23:59
Air Blank	0.000	00:00
Control Test	0.078	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: HECKERMAN

DATE/TIME OF TEST: 08/13/2021 02:03:16

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.139	01:58
Breath Volume	2.597	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.129	02:01
Breath Volume	2.449	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WILSON,MATT
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BUSCHHORN

DATE/TIME OF TEST: 08/16/2021 03:05:17

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.135	03:00
Breath Volume	2.609	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.133	03:03
Breath Volume	2.925	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: WOLVERTON

DATE/TIME OF TEST: 08/16/2021 22:54:12

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.000	22:49
Breath Volume	1.761	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.000	22:52
Breath Volume	2.828	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACK,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: YACK,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 08/20/2021 05:20:11

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.079	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.184	05:14
Breath Volume	1.644	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.185	05:18
Breath Volume	3.109	05:18
Air Blank	0.000	05:19
Control Test	0.078	05:19
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ANTUNEZ

DATE/TIME OF TEST: 08/21/2021 19:10:26

Diagnostics Check	OK	18:57	
Air Blank	0.000	18:57	
Control Test	0.080	18:58	
Air Blank	0.000	18:58	
Subject Sample	# 1 0.162	19:01	Volume Not Met
Breath Volume	0.078	19:01	
Air Blank	0.000	19:02	
Air Blank	0.000	19:04	
Subject Sample	# 2 0.187	19:05	
Breath Volume	2.722	19:05	
Air Blank	0.000	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 3 0.190	19:08	
Breath Volume	2.390	19:08	
Air Blank	0.000	19:09	
Control Test	0.074	19:09	Control Outside Tolerance
Air Blank	0.000	19:10	

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: KICKER

DATE/TIME OF TEST: 08/24/2021 03:11:07

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.176	03:06
Breath Volume	1.933	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.172	03:09
Breath Volume	1.609	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 08/26/2021 01:32:05

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.265	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLETON,DANIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,GERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: LIBBY

DATE/TIME OF TEST: 08/26/2021 16:35:56

Diagnostics Check	OK	16:29
Air Blank	0.000	16:29
Control Test	0.080	16:30
Air Blank	0.000	16:30
Subject Sample # 1	0.077	16:31
Breath Volume	2.867	16:31
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample # 2	0.080	16:34
Breath Volume	3.214	16:34
Air Blank	0.000	16:35
Control Test	0.080	16:35
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

Data Download Date/Time: 09/13/2021 14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ARMSTRONG JR

DATE/TIME OF TEST: 08/01/2021 01:55:19

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.103	01:50
Breath Volume	1.589	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.095	01:53
Breath Volume	1.472	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 08/11/2021 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ROSALES-ANTONIO

DATE/TIME OF TEST: 08/01/2021 03:10:09

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.162	03:04
Breath Volume	1.714	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.158	03:08
Breath Volume	1.468	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 08/11/2021 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TIBURCIO,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CORDOVA

DATE/TIME OF TEST: 08/01/2021 20:45:12

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.081	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.184	20:37	Volume Not Met
Breath Volume	0.554	20:37	
Air Blank	0.000	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 2 0.212	20:40	
Breath Volume	1.183	20:40	
Air Blank	0.000	20:41	
Air Blank	0.000	20:42	
Subject Sample	# 3 0.218	20:43	
Breath Volume	2.128	20:43	
Air Blank	0.000	20:44	
Control Test	0.079	20:44	
Air Blank	0.000	20:45	
Diagnostics Check	OK	20:45	

Data Download Date/Time: 08/11/2021 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 08/07/2021 03:13:14

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.216	03:08
Breath Volume	2.539	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.201	03:11
Breath Volume	3.003	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 08/11/2021 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: ASHBROOK

DATE/TIME OF TEST: 08/07/2021 03:49:22

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.160	03:44
Breath Volume	1.554	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.162	03:47
Breath Volume	1.585	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 08/11/2021 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: REYES JR.

DATE/TIME OF TEST: 08/08/2021 03:37:34

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.230	03:33
Breath Volume	1.207	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.242	03:35
Breath Volume	1.460	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 08/11/2021 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HORNE II

DATE/TIME OF TEST: 08/08/2021 03:54:49

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.081	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 08/11/2021 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: COOMES JR.

DATE/TIME OF TEST: 08/08/2021 04:31:08

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.194	04:26
Breath Volume	2.414	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.186	04:29
Breath Volume	1.894	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 08/11/2021 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR, BARBARA, H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MANGANARO, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 08/08/2021 08:19:28

Diagnostics Check	OK	08:12
Air Blank	0.000	08:13
Control Test	0.080	08:13
Air Blank	0.000	08:14
Subject Sample # 1	0.097	08:14
Breath Volume	2.910	08:14
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.096	08:17
Breath Volume	2.835	08:17
Air Blank	0.000	08:18
Control Test	0.080	08:18
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 08/11/2021 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/14/2021 04:02:12

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.185	03:57
Breath Volume	2.410	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.176	04:00
Breath Volume	2.289	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 09/10/2021 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 08/15/2021 00:27:52

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.081	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.166	00:22
Breath Volume	1.585	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.163	00:26
Breath Volume	1.566	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 09/10/2021 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/15/2021 02:26:26

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.081	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1 0.112	02:17	Volume Not Met
Breath Volume	0.542	02:17	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 2 0.123	02:21	
Breath Volume	1.234	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 3 0.120	02:24	
Breath Volume	1.285	02:24	
Air Blank	0.000	02:25	
Control Test	0.079	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 09/10/2021 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/15/2021 04:58:46

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.147	04:53
Breath Volume	1.875	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.148	04:57
Breath Volume	2.585	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 08/21/2021 00:55:10

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.183	00:50
Breath Volume	2.082	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.174	00:53
Breath Volume	1.449	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PIEDRAHITA ATEHORTUA

DATE/TIME OF TEST: 08/21/2021 03:53:26

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.046	03:48
Breath Volume	2.445	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.044	03:51
Breath Volume	2.609	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 08/21/2021 22:51:09

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1 0.000	22:43	Volume Not Met
Breath Volume	0.000	22:43	
Air Blank	0.000	22:43	
Air Blank	0.000	22:45	
Subject Sample	# 2 0.000	22:46	
Breath Volume	1.582	22:46	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 3 0.000	22:49	
Breath Volume	1.316	22:49	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 08/22/2021 01:10:46

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.151	01:06
Breath Volume	2.000	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.158	01:09
Breath Volume	2.812	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MATUSCHEWSKY

DATE/TIME OF TEST: 08/22/2021 04:27:59

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.094	04:23
Breath Volume	1.328	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.095	04:26
Breath Volume	1.152	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SHIVER

DATE/TIME OF TEST: 08/27/2021 03:14:51

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.132	03:09
Breath Volume	1.843	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.140	03:13
Breath Volume	2.777	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/29/2021 03:13:57

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.081	03:00	
Air Blank	0.000	03:01	
Subject Sample	# 1 0.075	03:04	Volume Not Met
Breath Volume	0.425	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 2 0.094	03:07	
Breath Volume	1.472	03:07	
Air Blank	0.000	03:08	
Air Blank	0.000	03:10	
Subject Sample	# 3 0.093	03:12	
Breath Volume	1.578	03:12	
Air Blank	0.000	03:13	
Control Test	0.078	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 08/29/2021 04:07:15

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.000	04:02
Breath Volume	2.121	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.000	04:05
Breath Volume	1.718	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 09/10/2021 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DAWSON,MARCUS
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: ALVARADO ESTRADA

DATE/TIME OF TEST: 08/01/2021 04:12:33

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.209	04:07
Breath Volume	1.832	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.205	04:10
Breath Volume	1.652	04:10
Air Blank	0.000	04:11
Control Test	0.076	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/19/2021 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: BAUMAN

DATE/TIME OF TEST: 08/05/2021 21:07:15

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.289	21:02
Breath Volume	1.656	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.283	21:05
Breath Volume	1.542	21:05
Air Blank	0.000	21:06
Control Test	0.079	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 08/19/2021 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BEANE,BRANDON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: ARAUJO

DATE/TIME OF TEST: 08/07/2021 05:27:42

Diagnostics Check	OK	05:24	
Air Blank	0.000	05:25	
Control Test	0.081	05:25	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:26	Subject Test Refused
Breath Volume	0.000	05:26	
Air Blank	0.000	05:26	
Control Test	0.080	05:27	
Air Blank	0.000	05:27	
Diagnostics Check	OK	05:27	

Data Download Date/Time: 08/19/2021 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: YOUNG,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: WHITEWAY

DATE/TIME OF TEST: 08/07/2021 19:55:40

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.083	19:50
Breath Volume	2.207	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:52
Subject Sample # 2	0.083	19:54
Breath Volume	1.265	19:54
Air Blank	0.000	19:54
Control Test	0.079	19:55
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 08/19/2021 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/11/2021 19:13:12

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:10	
Control Test	0.080	19:11	
Air Blank	0.000	19:11	
Subject Sample	# 1	19:11	Subject Test Refused
Breath Volume	0.000	19:11	
Air Blank	0.000	19:12	
Control Test	0.079	19:12	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 08/19/2021 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WILKINSON,RYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WILK

DATE/TIME OF TEST: 08/12/2021 01:41:36

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.117	01:36
Breath Volume	3.085	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.112	01:39
Breath Volume	2.078	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 08/19/2021 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: SCARPA

DATE/TIME OF TEST: 08/12/2021 16:41:05

Diagnostics Check	OK	16:34
Air Blank	0.000	16:35
Control Test	0.081	16:35
Air Blank	0.000	16:36
Subject Sample # 1	0.000	16:36
Breath Volume	3.027	16:36
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.000	16:39
Breath Volume	2.226	16:39
Air Blank	0.000	16:40
Control Test	0.081	16:40
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 08/19/2021 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/13/2021 04:12:28

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.081	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:10	Slope Not Met
Breath Volume	0.851	04:10	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 08/19/2021 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRACK,ROBERT,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 08/13/2021 04:22:03

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.206	04:15
Breath Volume	1.531	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.206	04:20
Breath Volume	2.121	04:20
Air Blank	0.000	04:21
Control Test	0.077	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:22

Data Download Date/Time: 08/19/2021 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI, PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: PIPERSBURGH

DATE/TIME OF TEST: 08/13/2021 21:35:40

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:30	No Sample Provided
Breath Volume	0.000	21:30	
Air Blank	0.000	21:30	
Air Blank	0.000	21:32	
Subject Sample	# 2	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:34	
Control Test	0.081	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 08/19/2021 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE, KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HANOLD

DATE/TIME OF TEST: 08/14/2021 01:34:59

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.153	01:30
Breath Volume	1.878	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.151	01:33
Breath Volume	2.199	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 08/19/2021 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI, PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 08/15/2021 22:03:00

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.081	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2	0.187	21:57
Breath Volume	1.417	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample	# 3	0.180	22:01
Breath Volume	1.273	22:01	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 08/19/2021 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ZALAK,JONATHAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: KUHLMAN

DATE/TIME OF TEST: 08/19/2021 02:48:19

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.145	02:36	No .020 agreement
Breath Volume	1.316	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.117	02:42	Volume Not Met
Breath Volume	0.453	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 3 0.117	02:46	No .020 agreement
Breath Volume	1.929	02:46	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 08/19/2021 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ZALAK, JONATHAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: KUHLMAN

DATE/TIME OF TEST: 08/19/2021 02:59:04

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.131	02:53
Breath Volume	1.839	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.126	02:57
Breath Volume	2.183	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:59

Data Download Date/Time: 08/19/2021 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WILKINSON,RYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/23/2021 00:10:29

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.080	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 09/29/2021 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOGDON,BRAD,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SPINA,TYLOR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: ARSLANIAN

DATE/TIME OF TEST: 08/29/2021 18:12:19

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.080	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.243	18:07
Breath Volume	1.636	18:07
Air Blank	0.000	18:08
Air Blank	0.000	18:10
Subject Sample # 2	0.248	18:10
Breath Volume	1.550	18:10
Air Blank	0.000	18:11
Control Test	0.079	18:11
Air Blank	0.000	18:12
Diagnostics Check	OK	18:12

Data Download Date/Time: 09/29/2021 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GOLDEN,LISA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ROSICLAIR

DATE/TIME OF TEST: 08/01/2021 01:16:41

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.078	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.079	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 09/01/2021 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNE,ASHTON,D
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SUPER,MICHAEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: PORUSH

DATE/TIME OF TEST: 08/03/2021 17:47:46

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.078	17:42
Air Blank	0.000	17:42
Subject Sample # 1	0.261	17:43
Breath Volume	2.832	17:43
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.255	17:45
Breath Volume	3.007	17:45
Air Blank	0.000	17:46
Control Test	0.078	17:47
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 09/01/2021 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,A,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: MOLINA PAULINO

DATE/TIME OF TEST: 08/05/2021 19:53:43

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.078	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Control Test	0.079	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 09/01/2021 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MALONEY

DATE/TIME OF TEST: 08/06/2021 01:28:04

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.078	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 09/01/2021 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GOMES DE OLIVEIRA

DATE/TIME OF TEST: 08/07/2021 01:43:38

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.146	01:38
Breath Volume	3.250	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.149	01:41
Breath Volume	3.402	01:41
Air Blank	0.000	01:42
Control Test	0.077	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/01/2021 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISNER, ADAM,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BURNETTE, ASHLEY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 08/10/2021 14:41:23

Diagnostics Check	OK	14:34
Air Blank	0.000	14:34
Control Test	0.078	14:35
Air Blank	0.000	14:35
Subject Sample # 1	0.290	14:36
Breath Volume	2.031	14:36
Air Blank	0.000	14:37
Air Blank	0.000	14:39
Subject Sample # 2	0.285	14:39
Breath Volume	1.234	14:39
Air Blank	0.000	14:40
Control Test	0.078	14:40
Air Blank	0.000	14:41
Diagnostics Check	OK	14:41

Data Download Date/Time: 09/01/2021 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WALKER,KIANA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WULF

DATE/TIME OF TEST: 08/24/2021 00:24:31

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.078	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.079	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 09/01/2021 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,YUDIEL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ASECIO,JAY
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: MANZANAREZ

DATE/TIME OF TEST: 08/07/2021 09:33:08

Diagnostics Check	OK	09:30	
Air Blank	0.000	09:30	
Control Test	0.082	09:31	
Air Blank	0.000	09:31	
Subject Sample	# 1	09:31	Subject Test Refused
Breath Volume	0.000	09:31	
Air Blank	0.000	09:32	
Control Test	0.082	09:32	
Air Blank	0.000	09:33	
Diagnostics Check	OK	09:33	

Data Download Date/Time: 09/02/2021 01:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,YUDIEL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ASECIO,JAY
ARRESTING OFFICER AGENCY: MICCOSUKEE PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: MANZANAREZ

DATE/TIME OF TEST: 08/07/2021 09:46:00

Diagnostics Check	OK	09:39
Air Blank	0.000	09:40
Control Test	0.081	09:40
Air Blank	0.000	09:41
Subject Sample # 1	0.150	09:41
Breath Volume	1.347	09:41
Air Blank	0.000	09:42
Air Blank	0.000	09:43
Subject Sample # 2	0.148	09:44
Breath Volume	1.535	09:44
Air Blank	0.000	09:45
Control Test	0.081	09:45
Air Blank	0.000	09:45
Diagnostics Check	OK	09:46

Data Download Date/Time: 09/02/2021 01:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,TREVOR,D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SEGUI,YOSVANNY
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 08/10/2021 03:00:06

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.082	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Control Test	0.081	02:59	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 09/02/2021 01:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NANAN,KEIL,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: NANAN,KEIL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 08/15/2021 04:51:49

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:49	
Control Test	0.081	04:49	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:50	
Control Test	0.081	04:51	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 09/02/2021 02:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERO ORTIZ,JOHAN,S
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BARRERO ORTIZ,JOHAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: SIGARS

DATE/TIME OF TEST: 08/15/2021 17:59:48

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.081	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.087	17:55
Breath Volume	1.269	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.082	17:58
Breath Volume	1.203	17:58
Air Blank	0.000	17:58
Control Test	0.079	17:59
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 09/02/2021 02:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CANTELAR,JOAN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: CORTES DELGADO

DATE/TIME OF TEST: 08/17/2021 17:52:30

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.082	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.000	17:48
Breath Volume	2.375	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.000	17:51
Breath Volume	2.812	17:51
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 09/02/2021 02:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: MALTEZ PEREZ

DATE/TIME OF TEST: 08/22/2021 18:39:01

Diagnostics Check	OK	18:32
Air Blank	0.000	18:32
Control Test	0.081	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.212	18:34
Breath Volume	2.140	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:36
Subject Sample # 2	0.224	18:37
Breath Volume	2.757	18:37
Air Blank	0.000	18:38
Control Test	0.078	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:39

Data Download Date/Time: 09/02/2021 02:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANIO,PIERRE,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: LLANIO,PIERRE
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: CETO TORREZ

DATE/TIME OF TEST: 08/28/2021 07:03:16

Diagnostics Check	OK	07:00	
Air Blank	0.000	07:00	
Control Test	0.082	07:01	
Air Blank	0.000	07:01	
Subject Sample	# 1	07:01	Subject Test Refused
Breath Volume	0.000	07:01	
Air Blank	0.000	07:02	
Control Test	0.082	07:02	
Air Blank	0.000	07:03	
Diagnostics Check	OK	07:03	

Data Download Date/Time: 09/02/2021 02:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,GUILLERMO,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: ROJAS,GUILLERMO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: GILLIARD

DATE/TIME OF TEST: 08/28/2021 23:25:50

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.124	23:21
Breath Volume	3.957	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.122	23:24
Breath Volume	3.597	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 09/02/2021 02:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTELAR,JOAN,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: BODDEN,DANIEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MARTINEZ GONZALEZ

DATE/TIME OF TEST: 08/29/2021 03:59:36

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.231	03:54
Breath Volume	1.449	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.248	03:57
Breath Volume	1.687	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 09/02/2021 02:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JARRETT,RASHAHN,C
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: JARRETT,RASHAHN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: CORDOBA

DATE/TIME OF TEST: 08/29/2021 05:57:37

Diagnostics Check	OK	05:54	
Air Blank	0.000	05:54	
Control Test	0.082	05:55	
Air Blank	0.000	05:55	
Subject Sample	# 1	05:56	Subject Test Refused
Breath Volume	0.000	05:56	
Air Blank	0.000	05:56	
Control Test	0.082	05:57	
Air Blank	0.000	05:57	
Diagnostics Check	OK	05:57	

Data Download Date/Time: 09/02/2021 02:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP TRP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ-BONILLA, RONALD,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: MELENDEZ-BONILLA, RONALD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 08/30/2021 01:37:57

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.172	01:29
Breath Volume	1.292	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.193	01:32
Breath Volume	1.937	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 3	0.185	01:36
Breath Volume	2.488	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 09/02/2021 02:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BERSANI

DATE/TIME OF TEST: 08/08/2021 04:57:10

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.077	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.279	04:52
Breath Volume	3.046	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.276	04:55
Breath Volume	2.250	04:55
Air Blank	0.000	04:56
Control Test	0.077	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 09/01/2021 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: RIVERA VEGA

DATE/TIME OF TEST: 08/29/2021 03:23:27

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.076	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.079	03:18
Breath Volume	1.953	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.079	03:21
Breath Volume	1.949	03:21
Air Blank	0.000	03:22
Control Test	0.076	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 09/01/2021 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: BOHLANDER

DATE/TIME OF TEST: 08/22/2021 18:05:11

Diagnostics Check	OK	18:02	
Air Blank	0.000	18:02	
Control Test	0.078	18:03	
Air Blank	0.000	18:03	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:04	
Control Test	0.078	18:04	
Air Blank	0.000	18:05	
Diagnostics Check	OK	18:05	

Data Download Date/Time: 09/13/2021 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGBY,CHARLES,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BIGBY,CHARLES
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1303107
DRY GAS EXPIRY DATE: 03/28/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CHAPPUIES

DATE/TIME OF TEST: 08/22/2021 23:34:11

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.190	23:29
Breath Volume	4.089	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.177	23:32
Breath Volume	1.660	23:32
Air Blank	0.000	23:33
Control Test	0.077	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 09/13/2021 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 08/07/2021 01:05:17

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.157	01:00
Breath Volume	1.699	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.164	01:03
Breath Volume	1.578	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 08/09/2021 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1111579
DRY GAS EXPIRY DATE: 03/19/2022
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: RODRIGUEZ CASTILLO

DATE/TIME OF TEST: 08/21/2021 02:45:27

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.151	02:40
Breath Volume	1.570	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.144	02:43
Breath Volume	1.898	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 09/13/2021 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ESPINO

DATE/TIME OF TEST: 08/07/2021 01:17:39

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.078	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.076	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,M,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SEDENO-SISCAL

DATE/TIME OF TEST: 08/08/2021 04:08:38

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.043	04:04
Breath Volume	2.148	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.037	04:07
Breath Volume	1.804	04:07
Air Blank	0.000	04:07
Control Test	0.077	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: XAVIER,STEEVE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: FERIA REYES

DATE/TIME OF TEST: 08/08/2021 05:12:45

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.077	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.080	05:07
Breath Volume	1.347	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.077	05:11
Breath Volume	2.703	05:11
Air Blank	0.000	05:11
Control Test	0.077	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VALDES,L
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GERMAN

DATE/TIME OF TEST: 08/13/2021 03:22:12

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.077	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.254	03:12
Breath Volume	1.417	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.233	03:16
Breath Volume	1.906	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 3	0.233	03:20
Breath Volume	1.820	03:20
Air Blank	0.000	03:21
Control Test	0.077	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: WINLAND,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GARRIDO

DATE/TIME OF TEST: 08/14/2021 01:01:17

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1 0.087	00:53	Volume Not Met
Breath Volume	0.195	00:53	
Air Blank	0.000	00:53	
Air Blank	0.000	00:55	
Subject Sample	# 2 0.088	00:55	
Breath Volume	1.644	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 3 0.086	00:59	
Breath Volume	1.324	00:59	
Air Blank	0.000	01:00	
Control Test	0.078	01:00	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUENO,MELISSA,C
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BUENO

DATE/TIME OF TEST: 08/15/2021 01:26:14

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.077	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.205	01:20
Breath Volume	1.355	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.205	01:24
Breath Volume	1.457	01:24
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: MADUENO

DATE/TIME OF TEST: 08/15/2021 05:30:56

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.077	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.150	05:25
Breath Volume	1.601	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.151	05:29
Breath Volume	1.890	05:29
Air Blank	0.000	05:30
Control Test	0.077	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/17/2021 05:42:39

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.076	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.141	05:37
Breath Volume	2.804	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.136	05:41
Breath Volume	2.859	05:41
Air Blank	0.000	05:41
Control Test	0.077	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: UNDERWOOD

DATE/TIME OF TEST: 08/19/2021 02:37:55

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.076	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.154	02:32
Breath Volume	2.308	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.153	02:36
Breath Volume	2.335	02:36
Air Blank	0.000	02:37
Control Test	0.077	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 08/20/2021 01:52:09

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.077	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.099	01:47
Breath Volume	2.203	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample # 2	0.097	01:50
Breath Volume	1.246	01:50
Air Blank	0.000	01:51
Control Test	0.077	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: KYLES

DATE/TIME OF TEST: 08/20/2021 04:44:59

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:42	
Control Test	0.077	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.076	04:44	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ENAMORADO

DATE/TIME OF TEST: 08/20/2021 23:56:05

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.078	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Control Test	0.077	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 08/21/2021 06:51:49

Diagnostics Check	OK	06:45
Air Blank	0.000	06:45
Control Test	0.077	06:46
Air Blank	0.000	06:46
Subject Sample # 1	0.126	06:47
Breath Volume	2.539	06:47
Air Blank	0.000	06:47
Air Blank	0.000	06:49
Subject Sample # 2	0.127	06:50
Breath Volume	1.593	06:50
Air Blank	0.000	06:50
Control Test	0.077	06:51
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: WINLAND,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/22/2021 04:31:24

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.077	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.072	04:26
Breath Volume	2.203	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.069	04:29
Breath Volume	1.652	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LABORDE

DATE/TIME OF TEST: 08/25/2021 03:35:01

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.078	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.077	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 09/07/2021 06:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: TRINIDAD

DATE/TIME OF TEST: 08/26/2021 05:24:01

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.077	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.076	05:23	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 09/07/2021 06:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 08/26/2021 23:04:08

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.077	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.131	22:59
Breath Volume	3.128	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:01
Subject Sample # 2	0.120	23:02
Breath Volume	1.566	23:02
Air Blank	0.000	23:03
Control Test	0.077	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 09/07/2021 06:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BRIFENI

DATE/TIME OF TEST: 08/27/2021 01:20:08

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.150	01:15
Breath Volume	2.765	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.149	01:18
Breath Volume	2.644	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 09/07/2021 06:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 08/27/2021 22:43:55

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.077	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.076	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 09/07/2021 06:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LEMOS

DATE/TIME OF TEST: 08/29/2021 02:55:07

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.077	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.127	02:50
Breath Volume	2.667	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.125	02:53
Breath Volume	1.812	02:53
Air Blank	0.000	02:54
Control Test	0.077	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 09/07/2021 06:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HU,HAROLD
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: ARIZA

DATE/TIME OF TEST: 08/02/2021 05:00:52

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.172	04:56
Breath Volume	2.546	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.175	04:59
Breath Volume	2.152	04:59
Air Blank	0.000	04:59
Control Test	0.080	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 08/19/2021 01:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEAUREGARD,JEFF
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: BARBU

DATE/TIME OF TEST: 08/05/2021 23:15:07

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:10
Breath Volume	2.406	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.000	23:13
Breath Volume	1.589	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 08/19/2021 01:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MURRAY,ROBERT
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 08/07/2021 14:32:10

Diagnostics Check	OK	14:25
Air Blank	0.000	14:26
Control Test	0.080	14:26
Air Blank	0.000	14:26
Subject Sample # 1	0.334	14:27
Breath Volume	1.664	14:27
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.321	14:30
Breath Volume	1.363	14:30
Air Blank	0.000	14:31
Control Test	0.082	14:31
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 08/19/2021 01:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MOSES,SYLVIANNA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 08/14/2021 14:31:00

Diagnostics Check	OK	14:24
Air Blank	0.000	14:24
Control Test	0.081	14:25
Air Blank	0.000	14:25
Subject Sample # 1	0.000	14:26
Breath Volume	1.273	14:26
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.000	14:29
Breath Volume	1.414	14:29
Air Blank	0.000	14:30
Control Test	0.082	14:30
Air Blank	0.000	14:30
Diagnostics Check	OK	14:31

Data Download Date/Time: 08/19/2021 01:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: VERRO II

DATE/TIME OF TEST: 08/16/2021 03:57:24

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 08/19/2021 01:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MORALES,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:34

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 08/16/2021 12:17:38

Diagnostics Check	OK	12:11
Air Blank	0.000	12:11
Control Test	0.079	12:12
Air Blank	0.000	12:12
Subject Sample # 1	0.225	12:13
Breath Volume	1.613	12:13
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample # 2	0.215	12:16
Breath Volume	1.265	12:16
Air Blank	0.000	12:16
Control Test	0.080	12:17
Air Blank	0.000	12:17
Diagnostics Check	OK	12:17

Data Download Date/Time: 08/19/2021 01:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ZILKA

DATE/TIME OF TEST: 08/11/2021 04:18:02

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.151	04:13
Breath Volume	1.140	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.135	04:16
Breath Volume	1.425	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:18

Data Download Date/Time: 08/25/2021 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: RAU

DATE/TIME OF TEST: 08/21/2021 02:19:22

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.114	02:14
Breath Volume	1.332	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.108	02:17
Breath Volume	1.531	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 08/25/2021 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 08/21/2021 22:47:55

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.174	22:43
Breath Volume	2.273	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.166	22:46
Breath Volume	1.437	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 08/25/2021 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 08/21/2021 22:59:39

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.165	22:53
Breath Volume	1.261	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.166	22:58
Breath Volume	2.062	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 08/25/2021 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 08/22/2021 03:23:52

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.222	03:18
Breath Volume	1.167	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.242	03:22
Breath Volume	1.492	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 08/25/2021 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTHWAY,GEORGIA,K
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: NORTHWAY,GEORGIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/24/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 08/27/2021 03:11:34

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.082	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 09/28/2021 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CAIN,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 08/06/2021 02:01:05

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.082	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.083	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 09/02/2021 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CALLOWAY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 08/16/2021 00:40:15

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.082	00:37	
Air Blank	0.000	00:37	
Subject Sample # 1	0.217	00:38	
Breath Volume	2.722	00:38	
Air Blank		00:38	RFI Detect
Air Blank	0.000	00:39	
Control Test	0.082	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 09/02/2021 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CALLOWAY,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 08/16/2021 00:52:10

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.225	00:44
Breath Volume	2.679	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.195	00:47
Breath Volume	1.742	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 3	0.204	00:50
Breath Volume	2.484	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HAIL,WILLIAM
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: BLUDWORTH

DATE/TIME OF TEST: 08/21/2021 04:43:32

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.082	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.081	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: FOUST

DATE/TIME OF TEST: 08/22/2021 05:40:15

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.082	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.193	05:35
Breath Volume	3.476	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:37
Subject Sample # 2	0.182	05:38
Breath Volume	3.910	05:38
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: TWIDDY,LINDSEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BARAHONA-ALEMAN

DATE/TIME OF TEST: 08/23/2021 00:40:54

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.081	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.082	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CAIN,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BROTHERS

DATE/TIME OF TEST: 08/25/2021 21:11:46

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:09	
Control Test	0.082	21:09	
Air Blank	0.000	21:10	
Subject Sample	# 1	21:10	RFI Detect
Breath Volume	0.000	21:10	
Air Blank	0.000	21:10	
Control Test	0.082	21:11	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CAIN,DANIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BROTHERS

DATE/TIME OF TEST: 08/25/2021 21:20:41

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.082	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.000	21:15
Breath Volume	1.707	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.000	21:19
Breath Volume	1.234	21:19
Air Blank	0.000	21:19
Control Test	0.082	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: CRAWFORD,BENJAMIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ZIMMER

DATE/TIME OF TEST: 08/29/2021 02:38:31

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.178	02:33
Breath Volume	1.148	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.178	02:36
Breath Volume	1.156	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: BULUTOGLU

DATE/TIME OF TEST: 08/31/2021 19:36:20

Diagnostics Check	OK	19:24	
Air Blank	0.000	19:25	
Control Test	0.082	19:25	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:29	No Sample Provided
Breath Volume	0.000	19:29	
Air Blank	0.000	19:29	
Air Blank	0.000	19:31	
Subject Sample	# 2	19:34	No Sample Provided
Breath Volume	0.000	19:34	
Air Blank	0.000	19:35	
Control Test	0.081	19:35	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JAMES,E
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: BULUTOGLU

DATE/TIME OF TEST: 08/31/2021 20:02:31

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:59	
Control Test	0.082	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:01	Subject Test Refused
Breath Volume	0.000	20:01	
Air Blank	0.000	20:01	
Control Test	0.082	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 09/02/2021 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SWEARING

DATE/TIME OF TEST: 08/26/2021 20:58:21

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.173	20:52
Breath Volume	2.640	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.193	20:56
Breath Volume	2.179	20:56
Air Blank	0.000	20:57
Control Test	0.079	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 09/24/2021 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: LUBIN

DATE/TIME OF TEST: 08/11/2021 04:55:45

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.078	04:44	
Air Blank	0.000	04:45	
Subject Sample	# 1 0.150	04:48	Volume Not Met
Breath Volume	0.937	04:48	
Air Blank	0.000	04:49	
Air Blank	0.000	04:50	
Subject Sample	# 2 0.152	04:54	Volume Not Met
Breath Volume	0.019	04:54	
Air Blank	0.000	04:54	
Control Test	0.078	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 08/18/2021 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: CISMOSKI

DATE/TIME OF TEST: 08/12/2021 23:23:49

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.077	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.162	23:18
Breath Volume	3.417	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.162	23:22
Breath Volume	3.140	23:22
Air Blank	0.000	23:22
Control Test	0.077	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 08/18/2021 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KESSEL

DATE/TIME OF TEST: 08/14/2021 03:27:02

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.077	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1 0.136	03:15	
Breath Volume	1.492	03:15	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample	# 2 0.113	03:20	Volume Not Met
Breath Volume	0.492	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:23	
Subject Sample	# 3 0.131	03:26	Volume Not Met
Breath Volume	0.273	03:26	
Air Blank		03:26	RFI Detect
Air Blank		03:27	RFI Detect

Data Download Date/Time: 08/18/2021 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KESSEL

DATE/TIME OF TEST: 08/14/2021 03:35:57

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.076	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.147	03:31
Breath Volume	3.363	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.146	03:34
Breath Volume	2.812	03:34
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 08/18/2021 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FACCINI, DAVID, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PEREZ JIMENEZ

DATE/TIME OF TEST: 08/01/2021 03:59:39

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.241	03:55
Breath Volume	1.761	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.246	03:58
Breath Volume	2.238	03:58
Air Blank	0.000	03:58
Control Test	0.077	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 08/18/2021 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FACCINI, DAVID, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: DOCTOR

DATE/TIME OF TEST: 08/15/2021 22:24:55

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.078	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.150	22:20
Breath Volume	1.980	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.149	22:23
Breath Volume	2.097	22:23
Air Blank	0.000	22:23
Control Test	0.078	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 08/18/2021 12:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FACCINI, DAVID, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 08/30/2021 00:23:04

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.075	00:16	
Air Blank	0.000	00:17	
Subject Sample # 1	0.228	00:17	
Breath Volume	2.054	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample # 2	0.219	00:20	
Breath Volume	1.691	00:20	
Air Blank	0.000	00:21	
Control Test	0.078	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	Fail	00:22	Internal Printer Test
Air Blank		00:23	Ambient Fail

Data Download Date/Time: 09/08/2021 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FACCINI, DAVID, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 08/30/2021 00:27:39

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.073	00:27	Control Outside Tolerance
Air Blank	0.000	00:27	

Data Download Date/Time: 09/08/2021 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FACCINI, DAVID, R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI, DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 08/30/2021 00:32:15

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.075	00:31	Interferent Detect
Air Blank	0.000	00:32	

Data Download Date/Time: 09/08/2021 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CARRASCO BAZAN

DATE/TIME OF TEST: 08/08/2021 23:01:17

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.078	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.077	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 08/18/2021 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 08/30/2021 03:35:38

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.077	03:35	
Air Blank		03:35	RFI Detect
Air Blank	0.000	03:35	

Data Download Date/Time: 09/08/2021 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 01/19/2023
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 08/30/2021 03:40:13

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.077	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:39	
Control Test	0.077	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 09/08/2021 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: LARKINS

DATE/TIME OF TEST: 08/31/2021 23:20:13

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.077	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.196	23:14
Breath Volume	1.289	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.198	23:18
Breath Volume	1.421	23:18
Air Blank	0.000	23:19
Control Test	0.077	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 09/08/2021 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: CECIL

DATE/TIME OF TEST: 08/25/2021 22:49:37

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.183	22:44
Breath Volume	2.050	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.173	22:48
Breath Volume	1.191	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 09/08/2021 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 08/26/2021 23:07:19

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.079	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.262	22:56	Volume Not Met
Breath Volume	0.839	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 2 0.276	23:00	
Breath Volume	1.105	23:00	
Air Blank	0.000	23:01	
Air Blank	0.000	23:03	
Subject Sample	# 3 0.278	23:05	
Breath Volume	1.242	23:05	
Air Blank	0.000	23:06	
Control Test	0.078	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 09/08/2021 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MILLNER II

DATE/TIME OF TEST: 08/28/2021 00:10:47

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.216	00:01	Volume Not Met
Breath Volume	0.152	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample	# 2 0.251	00:05	No .020 agreement
Breath Volume	1.390	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 3 0.274	00:09	No .020 agreement
Breath Volume	2.804	00:09	
Air Blank	0.000	00:09	
Control Test	0.078	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 09/08/2021 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: MILLNER II

DATE/TIME OF TEST: 08/28/2021 00:21:38

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.257	00:16
Breath Volume	1.265	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.267	00:19
Breath Volume	1.695	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 09/08/2021 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JOSE,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BELTRAN,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 08/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 08/05/2021 16:11:28

Diagnostics Check	OK	16:02
Air Blank	0.000	16:03
Control Test	0.077	16:03
Air Blank	0.000	16:04
Subject Sample # 1	0.152	16:05
Breath Volume	2.101	16:05
Air Blank	0.000	16:06
Air Blank	0.000	16:08
Subject Sample # 2	0.172	16:09
Breath Volume	2.167	16:09
Air Blank	0.000	16:10
Control Test	0.076	16:10
Air Blank	0.000	16:11
Diagnostics Check	OK	16:11

Data Download Date/Time: 09/20/2021 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,C,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 08/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: LOPEZMORALES

DATE/TIME OF TEST: 08/14/2021 20:08:32

Diagnostics Check	OK	19:59
Air Blank	0.000	19:59
Control Test	0.078	20:00
Air Blank	0.000	20:00
Subject Sample # 1	0.348	20:01
Breath Volume	2.183	20:01
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.322	20:03
Breath Volume	2.312	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 3	0.317	20:06
Breath Volume	1.812	20:06
Air Blank	0.000	20:07
Control Test	0.076	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 09/20/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: ZERPA,EDWIN
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 08/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 08/22/2021 20:25:48

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.077	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.201	20:21
Breath Volume	3.050	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.195	20:24
Breath Volume	1.980	20:24
Air Blank	0.000	20:24
Control Test	0.078	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 09/20/2021 12:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,C,A
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: WOODS,H
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 08/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SANCHEZPEREZ

DATE/TIME OF TEST: 08/25/2021 00:42:40

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.078	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.206	00:38
Breath Volume	2.414	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.215	00:40
Breath Volume	2.695	00:40
Air Blank	0.000	00:41
Control Test	0.077	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 09/20/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BUONPANE,V
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 08/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ARTEAGA

DATE/TIME OF TEST: 08/28/2021 01:34:42

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.229	01:30
Breath Volume	1.503	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.237	01:33
Breath Volume	1.773	01:33
Air Blank	0.000	01:33
Control Test	0.077	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 09/20/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/29/2021 00:59:54

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.174	00:54
Breath Volume	1.136	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.176	00:58
Breath Volume	1.148	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 09/01/2021 07:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TEMS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/29/2021 01:40:32

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.179	01:35
Breath Volume	3.769	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.188	01:38
Breath Volume	1.894	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 09/01/2021 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SOSNOWSKI

DATE/TIME OF TEST: 08/29/2021 23:12:10

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.159	23:07
Breath Volume	2.339	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.164	23:10
Breath Volume	2.136	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 09/01/2021 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TURNER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MELNITCHOUK

DATE/TIME OF TEST: 08/30/2021 01:22:44

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:18
Breath Volume	4.085	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:21
Breath Volume	4.070	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 09/01/2021 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KLOER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GIUNTA

DATE/TIME OF TEST: 08/30/2021 01:52:22

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.160	01:47
Breath Volume	2.039	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.158	01:50
Breath Volume	1.757	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 09/01/2021 07:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CUADRA,MARIO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 08/31/2021 00:53:38

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.167	00:48
Breath Volume	2.621	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.164	00:52
Breath Volume	2.242	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 09/01/2021 07:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORITZ,DFC,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 08/31/2021 22:28:36

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.181	22:23
Breath Volume	1.953	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.184	22:26
Breath Volume	2.343	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 09/01/2021 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 08/08/2021 05:54:31

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:52	
Control Test	0.078	05:52	
Air Blank	0.000	05:52	
Subject Sample	# 1	05:53	RFI Detect
Breath Volume	0.000	05:53	
Air Blank	0.000	05:53	
Control Test	0.079	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 08/19/2021 01:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 08/08/2021 06:02:15

Diagnostics Check	OK	05:56	
Air Blank	0.000	05:56	
Control Test	0.077	05:57	
Air Blank	0.000	05:57	
Subject Sample	# 1 0.218	05:58	
Breath Volume	2.027	05:58	
Air Blank	0.000	05:58	
Air Blank	0.000	06:00	
Subject Sample	# 2 0.216	06:01	
Breath Volume	1.929	06:01	
Air Blank		06:01	RFI Detect
Air Blank	0.000	06:02	

Data Download Date/Time: 08/19/2021 01:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 08/08/2021 06:13:52

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.077	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.212	06:09
Breath Volume	1.781	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:11
Subject Sample # 2	0.204	06:12
Breath Volume	1.980	06:12
Air Blank	0.000	06:12
Control Test	0.077	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 08/19/2021 01:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 08/14/2021 04:04:33

Diagnostics Check	Fail	04:04	Internal Printer Test
Air Blank	0.000	04:04	

Data Download Date/Time: 08/19/2021 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 08/14/2021 04:12:42

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.153	04:08
Breath Volume	1.472	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.155	04:11
Breath Volume	1.550	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/19/2021 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BRETT

DATE/TIME OF TEST: 08/15/2021 03:44:43

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.077	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.084	03:39
Breath Volume	2.148	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.083	03:43
Breath Volume	1.718	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 08/19/2021 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROIZ TOLEDO

DATE/TIME OF TEST: 08/22/2021 04:04:12

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.000	03:59
Breath Volume	1.996	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.000	04:02
Breath Volume	2.167	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 09/08/2021 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNE,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: POLO PATINO

DATE/TIME OF TEST: 08/01/2021 02:40:22

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.166	02:33
Breath Volume	1.445	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.155	02:38
Breath Volume	1.601	02:38
Air Blank	0.000	02:39
Control Test	0.076	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 08/01/2021 04:39:29

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.076	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.093	04:35
Breath Volume	3.210	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.087	04:37
Breath Volume	1.285	04:37
Air Blank	0.000	04:38
Control Test	0.076	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 08/01/2021 05:23:39

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.078	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.177	05:19
Breath Volume	4.617	05:19
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.176	05:22
Breath Volume	3.058	05:22
Air Blank	0.000	05:22
Control Test	0.077	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HUTCHESON,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HARBOUR

DATE/TIME OF TEST: 08/02/2021 04:33:28

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.077	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.139	04:28
Breath Volume	3.523	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.134	04:31
Breath Volume	2.812	04:31
Air Blank	0.000	04:32
Control Test	0.075	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FISHER,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ROGERS II

DATE/TIME OF TEST: 08/04/2021 21:40:39

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.078	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.264	21:36
Breath Volume	1.914	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.266	21:39
Breath Volume	2.042	21:39
Air Blank	0.000	21:39
Control Test	0.077	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNE,S,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DIAW

DATE/TIME OF TEST: 08/05/2021 02:33:31

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.073	02:32	Control Outside Tolerance
Air Blank	0.000	02:33	

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: CETO

DATE/TIME OF TEST: 08/07/2021 04:27:37

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.111	04:22
Breath Volume	1.531	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.103	04:26
Breath Volume	1.464	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 08/07/2021 05:45:43

Diagnostics Check	OK	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:40
Subject Sample # 1	0.058	05:41
Breath Volume	1.207	05:41
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.058	05:44
Breath Volume	1.375	05:44
Air Blank	0.000	05:44
Control Test	0.076	05:45
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: DEGAGNE

DATE/TIME OF TEST: 08/08/2021 00:10:17

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.077	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.078	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: RESTREPO BARRIENTOS

DATE/TIME OF TEST: 08/08/2021 05:20:50

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.078	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.204	05:15
Breath Volume	1.945	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.201	05:19
Breath Volume	2.222	05:19
Air Blank	0.000	05:19
Control Test	0.076	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/08/2021 06:38:41

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.078	06:31
Air Blank	0.000	06:32
Subject Sample # 1	0.112	06:33
Breath Volume	2.656	06:33
Air Blank	0.000	06:34
Air Blank	0.000	06:36
Subject Sample # 2	0.106	06:37
Breath Volume	2.179	06:37
Air Blank	0.000	06:37
Control Test	0.076	06:38
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 08/11/2021 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, DAVID, E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A2
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WINGFIELD

DATE/TIME OF TEST: 08/29/2021 02:12:15

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.185	02:05
Breath Volume	2.855	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.184	02:10
Breath Volume	3.214	02:10
Air Blank	0.000	02:11
Control Test	0.077	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 09/10/2021 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: KHANG,ADAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 08/07/2021 01:28:50

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.147	01:23
Breath Volume	1.242	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.147	01:27
Breath Volume	1.578	01:27
Air Blank	0.000	01:27
Control Test	0.078	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 09/04/2021 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ZUNIGA,MANUEL
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TZOC SEN

DATE/TIME OF TEST: 08/15/2021 03:18:21

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.210	03:13
Breath Volume	2.164	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.208	03:16
Breath Volume	1.480	03:16
Air Blank	0.000	03:17
Control Test	0.078	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 09/04/2021 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANDY,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: GONZALEZ,JOEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 08/29/2021 03:31:10

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.000	03:26
Breath Volume	2.339	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.000	03:29
Breath Volume	2.050	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 09/04/2021 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANDY,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ALVAREZ,OCTAVIO
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/29/2021 04:29:54

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.079	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:28	
Control Test	0.079	04:29	
Air Blank	0.000	04:29	
Diagnostics Check	OK	04:29	

Data Download Date/Time: 09/04/2021 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/18/2021 06:22:26

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:16	
Control Test	0.079	06:16	
Air Blank	0.000	06:17	
Subject Sample # 1	0.000	06:17	
Breath Volume	2.476	06:17	
Air Blank	0.000	06:18	
Air Blank	0.000	06:20	
Subject Sample # 2	0.000	06:20	
Breath Volume	2.710	06:20	
Air Blank	0.000	06:21	
Control Test	0.102	06:21	Interferent Detect
Air Blank		06:22	Purge Fail

Data Download Date/Time: 09/08/2021 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BRINGAS HERNANDEZ

DATE/TIME OF TEST: 08/19/2021 01:12:20

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:08	
Control Test	0.079	01:08	
Air Blank	0.000	01:09	
Subject Sample	# 1	01:10	Interferent Detect
Breath Volume	1.367	01:10	
Air Blank	0.000	01:11	
Control Test	0.055	01:11	Interferent Detect
Air Blank	0.000	01:12	

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BRINGAS HERNANDEZ

DATE/TIME OF TEST: 08/19/2021 01:25:00

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.104	01:19
Breath Volume	1.355	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.104	01:23
Breath Volume	1.207	01:23
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:25

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: SABIDO FLEITAS

DATE/TIME OF TEST: 08/20/2021 20:59:30

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.078	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.139	20:53
Breath Volume	2.281	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.125	20:57
Breath Volume	2.226	20:57
Air Blank	0.000	20:58
Control Test	0.078	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: OJEDA-LEON

DATE/TIME OF TEST: 08/21/2021 01:55:57

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.205	01:50
Breath Volume	2.109	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.210	01:54
Breath Volume	1.519	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HAGEN

DATE/TIME OF TEST: 08/21/2021 04:05:30

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.039	04:00
Breath Volume	1.523	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.041	04:03
Breath Volume	1.527	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CORDONE

DATE/TIME OF TEST: 08/23/2021 04:23:39

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.000	04:19
Breath Volume	2.730	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.000	04:22
Breath Volume	2.324	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: IMUL TZOY

DATE/TIME OF TEST: 08/25/2021 22:21:05

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:17	
Control Test	0.079	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 1	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: CRUZ CASTILLO

DATE/TIME OF TEST: 08/29/2021 01:25:26

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.200	01:20
Breath Volume	2.898	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.207	01:23
Breath Volume	3.179	01:23
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GRECO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18220080A2
DRY GAS EXPIRY DATE: 08/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/29/2021 02:55:36

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.077	02:51
Breath Volume	1.429	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.075	02:53
Breath Volume	2.058	02:53
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 09/08/2021 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCKINLEY, STEVEN, J

BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)

ARRESTING OFFICER: BRUDER, ALEX

ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)

LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1380119

DRY GAS EXPIRY DATE: 09/22/2023

VIOLATION CODE: BUI

ARREST TIME: 20:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 08/21/2021 21:11:38

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.078	21:06	
Air Blank	0.000	21:06	
Subject Sample	# 1 0.172	21:07	
Breath Volume	1.835	21:07	
Air Blank	0.000	21:08	
Air Blank	0.000	21:09	
Subject Sample	# 2 0.169	21:10	
Breath Volume	1.703	21:10	
Air Blank		21:11	Purge Fail
Air Blank	0.000	21:11	

Data Download Date/Time: 08/25/2021 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1380119
BREATH TEST OPERATOR: PADILLA,ANTHONY,D	DRY GAS EXPIRY DATE: 09/22/2023
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)	VIOLATION CODE: BUI
ARRESTING OFFICER: BRUDER,ALEXANDER	ARREST TIME: 21:28
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)	GENDER: Male
LAST AGENCY INSPECTION: 07/31/2021	OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JIMENEZ RODRIGUEZ	DATE/TIME OF TEST: 08/22/2021 22:33:07
--	--

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.079	22:31	
Air Blank		22:32	Purge Fail
Air Blank	0.000	22:33	

Data Download Date/Time: 08/25/2021 15:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PADILLA,ANTHONY,D

BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)

ARRESTING OFFICER: BRUDER,ALEXANDER

ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)

LAST AGENCY INSPECTION: 07/31/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1380119

DRY GAS EXPIRY DATE: 09/22/2023

VIOLATION CODE: BUI

ARREST TIME: 21:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JIMENEZ RODRIGUEZ

DATE/TIME OF TEST: 08/22/2021 22:38:00

Diagnostics Check	OK	22:35	
Air Blank	0.000	22:35	
Control Test	0.079	22:35	
Air Blank	0.000	22:36	
Subject Sample	# 1	22:36	Subject Test Refused
Breath Volume	0.000	22:36	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 08/25/2021 15:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SECHRIST

DATE/TIME OF TEST: 08/01/2021 01:18:59

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.081	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.133	01:14
Breath Volume	3.988	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample # 2	0.131	01:17
Breath Volume	3.816	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ELIZABETH,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BOWER,RACHEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SOUDER

DATE/TIME OF TEST: 08/08/2021 23:43:36

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.080	23:40	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: AUER

DATE/TIME OF TEST: 08/12/2021 23:35:32

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.081	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.059	23:28	Volume Not Met
Breath Volume	0.375	23:28	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2 0.057	23:33	Volume Not Met
Breath Volume	0.644	23:33	
Air Blank	0.000	23:34	
Control Test	0.078	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: AUER

DATE/TIME OF TEST: 08/12/2021 23:55:28

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.078	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1 0.064	23:44	
Breath Volume	1.714	23:44	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 2 0.055	23:49	Volume Not Met
Breath Volume	0.246	23:49	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 3 0.056	23:53	
Breath Volume	1.695	23:53	
Air Blank	0.000	23:54	
Control Test	0.078	23:54	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: GALBO,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CASTRO AMADOR

DATE/TIME OF TEST: 08/16/2021 04:36:41

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.213	04:30
Breath Volume	2.968	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.209	04:34
Breath Volume	2.226	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: JERABEK

DATE/TIME OF TEST: 08/16/2021 20:09:44

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.079	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.279	20:04
Breath Volume	1.949	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.294	20:07
Breath Volume	3.222	20:07
Air Blank	0.000	20:08
Control Test	0.078	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,BRANDON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BRACKINS

DATE/TIME OF TEST: 08/16/2021 21:55:34

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.215	21:50
Breath Volume	3.082	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.223	21:53
Breath Volume	3.089	21:53
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 08/16/2021 22:33:06

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.079	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1 0.248	22:25	Volume Not Met
Breath Volume	0.125	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 2 0.181	22:31	Volume Not Met
Breath Volume	0.703	22:31	
Air Blank	0.000	22:32	
Control Test	0.076	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 08/16/2021 22:38:08

Diagnostics Check	OK	22:36	
Air Blank		22:37	Ambient Fail
Air Blank	0.000	22:38	

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 08/16/2021 22:45:36

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.077	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1 0.202	22:44	Volume Not Met
Breath Volume	0.484	22:44	
Air Blank		22:44	Purge Fail
Air Blank	0.000	22:45	

Data Download Date/Time: 08/17/2021 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 08/16/2021 22:50:01

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.070	22:49	Control Outside Tolerance
Air Blank	0.000	22:50	

Data Download Date/Time: 08/17/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 08/16/2021 23:01:58

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.077	22:52	
Air Blank	0.000	22:52	
Subject Sample	# 1 0.223	22:56	Volume Not Met
Breath Volume	0.855	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 2 0.211	23:00	
Breath Volume	1.582	23:00	
Air Blank		23:01	Purge Fail
Air Blank	0.000	23:01	

Data Download Date/Time: 08/17/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 08/16/2021 23:17:14

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:03	
Control Test	0.076	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1 0.218	23:06	
Breath Volume	2.195	23:06	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 2 0.191	23:12	Volume Not Met
Breath Volume	0.312	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 3 0.220	23:15	
Breath Volume	2.980	23:15	
Air Blank	0.000	23:16	
Control Test	0.075	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 08/17/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,KENNETH,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WAITE,KENNETH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 08/16/2021 23:41:48

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.078	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Control Test	0.079	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 08/17/2021 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BARTOLO

DATE/TIME OF TEST: 08/21/2021 03:38:05

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ASHE

DATE/TIME OF TEST: 08/21/2021 03:57:19

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.079	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BARTOW

DATE/TIME OF TEST: 08/21/2021 20:59:15

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.181	20:54
Breath Volume	2.105	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.166	20:57
Breath Volume	1.144	20:57
Air Blank	0.000	20:58
Control Test	0.078	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HERNANDEZ,LEONARDO
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/22/2021 03:54:13

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.138	03:49
Breath Volume	2.121	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.134	03:52
Breath Volume	2.062	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LITTON,A
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 08/22/2021 23:10:12

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.080	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,ANDREW,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HUDSON,ANDREW
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 08/23/2021 23:42:47

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.000	23:36
Breath Volume	1.398	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.000	23:41
Breath Volume	1.292	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/28/2021 02:39:06

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,R
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 08/28/2021 23:45:19

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.168	23:32	No .020 agreement
Breath Volume	1.593	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2 0.142	23:37	No .020 agreement
Breath Volume	1.187	23:37	
Air Blank	0.000	23:38	
Air Blank	0.000	23:40	
Subject Sample	# 3 0.122	23:43	Volume Not Met
Breath Volume	0.554	23:43	
Air Blank	0.000	23:44	
Control Test	0.078	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 08/28/2021 23:54:07

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.156	23:48
Breath Volume	1.312	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.151	23:52
Breath Volume	1.507	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: GADONNIEX

DATE/TIME OF TEST: 08/29/2021 00:52:19

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 08/31/2021 03:49:25

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1 0.214	03:39	Volume Not Met
Breath Volume	0.789	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 2 0.248	03:43	
Breath Volume	2.011	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:45	
Subject Sample	# 3	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.075	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 09/09/2021 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/02/2021 03:31:50

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.130	03:27
Breath Volume	1.218	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.133	03:30
Breath Volume	1.933	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 08/16/2021 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 08/03/2021 23:23:07

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.080	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 08/16/2021 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 08/07/2021 01:29:46

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.080	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 08/16/2021 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CRESTEA, VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 08/08/2021 21:20:34

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:18	
Control Test	0.079	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 08/16/2021 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS,WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: STAAS

DATE/TIME OF TEST: 08/10/2021 19:33:46

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.079	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.182	19:29
Breath Volume	2.167	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.178	19:32
Breath Volume	1.921	19:32
Air Blank	0.000	19:32
Control Test	0.079	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 08/16/2021 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEMIEC,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Key Colony Beach PD
ARRESTING OFFICER: CRISTEA,VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 08/11/2021 21:58:55

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:55	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:57	
Control Test	0.080	21:58	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 08/16/2021 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FOSTER,MEGHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 08/15/2021 03:15:10

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.163	03:10
Breath Volume	1.851	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.162	03:13
Breath Volume	1.660	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 08/16/2021 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 13:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: MENDIETA

DATE/TIME OF TEST: 08/19/2021 14:24:07

Diagnostics Check	OK	14:17
Air Blank	0.000	14:18
Control Test	0.079	14:18
Air Blank	0.000	14:19
Subject Sample # 1	0.117	14:19
Breath Volume	3.183	14:19
Air Blank	0.000	14:20
Air Blank	0.000	14:22
Subject Sample # 2	0.115	14:22
Breath Volume	2.648	14:22
Air Blank	0.000	14:23
Control Test	0.079	14:23
Air Blank	0.000	14:24
Diagnostics Check	OK	14:24

Data Download Date/Time: 09/08/2021 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CHEYNE

DATE/TIME OF TEST: 08/21/2021 19:51:21

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:50	
Control Test	0.080	19:50	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

Data Download Date/Time: 09/08/2021 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO, JUAN-CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 08/22/2021 22:04:27

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:02	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 09/08/2021 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHANDLER, KYLE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/25/2021 20:30:28

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:25
Subject Sample # 1	0.181	20:25
Breath Volume	1.851	20:25
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.180	20:28
Breath Volume	1.714	20:28
Air Blank	0.000	20:29
Control Test	0.079	20:29
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 09/08/2021 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODNOW,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: ZAINO,JACOB
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: ADKISON

DATE/TIME OF TEST: 08/03/2021 23:32:16

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:27
Breath Volume	1.902	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.000	23:30
Breath Volume	1.609	23:30
Air Blank	0.000	23:31
Control Test	0.077	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HELGELAND

DATE/TIME OF TEST: 08/05/2021 00:22:41

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.171	00:17
Breath Volume	1.808	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.156	00:20
Breath Volume	1.144	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 08/05/2021 20:10:56

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.080	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.000	20:06
Breath Volume	3.656	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.000	20:09
Breath Volume	2.304	20:09
Air Blank	0.000	20:09
Control Test	0.080	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: KAY

DATE/TIME OF TEST: 08/06/2021 00:06:41

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.079	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 3	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	
Breath Volume	2.109	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 2	00:04	
Breath Volume	2.136	00:04	
Air Blank	0.000	00:05	
Control Test	0.073	00:06	Control Outside Tolerance
Air Blank	0.000	00:06	

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 08/06/2021 03:31:28

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.000	03:27
Breath Volume	1.539	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.000	03:30
Breath Volume	1.671	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLER,JESSICA,G
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: GALLER,JESSICA
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 08/07/2021 02:14:26

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1 0.047	02:07	Volume Not Met
Breath Volume	0.238	02:07	
Air Blank	0.000	02:08	
Air Blank	0.000	02:10	
Subject Sample	# 2	02:12	Subject Test Refused
Breath Volume	0.019	02:12	
Air Blank	0.000	02:13	
Control Test	0.078	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 08/07/2021 02:59:47

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.198	02:54
Breath Volume	2.988	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.200	02:58
Breath Volume	2.292	02:58
Air Blank	0.000	02:58
Control Test	0.077	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JAMES,W
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BURGESS,JAMES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: BIBB

DATE/TIME OF TEST: 08/08/2021 00:31:56

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.231	00:27
Breath Volume	1.609	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.220	00:30
Breath Volume	2.074	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,M,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MOORE,M
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CRETUL

DATE/TIME OF TEST: 08/08/2021 03:34:06

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.281	03:29
Breath Volume	1.882	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.278	03:32
Breath Volume	1.746	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: COLES

DATE/TIME OF TEST: 08/11/2021 02:58:47

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.193	02:53
Breath Volume	1.343	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.185	02:56
Breath Volume	1.730	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 08/14/2021 02:19:31

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.247	02:14
Breath Volume	3.257	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.238	02:17
Breath Volume	2.957	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALVET,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BERTALAN,DOMINIC
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: POSTLETHWAITE

DATE/TIME OF TEST: 08/18/2021 11:21:37

Diagnostics Check	OK	11:15
Air Blank	0.000	11:15
Control Test	0.080	11:15
Air Blank	0.000	11:16
Subject Sample # 1	0.000	11:17
Breath Volume	2.175	11:17
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.000	11:20
Breath Volume	1.117	11:20
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 08/18/2021 19:32:13

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.000	19:27
Breath Volume	1.871	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.000	19:30
Breath Volume	1.914	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHL,ASHLYN,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHL,ASHLYN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: HORST

DATE/TIME OF TEST: 08/21/2021 00:39:23

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.137	00:34
Breath Volume	3.800	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:36
Subject Sample # 2	0.137	00:37
Breath Volume	3.703	00:37
Air Blank	0.000	00:38
Control Test	0.078	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,R,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEST,R
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CRUMBY

DATE/TIME OF TEST: 08/21/2021 01:12:37

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.164	01:07
Breath Volume	2.765	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.148	01:10
Breath Volume	2.425	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GURRIERI,JOHN,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: GURRIERI,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: FAJARDO SAGASTUME

DATE/TIME OF TEST: 08/22/2021 01:50:59

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: BENTLEY CARROLL

DATE/TIME OF TEST: 08/22/2021 23:04:28

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.081	22:59
Breath Volume	1.500	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.077	23:02
Breath Volume	1.152	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SOMWARU

DATE/TIME OF TEST: 08/28/2021 03:15:24

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.162	03:10
Breath Volume	1.910	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.151	03:13
Breath Volume	1.406	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,JOSIE,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WELLS,JOSIE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 13:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: GOATER

DATE/TIME OF TEST: 08/30/2021 14:27:40

Diagnostics Check	OK	14:20
Air Blank	0.000	14:21
Control Test	0.080	14:21
Air Blank	0.000	14:22
Subject Sample # 1	0.000	14:22
Breath Volume	1.253	14:22
Air Blank	0.000	14:23
Air Blank	0.000	14:25
Subject Sample # 2	0.000	14:26
Breath Volume	1.914	14:26
Air Blank	0.000	14:26
Control Test	0.080	14:27
Air Blank	0.000	14:27
Diagnostics Check	OK	14:27

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELCHER,ASHLYN,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WHELCHER,ASHLYN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/03/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 08/31/2021 00:03:29

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.183	23:55
Breath Volume	1.980	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 3	0.162	23:58
Breath Volume	1.984	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.179	00:01
Breath Volume	2.957	00:01
Air Blank	0.000	00:02
Control Test	0.075	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 09/23/2021 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1060911
BREATH TEST OPERATOR: ELLIS,ALEXANDER,	DRY GAS EXPIRY DATE: 12/11/2021
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ELLIS,ALEXANDER	ARREST TIME: 23:02
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/28/2021	OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: OLAES DATE/TIME OF TEST: 08/04/2021 23:23:35

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.078	23:17	
Air Blank	0.000	23:18	
Subject Sample # 1	0.000	23:19	
Breath Volume	2.261	23:19	
Air Blank	0.000	23:19	
Air Blank	0.000	23:21	
Subject Sample # 2	0.000	23:22	
Breath Volume	2.019	23:22	
Air Blank	0.000	23:22	
Control Test	0.074	23:22	Control Outside Tolerance
Air Blank	0.000	23:23	

Data Download Date/Time: 09/30/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1060911
BREATH TEST OPERATOR: ELLIS,ALEXANDER,	DRY GAS EXPIRY DATE: 12/11/2021
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: CRAWFORD,JUDSON	ARREST TIME: 23:19
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 07/28/2021	OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: FARRELL DATE/TIME OF TEST: 08/05/2021 23:50:53

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.076	23:45	
Air Blank	0.000	23:45	
Subject Sample # 1	0.274	23:46	
Breath Volume	1.921	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:48	
Subject Sample # 2	0.266	23:49	
Breath Volume	1.464	23:49	
Air Blank	0.000	23:49	
Control Test	0.074	23:50	Control Outside Tolerance
Air Blank	0.000	23:50	

Data Download Date/Time: 09/30/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BARRERA,ALEXANDER
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: YOUNGS

DATE/TIME OF TEST: 08/08/2021 01:34:18

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 09/30/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MASON,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CALLAWAY,DANIEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 08/17/2021 22:23:32

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.149	22:18
Breath Volume	1.152	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.144	22:22
Breath Volume	1.289	22:22
Air Blank	0.000	22:22
Control Test	0.076	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 09/30/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1060911
BREATH TEST OPERATOR: BURKE,LYLE,J	DRY GAS EXPIRY DATE: 12/11/2021
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: BURKE,LYLE	ARREST TIME: 21:52
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 07/28/2021	OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: PACETTI DATE/TIME OF TEST: 08/20/2021 23:06:14

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.246	23:01
Breath Volume	3.187	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.248	23:04
Breath Volume	2.554	23:04
Air Blank	0.000	23:05
Control Test	0.078	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 09/30/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 07/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 08/21/2021 05:57:07

Diagnostics Check	OK	05:42	
Air Blank	0.000	05:42	
Control Test	0.079	05:43	
Air Blank	0.000	05:43	
Subject Sample	# 1 0.000	05:46	Volume Not Met
Breath Volume	0.468	05:46	
Air Blank	0.000	05:47	
Air Blank	0.000	05:49	
Subject Sample	# 2 0.000	05:50	
Breath Volume	1.199	05:50	
Air Blank	0.000	05:50	
Air Blank	0.000	05:52	
Subject Sample	# 3 0.000	05:55	Volume Not Met
Breath Volume	0.777	05:55	
Air Blank	0.000	05:56	
Control Test	0.080	05:56	
Air Blank	0.000	05:57	
Diagnostics Check	OK	05:57	

Data Download Date/Time: 09/30/2021 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WAXMAN

DATE/TIME OF TEST: 08/19/2021 04:08:44

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.000	04:04
Breath Volume	2.843	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.000	04:07
Breath Volume	2.941	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JEANTY,JEAN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 08/20/2021 04:28:39

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.159	04:23
Breath Volume	1.929	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.159	04:26
Breath Volume	2.093	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 08/20/2021 05:43:51

Diagnostics Check	OK	05:37
Air Blank	0.000	05:37
Control Test	0.081	05:37
Air Blank	0.000	05:38
Subject Sample # 1	0.095	05:39
Breath Volume	1.570	05:39
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.098	05:42
Breath Volume	1.699	05:42
Air Blank	0.000	05:42
Control Test	0.081	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JULIEN

DATE/TIME OF TEST: 08/21/2021 02:51:34

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.083	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.202	02:45
Breath Volume	1.394	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.204	02:49
Breath Volume	2.617	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: KILLEEN

DATE/TIME OF TEST: 08/21/2021 04:57:42

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.082	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.206	04:52
Breath Volume	1.878	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.201	04:55
Breath Volume	1.832	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOMES,BRUNO
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 08/22/2021 02:16:04

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.082	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.156	02:10
Breath Volume	1.468	02:10
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.151	02:14
Breath Volume	1.359	02:14
Air Blank	0.000	02:15
Control Test	0.077	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:16

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LISOVA,INNA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MINNIS

DATE/TIME OF TEST: 08/22/2021 04:25:08

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.019	04:20
Breath Volume	2.035	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.017	04:23
Breath Volume	1.578	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HIGUITA,WILLIAM
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: PUERTAS BRAVO

DATE/TIME OF TEST: 08/22/2021 05:08:29

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.207	05:03
Breath Volume	1.492	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.194	05:06
Breath Volume	1.871	05:06
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: QUINTEROS MONJARAS

DATE/TIME OF TEST: 08/22/2021 06:52:26

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.082	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.232	06:47
Breath Volume	3.125	06:47
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.231	06:50
Breath Volume	3.402	06:50
Air Blank	0.000	06:51
Control Test	0.081	06:51
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BUDYONNY

DATE/TIME OF TEST: 08/24/2021 01:13:28

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.158	01:08
Breath Volume	1.308	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.170	01:11
Breath Volume	1.421	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JEANTY,JEAN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/29/2021 04:08:11

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.082	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.171	04:03
Breath Volume	2.289	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.175	04:06
Breath Volume	1.816	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LIPTAK,LAUREN
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 08/30/2021 07:23:03

Diagnostics Check	OK	07:16
Air Blank	0.000	07:16
Control Test	0.081	07:17
Air Blank	0.000	07:17
Subject Sample # 1	0.066	07:18
Breath Volume	2.289	07:18
Air Blank	0.000	07:19
Air Blank	0.000	07:21
Subject Sample # 2	0.067	07:21
Breath Volume	3.066	07:21
Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:22
Diagnostics Check	OK	07:23

Data Download Date/Time: 09/08/2021 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 08/22/2021 04:02:58

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.197	03:58
Breath Volume	2.265	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.190	04:01
Breath Volume	2.226	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 09/08/2021 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: O'SULLIVAN

DATE/TIME OF TEST: 08/22/2021 21:36:54

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.208	21:31
Breath Volume	2.039	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.203	21:35
Breath Volume	2.187	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 09/08/2021 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL,T'KEYAH
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: RIVERA-NOVOA

DATE/TIME OF TEST: 08/23/2021 02:30:03

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1 0.169	02:26	
Breath Volume	2.640	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 2 0.169	02:29	
Breath Volume	2.089	02:29	
Air Blank		02:29	RFI Detect
Air Blank		02:30	RFI Detect

Data Download Date/Time: 09/08/2021 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL,T'KEYAH
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: RIVERA-NOVOA

DATE/TIME OF TEST: 08/23/2021 02:39:53

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.166	02:35
Breath Volume	1.867	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.158	02:38
Breath Volume	1.281	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 09/08/2021 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PETERKIN,MARSHANE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 08/30/2021 05:37:52

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.033	05:31
Breath Volume	1.304	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.031	05:36
Breath Volume	1.199	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 09/08/2021 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: TOKAY

DATE/TIME OF TEST: 08/31/2021 21:16:24

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.081	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:09	No Sample Provided
Breath Volume	0.000	21:09	
Air Blank	0.000	21:09	
Air Blank	0.000	21:11	
Subject Sample	# 2	21:15	No Sample Provided
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.082	21:15	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 09/08/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: HERMAN

DATE/TIME OF TEST: 08/31/2021 23:32:03

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.081	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.175	23:26
Breath Volume	1.273	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.171	23:30
Breath Volume	1.328	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/08/2021 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HALES,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: RISCACCHE QUIEJ

DATE/TIME OF TEST: 08/01/2021 05:13:04

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.163	05:08
Breath Volume	2.085	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.148	05:11
Breath Volume	1.472	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:13

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BOUND

DATE/TIME OF TEST: 08/03/2021 01:57:22

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.183	01:52
Breath Volume	2.015	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.185	01:55
Breath Volume	2.644	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEPPEL,ZACHARY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: STEDDING JR

DATE/TIME OF TEST: 08/08/2021 01:28:43

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.120	01:22
Breath Volume	1.562	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.125	01:26
Breath Volume	1.820	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: COOK JR

DATE/TIME OF TEST: 08/09/2021 23:38:00

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.081	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.024	23:32
Breath Volume	1.136	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.023	23:36
Breath Volume	1.667	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:38

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASAMIRO
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/12/2021 02:15:21

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.185	02:10
Breath Volume	1.570	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.182	02:13
Breath Volume	2.066	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GILLETTE

DATE/TIME OF TEST: 08/13/2021 03:36:36

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.191	03:30
Breath Volume	1.582	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.181	03:34
Breath Volume	2.019	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: D'HAITI,DWAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 08/14/2021 01:27:25

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.170	01:22
Breath Volume	2.621	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.161	01:25
Breath Volume	1.359	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,JASIEL
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: VACQUEZ

DATE/TIME OF TEST: 08/14/2021 02:28:09

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.260	02:23
Breath Volume	2.203	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.248	02:26
Breath Volume	2.410	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 08/19/2021 01:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MERCADO,J
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: CARLONE

DATE/TIME OF TEST: 08/14/2021 20:05:10

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:53	
Control Test	0.081	19:54	
Air Blank	0.000	19:54	
Subject Sample	# 1 0.043	19:57	Volume Not Met
Breath Volume	0.382	19:57	
Air Blank	0.000	19:58	
Air Blank	0.000	20:00	
Subject Sample	# 2 0.046	20:03	Volume Not Met
Breath Volume	0.054	20:03	
Air Blank	0.000	20:04	
Control Test	0.081	20:04	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:05	

Data Download Date/Time: 08/19/2021 01:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: WARING

DATE/TIME OF TEST: 08/15/2021 00:31:01

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.147	00:26
Breath Volume	2.480	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.153	00:29
Breath Volume	2.351	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:31

Data Download Date/Time: 08/19/2021 01:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SILER

DATE/TIME OF TEST: 08/15/2021 04:04:22

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.158	03:59
Breath Volume	3.078	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.145	04:02
Breath Volume	1.585	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/19/2021 01:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LA'TECE,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALSHAIER,ELIAS
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 08/16/2021 00:18:20

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.203	00:10
Breath Volume	1.589	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.187	00:16
Breath Volume	1.421	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 08/19/2021 01:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/16/2021 03:44:41

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.028	03:40
Breath Volume	1.121	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.028	03:43
Breath Volume	1.234	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 08/19/2021 01:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIS,JASHAUNIE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: ROEHRIG HUBERT

DATE/TIME OF TEST: 08/17/2021 23:30:48

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.081	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.000	23:26
Breath Volume	2.968	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.000	23:29
Breath Volume	3.089	23:29
Air Blank	0.000	23:29
Control Test	0.081	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 08/19/2021 01:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 08/27/2021 23:01:09

Diagnostics Check	OK	22:57	
Air Blank	0.000	22:58	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Subject Sample # 1	0.000	22:59	
Breath Volume	1.285	22:59	
Air Blank		22:59	RFI Detect
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 09/08/2021 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARMSTRONG,BRITTANY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/18/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 08/27/2021 23:11:17

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:06
Breath Volume	1.148	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:09
Breath Volume	1.292	23:09
Air Blank	0.000	23:10
Control Test	0.078	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 09/08/2021 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARCUM,JOSEPH,K

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: MARCUM,JOSEPH

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 07/16/2021

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 03521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 10:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: CECIL

DATE/TIME OF TEST: 08/14/2021 10:40:42

Diagnostics Check	Fail	10:40	Temperature Regulation Test
Air Blank	0.000	10:40	

Data Download Date/Time: 08/21/2021 02:34

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 202303
BREATH TEST OPERATOR: MARCUM,JOSEPH,K	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: MARCUM,JOSEPH	ARREST TIME: 19:27
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 08/21/2021	OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: HIGDON	DATE/TIME OF TEST: 08/24/2021 19:42:25
-----------------------------	--

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:37
Breath Volume	3.578	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.000	19:41
Breath Volume	3.691	19:41
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 09/08/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 202303
BREATH TEST OPERATOR: MARCUM,JOSEPH,K	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: MCMAHON,KENNETH	ARREST TIME: 20:40
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Female
LAST AGENCY INSPECTION: 08/21/2021	OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: NICKERSON	DATE/TIME OF TEST: 08/27/2021	20:57:48
--------------------------------	-------------------------------	----------

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:52	
Subject Sample # 1	0.000	20:52	
Breath Volume	1.382	20:52	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample # 2	0.000	20:55	
Breath Volume	2.343	20:55	
Air Blank	0.000	20:56	
Control Test	0.078	20:56	
Air Blank	0.000	20:57	
Diagnostics Check	Fail	20:57	Analytical Stability Test
Air Blank	0.000	20:57	

Data Download Date/Time: 09/08/2021 17:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 202303
BREATH TEST OPERATOR: MARCUM,JOSEPH,K	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: O'QUINN,KEVIN	ARREST TIME: 16:45
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 08/21/2021	OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: SMITH	DATE/TIME OF TEST: 08/28/2021	17:04:35
----------------------------	-------------------------------	----------

Diagnostics Check	OK	16:58
Air Blank	0.000	16:58
Control Test	0.080	16:58
Air Blank	0.000	16:59
Subject Sample # 1	0.171	16:59
Breath Volume	2.218	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.169	17:02
Breath Volume	2.210	17:02
Air Blank	0.000	17:03
Control Test	0.079	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 09/08/2021 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SISTRUNK,TRENT,D

BREATH TEST OPERATOR AGENCY: FHP Troop B

ARRESTING OFFICER: SISTRUNK,TRENT

ARRESTING OFFICER AGENCY: FHP Troop B

LAST AGENCY INSPECTION: 07/30/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 03521080A3

DRY GAS EXPIRY DATE: 03/05/2023

VIOLATION CODE: DUI

ARREST TIME: 17:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 08/07/2021 17:21:16

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.078	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.201	17:16
Breath Volume	2.851	17:16
Air Blank	0.000	17:16
Air Blank	0.000	17:18
Subject Sample # 2	0.206	17:19
Breath Volume	1.941	17:19
Air Blank	0.000	17:20
Control Test	0.077	17:20
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 09/01/2021 21:36

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 03521080A3
BREATH TEST OPERATOR: PATTERSON,KYLE,Q	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop G	VIOLATION CODE: DUI
ARRESTING OFFICER: PATTERSON,KYLE	ARREST TIME: 19:15
ARRESTING OFFICER AGENCY: FHP Troop G	GENDER: Male
LAST AGENCY INSPECTION: 07/01/2021	OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JOYNER DATE/TIME OF TEST: 08/20/2021 19:27:55

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.082	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.000	19:23
Breath Volume	1.972	19:23
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.000	19:26
Breath Volume	2.812	19:26
Air Blank	0.000	19:27
Control Test	0.083	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 08/30/2021 17:07

Subject Test Electronic Data

DATE/TIME OF TEST: 08/14/2021 19:38:27

Printed: 15 November 2021
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 03521080A3
BREATH TEST OPERATOR: RIVERA,VICTOR,J	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop D	VIOLATION CODE: DUI
ARRESTING OFFICER: RIVERA,VICTOR	ARREST TIME: 03:56
ARRESTING OFFICER AGENCY: FHP Troop D	GENDER: Male
LAST AGENCY INSPECTION: 08/06/2021	OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CORTES GONZALEZ	DATE/TIME OF TEST: 08/22/2021 04:29:19
--------------------------------------	--

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.195	04:24
Breath Volume	3.093	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.188	04:27
Breath Volume	2.625	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 10/13/2021 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICKISS,BRANT,C
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: LICKISS,BRANT
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13921080A2
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 08/13/2021 22:58:38

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.133	22:54
Breath Volume	2.042	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.151	22:57
Breath Volume	3.250	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 09/28/2021 10:54