

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 22:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 09/06/2021 23:02:38

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.077	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.205	22:58
Breath Volume	1.335	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.221	23:01
Breath Volume	2.035	23:01
Air Blank	0.000	23:01
Control Test	0.076	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 09/16/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VAN HOUTEN, JONATHAN, J

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: ROBISON, CASEY

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: ESCALONA MOREY

DATE/TIME OF TEST: 09/07/2021 00:19:17

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.077	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:12	No Sample Provided
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Control Test	0.077	00:18	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 09/16/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: BISHOP,CHRISTOPHER

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 09/07/2021 19:34:16

Diagnostics Check	OK	19:28
Air Blank	0.000	19:28
Control Test	0.077	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.186	19:29
Breath Volume	1.437	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.178	19:32
Breath Volume	1.429	19:32
Air Blank	0.000	19:33
Control Test	0.076	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 09/16/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROBISON,CASEY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: PELLEGRINO

DATE/TIME OF TEST: 09/08/2021 01:15:37

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.077	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:08	No Sample Provided
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2	01:14	No Sample Provided
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.077	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 09/16/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ROBISON,CASEY,M

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: PEREZ,SAMANTHA

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: LOPERENA

DATE/TIME OF TEST: 09/08/2021 19:33:02

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.078	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.233	19:28
Breath Volume	2.726	19:28
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.245	19:31
Breath Volume	2.343	19:31
Air Blank	0.000	19:32
Control Test	0.077	19:32
Air Blank	0.000	19:32
Diagnostics Check	OK	19:33

Data Download Date/Time: 09/16/2021 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCULLOUGH, SEAN, J

BREATH TEST OPERATOR AGENCY: St. Petersburg PD

ARRESTING OFFICER: ROBISON, CASEY

ARRESTING OFFICER AGENCY: St. Petersburg PD

LAST AGENCY INSPECTION: 09/16/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 09/30/2021 00:49:04

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.077	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.000	00:44
Breath Volume	5.425	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:47
Breath Volume	5.195	00:47
Air Blank	0.000	00:48
Control Test	0.078	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:49

Data Download Date/Time: 10/28/2021 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON,ERIC,G

BREATH TEST OPERATOR AGENCY: Satellite Beach PD

ARRESTING OFFICER: YBARRA,AIMEE

ARRESTING OFFICER AGENCY: Satellite Beach PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1327124

DRY GAS EXPIRY DATE: 05/14/2023

VIOLATION CODE: DUI

ARREST TIME: 01:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: FARRELL HANSEN

DATE/TIME OF TEST: 09/04/2021 02:06:18

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.227	02:01
Breath Volume	3.257	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.225	02:04
Breath Volume	3.253	02:04
Air Blank	0.000	02:05
Control Test	0.077	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 09/24/2021 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NEILL,MICHAEL,A

BREATH TEST OPERATOR AGENCY: Brevard County SO

ARRESTING OFFICER: NEILL,MICHAEL

ARRESTING OFFICER AGENCY: Brevard County SO

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1327124

DRY GAS EXPIRY DATE: 05/14/2023

VIOLATION CODE: DUI

ARREST TIME: 19:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/08/2021 21:40:08

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.018	21:35
Breath Volume	1.445	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.016	21:38
Breath Volume	1.816	21:38
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 09/24/2021 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HOUSMAN,JOHN,D

BREATH TEST OPERATOR AGENCY: Satellite Beach PD

ARRESTING OFFICER: WRIGHT,TIMOTHY

ARRESTING OFFICER AGENCY: Satellite Beach PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1327124

DRY GAS EXPIRY DATE: 05/14/2023

VIOLATION CODE: DUI

ARREST TIME: 23:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 09/11/2021 00:02:14

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.139	23:56
Breath Volume	1.488	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 1	0.126	00:00
Breath Volume	1.757	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 09/24/2021 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON,ERIC,G

BREATH TEST OPERATOR AGENCY: Satellite Beach PD

ARRESTING OFFICER: WHITE,CHRISTOPHER

ARRESTING OFFICER AGENCY: Satellite Beach PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1327124

DRY GAS EXPIRY DATE: 05/14/2023

VIOLATION CODE: DUI

ARREST TIME: 21:32

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MATTSON

DATE/TIME OF TEST: 09/13/2021 22:09:50

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.186	22:04
Breath Volume	1.425	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.167	22:08
Breath Volume	1.515	22:08
Air Blank	0.000	22:08
Control Test	0.077	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 09/24/2021 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ANDERSON,ERIC,G

BREATH TEST OPERATOR AGENCY: Satellite Beach PD

ARRESTING OFFICER: WHITE,CHRISTOPHER

ARRESTING OFFICER AGENCY: Satellite Beach PD

LAST AGENCY INSPECTION: 09/03/2021

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1327124

DRY GAS EXPIRY DATE: 05/14/2023

VIOLATION CODE: DUI

ARREST TIME: 00:57

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: IZUMI

DATE/TIME OF TEST: 09/14/2021 01:36:33

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.000	01:31
Breath Volume	2.820	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:35
Breath Volume	3.421	01:35
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 09/24/2021 19:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMALL,BRANNON,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 09/12/2021 18:47:31

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.079	18:42
Air Blank	0.000	18:42
Subject Sample # 1	0.213	18:43
Breath Volume	1.738	18:43
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.217	18:45
Breath Volume	3.109	18:45
Air Blank	0.000	18:46
Control Test	0.078	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 09/15/2021 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: CUCUL TZALAM

DATE/TIME OF TEST: 09/15/2021 23:22:03

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.154	23:17
Breath Volume	2.167	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.157	23:20
Breath Volume	2.050	23:20
Air Blank	0.000	23:21
Control Test	0.076	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:22

Data Download Date/Time: 10/21/2021 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: CUMMINS

DATE/TIME OF TEST: 09/16/2021 18:40:51

Diagnostics Check	OK	18:32
Air Blank	0.000	18:33
Control Test	0.080	18:33
Air Blank	0.000	18:34
Subject Sample # 1	0.109	18:36
Breath Volume	2.667	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:38
Subject Sample # 2	0.106	18:39
Breath Volume	2.167	18:39
Air Blank	0.000	18:39
Control Test	0.078	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 10/21/2021 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO, WESTON, C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RILEY, D
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: DEROSIA

DATE/TIME OF TEST: 09/18/2021 21:25:22

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:19
Air Blank	0.000	21:19
Subject Sample # 1	0.233	21:20
Breath Volume	3.859	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:22
Subject Sample # 2	0.225	21:23
Breath Volume	3.535	21:23
Air Blank	0.000	21:24
Control Test	0.079	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 10/21/2021 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 09/20/2021 13:27:50

Diagnostics Check	OK	13:16	
Air Blank	0.000	13:16	
Control Test	0.079	13:17	
Air Blank	0.000	13:17	
Subject Sample	# 1	13:20	No Sample Provided
Breath Volume	0.000	13:20	
Air Blank	0.000	13:21	
Air Blank	0.000	13:23	
Subject Sample	# 2	13:26	No Sample Provided
Breath Volume	0.000	13:26	
Air Blank	0.000	13:26	
Control Test	0.079	13:27	
Air Blank	0.000	13:27	
Diagnostics Check	OK	13:27	

Data Download Date/Time: 10/21/2021 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDE,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: TISDALE,RACHEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 09:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:52

SUBJECT'S LAST NAME: PORTUGAL

DATE/TIME OF TEST: 09/21/2021 10:19:51

Diagnostics Check	OK	10:13
Air Blank	0.000	10:14
Control Test	0.079	10:14
Air Blank	0.000	10:14
Subject Sample # 1	0.267	10:15
Breath Volume	2.257	10:15
Air Blank	0.000	10:15
Air Blank	0.000	10:17
Subject Sample # 2	0.262	10:18
Breath Volume	1.628	10:18
Air Blank	0.000	10:18
Control Test	0.079	10:19
Air Blank	0.000	10:19
Diagnostics Check	OK	10:19

Data Download Date/Time: 10/21/2021 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 09/26/2021 23:06:00

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.077	23:01
Breath Volume	1.605	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.079	23:04
Breath Volume	1.789	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:06

Data Download Date/Time: 10/21/2021 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENRICH,DANIEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GENRICH,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: MARINELARENA CHAVIRA

DATE/TIME OF TEST: 09/28/2021 16:00:57

Diagnostics Check	OK	15:54
Air Blank	0.000	15:55
Control Test	0.080	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.185	15:56
Breath Volume	3.613	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:58
Subject Sample # 2	0.193	15:59
Breath Volume	3.687	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:00
Diagnostics Check	OK	16:00

Data Download Date/Time: 10/21/2021 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NABULSI,T,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NABULSI,T
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: WAMPLER

DATE/TIME OF TEST: 09/04/2021 19:40:49

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.081	19:33
Air Blank	0.000	19:33
Subject Sample # 1	0.202	19:35
Breath Volume	1.484	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.203	19:39
Breath Volume	1.988	19:39
Air Blank	0.000	19:39
Control Test	0.079	19:40
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 09/09/2021 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 09/09/2021 23:41:14

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.079	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.114	23:36
Breath Volume	2.125	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.110	23:39
Breath Volume	2.683	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 10/13/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PUTNAM

DATE/TIME OF TEST: 09/10/2021 03:15:02

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.274	03:07
Breath Volume	2.910	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.251	03:10
Breath Volume	1.683	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 3	0.240	03:13
Breath Volume	1.156	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 10/13/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/12/2021 02:37:32

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 2	02:36	No Sample Provided
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 10/13/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 09/15/2021 01:39:18

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.221	01:34
Breath Volume	1.601	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.219	01:37
Breath Volume	1.328	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 10/13/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: NILES

DATE/TIME OF TEST: 09/17/2021 04:21:37

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.116	04:17
Breath Volume	2.714	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.114	04:20
Breath Volume	1.156	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 10/13/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/18/2021 03:01:17

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.080	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 10/13/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ULLOA

DATE/TIME OF TEST: 09/19/2021 00:40:04

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.126	00:35
Breath Volume	3.097	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.129	00:38
Breath Volume	3.308	00:38
Air Blank	0.000	00:39
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:40

Data Download Date/Time: 10/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 09/19/2021 06:39:33

Diagnostics Check	OK	06:32
Air Blank	0.000	06:33
Control Test	0.081	06:33
Air Blank	0.000	06:34
Subject Sample # 1	0.117	06:34
Breath Volume	2.656	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.120	06:38
Breath Volume	3.226	06:38
Air Blank	0.000	06:38
Control Test	0.080	06:38
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 10/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: VRONDRAN

DATE/TIME OF TEST: 09/25/2021 01:24:51

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:23	
Control Test	0.081	01:24	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 10/13/2021 16:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RAMIREZ MIRANDA

DATE/TIME OF TEST: 09/26/2021 02:11:25

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.140	02:07
Breath Volume	1.628	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.145	02:09
Breath Volume	1.187	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 10/13/2021 16:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 09/01/2021 19:12:43

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:01	
Control Test	0.080	19:01	
Air Blank	0.000	19:02	
Subject Sample	# 1 0.026	19:05	Volume Not Met
Breath Volume	0.230	19:05	
Air Blank	0.000	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 2 0.026	19:11	Volume Not Met
Breath Volume	0.804	19:11	
Air Blank	0.000	19:11	
Control Test	0.080	19:12	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 09/03/2021 04:06:15

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:05	
Control Test	0.081	04:05	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS, JEREMIAH, S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS, JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 11:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: HERNANDEZ VASQUEZ

DATE/TIME OF TEST: 09/03/2021 12:37:47

Diagnostics Check	OK	12:31
Air Blank	0.000	12:31
Control Test	0.081	12:32
Air Blank	0.000	12:32
Subject Sample # 1	0.259	12:33
Breath Volume	3.496	12:33
Air Blank	0.000	12:33
Air Blank	0.000	12:35
Subject Sample # 2	0.248	12:36
Breath Volume	3.007	12:36
Air Blank	0.000	12:36
Control Test	0.078	12:37
Air Blank	0.000	12:37
Diagnostics Check	OK	12:37

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MARTINEZ CORCINO

DATE/TIME OF TEST: 09/04/2021 03:20:26

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.161	03:15
Breath Volume	1.261	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.150	03:18
Breath Volume	1.648	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 09/04/2021 11:27:29

Diagnostics Check	OK	11:21
Air Blank	0.000	11:21
Control Test	0.079	11:21
Air Blank	0.000	11:22
Subject Sample # 1	0.241	11:22
Breath Volume	2.132	11:22
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.226	11:25
Breath Volume	1.511	11:25
Air Blank	0.000	11:26
Control Test	0.081	11:26
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 09/05/2021 01:44:56

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.164	01:40
Breath Volume	2.625	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.159	01:43
Breath Volume	2.750	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DEE

DATE/TIME OF TEST: 09/05/2021 04:30:46

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.125	04:26
Breath Volume	2.421	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.122	04:29
Breath Volume	2.648	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,SEAN,C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAMBERT,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: BUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 09/05/2021 19:39:54

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.079	19:29	
Air Blank	0.000	19:29	
Subject Sample	# 1	19:32	No Sample Provided
Breath Volume	0.000	19:32	
Air Blank	0.000	19:33	
Air Blank	0.000	19:35	
Subject Sample	# 2	19:38	No Sample Provided
Breath Volume	0.000	19:38	
Air Blank	0.000	19:38	
Control Test	0.080	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:39	

Data Download Date/Time: 09/09/2021 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BARNETT,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: AHRENS

DATE/TIME OF TEST: 09/22/2021 01:24:40

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.081	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.198	01:19
Breath Volume	2.167	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.198	01:23
Breath Volume	1.878	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 10/13/2021 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000751

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDON,JEFFERSON,W
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: OWENS,HUNTER
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 09/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1114319
DRY GAS EXPIRY DATE: 03/27/2022
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 09/25/2021 23:01:25

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.081	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1 0.229	22:51	
Breath Volume	2.785	22:51	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 2 0.193	22:56	Volume Not Met
Breath Volume	0.101	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 3 0.232	22:59	
Breath Volume	1.843	22:59	
Air Blank	0.000	23:00	
Control Test	0.077	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 10/13/2021 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABDALE,JAMES,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BRAMI,AARON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:26

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 09/02/2021 12:07:42

Diagnostics Check	OK	11:57
Air Blank	0.000	11:57
Control Test	0.078	11:58
Air Blank	0.000	11:58
Subject Sample # 1	0.139	12:01
Breath Volume	1.609	12:01
Air Blank	0.000	12:02
Air Blank	0.000	12:04
Subject Sample # 2	0.127	12:06
Breath Volume	2.457	12:06
Air Blank	0.000	12:06
Control Test	0.076	12:07
Air Blank	0.000	12:07
Diagnostics Check	OK	12:07

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ARTICAS ARTICAS

DATE/TIME OF TEST: 09/03/2021 03:48:20

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.078	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.185	03:43
Breath Volume	2.941	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.185	03:46
Breath Volume	3.007	03:46
Air Blank	0.000	03:47
Control Test	0.077	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABDALE,JAMES,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FRICKER,JEFFERY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BOUCARD

DATE/TIME OF TEST: 09/03/2021 22:45:36

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.290	22:41
Breath Volume	3.085	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.280	22:43
Breath Volume	2.449	22:43
Air Blank	0.000	22:44
Control Test	0.077	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,DENZEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALEXANDER,DENZEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RIBBLE 2

DATE/TIME OF TEST: 09/09/2021 01:50:18

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.221	01:42
Breath Volume	3.566	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.192	01:45
Breath Volume	2.359	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 3	0.209	01:48
Breath Volume	1.890	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,JORDI,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BAEZ,JORDI
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MAXEY

DATE/TIME OF TEST: 09/09/2021 03:17:58

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: CARBALLO

DATE/TIME OF TEST: 09/11/2021 21:49:28

Diagnostics Check	OK	21:48	
Air Blank		21:49	RFI Detect
Air Blank	0.000	21:49	

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: CLEMENTS,TOBY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: CARBALLO

DATE/TIME OF TEST: 09/11/2021 21:59:41

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.206	21:53
Breath Volume	1.277	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.214	21:58
Breath Volume	2.175	21:58
Air Blank	0.000	21:58
Control Test	0.079	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCREE,GARRETT,L
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CARROLL,FRANK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CREWS

DATE/TIME OF TEST: 09/14/2021 04:27:21

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.237	04:22
Breath Volume	3.402	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.233	04:25
Breath Volume	2.574	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TYSON,KAIFA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: MAGARY

DATE/TIME OF TEST: 09/15/2021 23:01:22

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 1 0.151	22:54	Volume Not Met
Breath Volume	0.050	22:54	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2 0.217	22:59	Volume Not Met
Breath Volume	0.089	22:59	
Air Blank	0.000	23:00	
Control Test	0.078	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENICK, MARK, T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON, JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: EDELSTEIN

DATE/TIME OF TEST: 09/18/2021 02:54:59

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.387	02:46	No .020 agreement
Breath Volume	1.757	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 2 0.358	02:49	No .020 agreement
Breath Volume	2.187	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 3 0.416	02:53	No .020 agreement
Breath Volume	3.695	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENICK,MARK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: EDELSTEIN

DATE/TIME OF TEST: 09/18/2021 03:05:33

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.078	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1 0.398	03:01	
Breath Volume	2.019	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 2	03:04	RFI Detect
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.078	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENICK, MARK, T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON, JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: EDELSTEIN

DATE/TIME OF TEST: 09/18/2021 03:14:59

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.387	03:10
Breath Volume	2.367	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.400	03:13
Breath Volume	3.746	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARAFA,JOSHUA,D
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ESPINOSA,CLAUDIA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SOLIZ

DATE/TIME OF TEST: 09/18/2021 03:52:13

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.235	03:47
Breath Volume	3.175	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.228	03:50
Breath Volume	3.710	03:50
Air Blank	0.000	03:51
Control Test	0.077	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: IRONS

DATE/TIME OF TEST: 09/18/2021 04:26:53

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.078	04:21	
Air Blank	0.000	04:21	
Subject Sample # 1	0.186	04:22	
Breath Volume	2.160	04:22	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample # 2	0.190	04:25	
Breath Volume	1.781	04:25	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	RFI Detect
Air Blank	0.000	04:26	

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENK,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WENK,ANTHONY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 08/25/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HAMER

DATE/TIME OF TEST: 09/20/2021 03:16:31

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.080	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:15	No Sample Provided
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 09/21/2021 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAKUBSIN,JAMES,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: YAKUBSIN,JAMES
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 09:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 09/24/2021 10:52:07

Diagnostics Check	OK	10:45
Air Blank	0.000	10:46
Control Test	0.080	10:46
Air Blank	0.000	10:46
Subject Sample # 1	0.231	10:47
Breath Volume	3.921	10:47
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.238	10:50
Breath Volume	3.453	10:50
Air Blank	0.000	10:51
Control Test	0.079	10:51
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: CONTRERAS,JAVIER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/25/2021 02:33:23

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.078	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:26	No Sample Provided
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.079	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: MCFARLANE

DATE/TIME OF TEST: 09/25/2021 05:26:35

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.132	05:19
Breath Volume	1.406	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:21
Subject Sample # 2	0.131	05:24
Breath Volume	1.277	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,DENZEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALEXANDER,DENZEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: THEOBALD

DATE/TIME OF TEST: 09/26/2021 01:16:40

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.195	01:12
Breath Volume	2.082	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.194	01:15
Breath Volume	2.105	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: AVENER

DATE/TIME OF TEST: 09/26/2021 02:09:04

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/26/2021 07:53:11

Diagnostics Check	OK	07:49	
Air Blank	0.000	07:49	
Control Test	0.079	07:50	
Air Blank	0.000	07:50	
Subject Sample	# 1 0.181	07:52	
Breath Volume	1.968	07:52	
Air Blank		07:52	RFI Detect
Air Blank		07:52	RFI Detect
Air Blank		07:53	RFI Detect

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/26/2021 08:06:11

Diagnostics Check	OK	07:57
Air Blank	0.000	07:57
Control Test	0.079	07:58
Air Blank	0.000	07:58
Subject Sample # 1	0.182	08:00
Breath Volume	1.261	08:00
Air Blank	0.000	08:01
Air Blank	0.000	08:02
Subject Sample # 2	0.170	08:04
Breath Volume	1.968	08:04
Air Blank	0.000	08:05
Control Test	0.077	08:05
Air Blank	0.000	08:06
Diagnostics Check	OK	08:06

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,JAVON,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: MCCUMBER

DATE/TIME OF TEST: 09/28/2021 07:43:32

Diagnostics Check	OK	07:37
Air Blank	0.000	07:37
Control Test	0.079	07:37
Air Blank	0.000	07:38
Subject Sample # 1	0.051	07:38
Breath Volume	1.386	07:38
Air Blank	0.000	07:39
Air Blank	0.000	07:41
Subject Sample # 2	0.044	07:42
Breath Volume	2.050	07:42
Air Blank	0.000	07:42
Control Test	0.079	07:42
Air Blank	0.000	07:43
Diagnostics Check	OK	07:43

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CLAYTON,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LITZKOW,CLAYTON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2021
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/30/2021 01:31:07

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.079	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 2	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 10/14/2021 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARD,JONATHAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/03/2021 01:50:36

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.082	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUHLBACH,BRIAN,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATEs,AMY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: WADDELL

DATE/TIME OF TEST: 09/04/2021 17:40:02

Diagnostics Check	OK	17:32
Air Blank	0.000	17:32
Control Test	0.081	17:33
Air Blank	0.000	17:33
Subject Sample # 1	0.186	17:35
Breath Volume	1.539	17:35
Air Blank	0.000	17:36
Air Blank	0.000	17:37
Subject Sample # 2	0.192	17:38
Breath Volume	1.503	17:38
Air Blank	0.000	17:39
Control Test	0.080	17:39
Air Blank	0.000	17:39
Diagnostics Check	OK	17:40

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWAIN,ERIC,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SWAIN,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MITCHEM

DATE/TIME OF TEST: 09/05/2021 02:32:29

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.081	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:28	No Sample Provided
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.082	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELTON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 09/06/2021 21:19:29

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.082	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.189	21:14
Breath Volume	2.832	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.195	21:17
Breath Volume	3.257	21:17
Air Blank	0.000	21:18
Control Test	0.081	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ROBERT,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIS,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 09/08/2021 23:36:42

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	3.910	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	4.156	23:35
Air Blank	0.000	23:35
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARTER,LARRY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 09/11/2021 15:37:31

Diagnostics Check	OK	15:29
Air Blank	0.000	15:30
Control Test	0.081	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.281	15:32
Breath Volume	2.476	15:32
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.280	15:35
Breath Volume	1.699	15:35
Air Blank	0.000	15:36
Control Test	0.075	15:36
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: ARRIAGA-ESTRADA

DATE/TIME OF TEST: 09/12/2021 00:17:36

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.081	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: MEADORS

DATE/TIME OF TEST: 09/17/2021 22:49:39

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.184	22:44
Breath Volume	2.269	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.165	22:47
Breath Volume	1.976	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: SHOEMAKE

DATE/TIME OF TEST: 09/18/2021 18:11:48

Diagnostics Check	OK	18:02	
Air Blank	0.000	18:02	
Control Test	0.080	18:03	
Air Blank	0.000	18:03	
Subject Sample	# 1	18:06	No Sample Provided
Breath Volume	0.000	18:06	
Air Blank	0.000	18:07	
Air Blank	0.000	18:09	
Subject Sample	# 2	18:10	Subject Test Refused
Breath Volume	0.000	18:10	
Air Blank	0.000	18:10	
Control Test	0.080	18:11	
Air Blank	0.000	18:11	
Diagnostics Check	OK	18:11	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LONDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK, JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: WOODWARD

DATE/TIME OF TEST: 09/19/2021 18:13:17

Diagnostics Check	OK	18:07	
Air Blank	0.000	18:07	
Control Test	0.081	18:08	
Air Blank	0.000	18:08	
Subject Sample	# 1	18:11	Subject Test Refused
Breath Volume	0.828	18:11	
Air Blank	0.000	18:12	
Control Test	0.081	18:12	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/19/2021 21:30:22

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:26	No Sample Provided
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample	# 2	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:29	
Control Test	0.080	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAINES,BENJAMIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 08/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VEST

DATE/TIME OF TEST: 09/21/2021 00:34:45

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.081	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 09/23/2021 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 09/23/2021 23:45:14

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:37	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Control Test	0.081	23:44	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LICHTY,JAKE,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LICHTY,JAKE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BALLACK

DATE/TIME OF TEST: 09/25/2021 03:14:31

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.151	03:09
Breath Volume	1.230	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.143	03:12
Breath Volume	1.167	03:12
Air Blank	0.000	03:13
Control Test	0.077	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,BENJAMIN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAINES,BENJAMIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 12:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/26/2021 01:48:48

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.111	01:44
Breath Volume	3.171	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.102	01:47
Breath Volume	2.257	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SLOCUMB

DATE/TIME OF TEST: 09/26/2021 04:08:53

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.081	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAURICIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MAURICIO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VALERO-MEDRANDA

DATE/TIME OF TEST: 09/27/2021 01:42:08

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:37	No Sample Provided
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Air Blank	0.000	01:40	
Subject Sample	# 2	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOWERY,BRITTANY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/30/2021 05:42:46

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:40	
Control Test	0.081	05:40	
Air Blank	0.000	05:41	
Subject Sample	# 1	05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:41	
Control Test	0.080	05:42	
Air Blank	0.000	05:42	
Diagnostics Check	OK	05:42	

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: BRYSON

DATE/TIME OF TEST: 09/30/2021 21:01:03

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.081	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 1 0.082	20:53	Volume Not Met
Breath Volume	0.191	20:53	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 2 0.110	20:59	Volume Not Met
Breath Volume	0.343	20:59	
Air Blank	0.000	21:00	
Control Test	0.078	21:00	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSHALL,ALEXANDRIA,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: SLAY

DATE/TIME OF TEST: 09/30/2021 21:12:27

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.112	21:06
Breath Volume	2.007	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.112	21:10
Breath Volume	2.121	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELTON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELTON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/23/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1349789
DRY GAS EXPIRY DATE: 07/15/2023
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 09/30/2021 22:49:28

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.081	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.018	22:45
Breath Volume	3.269	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.020	22:47
Breath Volume	2.980	22:47
Air Blank	0.000	22:48
Control Test	0.081	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 10/14/2021 16:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: LOPEZ GLENN

DATE/TIME OF TEST: 09/08/2021 21:32:11

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.078	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.165	21:27
Breath Volume	1.246	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.166	21:30
Breath Volume	1.656	21:30
Air Blank	0.000	21:31
Control Test	0.077	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 10/31/2021 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRYDEN,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/04/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/17/2021 20:58:46

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:56	
Control Test	0.078	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 1	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:57	
Control Test	0.078	20:58	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 10/31/2021 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER, RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 09/01/2021 16:42:03

Diagnostics Check	OK	16:37	
Air Blank	0.000	16:38	
Control Test	0.079	16:38	
Air Blank	0.000	16:39	
Subject Sample	# 1	16:40	Subject Test Refused
Breath Volume	0.000	16:40	
Air Blank	0.000	16:41	
Control Test	0.080	16:41	
Air Blank	0.000	16:41	
Diagnostics Check	OK	16:42	

Data Download Date/Time: 09/04/2021 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: PINTO,FELIPE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GILGUN

DATE/TIME OF TEST: 09/02/2021 20:55:33

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Control Test	0.080	20:54	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 09/04/2021 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: MASHBURN

DATE/TIME OF TEST: 09/03/2021 19:04:48

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.081	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.000	19:00
Breath Volume	2.695	19:00
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.000	19:03
Breath Volume	1.746	19:03
Air Blank	0.000	19:03
Control Test	0.081	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 09/04/2021 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACHEWICZ,RILEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/05/2021 02:25:20

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.208	02:20
Breath Volume	1.183	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.210	02:23
Breath Volume	1.894	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 10/01/2021 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEISANSAL,HARMON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 210203
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: RANSOM

DATE/TIME OF TEST: 09/06/2021 22:29:21

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:24
Breath Volume	2.738	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.000	22:27
Breath Volume	2.785	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 10/01/2021 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/08/2021 23:32:18

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:27
Breath Volume	1.613	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.000	23:30
Breath Volume	1.324	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 10/01/2021 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/10/2021 04:49:16

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.081	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:40	No Sample Provided
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Air Blank	0.000	04:42	
Subject Sample	# 2	0.171	04:44
Breath Volume	1.507	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample	# 3	0.157	04:47
Breath Volume	1.117	04:47	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 10/01/2021 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMPSON, DAVID, A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SAMPSON, DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: FUSCO

DATE/TIME OF TEST: 09/11/2021 06:51:35

Diagnostics Check	OK	06:44
Air Blank	0.000	06:44
Control Test	0.081	06:45
Air Blank	0.000	06:45
Subject Sample # 1	0.154	06:46
Breath Volume	1.261	06:46
Air Blank	0.000	06:47
Air Blank	0.000	06:49
Subject Sample # 2	0.161	06:49
Breath Volume	1.378	06:49
Air Blank	0.000	06:50
Control Test	0.078	06:50
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYD,J,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOYD,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 09/11/2021 23:01:16

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1 0.188	22:53	Volume Not Met
Breath Volume	0.050	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2 0.169	22:59	Volume Not Met
Breath Volume	0.332	22:59	
Air Blank	0.000	23:00	
Control Test	0.078	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLHARD,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VILLHARD,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 09/12/2021 02:02:01

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.223	01:57
Breath Volume	1.660	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.226	02:00
Breath Volume	2.683	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:02

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 09/13/2021 01:37:53

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.080	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.081	01:37	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, RAYMOND, L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MIDDLETON, BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: MELGAR

DATE/TIME OF TEST: 09/14/2021 16:45:24

Diagnostics Check	OK	16:38
Air Blank	0.000	16:39
Control Test	0.081	16:39
Air Blank	0.000	16:40
Subject Sample # 1	0.000	16:41
Breath Volume	2.859	16:41
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.000	16:44
Breath Volume	2.292	16:44
Air Blank	0.000	16:44
Control Test	0.081	16:44
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,BOBBY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARRISON,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GREGG

DATE/TIME OF TEST: 09/14/2021 21:59:44

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.081	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.209	21:55
Breath Volume	2.460	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.201	21:58
Breath Volume	1.695	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 09/16/2021 00:09:37

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.081	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1 0.000	00:04	
Breath Volume	3.656	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Control Test	0.080	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,JAMISON,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 09/16/2021 22:44:32

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.081	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.000	22:39
Breath Volume	1.113	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.000	22:43
Breath Volume	1.804	22:43
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/17/2021 01:04:50

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.081	00:57	
Air Blank	0.000	00:57	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 2	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 09/17/2021 03:31:24

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.232	03:26
Breath Volume	2.671	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.234	03:29
Breath Volume	1.132	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: KIDWELL

DATE/TIME OF TEST: 09/18/2021 02:47:10

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACHEWICZ,RILEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 09/18/2021 03:57:25

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.180	03:52
Breath Volume	3.847	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.181	03:55
Breath Volume	4.089	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 09/18/2021 05:21:52

Diagnostics Check	OK	05:21	
Air Blank		05:21	RFI Detect
Air Blank	0.000	05:21	

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 09/18/2021 05:30:23

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.081	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.000	05:25
Breath Volume	2.074	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.000	05:29
Breath Volume	1.417	05:29
Air Blank	0.000	05:29
Control Test	0.081	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 09/18/2021 05:52:32

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.080	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.146	05:47
Breath Volume	4.675	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.140	05:50
Breath Volume	4.867	05:50
Air Blank	0.000	05:51
Control Test	0.079	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAUVET,NIKOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 210203
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: STUFF

DATE/TIME OF TEST: 09/19/2021 00:15:47

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.080	00:12	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.081	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 10/01/2021 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 09/19/2021 01:44:17

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.164	01:39
Breath Volume	3.875	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.172	01:42
Breath Volume	3.609	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCKER,JEREMY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DUCKER,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: FINNEGAN

DATE/TIME OF TEST: 09/19/2021 03:38:16

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.214	03:32
Breath Volume	2.617	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.222	03:36
Breath Volume	2.507	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 09/19/2021 04:17:55

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.080	04:07	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2	04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,JAMISON,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LYNN,JAMISON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: DOLE

DATE/TIME OF TEST: 09/22/2021 00:34:28

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.082	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.000	00:29
Breath Volume	2.488	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.000	00:33
Breath Volume	3.617	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEISANSAL,HARMON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: PHILLIPPI

DATE/TIME OF TEST: 09/24/2021 16:33:01

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.080	16:25
Air Blank	0.000	16:26
Subject Sample # 1	0.260	16:28
Breath Volume	1.328	16:28
Air Blank	0.000	16:29
Air Blank	0.000	16:30
Subject Sample # 2	0.279	16:31
Breath Volume	1.609	16:31
Air Blank	0.000	16:32
Control Test	0.075	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:33

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERA,DANTE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CHAMINEAK

DATE/TIME OF TEST: 09/24/2021 23:39:17

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MCKEAN

DATE/TIME OF TEST: 09/25/2021 03:19:47

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.000	03:15
Breath Volume	2.972	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.000	03:18
Breath Volume	2.878	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: WEDGE

DATE/TIME OF TEST: 09/25/2021 17:08:06

Diagnostics Check	OK	17:01
Air Blank	0.000	17:02
Control Test	0.081	17:02
Air Blank	0.000	17:03
Subject Sample # 1	0.000	17:03
Breath Volume	1.277	17:03
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample # 2	0.000	17:06
Breath Volume	1.714	17:06
Air Blank	0.000	17:07
Control Test	0.081	17:07
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSEY,BENJAMIN,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: POSEY,BENJAMIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HARWOOD

DATE/TIME OF TEST: 09/25/2021 20:44:24

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.081	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.161	20:38
Breath Volume	1.906	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.150	20:42
Breath Volume	2.250	20:42
Air Blank	0.000	20:43
Control Test	0.078	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WATERS,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HOUSER

DATE/TIME OF TEST: 09/26/2021 01:55:16

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.081	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:54	
Control Test	0.080	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WATERS,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HOUSER

DATE/TIME OF TEST: 09/26/2021 02:08:47

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.166	02:03
Breath Volume	3.253	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.170	02:07
Breath Volume	2.859	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: TURBINA

DATE/TIME OF TEST: 09/26/2021 22:16:49

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.080	22:04	
Air Blank	0.000	22:04	
Subject Sample	# 1 0.140	22:07	Volume Not Met
Breath Volume	0.234	22:07	
Air Blank	0.000	22:08	
Air Blank	0.000	22:10	
Subject Sample	# 2 0.154	22:12	
Breath Volume	1.406	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 3 0.160	22:15	
Breath Volume	1.972	22:15	
Air Blank	0.000	22:15	
Control Test	0.078	22:16	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERA,DANTE,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 09/27/2021 03:51:30

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.141	03:46
Breath Volume	2.140	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.134	03:49
Breath Volume	1.675	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 210203
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 07:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:34

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 09/27/2021 09:08:17

Diagnostics Check	OK	09:02
Air Blank	0.000	09:02
Control Test	0.080	09:03
Air Blank	0.000	09:03
Subject Sample # 1	0.000	09:04
Breath Volume	3.140	09:04
Air Blank	0.000	09:04
Air Blank	0.000	09:06
Subject Sample # 2	0.000	09:06
Breath Volume	3.601	09:06
Air Blank	0.000	09:07
Control Test	0.079	09:07
Air Blank	0.000	09:08
Diagnostics Check	OK	09:08

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 08/31/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COGBURN

DATE/TIME OF TEST: 09/28/2021 01:06:32

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.000	01:02
Breath Volume	2.828	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.000	01:05
Breath Volume	2.738	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 10/01/2021 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 08/12/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520090A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: ARREGUIN RODRIGUEZ

DATE/TIME OF TEST: 09/04/2021 06:41:13

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.078	06:35
Air Blank	0.000	06:36
Subject Sample # 1	0.175	06:36
Breath Volume	1.503	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.170	06:39
Breath Volume	2.316	06:39
Air Blank	0.000	06:40
Control Test	0.078	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 09/21/2021 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGRETTE,JOAN,P
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: NEGRETTE,JOAN
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 09/21/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 09/24/2021 23:01:01

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.079	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.078	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 10/26/2021 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYCHOFF, COLLIN, T
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: WAYCHOFF, COLLIN
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/02/2021 19:28:59

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.079	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.168	19:24
Breath Volume	3.660	19:24
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.167	19:27
Breath Volume	3.398	19:27
Air Blank	0.000	19:28
Control Test	0.076	19:28
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTEATR,RICHARD,W
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: POTEAT,RICHARD
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/04/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HANCOCK

DATE/TIME OF TEST: 09/04/2021 02:16:58

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.045	02:12
Breath Volume	1.632	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.040	02:15
Breath Volume	1.250	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 09/04/2021 03:58:14

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.131	03:53
Breath Volume	3.472	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.124	03:56
Breath Volume	2.933	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: RIX,ARIANA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 09/07/2021 03:39:31

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.000	03:35
Breath Volume	3.187	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.000	03:38
Breath Volume	2.632	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: STEPHENS,MAURICE
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: DEROCHE

DATE/TIME OF TEST: 09/07/2021 21:21:44

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.078	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.086	21:17
Breath Volume	2.503	21:17
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.088	21:20
Breath Volume	2.421	21:20
Air Blank	0.000	21:20
Control Test	0.078	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO, KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 09/10/2021 16:05:14

Diagnostics Check	OK	16:03	
Air Blank	0.000	16:03	
Control Test	0.079	16:04	
Air Blank	0.000	16:04	RFI Detect
Air Blank	0.000	16:05	

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO, KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 09/10/2021 16:15:28

Diagnostics Check	OK	16:06
Air Blank	0.000	16:07
Control Test	0.078	16:07
Air Blank	0.000	16:07
Subject Sample # 1	0.232	16:09
Breath Volume	1.804	16:09
Air Blank	0.000	16:10
Air Blank	0.000	16:12
Subject Sample # 2	0.213	16:13
Breath Volume	1.765	16:13
Air Blank	0.000	16:14
Control Test	0.078	16:14
Air Blank	0.000	16:15
Diagnostics Check	OK	16:15

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARISI,T,J
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PARISI,T
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KERSNHER

DATE/TIME OF TEST: 09/10/2021 22:17:36

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.117	22:13
Breath Volume	2.003	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.119	22:16
Breath Volume	1.949	22:16
Air Blank	0.000	22:16
Control Test	0.077	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLESBY, MICHAEL, L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO, KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 09/12/2021 12:43:06

Diagnostics Check	OK	12:36
Air Blank	0.000	12:37
Control Test	0.079	12:37
Air Blank	0.000	12:38
Subject Sample # 1	0.000	12:38
Breath Volume	2.273	12:38
Air Blank	0.000	12:39
Air Blank	0.000	12:41
Subject Sample # 2	0.000	12:41
Breath Volume	2.605	12:41
Air Blank	0.000	12:42
Control Test	0.077	12:42
Air Blank	0.000	12:43
Diagnostics Check	OK	12:43

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIX,ARIANA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/15/2021 00:12:44

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.000	00:12	RFI Detect
Air Blank	0.000	00:12	

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAURES,BRANDYN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIX,ARIANA
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/15/2021 00:22:27

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.079	00:16	
Air Blank	0.000	00:16	
Subject Sample # 1	0.000	00:18	
Breath Volume	1.199	00:18	
Air Blank	0.000	00:19	
Air Blank		00:21	RFI Detect
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,JOE,D
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOHNSTON,JOSEPH
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 08/17/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: BUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: JAHASKE

DATE/TIME OF TEST: 09/21/2021 19:26:08

Diagnostics Check	OK	19:19
Air Blank	0.000	19:20
Control Test	0.079	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.234	19:21
Breath Volume	1.761	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:23
Subject Sample # 2	0.225	19:24
Breath Volume	3.386	19:24
Air Blank	0.000	19:25
Control Test	0.078	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SINGLETARY,MATHEW
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/24/2021 07:39:07

Diagnostics Check	OK	07:32
Air Blank	0.000	07:33
Control Test	0.080	07:33
Air Blank	0.000	07:33
Subject Sample # 1	0.212	07:34
Breath Volume	1.648	07:34
Air Blank	0.000	07:35
Air Blank	0.000	07:37
Subject Sample # 2	0.225	07:37
Breath Volume	1.367	07:37
Air Blank	0.000	07:38
Control Test	0.079	07:38
Air Blank	0.000	07:39
Diagnostics Check	OK	07:39

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASTOR,JOSEPH,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KASTOR,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 09/25/2021 03:08:46

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.190	03:04
Breath Volume	2.007	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.184	03:07
Breath Volume	2.152	03:07
Air Blank	0.000	03:07
Control Test	0.075	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 10/15/2021 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YASUDA,SHOHEI,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YASUDA,SHOHEI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: LIGHTSEY

DATE/TIME OF TEST: 09/12/2021 08:24:08

Diagnostics Check	OK	08:17
Air Blank	0.000	08:18
Control Test	0.081	08:18
Air Blank	0.000	08:18
Subject Sample # 1	0.085	08:19
Breath Volume	1.980	08:19
Air Blank	0.000	08:20
Air Blank	0.000	08:21
Subject Sample # 2	0.090	08:22
Breath Volume	3.003	08:22
Air Blank	0.000	08:23
Control Test	0.081	08:23
Air Blank	0.000	08:24
Diagnostics Check	OK	08:24

Data Download Date/Time: 10/04/2021 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: WIGLEY

DATE/TIME OF TEST: 09/19/2021 23:39:15

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.202	23:34
Breath Volume	3.753	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.200	23:37
Breath Volume	3.339	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 10/04/2021 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 09/21/2021 03:32:00

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.082	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.082	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 10/04/2021 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUSTER,CORY,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GUSTER,CORY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/22/2021 18:32:18

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.081	18:27
Air Blank	0.000	18:27
Subject Sample # 1	0.011	18:28
Breath Volume	2.347	18:28
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.010	18:30
Breath Volume	2.484	18:30
Air Blank	0.000	18:31
Control Test	0.080	18:31
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 10/04/2021 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1076482
DRY GAS EXPIRY DATE: 01/15/2022
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: PADELSKY

DATE/TIME OF TEST: 09/25/2021 01:55:18

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.154	01:49
Breath Volume	4.070	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.155	01:53
Breath Volume	3.605	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 10/04/2021 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMS,KYLE,N
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: ERLEMAN, JOSHUA
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 08/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: SABO

DATE/TIME OF TEST: 09/01/2021 00:36:46

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.079	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:28	No Sample Provided
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2	00:32	
Breath Volume	2.382	00:32	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 3	00:35	
Breath Volume	2.460	00:35	
Air Blank	0.000	00:35	
Control Test	0.079	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 09/02/2021 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SHAFFER,ROBIN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: DELAROSA

DATE/TIME OF TEST: 09/04/2021 17:34:35

Diagnostics Check	OK	17:29	
Air Blank	0.000	17:29	
Control Test	0.079	17:30	
Air Blank	0.000	17:30	
Subject Sample	# 1	17:33	Subject Test Refused
Breath Volume	0.746	17:33	
Air Blank	0.000	17:33	
Control Test	0.082	17:33	
Air Blank	0.000	17:34	
Diagnostics Check	OK	17:34	

Data Download Date/Time: 10/24/2021 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHANKLE,CHASE,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BAKER,WESLEY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: BRAND

DATE/TIME OF TEST: 09/05/2021 20:31:54

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.078	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.108	20:27
Breath Volume	1.976	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:29
Subject Sample # 2	0.101	20:30
Breath Volume	1.355	20:30
Air Blank	0.000	20:30
Control Test	0.078	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 10/24/2021 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHANKLE,CHASE,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BAKER,WESLEY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: BRAND

DATE/TIME OF TEST: 09/05/2021 20:42:16

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.114	20:37
Breath Volume	2.640	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.099	20:40
Breath Volume	1.601	20:40
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 10/24/2021 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHANKLE,CHASE,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SHANKLE,CHASE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 09/06/2021 00:24:57

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.193	00:20
Breath Volume	2.304	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:22
Subject Sample # 2	0.200	00:23
Breath Volume	1.402	00:23
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 10/24/2021 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHANKLE,CHASE,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WATSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: DARAGO

DATE/TIME OF TEST: 09/18/2021 20:01:34

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.080	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.146	19:57
Breath Volume	2.632	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.146	20:00
Breath Volume	2.804	20:00
Air Blank	0.000	20:00
Control Test	0.080	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 10/24/2021 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WYMAN

DATE/TIME OF TEST: 09/20/2021 23:02:16

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.150	22:57
Breath Volume	2.105	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.168	23:00
Breath Volume	3.710	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 10/24/2021 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ROSAS

DATE/TIME OF TEST: 09/24/2021 03:04:05

Diagnostics Check	OK	03:03	
Air Blank		03:03	RFI Detect
Air Blank	0.000	03:04	

Data Download Date/Time: 10/24/2021 16:12

Subject Test Electronic Data

DATE/TIME OF TEST: 09/24/2021 03:12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MENDEZ,JORGE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: ABT

DATE/TIME OF TEST: 09/03/2021 05:08:02

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.191	05:02
Breath Volume	1.437	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.202	05:06
Breath Volume	1.750	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:08

Data Download Date/Time: 09/29/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: SUTTERFIELD

DATE/TIME OF TEST: 09/05/2021 17:44:52

Diagnostics Check	OK	17:38
Air Blank	0.000	17:39
Control Test	0.081	17:39
Air Blank	0.000	17:40
Subject Sample # 1	0.246	17:40
Breath Volume	3.019	17:40
Air Blank	0.000	17:41
Air Blank	0.000	17:42
Subject Sample # 2	0.247	17:43
Breath Volume	3.000	17:43
Air Blank	0.000	17:43
Control Test	0.080	17:44
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 09/29/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ-VELASQUEZ,MIGUEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: COLGAN,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PEREZ-VELASQUEZ

DATE/TIME OF TEST: 09/07/2021 22:46:28

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:42
Breath Volume	1.878	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.000	22:45
Breath Volume	1.292	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 09/29/2021 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/08/2021 03:18:40

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 09/29/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: CASTILLO LOPEZ

DATE/TIME OF TEST: 09/12/2021 22:30:07

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.211	22:25
Breath Volume	2.855	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.213	22:28
Breath Volume	1.863	22:28
Air Blank	0.000	22:29
Control Test	0.078	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 09/29/2021 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: COLGAN,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: WHITEHOUSE

DATE/TIME OF TEST: 09/15/2021 21:26:30

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.000	21:21
Breath Volume	2.191	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.000	21:25
Breath Volume	3.941	21:25
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: GILKEY,JOSHUA
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 09/19/2021 05:07:37

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.144	05:03
Breath Volume	1.460	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.139	05:06
Breath Volume	1.273	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTONUCCI,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ANTONUCCI JR

DATE/TIME OF TEST: 09/19/2021 05:50:50

Diagnostics Check	OK	05:48	
Air Blank	0.000	05:48	
Control Test	0.080	05:48	
Air Blank	0.000	05:49	
Subject Sample	# 1	05:49	Subject Test Refused
Breath Volume	0.000	05:49	
Air Blank	0.000	05:49	
Control Test	0.080	05:50	
Air Blank	0.000	05:50	
Diagnostics Check	OK	05:50	

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRIQUEZ ZUNIGA,MARCIO,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: HENRIQUEZ ZUNIGA,MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ANTONUCCI JR

DATE/TIME OF TEST: 09/19/2021 06:00:42

Diagnostics Check	OK	05:53
Air Blank	0.000	05:54
Control Test	0.080	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.000	05:56
Breath Volume	2.359	05:56
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.000	05:59
Breath Volume	1.832	05:59
Air Blank	0.000	05:59
Control Test	0.079	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRCHNER,MATTHEW,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BROCATO,S
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 09/20/2021 04:58:36

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.080	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.159	04:54
Breath Volume	1.707	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.156	04:57
Breath Volume	2.019	04:57
Air Blank	0.000	04:57
Control Test	0.076	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMB,HUNTER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LIMB,HUNTER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROYLE

DATE/TIME OF TEST: 09/27/2021 03:38:50

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.167	03:34
Breath Volume	2.234	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.165	03:37
Breath Volume	2.191	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BLAKE,XAVIAN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 09/28/2021 02:04:16

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.036	01:57
Breath Volume	1.167	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.032	02:02
Breath Volume	1.167	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYFORD,GREGG,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RODRIGUEZ,ANDRES
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 09/28/2021 17:01:14

Diagnostics Check	OK	16:54
Air Blank	0.000	16:55
Control Test	0.080	16:55
Air Blank	0.000	16:56
Subject Sample # 1	0.000	16:56
Breath Volume	3.468	16:56
Air Blank	0.000	16:57
Air Blank	0.000	16:59
Subject Sample # 2	0.000	16:59
Breath Volume	3.078	16:59
Air Blank	0.000	17:00
Control Test	0.080	17:00
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 09/29/2021 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIPFORD,JASON,V
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LIPFORD,JASON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/30/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: BUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: HOPPER

DATE/TIME OF TEST: 09/11/2021 20:09:55

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.078	20:04
Air Blank	0.000	20:04
Subject Sample # 1	0.176	20:05
Breath Volume	1.929	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:07
Subject Sample # 2	0.174	20:08
Breath Volume	1.816	20:08
Air Blank	0.000	20:09
Control Test	0.077	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 10/03/2021 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ALTHAUS

DATE/TIME OF TEST: 09/01/2021 03:22:12

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.175	03:17
Breath Volume	1.894	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.180	03:20
Breath Volume	1.773	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: IRAHETA RIVERA

DATE/TIME OF TEST: 09/04/2021 03:04:34

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.103	02:59
Breath Volume	1.472	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.105	03:03
Breath Volume	1.890	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/04/2021 04:36:54

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.081	04:23	
Air Blank	0.000	04:23	
Subject Sample	# 1 0.197	04:26	Volume Not Met
Breath Volume	0.648	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2 0.252	04:32	
Breath Volume	1.660	04:32	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 3 0.249	04:35	
Breath Volume	1.847	04:35	
Air Blank	0.000	04:35	
Control Test	0.078	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 09/04/2021 05:56:45

Diagnostics Check	OK	05:45	
Air Blank	0.000	05:45	
Control Test	0.080	05:45	
Air Blank	0.000	05:46	
Subject Sample	# 1 0.186	05:49	Volume Not Met
Breath Volume	0.000	05:49	
Air Blank	0.000	05:50	
Air Blank	0.000	05:52	
Subject Sample	# 2 0.237	05:55	Volume Not Met
Breath Volume	0.031	05:55	
Air Blank	0.000	05:55	
Control Test	0.080	05:56	
Air Blank	0.000	05:56	
Diagnostics Check	OK	05:56	

Data Download Date/Time: 09/10/2021 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEQUIGNOT,KYLIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 09/04/2021 06:09:36

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:58	
Control Test	0.079	05:58	
Air Blank	0.000	05:59	
Subject Sample	# 1 0.228	06:02	Volume Not Met
Breath Volume	0.378	06:02	
Air Blank	0.000	06:03	
Air Blank	0.000	06:04	
Subject Sample	# 2 0.252	06:07	Volume Not Met
Breath Volume	0.082	06:07	
Air Blank	0.000	06:08	
Control Test	0.079	06:08	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,ANIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 09/05/2021 03:06:44

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY, ANIA, L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 09/05/2021 03:17:26

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.130	03:12
Breath Volume	1.238	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.137	03:15
Breath Volume	1.718	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HAWK

DATE/TIME OF TEST: 09/05/2021 04:39:44

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.180	04:34
Breath Volume	2.785	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.178	04:38
Breath Volume	2.398	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: SARBACKER

DATE/TIME OF TEST: 09/06/2021 00:59:46

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.121	00:54
Breath Volume	2.042	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.115	00:58
Breath Volume	2.816	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: COY CAC

DATE/TIME OF TEST: 09/06/2021 03:08:48

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.174	03:04
Breath Volume	2.867	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.174	03:07
Breath Volume	1.664	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 09/07/2021 01:35:23

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.141	01:30
Breath Volume	3.304	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.138	01:33
Breath Volume	3.023	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/11/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 09/07/2021 02:13:44

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.082	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.207	02:09
Breath Volume	1.531	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.208	02:12
Breath Volume	1.511	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 09/10/2021 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PICKNELL

DATE/TIME OF TEST: 09/05/2021 02:49:12

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.194	02:44
Breath Volume	1.687	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.203	02:47
Breath Volume	1.968	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 09/22/2021 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/09/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22419080A3
DRY GAS EXPIRY DATE: 10/05/2021
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 09/18/2021 06:50:26

Diagnostics Check	OK	06:39	
Air Blank	0.000	06:39	
Control Test	0.078	06:39	
Air Blank	0.000	06:40	
Subject Sample	# 1	06:43	No Sample Provided
Breath Volume	0.000	06:43	
Air Blank	0.000	06:43	
Air Blank	0.000	06:45	
Subject Sample	# 2	06:49	No Sample Provided
Breath Volume	0.000	06:49	
Air Blank	0.000	06:49	
Control Test	0.079	06:49	
Air Blank	0.000	06:50	
Diagnostics Check	OK	06:50	

Data Download Date/Time: 09/22/2021 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES, JONATHAN, K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: BARNETTE

DATE/TIME OF TEST: 09/23/2021 21:16:07

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.026	21:11
Breath Volume	2.359	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.025	21:14
Breath Volume	2.289	21:14
Air Blank	0.000	21:15
Control Test	0.078	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 10/06/2021 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 09/26/2021 02:14:56

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.171	02:10
Breath Volume	2.730	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.186	02:13
Breath Volume	2.054	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 10/06/2021 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: UNDERHILL

DATE/TIME OF TEST: 09/28/2021 04:56:19

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.064	04:51
Breath Volume	2.253	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.069	04:54
Breath Volume	2.265	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 10/06/2021 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS, JAKE, M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: QUINLAN, CHARLES
ARRESTING OFFICER AGENCY: Dixie County SO
LAST AGENCY INSPECTION: 09/22/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 09/29/2021 21:16:20

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.175	21:11
Breath Volume	2.546	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.183	21:14
Breath Volume	2.496	21:14
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 10/06/2021 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BURGOS OCASIO

DATE/TIME OF TEST: 09/03/2021 03:27:04

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.137	03:19	Volume Not Met
Breath Volume	0.031	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2 0.124	03:25	Volume Not Met
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.076	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PETERS,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: FULMER

DATE/TIME OF TEST: 09/03/2021 22:36:12

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.080	22:28	
Air Blank	0.000	22:29	
Subject Sample # 1	0.101	22:30	
Breath Volume	1.406	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:32	
Subject Sample # 2	0.100	22:34	
Breath Volume	1.734	22:34	
Air Blank	0.000	22:34	
Control Test	0.077	22:35	
Air Blank	0.000	22:35	
Diagnostics Check	Fail	22:35	Temperature Regulation Test
Air Blank	0.000	22:36	

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WIERTSEMA

DATE/TIME OF TEST: 09/05/2021 03:52:36

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.243	03:47
Breath Volume	2.535	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.224	03:50
Breath Volume	1.507	03:50
Air Blank	0.000	03:51
Control Test	0.077	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT, JOSHUA, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 09/12/2021 04:47:10

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.098	04:42
Breath Volume	1.621	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.089	04:45
Breath Volume	1.398	04:45
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 09/12/2021 23:03:21

Diagnostics Check	OK	23:03	
Air Blank		23:03	RFI Detect
Air Blank		23:03	RFI Detect

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 09/12/2021 23:07:47

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.080	23:06	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:07	RFI Detect
Breath Volume	0.000	23:07	
Air Blank		23:07	RFI Detect
Air Blank		23:07	RFI Detect

Data Download Date/Time: 09/16/2021 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 08/14/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 09/12/2021 23:20:55

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.203	23:15
Breath Volume	1.863	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.207	23:19
Breath Volume	1.949	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 09/15/2021 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WEBER,MATTHEW
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: RIIDLE

DATE/TIME OF TEST: 09/18/2021 01:36:59

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.078	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.150	01:31
Breath Volume	1.535	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.151	01:35
Breath Volume	1.550	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 10/21/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JONATHON,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,CHELSEA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: KROM

DATE/TIME OF TEST: 09/19/2021 00:42:32

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.066	00:37
Breath Volume	2.113	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.064	00:40
Breath Volume	1.718	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 10/21/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER, COLLEEN, E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: RAMOS, DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: KELLS

DATE/TIME OF TEST: 09/21/2021 07:55:15

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.078	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.078	07:48
Breath Volume	1.121	07:48
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.075	07:53
Breath Volume	1.453	07:53
Air Blank	0.000	07:54
Control Test	0.079	07:54
Air Blank	0.000	07:55
Diagnostics Check	OK	07:55

Data Download Date/Time: 10/21/2021 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD, HEATHER, M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MOONEY, JAMES
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:24

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 09/21/2021 12:48:53

Diagnostics Check	OK	12:44	
Air Blank	0.000	12:44	
Control Test	0.079	12:45	
Air Blank	0.000	12:45	
Subject Sample	# 1	12:47	Subject Test Refused
Breath Volume	0.000	12:47	
Air Blank	0.000	12:47	
Control Test	0.079	12:48	
Air Blank	0.000	12:48	
Diagnostics Check	OK	12:48	

Data Download Date/Time: 10/21/2021 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/24/2021 23:25:42

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.079	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.000	23:17	Volume Not Met
Breath Volume	0.398	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 2 0.000	23:21	
Breath Volume	1.605	23:21	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample	# 3 0.000	23:24	
Breath Volume	2.007	23:24	
Air Blank	0.000	23:24	
Control Test	0.080	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 10/21/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 09/25/2021 21:56:49

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:54	
Control Test	0.079	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 1	21:55	Subject Test Refused
Breath Volume	0.000	21:55	
Air Blank	0.000	21:55	
Control Test	0.079	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 10/21/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CALDWELL,JOSHUA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: PABON-VEGA

DATE/TIME OF TEST: 09/26/2021 17:04:24

Diagnostics Check	OK	16:56
Air Blank	0.000	16:57
Control Test	0.079	16:57
Air Blank	0.000	16:58
Subject Sample # 1	0.246	16:58
Breath Volume	1.566	16:58
Air Blank	0.000	16:59
Air Blank	0.000	17:01
Subject Sample # 2	0.257	17:02
Breath Volume	1.859	17:02
Air Blank	0.000	17:03
Control Test	0.079	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 10/21/2021 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ELLIOTT,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: CRISPINO

DATE/TIME OF TEST: 09/04/2021 23:16:46

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:03	
Subject Sample	# 1 0.139	23:07	Slope Not Level
Breath Volume	1.234	23:07	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 2 0.150	23:12	
Breath Volume	1.628	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 3 0.157	23:16	
Breath Volume	2.207	23:16	
Air Blank		23:16	RFI Detect
Air Blank	0.000	23:16	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAVIS, D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HENSHAW, DONALD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 09/05/2021 04:44:48

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.127	04:38
Breath Volume	1.605	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.128	04:43
Breath Volume	1.789	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,TRAVIS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HARM,CONNER
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ROEBUCK

DATE/TIME OF TEST: 09/06/2021 02:13:02

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.077	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.078	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAVIS, D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SMITH, DONALD
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 09/06/2021 04:14:32

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.274	04:09
Breath Volume	1.457	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.279	04:12
Breath Volume	1.886	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCASTER,CODY,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 09/12/2021 01:55:39

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.078	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALE,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RAMOS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/10/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 09/16/2021 23:37:25

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.078	23:29	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.121	23:33	Volume Not Met
Breath Volume	0.089	23:33	
Air Blank	0.000	23:34	
Air Blank	0.000	23:35	
Subject Sample	# 2	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.077	23:36	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 09/24/2021 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PEDRO,L
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROBINSON,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 09/01/2021 20:29:18

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:26	
Control Test	0.077	20:26	
Air Blank	0.000	20:27	
Subject Sample	# 1	20:27	Subject Test Refused
Breath Volume	0.000	20:27	
Air Blank	0.000	20:28	
Control Test	0.077	20:28	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 09/24/2021 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A2
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: PINKSTON

DATE/TIME OF TEST: 09/02/2021 14:29:47

Diagnostics Check	OK	14:21
Air Blank	0.000	14:22
Control Test	0.078	14:22
Air Blank	0.000	14:22
Subject Sample # 1	0.136	14:24
Breath Volume	1.140	14:24
Air Blank	0.000	14:25
Air Blank	0.000	14:26
Subject Sample # 2	0.143	14:28
Breath Volume	1.246	14:28
Air Blank	0.000	14:28
Control Test	0.077	14:29
Air Blank	0.000	14:29
Diagnostics Check	OK	14:29

Data Download Date/Time: 09/24/2021 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLEDSOE,JACOB,B
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: PEARL,TEDDY
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 09/05/2021 17:49:57

Diagnostics Check	OK	17:46	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Subject Sample	# 1	17:48	Subject Test Refused
Breath Volume	0.000	17:48	
Air Blank	0.000	17:49	
Control Test	0.080	17:49	
Air Blank	0.000	17:49	
Diagnostics Check	OK	17:49	

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN, GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/08/2021 17:17:44

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.079	17:12
Air Blank	0.000	17:12
Subject Sample # 1	0.214	17:13
Breath Volume	2.933	17:13
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample # 2	0.215	17:16
Breath Volume	3.121	17:16
Air Blank	0.000	17:16
Control Test	0.079	17:17
Air Blank	0.000	17:17
Diagnostics Check	OK	17:17

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 09/08/2021 23:58:25

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.196	23:52
Breath Volume	1.382	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.191	23:56
Breath Volume	1.144	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS, DOUG, J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: DANIELS, D
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 09/09/2021 23:32:14

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.146	23:27
Breath Volume	1.441	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.155	23:30
Breath Volume	2.386	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUBI,UZIEL,Y
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BUEHLER,ANTHONY
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 09/10/2021 21:59:15

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:56	
Control Test	0.079	21:57	
Air Blank	0.000	21:57	
Subject Sample	# 1	21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Control Test	0.079	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: HELMS,J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 09/11/2021 09:02:26

Diagnostics Check	OK	08:56
Air Blank	0.000	08:56
Control Test	0.079	08:56
Air Blank	0.000	08:57
Subject Sample # 1	0.113	08:57
Breath Volume	3.074	08:57
Air Blank	0.000	08:58
Air Blank	0.000	09:00
Subject Sample # 2	0.110	09:00
Breath Volume	2.437	09:00
Air Blank	0.000	09:01
Control Test	0.080	09:01
Air Blank	0.000	09:02
Diagnostics Check	OK	09:02

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN, G
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: DAVIES

DATE/TIME OF TEST: 09/13/2021 22:57:39

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.074	22:51
Breath Volume	1.515	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.076	22:56
Breath Volume	1.367	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,3
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PARRY,R
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: PINA

DATE/TIME OF TEST: 09/14/2021 01:02:07

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:53	No Sample Provided
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Air Blank	0.000	00:55	
Subject Sample	# 2	00:57	
Breath Volume	1.902	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 3	01:00	
Breath Volume	1.582	01:00	
Air Blank	0.000	01:01	
Control Test	0.081	01:01	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, WILLIAM, S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCKENZIE, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 09/15/2021 07:58:35

Diagnostics Check	OK	07:52
Air Blank	0.000	07:52
Control Test	0.079	07:52
Air Blank	0.000	07:53
Subject Sample # 1	0.201	07:53
Breath Volume	3.601	07:53
Air Blank	0.000	07:54
Air Blank	0.000	07:56
Subject Sample # 2	0.206	07:57
Breath Volume	3.242	07:57
Air Blank	0.000	07:57
Control Test	0.080	07:57
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN, GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: COFFEE

DATE/TIME OF TEST: 09/19/2021 01:07:13

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.232	01:02
Breath Volume	3.207	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.230	01:05
Breath Volume	3.292	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 09/29/2021 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,MIGUEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROBLES,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/19/2021 01:21:25

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	RFI Detect
Breath Volume	0.183	01:19	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 09/29/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,MIGUEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROBLES,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/19/2021 01:33:17

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.138	01:27
Breath Volume	1.542	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.118	01:31
Breath Volume	1.390	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 09/29/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN, GERALD, E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: KUCZKA, J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CHYNOWETH

DATE/TIME OF TEST: 09/19/2021 01:46:59

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.078	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.174	01:42
Breath Volume	2.300	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.182	01:45
Breath Volume	2.421	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 09/29/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLAP,JASON,N
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: MORGAN,JOSEPH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ORIOLI

DATE/TIME OF TEST: 09/21/2021 04:19:06

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.202	04:14
Breath Volume	1.812	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.214	04:17
Breath Volume	2.437	04:17
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 09/29/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE, KRISTOPHER
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: MACE

DATE/TIME OF TEST: 09/23/2021 20:03:34

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:59	
Control Test	0.078	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:02	
Control Test	0.080	20:02	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:03	

Data Download Date/Time: 09/29/2021 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: GUY,KEVIN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CARNLEY

DATE/TIME OF TEST: 09/26/2021 04:02:37

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.081	03:57
Breath Volume	1.679	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.079	04:01
Breath Volume	2.898	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 09/29/2021 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONK, ADAM,
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: FRANZESE, J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/29/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: RUGER

DATE/TIME OF TEST: 09/30/2021 20:08:07

Diagnostics Check	OK	20:05	
Air Blank	0.000	20:05	
Control Test	0.079	20:06	
Air Blank	0.000	20:06	
Subject Sample	# 1	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:07	
Control Test	0.079	20:07	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 10/27/2021 10:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 09/02/2021 03:16:32

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.078	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 10/03/2021 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALSH,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: NAVARRO,D
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: FACIO GARCIA

DATE/TIME OF TEST: 09/04/2021 22:15:30

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.182	22:10
Breath Volume	1.964	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.179	22:13
Breath Volume	3.320	22:13
Air Blank	0.000	22:14
Control Test	0.078	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 10/03/2021 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: HEIDT,BRAD
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0720080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/11/2021 22:57:27

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:53
Breath Volume	2.015	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:56
Breath Volume	2.390	22:56
Air Blank	0.000	22:56
Control Test	0.077	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 10/03/2021 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MENACHO,NICK
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 08/27/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: HEROLD

DATE/TIME OF TEST: 09/26/2021 04:31:15

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.028	04:26
Breath Volume	1.875	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.023	04:29
Breath Volume	1.125	04:29
Air Blank	0.000	04:30
Control Test	0.077	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 10/03/2021 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FERRARI

DATE/TIME OF TEST: 09/24/2021 00:32:23

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	Interferent Detect
Air Blank		00:32	RFI Detect

Data Download Date/Time: 10/07/2021 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: FERRARI

DATE/TIME OF TEST: 09/24/2021 00:50:41

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.209	00:46
Breath Volume	2.332	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.201	00:49
Breath Volume	2.214	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 10/07/2021 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JUAN,A
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: GARCIA,JUAN
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TERLITZ

DATE/TIME OF TEST: 09/28/2021 03:57:52

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.225	03:53
Breath Volume	3.597	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.220	03:56
Breath Volume	3.683	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 10/07/2021 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KNOOP,KENNETH
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BLANEY

DATE/TIME OF TEST: 09/01/2021 02:24:29

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.246	02:19
Breath Volume	3.214	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.241	02:22
Breath Volume	2.023	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 09/09/2021 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGLIA,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BRIGLIA,NICHOLAS
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: GUSSE

DATE/TIME OF TEST: 09/03/2021 19:18:44

Diagnostics Check	OK	19:11
Air Blank	0.000	19:12
Control Test	0.080	19:12
Air Blank	0.000	19:13
Subject Sample # 1	0.181	19:13
Breath Volume	1.871	19:13
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.174	19:17
Breath Volume	2.589	19:17
Air Blank	0.000	19:17
Control Test	0.078	19:18
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 09/09/2021 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARROLL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/19/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 09/04/2021 19:10:28

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:07	
Control Test	0.080	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:09	Subject Test Refused
Breath Volume	0.000	19:09	
Air Blank	0.000	19:09	
Control Test	0.079	19:09	
Air Blank	0.000	19:10	
Diagnostics Check	OK	19:10	

Data Download Date/Time: 09/09/2021 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: GERLT

DATE/TIME OF TEST: 09/10/2021 20:13:31

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:57	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Subject Sample	# 1 0.021	20:00	
Breath Volume	1.179	20:00	
Air Blank	0.000	20:01	
Air Blank	0.000	20:03	
Subject Sample	# 2 0.020	20:06	Volume Not Met
Breath Volume	0.222	20:06	
Air Blank	0.000	20:06	
Air Blank	0.000	20:08	
Subject Sample	# 3 0.018	20:12	Volume Not Met
Breath Volume	0.238	20:12	
Air Blank	0.000	20:12	
Control Test	0.078	20:12	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:13	

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: GERLT

DATE/TIME OF TEST: 09/10/2021 20:40:20

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.079	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.011	20:33
Breath Volume	1.898	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.011	20:38
Breath Volume	1.730	20:38
Air Blank	0.000	20:39
Control Test	0.079	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: RAMIREZ-CHAVEZ

DATE/TIME OF TEST: 09/10/2021 23:29:15

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.219	23:23
Breath Volume	1.816	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.214	23:27
Breath Volume	2.507	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRAVO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MARTINEZ-MURILLO

DATE/TIME OF TEST: 09/12/2021 04:16:40

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRAVO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MARTINEZ-MURILLO

DATE/TIME OF TEST: 09/12/2021 04:28:51

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.168	04:24
Breath Volume	2.632	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample # 2	0.165	04:27
Breath Volume	1.785	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEDER,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: REEDER,MATTHEW
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: COTTRELL

DATE/TIME OF TEST: 09/14/2021 02:23:42

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.171	02:18
Breath Volume	2.562	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.174	02:22
Breath Volume	2.332	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,T
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 09/17/2021 00:24:15

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:19
Breath Volume	1.476	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:22
Breath Volume	1.652	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DACEY,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: DACEY,MICHAEL
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 09/18/2021 03:41:26

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.158	03:36
Breath Volume	2.691	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.159	03:39
Breath Volume	1.796	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 09/19/2021 15:37:56

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.080	15:32
Air Blank	0.000	15:32
Subject Sample # 1	0.000	15:33
Breath Volume	1.503	15:33
Air Blank	0.000	15:33
Air Blank	0.000	15:36
Subject Sample # 2	0.000	15:36
Breath Volume	1.566	15:36
Air Blank	0.000	15:37
Control Test	0.080	15:37
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 09/19/2021 16:02:50

Diagnostics Check	OK	15:56
Air Blank	0.000	15:56
Control Test	0.080	15:57
Air Blank	0.000	15:57
Subject Sample # 1	0.000	15:58
Breath Volume	1.339	15:58
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.000	16:01
Breath Volume	1.507	16:01
Air Blank	0.000	16:01
Control Test	0.080	16:02
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PEREZ-GASPAR

DATE/TIME OF TEST: 09/21/2021 00:16:35

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.198	00:11
Breath Volume	2.093	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.200	00:15
Breath Volume	2.601	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 10/07/2021 06:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,TREVOR,N
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/08/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 09/27/2021 20:37:12

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.075	20:25	
Air Blank	0.000	20:26	
Subject Sample	# 1 0.000	20:29	Volume Not Met
Breath Volume	0.945	20:29	
Air Blank	0.000	20:29	
Air Blank	0.000	20:31	
Subject Sample	# 2 0.000	20:32	
Breath Volume	1.511	20:32	
Air Blank	0.000	20:32	
Air Blank	0.000	20:34	
Subject Sample	# 3 0.000	20:35	
Breath Volume	1.808	20:35	
Air Blank	0.000	20:36	
Control Test	0.000	20:36	Control Outside Tolerance
Air Blank	0.000	20:37	

Data Download Date/Time: 10/07/2021 06:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HUTCHINSON,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/18/2021 06:34:15

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.079	06:26
Air Blank	0.000	06:26
Subject Sample # 1	0.205	06:29
Breath Volume	1.953	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.214	06:32
Breath Volume	2.144	06:32
Air Blank	0.000	06:33
Control Test	0.078	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 09/21/2021 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PAUGA,JOHN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ROCHE

DATE/TIME OF TEST: 09/21/2021 22:22:39

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.224	22:18
Breath Volume	2.246	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.211	22:21
Breath Volume	2.265	22:21
Air Blank	0.000	22:21
Control Test	0.078	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 10/22/2021 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MCQUAY,ZACHARY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 09/26/2021 04:12:32

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.201	04:07
Breath Volume	1.613	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.213	04:11
Breath Volume	1.953	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 10/22/2021 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTCHINSON,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FOSTER,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/15/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1421958
DRY GAS EXPIRY DATE: 12/16/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 09/26/2021 04:43:37

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.184	04:39
Breath Volume	1.527	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.194	04:42
Breath Volume	2.078	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 10/22/2021 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORGAN,JOSHUA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: JESSEE

DATE/TIME OF TEST: 09/01/2021 02:05:35

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.081	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.213	02:01
Breath Volume	3.125	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.202	02:04
Breath Volume	2.648	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ERB

DATE/TIME OF TEST: 09/01/2021 23:05:47

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.095	23:01
Breath Volume	2.007	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.093	23:04
Breath Volume	1.632	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: PIRANA

DATE/TIME OF TEST: 09/02/2021 03:32:03

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.114	03:19	Volume Not Met
Breath Volume	0.328	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2 0.115	03:24	
Breath Volume	1.464	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 3 0.114	03:30	Volume Not Met
Breath Volume	0.617	03:30	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/06/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: PIRANA

DATE/TIME OF TEST: 09/02/2021 03:42:11

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.124	03:36
Breath Volume	1.136	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.106	03:40
Breath Volume	1.335	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 09/03/2021 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 08/28/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 09/07/2021 22:15:28

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.078	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.157	22:07
Breath Volume	1.382	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.178	22:10
Breath Volume	2.156	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 3	0.151	22:13
Breath Volume	1.687	22:13
Air Blank	0.000	22:14
Control Test	0.078	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 09/21/2021 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEEKS, TYLER, A
BREATH TEST OPERATOR AGENCY: Sneads PD
ARRESTING OFFICER: ALDAY, SETH
ARRESTING OFFICER AGENCY: Sneads PD
LAST AGENCY INSPECTION: 09/05/2021

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 09/26/2021 11:37:27

Diagnostics Check	OK	11:30
Air Blank	0.000	11:31
Control Test	0.081	11:31
Air Blank	0.000	11:31
Subject Sample # 1	0.149	11:32
Breath Volume	3.738	11:32
Air Blank	0.000	11:33
Air Blank	0.000	11:35
Subject Sample # 2	0.151	11:35
Breath Volume	4.210	11:35
Air Blank	0.000	11:36
Control Test	0.080	11:36
Air Blank	0.000	11:37
Diagnostics Check	OK	11:37

Data Download Date/Time: 10/04/2021 12:59