

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO,WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: PAM

DATE/TIME OF TEST: 10/08/2022 00:13:21

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:58	
Control Test	0.080	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.156	00:02	Slope Not Level
Breath Volume	1.183	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 2 0.181	00:06	
Breath Volume	1.828	00:06	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 3 0.191	00:11	
Breath Volume	1.671	00:11	
Air Blank	0.000	00:12	
Control Test	0.077	00:12	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 10/23/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: YUERGENS,A

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 10/11/2022 03:14:25

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.216	03:09
Breath Volume	1.664	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.208	03:12
Breath Volume	1.433	03:12
Air Blank	0.000	03:13
Control Test	0.077	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIDGEON, ANDREW, W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HUMISTON, HALLIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/13/2022 09:05:34

Diagnostics Check	OK	08:59
Air Blank	0.000	08:59
Control Test	0.079	09:00
Air Blank	0.000	09:00
Subject Sample # 1	0.000	09:01
Breath Volume	1.175	09:01
Air Blank	0.000	09:01
Air Blank	0.000	09:03
Subject Sample # 2	0.000	09:04
Breath Volume	1.425	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:04
Air Blank	0.000	09:05
Diagnostics Check	OK	09:05

Data Download Date/Time: 10/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: THATCHER,MATTHEW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:51

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/16/2022 12:47:55

Diagnostics Check	OK	12:41
Air Blank	0.000	12:41
Control Test	0.079	12:41
Air Blank	0.000	12:42
Subject Sample # 1	0.095	12:43
Breath Volume	1.304	12:43
Air Blank	0.000	12:43
Air Blank	0.000	12:45
Subject Sample # 2	0.097	12:46
Breath Volume	1.109	12:46
Air Blank	0.000	12:47
Control Test	0.080	12:47
Air Blank	0.000	12:47
Diagnostics Check	OK	12:47

Data Download Date/Time: 10/23/2022 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,A

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO,WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PRIETO

DATE/TIME OF TEST: 10/23/2022 01:04:18

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.079	01:02	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.079	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 10/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,A

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: STEPP,V

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/18/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:16

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MASSETTI

DATE/TIME OF TEST: 10/23/2022 04:14:24

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.072	04:09
Breath Volume	1.921	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.069	04:12
Breath Volume	1.800	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 10/23/2022 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: TRIMBOLI,MELANIE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 10/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: KAMMERDIENER

DATE/TIME OF TEST: 10/26/2022 12:56:27

Diagnostics Check	OK	12:49
Air Blank	0.000	12:50
Control Test	0.079	12:50
Air Blank	0.000	12:51
Subject Sample # 1	0.181	12:51
Breath Volume	1.410	12:51
Air Blank	0.000	12:52
Air Blank	0.000	12:54
Subject Sample # 2	0.200	12:54
Breath Volume	2.359	12:54
Air Blank	0.000	12:55
Control Test	0.079	12:55
Air Blank	0.000	12:56
Diagnostics Check	OK	12:56

Data Download Date/Time: 11/23/2022 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: CASTRO,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 19:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: DOS REIS

DATE/TIME OF TEST: 10/30/2022 20:57:19

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.079	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.265	20:52
Breath Volume	1.773	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:54
Subject Sample # 2	0.262	20:55
Breath Volume	1.804	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 11/23/2022 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BATISTA, JORGE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LUNDEEN

DATE/TIME OF TEST: 10/02/2022 23:35:44

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.273	23:30
Breath Volume	2.566	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.266	23:34
Breath Volume	2.804	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARMSTRONG,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CARHART

DATE/TIME OF TEST: 10/03/2022 02:58:36

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROSPER,BENSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: COUCE

DATE/TIME OF TEST: 10/03/2022 06:03:49

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.079	05:58
Air Blank	0.000	05:58
Subject Sample # 1	0.159	05:59
Breath Volume	2.121	05:59
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.163	06:02
Breath Volume	2.617	06:02
Air Blank	0.000	06:02
Control Test	0.080	06:03
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,MICHAEL,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GENENALE,F
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: STRANGE

DATE/TIME OF TEST: 10/03/2022 20:18:42

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.080	20:13
Air Blank	0.000	20:13
Subject Sample # 1	0.221	20:14
Breath Volume	2.058	20:14
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.222	20:17
Breath Volume	1.437	20:17
Air Blank	0.000	20:17
Control Test	0.079	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 10/07/2022 19:59:01

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:50	
Control Test	0.080	19:50	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:54	No Sample Provided
Breath Volume	0.000	19:54	
Air Blank	0.000	19:55	
Air Blank	0.000	19:57	
Subject Sample	# 2	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REILLY,S.,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REILLY,S.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CASTRO PU

DATE/TIME OF TEST: 10/08/2022 04:23:16

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.176	04:18
Breath Volume	2.703	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.165	04:21
Breath Volume	1.785	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PHILLIPS,MERCEDES
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/09/2022 04:53:51

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:52	Subject Test Refused
Breath Volume	0.000	04:52	
Air Blank	0.000	04:52	
Control Test	0.081	04:53	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BROOKS,JOSEPH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: FRAISSINET

DATE/TIME OF TEST: 10/09/2022 22:52:42

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.080	22:40	
Air Blank	0.000	22:40	
Subject Sample # 1	0.172	22:41	
Breath Volume	1.218	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample # 2		22:47	No Sample Provided
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample # 3		22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 10/25/2022 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: TIU OSORIO

DATE/TIME OF TEST: 10/10/2022 01:56:51

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.142	01:52
Breath Volume	2.269	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.140	01:55
Breath Volume	1.382	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BARRIOS,MELANY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: NETRO CANO

DATE/TIME OF TEST: 10/10/2022 03:35:41

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.065	03:31
Breath Volume	3.203	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.067	03:34
Breath Volume	4.031	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: BROOKS,JOSEPH
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: REICHENWALLNER

DATE/TIME OF TEST: 10/11/2022 00:16:36

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.085	00:11
Breath Volume	1.128	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.082	00:14
Breath Volume	1.582	00:14
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SPECA,STEVEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ZULLICK

DATE/TIME OF TEST: 10/11/2022 03:48:47

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.166	03:43
Breath Volume	1.933	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.165	03:47
Breath Volume	2.546	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 10/13/2022 02:46:34

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.109	02:41
Breath Volume	4.648	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.113	02:44
Breath Volume	4.023	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,NATHANIEL,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WOLFF,NATHANIEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: SELIN BARTHELMESS

DATE/TIME OF TEST: 10/14/2022 11:59:46

Diagnostics Check	OK	11:51	
Air Blank	0.000	11:52	
Control Test	0.080	11:52	
Air Blank	0.000	11:53	
Subject Sample	# 1 0.240	11:54	
Breath Volume	1.941	11:54	
Air Blank	0.000	11:54	
Air Blank	0.000	11:56	
Subject Sample	# 2	11:58	Sequence Aborted
Breath Volume	0.031	11:58	
Air Blank	0.000	11:59	

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SEGUNDO MARCOS

DATE/TIME OF TEST: 10/15/2022 00:51:56

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.226	00:46
Breath Volume	1.992	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.220	00:50
Breath Volume	1.515	00:50
Air Blank	0.000	00:50
Control Test	0.077	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SPECA,STEVEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: DECKARD

DATE/TIME OF TEST: 10/15/2022 03:32:47

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.081	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/16/2022 06:38:20

Diagnostics Check	OK	06:32
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:33
Subject Sample # 1	0.058	06:33
Breath Volume	2.894	06:33
Air Blank	0.000	06:34
Air Blank	0.000	06:36
Subject Sample # 2	0.057	06:36
Breath Volume	2.921	06:36
Air Blank	0.000	06:37
Control Test	0.080	06:37
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: PAUGH

DATE/TIME OF TEST: 10/17/2022 22:46:53

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.124	22:41
Breath Volume	1.558	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.122	22:45
Breath Volume	1.906	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MOSES,JORDAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 10/18/2022 01:30:09

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.150	01:25
Breath Volume	1.597	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.147	01:28
Breath Volume	1.187	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIOS,MELANY,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SCHIERA,FERNANDO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BERNHARDT

DATE/TIME OF TEST: 10/21/2022 21:21:32

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.080	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.114	21:17
Breath Volume	3.175	21:17
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.106	21:19
Breath Volume	2.136	21:19
Air Blank	0.000	21:20
Control Test	0.078	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WRAY,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: WOLFF

DATE/TIME OF TEST: 10/22/2022 12:01:53

Diagnostics Check	OK	11:51
Air Blank	0.000	11:52
Control Test	0.080	11:52
Air Blank	0.000	11:52
Subject Sample # 1	0.254	11:53
Breath Volume	2.402	11:53
Air Blank	0.000	11:54
Air Blank	0.000	11:56
Subject Sample # 2	0.233	11:56
Breath Volume	2.777	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 3	0.251	12:00
Breath Volume	2.433	12:00
Air Blank	0.000	12:00
Control Test	0.079	12:01
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLIGAN,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: HARWOOD

DATE/TIME OF TEST: 10/23/2022 02:16:13

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.217	02:11
Breath Volume	1.199	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.217	02:14
Breath Volume	1.195	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CROWLEY

DATE/TIME OF TEST: 10/23/2022 02:29:48

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.119	02:25
Breath Volume	3.753	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.118	02:28
Breath Volume	3.351	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,TORI,D
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: JOHNSTON,TORI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 10/23/2022 21:47:23

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.077	21:42
Breath Volume	1.910	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.077	21:45
Breath Volume	2.472	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 10/24/2022 00:19:37

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.167	00:14
Breath Volume	3.085	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.184	00:17
Breath Volume	3.582	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, DAVID, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALLEN, JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: WINSTON

DATE/TIME OF TEST: 10/24/2022 16:22:59

Diagnostics Check	OK	16:11	
Air Blank	0.000	16:12	
Control Test	0.081	16:12	
Air Blank	0.000	16:12	
Subject Sample	# 1	16:16	No Sample Provided
Breath Volume	0.000	16:16	
Air Blank	0.000	16:16	
Air Blank	0.000	16:18	
Subject Sample	# 2	16:21	No Sample Provided
Breath Volume	0.000	16:21	
Air Blank	0.000	16:22	
Control Test	0.078	16:22	
Air Blank	0.000	16:22	
Diagnostics Check	OK	16:22	

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: WINSTON

DATE/TIME OF TEST: 10/24/2022 16:47:36

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:34	
Control Test	0.081	16:34	
Air Blank	0.000	16:34	
Subject Sample	# 1 0.038	16:38	Volume Not Met
Breath Volume	0.050	16:38	
Air Blank	0.000	16:38	
Air Blank	0.000	16:40	
Subject Sample	# 2 0.048	16:42	
Breath Volume	1.375	16:42	
Air Blank	0.000	16:43	
Air Blank	0.000	16:45	
Subject Sample	# 3 0.045	16:46	
Breath Volume	1.152	16:46	
Air Blank	0.000	16:46	
Control Test	0.078	16:46	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MAIN

DATE/TIME OF TEST: 10/24/2022 19:23:33

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.081	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.000	19:19
Breath Volume	1.441	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.000	19:22
Breath Volume	1.593	19:22
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: TOX BUENAFE

DATE/TIME OF TEST: 10/25/2022 01:29:22

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.174	01:22
Breath Volume	1.226	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.174	01:27
Breath Volume	1.804	01:27
Air Blank	0.000	01:28
Control Test	0.081	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GARLICK,ROBERT
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/25/2022 04:41:20

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.030	04:36
Breath Volume	1.429	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample # 2	0.031	04:39
Breath Volume	1.832	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 10/25/2022 08:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,DERICK
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: POLESKY

DATE/TIME OF TEST: 10/26/2022 04:24:58

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.206	04:20
Breath Volume	2.894	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.203	04:23
Breath Volume	2.265	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 11/16/2022 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FIXLER,T
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 10/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: SOLBACH

DATE/TIME OF TEST: 10/30/2022 18:28:56

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.081	18:23
Air Blank	0.000	18:23
Subject Sample # 1	0.242	18:24
Breath Volume	2.445	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:26
Subject Sample # 2	0.238	18:27
Breath Volume	2.179	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 11/16/2022 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/30/2022 20:29:02

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.262	20:23
Breath Volume	2.820	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.260	20:27
Breath Volume	2.031	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:29

Data Download Date/Time: 11/16/2022 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARCHMAN,MARCHMAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/01/2022 03:54:56

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.251	03:50
Breath Volume	4.125	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.237	03:53
Breath Volume	3.410	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCKER,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/01/2022 13:31:51

Diagnostics Check	OK	13:25
Air Blank	0.000	13:25
Control Test	0.080	13:26
Air Blank	0.000	13:26
Subject Sample # 1	0.041	13:27
Breath Volume	2.273	13:27
Air Blank	0.000	13:27
Air Blank	0.000	13:29
Subject Sample # 2	0.039	13:30
Breath Volume	3.441	13:30
Air Blank	0.000	13:30
Control Test	0.079	13:31
Air Blank	0.000	13:31
Diagnostics Check	OK	13:31

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUCKER,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BRUCKER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/01/2022 14:32:33

Diagnostics Check	OK	14:26
Air Blank	0.000	14:26
Control Test	0.080	14:27
Air Blank	0.000	14:27
Subject Sample # 1	0.018	14:28
Breath Volume	2.878	14:28
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.016	14:31
Breath Volume	2.453	14:31
Air Blank	0.000	14:31
Control Test	0.080	14:31
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RISI,ALEYA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/03/2022 23:43:51

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.182	23:38
Breath Volume	3.277	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.185	23:42
Breath Volume	4.031	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RISI,ALEYA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/04/2022 00:35:14

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.000	00:30
Breath Volume	3.503	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.000	00:33
Breath Volume	3.519	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RISI,ALEYA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/04/2022 04:30:47

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.082	04:26
Breath Volume	3.652	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.085	04:29
Breath Volume	4.402	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REILLY,S.,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REILLY,S.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: LANCELLE

DATE/TIME OF TEST: 10/04/2022 05:11:51

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.130	05:07
Breath Volume	3.363	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.136	05:10
Breath Volume	3.324	05:10
Air Blank	0.000	05:10
Control Test	0.081	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 10/04/2022 12:32:54

Diagnostics Check	OK	12:26
Air Blank	0.000	12:27
Control Test	0.081	12:27
Air Blank	0.000	12:28
Subject Sample # 1	0.000	12:28
Breath Volume	1.640	12:28
Air Blank	0.000	12:29
Air Blank	0.000	12:31
Subject Sample # 2	0.000	12:31
Breath Volume	1.882	12:31
Air Blank	0.000	12:32
Control Test	0.081	12:32
Air Blank	0.000	12:32
Diagnostics Check	OK	12:32

Data Download Date/Time: 11/02/2022 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARDWELL,OFC
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: CARMACK

DATE/TIME OF TEST: 10/06/2022 21:10:18

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.080	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.161	21:05
Breath Volume	1.914	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:07
Subject Sample # 2	0.162	21:08
Breath Volume	1.886	21:08
Air Blank	0.000	21:09
Control Test	0.080	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: TON DE LA CRUZ

DATE/TIME OF TEST: 10/09/2022 05:00:55

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.183	04:56
Breath Volume	1.894	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.182	04:59
Breath Volume	2.214	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,JONATHAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BAIZE

DATE/TIME OF TEST: 10/10/2022 23:10:59

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:06
Breath Volume	2.945	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:09
Breath Volume	2.914	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,ROBERT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/12/2022 22:46:25

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.000	22:42
Breath Volume	1.281	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.000	22:45
Breath Volume	1.734	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DZUGAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DZUGAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: KORN

DATE/TIME OF TEST: 10/13/2022 01:40:19

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.166	01:35
Breath Volume	2.980	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.170	01:38
Breath Volume	2.664	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,TROY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/13/2022 03:01:30

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.120	02:56
Breath Volume	1.546	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.120	03:00
Breath Volume	1.816	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAVANCE,LAVANCE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 06:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 10/13/2022 06:34:30

Diagnostics Check	OK	06:28
Air Blank	0.000	06:29
Control Test	0.080	06:29
Air Blank	0.000	06:29
Subject Sample # 1	0.217	06:30
Breath Volume	1.621	06:30
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.218	06:33
Breath Volume	1.500	06:33
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,MARDEE,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MANO,MARDEE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 10/13/2022 12:55:46

Diagnostics Check	OK	12:49
Air Blank	0.000	12:50
Control Test	0.080	12:50
Air Blank	0.000	12:51
Subject Sample # 1	0.092	12:51
Breath Volume	2.054	12:51
Air Blank	0.000	12:51
Air Blank	0.000	12:53
Subject Sample # 2	0.090	12:54
Breath Volume	2.195	12:54
Air Blank	0.000	12:54
Control Test	0.080	12:55
Air Blank	0.000	12:55
Diagnostics Check	OK	12:55

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VASQUEZ,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/14/2022 03:58:01

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.000	03:53
Breath Volume	2.261	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.000	03:56
Breath Volume	2.898	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMBRIA,STEVEN,J
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: CAMBRIA,STEVEN
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: ALUC

DATE/TIME OF TEST: 10/14/2022 04:18:47

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.000	04:14
Breath Volume	2.578	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.000	04:17
Breath Volume	1.953	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAVANCE,CODY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: RAMOS CASTILLO

DATE/TIME OF TEST: 10/14/2022 19:06:23

Diagnostics Check	OK	18:59
Air Blank	0.000	19:00
Control Test	0.081	19:00
Air Blank	0.000	19:01
Subject Sample # 1	0.087	19:01
Breath Volume	3.542	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.081	19:04
Breath Volume	2.757	19:04
Air Blank	0.000	19:05
Control Test	0.081	19:05
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 11/02/2022 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RISI,ALEYA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/15/2022 04:50:20

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.000	04:45
Breath Volume	3.359	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.000	04:48
Breath Volume	3.496	04:48
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/15/2022 10:06:16

Diagnostics Check	OK	10:00
Air Blank	0.000	10:00
Control Test	0.081	10:00
Air Blank	0.000	10:01
Subject Sample # 1	0.069	10:01
Breath Volume	2.617	10:01
Air Blank	0.000	10:02
Air Blank	0.000	10:04
Subject Sample # 2	0.070	10:04
Breath Volume	2.753	10:04
Air Blank	0.000	10:05
Control Test	0.081	10:05
Air Blank	0.000	10:06
Diagnostics Check	OK	10:06

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/15/2022 12:37:48

Diagnostics Check	OK	12:31
Air Blank	0.000	12:32
Control Test	0.080	12:32
Air Blank	0.000	12:33
Subject Sample # 1	0.034	12:33
Breath Volume	2.476	12:33
Air Blank	0.000	12:34
Air Blank	0.000	12:36
Subject Sample # 2	0.034	12:36
Breath Volume	2.804	12:36
Air Blank	0.000	12:36
Control Test	0.080	12:37
Air Blank	0.000	12:37
Diagnostics Check	OK	12:37

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARIN,OFC
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 10/16/2022 01:19:37

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.149	01:13
Breath Volume	2.699	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.142	01:18
Breath Volume	2.343	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KIMBERLIN,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/18/2022 00:53:01

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.000	00:48
Breath Volume	2.070	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.000	00:51
Breath Volume	1.824	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:53

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALVEREZ,H
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 10/18/2022 04:19:41

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.195	04:15
Breath Volume	2.421	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.197	04:18
Breath Volume	2.230	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GANK,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: CHAMBLISS

DATE/TIME OF TEST: 10/18/2022 05:16:05

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.000	05:11
Breath Volume	2.003	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.000	05:14
Breath Volume	1.832	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:16

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MILLER,DETECTIVE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/18/2022 22:20:19

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.000	22:15
Breath Volume	1.882	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.000	22:18
Breath Volume	2.269	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,SEBASTIAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/20/2022 02:37:08

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.244	02:32
Breath Volume	4.093	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.245	02:35
Breath Volume	4.152	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,HENRY,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LUPI,VICTOR
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: MAYER

DATE/TIME OF TEST: 10/21/2022 15:54:47

Diagnostics Check	OK	15:48
Air Blank	0.000	15:49
Control Test	0.081	15:49
Air Blank	0.000	15:49
Subject Sample # 1	0.213	15:50
Breath Volume	1.859	15:50
Air Blank	0.000	15:50
Air Blank	0.000	15:52
Subject Sample # 2	0.215	15:53
Breath Volume	2.257	15:53
Air Blank	0.000	15:53
Control Test	0.080	15:54
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 11/02/2022 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER, PAIGE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MASON, DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SOKALSKI

DATE/TIME OF TEST: 10/21/2022 21:20:42

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.082	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.312	21:15
Breath Volume	1.847	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.306	21:19
Breath Volume	2.367	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER, PAIGE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SAVIC, DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SOLKALSKI

DATE/TIME OF TEST: 10/21/2022 23:33:00

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.254	23:28
Breath Volume	2.144	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.263	23:31
Breath Volume	2.312	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:33

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER, PAIGE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT, ROBERT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: ST GEORGE

DATE/TIME OF TEST: 10/22/2022 00:33:59

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.084	00:29
Breath Volume	1.585	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.080	00:32
Breath Volume	1.992	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER, PAIGE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORGAN, J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: LAYMAN III

DATE/TIME OF TEST: 10/22/2022 04:23:19

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.189	04:18
Breath Volume	2.523	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.182	04:21
Breath Volume	1.441	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SHULTZ,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: BOSTIAN

DATE/TIME OF TEST: 10/22/2022 08:26:57

Diagnostics Check	OK	08:20
Air Blank	0.000	08:21
Control Test	0.081	08:21
Air Blank	0.000	08:21
Subject Sample # 1	0.000	08:22
Breath Volume	1.292	08:22
Air Blank	0.000	08:23
Air Blank	0.000	08:25
Subject Sample # 2	0.000	08:25
Breath Volume	1.816	08:25
Air Blank	0.000	08:26
Control Test	0.081	08:26
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER, PAIGE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER, C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PLEASANT

DATE/TIME OF TEST: 10/23/2022 01:17:57

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.141	01:13
Breath Volume	2.906	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.138	01:16
Breath Volume	2.832	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER, PAIGE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT, ROBERT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/23/2022 03:12:35

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.088	03:07
Breath Volume	1.976	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.087	03:11
Breath Volume	2.625	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Punta Gorda PD
ARRESTING OFFICER: GRAY,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/23/2022 18:51:27

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.346	18:43
Breath Volume	4.140	18:43
Air Blank	0.000	18:44
Air Blank	0.000	18:45
Subject Sample # 2	0.323	18:46
Breath Volume	3.855	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:49
Subject Sample # 3	0.329	18:49
Breath Volume	4.156	18:49
Air Blank	0.000	18:50
Control Test	0.080	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VASQUEZ,B
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/23/2022 21:12:09

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.181	21:07
Breath Volume	1.937	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.178	21:10
Breath Volume	2.750	21:10
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PGPD,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/23/2022 21:26:15

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.081	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.283	21:21
Breath Volume	4.128	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.276	21:24
Breath Volume	4.390	21:24
Air Blank	0.000	21:25
Control Test	0.079	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SAVIC,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 10/24/2022 02:00:30

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.000	01:56
Breath Volume	1.351	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.000	01:59
Breath Volume	1.921	01:59
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PGPD,OFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/24/2022 04:58:29

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:53	
Control Test	0.081	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1 0.129	04:54	
Breath Volume	3.382	04:54	
Air Blank	0.000	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 2	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.080	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 11/02/2022 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/24/2022 22:23:26

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.216	22:19
Breath Volume	2.265	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.234	22:21
Breath Volume	2.230	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KIMBERLIN,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: LECLERCQ

DATE/TIME OF TEST: 10/25/2022 02:59:53

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.171	02:55
Breath Volume	4.210	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.173	02:58
Breath Volume	3.523	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROSPER,BENSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PROSPER,BENSON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/26/2022 21:08:00

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.173	21:03
Breath Volume	2.093	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.158	21:06
Breath Volume	1.910	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:08

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALVAREZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 10/28/2022 05:01:41

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.146	04:56
Breath Volume	3.941	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.152	05:00
Breath Volume	3.847	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WHYMER,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: MENNA

DATE/TIME OF TEST: 10/28/2022 20:26:04

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.081	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.000	20:21
Breath Volume	1.285	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.000	20:24
Breath Volume	1.238	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:26

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WALDROP,P
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/29/2022 02:47:12

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.233	02:38
Breath Volume	1.675	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.256	02:42
Breath Volume	2.839	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 3	0.254	02:45
Breath Volume	2.789	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/29/2022 06:39:35

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.080	06:34
Air Blank	0.000	06:34
Subject Sample # 1	0.000	06:35
Breath Volume	2.070	06:35
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.000	06:38
Breath Volume	2.214	06:38
Air Blank	0.000	06:38
Control Test	0.080	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/29/2022 10:43:07

Diagnostics Check	OK	10:36
Air Blank	0.000	10:36
Control Test	0.081	10:37
Air Blank	0.000	10:37
Subject Sample # 1	0.103	10:38
Breath Volume	2.000	10:38
Air Blank	0.000	10:38
Air Blank	0.000	10:40
Subject Sample # 2	0.103	10:41
Breath Volume	2.132	10:41
Air Blank	0.000	10:42
Control Test	0.080	10:42
Air Blank	0.000	10:43
Diagnostics Check	OK	10:43

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/29/2022 14:18:14

Diagnostics Check	OK	14:12
Air Blank	0.000	14:12
Control Test	0.081	14:12
Air Blank	0.000	14:13
Subject Sample # 1	0.048	14:13
Breath Volume	1.820	14:13
Air Blank	0.000	14:14
Air Blank	0.000	14:16
Subject Sample # 2	0.045	14:16
Breath Volume	1.359	14:16
Air Blank	0.000	14:17
Control Test	0.081	14:17
Air Blank	0.000	14:18
Diagnostics Check	OK	14:18

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENCIC,KENNETH,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PENCIC,KENNETH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/29/2022 15:15:44

Diagnostics Check	OK	15:09
Air Blank	0.000	15:10
Control Test	0.081	15:10
Air Blank	0.000	15:10
Subject Sample # 1	0.036	15:11
Breath Volume	1.960	15:11
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.033	15:14
Breath Volume	1.433	15:14
Air Blank	0.000	15:14
Control Test	0.080	15:15
Air Blank	0.000	15:15
Diagnostics Check	OK	15:15

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HANSOM

DATE/TIME OF TEST: 10/29/2022 19:12:28

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.174	19:04
Breath Volume	1.132	19:04
Air Blank	0.000	19:05
Air Blank	0.000	19:06
Subject Sample # 2	0.150	19:07
Breath Volume	1.449	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 3	0.174	19:11
Breath Volume	1.414	19:11
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 11/02/2022 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: SPAIN

DATE/TIME OF TEST: 10/29/2022 21:50:39

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Subject Sample # 1	0.223	21:46
Breath Volume	1.761	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.218	21:49
Breath Volume	2.191	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO, MONICA, S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO, CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: HANSCOM

DATE/TIME OF TEST: 10/30/2022 09:32:26

Diagnostics Check	OK	09:26
Air Blank	0.000	09:26
Control Test	0.080	09:27
Air Blank	0.000	09:27
Subject Sample # 1	0.000	09:28
Breath Volume	2.699	09:28
Air Blank	0.000	09:28
Air Blank	0.000	09:30
Subject Sample # 2	0.000	09:31
Breath Volume	2.136	09:31
Air Blank	0.000	09:31
Control Test	0.081	09:31
Air Blank	0.000	09:32
Diagnostics Check	OK	09:32

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO, MONICA, S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANDERSON, DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: INGARGIOLA

DATE/TIME OF TEST: 10/30/2022 10:35:58

Diagnostics Check	OK	10:29
Air Blank	0.000	10:30
Control Test	0.081	10:30
Air Blank	0.000	10:30
Subject Sample # 1	0.209	10:31
Breath Volume	2.710	10:31
Air Blank	0.000	10:32
Air Blank	0.000	10:33
Subject Sample # 2	0.206	10:34
Breath Volume	2.457	10:34
Air Blank	0.000	10:35
Control Test	0.081	10:35
Air Blank	0.000	10:35
Diagnostics Check	OK	10:35

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PARRISH,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 10/30/2022 18:46:03

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.081	18:39
Air Blank	0.000	18:40
Subject Sample # 1	0.145	18:41
Breath Volume	1.226	18:41
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.156	18:44
Breath Volume	2.429	18:44
Air Blank	0.000	18:45
Control Test	0.080	18:45
Air Blank	0.000	18:45
Diagnostics Check	OK	18:46

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,T,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARVER,CARVER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 10/30/2022 19:29:27

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.256	19:25
Breath Volume	1.382	19:25
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.257	19:27
Breath Volume	2.289	19:27
Air Blank	0.000	19:28
Control Test	0.079	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OWENS,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: HALKA

DATE/TIME OF TEST: 10/30/2022 20:15:23

Diagnostics Check	OK	20:07
Air Blank	0.000	20:07
Control Test	0.080	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.140	20:10
Breath Volume	3.429	20:10
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.129	20:13
Breath Volume	2.109	20:13
Air Blank	0.000	20:14
Control Test	0.081	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ZENONIANI,L
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HUTCHISON

DATE/TIME OF TEST: 10/30/2022 22:39:16

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.320	22:34
Breath Volume	2.179	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.314	22:37
Breath Volume	2.281	22:37
Air Blank	0.000	22:38
Control Test	0.079	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SKIBA

DATE/TIME OF TEST: 10/31/2022 02:05:55

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.270	02:01
Breath Volume	3.101	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.265	02:04
Breath Volume	3.042	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BOSZE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: GREBENUK

DATE/TIME OF TEST: 10/31/2022 07:34:58

Diagnostics Check	OK	07:28
Air Blank	0.000	07:29
Control Test	0.079	07:29
Air Blank	0.000	07:30
Subject Sample # 1	0.000	07:30
Breath Volume	1.335	07:30
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample # 2	0.000	07:33
Breath Volume	1.554	07:33
Air Blank	0.000	07:34
Control Test	0.079	07:34
Air Blank	0.000	07:34
Diagnostics Check	OK	07:34

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BOSZE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: MAIR

DATE/TIME OF TEST: 10/31/2022 14:06:17

Diagnostics Check	OK	13:59
Air Blank	0.000	14:00
Control Test	0.080	14:00
Air Blank	0.000	14:01
Subject Sample # 1	0.000	14:02
Breath Volume	1.914	14:02
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample # 2	0.000	14:04
Breath Volume	1.554	14:04
Air Blank	0.000	14:05
Control Test	0.080	14:05
Air Blank	0.000	14:06
Diagnostics Check	OK	14:06

Data Download Date/Time: 11/02/2022 20:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KLOYER,P
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 10/31/2022 22:52:56

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.272	22:48
Breath Volume	1.933	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.284	22:51
Breath Volume	2.152	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 11/02/2022 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALVAREZ,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 10/31/2022 23:30:40

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.000	23:26
Breath Volume	2.003	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.000	23:29
Breath Volume	2.132	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 11/02/2022 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, ANTHONY, L
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MAYES, DARRON
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HEBERT

DATE/TIME OF TEST: 10/08/2022 00:22:28

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.199	00:14
Breath Volume	4.125	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.177	00:17
Breath Volume	2.574	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 3	0.184	00:20
Breath Volume	3.886	00:20
Air Blank	0.000	00:21
Control Test	0.077	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 10/17/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 10/13/2022 18:59:29

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:54
Subject Sample # 1	0.009	18:55
Breath Volume	2.140	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.008	18:58
Breath Volume	1.632	18:58
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 10/17/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: AGUILAR BERNARDO

DATE/TIME OF TEST: 10/22/2022 01:08:01

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.209	00:59
Breath Volume	1.757	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.232	01:03
Breath Volume	2.910	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 3	0.207	01:06
Breath Volume	1.808	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:08

Data Download Date/Time: 11/14/2022 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: CASTELLANO

DATE/TIME OF TEST: 10/24/2022 02:20:11

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.110	02:15
Breath Volume	1.140	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.114	02:18
Breath Volume	1.328	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 11/14/2022 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 10/28/2022 11:08:44

Diagnostics Check	OK	11:02
Air Blank	0.000	11:03
Control Test	0.079	11:03
Air Blank	0.000	11:04
Subject Sample # 1	0.010	11:04
Breath Volume	1.277	11:04
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.010	11:07
Breath Volume	1.421	11:07
Air Blank	0.000	11:07
Control Test	0.080	11:08
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 11/14/2022 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: DELGADO CERRO

DATE/TIME OF TEST: 10/21/2022 03:24:32

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.000	03:20
Breath Volume	1.113	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.000	03:23
Breath Volume	1.136	03:23
Air Blank	0.000	03:23
Control Test	0.082	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 11/09/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: LIZARDI

DATE/TIME OF TEST: 10/29/2022 03:53:06

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.193	03:48
Breath Volume	2.230	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.184	03:51
Breath Volume	2.605	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 11/09/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 21:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 10/29/2022 22:30:00

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.171	22:24
Breath Volume	1.328	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.177	22:28
Breath Volume	1.445	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:30

Data Download Date/Time: 11/09/2022 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: MULDERRIG,SEAN

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 11821080A2

DRY GAS EXPIRY DATE: 06/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:36

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: OCASIO

DATE/TIME OF TEST: 10/30/2022 03:15:56

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.151	03:08
Breath Volume	1.140	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.165	03:14
Breath Volume	2.398	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 11/09/2022 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: SILVERA, SEBASTIAN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: LYKEN

DATE/TIME OF TEST: 10/09/2022 00:07:37

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.149	00:03
Breath Volume	2.187	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.135	00:06
Breath Volume	1.320	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: PORTILLO, SCARLYN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 10/09/2022 03:52:33

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.081	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.081	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSNER,S,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ARBELO,R
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SALINAS

DATE/TIME OF TEST: 10/11/2022 04:13:41

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:12	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ARBELO,RYAN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GABISKIRIYA

DATE/TIME OF TEST: 10/15/2022 02:47:00

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.079	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BLANDON,ALLAN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: LARGAESPADA

DATE/TIME OF TEST: 10/16/2022 00:03:43

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.081	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:02	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1224982
DRY GAS EXPIRY DATE: 10/31/2022
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 10/20/2022 03:40:04

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:39	
Control Test	0.080	03:39	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 10/20/2022 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,ALEXIS,
BREATH TEST OPERATOR AGENCY: HIALEAH PD
ARRESTING OFFICER: NATARENO,BRIANNE
ARRESTING OFFICER AGENCY: HIALEAH PD
LAST AGENCY INSPECTION: 09/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: SULEWSKI

DATE/TIME OF TEST: 10/22/2022 06:52:54

Diagnostics Check	OK	06:46
Air Blank	0.000	06:46
Control Test	0.081	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.214	06:48
Breath Volume	1.843	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:50
Subject Sample # 2	0.207	06:51
Breath Volume	2.363	06:51
Air Blank	0.000	06:51
Control Test	0.079	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 10/26/2022 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,ALEXIS,
BREATH TEST OPERATOR AGENCY: HIALEAH PD
ARRESTING OFFICER: ARTIME,JOEL
ARRESTING OFFICER AGENCY: HIALEAH PD
LAST AGENCY INSPECTION: 09/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: PINO MIRANDA

DATE/TIME OF TEST: 10/22/2022 07:30:25

Diagnostics Check	OK	07:23
Air Blank	0.000	07:24
Control Test	0.080	07:24
Air Blank	0.000	07:25
Subject Sample # 1	0.150	07:25
Breath Volume	1.445	07:25
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.165	07:28
Breath Volume	1.515	07:28
Air Blank	0.000	07:29
Control Test	0.079	07:29
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 10/26/2022 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SHACKELFORD

DATE/TIME OF TEST: 10/08/2022 02:35:27

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.126	02:27
Breath Volume	1.324	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.148	02:30
Breath Volume	1.503	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 3	0.134	02:33
Breath Volume	1.531	02:33
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARWOOD,L
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 10/16/2022 00:54:47

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.158	00:50
Breath Volume	5.109	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.163	00:53
Breath Volume	5.355	00:53
Air Blank	0.000	00:53
Control Test	0.078	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LAWLER,ADDISON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: DUBOIS

DATE/TIME OF TEST: 10/16/2022 01:32:30

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.205	01:23
Breath Volume	1.425	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.179	01:27
Breath Volume	1.687	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 3	0.175	01:30
Breath Volume	1.644	01:30
Air Blank	0.000	01:31
Control Test	0.077	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 10/20/2022 00:31:34

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.117	00:26
Breath Volume	3.679	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.119	00:29
Breath Volume	3.300	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: GOWAN

DATE/TIME OF TEST: 10/22/2022 22:55:56

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.106	22:51
Breath Volume	1.699	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.089	22:54
Breath Volume	1.207	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/27/2022 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: WASDIN

DATE/TIME OF TEST: 10/28/2022 21:04:40

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.078	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.102	20:59
Breath Volume	1.660	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.101	21:03
Breath Volume	1.968	21:03
Air Blank	0.000	21:03
Control Test	0.078	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 11/23/2022 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/29/2022 20:38:05

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.203	20:33
Breath Volume	2.324	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:35
Subject Sample # 2	0.218	20:36
Breath Volume	2.781	20:36
Air Blank	0.000	20:37
Control Test	0.078	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:38

Data Download Date/Time: 11/23/2022 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,MELISSA,A
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: NUNEZ,BENJAMIN
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/20/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CARPINETA

DATE/TIME OF TEST: 10/27/2022 01:33:42

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 11/10/2022 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BARRETT, DEVON
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 10/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: COPE II

DATE/TIME OF TEST: 10/29/2022 21:17:34

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:14	
Control Test	0.079	21:15	
Air Blank	0.000	21:15	
Subject Sample	# 1	21:16	Subject Test Refused
Breath Volume	0.000	21:16	
Air Blank	0.000	21:16	
Control Test	0.079	21:16	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

Data Download Date/Time: 11/10/2022 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: PEREZ,GUERSON
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: JIMENEZPINEIRO

DATE/TIME OF TEST: 10/15/2022 00:28:42

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 10/26/2022 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/21/2022 00:19:45

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Control Test	0.080	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 10/26/2022 08:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,FERNANDEZ
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 10/27/2022 22:57:38

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:54	
Control Test	0.081	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:56	
Control Test	0.081	22:57	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 12/01/2022 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: AGUIRRE,N
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: SOMARRIBAROSALES

DATE/TIME OF TEST: 10/28/2022 20:38:44

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.080	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.243	20:33
Breath Volume	1.718	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.245	20:37
Breath Volume	1.300	20:37
Air Blank	0.000	20:37
Control Test	0.078	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 12/01/2022 14:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SACIA

DATE/TIME OF TEST: 10/15/2022 23:46:04

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.077	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.193	23:41
Breath Volume	2.683	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.189	23:44
Breath Volume	1.707	23:44
Air Blank	0.000	23:45
Control Test	0.077	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:46

Data Download Date/Time: 11/01/2022 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MARJI

DATE/TIME OF TEST: 10/30/2022 04:15:07

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.078	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.101	04:10
Breath Volume	1.597	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:12
Subject Sample # 2	0.099	04:13
Breath Volume	1.640	04:13
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 11/01/2022 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MARCELLUS

DATE/TIME OF TEST: 10/31/2022 04:02:06

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.058	03:57
Breath Volume	2.667	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.063	04:00
Breath Volume	2.933	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 11/01/2022 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GRAVA

DATE/TIME OF TEST: 10/02/2022 04:03:32

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.079	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1 0.163	03:56	Volume Not Met
Breath Volume	0.136	03:56	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2 0.155	04:01	Volume Not Met
Breath Volume	0.410	04:01	
Air Blank	0.000	04:02	
Control Test	0.079	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GRAVA

DATE/TIME OF TEST: 10/02/2022 04:22:55

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.078	04:07	
Air Blank	0.000	04:07	
Subject Sample	# 1 0.121	04:10	Volume Not Met
Breath Volume	0.289	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2 0.133	04:15	
Breath Volume	1.410	04:15	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample	# 3 0.132	04:21	Volume Not Met
Breath Volume	1.082	04:21	
Air Blank	0.000	04:21	
Control Test	0.077	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GRAVA

DATE/TIME OF TEST: 10/02/2022 04:43:29

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.078	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.113	04:34	Slope Not Level
Breath Volume	1.191	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2 0.137	04:38	
Breath Volume	1.648	04:38	
Air Blank	0.000	04:38	
Air Blank	0.000	04:40	
Subject Sample	# 3 0.121	04:41	
Breath Volume	1.433	04:41	
Air Blank	0.000	04:42	
Control Test	0.078	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BALONICK

DATE/TIME OF TEST: 10/04/2022 22:19:28

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.216	22:14
Breath Volume	2.921	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.214	22:17
Breath Volume	3.214	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 10/07/2022 05:34:21

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.178	05:29
Breath Volume	2.035	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.164	05:32
Breath Volume	1.226	05:32
Air Blank	0.000	05:33
Control Test	0.078	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MESA MARTINEZ

DATE/TIME OF TEST: 10/08/2022 05:01:20

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.114	04:56
Breath Volume	1.417	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.105	04:59
Breath Volume	1.417	04:59
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: ZAMARRIPA

DATE/TIME OF TEST: 10/08/2022 05:13:51

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.079	05:08	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:12	Subject Test Refused
Breath Volume	0.542	05:12	
Air Blank	0.000	05:12	
Control Test	0.079	05:13	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COOPER,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GONCALVES WILSON

DATE/TIME OF TEST: 10/09/2022 01:55:32

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.124	01:50
Breath Volume	1.449	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.129	01:53
Breath Volume	1.703	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: GARCIA MARTINEZ

DATE/TIME OF TEST: 10/09/2022 05:19:42

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:16	
Control Test	0.080	05:16	
Air Blank	0.000	05:16	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Control Test	0.079	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/12/2022 22:20:22

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.079	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1 0.126	22:10	Volume Not Met
Breath Volume	0.019	22:10	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2 0.142	22:14	
Breath Volume	1.480	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:17	
Subject Sample	# 3 0.152	22:18	
Breath Volume	1.230	22:18	
Air Blank	0.000	22:19	
Control Test	0.078	22:19	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PADILLA MARTINEZ

DATE/TIME OF TEST: 10/13/2022 01:43:30

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.266	01:38
Breath Volume	2.128	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.253	01:41
Breath Volume	1.550	01:41
Air Blank	0.000	01:42
Control Test	0.077	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 10/13/2022 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLOOD,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: KATHAVATE

DATE/TIME OF TEST: 10/14/2022 02:08:59

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.146	02:03
Breath Volume	2.492	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.148	02:07
Breath Volume	2.382	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 10/14/2022 04:31:29

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.150	04:27
Breath Volume	1.304	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.155	04:29
Breath Volume	1.726	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 10/14/2022 21:28:33

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.127	21:23
Breath Volume	1.925	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.125	21:27
Breath Volume	1.464	21:27
Air Blank	0.000	21:27
Control Test	0.078	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STILES,MEGAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: BABA

DATE/TIME OF TEST: 10/15/2022 11:47:50

Diagnostics Check	OK	11:41
Air Blank	0.000	11:41
Control Test	0.080	11:41
Air Blank	0.000	11:42
Subject Sample # 1	0.152	11:43
Breath Volume	3.449	11:43
Air Blank	0.000	11:43
Air Blank	0.000	11:45
Subject Sample # 2	0.152	11:46
Breath Volume	3.414	11:46
Air Blank	0.000	11:46
Control Test	0.079	11:47
Air Blank	0.000	11:47
Diagnostics Check	OK	11:47

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PATINO DUQUE

DATE/TIME OF TEST: 10/15/2022 23:34:54

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.278	23:30
Breath Volume	1.867	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.277	23:33
Breath Volume	1.480	23:33
Air Blank	0.000	23:33
Control Test	0.078	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: NOGUEZ

DATE/TIME OF TEST: 10/16/2022 01:04:34

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.154	00:59
Breath Volume	1.859	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.155	01:02
Breath Volume	1.976	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/16/2022 03:58:53

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.075	03:54
Breath Volume	1.210	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.074	03:57
Breath Volume	1.222	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: ALVARADO OYUELA

DATE/TIME OF TEST: 10/16/2022 06:48:22

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.080	06:41
Air Blank	0.000	06:41
Subject Sample # 1	0.116	06:42
Breath Volume	1.250	06:42
Air Blank	0.000	06:43
Air Blank	0.000	06:45
Subject Sample # 2	0.112	06:46
Breath Volume	1.875	06:46
Air Blank	0.000	06:47
Control Test	0.079	06:47
Air Blank	0.000	06:48
Diagnostics Check	OK	06:48

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TABOAOA, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: COURSON

DATE/TIME OF TEST: 10/16/2022 19:20:06

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:07	
Control Test	0.080	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1 0.151	19:09	
Breath Volume	1.121	19:09	
Air Blank	0.000	19:10	
Air Blank	0.000	19:12	
Subject Sample	# 2 0.139	19:15	Volume Not Met
Breath Volume	0.421	19:15	
Air Blank	0.000	19:15	
Air Blank	0.000	19:17	
Subject Sample	# 3 0.160	19:18	
Breath Volume	1.105	19:18	
Air Blank	0.000	19:19	
Control Test	0.077	19:19	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/16/2022 20:54:40

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.173	20:50
Breath Volume	3.242	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.178	20:53
Breath Volume	3.144	20:53
Air Blank	0.000	20:53
Control Test	0.079	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 11/01/2022 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CASTILLO DE ZAYAS

DATE/TIME OF TEST: 10/17/2022 03:27:14

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.195	03:22
Breath Volume	1.640	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.185	03:25
Breath Volume	1.281	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEMAN,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: WITHUM

DATE/TIME OF TEST: 10/18/2022 16:50:37

Diagnostics Check	OK	16:44
Air Blank	0.000	16:44
Control Test	0.079	16:45
Air Blank	0.000	16:45
Subject Sample # 1	0.199	16:46
Breath Volume	2.554	16:46
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 2	0.192	16:49
Breath Volume	2.679	16:49
Air Blank	0.000	16:49
Control Test	0.079	16:50
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MARONE

DATE/TIME OF TEST: 10/18/2022 21:27:38

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.130	21:22
Breath Volume	1.253	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:24
Subject Sample # 2	0.130	21:26
Breath Volume	2.996	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STRAIN

DATE/TIME OF TEST: 10/19/2022 02:26:43

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.234	02:22
Breath Volume	2.585	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.230	02:25
Breath Volume	2.000	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 10/22/2022 00:59:21

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.081	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.258	00:54
Breath Volume	3.468	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.254	00:57
Breath Volume	2.160	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/22/2022 01:21:26

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.081	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: HARDESTY

DATE/TIME OF TEST: 10/22/2022 02:41:20

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.119	02:36
Breath Volume	1.886	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:38
Subject Sample # 2	0.126	02:39
Breath Volume	1.656	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/22/2022 04:28:12

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.133	04:23
Breath Volume	2.152	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.122	04:26
Breath Volume	1.230	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEWART,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: TONARELLI

DATE/TIME OF TEST: 10/22/2022 18:26:03

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:23	
Control Test	0.080	18:23	
Air Blank	0.000	18:24	
Subject Sample	# 1	18:24	Subject Test Refused
Breath Volume	0.000	18:24	
Air Blank	0.000	18:25	
Control Test	0.079	18:25	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:26	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: ALONZO BATEN

DATE/TIME OF TEST: 10/23/2022 01:34:31

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.199	01:29
Breath Volume	1.488	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.199	01:32
Breath Volume	1.746	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: GUTIERREZ DUENAS

DATE/TIME OF TEST: 10/23/2022 04:23:00

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.148	04:18
Breath Volume	1.371	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.152	04:21
Breath Volume	1.570	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:23

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: FLETCHER III

DATE/TIME OF TEST: 10/23/2022 04:43:50

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.099	04:39
Breath Volume	1.699	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.103	04:42
Breath Volume	1.273	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: COMM VEHIC
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 10/23/2022 06:05:16

Diagnostics Check	OK	06:01	
Air Blank	0.000	06:01	
Control Test	0.079	06:02	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:03	Subject Test Refused
Breath Volume	0.000	06:03	
Air Blank	0.000	06:04	
Control Test	0.079	06:04	
Air Blank	0.000	06:05	
Diagnostics Check	OK	06:05	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 10/23/2022 18:37:15

Diagnostics Check	OK	18:30
Air Blank	0.000	18:30
Control Test	0.079	18:31
Air Blank	0.000	18:31
Subject Sample # 1	0.187	18:32
Breath Volume	2.652	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.195	18:35
Breath Volume	2.746	18:35
Air Blank	0.000	18:36
Control Test	0.079	18:36
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 10/25/2022 01:02:25

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.136	00:57
Breath Volume	1.957	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.133	01:00
Breath Volume	1.210	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 10/25/2022 03:35:10

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.105	03:30
Breath Volume	1.664	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.113	03:33
Breath Volume	1.296	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ZACKERY

DATE/TIME OF TEST: 10/25/2022 21:36:57

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:29	
Control Test	0.080	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:33	No Sample Provided
Breath Volume	0.000	21:33	
Air Blank	0.000	21:34	
Air Blank	0.000	21:36	
Subject Sample	# 2	21:36	Sequence Aborted
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ARRIAGA GONZALEZ

DATE/TIME OF TEST: 10/26/2022 00:22:20

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.234	00:17
Breath Volume	2.031	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.238	00:20
Breath Volume	1.824	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: SQUIRES

DATE/TIME OF TEST: 10/26/2022 09:06:04

Diagnostics Check	OK	08:54	
Air Blank	0.000	08:54	
Control Test	0.079	08:55	
Air Blank	0.000	08:55	
Subject Sample	# 1 0.185	08:58	Volume Not Met
Breath Volume	0.222	08:58	
Air Blank	0.000	08:59	
Air Blank	0.000	09:01	
Subject Sample	# 2 0.173	09:04	Volume Not Met
Breath Volume	0.066	09:04	
Air Blank	0.000	09:05	
Control Test	0.079	09:05	
Air Blank	0.000	09:05	
Diagnostics Check	OK	09:06	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENDY,HALEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: SQUIRES

DATE/TIME OF TEST: 10/26/2022 09:19:05

Diagnostics Check	OK	09:07	
Air Blank	0.000	09:07	
Control Test	0.077	09:08	
Air Blank	0.000	09:08	
Subject Sample	# 1 0.175	09:11	Volume Not Met
Breath Volume	0.160	09:11	
Air Blank	0.000	09:12	
Air Blank	0.000	09:14	
Subject Sample	# 2 0.147	09:17	Volume Not Met
Breath Volume	0.183	09:17	
Air Blank	0.000	09:18	
Control Test	0.077	09:18	
Air Blank	0.000	09:19	
Diagnostics Check	OK	09:19	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: SANCHEZ DELGADO

DATE/TIME OF TEST: 10/26/2022 23:09:32

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.078	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.113	23:04
Breath Volume	2.417	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.114	23:08
Breath Volume	1.691	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE, G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MARTE

DATE/TIME OF TEST: 10/27/2022 03:33:42

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.167	03:27
Breath Volume	1.601	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.168	03:32
Breath Volume	1.503	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ODAMTTEN

DATE/TIME OF TEST: 10/28/2022 03:40:27

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:39	
Control Test	0.079	03:39	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRAN,N
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/28/2022 16:16:05

Diagnostics Check	OK	16:07
Air Blank	0.000	16:07
Control Test	0.080	16:07
Air Blank	0.000	16:08
Subject Sample # 1	0.392	16:10
Breath Volume	1.542	16:10
Air Blank	0.000	16:11
Air Blank	0.000	16:13
Subject Sample # 2	0.372	16:14
Breath Volume	1.578	16:14
Air Blank	0.000	16:15
Control Test	0.077	16:15
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: TORRES ALVARADO

DATE/TIME OF TEST: 10/28/2022 22:08:13

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.161	22:03
Breath Volume	1.839	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 2	0.154	22:06
Breath Volume	1.296	22:06
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 10/29/2022 04:40:01

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.101	04:34
Breath Volume	1.304	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.094	04:38
Breath Volume	2.656	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:40

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GERONIMO SALES

DATE/TIME OF TEST: 10/29/2022 04:54:46

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:53	Slope Not Met
Breath Volume	1.062	04:53	
Air Blank	0.000	04:53	
Control Test	0.078	04:54	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: GERONIMO SALES

DATE/TIME OF TEST: 10/29/2022 05:23:06

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.079	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.159	05:18
Breath Volume	2.054	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:20
Subject Sample # 2	0.165	05:21
Breath Volume	2.371	05:21
Air Blank	0.000	05:22
Control Test	0.078	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SCHAPPER,ANDREA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:06

SUBJECT'S LAST NAME: FRANQUIZ ROSALES

DATE/TIME OF TEST: 10/29/2022 09:34:04

Diagnostics Check	OK	09:27
Air Blank	0.000	09:27
Control Test	0.078	09:27
Air Blank	0.000	09:28
Subject Sample # 1	0.127	09:29
Breath Volume	2.859	09:29
Air Blank	0.000	09:30
Air Blank	0.000	09:32
Subject Sample # 2	0.130	09:32
Breath Volume	2.644	09:32
Air Blank	0.000	09:33
Control Test	0.079	09:33
Air Blank	0.000	09:33
Diagnostics Check	OK	09:34

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RODRIGUEZ-FERNANDEZ

DATE/TIME OF TEST: 10/30/2022 02:41:25

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.218	02:36
Breath Volume	2.214	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.214	02:39
Breath Volume	1.511	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: FRYE

DATE/TIME OF TEST: 10/30/2022 03:40:24

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.176	03:35
Breath Volume	1.984	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.160	03:38
Breath Volume	1.863	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MAMADOULTAIBOU

DATE/TIME OF TEST: 10/30/2022 03:59:21

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.223	03:54
Breath Volume	1.449	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.204	03:57
Breath Volume	1.636	03:57
Air Blank	0.000	03:58
Control Test	0.077	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 10/30/2022 21:12:19

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
Control Test	0.079	21:08	
Air Blank	0.000	21:09	
Subject Sample	# 1	21:10	Subject Test Refused
Breath Volume	0.000	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

Data Download Date/Time: 11/01/2022 11:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GOMEZ LOURIDO

DATE/TIME OF TEST: 10/30/2022 22:33:45

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.236	22:28
Breath Volume	2.335	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.228	22:32
Breath Volume	2.886	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 11/01/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: CAPOTE PEREZ

DATE/TIME OF TEST: 10/31/2022 04:28:52

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.248	04:24
Breath Volume	3.109	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.261	04:27
Breath Volume	2.761	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 11/01/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GUERRERO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: VEGA HERNANDEZ

DATE/TIME OF TEST: 10/31/2022 23:47:32

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.080	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.183	23:42
Breath Volume	1.722	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.173	23:45
Breath Volume	1.476	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 11/01/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: ESCOBAR GARCIA

DATE/TIME OF TEST: 10/09/2022 23:24:47

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.170	23:20
Breath Volume	2.500	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.168	23:23
Breath Volume	2.761	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 10/31/2022 16:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: DRAPAK

DATE/TIME OF TEST: 10/14/2022 02:52:29

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.078	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 10/31/2022 16:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: FERRAS

DATE/TIME OF TEST: 10/23/2022 06:43:12

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.081	06:37
Air Blank	0.000	06:37
Subject Sample # 1	0.075	06:38
Breath Volume	2.613	06:38
Air Blank	0.000	06:39
Air Blank	0.000	06:41
Subject Sample # 2	0.074	06:41
Breath Volume	3.703	06:41
Air Blank	0.000	06:42
Control Test	0.079	06:42
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 10/31/2022 16:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: MCALISTER

DATE/TIME OF TEST: 10/23/2022 21:45:26

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.076	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.100	21:40
Breath Volume	1.718	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.097	21:43
Breath Volume	1.683	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 10/31/2022 16:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK,DAVID,J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BIECK,DAVID
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BRASCOMB

DATE/TIME OF TEST: 10/02/2022 02:07:23

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: WHELCHER

DATE/TIME OF TEST: 10/04/2022 00:32:55

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.081	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: FLORENCE

DATE/TIME OF TEST: 10/07/2022 03:24:41

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: NESS

DATE/TIME OF TEST: 10/15/2022 02:36:57

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.158	02:32
Breath Volume	3.632	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.144	02:35
Breath Volume	3.031	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ALBRECHT,SYDNEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MEZA CRUZ

DATE/TIME OF TEST: 10/15/2022 04:03:56

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.151	03:59
Breath Volume	3.968	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.146	04:02
Breath Volume	2.128	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: YAKOUB

DATE/TIME OF TEST: 10/16/2022 00:50:12

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.081	00:40	
Air Blank	0.000	00:41	
Subject Sample	# 1 0.112	00:44	Volume Not Met
Breath Volume	0.281	00:44	
Air Blank	0.000	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 2	00:48	Subject Test Refused
Breath Volume	0.363	00:48	
Air Blank	0.000	00:49	
Control Test	0.080	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TYLER, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS, TYLER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 10/19/2022 00:20:46

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.081	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Control Test	0.082	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BLAND

DATE/TIME OF TEST: 10/23/2022 03:59:35

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.160	03:55
Breath Volume	2.296	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.158	03:58
Breath Volume	1.765	03:58
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 11/05/2022 22:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK, DAVID, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BIECK, DAVID
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/23/2022 05:22:34

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.118	05:18
Breath Volume	1.609	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.135	05:21
Breath Volume	2.281	05:21
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 11/05/2022 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CHAVARRO

DATE/TIME OF TEST: 10/27/2022 02:38:56

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 11/05/2022 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TYLER, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: STIRLING, MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GARCIA PRADO

DATE/TIME OF TEST: 10/29/2022 04:39:39

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:35	
Control Test	0.081	04:36	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 11/05/2022 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: STIRLING,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GOMEZ PAGAN

DATE/TIME OF TEST: 10/30/2022 04:27:29

Diagnostics Check	OK	04:26	
Air Blank		04:27	RFI Detect
Air Blank	0.000	04:27	

Data Download Date/Time: 11/05/2022 22:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: STIRLING,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 09/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GOMEZ PAGAN

DATE/TIME OF TEST: 10/30/2022 04:35:13

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.209	04:30
Breath Volume	1.878	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.207	04:33
Breath Volume	2.417	04:33
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 11/05/2022 22:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: VICI

DATE/TIME OF TEST: 10/02/2022 05:18:00

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.178	05:13
Breath Volume	2.281	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:15
Subject Sample # 2	0.175	05:16
Breath Volume	2.621	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:18

Data Download Date/Time: 10/28/2022 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLENECK,JORDAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK,JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 10/02/2022 05:52:32

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.081	05:46
Air Blank	0.000	05:46
Subject Sample # 1	0.148	05:47
Breath Volume	3.222	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.139	05:51
Breath Volume	2.488	05:51
Air Blank	0.000	05:51
Control Test	0.080	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 10/28/2022 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLENECK,JORDAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GLENECK,JORDAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 10/02/2022 23:38:38

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.081	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.080	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 10/28/2022 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:29

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 10/04/2022 13:54:44

Diagnostics Check	OK	13:50	
Air Blank	0.000	13:51	
Control Test	0.080	13:51	
Air Blank	0.000	13:52	
Subject Sample	# 1	13:53	Subject Test Refused
Breath Volume	0.000	13:53	
Air Blank	0.000	13:53	
Control Test	0.080	13:54	
Air Blank	0.000	13:54	
Diagnostics Check	OK	13:54	

Data Download Date/Time: 10/28/2022 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 10/04/2022 20:33:51

Diagnostics Check	OK	20:31	
Air Blank	0.000	20:31	
Control Test	0.080	20:31	
Air Blank	0.000	20:32	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.080	20:33	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 10/28/2022 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,LACY,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GEORGE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 06:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/05/2022 08:06:04

Diagnostics Check	OK	08:00
Air Blank	0.000	08:00
Control Test	0.081	08:00
Air Blank	0.000	08:01
Subject Sample # 1	0.000	08:01
Breath Volume	3.707	08:01
Air Blank	0.000	08:02
Air Blank	0.000	08:04
Subject Sample # 2	0.000	08:04
Breath Volume	3.117	08:04
Air Blank	0.000	08:05
Control Test	0.081	08:05
Air Blank	0.000	08:05
Diagnostics Check	OK	08:06

Data Download Date/Time: 10/28/2022 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PACK

DATE/TIME OF TEST: 10/06/2022 01:52:22

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.228	01:47
Breath Volume	2.613	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.221	01:50
Breath Volume	2.410	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 10/28/2022 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: TAN

DATE/TIME OF TEST: 10/07/2022 03:02:59

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.012	02:58
Breath Volume	2.425	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.012	03:01
Breath Volume	2.488	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 10/28/2022 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/08/2022 03:17:03

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.076	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 10/28/2022 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DAVEY

DATE/TIME OF TEST: 10/10/2022 00:07:41

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.081	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 10/28/2022 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/14/2022 23:00:12

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.203	22:52
Breath Volume	1.570	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.239	22:55
Breath Volume	1.953	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 3	0.219	22:58
Breath Volume	1.617	22:58
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 10/28/2022 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ANNABLE

DATE/TIME OF TEST: 10/15/2022 02:43:31

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 10/28/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINT,PATRICK,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REGISTER,JASON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-7
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 10/15/2022 03:14:31

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.127	03:10
Breath Volume	2.597	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.123	03:13
Breath Volume	2.070	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 10/28/2022 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: SABINO

DATE/TIME OF TEST: 10/15/2022 05:38:38

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.176	05:33
Breath Volume	3.582	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.172	05:37
Breath Volume	2.871	05:37
Air Blank	0.000	05:37
Control Test	0.079	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 10/28/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPEE,WYATT,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LAPEE,WYATT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/15/2022 09:25:19

Diagnostics Check	OK	09:15
Air Blank	0.000	09:16
Control Test	0.081	09:16
Air Blank	0.000	09:17
Subject Sample # 1	0.129	09:17
Breath Volume	4.011	09:17
Air Blank	0.000	09:18
Air Blank	0.000	09:20
Subject Sample # 2	0.108	09:20
Breath Volume	1.585	09:20
Air Blank	0.000	09:21
Air Blank	0.000	09:23
Subject Sample # 3	0.133	09:23
Breath Volume	3.148	09:23
Air Blank	0.000	09:24
Control Test	0.080	09:24
Air Blank	0.000	09:25
Diagnostics Check	OK	09:25

Data Download Date/Time: 10/28/2022 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 10/15/2022 22:11:42

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.119	22:07
Breath Volume	2.679	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.125	22:10
Breath Volume	4.105	22:10
Air Blank	0.000	22:10
Control Test	0.080	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 10/28/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF, KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: TRAINOR

DATE/TIME OF TEST: 10/16/2022 01:48:28

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.081	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 10/28/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUN DORIA,NORMA,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAUN DORIA,NORMA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: O325342-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 10/16/2022 03:35:37

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.184	03:31
Breath Volume	1.250	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.181	03:34
Breath Volume	2.039	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 10/28/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-5
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/16/2022 04:36:54

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1 0.134	04:31	Volume Not Met
Breath Volume	0.398	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 2	04:35	Subject Test Refused
Breath Volume	0.179	04:35	
Air Blank	0.000	04:35	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 10/28/2022 15:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,AARON,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CHAPMAN,AARON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 10/17/2022 05:47:39

Diagnostics Check	OK	05:47	
Air Blank		05:47	RFI Detect
Air Blank	0.000	05:47	

Data Download Date/Time: 10/28/2022 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,AARON,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CHAPMAN,AARON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 10/17/2022 05:55:22

Diagnostics Check	OK	05:49
Air Blank	0.000	05:49
Control Test	0.081	05:50
Air Blank	0.000	05:50
Subject Sample # 1	0.000	05:51
Breath Volume	2.414	05:51
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.000	05:54
Breath Volume	1.921	05:54
Air Blank	0.000	05:54
Control Test	0.080	05:54
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 10/28/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 10/21/2022 23:40:58

Diagnostics Check	OK	23:40	
Air Blank		23:40	RFI Detect
Air Blank		23:40	RFI Detect

Data Download Date/Time: 10/28/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 10/21/2022 23:46:00

Diagnostics Check	OK	23:45	
Air Blank		23:45	RFI Detect
Air Blank		23:46	RFI Detect

Data Download Date/Time: 10/28/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 10/21/2022 23:53:14

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank		23:53	RFI Detect
Air Blank		23:53	RFI Detect

Data Download Date/Time: 10/28/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 10/22/2022 00:00:19

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.081	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.081	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 10/28/2022 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUNDIN,GABREAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 10/22/2022 06:21:51

Diagnostics Check	OK	06:06	
Air Blank	0.000	06:06	
Control Test	0.081	06:06	
Air Blank	0.000	06:07	
Subject Sample	# 1 0.225	06:10	Volume Not Met
Breath Volume	0.109	06:10	
Air Blank	0.000	06:11	
Air Blank	0.000	06:13	
Subject Sample	# 2 0.235	06:14	
Breath Volume	2.257	06:14	
Air Blank	0.000	06:15	
Air Blank	0.000	06:17	
Subject Sample	# 3 0.238	06:20	Volume Not Met
Breath Volume	0.546	06:20	
Air Blank	0.000	06:20	
Control Test	0.080	06:21	
Air Blank	0.000	06:21	
Diagnostics Check	OK	06:21	

Data Download Date/Time: 10/28/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNIS,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HENNIS,TIMOTHY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 10/27/2022 23:37:07

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.150	23:32
Breath Volume	2.460	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:34
Subject Sample # 2	0.158	23:35
Breath Volume	3.117	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 10/28/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO
HERNANDEZ,GONZALO,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024

VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MORENO HERNANDEZ

DATE/TIME OF TEST: 10/27/2022 23:51:54

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.158	23:47
Breath Volume	1.800	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.163	23:50
Breath Volume	2.539	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/28/2022 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 09/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: SHIRAH

DATE/TIME OF TEST: 10/28/2022 02:14:19

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.175	02:09
Breath Volume	1.707	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.167	02:12
Breath Volume	1.566	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 10/28/2022 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: SALAZAR DIAZ

DATE/TIME OF TEST: 10/29/2022 05:13:55

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:02	
Control Test	0.080	05:03	
Air Blank	0.000	05:03	
Subject Sample	# 1 0.167	05:06	Volume Not Met
Breath Volume	0.136	05:06	
Air Blank	0.000	05:07	
Air Blank	0.000	05:09	
Subject Sample	# 2	05:12	No Sample Provided
Breath Volume	0.000	05:12	
Air Blank	0.000	05:12	
Control Test	0.078	05:13	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 11/23/2022 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: NEYRA

DATE/TIME OF TEST: 10/29/2022 21:24:36

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:16	
Control Test	0.080	21:16	
Air Blank	0.000	21:17	
Subject Sample	# 1	21:20	No Sample Provided
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.080	21:24	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 11/23/2022 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 10/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CANTIN

DATE/TIME OF TEST: 10/31/2022 00:14:09

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.109	00:09
Breath Volume	1.890	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.106	00:12
Breath Volume	2.234	00:12
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 11/23/2022 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PETERS,KRISTY,L

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: PETERS,KRISTY

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: DUI

ARREST TIME: 11:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/13/2022 00:30:19

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.083	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:26
Breath Volume	3.667	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:28
Breath Volume	3.621	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 12/20/2022 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PETERS,KRISTY,L

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: PETERS,KRISTY

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: DUI

ARREST TIME: 19:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/18/2022 20:23:33

Diagnostics Check	OK	20:20	
Air Blank	0.000	20:20	
Control Test	0.083	20:21	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:22	Subject Test Refused
Breath Volume	0.000	20:22	
Air Blank	0.000	20:22	
Control Test	0.082	20:22	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:23	

Data Download Date/Time: 12/20/2022 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PETERS,KRISTY,L

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: PETERS,KRISTY

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 09/22/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1249794

DRY GAS EXPIRY DATE: 12/13/2022

VIOLATION CODE: DUI

ARREST TIME: 00:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 10/22/2022 01:21:24

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.205	01:16
Breath Volume	3.007	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.207	01:19
Breath Volume	1.972	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 12/20/2022 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 10/19/2022 01:40:10

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.114	01:35
Breath Volume	1.894	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.114	01:38
Breath Volume	2.070	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 10/23/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: GOSLER

DATE/TIME OF TEST: 10/27/2022 21:01:05

Diagnostics Check	OK	20:54
Air Blank	0.000	20:55
Control Test	0.080	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.194	20:56
Breath Volume	2.984	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.196	20:59
Breath Volume	3.500	20:59
Air Blank	0.000	21:00
Control Test	0.080	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 11/29/2022 23:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 10/28/2022 02:40:40

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.163	02:36
Breath Volume	1.839	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.162	02:39
Breath Volume	1.339	02:39
Air Blank	0.000	02:39
Control Test	0.078	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 11/29/2022 23:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: BATISTA-PETROS

DATE/TIME OF TEST: 10/31/2022 22:29:34

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.201	22:25
Breath Volume	2.281	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.202	22:28
Breath Volume	2.531	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 11/29/2022 23:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 10/14/2022 00:11:46

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.201	00:07
Breath Volume	1.640	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.218	00:10
Breath Volume	2.179	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MULKEY

DATE/TIME OF TEST: 10/14/2022 02:23:41

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.233	02:19
Breath Volume	1.902	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.241	02:22
Breath Volume	1.539	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: KAVCSAK

DATE/TIME OF TEST: 10/14/2022 22:30:43

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.000	22:26
Breath Volume	2.367	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:29
Breath Volume	1.613	22:29
Air Blank	0.000	22:29
Control Test	0.081	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 10/15/2022 03:44:55

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:37	No Sample Provided
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:43	No Sample Provided
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: GONZALEZ JR

DATE/TIME OF TEST: 10/15/2022 22:07:12

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.185	22:02
Breath Volume	2.449	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.180	22:05
Breath Volume	2.207	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 10/15/2022 23:45:47

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.123	23:41
Breath Volume	2.789	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.113	23:44
Breath Volume	1.328	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: LORDE

DATE/TIME OF TEST: 10/16/2022 03:01:03

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.167	02:51	Volume Not Met
Breath Volume	0.105	02:51	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 2 0.188	02:55	
Breath Volume	2.070	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 3 0.177	02:59	
Breath Volume	2.343	02:59	
Air Blank	0.000	03:00	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 10/16/2022 05:07:15

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.130	05:02
Breath Volume	2.832	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.128	05:05
Breath Volume	2.601	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: GUNNIN

DATE/TIME OF TEST: 10/17/2022 00:12:31

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.142	00:07
Breath Volume	1.273	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.144	00:10
Breath Volume	1.285	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MACARIO GONZALEZ

DATE/TIME OF TEST: 10/17/2022 04:52:24

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.186	04:47
Breath Volume	1.515	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.203	04:50
Breath Volume	1.910	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 10/18/2022 23:49:16

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.081	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.012	23:44
Breath Volume	1.414	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.013	23:47
Breath Volume	1.941	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MARTINEZ MARTINEZ

DATE/TIME OF TEST: 10/19/2022 04:24:35

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.165	04:19
Breath Volume	2.539	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.152	04:23
Breath Volume	1.628	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 10/19/2022 23:04:45

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.081	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.198	22:59
Breath Volume	1.746	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.198	23:03
Breath Volume	1.550	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 11/01/2022 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BAUTISTA BRITO

DATE/TIME OF TEST: 10/22/2022 02:36:26

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.116	02:31
Breath Volume	1.722	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.109	02:34
Breath Volume	1.671	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 10/22/2022 02:55:41

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.091	02:51
Breath Volume	2.816	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.090	02:54
Breath Volume	2.785	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: PARCHMON

DATE/TIME OF TEST: 10/22/2022 04:25:37

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.218	04:19
Breath Volume	1.351	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.220	04:24
Breath Volume	1.687	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 10/22/2022 06:32:57

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.080	06:26
Air Blank	0.000	06:26
Subject Sample # 1	0.102	06:28
Breath Volume	1.500	06:28
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.102	06:31
Breath Volume	1.355	06:31
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: LOZANO-TRUJILLO

DATE/TIME OF TEST: 10/23/2022 01:36:43

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.199	01:28
Breath Volume	1.351	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.220	01:31
Breath Volume	1.593	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 3	0.219	01:35
Breath Volume	1.253	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RIVERA,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: AVILA GONZALEZ

DATE/TIME OF TEST: 10/23/2022 05:05:31

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:51	
Subject Sample	# 1 0.209	04:55	Volume Not Met
Breath Volume	0.906	04:55	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 2 0.215	05:00	
Breath Volume	1.695	05:00	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample	# 3 0.215	05:03	
Breath Volume	1.609	05:03	
Air Blank	0.000	05:04	
Control Test	0.078	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RESENDIZ DE SANTIAGO

DATE/TIME OF TEST: 10/24/2022 02:29:06

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.192	02:24
Breath Volume	2.378	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.193	02:27
Breath Volume	2.382	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: STARNES

DATE/TIME OF TEST: 10/24/2022 04:44:13

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.194	04:36
Breath Volume	3.011	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample # 2	0.172	04:39
Breath Volume	2.359	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 3	0.200	04:42
Breath Volume	4.480	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BARLAMAS

DATE/TIME OF TEST: 10/26/2022 03:10:20

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILCHER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LAYNE

DATE/TIME OF TEST: 10/27/2022 01:35:23

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:27	No Sample Provided
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Air Blank	0.000	01:29	
Subject Sample	# 2	0.139	01:30
Breath Volume	2.226	01:30	
Air Blank	0.000	01:31	
Air Blank	0.000	01:33	
Subject Sample	# 3	0.128	01:33
Breath Volume	1.789	01:33	
Air Blank	0.000	01:34	
Control Test	0.079	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WELLS,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: BAPTIST JR

DATE/TIME OF TEST: 10/27/2022 04:47:59

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.000	04:43
Breath Volume	1.292	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.000	04:46
Breath Volume	2.031	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 10/28/2022 22:21:26

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.250	22:13
Breath Volume	1.839	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.219	22:16
Breath Volume	1.683	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 3	0.226	22:19
Breath Volume	1.851	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: HILL III

DATE/TIME OF TEST: 10/29/2022 00:41:20

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.000	00:36
Breath Volume	2.171	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.000	00:39
Breath Volume	1.332	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: VESANILLA GUERRERO

DATE/TIME OF TEST: 10/29/2022 05:56:40

Diagnostics Check	OK	05:49
Air Blank	0.000	05:50
Control Test	0.080	05:50
Air Blank	0.000	05:51
Subject Sample # 1	0.147	05:51
Breath Volume	1.535	05:51
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.156	05:55
Breath Volume	1.640	05:55
Air Blank	0.000	05:55
Control Test	0.080	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 10/30/2022 05:45:27

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.079	05:39
Air Blank	0.000	05:39
Subject Sample # 1	0.158	05:40
Breath Volume	2.382	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:42
Subject Sample # 2	0.163	05:43
Breath Volume	2.093	05:43
Air Blank	0.000	05:44
Control Test	0.079	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: PEREZ DIAZ

DATE/TIME OF TEST: 10/30/2022 19:50:32

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:39	
Control Test	0.080	19:39	
Air Blank	0.000	19:40	
Subject Sample	# 1	19:43	No Sample Provided
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Air Blank	0.000	19:46	
Subject Sample	# 2	19:49	No Sample Provided
Breath Volume	0.000	19:49	
Air Blank	0.000	19:49	
Control Test	0.080	19:49	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 10/31/2022 06:33:14

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.080	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.205	06:28
Breath Volume	2.941	06:28
Air Blank	0.000	06:29
Air Blank	0.000	06:31
Subject Sample # 2	0.204	06:31
Breath Volume	2.394	06:31
Air Blank	0.000	06:32
Control Test	0.079	06:32
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 10/31/2022 22:07:50

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.222	22:03
Breath Volume	1.328	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.234	22:06
Breath Volume	1.906	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 11/01/2022 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BREWSTER

DATE/TIME OF TEST: 10/01/2022 02:54:01

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 2	0.197	02:49
Breath Volume	2.332	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 3	0.208	02:52
Breath Volume	2.464	02:52	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 10/25/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BAR-JOSEPH

DATE/TIME OF TEST: 10/01/2022 04:03:29

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:56	No Sample Provided
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2	04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOND,W
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/04/2022 14:18:36

Diagnostics Check	OK	14:09	
Air Blank	0.000	14:09	
Control Test	0.080	14:10	
Air Blank	0.000	14:10	
Subject Sample	# 1	14:13	No Sample Provided
Breath Volume	0.000	14:13	
Air Blank	0.000	14:14	
Air Blank	0.000	14:16	
Subject Sample	# 2	14:17	Subject Test Refused
Breath Volume	0.000	14:17	
Air Blank	0.000	14:17	
Control Test	0.081	14:18	
Air Blank	0.000	14:18	
Diagnostics Check	OK	14:18	

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEMARIA,JAMES
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/05/2022 02:59:54

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.232	02:54
Breath Volume	1.339	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.246	02:58
Breath Volume	1.875	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: VONGKANYA

DATE/TIME OF TEST: 10/06/2022 15:44:18

Diagnostics Check	OK	15:37
Air Blank	0.000	15:38
Control Test	0.080	15:38
Air Blank	0.000	15:39
Subject Sample # 1	0.000	15:39
Breath Volume	1.167	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:42
Subject Sample # 2	0.000	15:42
Breath Volume	2.562	15:42
Air Blank	0.000	15:43
Control Test	0.079	15:43
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BRADDY

DATE/TIME OF TEST: 10/07/2022 02:00:11

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/07/2022 21:57:41

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.208	21:53
Breath Volume	1.718	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.206	21:56
Breath Volume	2.070	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MAYNARD JR

DATE/TIME OF TEST: 10/08/2022 00:13:32

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.222	00:08
Breath Volume	1.277	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.212	00:11
Breath Volume	1.179	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: MORMON

DATE/TIME OF TEST: 10/08/2022 23:30:35

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:19	
Control Test	0.080	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:23	No Sample Provided
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample	# 2	23:29	No Sample Provided
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKASAL,TYVON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JACKASAL,TYVON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BARTHELEMY

DATE/TIME OF TEST: 10/10/2022 23:46:54

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:42	
Control Test	0.081	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:45	Interferent Detect
Breath Volume	1.464	23:45	
Air Blank	0.000	23:45	
Control Test	0.081	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKASAL, TYVON, T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JACKASAL, TYVON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BARTHELEMY

DATE/TIME OF TEST: 10/10/2022 23:57:15

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.080	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 1	23:56	Interferent Detect
Breath Volume	2.242	23:56	
Air Blank	0.000	23:56	
		23:56	RFI Detect
Air Blank	0.000	23:57	

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU, A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: RIVERA-ARELLANO

DATE/TIME OF TEST: 10/12/2022 01:50:38

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.081	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.246	01:46
Breath Volume	2.746	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.247	01:48
Breath Volume	2.750	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/12/2022 17:58:23

Diagnostics Check	OK	17:47	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Subject Sample	# 1	17:51	No Sample Provided
Breath Volume	0.000	17:51	
Air Blank	0.000	17:51	
Air Blank	0.000	17:53	
Subject Sample	# 2	17:57	No Sample Provided
Breath Volume	0.000	17:57	
Air Blank	0.000	17:57	
Control Test	0.080	17:57	
Air Blank	0.000	17:58	
Diagnostics Check	OK	17:58	

Data Download Date/Time: 10/25/2022 13:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BRANDER

DATE/TIME OF TEST: 10/13/2022 02:15:21

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:04	
Control Test	0.081	02:04	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:08	No Sample Provided
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Air Blank	0.000	02:10	
Subject Sample	# 2	02:13	No Sample Provided
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.082	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: STUBBINGS

DATE/TIME OF TEST: 10/13/2022 20:47:30

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.143	20:42
Breath Volume	1.679	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.142	20:45
Breath Volume	1.687	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MCAFFEE

DATE/TIME OF TEST: 10/13/2022 23:57:51

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.000	23:53
Breath Volume	2.476	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:56
Breath Volume	1.851	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: PETITT

DATE/TIME OF TEST: 10/14/2022 06:18:17

Diagnostics Check	OK	06:15	
Air Blank	0.000	06:15	
Control Test	0.081	06:15	
Air Blank	0.000	06:16	
Subject Sample	# 1	06:16	Subject Test Refused
Breath Volume	0.000	06:16	
Air Blank	0.000	06:17	
Control Test	0.081	06:17	
Air Blank	0.000	06:18	
Diagnostics Check	OK	06:18	

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:27

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 10/14/2022 13:53:23

Diagnostics Check	OK	13:47
Air Blank	0.000	13:47
Control Test	0.081	13:47
Air Blank	0.000	13:48
Subject Sample # 1	0.000	13:49
Breath Volume	2.933	13:49
Air Blank	0.000	13:49
Air Blank	0.000	13:51
Subject Sample # 2	0.000	13:52
Breath Volume	2.875	13:52
Air Blank	0.000	13:52
Control Test	0.078	13:52
Air Blank	0.000	13:53
Diagnostics Check	OK	13:53

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY III,KEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BARTOLOMEI

DATE/TIME OF TEST: 10/15/2022 03:44:29

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.081	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SESSIONS,C,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/15/2022 22:10:31

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.336	22:01
Breath Volume	1.457	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.312	22:05
Breath Volume	1.585	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 3	0.315	22:08
Breath Volume	1.578	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSAM,MATTHEW,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K.
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 10/16/2022 00:40:57

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.208	00:35
Breath Volume	2.671	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.188	00:39
Breath Volume	1.851	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 10/16/2022 09:46:54

Diagnostics Check	OK	09:40
Air Blank	0.000	09:41
Control Test	0.081	09:41
Air Blank	0.000	09:42
Subject Sample # 1	0.000	09:42
Breath Volume	1.679	09:42
Air Blank	0.000	09:43
Air Blank	0.000	09:45
Subject Sample # 2	0.000	09:45
Breath Volume	1.519	09:45
Air Blank	0.000	09:45
Control Test	0.079	09:46
Air Blank	0.000	09:46
Diagnostics Check	OK	09:46

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SCOTT,BRANDON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: MENDONCA

DATE/TIME OF TEST: 10/17/2022 02:32:27

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.144	02:22
Breath Volume	1.648	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.165	02:26
Breath Volume	1.519	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 3	0.169	02:30
Breath Volume	2.535	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWEIGER,JONATHAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/18/2022 00:07:11

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.115	00:02
Breath Volume	3.355	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.114	00:05
Breath Volume	3.218	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/19/2022 22:56:57

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.082	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.209	22:52
Breath Volume	2.054	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.194	22:55
Breath Volume	1.433	22:55
Air Blank	0.000	22:56
Control Test	0.082	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 10/25/2022 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH, KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS, KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SHIRLEY

DATE/TIME OF TEST: 10/20/2022 02:33:01

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.081	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.000	02:28
Breath Volume	1.910	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.000	02:31
Breath Volume	2.503	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:33

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: OSORIO

DATE/TIME OF TEST: 10/21/2022 01:57:43

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.152	01:53
Breath Volume	1.437	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.155	01:56
Breath Volume	1.703	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DEMARIA,JAMES
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 10/21/2022 04:04:39

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.081	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.082	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: O'CONNELL,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 10/21/2022 10:49:03

Diagnostics Check	OK	10:42
Air Blank	0.000	10:43
Control Test	0.081	10:43
Air Blank	0.000	10:44
Subject Sample # 1	0.053	10:44
Breath Volume	1.746	10:44
Air Blank	0.000	10:45
Air Blank	0.000	10:47
Subject Sample # 2	0.049	10:47
Breath Volume	2.578	10:47
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:48
Diagnostics Check	OK	10:49

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/21/2022 22:56:52

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:48	
Control Test	0.081	22:49	
Air Blank	0.000	22:49	
Subject Sample	# 1	22:52	No Sample Provided
Breath Volume	0.000	22:52	
Air Blank	0.000	22:53	
Air Blank	0.000	22:55	
Subject Sample	# 2	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:55	
Control Test	0.082	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,N
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: PARTON

DATE/TIME OF TEST: 10/22/2022 00:05:22

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.081	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:01
Breath Volume	2.445	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.000	00:04
Breath Volume	2.589	00:04
Air Blank	0.000	00:04
Control Test	0.082	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 10/22/2022 00:34:59

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.081	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HERNANDEZ RAMIREZ

DATE/TIME OF TEST: 10/22/2022 03:41:15

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.081	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACA,E,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MCBETH

DATE/TIME OF TEST: 10/22/2022 03:56:04

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.081	03:54	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Control Test	0.081	03:55	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 10/25/2022 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 10/23/2022 00:49:09

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.081	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:42	No Sample Provided
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Air Blank	0.000	00:44	
Subject Sample	# 2	00:47	No Sample Provided
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 10/25/2022 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/23/2022 02:08:34

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.000	02:04
Breath Volume	1.414	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.000	02:07
Breath Volume	1.328	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 10/25/2022 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 10/23/2022 15:51:10

Diagnostics Check	OK	15:44
Air Blank	0.000	15:45
Control Test	0.081	15:45
Air Blank	0.000	15:45
Subject Sample # 1	0.212	15:46
Breath Volume	2.175	15:46
Air Blank	0.000	15:47
Air Blank	0.000	15:49
Subject Sample # 2	0.230	15:49
Breath Volume	2.417	15:49
Air Blank	0.000	15:50
Control Test	0.081	15:50
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

Data Download Date/Time: 10/25/2022 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON, DAVID, J
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: HOFFMAN, CHRISTINE
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD
LAST AGENCY INSPECTION: 10/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: GUERRERO SALINAS

DATE/TIME OF TEST: 10/30/2022 05:25:33

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.082	05:19
Air Blank	0.000	05:20
Subject Sample # 1	0.060	05:20
Breath Volume	4.148	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.059	05:23
Breath Volume	4.199	05:23
Air Blank	0.000	05:24
Control Test	0.078	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 11/17/2022 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: BEISELL, BRYAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: ARDUCA

DATE/TIME OF TEST: 10/06/2022 05:13:35

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.230	05:08
Breath Volume	2.261	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.232	05:11
Breath Volume	1.195	05:11
Air Blank	0.000	05:12
Control Test	0.078	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 10/18/2022 17:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AMATRUDA,CHRIS,G

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: SERRANO,VICENTE

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 00:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MOCTEZUMA

DATE/TIME OF TEST: 10/16/2022 00:44:24

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.215	00:39
Breath Volume	1.500	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.214	00:42
Breath Volume	2.933	00:42
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/18/2022 17:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AMATRUDA,CHRIS,G

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: STEMOM,ANDREW

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 09/15/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 03:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ORTUNO

DATE/TIME OF TEST: 10/16/2022 04:45:12

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.184	04:40
Breath Volume	1.839	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.187	04:43
Breath Volume	2.804	04:43
Air Blank	0.000	04:44
Control Test	0.078	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 10/18/2022 17:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: CAMPBELL,RYAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: OUSLEY

DATE/TIME OF TEST: 10/29/2022 03:47:38

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.157	03:42
Breath Volume	1.316	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.161	03:46
Breath Volume	1.539	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 11/16/2022 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 09/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SCPIO

DATE/TIME OF TEST: 10/01/2022 03:08:39

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.138	03:03
Breath Volume	1.640	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.128	03:06
Breath Volume	1.296	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 10/06/2022 15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/09/2022 03:05:51

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.167	03:00
Breath Volume	1.835	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.165	03:04
Breath Volume	1.281	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 11/21/2022 03:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE, AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DURRENCE

DATE/TIME OF TEST: 10/09/2022 03:15:58

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 11/21/2022 03:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: EMIG

DATE/TIME OF TEST: 10/15/2022 03:07:20

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.132	03:02
Breath Volume	3.621	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.127	03:05
Breath Volume	3.785	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 11/21/2022 03:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MATULICH

DATE/TIME OF TEST: 10/16/2022 21:08:25

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.081	21:06	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Control Test	0.080	21:07	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

Data Download Date/Time: 11/21/2022 03:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/17/2022 03:54:00

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Subject Sample # 1	0.146	03:42	
Breath Volume	1.902	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample # 2	0.123	03:48	Volume Not Met
Breath Volume	0.039	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample # 3	0.160	03:52	
Breath Volume	3.300	03:52	
Air Blank	0.000	03:53	
Control Test	0.077	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 11/21/2022 03:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,M
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/30/2022 01:58:00

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.205	01:52
Breath Volume	1.250	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.197	01:56
Breath Volume	1.734	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:58

Data Download Date/Time: 11/21/2022 03:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: VAZQUEZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/31/2022 03:08:52

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.009	03:04
Breath Volume	1.488	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.008	03:07
Breath Volume	1.589	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 11/21/2022 03:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CALVERT,P
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 10/04/2022 04:31:44

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.076	04:29	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Control Test	0.077	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 10/06/2022 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PLAZA

DATE/TIME OF TEST: 10/07/2022 05:27:01

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.077	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.127	05:22
Breath Volume	1.140	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.142	05:25
Breath Volume	1.128	05:25
Air Blank	0.000	05:26
Control Test	0.076	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:27

Data Download Date/Time: 11/03/2022 03:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SOTOLONGO

DATE/TIME OF TEST: 10/08/2022 01:05:28

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.078	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.217	01:00
Breath Volume	3.023	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.216	01:03
Breath Volume	2.597	01:03
Air Blank	0.000	01:04
Control Test	0.077	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 11/03/2022 03:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGLIA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ROSAL

DATE/TIME OF TEST: 10/09/2022 01:23:47

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.077	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.076	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 11/03/2022 03:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BORRELL,J
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: HERNANDEZ MARTINEZ

DATE/TIME OF TEST: 10/09/2022 04:53:20

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.109	04:48
Breath Volume	1.902	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.107	04:51
Breath Volume	1.492	04:51
Air Blank	0.000	04:52
Control Test	0.078	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 11/03/2022 03:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SANCHEZ,ELIZABETH
ARRESTING OFFICER AGENCY: KEY BISCAYNE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VINGERHOETS

DATE/TIME OF TEST: 10/16/2022 01:06:34

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.076	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.198	01:01
Breath Volume	2.250	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.192	01:04
Breath Volume	2.019	01:04
Air Blank	0.000	01:05
Control Test	0.076	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 11/03/2022 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CHERASIA,E
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 10/22/2022 07:25:34

Diagnostics Check	OK	07:22	
Air Blank	0.000	07:22	
Control Test	0.078	07:23	
Air Blank	0.000	07:23	
Subject Sample	# 1	07:24	Subject Test Refused
Breath Volume	0.000	07:24	
Air Blank	0.000	07:24	
Control Test	0.078	07:24	
Air Blank	0.000	07:25	
Diagnostics Check	OK	07:25	

Data Download Date/Time: 11/03/2022 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGILA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/28/2022 03:44:38

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.077	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.148	03:39
Breath Volume	1.171	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.156	03:42
Breath Volume	1.363	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 11/03/2022 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CARO

DATE/TIME OF TEST: 10/29/2022 03:55:16

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.120	03:50
Breath Volume	2.386	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.111	03:53
Breath Volume	2.253	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 11/03/2022 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RODRIGUEZ BALLATE

DATE/TIME OF TEST: 10/30/2022 00:44:33

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.078	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.167	00:39
Breath Volume	1.960	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.161	00:42
Breath Volume	1.531	00:42
Air Blank	0.000	00:43
Control Test	0.078	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 11/03/2022 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: DERAS

DATE/TIME OF TEST: 10/30/2022 02:47:20

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.077	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.078	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 11/03/2022 03:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CASTILLO CHAVARRIA

DATE/TIME OF TEST: 10/07/2022 02:52:50

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.080	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.080	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 11/03/2022 03:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: BETTER GONZALEZ

DATE/TIME OF TEST: 10/09/2022 06:22:19

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:19	
Control Test	0.079	06:20	
Air Blank	0.000	06:20	
Subject Sample	# 1	06:20	Subject Test Refused
Breath Volume	0.000	06:20	
Air Blank	0.000	06:21	
Control Test	0.079	06:21	
Air Blank	0.000	06:22	
Diagnostics Check	OK	06:22	

Data Download Date/Time: 11/03/2022 03:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,ADRIAN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: PUERTA

DATE/TIME OF TEST: 10/01/2022 00:07:59

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.078	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.141	00:03
Breath Volume	2.578	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.149	00:06
Breath Volume	3.031	00:06
Air Blank	0.000	00:07
Control Test	0.078	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 10/06/2022 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: SANTELISES

DATE/TIME OF TEST: 10/06/2022 06:37:03

Diagnostics Check	OK	06:22	
Air Blank	0.000	06:23	
Control Test	0.079	06:23	
Air Blank	0.000	06:24	
Subject Sample	# 1	06:27	No Sample Provided
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Air Blank	0.000	06:29	
Subject Sample	# 2	0.180	06:32
Breath Volume	2.218	06:32	
Air Blank	0.000	06:32	
Air Blank	0.000	06:34	
Subject Sample	# 3	0.184	06:35
Breath Volume	2.070	06:35	
Air Blank	0.000	06:36	
Control Test	0.077	06:36	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:37	

Data Download Date/Time: 11/10/2022 06:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 10/08/2022 01:27:58

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.205	01:22
Breath Volume	1.156	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.187	01:26
Breath Volume	1.253	01:26
Air Blank	0.000	01:26
Control Test	0.077	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 11/10/2022 06:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 10/09/2022 04:34:24

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.077	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.240	04:27
Breath Volume	1.359	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.259	04:32
Breath Volume	2.312	04:32
Air Blank	0.000	04:33
Control Test	0.076	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 11/10/2022 06:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/14/2022 03:37:44

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1 0.152	03:29	Volume Not Met
Breath Volume	0.070	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2 0.165	03:32	
Breath Volume	1.699	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample	# 3 0.172	03:36	
Breath Volume	1.464	03:36	
Air Blank	0.000	03:36	
Control Test	0.075	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 11/10/2022 06:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: ROJAS-FIESTAS

DATE/TIME OF TEST: 10/16/2022 03:20:50

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.180	03:16
Breath Volume	1.164	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.176	03:19
Breath Volume	1.257	03:19
Air Blank	0.000	03:19
Control Test	0.077	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 11/10/2022 06:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: VILABOA

DATE/TIME OF TEST: 10/28/2022 01:13:31

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.078	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.330	01:08
Breath Volume	3.750	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.323	01:11
Breath Volume	3.496	01:11
Air Blank	0.000	01:12
Control Test	0.077	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 11/10/2022 06:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CAMPUZANO-CARDOZO

DATE/TIME OF TEST: 10/28/2022 05:22:01

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.078	05:14	
Air Blank	0.000	05:14	
Subject Sample	# 1 0.180	05:17	
Breath Volume	2.695	05:17	
Air Blank	0.000	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 2 0.182	05:20	
Breath Volume	1.707	05:20	
Air Blank	0.000	05:20	
Control Test	0.076	05:21	
Air Blank		05:21	RFI Detect
Air Blank	0.000	05:22	

Data Download Date/Time: 11/10/2022 06:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CAMPUZANO-CARDOZO

DATE/TIME OF TEST: 10/28/2022 05:31:18

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.076	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.179	05:26
Breath Volume	2.343	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:28
Subject Sample # 2	0.173	05:29
Breath Volume	2.500	05:29
Air Blank	0.000	05:30
Control Test	0.076	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 11/10/2022 06:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 10/29/2022 02:17:08

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.077	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.130	02:12
Breath Volume	3.761	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.120	02:15
Breath Volume	2.234	02:15
Air Blank	0.000	02:16
Control Test	0.077	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 11/10/2022 06:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIST,JALEN,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LEWIS,MARIO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: GUZMAN SELPA

DATE/TIME OF TEST: 10/14/2022 06:02:29

Diagnostics Check	OK	05:54
Air Blank	0.000	05:55
Control Test	0.082	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.157	05:57
Breath Volume	1.285	05:57
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.152	06:00
Breath Volume	2.078	06:00
Air Blank	0.000	06:01
Control Test	0.077	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 11/23/2022 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIST,JALEN,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: PEPIN,MANUEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: CRUZ BERRIOS

DATE/TIME OF TEST: 10/14/2022 06:14:31

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.080	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.186	06:09
Breath Volume	2.644	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.168	06:12
Breath Volume	1.972	06:12
Air Blank	0.000	06:13
Control Test	0.076	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 11/23/2022 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,COLLIN,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MCDONALD,K
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SENTOWSKI

DATE/TIME OF TEST: 10/24/2022 03:29:22

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.081	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:28	Subject Test Refused
Breath Volume	0.000	03:28	
Air Blank	0.000	03:28	
Control Test	0.081	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 11/23/2022 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,DEPUTY G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: PARADIS

DATE/TIME OF TEST: 10/10/2022 21:43:32

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.079	21:40	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.080	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 10/23/2022 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: AJANEL SONTAY

DATE/TIME OF TEST: 10/01/2022 02:24:14

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.220	02:19
Breath Volume	2.339	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.221	02:22
Breath Volume	2.242	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 10/21/2022 15:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: GONZALEZ LUCAS

DATE/TIME OF TEST: 10/01/2022 22:26:00

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.278	22:21
Breath Volume	2.343	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample # 2	0.273	22:24
Breath Volume	2.808	22:24
Air Blank	0.000	22:25
Control Test	0.085	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:26

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: TWOHIG

DATE/TIME OF TEST: 10/02/2022 04:15:54

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.334	04:11
Breath Volume	1.164	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.354	04:14
Breath Volume	1.746	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROWE, RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/02/2022 12:32:40

Diagnostics Check	OK	12:26
Air Blank	0.000	12:26
Control Test	0.081	12:27
Air Blank	0.000	12:27
Subject Sample # 1	0.218	12:28
Breath Volume	1.695	12:28
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 2	0.215	12:31
Breath Volume	1.707	12:31
Air Blank	0.000	12:31
Control Test	0.080	12:32
Air Blank	0.000	12:32
Diagnostics Check	OK	12:32

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BRUMBAUGH

DATE/TIME OF TEST: 10/05/2022 04:30:39

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.161	04:26
Breath Volume	1.785	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.173	04:29
Breath Volume	1.855	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD,RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PEREZ RAMIREZ

DATE/TIME OF TEST: 10/06/2022 03:38:37

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.216	03:33
Breath Volume	1.519	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.209	03:36
Breath Volume	1.640	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HOER

DATE/TIME OF TEST: 10/06/2022 22:37:05

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.128	22:32
Breath Volume	1.953	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:34
Subject Sample # 2	0.135	22:35
Breath Volume	2.511	22:35
Air Blank	0.000	22:36
Control Test	0.081	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: AJIATAZ CASTRO

DATE/TIME OF TEST: 10/06/2022 23:42:34

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.144	23:37
Breath Volume	1.375	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.154	23:40
Breath Volume	1.843	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ROSSO

DATE/TIME OF TEST: 10/07/2022 01:40:34

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.000	01:36
Breath Volume	1.851	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.000	01:39
Breath Volume	2.128	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: DEVITT

DATE/TIME OF TEST: 10/08/2022 05:17:11

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.097	05:12
Breath Volume	2.894	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.094	05:15
Breath Volume	3.394	05:15
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 10/08/2022 22:40:42

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.081	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.084	22:36
Breath Volume	3.351	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.083	22:39
Breath Volume	3.800	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MATUTE

DATE/TIME OF TEST: 10/09/2022 02:45:50

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.225	02:40
Breath Volume	1.339	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.230	02:44
Breath Volume	1.820	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,EDWIN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: VELAZQUEZ RUIZ

DATE/TIME OF TEST: 10/13/2022 05:40:57

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.081	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.161	05:36
Breath Volume	1.824	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:38
Subject Sample # 2	0.156	05:39
Breath Volume	1.378	05:39
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOBITKA,NICHOLAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: SANCHEZ GERONIMO

DATE/TIME OF TEST: 10/13/2022 22:25:44

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.081	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.198	22:20
Breath Volume	1.656	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.187	22:24
Breath Volume	2.148	22:24
Air Blank	0.000	22:24
Control Test	0.078	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 10/14/2022 01:54:52

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.224	01:50
Breath Volume	1.367	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.231	01:53
Breath Volume	1.593	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 10/14/2022 22:29:40

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.204	22:24
Breath Volume	2.464	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.201	22:28
Breath Volume	2.828	22:28
Air Blank	0.000	22:28
Control Test	0.078	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: KHYTROV

DATE/TIME OF TEST: 10/15/2022 00:50:13

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.216	00:45
Breath Volume	3.089	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.208	00:48
Breath Volume	2.457	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MONDELUS

DATE/TIME OF TEST: 10/15/2022 02:36:20

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.188	02:30
Breath Volume	1.519	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.172	02:34
Breath Volume	1.394	02:34
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FESTA, JESSICA
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: GIUS

DATE/TIME OF TEST: 10/16/2022 01:00:46

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.080	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:56	No Sample Provided
Breath Volume	0.000	00:56	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.077	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HENRIQUEZ HERNANDEZ

DATE/TIME OF TEST: 10/16/2022 03:05:34

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.225	03:00
Breath Volume	2.812	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.223	03:03
Breath Volume	2.593	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: VASQUEZ RAMOS

DATE/TIME OF TEST: 10/16/2022 05:08:18

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.285	05:03
Breath Volume	3.277	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:05
Subject Sample # 2	0.282	05:06
Breath Volume	3.066	05:06
Air Blank	0.000	05:07
Control Test	0.078	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WOOD, DERRICK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: PESANTEZ

DATE/TIME OF TEST: 10/16/2022 08:54:23

Diagnostics Check	OK	08:47
Air Blank	0.000	08:47
Control Test	0.081	08:48
Air Blank	0.000	08:48
Subject Sample # 1	0.113	08:49
Breath Volume	1.167	08:49
Air Blank	0.000	08:50
Air Blank	0.000	08:51
Subject Sample # 2	0.117	08:52
Breath Volume	1.132	08:52
Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:54
Diagnostics Check	OK	08:54

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: JUAN PEDRO

DATE/TIME OF TEST: 10/17/2022 00:44:09

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.186	00:39
Breath Volume	1.609	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.182	00:42
Breath Volume	2.441	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FONGER, WALTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DELA ROSA MICHEL

DATE/TIME OF TEST: 10/17/2022 01:45:10

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.120	01:40
Breath Volume	1.164	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.118	01:43
Breath Volume	1.421	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: RAMIREZ SANDOVAL

DATE/TIME OF TEST: 10/17/2022 03:17:34

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.149	03:12
Breath Volume	1.621	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.153	03:15
Breath Volume	3.882	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 10/20/2022 03:07:09

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.136	03:02
Breath Volume	1.769	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.140	03:05
Breath Volume	2.839	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: BERNATCHEZ

DATE/TIME OF TEST: 10/20/2022 23:15:28

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.167	23:08
Breath Volume	1.367	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.162	23:13
Breath Volume	1.347	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 10/21/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 02/05/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 01:04
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GODBOLT DATE/TIME OF TEST: 10/03/2022 01:31:15

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.080	01:28	
Air Blank	0.000	01:28	
Subject Sample # 1	0.176	01:29	
Breath Volume	1.132	01:29	
Air Blank		01:30	Purge Fail
Air Blank		01:31	Purge Fail

Data Download Date/Time: 10/31/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 02/05/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 01:04
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GODBOLT DATE/TIME OF TEST: 10/03/2022 01:40:05

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:34	
Subject Sample # 1	0.180	01:35	
Breath Volume	1.351	01:35	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample # 2	0.168	01:38	
Breath Volume	1.285	01:38	
Air Blank	0.000	01:39	
Control Test	0.076	01:39	Interferent Detect
Air Blank	0.000	01:40	

Data Download Date/Time: 10/31/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 02/05/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 01:04
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GODBOLT DATE/TIME OF TEST: 10/03/2022 01:52:59

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.164	01:48
Breath Volume	1.394	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.172	01:51
Breath Volume	1.351	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 10/31/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 21:16
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Female
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: RAPHAEL DATE/TIME OF TEST: 10/03/2022 21:45:22

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.082	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.007	21:41
Breath Volume	1.339	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.000	21:43
Breath Volume	2.238	21:43
Air Blank	0.000	21:44
Control Test	0.084	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 10/31/2022 14:59

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 03/05/2023
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 19:28
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: RIVERA ACEVEDO	DATE/TIME OF TEST: 10/05/2022 19:54:37
-------------------------------------	--

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.080	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.039	19:50
Breath Volume	2.906	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.040	19:53
Breath Volume	2.441	19:53
Air Blank	0.000	19:53
Control Test	0.081	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 10/31/2022 14:59

Subject Test Electronic Data

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 02/05/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 03:50
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JIMENEZ DATE/TIME OF TEST: 10/07/2022 04:22:45

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:11	
Air Blank	0.000	04:11	
Subject Sample	# 1 0.184	04:14	Volume Not Met
Breath Volume	0.015	04:14	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample	# 2 0.202	04:19	
Breath Volume	1.488	04:19	
Air Blank	0.000	04:20	
Air Blank		04:22	Ambient Fail
Air Blank	0.000	04:22	

Data Download Date/Time: 10/31/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 02/05/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,AUSTIN	ARREST TIME: 03:50
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: JIMENEZ DATE/TIME OF TEST: 10/07/2022 04:30:53

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:25	
Air Blank	0.000	04:25	
Subject Sample # 1	0.195	04:27	
Breath Volume	1.531	04:27	
Air Blank	0.000	04:28	
Air Blank		04:30	Ambient Fail
Air Blank	0.000	04:30	

Data Download Date/Time: 10/31/2022 14:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24229080A1
BREATH TEST OPERATOR: WELLS,AUSTIN,D	DRY GAS EXPIRY DATE: 02/05/2024
BREATH TEST OPERATOR AGENCY: FHP Troop C	VIOLATION CODE: DUI
ARRESTING OFFICER: EARLYWINE,GARRETT	ARREST TIME: 02:52
ARRESTING OFFICER AGENCY: FHP Troop C	GENDER: Male
LAST AGENCY INSPECTION: 09/25/2022	OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CARR	DATE/TIME OF TEST: 10/08/2022 03:38:15
---------------------------	--

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.140	03:33
Breath Volume	2.097	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.145	03:36
Breath Volume	2.511	03:36
Air Blank	0.000	03:37
Control Test	0.082	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 10/31/2022 14:59