

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: MATIAS-TOMAS

DATE/TIME OF TEST: 11/19/2022 21:07:50

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.080	21:00
Air Blank	0.000	21:00
Subject Sample # 1	0.172	21:03
Breath Volume	1.722	21:03
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.163	21:06
Breath Volume	1.636	21:06
Air Blank	0.000	21:06
Control Test	0.080	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 12/05/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 11/26/2022 01:15:44

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.204	01:11
Breath Volume	2.148	01:11
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.209	01:14
Breath Volume	2.515	01:14
Air Blank	0.000	01:14
Control Test	0.082	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 12/05/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 10/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: VICENTE HERNANDEZ

DATE/TIME OF TEST: 11/26/2022 22:45:16

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.272	22:40
Breath Volume	1.902	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:42
Subject Sample # 2	0.262	22:43
Breath Volume	2.578	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 12/01/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STRANGE,BOBBY
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 11/12/2022 23:30:13

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.079	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1 0.262	23:25	
Breath Volume	4.085	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2 0.247	23:28	
Breath Volume	3.113	23:28	
Air Blank	0.000	23:29	
Control Test	0.069	23:29	Control Outside Tolerance
Air Blank	0.000	23:30	

Data Download Date/Time: 11/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STRANGE,BOBBY
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 11/12/2022 23:39:00

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.077	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.248	23:34
Breath Volume	2.992	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.244	23:37
Breath Volume	3.449	23:37
Air Blank	0.000	23:38
Control Test	0.078	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:39

Data Download Date/Time: 11/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: YANICK,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 11/13/2022 00:11:46

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.078	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.226	00:07
Breath Volume	1.343	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.239	00:10
Breath Volume	4.281	00:10
Air Blank	0.000	00:10
Control Test	0.078	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 11/28/2022 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MIRANDA,J
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: GOMEZ CARDONA

DATE/TIME OF TEST: 11/20/2022 01:23:36

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.138	01:19
Breath Volume	3.328	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.137	01:21
Breath Volume	3.039	01:21
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 11/28/2022 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: RADCLIFF,AARON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: RAUSTAD

DATE/TIME OF TEST: 11/21/2022 20:43:58

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:41	
Control Test	0.077	20:41	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:42	Subject Test Refused
Breath Volume	0.000	20:42	
Air Blank	0.000	20:43	
Control Test	0.079	20:43	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 11/28/2022 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JIMENEZ, WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JIMENEZ, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: GRICE

DATE/TIME OF TEST: 11/01/2022 19:33:38

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.079	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.203	19:28
Breath Volume	3.066	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.199	19:31
Breath Volume	2.960	19:31
Air Blank	0.000	19:32
Control Test	0.076	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 11/14/2022 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JABLONSKI, MARK, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JABLONSKI, MARK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: HERNANDEZ CARRANZA

DATE/TIME OF TEST: 11/08/2022 02:45:07

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 11/14/2022 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FLYNN,H
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/01/2022 22:15:14

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:03	
Subject Sample	# 1 0.146	22:06	Volume Not Met
Breath Volume	0.832	22:06	
Air Blank	0.000	22:07	
Air Blank	0.000	22:09	
Subject Sample	# 2 0.166	22:10	
Breath Volume	2.074	22:10	
Air Blank	0.000	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 3 0.161	22:13	
Breath Volume	1.507	22:13	
Air Blank	0.000	22:14	
Control Test	0.074	22:14	Control Outside Tolerance
Air Blank	0.000	22:15	

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAERTZ,MARK,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 11/03/2022 19:54:39

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.000	19:50
Breath Volume	2.105	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.000	19:53
Breath Volume	1.964	19:53
Air Blank	0.000	19:53
Control Test	0.079	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,TARIQ
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: TRENCH RIVERA

DATE/TIME OF TEST: 11/06/2022 23:04:14

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1 0.127	22:55	Volume Not Met
Breath Volume	0.664	22:55	
Air Blank	0.000	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 2 0.143	22:58	
Breath Volume	1.714	22:58	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 3 0.143	23:02	
Breath Volume	1.691	23:02	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CROSBY,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/11/2022 20:10:02

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.080	20:02	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:05	No Sample Provided
Breath Volume	0.000	20:05	
Air Blank	0.000	20:06	
Air Blank	0.000	20:08	
Subject Sample	# 2	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: JURKIEWICZ

DATE/TIME OF TEST: 11/12/2022 21:11:16

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.080	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.134	21:05
Breath Volume	1.199	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.116	21:09
Breath Volume	1.414	21:09
Air Blank	0.000	21:10
Control Test	0.078	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARES,J
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: TRUE

DATE/TIME OF TEST: 11/13/2022 01:58:10

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILLIAMS,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 11/17/2022 19:24:08

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:16	
Control Test	0.080	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:20	No Sample Provided
Breath Volume	0.000	19:20	
Air Blank	0.000	19:20	
Air Blank	0.000	19:22	
Subject Sample	# 2	19:22	Subject Test Refused
Breath Volume	0.000	19:22	
Air Blank	0.000	19:23	
Control Test	0.080	19:23	
Air Blank	0.000	19:24	
Diagnostics Check	OK	19:24	

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANTO

DATE/TIME OF TEST: 11/22/2022 04:06:42

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.078	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.000	04:02
Breath Volume	2.414	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.000	04:05
Breath Volume	2.007	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 11/25/2022 01:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: ANG

DATE/TIME OF TEST: 11/22/2022 23:38:18

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.189	23:33
Breath Volume	1.539	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.193	23:36
Breath Volume	1.847	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 11/25/2022 01:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECK,S,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: KOEHLER,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 11/23/2022 12:54:38

Diagnostics Check	OK	12:48
Air Blank	0.000	12:49
Control Test	0.079	12:49
Air Blank	0.000	12:49
Subject Sample # 1	0.000	12:50
Breath Volume	1.937	12:50
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample # 2	0.000	12:53
Breath Volume	1.773	12:53
Air Blank	0.000	12:53
Control Test	0.079	12:54
Air Blank	0.000	12:54
Diagnostics Check	OK	12:54

Data Download Date/Time: 11/25/2022 01:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PARAMORE,TERRENCE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: BLOOMQUIST

DATE/TIME OF TEST: 11/01/2022 18:41:55

Diagnostics Check	OK	18:35
Air Blank	0.000	18:35
Control Test	0.079	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.280	18:36
Breath Volume	2.269	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.267	18:40
Breath Volume	3.375	18:40
Air Blank	0.000	18:40
Control Test	0.079	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 11/18/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 11/02/2022 21:02:44

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:55	
Control Test	0.079	20:55	
Air Blank	0.000	20:55	
Subject Sample	# 1 0.127	20:56	
Breath Volume	1.472	20:56	
Air Blank	0.000	20:57	
Air Blank	0.000	20:59	
Subject Sample	# 2	21:01	Subject Test Refused
Breath Volume	0.339	21:01	
Air Blank	0.000	21:01	
Control Test	0.078	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 11/18/2022 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORGAN,DAVID
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/03/2022 22:39:47

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.079	22:36	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:38	Interferent Detect
Breath Volume	2.273	22:38	
Air Blank	0.000	22:38	
Control Test	0.079	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 11/18/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORGAN, DAVID
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/03/2022 23:06:54

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:05	Interferent Detect
Breath Volume	2.125	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:06	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 11/18/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/04/2022 20:29:50

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.000	20:25
Breath Volume	1.117	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.000	20:28
Breath Volume	1.726	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 11/18/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 11/05/2022 00:55:38

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.123	00:50
Breath Volume	1.925	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.131	00:54
Breath Volume	1.343	00:54
Air Blank	0.000	00:54
Control Test	0.076	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 11/18/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CALDERON, ULISES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 11/05/2022 01:54:29

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1 0.000	01:47	Volume Not Met
Breath Volume	0.171	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2 0.000	01:53	Volume Not Met
Breath Volume	0.050	01:53	
Air Blank	0.000	01:53	
Control Test	0.078	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 11/18/2022 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BRAVO MORALES

DATE/TIME OF TEST: 11/05/2022 02:45:42

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.211	02:41
Breath Volume	3.214	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.199	02:44
Breath Volume	3.062	02:44
Air Blank	0.000	02:44
Control Test	0.077	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCLUSKEY, RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: DAZA

DATE/TIME OF TEST: 11/05/2022 05:59:14

Diagnostics Check	OK	05:52
Air Blank	0.000	05:53
Control Test	0.079	05:53
Air Blank	0.000	05:54
Subject Sample # 1	0.132	05:54
Breath Volume	1.964	05:54
Air Blank	0.000	05:55
Air Blank	0.000	05:57
Subject Sample # 2	0.127	05:57
Breath Volume	1.441	05:57
Air Blank	0.000	05:58
Control Test	0.078	05:58
Air Blank	0.000	05:59
Diagnostics Check	OK	05:59

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LERNER, EVERETT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: CHAVARRIA GUEVARA

DATE/TIME OF TEST: 11/05/2022 19:22:08

Diagnostics Check	OK	19:15
Air Blank	0.000	19:15
Control Test	0.078	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.183	19:17
Breath Volume	1.632	19:17
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.181	19:20
Breath Volume	1.937	19:20
Air Blank	0.000	19:21
Control Test	0.078	19:21
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM, CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: AKOUNDI

DATE/TIME OF TEST: 11/06/2022 01:18:25

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.156	01:14
Breath Volume	2.472	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.151	01:16
Breath Volume	2.246	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES, FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 11/06/2022 01:03:50

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Subject Sample # 2	0.193	01:58
Breath Volume	1.429	01:58
Air Blank	0.000	01:58
Air Blank	0.000	01:00
Subject Sample # 1	0.193	01:02
Breath Volume	2.148	01:02
Air Blank	0.000	01:02
Control Test	0.077	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS, THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: JIMENEZ SILVANO

DATE/TIME OF TEST: 11/06/2022 01:46:49

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.077	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.163	01:42
Breath Volume	2.160	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.164	01:45
Breath Volume	2.261	01:45
Air Blank	0.000	01:45
Control Test	0.077	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FULCIDOR

DATE/TIME OF TEST: 11/06/2022 02:26:29

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.130	02:22
Breath Volume	2.394	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.130	02:24
Breath Volume	1.292	02:24
Air Blank	0.000	02:25
Control Test	0.077	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: EAMES

DATE/TIME OF TEST: 11/06/2022 21:37:26

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:29	
Control Test	0.079	21:29	
Air Blank	0.000	21:30	
Subject Sample	# 1 0.160	21:33	Volume Not Met
Breath Volume	0.660	21:33	
Air Blank	0.000	21:34	
Air Blank	0.000	21:35	
Subject Sample	# 2	21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	
Control Test	0.078	21:36	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JENNINGS,DAVON
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: GALLIANI

DATE/TIME OF TEST: 11/06/2022 22:37:16

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.079	22:24	
Air Blank	0.000	22:25	
Subject Sample	# 1 0.077	22:28	Volume Not Met
Breath Volume	0.761	22:28	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 2 0.074	22:32	
Breath Volume	1.363	22:32	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 3 0.069	22:35	
Breath Volume	1.566	22:35	
Air Blank	0.000	22:36	
Control Test	0.077	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SIERRA BRIONES

DATE/TIME OF TEST: 11/07/2022 00:37:49

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.197	00:33
Breath Volume	2.074	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.201	00:36
Breath Volume	1.882	00:36
Air Blank	0.000	00:36
Control Test	0.078	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: MACGARVEY

DATE/TIME OF TEST: 11/07/2022 21:28:02

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.187	21:22
Breath Volume	1.488	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:24
Subject Sample # 2	0.187	21:26
Breath Volume	1.460	21:26
Air Blank	0.000	21:27
Control Test	0.078	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:28

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: DONHAM

DATE/TIME OF TEST: 11/08/2022 00:34:31

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.268	00:30
Breath Volume	1.312	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.271	00:32
Breath Volume	3.082	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: SCHVARTZMAN

DATE/TIME OF TEST: 11/08/2022 04:52:55

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.168	04:48
Breath Volume	1.218	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.167	04:51
Breath Volume	1.199	04:51
Air Blank	0.000	04:52
Control Test	0.078	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: JOSMA

DATE/TIME OF TEST: 11/08/2022 19:41:34

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.079	19:28	
Air Blank	0.000	19:28	
Subject Sample	# 1 0.238	19:31	
Breath Volume	1.507	19:31	
Air Blank	0.000	19:32	
Air Blank	0.000	19:34	
Subject Sample	# 2 0.240	19:37	Volume Not Met
Breath Volume	0.121	19:37	
Air Blank	0.000	19:38	
Air Blank	0.000	19:39	
Subject Sample	# 3	19:40	Subject Test Refused
Breath Volume	0.000	19:40	
Air Blank	0.000	19:40	
Control Test	0.076	19:40	
Air Blank	0.000	19:41	
Diagnostics Check	OK	19:41	

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: SHOULDERS

DATE/TIME OF TEST: 11/08/2022 22:16:44

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.184	22:12
Breath Volume	1.554	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:14
Subject Sample # 2	0.178	22:15
Breath Volume	1.207	22:15
Air Blank	0.000	22:15
Control Test	0.077	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SIMMONDS

DATE/TIME OF TEST: 11/11/2022 03:02:17

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.129	02:57
Breath Volume	2.621	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.126	03:00
Breath Volume	2.664	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CHOVON GUARCHOJ

DATE/TIME OF TEST: 11/12/2022 04:20:11

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.267	04:14
Breath Volume	2.289	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.256	04:18
Breath Volume	2.527	04:18
Air Blank	0.000	04:19
Control Test	0.076	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: RUA SANCHEZ

DATE/TIME OF TEST: 11/12/2022 23:30:12

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.179	23:25
Breath Volume	2.898	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.178	23:28
Breath Volume	3.625	23:28
Air Blank	0.000	23:29
Control Test	0.078	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ADAY

DATE/TIME OF TEST: 11/13/2022 00:08:40

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.078	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.154	00:03
Breath Volume	1.335	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.150	00:07
Breath Volume	1.261	00:07
Air Blank	0.000	00:07
Control Test	0.077	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: PEREZ PASCUAL

DATE/TIME OF TEST: 11/13/2022 19:17:21

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.327	19:12
Breath Volume	2.964	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.323	19:15
Breath Volume	2.636	19:15
Air Blank	0.000	19:16
Control Test	0.078	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: PEREDA GONZALEZ

DATE/TIME OF TEST: 11/13/2022 20:58:50

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.219	20:54
Breath Volume	2.464	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.218	20:57
Breath Volume	2.781	20:57
Air Blank	0.000	20:57
Control Test	0.077	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANEGO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 11/13/2022 22:24:37

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.087	22:18
Breath Volume	1.218	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.088	22:23
Breath Volume	1.296	22:23
Air Blank	0.000	22:23
Control Test	0.078	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: CASTANON GONZALEZ

DATE/TIME OF TEST: 11/13/2022 23:31:09

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.078	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.203	23:26
Breath Volume	2.687	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.192	23:29
Breath Volume	2.140	23:29
Air Blank	0.000	23:30
Control Test	0.077	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HIGIER

DATE/TIME OF TEST: 11/15/2022 00:48:14

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:43	
Control Test	0.079	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.074	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 11/17/2022 01:03:15

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.256	00:58
Breath Volume	2.164	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.252	01:01
Breath Volume	2.285	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BATEN RAMOS

DATE/TIME OF TEST: 11/17/2022 03:14:45

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.208	03:10
Breath Volume	1.824	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.223	03:13
Breath Volume	1.593	03:13
Air Blank	0.000	03:13
Control Test	0.077	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILORD, JHAMIL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 11/17/2022 22:17:29

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.128	22:13
Breath Volume	3.140	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.124	22:15
Breath Volume	3.097	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: PEREZ ARANDA

DATE/TIME OF TEST: 11/18/2022 03:17:06

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.123	03:12
Breath Volume	1.644	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.122	03:15
Breath Volume	1.992	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 11/18/2022 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES, ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: GARCIA CARBALLO

DATE/TIME OF TEST: 11/18/2022 18:00:33

Diagnostics Check	OK	17:54
Air Blank	0.000	17:54
Control Test	0.080	17:54
Air Blank	0.000	17:55
Subject Sample # 1	0.165	17:55
Breath Volume	1.699	17:55
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample # 2	0.168	17:58
Breath Volume	1.933	17:58
Air Blank	0.000	17:59
Control Test	0.080	17:59
Air Blank	0.000	18:00
Diagnostics Check	OK	18:00

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 11/18/2022 20:45:30

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.080	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.204	20:40
Breath Volume	1.390	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.196	20:43
Breath Volume	1.441	20:43
Air Blank	0.000	20:44
Control Test	0.079	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO, ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CALDERON CASTILLO

DATE/TIME OF TEST: 11/19/2022 02:27:18

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.209	02:22
Breath Volume	2.332	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.197	02:25
Breath Volume	2.511	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD, EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LOPEZ BERNARDO

DATE/TIME OF TEST: 11/19/2022 21:18:08

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.081	21:12
Air Blank	0.000	21:12
Subject Sample # 1	0.181	21:13
Breath Volume	1.625	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:15
Subject Sample # 2	0.187	21:16
Breath Volume	2.507	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/20/2022 00:27:48

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.136	00:23
Breath Volume	2.535	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.127	00:26
Breath Volume	1.539	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO, LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: BARONE

DATE/TIME OF TEST: 11/20/2022 03:17:46

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1 0.000	03:10	Volume Not Met
Breath Volume	0.226	03:10	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2 0.000	03:13	
Breath Volume	1.457	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 3 0.000	03:16	
Breath Volume	1.355	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO, ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: GONZALEZ BRACHO

DATE/TIME OF TEST: 11/20/2022 06:23:37

Diagnostics Check	OK	06:15
Air Blank	0.000	06:16
Control Test	0.080	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.172	06:18
Breath Volume	3.011	06:18
Air Blank	0.000	06:19
Air Blank	0.000	06:20
Subject Sample # 2	0.168	06:22
Breath Volume	2.175	06:22
Air Blank	0.000	06:22
Control Test	0.078	06:23
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: QUIROZ RIVAS

DATE/TIME OF TEST: 11/22/2022 03:53:59

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.000	03:49
Breath Volume	1.835	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.000	03:52
Breath Volume	1.480	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JOHNTHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 11/23/2022 05:02:33

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.000	04:58
Breath Volume	2.093	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.000	05:01
Breath Volume	1.375	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEONARDI, KAILEY
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 11/24/2022 00:26:31

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.188	00:22
Breath Volume	2.597	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.194	00:24
Breath Volume	2.328	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: TADDY

DATE/TIME OF TEST: 11/24/2022 00:59:20

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.188	00:52
Breath Volume	1.632	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.213	00:54
Breath Volume	2.753	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 3	0.207	00:57
Breath Volume	2.238	00:57
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CHAVEZ SOTO

DATE/TIME OF TEST: 11/24/2022 02:56:16

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.120	02:51
Breath Volume	1.605	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.114	02:54
Breath Volume	1.625	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANTANA, DELVIS
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: AGUILAR JACINTO

DATE/TIME OF TEST: 11/24/2022 03:22:08

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.245	03:17
Breath Volume	1.222	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.243	03:20
Breath Volume	1.125	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: VASQUEZ LELASQUEZ

DATE/TIME OF TEST: 11/24/2022 04:15:59

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.189	04:11
Breath Volume	1.703	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.195	04:14
Breath Volume	1.363	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: HONORE

DATE/TIME OF TEST: 11/25/2022 00:37:57

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.247	00:32
Breath Volume	1.222	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.243	00:36
Breath Volume	1.972	00:36
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RICHELIEU

DATE/TIME OF TEST: 11/25/2022 02:11:12

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.167	02:05
Breath Volume	1.695	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.182	02:09
Breath Volume	3.402	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: MOMPOINT

DATE/TIME OF TEST: 11/25/2022 05:25:19

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.078	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.112	05:20
Breath Volume	1.453	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.122	05:23
Breath Volume	3.015	05:23
Air Blank	0.000	05:24
Control Test	0.078	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VIANUEVA, BRANDON
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: MATA

DATE/TIME OF TEST: 11/25/2022 07:49:55

Diagnostics Check	OK	07:42
Air Blank	0.000	07:42
Control Test	0.079	07:43
Air Blank	0.000	07:43
Subject Sample # 1	0.179	07:44
Breath Volume	2.527	07:44
Air Blank	0.000	07:45
Air Blank	0.000	07:46
Subject Sample # 2	0.193	07:48
Breath Volume	2.667	07:48
Air Blank	0.000	07:49
Control Test	0.078	07:49
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILORD, JHAMIL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: CRAVEN

DATE/TIME OF TEST: 11/25/2022 20:14:30

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.189	20:09
Breath Volume	2.785	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.197	20:12
Breath Volume	3.738	20:12
Air Blank	0.000	20:13
Control Test	0.079	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WIL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: ZACARIAS RAMIREZ

DATE/TIME OF TEST: 11/25/2022 20:48:34

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.205	20:43
Breath Volume	2.570	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.197	20:47
Breath Volume	1.773	20:47
Air Blank	0.000	20:47
Control Test	0.078	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: VALDES PUGA

DATE/TIME OF TEST: 11/26/2022 00:42:47

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.143	00:38
Breath Volume	2.386	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.145	00:41
Breath Volume	2.414	00:41
Air Blank	0.000	00:41
Control Test	0.078	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: LOPEZ ADRIANO

DATE/TIME OF TEST: 11/26/2022 21:50:59

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.079	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.174	21:46
Breath Volume	2.550	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:48
Subject Sample # 2	0.172	21:49
Breath Volume	3.039	21:49
Air Blank	0.000	21:50
Control Test	0.078	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRUNTZ, NICHOLAS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 11/27/2022 03:36:55

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.217	03:32
Breath Volume	2.265	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.204	03:35
Breath Volume	3.500	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBLERO,ARTURO
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: PARRA TREVIZO

DATE/TIME OF TEST: 11/27/2022 23:43:18

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1 0.201	23:36	Volume Not Met
Breath Volume	0.066	23:36	
Air Blank	0.000	23:36	
Air Blank	0.000	23:38	
Subject Sample	# 2 0.175	23:41	Volume Not Met
Breath Volume	0.070	23:41	
Air Blank	0.000	23:42	
Control Test	0.076	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBLERO,ARTURO
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: PARRA TREVIZO

DATE/TIME OF TEST: 11/27/2022 23:50:12

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.077	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1	23:48	Subject Test Refused
Breath Volume	0.000	23:48	
Air Blank	0.000	23:49	
Control Test	0.078	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM, CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: BRINZO

DATE/TIME OF TEST: 11/28/2022 05:03:10

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.208	04:58
Breath Volume	2.019	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.215	05:01
Breath Volume	2.593	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 11/28/2022 05:17:27

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.078	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.141	05:13
Breath Volume	1.718	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.145	05:15
Breath Volume	2.062	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RINI, LAWRENCE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: FRIEDMAN

DATE/TIME OF TEST: 11/29/2022 11:16:22

Diagnostics Check	OK	11:09
Air Blank	0.000	11:10
Control Test	0.080	11:10
Air Blank	0.000	11:11
Subject Sample # 1	0.000	11:12
Breath Volume	1.292	11:12
Air Blank	0.000	11:12
Air Blank	0.000	11:14
Subject Sample # 2	0.000	11:15
Breath Volume	1.414	11:15
Air Blank	0.000	11:15
Control Test	0.080	11:15
Air Blank	0.000	11:16
Diagnostics Check	OK	11:16

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAGET, HORLKINS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: ALLI

DATE/TIME OF TEST: 11/30/2022 11:27:45

Diagnostics Check	OK	11:21
Air Blank	0.000	11:22
Control Test	0.079	11:22
Air Blank	0.000	11:22
Subject Sample # 1	0.000	11:23
Breath Volume	2.726	11:23
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.000	11:26
Breath Volume	2.824	11:26
Air Blank	0.000	11:26
Control Test	0.079	11:27
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MORALES DIAZ

DATE/TIME OF TEST: 11/19/2022 01:47:31

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.201	01:42
Breath Volume	2.132	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.200	01:45
Breath Volume	2.035	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: CEDENO GOMEZ

DATE/TIME OF TEST: 11/19/2022 22:32:22

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.152	22:27
Breath Volume	2.046	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.150	22:30
Breath Volume	1.863	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HENRIQUEZZUNIGA, MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PEREZ CAIXBA

DATE/TIME OF TEST: 11/20/2022 03:40:13

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.193	03:35
Breath Volume	1.660	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.187	03:38
Breath Volume	1.835	03:38
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/20/2022 05:41:55

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.079	05:36
Air Blank	0.000	05:36
Subject Sample # 1	0.112	05:37
Breath Volume	1.433	05:37
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.120	05:40
Breath Volume	1.921	05:40
Air Blank	0.000	05:41
Control Test	0.078	05:41
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES, A
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 11/20/2022 21:11:26

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.148	21:06
Breath Volume	2.441	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.146	21:09
Breath Volume	3.503	21:09
Air Blank	0.000	21:10
Control Test	0.078	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLARKE,ROMAINE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/20/2022 23:27:57

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.079	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.243	23:15	No .020 agreement
Breath Volume	1.207	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 2 0.233	23:21	Volume Not Met
Breath Volume	0.031	23:21	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample	# 3 0.215	23:26	No .020 agreement
Breath Volume	1.722	23:26	
Air Blank	0.000	23:26	
Control Test	0.076	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLARKE,ROMAINE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/21/2022 00:35:29

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.078	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.191	00:30
Breath Volume	2.246	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.197	00:33
Breath Volume	1.886	00:33
Air Blank	0.000	00:34
Control Test	0.077	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: SHAMBO

DATE/TIME OF TEST: 11/21/2022 02:50:09

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.245	02:45
Breath Volume	2.382	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.248	02:48
Breath Volume	2.496	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEVIS, MILJEN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: MCANLY

DATE/TIME OF TEST: 11/22/2022 17:51:39

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.079	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.208	17:46
Breath Volume	1.343	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.220	17:50
Breath Volume	2.285	17:50
Air Blank	0.000	17:50
Control Test	0.079	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRIS PHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CASTILLO GUZMAN

DATE/TIME OF TEST: 11/24/2022 23:22:42

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.079	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.126	23:18
Breath Volume	2.753	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.127	23:21
Breath Volume	2.625	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: PEREZVASQUEZ

DATE/TIME OF TEST: 11/25/2022 01:52:54

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.196	01:48
Breath Volume	1.589	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.196	01:51
Breath Volume	1.757	01:51
Air Blank	0.000	01:51
Control Test	0.079	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: RUIZ ZELAIA

DATE/TIME OF TEST: 11/25/2022 06:01:51

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.079	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.226	05:57
Breath Volume	1.269	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.231	06:00
Breath Volume	1.527	06:00
Air Blank	0.000	06:00
Control Test	0.079	06:01
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DIGHTON

DATE/TIME OF TEST: 11/26/2022 01:42:58

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.197	01:38
Breath Volume	1.968	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.198	01:41
Breath Volume	2.261	01:41
Air Blank	0.000	01:42
Control Test	0.078	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: PEREZ MEDINA

DATE/TIME OF TEST: 11/26/2022 22:31:37

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:23	
Control Test	0.079	22:23	
Air Blank	0.000	22:24	
Subject Sample	# 1 0.148	22:27	Volume Not Met
Breath Volume	1.089	22:27	
Air Blank	0.000	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 2	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.076	22:31	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: GUZMAN COBO

DATE/TIME OF TEST: 11/27/2022 02:12:10

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.176	02:07
Breath Volume	1.445	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.173	02:10
Breath Volume	1.609	02:10
Air Blank	0.000	02:11
Control Test	0.077	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANEGO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 11/30/2022 19:56:10

Diagnostics Check	OK	19:52	
Air Blank	0.000	19:52	
Control Test	0.080	19:52	
Air Blank	0.000	19:53	
Subject Sample	# 1	19:54	Subject Test Refused
Breath Volume	0.000	19:54	
Air Blank	0.000	19:55	
Control Test	0.079	19:55	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MONTAGUE

DATE/TIME OF TEST: 11/01/2022 03:34:46

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.098	03:30
Breath Volume	3.093	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.094	03:33
Breath Volume	2.433	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOBITKA, NICHOLAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 11/05/2022 00:21:55

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.183	00:16
Breath Volume	1.421	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.187	00:20
Breath Volume	1.683	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MAZAREIGOS ORTIZ

DATE/TIME OF TEST: 11/05/2022 02:29:47

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.183	02:24
Breath Volume	2.628	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.200	02:28
Breath Volume	2.378	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/06/2022 01:24:54

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.183	01:19
Breath Volume	1.531	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.189	01:23
Breath Volume	1.871	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ETORRE, COLTON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RODRIQUES

DATE/TIME OF TEST: 11/06/2022 01:27:41

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.185	01:22
Breath Volume	1.433	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.178	01:26
Breath Volume	2.105	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: VILLARREAL

DATE/TIME OF TEST: 11/06/2022 01:53:50

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.102	01:49
Breath Volume	1.433	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.102	01:52
Breath Volume	1.378	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HENRIQUEZ ZUNIGA, MARCIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 11/06/2022 03:52:26

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.201	03:44
Breath Volume	3.460	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.178	03:47
Breath Volume	1.484	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 3	0.201	03:50
Breath Volume	3.613	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: THRALL

DATE/TIME OF TEST: 11/06/2022 04:56:57

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1 0.188	04:51	Volume Not Met
Breath Volume	0.039	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2	04:55	Subject Test Refused
Breath Volume	0.207	04:55	
Air Blank	0.000	04:56	
Control Test	0.078	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES, FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 11/07/2022 05:33:18

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.176	05:28
Breath Volume	1.230	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:30
Subject Sample # 2	0.185	05:31
Breath Volume	2.933	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PAQUETTE, JASON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:17

SUBJECT'S LAST NAME: SILVIEUS

DATE/TIME OF TEST: 11/07/2022 13:53:55

Diagnostics Check	OK	13:47
Air Blank	0.000	13:48
Control Test	0.081	13:48
Air Blank	0.000	13:49
Subject Sample # 1	0.067	13:49
Breath Volume	1.652	13:49
Air Blank	0.000	13:50
Air Blank	0.000	13:51
Subject Sample # 2	0.070	13:52
Breath Volume	1.777	13:52
Air Blank	0.000	13:53
Control Test	0.081	13:53
Air Blank	0.000	13:53
Diagnostics Check	OK	13:53

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CUSSON,MATTHEW
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: DE LEON HERNANDEZ

DATE/TIME OF TEST: 11/09/2022 02:40:50

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.212	02:36
Breath Volume	3.250	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.223	02:39
Breath Volume	3.457	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 11/11/2022 03:16:13

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.196	03:11
Breath Volume	1.515	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:13
Subject Sample # 2	0.196	03:14
Breath Volume	1.488	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITLOCK

DATE/TIME OF TEST: 11/11/2022 22:04:40

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.081	21:59
Air Blank	0.000	21:59
Subject Sample # 1	0.000	22:00
Breath Volume	1.707	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.000	22:03
Breath Volume	3.511	22:03
Air Blank	0.000	22:03
Control Test	0.080	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MACARIO PUAC

DATE/TIME OF TEST: 11/12/2022 03:58:27

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.128	03:54
Breath Volume	1.949	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.128	03:56
Breath Volume	2.101	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILORD, JHAMIL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: GUNDERSON

DATE/TIME OF TEST: 11/12/2022 20:34:01

Diagnostics Check	OK	20:24
Air Blank	0.000	20:24
Control Test	0.080	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.208	20:27
Breath Volume	1.757	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.206	20:32
Breath Volume	1.511	20:32
Air Blank	0.000	20:33
Control Test	0.079	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:34

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES, ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: KOSTA

DATE/TIME OF TEST: 11/12/2022 21:32:33

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.000	21:28
Breath Volume	2.609	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.000	21:31
Breath Volume	1.804	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD, EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: MASSIMO

DATE/TIME OF TEST: 11/12/2022 23:46:33

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.091	23:41
Breath Volume	1.304	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.092	23:44
Breath Volume	2.101	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SIMMONS, TIMOTHY
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 11/13/2022 06:53:02

Diagnostics Check	OK	06:45
Air Blank	0.000	06:45
Control Test	0.080	06:46
Air Blank	0.000	06:46
Subject Sample # 1	0.138	06:48
Breath Volume	3.097	06:48
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.135	06:51
Breath Volume	2.933	06:51
Air Blank	0.000	06:52
Control Test	0.079	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:53

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER, CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 11/14/2022 00:29:29

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.000	00:25
Breath Volume	1.906	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.000	00:28
Breath Volume	2.347	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FESTA, JESS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PASCAL

DATE/TIME OF TEST: 11/18/2022 02:36:24

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.081	02:35	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 11/18/2022 16:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: OSHAUGHNESSY

DATE/TIME OF TEST: 11/03/2022 02:26:56

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.101	02:21
Breath Volume	1.753	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample # 2	0.101	02:25
Breath Volume	2.480	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/16/2022 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: DACOSTA

DATE/TIME OF TEST: 11/08/2022 20:40:55

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.078	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.211	20:32
Breath Volume	1.640	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:34
Subject Sample # 2	0.176	20:35
Breath Volume	1.898	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 3	0.215	20:39
Breath Volume	1.601	20:39
Air Blank	0.000	20:39
Control Test	0.078	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 11/16/2022 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BING

DATE/TIME OF TEST: 11/09/2022 01:51:18

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.115	01:46
Breath Volume	3.457	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.115	01:49
Breath Volume	2.550	01:49
Air Blank	0.000	01:50
Control Test	0.077	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 11/16/2022 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/13/2022 23:02:37

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.000	22:58
Breath Volume	2.593	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:01
Breath Volume	1.437	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 11/16/2022 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: LITTLEFIELD

DATE/TIME OF TEST: 11/17/2022 02:49:02

Diagnostics Check	OK	02:48	
Air Blank		02:48	RFI Detect
Air Blank		02:49	RFI Detect

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: TTLEFIELD

DATE/TIME OF TEST: 11/17/2022 02:58:07

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.157	02:53
Breath Volume	2.019	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.160	02:56
Breath Volume	2.324	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RAYMAKER

DATE/TIME OF TEST: 11/17/2022 23:57:58

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.184	23:53
Breath Volume	3.035	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.191	23:56
Breath Volume	2.796	23:56
Air Blank	0.000	23:57
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RUEN

DATE/TIME OF TEST: 11/27/2022 02:32:02

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.078	02:31	
Air Blank		02:31	RFI Detect
Air Blank	0.000	02:32	

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RUEN

DATE/TIME OF TEST: 11/27/2022 02:40:18

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.077	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.097	02:35
Breath Volume	3.875	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.092	02:38
Breath Volume	1.203	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FRENCH

DATE/TIME OF TEST: 11/28/2022 02:32:27

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.181	02:27
Breath Volume	4.121	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.175	02:30
Breath Volume	3.433	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEIMIG,EMORIE,I
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: HARRIS,MICHAEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 10/29/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: SANCHEZ BALDOQUIN

DATE/TIME OF TEST: 11/04/2022 22:41:04

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.078	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.131	22:36
Breath Volume	1.355	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.135	22:39
Breath Volume	1.960	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:41

Data Download Date/Time: 11/09/2022 21:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,PABLO,
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SEVOR,ETHAN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 11/12/2022 01:22:13

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.248	01:17
Breath Volume	2.589	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.256	01:20
Breath Volume	3.437	01:20
Air Blank	0.000	01:21
Control Test	0.077	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 12/29/2022 00:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, MAURQUICE, D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: HENRY, MONICA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WILDMAN

DATE/TIME OF TEST: 11/12/2022 04:27:56

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.133	04:22
Breath Volume	2.921	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.131	04:26
Breath Volume	3.117	04:26
Air Blank	0.000	04:26
Control Test	0.077	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 12/29/2022 00:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: EARNEST

DATE/TIME OF TEST: 11/16/2022 20:55:26

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.179	20:50
Breath Volume	3.300	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.177	20:53
Breath Volume	2.933	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 12/06/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: PIKE

DATE/TIME OF TEST: 11/18/2022 01:13:17

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.159	01:08
Breath Volume	2.906	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.167	01:11
Breath Volume	3.101	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 12/06/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CLARK,COLTIN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/18/2022 03:28:40

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.082	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.029	03:24
Breath Volume	1.148	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.030	03:27
Breath Volume	2.531	03:27
Air Blank	0.000	03:27
Control Test	0.078	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/06/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORELLANA, ADALBERTO, L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ORELLANA, ADALBERTO
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: MENDEZ SALES

DATE/TIME OF TEST: 11/19/2022 13:27:54

Diagnostics Check	OK	13:20
Air Blank	0.000	13:20
Control Test	0.080	13:20
Air Blank	0.000	13:21
Subject Sample # 1	0.191	13:22
Breath Volume	1.183	13:22
Air Blank	0.000	13:22
Air Blank	0.000	13:24
Subject Sample # 2	0.179	13:26
Breath Volume	1.496	13:26
Air Blank	0.000	13:26
Control Test	0.079	13:27
Air Blank	0.000	13:27
Diagnostics Check	OK	13:27

Data Download Date/Time: 12/06/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/25/2022 03:17:38

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.251	03:13
Breath Volume	1.207	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.246	03:15
Breath Volume	2.046	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 12/06/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINFREE, SAIGE, D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MANNING, CARL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: AVIANEDA

DATE/TIME OF TEST: 11/29/2022 23:21:13

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.079	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1 0.215	23:15	
Breath Volume	5.074	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 2	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:20	
Control Test	0.077	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 12/06/2022 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: CRISP

DATE/TIME OF TEST: 11/02/2022 20:55:51

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.078	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.089	20:51
Breath Volume	2.257	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:53
Subject Sample # 2	0.090	20:54
Breath Volume	2.859	20:54
Air Blank	0.000	20:54
Control Test	0.079	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 11/09/2022 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: AGUILAR CORDERO

DATE/TIME OF TEST: 11/05/2022 23:41:30

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.078	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.113	23:37
Breath Volume	1.875	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.113	23:39
Breath Volume	1.601	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 11/09/2022 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CRISTEA,VICTOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 11/08/2022 03:15:41

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.213	03:07
Breath Volume	2.023	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.238	03:11
Breath Volume	2.496	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 3	0.220	03:14
Breath Volume	2.144	03:14
Air Blank	0.000	03:14
Control Test	0.077	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 11/09/2022 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: CUELLAR LEZCANO

DATE/TIME OF TEST: 11/17/2022 03:28:17

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.214	03:23
Breath Volume	1.242	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.212	03:26
Breath Volume	2.566	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MATHEWSON

DATE/TIME OF TEST: 11/17/2022 20:09:58

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:07	
Control Test	0.079	20:07	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MATHEWSON

DATE/TIME OF TEST: 11/17/2022 20:21:50

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.079	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.217	20:17
Breath Volume	1.628	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:19
Subject Sample # 2	0.215	20:20
Breath Volume	1.824	20:20
Air Blank	0.000	20:20
Control Test	0.078	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATHEWS,MARY ANN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KEEGAN

DATE/TIME OF TEST: 11/18/2022 03:36:54

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.167	03:32
Breath Volume	2.828	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.170	03:35
Breath Volume	2.429	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 11/18/2022 21:53:17

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.192	21:48
Breath Volume	3.144	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.199	21:51
Breath Volume	2.437	21:51
Air Blank	0.000	21:52
Control Test	0.078	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA,WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 11/19/2022 00:51:36

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.148	00:47
Breath Volume	2.210	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.144	00:50
Breath Volume	1.703	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/19/2022 20:50:22

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.078	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:43	No Sample Provided
Breath Volume	0.000	20:43	
Air Blank	0.000	20:43	
Air Blank	0.000	20:45	
Subject Sample	# 2	20:48	No Sample Provided
Breath Volume	0.000	20:48	
Air Blank	0.000	20:49	
Control Test	0.078	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODRICH,SARAH,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ASKINS,EDWARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 11/19/2022 21:39:31

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:37	
Control Test	0.079	21:37	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:38	Subject Test Refused
Breath Volume	0.000	21:38	
Air Blank	0.000	21:38	
Control Test	0.079	21:38	
Air Blank	0.000	21:39	
Diagnostics Check	OK	21:39	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DARDEN

DATE/TIME OF TEST: 11/21/2022 04:03:35

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.134	03:58
Breath Volume	1.605	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.139	04:01
Breath Volume	1.484	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: KILPATRICK

DATE/TIME OF TEST: 11/26/2022 03:07:58

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.078	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.079	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARR,JASON,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 11/26/2022 06:39:47

Diagnostics Check	OK	06:31	
Air Blank	0.000	06:31	
Control Test	0.079	06:31	
Air Blank	0.000	06:32	
Subject Sample	# 1	06:35	No Sample Provided
Breath Volume	0.000	06:35	
Air Blank	0.000	06:36	
Air Blank	0.000	06:38	
Subject Sample	# 2	06:38	Subject Test Refused
Breath Volume	0.000	06:38	
Air Blank	0.000	06:38	
Control Test	0.079	06:39	
Air Blank	0.000	06:39	
Diagnostics Check	OK	06:39	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 11/27/2022 02:12:03

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.179	02:07
Breath Volume	2.003	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.182	02:10
Breath Volume	1.742	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 11/29/2022 00:33:04

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.078	00:30	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 11/05/2022 03:15:45

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.081	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.081	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: BACA APON

DATE/TIME OF TEST: 11/06/2022 02:55:00

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.181	02:50
Breath Volume	1.335	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.198	02:53
Breath Volume	2.613	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GARCIA SALES

DATE/TIME OF TEST: 11/12/2022 03:54:25

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.260	03:49
Breath Volume	2.511	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 2	0.260	03:52
Breath Volume	3.789	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 11/21/2022 00:07:25

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.082	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:02
Breath Volume	2.585	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:06
Breath Volume	2.476	00:06
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRIGA,JUAN,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: BARRIGA,JUAN
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: LISBOA FELICIANO

DATE/TIME OF TEST: 11/23/2022 20:55:10

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.081	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.000	20:50
Breath Volume	4.414	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.000	20:53
Breath Volume	2.515	20:53
Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: GAUNT

DATE/TIME OF TEST: 11/24/2022 01:47:06

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HERNANDEZ GIL

DATE/TIME OF TEST: 11/25/2022 04:05:38

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.134	04:00
Breath Volume	1.558	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.126	04:03
Breath Volume	1.613	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 11/26/2022 05:52:10

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:39	
Control Test		0.081	05:39	
Air Blank		0.000	05:40	
Subject Sample	# 1	0.165	05:43	Volume Not Met
Breath Volume		0.457	05:43	
Air Blank		0.000	05:44	
Air Blank		0.000	05:46	
Subject Sample	# 2	0.213	05:47	
Breath Volume		1.832	05:47	
Air Blank		0.000	05:48	
Air Blank		0.000	05:49	
Subject Sample	# 3	0.221	05:50	
Breath Volume		1.222	05:50	
Air Blank		0.000	05:51	
Control Test		0.078	05:51	
Air Blank		0.000	05:52	
Diagnostics Check		OK	05:52	

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ZAYDON

DATE/TIME OF TEST: 11/27/2022 02:00:21

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.081	01:46	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:49	No Sample Provided
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample	# 2	01:54	
Breath Volume	1.875	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 3	01:58	
Breath Volume	2.007	01:58	
Air Blank	0.000	01:59	
Control Test	0.076	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOREJON,BRYAN,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: MOREJON,BRYAN
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:02

SUBJECT'S LAST NAME: ZELEDON

DATE/TIME OF TEST: 11/28/2022 10:33:36

Diagnostics Check	OK	10:27
Air Blank	0.000	10:28
Control Test	0.081	10:28
Air Blank	0.000	10:29
Subject Sample # 1	0.000	10:29
Breath Volume	1.117	10:29
Air Blank	0.000	10:29
Air Blank	0.000	10:31
Subject Sample # 2	0.000	10:32
Breath Volume	1.375	10:32
Air Blank	0.000	10:32
Control Test	0.079	10:32
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: SOSA GARCIA

DATE/TIME OF TEST: 11/07/2022 03:02:42

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.077	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1 0.226	02:57	
Breath Volume	3.121	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2 0.235	03:00	
Breath Volume	3.003	03:00	
Air Blank	0.000	03:01	
Control Test	0.077	03:01	
Air Blank		03:02	RFI Detect
Air Blank	0.000	03:02	

Data Download Date/Time: 12/14/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1273831
DRY GAS EXPIRY DATE: 01/31/2023
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MCKENNA

DATE/TIME OF TEST: 11/22/2022 04:33:57

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.076	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Control Test	0.077	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 12/14/2022 14:28