

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEMAN,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LARICCI,STEPHANIE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CASTILO

DATE/TIME OF TEST: 12/17/2022 03:45:17

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.078	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.079	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 12/19/2022 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,ANDREW,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: FRIED,DANIEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: HARRINANAN

DATE/TIME OF TEST: 12/20/2022 22:11:11

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.079	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.202	22:06
Breath Volume	1.863	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.194	22:09
Breath Volume	1.968	22:09
Air Blank	0.000	22:10
Control Test	0.077	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 01/11/2023 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALZATE,MELENNI,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ALZATE,MELENNI
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SEHARA

DATE/TIME OF TEST: 12/24/2022 03:16:34

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.137	03:11
Breath Volume	3.765	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.150	03:14
Breath Volume	3.937	03:14
Air Blank	0.000	03:15
Control Test	0.077	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 01/11/2023 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSICK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: BUSICK,ROBERT
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: BLANKS

DATE/TIME OF TEST: 12/26/2022 00:49:27

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.082	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.083	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 01/18/2023 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSICK,ROBERT,C
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: BUSICK,ROBERT
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: BLANKS

DATE/TIME OF TEST: 12/26/2022 00:54:42

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.083	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Control Test	0.083	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 01/18/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 12/26/2022 16:56:50

Diagnostics Check	OK	16:53	
Air Blank	0.000	16:54	
Control Test	0.082	16:54	
Air Blank	0.000	16:55	
Subject Sample	# 1	16:55	Subject Test Refused
Breath Volume	0.000	16:55	
Air Blank	0.000	16:55	
Control Test	0.082	16:56	
Air Blank	0.000	16:56	
Diagnostics Check	OK	16:56	

Data Download Date/Time: 01/18/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: BATTAGLIA,PAUL
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/11/2022 04:05:18

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.202	03:57
Breath Volume	3.925	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.181	04:00
Breath Volume	2.753	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 3	0.181	04:03
Breath Volume	1.890	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJERA,M,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: TEJERA,MARTHA
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 12/16/2022 17:36:48

Diagnostics Check	OK	17:29
Air Blank	0.000	17:30
Control Test	0.081	17:30
Air Blank	0.000	17:31
Subject Sample # 1	0.152	17:31
Breath Volume	3.382	17:31
Air Blank	0.000	17:32
Air Blank	0.000	17:33
Subject Sample # 2	0.139	17:35
Breath Volume	1.953	17:35
Air Blank	0.000	17:35
Control Test	0.080	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 12/17/2022 01:52:40

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.079	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: BATTAGLIA,PAUL
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: DUARTE GUERRA

DATE/TIME OF TEST: 12/17/2022 03:40:57

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.137	03:36
Breath Volume	3.511	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.146	03:39
Breath Volume	3.644	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTLE,GLENNA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GREER,SYLVIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 12/22/2022 16:16:19

Diagnostics Check	OK	16:07
Air Blank	0.000	16:07
Control Test	0.081	16:08
Air Blank	0.000	16:08
Subject Sample # 1	0.000	16:10
Breath Volume	1.316	16:10
Air Blank	0.000	16:11
Air Blank	0.000	16:13
Subject Sample # 2	0.000	16:14
Breath Volume	1.167	16:14
Air Blank	0.000	16:15
Control Test	0.080	16:15
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTLE,GLENNA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GREER,SYLVIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 12/22/2022 16:36:26

Diagnostics Check	OK	16:29
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:30
Subject Sample # 1	0.000	16:31
Breath Volume	1.335	16:31
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample # 2	0.000	16:35
Breath Volume	1.203	16:35
Air Blank	0.000	16:35
Control Test	0.079	16:35
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: BATTAGLIA,PAUL
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: URBAY

DATE/TIME OF TEST: 12/23/2022 02:18:31

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.250	02:13
Breath Volume	1.542	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.245	02:16
Breath Volume	3.644	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: BATTAGLIA,PAUL
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: QUINTERO PENAGOS

DATE/TIME OF TEST: 12/24/2022 02:29:26

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.018	02:24
Breath Volume	2.371	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.017	02:27
Breath Volume	2.914	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILANES,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GONZALEZ,THALIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 12/25/2022 15:22:34

Diagnostics Check	OK	15:19	
Air Blank	0.000	15:20	
Control Test	0.082	15:20	
Air Blank	0.000	15:20	
Subject Sample	# 1	15:21	Subject Test Refused
Breath Volume	0.000	15:21	
Air Blank	0.000	15:21	
Control Test	0.081	15:21	
Air Blank	0.000	15:22	
Diagnostics Check	OK	15:22	

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE JESUS,RAY,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: CHAVES,JULIAN
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: FREDERIC

DATE/TIME OF TEST: 12/25/2022 23:51:19

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.170	23:46
Breath Volume	1.339	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.165	23:49
Breath Volume	1.542	23:49
Air Blank	0.000	23:50
Control Test	0.081	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA, IAN, O
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: SANTANA, IAN
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SORGE

DATE/TIME OF TEST: 12/31/2022 03:17:59

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.200	03:13
Breath Volume	2.894	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.199	03:16
Breath Volume	2.453	03:16
Air Blank	0.000	03:17
Control Test	0.077	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 01/05/2023 07:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GALLARDO

DATE/TIME OF TEST: 12/22/2022 03:04:02

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 01/05/2023 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GALLARDO

DATE/TIME OF TEST: 12/22/2022 03:30:28

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.150	03:18	
Breath Volume	1.730	03:18	
Air Blank	0.000	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 2 0.157	03:23	Volume Not Met
Breath Volume	0.050	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 3 0.152	03:28	
Breath Volume	1.902	03:28	
Air Blank	0.000	03:29	
Control Test	0.081	03:29	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 01/05/2023 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MATON-MARCOS

DATE/TIME OF TEST: 12/30/2022 00:15:25

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.190	00:10
Breath Volume	2.660	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.183	00:13
Breath Volume	1.605	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 01/05/2023 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/18/2022 19:39:24

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.079	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.116	19:35
Breath Volume	1.230	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.115	19:37
Breath Volume	1.296	19:37
Air Blank	0.000	19:38
Control Test	0.079	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 01/05/2023 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SMITH-FARMER

DATE/TIME OF TEST: 12/20/2022 21:51:44

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:34	
Control Test	0.080	21:35	
Air Blank	0.000	21:35	
Subject Sample	# 1 0.234	21:38	Volume Not Met
Breath Volume	0.691	21:38	
Air Blank	0.000	21:39	
Air Blank	0.000	21:41	
Subject Sample	# 2 0.220	21:44	
Breath Volume	1.117	21:44	
Air Blank	0.000	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 3 0.225	21:49	
Breath Volume	1.199	21:49	
Air Blank	0.000	21:50	
Control Test	0.079	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 01/05/2023 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ-LOPEZ,RICHARD,A
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GOMEZ-LOPEZ,RICHARD
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CASTRO PEREZ

DATE/TIME OF TEST: 12/10/2022 21:42:07

Diagnostics Check	OK	21:41	
Air Blank		21:41	Ambient Fail
Air Blank	0.000	21:42	

Data Download Date/Time: 01/05/2023 20:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ-LOPEZ,RICHARD,A
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: GOMEZ-LOPEZ,RICHARD
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CASTRO-PEREZ

DATE/TIME OF TEST: 12/10/2022 21:50:25

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.077	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.205	21:45
Breath Volume	2.039	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.201	21:48
Breath Volume	2.046	21:48
Air Blank	0.000	21:49
Control Test	0.076	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 01/05/2023 20:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, MAURICIO,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: COOPER, HENRIQUE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 12/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00521080A2
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 12/13/2022 22:21:11

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:12	
Control Test	0.079	22:13	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:16	No Sample Provided
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 01/05/2023 20:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JONES,DALLAS,L

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: LEWIS,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1482048

DRY GAS EXPIRY DATE: 04/21/2024

VIOLATION CODE: DUI

ARREST TIME: 22:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: ALMARAZ

DATE/TIME OF TEST: 12/16/2022 22:52:12

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.079	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 01/12/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FHP TROOP A	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1482048
BREATH TEST OPERATOR: JONES,DALLAS,L	DRY GAS EXPIRY DATE: 04/21/2024
BREATH TEST OPERATOR AGENCY: FHP Troop A	VIOLATION CODE: DUI
ARRESTING OFFICER: LEWIS,TIMOTHY	ARREST TIME: 21:12
ARRESTING OFFICER AGENCY: FHP Troop A	GENDER: Male
LAST AGENCY INSPECTION: 12/01/2022	OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: KING	DATE/TIME OF TEST: 12/22/2022	21:41:05
---------------------------	-------------------------------	----------

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.079	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.079	21:36
Breath Volume	2.910	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:38
Subject Sample # 2	0.079	21:39
Breath Volume	3.269	21:39
Air Blank	0.000	21:40
Control Test	0.078	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:41

Data Download Date/Time: 01/12/2023 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/01/2022 01:01:55

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.082	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUINN,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DANIELLE

DATE/TIME OF TEST: 12/01/2022 23:18:34

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.081	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1 0.220	23:07	
Breath Volume	1.535	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 2 0.199	23:13	Volume Not Met
Breath Volume	0.832	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 3 0.215	23:17	
Breath Volume	2.082	23:17	
Air Blank	0.000	23:17	
Control Test	0.079	23:17	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:18	

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 12/02/2022 00:50:37

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.194	00:45
Breath Volume	1.164	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.198	00:49
Breath Volume	1.468	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/02/2022 02:53:02

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.164	02:48
Breath Volume	1.355	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.154	02:51
Breath Volume	1.582	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:53

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: DIMITROV

DATE/TIME OF TEST: 12/02/2022 04:23:03

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.156	04:18
Breath Volume	1.375	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.150	04:21
Breath Volume	2.113	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:23

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,K,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/02/2022 07:06:30

Diagnostics Check	OK	07:00
Air Blank	0.000	07:00
Control Test	0.081	07:00
Air Blank	0.000	07:01
Subject Sample # 1	0.081	07:02
Breath Volume	1.734	07:02
Air Blank	0.000	07:02
Air Blank	0.000	07:04
Subject Sample # 2	0.083	07:05
Breath Volume	2.742	07:05
Air Blank	0.000	07:05
Control Test	0.080	07:05
Air Blank	0.000	07:06
Diagnostics Check	OK	07:06

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KNOPICK,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/02/2022 23:24:37

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.164	23:20
Breath Volume	3.628	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.166	23:23
Breath Volume	3.265	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 12/03/2022 01:11:08

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.081	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.081	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASSAFUME,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 12/03/2022 02:19:09

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.165	02:14
Breath Volume	1.710	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.171	02:17
Breath Volume	2.128	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 12/03/2022 04:36:00

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:35	
Control Test	0.081	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 12/07/2022 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: DITMAN

DATE/TIME OF TEST: 12/03/2022 05:17:41

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:12	
Control Test	0.080	05:12	
Air Blank	0.000	05:13	
Subject Sample # 1	0.165	05:13	
Breath Volume	2.753	05:13	
Air Blank	0.000	05:14	
Air Blank	0.000	05:16	
Subject Sample # 2		05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.080	05:17	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: DITMAN

DATE/TIME OF TEST: 12/03/2022 05:26:40

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.174	05:22
Breath Volume	3.238	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.172	05:25
Breath Volume	1.625	05:25
Air Blank	0.000	05:25
Control Test	0.081	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FOUST

DATE/TIME OF TEST: 12/03/2022 23:35:10

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.081	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.000	23:27	Volume Not Met
Breath Volume	0.035	23:27	
Air Blank	0.000	23:27	
Air Blank	0.000	23:29	
Subject Sample	# 2 0.000	23:30	
Breath Volume	1.859	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 3 0.000	23:33	
Breath Volume	1.542	23:33	
Air Blank	0.000	23:34	
Control Test	0.081	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CHOATE

DATE/TIME OF TEST: 12/03/2022 23:54:26

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.082	23:49
Breath Volume	2.660	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.084	23:52
Breath Volume	2.234	23:52
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SALAFIA

DATE/TIME OF TEST: 12/04/2022 00:40:16

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.000	00:35
Breath Volume	1.976	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.000	00:38
Breath Volume	2.242	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 12/04/2022 05:16:22

Diagnostics Check	OK	05:11	
Air Blank	0.000	05:12	
Control Test	0.081	05:12	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.082	05:15	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: IKIRIMA

DATE/TIME OF TEST: 12/04/2022 05:26:17

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.082	05:24	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:24	Subject Test Refused
Breath Volume	0.000	05:24	
Air Blank	0.000	05:25	
Control Test	0.081	05:25	
Air Blank	0.000	05:26	
Diagnostics Check	OK	05:26	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: DRISCOLL

DATE/TIME OF TEST: 12/04/2022 19:40:02

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:35	
Control Test	0.081	19:35	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:39	
Control Test	0.081	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RILL, DEP AUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/04/2022 21:18:32

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.081	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.167	21:14
Breath Volume	3.382	21:14
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.166	21:17
Breath Volume	2.632	21:17
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM, DEP DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 12/04/2022 22:36:27

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.081	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.174	22:31
Breath Volume	1.675	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.169	22:34
Breath Volume	1.925	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: KORDIC

DATE/TIME OF TEST: 12/05/2022 00:54:12

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.081	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:53	
Control Test	0.081	00:53	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/05/2022 04:35:09

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:34	
Control Test	0.081	04:34	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHEELER,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 12/06/2022 01:37:26

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Control Test	0.081	01:36	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY, TYLER,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF, R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MAJORS

DATE/TIME OF TEST: 12/06/2022 03:17:34

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.218	03:12
Breath Volume	3.031	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.212	03:16
Breath Volume	2.511	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAVER,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: NICKLAUS

DATE/TIME OF TEST: 12/07/2022 18:39:22

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:36	
Control Test	0.081	18:37	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:38	Subject Test Refused
Breath Volume	0.000	18:38	
Air Blank	0.000	18:38	
Control Test	0.081	18:38	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:39	

Data Download Date/Time: 12/07/2022 22:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 12/08/2022 00:09:05

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:08	
Control Test	0.081	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 01/23/2023 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EASTTY,MARK,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAVER,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 12/08/2022 17:31:54

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.082	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:27
Breath Volume	2.117	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:30
Breath Volume	2.218	17:30
Air Blank	0.000	17:31
Control Test	0.081	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 01/23/2023 09:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RILL,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: NAYLOR

DATE/TIME OF TEST: 12/08/2022 23:33:38

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.081	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.166	23:28
Breath Volume	1.136	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:30
Subject Sample # 2	0.160	23:32
Breath Volume	1.113	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JENKINS,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KARLS

DATE/TIME OF TEST: 12/10/2022 01:53:05

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:48
Breath Volume	2.597	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.000	01:51
Breath Volume	2.773	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ASHTON

DATE/TIME OF TEST: 12/10/2022 02:25:32

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.081	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PENNAMON,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/10/2022 03:37:03

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.116	03:32
Breath Volume	2.363	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.113	03:35
Breath Volume	2.945	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBOGHINI,P
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SCOTTO

DATE/TIME OF TEST: 12/10/2022 03:58:16

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.081	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MCKEON

DATE/TIME OF TEST: 12/10/2022 04:18:51

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.216	04:14
Breath Volume	1.246	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.202	04:17
Breath Volume	1.320	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAUGHERTY,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SANTOS-SANTIAGO

DATE/TIME OF TEST: 12/10/2022 04:38:12

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.145	04:33
Breath Volume	1.781	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.149	04:36
Breath Volume	1.710	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: LOVELAND

DATE/TIME OF TEST: 12/11/2022 05:13:45

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.206	05:09
Breath Volume	1.964	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.210	05:12
Breath Volume	2.011	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 12/11/2022 22:00:05

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:57	
Control Test	0.081	21:57	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.081	21:59	
Air Blank	0.000	21:59	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRIM,TYRONE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: LUCAS SANTOS

DATE/TIME OF TEST: 12/12/2022 07:31:07

Diagnostics Check	OK	07:26	
Air Blank	0.000	07:26	
Control Test	0.081	07:27	
Air Blank	0.000	07:27	
Subject Sample	# 1	07:29	Subject Test Refused
Breath Volume	0.000	07:29	
Air Blank	0.000	07:30	
Control Test	0.081	07:30	
Air Blank	0.000	07:31	
Diagnostics Check	OK	07:31	

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASSAFUME,ANTHONY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: STYERS

DATE/TIME OF TEST: 12/13/2022 03:51:43

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.116	03:46
Breath Volume	1.472	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.120	03:50
Breath Volume	2.035	03:50
Air Blank	0.000	03:50
Control Test	0.078	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 12/13/2022 22:57:38

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:53
Breath Volume	2.683	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:56
Breath Volume	2.882	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 01/23/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/14/2022 03:10:17

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.081	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 01/23/2023 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: RAKOCINSKI

DATE/TIME OF TEST: 12/15/2022 04:30:54

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.000	04:26
Breath Volume	1.386	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.000	04:29
Breath Volume	1.175	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 01/23/2023 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 12/15/2022 21:36:14

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.143	21:31
Breath Volume	2.617	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.137	21:34
Breath Volume	1.632	21:34
Air Blank	0.000	21:35
Control Test	0.081	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 01/23/2023 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 12/15/2022 22:06:45

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.165	22:01
Breath Volume	2.398	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.160	22:05
Breath Volume	2.746	22:05
Air Blank	0.000	22:05
Control Test	0.081	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 01/23/2023 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 12/16/2022 03:21:29

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.000	03:17
Breath Volume	1.730	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.000	03:20
Breath Volume	2.015	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 01/23/2023 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRENNAN,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: GIMGELESKI

DATE/TIME OF TEST: 12/17/2022 04:36:30

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.128	04:31
Breath Volume	1.425	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.128	04:34
Breath Volume	2.261	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 01/23/2023 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ROSADO CINTRON

DATE/TIME OF TEST: 12/17/2022 21:44:40

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.081	21:42	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:43	
Control Test	0.081	21:44	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 01/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MATHEWS,J
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 12/17/2022 22:16:37

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.203	22:11
Breath Volume	3.125	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.201	22:15
Breath Volume	2.558	22:15
Air Blank	0.000	22:15
Control Test	0.081	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 01/23/2023 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBLEY, TRAVIS, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIBLEY, TRAVIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 12/17/2022 23:16:41

Diagnostics Check	OK	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.000	23:12
Breath Volume	3.121	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.000	23:15
Breath Volume	3.351	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 01/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIIGGS,J
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: ZAWACKI

DATE/TIME OF TEST: 12/18/2022 00:53:12

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 01/23/2023 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: STEFKE

DATE/TIME OF TEST: 12/18/2022 03:34:16

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.198	03:29
Breath Volume	1.621	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.190	03:32
Breath Volume	1.917	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/23/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: UPSHAW

DATE/TIME OF TEST: 12/18/2022 05:43:57

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.081	05:41	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:42	Subject Test Refused
Breath Volume	0.000	05:42	
Air Blank	0.000	05:43	
Control Test	0.080	05:43	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 01/23/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/18/2022 06:22:28

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:11	
Control Test	0.081	06:11	
Air Blank	0.000	06:12	
Subject Sample	# 1 0.207	06:15	Volume Not Met
Breath Volume	0.000	06:15	
Air Blank	0.000	06:15	
Air Blank	0.000	06:17	
Subject Sample	# 2 0.195	06:20	Volume Not Met
Breath Volume	0.644	06:20	
Air Blank	0.000	06:21	
Control Test	0.080	06:21	
Air Blank	0.000	06:22	
Diagnostics Check	OK	06:22	

Data Download Date/Time: 01/23/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/18/2022 06:33:59

Diagnostics Check	OK	06:26
Air Blank	0.000	06:26
Control Test	0.080	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.209	06:27
Breath Volume	1.714	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.191	06:32
Breath Volume	1.523	06:32
Air Blank	0.000	06:33
Control Test	0.081	06:33
Air Blank	0.000	06:33
Diagnostics Check	OK	06:33

Data Download Date/Time: 01/23/2023 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: GREENFIELD

DATE/TIME OF TEST: 12/19/2022 04:01:19

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.080	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 01/23/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCFERRIN,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/21/2022 03:45:35

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 01/23/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: WASSON OLDEN

DATE/TIME OF TEST: 12/21/2022 22:06:17

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.081	22:04	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:05	
Control Test	0.081	22:05	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 01/23/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 12/22/2022 03:13:44

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.145	03:09
Breath Volume	2.281	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.141	03:12
Breath Volume	3.097	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/23/2023 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 12/22/2022 04:12:30

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.081	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 01/23/2023 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: DIBIASE

DATE/TIME OF TEST: 12/22/2022 21:04:11

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.000	20:59
Breath Volume	1.871	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:02
Breath Volume	1.609	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAPPIELLO,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 12/23/2022 05:02:44

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.000	04:58
Breath Volume	3.039	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.000	05:01
Breath Volume	2.554	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: OCONNELL

DATE/TIME OF TEST: 12/23/2022 06:27:16

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:24	
Control Test	0.080	06:25	
Air Blank	0.000	06:25	
Subject Sample	# 1	06:25	Subject Test Refused
Breath Volume	0.000	06:25	
Air Blank	0.000	06:26	
Control Test	0.080	06:26	
Air Blank	0.000	06:27	
Diagnostics Check	OK	06:27	

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: DUCHARME

DATE/TIME OF TEST: 12/23/2022 21:51:15

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:48	
Control Test	0.082	21:49	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:49	Subject Test Refused
Breath Volume	0.000	21:49	
Air Blank	0.000	21:50	
Control Test	0.081	21:50	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARQUEZ,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 12/23/2022 23:58:53

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.081	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:58	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/24/2022 00:38:12

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.081	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: VIAMONTES ALVAREZ

DATE/TIME OF TEST: 12/24/2022 03:56:38

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.203	03:52
Breath Volume	2.843	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.203	03:55
Breath Volume	2.722	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAVIS,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 12/25/2022 05:20:41

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.081	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.160	05:10
Breath Volume	2.019	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.181	05:14
Breath Volume	2.691	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 3	0.173	05:19
Breath Volume	1.894	05:19
Air Blank	0.000	05:19
Control Test	0.081	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: HOWARD JR

DATE/TIME OF TEST: 12/28/2022 20:41:12

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.122	20:36
Breath Volume	1.968	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.123	20:39
Breath Volume	1.941	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERY,DIEDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 12/28/2022 21:22:20

Diagnostics Check	OK	21:19	
Air Blank	0.000	21:19	
Control Test	0.081	21:20	
Air Blank	0.000	21:20	
Subject Sample	# 1	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 01/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PURIC

DATE/TIME OF TEST: 12/29/2022 23:14:07

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.020	23:09
Breath Volume	2.476	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.020	23:12
Breath Volume	2.117	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/23/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: TRIPI

DATE/TIME OF TEST: 12/30/2022 16:22:34

Diagnostics Check	OK	16:12	
Air Blank	0.000	16:13	
Control Test	0.080	16:13	
Air Blank	0.000	16:14	
Subject Sample	# 1 0.032	16:17	Volume Not Met
Breath Volume	0.953	16:17	
Air Blank	0.000	16:17	
Air Blank	0.000	16:19	
Subject Sample	# 2	16:21	Subject Test Refused
Breath Volume	0.750	16:21	
Air Blank	0.000	16:21	
Control Test	0.080	16:21	
Air Blank	0.000	16:22	
Diagnostics Check	OK	16:22	

Data Download Date/Time: 01/23/2023 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/30/2022 16:33:40

Diagnostics Check	OK	16:27
Air Blank	0.000	16:28
Control Test	0.081	16:28
Air Blank	0.000	16:28
Subject Sample # 1	0.000	16:29
Breath Volume	2.132	16:29
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample # 2	0.000	16:32
Breath Volume	1.984	16:32
Air Blank	0.000	16:32
Control Test	0.081	16:33
Air Blank	0.000	16:33
Diagnostics Check	OK	16:33

Data Download Date/Time: 01/23/2023 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STICKNEY,MELANIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 12/30/2022 18:41:08

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:38	
Control Test	0.081	18:38	
Air Blank	0.000	18:39	
Subject Sample	# 1	18:39	Subject Test Refused
Breath Volume	0.000	18:39	
Air Blank	0.000	18:40	
Control Test	0.080	18:40	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:41	

Data Download Date/Time: 01/23/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 12/31/2022 01:51:39

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.190	01:47
Breath Volume	2.222	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.196	01:50
Breath Volume	1.628	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 01/23/2023 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 12/31/2022 03:10:47

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 01/23/2023 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: GALTO

DATE/TIME OF TEST: 12/31/2022 21:34:42

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.081	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:30
Breath Volume	1.898	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:33
Breath Volume	2.445	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 01/23/2023 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BACK,REID,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ESPINO

DATE/TIME OF TEST: 12/01/2022 01:25:48

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.082	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.000	01:21
Breath Volume	2.417	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.000	01:24
Breath Volume	2.335	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 12/07/2022 22:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLINS,JOHN,W
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MULLINS,JOHN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: STUART

DATE/TIME OF TEST: 12/02/2022 04:47:02

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.082	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.172	04:42
Breath Volume	1.296	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.168	04:45
Breath Volume	1.960	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 12/07/2022 22:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MATTOX,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 12/02/2022 08:04:48

Diagnostics Check	OK	08:00	
Air Blank	0.000	08:00	
Control Test	0.081	08:01	
Air Blank	0.000	08:01	
Subject Sample	# 1	08:03	Subject Test Refused
Breath Volume	0.000	08:03	
Air Blank	0.000	08:03	
Control Test	0.081	08:04	
Air Blank	0.000	08:04	
Diagnostics Check	OK	08:04	

Data Download Date/Time: 12/07/2022 22:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 12/02/2022 23:24:51

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.327	23:20
Breath Volume	2.195	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.320	23:23
Breath Volume	1.675	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 12/07/2022 22:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REDZEPI,IMAN
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 12/03/2022 10:44:07

Diagnostics Check	OK	10:30	
Air Blank	0.000	10:31	
Control Test	0.081	10:31	
Air Blank	0.000	10:32	
Subject Sample	# 1 0.264	10:35	Volume Not Met
Breath Volume	0.398	10:35	
Air Blank	0.000	10:36	
Air Blank	0.000	10:37	
Subject Sample	# 2 0.321	10:38	
Breath Volume	1.843	10:38	
Air Blank	0.000	10:39	
Air Blank	0.000	10:41	
Subject Sample	# 3 0.305	10:42	
Breath Volume	1.730	10:42	
Air Blank	0.000	10:43	
Control Test	0.079	10:43	
Air Blank	0.000	10:44	
Diagnostics Check	OK	10:44	

Data Download Date/Time: 12/07/2022 22:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/04/2022 00:32:22

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.107	00:27
Breath Volume	1.632	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.107	00:30
Breath Volume	1.734	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 12/07/2022 22:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: THOMAS, DEP ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/04/2022 21:20:33

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:18	
Control Test	0.080	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 12/07/2022 22:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HAWLEY

DATE/TIME OF TEST: 12/11/2022 03:32:45

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.188	03:28
Breath Volume	1.406	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.199	03:31
Breath Volume	2.480	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 01/26/2023 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAPPIELLO,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: STONGE

DATE/TIME OF TEST: 12/13/2022 01:03:52

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.183	00:58
Breath Volume	3.269	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.174	01:02
Breath Volume	2.347	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 01/26/2023 23:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERNANDES,TRAVIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: VALDIVIESO

DATE/TIME OF TEST: 12/15/2022 18:11:51

Diagnostics Check	OK	18:06	
Air Blank	0.000	18:06	
Control Test	0.081	18:06	
Air Blank	0.000	18:07	
Subject Sample	# 1	18:10	Subject Test Refused
Breath Volume	0.082	18:10	
Air Blank	0.000	18:10	
Control Test	0.081	18:11	
Air Blank	0.000	18:11	
Diagnostics Check	OK	18:11	

Data Download Date/Time: 01/26/2023 23:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: BRUCKLER

DATE/TIME OF TEST: 12/17/2022 05:05:41

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:03	
Control Test	0.081	05:03	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSEN,BRANDIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: WEILMANN

DATE/TIME OF TEST: 12/19/2022 00:39:28

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.219	00:34
Breath Volume	4.171	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.223	00:37
Breath Volume	3.582	00:37
Air Blank	0.000	00:38
Control Test	0.082	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DELLECHIAIE,DAKOTA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/19/2022 08:13:41

Diagnostics Check	OK	08:10	
Air Blank	0.000	08:11	
Control Test	0.081	08:11	
Air Blank	0.000	08:12	
Subject Sample	# 1	08:12	Subject Test Refused
Breath Volume	0.000	08:12	
Air Blank	0.000	08:12	
Control Test	0.082	08:13	
Air Blank	0.000	08:13	
Diagnostics Check	OK	08:13	

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILLIAMS,GARY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: KARST

DATE/TIME OF TEST: 12/19/2022 16:15:02

Diagnostics Check	OK	16:08
Air Blank	0.000	16:08
Control Test	0.080	16:09
Air Blank	0.000	16:09
Subject Sample # 1	0.285	16:10
Breath Volume	1.453	16:10
Air Blank	0.000	16:11
Air Blank	0.000	16:12
Subject Sample # 2	0.289	16:13
Breath Volume	1.464	16:13
Air Blank	0.000	16:14
Control Test	0.081	16:14
Air Blank	0.000	16:14
Diagnostics Check	OK	16:15

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VANDERMOLIN,CORY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/22/2022 16:46:20

Diagnostics Check	OK	16:39
Air Blank	0.000	16:40
Control Test	0.082	16:40
Air Blank	0.000	16:41
Subject Sample # 1	0.286	16:41
Breath Volume	3.226	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.301	16:44
Breath Volume	3.476	16:44
Air Blank	0.000	16:45
Control Test	0.081	16:45
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: HOM

DATE/TIME OF TEST: 12/24/2022 03:38:51

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.082	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.180	03:34
Breath Volume	1.484	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.177	03:37
Breath Volume	1.394	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: DUNPHY

DATE/TIME OF TEST: 12/24/2022 23:23:11

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.082	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:22	
Control Test	0.082	23:22	
Air Blank	0.000	23:23	
Diagnostics Check	OK	23:23	

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEALL,C
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/27/2022 03:52:18

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.082	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:51	
Control Test	0.082	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHRISTIE,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CACHO-CASTRO

DATE/TIME OF TEST: 12/27/2022 20:12:54

Diagnostics Check	OK	20:06
Air Blank	0.000	20:06
Control Test	0.082	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.210	20:08
Breath Volume	3.554	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:10
Subject Sample # 2	0.212	20:11
Breath Volume	3.058	20:11
Air Blank	0.000	20:11
Control Test	0.079	20:12
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 01/26/2023 23:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PEREZ CAZUELA

DATE/TIME OF TEST: 12/30/2022 00:39:04

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1 0.174	00:31	Volume Not Met
Breath Volume	0.062	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.199	00:37	Volume Not Met
Breath Volume	1.023	00:37	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 01/26/2023 23:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DENUCCIO,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: DUGGAN

DATE/TIME OF TEST: 12/30/2022 08:17:23

Diagnostics Check	OK	08:15	
Air Blank	0.000	08:16	
Control Test	0.081	08:16	
Air Blank		08:16	RFI Detect
Air Blank	0.000	08:17	

Data Download Date/Time: 01/26/2023 23:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DENUCCIO,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: DUGGAN

DATE/TIME OF TEST: 12/30/2022 08:22:26

Diagnostics Check	OK	08:18	
Air Blank	0.000	08:19	
Control Test	0.081	08:19	
Air Blank	0.000	08:19	
Subject Sample	# 1	08:21	Subject Test Refused
Breath Volume	0.000	08:21	
Air Blank	0.000	08:21	
Control Test	0.082	08:21	
Air Blank	0.000	08:22	
Diagnostics Check	OK	08:22	

Data Download Date/Time: 01/26/2023 23:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 12/31/2022 03:36:32

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.129	03:31
Breath Volume	2.023	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.124	03:34
Breath Volume	1.394	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/26/2023 23:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBLEY, TRAVIS, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIBLEY, TRAVIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: MIZELL

DATE/TIME OF TEST: 12/31/2022 06:16:24

Diagnostics Check	OK	06:10
Air Blank	0.000	06:10
Control Test	0.080	06:11
Air Blank	0.000	06:11
Subject Sample # 1	0.091	06:12
Breath Volume	2.710	06:12
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.089	06:14
Breath Volume	2.480	06:14
Air Blank	0.000	06:15
Control Test	0.081	06:15
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 01/26/2023 23:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/31/2022 23:19:32

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.030	23:14
Breath Volume	1.492	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.029	23:18
Breath Volume	1.292	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 01/26/2023 23:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ
VILLANUEVA,ANTONIO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FERNANDEZ
VILLANUEVA,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024

VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female

OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 12/03/2022 22:44:37

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.146	22:39
Breath Volume	1.273	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.145	22:43
Breath Volume	1.355	22:43
Air Blank	0.000	22:43
Control Test	0.080	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 12/14/2022 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: BOUTS

DATE/TIME OF TEST: 12/04/2022 06:03:28

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.081	06:01	
Air Blank	0.000	06:01	
Subject Sample	# 1	06:02	Subject Test Refused
Breath Volume	0.000	06:02	
Air Blank	0.000	06:02	
Control Test	0.080	06:02	
Air Blank	0.000	06:03	
Diagnostics Check	OK	06:03	

Data Download Date/Time: 12/14/2022 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MATZ

DATE/TIME OF TEST: 12/09/2022 02:45:10

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:40
Breath Volume	2.308	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:43
Breath Volume	2.671	02:43
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 12/14/2022 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HANEY

DATE/TIME OF TEST: 12/17/2022 22:31:57

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.000	22:27
Breath Volume	2.167	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.000	22:30
Breath Volume	1.750	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 01/26/2023 23:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/18/2022 03:54:44

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 01/26/2023 23:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: MARACH

DATE/TIME OF TEST: 12/31/2022 22:07:44

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.231	22:03
Breath Volume	1.261	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.226	22:06
Breath Volume	1.234	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 01/26/2023 23:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 12/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1327124
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 12/24/2022 01:56:00

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 01/17/2023 21:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: PLASENCIA LAZO

DATE/TIME OF TEST: 12/10/2022 03:16:18

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
		03:15	RFI Detect
Air Blank	0.000	03:16	

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: PLASENCIA LAZO

DATE/TIME OF TEST: 12/10/2022 03:23:12

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.082	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: MORENO,JENEFFER
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ANDRE

DATE/TIME OF TEST: 12/11/2022 02:53:43

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.082	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1 0.201	02:45	Volume Not Met
Breath Volume	0.750	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 2 0.207	02:48	
Breath Volume	1.515	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 3 0.210	02:52	
Breath Volume	1.375	02:52	
Air Blank	0.000	02:52	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: CARRILLO PEREZ

DATE/TIME OF TEST: 12/11/2022 05:21:02

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.082	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.239	05:16
Breath Volume	2.031	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.239	05:19
Breath Volume	2.890	05:19
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:21

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: FERREIRA ALVES

DATE/TIME OF TEST: 12/23/2022 04:37:08

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:34	
Control Test	0.081	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: BRAVO,GEORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: LOPEZ CIGARROA

DATE/TIME OF TEST: 12/23/2022 23:43:47

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.082	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.159	23:38
Breath Volume	2.398	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.149	23:42
Breath Volume	1.875	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: VEGA,OBEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PEDROZA

DATE/TIME OF TEST: 12/24/2022 00:44:04

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.083	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.146	00:39
Breath Volume	3.246	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.150	00:42
Breath Volume	4.640	00:42
Air Blank	0.000	00:43
Control Test	0.083	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:44

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: SALGADO,NELSON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 12/24/2022 06:30:03

Diagnostics Check	OK	06:27	
Air Blank	0.000	06:27	
Control Test	0.085	06:27	
Air Blank	0.000	06:28	
Subject Sample	# 1	06:28	Subject Test Refused
Breath Volume	0.000	06:28	
Air Blank	0.000	06:29	
Control Test	0.084	06:29	
Air Blank	0.000	06:29	
Diagnostics Check	OK	06:30	

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: FROST

DATE/TIME OF TEST: 12/31/2022 04:33:21

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.081	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 01/30/2023 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,DANIEL,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN HINOJOSA,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LARIA

DATE/TIME OF TEST: 12/08/2022 03:59:16

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.205	03:54
Breath Volume	1.722	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.197	03:57
Breath Volume	1.113	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 12/08/2022 04:11:24

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.078	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.197	04:06
Breath Volume	2.730	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.186	04:09
Breath Volume	3.550	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,ERIC,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WENNLUND,F
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: CUGINI

DATE/TIME OF TEST: 12/09/2022 20:35:51

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.227	20:31
Breath Volume	1.539	20:31
Air Blank	0.000	20:32
Air Blank	0.000	20:33
Subject Sample # 2	0.240	20:34
Breath Volume	1.328	20:34
Air Blank	0.000	20:34
Control Test	0.077	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,T,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PURSLEY,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: ANAYA

DATE/TIME OF TEST: 12/09/2022 22:31:56

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.078	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.253	22:27
Breath Volume	2.757	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.256	22:30
Breath Volume	3.203	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HENNESSEY,R
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 12/11/2022 03:10:42

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.210	03:06
Breath Volume	3.585	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.204	03:09
Breath Volume	2.703	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VALENTIN,E
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HERRERA PEREZ

DATE/TIME OF TEST: 12/12/2022 00:47:13

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.256	00:42
Breath Volume	1.800	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.261	00:45
Breath Volume	1.640	00:45
Air Blank	0.000	00:46
Control Test	0.078	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 12/13/2022 23:52:06

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.078	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.079	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,J.,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VILLANUEVA,H.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: CALITO-CHUGA

DATE/TIME OF TEST: 12/17/2022 05:06:13

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.160	05:01
Breath Volume	1.824	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.155	05:04
Breath Volume	2.054	05:04
Air Blank	0.000	05:05
Control Test	0.077	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,S,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 12/17/2022 23:33:33

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.015	23:27
Breath Volume	1.367	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.015	23:32
Breath Volume	1.320	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MOORE,CRAIG
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: POPSUB

DATE/TIME OF TEST: 12/18/2022 15:48:11

Diagnostics Check	OK	15:44	
Air Blank	0.000	15:45	
Control Test	0.078	15:45	
Air Blank	0.000	15:45	
Subject Sample	# 1	15:46	Slope Not Met
Breath Volume	2.503	15:46	
Air Blank	0.000	15:47	
Control Test	0.079	15:47	
Air Blank	0.000	15:48	
Diagnostics Check	OK	15:48	

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECORELLA,R,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RIETZ,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: COLDWELL

DATE/TIME OF TEST: 12/19/2022 21:32:01

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.240	21:27
Breath Volume	2.199	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.235	21:30
Breath Volume	2.164	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:32

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAKUBEK,THOMAS,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CRAMER,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/21/2022 20:51:57

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.210	20:47
Breath Volume	1.238	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.219	20:50
Breath Volume	1.527	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLOAT,LUCAS,G
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PURSLEY,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ESTEVES

DATE/TIME OF TEST: 12/22/2022 22:53:27

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.078	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.109	22:49
Breath Volume	2.648	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.109	22:51
Breath Volume	3.621	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: BARFIELD

DATE/TIME OF TEST: 12/22/2022 23:27:56

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:19	
Control Test	0.078	23:20	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:23	No Sample Provided
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample	# 2	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:27	
Control Test	0.078	23:27	
Air Blank	0.000	23:27	
Diagnostics Check	OK	23:27	

Data Download Date/Time: 01/12/2023 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASALA,J,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GORDILLO JUAN

DATE/TIME OF TEST: 12/23/2022 22:45:16

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.131	22:40
Breath Volume	3.324	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:42
Subject Sample # 2	0.131	22:43
Breath Volume	1.859	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 01/12/2023 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEBEL,B,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VELEZ,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 12/24/2022 03:01:10

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
		03:00	RFI Detect
Air Blank	0.000	03:01	

Data Download Date/Time: 01/12/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEBEL,B,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VELEZ,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 12/24/2022 03:11:02

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.142	03:06
Breath Volume	1.269	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.151	03:09
Breath Volume	2.070	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:11

Data Download Date/Time: 01/12/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PATTEN,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 12/24/2022 03:23:14

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.160	03:18
Breath Volume	1.867	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.152	03:21
Breath Volume	1.210	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 01/12/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MACIQUES,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 12/27/2022 03:36:19

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.201	03:31
Breath Volume	1.750	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.200	03:34
Breath Volume	1.636	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/12/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERUZ,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TERUZ,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SHAIN

DATE/TIME OF TEST: 12/27/2022 20:13:54

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.200	20:09
Breath Volume	2.253	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:11
Subject Sample # 2	0.195	20:12
Breath Volume	1.953	20:12
Air Blank	0.000	20:13
Control Test	0.079	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 01/12/2023 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,E,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PALMER,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: CONCEPCONFLORES

DATE/TIME OF TEST: 12/30/2022 05:58:06

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.150	05:53
Breath Volume	1.121	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.152	05:56
Breath Volume	1.132	05:56
Air Blank	0.000	05:57
Control Test	0.078	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 01/12/2023 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RODRIGUEZ,LARRY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/24/2022 22:16:52

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.259	22:12
Breath Volume	1.554	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:14
Subject Sample # 2	0.261	22:15
Breath Volume	2.554	22:15
Air Blank	0.000	22:15
Control Test	0.084	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 01/24/2023 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARL, CLIFFORD, O
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: THIMOTHEE, CHARLES
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: AGUIREE RODRIGUEZ

DATE/TIME OF TEST: 12/25/2022 17:16:37

Diagnostics Check	OK	17:11	
Air Blank	0.000	17:12	
Control Test	0.082	17:12	
Air Blank	0.000	17:13	
Subject Sample # 1	0.150	17:15	
Breath Volume	1.195	17:15	
Air Blank		17:15	Purge Fail
Air Blank	0.000	17:16	

Data Download Date/Time: 01/24/2023 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARL,CLIFFORD,O
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: THIMOTHEE,CHARLES
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: AGUIRRE RODRIGUEZ

DATE/TIME OF TEST: 12/25/2022 17:26:11

Diagnostics Check	OK	17:19
Air Blank	0.000	17:20
Control Test	0.082	17:20
Air Blank	0.000	17:21
Subject Sample # 1	0.158	17:21
Breath Volume	1.371	17:21
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample # 2	0.152	17:24
Breath Volume	2.574	17:24
Air Blank	0.000	17:25
Control Test	0.084	17:25
Air Blank	0.000	17:26
Diagnostics Check	OK	17:26

Data Download Date/Time: 01/24/2023 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIENTOS,KENNETH
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 12/27/2022 22:54:11

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.082	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1	22:52	Subject Test Refused
Breath Volume	0.000	22:52	
Air Blank	0.000	22:53	
Control Test	0.083	22:53	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 01/24/2023 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GARCIA-SANTOS

DATE/TIME OF TEST: 12/28/2022 01:51:15

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.082	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.135	01:46
Breath Volume	2.964	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.131	01:49
Breath Volume	2.605	01:49
Air Blank	0.000	01:50
Control Test	0.082	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 01/24/2023 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIENTOS,KENNETH
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 12/31/2022 01:00:09

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.082	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 01/24/2023 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMM,JAMES,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GAY,TYLER
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 12/10/2022 23:19:39

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:16	
Control Test	0.079	23:16	
Air Blank	0.000	23:17	
Subject Sample	# 1 0.184	23:18	
Breath Volume	1.390	23:18	
Air Blank		23:19	Purge Fail
Air Blank	0.000	23:19	

Data Download Date/Time: 12/21/2022 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMM,JAMES,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GAY,TYLER
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 12/10/2022 23:28:56

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.173	23:24
Breath Volume	1.945	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.163	23:27
Breath Volume	1.511	23:27
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 12/21/2022 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: SCelfo,ZACHARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 12/07/2022 00:23:38

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.083	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	
Control Test	0.081	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 12/16/2022 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J

BREATH TEST OPERATOR AGENCY: Hendry County SO

ARRESTING OFFICER: SCelfo,ZACHARY

ARRESTING OFFICER AGENCY: Hendry County SO

LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1298240

DRY GAS EXPIRY DATE: 03/19/2023

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 12/07/2022 00:31:51

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.115	00:27
Breath Volume	2.593	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.121	00:30
Breath Volume	2.289	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 12/16/2022 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 12/11/2022 20:02:35

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:48	
Control Test	0.081	19:48	
Air Blank	0.000	19:48	
Subject Sample	# 1 0.208	19:49	
Breath Volume	1.363	19:49	
Air Blank	0.000	19:50	
Air Blank	0.000	19:52	
Subject Sample	# 2 0.189	19:55	Volume Not Met
Breath Volume	0.921	19:55	
Air Blank	0.000	19:55	
Air Blank	0.000	19:57	
Subject Sample	# 3 0.181	20:00	Volume Not Met
Breath Volume	0.433	20:00	
Air Blank	0.000	20:01	
Control Test	0.080	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 12/16/2022 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1298240
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 12/11/2022 20:12:24

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.079	20:05
Air Blank	0.000	20:06
Subject Sample # 1	0.219	20:07
Breath Volume	1.503	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.214	20:10
Breath Volume	1.953	20:10
Air Blank	0.000	20:11
Control Test	0.079	20:11
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 12/16/2022 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA, COLBY, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA, COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: NICHELOS

DATE/TIME OF TEST: 12/28/2022 11:36:43

Diagnostics Check	OK	11:30	
Air Blank	0.000	11:30	
Control Test	0.078	11:30	
Air Blank	0.000	11:31	
Subject Sample # 1	0.000	11:32	
Breath Volume	3.296	11:32	
Air Blank	0.000	11:32	
Air Blank	0.000	11:34	
Subject Sample # 2	0.000	11:35	
Breath Volume	3.222	11:35	
Air Blank	0.000	11:35	
Control Test	0.078	11:35	
Air Blank		11:36	RFI Detect
Air Blank	0.000	11:36	

Data Download Date/Time: 01/11/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA,COLBY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA,COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: NICHELOS

DATE/TIME OF TEST: 12/28/2022 11:46:44

Diagnostics Check	OK	11:40
Air Blank	0.000	11:41
Control Test	0.078	11:41
Air Blank	0.000	11:41
Subject Sample # 1	0.000	11:42
Breath Volume	3.132	11:42
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.000	11:45
Breath Volume	3.066	11:45
Air Blank	0.000	11:45
Control Test	0.078	11:46
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 01/11/2023 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FOLTZ

DATE/TIME OF TEST: 12/01/2022 01:39:48

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.168	01:35
Breath Volume	1.121	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.154	01:38
Breath Volume	1.679	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 12/27/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DIAZ, DIANA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WINGFIELD

DATE/TIME OF TEST: 12/04/2022 03:43:25

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.156	03:38
Breath Volume	1.500	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.157	03:41
Breath Volume	2.054	03:41
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 12/27/2022 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,BRANDON,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SHEFFIELD,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BARRAGE

DATE/TIME OF TEST: 12/04/2022 04:41:34

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.182	04:33
Breath Volume	1.257	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.157	04:36
Breath Volume	1.835	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 3	0.173	04:40
Breath Volume	2.339	04:40
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 12/27/2022 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,LEE,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MAGEAU,NATHAN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 12/06/2022 21:15:25

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.229	21:10
Breath Volume	1.734	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.222	21:13
Breath Volume	1.160	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 12/27/2022 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WONSCH,J,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WONSCH,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 12/12/2022 00:03:23

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.079	23:49	
Air Blank	0.000	23:49	
Subject Sample	# 2 0.184	23:52	Volume Not Met
Breath Volume	0.695	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.184	23:56	
Breath Volume	1.417	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.182	00:01	Volume Not Met
Breath Volume	0.246	00:01	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 12/27/2022 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLY, WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SLAPPEY

DATE/TIME OF TEST: 12/31/2022 04:36:10

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.127	04:31
Breath Volume	1.480	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.126	04:34
Breath Volume	1.472	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 01/10/2023 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEFFIELD,BRANDON,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SHEFFIELD,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LANDAVERDE

DATE/TIME OF TEST: 12/18/2022 04:01:47

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.084	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1 0.104	03:55	Volume Not Met
Breath Volume	0.042	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2	04:00	Subject Test Refused
Breath Volume	0.035	04:00	
Air Blank	0.000	04:00	
Control Test	0.082	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 01/24/2023 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSCHETTO,ANTHONY,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MOSCHETTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/22/2022 21:04:46

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.082	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.058	21:00
Breath Volume	1.417	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.057	21:03
Breath Volume	1.523	21:03
Air Blank	0.000	21:03
Control Test	0.082	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 01/24/2023 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 12/16/2022 02:18:08

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.256	02:16	
Breath Volume	1.960	02:16	
Air Blank		02:17	Purge Fail
Air Blank	0.000	02:18	

Data Download Date/Time: 12/21/2022 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 12/16/2022 02:31:15

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.261	02:22
Breath Volume	2.183	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.229	02:26
Breath Volume	1.867	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 3	0.253	02:29
Breath Volume	2.058	02:29
Air Blank	0.000	02:30
Control Test	0.082	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 12/21/2022 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: VOLERSTINE, CRAIG
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: STORY

DATE/TIME OF TEST: 12/20/2022 11:16:57

Diagnostics Check	OK	11:10
Air Blank	0.000	11:11
Control Test	0.081	11:11
Air Blank	0.000	11:12
Subject Sample # 1	0.214	11:12
Breath Volume	2.960	11:12
Air Blank	0.000	11:13
Air Blank	0.000	11:14
Subject Sample # 2	0.213	11:15
Breath Volume	2.691	11:15
Air Blank	0.000	11:16
Control Test	0.083	11:16
Air Blank	0.000	11:16
Diagnostics Check	OK	11:16

Data Download Date/Time: 12/21/2022 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORAN,ETHAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FORAN,ETHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 12/31/2022 02:16:46

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.184	02:12
Breath Volume	3.875	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.185	02:15
Breath Volume	2.968	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 01/25/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARC, VUKANSON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BAPTIST

DATE/TIME OF TEST: 12/03/2022 01:40:30

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.228	01:36
Breath Volume	1.871	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.215	01:39
Breath Volume	1.894	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 12/21/2022 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/03/2022 02:08:15

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.183	02:03
Breath Volume	2.273	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.180	02:06
Breath Volume	1.453	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 12/21/2022 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: ROSARIO JR

DATE/TIME OF TEST: 12/06/2022 18:13:27

Diagnostics Check	OK	18:07
Air Blank	0.000	18:07
Control Test	0.079	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.000	18:09
Breath Volume	1.410	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.000	18:12
Breath Volume	1.269	18:12
Air Blank	0.000	18:12
Control Test	0.080	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 12/21/2022 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/08/2022 01:14:07

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.131	01:09
Breath Volume	1.480	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.129	01:12
Breath Volume	1.328	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSTON,TANNER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: MCLACHLAN II

DATE/TIME OF TEST: 12/08/2022 03:31:48

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.083	03:26
Breath Volume	1.312	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.081	03:30
Breath Volume	1.148	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 12/08/2022 23:58:15

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.000	23:54
Breath Volume	1.835	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:56
Breath Volume	1.898	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/09/2022 19:23:28

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:16	
Control Test	0.079	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1 0.364	19:17	
Breath Volume	1.199	19:17	
Air Blank	0.000	19:18	
Air Blank	0.000	19:20	
Subject Sample	# 2	19:22	Subject Test Refused
Breath Volume	0.000	19:22	
Air Blank	0.000	19:22	
Control Test	0.078	19:22	
Air Blank	0.000	19:23	
Diagnostics Check	OK	19:23	

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: TEJADA

DATE/TIME OF TEST: 12/09/2022 23:12:25

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.177	23:08
Breath Volume	3.078	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.193	23:10
Breath Volume	1.210	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: RAMIREZ CHAVEZ

DATE/TIME OF TEST: 12/10/2022 02:18:19

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.156	02:13
Breath Volume	2.238	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:15
Subject Sample # 2	0.145	02:16
Breath Volume	2.339	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,\
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MANNS,HUNTER
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SACUL

DATE/TIME OF TEST: 12/11/2022 00:12:52

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.168	00:08
Breath Volume	2.171	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.169	00:11
Breath Volume	2.109	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CHIMBO PINEDA

DATE/TIME OF TEST: 12/11/2022 04:18:07

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.158	04:13
Breath Volume	1.125	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.162	04:16
Breath Volume	1.285	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: DERUDDER

DATE/TIME OF TEST: 12/13/2022 21:16:57

Diagnostics Check	OK	21:11
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.000	21:12
Breath Volume	1.675	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.000	21:15
Breath Volume	1.718	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACQUES,JOSHUA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: KAZAROS

DATE/TIME OF TEST: 12/14/2022 02:38:53

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1 0.259	02:30	Volume Not Met
Breath Volume	0.726	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 2 0.257	02:33	
Breath Volume	1.562	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 3 0.264	02:37	
Breath Volume	1.660	02:37	
Air Blank	0.000	02:37	
Control Test	0.078	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAGUER,GUSTAVO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: MIRO CORTEZ

DATE/TIME OF TEST: 12/14/2022 22:21:15

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.204	22:16
Breath Volume	1.503	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.206	22:19
Breath Volume	1.484	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 12/21/2022 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/15/2022 19:42:59

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.079	19:37
Air Blank	0.000	19:38
Subject Sample # 1	0.357	19:38
Breath Volume	2.363	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.367	19:41
Breath Volume	2.207	19:41
Air Blank	0.000	19:42
Control Test	0.079	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 12/21/2022 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZUKOWSKI,ELIJAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GUITERREZ

DATE/TIME OF TEST: 12/15/2022 20:51:26

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.079	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	1.812	20:49	
Air Blank	0.000	20:50	
Control Test	0.080	20:50	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

Data Download Date/Time: 12/21/2022 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: RAYA

DATE/TIME OF TEST: 12/17/2022 01:39:12

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.138	01:34
Breath Volume	2.554	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.133	01:37
Breath Volume	2.011	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 12/21/2022 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNOWLES,ZOE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: OKARI

DATE/TIME OF TEST: 12/17/2022 03:03:53

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.167	02:59
Breath Volume	1.257	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.163	03:02
Breath Volume	1.441	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 12/21/2022 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORDON,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: CANCEL

DATE/TIME OF TEST: 12/22/2022 04:47:29

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.159	04:43
Breath Volume	1.953	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.153	04:46
Breath Volume	2.328	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 01/19/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: ALTENOR

DATE/TIME OF TEST: 12/23/2022 00:32:41

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.117	00:28
Breath Volume	2.156	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.115	00:31
Breath Volume	1.398	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/19/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COMPTON,LARRY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: GILMONIO

DATE/TIME OF TEST: 12/23/2022 03:43:48

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.198	03:39
Breath Volume	1.667	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.211	03:42
Breath Volume	2.398	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 01/19/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: MONTANEZ JR

DATE/TIME OF TEST: 12/23/2022 04:33:20

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.083	04:29
Breath Volume	3.421	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.084	04:31
Breath Volume	3.824	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 01/19/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ADAMS,JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: BARRIOS

DATE/TIME OF TEST: 12/23/2022 09:11:42

Diagnostics Check	OK	09:05
Air Blank	0.000	09:06
Control Test	0.080	09:06
Air Blank	0.000	09:07
Subject Sample # 1	0.000	09:07
Breath Volume	1.515	09:07
Air Blank	0.000	09:07
Air Blank	0.000	09:09
Subject Sample # 2	0.000	09:10
Breath Volume	1.308	09:10
Air Blank	0.000	09:10
Control Test	0.080	09:11
Air Blank	0.000	09:11
Diagnostics Check	OK	09:11

Data Download Date/Time: 01/19/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZVELASQUEZ,JORGE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 12/23/2022 16:55:23

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.080	16:50	
Air Blank	0.000	16:51	
Subject Sample	# 1	16:54	Subject Test Refused
Breath Volume	0.000	16:54	
Air Blank	0.000	16:54	
Control Test	0.080	16:54	
Air Blank	0.000	16:55	
Diagnostics Check	OK	16:55	

Data Download Date/Time: 01/19/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYES,WILFREDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MARTINEZ TEVERA

DATE/TIME OF TEST: 12/24/2022 01:05:51

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Subject Sample # 1	0.126	01:01
Breath Volume	2.519	01:01
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.131	01:04
Breath Volume	3.117	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 01/19/2023 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRESPO, BRYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 12/24/2022 01:49:37

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.042	01:45
Breath Volume	1.128	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.040	01:48
Breath Volume	1.531	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON, KELVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/24/2022 04:43:04

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.226	04:38
Breath Volume	2.277	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.220	04:41
Breath Volume	2.363	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:43

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AMARAL,MARCELO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MIJANDRO RAMIREZ

DATE/TIME OF TEST: 12/24/2022 21:47:49

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1 0.185	21:40	Volume Not Met
Breath Volume	0.644	21:40	
Air Blank	0.000	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 2 0.215	21:43	
Breath Volume	1.351	21:43	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 3 0.196	21:46	
Breath Volume	1.125	21:46	
Air Blank	0.000	21:46	
Control Test	0.080	21:47	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FIGUEROA,ISIAH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SILVESTRE SOSA

DATE/TIME OF TEST: 12/25/2022 02:11:48

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.081	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.210	01:59	
Breath Volume	1.394	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:04	No Sample Provided
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 3	02:10	No Sample Provided
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.081	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCLEAN, NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/25/2022 04:46:31

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.188	04:42
Breath Volume	2.460	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.186	04:45
Breath Volume	2.914	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: JIMENEZ SANTOS

DATE/TIME OF TEST: 12/25/2022 05:18:43

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.081	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.090	05:14
Breath Volume	2.476	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.085	05:17
Breath Volume	1.250	05:17
Air Blank	0.000	05:17
Control Test	0.081	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/25/2022 22:59:07

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.081	22:57	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.081	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA, VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LAUREANO ANDRES

DATE/TIME OF TEST: 12/26/2022 02:22:40

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.225	02:18
Breath Volume	3.300	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.229	02:21
Breath Volume	3.679	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 01/19/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL, GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: HAIDER

DATE/TIME OF TEST: 12/26/2022 07:23:35

Diagnostics Check	OK	07:17
Air Blank	0.000	07:17
Control Test	0.081	07:18
Air Blank	0.000	07:18
Subject Sample # 1	0.000	07:18
Breath Volume	1.281	07:18
Air Blank	0.000	07:19
Air Blank	0.000	07:21
Subject Sample # 2	0.000	07:22
Breath Volume	2.070	07:22
Air Blank	0.000	07:22
Control Test	0.082	07:23
Air Blank	0.000	07:23
Diagnostics Check	OK	07:23

Data Download Date/Time: 01/19/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 12/26/2022 20:21:55

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.081	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.138	20:16
Breath Volume	2.277	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.140	20:20
Breath Volume	1.277	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 01/19/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,JEFFREY
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SHAKUR

DATE/TIME OF TEST: 12/26/2022 22:02:27

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.151	21:57
Breath Volume	1.125	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.167	22:01
Breath Volume	1.316	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 01/19/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/27/2022 03:37:24

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.106	03:32
Breath Volume	1.269	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.102	03:35
Breath Volume	1.238	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/19/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: VIERING

DATE/TIME OF TEST: 12/28/2022 04:08:12

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.198	04:03
Breath Volume	1.476	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.188	04:06
Breath Volume	1.644	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 01/19/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO, GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 12/29/2022 00:32:24

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.187	00:28
Breath Volume	2.078	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.183	00:30
Breath Volume	1.953	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/19/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MARSON FRACARI

DATE/TIME OF TEST: 12/29/2022 02:18:50

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.080	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Air Blank	0.000	02:14	
Subject Sample	# 2	02:17	No Sample Provided
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:18	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 01/19/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 12/29/2022 04:17:27

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.138	04:13
Breath Volume	1.734	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.145	04:16
Breath Volume	3.191	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 01/19/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PETERSON,THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GROMLICH

DATE/TIME OF TEST: 12/30/2022 02:50:10

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.166	02:45
Breath Volume	1.304	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.169	02:48
Breath Volume	1.136	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 01/19/2023 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MOURA MACEDO

DATE/TIME OF TEST: 12/30/2022 03:46:17

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.174	03:41
Breath Volume	1.769	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.172	03:44
Breath Volume	1.289	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 01/19/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO, GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MARTINEZ PEREZ

DATE/TIME OF TEST: 12/31/2022 02:31:58

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:18	
Subject Sample # 1	0.231	02:19	
Breath Volume	1.878	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample # 2	0.227	02:25	Volume Not Met
Breath Volume	0.093	02:25	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample # 3	0.223	02:30	
Breath Volume	1.261	02:30	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 01/19/2023 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH,RODGER
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BURCHFIELD

DATE/TIME OF TEST: 12/31/2022 03:34:12

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.137	03:24
Breath Volume	1.347	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.115	03:28
Breath Volume	2.199	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 3	0.133	03:32
Breath Volume	1.742	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/19/2023 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON,KELVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RORIE

DATE/TIME OF TEST: 12/31/2022 04:01:26

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.045	03:56
Breath Volume	1.656	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.042	04:00
Breath Volume	1.214	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 01/19/2023 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARAMIELLO,CATEY,A
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: MCCONNELL,JEFFERY
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RODRIGDUEZ

DATE/TIME OF TEST: 12/22/2022 00:12:56

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.148	00:08
Breath Volume	1.210	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.147	00:11
Breath Volume	1.988	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 01/29/2023 18:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,BRANDON,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CADY,PETER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 12/23/2022 02:22:00

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.078	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 01/03/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRON,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JAROSZEWSKI,THIAGO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: PEREZ-PEREZ

DATE/TIME OF TEST: 12/23/2022 18:29:34

Diagnostics Check	OK	18:22
Air Blank	0.000	18:22
Control Test	0.079	18:23
Air Blank	0.000	18:23
Subject Sample # 1	0.161	18:24
Breath Volume	1.402	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.162	18:27
Breath Volume	1.906	18:27
Air Blank	0.000	18:28
Control Test	0.079	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 01/03/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRON,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CULPEPPER,DARIUS
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 12/23/2022 18:41:23

Diagnostics Check	OK	18:34
Air Blank	0.000	18:35
Control Test	0.078	18:35
Air Blank	0.000	18:35
Subject Sample # 1	0.304	18:36
Breath Volume	1.457	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.310	18:39
Breath Volume	2.003	18:39
Air Blank	0.000	18:40
Control Test	0.077	18:40
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 01/03/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBERI, FRANCIS, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: VILLEGAS, BRIAN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: VALERA LUCENA

DATE/TIME OF TEST: 12/26/2022 01:13:08

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.080	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1 0.182	01:01	
Breath Volume	2.046	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2 0.165	01:07	Volume Not Met
Breath Volume	0.085	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:09	
Subject Sample	# 3 0.175	01:11	
Breath Volume	2.136	01:11	
Air Blank	0.000	01:12	
Control Test	0.073	01:12	Control Outside Tolerance
Air Blank	0.000	01:13	

Data Download Date/Time: 01/03/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTANEZ,YOMARI,I
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: FONTANEZ,YOMARI
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: CUNDIFF

DATE/TIME OF TEST: 12/28/2022 02:01:10

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.115	01:56
Breath Volume	2.917	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:58
Subject Sample # 2	0.115	01:59
Breath Volume	2.417	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 01/03/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBERI, FRANCIS, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: OWENS, JAIQUAN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FLORES ALVAREZ

DATE/TIME OF TEST: 12/29/2022 04:48:25

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.173	04:43
Breath Volume	2.250	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.157	04:46
Breath Volume	1.800	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 01/03/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ROMAN VILLAMIZAR

DATE/TIME OF TEST: 12/31/2022 03:36:52

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.066	03:32
Breath Volume	1.554	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.073	03:35
Breath Volume	2.832	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/03/2023 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: QUILES-DAVILA,DARIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: DOE 22011162

DATE/TIME OF TEST: 12/31/2022 23:26:16

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.137	23:21
Breath Volume	1.742	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.127	23:24
Breath Volume	1.304	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 01/03/2023 13:36