

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: FAUCI

DATE/TIME OF TEST: 12/08/2022 00:08:37

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.081	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.143	00:04
Breath Volume	4.101	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.134	00:07
Breath Volume	2.675	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 01/03/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DRANGUET

DATE/TIME OF TEST: 12/09/2022 01:35:55

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.046	01:31
Breath Volume	3.613	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.044	01:34
Breath Volume	3.644	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 01/03/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 12/11/2022 01:33:37

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.055	01:29
Breath Volume	2.707	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.054	01:32
Breath Volume	3.679	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 01/03/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 12/17/2022 23:57:42

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.079	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.086	23:52
Breath Volume	2.210	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.087	23:56
Breath Volume	2.230	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 01/03/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/18/2022 01:35:25

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.065	01:30
Breath Volume	3.894	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.064	01:33
Breath Volume	3.507	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 01/03/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 12/18/2022 02:55:12

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.082	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.023	02:50
Breath Volume	3.218	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.020	02:53
Breath Volume	3.050	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/03/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 12/22/2022 22:50:41

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.081	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.026	22:46
Breath Volume	3.167	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.025	22:49
Breath Volume	3.656	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:50
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 01/03/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HURR

DATE/TIME OF TEST: 12/28/2022 22:20:56

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.185	22:16
Breath Volume	4.363	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:18
Subject Sample # 2	0.188	22:19
Breath Volume	3.648	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 01/03/2023 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 12/31/2022 23:33:11

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.007	23:29
Breath Volume	2.300	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:31
Breath Volume	1.343	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 01/03/2023 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PRICE,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: SAMAROO

DATE/TIME OF TEST: 12/01/2022 04:32:31

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.080	04:30	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Control Test	0.081	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 12/26/2022 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: HARELL,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/03/2022 22:16:33

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.173	22:08
Breath Volume	2.046	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.198	22:11
Breath Volume	1.636	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 3	0.190	22:14
Breath Volume	2.042	22:14
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 12/26/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORREST,RYAN,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: FORREST,RYAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: NORENA

DATE/TIME OF TEST: 12/04/2022 23:15:29

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.080	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1 0.168	23:08	Volume Not Met
Breath Volume	0.148	23:08	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 2 0.158	23:13	Volume Not Met
Breath Volume	0.285	23:13	
Air Blank	0.000	23:14	
Control Test	0.079	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 12/26/2022 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CONLEY,SHAWNA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MACKER

DATE/TIME OF TEST: 12/10/2022 03:26:59

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 12/26/2022 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: QUELIZ,CARLOS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: MCGINLEY

DATE/TIME OF TEST: 12/10/2022 20:53:34

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.081	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.179	20:48
Breath Volume	3.957	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.177	20:52
Breath Volume	3.289	20:52
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 12/26/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRABILL,JAMES,H
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: KNOX,KATELYN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/21/2022 01:14:35

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.197	01:10
Breath Volume	1.824	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.199	01:13
Breath Volume	1.347	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/26/2022 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ANDERSON,NATHANIEL
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1475154
DRY GAS EXPIRY DATE: 03/17/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: TEITLOFF

DATE/TIME OF TEST: 12/26/2022 04:03:58

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.253	03:58
Breath Volume	3.183	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.245	04:02
Breath Volume	4.062	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 12/26/2022 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CALEB,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OLIVA,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GALEANO

DATE/TIME OF TEST: 12/01/2022 21:23:25

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEROUX,MARC,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LEROUX,MARC
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JEFF-MCKINLEY

DATE/TIME OF TEST: 12/02/2022 04:08:14

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.075	04:03
Breath Volume	1.562	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.070	04:06
Breath Volume	1.539	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: MCCLENATHAN

DATE/TIME OF TEST: 12/03/2022 18:41:45

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:39	
Control Test	0.081	18:39	
Air Blank	0.000	18:40	
Subject Sample	# 1	18:40	Subject Test Refused
Breath Volume	0.000	18:40	
Air Blank	0.000	18:40	
Control Test	0.081	18:41	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:41	

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARPER, TYRIQUE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SAUCO MALDONADO

DATE/TIME OF TEST: 12/04/2022 04:43:04

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.087	04:38
Breath Volume	2.468	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.090	04:41
Breath Volume	2.000	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:43

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: VARGAS SALINAS

DATE/TIME OF TEST: 12/04/2022 22:27:13

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.208	22:21
Breath Volume	2.542	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.212	22:25
Breath Volume	1.906	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COZZONE,GAETANO
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROTELLA

DATE/TIME OF TEST: 12/04/2022 22:58:34

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.138	22:53
Breath Volume	1.781	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.143	22:57
Breath Volume	1.808	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WICHMAN, MICHELLE
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/07/2022 23:14:21

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.187	23:09
Breath Volume	2.402	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.197	23:12
Breath Volume	2.382	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 12/09/2022 01:55:42

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.167	01:51
Breath Volume	2.347	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.174	01:54
Breath Volume	2.843	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHMIDT, J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: ROKHVARG

DATE/TIME OF TEST: 12/09/2022 09:45:42

Diagnostics Check	OK	09:39
Air Blank	0.000	09:39
Control Test	0.081	09:39
Air Blank	0.000	09:40
Subject Sample # 1	0.211	09:41
Breath Volume	1.503	09:41
Air Blank	0.000	09:41
Air Blank	0.000	09:43
Subject Sample # 2	0.227	09:44
Breath Volume	1.570	09:44
Air Blank	0.000	09:44
Control Test	0.080	09:45
Air Blank	0.000	09:45
Diagnostics Check	OK	09:45

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CARTER, B
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 12/09/2022 19:34:15

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.185	19:27
Breath Volume	1.378	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.167	19:32
Breath Volume	1.390	19:32
Air Blank	0.000	19:33
Control Test	0.081	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 12/10/2022 02:27:37

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:27	Sequence Aborted
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	

Data Download Date/Time: 01/05/2023 09:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 12/10/2022 02:46:32

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:41	
Subject Sample # 1	0.144	02:42	
Breath Volume	3.746	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:44	
Subject Sample # 2	0.143	02:45	
Breath Volume	2.519	02:45	
Air Blank		02:46	RFI Detect
Air Blank	0.000	02:46	

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 12/10/2022 02:55:48

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.138	02:51
Breath Volume	4.140	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.141	02:54
Breath Volume	3.449	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: QUINTERI, D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: HAINES

DATE/TIME OF TEST: 12/11/2022 16:32:00

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.081	16:25
Air Blank	0.000	16:26
Subject Sample # 1	0.263	16:26
Breath Volume	1.550	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.257	16:30
Breath Volume	1.789	16:30
Air Blank	0.000	16:31
Control Test	0.080	16:31
Air Blank	0.000	16:31
Diagnostics Check	OK	16:32

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/15/2022 03:03:24

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.000	02:58
Breath Volume	2.148	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.000	03:02
Breath Volume	1.996	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CALEB,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HUZIOR,N
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SAINTASSE

DATE/TIME OF TEST: 12/16/2022 00:41:53

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1 0.136	00:31	Volume Not Met
Breath Volume	0.660	00:31	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2 0.144	00:36	
Breath Volume	3.042	00:36	
Air Blank	0.000	00:37	
Air Blank	0.000	00:39	
Subject Sample	# 3 0.146	00:40	
Breath Volume	1.671	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAMPS,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PERKINS,KAYLEIGH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: COUSLIN

DATE/TIME OF TEST: 12/22/2022 18:22:50

Diagnostics Check	OK	18:15
Air Blank	0.000	18:15
Control Test	0.080	18:15
Air Blank	0.000	18:16
Subject Sample # 1	0.301	18:17
Breath Volume	2.421	18:17
Air Blank	0.000	18:18
Air Blank	0.000	18:20
Subject Sample # 2	0.301	18:21
Breath Volume	1.792	18:21
Air Blank	0.000	18:21
Control Test	0.080	18:22
Air Blank	0.000	18:22
Diagnostics Check	OK	18:22

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARTLEY,BRANDEN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 12/24/2022 02:30:43

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.081	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.082	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/24/2022 23:47:34

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.082	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.266	23:43
Breath Volume	2.277	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.260	23:46
Breath Volume	2.367	23:46
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: INKMAN

DATE/TIME OF TEST: 12/28/2022 00:10:09

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	RFI Detect
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: INKMAN

DATE/TIME OF TEST: 12/28/2022 00:19:30

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.081	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1 0.254	00:15	
Breath Volume	2.480	00:15	
Air Blank	0.000	00:15	
Air Blank	0.000	00:17	
Subject Sample	# 2 0.248	00:18	
Breath Volume	2.531	00:18	
Air Blank		00:19	RFI Detect
Air Blank	0.000	00:19	

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: INKMAN

DATE/TIME OF TEST: 12/28/2022 00:42:54

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.245	00:37
Breath Volume	1.574	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.241	00:41
Breath Volume	2.265	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICOLAS,MITA,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LOGSDON,DEREK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 12/28/2022 11:59:59

Diagnostics Check	OK	11:54
Air Blank	0.000	11:54
Control Test	0.080	11:54
Air Blank	0.000	11:55
Subject Sample # 1	0.356	11:55
Breath Volume	1.199	11:55
Air Blank	0.000	11:56
Air Blank	0.000	11:58
Subject Sample # 2	0.341	11:58
Breath Volume	1.230	11:58
Air Blank	0.000	11:59
Control Test	0.081	11:59
Air Blank	0.000	11:59
Diagnostics Check	OK	11:59

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON, CALEB, M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SCHRAGER, O
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MASON JR.

DATE/TIME OF TEST: 12/29/2022 19:34:33

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.163	19:28
Breath Volume	1.480	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.157	19:33
Breath Volume	1.359	19:33
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: KITCHENS

DATE/TIME OF TEST: 12/29/2022 20:36:39

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.000	20:32
Breath Volume	1.746	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.000	20:35
Breath Volume	2.050	20:35
Air Blank	0.000	20:35
Control Test	0.081	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/31/2022 02:46:02

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.081	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:45	
Control Test	0.081	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/31/2022 17:22:36

Diagnostics Check	OK	17:16
Air Blank	0.000	17:16
Control Test	0.081	17:17
Air Blank	0.000	17:17
Subject Sample # 1	0.000	17:18
Breath Volume	2.238	17:18
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.000	17:21
Breath Volume	2.089	17:21
Air Blank	0.000	17:21
Control Test	0.080	17:22
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARPER, TYRIQUE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WENTWORTH

DATE/TIME OF TEST: 12/31/2022 21:27:23

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.181	21:22
Breath Volume	1.535	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:24
Subject Sample # 2	0.175	21:25
Breath Volume	1.859	21:25
Air Blank	0.000	21:26
Control Test	0.081	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 01/05/2023 09:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 11/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: NOSSA

DATE/TIME OF TEST: 12/10/2022 01:35:42

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.141	01:31
Breath Volume	2.359	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.141	01:34
Breath Volume	2.125	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 12/14/2022 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 11/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31420080A2
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GAMACHE

DATE/TIME OF TEST: 12/11/2022 03:13:15

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.098	03:07
Breath Volume	1.121	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.098	03:11
Breath Volume	1.355	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 12/14/2022 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,JOSE,M
BREATH TEST OPERATOR AGENCY: BROWARD COUNTY SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: BROWARD COUNTY SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/19/2022 01:33:24

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.081	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.186	01:28
Breath Volume	1.929	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.183	01:31
Breath Volume	2.250	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 01/04/2023 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,TREVOR,W
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: JOHNSON,CURT
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: SNODDY

DATE/TIME OF TEST: 12/04/2022 22:26:27

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.218	22:21
Breath Volume	2.238	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.227	22:24
Breath Volume	1.390	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 12/27/2022 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOUGERE,MARCEL,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: FOUGERE,MARCEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A2
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 12/24/2022 01:07:38

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.158	01:02
Breath Volume	2.347	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.160	01:06
Breath Volume	2.621	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 12/27/2022 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/03/2022 02:06:39

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:02	No Sample Provided
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 2	02:05	RFI Detect
Breath Volume	0.000	02:05	
Air Blank	0.000	02:05	
Control Test	0.081	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 12/21/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/03/2022 02:12:50

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.080	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 12/21/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: WISEMAN

DATE/TIME OF TEST: 12/03/2022 03:28:58

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.139	03:24
Breath Volume	2.183	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.132	03:27
Breath Volume	2.558	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/21/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/17/2022 03:53:38

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.256	03:47
Breath Volume	1.515	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.247	03:51
Breath Volume	1.195	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 12/21/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: FRIELER

DATE/TIME OF TEST: 12/29/2022 02:59:23

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.081	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Air Blank	0.000	02:54	
Subject Sample	# 2	02:58	No Sample Provided
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.081	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 01/17/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JENKINS,ANDRE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 11/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 12/01/2022 23:40:50

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.081	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.153	23:35
Breath Volume	1.335	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.154	23:39
Breath Volume	2.058	23:39
Air Blank	0.000	23:39
Control Test	0.078	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 12/25/2022 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HELMS,WILLIAM
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 11/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 12/16/2022 21:16:09

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.132	21:11
Breath Volume	2.316	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.128	21:14
Breath Volume	2.441	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 12/25/2022 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JENKINS,ANDRE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 11/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 12/20/2022 19:57:09

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.081	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.147	19:51
Breath Volume	1.285	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.146	19:55
Breath Volume	1.542	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 12/25/2022 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HARRIS,LODRICK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: CORLETT

DATE/TIME OF TEST: 12/23/2022 23:00:40

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.313	22:55
Breath Volume	2.058	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.327	22:59
Breath Volume	1.285	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 12/25/2022 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACZAK, DAVID, M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK, AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: DRINKWATER

DATE/TIME OF TEST: 12/11/2022 03:07:18

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.167	03:02
Breath Volume	2.816	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.163	03:05
Breath Volume	2.554	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/27/2022 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MORALES CANO

DATE/TIME OF TEST: 12/15/2022 01:17:21

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.210	01:12
Breath Volume	1.222	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.218	01:15
Breath Volume	1.144	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/27/2022 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SAULTER

DATE/TIME OF TEST: 12/16/2022 04:30:29

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.082	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.087	04:25
Breath Volume	2.921	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.087	04:29
Breath Volume	3.109	04:29
Air Blank	0.000	04:29
Control Test	0.082	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 12/27/2022 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 12/16/2022 06:47:08

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.081	06:41
Air Blank	0.000	06:41
Subject Sample # 1	0.117	06:42
Breath Volume	1.390	06:42
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.121	06:45
Breath Volume	2.867	06:45
Air Blank	0.000	06:46
Control Test	0.082	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 12/27/2022 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLY, WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A4
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SLAPPEY

DATE/TIME OF TEST: 12/31/2022 04:26:36

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.128	04:19
Breath Volume	1.707	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.141	04:22
Breath Volume	1.144	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:26

Data Download Date/Time: 01/10/2023 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LAFLEUR,ANTHONY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: BARNETTE

DATE/TIME OF TEST: 12/18/2022 23:04:52

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.082	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Control Test	0.082	23:04	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 12/28/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: YANICK, M
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RYMER

DATE/TIME OF TEST: 12/10/2022 23:19:35

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.137	23:14
Breath Volume	1.792	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.139	23:18
Breath Volume	2.308	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 01/04/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,BRADLEY,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: DEFALCO,CHRIS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: NACIMIENTO

DATE/TIME OF TEST: 12/11/2022 06:22:42

Diagnostics Check	OK	06:15
Air Blank	0.000	06:16
Control Test	0.083	06:16
Air Blank	0.000	06:17
Subject Sample # 1	0.185	06:17
Breath Volume	2.195	06:17
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.183	06:21
Breath Volume	2.656	06:21
Air Blank	0.000	06:21
Control Test	0.078	06:22
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 01/04/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: COCOA BEACH PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: COCOA BEACH PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 12/16/2022 00:21:56

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.160	00:17
Breath Volume	1.812	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample # 2	0.157	00:20
Breath Volume	1.355	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 01/04/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: COCOA BEACH PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 12/18/2022 20:10:13

Diagnostics Check	OK	20:01
Air Blank	0.000	20:02
Control Test	0.082	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.262	20:04
Breath Volume	1.515	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.263	20:08
Breath Volume	1.496	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 01/04/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: ROSA,EMILIO
ARRESTING OFFICER AGENCY: COCOA BEACH PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: STAUSS

DATE/TIME OF TEST: 12/22/2022 22:11:19

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.185	22:06
Breath Volume	3.222	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.181	22:09
Breath Volume	1.867	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 01/04/2023 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BURGOS, ANTHONY
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: BLYTHE

DATE/TIME OF TEST: 12/01/2022 19:18:07

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:06	
Control Test	0.079	19:06	
Air Blank	0.000	19:07	
Subject Sample	# 1	19:10	No Sample Provided
Breath Volume	0.000	19:10	
Air Blank	0.000	19:10	
Air Blank	0.000	19:12	
Subject Sample	# 2	0.320	19:13
Breath Volume	1.296	19:13	
Air Blank	0.000	19:14	
Air Blank	0.000	19:16	
Subject Sample	# 3	0.324	19:16
Breath Volume	1.472	19:16	
Air Blank	0.000	19:17	
Control Test	0.080	19:17	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,ALEX
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LUNSFORD

DATE/TIME OF TEST: 12/07/2022 23:45:53

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.135	23:41
Breath Volume	1.765	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.131	23:44
Breath Volume	1.640	23:44
Air Blank	0.000	23:44
Control Test	0.078	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BEENER,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GLINKA

DATE/TIME OF TEST: 12/10/2022 04:42:05

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.080	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:41	RFI Detect
Breath Volume	0.074	04:41	
Air Blank		04:41	RFI Detect
Air Blank	0.000	04:42	

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BEENER,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GLINKA

DATE/TIME OF TEST: 12/10/2022 04:57:53

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.079	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.144	04:50	Volume Not Met
Breath Volume	0.828	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.125	04:56	Volume Not Met
Breath Volume	0.250	04:56	
Air Blank	0.000	04:56	
Control Test	0.078	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ, CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 12/10/2022 05:15:34

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.079	05:04	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:08	No Sample Provided
Breath Volume	0.000	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2	05:14	No Sample Provided
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.080	05:14	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PABON,G,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BEENER,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GLINKA

DATE/TIME OF TEST: 12/10/2022 05:25:40

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.080	05:20	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:24	RFI Detect
Breath Volume	0.000	05:24	
Air Blank	0.000	05:24	
Control Test	0.080	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,JIMMIE,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BEENER,D

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 12/11/2022 02:58:26

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.182	02:53
Breath Volume	3.269	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.184	02:56
Breath Volume	3.066	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: TOWNSON, TYLER
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MCCOY JR

DATE/TIME OF TEST: 12/11/2022 05:23:20

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:15	
Control Test	0.080	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 1	05:19	No Sample Provided
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Air Blank	0.000	05:21	
Subject Sample	# 2	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: ESKIN,DEVREN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/11/2022 05:47:52

Diagnostics Check	OK	05:40
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.202	05:43
Breath Volume	1.601	05:43
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.199	05:46
Breath Volume	2.093	05:46
Air Blank	0.000	05:46
Control Test	0.080	05:47
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 12/14/2022 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CAMACHO,NICOLE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ESTRADA LOPEZ

DATE/TIME OF TEST: 12/16/2022 23:49:07

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.193	23:44
Breath Volume	1.652	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.197	23:47
Breath Volume	1.726	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 01/10/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE,KRISTOPER,D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BEENER,KYLE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: NEWCOMB

DATE/TIME OF TEST: 12/24/2022 03:32:09

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.204	03:27
Breath Volume	1.152	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.209	03:30
Breath Volume	2.179	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 01/10/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PABON,G,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GEOGHAGEN

DATE/TIME OF TEST: 12/26/2022 02:27:30

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.164	02:23
Breath Volume	3.683	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.170	02:25
Breath Volume	3.906	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 01/10/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: PEREZ, CALEB
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 12/28/2022 22:20:31

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.115	22:15
Breath Volume	2.277	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.109	22:19
Breath Volume	1.695	22:19
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 01/10/2023 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRADER,E,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SCHRADER,E
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 12/04/2022 13:19:49

Diagnostics Check	OK	13:16	
Air Blank	0.000	13:16	
Control Test	0.080	13:17	
Air Blank	0.000	13:17	
Subject Sample	# 1	13:18	Subject Test Refused
Breath Volume	0.000	13:18	
Air Blank	0.000	13:18	
Control Test	0.080	13:19	
Air Blank	0.000	13:19	
Diagnostics Check	OK	13:19	

Data Download Date/Time: 12/09/2022 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,R,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PENA,R
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 12/06/2022 22:32:03

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.000	22:27
Breath Volume	3.289	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.000	22:30
Breath Volume	2.882	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:32

Data Download Date/Time: 12/09/2022 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, RONALD,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: CLARK, KURTIS
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: SANSBURY

DATE/TIME OF TEST: 12/08/2022 22:48:35

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.080	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Control Test	0.080	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 12/09/2022 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,DANIEL,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ESTEBAN GASPAR

DATE/TIME OF TEST: 12/01/2022 03:18:46

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.161	03:13
Breath Volume	1.234	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.158	03:17
Breath Volume	1.765	03:17
Air Blank	0.000	03:17
Control Test	0.081	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 12/05/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOGER,AARON,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KOGER,AARON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CHARKOWICK

DATE/TIME OF TEST: 12/02/2022 02:56:59

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.222	02:52
Breath Volume	1.242	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.228	02:55
Breath Volume	1.992	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/05/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,DILLAN,R
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: STILTNER,JARRED
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LOREDO

DATE/TIME OF TEST: 12/06/2022 22:54:52

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.082	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.030	22:50
Breath Volume	3.425	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.028	22:53
Breath Volume	3.082	22:53
Air Blank	0.000	22:53
Control Test	0.082	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: PABLO

DATE/TIME OF TEST: 12/08/2022 17:46:03

Diagnostics Check	OK	17:38
Air Blank	0.000	17:38
Control Test	0.082	17:39
Air Blank	0.000	17:39
Subject Sample # 1	0.253	17:40
Breath Volume	2.464	17:40
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 2	0.258	17:44
Breath Volume	2.953	17:44
Air Blank	0.000	17:45
Control Test	0.082	17:45
Air Blank	0.000	17:45
Diagnostics Check	OK	17:46

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDARELLI,CARA,G
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: CARDARELLI,CARA
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: POUNCEY

DATE/TIME OF TEST: 12/10/2022 03:45:50

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.082	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.082	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 12/11/2022 03:55:42

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.083	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.237	03:51
Breath Volume	1.761	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.242	03:54
Breath Volume	1.441	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: SUEN

DATE/TIME OF TEST: 12/11/2022 20:57:08

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.081	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1	20:55	Subject Test Refused
Breath Volume	0.000	20:55	
Air Blank	0.000	20:56	
Control Test	0.082	20:56	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/13/2022 03:48:45

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.201	03:44
Breath Volume	1.355	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.187	03:47
Breath Volume	1.132	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/13/2022 19:24:05

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.082	19:18
Air Blank	0.000	19:19
Subject Sample # 1	0.215	19:19
Breath Volume	3.488	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.222	19:22
Breath Volume	2.789	19:22
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HATCH

DATE/TIME OF TEST: 12/18/2022 02:10:35

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.082	02:09	
Air Blank		02:10	RFI Detect
Air Blank	0.000	02:10	

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HATCH

DATE/TIME OF TEST: 12/18/2022 02:20:06

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.082	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.060	02:15
Breath Volume	3.585	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.058	02:18
Breath Volume	2.988	02:18
Air Blank	0.000	02:19
Control Test	0.082	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PEDRO

DATE/TIME OF TEST: 12/18/2022 23:20:53

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.082	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.205	23:16
Breath Volume	3.855	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.210	23:19
Breath Volume	3.777	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SANTAMARIA,LEO
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WHIPKEY

DATE/TIME OF TEST: 12/20/2022 02:54:01

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.082	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.082	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CLERC

DATE/TIME OF TEST: 12/22/2022 01:46:38

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.082	01:39	
Air Blank	0.000	01:40	
Subject Sample # 1	0.259	01:41	
Breath Volume	2.800	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample # 2		01:45	Slope Not Met
Breath Volume	0.597	01:45	
Air Blank	0.000	01:45	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CLERC

DATE/TIME OF TEST: 12/22/2022 01:58:01

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.250	01:50
Breath Volume	2.335	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.272	01:53
Breath Volume	2.226	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 3	0.265	01:56
Breath Volume	2.382	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:58

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA, ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAMBY

DATE/TIME OF TEST: 12/23/2022 02:29:06

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.082	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.000	02:24
Breath Volume	1.207	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.000	02:27
Breath Volume	2.214	02:27
Air Blank	0.000	02:28
Control Test	0.082	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,JAMES,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 12/23/2022 15:40:44

Diagnostics Check	OK	15:32
Air Blank	0.000	15:32
Control Test	0.081	15:32
Air Blank	0.000	15:33
Subject Sample # 1	0.204	15:35
Breath Volume	2.308	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample # 2	0.193	15:39
Breath Volume	1.972	15:39
Air Blank	0.000	15:39
Control Test	0.078	15:40
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: LIFGREN

DATE/TIME OF TEST: 12/23/2022 22:15:49

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.083	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.012	22:10
Breath Volume	3.574	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.014	22:14
Breath Volume	1.937	22:14
Air Blank	0.000	22:14
Control Test	0.083	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: MANTECON

DATE/TIME OF TEST: 12/26/2022 20:37:27

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.083	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.150	20:32
Breath Volume	1.843	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.142	20:35
Breath Volume	3.273	20:35
Air Blank	0.000	20:36
Control Test	0.082	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: PRETZANTZIN

DATE/TIME OF TEST: 12/26/2022 23:31:22

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.082	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.161	23:26
Breath Volume	1.792	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.155	23:29
Breath Volume	1.265	23:29
Air Blank	0.000	23:30
Control Test	0.082	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 12/27/2022 01:50:46

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.082	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.184	01:46
Breath Volume	1.386	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.189	01:49
Breath Volume	2.199	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPUGNARDI,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SPUGNARDI,TIMOTHY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: COTTRELL

DATE/TIME OF TEST: 12/29/2022 21:54:34

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.082	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.093	21:50
Breath Volume	1.718	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.097	21:53
Breath Volume	1.933	21:53
Air Blank	0.000	21:53
Control Test	0.082	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: VELDEN

DATE/TIME OF TEST: 12/30/2022 22:57:42

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.082	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.133	22:52
Breath Volume	1.246	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.135	22:56
Breath Volume	1.371	22:56
Air Blank	0.000	22:56
Control Test	0.082	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: FRANCISCO-DELEON

DATE/TIME OF TEST: 12/31/2022 00:08:32

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.083	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.163	00:03
Breath Volume	1.777	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.164	00:07
Breath Volume	2.031	00:07
Air Blank	0.000	00:07
Control Test	0.082	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CROFOOT

DATE/TIME OF TEST: 12/31/2022 21:07:58

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.083	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.143	21:03
Breath Volume	2.769	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.142	21:06
Breath Volume	2.859	21:06
Air Blank	0.000	21:07
Control Test	0.082	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SONTAY-ELIAS

DATE/TIME OF TEST: 12/31/2022 23:13:12

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.195	23:08
Breath Volume	2.843	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.193	23:11
Breath Volume	2.546	23:11
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 01/16/2023 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: PRICE III

DATE/TIME OF TEST: 12/01/2022 04:33:14

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.083	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.000	04:28
Breath Volume	2.687	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.000	04:31
Breath Volume	1.406	04:31
Air Blank	0.000	04:32
Control Test	0.083	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 12/18/2022 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 12/06/2022 01:08:23

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.236	01:03
Breath Volume	1.402	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.238	01:06
Breath Volume	1.175	01:06
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/18/2022 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PELFREY,PAUL,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MCCARLEY

DATE/TIME OF TEST: 12/08/2022 02:56:41

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.156	02:52
Breath Volume	1.113	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.158	02:55
Breath Volume	1.664	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/18/2022 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHNATHAN,F
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: YOUNG,JOHNATHAN
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 12/09/2022 02:59:58

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.082	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.020	02:54
Breath Volume	1.425	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.020	02:58
Breath Volume	1.574	02:58
Air Blank	0.000	02:59
Control Test	0.083	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 12/18/2022 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHNATHAN,F
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 12/09/2022 03:14:12

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.189	03:09
Breath Volume	2.343	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.190	03:12
Breath Volume	2.250	03:12
Air Blank	0.000	03:13
Control Test	0.081	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 12/18/2022 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 12/10/2022 01:46:43

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.082	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.209	01:41
Breath Volume	2.343	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.197	01:45
Breath Volume	1.652	01:45
Air Blank	0.000	01:45
Control Test	0.082	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 12/18/2022 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAFA,YUSUF,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CONGER

DATE/TIME OF TEST: 12/11/2022 02:40:20

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.082	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.182	02:35
Breath Volume	2.773	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.179	02:38
Breath Volume	2.382	02:38
Air Blank	0.000	02:39
Control Test	0.082	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 12/18/2022 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HENTSHEL

DATE/TIME OF TEST: 12/16/2022 23:37:21

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.082	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.068	23:32
Breath Volume	2.476	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.071	23:35
Breath Volume	3.710	23:35
Air Blank	0.000	23:36
Control Test	0.083	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 12/18/2022 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HATLEY,KENDALL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BERGMAN

DATE/TIME OF TEST: 12/17/2022 01:09:44

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.082	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.154	01:03
Breath Volume	1.738	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.153	01:08
Breath Volume	1.546	01:08
Air Blank	0.000	01:08
Control Test	0.083	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 12/18/2022 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COUPE,JESSICA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: AGUIRRE ORTEGA

DATE/TIME OF TEST: 12/18/2022 04:10:33

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.082	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.156	04:06
Breath Volume	1.109	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.151	04:09
Breath Volume	1.406	04:09
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 12/18/2022 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HAGLER

DATE/TIME OF TEST: 12/25/2022 04:28:45

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.083	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.171	04:23
Breath Volume	1.234	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.170	04:27
Breath Volume	1.203	04:27
Air Blank	0.000	04:27
Control Test	0.083	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 01/26/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: NALLEY

DATE/TIME OF TEST: 12/28/2022 22:47:24

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.082	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.014	22:42
Breath Volume	1.625	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.014	22:45
Breath Volume	1.250	22:45
Air Blank	0.000	22:46
Control Test	0.082	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 01/26/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MELCHER,ANDREW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CASEY

DATE/TIME OF TEST: 12/30/2022 03:38:57

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.083	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.067	03:34
Breath Volume	2.460	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.067	03:37
Breath Volume	2.371	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 01/26/2023 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,J.,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,A.
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ROJAS-MARIN

DATE/TIME OF TEST: 12/03/2022 21:34:18

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.216	21:29
Breath Volume	1.406	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.206	21:32
Breath Volume	1.652	21:32
Air Blank	0.000	21:33
Control Test	0.079	21:33
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 12/07/2022 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BENOIT,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MORREALE

DATE/TIME OF TEST: 12/04/2022 03:01:56

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.102	02:57
Breath Volume	3.363	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.099	03:00
Breath Volume	3.437	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/07/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENRRIQUEZ,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VILLANUEVA,H
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ALONZOPEREZ

DATE/TIME OF TEST: 12/04/2022 04:10:55

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1 0.167	04:05	
Breath Volume	1.601	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2	04:09	RFI Detect
Breath Volume	0.156	04:09	
Air Blank	0.000	04:09	
Control Test	0.080	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 12/07/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENRRIQUEZ,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VILLANUEVA,H
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ALONZOPEREZ

DATE/TIME OF TEST: 12/04/2022 04:23:23

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 1 0.155	04:16	
Breath Volume	1.480	04:16	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 2	04:21	Subject Test Refused
Breath Volume	0.152	04:21	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 12/07/2022 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PEDRO PEDRO

DATE/TIME OF TEST: 12/09/2022 03:54:21

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.077	03:49
Breath Volume	2.046	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.077	03:52
Breath Volume	2.898	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENCE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SUDANO,KYLE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 12/10/2022 07:22:25

Diagnostics Check	OK	07:15
Air Blank	0.000	07:15
Control Test	0.080	07:16
Air Blank	0.000	07:16
Subject Sample # 1	0.255	07:17
Breath Volume	2.949	07:17
Air Blank	0.000	07:18
Air Blank	0.000	07:19
Subject Sample # 2	0.256	07:20
Breath Volume	3.121	07:20
Air Blank	0.000	07:21
Control Test	0.079	07:21
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARIMON,KARLHEINZ,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TARAZONA,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: VEGAPADRON

DATE/TIME OF TEST: 12/11/2022 07:08:26

Diagnostics Check	OK	07:00
Air Blank	0.000	07:00
Control Test	0.080	07:01
Air Blank	0.000	07:01
Subject Sample # 1	0.120	07:03
Breath Volume	2.992	07:03
Air Blank	0.000	07:04
Air Blank	0.000	07:06
Subject Sample # 2	0.120	07:06
Breath Volume	2.632	07:06
Air Blank	0.000	07:07
Control Test	0.080	07:07
Air Blank	0.000	07:08
Diagnostics Check	OK	07:08

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,JONATHAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GNERRE

DATE/TIME OF TEST: 12/16/2022 01:15:04

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.179	01:10
Breath Volume	3.433	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.192	01:13
Breath Volume	3.519	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:15

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVA,MIGUEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: SHMINA

DATE/TIME OF TEST: 12/16/2022 21:43:49

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.166	21:39
Breath Volume	2.304	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:41
Subject Sample # 2	0.171	21:42
Breath Volume	2.203	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND, W, R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WHITE, D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SHODA

DATE/TIME OF TEST: 12/17/2022 04:47:03

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.160	04:42
Breath Volume	2.570	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.161	04:45
Breath Volume	2.644	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: DE JESUS GONZALEZ

DATE/TIME OF TEST: 12/18/2022 03:46:56

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.158	03:42
Breath Volume	1.464	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.156	03:45
Breath Volume	1.164	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA, URIEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LAFLOWER, J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NAVARRETE PATRICIO

DATE/TIME OF TEST: 12/23/2022 03:16:49

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.233	03:12
Breath Volume	3.500	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.229	03:15
Breath Volume	3.320	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUZMAN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 12/24/2022 04:35:39

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.170	04:30
Breath Volume	1.519	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.171	04:34
Breath Volume	1.476	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RODRIGUEZ MORALES

DATE/TIME OF TEST: 12/30/2022 22:38:45

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.080	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.131	22:34
Breath Volume	1.222	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.132	22:37
Breath Volume	1.753	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MYKHASKIV,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: PEREZGOMEZ

DATE/TIME OF TEST: 12/31/2022 00:50:51

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.202	00:46
Breath Volume	1.621	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.199	00:49
Breath Volume	1.117	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND,W,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENDOZA,U
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SANCHEZTOLENTINO

DATE/TIME OF TEST: 12/31/2022 02:42:50

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.156	02:38
Breath Volume	2.195	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.151	02:41
Breath Volume	1.847	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 01/14/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER,TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/03/2022 13:27:08

Diagnostics Check	OK	13:20
Air Blank	0.000	13:21
Control Test	0.078	13:21
Air Blank	0.000	13:22
Subject Sample # 1	0.046	13:22
Breath Volume	3.402	13:22
Air Blank	0.000	13:23
Air Blank	0.000	13:25
Subject Sample # 2	0.045	13:25
Breath Volume	3.359	13:25
Air Blank	0.000	13:26
Control Test	0.078	13:26
Air Blank	0.000	13:27
Diagnostics Check	OK	13:27

Data Download Date/Time: 01/23/2023 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PAUGA,JOHN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: MARTINEZ ACUNA

DATE/TIME OF TEST: 12/06/2022 18:49:07

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.078	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.192	18:44
Breath Volume	2.089	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.192	18:47
Breath Volume	1.890	18:47
Air Blank	0.000	18:48
Control Test	0.077	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 01/23/2023 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BARRAS,HOPE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SNAVELY

DATE/TIME OF TEST: 12/09/2022 19:10:22

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:06	
Control Test	0.079	19:07	
Air Blank	0.000	19:07	
Subject Sample	# 1	19:09	Subject Test Refused
Breath Volume	0.000	19:09	
Air Blank	0.000	19:09	
Control Test	0.076	19:09	
Air Blank	0.000	19:10	
Diagnostics Check	OK	19:10	

Data Download Date/Time: 01/23/2023 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: TILLER,HANNAH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 10:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 12/12/2022 11:05:50

Diagnostics Check	OK	10:57
Air Blank	0.000	10:58
Control Test	0.077	10:58
Air Blank	0.000	10:58
Subject Sample # 1	0.292	11:00
Breath Volume	1.187	11:00
Air Blank	0.000	11:00
Air Blank	0.000	11:02
Subject Sample # 2	0.273	11:04
Breath Volume	1.398	11:04
Air Blank	0.000	11:04
Control Test	0.077	11:05
Air Blank	0.000	11:05
Diagnostics Check	OK	11:05

Data Download Date/Time: 01/23/2023 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONAHAN,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MONAHAN,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: XON-GARCIA

DATE/TIME OF TEST: 12/24/2022 20:54:56

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.078	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.211	20:49
Breath Volume	2.042	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.215	20:53
Breath Volume	3.042	20:53
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 01/23/2023 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GERARDO PEREZ

DATE/TIME OF TEST: 12/01/2022 02:03:37

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.207	01:55
Breath Volume	3.093	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.184	01:59
Breath Volume	1.386	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 3	0.199	02:02
Breath Volume	2.730	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 12/29/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ROBLERO

DATE/TIME OF TEST: 12/03/2022 01:22:05

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.234	01:17
Breath Volume	2.742	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.234	01:20
Breath Volume	3.281	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:22

Data Download Date/Time: 12/29/2022 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: CORDREY

DATE/TIME OF TEST: 12/03/2022 23:50:44

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.082	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.199	23:45
Breath Volume	2.242	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.182	23:49
Breath Volume	1.195	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,EDWARD,E
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 12/06/2022 22:01:43

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.081	21:54	
Air Blank	0.000	21:54	
Subject Sample	# 1 0.168	21:57	Volume Not Met
Breath Volume	0.132	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample	# 2	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.081	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MCINDOO

DATE/TIME OF TEST: 12/08/2022 23:46:58

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.082	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:42
Breath Volume	1.550	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.000	23:45
Breath Volume	1.679	23:45
Air Blank	0.000	23:46
Control Test	0.082	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PEREZ BRAVO

DATE/TIME OF TEST: 12/11/2022 00:24:31

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.081	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.224	00:19
Breath Volume	2.738	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.218	00:22
Breath Volume	2.816	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PILLOT

DATE/TIME OF TEST: 12/14/2022 02:02:26

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.081	02:00	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.082	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 12/14/2022 23:44:02

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.082	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.197	23:39
Breath Volume	2.011	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.203	23:42
Breath Volume	2.007	23:42
Air Blank	0.000	23:43
Control Test	0.082	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,EDWARD,E
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 12/16/2022 19:14:51

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.082	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.000	19:10
Breath Volume	3.792	19:10
Air Blank	0.000	19:10
Air Blank	0.000	19:12
Subject Sample # 2	0.000	19:13
Breath Volume	3.859	19:13
Air Blank	0.000	19:13
Control Test	0.081	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: VILLANUNEZ

DATE/TIME OF TEST: 12/19/2022 15:21:11

Diagnostics Check	OK	15:18	
Air Blank	0.000	15:18	
Control Test	0.081	15:19	
Air Blank	0.000	15:19	
Subject Sample	# 1	15:19	Subject Test Refused
Breath Volume	0.000	15:19	
Air Blank	0.000	15:20	
Control Test	0.081	15:20	
Air Blank	0.000	15:21	
Diagnostics Check	OK	15:21	

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 12/19/2022 22:33:10

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.193	22:28
Breath Volume	2.449	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.191	22:31
Breath Volume	2.019	22:31
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 12/29/2022 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLARD,DANIEL,C
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 12/28/2022 19:02:27

Diagnostics Check	OK	18:56
Air Blank	0.000	18:56
Control Test	0.082	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.230	18:57
Breath Volume	2.925	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.245	19:00
Breath Volume	3.496	19:00
Air Blank	0.000	19:01
Control Test	0.081	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 12/29/2022 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLARD,DANIEL,C
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DEFILIPPIS,MICHAEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DIAZ BENITEZ

DATE/TIME OF TEST: 12/30/2022 02:44:12

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.164	02:39
Breath Volume	2.796	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.168	02:42
Breath Volume	3.636	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 01/25/2023 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADEMACHER,RICHARD,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BILLUE,MOLLY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: VALDEZ

DATE/TIME OF TEST: 12/03/2022 03:24:36

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.133	03:19
Breath Volume	1.285	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.130	03:22
Breath Volume	1.542	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 12/13/2022 10:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILLUE,MOLLY,S
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 12/24/2022 05:10:17

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.082	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.193	05:05
Breath Volume	1.191	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.187	05:08
Breath Volume	1.507	05:08
Air Blank	0.000	05:09
Control Test	0.081	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 01/18/2023 00:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: CHIAPPA,JOSEPH
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BAIAMONTE

DATE/TIME OF TEST: 12/06/2022 22:38:20

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.000	22:33
Breath Volume	2.628	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.000	22:36
Breath Volume	1.957	22:36
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 12/15/2022 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: WELLS,KYLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: VILLENEUVE

DATE/TIME OF TEST: 12/12/2022 20:15:58

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.080	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.188	20:11
Breath Volume	2.132	20:11
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.169	20:14
Breath Volume	2.000	20:14
Air Blank	0.000	20:15
Control Test	0.081	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 12/15/2022 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: WELLS,KYLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/18/2022 23:20:31

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.082	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.291	23:16
Breath Volume	2.886	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.288	23:18
Breath Volume	2.273	23:18
Air Blank	0.000	23:19
Control Test	0.082	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 01/27/2023 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHMIEDER, NICOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CHAPKIS

DATE/TIME OF TEST: 12/10/2022 03:52:12

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.164	03:47
Breath Volume	1.621	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.157	03:50
Breath Volume	1.382	03:50
Air Blank	0.000	03:51
Control Test	0.077	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 12/21/2022 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/10/2022 04:13:31

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.027	04:09
Breath Volume	4.671	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.029	04:12
Breath Volume	4.769	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/21/2022 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO, GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A3
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/17/2022 04:02:26

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 12/21/2022 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: LOPEZ MENDEZ

DATE/TIME OF TEST: 12/01/2022 01:04:47

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.080	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.178	01:00
Breath Volume	3.222	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.174	01:03
Breath Volume	3.125	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 12/21/2022 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCALES,BRANDON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DOBSON

DATE/TIME OF TEST: 12/01/2022 03:26:26

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.095	03:22
Breath Volume	2.414	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.094	03:24
Breath Volume	2.464	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 12/21/2022 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PITMAN,DAVID
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: O MASTA

DATE/TIME OF TEST: 12/01/2022 03:47:07

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.227	03:42
Breath Volume	3.160	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.214	03:45
Breath Volume	2.707	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH,RODGER
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: SMETHURST

DATE/TIME OF TEST: 12/01/2022 04:14:42

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.141	04:10
Breath Volume	1.605	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.143	04:13
Breath Volume	2.117	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AMARAL,WILKER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: ROSARIO BURGOS

DATE/TIME OF TEST: 12/02/2022 05:06:03

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.082	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.139	05:01
Breath Volume	1.117	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:03
Subject Sample # 2	0.146	05:04
Breath Volume	1.535	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:06

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 12/02/2022 21:26:57

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
Control Test	0.081	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:19	No Sample Provided
Breath Volume	0.000	21:19	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2	21:25	No Sample Provided
Breath Volume	0.000	21:25	
Air Blank	0.000	21:26	
Control Test	0.082	21:26	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 12/02/2022 21:49:40

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.081	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.266	21:44
Breath Volume	1.695	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.274	21:48
Breath Volume	1.992	21:48
Air Blank	0.000	21:48
Control Test	0.079	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 12/03/2022 01:34:06

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.297	01:29
Breath Volume	2.488	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.302	01:32
Breath Volume	3.089	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,RICHARD
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: VAGHENI

DATE/TIME OF TEST: 12/03/2022 03:35:57

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.228	03:31
Breath Volume	1.515	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.232	03:34
Breath Volume	1.976	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD, MIGUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GONZALEZ QUETELL

DATE/TIME OF TEST: 12/03/2022 04:13:42

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.200	04:09
Breath Volume	1.875	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.194	04:12
Breath Volume	1.441	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TERRETT,NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 12/03/2022 05:10:30

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.208	05:06
Breath Volume	2.167	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.216	05:08
Breath Volume	3.046	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 12/21/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KACIN

DATE/TIME OF TEST: 12/03/2022 23:49:17

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.256	23:41
Breath Volume	2.777	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.233	23:44
Breath Volume	1.125	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 3	0.260	23:47
Breath Volume	1.531	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: VAZQUEZ RAMIREZ

DATE/TIME OF TEST: 12/04/2022 00:28:48

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.204	00:24
Breath Volume	3.230	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:26
Subject Sample # 2	0.191	00:27
Breath Volume	1.570	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE, ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BARIE

DATE/TIME OF TEST: 12/04/2022 02:52:47

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.088	02:48
Breath Volume	1.996	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.097	02:51
Breath Volume	2.347	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ALVARADO MUNGUIA

DATE/TIME OF TEST: 12/04/2022 03:17:03

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.082	03:14	
Air Blank	0.000	03:14	
Subject Sample # 1	0.146	03:15	
Breath Volume	1.453	03:15	
Air Blank		03:15	RFI Detect
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ALVARADO MUNGUIA

DATE/TIME OF TEST: 12/04/2022 03:24:24

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.141	03:19
Breath Volume	2.382	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.140	03:22
Breath Volume	1.257	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/04/2022 04:59:56

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.131	04:55
Breath Volume	1.160	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.131	04:58
Breath Volume	1.390	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GANDIA ARES, YORTSAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PEREZ JIMENEZ

DATE/TIME OF TEST: 12/04/2022 05:14:17

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.138	05:09
Breath Volume	2.105	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.132	05:12
Breath Volume	2.578	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ASHLEY JR

DATE/TIME OF TEST: 12/08/2022 01:34:13

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.212	01:29
Breath Volume	2.214	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.206	01:32
Breath Volume	1.378	01:32
Air Blank	0.000	01:33
Control Test	0.078	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SANGSTER

DATE/TIME OF TEST: 12/08/2022 02:52:39

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.082	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.136	02:48
Breath Volume	1.265	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.136	02:51
Breath Volume	1.695	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/08/2022 04:54:36

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.112	04:50
Breath Volume	1.234	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.119	04:53
Breath Volume	1.171	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLIAMS,MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 12/08/2022 06:17:57

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.081	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.171	06:13
Breath Volume	2.054	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.165	06:16
Breath Volume	1.343	06:16
Air Blank	0.000	06:17
Control Test	0.080	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARRAGAN GARCIA,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:13

SUBJECT'S LAST NAME: LEITCH

DATE/TIME OF TEST: 12/09/2022 13:44:49

Diagnostics Check	OK	13:37
Air Blank	0.000	13:38
Control Test	0.081	13:38
Air Blank	0.000	13:39
Subject Sample # 1	0.000	13:40
Breath Volume	1.140	13:40
Air Blank	0.000	13:41
Air Blank	0.000	13:43
Subject Sample # 2	0.000	13:43
Breath Volume	1.476	13:43
Air Blank	0.000	13:43
Control Test	0.080	13:44
Air Blank	0.000	13:44
Diagnostics Check	OK	13:44

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OJEDA COLON,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: JAIME

DATE/TIME OF TEST: 12/09/2022 20:30:39

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.080	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.190	20:26
Breath Volume	3.281	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.199	20:29
Breath Volume	3.238	20:29
Air Blank	0.000	20:29
Control Test	0.080	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,CHRISTIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SALAZAR OBANDO

DATE/TIME OF TEST: 12/10/2022 03:40:45

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.168	03:36
Breath Volume	3.898	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.164	03:39
Breath Volume	4.230	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/21/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PEREZ ESCALANTE

DATE/TIME OF TEST: 12/10/2022 04:08:15

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.204	04:03
Breath Volume	1.132	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.219	04:06
Breath Volume	2.250	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOSS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: CZARNY

DATE/TIME OF TEST: 12/10/2022 21:34:24

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:23	
Control Test	0.081	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1 0.134	21:27	Volume Not Met
Breath Volume	0.777	21:27	
Air Blank	0.000	21:27	
Air Blank	0.000	21:29	
Subject Sample	# 2 0.136	21:32	Volume Not Met
Breath Volume	0.753	21:32	
Air Blank	0.000	21:33	
Control Test	0.078	21:33	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOSS,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: CZARNY

DATE/TIME OF TEST: 12/10/2022 21:42:22

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:36
Subject Sample # 1	0.167	21:37
Breath Volume	1.457	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 2	0.156	21:40
Breath Volume	1.878	21:40
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: ALSTON

DATE/TIME OF TEST: 12/11/2022 05:17:01

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.081	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.087	05:12
Breath Volume	1.183	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.085	05:15
Breath Volume	1.503	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:17

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,JUAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: DEFRIASRODRIGUEZ

DATE/TIME OF TEST: 12/11/2022 07:27:38

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.209	07:23
Breath Volume	1.171	07:23
Air Blank	0.000	07:23
Air Blank	0.000	07:25
Subject Sample # 2	0.212	07:26
Breath Volume	1.457	07:26
Air Blank	0.000	07:26
Control Test	0.080	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:27

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MITCHELL,CEDRIC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: FELIXLEON

DATE/TIME OF TEST: 12/11/2022 08:15:32

Diagnostics Check	OK	08:08
Air Blank	0.000	08:09
Control Test	0.081	08:09
Air Blank	0.000	08:10
Subject Sample # 1	0.140	08:10
Breath Volume	1.148	08:10
Air Blank	0.000	08:11
Air Blank	0.000	08:13
Subject Sample # 2	0.152	08:13
Breath Volume	3.023	08:13
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:15
Diagnostics Check	OK	08:15

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PAGAN,KEVIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: FALANGAMARTINS

DATE/TIME OF TEST: 12/11/2022 09:07:55

Diagnostics Check	OK	09:01
Air Blank	0.000	09:01
Control Test	0.081	09:02
Air Blank	0.000	09:02
Subject Sample # 1	0.057	09:03
Breath Volume	1.539	09:03
Air Blank	0.000	09:03
Air Blank	0.000	09:05
Subject Sample # 2	0.062	09:06
Breath Volume	3.230	09:06
Air Blank	0.000	09:07
Control Test	0.081	09:07
Air Blank	0.000	09:07
Diagnostics Check	OK	09:07

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WISECUP,AUSTIN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: GALLANT

DATE/TIME OF TEST: 12/11/2022 21:40:46

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.007	21:36
Breath Volume	1.218	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.007	21:39
Breath Volume	2.183	21:39
Air Blank	0.000	21:39
Control Test	0.080	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOPKINS,JEFFREY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: FERRARA

DATE/TIME OF TEST: 12/13/2022 02:19:31

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.007	02:15
Breath Volume	1.660	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.007	02:18
Breath Volume	2.945	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,RICHARD
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 12/13/2022 03:13:06

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.082	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.206	03:08
Breath Volume	2.257	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.202	03:11
Breath Volume	2.062	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRISINTE,ANGELO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SIRJU

DATE/TIME OF TEST: 12/13/2022 03:46:56

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.216	03:42
Breath Volume	1.960	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:44
Subject Sample # 2	0.215	03:45
Breath Volume	1.804	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALLE,OMAR
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: VANDEVENTER

DATE/TIME OF TEST: 12/14/2022 22:46:45

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.142	22:42
Breath Volume	2.687	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.147	22:45
Breath Volume	3.972	22:45
Air Blank	0.000	22:45
Control Test	0.081	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 12/21/2022 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHAND,CHRIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 12/16/2022 03:00:42

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.167	02:56
Breath Volume	3.367	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.167	02:59
Breath Volume	3.613	02:59
Air Blank	0.000	02:59
Control Test	0.081	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUNT,FRANCIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: RUDMAN

DATE/TIME OF TEST: 12/16/2022 04:18:57

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.211	04:14
Breath Volume	3.250	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:16
Subject Sample # 2	0.210	04:17
Breath Volume	3.550	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STONE,NICOLAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 12/16/2022 06:25:23

Diagnostics Check	OK	06:19
Air Blank	0.000	06:19
Control Test	0.082	06:20
Air Blank	0.000	06:20
Subject Sample # 1	0.000	06:21
Breath Volume	1.878	06:21
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.000	06:24
Breath Volume	2.078	06:24
Air Blank	0.000	06:24
Control Test	0.081	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,JONATHAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: TORCATMARCANO

DATE/TIME OF TEST: 12/16/2022 07:43:08

Diagnostics Check	OK	07:36
Air Blank	0.000	07:37
Control Test	0.082	07:37
Air Blank	0.000	07:37
Subject Sample # 1	0.175	07:38
Breath Volume	2.406	07:38
Air Blank	0.000	07:39
Air Blank	0.000	07:40
Subject Sample # 2	0.173	07:41
Breath Volume	2.703	07:41
Air Blank	0.000	07:42
Control Test	0.081	07:42
Air Blank	0.000	07:43
Diagnostics Check	OK	07:43

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NDOUMBE, FREDERIC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: OHAREK

DATE/TIME OF TEST: 12/16/2022 18:23:41

Diagnostics Check	OK	18:17
Air Blank	0.000	18:18
Control Test	0.082	18:18
Air Blank	0.000	18:18
Subject Sample # 1	0.187	18:19
Breath Volume	3.179	18:19
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.182	18:22
Breath Volume	2.804	18:22
Air Blank	0.000	18:22
Control Test	0.081	18:23
Air Blank	0.000	18:23
Diagnostics Check	OK	18:23

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYER,ADAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: JIRON TORREZ

DATE/TIME OF TEST: 12/17/2022 01:30:27

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.082	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.133	01:25
Breath Volume	2.074	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.133	01:28
Breath Volume	1.757	01:28
Air Blank	0.000	01:29
Control Test	0.082	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORDON,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MCNULTY

DATE/TIME OF TEST: 12/17/2022 03:54:24

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.081	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.157	03:49
Breath Volume	1.769	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.158	03:52
Breath Volume	1.312	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 12/18/2022 06:18:18

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.082	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.149	06:13
Breath Volume	1.960	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.149	06:16
Breath Volume	2.019	06:16
Air Blank	0.000	06:17
Control Test	0.080	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 12/21/2022 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEANBAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: DIAZ QUINONES

DATE/TIME OF TEST: 12/18/2022 07:31:55

Diagnostics Check	OK	07:25
Air Blank	0.000	07:26
Control Test	0.081	07:26
Air Blank	0.000	07:27
Subject Sample # 1	0.093	07:27
Breath Volume	1.386	07:27
Air Blank	0.000	07:28
Air Blank	0.000	07:30
Subject Sample # 2	0.099	07:30
Breath Volume	2.421	07:30
Air Blank	0.000	07:31
Control Test	0.080	07:31
Air Blank	0.000	07:31
Diagnostics Check	OK	07:31

Data Download Date/Time: 12/21/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GIBSON,CASEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/18/2022 09:19:47

Diagnostics Check	OK	09:13
Air Blank	0.000	09:14
Control Test	0.082	09:14
Air Blank	0.000	09:14
Subject Sample # 1	0.008	09:15
Breath Volume	2.964	09:15
Air Blank	0.000	09:15
Air Blank	0.000	09:17
Subject Sample # 2	0.008	09:18
Breath Volume	3.218	09:18
Air Blank	0.000	09:18
Control Test	0.081	09:19
Air Blank	0.000	09:19
Diagnostics Check	OK	09:19

Data Download Date/Time: 12/21/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,JOSUE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CLICK

DATE/TIME OF TEST: 12/18/2022 20:31:18

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.082	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.315	20:26
Breath Volume	1.437	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.324	20:29
Breath Volume	1.562	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 12/21/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DESCHRYVER, KYLE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 12/19/2022 01:57:49

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.219	01:50
Breath Volume	1.679	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.240	01:53
Breath Volume	3.355	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 3	0.232	01:56
Breath Volume	3.128	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 12/21/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLACKWELL,ROBERT
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 12/21/2022 02:38:40

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.081	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.152	02:34
Breath Volume	3.238	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.157	02:37
Breath Volume	3.269	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 12/21/2022 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUCIANO MONTES,NICOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: SMELTZER

DATE/TIME OF TEST: 12/22/2022 22:36:39

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.082	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.148	22:32
Breath Volume	2.246	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.145	22:35
Breath Volume	2.546	22:35
Air Blank	0.000	22:35
Control Test	0.081	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 01/20/2023 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARSLUND,BRENT
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: BRAZZELL

DATE/TIME OF TEST: 12/23/2022 01:25:14

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:22	
Control Test	0.081	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:23	Subject Test Refused
Breath Volume	0.000	01:23	
Air Blank	0.000	01:24	
Control Test	0.080	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BAGBY

DATE/TIME OF TEST: 12/23/2022 03:57:03

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.081	03:52
Breath Volume	1.632	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.082	03:55
Breath Volume	1.425	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MOJZAK

DATE/TIME OF TEST: 12/23/2022 04:26:48

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.137	04:22
Breath Volume	1.132	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.131	04:25
Breath Volume	2.210	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: TRINIDAD SILVA

DATE/TIME OF TEST: 12/23/2022 21:21:35

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.082	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.046	21:17
Breath Volume	1.277	21:17
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.042	21:20
Breath Volume	1.160	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 12/24/2022 03:43:54

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.082	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.082	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 12/27/2022 23:56:44

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.083	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.046	23:52
Breath Volume	1.394	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.043	23:55
Breath Volume	1.191	23:55
Air Blank	0.000	23:55
Control Test	0.082	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MANGUM

DATE/TIME OF TEST: 12/28/2022 20:23:46

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.082	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.134	20:19
Breath Volume	1.562	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:21
Subject Sample # 2	0.130	20:22
Breath Volume	1.472	20:22
Air Blank	0.000	20:22
Control Test	0.081	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/29/2022 00:57:53

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.082	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.108	00:53
Breath Volume	1.574	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.111	00:56
Breath Volume	1.164	00:56
Air Blank	0.000	00:56
Control Test	0.080	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRESPO,BRYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 12/30/2022 01:40:45

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.194	01:36
Breath Volume	1.398	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.193	01:39
Breath Volume	1.101	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KOPINSKI,PATRYK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: LORUSSO

DATE/TIME OF TEST: 12/30/2022 22:27:55

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.082	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.091	22:23
Breath Volume	2.839	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.091	22:26
Breath Volume	2.628	22:26
Air Blank	0.000	22:27
Control Test	0.081	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ALEXIS

DATE/TIME OF TEST: 12/31/2022 02:46:46

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.082	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1 0.195	02:35	Volume Not Met
Breath Volume	0.828	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample	# 2 0.186	02:41	
Breath Volume	1.734	02:41	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample	# 3 0.176	02:45	
Breath Volume	1.199	02:45	
Air Blank	0.000	02:45	
Control Test	0.077	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WHITTED

DATE/TIME OF TEST: 12/31/2022 03:39:31

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.067	03:35
Breath Volume	1.906	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.063	03:38
Breath Volume	1.398	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 01/20/2023 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BOLDT,PAIGE
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 12/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/15/2022 16:42:16

Diagnostics Check	OK	16:36
Air Blank	0.000	16:36
Control Test	0.079	16:36
Air Blank	0.000	16:37
Subject Sample # 1	0.000	16:37
Breath Volume	1.949	16:37
Air Blank	0.000	16:38
Air Blank	0.000	16:40
Subject Sample # 2	0.000	16:40
Breath Volume	1.285	16:40
Air Blank	0.000	16:41
Control Test	0.080	16:41
Air Blank	0.000	16:42
Diagnostics Check	OK	16:42

Data Download Date/Time: 01/20/2023 04:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOREHEAD,AVERY,F
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: KRAMER,TREVOR
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 12/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: KEROUAC

DATE/TIME OF TEST: 12/18/2022 02:22:49

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.211	02:17
Breath Volume	1.968	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.202	02:21
Breath Volume	1.687	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 01/20/2023 04:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CACERES,JACOB,D
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: KRAMER,TREVOR
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 12/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 12/21/2022 22:59:02

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.079	22:47	
Air Blank	0.000	22:48	
Subject Sample	# 1 0.223	22:51	Volume Not Met
Breath Volume	0.089	22:51	
Air Blank	0.000	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 2 0.202	22:54	
Breath Volume	1.789	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 3 0.183	22:57	
Breath Volume	1.300	22:57	
Air Blank	0.000	22:58	
Control Test	0.077	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 01/20/2023 04:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIND,CHRISTIAN,G
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BUCKLEY,DREW
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 12/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 12/24/2022 01:28:50

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.081	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.080	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 01/20/2023 04:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDMAN,BAYLEE,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PEARSON,KEANDRE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GAGNON

DATE/TIME OF TEST: 12/02/2022 01:55:36

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:48	No Sample Provided
Breath Volume	0.000	01:48	
Air Blank	0.000	01:49	
Air Blank	0.000	01:51	
Subject Sample	# 2	01:54	No Sample Provided
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.081	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 12/29/2022 07:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,FELICIA,N
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MOFFATT,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: GILES JR

DATE/TIME OF TEST: 12/06/2022 16:41:49

Diagnostics Check	OK	16:29	
Air Blank	0.000	16:29	
Control Test	0.082	16:30	
Air Blank	0.000	16:30	
Subject Sample	# 1 0.024	16:33	Volume Not Met
Breath Volume	0.226	16:33	
Air Blank	0.000	16:34	
Air Blank	0.000	16:36	
Subject Sample	# 2 0.018	16:37	
Breath Volume	1.257	16:37	
Air Blank	0.000	16:37	
Air Blank	0.000	16:39	
Subject Sample	# 3 0.015	16:40	
Breath Volume	1.144	16:40	
Air Blank	0.000	16:40	
Control Test	0.079	16:41	
Air Blank	0.000	16:41	
Diagnostics Check	OK	16:41	

Data Download Date/Time: 12/29/2022 07:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDMAN,BAYLEE,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MOFFATT,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: LEAR

DATE/TIME OF TEST: 12/11/2022 05:18:08

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.081	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 1 0.143	05:16	Volume Not Met
Breath Volume	0.175	05:16	
Air Blank		05:17	Purge Fail
Air Blank		05:18	Purge Fail

Data Download Date/Time: 12/29/2022 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: WILSON,JORDAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/13/2022 02:54:02

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.081	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:47	No Sample Provided
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.082	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 12/29/2022 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DASQUE,D
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 12/01/2022 00:33:32

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.195	00:28
Breath Volume	3.054	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.189	00:31
Breath Volume	2.121	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: AURIENCE

DATE/TIME OF TEST: 12/03/2022 01:27:57

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.137	01:23
Breath Volume	2.769	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.132	01:26
Breath Volume	2.542	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KILLON,JONATHAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BAPTISTE

DATE/TIME OF TEST: 12/04/2022 02:52:40

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.079	02:48
Breath Volume	2.500	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.079	02:51
Breath Volume	2.691	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GALLAWAY

DATE/TIME OF TEST: 12/04/2022 03:47:07

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.079	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REDICK,D
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/05/2022 02:58:53

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.000	02:54
Breath Volume	1.898	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.000	02:57
Breath Volume	1.988	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: MANLEY

DATE/TIME OF TEST: 12/09/2022 00:10:24

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.212	00:04
Breath Volume	1.980	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.211	00:08
Breath Volume	2.093	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: SEIBEL

DATE/TIME OF TEST: 12/10/2022 17:09:20

Diagnostics Check	OK	17:02
Air Blank	0.000	17:03
Control Test	0.080	17:03
Air Blank	0.000	17:04
Subject Sample # 1	0.036	17:04
Breath Volume	2.144	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.036	17:07
Breath Volume	2.296	17:07
Air Blank	0.000	17:08
Control Test	0.079	17:08
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURKE,ROBERT
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GARCIA GABRIEL

DATE/TIME OF TEST: 12/11/2022 01:39:41

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.040	01:35
Breath Volume	2.398	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.038	01:38
Breath Volume	2.109	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOXLEY,MICHAEL
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 12/11/2022 05:03:56

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.090	04:59
Breath Volume	3.062	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.093	05:02
Breath Volume	2.929	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANROMPA,V,
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: DE LA ROSA,F
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 12/11/2022 05:54:06

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.080	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.324	05:46
Breath Volume	1.648	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.361	05:49
Breath Volume	1.199	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:51
Subject Sample # 3	0.379	05:52
Breath Volume	1.308	05:52
Air Blank	0.000	05:53
Control Test	0.078	05:53
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KOUROUMOUSIS,G
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 12/14/2022 00:02:09

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.250	23:57
Breath Volume	2.128	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.242	00:00
Breath Volume	3.691	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DASQUE,DARREN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: FERNADEZ

DATE/TIME OF TEST: 12/15/2022 21:48:11

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.079	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.196	21:41
Breath Volume	2.976	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.199	21:46
Breath Volume	2.843	21:46
Air Blank	0.000	21:47
Control Test	0.080	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FENTY,MICHAELA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ESTEVES GARCIA

DATE/TIME OF TEST: 12/16/2022 03:13:54

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.151	03:08
Breath Volume	1.156	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.145	03:12
Breath Volume	2.511	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SADOWSKI,ERICKA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MICHALOWSKI

DATE/TIME OF TEST: 12/17/2022 01:11:48

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.000	01:07
Breath Volume	1.738	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.000	01:10
Breath Volume	2.007	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIORIO,JAMES
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: HAYZILL

DATE/TIME OF TEST: 12/17/2022 01:38:17

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.180	01:33
Breath Volume	2.753	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.172	01:36
Breath Volume	2.832	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/17/2022 03:58:49

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.152	03:54
Breath Volume	2.726	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.146	03:57
Breath Volume	2.546	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,VICTORIA
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: GALINDO GALVEZ

DATE/TIME OF TEST: 12/17/2022 19:36:21

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:33	
Control Test	0.080	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 1	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:35	
Control Test	0.081	19:35	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,VICTORIA
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: GALINDO GALVEZ

DATE/TIME OF TEST: 12/17/2022 19:46:02

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.133	19:41
Breath Volume	1.464	19:41
Air Blank	0.000	19:41
Air Blank	0.000	19:43
Subject Sample # 2	0.141	19:44
Breath Volume	1.625	19:44
Air Blank	0.000	19:45
Control Test	0.079	19:45
Air Blank	0.000	19:45
Diagnostics Check	OK	19:46

Data Download Date/Time: 01/02/2023 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,SEAN,
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: HILL,SEAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 12/17/2022 20:24:31

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.152	20:20
Breath Volume	2.324	20:20
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.151	20:23
Breath Volume	2.875	20:23
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANCHEZ,J
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BINNEY

DATE/TIME OF TEST: 12/18/2022 00:23:55

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.058	00:18
Breath Volume	1.566	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.057	00:22
Breath Volume	1.542	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HARMAN,R
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CURTIN

DATE/TIME OF TEST: 12/18/2022 01:27:25

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.000	01:22
Breath Volume	1.750	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.000	01:26
Breath Volume	1.507	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIAL,V
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 12/18/2022 02:13:23

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.197	02:08
Breath Volume	4.363	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.185	02:11
Breath Volume	4.421	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZINA,MEGAN,N
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 12/18/2022 17:56:12

Diagnostics Check	OK	17:50	
Air Blank	0.000	17:50	
Control Test	0.079	17:50	
Air Blank	0.000	17:51	
Subject Sample	# 1 0.241	17:52	
Breath Volume	1.906	17:52	
Air Blank	0.000	17:52	
Air Blank	0.000	17:54	
Subject Sample	# 2	17:55	RFI Detect
Breath Volume	0.855	17:55	
Air Blank		17:55	RFI Detect
Air Blank	0.000	17:56	

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZINA,MEGAN,N
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLAZINA,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 12/18/2022 18:05:47

Diagnostics Check	OK	17:59
Air Blank	0.000	17:59
Control Test	0.079	18:00
Air Blank	0.000	18:00
Subject Sample # 1	0.244	18:01
Breath Volume	1.980	18:01
Air Blank	0.000	18:01
Air Blank	0.000	18:03
Subject Sample # 2	0.237	18:04
Breath Volume	1.898	18:04
Air Blank	0.000	18:04
Control Test	0.080	18:05
Air Blank	0.000	18:05
Diagnostics Check	OK	18:05

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,LOGAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: LEVERING

DATE/TIME OF TEST: 12/19/2022 02:05:29

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.162	02:00
Breath Volume	3.359	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample # 2	0.158	02:03
Breath Volume	3.144	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TEEHAN,M
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/19/2022 02:48:44

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.288	02:44
Breath Volume	3.402	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.298	02:47
Breath Volume	3.957	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SAAVEDRA,J
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WIMBLEY

DATE/TIME OF TEST: 12/19/2022 05:00:55

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.180	04:56
Breath Volume	1.812	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.179	04:59
Breath Volume	1.464	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODARD,DALLESSANDRO,D
BREATH TEST OPERATOR AGENCY: Winter Springs PD
ARRESTING OFFICER: HARTNER,L
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: CARABALLO

DATE/TIME OF TEST: 12/19/2022 08:28:53

Diagnostics Check	OK	08:16	
Air Blank	0.000	08:16	
Control Test	0.080	08:17	
Air Blank	0.000	08:17	
Subject Sample	# 1 0.045	08:20	Volume Not Met
Breath Volume	0.054	08:20	
Air Blank	0.000	08:21	
Air Blank	0.000	08:23	
Subject Sample	# 2 0.049	08:24	
Breath Volume	2.648	08:24	
Air Blank	0.000	08:24	
Air Blank	0.000	08:26	
Subject Sample	# 3 0.050	08:27	
Breath Volume	2.714	08:27	
Air Blank	0.000	08:27	
Control Test	0.080	08:28	
Air Blank	0.000	08:28	
Diagnostics Check	OK	08:28	

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GUILLEN-LARA,F
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: REAGAN

DATE/TIME OF TEST: 12/20/2022 02:36:31

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.201	02:31
Breath Volume	1.875	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.196	02:34
Breath Volume	1.703	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 01/02/2023 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GANT,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MURTHA

DATE/TIME OF TEST: 12/20/2022 03:25:24

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.072	03:20
Breath Volume	1.925	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.079	03:23
Breath Volume	1.960	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 01/03/2023 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SHERROD,B
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: DEITCH

DATE/TIME OF TEST: 12/20/2022 23:08:21

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.118	23:03
Breath Volume	2.757	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.117	23:06
Breath Volume	2.468	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 01/03/2023 14:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HIDEROA,K
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CLEVENGER

DATE/TIME OF TEST: 12/22/2022 04:18:28

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.079	04:04	
Air Blank	0.000	04:04	
Subject Sample	# 1 0.137	04:07	Volume Not Met
Breath Volume	0.269	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2 0.138	04:11	
Breath Volume	1.347	04:11	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 3 0.147	04:16	Volume Not Met
Breath Volume	0.292	04:16	
Air Blank	0.000	04:17	
Control Test	0.078	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 01/03/2023 14:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HIDEROA,K
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CLEVENGER

DATE/TIME OF TEST: 12/22/2022 04:32:11

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.145	04:25
Breath Volume	1.710	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.132	04:30
Breath Volume	2.078	04:30
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 01/03/2023 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HILL,SEAN,

BREATH TEST OPERATOR AGENCY: Sanford PD

ARRESTING OFFICER: HILL,SEAN

ARRESTING OFFICER AGENCY: Sanford PD

LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 19:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/22/2022 21:05:28

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.080	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1 0.043	20:58	Volume Not Met
Breath Volume	0.253	20:58	
Air Blank	0.000	20:58	
Air Blank	0.000	21:00	
Subject Sample	# 2 0.042	21:03	Volume Not Met
Breath Volume	0.015	21:03	
Air Blank	0.000	21:04	
Control Test	0.079	21:04	
Air Blank	0.000	21:05	
Diagnostics Check	OK	21:05	

Data Download Date/Time: 01/03/2023 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALLEY,BROOKE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 12/23/2022 04:00:01

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1 0.132	03:52	Volume Not Met
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2 0.157	03:58	Volume Not Met
Breath Volume	0.019	03:58	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 01/03/2023 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOWARD,ERIC
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: BROYLES

DATE/TIME OF TEST: 12/23/2022 21:45:55

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.184	21:41
Breath Volume	3.507	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.188	21:44
Breath Volume	3.765	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 01/03/2023 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLUMSTEIN,REBEKAH
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MARTIN III

DATE/TIME OF TEST: 12/24/2022 01:54:13

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.140	01:49
Breath Volume	3.335	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.134	01:52
Breath Volume	1.648	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 01/03/2023 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURKS,RICKY
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 12/24/2022 03:58:24

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.210	03:53
Breath Volume	2.539	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.200	03:56
Breath Volume	2.472	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/03/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: KRECEK

DATE/TIME OF TEST: 12/24/2022 04:16:49

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.125	04:12
Breath Volume	2.546	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.117	04:15
Breath Volume	1.398	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 01/03/2023 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FERRARA

DATE/TIME OF TEST: 12/24/2022 23:28:38

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.016	23:23
Breath Volume	2.824	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.015	23:27
Breath Volume	3.464	23:27
Air Blank	0.000	23:27
Control Test	0.081	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CHAMBERS,AARON
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 12/25/2022 05:10:34

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.193	05:05
Breath Volume	1.371	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.195	05:09
Breath Volume	1.289	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,LOGAN
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: VETTER

DATE/TIME OF TEST: 12/27/2022 02:37:51

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.204	02:32
Breath Volume	1.726	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.201	02:36
Breath Volume	1.300	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 12/27/2022 03:27:28

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.118	03:22
Breath Volume	1.574	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.110	03:25
Breath Volume	2.140	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REDICK,D
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 12/28/2022 06:03:40

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.081	05:56
Air Blank	0.000	05:57
Subject Sample # 1	0.000	05:59
Breath Volume	2.367	05:59
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.000	06:02
Breath Volume	1.187	06:02
Air Blank	0.000	06:02
Control Test	0.080	06:03
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GARCIA-CHOW,L
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: FONTANE

DATE/TIME OF TEST: 12/28/2022 23:02:21

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.081	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.172	22:57
Breath Volume	1.601	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.170	23:00
Breath Volume	1.566	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHINEA,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SIMPSON,J
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: TOATES

DATE/TIME OF TEST: 12/29/2022 02:31:14

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.183	02:26
Breath Volume	2.132	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.186	02:29
Breath Volume	2.390	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: VAZ

DATE/TIME OF TEST: 12/30/2022 04:22:57

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.079	04:14	
Air Blank	0.000	04:15	
Subject Sample	# 1 0.183	04:18	Volume Not Met
Breath Volume	0.085	04:18	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.078	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ESCOBAR ORELLANA

DATE/TIME OF TEST: 12/31/2022 03:44:32

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1 0.101	03:37	Volume Not Met
Breath Volume	0.136	03:37	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2 0.105	03:42	Volume Not Met
Breath Volume	0.808	03:42	
Air Blank	0.000	03:43	
Control Test	0.078	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: EXCOBAR ORELLANA

DATE/TIME OF TEST: 12/31/2022 03:54:12

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.097	03:48
Breath Volume	1.250	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.090	03:52
Breath Volume	1.296	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 01/03/2023 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/03/2022 01:28:13

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.184	01:22
Breath Volume	1.199	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.166	01:26
Breath Volume	1.250	01:26
Air Blank	0.000	01:27
Control Test	0.078	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/08/2022 21:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,B,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: LACEY

DATE/TIME OF TEST: 12/15/2022 02:14:12

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.099	02:08
Breath Volume	1.496	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.097	02:12
Breath Volume	1.410	02:12
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 01/27/2023 20:09