

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 12/10/2022 06:36:06

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.079	06:29
Air Blank	0.000	06:30
Subject Sample # 1	0.156	06:30
Breath Volume	3.523	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 2	0.154	06:34
Breath Volume	3.625	06:34
Air Blank	0.000	06:35
Control Test	0.079	06:35
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 12/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HODGDON,TODD
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: DEBLASI

DATE/TIME OF TEST: 12/11/2022 16:51:22

Diagnostics Check	OK	16:41
Air Blank	0.000	16:41
Control Test	0.078	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.222	16:44
Breath Volume	1.714	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.215	16:49
Breath Volume	1.812	16:49
Air Blank	0.000	16:50
Control Test	0.078	16:50
Air Blank	0.000	16:51
Diagnostics Check	OK	16:51

Data Download Date/Time: 12/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: STEPP,V

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 01:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: FORMAN

DATE/TIME OF TEST: 12/15/2022 01:59:28

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.162	01:55
Breath Volume	3.265	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.159	01:57
Breath Volume	1.421	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 12/28/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: ROMANDETTI,M

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: TURNBULL

DATE/TIME OF TEST: 12/23/2022 20:40:04

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.079	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.079	20:35
Breath Volume	1.902	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.081	20:38
Breath Volume	2.722	20:38
Air Blank	0.000	20:39
Control Test	0.079	20:39
Air Blank	0.000	20:39
Diagnostics Check	OK	20:40

Data Download Date/Time: 12/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: JAMES,JOLENE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: PIETRO

DATE/TIME OF TEST: 12/24/2022 18:50:11

Diagnostics Check	OK	18:44
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Subject Sample # 1	0.000	18:45
Breath Volume	2.570	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.000	18:48
Breath Volume	2.523	18:48
Air Blank	0.000	18:49
Control Test	0.080	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 12/28/2022 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MOORE,GREG,C

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MEEHAN,ROBERT

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 13:24

GENDER: Female

OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: JAMISON

DATE/TIME OF TEST: 12/29/2022 14:32:51

Diagnostics Check	OK	14:26	
Air Blank	0.000	14:26	
Control Test	0.080	14:27	
Air Blank	0.000	14:27	
Subject Sample	# 1 0.000	14:28	
Breath Volume	1.316	14:28	
Air Blank	0.000	14:28	
Air Blank	0.000	14:30	
Subject Sample	# 2 0.000	14:31	
Breath Volume	2.097	14:31	
Air Blank	0.000	14:31	
Control Test	0.079	14:32	
Air Blank		14:32	RFI Detect
Air Blank	0.000	14:32	

Data Download Date/Time: 01/27/2023 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,K,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TISDALE,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 12/03/2022 08:46:07

Diagnostics Check	OK	08:39
Air Blank	0.000	08:40
Control Test	0.081	08:40
Air Blank	0.000	08:40
Subject Sample # 1	0.104	08:41
Breath Volume	2.316	08:41
Air Blank	0.000	08:42
Air Blank	0.000	08:43
Subject Sample # 2	0.104	08:44
Breath Volume	2.410	08:44
Air Blank	0.000	08:45
Control Test	0.081	08:45
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHIERA,FERNANDO,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CORNIER,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 12/03/2022 14:14:48

Diagnostics Check	OK	14:08
Air Blank	0.000	14:08
Control Test	0.082	14:09
Air Blank	0.000	14:09
Subject Sample # 1	0.078	14:10
Breath Volume	3.066	14:10
Air Blank	0.000	14:10
Air Blank	0.000	14:12
Subject Sample # 2	0.075	14:13
Breath Volume	3.476	14:13
Air Blank	0.000	14:13
Control Test	0.080	14:14
Air Blank	0.000	14:14
Diagnostics Check	OK	14:14

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS,JONATHAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: DIXON KINSER

DATE/TIME OF TEST: 12/03/2022 16:56:05

Diagnostics Check	OK	16:49
Air Blank	0.000	16:49
Control Test	0.081	16:50
Air Blank	0.000	16:50
Subject Sample # 1	0.161	16:51
Breath Volume	2.261	16:51
Air Blank	0.000	16:52
Air Blank	0.000	16:53
Subject Sample # 2	0.167	16:54
Breath Volume	1.578	16:54
Air Blank	0.000	16:55
Control Test	0.080	16:55
Air Blank	0.000	16:55
Diagnostics Check	OK	16:56

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAK,KYLE,F
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOYOS,JONATHAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:19

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 12/04/2022 14:47:19

Diagnostics Check	OK	14:40
Air Blank	0.000	14:40
Control Test	0.080	14:40
Air Blank	0.000	14:41
Subject Sample # 1	0.300	14:41
Breath Volume	1.113	14:41
Air Blank	0.000	14:42
Air Blank	0.000	14:44
Subject Sample # 2	0.299	14:45
Breath Volume	1.472	14:45
Air Blank	0.000	14:46
Control Test	0.080	14:46
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GUADGRRAMA CASTANEDA

DATE/TIME OF TEST: 12/04/2022 21:43:35

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.101	21:39
Breath Volume	2.089	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.097	21:41
Breath Volume	1.695	21:41
Air Blank	0.000	21:42
Control Test	0.079	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 12/04/2022 23:16:48

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.052	23:12
Breath Volume	1.968	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.050	23:15
Breath Volume	1.789	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 12/06/2022 19:35:27

Diagnostics Check	OK	19:28
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.285	19:30
Breath Volume	3.050	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:32
Subject Sample # 2	0.281	19:33
Breath Volume	3.238	19:33
Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LUMLEY

DATE/TIME OF TEST: 12/06/2022 22:58:55

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.167	22:54
Breath Volume	1.445	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:56
Subject Sample # 2	0.174	22:57
Breath Volume	1.386	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 12/07/2022 02:25:53

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.173	02:20
Breath Volume	1.304	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.178	02:24
Breath Volume	1.582	02:24
Air Blank	0.000	02:24
Control Test	0.078	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, DAVID, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE, DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 12/07/2022 05:36:34

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.096	05:31
Breath Volume	2.894	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.100	05:34
Breath Volume	2.078	05:34
Air Blank	0.000	05:35
Control Test	0.080	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: PORRO RODRIGUEZ

DATE/TIME OF TEST: 12/08/2022 00:11:55

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.203	00:07
Breath Volume	2.261	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:09
Subject Sample # 2	0.193	00:10
Breath Volume	2.527	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: NORTH

DATE/TIME OF TEST: 12/11/2022 01:34:44

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.156	01:30
Breath Volume	1.828	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.143	01:33
Breath Volume	1.441	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARMSTRONG,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/11/2022 05:54:51

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:52	
Control Test	0.080	05:52	
Air Blank	0.000	05:53	
Subject Sample	# 1	05:53	Subject Test Refused
Breath Volume	0.000	05:53	
Air Blank	0.000	05:53	
Control Test	0.080	05:54	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/11/2022 21:07:50

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:56	
Control Test	0.080	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1 0.146	20:57	
Breath Volume	2.703	20:57	
Air Blank	0.000	20:58	
Air Blank	0.000	20:59	
Subject Sample	# 2	21:03	No Sample Provided
Breath Volume	0.000	21:03	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample	# 3 0.131	21:06	
Breath Volume	1.777	21:06	
Air Blank	0.000	21:06	
Control Test	0.079	21:07	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GARCIA MENDEZ

DATE/TIME OF TEST: 12/12/2022 01:15:21

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.011	01:10
Breath Volume	1.457	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.008	01:13
Breath Volume	3.312	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 12/12/2022 02:09:39

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.067	02:05
Breath Volume	3.042	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.064	02:08
Breath Volume	1.851	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 12/13/2022 09:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRAGO,JAMES,A
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: SERRAGO,JAMES
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: SMETANA

DATE/TIME OF TEST: 12/13/2022 23:04:20

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.055	22:59
Breath Volume	2.609	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.050	23:02
Breath Volume	1.371	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: D240513995490

DATE/TIME OF TEST: 12/14/2022 04:17:32

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.140	04:12
Breath Volume	1.207	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.145	04:15
Breath Volume	2.140	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACH,B,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MARTINEZ,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: BRITTEN

DATE/TIME OF TEST: 12/14/2022 13:34:13

Diagnostics Check	OK	13:27
Air Blank	0.000	13:28
Control Test	0.080	13:28
Air Blank	0.000	13:29
Subject Sample # 1	0.225	13:29
Breath Volume	3.046	13:29
Air Blank	0.000	13:30
Air Blank	0.000	13:32
Subject Sample # 2	0.219	13:32
Breath Volume	2.542	13:32
Air Blank	0.000	13:33
Control Test	0.077	13:33
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/15/2022 03:11:27

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.000	03:07
Breath Volume	2.121	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:10
Breath Volume	1.687	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MARTINEAU

DATE/TIME OF TEST: 12/16/2022 00:48:24

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.146	00:43
Breath Volume	1.511	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.144	00:46
Breath Volume	1.929	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: DIAZ CHAVEZ

DATE/TIME OF TEST: 12/17/2022 00:15:32

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.179	00:11
Breath Volume	2.855	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.177	00:13
Breath Volume	2.757	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/17/2022 02:53:20

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.126	02:48
Breath Volume	1.859	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.122	02:51
Breath Volume	1.855	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 12/17/2022 03:12:14

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.181	03:07
Breath Volume	3.007	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.178	03:10
Breath Volume	2.648	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO, JUAN, M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH, TYLER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PEDRO FRANCISCO

DATE/TIME OF TEST: 12/18/2022 01:46:59

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.216	01:42
Breath Volume	2.832	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.231	01:45
Breath Volume	3.312	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VACHA

DATE/TIME OF TEST: 12/18/2022 01:56:01

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.250	01:51
Breath Volume	1.410	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.251	01:54
Breath Volume	1.617	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:56

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHIERA,FERNANDO,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TORRES,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: GOMEZ LOPEZ

DATE/TIME OF TEST: 12/18/2022 18:01:08

Diagnostics Check	OK	17:54
Air Blank	0.000	17:54
Control Test	0.080	17:54
Air Blank	0.000	17:55
Subject Sample # 1	0.148	17:56
Breath Volume	2.332	17:56
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample # 2	0.151	17:59
Breath Volume	2.082	17:59
Air Blank	0.000	18:00
Control Test	0.080	18:00
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,P,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ARMSTRONG,P
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 12/21/2022 00:28:28

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.105	00:23
Breath Volume	1.605	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.107	00:26
Breath Volume	2.003	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/21/2022 03:49:25

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.150	03:44
Breath Volume	1.750	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.155	03:47
Breath Volume	1.414	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: KAMINSKI

DATE/TIME OF TEST: 12/22/2022 21:16:23

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.081	21:14	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.081	21:15	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: LUIS ALONZO

DATE/TIME OF TEST: 12/24/2022 03:58:24

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.208	03:53
Breath Volume	2.785	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.204	03:56
Breath Volume	2.214	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: ZAWACKI

DATE/TIME OF TEST: 12/24/2022 21:12:00

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.081	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.128	21:07
Breath Volume	1.851	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:09
Subject Sample # 2	0.124	21:10
Breath Volume	1.472	21:10
Air Blank	0.000	21:11
Control Test	0.080	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:12

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JAYDA,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JACKSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: KOHNKE

DATE/TIME OF TEST: 12/29/2022 18:58:32

Diagnostics Check	OK	18:48
Air Blank	0.000	18:49
Control Test	0.080	18:49
Air Blank	0.000	18:50
Subject Sample # 1	0.278	18:50
Breath Volume	3.175	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.245	18:53
Breath Volume	2.328	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 3	0.272	18:56
Breath Volume	2.035	18:56
Air Blank	0.000	18:57
Control Test	0.075	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROSPER,BENSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CORNIER,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: MCVEY

DATE/TIME OF TEST: 12/30/2022 10:06:48

Diagnostics Check	OK	10:03	
Air Blank	0.000	10:04	
Control Test	0.080	10:04	
Air Blank	0.000	10:04	
Subject Sample	# 1	10:05	Subject Test Refused
Breath Volume	0.000	10:05	
Air Blank	0.000	10:05	
Control Test	0.081	10:06	
Air Blank	0.000	10:06	
Diagnostics Check	OK	10:06	

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAY,JOSEPH,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WAY,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A3
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/30/2022 18:43:45

Diagnostics Check	OK	18:37
Air Blank	0.000	18:37
Control Test	0.079	18:38
Air Blank	0.000	18:38
Subject Sample # 1	0.000	18:39
Breath Volume	4.855	18:39
Air Blank	0.000	18:39
Air Blank	0.000	18:41
Subject Sample # 2	0.000	18:42
Breath Volume	5.148	18:42
Air Blank	0.000	18:42
Control Test	0.080	18:43
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 01/12/2023 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/08/2022 02:57:11

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.077	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.125	02:52
Breath Volume	1.261	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.125	02:55
Breath Volume	1.457	02:55
Air Blank	0.000	02:56
Control Test	0.077	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 01/03/2023 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,J
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: JUNOT

DATE/TIME OF TEST: 12/09/2022 01:22:53

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.078	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Control Test	0.077	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 01/03/2023 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PAVA

DATE/TIME OF TEST: 12/09/2022 02:33:19

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.078	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.077	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 01/03/2023 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,DETRANO
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: WAINSCOTT

DATE/TIME OF TEST: 12/10/2022 19:53:56

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.077	19:48
Air Blank	0.000	19:48
Subject Sample # 1	0.298	19:49
Breath Volume	2.003	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:51
Subject Sample # 2	0.305	19:52
Breath Volume	1.773	19:52
Air Blank	0.000	19:52
Control Test	0.077	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 01/03/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: WEIXEL

DATE/TIME OF TEST: 12/10/2022 21:01:02

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.236	20:56
Breath Volume	1.742	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:58
Subject Sample # 2	0.233	20:59
Breath Volume	1.667	20:59
Air Blank	0.000	21:00
Control Test	0.076	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:01

Data Download Date/Time: 01/03/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A2
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PIOTROWSKI

DATE/TIME OF TEST: 12/29/2022 04:13:52

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.117	04:09
Breath Volume	1.953	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.122	04:12
Breath Volume	2.042	04:12
Air Blank	0.000	04:12
Control Test	0.077	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 01/03/2023 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COURTHOUSE,CHARLOTTE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 12/01/2022 15:00:03

Diagnostics Check	OK	14:53
Air Blank	0.000	14:54
Control Test	0.079	14:54
Air Blank	0.000	14:54
Subject Sample # 1	0.121	14:55
Breath Volume	3.097	14:55
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.120	14:58
Breath Volume	2.988	14:58
Air Blank	0.000	14:59
Control Test	0.079	14:59
Air Blank	0.000	14:59
Diagnostics Check	OK	15:00

Data Download Date/Time: 12/29/2022 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,SEBASTIAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: TOLENTINO CASTILLO

DATE/TIME OF TEST: 12/02/2022 01:18:11

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.116	01:13
Breath Volume	3.148	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.103	01:16
Breath Volume	3.535	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 12/29/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DUSTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PITMAN

DATE/TIME OF TEST: 12/02/2022 02:25:12

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.150	02:20
Breath Volume	3.652	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.151	02:23
Breath Volume	3.843	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 12/29/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GUEVARA,W
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MATEO-BARAHONA

DATE/TIME OF TEST: 12/03/2022 01:53:48

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.129	01:49
Breath Volume	2.609	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.125	01:52
Breath Volume	2.019	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 12/29/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CONANT,CORPORAL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 12/03/2022 15:52:38

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.080	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.000	15:48
Breath Volume	2.011	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.000	15:51
Breath Volume	2.425	15:51
Air Blank	0.000	15:51
Control Test	0.079	15:52
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 12/29/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRITES,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,JOSEPH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 12/06/2022 04:04:08

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.048	03:59
Breath Volume	2.804	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.047	04:02
Breath Volume	2.253	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 12/29/2022 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CATES

DATE/TIME OF TEST: 12/07/2022 00:35:14

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.117	00:30
Breath Volume	3.316	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.112	00:33
Breath Volume	3.195	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: ZUKOWSKI

DATE/TIME OF TEST: 12/07/2022 15:53:06

Diagnostics Check	OK	15:46
Air Blank	0.000	15:47
Control Test	0.079	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.000	15:48
Breath Volume	1.750	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.000	15:51
Breath Volume	1.968	15:51
Air Blank	0.000	15:52
Control Test	0.080	15:52
Air Blank	0.000	15:53
Diagnostics Check	OK	15:53

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: CRONKHITE

DATE/TIME OF TEST: 12/10/2022 00:33:14

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.079	00:19	
Air Blank	0.000	00:19	
Subject Sample	# 1 0.051	00:22	Volume Not Met
Breath Volume	0.058	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 2 0.065	00:26	
Breath Volume	1.832	00:26	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample	# 3 0.070	00:31	
Breath Volume	2.042	00:31	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KIMBERLIN,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/10/2022 02:10:24

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.161	02:05
Breath Volume	1.542	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.158	02:08
Breath Volume	3.054	02:08
Air Blank	0.000	02:09
Control Test	0.078	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: OWENS,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: TREAT

DATE/TIME OF TEST: 12/10/2022 23:39:29

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.215	23:34
Breath Volume	4.144	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.207	23:38
Breath Volume	3.671	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SIDES,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/11/2022 02:37:00

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.000	02:32
Breath Volume	3.351	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.000	02:35
Breath Volume	3.511	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REILLY,S.,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REILLY,S.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: PEREZ SUAREZ

DATE/TIME OF TEST: 12/11/2022 19:44:39

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.080	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.103	19:40
Breath Volume	2.320	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.100	19:43
Breath Volume	4.101	19:43
Air Blank	0.000	19:43
Control Test	0.079	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WALDROP,P
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 12/11/2022 23:22:49

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.000	23:18
Breath Volume	2.574	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.000	23:21
Breath Volume	2.632	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 12/29/2022 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WELCH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: FINEMAN

DATE/TIME OF TEST: 12/12/2022 17:59:11

Diagnostics Check	OK	17:52
Air Blank	0.000	17:53
Control Test	0.079	17:53
Air Blank	0.000	17:54
Subject Sample # 1	0.262	17:54
Breath Volume	1.851	17:54
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.255	17:57
Breath Volume	2.031	17:57
Air Blank	0.000	17:58
Control Test	0.076	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EGAN,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: FINEMAN

DATE/TIME OF TEST: 12/12/2022 23:52:34

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.162	23:47
Breath Volume	2.230	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.165	23:51
Breath Volume	2.257	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSZE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BOSZE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:34

SUBJECT'S LAST NAME: FINEMAN

DATE/TIME OF TEST: 12/13/2022 10:02:37

Diagnostics Check	OK	09:56
Air Blank	0.000	09:57
Control Test	0.080	09:57
Air Blank	0.000	09:58
Subject Sample # 1	0.000	09:58
Breath Volume	2.117	09:58
Air Blank	0.000	09:58
Air Blank	0.000	10:00
Subject Sample # 2	0.000	10:01
Breath Volume	2.125	10:01
Air Blank	0.000	10:01
Control Test	0.080	10:02
Air Blank	0.000	10:02
Diagnostics Check	OK	10:02

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 12/13/2022 20:54:11

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.080	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.158	20:49
Breath Volume	2.695	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.152	20:52
Breath Volume	2.207	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SALISBURY,N
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WATTLES

DATE/TIME OF TEST: 12/15/2022 00:01:42

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:54
Air Blank	0.000	23:54
Subject Sample # 2	0.209	23:57
Breath Volume	2.765	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.205	00:00
Breath Volume	2.769	00:00
Air Blank	0.000	00:00
Control Test	0.078	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/15/2022 03:05:28

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.176	03:00
Breath Volume	1.535	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.163	03:03
Breath Volume	1.816	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SWCHARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 12/15/2022 23:33:56

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.209	23:29
Breath Volume	1.921	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.211	23:32
Breath Volume	1.910	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DZUGAN,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DZUGAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: HOEL

DATE/TIME OF TEST: 12/17/2022 01:49:33

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.096	01:45
Breath Volume	2.757	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.095	01:48
Breath Volume	3.031	01:48
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,KANADA,E
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 12/17/2022 20:14:27

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.173	20:09
Breath Volume	3.863	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.176	20:12
Breath Volume	3.359	20:12
Air Blank	0.000	20:13
Control Test	0.079	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 12/29/2022 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,MARDEE,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MANO,MARDEE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 12/18/2022 15:15:32

Diagnostics Check	OK	15:09
Air Blank	0.000	15:09
Control Test	0.080	15:10
Air Blank	0.000	15:10
Subject Sample # 1	0.000	15:11
Breath Volume	3.550	15:11
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.000	15:14
Breath Volume	2.929	15:14
Air Blank	0.000	15:14
Control Test	0.080	15:14
Air Blank	0.000	15:15
Diagnostics Check	OK	15:15

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MATZICOFF,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/19/2022 03:01:17

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.092	02:56
Breath Volume	3.636	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.094	02:59
Breath Volume	4.281	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORITZ,JONATHAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 12/20/2022 00:28:16

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.000	00:23
Breath Volume	1.386	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.000	00:26
Breath Volume	2.117	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,JOSEPH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 12/20/2022 23:29:31

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.000	23:24
Breath Volume	3.011	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.000	23:28
Breath Volume	2.921	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MATZICOFF,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: BITTNER

DATE/TIME OF TEST: 12/21/2022 20:23:01

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.201	20:18
Breath Volume	2.109	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:20
Subject Sample # 2	0.186	20:21
Breath Volume	1.621	20:21
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:23

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PATTERSON,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: NAVE

DATE/TIME OF TEST: 12/23/2022 01:08:00

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.212	01:03
Breath Volume	2.250	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.205	01:06
Breath Volume	1.421	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VETRI,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HARTSOCK

DATE/TIME OF TEST: 12/23/2022 22:33:40

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.079	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.216	22:28
Breath Volume	2.214	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.213	22:32
Breath Volume	1.742	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SIDES,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MARYON

DATE/TIME OF TEST: 12/24/2022 02:28:54

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.159	02:22
Breath Volume	1.164	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.153	02:27
Breath Volume	1.234	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 12/29/2022 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SIDES,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 12/25/2022 00:00:18

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.119	23:55
Breath Volume	2.925	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.119	23:58
Breath Volume	3.074	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 12/29/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RISI,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,JONATHAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: YUNKER

DATE/TIME OF TEST: 12/28/2022 21:30:38

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.022	21:26
Breath Volume	1.433	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.021	21:29
Breath Volume	1.316	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 12/29/2022 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ORDAZ-GARCIA

DATE/TIME OF TEST: 12/30/2022 00:38:32

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1 0.189	00:30	Volume Not Met
Breath Volume	0.015	00:30	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 2 0.190	00:33	
Breath Volume	2.031	00:33	
Air Blank	0.000	00:34	
Air Blank	0.000	00:36	
Subject Sample	# 3 0.198	00:37	
Breath Volume	2.750	00:37	
Air Blank	0.000	00:37	
Control Test	0.078	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 01/30/2023 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JANKOWSKI,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 12/30/2022 04:59:30

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.122	04:54
Breath Volume	2.640	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.121	04:58
Breath Volume	3.578	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 01/30/2023 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,JAMIE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BRETT

DATE/TIME OF TEST: 12/31/2022 01:33:36

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.203	01:28
Breath Volume	2.914	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.203	01:32
Breath Volume	2.863	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 01/30/2023 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRIBBEN,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/31/2022 03:36:39

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.222	03:31
Breath Volume	1.914	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.237	03:35
Breath Volume	1.652	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/30/2023 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/31/2022 23:19:34

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.156	23:14
Breath Volume	1.812	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.152	23:18
Breath Volume	1.312	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 01/30/2023 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 12/02/2022 03:27:16

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.107	03:21
Breath Volume	1.441	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.104	03:25
Breath Volume	2.191	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 12/19/2022 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FALLUCCA,MARIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/03/2022 18:07:24

Diagnostics Check	OK	18:00
Air Blank	0.000	18:01
Control Test	0.080	18:01
Air Blank	0.000	18:02
Subject Sample # 1	0.147	18:02
Breath Volume	3.335	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.144	18:05
Breath Volume	3.312	18:05
Air Blank	0.000	18:06
Control Test	0.080	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time: 12/19/2022 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: DECK II

DATE/TIME OF TEST: 12/03/2022 22:58:26

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:53
Breath Volume	3.429	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:57
Breath Volume	3.589	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 12/19/2022 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: TORRES HERNANDEZ

DATE/TIME OF TEST: 12/04/2022 18:54:18

Diagnostics Check	OK	18:42	
Air Blank	0.000	18:43	
Control Test	0.080	18:43	
Air Blank	0.000	18:44	
Subject Sample	# 1	18:47	No Sample Provided
Breath Volume	0.000	18:47	
Air Blank	0.000	18:47	
Air Blank	0.000	18:49	
Subject Sample	# 2	18:52	No Sample Provided
Breath Volume	0.000	18:52	
Air Blank	0.000	18:53	
Control Test	0.080	18:53	
Air Blank	0.000	18:54	
Diagnostics Check	OK	18:54	

Data Download Date/Time: 12/19/2022 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: TORRES HERNANDEZ

DATE/TIME OF TEST: 12/04/2022 19:17:00

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:05	
Control Test	0.080	19:06	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:09	No Sample Provided
Breath Volume	0.000	19:09	
Air Blank	0.000	19:10	
Air Blank	0.000	19:12	
Subject Sample	# 2	19:15	Volume Not Met
Breath Volume	0.601	19:15	
Air Blank	0.000	19:16	
Control Test	0.080	19:16	
Air Blank	0.000	19:16	
Diagnostics Check	OK	19:17	

Data Download Date/Time: 12/19/2022 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/07/2022 00:45:29

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.185	00:39
Breath Volume	1.449	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.195	00:43
Breath Volume	1.750	00:43
Air Blank	0.000	00:44
Control Test	0.078	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: NERO

DATE/TIME OF TEST: 12/10/2022 00:52:30

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
		00:52	RFI Detect
Air Blank	0.000	00:52	

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: NERO

DATE/TIME OF TEST: 12/10/2022 01:05:06

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CARRANZA IBARRA

DATE/TIME OF TEST: 12/10/2022 23:09:05

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.079	23:08	RFI Detect
Air Blank	0.000	23:09	

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CARRANZA IBARRA

DATE/TIME OF TEST: 12/10/2022 23:19:47

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.328	23:12
Breath Volume	3.156	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.305	23:15
Breath Volume	3.476	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 3	0.292	23:18
Breath Volume	3.218	23:18
Air Blank	0.000	23:18
Control Test	0.078	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: GRANDADOS GONZALEZ

DATE/TIME OF TEST: 12/12/2022 22:03:29

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.106	21:58
Breath Volume	3.972	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.103	22:01
Breath Volume	3.054	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: BRINKER

DATE/TIME OF TEST: 12/13/2022 21:04:55

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.080	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.000	21:00
Breath Volume	2.992	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:03
Breath Volume	3.722	21:03
Air Blank	0.000	21:03
Control Test	0.080	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DESERIO

DATE/TIME OF TEST: 12/13/2022 23:44:21

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.104	23:39
Breath Volume	1.628	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.097	23:42
Breath Volume	2.054	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LOPEZ GOMEZ

DATE/TIME OF TEST: 12/14/2022 21:30:40

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:22	
Subject Sample	# 1 0.197	21:25	Volume Not Met
Breath Volume	0.035	21:25	
Air Blank	0.000	21:26	
Air Blank	0.000	21:27	
Subject Sample	# 2	21:28	Subject Test Refused
Breath Volume	0.070	21:28	
Air Blank	0.000	21:29	
Control Test	0.076	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 12/16/2022 21:04:00

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.141	20:59
Breath Volume	2.953	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:01
Subject Sample # 2	0.124	21:02
Breath Volume	1.726	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:04

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSAKOWSKI,PETER,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KOSAKOWSKI,PETER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: DMUCHOWSKI

DATE/TIME OF TEST: 12/17/2022 00:36:34

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.081	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CARRANZA DOMINGUEZ

DATE/TIME OF TEST: 12/17/2022 01:06:25

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.080	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1 0.201	00:57	Volume Not Met
Breath Volume	0.035	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	00:59	
Subject Sample	# 2 0.218	01:00	
Breath Volume	2.554	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 3 0.202	01:04	
Breath Volume	1.234	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 12/19/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: VELASQUEZ VELASQUEZ

DATE/TIME OF TEST: 12/01/2022 01:31:32

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.164	01:27
Breath Volume	2.828	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.159	01:30
Breath Volume	2.644	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 12/12/2022 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBER,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/01/2022 04:55:15

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.089	04:50
Breath Volume	1.519	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.086	04:53
Breath Volume	1.996	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 12/02/2022 00:48:32

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.079	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.093	00:43
Breath Volume	1.554	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.093	00:47
Breath Volume	1.871	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PUTINTSEV

DATE/TIME OF TEST: 12/02/2022 03:48:14

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.079	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAYLOR,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/02/2022 06:15:37

Diagnostics Check	OK	06:09
Air Blank	0.000	06:09
Control Test	0.079	06:09
Air Blank	0.000	06:10
Subject Sample # 1	0.232	06:11
Breath Volume	2.046	06:11
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.234	06:14
Breath Volume	1.566	06:14
Air Blank	0.000	06:14
Control Test	0.079	06:15
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FELIZ,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: WILHELM

DATE/TIME OF TEST: 12/02/2022 23:02:45

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.079	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.079	23:02	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: DOSS

DATE/TIME OF TEST: 12/03/2022 04:16:20

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.124	04:11
Breath Volume	2.398	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.121	04:14
Breath Volume	2.613	04:14
Air Blank	0.000	04:15
Control Test	0.079	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: YOUNG,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SIBOUNHEUANG

DATE/TIME OF TEST: 12/03/2022 22:59:48

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.249	22:53
Breath Volume	1.824	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.233	22:58
Breath Volume	1.351	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SMALLWOOD

DATE/TIME OF TEST: 12/04/2022 00:07:11

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.080	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 12/04/2022 03:19:46

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Subject Sample # 1	0.008	03:14	
Breath Volume	1.183	03:14	
Air Blank	0.000	03:15	
Air Blank	0.000	03:17	
Subject Sample # 2		03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: HERNANDEZ MARTINEZ

DATE/TIME OF TEST: 12/04/2022 20:02:09

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.078	19:59	
Air Blank	0.000	20:00	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:01	
Control Test	0.079	20:01	
Air Blank	0.000	20:02	
Diagnostics Check	OK	20:02	

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,Y
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MALDONADO MESA

DATE/TIME OF TEST: 12/04/2022 22:04:07

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.079	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.221	21:59
Breath Volume	1.191	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.216	22:02
Breath Volume	1.953	22:02
Air Blank	0.000	22:03
Control Test	0.079	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 12/04/2022 23:45:58

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:30	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.137	23:33	
Breath Volume	1.363	23:33	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2 0.116	23:38	Volume Not Met
Breath Volume	0.242	23:38	
Air Blank	0.000	23:39	
Air Blank	0.000	23:41	
Subject Sample	# 3 0.134	23:44	Volume Not Met
Breath Volume	0.722	23:44	
Air Blank	0.000	23:45	
Control Test	0.078	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 12/05/2022 02:22:37

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.145	02:18
Breath Volume	2.257	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.136	02:21
Breath Volume	1.828	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: COLUMBUS

DATE/TIME OF TEST: 12/06/2022 01:02:56

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.078	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.013	00:58
Breath Volume	2.496	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.013	01:01
Breath Volume	2.648	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: PONDER

DATE/TIME OF TEST: 12/07/2022 02:56:31

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.138	02:51
Breath Volume	1.773	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.144	02:54
Breath Volume	1.167	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, PHINEAS, L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILEY, K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: ESSING

DATE/TIME OF TEST: 12/07/2022 10:45:24

Diagnostics Check	OK	10:39
Air Blank	0.000	10:39
Control Test	0.078	10:39
Air Blank	0.000	10:40
Subject Sample # 1	0.000	10:41
Breath Volume	3.265	10:41
Air Blank	0.000	10:41
Air Blank	0.000	10:43
Subject Sample # 2	0.000	10:44
Breath Volume	2.937	10:44
Air Blank	0.000	10:44
Control Test	0.079	10:44
Air Blank	0.000	10:45
Diagnostics Check	OK	10:45

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MARILLE

DATE/TIME OF TEST: 12/08/2022 01:45:19

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.133	01:40
Breath Volume	1.472	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.140	01:43
Breath Volume	1.425	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 12/08/2022 22:54:06

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.079	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.154	22:49
Breath Volume	2.464	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.155	22:52
Breath Volume	1.875	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 12/09/2022 03:08:51

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.083	03:04
Breath Volume	2.398	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.079	03:07
Breath Volume	1.644	03:07
Air Blank	0.000	03:07
Control Test	0.079	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 12/12/2022 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: CORALIC, ALEXANDAR
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: AMIDON

DATE/TIME OF TEST: 12/08/2022 02:01:19

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.190	01:55
Breath Volume	2.007	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.180	01:59
Breath Volume	2.203	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 01/25/2023 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: LAFONTANT, SEDONA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 12/08/2022 06:15:31

Diagnostics Check	OK	06:09
Air Blank	0.000	06:09
Control Test	0.079	06:09
Air Blank	0.000	06:10
Subject Sample # 1	0.092	06:10
Breath Volume	4.609	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.090	06:14
Breath Volume	4.527	06:14
Air Blank	0.000	06:14
Control Test	0.079	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 01/25/2023 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LORENZO,FRANCISCO,

BREATH TEST OPERATOR AGENCY: HIALEAH PD

ARRESTING OFFICER: FERRER,YONY

ARRESTING OFFICER AGENCY: HIALEAH PD

LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402370890

DRY GAS EXPIRY DATE: 05/09/2025

VIOLATION CODE: DUI

ARREST TIME: 04:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 12/11/2022 05:18:57

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:18	
Control Test	0.079	05:18	
Air Blank		05:18	RFI Detect
Air Blank	0.000	05:18	

Data Download Date/Time: 01/26/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: HIALEAH PD
ARRESTING OFFICER: FERRER,YONY
ARRESTING OFFICER AGENCY: HIALEAH PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 12/11/2022 05:24:45

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1	05:23	Subject Test Refused
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.080	05:24	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 01/26/2023 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,ALEXIS,
BREATH TEST OPERATOR AGENCY: HIALEAH PD
ARRESTING OFFICER: CARDOSO,MARCOS
ARRESTING OFFICER AGENCY: HIALEAH PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MORENO SUAREZ

DATE/TIME OF TEST: 12/10/2022 02:11:37

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.080	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 01/26/2023 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DEARMAS,ROBERTO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 05/11/2022
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 12/10/2022 00:12:11

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:09	
Control Test	0.079	00:09	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Control Test	0.078	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 12/27/2022 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: LEONARD,J
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 12/31/2022 06:55:29

Diagnostics Check	OK	06:50	
Air Blank	0.000	06:51	
Control Test	0.080	06:51	
Air Blank	0.000	06:52	
Subject Sample	# 1	06:54	Subject Test Refused
Breath Volume	0.000	06:54	
Air Blank	0.000	06:54	
Control Test	0.079	06:54	
Air Blank	0.000	06:55	
Diagnostics Check	OK	06:55	

Data Download Date/Time: 01/30/2023 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,GERALD,N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON,GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CROOMS

DATE/TIME OF TEST: 12/10/2022 02:03:24

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.077	02:02	
Air Blank		02:02	RFI Detect
Air Blank	0.000	02:03	

Data Download Date/Time: 01/26/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CROOMS

DATE/TIME OF TEST: 12/10/2022 02:11:54

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.084	02:07
Breath Volume	1.898	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.080	02:10
Breath Volume	1.105	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/26/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TAGIKHANI

DATE/TIME OF TEST: 12/11/2022 01:51:38

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.128	01:46
Breath Volume	1.187	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.124	01:49
Breath Volume	1.246	01:49
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 01/26/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/15/2022 00:32:49

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.077	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.203	00:28
Breath Volume	2.015	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.190	00:31
Breath Volume	1.554	00:31
Air Blank	0.000	00:31
Control Test	0.077	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/26/2023 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MALLOY

DATE/TIME OF TEST: 12/28/2022 00:49:32

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:44	
Subject Sample	# 1 0.244	00:48	Volume Not Met
Breath Volume	0.015	00:48	
Air Blank		00:48	Purge Fail
Air Blank	0.000	00:49	

Data Download Date/Time: 01/26/2023 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RODRIGUEZ, MIGUEL
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 11/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: TOLENTINO

DATE/TIME OF TEST: 12/01/2022 03:21:24

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.078	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.077	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 12/21/2022 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DARIAS, SOFIA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 11/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: HERRERA QUIJADA

DATE/TIME OF TEST: 12/01/2022 03:45:27

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	0.102	03:39
Breath Volume	1.988	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 3	0.106	03:43
Breath Volume	3.195	03:43	
Air Blank	0.000	03:44	
Control Test	0.080	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 12/21/2022 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1355293
BREATH TEST OPERATOR: BATTAGLIA,PAUL,M	DRY GAS EXPIRY DATE: 07/29/2023
BREATH TEST OPERATOR AGENCY: NORTH BAY VILLAGE PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BATTAGLIA,PAUL	ARREST TIME: 19:29
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD	GENDER: Female
LAST AGENCY INSPECTION: 11/10/2022	OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: PAZ	DATE/TIME OF TEST: 12/03/2022	20:56:26
--------------------------	-------------------------------	----------

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:44	
Control Test	0.079	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1 0.211	20:48	Volume Not Met
Breath Volume	0.269	20:48	
Air Blank	0.000	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 2 0.220	20:52	
Breath Volume	2.457	20:52	
Air Blank	0.000	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 3	20:55	Subject Test Refused
Breath Volume	0.000	20:55	
Air Blank	0.000	20:55	
Control Test	0.078	20:55	
Air Blank	0.000	20:56	
Diagnostics Check	OK	20:56	

Data Download Date/Time: 12/21/2022 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DARIAS, SOFIA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 11/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/15/2022 00:04:59

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.081	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:03	Subject Test Refused
Breath Volume	0.000	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 12/21/2022 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: MONTENEGRO, ANDREA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 11/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BASTIDAS PEREZ

DATE/TIME OF TEST: 12/17/2022 21:54:20

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.079	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.099	21:49
Breath Volume	1.582	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.094	21:52
Breath Volume	2.195	21:52
Air Blank	0.000	21:53
Control Test	0.078	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 12/21/2022 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: KAMYCZEK, ANETA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 11/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: BAJCAR

DATE/TIME OF TEST: 12/18/2022 01:07:05

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 12/21/2022 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS, GUILLERMO, E
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: BENITEZ, NORLAN
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 11/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HECKEL

DATE/TIME OF TEST: 12/19/2022 02:09:12

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.253	01:58	
Breath Volume	1.609	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2 0.211	02:04	Slope Not Level
Breath Volume	1.167	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 3	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.076	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 12/21/2022 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS,GUILLERMO,E
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: GUEITS,JOSHUA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/31/2022 00:21:49

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.079	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 01/12/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: DARIAS,SOFIA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MENDEZ GARCIA

DATE/TIME OF TEST: 12/31/2022 01:13:51

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:11	
Control Test	0.080	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.079	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 01/12/2023 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,GIRANNY,
BREATH TEST OPERATOR AGENCY: MIAMI BEACH PD
ARRESTING OFFICER: RANSON,CHAKA
ARRESTING OFFICER AGENCY: MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: MARCELIN

DATE/TIME OF TEST: 12/31/2022 07:38:43

Diagnostics Check	OK	07:31
Air Blank	0.000	07:32
Control Test	0.078	07:32
Air Blank	0.000	07:32
Subject Sample # 1	0.140	07:34
Breath Volume	2.636	07:34
Air Blank	0.000	07:34
Air Blank	0.000	07:36
Subject Sample # 2	0.140	07:37
Breath Volume	1.898	07:37
Air Blank	0.000	07:37
Control Test	0.079	07:38
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 01/12/2023 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: ORAMAS

DATE/TIME OF TEST: 12/15/2022 22:17:06

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank		22:16	RFI Detect
Air Blank	0.000	22:17	

Data Download Date/Time: 02/01/2023 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: ORAMAS

DATE/TIME OF TEST: 12/15/2022 22:27:02

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Subject Sample # 1	0.126	22:22	
Breath Volume	1.804	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample # 2	0.130	22:25	
Breath Volume	1.679	22:25	
Air Blank	0.000	22:26	
Control Test	0.008	22:26	RFI Detect
Air Blank	0.000	22:27	

Data Download Date/Time: 02/01/2023 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: RIVERO,N
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DELEONPEREZ

DATE/TIME OF TEST: 12/22/2022 02:56:02

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.215	02:51
Breath Volume	1.781	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.203	02:54
Breath Volume	2.617	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

Data Download Date/Time: 02/01/2023 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ, ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: RISCO, DARIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: PARDOGUTIERREZ

DATE/TIME OF TEST: 12/24/2022 02:54:04

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.105	02:49
Breath Volume	2.390	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:51
Subject Sample # 2	0.105	02:52
Breath Volume	2.765	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:54

Data Download Date/Time: 02/01/2023 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: GUZMAN,RICHARD
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:46

SUBJECT'S LAST NAME: RAMIREZ DE DIAZ

DATE/TIME OF TEST: 12/24/2022 11:13:26

Diagnostics Check	OK	11:07
Air Blank	0.000	11:07
Control Test	0.081	11:08
Air Blank	0.000	11:08
Subject Sample # 1	0.195	11:08
Breath Volume	2.207	11:08
Air Blank	0.000	11:09
Air Blank	0.000	11:11
Subject Sample # 2	0.197	11:11
Breath Volume	1.660	11:11
Air Blank	0.000	11:12
Control Test	0.081	11:12
Air Blank	0.000	11:13
Diagnostics Check	OK	11:13

Data Download Date/Time: 02/01/2023 18:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: RISCO,D
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: LEPEROJO

DATE/TIME OF TEST: 12/30/2022 03:32:49

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.078	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 02/01/2023 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,JOSHUA,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HARRISON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: GODBOLT

DATE/TIME OF TEST: 12/10/2022 06:20:03

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.083	06:12
Air Blank	0.000	06:12
Subject Sample # 1	0.168	06:13
Breath Volume	1.671	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.169	06:18
Breath Volume	2.078	06:18
Air Blank	0.000	06:19
Control Test	0.082	06:19
Air Blank	0.000	06:19
Diagnostics Check	OK	06:20

Data Download Date/Time: 01/04/2023 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,ABIGAYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS,ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 12/24/2022 21:17:28

Diagnostics Check	OK	21:16	
Air Blank		21:17	RFI Detect
Air Blank	0.000	21:17	

Data Download Date/Time: 12/29/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 12/24/2022 21:21:29

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:20	RFI Detect
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Control Test	0.081	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 12/29/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,ABIGAYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS,ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 12/24/2022 21:22:49

Diagnostics Check	OK	21:22	
Air Blank		21:22	RFI Detect
Air Blank		21:22	RFI Detect

Data Download Date/Time: 12/29/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/25/2022 17:00:56

Diagnostics Check	OK	16:57	
Air Blank	0.000	16:57	
Control Test	0.082	16:57	
Air Blank	0.000	16:58	
Subject Sample	# 1 0.201	16:59	
Breath Volume	1.941	16:59	
Air Blank		17:00	Purge Fail
Air Blank	0.000	17:00	

Data Download Date/Time: 12/29/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/25/2022 17:07:53

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:03	
Control Test	0.082	17:03	
Air Blank	0.000	17:04	
Subject Sample # 1	0.187	17:06	
Breath Volume	2.093	17:06	
Air Blank		17:07	Purge Fail
Air Blank	0.000	17:07	

Data Download Date/Time: 12/29/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/25/2022 17:16:10

Diagnostics Check	OK	17:09	
Air Blank	0.000	17:09	
Control Test	0.081	17:10	
Air Blank	0.000	17:10	
Subject Sample # 1	0.181	17:11	
Breath Volume	1.320	17:11	
Air Blank	0.000	17:12	
Air Blank	0.000	17:13	
Subject Sample # 2	0.196	17:14	
Breath Volume	1.781	17:14	
Air Blank		17:15	Purge Fail
Air Blank	0.000	17:16	

Data Download Date/Time: 12/29/2022 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 12/15/2022 04:41:22

Diagnostics Check	OK	04:41	
Air Blank		04:41	RFI Detect
Air Blank		04:41	RFI Detect

Data Download Date/Time: 02/01/2023 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 12/18/2022 02:34:15

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.157	02:29	
Breath Volume	1.746	02:29	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 2 0.168	02:32	
Breath Volume	2.003	02:32	
Air Blank	0.000	02:33	
Control Test	0.078	02:33	
Air Blank		02:33	RFI Detect
Air Blank	0.000	02:34	

Data Download Date/Time: 02/01/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 12/18/2022 02:45:27

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.166	02:40
Breath Volume	1.679	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.169	02:43
Breath Volume	1.640	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/01/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SADIK

DATE/TIME OF TEST: 12/18/2022 03:54:26

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
		03:54	RFI Detect
Air Blank		03:54	RFI Detect

Data Download Date/Time: 02/01/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SADIK

DATE/TIME OF TEST: 12/18/2022 03:56:52

Diagnostics Check	OK	03:56	
Air Blank		03:56	RFI Detect
Air Blank	0.000	03:56	

Data Download Date/Time: 02/01/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,AUSTIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SADIK

DATE/TIME OF TEST: 12/18/2022 04:08:48

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.078	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.117	04:03
Breath Volume	1.492	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.106	04:07
Breath Volume	1.417	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 02/01/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLEN,JUSTIN,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WALLEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: SHARRON

DATE/TIME OF TEST: 12/30/2022 19:36:49

Diagnostics Check	Fail	19:36	Voltage/Current Test
Air Blank	0.000	19:36	

Data Download Date/Time: 01/06/2023 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: ROMIG

DATE/TIME OF TEST: 12/01/2022 01:37:44

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.082	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.122	01:33
Breath Volume	2.972	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.121	01:36
Breath Volume	3.343	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: KILLEN

DATE/TIME OF TEST: 12/02/2022 03:12:10

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.222	03:07
Breath Volume	2.535	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.210	03:10
Breath Volume	1.312	03:10
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK, DAVID, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: LARGEN, JACK
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: UDDIN

DATE/TIME OF TEST: 12/03/2022 03:10:26

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.097	03:06
Breath Volume	1.652	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.097	03:08
Breath Volume	2.386	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TYLER, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS, TYLER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DINSMORE-NASSAR

DATE/TIME OF TEST: 12/04/2022 02:40:44

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.159	02:36
Breath Volume	1.949	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.163	02:39
Breath Volume	2.660	02:39
Air Blank	0.000	02:39
Control Test	0.078	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,S,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,S
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 12/07/2022 20:07:30

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:03	
Subject Sample	# 1	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:06	
Control Test	0.081	20:06	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGER, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ALBRECHT, SYDNEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 12/10/2022 01:29:53

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.262	01:25
Breath Volume	2.390	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.253	01:28
Breath Volume	2.550	01:28
Air Blank	0.000	01:28
Control Test	0.078	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGER, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ALBRECHT, SYDNEY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 12/10/2022 02:07:53

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.081	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/11/2022 03:33:34

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.083	03:29
Breath Volume	3.574	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.081	03:32
Breath Volume	3.531	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: NOLASCO-MEJORGA

DATE/TIME OF TEST: 12/17/2022 23:19:39

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.082	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:15
Breath Volume	4.027	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:18
Breath Volume	4.871	23:18
Air Blank	0.000	23:18
Control Test	0.082	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/20/2022 02:52:33

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.167	02:47
Breath Volume	2.609	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.175	02:51
Breath Volume	2.382	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 12/21/2022 22:36:00

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:33	
Control Test	0.080	22:33	
Air Blank	0.000	22:34	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: DUCHARME

DATE/TIME OF TEST: 12/22/2022 02:31:04

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.143	02:26
Breath Volume	3.347	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.148	02:29
Breath Volume	3.257	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: STIRLING,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/22/2022 15:59:24

Diagnostics Check	OK	15:52
Air Blank	0.000	15:52
Control Test	0.081	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.126	15:54
Breath Volume	3.835	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.121	15:57
Breath Volume	2.292	15:57
Air Blank	0.000	15:58
Control Test	0.078	15:58
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: STIRLING,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BONANNO

DATE/TIME OF TEST: 12/24/2022 01:24:15

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.082	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.082	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BARNES JR

DATE/TIME OF TEST: 12/24/2022 03:59:53

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.082	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.129	03:54
Breath Volume	1.207	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.132	03:58
Breath Volume	1.675	03:58
Air Blank	0.000	03:58
Control Test	0.083	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 12/30/2022 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: STIRLING,MORGAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/25/2022 00:15:16

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.083	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.000	00:11
Breath Volume	3.789	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.000	00:13
Breath Volume	3.914	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 12/30/2022 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 12/27/2022 04:00:41

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:58	
Control Test	0.082	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.083	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 12/30/2022 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 12/28/2022 00:01:59

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.082	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 2 0.244	23:49	
Breath Volume	1.500	23:49	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 3 0.240	23:54	Volume Not Met
Breath Volume	0.019	23:54	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 1 0.217	00:00	Volume Not Met
Breath Volume	0.476	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 12/30/2022 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/29/2022 02:50:52

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.132	02:45
Breath Volume	1.535	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.140	02:49
Breath Volume	1.488	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 12/30/2022 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/30/2022 04:46:10

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.081	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.142	04:40
Breath Volume	1.878	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.137	04:44
Breath Volume	1.652	04:44
Air Blank	0.000	04:45
Control Test	0.077	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 12/30/2022 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSFORD,CARLDERIUS,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANSFORD,CARLDERIUS
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CABALLA

DATE/TIME OF TEST: 12/31/2022 21:21:58

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
Control Test	0.081	21:16	
Air Blank	0.000	21:17	
Subject Sample # 1	0.000	21:17	
Breath Volume	1.128	21:17	
Air Blank	0.000	21:18	
Air Blank	0.000	21:20	
Subject Sample # 2		21:20	RFI Detect
Breath Volume	0.000	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 02/02/2023 00:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: KLOPFENSTEIN

DATE/TIME OF TEST: 12/02/2022 15:48:46

Diagnostics Check	OK	15:42
Air Blank	0.000	15:43
Control Test	0.081	15:43
Air Blank	0.000	15:43
Subject Sample # 1	0.212	15:44
Breath Volume	2.300	15:44
Air Blank	0.000	15:45
Air Blank	0.000	15:46
Subject Sample # 2	0.216	15:47
Breath Volume	2.441	15:47
Air Blank	0.000	15:47
Control Test	0.080	15:48
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUNDIN,GABREAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MASON,NICHOLAS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 12/03/2022 00:30:33

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:25	No Sample Provided
Breath Volume	0.000	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample	# 2	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.081	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 12/03/2022 04:40:45

Diagnostics Check	OK	04:40	
Air Blank		04:40	RFI Detect
Air Blank	0.000	04:40	

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 12/03/2022 04:51:10

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.081	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.205	04:45
Breath Volume	2.433	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample # 2	0.200	04:49
Breath Volume	2.558	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 12/04/2022 00:29:12

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.178	00:24
Breath Volume	1.468	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.182	00:27
Breath Volume	1.976	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHAGAN,ANDREW,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: JENNINGS,A
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LOPEZ-VASQUEZ

DATE/TIME OF TEST: 12/04/2022 03:02:08

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:55	No Sample Provided
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 2	03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHAGAN,ANDREW,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: JENNINGS,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LOPEZ-VASQUEZ

DATE/TIME OF TEST: 12/04/2022 03:49:07

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1 0.021	03:39	Volume Not Met
Breath Volume	0.031	03:39	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample	# 2 0.204	03:42	
Breath Volume	1.367	03:42	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 3	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHAGAN,ANDREW,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GEOGHAGAN,ANDREW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/05/2022 02:26:40

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.079	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:20	No Sample Provided
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA MEJIA,JOSE,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOWERY,BRITTANY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: AVILA MEJIA

DATE/TIME OF TEST: 12/08/2022 03:15:16

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1 0.158	03:09	
Breath Volume	1.261	03:09	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2	03:13	Subject Test Refused
Breath Volume	0.386	03:13	
Air Blank	0.000	03:14	
Control Test	0.079	03:14	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: MASICH

DATE/TIME OF TEST: 12/09/2022 19:52:03

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.240	19:47
Breath Volume	2.894	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:49
Subject Sample # 2	0.236	19:50
Breath Volume	1.910	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:52

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER, LANDON, L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEEBECK, JENNA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: CASSIANODESOUZA

DATE/TIME OF TEST: 12/09/2022 22:14:41

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.080	22:12	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.080	22:14	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 12/27/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITTRELL, WILLIAM, G
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KITTRELL, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: YAC HERNANDEZ

DATE/TIME OF TEST: 12/11/2022 00:41:10

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.214	00:36
Breath Volume	2.453	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.222	00:39
Breath Volume	3.035	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: WESSELLS

DATE/TIME OF TEST: 12/13/2022 17:22:13

Diagnostics Check	OK	17:15
Air Blank	0.000	17:16
Control Test	0.081	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.156	17:17
Breath Volume	1.464	17:17
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.153	17:20
Breath Volume	1.160	17:20
Air Blank	0.000	17:21
Control Test	0.080	17:21
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHAGAN,ANDREW,C
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RANDOLPH,BRENT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 12/14/2022 03:07:55

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.080	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:02	No Sample Provided
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 2	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: GAGE

DATE/TIME OF TEST: 12/16/2022 04:43:29

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.163	04:37
Breath Volume	1.660	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.151	04:41
Breath Volume	1.242	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOONEY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOONEY,KEVIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CROWLEY

DATE/TIME OF TEST: 12/16/2022 23:13:43

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.081	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1 0.228	23:03	
Breath Volume	1.855	23:03	
Air Blank	0.000	23:03	
Air Blank	0.000	23:05	
Subject Sample	# 2	23:08	No Sample Provided
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Air Blank	0.000	23:11	
Subject Sample	# 3 0.212	23:12	
Breath Volume	1.699	23:12	
Air Blank	0.000	23:12	
Control Test	0.080	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITTRELL, WILLIAM, G
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE, RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MC MANUS

DATE/TIME OF TEST: 12/17/2022 00:17:39

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.159	00:13
Breath Volume	1.851	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.157	00:16
Breath Volume	2.167	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWERY,BRITTANY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOWERY,BRITTANY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 11:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 12/17/2022 01:11:13

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:07	
Control Test	0.082	01:07	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:09	Subject Test Refused
Breath Volume	0.000	01:09	
Air Blank	0.000	01:10	
Control Test	0.081	01:10	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOONEY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOONEY,KEVIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 12/17/2022 05:15:35

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:11	
Control Test	0.081	05:11	
Air Blank	0.000	05:11	
Subject Sample	# 1 0.181	05:12	
Breath Volume	3.003	05:12	
Air Blank	0.000	05:13	
Air Blank	0.000	05:15	
Subject Sample	# 2	05:15	RFI Detect
Breath Volume	0.000	05:15	
Air Blank		05:15	RFI Detect
Air Blank		05:15	RFI Detect

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWERY,BRITTANY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOWERY,BRITTANY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: VAN GRAAN

DATE/TIME OF TEST: 12/17/2022 06:16:20

Diagnostics Check	OK	06:09
Air Blank	0.000	06:10
Control Test	0.080	06:10
Air Blank	0.000	06:10
Subject Sample # 1	0.190	06:11
Breath Volume	2.515	06:11
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.184	06:14
Breath Volume	4.777	06:14
Air Blank	0.000	06:15
Control Test	0.079	06:15
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRANE,RICHARD,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DRANE,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 12/18/2022 05:23:31

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.082	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:16	No Sample Provided
Breath Volume	0.000	05:16	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample	# 2	05:22	No Sample Provided
Breath Volume	0.000	05:22	
Air Blank	0.000	05:22	
Control Test	0.082	05:22	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELDON,ERIC,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SHELDON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/18/2022 17:18:07

Diagnostics Check	OK	17:15	
Air Blank	0.000	17:15	
Control Test	0.082	17:16	
Air Blank	0.000	17:16	
Subject Sample	# 1	17:16	Subject Test Refused
Breath Volume	0.000	17:16	
Air Blank	0.000	17:17	
Control Test	0.082	17:17	
Air Blank	0.000	17:18	
Diagnostics Check	OK	17:18	

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 12/18/2022 18:32:14

Diagnostics Check	OK	18:24
Air Blank	0.000	18:24
Control Test	0.081	18:24
Air Blank	0.000	18:25
Subject Sample # 1	0.191	18:27
Breath Volume	1.175	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:29
Subject Sample # 2	0.199	18:30
Breath Volume	1.281	18:30
Air Blank	0.000	18:31
Control Test	0.080	18:31
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORTON,JAMES,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MORTON,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 12/20/2022 22:22:01

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.160	22:16
Breath Volume	2.421	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.150	22:20
Breath Volume	2.089	22:20
Air Blank	0.000	22:21
Control Test	0.078	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:22

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,LANDON,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: WARDEN

DATE/TIME OF TEST: 12/23/2022 20:22:04

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.082	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.000	20:17
Breath Volume	1.476	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.000	20:20
Breath Volume	1.968	20:20
Air Blank	0.000	20:21
Control Test	0.082	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:22

Data Download Date/Time: 12/27/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEEBECK, JENNA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CANIZALES

DATE/TIME OF TEST: 12/29/2022 00:16:35

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.184	00:11
Breath Volume	1.765	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.171	00:14
Breath Volume	2.535	00:14
Air Blank	0.000	00:15
Control Test	0.078	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 01/30/2023 16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,DAVID,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DIXON,EZEKIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TRAMMELL

DATE/TIME OF TEST: 12/30/2022 01:08:44

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.081	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 01/30/2023 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ERIC,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: ORTIZ DIEGO

DATE/TIME OF TEST: 12/30/2022 15:29:33

Diagnostics Check	OK	15:23
Air Blank	0.000	15:23
Control Test	0.081	15:24
Air Blank	0.000	15:24
Subject Sample # 1	0.294	15:25
Breath Volume	1.195	15:25
Air Blank	0.000	15:25
Air Blank	0.000	15:27
Subject Sample # 2	0.298	15:27
Breath Volume	1.492	15:27
Air Blank	0.000	15:28
Control Test	0.080	15:28
Air Blank	0.000	15:29
Diagnostics Check	OK	15:29

Data Download Date/Time: 01/30/2023 16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOONEY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HACKETT,TRAVIS
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/31/2022 05:25:44

Diagnostics Check	OK	05:25	
Air Blank		05:25	RFI Detect
Air Blank	0.000	05:25	

Data Download Date/Time: 01/30/2023 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOONEY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HACKETT,TRAVIS
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 12/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: O325341-9
DRY GAS EXPIRY DATE: 11/18/2024
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/31/2022 05:33:56

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.045	05:29
Breath Volume	1.792	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:31
Subject Sample # 2	0.042	05:32
Breath Volume	3.132	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 01/30/2023 16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: CULBERTSON,TYLER
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: COPELAND 3

DATE/TIME OF TEST: 12/11/2022 02:22:15

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.082	02:20	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 12/20/2022 16:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: LEWIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1249794
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: VAN COTT

DATE/TIME OF TEST: 12/12/2022 19:42:24

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.220	19:37
Breath Volume	2.085	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.216	19:40
Breath Volume	2.296	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 12/20/2022 16:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/02/2022 20:41:38

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.079	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.206	20:37
Breath Volume	2.148	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.200	20:40
Breath Volume	2.328	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 12/14/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: ADDERLEY

DATE/TIME OF TEST: 12/03/2022 22:38:58

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.080	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.139	22:34
Breath Volume	1.785	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.129	22:37
Breath Volume	1.679	22:37
Air Blank	0.000	22:38
Control Test	0.078	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 12/14/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HEDBERG

DATE/TIME OF TEST: 12/04/2022 04:02:02

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.250	03:57
Breath Volume	2.937	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.254	04:00
Breath Volume	2.976	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:02

Data Download Date/Time: 12/14/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: NUCCIO

DATE/TIME OF TEST: 12/30/2022 22:03:09

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.161	21:58
Breath Volume	2.023	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.154	22:01
Breath Volume	1.410	22:01
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 02/01/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 12/02/2022 05:06:17

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.081	04:52	
Air Blank	0.000	04:53	
Subject Sample	# 1 0.110	04:56	Volume Not Met
Breath Volume	0.949	04:56	
Air Blank	0.000	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 2 0.114	05:00	
Breath Volume	1.605	05:00	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample	# 3 0.110	05:04	
Breath Volume	1.164	05:04	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 12/09/2022 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENLEE,NANCY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMSON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: KREMKAU

DATE/TIME OF TEST: 12/02/2022 18:37:03

Diagnostics Check	OK	18:30
Air Blank	0.000	18:30
Control Test	0.081	18:31
Air Blank	0.000	18:31
Subject Sample # 1	0.083	18:32
Breath Volume	2.761	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.081	18:35
Breath Volume	1.750	18:35
Air Blank	0.000	18:36
Control Test	0.080	18:36
Air Blank	0.000	18:36
Diagnostics Check	OK	18:37

Data Download Date/Time: 12/09/2022 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 12/02/2022 23:39:54

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:32	No Sample Provided
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Air Blank	0.000	23:34	
Subject Sample	# 2	0.278	23:35
Breath Volume	2.886	23:35	
Air Blank	0.000	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 3	0.289	23:38
Breath Volume	2.316	23:38	
Air Blank	0.000	23:38	
Control Test	0.080	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 12/09/2022 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WELLS, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PENA MOJENA

DATE/TIME OF TEST: 12/03/2022 03:08:51

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.148	03:04
Breath Volume	2.339	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.140	03:07
Breath Volume	1.562	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 12/09/2022 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 12/03/2022 21:19:41

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.380	21:15
Breath Volume	1.625	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.374	21:18
Breath Volume	1.753	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASON,FRANCESCA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAASE,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: CROWDER-SCHOPPMAN

DATE/TIME OF TEST: 12/03/2022 23:22:31

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.080	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.165	23:18
Breath Volume	1.484	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.179	23:20
Breath Volume	1.476	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: AUBERTIN

DATE/TIME OF TEST: 12/04/2022 00:18:31

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.314	00:14
Breath Volume	1.730	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.313	00:16
Breath Volume	1.609	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHIRELING,S
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: NEESE

DATE/TIME OF TEST: 12/04/2022 00:39:08

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.242	00:33
Breath Volume	1.199	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.232	00:37
Breath Volume	2.394	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: AGRONT

DATE/TIME OF TEST: 12/04/2022 06:51:52

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.080	06:46
Air Blank	0.000	06:46
Subject Sample # 1	0.118	06:47
Breath Volume	3.445	06:47
Air Blank	0.000	06:47
Air Blank	0.000	06:49
Subject Sample # 2	0.117	06:50
Breath Volume	2.957	06:50
Air Blank	0.000	06:50
Control Test	0.080	06:51
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOX,STEPHANIE,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOGOLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: KIMBLE

DATE/TIME OF TEST: 12/04/2022 17:08:01

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:56	
Control Test	0.080	16:57	
Air Blank	0.000	16:57	
Subject Sample	# 1 0.132	17:00	Volume Not Met
Breath Volume	0.125	17:00	
Air Blank	0.000	17:01	
Air Blank	0.000	17:03	
Subject Sample	# 2 0.146	17:06	Volume Not Met
Breath Volume	0.179	17:06	
Air Blank	0.000	17:07	
Control Test	0.079	17:07	
Air Blank	0.000	17:07	
Diagnostics Check	OK	17:08	

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: NOGALES CLAROS

DATE/TIME OF TEST: 12/05/2022 05:08:50

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.191	05:03
Breath Volume	1.210	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.175	05:07
Breath Volume	1.203	05:07
Air Blank	0.000	05:07
Control Test	0.080	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CHOI,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 12/05/2022 20:41:02

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.079	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.193	20:36
Breath Volume	1.988	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.196	20:39
Breath Volume	1.757	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROBERTS,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MUNSON

DATE/TIME OF TEST: 12/05/2022 21:03:34

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.080	20:58
Air Blank	0.000	20:58
Subject Sample # 1	0.218	20:59
Breath Volume	2.640	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.198	21:02
Breath Volume	2.105	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: LOPEZ,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: VISMANOS

DATE/TIME OF TEST: 12/07/2022 00:12:55

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.179	00:08
Breath Volume	1.449	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.183	00:11
Breath Volume	1.597	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GALLOWAY,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: CASTILLA

DATE/TIME OF TEST: 12/07/2022 03:48:44

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.107	03:44
Breath Volume	1.804	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.112	03:47
Breath Volume	1.781	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CRUZ AGUIRRE

DATE/TIME OF TEST: 12/08/2022 01:18:23

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.173	01:12
Breath Volume	1.296	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.165	01:16
Breath Volume	1.382	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 12/09/2022 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MORINA

DATE/TIME OF TEST: 12/17/2022 02:31:08

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.019	02:26
Breath Volume	1.902	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.016	02:29
Breath Volume	1.222	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BOBOTAS

DATE/TIME OF TEST: 12/17/2022 22:56:27

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.110	22:51
Breath Volume	1.878	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.110	22:54
Breath Volume	1.820	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,SETH,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,SETH
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 12/17/2022 23:35:48

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.055	23:30
Breath Volume	1.363	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.054	23:34
Breath Volume	1.664	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MOLENA

DATE/TIME OF TEST: 12/18/2022 00:20:22

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.100	00:15
Breath Volume	1.812	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.101	00:18
Breath Volume	2.292	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 12/18/2022 01:16:13

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.159	01:11
Breath Volume	2.531	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.159	01:14
Breath Volume	2.917	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: KRAGTHORPE

DATE/TIME OF TEST: 12/18/2022 01:46:13

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.000	01:41
Breath Volume	2.214	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:44
Breath Volume	3.734	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DETRANO,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: CORREIA

DATE/TIME OF TEST: 12/18/2022 02:26:17

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.217	02:21
Breath Volume	2.066	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.212	02:24
Breath Volume	1.660	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: GILCHRIST

DATE/TIME OF TEST: 12/18/2022 03:33:38

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 01/23/2023 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,M
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 11/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: ADMIN
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: CHAFFIN

DATE/TIME OF TEST: 12/05/2022 06:52:21

Diagnostics Check	OK	06:46
Air Blank	0.000	06:47
Control Test	0.079	06:47
Air Blank	0.000	06:47
Subject Sample # 1	0.000	06:48
Breath Volume	2.769	06:48
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.000	06:51
Breath Volume	1.960	06:51
Air Blank	0.000	06:51
Control Test	0.078	06:51
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 12/15/2022 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINDHAM,AUSTIN,K
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/10/2022 01:14:08

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.139	01:09
Breath Volume	2.593	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.132	01:12
Breath Volume	2.628	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/15/2022 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEYER, CHARLES, L
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 12/15/2022 23:37:02

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.079	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:30	No Sample Provided
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2	23:35	No Sample Provided
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 01/26/2023 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: SMITH,MATT
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: ADMIN
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 12/19/2022 21:38:54

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.080	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.000	21:34
Breath Volume	2.453	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:37
Breath Volume	2.484	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 01/26/2023 12:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MEANEY,TREVOR
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 12/05/2022 12:56:03

Diagnostics Check	OK	12:49
Air Blank	0.000	12:50
Control Test	0.081	12:50
Air Blank	0.000	12:51
Subject Sample # 1	0.180	12:51
Breath Volume	2.792	12:51
Air Blank	0.000	12:52
Air Blank	0.000	12:54
Subject Sample # 2	0.178	12:54
Breath Volume	2.457	12:54
Air Blank	0.000	12:55
Control Test	0.080	12:55
Air Blank	0.000	12:55
Diagnostics Check	OK	12:56

Data Download Date/Time: 12/30/2022 15:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JASON,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: AVILA,NICOLE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PERLA LAZO

DATE/TIME OF TEST: 12/10/2022 04:17:46

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.190	04:13
Breath Volume	2.023	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.187	04:16
Breath Volume	1.886	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 12/30/2022 15:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCHESTER,KAYLIE,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MILLS,JACOB
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: DELAURENTIS

DATE/TIME OF TEST: 12/22/2022 00:22:22

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.081	00:10	
Air Blank	0.000	00:10	
Subject Sample	# 1 0.181	00:14	Volume Not Met
Breath Volume	0.132	00:14	
Air Blank	0.000	00:14	
Air Blank	0.000	00:16	
Subject Sample	# 2 0.180	00:17	
Breath Volume	1.535	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample	# 3 0.184	00:20	
Breath Volume	1.167	00:20	
Air Blank	0.000	00:21	
Control Test	0.078	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 12/30/2022 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: PETERSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/25/2025
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 12/24/2022 00:27:12

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.184	00:22
Breath Volume	1.933	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.186	00:25
Breath Volume	2.171	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 12/30/2022 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: SERRANO, VICENTE
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: VEGA-SILVA

DATE/TIME OF TEST: 12/10/2022 02:48:37

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.050	02:44
Breath Volume	2.089	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.046	02:47
Breath Volume	2.136	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 12/15/2022 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: SERRANO, VICENTE
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: PEREZ-ALMANZA

DATE/TIME OF TEST: 12/24/2022 01:26:42

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.115	01:22
Breath Volume	2.765	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.109	01:25
Breath Volume	1.914	01:25
Air Blank	0.000	01:25
Control Test	0.081	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/17/2023 16:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: SERRANO, VICENTE
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PADILLA-RAMIREZ

DATE/TIME OF TEST: 12/25/2022 00:48:21

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.155	00:43
Breath Volume	2.742	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.160	00:46
Breath Volume	3.085	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 01/17/2023 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: SERRANO, VICENTE
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 12/09/2022
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/26/2022 02:15:48

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.073	02:11
Breath Volume	1.500	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.074	02:14
Breath Volume	1.375	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 01/17/2023 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: SERRANO,VICENTE
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 12/29/2022 01:20:15

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.081	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:19	
Control Test	0.082	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 01/17/2023 16:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 12/03/2022 02:13:25

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 12/13/2022 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAETSCH,JOSH,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSH
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 12/09/2022 01:08:31

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.081	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 1 0.095	01:00	Volume Not Met
Breath Volume	0.996	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2 0.104	01:03	
Breath Volume	1.347	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 3 0.104	01:06	
Breath Volume	1.304	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 12/13/2022 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAZURYK, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE, AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 11/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: GRAND PRE

DATE/TIME OF TEST: 12/10/2022 20:40:33

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.081	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Control Test	0.080	20:39	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

Data Download Date/Time: 12/13/2022 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: FARRIS,MICHAEL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: MAZAIEGOS PERDOMO

DATE/TIME OF TEST: 12/24/2022 18:25:31

Diagnostics Check	OK	18:19
Air Blank	0.000	18:19
Control Test	0.083	18:20
Air Blank	0.000	18:20
Subject Sample # 1	0.206	18:21
Breath Volume	1.781	18:21
Air Blank	0.000	18:21
Air Blank	0.000	18:23
Subject Sample # 2	0.194	18:23
Breath Volume	1.753	18:23
Air Blank	0.000	18:24
Control Test	0.079	18:24
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 01/20/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: MITTS JR

DATE/TIME OF TEST: 12/31/2022 19:09:50

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:07	
Control Test	0.080	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:08	Subject Test Refused
Breath Volume	0.000	19:08	
Air Blank	0.000	19:08	
Control Test	0.079	19:09	
Air Blank	0.000	19:09	
Diagnostics Check	OK	19:09	

Data Download Date/Time: 01/20/2023 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HERNANDEZ-MACHADO

DATE/TIME OF TEST: 12/30/2022 23:30:03

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.076	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	RFI Detect
Breath Volume	0.000	23:28	
Air Blank	0.000	23:29	
Control Test	0.076	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 01/05/2023 07:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HERNANDEZ-MACHADO

DATE/TIME OF TEST: 12/30/2022 23:38:11

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.076	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.076	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 01/05/2023 07:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADURO,VICTOR,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MADURO,VICTOR
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/31/2022 03:23:25

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.078	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.077	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 01/05/2023 07:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LORIGA,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 12/01/2022 00:12:05

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.190	00:07
Breath Volume	2.945	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.195	00:10
Breath Volume	3.183	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/01/2022 05:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGILA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 12/02/2022 02:52:41

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.078	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:48
Breath Volume	1.742	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.000	02:51
Breath Volume	1.257	02:51
Air Blank	0.000	02:51
Control Test	0.078	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/05/2023 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: JEAN BAPTIST,D
ARRESTING OFFICER AGENCY: NORTH MIAMI BEACH PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 12/03/2022 23:18:04

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.077	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.303	23:13
Breath Volume	1.937	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.310	23:16
Breath Volume	2.828	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:18

Data Download Date/Time: 01/05/2023 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: COLVIN

DATE/TIME OF TEST: 12/04/2022 05:50:28

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.078	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.164	05:45
Breath Volume	1.777	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.160	05:48
Breath Volume	1.343	05:48
Air Blank	0.000	05:49
Control Test	0.078	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 01/05/2023 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 12/08/2022 00:50:20

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.077	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.078	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 01/05/2023 07:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATTAGILA,P
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: VILLAVICENCIO PENA

DATE/TIME OF TEST: 12/09/2022 03:54:28

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.077	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1 0.136	03:46	Volume Not Met
Breath Volume	0.187	03:46	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 2 0.184	03:49	
Breath Volume	3.312	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:52	
Subject Sample	# 3 0.179	03:52	
Breath Volume	1.531	03:52	
Air Blank	0.000	03:53	
Control Test	0.076	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 01/05/2023 08:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CHAVEZ LOPEZ

DATE/TIME OF TEST: 12/10/2022 01:36:07

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.077	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.168	01:31
Breath Volume	1.835	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.161	01:34
Breath Volume	2.480	01:34
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 01/05/2023 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 12/11/2022 05:17:33

Diagnostics Check	OK	05:14	
Air Blank	0.000	05:14	
Control Test	0.078	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.077	05:16	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 01/05/2023 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 12/03/2022 05:51:24

Diagnostics Check	OK	05:44
Air Blank	0.000	05:44
Control Test	0.079	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.124	05:46
Breath Volume	1.417	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.127	05:49
Breath Volume	1.761	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 12/22/2022 08:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,RYAN,T
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RODRIGUEZ,E
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 01/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 12/20/2022 05:32:29

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.133	05:27
Breath Volume	1.812	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.136	05:30
Breath Volume	1.570	05:30
Air Blank	0.000	05:31
Control Test	0.079	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 12/22/2022 08:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 12/17/2022 08:42:48

Diagnostics Check	OK	08:36
Air Blank	0.000	08:37
Control Test	0.080	08:37
Air Blank	0.000	08:37
Subject Sample # 1	0.111	08:38
Breath Volume	1.519	08:38
Air Blank	0.000	08:39
Air Blank	0.000	08:40
Subject Sample # 2	0.122	08:41
Breath Volume	3.578	08:41
Air Blank	0.000	08:41
Control Test	0.081	08:42
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

Data Download Date/Time: 12/22/2022 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BRINKER

DATE/TIME OF TEST: 12/01/2022 00:16:56

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 12/14/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: KARODE

DATE/TIME OF TEST: 12/16/2022 21:55:34

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.079	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.158	21:50
Breath Volume	2.265	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.162	21:53
Breath Volume	1.691	21:53
Air Blank	0.000	21:54
Control Test	0.079	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 12/22/2022 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: PACOJOJ

DATE/TIME OF TEST: 12/17/2022 22:35:07

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.341	22:30
Breath Volume	1.195	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.344	22:33
Breath Volume	2.039	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 12/22/2022 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19021080A2
DRY GAS EXPIRY DATE: 09/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: HERNANDEZ ALVARADO

DATE/TIME OF TEST: 12/17/2022 23:33:58

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:28
Subject Sample # 1	0.212	23:29
Breath Volume	2.261	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.196	23:32
Breath Volume	2.558	23:32
Air Blank	0.000	23:33
Control Test	0.077	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 12/22/2022 14:24