

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A3
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: OUELLETE

DATE/TIME OF TEST: 12/03/2022 00:42:57

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.129	00:38
Breath Volume	1.941	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.134	00:41
Breath Volume	2.871	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 12/05/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 12/05/2022 21:44:38

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.080	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:40	No Sample Provided
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:43	
Control Test	0.081	21:44	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 12/30/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: SOLIS PEREZ

DATE/TIME OF TEST: 12/08/2022 21:17:41

Diagnostics Check	OK	21:08
Air Blank	0.000	21:09
Control Test	0.079	21:09
Air Blank	0.000	21:10
Subject Sample # 1	0.230	21:11
Breath Volume	2.050	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.222	21:16
Breath Volume	1.964	21:16
Air Blank	0.000	21:16
Control Test	0.078	21:17
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 12/30/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ROCA,LUIS
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 12/09/2022 22:17:02

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.078	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.204	22:11
Breath Volume	1.582	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.187	22:15
Breath Volume	1.500	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:17

Data Download Date/Time: 12/30/2022 11:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: AGUILAR JUAREZ

DATE/TIME OF TEST: 12/25/2022 20:14:24

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.082	20:08
Air Blank	0.000	20:08
Subject Sample # 1	0.247	20:09
Breath Volume	1.953	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.246	20:12
Breath Volume	1.625	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 12/30/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCA,LUIS,C
BREATH TEST OPERATOR AGENCY: HOMESTEAD PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: HOMESTEAD PD
LAST AGENCY INSPECTION: 12/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/25/2022 20:35:09

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.081	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.260	20:30
Breath Volume	1.691	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.273	20:33
Breath Volume	2.855	20:33
Air Blank	0.000	20:34
Control Test	0.081	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 12/30/2022 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MARK,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BUSTILLO,ZANE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: LAUFTER

DATE/TIME OF TEST: 12/04/2022 12:59:35

Diagnostics Check	OK	12:58	
Air Blank		12:59	RFI Detect
Air Blank	0.000	12:59	

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, MARK, E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BUSTILLO, ZANE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: LAUFTER

DATE/TIME OF TEST: 12/04/2022 13:09:29

Diagnostics Check	OK	13:02
Air Blank	0.000	13:02
Control Test	0.079	13:02
Air Blank	0.000	13:03
Subject Sample # 1	0.270	13:04
Breath Volume	1.566	13:04
Air Blank	0.000	13:05
Air Blank	0.000	13:06
Subject Sample # 2	0.265	13:07
Breath Volume	1.671	13:07
Air Blank	0.000	13:08
Control Test	0.078	13:08
Air Blank	0.000	13:09
Diagnostics Check	OK	13:09

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: REINHARDT,KYLE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BERLIN

DATE/TIME OF TEST: 12/11/2022 23:57:03

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.000	23:52
Breath Volume	1.824	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:55
Breath Volume	2.394	23:55
Air Blank	0.000	23:56
Control Test	0.077	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:57

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MIRANDA,J
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/12/2022 21:23:32

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.179	21:18
Breath Volume	2.476	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.180	21:21
Breath Volume	2.351	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DEFALCO,CHRIS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: NAEYAERT

DATE/TIME OF TEST: 12/15/2022 20:16:27

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.010	20:12
Breath Volume	3.507	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.008	20:15
Breath Volume	3.582	20:15
Air Blank	0.000	20:15
Control Test	0.076	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STAHRE,C
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: KUZNICKI

DATE/TIME OF TEST: 12/17/2022 02:11:39

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.079	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:10	Sequence Aborted
Breath Volume	0.660	02:10	
Air Blank	0.000	02:11	

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: REINHARDT,KYLE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SODERBERG

DATE/TIME OF TEST: 12/20/2022 03:51:40

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.238	03:47
Breath Volume	2.410	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.242	03:50
Breath Volume	2.570	03:50
Air Blank	0.000	03:50
Control Test	0.075	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY,MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RANNO

DATE/TIME OF TEST: 12/21/2022 03:42:13

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.118	03:35
Breath Volume	1.449	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.108	03:40
Breath Volume	1.546	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: YANICK,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 12/25/2022 03:12:54

Diagnostics Check	OK	03:11	
Air Blank		03:12	Ambient Fail
Air Blank	0.000	03:12	

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: YANICK,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 12/25/2022 03:22:01

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.151	03:17
Breath Volume	3.187	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.150	03:20
Breath Volume	2.871	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFLEUR, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: LAFLEUR, MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: DANGLER

DATE/TIME OF TEST: 12/27/2022 19:54:00

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.201	19:48
Breath Volume	1.910	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.198	19:52
Breath Volume	2.757	19:52
Air Blank	0.000	19:53
Control Test	0.078	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:54

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BENNETT,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/28/2022 21:27:41

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.078	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.175	21:22
Breath Volume	2.871	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.162	21:26
Breath Volume	1.617	21:26
Air Blank	0.000	21:26
Control Test	0.076	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 01/04/2023 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BLISS

DATE/TIME OF TEST: 12/26/2022 18:37:22

Diagnostics Check	OK	18:20	
Air Blank	0.000	18:20	
Control Test	0.081	18:21	
Air Blank	0.000	18:21	
Subject Sample	# 1 0.210	18:24	
Breath Volume	1.160	18:24	
Air Blank	0.000	18:25	
Air Blank	0.000	18:26	
Subject Sample	# 2 0.178	18:29	Volume Not Met
Breath Volume	0.109	18:29	
Air Blank	0.000	18:30	
Air Blank	0.000	18:32	
Subject Sample	# 3 0.169	18:35	Volume Not Met
Breath Volume	0.042	18:35	
Air Blank	0.000	18:36	
Control Test	0.073	18:36	Control Outside Tolerance
Air Blank	0.000	18:37	

Data Download Date/Time: 01/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BLISS

DATE/TIME OF TEST: 12/26/2022 18:49:03

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.078	18:41
Air Blank	0.000	18:41
Subject Sample # 1	0.178	18:42
Breath Volume	1.324	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.183	18:47
Breath Volume	1.414	18:47
Air Blank	0.000	18:48
Control Test	0.078	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:49

Data Download Date/Time: 01/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 12/27/2022 21:25:02

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.081	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.066	21:20
Breath Volume	3.304	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:22
Subject Sample # 2	0.063	21:23
Breath Volume	2.289	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:25

Data Download Date/Time: 01/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELASQUEZ,ARTURO,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MORALES PANTALEON

DATE/TIME OF TEST: 12/31/2022 00:07:15

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.181	00:02	
Breath Volume	1.273	00:02	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample	# 2 0.162	00:05	
Breath Volume	1.421	00:05	
Air Blank	0.000	00:06	
Control Test	0.056	00:06	Control Outside Tolerance
Air Blank	0.000	00:07	

Data Download Date/Time: 01/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: HUFTON

DATE/TIME OF TEST: 12/31/2022 01:09:21

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1 0.150	01:03	Volume Not Met
Breath Volume	0.750	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 2	01:07	Subject Test Refused
Breath Volume	0.250	01:07	
Air Blank	0.000	01:08	
Control Test	0.077	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 01/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ARRIAGA RODRIGUEZ

DATE/TIME OF TEST: 12/31/2022 02:40:29

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.078	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1 0.251	02:27	
Breath Volume	1.351	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 2 0.187	02:32	Volume Not Met
Breath Volume	0.062	02:32	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample	# 3 0.181	02:38	Volume Not Met
Breath Volume	0.718	02:38	
Air Blank	0.000	02:39	
Control Test	0.076	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 01/09/2023 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ARRIAGA RODRIGUEZ

DATE/TIME OF TEST: 12/31/2022 02:48:52

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.076	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.234	02:44
Breath Volume	2.191	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.233	02:47
Breath Volume	1.382	02:47
Air Blank	0.000	02:47
Control Test	0.077	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 01/09/2023 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STONE,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: NEESE

DATE/TIME OF TEST: 12/04/2022 06:40:11

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.080	06:34
Air Blank	0.000	06:34
Subject Sample # 1	0.238	06:35
Breath Volume	1.480	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:37
Subject Sample # 2	0.244	06:38
Breath Volume	1.472	06:38
Air Blank	0.000	06:39
Control Test	0.080	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 12/28/2022 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,J,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: ADAMS,S
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: COFFIN

DATE/TIME OF TEST: 12/06/2022 07:40:44

Diagnostics Check	OK	07:34
Air Blank	0.000	07:34
Control Test	0.080	07:34
Air Blank	0.000	07:35
Subject Sample # 1	0.000	07:36
Breath Volume	1.984	07:36
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample # 2	0.000	07:39
Breath Volume	1.851	07:39
Air Blank	0.000	07:39
Control Test	0.080	07:40
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

Data Download Date/Time: 12/28/2022 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: CAVINS

DATE/TIME OF TEST: 12/17/2022 07:57:10

Diagnostics Check	OK	07:50
Air Blank	0.000	07:50
Control Test	0.081	07:51
Air Blank	0.000	07:51
Subject Sample # 1	0.229	07:52
Breath Volume	2.101	07:52
Air Blank	0.000	07:53
Air Blank	0.000	07:54
Subject Sample # 2	0.233	07:55
Breath Volume	1.875	07:55
Air Blank	0.000	07:56
Control Test	0.080	07:56
Air Blank	0.000	07:57
Diagnostics Check	OK	07:57

Data Download Date/Time: 12/28/2022 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FLYNN,H
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 11/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HOLE

DATE/TIME OF TEST: 12/18/2022 22:13:15

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:05	
Control Test	0.081	22:05	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:09	No Sample Provided
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample	# 2	22:11	Subject Test Refused
Breath Volume	0.000	22:11	
Air Blank	0.000	22:12	
Control Test	0.081	22:12	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 12/28/2022 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD,RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 12/01/2022 00:43:07

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.122	00:38
Breath Volume	1.371	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.128	00:41
Breath Volume	1.871	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 12/01/2022 23:15:00

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1 0.153	23:09	Volume Not Met
Breath Volume	0.136	23:09	
Air Blank	0.000	23:10	
Air Blank	0.000	23:12	
Subject Sample	# 2	23:13	Subject Test Refused
Breath Volume	0.187	23:13	
Air Blank	0.000	23:14	
Control Test	0.077	23:14	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: KRSAK

DATE/TIME OF TEST: 12/02/2022 18:41:50

Diagnostics Check	OK	18:35
Air Blank	0.000	18:35
Control Test	0.080	18:35
Air Blank	0.000	18:36
Subject Sample # 1	0.260	18:37
Breath Volume	3.019	18:37
Air Blank	0.000	18:37
Air Blank	0.000	18:39
Subject Sample # 2	0.253	18:40
Breath Volume	2.906	18:40
Air Blank	0.000	18:40
Control Test	0.079	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES, ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: DOMINGO MENDEZ

DATE/TIME OF TEST: 12/03/2022 02:37:38

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.183	02:33
Breath Volume	2.656	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.187	02:36
Breath Volume	2.394	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FEINSTEIN

DATE/TIME OF TEST: 12/03/2022 04:45:40

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.156	04:40
Breath Volume	1.621	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.156	04:44
Breath Volume	1.242	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PHILIPPE, BENDJY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: FLORES ARIAS

DATE/TIME OF TEST: 12/03/2022 12:21:26

Diagnostics Check	OK	12:13
Air Blank	0.000	12:14
Control Test	0.079	12:14
Air Blank	0.000	12:15
Subject Sample # 1	0.211	12:16
Breath Volume	2.859	12:16
Air Blank	0.000	12:16
Air Blank	0.000	12:18
Subject Sample # 2	0.219	12:19
Breath Volume	2.183	12:19
Air Blank	0.000	12:20
Control Test	0.079	12:20
Air Blank	0.000	12:21
Diagnostics Check	OK	12:21

Data Download Date/Time: 12/22/2022 16:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: SALES JIMENEZ

DATE/TIME OF TEST: 12/04/2022 00:52:40

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.218	00:48
Breath Volume	1.539	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.220	00:51
Breath Volume	2.632	00:51
Air Blank	0.000	00:51
Control Test	0.079	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: GONZALEZ GONZALEZ

DATE/TIME OF TEST: 12/04/2022 04:01:21

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.225	03:56
Breath Volume	3.062	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.234	03:59
Breath Volume	2.988	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 12/04/2022 04:56:20

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.075	04:51
Breath Volume	1.363	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.075	04:54
Breath Volume	1.847	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: FARR

DATE/TIME OF TEST: 12/04/2022 22:35:24

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.126	22:30
Breath Volume	2.097	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.125	22:33
Breath Volume	1.511	22:33
Air Blank	0.000	22:34
Control Test	0.078	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GONZALEZ CASTRO

DATE/TIME OF TEST: 12/05/2022 00:43:35

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Subject Sample # 1	0.264	00:39
Breath Volume	2.054	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.252	00:41
Breath Volume	2.906	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COSTALES,ALEXANDRO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MANGALIE

DATE/TIME OF TEST: 12/05/2022 03:15:58

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.079	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1 0.116	03:06	Volume Not Met
Breath Volume	0.039	03:06	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2 0.115	03:10	
Breath Volume	1.210	03:10	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 3 0.116	03:14	
Breath Volume	1.269	03:14	
Air Blank	0.000	03:15	
Control Test	0.078	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/05/2022 23:37:24

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.000	23:33
Breath Volume	1.503	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:36
Breath Volume	1.132	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: THOMPSON JR

DATE/TIME OF TEST: 12/07/2022 04:09:43

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.000	04:05
Breath Volume	1.468	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.000	04:08
Breath Volume	1.175	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDRE, DOMINIC
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: NAZARIEGOS LOPEZ

DATE/TIME OF TEST: 12/07/2022 23:24:30

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.106	23:20
Breath Volume	1.589	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.108	23:22
Breath Volume	2.160	23:22
Air Blank	0.000	23:23
Control Test	0.078	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: CAYAX JUAREZ

DATE/TIME OF TEST: 12/09/2022 22:45:10

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.187	22:40
Breath Volume	1.656	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.186	22:43
Breath Volume	2.062	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BONILLA COREAS

DATE/TIME OF TEST: 12/10/2022 00:47:46

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.194	00:43
Breath Volume	3.460	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.191	00:46
Breath Volume	3.355	00:46
Air Blank	0.000	00:46
Control Test	0.078	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SEIER,BRADLEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: FORTMAN

DATE/TIME OF TEST: 12/10/2022 22:59:18

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.269	22:54
Breath Volume	1.785	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:56
Subject Sample # 2	0.272	22:57
Breath Volume	1.761	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,RYAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:53

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/11/2022 12:32:27

Diagnostics Check	OK	12:26
Air Blank	0.000	12:26
Control Test	0.080	12:27
Air Blank	0.000	12:27
Subject Sample # 1	0.111	12:27
Breath Volume	1.441	12:27
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 2	0.113	12:30
Breath Volume	1.695	12:30
Air Blank	0.000	12:31
Control Test	0.079	12:31
Air Blank	0.000	12:32
Diagnostics Check	OK	12:32

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JUAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JIMENEZ REYES

DATE/TIME OF TEST: 12/12/2022 01:32:27

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.167	01:27
Breath Volume	1.117	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.176	01:30
Breath Volume	1.320	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: OLVERA FLORES

DATE/TIME OF TEST: 12/12/2022 02:46:27

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.204	02:41
Breath Volume	2.871	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.188	02:44
Breath Volume	2.066	02:44
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: BOGAN

DATE/TIME OF TEST: 12/12/2022 10:06:34

Diagnostics Check	OK	09:58	
Air Blank	0.000	09:58	
Control Test	0.080	09:58	
Air Blank	0.000	09:59	
Subject Sample	# 1 0.000	10:02	Volume Not Met
Breath Volume	0.527	10:02	
Air Blank	0.000	10:03	
Air Blank	0.000	10:05	
Subject Sample	# 2	10:05	Subject Test Refused
Breath Volume	0.000	10:05	
Air Blank	0.000	10:05	
Control Test	0.079	10:05	
Air Blank	0.000	10:06	
Diagnostics Check	OK	10:06	

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SHORTELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: NOWACZYK

DATE/TIME OF TEST: 12/12/2022 22:45:05

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.179	22:40
Breath Volume	2.695	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:42
Subject Sample # 2	0.175	22:43
Breath Volume	2.894	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:45

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GWYNN

DATE/TIME OF TEST: 12/15/2022 01:55:32

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.124	01:47
Breath Volume	1.128	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.152	01:50
Breath Volume	2.402	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 3	0.150	01:53
Breath Volume	2.402	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: ORDONEZ PEREZ

DATE/TIME OF TEST: 12/16/2022 00:08:45

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.213	00:04
Breath Volume	1.550	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.218	00:07
Breath Volume	1.828	00:07
Air Blank	0.000	00:07
Control Test	0.078	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: TZUNUN TAX

DATE/TIME OF TEST: 12/16/2022 22:30:39

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.080	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.197	22:26
Breath Volume	2.371	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.201	22:29
Breath Volume	2.390	22:29
Air Blank	0.000	22:29
Control Test	0.079	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOVETT,JAMES
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/16/2022 23:44:29

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1 0.189	23:36	Volume Not Met
Breath Volume	0.804	23:36	
Air Blank	0.000	23:36	
Air Blank	0.000	23:38	
Subject Sample	# 2 0.197	23:39	
Breath Volume	1.882	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 3 0.190	23:42	
Breath Volume	1.449	23:42	
Air Blank	0.000	23:43	
Control Test	0.077	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: PEREZ LARIOS

DATE/TIME OF TEST: 12/17/2022 00:19:13

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.234	00:14
Breath Volume	1.531	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.239	00:17
Breath Volume	1.519	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: MARTIN JIMENEZ

DATE/TIME OF TEST: 12/17/2022 04:44:42

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.105	04:40
Breath Volume	2.218	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.106	04:43
Breath Volume	2.250	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 12/17/2022 07:17:38

Diagnostics Check	OK	07:09
Air Blank	0.000	07:10
Control Test	0.079	07:10
Air Blank	0.000	07:10
Subject Sample # 1	0.126	07:11
Breath Volume	1.359	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.131	07:16
Breath Volume	1.992	07:16
Air Blank	0.000	07:16
Control Test	0.079	07:17
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VIANUEVA,BRANDON
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: PASSOS

DATE/TIME OF TEST: 12/18/2022 23:30:08

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.201	23:25
Breath Volume	1.832	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.206	23:28
Breath Volume	2.476	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRANT,JEROME
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GOMEZ LOPEZ

DATE/TIME OF TEST: 12/19/2022 02:16:40

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.143	02:11
Breath Volume	1.457	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.149	02:15
Breath Volume	1.531	02:15
Air Blank	0.000	02:15
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SANDOVAL RAMIREZ

DATE/TIME OF TEST: 12/19/2022 03:32:10

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.176	03:27
Breath Volume	1.753	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.177	03:30
Breath Volume	1.152	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: RANDAZZO

DATE/TIME OF TEST: 12/19/2022 23:38:40

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.081	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1 0.209	23:34	Volume Not Met
Breath Volume	0.140	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2	23:37	Subject Test Refused
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Control Test	0.076	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: GONZALEZ MATUL

DATE/TIME OF TEST: 12/21/2022 22:38:19

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.257	22:33
Breath Volume	2.117	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.259	22:36
Breath Volume	2.148	22:36
Air Blank	0.000	22:37
Control Test	0.079	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BURKE,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 10:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:36

SUBJECT'S LAST NAME: MOERINGS

DATE/TIME OF TEST: 12/22/2022 12:13:58

Diagnostics Check	OK	12:07
Air Blank	0.000	12:07
Control Test	0.079	12:07
Air Blank	0.000	12:08
Subject Sample # 1	0.000	12:09
Breath Volume	1.386	12:09
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample # 2	0.000	12:12
Breath Volume	1.378	12:12
Air Blank	0.000	12:13
Control Test	0.079	12:13
Air Blank	0.000	12:13
Diagnostics Check	OK	12:13

Data Download Date/Time: 12/22/2022 16:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: GECK

DATE/TIME OF TEST: 12/01/2022 01:23:36

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.073	01:18
Breath Volume	3.574	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.068	01:22
Breath Volume	3.425	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/02/2022 02:04:37

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.126	02:00
Breath Volume	1.382	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.131	02:03
Breath Volume	1.656	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAYLES, RASHAD
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: SONTAY GONZALEZ

DATE/TIME OF TEST: 12/02/2022 06:27:43

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.079	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.216	06:23
Breath Volume	2.898	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.212	06:26
Breath Volume	2.992	06:26
Air Blank	0.000	06:26
Control Test	0.077	06:27
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: SOTO-PEREZ

DATE/TIME OF TEST: 12/03/2022 00:37:26

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.178	00:32
Breath Volume	3.320	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.176	00:35
Breath Volume	2.652	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 12/03/2022 02:31:28

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.178	02:26
Breath Volume	1.613	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.173	02:29
Breath Volume	1.726	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: REINHART

DATE/TIME OF TEST: 12/03/2022 04:17:48

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.000	04:13
Breath Volume	1.277	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.000	04:16
Breath Volume	1.281	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES, KERWIN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: THIMOT

DATE/TIME OF TEST: 12/03/2022 06:10:52

Diagnostics Check	OK	06:04
Air Blank	0.000	06:05
Control Test	0.079	06:05
Air Blank	0.000	06:05
Subject Sample # 1	0.120	06:06
Breath Volume	2.082	06:06
Air Blank	0.000	06:07
Air Blank	0.000	06:08
Subject Sample # 2	0.121	06:09
Breath Volume	2.031	06:09
Air Blank	0.000	06:09
Control Test	0.079	06:10
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MARTINEZ ORDONEZ

DATE/TIME OF TEST: 12/04/2022 23:12:33

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.079	23:00	
Air Blank	0.000	23:00	
Subject Sample	# 1 0.284	23:03	Volume Not Met
Breath Volume	0.906	23:03	
Air Blank	0.000	23:04	
Air Blank	0.000	23:06	
Subject Sample	# 2 0.285	23:07	
Breath Volume	2.179	23:07	
Air Blank	0.000	23:08	
Air Blank	0.000	23:10	
Subject Sample	# 3 0.291	23:10	
Breath Volume	1.332	23:10	
Air Blank	0.000	23:11	
Control Test	0.075	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SAVOIE

DATE/TIME OF TEST: 12/05/2022 04:18:03

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.215	04:10
Breath Volume	1.332	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.238	04:13
Breath Volume	2.378	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 3	0.243	04:16
Breath Volume	4.093	04:16
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: QUIROZ CARDONA

DATE/TIME OF TEST: 12/07/2022 03:21:03

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.235	03:16
Breath Volume	2.660	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.233	03:19
Breath Volume	1.917	03:19
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:21

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/07/2022 04:18:23

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.129	04:13
Breath Volume	1.308	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.132	04:16
Breath Volume	2.351	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: DOMINGO CINTO

DATE/TIME OF TEST: 12/07/2022 05:26:49

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.079	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.184	05:21
Breath Volume	1.984	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.175	05:25
Breath Volume	1.941	05:25
Air Blank	0.000	05:25
Control Test	0.078	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: GREENWALD-DERET

DATE/TIME OF TEST: 12/08/2022 01:56:57

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.090	01:52
Breath Volume	1.664	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.089	01:55
Breath Volume	2.226	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON, MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 12/09/2022 22:30:32

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.167	22:25
Breath Volume	1.687	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.162	22:28
Breath Volume	1.734	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 12/10/2022 00:40:35

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.161	00:35
Breath Volume	1.289	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.169	00:38
Breath Volume	1.816	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: PETELA

DATE/TIME OF TEST: 12/10/2022 05:35:28

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.206	05:30
Breath Volume	2.769	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.211	05:33
Breath Volume	2.914	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE, DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: ESTRELLA SURIEL

DATE/TIME OF TEST: 12/10/2022 07:00:04

Diagnostics Check	OK	06:53
Air Blank	0.000	06:53
Control Test	0.080	06:53
Air Blank	0.000	06:54
Subject Sample # 1	0.134	06:55
Breath Volume	2.199	06:55
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.128	06:58
Breath Volume	2.273	06:58
Air Blank	0.000	06:59
Control Test	0.079	06:59
Air Blank	0.000	06:59
Diagnostics Check	OK	07:00

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE, RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: RUTH

DATE/TIME OF TEST: 12/11/2022 00:02:59

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.161	23:57
Breath Volume	1.164	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.167	00:01
Breath Volume	1.546	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: FUENMAYOR CONTRERAS

DATE/TIME OF TEST: 12/11/2022 03:26:03

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.189	03:21
Breath Volume	2.742	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.186	03:24
Breath Volume	2.496	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:26

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ\],MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/11/2022 23:49:50

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.230	23:45
Breath Volume	2.902	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.223	23:48
Breath Volume	1.503	23:48
Air Blank	0.000	23:48
Control Test	0.078	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 12/12/2022 02:16:49

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.079	02:13	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MENDOZA DIAZ

DATE/TIME OF TEST: 12/16/2022 01:51:45

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.204	01:47
Breath Volume	1.558	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.221	01:50
Breath Volume	2.136	01:50
Air Blank	0.000	01:50
Control Test	0.077	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 12/16/2022 22:51:41

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.199	22:47
Breath Volume	3.179	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.196	22:50
Breath Volume	2.726	22:50
Air Blank	0.000	22:50
Control Test	0.078	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ORTEGA INTURIAS

DATE/TIME OF TEST: 12/17/2022 00:12:05

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.164	00:07
Breath Volume	3.777	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:09
Subject Sample # 2	0.161	00:10
Breath Volume	3.621	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: DIAZ ROBLERO

DATE/TIME OF TEST: 12/17/2022 00:56:59

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.200	00:52
Breath Volume	2.582	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.194	00:55
Breath Volume	2.695	00:55
Air Blank	0.000	00:56
Control Test	0.077	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ, ROBERTO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: MONROIG SANTIAGO

DATE/TIME OF TEST: 12/17/2022 03:46:12

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.105	03:41
Breath Volume	2.082	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.106	03:44
Breath Volume	2.046	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MORIARTY

DATE/TIME OF TEST: 12/18/2022 00:13:13

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.267	00:08
Breath Volume	2.035	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.268	00:11
Breath Volume	1.292	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: DOMINGO-ESCOBAR

DATE/TIME OF TEST: 12/18/2022 20:02:53

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.079	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.171	19:58
Breath Volume	2.156	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.160	20:01
Breath Volume	1.421	20:01
Air Blank	0.000	20:01
Control Test	0.079	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FERNANDES CUNHA

DATE/TIME OF TEST: 12/18/2022 20:24:45

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.115	20:20
Breath Volume	1.773	20:20
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.113	20:23
Breath Volume	2.132	20:23
Air Blank	0.000	20:23
Control Test	0.079	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: PEREZ JUAREZ

DATE/TIME OF TEST: 12/18/2022 23:36:07

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.263	23:31
Breath Volume	1.960	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.260	23:34
Breath Volume	1.976	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 12/19/2022 00:55:16

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.031	00:50
Breath Volume	1.781	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.031	00:53
Breath Volume	1.902	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: KAIRALLA

DATE/TIME OF TEST: 12/19/2022 02:54:10

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.054	02:49
Breath Volume	3.121	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.052	02:52
Breath Volume	2.566	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: PEREZ VASQUEZ

DATE/TIME OF TEST: 12/21/2022 20:24:18

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.163	20:19
Breath Volume	2.636	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:21
Subject Sample # 2	0.168	20:22
Breath Volume	2.785	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JEANBAPTISTE,SAMUEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: KOSKUBA

DATE/TIME OF TEST: 12/21/2022 23:32:26

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:28
Breath Volume	3.800	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.000	23:31
Breath Volume	3.902	23:31
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ, JUAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: CUEVA PAUTA

DATE/TIME OF TEST: 12/22/2022 00:22:40

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.159	00:18
Breath Volume	1.933	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.157	00:21
Breath Volume	1.535	00:21
Air Blank	0.000	00:21
Control Test	0.077	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 12/22/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: CAVALLINI RAMIREZ

DATE/TIME OF TEST: 12/22/2022 17:09:24

Diagnostics Check	OK	17:02
Air Blank	0.000	17:03
Control Test	0.079	17:03
Air Blank	0.000	17:04
Subject Sample # 1	0.136	17:04
Breath Volume	2.761	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.139	17:07
Breath Volume	2.574	17:07
Air Blank	0.000	17:08
Control Test	0.078	17:08
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SEMANAT PEREZ

DATE/TIME OF TEST: 12/23/2022 00:02:40

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.121	23:58
Breath Volume	1.992	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.121	00:01
Breath Volume	1.804	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: MIRANDA HERNANDEZ

DATE/TIME OF TEST: 12/23/2022 01:09:11

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.167	01:04
Breath Volume	2.093	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.171	01:07
Breath Volume	2.808	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/23/2022 22:26:08

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.158	22:21
Breath Volume	1.933	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.163	22:24
Breath Volume	1.468	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: LAPITSKIY

DATE/TIME OF TEST: 12/23/2022 22:41:26

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.071	22:36
Breath Volume	1.929	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.073	22:39
Breath Volume	2.359	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 12/24/2022 22:19:52

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.253	22:14
Breath Volume	1.882	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.241	22:18
Breath Volume	1.398	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: HABANSKY

DATE/TIME OF TEST: 12/25/2022 00:11:18

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.119	00:06
Breath Volume	1.324	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.125	00:09
Breath Volume	2.339	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MAIDA

DATE/TIME OF TEST: 12/25/2022 01:18:23

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.026	01:13
Breath Volume	1.726	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.025	01:16
Breath Volume	1.386	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VASQUEZ MENDEZ

DATE/TIME OF TEST: 12/25/2022 02:47:45

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.160	02:43
Breath Volume	3.015	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.164	02:46
Breath Volume	3.128	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/25/2022 06:04:24

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.081	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.171	05:59
Breath Volume	1.148	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.173	06:02
Breath Volume	2.414	06:02
Air Blank	0.000	06:03
Control Test	0.082	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: LINDSTROM

DATE/TIME OF TEST: 12/26/2022 00:04:27

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.202	23:59
Breath Volume	1.429	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.203	00:02
Breath Volume	2.714	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: RAMOS ORDONEZ

DATE/TIME OF TEST: 12/29/2022 01:03:36

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.129	00:58
Breath Volume	1.992	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.127	01:02
Breath Volume	2.421	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: AGUILERA

DATE/TIME OF TEST: 12/29/2022 23:23:33

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.176	23:19
Breath Volume	3.828	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.171	23:21
Breath Volume	3.820	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/30/2022 01:56:14

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.208	01:51
Breath Volume	2.527	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.203	01:54
Breath Volume	2.503	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: STROGONOVA

DATE/TIME OF TEST: 12/31/2022 01:16:38

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.163	01:11
Breath Volume	1.117	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.155	01:15
Breath Volume	1.281	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MORALES OCHOA

DATE/TIME OF TEST: 12/31/2022 03:22:09

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.144	03:17
Breath Volume	1.972	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.141	03:20
Breath Volume	2.285	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ZACARIAS-FIGUEROA

DATE/TIME OF TEST: 12/31/2022 03:36:22

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.228	03:31
Breath Volume	2.480	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.223	03:34
Breath Volume	2.441	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCLUSKEY,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: VIRULA

DATE/TIME OF TEST: 12/31/2022 21:19:52

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.241	21:14
Breath Volume	1.464	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.236	21:18
Breath Volume	1.769	21:18
Air Blank	0.000	21:18
Control Test	0.079	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOLLIGAN,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 12/31/2022 21:54:40

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.124	21:49
Breath Volume	1.859	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.120	21:53
Breath Volume	1.531	21:53
Air Blank	0.000	21:53
Control Test	0.078	21:54
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 01/13/2023 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 12/01/2022 23:04:44

Diagnostics Check	OK	23:03	
Air Blank		23:04	Sequence Aborted
Air Blank	0.000	23:04	

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 12/01/2022 23:18:03

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.079	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.000	23:13
Breath Volume	2.109	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.000	23:16
Breath Volume	2.210	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:18

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: SHIPP

DATE/TIME OF TEST: 12/10/2022 20:13:39

Diagnostics Check	OK	20:06
Air Blank	0.000	20:07
Control Test	0.079	20:07
Air Blank	0.000	20:07
Subject Sample # 1	0.150	20:08
Breath Volume	2.152	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.148	20:12
Breath Volume	2.609	20:12
Air Blank	0.000	20:12
Control Test	0.078	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CLENDANIEL

DATE/TIME OF TEST: 12/11/2022 02:45:58

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.146	02:41
Breath Volume	2.945	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.141	02:44
Breath Volume	2.464	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 12/11/2022 23:23:11

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.000	23:18
Breath Volume	2.117	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:21
Breath Volume	2.890	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SAMILE

DATE/TIME OF TEST: 12/12/2022 03:22:02

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.081	03:17
Breath Volume	3.167	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.077	03:20
Breath Volume	2.527	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: NEEDHAM

DATE/TIME OF TEST: 12/14/2022 23:25:31

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.078	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.132	23:20
Breath Volume	1.765	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.133	23:23
Breath Volume	2.015	23:23
Air Blank	0.000	23:24
Control Test	0.076	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROCKWELL

DATE/TIME OF TEST: 12/15/2022 03:36:52

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.120	03:32
Breath Volume	1.152	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.114	03:35
Breath Volume	1.441	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/20/2022 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 12/24/2022 01:08:14

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.083	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.169	01:03
Breath Volume	1.800	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.172	01:06
Breath Volume	2.296	01:06
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 01/25/2023 18:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: THOUBBORON

DATE/TIME OF TEST: 12/29/2022 01:51:50

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.079	01:51	
Air Blank		01:51	RFI Detect
Air Blank	0.000	01:51	

Data Download Date/Time: 01/25/2023 18:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: THOUBBORON

DATE/TIME OF TEST: 12/29/2022 01:56:44

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:54	
Control Test	0.079	01:54	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:55	RFI Detect
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Control Test	0.078	01:56	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 01/25/2023 18:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: THOUBBORON

DATE/TIME OF TEST: 12/29/2022 02:00:25

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.079	01:59	
Air Blank	0.000	01:59	
		02:00	RFI Detect
Air Blank		02:00	RFI Detect

Data Download Date/Time: 01/25/2023 18:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: THOUBBORON

DATE/TIME OF TEST: 12/29/2022 02:03:25

Diagnostics Check	OK	02:03	
Air Blank		02:03	RFI Detect
Air Blank		02:03	RFI Detect

Data Download Date/Time: 01/25/2023 18:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 311548
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: THOUBBORON

DATE/TIME OF TEST: 12/29/2022 02:07:01

Diagnostics Check	OK	02:06	
Air Blank		02:06	RFI Detect
Air Blank	0.000	02:07	

Data Download Date/Time: 01/25/2023 18:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KAPTEN,NALELL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PFEIFER

DATE/TIME OF TEST: 12/07/2022 23:22:52

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.083	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.158	23:18
Breath Volume	1.660	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.159	23:21
Breath Volume	1.949	23:21
Air Blank	0.000	23:21
Control Test	0.083	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 12/30/2022 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 12/08/2022 01:37:47

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.083	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.141	01:32
Breath Volume	2.015	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.141	01:36
Breath Volume	1.542	01:36
Air Blank	0.000	01:36
Control Test	0.083	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 12/30/2022 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,TYLER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CRUZ RAMOS

DATE/TIME OF TEST: 12/18/2022 01:03:01

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.083	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.081	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 12/30/2022 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANCY,JARRELL,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: RIPLEY,DALTON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 12/13/2022 20:11:07

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
		20:10	RFI Detect
Air Blank		20:11	Improper Sample

Data Download Date/Time: 12/29/2022 00:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANCY,JARRELL,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: RIPLEY,DALTON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 12/13/2022 20:15:04

Diagnostics Check	OK	20:14	
Air Blank		20:14	Improper Sample
Air Blank		20:15	Improper Sample

Data Download Date/Time: 12/29/2022 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANCY,JARRELL,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: RIPLEY,DALTON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 12/13/2022 20:19:46

Diagnostics Check	OK	20:18	
Air Blank		20:19	Ambient Fail
Air Blank	0.000	20:19	

Data Download Date/Time: 12/29/2022 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHOTT,MATTHEW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MARTINEZ VARELA

DATE/TIME OF TEST: 12/18/2022 00:45:23

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.212	00:40
Breath Volume	2.480	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 2	0.212	00:43
Breath Volume	2.660	00:43
Air Blank	0.000	00:44
Control Test	0.078	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 12/29/2022 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUTTERLY,BROOKE,L
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHUTTERLY,BROOKE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/26/2022 04:02:32

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.178	03:58
Breath Volume	1.625	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.174	04:00
Breath Volume	1.601	04:00
Air Blank	0.000	04:01
Control Test	0.076	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 12/29/2022 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: VIZCAYA SANCHEZ

DATE/TIME OF TEST: 12/04/2022 07:04:41

Diagnostics Check	OK	06:58
Air Blank	0.000	06:58
Control Test	0.080	06:59
Air Blank	0.000	06:59
Subject Sample # 1	0.175	07:00
Breath Volume	2.878	07:00
Air Blank	0.000	07:01
Air Blank	0.000	07:02
Subject Sample # 2	0.173	07:03
Breath Volume	1.121	07:03
Air Blank	0.000	07:03
Control Test	0.077	07:04
Air Blank	0.000	07:04
Diagnostics Check	OK	07:04

Data Download Date/Time: 12/06/2022 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/05/2022 20:22:12

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.079	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.224	20:17
Breath Volume	2.566	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.231	20:20
Breath Volume	2.500	20:20
Air Blank	0.000	20:21
Control Test	0.076	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 12/06/2022 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ACEVEDO,JEILANI
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SOZA-CASTILLO

DATE/TIME OF TEST: 12/12/2022 02:45:23

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.148	02:37
Breath Volume	1.593	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.124	02:40
Breath Volume	1.460	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 3	0.139	02:43
Breath Volume	2.308	02:43
Air Blank	0.000	02:44
Control Test	0.075	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 01/17/2023 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: CARPENTIER

DATE/TIME OF TEST: 12/18/2022 03:24:34

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.186	03:19
Breath Volume	2.195	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.181	03:23
Breath Volume	2.253	03:23
Air Blank	0.000	03:23
Control Test	0.075	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 01/17/2023 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEIB,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: BARTHOLOMEW

DATE/TIME OF TEST: 12/19/2022 18:56:51

Diagnostics Check	OK	18:50
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:51
Subject Sample # 1	0.118	18:52
Breath Volume	1.511	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:54
Subject Sample # 2	0.116	18:55
Breath Volume	1.804	18:55
Air Blank	0.000	18:55
Control Test	0.075	18:56
Air Blank	0.000	18:56
Diagnostics Check	OK	18:56

Data Download Date/Time: 01/17/2023 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN JACQUES,WEDLET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 12/01/2022 00:40:23

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.079	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: KELMANN

DATE/TIME OF TEST: 12/04/2022 22:03:22

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.079	22:01	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:02	Subject Test Refused
Breath Volume	0.000	22:02	
Air Blank	0.000	22:02	
Control Test	0.079	22:02	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: BOBO,GILBERT
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 12/06/2022 23:01:27

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.035	22:57
Breath Volume	2.562	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.036	22:59
Breath Volume	2.585	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGE,DANIEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: DIAZ,JESSICA
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: THIELE

DATE/TIME OF TEST: 12/07/2022 21:40:58

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.078	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Sequence Aborted
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGE,DANIEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: DIAZ,JESSICA
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: THIELE

DATE/TIME OF TEST: 12/07/2022 22:02:55

Diagnostics Check	OK	21:59	
Air Blank	0.000	22:00	
Control Test	0.079	22:00	
Air Blank	0.000	22:01	
Subject Sample	# 1	22:02	Sequence Aborted
Breath Volume	0.000	22:02	
Air Blank	0.000	22:02	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGE,DANIEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: DIAZ,JESSICA
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: THIELE

DATE/TIME OF TEST: 12/07/2022 22:48:10

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.076	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:46	Subject Test Refused
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	
Control Test	0.075	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ELMORE,BENJAMIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GINTHER

DATE/TIME OF TEST: 12/10/2022 03:26:37

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.079	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CRUZET

DATE/TIME OF TEST: 12/10/2022 22:56:52

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.155	22:52
Breath Volume	1.609	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.159	22:55
Breath Volume	1.679	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONACO,ASHLEY,L
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MONACO,ASHLEY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/11/2022 00:49:35

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.165	00:44
Breath Volume	3.039	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.157	00:48
Breath Volume	2.027	00:48
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ, ANDRES,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARCIA, JULIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: NODA FUENTES

DATE/TIME OF TEST: 12/11/2022 21:30:14

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.078	21:28	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:29	
Control Test	0.078	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MATTHEWS,MARRY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 11/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: LABRADA

DATE/TIME OF TEST: 12/12/2022 00:32:24

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.078	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Control Test	0.078	00:31	
Air Blank		00:31	RFI Detect
Air Blank	0.000	00:32	

Data Download Date/Time: 12/15/2022 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: BOBO,GILBERT
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/16/2022 02:05:17

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.078	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Control Test	0.078	02:04	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 01/12/2023 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: WARE ROGERS,BRANDON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: GUTIERREZ JIMENEZ

DATE/TIME OF TEST: 12/22/2022 21:02:51

Diagnostics Check	OK	20:59	
Air Blank	0.000	21:00	
Control Test	0.079	21:00	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.079	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 01/12/2023 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARCIA,JULIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: FERRARO

DATE/TIME OF TEST: 12/24/2022 00:54:21

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:53	
Control Test	0.079	00:53	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 01/12/2023 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 12/28/2022 07:01:13

Diagnostics Check	OK	06:58	
Air Blank	0.000	06:58	
Control Test	0.079	06:59	
Air Blank	0.000	06:59	
Subject Sample	# 1	06:59	RFI Detect
Breath Volume	0.000	06:59	
Air Blank	0.000	07:00	
Control Test	0.079	07:00	
Air Blank	0.000	07:01	
Diagnostics Check	OK	07:01	

Data Download Date/Time: 01/12/2023 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,NERELYS,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA,NERELYS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: HAMM

DATE/TIME OF TEST: 12/28/2022 07:10:15

Diagnostics Check	OK	07:03
Air Blank	0.000	07:03
Control Test	0.080	07:04
Air Blank	0.000	07:04
Subject Sample # 1	0.099	07:05
Breath Volume	1.625	07:05
Air Blank	0.000	07:06
Air Blank	0.000	07:07
Subject Sample # 2	0.109	07:08
Breath Volume	2.660	07:08
Air Blank	0.000	07:09
Control Test	0.080	07:09
Air Blank	0.000	07:10
Diagnostics Check	OK	07:10

Data Download Date/Time: 01/12/2023 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: HERNANDEZ ECHEVERIA

DATE/TIME OF TEST: 12/28/2022 22:06:01

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:05	
Control Test	0.079	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 01/12/2023 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLINA,IGNACIO,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: OKEEFE,VAUGHN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 13:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: RAMOS GARCIA

DATE/TIME OF TEST: 12/30/2022 13:52:20

Diagnostics Check	OK	13:45
Air Blank	0.000	13:46
Control Test	0.078	13:46
Air Blank	0.000	13:47
Subject Sample # 1	0.208	13:47
Breath Volume	2.152	13:47
Air Blank	0.000	13:48
Air Blank	0.000	13:50
Subject Sample # 2	0.212	13:50
Breath Volume	2.496	13:50
Air Blank	0.000	13:51
Control Test	0.078	13:51
Air Blank	0.000	13:52
Diagnostics Check	OK	13:52

Data Download Date/Time: 01/12/2023 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,LANCE,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,LANCE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 12/31/2022 16:00:02

Diagnostics Check	OK	15:53
Air Blank	0.000	15:53
Control Test	0.078	15:54
Air Blank	0.000	15:54
Subject Sample # 1	0.134	15:55
Breath Volume	1.636	15:55
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.139	15:58
Breath Volume	2.078	15:58
Air Blank	0.000	15:59
Control Test	0.078	15:59
Air Blank	0.000	15:59
Diagnostics Check	OK	16:00

Data Download Date/Time: 01/12/2023 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 12/31/2022 19:16:04

Diagnostics Check	OK	19:04
Air Blank	0.000	19:05
Control Test	0.079	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.339	19:07
Breath Volume	1.804	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:09
Subject Sample # 2	0.312	19:11
Breath Volume	1.562	19:11
Air Blank	0.000	19:12
Air Blank	0.000	19:13
Subject Sample # 3	0.313	19:14
Breath Volume	1.535	19:14
Air Blank	0.000	19:15
Control Test	0.076	19:15
Air Blank	0.000	19:15
Diagnostics Check	OK	19:16

Data Download Date/Time: 01/12/2023 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,ROBERT,A
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WHITE,NICHEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 11/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: PEREZCARRILLO

DATE/TIME OF TEST: 12/06/2022 01:16:33

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 12/07/2022 18:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: CABRERA CORREA

DATE/TIME OF TEST: 12/16/2022 04:10:49

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.097	04:06
Breath Volume	3.511	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.100	04:09
Breath Volume	3.523	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 01/30/2023 16:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: REMON JAMES

DATE/TIME OF TEST: 12/18/2022 22:21:19

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.080	22:19	
Air Blank	0.000	22:19	
Subject Sample	# 1	22:19	Subject Test Refused
Breath Volume	0.000	22:19	
Air Blank	0.000	22:20	
Control Test	0.081	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 01/30/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FUENTES PABLO

DATE/TIME OF TEST: 12/19/2022 03:44:48

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.127	03:39
Breath Volume	1.875	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.126	03:43
Breath Volume	2.617	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 01/30/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ROMERO ZELEDON

DATE/TIME OF TEST: 12/24/2022 03:28:47

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.187	03:23
Breath Volume	2.726	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.170	03:27
Breath Volume	1.609	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 01/30/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: DE PAZ

DATE/TIME OF TEST: 12/25/2022 02:27:58

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.083	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Control Test	0.083	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 01/30/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PHIFER

DATE/TIME OF TEST: 12/26/2022 02:54:09

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.083	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.250	02:48
Breath Volume	2.414	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.233	02:52
Breath Volume	1.195	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 01/30/2023 16:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/30/2022 03:21:32

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.077	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.180	03:16
Breath Volume	1.238	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.168	03:19
Breath Volume	1.605	03:19
Air Blank	0.000	03:20
Control Test	0.076	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 01/04/2023 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FARDELLA, BLAKE
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 12/23/2022 00:27:02

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.169	00:21
Breath Volume	1.343	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.170	00:25
Breath Volume	2.480	00:25
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:27

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PARENT,AUSTIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: LEVITON

DATE/TIME OF TEST: 12/23/2022 22:26:26

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1 0.000	22:18	Volume Not Met
Breath Volume	0.878	22:18	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2 0.000	22:21	
Breath Volume	1.390	22:21	
Air Blank	0.000	22:22	
Air Blank	0.000	22:24	
Subject Sample	# 3 0.000	22:25	
Breath Volume	1.312	22:25	
Air Blank	0.000	22:25	
Control Test	0.080	22:25	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 12/23/2022 23:04:05

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.156	22:59
Breath Volume	1.222	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.143	23:02
Breath Volume	1.640	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: MERISEME

DATE/TIME OF TEST: 12/24/2022 22:54:15

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.082	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.244	22:48
Breath Volume	1.226	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.245	22:52
Breath Volume	1.738	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: MORENO HERNANDEZ

DATE/TIME OF TEST: 12/25/2022 21:45:17

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.082	21:41	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:43	Subject Test Refused
Breath Volume	0.210	21:43	
Air Blank	0.000	21:44	
Control Test	0.082	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/26/2022 00:16:16

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.174	00:11
Breath Volume	2.492	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.171	00:14
Breath Volume	3.292	00:14
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,P
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 12/26/2022 03:15:18

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.081	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.199	03:10
Breath Volume	2.101	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.181	03:13
Breath Volume	1.226	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BUTZBACH,WILLIAM
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: TRUBCHIKOV

DATE/TIME OF TEST: 12/26/2022 17:46:52

Diagnostics Check	OK	17:39
Air Blank	0.000	17:40
Control Test	0.080	17:40
Air Blank	0.000	17:41
Subject Sample # 1	0.203	17:41
Breath Volume	1.898	17:41
Air Blank	0.000	17:42
Air Blank	0.000	17:43
Subject Sample # 2	0.215	17:45
Breath Volume	2.421	17:45
Air Blank	0.000	17:45
Control Test	0.079	17:46
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: CENELIEN

DATE/TIME OF TEST: 12/26/2022 20:37:26

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.248	20:30
Breath Volume	1.386	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.240	20:35
Breath Volume	3.320	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CAJUSTE, AIRD
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/29/2022 23:14:27

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.081	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.163	23:09
Breath Volume	1.519	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.157	23:12
Breath Volume	1.421	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ, ISIAH
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: KIRCHBAUM

DATE/TIME OF TEST: 12/29/2022 23:58:08

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.188	23:53
Breath Volume	3.164	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.191	23:56
Breath Volume	3.144	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KEEL, SHELDON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: VELASQUEZ PEREZ

DATE/TIME OF TEST: 12/31/2022 02:10:05

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.192	02:05
Breath Volume	3.132	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.177	02:08
Breath Volume	2.925	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 01/13/2023 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES, FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ACUNA HERNANDEZ

DATE/TIME OF TEST: 12/31/2022 04:15:10

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1 0.095	04:10	Volume Not Met
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:13	Subject Test Refused
Breath Volume	0.082	04:13	
Air Blank	0.000	04:14	
Control Test	0.078	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 01/13/2023 13:02