

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DOWLING

DATE/TIME OF TEST: 02/03/2022 01:41:55

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	1.871	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:40
Breath Volume	2.546	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,LYLE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BURKE,LYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/05/2022 03:14:17

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	RFI Detect
Air Blank	0.000	03:14	

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,LYLE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BURKE,LYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/05/2022 03:23:49

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.216	03:18
Breath Volume	1.402	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.212	03:22
Breath Volume	1.625	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HIGH

DATE/TIME OF TEST: 02/05/2022 23:17:21

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Subject Sample # 1	0.115	23:12	
Breath Volume	1.824	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample # 2		23:15	Interferent Detect
Breath Volume	1.886	23:15	
Air Blank	0.000	23:16	
Control Test	0.082	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HIGH

DATE/TIME OF TEST: 02/05/2022 23:29:14

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.078	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.123	23:24
Breath Volume	2.066	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample # 2	0.116	23:27
Breath Volume	1.726	23:27
Air Blank	0.000	23:28
Control Test	0.082	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TATEM

DATE/TIME OF TEST: 02/10/2022 20:27:57

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.080	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.206	20:23
Breath Volume	2.457	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.210	20:26
Breath Volume	1.402	20:26
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/12/2022 01:19:56

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.128	01:15
Breath Volume	3.503	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:17
Subject Sample # 2	0.127	01:18
Breath Volume	3.605	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 02/18/2022 00:06:59

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.077	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.151	00:02
Breath Volume	1.886	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.148	00:05
Breath Volume	1.355	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 02/19/2022 00:56:56

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.126	00:52
Breath Volume	2.511	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.120	00:55
Breath Volume	1.882	00:55
Air Blank	0.000	00:55
Control Test	0.078	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 02/19/2022 01:56:48

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.078	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:55	Interferent Detect
Breath Volume	1.667	01:55	
Air Blank	0.000	01:55	
Control Test	0.084	01:56	Interferent Detect
Air Blank	0.000	01:56	

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 02/19/2022 02:01:44

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.078	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Interferent Detect
Breath Volume	2.140	02:00	
Air Blank	0.000	02:00	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 02/19/2022 02:33:39

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.082	02:28
Breath Volume	1.738	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.077	02:32
Breath Volume	1.175	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SAYLOR

DATE/TIME OF TEST: 02/24/2022 00:40:14

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.000	00:36
Breath Volume	3.511	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.000	00:38
Breath Volume	2.570	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 02/25/2022 21:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: BLOSSE

DATE/TIME OF TEST: 02/24/2022 01:28:40

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.205	01:24
Breath Volume	4.171	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.217	01:27
Breath Volume	3.554	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 02/25/2022 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MCCONNAUGHEY

DATE/TIME OF TEST: 02/25/2022 23:33:59

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.281	23:29
Breath Volume	1.660	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.283	23:32
Breath Volume	1.742	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 04/04/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HINES,R
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 02/26/2022 03:04:59

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.115	03:00
Breath Volume	1.460	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.110	03:03
Breath Volume	1.570	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/04/2022 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 02/26/2022 03:17:17

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.109	03:12
Breath Volume	1.980	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.099	03:15
Breath Volume	1.648	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 04/04/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MACHIN

DATE/TIME OF TEST: 02/26/2022 23:00:59

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.119	22:56
Breath Volume	3.437	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 2	0.116	22:59
Breath Volume	3.765	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 04/04/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 193173
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 02/27/2022 02:20:22

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.182	02:16
Breath Volume	2.074	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.184	02:18
Breath Volume	1.675	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 04/04/2022 15:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGLE, TONY,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: LEARY, DANTE
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: DOPART

DATE/TIME OF TEST: 02/03/2022 00:11:11

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.082	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 02/28/2022 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BERNAL,ANTHONY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DEGRAFF

DATE/TIME OF TEST: 02/05/2022 04:00:27

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.082	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.236	03:55
Breath Volume	2.304	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.235	03:58
Breath Volume	2.566	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 02/28/2022 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FLORES GALLEGOS

DATE/TIME OF TEST: 02/09/2022 23:29:38

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.082	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.082	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 02/28/2022 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARRIS,MICHAEL,W

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: JAMES,ROYCE

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 22:35

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FLORES GALLEGOS

DATE/TIME OF TEST: 02/09/2022 23:36:24

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.082	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Control Test	0.082	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 02/28/2022 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 02/03/2022 00:12:58

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.206	00:07
Breath Volume	3.113	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.195	00:11
Breath Volume	1.199	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 02/14/2022 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, HECTOR, A
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: VILLANUEVA, HECTOR
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GONZALEZ CHOYX

DATE/TIME OF TEST: 02/03/2022 21:12:14

Diagnostics Check	OK	21:11	
Air Blank		21:12	RFI Detect
Air Blank		21:12	RFI Detect

Data Download Date/Time: 02/14/2022 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, HECTOR, A
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: VILLANUEVA, HECTOR
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GONZALEZ CHOYX

DATE/TIME OF TEST: 02/03/2022 21:20:11

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:16	
Control Test	0.081	21:16	
Air Blank	0.000	21:17	
Subject Sample # 1	0.102	21:18	
Breath Volume	4.367	21:18	
Air Blank		21:18	RFI Detect
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 02/14/2022 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, HECTOR, A
BREATH TEST OPERATOR AGENCY: FHP TROOP F
ARRESTING OFFICER: VILLANUEVA, HECTOR
ARRESTING OFFICER AGENCY: FHP TROOP F
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GONZALEZ CHOYX

DATE/TIME OF TEST: 02/03/2022 21:32:06

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.101	21:27
Breath Volume	3.785	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.096	21:30
Breath Volume	3.230	21:30
Air Blank	0.000	21:31
Control Test	0.081	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 02/14/2022 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: FARIA CASTRO

DATE/TIME OF TEST: 02/12/2022 00:38:12

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.157	00:33
Breath Volume	1.371	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:35
Subject Sample # 2	0.157	00:36
Breath Volume	1.449	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 02/14/2022 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRILL,JACOB,M
BREATH TEST OPERATOR AGENCY: MANATEE COUNTY SO
ARRESTING OFFICER: ESSEK,JAKE
ARRESTING OFFICER AGENCY: MANATEE COUNTY SO
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211933
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BOWLES

DATE/TIME OF TEST: 02/13/2022 01:00:58

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1 0.202	00:55	
Breath Volume	2.750	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 2	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.080	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 02/14/2022 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HUNTER,RACHEL,R

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: DAVIS,CARL

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1283438

DRY GAS EXPIRY DATE: 02/25/2023

VIOLATION CODE: DUI

ARREST TIME: 21:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: FOERDE

DATE/TIME OF TEST: 02/06/2022 23:12:05

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.150	23:07
Breath Volume	1.636	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:09
Subject Sample # 2	0.152	23:10
Breath Volume	2.117	23:10
Air Blank	0.000	23:11
Control Test	0.082	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 02/21/2022 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: BROWN-WILSON,TREVON
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: ADMIN
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: SEGRAVES

DATE/TIME OF TEST: 02/20/2022 00:44:06

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.083	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.189	00:39
Breath Volume	4.832	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.196	00:42
Breath Volume	5.214	00:42
Air Blank	0.000	00:43
Control Test	0.083	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 02/21/2022 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/08/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 02/12/2022 19:57:25

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:55	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:56	Subject Test Refused
Breath Volume	0.000	19:56	
Air Blank	0.000	19:56	
Control Test	0.081	19:56	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:57	

Data Download Date/Time: 03/29/2022 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LENIHAN,BRIAN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 02/12/2022 20:03:58

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.080	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:03	

Data Download Date/Time: 03/29/2022 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,ROBERT,B
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LABRIE,DAVID
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 02/20/2022 10:46:52

Diagnostics Check	OK	10:44	
Air Blank	0.000	10:44	
Control Test	0.080	10:44	
Air Blank	0.000	10:45	
Subject Sample	# 1	10:45	Subject Test Refused
Breath Volume	0.000	10:45	
Air Blank	0.000	10:45	
Control Test	0.080	10:46	
Air Blank	0.000	10:46	
Diagnostics Check	OK	10:46	

Data Download Date/Time: 03/29/2022 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: CLINE,JEREMY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: SALISBURY

DATE/TIME OF TEST: 02/25/2022 23:39:43

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.160	23:35
Breath Volume	1.742	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.166	23:38
Breath Volume	2.386	23:38
Air Blank	0.000	23:38
Control Test	0.081	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 03/29/2022 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REASONER,DEBBIE,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,ALEX
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1149031
DRY GAS EXPIRY DATE: 05/30/2022
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 02/04/2022 00:00:48

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.051	23:56
Breath Volume	2.492	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.050	23:59
Breath Volume	2.406	23:59
Air Blank	0.000	23:59
Control Test	0.080	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 02/10/2022 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KRUSE, KRISTOPHER, D

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LAWLES, BILLY

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 00:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 02/07/2022 01:22:31

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.000	01:18
Breath Volume	3.386	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:21
Breath Volume	1.839	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 02/10/2022 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,JIMMIE,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MORALES,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1149031

DRY GAS EXPIRY DATE: 05/30/2022

VIOLATION CODE: DUI

ARREST TIME: 20:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/09/2022 21:20:36

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.081	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.208	21:16
Breath Volume	2.261	21:16
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.211	21:19
Breath Volume	2.117	21:19
Air Blank	0.000	21:19
Control Test	0.081	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 02/10/2022 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REASONER,DEBBIE,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,ALEX
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/11/2022 23:06:56

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.111	23:02
Breath Volume	2.531	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.101	23:05
Breath Volume	3.261	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 03/03/2022 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GUY,KEVIN,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: KRUSE,KRIS

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 03:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: TAYLOR-HODGE

DATE/TIME OF TEST: 02/19/2022 03:47:19

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.000	03:42
Breath Volume	1.558	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.000	03:45
Breath Volume	2.175	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/03/2022 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GUY,KEVIN,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MORALES,KAYLA

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 02422080A1

DRY GAS EXPIRY DATE: 02/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 02/20/2022 00:28:39

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.081	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.000	00:24
Breath Volume	1.117	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.000	00:27
Breath Volume	1.285	00:27
Air Blank	0.000	00:27
Control Test	0.081	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 03/03/2022 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,RUSSEL,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WHITEHEAD,K
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: BLYDENBURGH

DATE/TIME OF TEST: 02/06/2022 16:33:46

Diagnostics Check	OK	16:22	
Air Blank	0.000	16:22	
Control Test	0.080	16:23	
Air Blank	0.000	16:23	
Subject Sample	# 1	16:26	No Sample Provided
Breath Volume	0.000	16:26	
Air Blank	0.000	16:27	
Air Blank	0.000	16:29	
Subject Sample	# 2	16:32	No Sample Provided
Breath Volume	0.000	16:32	
Air Blank	0.000	16:32	
Control Test	0.081	16:33	
Air Blank	0.000	16:33	
Diagnostics Check	OK	16:33	

Data Download Date/Time: 02/17/2022 18:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: RICHARDSON,V
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ARTIAGA

DATE/TIME OF TEST: 02/13/2022 01:38:18

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.165	01:32
Breath Volume	1.578	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.166	01:36
Breath Volume	1.148	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 02/17/2022 18:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MAVES,JEFFREY
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SUBJECT

DATE/TIME OF TEST: 02/17/2022 16:36:33

Diagnostics Check	OK	16:33	
Air Blank	0.000	16:33	
Control Test	0.081	16:33	
Air Blank	0.000	16:34	
Subject Sample	# 1	16:35	Subject Test Refused
Breath Volume	0.000	16:35	
Air Blank	0.000	16:35	
Control Test	0.081	16:35	
Air Blank	0.000	16:36	
Diagnostics Check	OK	16:36	

Data Download Date/Time: 02/17/2022 18:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,TARA,E
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LAWSON,TARA
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SUBJECT

DATE/TIME OF TEST: 02/17/2022 17:00:26

Diagnostics Check	OK	16:50
Air Blank	0.000	16:51
Control Test	0.081	16:51
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:55
Breath Volume	2.144	16:55
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.000	16:59
Breath Volume	1.843	16:59
Air Blank	0.000	16:59
Control Test	0.080	16:59
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 02/17/2022 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: REYES CRUZ

DATE/TIME OF TEST: 02/13/2022 03:57:04

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.198	03:49
Breath Volume	1.984	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.177	03:52
Breath Volume	2.078	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 3	0.174	03:55
Breath Volume	2.132	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/15/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/14/2022 01:09:53

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.082	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1 0.089	01:01	Volume Not Met
Breath Volume	0.703	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2 0.089	01:04	
Breath Volume	1.472	01:04	
Air Blank	0.000	01:05	
Air Blank	0.000	01:07	
Subject Sample	# 3 0.085	01:08	
Breath Volume	1.222	01:08	
Air Blank	0.000	01:08	
Control Test	0.082	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 03/15/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MARBLE

DATE/TIME OF TEST: 02/14/2022 04:21:20

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.090	04:16
Breath Volume	3.578	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.093	04:19
Breath Volume	3.214	04:19
Air Blank	0.000	04:20
Control Test	0.082	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 03/15/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: OHAGAN

DATE/TIME OF TEST: 02/23/2022 20:16:48

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.000	20:12
Breath Volume	2.816	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.000	20:15
Breath Volume	2.394	20:15
Air Blank	0.000	20:15
Control Test	0.082	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 03/15/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 02/24/2022 01:08:30

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.270	01:03
Breath Volume	2.453	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.268	01:06
Breath Volume	2.398	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 03/15/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DRESSLER

DATE/TIME OF TEST: 02/26/2022 00:12:11

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.153	00:07
Breath Volume	1.980	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.160	00:10
Breath Volume	1.722	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/15/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: IVANOVA

DATE/TIME OF TEST: 02/26/2022 06:51:09

Diagnostics Check	OK	06:44
Air Blank	0.000	06:45
Control Test	0.081	06:45
Air Blank	0.000	06:46
Subject Sample # 1	0.169	06:46
Breath Volume	1.511	06:46
Air Blank	0.000	06:47
Air Blank	0.000	06:49
Subject Sample # 2	0.166	06:49
Breath Volume	1.476	06:49
Air Blank	0.000	06:50
Control Test	0.081	06:50
Air Blank	0.000	06:51
Diagnostics Check	OK	06:51

Data Download Date/Time: 03/15/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MCKIVER

DATE/TIME OF TEST: 02/27/2022 00:57:13

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.081	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1 0.128	00:50	Volume Not Met
Breath Volume	0.906	00:50	
Air Blank	0.000	00:50	
Air Blank	0.000	00:52	
Subject Sample	# 2	00:55	No Sample Provided
Breath Volume	0.000	00:55	
Air Blank	0.000	00:56	
Control Test	0.080	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 03/15/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKEEBY,WADE,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: SANTOS PACHECO

DATE/TIME OF TEST: 02/27/2022 06:45:19

Diagnostics Check	OK	06:37
Air Blank	0.000	06:37
Control Test	0.079	06:38
Air Blank	0.000	06:38
Subject Sample # 1	0.100	06:40
Breath Volume	1.218	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:43
Subject Sample # 2	0.100	06:43
Breath Volume	1.472	06:43
Air Blank	0.000	06:44
Control Test	0.080	06:44
Air Blank	0.000	06:45
Diagnostics Check	OK	06:45

Data Download Date/Time: 03/15/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/27/2022 23:06:58

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.036	23:02
Breath Volume	1.453	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.039	23:05
Breath Volume	1.101	23:05
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 03/15/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA, URIEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BELTRAN, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LOPEZ CHAVEZ

DATE/TIME OF TEST: 02/04/2022 05:05:51

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.094	05:01
Breath Volume	1.792	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.095	05:04
Breath Volume	2.699	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 02/04/2022 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALUSMA,JOE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HOUSEMAN

DATE/TIME OF TEST: 02/19/2022 22:30:02

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.078	22:25
Breath Volume	1.265	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.072	22:28
Breath Volume	1.585	22:28
Air Blank	0.000	22:29
Control Test	0.077	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:30

Data Download Date/Time: 03/11/2022 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: MATEODIEGO

DATE/TIME OF TEST: 02/20/2022 21:26:06

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.081	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.197	21:21
Breath Volume	2.089	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:23
Subject Sample # 2	0.189	21:24
Breath Volume	1.636	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 03/11/2022 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZHU,HENRY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: ALMARAZ

DATE/TIME OF TEST: 02/02/2022 22:30:24

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.080	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 1 0.192	22:19	Volume Not Met
Breath Volume	1.140	22:19	
Air Blank	0.000	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 2 0.209	22:23	
Breath Volume	1.593	22:23	
Air Blank	0.000	22:24	
Air Blank	0.000	22:26	
Subject Sample	# 3 0.203	22:28	
Breath Volume	1.746	22:28	
Air Blank	0.000	22:29	
Control Test	0.079	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 03/03/2022 17:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MANSON,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: TOMKINSON

DATE/TIME OF TEST: 02/03/2022 08:31:57

Diagnostics Check	OK	08:25
Air Blank	0.000	08:26
Control Test	0.080	08:26
Air Blank	0.000	08:27
Subject Sample # 1	0.000	08:27
Breath Volume	3.777	08:27
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.000	08:30
Breath Volume	2.339	08:30
Air Blank	0.000	08:31
Control Test	0.079	08:31
Air Blank	0.000	08:31
Diagnostics Check	OK	08:31

Data Download Date/Time: 03/03/2022 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTER,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MCQUAY,ZACH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 02/04/2022 03:39:13

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.147	03:34
Breath Volume	2.351	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.157	03:37
Breath Volume	1.996	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTER,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PORTER,ALEXANDER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: RODRIGUEZ HERNANDEZ

DATE/TIME OF TEST: 02/06/2022 23:21:05

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.222	23:12
Breath Volume	2.488	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.198	23:16
Breath Volume	1.195	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 3	0.225	23:19
Breath Volume	1.648	23:19
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 02/08/2022 21:07:30

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.080	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.079	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 02/22/2022 03:28:12

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PAUGA,JOHN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: HAINES

DATE/TIME OF TEST: 02/22/2022 16:48:20

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:46	
Control Test	0.081	16:46	
Air Blank	0.000	16:46	
Subject Sample	# 1	16:47	Sequence Aborted
Breath Volume	0.000	16:47	
Air Blank	0.000	16:48	

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 02/23/2022 23:04:56

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.132	23:00
Breath Volume	3.847	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.126	23:03
Breath Volume	3.851	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTER,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CALASSO,A
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: BANE

DATE/TIME OF TEST: 02/25/2022 19:26:21

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.079	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.126	19:21
Breath Volume	1.894	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.125	19:24
Breath Volume	1.773	19:24
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 02/26/2022 01:53:27

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.079	01:51	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Control Test	0.079	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/26/2022 05:06:03

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.182	04:58
Breath Volume	1.445	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:00
Subject Sample # 2	0.216	05:01
Breath Volume	2.972	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 3	0.207	05:04
Breath Volume	2.917	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:06

Data Download Date/Time: 04/05/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZHU,HENRY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: ALVARO

DATE/TIME OF TEST: 02/26/2022 05:16:49

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:11
Subject Sample # 1	0.084	05:12
Breath Volume	2.531	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.084	05:15
Breath Volume	3.199	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:16
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 04/05/2022 13:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSH,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ROMAN,RYAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MERICK

DATE/TIME OF TEST: 02/24/2022 20:30:01

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:26	
Control Test	0.079	20:26	
Air Blank	0.000	20:27	
Subject Sample	# 1	20:29	Sequence Aborted
Breath Volume	0.140	20:29	
Air Blank	0.000	20:30	

Data Download Date/Time: 03/08/2022 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSH,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ROMAN,RYAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MERICK

DATE/TIME OF TEST: 02/24/2022 20:41:38

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.077	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.181	20:37	Volume Not Met
Breath Volume	0.429	20:37	
Air Blank	0.000	20:37	
Air Blank	0.000	20:39	
Subject Sample	# 2	20:41	Sequence Aborted
Breath Volume	0.000	20:41	
Air Blank	0.000	20:41	

Data Download Date/Time: 03/08/2022 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSH, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ROMAN, RYAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MERICK

DATE/TIME OF TEST: 02/24/2022 20:52:23

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:44	
Control Test	0.078	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1 0.148	20:48	Volume Not Met
Breath Volume	0.566	20:48	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 2	20:51	Sequence Aborted
Breath Volume	0.476	20:51	
Air Blank	0.000	20:52	

Data Download Date/Time: 03/08/2022 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSH, MATTHEW, E
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ROMAN, RYAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 02/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20520080A1
DRY GAS EXPIRY DATE: 09/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: MERICK

DATE/TIME OF TEST: 02/24/2022 21:06:24

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.235	21:01
Breath Volume	4.460	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.236	21:04
Breath Volume	2.691	21:04
Air Blank	0.000	21:05
Control Test	0.077	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 03/08/2022 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGRASSIA,JAMES,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: KAZAKOV

DATE/TIME OF TEST: 02/20/2022 03:09:10

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.094	03:00	Volume Not Met
Breath Volume	0.148	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.097	03:03	
Breath Volume	2.335	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 3 0.097	03:07	
Breath Volume	1.992	03:07	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 03/20/2022 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DAVID,A
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LACEN,TYLER
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A3
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 02/25/2022 00:57:29

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.198	00:52
Breath Volume	1.125	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.192	00:55
Breath Volume	1.355	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 03/29/2022 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH, LAURA, R
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WALCUTT, BRIAN
ARRESTING OFFICER AGENCY: Volusia County Beach
LAST AGENCY INSPECTION: 01/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: RITZ

DATE/TIME OF TEST: 02/11/2022 02:24:19

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.000	02:20
Breath Volume	3.445	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.000	02:22
Breath Volume	2.628	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 02/28/2022 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH, LAURA, R
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WELLS, KYLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 01/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: GOULD

DATE/TIME OF TEST: 02/18/2022 03:55:46

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.212	03:50
Breath Volume	2.843	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.230	03:54
Breath Volume	2.082	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 02/28/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CAMERON,P
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: ROONEY,Z
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 01/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/23/2022 09:00:56

Diagnostics Check	OK	08:54
Air Blank	0.000	08:54
Control Test	0.080	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.000	08:56
Breath Volume	2.425	08:56
Air Blank	0.000	08:56
Air Blank	0.000	08:58
Subject Sample # 2	0.000	08:59
Breath Volume	2.531	08:59
Air Blank	0.000	09:00
Control Test	0.078	09:00
Air Blank	0.000	09:00
Diagnostics Check	OK	09:00

Data Download Date/Time: 02/28/2022 14:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1217042
DRY GAS EXPIRY DATE: 10/10/2022
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: YARCHUK

DATE/TIME OF TEST: 02/21/2022 16:04:12

Diagnostics Check	OK	15:57
Air Blank	0.000	15:58
Control Test	0.076	15:58
Air Blank	0.000	15:59
Subject Sample # 1	0.000	15:59
Breath Volume	2.058	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:02
Subject Sample # 2	0.000	16:02
Breath Volume	1.910	16:02
Air Blank	0.000	16:03
Control Test	0.075	16:03
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

Data Download Date/Time: 04/08/2022 17:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: DE LA CRUZ,CRISTIAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MARIANI

DATE/TIME OF TEST: 02/12/2022 23:05:43

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.083	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.259	23:01
Breath Volume	1.562	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.244	23:04
Breath Volume	1.875	23:04
Air Blank	0.000	23:04
Control Test	0.082	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 02/25/2022 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,ZAKERIE,D
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: CALEND,NICHOLAS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 09:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:42

SUBJECT'S LAST NAME: ROYER

DATE/TIME OF TEST: 02/22/2022 10:21:29

Diagnostics Check	OK	10:14
Air Blank	0.000	10:15
Control Test	0.083	10:15
Air Blank	0.000	10:16
Subject Sample # 1	0.117	10:17
Breath Volume	2.792	10:17
Air Blank	0.000	10:17
Air Blank	0.000	10:19
Subject Sample # 2	0.118	10:20
Breath Volume	2.500	10:20
Air Blank	0.000	10:20
Control Test	0.081	10:20
Air Blank	0.000	10:21
Diagnostics Check	OK	10:21

Data Download Date/Time: 02/25/2022 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS, E
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: VELEZ, ED
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 02/23/2022 03:05:28

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.184	03:00
Breath Volume	2.402	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.193	03:03
Breath Volume	1.378	03:03
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 02/25/2022 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: REAHL,TIMOTHY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1422348
DRY GAS EXPIRY DATE: 12/24/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: CAMPAGNUOLO

DATE/TIME OF TEST: 02/25/2022 00:12:58

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.082	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.050	00:08
Breath Volume	1.167	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.048	00:11
Breath Volume	1.246	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 02/25/2022 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A3
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MARTINEZ JR

DATE/TIME OF TEST: 02/19/2022 04:14:42

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.090	04:10
Breath Volume	4.007	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.088	04:13
Breath Volume	4.199	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 03/10/2022 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CLARK,JAMES,P

BREATH TEST OPERATOR AGENCY: Rockledge PD

ARRESTING OFFICER: LEITCH,RICKFORD

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 25721080A2

DRY GAS EXPIRY DATE: 11/05/2023

VIOLATION CODE: DUI

ARREST TIME: 13:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 02/12/2022 14:19:09

Diagnostics Check	OK	14:12
Air Blank	0.000	14:13
Control Test	0.080	14:13
Air Blank	0.000	14:14
Subject Sample # 1	0.000	14:14
Breath Volume	4.203	14:14
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.000	14:17
Breath Volume	3.734	14:17
Air Blank	0.000	14:18
Control Test	0.078	14:18
Air Blank	0.000	14:19
Diagnostics Check	OK	14:19

Data Download Date/Time: 03/17/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MYNY,E
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HERNANDEZ PEREZ

DATE/TIME OF TEST: 02/13/2022 00:02:31

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.190	23:57
Breath Volume	2.687	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.183	00:00
Breath Volume	2.847	00:00
Air Blank	0.000	00:01
Control Test	0.077	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/17/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BRUNING,K
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: PEREZ-GUZMAN

DATE/TIME OF TEST: 02/19/2022 21:27:23

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.188	21:22
Breath Volume	1.464	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.192	21:25
Breath Volume	1.703	21:25
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 03/17/2022 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BELL,R
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: VAN DYNE

DATE/TIME OF TEST: 02/23/2022 19:00:16

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.080	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.000	18:56
Breath Volume	1.300	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.000	18:58
Breath Volume	2.343	18:58
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 03/17/2022 22:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, MARK, V

BREATH TEST OPERATOR AGENCY: Rockledge PD

ARRESTING OFFICER: MOREHEAD, AVERY

ARRESTING OFFICER AGENCY: Rockledge PD

LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1175968

DRY GAS EXPIRY DATE: 07/22/2022

VIOLATION CODE: DUI

ARREST TIME: 05:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/09/2022 06:02:12

Diagnostics Check	OK	05:50	
Air Blank	0.000	05:51	
Control Test	0.080	05:51	
Air Blank	0.000	05:51	
Subject Sample	# 1	05:55	No Sample Provided
Breath Volume	0.000	05:55	
Air Blank	0.000	05:55	
Air Blank	0.000	05:57	
Subject Sample	# 2	06:00	Volume Not Met
Breath Volume	0.335	06:00	
Air Blank	0.000	06:01	
Control Test	0.079	06:01	
Air Blank	0.000	06:02	
Diagnostics Check	OK	06:02	

Data Download Date/Time: 02/28/2022 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOREHEAD,AVERY,F
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: MOREHEAD,AVERY
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/09/2022 06:13:46

Diagnostics Check	OK	06:08	
Air Blank	0.000	06:09	
Control Test	0.079	06:09	
Air Blank	0.000	06:09	
Subject Sample	# 1	06:12	Subject Test Refused
Breath Volume	0.382	06:12	
Air Blank	0.000	06:12	
Control Test	0.079	06:13	
Air Blank	0.000	06:13	
Diagnostics Check	OK	06:13	

Data Download Date/Time: 02/28/2022 09:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LIVINGSTON,JAVON
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 13:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:09

SUBJECT'S LAST NAME: PARTIN

DATE/TIME OF TEST: 02/10/2022 15:33:18

Diagnostics Check	OK	15:29	
Air Blank	0.000	15:30	
Control Test	0.079	15:30	
Air Blank	0.000	15:30	
Subject Sample	# 1	15:31	Subject Test Refused
Breath Volume	0.000	15:31	
Air Blank	0.000	15:32	
Control Test	0.079	15:32	
Air Blank	0.000	15:33	
Diagnostics Check	OK	15:33	

Data Download Date/Time: 02/28/2022 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1175968
BREATH TEST OPERATOR: GOMEZ,IVETTE,	DRY GAS EXPIRY DATE: 07/22/2022
BREATH TEST OPERATOR AGENCY: Rockledge PD	VIOLATION CODE: DUI
ARRESTING OFFICER: SPIVEY,MAKENZIE	ARREST TIME: 01:41
ARRESTING OFFICER AGENCY: Rockledge PD	GENDER: Male
LAST AGENCY INSPECTION: 01/27/2022	OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DIAZ-SANCHEZ	DATE/TIME OF TEST: 02/17/2022	02:15:04
-----------------------------------	-------------------------------	----------

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.211	02:10
Breath Volume	2.355	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.207	02:13
Breath Volume	2.500	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:15

Data Download Date/Time: 02/28/2022 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: PAVLAKOS,PAVLOS
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/22/2022
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: STOUT BROWN

DATE/TIME OF TEST: 02/19/2022 12:22:17

Diagnostics Check	OK	12:13
Air Blank	0.000	12:14
Control Test	0.079	12:14
Air Blank	0.000	12:15
Subject Sample # 1	0.000	12:17
Breath Volume	1.257	12:17
Air Blank	0.000	12:18
Air Blank	0.000	12:20
Subject Sample # 2	0.000	12:20
Breath Volume	1.125	12:20
Air Blank	0.000	12:21
Control Test	0.078	12:21
Air Blank	0.000	12:22
Diagnostics Check	OK	12:22

Data Download Date/Time: 02/28/2022 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: FITZGERALD,KYLE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 02/02/2022 21:16:06

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.167	21:11
Breath Volume	2.683	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.169	21:14
Breath Volume	2.445	21:14
Air Blank	0.000	21:15
Control Test	0.079	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KAPLAN,MICHAEL
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FOLSTEIN

DATE/TIME OF TEST: 02/04/2022 02:41:25

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.204	02:36
Breath Volume	3.351	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.187	02:39
Breath Volume	2.574	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,FELICIA,N
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: GIBBONS,LOGAN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:53

SUBJECT'S LAST NAME: ALAMRI

DATE/TIME OF TEST: 02/06/2022 09:21:57

Diagnostics Check	OK	09:14
Air Blank	0.000	09:15
Control Test	0.081	09:15
Air Blank	0.000	09:15
Subject Sample # 1	0.208	09:17
Breath Volume	1.609	09:17
Air Blank	0.000	09:18
Air Blank	0.000	09:19
Subject Sample # 2	0.204	09:20
Breath Volume	1.640	09:20
Air Blank	0.000	09:21
Control Test	0.079	09:21
Air Blank	0.000	09:21
Diagnostics Check	OK	09:21

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,A,M
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PEARSON,K
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 02/11/2022 14:37:43

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.081	14:32
Air Blank	0.000	14:33
Subject Sample # 1	0.000	14:33
Breath Volume	2.574	14:33
Air Blank	0.000	14:33
Air Blank	0.000	14:35
Subject Sample # 2	0.000	14:36
Breath Volume	2.289	14:36
Air Blank	0.000	14:36
Control Test	0.079	14:37
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZEWICZ,ALISON,E
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MORENO PEARSON,KEANDRE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 02/12/2022 09:33:05

Diagnostics Check	OK	09:26
Air Blank	0.000	09:27
Control Test	0.080	09:27
Air Blank	0.000	09:28
Subject Sample # 1	0.000	09:28
Breath Volume	1.570	09:28
Air Blank	0.000	09:29
Air Blank	0.000	09:31
Subject Sample # 2	0.000	09:31
Breath Volume	2.242	09:31
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:33
Diagnostics Check	OK	09:33

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: VAZQUEZ,ANDREW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: TIPTON

DATE/TIME OF TEST: 02/13/2022 21:25:00

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.081	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.000	21:20
Breath Volume	1.683	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.000	21:23
Breath Volume	1.730	21:23
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:25

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PEARSON,KEANDRE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: GANESH

DATE/TIME OF TEST: 02/13/2022 22:02:01

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.000	21:57
Breath Volume	2.539	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.000	22:00
Breath Volume	2.015	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:02

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: TUBAJ,NICHOLAS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DWYER

DATE/TIME OF TEST: 02/23/2022 02:43:23

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:38
Breath Volume	1.394	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:42
Breath Volume	1.457	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAZEWICZ,ALISON,E
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAZEWICZ,ALISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/25/2022 12:36:32

Diagnostics Check	OK	12:29
Air Blank	0.000	12:30
Control Test	0.081	12:30
Air Blank	0.000	12:30
Subject Sample # 1	0.228	12:32
Breath Volume	1.113	12:32
Air Blank	0.000	12:32
Air Blank	0.000	12:34
Subject Sample # 2	0.224	12:34
Breath Volume	1.105	12:34
Air Blank	0.000	12:35
Control Test	0.079	12:35
Air Blank	0.000	12:36
Diagnostics Check	OK	12:36

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: TUBAJ,NICHOLAS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/20/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: POSTELL

DATE/TIME OF TEST: 02/26/2022 05:37:24

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.082	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.127	05:32
Breath Volume	1.281	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.121	05:35
Breath Volume	1.398	05:35
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 03/01/2022 06:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,TUCKER,W
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: PROCTOR,ZAK
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 02/06/2022 03:01:00

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	Interferent Detect
Air Blank	0.000	03:01	

Data Download Date/Time: 02/22/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,TUCKER,W
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: PROCTER,ZAK
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MCINTOSH

DATE/TIME OF TEST: 02/06/2022 03:05:22

Diagnostics Check	OK	03:04	
Air Blank		03:04	Ambient Fail
Air Blank		03:05	Ambient Fail

Data Download Date/Time: 02/22/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,CHRISTOPHER,C
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: BADIKYAN,DAVID
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28719080A1
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: KEAR

DATE/TIME OF TEST: 02/14/2022 00:09:59

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.080	00:04	
Air Blank	0.000	00:05	
Subject Sample # 1	0.146	00:05	
Breath Volume	1.871	00:05	
Air Blank	0.000	00:06	
Air Blank	0.000	00:08	
Subject Sample # 2	0.133	00:08	
Breath Volume	1.238	00:08	
Air Blank	0.000	00:09	
Control Test	0.071	00:09	Control Outside Tolerance
Air Blank	0.000	00:09	

Data Download Date/Time: 02/22/2022 15:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WALTERS,KENRICK
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PIETRO

DATE/TIME OF TEST: 02/01/2022 23:14:35

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:10
Breath Volume	2.949	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:13
Breath Volume	3.105	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 02/24/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/23/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: GEDDES

DATE/TIME OF TEST: 02/14/2022 00:10:59

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.153	00:03	Volume Not Met
Breath Volume	0.050	00:03	
Air Blank	0.000	00:04	
Air Blank	0.000	00:06	
Subject Sample	# 2 0.142	00:09	Volume Not Met
Breath Volume	0.171	00:09	
Air Blank	0.000	00:10	
Control Test	0.077	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 02/24/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: EVERHART,BRANDON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/17/2022 02:29:04

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.194	02:24
Breath Volume	1.632	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.200	02:27
Breath Volume	1.187	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:29

Data Download Date/Time: 02/24/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLACRES,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GONZALES,EDUARDO
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 02/17/2022 03:11:32

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.000	03:07
Breath Volume	2.453	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:10
Breath Volume	2.558	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 02/24/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIDNER,CODY,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FEARNOW,SAMANTHA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MCCASLAND

DATE/TIME OF TEST: 02/21/2022 02:21:57

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.000	02:16
Breath Volume	1.316	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:20
Breath Volume	1.820	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 02/24/2022 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 02/27/2022 07:19:06

Diagnostics Check	OK	07:12
Air Blank	0.000	07:13
Control Test	0.080	07:13
Air Blank	0.000	07:14
Subject Sample # 1	0.140	07:14
Breath Volume	1.800	07:14
Air Blank	0.000	07:15
Air Blank	0.000	07:17
Subject Sample # 2	0.138	07:17
Breath Volume	1.113	07:17
Air Blank	0.000	07:18
Control Test	0.079	07:18
Air Blank	0.000	07:19
Diagnostics Check	OK	07:19

Data Download Date/Time: 04/07/2022 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: VENTRE

DATE/TIME OF TEST: 02/28/2022 05:27:28

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.167	05:23
Breath Volume	2.015	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.161	05:25
Breath Volume	1.496	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 04/07/2022 13:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MCCUNE

DATE/TIME OF TEST: 02/03/2022 01:31:44

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.069	01:27
Breath Volume	2.230	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.074	01:30
Breath Volume	1.734	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 02/23/2022 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 00:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CADE-BOYNTON

DATE/TIME OF TEST: 02/08/2022 00:53:30

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.041	00:49
Breath Volume	1.699	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.042	00:52
Breath Volume	1.617	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 02/23/2022 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 02/10/2022 02:31:56

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.134	02:27
Breath Volume	2.656	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.133	02:30
Breath Volume	2.582	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/23/2022 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 18:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: DAVEY

DATE/TIME OF TEST: 02/12/2022 19:05:21

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.079	19:00
Air Blank	0.000	19:00
Subject Sample # 1	0.144	19:01
Breath Volume	3.027	19:01
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.144	19:03
Breath Volume	2.406	19:03
Air Blank	0.000	19:04
Control Test	0.079	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 02/23/2022 15:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 03:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ZUNIGA FLORES

DATE/TIME OF TEST: 02/13/2022 03:31:26

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.225	03:26
Breath Volume	1.800	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.212	03:29
Breath Volume	1.914	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 02/23/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: TRIMBOLI,MELANIE,M

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: TRIMBOLI,MELANIE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 23:45

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: PIMENTEL

DATE/TIME OF TEST: 02/14/2022 00:19:58

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.176	00:15
Breath Volume	1.171	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.177	00:18
Breath Volume	1.578	00:18
Air Blank	0.000	00:19
Control Test	0.081	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 02/23/2022 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: PEREZ LOPEZ,J
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: VAN SANDT

DATE/TIME OF TEST: 02/14/2022 01:56:10

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.196	01:51
Breath Volume	1.699	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.199	01:54
Breath Volume	1.664	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 02/23/2022 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BARRETTE

DATE/TIME OF TEST: 02/14/2022 03:22:39

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.081	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:15	No Sample Provided
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Air Blank	0.000	03:18	
Subject Sample	# 2	03:21	No Sample Provided
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 02/23/2022 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 02:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BARRETTE

DATE/TIME OF TEST: 02/14/2022 03:30:27

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.193	03:25
Breath Volume	1.781	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.190	03:28
Breath Volume	1.363	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 02/23/2022 16:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: SPALDING

DATE/TIME OF TEST: 02/19/2022 01:42:42

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.193	01:38
Breath Volume	2.488	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.184	01:41
Breath Volume	1.851	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 02/23/2022 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/20/2022 01:37:34

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.142	01:33
Breath Volume	1.742	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.149	01:36
Breath Volume	2.210	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 02/24/2022 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: AUSBROOK

DATE/TIME OF TEST: 02/25/2022 01:44:59

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.104	01:40
Breath Volume	3.164	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.103	01:43
Breath Volume	2.996	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 03/27/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24620080A2

DRY GAS EXPIRY DATE: 11/05/2022

VIOLATION CODE: DUI

ARREST TIME: 04:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/25/2022 04:53:38

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.176	04:49
Breath Volume	1.582	04:49
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.182	04:52
Breath Volume	1.605	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 03/27/2022 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,JERRY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CLEBORNE

DATE/TIME OF TEST: 02/01/2022 01:03:32

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.000	00:58
Breath Volume	2.457	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.000	01:02
Breath Volume	3.585	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 02/02/2022 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: MORRISSEY,ZACHARY
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 02/02/2022 20:43:33

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:41	
Control Test	0.081	20:41	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:42	Subject Test Refused
Breath Volume	0.000	20:42	
Air Blank	0.000	20:42	
Control Test	0.081	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 02/02/2022 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: TORANZO

DATE/TIME OF TEST: 02/03/2022 02:31:18

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.320	02:26
Breath Volume	2.589	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.311	02:29
Breath Volume	3.546	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHRI,BRENDAN,R
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NOTARIO,FREDERICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GROENENDAAL

DATE/TIME OF TEST: 02/04/2022 22:43:17

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.174	22:38
Breath Volume	2.527	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.166	22:41
Breath Volume	2.527	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHRI,BRENDAN,R
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOXLEY,MICHAEL
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 02/05/2022 01:52:05

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.081	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHRI,BRENDAN,R
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SINGH,T
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: FERANDEZ

DATE/TIME OF TEST: 02/05/2022 05:55:49

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:53	
Control Test	0.081	05:53	
Air Blank	0.000	05:54	
Subject Sample	# 1	05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:54	
Control Test	0.081	05:55	
Air Blank	0.000	05:55	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: NEWBERG

DATE/TIME OF TEST: 02/06/2022 02:57:49

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.163	02:53
Breath Volume	3.050	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.149	02:56
Breath Volume	1.816	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLUMSTEIN,REBEKAH
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BEJERANO

DATE/TIME OF TEST: 02/06/2022 03:26:57

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.183	03:22
Breath Volume	3.417	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.169	03:25
Breath Volume	1.714	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VANROMPA,VITO
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: RITTENHOUSE

DATE/TIME OF TEST: 02/22/2022 22:18:04

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.082	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.116	22:13
Breath Volume	2.171	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.113	22:16
Breath Volume	1.699	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:18

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BOSCO

DATE/TIME OF TEST: 02/23/2022 02:58:33

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.100	02:53
Breath Volume	1.433	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.107	02:57
Breath Volume	2.546	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DASQUE,DARREN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: RUBIANO BENAVIDES

DATE/TIME OF TEST: 02/24/2022 02:12:08

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1 0.158	02:02	Volume Not Met
Breath Volume	0.792	02:02	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample	# 2 0.159	02:06	
Breath Volume	1.769	02:06	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 3 0.158	02:10	
Breath Volume	1.218	02:10	
Air Blank	0.000	02:11	
Control Test	0.078	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DIMELLA,PABLO
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LONDONO GOMEZ

DATE/TIME OF TEST: 02/24/2022 02:48:07

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.121	02:43
Breath Volume	2.152	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.115	02:46
Breath Volume	2.031	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BELT,WILLIAM
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 02/25/2022 00:29:53

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.081	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1 0.145	00:17	Volume Not Met
Breath Volume	0.031	00:17	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 2 0.154	00:22	
Breath Volume	1.992	00:22	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 3 0.144	00:28	Volume Not Met
Breath Volume	0.214	00:28	
Air Blank	0.000	00:28	
Control Test	0.079	00:29	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 02/28/2022 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REYES-CENTENO,SERGIO
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PIVERAL

DATE/TIME OF TEST: 02/25/2022 03:41:59

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.170	03:37
Breath Volume	2.726	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.162	03:40
Breath Volume	2.085	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 02/28/2022 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MATEO,YORDAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/26/2022 01:08:23

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.184	01:04
Breath Volume	2.214	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.179	01:06
Breath Volume	1.703	01:06
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 02/28/2022 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UCELO,BRYAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 02/26/2022 04:38:26

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.172	04:33
Breath Volume	3.578	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.171	04:36
Breath Volume	3.726	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 02/28/2022 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,BRIAN,P
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SMITH,MEGAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: KOHLER

DATE/TIME OF TEST: 02/27/2022 08:13:47

Diagnostics Check	OK	08:06
Air Blank	0.000	08:07
Control Test	0.080	08:07
Air Blank	0.000	08:08
Subject Sample # 1	0.100	08:09
Breath Volume	4.398	08:09
Air Blank	0.000	08:09
Air Blank	0.000	08:11
Subject Sample # 2	0.083	08:12
Breath Volume	2.207	08:12
Air Blank	0.000	08:12
Control Test	0.079	08:13
Air Blank	0.000	08:13
Diagnostics Check	OK	08:13

Data Download Date/Time: 02/28/2022 21:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LARRABEE,JEREMY
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/01/2022 03:55:50

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.174	03:50
Breath Volume	2.464	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.160	03:54
Breath Volume	1.480	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JAMES,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 02/01/2022 18:28:31

Diagnostics Check	OK	18:19
Air Blank	0.000	18:19
Control Test	0.080	18:20
Air Blank	0.000	18:20
Subject Sample # 1	0.101	18:23
Breath Volume	1.195	18:23
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.100	18:26
Breath Volume	1.863	18:26
Air Blank	0.000	18:27
Control Test	0.079	18:27
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/02/2022 01:12:53

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.000	01:08
Breath Volume	1.917	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.000	01:11
Breath Volume	1.839	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BAKOMIHALIS,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/02/2022 19:07:04

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:04	
Control Test	0.080	19:04	
Air Blank	0.000	19:04	
Subject Sample	# 1	19:05	Subject Test Refused
Breath Volume	0.000	19:05	
Air Blank	0.000	19:06	
Control Test	0.080	19:06	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:07	

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/02/2022 20:50:41

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.079	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.080	20:50	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: PANTOJA

DATE/TIME OF TEST: 02/02/2022 22:45:14

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.079	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.172	22:40
Breath Volume	3.132	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.167	22:43
Breath Volume	2.628	22:43
Air Blank	0.000	22:44
Control Test	0.080	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: KRAUSS

DATE/TIME OF TEST: 02/03/2022 04:22:08

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ADAIR

DATE/TIME OF TEST: 02/03/2022 21:07:24

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.251	21:02
Breath Volume	2.625	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:04
Subject Sample # 2	0.245	21:05
Breath Volume	2.515	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 02/03/2022 22:55:39

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:43	
Control Test	0.079	22:43	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:47	No Sample Provided
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 2	0.016	22:50
Breath Volume	1.621	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:53	
Subject Sample	# 3	0.017	22:53
Breath Volume	1.984	22:53	
Air Blank	0.000	22:54	
Control Test	0.080	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 02/04/2022 04:22:27

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.187	04:17
Breath Volume	1.925	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.197	04:20
Breath Volume	2.195	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/04/2022 04:32:02

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:29	
Control Test	0.080	04:29	
Air Blank	0.000	04:30	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.079	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKYPACK,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: KELSCH

DATE/TIME OF TEST: 02/04/2022 18:19:46

Diagnostics Check	OK	18:08	
Air Blank	0.000	18:08	
Control Test	0.080	18:09	
Air Blank	0.000	18:09	
Subject Sample	# 1	18:12	No Sample Provided
Breath Volume	0.000	18:12	
Air Blank	0.000	18:13	
Air Blank	0.000	18:15	
Subject Sample	# 2	18:18	No Sample Provided
Breath Volume	0.000	18:18	
Air Blank	0.000	18:18	
Control Test	0.079	18:19	
Air Blank	0.000	18:19	
Diagnostics Check	OK	18:19	

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 02/05/2022 03:42:45

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.115	03:38
Breath Volume	2.312	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.122	03:41
Breath Volume	2.484	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: BETZ

DATE/TIME OF TEST: 02/05/2022 08:04:53

Diagnostics Check	OK	07:57
Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:59
Subject Sample # 1	0.139	07:59
Breath Volume	4.871	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.134	08:03
Breath Volume	3.710	08:03
Air Blank	0.000	08:03
Control Test	0.079	08:04
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 02/17/2022 20:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARLINGTON,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 02/05/2022 22:07:46

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:05	
Control Test	0.080	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:06	Subject Test Refused
Breath Volume	0.000	22:06	
Air Blank	0.000	22:06	
Control Test	0.080	22:07	
Air Blank	0.000	22:07	
Diagnostics Check	OK	22:07	

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOODY,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/06/2022 01:59:58

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.112	01:55
Breath Volume	2.253	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.107	01:58
Breath Volume	2.152	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 02/06/2022 03:35:23

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.258	03:25
Breath Volume	1.464	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.236	03:28
Breath Volume	1.394	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 3	0.229	03:33
Breath Volume	1.671	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FLORENCE

DATE/TIME OF TEST: 02/06/2022 04:02:54

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTELLA,D
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/07/2022 03:23:46

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:16	No Sample Provided
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:19	
Subject Sample	# 2	03:22	No Sample Provided
Breath Volume	0.000	03:22	
Air Blank	0.000	03:22	
Control Test	0.081	03:23	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCKEE,KATHLEEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: SEGRAVES

DATE/TIME OF TEST: 02/07/2022 05:13:17

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.145	05:08
Breath Volume	3.871	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.142	05:11
Breath Volume	3.812	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LIRA RANGEL,CHRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: TURCO

DATE/TIME OF TEST: 02/08/2022 05:33:38

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.081	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:32	Subject Test Refused
Breath Volume	0.000	05:32	
Air Blank	0.000	05:32	
Control Test	0.080	05:33	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 02/09/2022 22:00:55

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.000	21:56
Breath Volume	2.167	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.000	21:59
Breath Volume	2.175	21:59
Air Blank	0.000	21:59
Control Test	0.080	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: UNION

DATE/TIME OF TEST: 02/10/2022 02:23:48

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.198	02:14	Volume Not Met
Breath Volume	0.039	02:14	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2 0.212	02:18	
Breath Volume	1.781	02:18	
Air Blank	0.000	02:19	
Air Blank	0.000	02:21	
Subject Sample	# 3 0.204	02:22	
Breath Volume	1.855	02:22	
Air Blank	0.000	02:22	
Control Test	0.076	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 02/17/2022 20:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: CARRIGAN

DATE/TIME OF TEST: 02/10/2022 21:55:31

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.081	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1 0.296	21:42	
Breath Volume	2.289	21:42	
Air Blank	0.000	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 2 0.226	21:47	Volume Not Met
Breath Volume	0.140	21:47	
Air Blank	0.000	21:49	
Air Blank	0.000	21:50	
Subject Sample	# 3 0.297	21:53	
Breath Volume	1.242	21:53	
Air Blank	0.000	21:54	
Control Test	0.075	21:54	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 02/17/2022 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KING,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 02/11/2022 00:03:22

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.077	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.221	23:58
Breath Volume	1.582	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.214	00:01
Breath Volume	1.628	00:01
Air Blank	0.000	00:02
Control Test	0.078	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 02/17/2022 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/11/2022 02:11:22

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.078	02:06
Breath Volume	2.734	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.078	02:09
Breath Volume	3.503	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 02/17/2022 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 02/11/2022 03:34:13

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 02/17/2022 21:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 02/11/2022 03:51:54

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.079	03:51	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 02/17/2022 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ELLIS JR

DATE/TIME OF TEST: 02/12/2022 04:48:46

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.079	04:46	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.079	04:48	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 02/17/2022 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,KENNETH,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 02/12/2022 05:43:27

Diagnostics Check	OK	05:36
Air Blank	0.000	05:37
Control Test	0.079	05:37
Air Blank	0.000	05:37
Subject Sample # 1	0.126	05:38
Breath Volume	4.355	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 2	0.123	05:41
Breath Volume	4.226	05:41
Air Blank	0.000	05:42
Control Test	0.078	05:42
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 02/17/2022 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,STEVEN,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROFF,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: BELLEMARE

DATE/TIME OF TEST: 02/12/2022 15:18:11

Diagnostics Check	OK	15:11	
Air Blank	0.000	15:11	
Control Test	0.079	15:11	
Air Blank	0.000	15:12	
Subject Sample	# 1 0.159	15:13	
Breath Volume	3.042	15:13	
Air Blank	0.000	15:14	
Air Blank	0.000	15:15	
Subject Sample	# 2 0.154	15:16	
Breath Volume	1.605	15:16	
Air Blank	0.000	15:17	
Control Test	0.079	15:17	
Air Blank		15:17	RFI Detect
Air Blank	0.000	15:18	

Data Download Date/Time: 02/17/2022 21:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEAVER,N
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GORNY

DATE/TIME OF TEST: 02/12/2022 23:23:51

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.079	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.145	23:18
Breath Volume	1.722	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.149	23:22
Breath Volume	2.316	23:22
Air Blank	0.000	23:22
Control Test	0.079	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BIGLIN,A
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: OVERTON

DATE/TIME OF TEST: 02/13/2022 00:39:07

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.080	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.079	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: TENNANT

DATE/TIME OF TEST: 02/13/2022 00:44:56

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GILES

DATE/TIME OF TEST: 02/13/2022 01:10:15

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.079	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: FRANCES

DATE/TIME OF TEST: 02/13/2022 05:16:46

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.079	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.079	05:16	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: NOCERINO

DATE/TIME OF TEST: 02/14/2022 12:40:34

Diagnostics Check	OK	12:34
Air Blank	0.000	12:34
Control Test	0.080	12:35
Air Blank	0.000	12:35
Subject Sample # 1	0.000	12:36
Breath Volume	1.945	12:36
Air Blank	0.000	12:36
Air Blank	0.000	12:38
Subject Sample # 2	0.000	12:39
Breath Volume	1.960	12:39
Air Blank	0.000	12:39
Control Test	0.080	12:39
Air Blank	0.000	12:40
Diagnostics Check	OK	12:40

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IFFT,GARRETT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WELLS,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 02/15/2022 01:50:40

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.000	01:46
Breath Volume	2.367	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.000	01:49
Breath Volume	1.433	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 02/15/2022 20:14:29

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:02	
Control Test	0.080	20:03	
Air Blank	0.000	20:03	
Subject Sample	# 1 0.302	20:06	Volume Not Met
Breath Volume	0.113	20:06	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample	# 2 0.288	20:12	Volume Not Met
Breath Volume	0.253	20:12	
Air Blank	0.000	20:13	
Control Test	0.075	20:13	
Air Blank	0.000	20:14	
Diagnostics Check	OK	20:14	

Data Download Date/Time: 02/17/2022 21:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 02/15/2022 20:25:56

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.076	20:18
Air Blank	0.000	20:18
Subject Sample # 1	0.322	20:19
Breath Volume	1.644	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:21
Subject Sample # 2	0.303	20:23
Breath Volume	1.871	20:23
Air Blank	0.000	20:24
Control Test	0.077	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 02/17/2022 21:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COLMAN, DEPUTY M.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: YAGMIN

DATE/TIME OF TEST: 02/15/2022 23:08:20

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.079	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:07	
Control Test	0.079	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 02/17/2022 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,ZOE,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FOSSITT

DATE/TIME OF TEST: 02/16/2022 03:57:46

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.136	03:53
Breath Volume	3.738	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.133	03:56
Breath Volume	2.500	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 02/17/2022 21:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CROWELL

DATE/TIME OF TEST: 02/16/2022 05:06:10

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.000	05:01
Breath Volume	3.199	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.000	05:04
Breath Volume	2.761	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 02/17/2022 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMIN,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: AREVALO

DATE/TIME OF TEST: 02/17/2022 22:58:52

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.080	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.155	22:53
Breath Volume	2.160	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.149	22:57
Breath Volume	1.199	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 03/23/2022 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ALHUDARE

DATE/TIME OF TEST: 02/18/2022 03:18:49

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.016	03:13
Breath Volume	1.207	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.016	03:17
Breath Volume	1.351	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/23/2022 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORPHYADES,V
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: HERNANDEZ-DOTHI

DATE/TIME OF TEST: 02/18/2022 04:14:31

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.079	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 2	0.110	04:09
Breath Volume	2.054	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 3	0.103	04:12
Breath Volume	2.441	04:12	
Air Blank	0.000	04:13	
Control Test	0.078	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 03/23/2022 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARLINGTON,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COBAS

DATE/TIME OF TEST: 02/19/2022 01:23:30

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.079	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:22	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 03/23/2022 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RILL,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TUKES

DATE/TIME OF TEST: 02/19/2022 02:31:36

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:27
Breath Volume	1.167	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.000	02:30
Breath Volume	1.425	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,Y
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: FALLEUR

DATE/TIME OF TEST: 02/20/2022 03:10:07

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:04	No Sample Provided
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 02/20/2022 03:17:25

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/20/2022 22:16:47

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:09	No Sample Provided
Breath Volume	0.000	22:09	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2	22:15	No Sample Provided
Breath Volume	0.000	22:15	
Air Blank	0.000	22:15	
Control Test	0.080	22:16	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVANOVIC,M,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 11:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/21/2022 02:05:37

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:02	
Control Test	0.080	02:02	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BIGLIN,A
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: TURMAN

DATE/TIME OF TEST: 02/21/2022 03:07:37

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.080	03:05	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVANOVIC,M,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVIC,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GIZATULIN

DATE/TIME OF TEST: 02/21/2022 04:11:43

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.229	04:05	Volume Not Met
Breath Volume	0.187	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 2	04:09	Subject Test Refused
Breath Volume	0.066	04:09	
Air Blank	0.000	04:10	
Control Test	0.076	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,KENNETH,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,KENNETH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 02/22/2022 18:08:33

Diagnostics Check	OK	18:01
Air Blank	0.000	18:02
Control Test	0.079	18:02
Air Blank	0.000	18:03
Subject Sample # 1	0.201	18:03
Breath Volume	1.703	18:03
Air Blank	0.000	18:04
Air Blank	0.000	18:06
Subject Sample # 2	0.187	18:06
Breath Volume	1.968	18:06
Air Blank	0.000	18:07
Control Test	0.078	18:07
Air Blank	0.000	18:08
Diagnostics Check	OK	18:08

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: NOCE

DATE/TIME OF TEST: 02/23/2022 22:58:44

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.155	22:54
Breath Volume	2.097	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.145	22:57
Breath Volume	1.710	22:57
Air Blank	0.000	22:57
Control Test	0.079	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVANOVICLM,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: IVANOVICLM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: BALL

DATE/TIME OF TEST: 02/23/2022 23:51:15

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:50	
Control Test	0.080	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CITRANO,A
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 02/24/2022 04:23:22

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.184	04:18
Breath Volume	2.105	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.180	04:21
Breath Volume	1.148	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHWORTH,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: WELDEN

DATE/TIME OF TEST: 02/24/2022 19:49:14

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.079	19:43
Air Blank	0.000	19:43
Subject Sample # 1	0.245	19:44
Breath Volume	1.707	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:46
Subject Sample # 2	0.233	19:47
Breath Volume	1.753	19:47
Air Blank	0.000	19:48
Control Test	0.079	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 02/26/2022 05:34:20

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.136	05:29
Breath Volume	3.222	05:29
Air Blank	0.000	05:30
Air Blank	0.000	05:31
Subject Sample # 2	0.135	05:32
Breath Volume	3.429	05:32
Air Blank	0.000	05:33
Control Test	0.078	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EALY,S,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EALY,STEVEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: LEIHER

DATE/TIME OF TEST: 02/26/2022 16:59:00

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:56	
Control Test	0.080	16:56	
Air Blank	0.000	16:57	
Subject Sample	# 1	16:57	Subject Test Refused
Breath Volume	0.000	16:57	
Air Blank	0.000	16:58	
Control Test	0.080	16:58	
Air Blank	0.000	16:58	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: RODRIGUES

DATE/TIME OF TEST: 02/27/2022 04:52:40

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.093	04:47
Breath Volume	3.773	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.093	04:50
Breath Volume	3.070	04:50
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 03/23/2022 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: COPPETT

DATE/TIME OF TEST: 02/27/2022 23:14:23

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:09
Breath Volume	1.453	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:12
Breath Volume	1.414	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 03/23/2022 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE, CORPORAL W.,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE, CORPORAL W.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 02/28/2022 01:31:25

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.025	01:26
Breath Volume	4.031	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.025	01:29
Breath Volume	3.371	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 03/23/2022 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GUTTENBERGER

DATE/TIME OF TEST: 02/28/2022 22:33:48

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Control Test	0.079	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 03/23/2022 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GUTTENBERGER

DATE/TIME OF TEST: 02/28/2022 22:53:27

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.138	22:48
Breath Volume	1.531	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.133	22:51
Breath Volume	1.203	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 03/23/2022 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,STEVEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TURNER,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 02/02/2022 21:48:35

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:46	
Control Test	0.080	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:47	Subject Test Refused
Breath Volume	0.000	21:47	
Air Blank	0.000	21:47	
Control Test	0.081	21:47	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 02/02/2022 23:20:34

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.080	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.080	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 02/02/2022 23:34:26

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:26	
Control Test	0.080	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1 0.170	23:30	Volume Not Met
Breath Volume	0.191	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 2	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.078	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: JADRNICEK

DATE/TIME OF TEST: 02/03/2022 03:01:42

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.000	02:57
Breath Volume	1.398	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.000	03:00
Breath Volume	1.371	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MENDOZA ESPINOZA

DATE/TIME OF TEST: 02/06/2022 01:47:27

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.078	01:43
Breath Volume	2.398	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.078	01:45
Breath Volume	2.207	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MCGAN

DATE/TIME OF TEST: 02/06/2022 02:55:13

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.098	02:50
Breath Volume	1.789	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.094	02:53
Breath Volume	1.250	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: HURT

DATE/TIME OF TEST: 02/07/2022 20:25:43

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.080	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.000	20:21
Breath Volume	2.660	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.000	20:24
Breath Volume	2.160	20:24
Air Blank	0.000	20:24
Control Test	0.080	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRIM,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BOURGEOIS

DATE/TIME OF TEST: 02/09/2022 02:16:44

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:13	
Control Test	0.081	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 02/11/2022 02:00:47

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.000	01:56
Breath Volume	1.234	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.000	01:59
Breath Volume	1.132	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: OHR

DATE/TIME OF TEST: 02/11/2022 02:09:45

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.080	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: FITZGIBBON

DATE/TIME OF TEST: 02/11/2022 04:05:14

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.016	04:00
Breath Volume	1.703	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.017	04:03
Breath Volume	2.363	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 02/12/2022 04:21:00

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.081	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 02/12/2022 04:50:36

Diagnostics Check	OK	04:47	
Air Blank	0.000	04:48	
Control Test	0.081	04:48	
Air Blank	0.000	04:49	
Subject Sample	# 1	04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.081	04:50	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: LOCK

DATE/TIME OF TEST: 02/12/2022 05:21:41

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.081	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:20	
Control Test	0.081	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DAMBROSIO,NICHOLAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: AMNAY

DATE/TIME OF TEST: 02/12/2022 06:01:00

Diagnostics Check	OK	05:51
Air Blank	0.000	05:51
Control Test	0.081	05:51
Air Blank	0.000	05:52
Subject Sample # 1	0.186	05:55
Breath Volume	1.285	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:57
Subject Sample # 2	0.187	05:59
Breath Volume	1.339	05:59
Air Blank	0.000	06:00
Control Test	0.082	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:01

Data Download Date/Time: 02/17/2022 01:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,STEVEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TURNER,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 02/13/2022 20:27:48

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.080	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.194	20:23
Breath Volume	1.500	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.199	20:26
Breath Volume	1.250	20:26
Air Blank	0.000	20:26
Control Test	0.079	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 02/17/2022 01:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: QUIROS NOLASCO

DATE/TIME OF TEST: 02/14/2022 03:05:53

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.000	03:01
Breath Volume	1.832	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.000	03:04
Breath Volume	1.410	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 02/17/2022 01:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DEANGELO,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 02/14/2022 04:12:31

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.081	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1 0.184	04:08	
Breath Volume	1.394	04:08	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2	04:11	RFI Detect
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 02/17/2022 01:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DEANGELO,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 02/14/2022 04:22:38

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.080	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.185	04:18
Breath Volume	1.261	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.196	04:21
Breath Volume	1.171	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 02/17/2022 01:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHOMP,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/14/2022 04:35:06

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.237	04:30
Breath Volume	2.039	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.232	04:33
Breath Volume	2.039	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 02/17/2022 01:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTIFER,JAMES,P
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 02/15/2022 16:42:27

Diagnostics Check	OK	16:36
Air Blank	0.000	16:36
Control Test	0.080	16:36
Air Blank	0.000	16:37
Subject Sample # 1	0.051	16:37
Breath Volume	2.320	16:37
Air Blank	0.000	16:38
Air Blank	0.000	16:40
Subject Sample # 2	0.047	16:40
Breath Volume	1.796	16:40
Air Blank	0.000	16:41
Control Test	0.080	16:41
Air Blank	0.000	16:42
Diagnostics Check	OK	16:42

Data Download Date/Time: 02/17/2022 01:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FLORES MURALLES

DATE/TIME OF TEST: 02/17/2022 02:37:06

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.226	02:32
Breath Volume	2.125	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.233	02:35
Breath Volume	1.781	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/23/2022 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 02/17/2022 03:21:31

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.079	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 03/23/2022 15:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DANG

DATE/TIME OF TEST: 02/17/2022 19:36:55

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:34	
Control Test	0.080	19:34	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:35	Subject Test Refused
Breath Volume	0.000	19:35	
Air Blank	0.000	19:36	
Control Test	0.080	19:36	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 02/17/2022 23:10:13

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.193	23:04
Breath Volume	1.449	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.209	23:08
Breath Volume	1.570	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GARSIDE

DATE/TIME OF TEST: 02/18/2022 01:09:27

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.111	01:04
Breath Volume	1.691	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.105	01:07
Breath Volume	1.609	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/18/2022 01:54:51

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:53	Subject Test Refused
Breath Volume	0.000	01:53	
Air Blank	0.000	01:53	
Control Test	0.080	01:54	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GALKA

DATE/TIME OF TEST: 02/19/2022 01:38:26

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.000	01:34
Breath Volume	1.691	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.000	01:37
Breath Volume	1.402	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG113403
DRY GAS EXPIRY DATE: 05/14/2023
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CORMIER

DATE/TIME OF TEST: 02/20/2022 03:35:17

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RAYMOND,C
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: WJITE

DATE/TIME OF TEST: 02/21/2022 21:23:09

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.080	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: KALEDA

DATE/TIME OF TEST: 02/21/2022 21:33:49

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.135	21:29
Breath Volume	1.308	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:31
Subject Sample # 2	0.130	21:32
Breath Volume	1.316	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASSAFUME,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KORDELSKI

DATE/TIME OF TEST: 02/22/2022 01:23:47

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RADDER,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 02/23/2022 01:55:13

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.163	01:50
Breath Volume	1.191	01:50
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.156	01:53
Breath Volume	1.179	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: BAUMGARTNER

DATE/TIME OF TEST: 02/24/2022 19:51:50

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:50	
Control Test	0.079	19:50	
Air Blank	0.000	19:50	
Subject Sample	# 1	19:51	RFI Detect
Breath Volume	0.000	19:51	
Air Blank		19:51	RFI Detect
Air Blank	0.000	19:51	

Data Download Date/Time: 03/23/2022 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZELL,BENJAMIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROZZELL,BENJAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: BAUMGARTNER

DATE/TIME OF TEST: 02/24/2022 20:07:12

Diagnostics Check	OK	19:55	
Air Blank	0.000	19:56	
Control Test	0.079	19:56	
Air Blank	0.000	19:57	
Subject Sample	# 1	20:00	No Sample Provided
Breath Volume	0.000	20:00	
Air Blank	0.000	20:00	
Air Blank	0.000	20:02	
Subject Sample	# 2	20:05	No Sample Provided
Breath Volume	0.000	20:05	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 03/23/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ONEILL

DATE/TIME OF TEST: 02/26/2022 03:26:05

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.000	03:21
Breath Volume	2.132	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.000	03:24
Breath Volume	1.410	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/23/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: KURMAY

DATE/TIME OF TEST: 02/27/2022 21:33:03

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:30	
Control Test	0.080	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:31	Subject Test Refused
Breath Volume	0.000	21:31	
Air Blank	0.000	21:32	
Control Test	0.081	21:32	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 03/23/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: DILAN MORALES

DATE/TIME OF TEST: 02/27/2022 22:06:52

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:04	
Control Test	0.080	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:05	Subject Test Refused
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:06	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 03/23/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KOWALSKI,BRANDON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MANGUM

DATE/TIME OF TEST: 02/27/2022 22:52:06

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.080	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1 0.292	22:47	
Breath Volume	2.394	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample	# 2	22:50	Slope Not Met
Breath Volume	1.011	22:50	
Air Blank	0.000	22:51	
Control Test	0.079	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 03/23/2022 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: VANSTEMPVOORT

DATE/TIME OF TEST: 02/28/2022 01:41:08

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.159	01:36
Breath Volume	2.335	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.153	01:39
Breath Volume	3.046	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 03/23/2022 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PALAZZOLO,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: TYRKALA

DATE/TIME OF TEST: 02/28/2022 04:51:14

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.151	04:46
Breath Volume	1.242	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.155	04:49
Breath Volume	1.406	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 03/23/2022 15:15