Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: REYES,ISRAEL
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CHICAS HERNANDEZ

DATE/TIME OF TEST: 02/06/2022 22:33:25

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.081	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.149	22:28
Breath Volume		2.089	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.157	22:31
Breath Volume		1.695	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 03/01/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: CHAVARRIA,DEGLYS
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/20/2022 00:14:54

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.081	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.131	00:10
Breath Volume		2.503	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:12
Subject Sample	# 2	0.137	00:13
Breath Volume		2.187	00:13
Air Blank		0.000	00:14
Control Test		0.081	00:14
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 03/01/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: RUIZ,TOMMY
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 02/25/2022 00:36:52

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.180	00:32
Breath Volume		2.386	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.184	00:35
Breath Volume		3.140	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 03/01/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: DE MELO FERREIRA

DATE/TIME OF TEST: 02/04/2022 22:08:06

Diagnostics Check		OK	22:02
Air Blank		0.000	22:02
Control Test		0.077	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.085	22:03
Breath Volume		3.535	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.083	22:06
Breath Volume		3.476	22:06
Air Blank		0.000	22:07
Control Test		0.078	22:07
Air Blank		0.000	22:07
Diagnostics Check		OK	22:08

Data Download Date/Time: 03/04/2022 05:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MURALLES

DATE/TIME OF TEST: 02/05/2022 03:23:52

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.077	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.240	03:19
Breath Volume		1.640	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.249	03:22
Breath Volume		2.878	03:22
Air Blank		0.000	03:22
Control Test		0.077	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 03/04/2022 05:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: LAUSELL

DATE/TIME OF TEST: 02/05/2022 05:49:48

Diagnostics Check		OK	05:43
Air Blank		0.000	05:44
Control Test		0.077	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.114	05:45
Breath Volume		2.148	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:47
Subject Sample	# 2	0.108	05:48
Breath Volume		2.109	05:48
Air Blank		0.000	05:48
Control Test		0.077	05:49
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 03/04/2022 05:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: CARRIZOSA-ZAPATA

DATE/TIME OF TEST: 02/05/2022 06:04:09

Diagnostics Check		OK	05:57
Air Blank		0.000	05:57
Control Test		0.077	05:58
Air Blank		0.000	05:58
Subject Sample	# 1	0.144	05:59
Breath Volume		1.898	05:59
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 2	0.136	06:02
Breath Volume		2.546	06:02
Air Blank		0.000	06:03
Control Test		0.077	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 03/04/2022 05:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,R,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DUARTE,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 02/01/2022 17:35:08

Diagnostics Check		OK	17:28
Air Blank		0.000	17:29
Control Test		0.081	17:29
Air Blank		0.000	17:30
Subject Sample	# 1	0.226	17:30
Breath Volume		1.402	17:30
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.216	17:33
Breath Volume		2.363	17:33
Air Blank		0.000	17:34
Control Test		0.079	17:34
Air Blank		0.000	17:35
Diagnostics Check		OK	17:35

Data Download Date/Time: 02/02/2022 06:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,R,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DUARTE,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/12/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 02/01/2022 17:48:38

Diagnostics Check		OK	17:42
Air Blank		0.000	17:42
Control Test		0.081	17:42
Air Blank		0.000	17:43
Subject Sample	# 1	0.082	17:43
Breath Volume		4.246	17:43
Air Blank		0.000	17:44
Air Blank		0.000	17:46
Subject Sample	# 2	0.084	17:46
Breath Volume		4.656	17:46
Air Blank		0.000	17:47
Control Test		0.079	17:48
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 02/02/2022 06:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,OFC R.,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MENDOZA,SGT. R.
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 02/02/2022 15:40:17

Diagnostics Check		OK	15:31	
Air Blank		0.000	15:31	
Control Test		0.080	15:31	
Air Blank		0.000	15:32	
Subject Sample	# 1	0.333	15:32	
Breath Volume		1.312	15:32	
Air Blank		0.000	15:33	
Air Blank		0.000	15:35	
Subject Sample	# 2	0.287	15:35	
Breath Volume		1.183	15:35	
Air Blank		0.000	15:36	
Air Blank		0.000	15:38	
Subject Sample	# 3	0.316	15:39	
Breath Volume		2.195	15:39	
Air Blank			15:39	Purge Fail
Air Blank		0.000	15:40	

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BARQUERO,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: OCHOA JR

DATE/TIME OF TEST: 02/02/2022 22:53:57

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:51	
Control Test		0.080	22:51	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:52	Subject Test Refused
Breath Volume		0.000	22:52	
Air Blank		0.000	22:53	
Control Test		0.081	22:53	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: PARDO

DATE/TIME OF TEST: 02/03/2022 03:25:50

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.081	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.043	03:21
Breath Volume		3.433	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.044	03:24
Breath Volume		3.003	03:24
Air Blank		0.000	03:24
Control Test		0.081	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VALLE,J
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: PARRA SANCHEZ

DATE/TIME OF TEST: 02/05/2022 06:02:42

Diagnostics Check		OK	05:54
Air Blank		0.000	05:55
Control Test		0.080	05:55
Air Blank		0.000	05:56
Subject Sample	# 1	0.190	05:57
Breath Volume		2.382	05:57
Air Blank		0.000	05:58
Air Blank		0.000	05:59
Subject Sample	# 2	0.185	06:01
Breath Volume		2.210	06:01
Air Blank		0.000	06:01
Control Test		0.080	06:02
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,OFC. R.,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: TAMARGO,DET. R.
ARRESTING OFFICER AGENCY: MIAMI SPRINGS PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 02/09/2022 18:42:27

Diagnostics Check		OK	18:35
Air Blank		0.000	18:36
Control Test		0.081	18:36
Air Blank		0.000	18:36
Subject Sample	# 1	0.000	18:38
Breath Volume		2.082	18:38
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.000	18:41
Breath Volume		2.429	18:41
Air Blank		0.000	18:41
Control Test		0.080	18:41
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/13/2022 07:49:38

Diagnostics Check		OK	07:37	
Air Blank		0.000	07:38	
Control Test		0.081	07:38	
Air Blank		0.000	07:39	
Subject Sample	# 1	0.152	07:42	Volume Not Met
Breath Volume		0.246	07:42	
Air Blank		0.000	07:43	
Air Blank		0.000	07:44	
Subject Sample	# 2	0.155	07:47	Volume Not Met
Breath Volume		0.054	07:47	
Air Blank		0.000	07:48	
Control Test		0.079	07:48	
Air Blank		0.000	07:49	
Diagnostics Check		OK	07:49	

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/13/2022 07:54:23

Diagnostics Check		OK	07:51	
Air Blank		0.000	07:51	
Control Test		0.079	07:52	
Air Blank		0.000	07:52	
Subject Sample	# 1		07:52	Subject Test Refused
Breath Volume		0.000	07:52	
Air Blank		0.000	07:53	
Control Test		0.079	07:53	
Air Blank		0.000	07:54	
Diagnostics Check		OK	07:54	

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,D,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RAMIREZ,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: RODRIGUEZ TORRES

DATE/TIME OF TEST: 02/18/2022 08:57:56

Diagnostics Check		OK	08:49	
Air Blank		0.000	08:49	
Control Test		0.080	08:50	
Air Blank		0.000	08:50	
Subject Sample	# 1		08:53	No Sample Provided
Breath Volume		0.000	08:53	
Air Blank		0.000	08:54	
Air Blank		0.000	08:56	
Subject Sample	# 2		08:56	Subject Test Refused
Breath Volume		0.000	08:56	
Air Blank		0.000	08:57	
Control Test		0.081	08:57	
Air Blank		0.000	08:57	
Diagnostics Check		OK	08:57	

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 02/20/2022 03:35:53

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.081	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.135	03:31
Breath Volume		1.656	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:33
Subject Sample	# 2	0.127	03:34
Breath Volume		1.238	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,M,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VALDEZ,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: TORRES OSUNA

DATE/TIME OF TEST: 02/24/2022 20:42:25

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:39	
Control Test		0.081	20:39	
Air Blank		0.000	20:40	
Subject Sample	# 1		20:41	Subject Test Refused
Breath Volume		0.000	20:41	
Air Blank		0.000	20:41	
Control Test		0.082	20:41	
Air Blank		0.000	20:42	
Diagnostics Check		OK	20:42	

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: PEREZ,ANTONIO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 02/25/2022 23:56:39

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.081	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.200	23:51
Breath Volume		2.343	23:51
Air Blank		0.000	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.195	23:54
Breath Volume		2.601	23:54
Air Blank		0.000	23:55
Control Test		0.077	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 02/01/2022 01:58:03

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:52
Subject Sample	# 1	0.097	01:53
Breath Volume		1.500	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.095	01:56
Breath Volume		1.539	01:56
Air Blank		0.000	01:57
Control Test		0.078	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:58

Data Download Date/Time: 02/05/2022 15:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: LEMUS DUBON

DATE/TIME OF TEST: 02/02/2022 01:49:09

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.078	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.079	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 02/05/2022 15:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DORSEY-BANKS

DATE/TIME OF TEST: 02/02/2022 22:34:07

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:31	
Control Test		0.078	22:32	
Air Blank		0.000	22:32	
Subject Sample	# 1		22:32	Subject Test Refused
Breath Volume		0.000	22:32	
Air Blank		0.000	22:33	
Control Test		0.078	22:33	
Air Blank		0.000	22:34	
Diagnostics Check		OK	22:34	

Data Download Date/Time: 02/05/2022 15:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 02/03/2022 03:03:59

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.078	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:02	Subject Test Refused
Breath Volume		0.000	03:02	
Air Blank		0.000	03:03	
Control Test		0.078	03:03	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 02/05/2022 15:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: THACKER

DATE/TIME OF TEST: 02/03/2022 06:04:42

Diagnostics Check		OK	05:57
Air Blank		0.000	05:58
Control Test		0.077	05:58
Air Blank		0.000	05:59
Subject Sample	# 1	0.136	06:00
Breath Volume		1.953	06:00
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 2	0.146	06:03
Breath Volume		1.433	06:03
Air Blank		0.000	06:03
Control Test		0.078	06:04
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 02/05/2022 15:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MUNGUIA RIVERA

DATE/TIME OF TEST: 02/05/2022 05:29:24

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.078	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.133	05:24
Breath Volume		1.859	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.132	05:27
Breath Volume		2.082	05:27
Air Blank		0.000	05:28
Control Test		0.078	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 02/05/2022 15:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/06/2022 02:52:31

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.202	02:47
Breath Volume		3.101	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.195	02:51
Breath Volume		2.304	02:51
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 02/23/2022 11:21

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: DINGLER

DATE/TIME OF TEST: 02/06/2022 05:39:53

Diagnostics Check		OK	05:33
Air Blank		0.000	05:33
Control Test		0.079	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.188	05:35
Breath Volume		1.910	05:35
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.195	05:38
Breath Volume		1.875	05:38
Air Blank		0.000	05:38
Control Test		0.078	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 02/23/2022 11:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ILLAS

DATE/TIME OF TEST: 02/07/2022 01:35:06

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.078	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.097	01:30
Breath Volume		1.667	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.093	01:33
Breath Volume		1.687	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 02/23/2022 11:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 02/10/2022 05:39:22

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:37	
Control Test		0.079	05:37	
Air Blank		0.000	05:37	
Subject Sample	# 1		05:38	Subject Test Refused
Breath Volume		0.000	05:38	
Air Blank		0.000	05:38	
Control Test		0.079	05:38	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 02/23/2022 11:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LOMBARDO

DATE/TIME OF TEST: 02/11/2022 02:16:34

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:14	
Control Test		0.079	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:15	Subject Test Refused
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 02/23/2022 11:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MACMILLAN

DATE/TIME OF TEST: 02/12/2022 01:32:32

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.164	01:27
Breath Volume		3.460	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.163	01:31
Breath Volume		2.878	01:31
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 02/23/2022 11:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 02/12/2022 04:41:37

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.079	04:35
Air Blank		0.000	04:36
Subject Sample	# 1	0.159	04:36
Breath Volume		1.480	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.153	04:40
Breath Volume		2.027	04:40
Air Blank		0.000	04:40
Control Test		0.078	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 02/23/2022 11:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 02/12/2022 19:57:08

Diagnostics Check		OK	19:50
Air Blank		0.000	19:51
Control Test		0.078	19:51
Air Blank		0.000	19:51
Subject Sample	# 1	0.110	19:52
Breath Volume		1.250	19:52
Air Blank		0.000	19:53
Air Blank		0.000	19:54
Subject Sample	# 2	0.111	19:55
Breath Volume		1.425	19:55
Air Blank		0.000	19:56
Control Test		0.078	19:56
Air Blank		0.000	19:57
Diagnostics Check		OK	19:57

Data Download Date/Time: 02/23/2022 11:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 02/12/2022 22:30:11

Diagnostics Check		OK	22:27	
Air Blank		0.000	22:27	
Control Test		0.078	22:28	
Air Blank		0.000	22:28	
Subject Sample	# 1		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:29	
Control Test		0.078	22:29	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 02/23/2022 11:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARDWELL,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CARRINGTON

DATE/TIME OF TEST: 02/13/2022 02:40:05

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.078	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:39	
Control Test		0.078	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 02/13/2022 03:40:25

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.165	03:35
Breath Volume		1.546	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.170	03:38
Breath Volume		1.539	03:38
Air Blank		0.000	03:39
Control Test		0.077	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: PENA LOPEZ

DATE/TIME OF TEST: 02/14/2022 03:12:42

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.134	03:08
Breath Volume		2.257	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.134	03:11
Breath Volume		3.355	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 02/17/2022 02:41:11

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.125	02:36
Breath Volume		1.484	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.123	02:39
Breath Volume		1.371	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HALBRITTER,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 02/18/2022 06:09:01

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:06	
Control Test		0.077	06:06	
Air Blank		0.000	06:07	
Subject Sample	# 1		06:07	Subject Test Refused
Breath Volume		0.000	06:07	
Air Blank		0.000	06:08	
Control Test		0.076	06:08	
Air Blank		0.000	06:08	
Diagnostics Check		OK	06:09	

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MATTIACE,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 02/19/2022 02:37:39

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.145	02:32
Breath Volume		1.429	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.143	02:36
Breath Volume		1.324	02:36
Air Blank		0.000	02:36
Control Test		0.078	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/19/2022 22:57:02

Diagnostics Check		OK	22:54	
Air Blank		0.000	22:54	
Control Test		0.079	22:54	
Air Blank		0.000	22:55	
Subject Sample	# 1		22:55	Subject Test Refused
Breath Volume		0.000	22:55	
Air Blank		0.000	22:56	
Control Test		0.079	22:56	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:57	

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: MELEAN REYES

DATE/TIME OF TEST: 02/20/2022 23:24:49

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.077	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.145	23:20
Breath Volume		1.191	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:22
Subject Sample	# 2	0.145	23:23
Breath Volume		1.601	23:23
Air Blank		0.000	23:23
Control Test		0.077	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/05/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 02/22/2022 04:22:13

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:19	
Control Test		0.077	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:20	Subject Test Refused
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 02/23/2022 11:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LEVY

DATE/TIME OF TEST: 02/25/2022 03:08:29

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.169	03:03
Breath Volume		2.808	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.161	03:06
Breath Volume		2.730	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 03/10/2022 04:41

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPURRIER,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 02/25/2022 22:00:49

Diagnostics Check		OK	21:54
Air Blank		0.000	21:54
Control Test		0.078	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.145	21:56
Breath Volume		1.945	21:56
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.144	21:59
Breath Volume		1.738	21:59
Air Blank		0.000	21:59
Control Test		0.077	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 03/10/2022 04:41

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: PEGG

DATE/TIME OF TEST: 02/26/2022 02:52:45

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.000	02:48
Breath Volume		3.804	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.000	02:51
Breath Volume		3.945	02:51
Air Blank		0.000	02:51
Control Test		0.078	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 03/10/2022 04:41

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 02/26/2022 20:32:45

Diagnostics Check		OK	20:29	
Air Blank		0.000	20:30	
Control Test		0.078	20:30	
Air Blank		0.000	20:31	
Subject Sample	# 1		20:31	Subject Test Refused
Breath Volume		0.000	20:31	
Air Blank		0.000	20:31	
Control Test		0.078	20:32	
Air Blank		0.000	20:32	
Diagnostics Check		OK	20:32	

Data Download Date/Time: 03/10/2022 04:41

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 02/26/2022 22:16:21

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.076	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.181	22:11
Breath Volume		2.269	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.184	22:14
Breath Volume		1.972	22:14
Air Blank		0.000	22:15
Control Test		0.078	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SKRAMSTAD

DATE/TIME OF TEST: 02/27/2022 02:50:02

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.078	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.196	02:45
Breath Volume		2.066	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.191	02:48
Breath Volume		1.988	02:48
Air Blank		0.000	02:49
Control Test		0.078	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:50

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPURRIER,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/27/2022 04:57:32

Diagnostics Check		OK	04:54	
Air Blank		0.000	04:55	
Control Test		0.078	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 1		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Control Test		0.078	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: HERNANDEZ ALMARAZ

DATE/TIME OF TEST: 02/27/2022 05:39:50

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.078	05:33
Air Blank		0.000	05:33
Subject Sample	# 1	0.132	05:35
Breath Volume		2.117	05:35
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.127	05:38
Breath Volume		2.539	05:38
Air Blank		0.000	05:38
Control Test		0.078	05:39
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 02/27/2022 19:31:42

Diagnostics Check		OK	19:28	
Air Blank		0.000	19:29	
Control Test		0.078	19:29	
Air Blank		0.000	19:30	
Subject Sample	# 1		19:30	Subject Test Refused
Breath Volume		0.000	19:30	
Air Blank		0.000	19:30	
Control Test		0.077	19:31	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:31	

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: SAULSBY

DATE/TIME OF TEST: 02/27/2022 23:29:01

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:14	
Control Test		0.078	23:14	
Air Blank		0.000	23:15	
Subject Sample	# 1	0.059	23:16	
Breath Volume		1.546	23:16	
Air Blank		0.000	23:16	
Air Blank		0.000	23:18	
Subject Sample	# 2	0.054	23:21	Volume Not Met
Breath Volume		0.058	23:21	
Air Blank		0.000	23:22	
Air Blank		0.000	23:24	
Subject Sample	# 3	0.058	23:27	Volume Not Met
Breath Volume		0.207	23:27	
Air Blank		0.000	23:28	
Control Test		0.077	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: LAMPLEY

DATE/TIME OF TEST: 02/28/2022 01:50:13

Diagnostics Check		OK	01:43
Air Blank		0.000	01:44
Control Test		0.077	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.040	01:45
Breath Volume		1.531	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.037	01:48
Breath Volume		1.488	01:48
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 02/28/2022 04:21:24

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.124	04:13
Breath Volume		3.050	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.145	04:16
Breath Volume		3.535	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 3	0.137	04:19
Breath Volume		4.656	04:19
Air Blank		0.000	04:20
Control Test		0.077	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 03/10/2022 04:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,C,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: MIREE

DATE/TIME OF TEST: 02/28/2022 16:55:38

Diagnostics Check		OK	16:49
Air Blank		0.000	16:50
Control Test		0.078	16:50
Air Blank		0.000	16:50
Subject Sample	# 1	0.000	16:51
Breath Volume		2.441	16:51
Air Blank		0.000	16:51
Air Blank		0.000	16:53
Subject Sample	# 2	0.000	16:54
Breath Volume		1.722	16:54
Air Blank		0.000	16:54
Control Test		0.078	16:55
Air Blank		0.000	16:55
Diagnostics Check		OK	16:55

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: LOPEZ FERNANDEZ

DATE/TIME OF TEST: 02/25/2022 04:53:34

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.000	04:53	Control Outside Tolerance
Air Blank	0.000	04:53	

Data Download Date/Time: 03/10/2022 03:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: HERNANDEZ ALMARAZ

DATE/TIME OF TEST: 02/27/2022 05:28:04

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.000	05:27	Control Outside Tolerance
Air Blank	0.000	05:28	

Data Download Date/Time: 03/10/2022 03:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1048227
DRY GAS EXPIRY DATE: 11/15/2021
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/02/2022 21:28:36

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:26	
Control Test		0.081	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 1		21:27	Subject Test Refused
Breath Volume		0.000	21:27	
Air Blank		0.000	21:27	
Control Test		0.080	21:28	
Air Blank		0.000	21:28	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 02/04/2022 08:38

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: CAROTTI,DEPUTY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 02/03/2022 23:48:51

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.000	23:44
Breath Volume		1.589	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.000	23:47
Breath Volume		1.898	23:47
Air Blank		0.000	23:47
Control Test		0.077	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 04/28/2022 09:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: LEWIS,L
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 02/06/2022 23:06:06

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.079	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:04	Subject Test Refused
Breath Volume		0.000	23:04	
Air Blank		0.000	23:05	
Control Test		0.079	23:05	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 04/28/2022 09:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: OTHER
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 02/11/2022 02:40:48

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1	0.215	02:33	Volume Not Met
Breath Volume		0.031	02:33	
Air Blank		0.000	02:34	
Air Blank		0.000	02:36	
Subject Sample	# 2	0.000	02:39	Volume Not Met
Breath Volume		0.046	02:39	
Air Blank		0.000	02:39	
Control Test		0.075	02:40	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 04/28/2022 09:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 02/15/2022 18:41:59

Diagnostics Check		OK	18:39	
Air Blank		0.000	18:39	
Control Test		0.081	18:39	
Air Blank		0.000	18:40	
Subject Sample	# 1		18:40	Subject Test Refused
Breath Volume		0.000	18:40	
Air Blank		0.000	18:41	
Control Test		0.080	18:41	
Air Blank		0.000	18:41	
Diagnostics Check		OK	18:41	

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,DEPUTY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/16/2022 00:44:31

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.080	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.179	00:39
Breath Volume		3.839	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.175	00:42
Breath Volume		3.679	00:42
Air Blank		0.000	00:43
Control Test		0.078	00:43
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,DEPUTY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: VARNES

DATE/TIME OF TEST: 02/20/2022 23:56:50

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.079	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:52
Breath Volume		1.449	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		1.867	23:55
Air Blank		0.000	23:55
Control Test		0.078	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 02/23/2022 21:10:33

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.079	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.218	21:05
Breath Volume		1.128	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.201	21:08
Breath Volume		1.367	21:08
Air Blank		0.000	21:09
Control Test		0.076	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,DEPUTY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/26/2022 03:32:10

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.072	03:26
Breath Volume		1.230	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.068	03:30
Breath Volume		1.687	03:30
Air Blank		0.000	03:31
Control Test		0.076	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WYCHUNAS,REIDE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/02/2022 21:03:31

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:02	
Control Test	0.082	21:02	
Air Blank		21:03	RFI Detect
Air Blank	0.000	21:03	

Data Download Date/Time: 02/28/2022 17:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WYCHUNAS,REIDE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/02/2022 21:13:48

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.081	21:07
Air Blank		0.000	21:07
Subject Sample	# 1	0.216	21:09
Breath Volume		1.914	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.222	21:12
Breath Volume		1.625	21:12
Air Blank		0.000	21:12
Control Test		0.080	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 02/28/2022 17:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 02/05/2022 02:57:17

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.082	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.014	02:53
Breath Volume		3.488	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.013	02:55
Breath Volume		3.015	02:55
Air Blank		0.000	02:56
Control Test		0.082	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 02/28/2022 17:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/06/2022 00:29:06

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.083	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.000	00:24
Breath Volume		1.789	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.000	00:27
Breath Volume		2.050	00:27
Air Blank		0.000	00:28
Control Test		0.083	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/28/2022 17:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORENE,AIDAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NORENE,AIDAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FELIX VILLAMAR

DATE/TIME OF TEST: 02/06/2022 01:06:36

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.083	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.196	01:02
Breath Volume		4.082	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.199	01:05
Breath Volume		3.796	01:05
Air Blank		0.000	01:05
Control Test		0.082	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 02/28/2022 17:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: COOLEY,LOGAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: HENDLER

DATE/TIME OF TEST: 02/07/2022 04:01:27

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:49	
Control Test		0.083	03:49	
Air Blank		0.000	03:50	
Subject Sample	# 1		03:53	No Sample Provided
Breath Volume		0.000	03:53	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.021	03:56	
Breath Volume		2.511	03:56	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 3	0.018	04:00	
Breath Volume		2.195	04:00	
Air Blank		0.000	04:00	
Control Test		0.082	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 02/28/2022 17:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNGHANSE,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:38

SUBJECT'S LAST NAME: MEGGINSON

DATE/TIME OF TEST: 02/07/2022 10:05:31

Diagnostics Check		OK	09:58
Air Blank		0.000	09:59
Control Test		0.081	09:59
Air Blank		0.000	09:59
Subject Sample	# 1	0.337	10:00
Breath Volume		3.414	10:00
Air Blank		0.000	10:01
Air Blank		0.000	10:03
Subject Sample	# 2	0.335	10:03
Breath Volume		3.203	10:03
Air Blank		0.000	10:04
Control Test		0.081	10:04
Air Blank		0.000	10:05
Diagnostics Check		OK	10:05

Data Download Date/Time: 02/28/2022 17:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NASH,JUSTIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 02/07/2022 16:23:04

Diagnostics Check		OK	16:20	
Air Blank		0.000	16:20	
Control Test		0.081	16:21	
Air Blank		0.000	16:21	
Subject Sample	# 1		16:21	Subject Test Refused
Breath Volume		0.000	16:21	
Air Blank		0.000	16:22	
Control Test		0.081	16:22	
Air Blank		0.000	16:22	
Diagnostics Check		OK	16:23	

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 11:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 02/09/2022 00:31:41

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.082	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.247	00:27
Breath Volume		2.054	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.246	00:30
Breath Volume		1.992	00:30
Air Blank		0.000	00:30
Control Test		0.082	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCKER,JEREMY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DUCKER,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: VOIT

DATE/TIME OF TEST: 02/11/2022 03:56:28

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.081	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.212	03:51
Breath Volume		2.031	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.210	03:54
Breath Volume		2.863	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SCAPIN,BRANDON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/11/2022 04:54:23

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.082	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.131	04:49
Breath Volume		2.429	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.122	04:52
Breath Volume		1.355	04:52
Air Blank		0.000	04:53
Control Test		0.082	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNN,JAMISON,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELCH,BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 02/12/2022 00:24:39

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:16	
Control Test		0.081	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 1		00:20	No Sample Provided
Breath Volume		0.000	00:20	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.081	00:24	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/12/2022 02:17:04

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.082	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.053	02:12
Breath Volume		4.218	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.049	02:15
Breath Volume		3.593	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:17

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/12/2022 02:52:26

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.081	02:49	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Control Test		0.081	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 02/12/2022 22:55:28

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:52	
Control Test		0.081	22:52	
Air Blank		0.000	22:53	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:54	
Control Test		0.081	22:54	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: PFEFFER

DATE/TIME OF TEST: 02/13/2022 00:04:13

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.082	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.121	23:59
Breath Volume		3.515	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.119	00:02
Breath Volume		3.183	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 02/13/2022 04:59:45

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.082	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:52	No Sample Provided
Breath Volume		0.000	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 2		04:58	No Sample Provided
Breath Volume		0.000	04:58	
Air Blank		0.000	04:58	
Control Test		0.082	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: CAZARES VALDEZ

DATE/TIME OF TEST: 02/13/2022 05:29:32

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.082	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.132	05:25
Breath Volume		3.074	05:25
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.131	05:28
Breath Volume		2.457	05:28
Air Blank		0.000	05:28
Control Test		0.081	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: PETITT

DATE/TIME OF TEST: 02/18/2022 02:31:30

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.082	02:30	
Air Blank		02:31	RFI Detect
Air Blank	0.000	02:31	

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZAHCARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: PETITT

DATE/TIME OF TEST: 02/18/2022 02:39:52

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.082	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.000	02:35
Breath Volume		1.792	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.000	02:38
Breath Volume		1.449	02:38
Air Blank		0.000	02:38
Control Test		0.082	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLE,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LOCKETT

DATE/TIME OF TEST: 02/19/2022 00:49:57

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.083	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.121	00:45
Breath Volume		2.179	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.115	00:48
Breath Volume		2.324	00:48
Air Blank		0.000	00:49
Control Test		0.082	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 02/28/2022 17:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,PAUL,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RODRIGUEZ,PAUL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SCHULZ

DATE/TIME OF TEST: 02/19/2022 01:55:55

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.082	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.110	01:51
Breath Volume		2.734	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.105	01:54
Breath Volume		2.425	01:54
Air Blank		0.000	01:55
Control Test		0.082	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/21/2022 01:18:45

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.082	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Control Test		0.082	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 02/22/2022 00:47:43

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.081	00:45	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:46	
Control Test		0.081	00:47	
Air Blank		0.000	00:47	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MIXON

DATE/TIME OF TEST: 02/23/2022 01:07:49

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:06	
Control Test		0.082	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank			01:07	RFI Detect
Air Blank		0.000	01:07	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MIXON

DATE/TIME OF TEST: 02/23/2022 01:14:04

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.081	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:13	
Control Test		0.082	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,RICKY,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 02/23/2022 02:03:03

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.082	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.000	01:58
Breath Volume		1.339	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.000	02:01
Breath Volume		1.652	02:01
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:03

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WATERS,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MCELROY

DATE/TIME OF TEST: 02/24/2022 00:29:40

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.081	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Control Test		0.081	00:29	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SPANO

DATE/TIME OF TEST: 02/24/2022 02:43:02

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.079	02:42	RFI Detect
Air Blank	0.000	02:43	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SPANO

DATE/TIME OF TEST: 02/24/2022 02:46:53

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.081	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.081	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/24/2022 21:44:59

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.082	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.092	21:39
Breath Volume		1.195	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.093	21:43
Breath Volume		1.667	21:43
Air Blank		0.000	21:44
Control Test		0.080	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ACRE

DATE/TIME OF TEST: 02/25/2022 02:08:57

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.081	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.000	02:04
Breath Volume		3.421	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.000	02:07
Breath Volume		3.003	02:07
Air Blank		0.000	02:08
Control Test		0.081	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELCH,BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 02/25/2022 02:55:11

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.081	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.117	02:50
Breath Volume		1.308	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.112	02:53
Breath Volume		1.277	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACHEWICZ,RILEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HOLLIS

DATE/TIME OF TEST: 02/26/2022 00:06:18

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.082	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.200	00:01
Breath Volume		1.988	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.202	00:04
Breath Volume		1.820	00:04
Air Blank		0.000	00:05
Control Test		0.081	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 02/26/2022 03:27:17

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.082	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:25	Subject Test Refused
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Control Test		0.082	03:26	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 02/28/2022 17:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACHEWICZ,RILEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CARSTENS

DATE/TIME OF TEST: 02/27/2022 02:07:58

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.081	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		02:01	No Sample Provided
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 2		02:06	No Sample Provided
Breath Volume		0.000	02:06	
Air Blank		0.000	02:07	
Control Test		0.082	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 02/28/2022 17:38

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,LLOYD,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ALEXANDER,LLOYD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: JABLONSKI

DATE/TIME OF TEST: 02/27/2022 15:48:05

Diagnostics Check		OK	15:41
Air Blank		0.000	15:42
Control Test		0.081	15:42
Air Blank		0.000	15:42
Subject Sample	# 1	0.259	15:43
Breath Volume		4.730	15:43
Air Blank		0.000	15:44
Air Blank		0.000	15:46
Subject Sample	# 2	0.247	15:46
Breath Volume		3.398	15:46
Air Blank		0.000	15:47
Control Test		0.080	15:47
Air Blank		0.000	15:47
Diagnostics Check		OK	15:48

Data Download Date/Time: 02/28/2022 17:38

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPKINS,TYLER,C
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: NASH,JARRED
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MARINO CASTILLO

DATE/TIME OF TEST: 02/14/2022 02:41:54

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.079	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 03/10/2022 11:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,KENNETH,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PHILLIPS,KENNETH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: IACOVINO

DATE/TIME OF TEST: 02/01/2022 15:51:33

Diagnostics Check		OK	15:45
Air Blank		0.000	15:45
Control Test		0.082	15:46
Air Blank		0.000	15:46
Subject Sample	# 1	0.000	15:47
Breath Volume		3.207	15:47
Air Blank		0.000	15:47
Air Blank		0.000	15:49
Subject Sample	# 2	0.000	15:50
Breath Volume		3.242	15:50
Air Blank		0.000	15:50
Control Test		0.080	15:50
Air Blank		0.000	15:51
Diagnostics Check		OK	15:51

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RUANO

DATE/TIME OF TEST: 02/02/2022 01:01:12

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.082	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.000	00:56
Breath Volume		3.449	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.000	00:59
Breath Volume		2.457	00:59
Air Blank		0.000	01:00
Control Test		0.082	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:49

SUBJECT'S LAST NAME: CASTAN CUTINO

DATE/TIME OF TEST: 02/05/2022 10:15:32

Diagnostics Check		OK	10:09
Air Blank		0.000	10:09
Control Test		0.082	10:10
Air Blank		0.000	10:10
Subject Sample	# 1	0.106	10:11
Breath Volume		1.359	10:11
Air Blank		0.000	10:11
Air Blank		0.000	10:13
Subject Sample	# 2	0.104	10:14
Breath Volume		2.968	10:14
Air Blank		0.000	10:14
Control Test		0.083	10:14
Air Blank		0.000	10:15
Diagnostics Check		OK	10:15

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MADDOX ANASTASI

DATE/TIME OF TEST: 02/06/2022 02:51:01

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.082	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.143	02:45
Breath Volume		1.441	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.149	02:49
Breath Volume		1.609	02:49
Air Blank		0.000	02:50
Control Test		0.082	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:51

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 02/07/2022 03:47:50

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.082	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.000	03:43
Breath Volume		4.625	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.000	03:46
Breath Volume		4.644	03:46
Air Blank		0.000	03:46
Control Test		0.082	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 15:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/07/2022 17:11:55

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.081	17:06
Air Blank		0.000	17:06
Subject Sample	# 1	0.000	17:07
Breath Volume		2.574	17:07
Air Blank		0.000	17:07
Air Blank		0.000	17:09
Subject Sample	# 2	0.000	17:10
Breath Volume		1.910	17:10
Air Blank		0.000	17:11
Control Test		0.082	17:11
Air Blank		0.000	17:11
Diagnostics Check		OK	17:11

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARGROVE,ERIK,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HARGROVE,ERIK
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: KEANE

DATE/TIME OF TEST: 02/08/2022 22:05:34

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.081	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.253	22:00
Breath Volume		4.363	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.246	22:04
Breath Volume		3.984	22:04
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CUE

DATE/TIME OF TEST: 02/11/2022 03:12:57

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.082	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.154	03:08
Breath Volume		1.988	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.149	03:11
Breath Volume		1.921	03:11
Air Blank		0.000	03:12
Control Test		0.082	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHELL,GARY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MOND

DATE/TIME OF TEST: 02/12/2022 04:11:48

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.081	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2		04:10	Subject Test Refused
Breath Volume		0.000	04:10	
Air Blank		0.000	04:10	
Control Test		0.082	04:11	
Air Blank		0.000	04:11	
Diagnostics Check		OK	04:11	

Data Download Date/Time: 02/25/2022 14:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHELL,GARY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 02/12/2022 22:11:55

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.081	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.100	22:06
Breath Volume		1.304	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.100	22:10
Breath Volume		2.667	22:10
Air Blank		0.000	22:11
Control Test		0.081	22:11
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARTER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: TAYLOR,JONATHAN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 02/16/2022 01:13:57

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.082	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.300	01:09
Breath Volume		2.062	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:11
Subject Sample	# 2	0.303	01:12
Breath Volume		2.320	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,KENNETH,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PHILLIPS,KENNETH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: IACOVINO

DATE/TIME OF TEST: 02/18/2022 17:16:57

Diagnostics Check		OK	17:10
Air Blank		0.000	17:11
Control Test		0.081	17:11
Air Blank		0.000	17:12
Subject Sample	# 1	0.000	17:12
Breath Volume		2.863	17:12
Air Blank		0.000	17:13
Air Blank		0.000	17:15
Subject Sample	# 2	0.000	17:15
Breath Volume		3.050	17:15
Air Blank		0.000	17:16
Control Test		0.080	17:16
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ROBERTS,BARRY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MORNINGSTAR

DATE/TIME OF TEST: 02/19/2022 01:18:27

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.082	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.176	01:13
Breath Volume		1.738	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.171	01:16
Breath Volume		1.925	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NORRIS,CHASE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 02/19/2022 03:03:39

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.083	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.174	02:59
Breath Volume		1.929	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.172	03:02
Breath Volume		2.828	03:02
Air Blank		0.000	03:02
Control Test		0.082	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ARAFA,YUSUF
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: FENNER

DATE/TIME OF TEST: 02/19/2022 03:19:31

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.083	03:15
Breath Volume		2.398	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.084	03:18
Breath Volume		2.410	03:18
Air Blank		0.000	03:18
Control Test		0.082	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,T A
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 02/19/2022 04:46:06

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:38	
Control Test		0.082	04:38	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:42	No Sample Provided
Breath Volume		0.000	04:42	
Air Blank		0.000	04:42	
Air Blank		0.000	04:44	
Subject Sample	# 2		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Control Test		0.082	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 05:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: TRACY II

DATE/TIME OF TEST: 02/19/2022 06:43:13

Diagnostics Check		OK	06:37
Air Blank		0.000	06:37
Control Test		0.082	06:37
Air Blank		0.000	06:38
Subject Sample	# 1	0.087	06:38
Breath Volume		3.519	06:38
Air Blank		0.000	06:39
Air Blank		0.000	06:41
Subject Sample	# 2	0.086	06:41
Breath Volume		3.171	06:41
Air Blank		0.000	06:42
Control Test		0.083	06:42
Air Blank		0.000	06:43
Diagnostics Check		OK	06:43

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: VERGE

DATE/TIME OF TEST: 02/24/2022 04:15:37

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.084	04:10
Breath Volume		1.871	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.091	04:14
Breath Volume		1.863	04:14
Air Blank		0.000	04:14
Control Test		0.081	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 02/25/2022 14:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,KENNETH,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PHILLIPS,KENNETH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: COURT ORD
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: IACOVINO

DATE/TIME OF TEST: 02/28/2022 18:10:55

Diagnostics Check		OK	18:04
Air Blank		0.000	18:05
Control Test		0.081	18:05
Air Blank		0.000	18:06
Subject Sample	# 1	0.000	18:06
Breath Volume		2.753	18:06
Air Blank		0.000	18:07
Air Blank		0.000	18:09
Subject Sample	# 2	0.000	18:09
Breath Volume		3.019	18:09
Air Blank		0.000	18:10
Control Test		0.081	18:10
Air Blank		0.000	18:10
Diagnostics Check		OK	18:10

Data Download Date/Time: 03/09/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KITTRELL,WILLIAM,G
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: PEREZ CHUJ

DATE/TIME OF TEST: 02/13/2022 00:29:49

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.205	00:25
Breath Volume		3.324	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.200	00:28
Breath Volume		3.425	00:28
Air Blank		0.000	00:28
Control Test		0.079	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 02/16/2022 15:45

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 02/27/2022 20:01:26

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:59	
Control Test		0.079	19:59	
Air Blank		0.000	19:59	
Subject Sample	# 1		20:00	Subject Test Refused
Breath Volume		0.000	20:00	
Air Blank		0.000	20:00	
Control Test		0.080	20:00	
Air Blank		0.000	20:01	
Diagnostics Check		OK	20:01	

Data Download Date/Time: 03/21/2022 19:39

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CASIAN,SKYLAR
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: MILEK

DATE/TIME OF TEST: 02/03/2022 23:15:04

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:12	
Control Test		0.081	23:13	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:13	Subject Test Refused
Breath Volume		0.000	23:13	
Air Blank		0.000	23:14	
Control Test		0.081	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 02/28/2022 13:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/04/2022 23:49:49

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:47	
Control Test		0.079	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:48	Subject Test Refused
Breath Volume		0.000	23:48	
Air Blank		0.000	23:48	
Control Test		0.079	23:49	
Air Blank		0.000	23:49	
Diagnostics Check		OK	23:49	

Data Download Date/Time: 02/28/2022 13:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/05/2022 02:01:08

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.081	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 02/28/2022 13:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CASIAN,SKYLAR
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 02/06/2022 04:00:49

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.081	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 02/28/2022 13:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 02/13/2022 22:01:14

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:58	
Control Test		0.081	21:59	
Air Blank		0.000	21:59	
Subject Sample	# 1		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	22:00	
Control Test		0.081	22:00	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 02/28/2022 13:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 02/19/2022 02:00:22

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.104	01:55
Breath Volume		3.566	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.097	01:58
Breath Volume		2.742	01:58
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 02/28/2022 13:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BEDDOE

DATE/TIME OF TEST: 02/19/2022 03:45:58

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.211	03:41
Breath Volume		2.472	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.204	03:44
Breath Volume		3.523	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO,MERCEDES,T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: BUNTING,JESSICA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MENDOZACHAVEZ

DATE/TIME OF TEST: 02/20/2022 02:54:33

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.177	02:49
Breath Volume		2.140	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.166	02:52
Breath Volume		1.105	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: BIELEFELDT

DATE/TIME OF TEST: 02/20/2022 20:07:09

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.080	20:00
Air Blank		0.000	20:01
Subject Sample	# 1	0.110	20:02
Breath Volume		1.578	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:04
Subject Sample	# 2	0.101	20:05
Breath Volume		2.289	20:05
Air Blank		0.000	20:06
Control Test		0.081	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ABDELQUADER,DYLAN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 02/22/2022 02:19:50

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.081	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.081	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO,MERCEDES,T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: BUNTING,JESSICA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: PYLE

DATE/TIME OF TEST: 02/23/2022 22:10:29

Diagnostics Check		OK	22:07	
Air Blank		0.000	22:08	
Control Test		0.081	22:08	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:09	
Control Test		0.081	22:09	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: LOPES-GONZALEZ

DATE/TIME OF TEST: 02/25/2022 22:02:48

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.182	21:57
Breath Volume		1.875	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.192	22:01
Breath Volume		2.031	22:01
Air Blank		0.000	22:01
Control Test		0.080	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 02/26/2022 03:12:26

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.082	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Control Test		0.081	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 02/26/2022 20:59:11

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.093	20:53
Breath Volume		1.714	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:55
Subject Sample	# 2	0.090	20:57
Breath Volume		1.683	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: CARRIER

DATE/TIME OF TEST: 02/27/2022 00:16:01

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.081	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.108	00:11
Breath Volume		2.003	00:11
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.113	00:14
Breath Volume		1.773	00:14
Air Blank		0.000	00:15
Control Test		0.081	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:16

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: ZEIMET

DATE/TIME OF TEST: 02/27/2022 03:16:09

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.154	03:11
Breath Volume		3.246	03:11
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.151	03:14
Breath Volume		2.441	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CASIAN,SKYLAR
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: RATCLIFF

DATE/TIME OF TEST: 02/27/2022 16:55:54

Diagnostics Check		OK	16:52	
Air Blank		0.000	16:53	
Control Test		0.080	16:53	
Air Blank		0.000	16:54	
Subject Sample	# 1		16:54	Subject Test Refused
Breath Volume		0.000	16:54	
Air Blank		0.000	16:54	
Control Test		0.080	16:55	
Air Blank		0.000	16:55	
Diagnostics Check		OK	16:55	

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ABDELQUADER,DYLAN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: GODDARD

DATE/TIME OF TEST: 02/27/2022 18:09:35

Diagnostics Check		OK	18:02
Air Blank		0.000	18:03
Control Test		0.081	18:03
Air Blank		0.000	18:04
Subject Sample	# 1	0.000	18:04
Breath Volume		1.285	18:04
Air Blank		0.000	18:05
Air Blank		0.000	18:07
Subject Sample	# 2	0.000	18:08
Breath Volume		1.746	18:08
Air Blank		0.000	18:08
Control Test		0.080	18:08
Air Blank		0.000	18:09
Diagnostics Check		OK	18:09

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/17/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 02/27/2022 21:39:44

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:37	
Control Test		0.080	21:37	
Air Blank		0.000	21:38	
Subject Sample	# 1		21:38	Subject Test Refused
Breath Volume		0.000	21:38	
Air Blank		0.000	21:38	
Control Test		0.080	21:39	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 02/28/2022 13:29

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,BENJAMIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 02/01/2022 12:17:33

Diagnostics Check		OK	12:10
Air Blank		0.000	12:10
Control Test		0.081	12:11
Air Blank		0.000	12:11
Subject Sample	# 1	0.187	12:12
Breath Volume		1.246	12:12
Air Blank		0.000	12:13
Air Blank		0.000	12:15
Subject Sample	# 2	0.183	12:16
Breath Volume		1.683	12:16
Air Blank		0.000	12:16
Control Test		0.081	12:16
Air Blank		0.000	12:17
Diagnostics Check		OK	12:17

Data Download Date/Time: 02/28/2022 16:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CODY,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BROWN,CODY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ARMSTEAD

DATE/TIME OF TEST: 02/02/2022 01:47:05

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.080	01:41	
Air Blank		0.000	01:41	
Subject Sample	# 1	0.156	01:42	
Breath Volume		1.332	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 2		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:46	
Control Test		0.079	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 02/28/2022 16:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CRAWFORD,BRENT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: MACKENZIE

DATE/TIME OF TEST: 02/05/2022 22:49:02

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.080	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.134	22:44
Breath Volume		1.707	22:44
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.139	22:47
Breath Volume		1.441	22:47
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:48
Diagnostics Check		OK	22:49

Data Download Date/Time: 02/28/2022 16:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONCALVES,JOAO,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: FALEIROS,JULIANO
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: KIMBALL

DATE/TIME OF TEST: 02/12/2022 20:14:18

Diagnostics Check		OK	20:06
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.149	20:09
Breath Volume		3.402	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:12
Subject Sample	# 2	0.131	20:12
Breath Volume		1.640	20:12
Air Blank		0.000	20:13
Control Test		0.078	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 02/28/2022 16:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,BENJAMIN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: PRIBYL

DATE/TIME OF TEST: 02/14/2022 16:22:39

Diagnostics Check		OK	16:16
Air Blank		0.000	16:16
Control Test		0.081	16:17
Air Blank		0.000	16:17
Subject Sample	# 1	0.000	16:18
Breath Volume		1.210	16:18
Air Blank		0.000	16:18
Air Blank		0.000	16:20
Subject Sample	# 2	0.000	16:21
Breath Volume		1.113	16:21
Air Blank		0.000	16:21
Control Test		0.081	16:22
Air Blank		0.000	16:22
Diagnostics Check		OK	16:22

Data Download Date/Time: 02/28/2022 16:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,B
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: HABERLANDT

DATE/TIME OF TEST: 02/20/2022 06:52:36

Diagnostics Check		OK	06:48	
Air Blank		0.000	06:49	
Control Test		0.081	06:49	
Air Blank		0.000	06:50	
Subject Sample	# 1	0.118	06:50	
Breath Volume		1.886	06:50	
Air Blank			06:51	RFI Detect
Air Blank		0.000	06:51	
Control Test		0.080	06:52	
Air Blank		0.000	06:52	
Diagnostics Check		OK	06:52	

Data Download Date/Time: 02/28/2022 16:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GASPAR,B
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: HABERLANDT

DATE/TIME OF TEST: 02/20/2022 07:01:10

Diagnostics Check		OK	06:54
Air Blank		0.000	06:55
Control Test		0.081	06:55
Air Blank		0.000	06:56
Subject Sample	# 1	0.114	06:56
Breath Volume		2.003	06:56
Air Blank		0.000	06:57
Air Blank		0.000	06:59
Subject Sample	# 2	0.119	06:59
Breath Volume		1.714	06:59
Air Blank		0.000	07:00
Control Test		0.081	07:00
Air Blank		0.000	07:01
Diagnostics Check		OK	07:01

Data Download Date/Time: 02/28/2022 16:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RODGERS,BRANDON
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 02/23/2022 03:57:38

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.141	03:52
Breath Volume		1.210	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.140	03:56
Breath Volume		1.500	03:56
Air Blank		0.000	03:56
Control Test		0.079	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 02/28/2022 16:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: LIEVANO

DATE/TIME OF TEST: 02/26/2022 22:56:23

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.183	22:51
Breath Volume		1.261	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.189	22:54
Breath Volume		2.011	22:54
Air Blank		0.000	22:55
Control Test		0.080	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 02/28/2022 16:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WOODRUFF,MARK
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MOYLAN

DATE/TIME OF TEST: 02/28/2022 01:00:29

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.231	00:56
Breath Volume		1.878	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.238	00:58
Breath Volume		1.605	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 02/28/2022 16:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: REICHELDERFER

DATE/TIME OF TEST: 02/01/2022 00:56:36

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:45	
Control Test		0.083	00:45	
Air Blank		0.000	00:45	
Subject Sample	# 1	0.181	00:48	Volume Not Met
Breath Volume		0.132	00:48	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2	0.154	00:54	Volume Not Met
Breath Volume		0.101	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 03/01/2022 09:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: REICHEKDERFER

DATE/TIME OF TEST: 02/01/2022 01:17:01

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.079	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1	0.165	01:03	
Breath Volume		1.179	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.193	01:09	Volume Not Met
Breath Volume		0.113	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 3	0.161	01:15	Volume Not Met
Breath Volume		1.066	01:15	
Air Blank		0.000	01:16	
Control Test		0.079	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 03/01/2022 09:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,THERESA,K
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OOLEY,J
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 02/03/2022 16:53:41

Diagnostics Check		OK	16:47
Air Blank		0.000	16:48
Control Test		0.081	16:48
Air Blank		0.000	16:48
Subject Sample	# 1	0.090	16:49
Breath Volume		4.359	16:49
Air Blank		0.000	16:49
Air Blank		0.000	16:51
Subject Sample	# 2	0.090	16:52
Breath Volume		3.964	16:52
Air Blank		0.000	16:52
Control Test		0.081	16:53
Air Blank		0.000	16:53
Diagnostics Check		OK	16:53

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/03/2022 21:14:10

Diagnostics Check		OK	21:02	
Air Blank		0.000	21:02	
Control Test		0.081	21:03	
Air Blank		0.000	21:03	
Subject Sample	# 1	0.072	21:04	
Breath Volume		1.144	21:04	
Air Blank		0.000	21:04	
Air Blank		0.000	21:06	
Subject Sample	# 2	0.067	21:09	Volume Not Met
Breath Volume		0.585	21:09	
Air Blank		0.000	21:10	
Air Blank		0.000	21:12	
Subject Sample	# 3	0.062	21:12	
Breath Volume		1.246	21:12	
Air Blank		0.000	21:13	
Control Test		0.080	21:13	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTMAN,KEVIN,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARACARO,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 02/04/2022 18:00:04

Diagnostics Check		OK	17:53
Air Blank		0.000	17:54
Control Test		0.081	17:54
Air Blank		0.000	17:54
Subject Sample	# 1	0.252	17:55
Breath Volume		2.218	17:55
Air Blank		0.000	17:56
Air Blank		0.000	17:57
Subject Sample	# 2	0.264	17:58
Breath Volume		2.812	17:58
Air Blank		0.000	17:59
Control Test		0.080	17:59
Air Blank		0.000	17:59
Diagnostics Check		OK	18:00

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/04/2022 19:24:37

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:21	
Control Test		0.081	19:22	
Air Blank		0.000	19:22	
Subject Sample	# 1		19:23	Subject Test Refused
Breath Volume		0.000	19:23	
Air Blank		0.000	19:23	
Control Test		0.081	19:24	
Air Blank		0.000	19:24	
Diagnostics Check		OK	19:24	

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MASLAN,JESSE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: KONKO

DATE/TIME OF TEST: 02/05/2022 00:00:32

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:58	
Control Test		0.081	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Control Test		0.082	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PRYOR

DATE/TIME OF TEST: 02/05/2022 03:22:08

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:19	
Control Test		0.081	03:20	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:21	
Control Test		0.082	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTY,K,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MARTY,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: DASRATH

DATE/TIME OF TEST: 02/05/2022 04:35:19

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.134	04:29
Breath Volume		1.371	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:31
Subject Sample	# 2	0.118	04:33
Breath Volume		1.695	04:33
Air Blank		0.000	04:34
Control Test		0.082	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: VICKERS

DATE/TIME OF TEST: 02/05/2022 21:43:07

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.082	21:31	
Air Blank		0.000	21:32	
Subject Sample	# 1	0.195	21:35	Volume Not Met
Breath Volume		0.792	21:35	
Air Blank		0.000	21:36	
Air Blank		0.000	21:38	
Subject Sample	# 2	0.204	21:41	Volume Not Met
Breath Volume		0.796	21:41	
Air Blank		0.000	21:42	
Control Test		0.080	21:42	
Air Blank		0.000	21:43	
Diagnostics Check		OK	21:43	

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: VICKERS

DATE/TIME OF TEST: 02/05/2022 21:53:07

Diagnostics Check		OK	21:45
Air Blank		0.000	21:45
Control Test		0.079	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.223	21:47
Breath Volume		1.121	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:49
Subject Sample	# 2	0.222	21:51
Breath Volume		1.531	21:51
Air Blank		0.000	21:52
Control Test		0.080	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: KAY

DATE/TIME OF TEST: 02/06/2022 01:23:21

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:20	
Control Test		0.082	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:22	
Control Test		0.081	01:22	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,CPL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 02/06/2022 02:07:31

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.110	02:02
Breath Volume		2.664	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.108	02:06
Breath Volume		2.601	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 03/01/2022 09:58

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HARTMAN,RYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 02/06/2022 07:14:46

Diagnostics Check		OK	07:08
Air Blank		0.000	07:08
Control Test		0.079	07:09
Air Blank		0.000	07:09
Subject Sample	# 1	0.098	07:10
Breath Volume		2.347	07:10
Air Blank		0.000	07:10
Air Blank		0.000	07:12
Subject Sample	# 2	0.104	07:13
Breath Volume		2.515	07:13
Air Blank		0.000	07:13
Control Test		0.079	07:14
Air Blank		0.000	07:14
Diagnostics Check		OK	07:14

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 02/07/2022 01:19:33

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.079	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.093	01:14
Breath Volume		1.246	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.093	01:18
Breath Volume		1.644	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,CPL
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/07/2022 02:37:48

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:35	
Control Test		0.079	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.080	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DEPUTY,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PIZZUTI,OFC.
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BRECHT

DATE/TIME OF TEST: 02/07/2022 03:55:07

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:44	
Subject Sample	# 1	0.182	03:47	Volume Not Met
Breath Volume		0.761	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.156	03:50	
Breath Volume		1.367	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 3	0.163	03:53	
Breath Volume		1.765	03:53	
Air Blank		0.000	03:54	
Control Test		0.078	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 02/08/2022 00:34:58

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.206	00:30
Breath Volume		2.304	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.220	00:33
Breath Volume		1.683	00:33
Air Blank		0.000	00:34
Control Test		0.079	00:34
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROACHE,H
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SLEDJESKI

DATE/TIME OF TEST: 02/09/2022 02:25:30

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Control Test		0.079	02:24	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LARABELL

DATE/TIME OF TEST: 02/10/2022 22:47:48

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.080	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.070	22:43
Breath Volume		1.976	22:43
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 2	0.070	22:46
Breath Volume		1.527	22:46
Air Blank		0.000	22:46
Control Test		0.079	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 02/12/2022 02:51:50

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.079	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARKWELL,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: GARCIA HURTADO

DATE/TIME OF TEST: 02/12/2022 15:40:45

Diagnostics Check		OK	15:37	
Air Blank		0.000	15:38	
Control Test		0.080	15:38	
Air Blank		0.000	15:39	
Subject Sample	# 1		15:39	Subject Test Refused
Breath Volume		0.000	15:39	
Air Blank		0.000	15:39	
Control Test		0.079	15:40	
Air Blank		0.000	15:40	
Diagnostics Check		OK	15:40	

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RICKY,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PEREZ,RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: KARNATZ

DATE/TIME OF TEST: 02/12/2022 18:58:46

Diagnostics Check		OK	18:52
Air Blank		0.000	18:52
Control Test		0.079	18:53
Air Blank		0.000	18:53
Subject Sample	# 1	0.076	18:54
Breath Volume		1.531	18:54
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.075	18:57
Breath Volume		1.597	18:57
Air Blank		0.000	18:57
Control Test		0.079	18:58
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 02/12/2022 23:33:44

Diagnostics Check		OK	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.151	23:28
Breath Volume		1.460	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.151	23:32
Breath Volume		1.828	23:32
Air Blank		0.000	23:32
Control Test		0.079	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 03/01/2022 09:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NEWMAN,P
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: MCDEW

DATE/TIME OF TEST: 02/13/2022 19:19:27

Diagnostics Check		OK	19:13
Air Blank		0.000	19:13
Control Test		0.079	19:14
Air Blank		0.000	19:14
Subject Sample	# 1	0.241	19:14
Breath Volume		1.585	19:14
Air Blank		0.000	19:15
Air Blank		0.000	19:17
Subject Sample	# 2	0.240	19:17
Breath Volume		1.847	19:17
Air Blank		0.000	19:18
Control Test		0.079	19:18
Air Blank		0.000	19:19
Diagnostics Check		OK	19:19

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SELDES,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 02/13/2022 22:59:37

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.080	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.344	22:54
Breath Volume		1.820	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.335	22:57
Breath Volume		1.718	22:57
Air Blank		0.000	22:58
Control Test		0.076	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/14/2022 04:43:31

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:42	
Control Test		0.080	04:42	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINTRON,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RISPOLI,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: ANTHOS

DATE/TIME OF TEST: 02/16/2022 14:13:32

Diagnostics Check		OK	14:10	
Air Blank		0.000	14:11	
Control Test		0.080	14:11	
Air Blank		0.000	14:12	
Subject Sample	# 1		14:12	Subject Test Refused
Breath Volume		0.000	14:12	
Air Blank		0.000	14:12	
Control Test		0.080	14:12	
Air Blank		0.000	14:13	
Diagnostics Check		OK	14:13	

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KESKINEN,D
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 02/18/2022 23:25:49

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.079	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.000	23:21
Breath Volume		2.312	23:21
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.000	23:24
Breath Volume		3.246	23:24
Air Blank		0.000	23:24
Control Test		0.079	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PIZZUTI,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 02/19/2022 02:49:04

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTY,K,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MARTY,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ORNELAS

DATE/TIME OF TEST: 02/19/2022 04:27:44

Diagnostics Check	OK	04:27	
Air Blank		04:27	Improper Sample
Air Blank		04:27	Improper Sample

Data Download Date/Time: 03/01/2022 10:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTY,K,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MARTY,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ORNELAS

DATE/TIME OF TEST: 02/19/2022 04:30:46

Diagnostics Check	OK	04:30	
Air Blank		04:30	Improper Sample
Air Blank	0.000	04:30	

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTY,K,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MARTY,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ORNELAS

DATE/TIME OF TEST: 02/19/2022 04:39:14

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.154	04:34
Breath Volume		2.074	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.148	04:37
Breath Volume		1.375	04:37
Air Blank		0.000	04:38
Control Test		0.078	04:38
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 02/19/2022 19:05:14

Diagnostics Check		OK	18:58
Air Blank		0.000	18:59
Control Test		0.079	18:59
Air Blank		0.000	19:00
Subject Sample	# 1	0.000	19:00
Breath Volume		2.582	19:00
Air Blank		0.000	19:01
Air Blank		0.000	19:03
Subject Sample	# 2	0.000	19:03
Breath Volume		2.917	19:03
Air Blank		0.000	19:04
Control Test		0.079	19:04
Air Blank		0.000	19:05
Diagnostics Check		OK	19:05

Data Download Date/Time: 03/01/2022 10:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MARROQUIN,RICHARD
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: BUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: HEWITT

DATE/TIME OF TEST: 02/19/2022 20:25:29

Diagnostics Check		OK	20:19
Air Blank		0.000	20:19
Control Test		0.080	20:20
Air Blank		0.000	20:20
Subject Sample	# 1	0.000	20:21
Breath Volume		3.101	20:21
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.000	20:24
Breath Volume		2.199	20:24
Air Blank		0.000	20:24
Control Test		0.080	20:24
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 03/01/2022 10:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RADOCK,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 02/20/2022 00:23:39

Diagnostics Check		OK	00:15
Air Blank		0.000	00:16
Control Test		0.081	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.131	00:18
Breath Volume		1.785	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:20
Subject Sample	# 2	0.133	00:22
Breath Volume		1.980	00:22
Air Blank		0.000	00:22
Control Test		0.079	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 03/01/2022 10:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 02/20/2022 03:33:59

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 03/01/2022 11:55

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SZOTT,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: BLATCH

DATE/TIME OF TEST: 02/20/2022 04:27:06

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.228	04:21
Breath Volume		1.554	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.228	04:25
Breath Volume		1.703	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 03/01/2022 11:55

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOUINARD,JEFFREY,M
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CHOUINARD,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 08:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/20/2022 11:05:59

Diagnostics Check		OK	10:50	
Air Blank		0.000	10:50	
Control Test		0.079	10:50	
Air Blank		0.000	10:51	
Subject Sample	# 1	0.154	10:54	Volume Not Met
Breath Volume		0.613	10:54	
Air Blank		0.000	10:55	
Air Blank		0.000	10:57	
Subject Sample	# 2	0.168	10:58	
Breath Volume		1.386	10:58	
Air Blank		0.000	10:59	
Air Blank		0.000	11:01	
Subject Sample	# 3	0.124	11:04	Volume Not Met
Breath Volume		0.515	11:04	
Air Blank		0.000	11:05	
Control Test		0.078	11:05	
Air Blank		0.000	11:05	
Diagnostics Check		OK	11:05	

Data Download Date/Time: 03/01/2022 11:55

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211953
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GUTIERREZ-HERNANDEZ

DATE/TIME OF TEST: 02/21/2022 00:41:50

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.123	00:37
Breath Volume		2.007	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:39
Subject Sample	# 2	0.132	00:40
Breath Volume		2.757	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 03/01/2022 11:55

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: LI,QUANG
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 192223
DRY GAS EXPIRY DATE: 03/19/2023
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 02/24/2022 16:01:55

Diagnostics Check		OK	15:56	
Air Blank		0.000	15:56	
Control Test		0.080	15:57	
Air Blank		0.000	15:57	
Subject Sample	# 1		16:00	Subject Test Refused
Breath Volume		0.148	16:00	
Air Blank		0.000	16:01	
Control Test		0.080	16:01	
Air Blank		0.000	16:01	
Diagnostics Check		OK	16:01	

Data Download Date/Time: 03/01/2022 11:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/27/2022 03:01:59

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:01	
Control Test		0.080	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 03/01/2022 11:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: HANNA

DATE/TIME OF TEST: 02/27/2022 18:43:01

Diagnostics Check		OK	18:36
Air Blank		0.000	18:37
Control Test		0.079	18:37
Air Blank		0.000	18:37
Subject Sample	# 1	0.155	18:38
Breath Volume		1.710	18:38
Air Blank		0.000	18:39
Air Blank		0.000	18:40
Subject Sample	# 2	0.151	18:41
Breath Volume		1.675	18:41
Air Blank		0.000	18:42
Control Test		0.078	18:42
Air Blank		0.000	18:42
Diagnostics Check		OK	18:43

Data Download Date/Time: 03/01/2022 11:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: AGUILAR,R
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MAGGIACOMO

DATE/TIME OF TEST: 02/27/2022 23:57:56

Diagnostics Check		OK	23:51
Air Blank		0.000	23:52
Control Test		0.080	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.171	23:53
Breath Volume		1.468	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:55
Subject Sample	# 2	0.171	23:56
Breath Volume		1.953	23:56
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 03/01/2022 11:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/28/2022 01:33:38

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.080	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.170	01:29
Breath Volume		1.632	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.168	01:32
Breath Volume		1.949	01:32
Air Blank		0.000	01:32
Control Test		0.079	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 03/01/2022 11:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRANDES,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/28/2022 17:01:31

Diagnostics Check		OK	16:55	
Air Blank		0.000	16:56	
Control Test		0.080	16:56	
Air Blank		0.000	16:56	
Subject Sample	# 1		17:00	Subject Test Refused
Breath Volume		0.531	17:00	
Air Blank		0.000	17:00	
Control Test		0.080	17:00	
Air Blank		0.000	17:01	
Diagnostics Check		OK	17:01	

Data Download Date/Time: 03/01/2022 11:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WHITT

DATE/TIME OF TEST: 02/28/2022 23:16:42

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:11	
Control Test		0.080	23:12	
Air Blank		0.000	23:12	
Subject Sample	# 1		23:15	Subject Test Refused
Breath Volume		0.000	23:15	
Air Blank		0.000	23:15	
Control Test		0.080	23:16	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 03/01/2022 11:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KRONDAHL,CARL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 02/28/2022 01:32:28

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.000	01:28
Breath Volume		3.917	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.000	01:31
Breath Volume		3.851	01:31
Air Blank		0.000	01:31
Control Test		0.082	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 04/19/2022 09:31

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: KENDALL

DATE/TIME OF TEST: 02/28/2022 02:41:30

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.080	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 04/19/2022 09:31

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIETRON,M
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 02/01/2022 05:09:11

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.081	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.123	05:04
Breath Volume		1.800	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.124	05:07
Breath Volume		2.039	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 02/03/2022 17:38

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: VOLAK

DATE/TIME OF TEST: 02/02/2022 00:47:29

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.183	00:43
Breath Volume		3.929	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.180	00:46
Breath Volume		3.355	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 02/03/2022 17:38

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: REICH

DATE/TIME OF TEST: 02/03/2022 17:23:34

Diagnostics Check		OK	17:17
Air Blank		0.000	17:17
Control Test		0.080	17:18
Air Blank		0.000	17:18
Subject Sample	# 1	0.000	17:19
Breath Volume		3.480	17:19
Air Blank		0.000	17:19
Air Blank		0.000	17:21
Subject Sample	# 2	0.000	17:22
Breath Volume		2.867	17:22
Air Blank		0.000	17:22
Control Test		0.080	17:22
Air Blank		0.000	17:23
Diagnostics Check		OK	17:23

Data Download Date/Time: 02/03/2022 17:38

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 02/04/2022 01:40:14

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Subject Sample	# 1	0.161	01:35
Breath Volume		1.250	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.161	01:38
Breath Volume		1.406	01:38
Air Blank		0.000	01:39
Control Test		0.078	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHELTON,PATRICIA,M
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI,OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: WHELTON

DATE/TIME OF TEST: 02/04/2022 18:26:34

Diagnostics Check		OK	18:20
Air Blank		0.000	18:20
Control Test		0.079	18:21
Air Blank		0.000	18:21
Subject Sample	# 1	0.175	18:22
Breath Volume		1.152	18:22
Air Blank		0.000	18:22
Air Blank		0.000	18:24
Subject Sample	# 2	0.174	18:25
Breath Volume		1.589	18:25
Air Blank		0.000	18:25
Control Test		0.078	18:25
Air Blank		0.000	18:26
Diagnostics Check		OK	18:26

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLADVANDI,OFC. M.,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: FOLADVANDI,OFC. M.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: WHELTON

DATE/TIME OF TEST: 02/04/2022 18:49:25

Diagnostics Check		OK	18:42
Air Blank		0.000	18:42
Control Test		0.079	18:43
Air Blank		0.000	18:43
Subject Sample	# 1	0.164	18:45
Breath Volume		1.562	18:45
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.167	18:47
Breath Volume		1.531	18:47
Air Blank		0.000	18:48
Control Test		0.078	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: VERDUGO-ORTIZ

DATE/TIME OF TEST: 02/04/2022 23:35:06

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.129	23:30
Breath Volume		4.156	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.128	23:33
Breath Volume		4.089	23:33
Air Blank		0.000	23:34
Control Test		0.079	23:34
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MITCHELL,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KIRKPATRICK

DATE/TIME OF TEST: 02/05/2022 02:44:13

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.080	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1	0.119	02:34	Volume Not Met
Breath Volume		0.789	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.117	02:39	
Breath Volume		1.328	02:39	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 3	0.109	02:42	
Breath Volume		1.644	02:42	
Air Blank		0.000	02:43	
Control Test		0.077	02:43	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIETRON,OFFICER M.
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: AFFINITO

DATE/TIME OF TEST: 02/05/2022 02:58:27

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.079	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.118	02:54
Breath Volume		1.753	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.115	02:56
Breath Volume		1.789	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: PICKETT

DATE/TIME OF TEST: 02/05/2022 04:01:52

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.146	03:57
Breath Volume		1.718	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.137	04:00
Breath Volume		2.832	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: JANOTKA

DATE/TIME OF TEST: 02/06/2022 02:56:35

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.182	02:52
Breath Volume		2.402	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.183	02:55
Breath Volume		2.621	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KLEIN,OFFICER B.
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 02/06/2022 03:27:54

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.164	03:23
Breath Volume		2.960	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.158	03:26
Breath Volume		1.816	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: TOMKINSON

DATE/TIME OF TEST: 02/06/2022 19:32:31

Diagnostics Check		OK	19:21	
Air Blank		0.000	19:21	
Control Test		0.079	19:21	
Air Blank		0.000	19:22	
Subject Sample	# 1	0.213	19:25	Volume Not Met
Breath Volume		0.015	19:25	
Air Blank		0.000	19:26	
Air Blank		0.000	19:27	
Subject Sample	# 2	0.141	19:31	Volume Not Met
Breath Volume		0.058	19:31	
Air Blank		0.000	19:31	
Control Test		0.077	19:31	
Air Blank		0.000	19:32	
Diagnostics Check		OK	19:32	

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BENTZ,WHITNEY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 02/07/2022 06:31:01

Diagnostics Check		OK	06:24
Air Blank		0.000	06:25
Control Test		0.079	06:25
Air Blank		0.000	06:26
Subject Sample	# 1	0.149	06:26
Breath Volume		1.492	06:26
Air Blank		0.000	06:27
Air Blank		0.000	06:29
Subject Sample	# 2	0.151	06:29
Breath Volume		1.722	06:29
Air Blank		0.000	06:30
Control Test		0.079	06:30
Air Blank		0.000	06:30
Diagnostics Check		OK	06:31

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/09/2022 20:23:05

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:11	
Control Test		0.079	20:12	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:15	No Sample Provided
Breath Volume		0.000	20:15	
Air Blank		0.000	20:16	
Air Blank		0.000	20:18	
Subject Sample	# 2	0.013	20:21	Volume Not Met
Breath Volume		0.117	20:21	
Air Blank		0.000	20:22	
Control Test		0.080	20:22	
Air Blank		0.000	20:22	
Diagnostics Check		OK	20:23	

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: PEREIRA-ROCHA

DATE/TIME OF TEST: 02/11/2022 04:14:18

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.185	04:09
Breath Volume		1.347	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.196	04:12
Breath Volume		1.351	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TAYLOR,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: DEMERS

DATE/TIME OF TEST: 02/12/2022 18:30:37

Diagnostics Check		OK	18:18	
Air Blank		0.000	18:18	
Control Test		0.080	18:18	
Air Blank		0.000	18:19	
Subject Sample	# 1	0.000	18:22	Volume Not Met
Breath Volume		0.171	18:22	
Air Blank		0.000	18:23	
Air Blank		0.000	18:25	
Subject Sample	# 2	0.000	18:25	
Breath Volume		1.542	18:25	
Air Blank		0.000	18:26	
Air Blank		0.000	18:28	
Subject Sample	# 3	0.000	18:29	
Breath Volume		1.523	18:29	
Air Blank		0.000	18:29	
Control Test		0.079	18:30	
Air Blank		0.000	18:30	
Diagnostics Check		OK	18:30	

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ROMAGNOLO

DATE/TIME OF TEST: 02/13/2022 03:14:06

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.066	03:09
Breath Volume		2.113	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.067	03:12
Breath Volume		2.429	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDOREK,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: ORMEDO

DATE/TIME OF TEST: 02/13/2022 06:32:44

Diagnostics Check		OK	06:26
Air Blank		0.000	06:27
Control Test		0.079	06:27
Air Blank		0.000	06:27
Subject Sample	# 1	0.139	06:28
Breath Volume		1.687	06:28
Air Blank		0.000	06:29
Air Blank		0.000	06:30
Subject Sample	# 2	0.136	06:31
Breath Volume		2.070	06:31
Air Blank		0.000	06:31
Control Test		0.079	06:32
Air Blank		0.000	06:32
Diagnostics Check		OK	06:32

Data Download Date/Time: 02/13/2022 10:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 02/13/2022 23:48:53

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.080	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.237	23:44
Breath Volume		2.828	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:46
Subject Sample	# 2	0.232	23:47
Breath Volume		1.878	23:47
Air Blank		0.000	23:47
Control Test		0.078	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: CARSTENS

DATE/TIME OF TEST: 02/14/2022 01:31:58

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.146	01:27
Breath Volume		1.789	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.147	01:30
Breath Volume		2.031	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/16/2022 01:10:32

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.080	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: GASKELL

DATE/TIME OF TEST: 02/18/2022 22:10:59

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:57	
Control Test		0.079	21:57	
Air Blank		0.000	21:58	
Subject Sample	# 1	0.161	22:01	Volume Not Met
Breath Volume		0.414	22:01	
Air Blank		0.000	22:01	
Air Blank		0.000	22:03	
Subject Sample	# 2	0.176	22:06	
Breath Volume		1.605	22:06	
Air Blank		0.000	22:06	
Air Blank		0.000	22:08	
Subject Sample	# 3	0.168	22:09	
Breath Volume		2.257	22:09	
Air Blank		0.000	22:10	
Control Test		0.077	22:10	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: PANSINI

DATE/TIME OF TEST: 02/19/2022 00:38:01

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.000	00:33
Breath Volume		2.542	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.000	00:36
Breath Volume		2.335	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:38

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/20/2022 03:18:04

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.196	03:13
Breath Volume		3.332	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.185	03:16
Breath Volume		1.898	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FRANCHUK

DATE/TIME OF TEST: 02/20/2022 04:52:11

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.081	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.192	04:47
Breath Volume		1.347	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.192	04:50
Breath Volume		1.968	04:50
Air Blank		0.000	04:51
Control Test		0.081	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 02/20/2022 23:39:48

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.079	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.160	23:34
Breath Volume		1.429	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.144	23:38
Breath Volume		1.105	23:38
Air Blank		0.000	23:38
Control Test		0.080	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 02/21/2022 04:15:43

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.245	04:11
Breath Volume		2.859	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.240	04:14
Breath Volume		3.195	04:14
Air Blank		0.000	04:14
Control Test		0.079	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIETRON,MICHAEL
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 02/21/2022 04:58:30

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.081	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.127	04:54
Breath Volume		2.664	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.123	04:57
Breath Volume		2.843	04:57
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GILLESPIE

DATE/TIME OF TEST: 02/22/2022 01:07:53

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.183	01:03
Breath Volume		1.917	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.184	01:06
Breath Volume		1.285	01:06
Air Blank		0.000	01:06
Control Test		0.078	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 02/22/2022 02:12:09

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.200	02:07
Breath Volume		2.242	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.200	02:10
Breath Volume		2.125	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 03/13/2022 11:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: GLICK

DATE/TIME OF TEST: 02/22/2022 18:58:12

Diagnostics Check		OK	18:52
Air Blank		0.000	18:52
Control Test		0.079	18:52
Air Blank		0.000	18:53
Subject Sample	# 1	0.000	18:54
Breath Volume		2.812	18:54
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.000	18:56
Breath Volume		1.652	18:56
Air Blank		0.000	18:57
Control Test		0.079	18:57
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: QUINN-JONES

DATE/TIME OF TEST: 02/22/2022 19:34:59

Diagnostics Check		OK	19:27
Air Blank		0.000	19:28
Control Test		0.080	19:28
Air Blank		0.000	19:29
Subject Sample	# 1	0.210	19:30
Breath Volume		3.300	19:30
Air Blank		0.000	19:30
Air Blank		0.000	19:32
Subject Sample	# 2	0.206	19:33
Breath Volume		3.199	19:33
Air Blank		0.000	19:34
Control Test		0.079	19:34
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DOMINIS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: GAUTHIER

DATE/TIME OF TEST: 02/23/2022 12:52:38

Diagnostics Check		OK	12:47	
Air Blank		0.000	12:47	
Control Test		0.081	12:47	
Air Blank		0.000	12:48	
Subject Sample	# 1		12:51	Subject Test Refused
Breath Volume		0.000	12:51	
Air Blank		0.000	12:51	
Control Test		0.080	12:52	
Air Blank		0.000	12:52	
Diagnostics Check		OK	12:52	

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DOMINIS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: BUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: GAUTHIER

DATE/TIME OF TEST: 02/23/2022 13:25:19

Diagnostics Check		OK	13:18
Air Blank		0.000	13:19
Control Test		0.080	13:19
Air Blank		0.000	13:20
Subject Sample	# 1	0.188	13:20
Breath Volume		1.199	13:20
Air Blank		0.000	13:21
Air Blank		0.000	13:23
Subject Sample	# 2	0.186	13:23
Breath Volume		1.289	13:23
Air Blank		0.000	13:24
Control Test		0.078	13:24
Air Blank		0.000	13:25
Diagnostics Check		OK	13:25

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DUVALL

DATE/TIME OF TEST: 02/24/2022 00:48:25

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.154	00:43
Breath Volume		2.886	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.149	00:46
Breath Volume		2.890	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/24/2022 01:34:11

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.000	01:29
Breath Volume		2.328	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.000	01:32
Breath Volume		1.757	01:32
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GERKEN

DATE/TIME OF TEST: 02/24/2022 03:26:56

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.178	03:22
Breath Volume		2.179	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:24
Subject Sample	# 2	0.171	03:25
Breath Volume		2.734	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ESQUIVEL MARQUEZ

DATE/TIME OF TEST: 02/25/2022 01:53:54

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.080	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.166	01:49
Breath Volume		1.136	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.162	01:52
Breath Volume		1.460	01:52
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JANKOWSKI,OFFICER
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PIZZARDI

DATE/TIME OF TEST: 02/25/2022 05:14:31

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.079	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.133	05:10
Breath Volume		3.976	05:10
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.126	05:13
Breath Volume		2.964	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BALES,T
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PULLIAM

DATE/TIME OF TEST: 02/26/2022 02:57:38

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.079	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.206	02:53
Breath Volume		3.308	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.191	02:56
Breath Volume		2.507	02:56
Air Blank		0.000	02:56
Control Test		0.075	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STEVENS,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: CALDERON ROBLERO

DATE/TIME OF TEST: 02/26/2022 04:49:03

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.201	04:44
Breath Volume		2.308	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.201	04:47
Breath Volume		2.089	04:47
Air Blank		0.000	04:48
Control Test		0.079	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:49

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SWEAT

DATE/TIME OF TEST: 02/27/2022 00:51:35

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.141	00:47
Breath Volume		2.710	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.141	00:50
Breath Volume		2.488	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GREGORIO,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: PAPINEAU

DATE/TIME OF TEST: 02/27/2022 04:42:15

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.080	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.201	04:37
Breath Volume		2.289	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.199	04:40
Breath Volume		1.234	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,HECTOR,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: MAKI

DATE/TIME OF TEST: 02/05/2022 22:50:54

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:39	
Control Test		0.080	22:40	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:43	No Sample Provided
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Air Blank		0.000	22:46	
Subject Sample	# 2		22:49	No Sample Provided
Breath Volume		0.000	22:49	
Air Blank		0.000	22:50	
Control Test		0.082	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 03/03/2022 00:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROVENCHER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF,ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BRACKETT

DATE/TIME OF TEST: 02/18/2022 04:01:01

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.233	03:56
Breath Volume		1.457	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.229	03:59
Breath Volume		1.640	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:01

Data Download Date/Time: 03/03/2022 00:53

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JAMIE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MENDEZ,JAMIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/19/2022 07:43:36

Diagnostics Check		OK	07:35
Air Blank		0.000	07:36
Control Test		0.081	07:36
Air Blank		0.000	07:37
Subject Sample	# 1	0.139	07:38
Breath Volume		1.250	07:38
Air Blank		0.000	07:39
Air Blank		0.000	07:41
Subject Sample	# 2	0.141	07:42
Breath Volume		1.503	07:42
Air Blank		0.000	07:42
Control Test		0.081	07:43
Air Blank		0.000	07:43
Diagnostics Check		OK	07:43

Data Download Date/Time: 03/03/2022 00:53

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JAMIE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MENDEZ,JAMIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/22/2022 02:33:44

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.000	02:29
Breath Volume		1.855	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.000	02:32
Breath Volume		2.191	02:32
Air Blank		0.000	02:32
Control Test		0.082	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 03/03/2022 00:53

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCILLA,RICHARD,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MANCILLA,RICHARD
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/28/2022 12:02:52

Diagnostics Check		OK	11:57
Air Blank		0.000	11:57
Control Test		0.079	11:57
Air Blank		0.000	11:58
Subject Sample	# 1	0.000	11:58
Breath Volume		2.910	11:58
Air Blank		0.000	11:59
Air Blank		0.000	12:01
Subject Sample	# 2	0.000	12:01
Breath Volume		3.062	12:01
Air Blank		0.000	12:01
Control Test		0.078	12:02
Air Blank		0.000	12:02
Diagnostics Check		OK	12:02

Data Download Date/Time: 02/28/2022 12:41

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SHERRILL

DATE/TIME OF TEST: 02/01/2022 22:21:42

Diagnostics Check		OK	22:15
Air Blank		0.000	22:16
Control Test		0.081	22:16
Air Blank		0.000	22:16
Subject Sample	# 1	0.091	22:17
Breath Volume		1.277	22:17
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.093	22:20
Breath Volume		1.351	22:20
Air Blank		0.000	22:20
Control Test		0.080	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 02/03/2022 13:32

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKER,DEREK,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: WILLIAMS,GERARD
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 02/01/2022 23:19:16

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.081	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.080	23:18	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 02/03/2022 13:32

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JOSHUA,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SIMMONS,JOHN
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 02/05/2022 20:07:48

Diagnostics Check		OK	20:04	
Air Blank		0.000	20:05	
Control Test		0.081	20:05	
Air Blank		0.000	20:06	
Subject Sample	# 1		20:06	Subject Test Refused
Breath Volume		0.000	20:06	
Air Blank		0.000	20:06	
Control Test		0.081	20:07	
Air Blank		0.000	20:07	
Diagnostics Check		OK	20:07	

Data Download Date/Time: 03/10/2022 18:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 02/07/2022 23:17:36

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.081	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.117	23:12
Breath Volume		3.160	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.107	23:15
Breath Volume		2.539	23:15
Air Blank		0.000	23:16
Control Test		0.080	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 03/10/2022 18:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILLIAMS,JENNIFER
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 02/18/2022 18:45:04

Diagnostics Check		OK	18:41	
Air Blank		0.000	18:42	
Control Test		0.080	18:42	
Air Blank		0.000	18:42	
Subject Sample	# 1		18:43	Interferent Detect
Breath Volume		2.453	18:43	
Air Blank		0.000	18:44	
Control Test		0.080	18:44	
Air Blank		0.000	18:44	
Diagnostics Check		OK	18:45	

Data Download Date/Time: 03/10/2022 18:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILLIAMS,JENNIFER
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 02/18/2022 18:49:47

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:46	
Control Test		0.079	18:47	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:48	Interferent Detect
Breath Volume		2.171	18:48	
Air Blank		0.000	18:48	
Control Test		0.079	18:49	
Air Blank		0.000	18:49	
Diagnostics Check		OK	18:49	

Data Download Date/Time: 03/10/2022 18:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILLIAMS,JENNIFER
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 02/18/2022 18:54:57

Diagnostics Check		OK	18:51	
Air Blank		0.000	18:51	
Control Test		0.079	18:52	
Air Blank		0.000	18:52	
Subject Sample	# 1		18:53	Interferent Detect
Breath Volume		1.777	18:53	
Air Blank		0.000	18:54	
Control Test		0.080	18:54	
Air Blank		0.000	18:54	
Diagnostics Check		OK	18:54	

Data Download Date/Time: 03/10/2022 18:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MCNEIL,MICHAEL
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: WILLS

DATE/TIME OF TEST: 02/18/2022 21:40:38

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.080	21:34
Air Blank		0.000	21:35
Subject Sample	# 1	0.052	21:36
Breath Volume		2.109	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.054	21:39
Breath Volume		2.843	21:39
Air Blank		0.000	21:39
Control Test		0.079	21:40
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 03/10/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,L
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDFAIR,WILLIAM
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 01/31/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 02/26/2022 21:13:35

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.080	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Control Test		0.080	21:12	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 03/10/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FRALEY

DATE/TIME OF TEST: 02/04/2022 23:45:16

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:42	
Control Test		0.079	23:43	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:43	Subject Test Refused
Breath Volume		0.000	23:43	
Air Blank		0.000	23:44	
Control Test		0.080	23:44	
Air Blank		0.000	23:45	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 02/17/2022 00:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/05/2022 01:45:17

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.067	01:44	Control Outside Tolerance
Air Blank	0.000	01:45	

Data Download Date/Time: 02/17/2022 00:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/05/2022 01:55:27

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.080	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:54	
Control Test		0.079	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 02/17/2022 00:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 02/05/2022 02:49:52

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.224	02:45
Breath Volume		1.750	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.221	02:48
Breath Volume		1.339	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 02/05/2022 03:38:03

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.256	03:33
Breath Volume		2.312	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:35
Subject Sample	# 2	0.264	03:36
Breath Volume		1.652	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/05/2022 23:50:29

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.117	23:46
Breath Volume		1.988	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.116	23:48
Breath Volume		1.734	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: LEAL PEREZ

DATE/TIME OF TEST: 02/06/2022 01:31:59

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.140	01:27
Breath Volume		1.285	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:29
Subject Sample	# 2	0.134	01:30
Breath Volume		1.136	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BERNABE

DATE/TIME OF TEST: 02/06/2022 02:16:11

Diagnostics Check		OK	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.175	02:11
Breath Volume		1.238	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.165	02:14
Breath Volume		1.277	02:14
Air Blank		0.000	02:15
Control Test		0.079	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 02/06/2022 02:53:06

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.079	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:51	Subject Test Refused
Breath Volume		0.000	02:51	
Air Blank		0.000	02:51	
Control Test		0.080	02:52	
Air Blank			02:52	RFI Detect
Air Blank		0.000	02:53	

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 02/06/2022 02:56:58

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SFORZA

DATE/TIME OF TEST: 02/06/2022 03:21:09

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.079	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:19	Subject Test Refused
Breath Volume		0.000	03:19	
Air Blank		0.000	03:20	
Control Test		0.079	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 02/17/2022 00:36

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/09/2022 20:31:29

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:26
Subject Sample	# 1	0.000	20:27
Breath Volume		2.015	20:27
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.000	20:30
Breath Volume		1.343	20:30
Air Blank		0.000	20:30
Control Test		0.079	20:30
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 02/17/2022 00:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 02/10/2022 00:01:00

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.000	23:56
Breath Volume		1.800	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.000	23:59
Breath Volume		2.488	23:59
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:01

Data Download Date/Time: 02/17/2022 00:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SABATINO

DATE/TIME OF TEST: 02/12/2022 00:49:27

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.058	00:45
Breath Volume		2.210	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.055	00:47
Breath Volume		3.597	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 02/17/2022 00:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: PIRMAN

DATE/TIME OF TEST: 02/12/2022 01:29:00

Diagnostics Check		OK	01:26	
Air Blank		0.000	01:26	
Control Test		0.079	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:28	
Control Test		0.080	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 02/17/2022 00:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 02/12/2022 02:14:30

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.121	02:10
Breath Volume		1.589	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.124	02:12
Breath Volume		1.242	02:12
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 02/17/2022 00:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 02/12/2022 03:55:34

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.054	03:51
Breath Volume		1.648	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.053	03:54
Breath Volume		1.644	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 02/17/2022 00:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 02/12/2022 23:36:15

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.081	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1	0.168	23:31	
Breath Volume		1.574	23:31	
Air Blank		0.000	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 2	0.158	23:34	
Breath Volume		1.269	23:34	
Air Blank		0.000	23:35	
Control Test		0.079	23:35	
Air Blank			23:36	RFI Detect
Air Blank			23:36	RFI Detect

Data Download Date/Time: 02/17/2022 00:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 02/13/2022 00:40:02

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.079	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:35
Breath Volume		1.476	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.000	00:38
Breath Volume		2.023	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:40

Data Download Date/Time: 02/17/2022 00:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: DOODAN

DATE/TIME OF TEST: 02/13/2022 02:19:30

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.123	02:15
Breath Volume		2.656	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.116	02:17
Breath Volume		1.449	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 02/17/2022 00:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DURAN-PENA

DATE/TIME OF TEST: 02/13/2022 03:31:12

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.223	03:26
Breath Volume		2.156	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:28
Subject Sample	# 2	0.221	03:29
Breath Volume		2.453	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 02/17/2022 00:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: FORTINO III

DATE/TIME OF TEST: 02/18/2022 03:14:47

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.148	03:10
Breath Volume		3.527	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.151	03:13
Breath Volume		3.019	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: NOCE

DATE/TIME OF TEST: 02/23/2022 22:08:16

Diagnostics Check	OK	22:07	
Air Blank		22:07	RFI Detect
Air Blank	0.000	22:08	

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BOVA

DATE/TIME OF TEST: 02/24/2022 03:16:34

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 02/24/2022 04:11:30

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.191	04:07
Breath Volume		2.296	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.187	04:09
Breath Volume		1.687	04:09
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 02/25/2022 02:09:38

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.194	02:05
Breath Volume		3.019	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.180	02:08
Breath Volume		1.839	02:08
Air Blank		0.000	02:08
Control Test		0.078	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/25/2022 04:11:23

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:06
Subject Sample	# 1	0.000	04:06
Breath Volume		1.546	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.000	04:09
Breath Volume		1.734	04:09
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: NIETO

DATE/TIME OF TEST: 02/27/2022 00:12:09

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:09	
Control Test		0.079	00:10	
Air Blank		0.000	00:10	
Subject Sample	# 1		00:10	Subject Test Refused
Breath Volume		0.000	00:10	
Air Blank		0.000	00:11	
Control Test		0.079	00:11	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/27/2022 00:59:54

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.080	00:57	
Air Blank		0.000	00:58	
Subject Sample	# 1		00:58	Subject Test Refused
Breath Volume		0.000	00:58	
Air Blank		0.000	00:58	
Control Test		0.080	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STOESS

DATE/TIME OF TEST: 02/27/2022 01:08:28

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.000	01:04
Breath Volume		2.371	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.000	01:07
Breath Volume		2.152	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 02/27/2022 02:33:13

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.078	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.000	02:28
Breath Volume		1.269	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.000	02:31
Breath Volume		1.113	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 02/27/2022 03:48:52

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.079	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1	0.161	03:44	Volume Not Met
Breath Volume		0.632	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Control Test		0.077	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: GOMEZ GREGORIO

DATE/TIME OF TEST: 02/02/2022 20:41:33

Diagnostics Check		OK	20:35
Air Blank		0.000	20:35
Control Test		0.081	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.183	20:37
Breath Volume		2.500	20:37
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.180	20:40
Breath Volume		2.562	20:40
Air Blank		0.000	20:40
Control Test		0.080	20:40
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LIGHTSEY

DATE/TIME OF TEST: 02/03/2022 03:38:07

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.080	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.135	03:27	Volume Not Met
Breath Volume		0.046	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.157	03:33	
Breath Volume		1.703	03:33	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 3	0.165	03:36	
Breath Volume		1.160	03:36	
Air Blank		0.000	03:37	
Control Test		0.078	03:37	
Air Blank		0.000	03:38	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHILD,MIGUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: ALVARADO GARCIA

DATE/TIME OF TEST: 02/04/2022 01:25:09

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.214	01:20
Breath Volume		2.835	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.211	01:23
Breath Volume		3.683	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 02/04/2022 03:25:21

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.081	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.203	03:21
Breath Volume		1.914	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.208	03:23
Breath Volume		1.613	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ESTEP,CANDACE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/04/2022 05:33:00

Diagnostics Check		OK	05:24
Air Blank		0.000	05:25
Control Test		0.080	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.046	05:28
Breath Volume		1.453	05:28
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.042	05:31
Breath Volume		1.339	05:31
Air Blank		0.000	05:32
Control Test		0.080	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:33

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CURRAN,JASMINE
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: BLALOCK

DATE/TIME OF TEST: 02/04/2022 17:39:08

Diagnostics Check		OK	17:27	
Air Blank		0.000	17:27	
Control Test		0.080	17:28	
Air Blank		0.000	17:28	
Subject Sample	# 1	0.148	17:31	Volume Not Met
Breath Volume		0.175	17:31	
Air Blank		0.000	17:32	
Air Blank		0.000	17:34	
Subject Sample	# 2	0.170	17:34	
Breath Volume		2.562	17:34	
Air Blank		0.000	17:35	
Air Blank		0.000	17:37	
Subject Sample	# 3	0.174	17:37	
Breath Volume		2.800	17:37	
Air Blank		0.000	17:38	
Control Test		0.080	17:38	
Air Blank		0.000	17:39	
Diagnostics Check		OK	17:39	

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZ VELASQUEZ,JORGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: DIAZ TERATOL

DATE/TIME OF TEST: 02/04/2022 18:48:15

Diagnostics Check		OK	18:42
Air Blank		0.000	18:42
Control Test		0.080	18:42
Air Blank		0.000	18:43
Subject Sample	# 1	0.228	18:43
Breath Volume		1.921	18:43
Air Blank		0.000	18:44
Air Blank		0.000	18:46
Subject Sample	# 2	0.232	18:46
Breath Volume		3.742	18:46
Air Blank		0.000	18:47
Control Test		0.080	18:47
Air Blank		0.000	18:48
Diagnostics Check		OK	18:48

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRELAND,DAVID
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WHITEHILL

DATE/TIME OF TEST: 02/05/2022 01:39:30

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.088	01:35
Breath Volume		2.429	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.084	01:38
Breath Volume		2.132	01:38
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLZENDORF,ANDRE
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 02/05/2022 04:21:39

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.136	04:16
Breath Volume		1.292	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.137	04:20
Breath Volume		1.300	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SMITH,RODGER
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LAPUERTA

DATE/TIME OF TEST: 02/05/2022 07:30:38

Diagnostics Check		OK	07:24
Air Blank		0.000	07:24
Control Test		0.080	07:25
Air Blank		0.000	07:25
Subject Sample	# 1	0.085	07:26
Breath Volume		4.808	07:26
Air Blank		0.000	07:26
Air Blank		0.000	07:28
Subject Sample	# 2	0.089	07:29
Breath Volume		4.410	07:29
Air Blank		0.000	07:29
Control Test		0.081	07:30
Air Blank		0.000	07:30
Diagnostics Check		OK	07:30

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BROWN,DARRELL
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/06/2022 22:50:11

Diagnostics Check		OK	22:43
Air Blank		0.000	22:44
Control Test		0.081	22:44
Air Blank		0.000	22:44
Subject Sample	# 1	0.193	22:45
Breath Volume		1.847	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.191	22:48
Breath Volume		1.160	22:48
Air Blank		0.000	22:49
Control Test		0.080	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 02/08/2022 04:47:15

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.081	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.211	04:42
Breath Volume		1.917	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.213	04:45
Breath Volume		1.648	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 02/08/2022 23:07:26

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.082	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.159	23:02
Breath Volume		3.503	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.155	23:05
Breath Volume		3.015	23:05
Air Blank		0.000	23:06
Control Test		0.081	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAKSH,SAEED
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: MERECKI

DATE/TIME OF TEST: 02/11/2022 21:12:25

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.081	21:06
Air Blank		0.000	21:07
Subject Sample	# 1	0.158	21:08
Breath Volume		1.496	21:08
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.167	21:10
Breath Volume		1.406	21:10
Air Blank		0.000	21:11
Control Test		0.081	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: ALICEA JR

DATE/TIME OF TEST: 02/11/2022 23:07:51

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.081	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.214	23:03
Breath Volume		2.156	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:05
Subject Sample	# 2	0.214	23:06
Breath Volume		1.949	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: VENTURA GAMEZ

DATE/TIME OF TEST: 02/12/2022 02:09:01

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.167	01:59	Volume Not Met
Breath Volume		0.957	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 2	0.174	02:03	
Breath Volume		1.234	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 3	0.165	02:07	
Breath Volume		1.687	02:07	
Air Blank		0.000	02:08	
Control Test		0.080	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 03/10/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PROFFER,KATIE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: VILARDI

DATE/TIME OF TEST: 02/12/2022 04:03:20

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.082	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.099	03:58
Breath Volume		3.101	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.086	04:01
Breath Volume		1.320	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KING,ARTHUR
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: GIOIA

DATE/TIME OF TEST: 02/12/2022 04:24:29

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.023	04:19
Breath Volume		1.230	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.025	04:23
Breath Volume		1.699	04:23
Air Blank		0.000	04:23
Control Test		0.082	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,JORDAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: VELASQUEZVASQUEZ

DATE/TIME OF TEST: 02/12/2022 05:59:55

Diagnostics Check		OK	05:48
Air Blank		0.000	05:49
Control Test		0.082	05:49
Air Blank		0.000	05:49
Subject Sample	# 1	0.252	05:52
Breath Volume		2.886	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.232	05:58
Breath Volume		2.019	05:58
Air Blank		0.000	05:58
Control Test		0.079	05:59
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 02/12/2022 19:29:16

Diagnostics Check		OK	19:21
Air Blank		0.000	19:22
Control Test		0.081	19:22
Air Blank		0.000	19:23
Subject Sample	# 1	0.281	19:24
Breath Volume		1.222	19:24
Air Blank		0.000	19:24
Air Blank		0.000	19:26
Subject Sample	# 2	0.290	19:27
Breath Volume		1.652	19:27
Air Blank		0.000	19:28
Control Test		0.081	19:28
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUHART,TOMMY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: NEGRON RODRIGUEZ

DATE/TIME OF TEST: 02/13/2022 00:27:45

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.047	00:23
Breath Volume		1.472	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.046	00:26
Breath Volume		1.453	00:26
Air Blank		0.000	00:26
Control Test		0.082	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOUGHERTY,JOSEPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 02/13/2022 02:29:40

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.148	02:25
Breath Volume		1.636	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.155	02:28
Breath Volume		1.886	02:28
Air Blank		0.000	02:28
Control Test		0.081	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATSON,ISAIAH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: ABNEY

DATE/TIME OF TEST: 02/13/2022 10:34:05

Diagnostics Check		OK	10:27
Air Blank		0.000	10:28
Control Test		0.081	10:28
Air Blank		0.000	10:29
Subject Sample	# 1	0.000	10:29
Breath Volume		2.414	10:29
Air Blank		0.000	10:30
Air Blank		0.000	10:32
Subject Sample	# 2	0.000	10:32
Breath Volume		2.734	10:32
Air Blank		0.000	10:33
Control Test		0.082	10:33
Air Blank		0.000	10:34
Diagnostics Check		OK	10:34

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZ,RICHARD
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: DOMINGUEZ DOMINGUEZ

DATE/TIME OF TEST: 02/13/2022 20:33:54

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.082	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.160	20:29
Breath Volume		1.664	20:29
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.153	20:32
Breath Volume		1.816	20:32
Air Blank		0.000	20:32
Control Test		0.081	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DEVOSS,EZEKIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/14/2022 01:39:55

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.082	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.226	01:35
Breath Volume		1.375	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.223	01:38
Breath Volume		1.531	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIEBKNECHT,KEVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MCCLAIN

DATE/TIME OF TEST: 02/14/2022 02:43:52

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.082	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.081	02:39
Breath Volume		1.351	02:39
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.081	02:42
Breath Volume		1.460	02:42
Air Blank		0.000	02:42
Control Test		0.081	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: ABRAMOWITZ

DATE/TIME OF TEST: 02/15/2022 00:26:13

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.082	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.098	00:21
Breath Volume		1.179	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.095	00:24
Breath Volume		1.378	00:24
Air Blank		0.000	00:25
Control Test		0.081	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 02/19/2022 18:19:33

Diagnostics Check		OK	18:13
Air Blank		0.000	18:14
Control Test		0.082	18:14
Air Blank		0.000	18:14
Subject Sample	# 1	0.175	18:15
Breath Volume		2.394	18:15
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.173	18:18
Breath Volume		1.851	18:18
Air Blank		0.000	18:18
Control Test		0.082	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOSA,EMMANUEL
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: RIGGS

DATE/TIME OF TEST: 02/19/2022 23:43:25

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.082	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:36	No Sample Provided
Breath Volume		0.000	23:36	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 2		23:42	No Sample Provided
Breath Volume		0.000	23:42	
Air Blank		0.000	23:42	
Control Test		0.082	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BRAUD

DATE/TIME OF TEST: 02/20/2022 02:01:06

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:49	
Control Test		0.082	01:50	
Air Blank		0.000	01:50	
Subject Sample	# 1	0.149	01:53	Volume Not Met
Breath Volume		0.898	01:53	
Air Blank		0.000	01:54	
Air Blank		0.000	01:56	
Subject Sample	# 2	0.165	01:56	
Breath Volume		2.023	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 3	0.179	01:59	
Breath Volume		1.378	01:59	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHOEN,ELLIS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: FRECONNA

DATE/TIME OF TEST: 02/20/2022 03:24:59

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.082	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.124	03:14	Volume Not Met
Breath Volume		0.843	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.133	03:18	
Breath Volume		1.406	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.132	03:23	
Breath Volume		1.242	03:23	
Air Blank		0.000	03:24	
Control Test		0.081	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIBEIRO,NICOLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 02/20/2022 05:15:35

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.081	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.141	05:11
Breath Volume		2.730	05:11
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.148	05:14
Breath Volume		3.035	05:14
Air Blank		0.000	05:14
Control Test		0.081	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EARLEY,VERNON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: RIVERA COLON

DATE/TIME OF TEST: 02/20/2022 20:36:04

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.082	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.178	20:31
Breath Volume		1.691	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.176	20:34
Breath Volume		1.855	20:34
Air Blank		0.000	20:35
Control Test		0.082	20:35
Air Blank		0.000	20:35
Diagnostics Check		OK	20:36

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 02/21/2022 00:03:26

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.082	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.204	23:58
Breath Volume		1.628	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.207	00:01
Breath Volume		1.191	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHAOUNI,KALEY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 02/21/2022 01:10:35

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.082	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.188	01:06
Breath Volume		1.800	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.188	01:09
Breath Volume		1.820	01:09
Air Blank		0.000	01:09
Control Test		0.081	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZSUERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:08

SUBJECT'S LAST NAME: GAYADEEN

DATE/TIME OF TEST: 02/21/2022 10:46:49

Diagnostics Check		OK	10:34	
Air Blank		0.000	10:34	
Control Test		0.082	10:35	
Air Blank		0.000	10:35	
Subject Sample	# 1	0.201	10:38	Volume Not Met
Breath Volume		0.222	10:38	
Air Blank		0.000	10:39	
Air Blank		0.000	10:41	
Subject Sample	# 2	0.206	10:41	
Breath Volume		1.386	10:41	
Air Blank		0.000	10:42	
Air Blank		0.000	10:44	
Subject Sample	# 3	0.197	10:45	
Breath Volume		1.414	10:45	
Air Blank		0.000	10:45	
Control Test		0.081	10:46	
Air Blank		0.000	10:46	
Diagnostics Check		OK	10:46	

Data Download Date/Time: 03/10/2022 12:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATSON,ISAIAH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LEMUS

DATE/TIME OF TEST: 02/21/2022 23:36:59

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.082	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.173	23:32
Breath Volume		1.367	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.160	23:35
Breath Volume		1.351	23:35
Air Blank		0.000	23:36
Control Test		0.082	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CARBALLO

DATE/TIME OF TEST: 02/22/2022 02:35:43

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.082	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.088	02:31
Breath Volume		1.398	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.084	02:34
Breath Volume		1.117	02:34
Air Blank		0.000	02:34
Control Test		0.081	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 02/24/2022 18:35:56

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.082	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.000	18:31
Breath Volume		3.968	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.000	18:34
Breath Volume		3.816	18:34
Air Blank		0.000	18:35
Control Test		0.081	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MARTINEZ TORO

DATE/TIME OF TEST: 02/25/2022 00:54:41

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.081	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.116	00:50
Breath Volume		1.253	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.118	00:53
Breath Volume		2.343	00:53
Air Blank		0.000	00:53
Control Test		0.081	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 02/26/2022 01:42:42

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:40	
Control Test		0.082	01:40	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Control Test		0.082	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUFFUS,DAVID
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/26/2022 21:29:56

Diagnostics Check		OK	21:24
Air Blank		0.000	21:24
Control Test		0.082	21:24
Air Blank		0.000	21:25
Subject Sample	# 1	0.210	21:25
Breath Volume		1.144	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.210	21:28
Breath Volume		1.238	21:28
Air Blank		0.000	21:29
Control Test		0.082	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIEL,CONNELL
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: LORENZANA YUMAN

DATE/TIME OF TEST: 02/27/2022 00:06:10

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.082	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.093	00:01
Breath Volume		2.066	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.085	00:04
Breath Volume		1.199	00:04
Air Blank		0.000	00:05
Control Test		0.082	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BORDEAUX,LAUREN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 02/27/2022 01:05:32

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.082	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.262	01:01
Breath Volume		1.730	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.246	01:03
Breath Volume		2.449	01:03
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUHART,TOMMY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: OSBORN

DATE/TIME OF TEST: 02/27/2022 01:40:56

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.081	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.000	01:36
Breath Volume		2.003	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.000	01:39
Breath Volume		3.554	01:39
Air Blank		0.000	01:40
Control Test		0.082	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SOTO,JONATHAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: GOMES PEREZ

DATE/TIME OF TEST: 02/27/2022 20:53:04

Diagnostics Check		OK	20:46
Air Blank		0.000	20:46
Control Test		0.083	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.250	20:48
Breath Volume		1.296	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:50
Subject Sample	# 2	0.248	20:51
Breath Volume		1.128	20:51
Air Blank		0.000	20:52
Control Test		0.081	20:52
Air Blank		0.000	20:52
Diagnostics Check		OK	20:53

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: RIDDLE

DATE/TIME OF TEST: 02/01/2022 17:17:02

Diagnostics Check		OK	17:05	
Air Blank		0.000	17:05	
Control Test		0.082	17:05	
Air Blank		0.000	17:06	
Subject Sample	# 1	0.056	17:09	Volume Not Met
Breath Volume		0.757	17:09	
Air Blank		0.000	17:10	
Air Blank		0.000	17:11	
Subject Sample	# 2	0.061	17:12	
Breath Volume		1.437	17:12	
Air Blank		0.000	17:13	
Air Blank		0.000	17:14	
Subject Sample	# 3	0.062	17:15	
Breath Volume		1.664	17:15	
Air Blank		0.000	17:16	
Control Test		0.077	17:16	RFI Detect
Air Blank		0.000	17:17	

Data Download Date/Time: 02/15/2022 21:14

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ARTHURS

DATE/TIME OF TEST: 02/05/2022 20:56:14

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.078	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.204	20:51
Breath Volume		2.398	20:51
Air Blank		0.000	20:52
Air Blank		0.000	20:53
Subject Sample	# 2	0.200	20:54
Breath Volume		1.710	20:54
Air Blank		0.000	20:55
Control Test		0.078	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 02/15/2022 21:15

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: PEOPLES

DATE/TIME OF TEST: 02/09/2022 19:43:07

Diagnostics Check		OK	19:36
Air Blank		0.000	19:37
Control Test		0.079	19:37
Air Blank		0.000	19:38
Subject Sample	# 1	0.155	19:38
Breath Volume		2.074	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.150	19:41
Breath Volume		1.171	19:41
Air Blank		0.000	19:42
Control Test		0.078	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 02/15/2022 21:15

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,JOSE,G
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 02/20/2022 01:04:31

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.254	00:59
Breath Volume		1.742	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.253	01:02
Breath Volume		2.949	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 03/06/2022 11:05

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,JOHN,D
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WILKES,JOHN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/15/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/28/2022 21:16:46

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.079	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.345	21:12
Breath Volume		2.640	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.336	21:15
Breath Volume		2.734	21:15
Air Blank		0.000	21:15
Control Test		0.079	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 03/06/2022 11:05

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/11/2022 04:48:02

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.080	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:40	No Sample Provided
Breath Volume		0.000	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2	0.161	04:46	Volume Not Met
Breath Volume		1.203	04:46	
Air Blank		0.000	04:47	
Control Test		0.079	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 02/22/2022 13:16

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: HARTSELLE,A
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 01/29/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: MANZO

DATE/TIME OF TEST: 02/04/2022 21:20:25

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.082	21:12	
Air Blank		0.000	21:12	
Subject Sample	# 1	0.237	21:14	
Breath Volume		1.601	21:14	
Air Blank		0.000	21:15	
Air Blank		0.000	21:16	
Subject Sample	# 2	0.235	21:19	
Breath Volume		1.480	21:19	
Air Blank			21:19	Purge Fail
Air Blank		0.000	21:20	

Data Download Date/Time: 02/23/2022 16:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PASZTOR,MICHAEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/05/2022 01:49:09

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.122	01:44
Breath Volume		3.214	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.123	01:47
Breath Volume		2.859	01:47
Air Blank		0.000	01:48
Control Test		0.079	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 03/16/2022 13:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,MEGAN,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PARMENTER,JACOB
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: GOMEZGOMEZ

DATE/TIME OF TEST: 02/07/2022 03:34:56

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.078	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.226	03:28
Breath Volume		1.210	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.217	03:33
Breath Volume		1.765	03:33
Air Blank		0.000	03:33
Control Test		0.076	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 03/16/2022 13:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,MEGAN,M
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PASZTOR,MICHAEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/07/2022 03:47:06

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.137	03:42
Breath Volume		4.156	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.136	03:45
Breath Volume		4.050	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 03/16/2022 13:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GONZALEZ GALVEZ

DATE/TIME OF TEST: 02/11/2022 23:56:30

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.118	23:52
Breath Volume		2.105	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.120	23:54
Breath Volume		3.164	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 03/16/2022 13:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND,WILLIAM,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BELTRAN,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LUIS-SANCHEZ

DATE/TIME OF TEST: 02/12/2022 03:07:41

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.217	03:03
Breath Volume		1.746	03:03
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.198	03:06
Breath Volume		3.078	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND,WILLIAM,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: KIRTLEY,C
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: MORALESGOMEZ

DATE/TIME OF TEST: 02/12/2022 18:19:26

Diagnostics Check		OK	18:12
Air Blank		0.000	18:13
Control Test		0.079	18:13
Air Blank		0.000	18:13
Subject Sample	# 1	0.275	18:14
Breath Volume		1.550	18:14
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.262	18:17
Breath Volume		1.574	18:17
Air Blank		0.000	18:18
Control Test		0.079	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND,WILLIAM,R
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TARAZONA,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: SALESPEREZ

DATE/TIME OF TEST: 02/12/2022 19:49:00

Diagnostics Check		OK	19:42
Air Blank		0.000	19:42
Control Test		0.078	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.241	19:43
Breath Volume		1.957	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.240	19:47
Breath Volume		1.792	19:47
Air Blank		0.000	19:48
Control Test		0.079	19:48
Air Blank		0.000	19:48
Diagnostics Check		OK	19:49

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 02/18/2022 03:15:48

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.078	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.169	03:10
Breath Volume		1.664	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.159	03:14
Breath Volume		1.121	03:14
Air Blank		0.000	03:14
Control Test		0.079	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: PASZTOR,MICHAEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/19/2022 22:09:29

Diagnostics Check		OK	22:01
Air Blank		0.000	22:02
Control Test		0.079	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.156	22:04
Breath Volume		3.410	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.160	22:07
Breath Volume		2.949	22:07
Air Blank		0.000	22:08
Control Test		0.078	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: MENDOZA,URIEL
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SANCHEZ LUIS

DATE/TIME OF TEST: 02/25/2022 21:41:44

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.079	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.197	21:37
Breath Volume		1.210	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.200	21:40
Breath Volume		1.945	21:40
Air Blank		0.000	21:40
Control Test		0.079	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: BELTRAN,S
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: DIEGO JOSE

DATE/TIME OF TEST: 02/25/2022 22:38:49

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.079	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.252	22:33
Breath Volume		1.882	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.261	22:37
Breath Volume		1.957	22:37
Air Blank		0.000	22:37
Control Test		0.079	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,ANDREW,B
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: JACOBS,RICHARD
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: AMORESGONZALEZ

DATE/TIME OF TEST: 02/28/2022 22:43:41

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.079	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.120	22:39
Breath Volume		2.050	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.121	22:42
Breath Volume		2.296	22:42
Air Blank		0.000	22:42
Control Test		0.077	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: STRAND

DATE/TIME OF TEST: 02/03/2022 00:52:47

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.081	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1	0.164	00:42	
Breath Volume		1.296	00:42	
Air Blank		0.000	00:43	
Air Blank		0.000	00:44	
Subject Sample	# 2	0.134	00:47	Volume Not Met
Breath Volume		0.660	00:47	
Air Blank		0.000	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 3	0.171	00:51	
Breath Volume		2.277	00:51	
Air Blank		0.000	00:51	
Control Test		0.077	00:52	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 02/28/2022 15:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADET,SHARLINE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARIAS,LUIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 02/05/2022 23:23:45

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.082	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.080	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 02/28/2022 15:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/06/2022 06:41:18

Diagnostics Check		OK	06:38	
Air Blank		0.000	06:38	
Control Test		0.081	06:39	
Air Blank		0.000	06:39	
Subject Sample	# 1		06:39	Subject Test Refused
Breath Volume		0.000	06:39	
Air Blank		0.000	06:40	
Control Test		0.080	06:40	
Air Blank		0.000	06:41	
Diagnostics Check		OK	06:41	

Data Download Date/Time: 02/28/2022 16:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: 27822,MAGDIEL,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 02/07/2022 10:28:16

Diagnostics Check		OK	10:25	
Air Blank		0.000	10:25	
Control Test		0.080	10:26	
Air Blank		0.000	10:26	
Subject Sample	# 1		10:26	Subject Test Refused
Breath Volume		0.000	10:26	
Air Blank		0.000	10:27	
Control Test		0.080	10:27	
Air Blank		0.000	10:28	
Diagnostics Check		OK	10:28	

Data Download Date/Time: 02/28/2022 16:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,ELOUSE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: JOSEPH,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: POMAR

DATE/TIME OF TEST: 02/08/2022 07:52:49

Diagnostics Check		OK	07:47	
Air Blank		0.000	07:48	
Control Test		0.081	07:48	
Air Blank		0.000	07:49	
Subject Sample	# 1		07:51	Subject Test Refused
Breath Volume		0.000	07:51	
Air Blank		0.000	07:51	
Control Test		0.081	07:52	
Air Blank		0.000	07:52	
Diagnostics Check		OK	07:52	

Data Download Date/Time: 02/28/2022 16:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEJERA,MARTHA,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: TEJERA,MARTHA
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DOMINIQUE

DATE/TIME OF TEST: 02/10/2022 23:04:03

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.082	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.178	22:59
Breath Volume		2.273	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:01
Subject Sample	# 2	0.168	23:02
Breath Volume		2.847	23:02
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:04

Data Download Date/Time: 02/28/2022 16:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,ALEXIA,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,LUIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: FALLON

DATE/TIME OF TEST: 02/11/2022 18:59:26

Diagnostics Check		OK	18:56	
Air Blank		0.000	18:57	
Control Test		0.081	18:57	
Air Blank		0.000	18:57	
Subject Sample	# 1		18:58	Subject Test Refused
Breath Volume		0.000	18:58	
Air Blank		0.000	18:58	
Control Test		0.081	18:58	
Air Blank		0.000	18:59	
Diagnostics Check		OK	18:59	

Data Download Date/Time: 02/28/2022 16:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARRERO PORTAL

DATE/TIME OF TEST: 02/12/2022 04:28:28

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.181	04:22
Breath Volume		1.132	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.183	04:26
Breath Volume		1.160	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 02/28/2022 16:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHESTER,CHARLES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: DELANCY

DATE/TIME OF TEST: 02/17/2022 21:35:50

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:33	
Control Test		0.081	21:33	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:34	Subject Test Refused
Breath Volume		0.000	21:34	
Air Blank		0.000	21:34	
Control Test		0.080	21:35	
Air Blank		0.000	21:35	
Diagnostics Check		OK	21:35	

Data Download Date/Time: 02/28/2022 16:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JORGE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: KLURMAN,JEREMY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/21/2022 00:06:26

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.082	00:04	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:05	RFI Detect
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.081	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 02/28/2022 17:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JORGE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: KLURMAN,JEREMY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/21/2022 00:21:07

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.081	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.166	00:16
Breath Volume		2.906	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.178	00:19
Breath Volume		3.457	00:19
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 02/28/2022 17:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PORRAS

DATE/TIME OF TEST: 02/23/2022 04:37:03

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.127	04:32
Breath Volume		1.835	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.127	04:35
Breath Volume		1.871	04:35
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:37

Data Download Date/Time: 02/28/2022 17:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JOSE,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VASQUEZ,JASON
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 07:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: CONCEPCION TAVARES

DATE/TIME OF TEST: 02/26/2022 08:19:04

Diagnostics Check		OK	08:11
Air Blank		0.000	08:11
Control Test		0.082	08:12
Air Blank		0.000	08:12
Subject Sample	# 1	0.221	08:13
Breath Volume		1.777	08:13
Air Blank		0.000	08:14
Air Blank		0.000	08:16
Subject Sample	# 2	0.220	08:17
Breath Volume		1.828	08:17
Air Blank		0.000	08:18
Control Test		0.079	08:18
Air Blank		0.000	08:18
Diagnostics Check		OK	08:19

Data Download Date/Time: 02/28/2022 17:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,MANUEL,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RODRIGUEZ,JENNIFER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: BISTER

DATE/TIME OF TEST: 02/27/2022 07:37:54

Diagnostics Check		OK	07:31
Air Blank		0.000	07:31
Control Test		0.082	07:31
Air Blank		0.000	07:32
Subject Sample	# 1	0.233	07:33
Breath Volume		3.726	07:33
Air Blank		0.000	07:33
Air Blank		0.000	07:35
Subject Sample	# 2	0.226	07:36
Breath Volume		2.695	07:36
Air Blank		0.000	07:36
Control Test		0.080	07:37
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

Data Download Date/Time: 02/28/2022 17:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: OMERAGIC

DATE/TIME OF TEST: 02/05/2022 04:27:01

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:24	
Control Test		0.081	04:25	
Air Blank		0.000	04:25	
Subject Sample	# 1		04:25	Subject Test Refused
Breath Volume		0.000	04:25	
Air Blank		0.000	04:26	
Control Test		0.081	04:26	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: BARKLEY

DATE/TIME OF TEST: 02/08/2022 21:54:10

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.081	21:48
Air Blank		0.000	21:49
Subject Sample	# 1	0.152	21:49
Breath Volume		1.589	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.146	21:52
Breath Volume		1.593	21:52
Air Blank		0.000	21:53
Control Test		0.078	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: MEANEY

DATE/TIME OF TEST: 02/09/2022 18:43:56

Diagnostics Check		OK	18:38
Air Blank		0.000	18:38
Control Test		0.080	18:38
Air Blank		0.000	18:39
Subject Sample	# 1	0.000	18:39
Breath Volume		2.316	18:39
Air Blank		0.000	18:40
Air Blank		0.000	18:42
Subject Sample	# 2	0.000	18:42
Breath Volume		3.457	18:42
Air Blank		0.000	18:43
Control Test		0.079	18:43
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: LINSON

DATE/TIME OF TEST: 02/09/2022 23:42:49

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.081	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.000	23:38
Breath Volume		2.570	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.000	23:41
Breath Volume		2.406	23:41
Air Blank		0.000	23:41
Control Test		0.080	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUSCIO,JESSE,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TRUSCIO,JESSE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SARNOWSKI

DATE/TIME OF TEST: 02/10/2022 01:11:43

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.109	01:07
Breath Volume		1.773	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.112	01:10
Breath Volume		1.601	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRILL,JOSHUA,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BICKHART,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080A1
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 02/11/2022 21:39:05

Diagnostics Check		OK	21:31
Air Blank		0.000	21:32
Control Test		0.080	21:32
Air Blank		0.000	21:33
Subject Sample	# 1	0.185	21:34
Breath Volume		2.359	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.177	21:37
Breath Volume		1.781	21:37
Air Blank		0.000	21:38
Control Test		0.079	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: TOWNSEND,T
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1060911
DRY GAS EXPIRY DATE: 12/11/2021
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MAUGER

DATE/TIME OF TEST: 02/19/2022 00:16:35

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.321	00:11
Breath Volume		3.488	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.335	00:15
Breath Volume		2.402	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02021080
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WILLIFORD

DATE/TIME OF TEST: 02/23/2022 01:27:48

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.056	01:27	Control Outside Tolerance
Air Blank	0.000	01:27	

Data Download Date/Time: 03/16/2022 21:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.