Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: LORENZO,FRANCISCO
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TIELES RODRIGUEZ

DATE/TIME OF TEST: 03/02/2022 03:07:48

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1	0.151	02:58	Volume Not Met
Breath Volume		0.199	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 2	0.182	03:02	
Breath Volume		1.175	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 3	0.186	03:06	
Breath Volume		1.945	03:06	
Air Blank		0.000	03:06	
Control Test		0.078	03:07	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 03/15/2022 11:33

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVARRIA,DEGLYS,E
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: RUIZ,TOMMY
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: CAOBI

DATE/TIME OF TEST: 03/06/2022 00:54:48

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:52	
Control Test		0.081	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:53	
Control Test		0.081	00:54	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 03/15/2022 11:33

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: LORENZO,FRANCISCO
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 03/01/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DOLORIT

DATE/TIME OF TEST: 03/10/2022 01:46:58

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:44	
Control Test		0.080	01:44	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:46	
Control Test		0.080	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 03/15/2022 11:33

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: LORENZO,FRANCISO
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 03/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ALOMA LOPEZ

DATE/TIME OF TEST: 03/19/2022 03:22:34

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.147	03:17
Breath Volume		1.449	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.139	03:20
Breath Volume		1.734	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 04/20/2022 23:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: HIALEAH GARDENS PD
ARRESTING OFFICER: REYES,ISRAEL
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 03/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LOPEZ ALVAREZ

DATE/TIME OF TEST: 03/20/2022 02:07:31

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.196	02:02
Breath Volume		1.890	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.192	02:05
Breath Volume		1.789	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 04/20/2022 23:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: GUERRA,YOEL
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 03/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 03/20/2022 21:54:10

Diagnostics Check		OK	21:47
Air Blank		0.000	21:47
Control Test		0.081	21:48
Air Blank		0.000	21:48
Subject Sample	# 1	0.214	21:49
Breath Volume		1.382	21:49
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.215	21:52
Breath Volume		1.351	21:52
Air Blank		0.000	21:53
Control Test		0.079	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 04/20/2022 23:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: JAMES,FRANCES
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: MESA

DATE/TIME OF TEST: 03/02/2022 23:36:22

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:33	
Control Test		0.081	23:33	
Air Blank		0.000	23:34	
Subject Sample	# 1		23:34	Subject Test Refused
Breath Volume		0.000	23:34	
Air Blank		0.000	23:35	
Control Test		0.081	23:35	
Air Blank		0.000	23:36	
Diagnostics Check		OK	23:36	

Data Download Date/Time: 03/04/2022 07:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANAS,JUAN,C
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DIEUDONNE,K
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 13:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 03/09/2022 14:12:06

Diagnostics Check		OK	14:09	
Air Blank		0.000	14:09	
Control Test		0.080	14:09	
Air Blank		0.000	14:10	
Subject Sample	# 1		14:10	Subject Test Refused
Breath Volume		0.000	14:10	
Air Blank		0.000	14:11	
Control Test		0.080	14:11	
Air Blank		0.000	14:12	
Diagnostics Check		OK	14:12	

Data Download Date/Time: 04/06/2022 07:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL JR,DERRICK,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: O'NEAL JR,D
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: BUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: STEVEN

DATE/TIME OF TEST: 03/12/2022 16:47:17

Diagnostics Check		OK	16:44	
Air Blank		0.000	16:44	
Control Test		0.081	16:44	
Air Blank		0.000	16:45	
Subject Sample	# 1		16:45	Subject Test Refused
Breath Volume		0.441	16:45	
Air Blank		0.000	16:46	
Control Test		0.081	16:46	
Air Blank		0.000	16:47	
Diagnostics Check		OK	16:47	

Data Download Date/Time: 04/06/2022 07:14

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: EDWARD PEREZ

DATE/TIME OF TEST: 03/16/2022 01:46:07

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.080	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.080	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BLOCKER

DATE/TIME OF TEST: 03/18/2022 03:26:08

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.184	03:20
Breath Volume		1.296	03:20
Air Blank		0.000	03:22
Air Blank		0.000	03:23
Subject Sample	# 2	0.183	03:24
Breath Volume		1.199	03:24
Air Blank		0.000	03:25
Control Test		0.075	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIS,SYLVIA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HOLLIS,SYLVIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: AMAYA ESPINOZA

DATE/TIME OF TEST: 03/19/2022 06:16:53

Diagnostics Check		OK	06:05	
Air Blank		0.000	06:05	
Control Test		0.080	06:06	
Air Blank		0.000	06:06	
Subject Sample	# 1		06:09	No Sample Provided
Breath Volume		0.000	06:09	
Air Blank		0.000	06:10	
Air Blank		0.000	06:12	
Subject Sample	# 2		06:15	No Sample Provided
Breath Volume		0.000	06:15	
Air Blank		0.000	06:15	
Control Test		0.081	06:16	
Air Blank		0.000	06:16	
Diagnostics Check		OK	06:16	

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOURIDO,MODESTO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOURIDO,MODESTO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:53

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/21/2022 13:28:15

Diagnostics Check		OK	13:19
Air Blank		0.000	13:19
Control Test		0.081	13:20
Air Blank		0.000	13:20
Subject Sample	# 1	0.207	13:22
Breath Volume		1.363	13:22
Air Blank		0.000	13:23
Air Blank		0.000	13:25
Subject Sample	# 2	0.212	13:26
Breath Volume		2.437	13:26
Air Blank		0.000	13:27
Control Test		0.076	13:27
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINAA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 03/22/2022 07:56:51

Diagnostics Check		OK	07:51	
Air Blank		0.000	07:52	
Control Test		0.081	07:52	
Air Blank		0.000	07:53	
Subject Sample	# 1	0.070	07:54	
Breath Volume		3.441	07:54	
Air Blank		0.000	07:55	
Air Blank			07:56	Improper Sample
Air Blank			07:56	Improper Sample
Air Blank			07:56	Improper Sample

Data Download Date/Time: 04/06/2022 07:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 03/22/2022 08:01:34

Diagnostics Check	OK	08:00	
Air Blank		08:01	Ambient Fail
Air Blank	0.000	08:01	

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 03/22/2022 08:09:30

Diagnostics Check		OK	08:03
Air Blank		0.000	08:03
Control Test		0.079	08:04
Air Blank		0.000	08:04
Subject Sample	# 1	0.069	08:05
Breath Volume		2.308	08:05
Air Blank		0.000	08:05
Air Blank		0.000	08:07
Subject Sample	# 2	0.065	08:07
Breath Volume		3.125	08:07
Air Blank		0.000	08:08
Control Test		0.079	08:08
Air Blank		0.000	08:09
Diagnostics Check		OK	08:09

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MARTINEZ,A
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: COLMENARES URBAEZ

DATE/TIME OF TEST: 03/26/2022 05:08:27

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.214	05:03
Breath Volume		1.347	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.203	05:06
Breath Volume		1.589	05:06
Air Blank		0.000	05:07
Control Test		0.079	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BARRIGA

DATE/TIME OF TEST: 03/28/2022 23:06:10

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:03	
Control Test		0.080	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:04	Subject Test Refused
Breath Volume		0.000	23:04	
Air Blank		0.000	23:05	
Control Test		0.081	23:05	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05620080A1
DRY GAS EXPIRY DATE: 04/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: CANONICO PAEZ

DATE/TIME OF TEST: 03/29/2022 03:35:38

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.081	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.080	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 04/06/2022 07:30

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: LAFASO

DATE/TIME OF TEST: 03/01/2022 22:08:29

Diagnostics Check		OK	22:02
Air Blank		0.000	22:02
Control Test		0.078	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.150	22:03
Breath Volume		3.519	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.155	22:06
Breath Volume		3.457	22:06
Air Blank		0.000	22:07
Control Test		0.078	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: NIX

DATE/TIME OF TEST: 03/02/2022 03:13:14

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.129	03:08
Breath Volume		2.828	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.110	03:11
Breath Volume		1.589	03:11
Air Blank		0.000	03:12
Control Test		0.077	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: HITZ

DATE/TIME OF TEST: 03/03/2022 00:20:28

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.079	00:18	
Air Blank		0.000	00:18	
Subject Sample	# 1		00:19	Subject Test Refused
Breath Volume		0.000	00:19	
Air Blank		0.000	00:19	
Control Test		0.079	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SORRELLS,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SHERWOOD

DATE/TIME OF TEST: 03/03/2022 03:11:46

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.079	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITMAR,V
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: MCVAY

DATE/TIME OF TEST: 03/03/2022 04:55:02

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.079	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.181	04:49
Breath Volume		1.851	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:51
Subject Sample	# 2	0.167	04:53
Breath Volume		1.480	04:53
Air Blank		0.000	04:54
Control Test		0.078	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:55

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVERETT,I,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CURRY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/03/2022 07:57:01

Diagnostics Check		OK	07:50
Air Blank		0.000	07:50
Control Test		0.079	07:51
Air Blank		0.000	07:51
Subject Sample	# 1	0.151	07:52
Breath Volume		1.656	07:52
Air Blank		0.000	07:53
Air Blank		0.000	07:55
Subject Sample	# 2	0.154	07:55
Breath Volume		2.140	07:55
Air Blank		0.000	07:56
Control Test		0.078	07:56
Air Blank		0.000	07:56
Diagnostics Check		OK	07:57

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 03/04/2022 03:20:42

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.079	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:19	Subject Test Refused
Breath Volume		0.000	03:19	
Air Blank		0.000	03:19	
Control Test		0.078	03:20	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: O'CONNELL,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/04/2022 21:00:31

Diagnostics Check		OK	20:53
Air Blank		0.000	20:54
Control Test		0.077	20:54
Air Blank		0.000	20:55
Subject Sample	# 1	0.308	20:56
Breath Volume		2.867	20:56
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.291	20:59
Breath Volume		2.457	20:59
Air Blank		0.000	20:59
Control Test		0.077	20:59
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEURER,D,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: GARTMAN

DATE/TIME OF TEST: 03/05/2022 04:58:42

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.078	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.173	04:54
Breath Volume		3.171	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.175	04:57
Breath Volume		2.703	04:57
Air Blank		0.000	04:57
Control Test		0.078	04:58
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ARZUAGA DIEPPA

DATE/TIME OF TEST: 03/07/2022 23:53:41

Diagnostics Check		OK	23:50	
Air Blank		0.000	23:51	
Control Test		0.077	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 1		23:52	Subject Test Refused
Breath Volume		0.000	23:52	
Air Blank		0.000	23:52	
Control Test		0.077	23:53	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/08/2022 03:15:22

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.076	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.097	03:10
Breath Volume		1.269	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.105	03:13
Breath Volume		1.863	03:13
Air Blank		0.000	03:14
Control Test		0.077	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: SACHEWICZ

DATE/TIME OF TEST: 03/08/2022 23:23:39

Diagnostics Check		OK	23:16
Air Blank		0.000	23:17
Control Test		0.077	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.210	23:18
Breath Volume		1.960	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.213	23:22
Breath Volume		2.679	23:22
Air Blank		0.000	23:22
Control Test		0.077	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 03/09/2022 02:24:59

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.078	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.138	02:20
Breath Volume		1.714	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.142	02:23
Breath Volume		1.339	02:23
Air Blank		0.000	02:24
Control Test		0.078	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/09/2022 20:45:41

Diagnostics Check		OK	20:38
Air Blank		0.000	20:38
Control Test		0.078	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.204	20:40
Breath Volume		2.785	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.194	20:44
Breath Volume		2.230	20:44
Air Blank		0.000	20:44
Control Test		0.077	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 03/10/2022 04:51

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GALLION

DATE/TIME OF TEST: 03/09/2022 23:02:06

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.077	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.040	22:57
Breath Volume		2.675	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.035	23:00
Breath Volume		1.195	23:00
Air Blank		0.000	23:01
Control Test		0.077	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 03/10/2022 04:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BREWIN

DATE/TIME OF TEST: 03/10/2022 03:15:43

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:13	
Control Test		0.078	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.078	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 03/10/2022 04:52

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BAEZ PACHECO

DATE/TIME OF TEST: 03/20/2022 02:37:35

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.184	02:32
Breath Volume		1.628	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.185	02:36
Breath Volume		1.589	02:36
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 03/23/2022 12:25

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: AMICA

DATE/TIME OF TEST: 03/21/2022 02:16:20

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.142	02:11
Breath Volume		1.632	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.137	02:14
Breath Volume		1.480	02:14
Air Blank		0.000	02:15
Control Test		0.078	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 03/23/2022 12:25

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: CASTILLA

DATE/TIME OF TEST: 03/21/2022 05:21:53

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:19	
Control Test		0.079	05:19	
Air Blank		0.000	05:20	
Subject Sample	# 1		05:20	Subject Test Refused
Breath Volume		0.000	05:20	
Air Blank		0.000	05:20	
Control Test		0.078	05:21	
Air Blank		0.000	05:21	
Diagnostics Check		OK	05:21	

Data Download Date/Time: 03/23/2022 12:25

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 03/24/2022 00:52:03

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:47	
Control Test		0.078	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Control Test		0.077	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 04/15/2022 16:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 03/24/2022 03:23:28

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.077	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.077	03:22	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 04/15/2022 16:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: VO,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 03/26/2022 02:16:42

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.077	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.237	02:12
Breath Volume		1.394	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.237	02:15
Breath Volume		1.433	02:15
Air Blank		0.000	02:15
Control Test		0.078	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 04/15/2022 16:59

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IZZI,NATASHA,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 03/26/2022 04:04:16

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.160	03:59
Breath Volume		1.199	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.164	04:02
Breath Volume		1.546	04:02
Air Blank		0.000	04:03
Control Test		0.077	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 04/15/2022 17:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SIMMONS,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: LEITNER

DATE/TIME OF TEST: 03/26/2022 07:00:51

Diagnostics Check		OK	06:52
Air Blank		0.000	06:52
Control Test		0.078	06:53
Air Blank		0.000	06:53
Subject Sample	# 1	0.199	06:55
Breath Volume		1.460	06:55
Air Blank		0.000	06:56
Air Blank		0.000	06:58
Subject Sample	# 2	0.198	06:59
Breath Volume		1.156	06:59
Air Blank		0.000	06:59
Control Test		0.078	07:00
Air Blank		0.000	07:00
Diagnostics Check		OK	07:00

Data Download Date/Time: 04/15/2022 17:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: TEMPLE

DATE/TIME OF TEST: 03/27/2022 20:01:28

Diagnostics Check		OK	19:54
Air Blank		0.000	19:55
Control Test		0.078	19:55
Air Blank		0.000	19:56
Subject Sample	# 1	0.148	19:56
Breath Volume		2.597	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.147	20:00
Breath Volume		2.277	20:00
Air Blank		0.000	20:00
Control Test		0.077	20:00
Air Blank		0.000	20:01
Diagnostics Check		OK	20:01

Data Download Date/Time: 04/15/2022 17:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: VAUGHNS

DATE/TIME OF TEST: 03/29/2022 00:24:43

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.078	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.078	00:24	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 04/15/2022 17:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DINKENS

DATE/TIME OF TEST: 03/31/2022 00:15:08

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.077	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.195	00:10
Breath Volume		2.703	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.202	00:13
Breath Volume		2.750	00:13
Air Blank		0.000	00:14
Control Test		0.077	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 04/15/2022 17:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: SYS CHECK
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/01/2022 10:14:26

Diagnostics Check	OK	10:12	
Air Blank	0.000	10:13	
Control Test	0.000	10:13	Control Outside Tolerance
Air Blank	0.000	10:14	

Data Download Date/Time: 03/10/2022 03:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,RICHARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/02/2022 13:44:43

Diagnostics Check		OK	13:38
Air Blank		0.000	13:39
Control Test		0.077	13:39
Air Blank		0.000	13:40
Subject Sample	# 1	0.000	13:40
Breath Volume		1.707	13:40
Air Blank		0.000	13:40
Air Blank		0.000	13:42
Subject Sample	# 2	0.000	13:43
Breath Volume		1.449	13:43
Air Blank		0.000	13:43
Control Test		0.078	13:44
Air Blank		0.000	13:44
Diagnostics Check		OK	13:44

Data Download Date/Time: 03/10/2022 03:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTHRAN,D,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FLORENCE,V
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: BLAZEJOWSKI

DATE/TIME OF TEST: 03/05/2022 08:02:35

Diagnostics Check		OK	07:58	
Air Blank		0.000	07:59	
Control Test		0.077	07:59	
Air Blank		0.000	08:00	
Subject Sample	# 1		08:01	Subject Test Refused
Breath Volume		0.000	08:01	
Air Blank		0.000	08:01	
Control Test		0.078	08:01	
Air Blank		0.000	08:02	
Diagnostics Check		OK	08:02	

Data Download Date/Time: 03/10/2022 03:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/09/2022 21:16:42

Diagnostics Check		OK	21:13	
Air Blank		0.000	21:14	
Control Test		0.078	21:14	
Air Blank		0.000	21:15	
Subject Sample	# 1		21:15	Subject Test Refused
Breath Volume		0.000	21:15	
Air Blank		0.000	21:15	
Control Test		0.078	21:16	
Air Blank		0.000	21:16	
Diagnostics Check		OK	21:16	

Data Download Date/Time: 03/10/2022 03:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LEFEVRE

DATE/TIME OF TEST: 03/10/2022 02:33:35

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.078	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.078	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 03/10/2022 03:43

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RALSTON

DATE/TIME OF TEST: 03/10/2022 04:52:00

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.077	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1		04:50	Subject Test Refused
Breath Volume		0.000	04:50	
Air Blank		0.000	04:51	
Control Test		0.077	04:51	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAMBERT,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: SATLER

DATE/TIME OF TEST: 03/10/2022 10:16:56

Diagnostics Check		OK	10:10
Air Blank		0.000	10:11
Control Test		0.077	10:11
Air Blank		0.000	10:12
Subject Sample	# 1	0.000	10:12
Breath Volume		4.183	10:12
Air Blank		0.000	10:13
Air Blank		0.000	10:15
Subject Sample	# 2	0.000	10:15
Breath Volume		4.375	10:15
Air Blank		0.000	10:16
Control Test		0.077	10:16
Air Blank		0.000	10:16
Diagnostics Check		OK	10:16

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 03/11/2022 00:26:02

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.077	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.122	00:20
Breath Volume		1.191	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.116	00:24
Breath Volume		1.765	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:26

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: BOSTICK

DATE/TIME OF TEST: 03/17/2022 23:54:36

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.077	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.000	23:48
Breath Volume		1.445	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.000	23:53
Breath Volume		1.199	23:53
Air Blank		0.000	23:53
Control Test		0.078	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 03/18/2022 02:10:53

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.078	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.046	02:06
Breath Volume		1.519	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.043	02:09
Breath Volume		2.015	02:09
Air Blank		0.000	02:09
Control Test		0.077	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: FOREMAN

DATE/TIME OF TEST: 03/19/2022 03:26:12

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.077	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.153	03:21
Breath Volume		1.171	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.144	03:24
Breath Volume		1.808	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HERNANDEZ VELAZUEZ

DATE/TIME OF TEST: 03/20/2022 02:03:00

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.078	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.132	01:58
Breath Volume		1.382	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.137	02:01
Breath Volume		2.058	02:01
Air Blank		0.000	02:02
Control Test		0.077	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:03

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MONDRAGON MENDOZA

DATE/TIME OF TEST: 03/21/2022 23:50:54

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.078	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.181	23:46
Breath Volume		1.566	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.176	23:49
Breath Volume		1.585	23:49
Air Blank		0.000	23:49
Control Test		0.077	23:50
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 03/22/2022 03:14:05

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.078	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	
Control Test		0.078	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 03/23/2022 12:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PETERSON,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/24/2022 01:10:29

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.078	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.077	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 04/16/2022 08:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURVIS,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: KIMBLE

DATE/TIME OF TEST: 03/25/2022 02:46:59

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.049	02:42
Breath Volume		3.386	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.048	02:45
Breath Volume		3.628	02:45
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 04/16/2022 08:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EAST,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 03/26/2022 18:35:53

Diagnostics Check		OK	18:28
Air Blank		0.000	18:29
Control Test		0.078	18:29
Air Blank		0.000	18:30
Subject Sample	# 1	0.169	18:30
Breath Volume		1.832	18:30
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.175	18:34
Breath Volume		2.812	18:34
Air Blank		0.000	18:34
Control Test		0.077	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 04/16/2022 08:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COREY,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SLONAKER

DATE/TIME OF TEST: 03/26/2022 23:12:01

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:09	
Control Test		0.078	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:11	
Control Test		0.078	23:11	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 04/16/2022 08:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LOPEZ OZUNA

DATE/TIME OF TEST: 03/27/2022 00:51:57

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.078	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.112	00:47
Breath Volume		3.472	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:49
Subject Sample	# 2	0.105	00:50
Breath Volume		2.437	00:50
Air Blank		0.000	00:51
Control Test		0.077	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 04/16/2022 08:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JIMENEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: KRISHNAKUMAR

DATE/TIME OF TEST: 03/27/2022 19:33:44

Diagnostics Check		OK	19:30	
Air Blank		0.000	19:30	
Control Test		0.077	19:31	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:32	Subject Test Refused
Breath Volume		0.000	19:32	
Air Blank		0.000	19:32	
Control Test		0.078	19:33	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 04/16/2022 08:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/27/2022 23:44:26

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.078	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.170	23:39
Breath Volume		1.871	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.166	23:42
Breath Volume		2.242	23:42
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 04/16/2022 08:48

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HERNANDEZ SANCHEZ

DATE/TIME OF TEST: 03/28/2022 00:57:30

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.078	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.123	00:52
Breath Volume		1.691	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.126	00:56
Breath Volume		1.808	00:56
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 04/16/2022 08:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IZZI,NATASHA,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: KLOEPPEL

DATE/TIME OF TEST: 03/28/2022 21:47:42

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:45	
Control Test		0.077	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:46	
Control Test		0.077	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 04/16/2022 08:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA-SANCHEZ,R,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WETZ

DATE/TIME OF TEST: 03/29/2022 02:07:54

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.078	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Control Test		0.077	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 04/16/2022 08:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IZZI,NATASHA,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08121080A1
DRY GAS EXPIRY DATE: 05/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 03/30/2022 05:00:28

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:58	
Control Test		0.078	04:58	
Air Blank		0.000	04:58	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	04:59	
Control Test		0.078	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 04/16/2022 08:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOELZER,K,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TOWNSEND,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SULFARO

DATE/TIME OF TEST: 03/18/2022 02:40:47

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.079	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.228	02:36
Breath Volume		2.105	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.241	02:39
Breath Volume		2.800	02:39
Air Blank		0.000	02:39
Control Test		0.078	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 03/29/2022 02:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNEAD,C,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PAETCSH,J
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: RANDOLPH

DATE/TIME OF TEST: 03/18/2022 21:26:26

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.078	21:18
Air Blank		0.000	21:19
Subject Sample	# 1	0.234	21:21
Breath Volume		1.781	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:23
Subject Sample	# 2	0.233	21:24
Breath Volume		2.617	21:24
Air Blank		0.000	21:25
Control Test		0.076	21:25
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 03/29/2022 02:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNEAD,C,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HOLMBERG,H
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/19/2022 05:32:20

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.000	05:27
Breath Volume		3.089	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.000	05:30
Breath Volume		2.367	05:30
Air Blank		0.000	05:31
Control Test		0.077	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 03/29/2022 02:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIPPINS,C,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,R
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SOUD

DATE/TIME OF TEST: 03/23/2022 02:01:28

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.080	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:00	
Control Test		0.078	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 03/29/2022 02:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HOLTON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 03/25/2022 02:10:20

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.080	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 03/29/2022 02:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: PARK

DATE/TIME OF TEST: 03/02/2022 03:12:41

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.183	03:07
Breath Volume		1.832	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.188	03:11
Breath Volume		1.753	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KING,PAUL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: AKRIDGE

DATE/TIME OF TEST: 03/05/2022 06:42:26

Diagnostics Check		OK	06:35
Air Blank		0.000	06:36
Control Test		0.078	06:36
Air Blank		0.000	06:37
Subject Sample	# 1	0.149	06:37
Breath Volume		1.898	06:37
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.147	06:40
Breath Volume		2.066	06:40
Air Blank		0.000	06:41
Control Test		0.078	06:41
Air Blank		0.000	06:42
Diagnostics Check		OK	06:42

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: LEWIS,L
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PHILIPS

DATE/TIME OF TEST: 03/05/2022 22:11:15

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.078	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.206	22:06
Breath Volume		2.988	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.192	22:09
Breath Volume		1.839	22:09
Air Blank		0.000	22:10
Control Test		0.077	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAEYAERT,JEFF,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: NAEYAERT,JEFF
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/06/2022 03:23:53

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.188	03:19
Breath Volume		1.929	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:21
Subject Sample	# 2	0.190	03:22
Breath Volume		1.765	03:22
Air Blank		0.000	03:22
Control Test		0.078	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DOUGHMAN

DATE/TIME OF TEST: 03/06/2022 03:58:37

Diagnostics Check	OK	03:57	
Air Blank		03:58	RFI Detect
Air Blank	0.000	03:58	

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DOUGHMAN

DATE/TIME OF TEST: 03/06/2022 04:01:52

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.077	04:00	
Air Blank	0.000	04:01	
		04:01	RFI Detect
Air Blank	0.000	04:01	

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DOUGHMAN

DATE/TIME OF TEST: 03/06/2022 04:04:02

Diagnostics Check	OK	04:03	
Air Blank		04:03	RFI Detect
Air Blank		04:04	RFI Detect

Data Download Date/Time: 04/28/2022 09:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DOUGHMAN

DATE/TIME OF TEST: 03/06/2022 04:08:21

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.078	04:06	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Control Test		0.078	04:07	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/06/2022 22:23:33

Diagnostics Check		OK	22:19	
Air Blank		0.000	22:20	
Control Test		0.078	22:20	
Air Blank		0.000	22:20	
Subject Sample	# 1	0.100	22:21	
Breath Volume		1.566	22:21	
Air Blank			22:22	RFI Detect
Air Blank		0.000	22:22	
Control Test		0.079	22:22	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,L,
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,G
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/06/2022 22:33:17

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.078	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.094	22:27
Breath Volume		1.164	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.091	22:31
Breath Volume		1.292	22:31
Air Blank		0.000	22:32
Control Test		0.077	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,DEPUTY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/12/2022 03:34:28

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.078	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.195	03:29
Breath Volume		1.484	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.191	03:32
Breath Volume		1.445	03:32
Air Blank		0.000	03:33
Control Test		0.077	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: CAROTTI,DEP
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 03/13/2022 21:33:58

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.231	21:29
Breath Volume		1.453	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:31
Subject Sample	# 2	0.245	21:32
Breath Volume		2.332	21:32
Air Blank		0.000	21:33
Control Test		0.078	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCHIBALD,BRIANNA,J
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ARCHIBALD,BRIANNA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 03/14/2022 11:42:10

Diagnostics Check		OK	11:35
Air Blank		0.000	11:36
Control Test		0.079	11:36
Air Blank		0.000	11:36
Subject Sample	# 1	0.146	11:37
Breath Volume		2.894	11:37
Air Blank		0.000	11:38
Air Blank		0.000	11:39
Subject Sample	# 2	0.134	11:40
Breath Volume		1.433	11:40
Air Blank		0.000	11:41
Control Test		0.077	11:41
Air Blank		0.000	11:42
Diagnostics Check		OK	11:42

Data Download Date/Time: 04/28/2022 09:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,BRANDON,A
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLEN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: CREEKMORE

DATE/TIME OF TEST: 03/23/2022 20:48:30

Diagnostics Check		OK	20:41
Air Blank		0.000	20:42
Control Test		0.079	20:42
Air Blank		0.000	20:43
Subject Sample	# 1	0.000	20:44
Breath Volume		3.937	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.000	20:47
Breath Volume		2.273	20:47
Air Blank		0.000	20:47
Control Test		0.079	20:47
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 04/28/2022 09:05

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,DEP
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MATCHETT

DATE/TIME OF TEST: 03/27/2022 03:44:00

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.079	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.183	03:39
Breath Volume		1.878	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.174	03:42
Breath Volume		2.152	03:42
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:44

Data Download Date/Time: 04/28/2022 09:05

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,BRANDON,A
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: SAPUNDJIEFF,THOMAS
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: DURRANCE

DATE/TIME OF TEST: 03/28/2022 23:03:55

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.078	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.238	22:57
Breath Volume		5.476	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.233	23:02
Breath Volume		1.933	23:02
Air Blank		0.000	23:02
Control Test		0.078	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 04/28/2022 09:05

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLUDSWORTH,SGT,J
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: CAROTTI,DEP
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/01/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/31/2022 00:36:08

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.078	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:32	No Sample Provided
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2		00:34	Subject Test Refused
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Control Test		0.078	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 04/28/2022 09:05

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 03/02/2022 21:31:14

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.232	21:26
Breath Volume		2.382	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.238	21:29
Breath Volume		2.417	21:29
Air Blank		0.000	21:30
Control Test		0.079	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 03/27/2022 10:27

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 03/01/2022 00:33:24

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:32	Subject Test Refused
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Control Test		0.080	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 03/25/2022 22:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WARD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/02/2022 01:25:01

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.081	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Control Test		0.081	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 03/25/2022 22:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: CRAFT

DATE/TIME OF TEST: 03/02/2022 19:30:29

Diagnostics Check		OK	19:23
Air Blank		0.000	19:24
Control Test		0.081	19:24
Air Blank		0.000	19:25
Subject Sample	# 1	0.000	19:25
Breath Volume		1.554	19:25
Air Blank		0.000	19:26
Air Blank		0.000	19:28
Subject Sample	# 2	0.000	19:29
Breath Volume		1.429	19:29
Air Blank		0.000	19:29
Control Test		0.081	19:29
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

Data Download Date/Time: 03/25/2022 22:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 03/04/2022 02:25:24

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.000	02:21
Breath Volume		1.648	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.000	02:24
Breath Volume		1.824	02:24
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 03/25/2022 22:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAND,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PETTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: FORREY

DATE/TIME OF TEST: 03/04/2022 16:42:58

Diagnostics Check		OK	16:39	
Air Blank		0.000	16:40	
Control Test		0.080	16:40	
Air Blank		0.000	16:41	
Subject Sample	# 1		16:41	Subject Test Refused
Breath Volume		0.000	16:41	
Air Blank		0.000	16:42	
Control Test		0.080	16:42	
Air Blank		0.000	16:42	
Diagnostics Check		OK	16:42	

Data Download Date/Time: 03/25/2022 22:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVIER,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GROENENBOOM,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 03/05/2022 19:30:34

Diagnostics Check		OK	19:27	
Air Blank		0.000	19:28	
Control Test		0.080	19:28	
Air Blank		0.000	19:29	
Subject Sample	# 1		19:29	Subject Test Refused
Breath Volume		0.000	19:29	
Air Blank		0.000	19:29	
Control Test		0.081	19:29	
Air Blank		0.000	19:30	
Diagnostics Check		OK	19:30	

Data Download Date/Time: 03/25/2022 22:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/06/2022 03:09:56

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.121	03:05
Breath Volume		2.285	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:07
Subject Sample	# 2	0.121	03:08
Breath Volume		2.046	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BROUSSARD

DATE/TIME OF TEST: 03/07/2022 07:17:34

Diagnostics Check		OK	07:10
Air Blank		0.000	07:10
Control Test		0.080	07:11
Air Blank		0.000	07:11
Subject Sample	# 1	0.010	07:12
Breath Volume		1.167	07:12
Air Blank		0.000	07:12
Air Blank		0.000	07:14
Subject Sample	# 2	0.009	07:16
Breath Volume		1.449	07:16
Air Blank		0.000	07:16
Control Test		0.080	07:16
Air Blank		0.000	07:17
Diagnostics Check		OK	07:17

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELCH,BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: STERCHI

DATE/TIME OF TEST: 03/09/2022 02:44:30

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.079	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.130	02:39
Breath Volume		2.449	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:41
Subject Sample	# 2	0.126	02:42
Breath Volume		2.535	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,COREY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CHESNESS

DATE/TIME OF TEST: 03/11/2022 02:51:57

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1	0.240	02:41	
Breath Volume		2.730	02:41	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 2	0.224	02:47	Volume Not Met
Breath Volume		0.664	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 3	0.236	02:50	
Breath Volume		2.429	02:50	
Air Blank		0.000	02:51	
Control Test		0.075	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PERRITT

DATE/TIME OF TEST: 03/11/2022 03:03:36

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.000	02:58
Breath Volume		3.476	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.000	03:02
Breath Volume		3.613	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 03/12/2022 04:19:05

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.000	04:14
Breath Volume		2.906	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.000	04:17
Breath Volume		2.890	04:17
Air Blank		0.000	04:18
Control Test		0.081	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:19

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/13/2022 00:17:37

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.235	00:13
Breath Volume		2.933	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.240	00:16
Breath Volume		2.984	00:16
Air Blank		0.000	00:16
Control Test		0.078	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOYE,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GERENCSER

DATE/TIME OF TEST: 03/13/2022 23:19:18

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.081	23:17	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.080	23:18	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: JONES,TYLER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 03/15/2022 01:44:33

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.154	01:38
Breath Volume		1.410	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.139	01:42
Breath Volume		2.750	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/15/2022 22:22:50

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.080	22:22	
Air Blank		22:22	RFI Detect
Air Blank	0.000	22:22	

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/15/2022 22:26:48

Diagnostics Check	OK	22:26	
Air Blank		22:26	Improper Sample
Air Blank		22:26	Improper Sample

Data Download Date/Time: 03/25/2022 22:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: C514439973350

DATE/TIME OF TEST: 03/15/2022 22:29:14

Diagnostics Check	OK	22:29	
Air Blank		22:29	Improper Sample
Air Blank		22:29	Improper Sample

Data Download Date/Time: 03/25/2022 22:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/15/2022 22:38:00

Diagnostics Check	OK	22:37	
Air Blank		22:37	Ambient Fail
Air Blank	0.000	22:38	

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KIM,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/15/2022 22:45:29

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.079	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.135	22:41
Breath Volume		2.242	22:41
Air Blank		0.000	22:41
Air Blank		0.000	22:43
Subject Sample	# 2	0.129	22:43
Breath Volume		1.656	22:43
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 03/25/2022 22:08

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: PIMBLE

DATE/TIME OF TEST: 03/16/2022 03:49:22

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:38	
Control Test		0.081	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:42	No Sample Provided
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Air Blank		0.000	03:44	
Subject Sample	# 2		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:48	
Control Test		0.080	03:48	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SIMMONS,JUSTN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: SHEALEY

DATE/TIME OF TEST: 03/16/2022 20:49:56

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.081	20:49	
Air Blank		20:49	RFI Detect
Air Blank	0.000	20:49	

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SIMMONS,JUSTIN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: SHEALEY

DATE/TIME OF TEST: 03/16/2022 21:14:41

Diagnostics Check		OK	21:09	
Air Blank		0.000	21:09	
Control Test		0.081	21:09	
Air Blank		0.000	21:10	
Subject Sample	# 1		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:13	
Control Test		0.080	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:14	

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STACHEWICZ,RILEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: DERNULC

DATE/TIME OF TEST: 03/19/2022 02:09:35

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.155	02:05
Breath Volume		3.445	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.152	02:08
Breath Volume		3.003	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: TILTON

DATE/TIME OF TEST: 03/19/2022 05:30:23

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.081	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.232	05:25
Breath Volume		2.187	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.231	05:28
Breath Volume		1.453	05:28
Air Blank		0.000	05:29
Control Test		0.080	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,BRITTNEY,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GOLDEN,RYAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 03/20/2022 22:42:57

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:40	
Control Test		0.080	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:42	
Control Test		0.081	22:42	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TRAVERS,LEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JORDEN

DATE/TIME OF TEST: 03/21/2022 01:03:18

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.080	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MACPHERSON,ZACHARY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/22/2022 01:49:07

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.000	01:44
Breath Volume		2.691	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.000	01:47
Breath Volume		1.890	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MIDDLETON,BLAKE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: BARNHILL

DATE/TIME OF TEST: 03/23/2022 20:41:44

Diagnostics Check		OK	20:34
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:36
Subject Sample	# 1	0.240	20:37
Breath Volume		2.425	20:37
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.238	20:40
Breath Volume		2.843	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPER,CADEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOPER,CADEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HALUSKA

DATE/TIME OF TEST: 03/23/2022 22:20:10

Diagnostics Check		OK	22:06	
Air Blank		0.000	22:06	
Control Test		0.080	22:07	
Air Blank		0.000	22:07	
Subject Sample	# 1	0.105	22:10	Volume Not Met
Breath Volume		0.687	22:10	
Air Blank		0.000	22:11	
Air Blank		0.000	22:13	
Subject Sample	# 2	0.121	22:14	
Breath Volume		1.468	22:14	
Air Blank		0.000	22:14	
Air Blank		0.000	22:16	
Subject Sample	# 3		22:18	RFI Detect
Breath Volume		0.199	22:18	
Air Blank		0.000	22:19	
Control Test		0.078	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 03/25/2022 22:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLER,J,G
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ZOLLER,J
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 02/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CROWDER

DATE/TIME OF TEST: 03/08/2022 02:13:49

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.082	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.202	02:06
Breath Volume		1.589	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.236	02:09
Breath Volume		4.093	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 3	0.234	02:12
Breath Volume		3.875	02:12
Air Blank		0.000	02:12
Control Test		0.078	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 03/25/2022 22:12

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOHNER,KYLER,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TRAVERS,LEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/27/2022 01:39:49

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.081	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Control Test		0.081	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 04/21/2022 19:53

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUIS,NORMAND,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MARQUIS,NORMAND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WILBURN

DATE/TIME OF TEST: 03/30/2022 00:38:54

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:27	
Control Test		0.080	00:28	
Air Blank		0.000	00:28	
Subject Sample	# 1	0.149	00:31	Volume Not Met
Breath Volume		0.128	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2	0.138	00:37	Volume Not Met
Breath Volume		0.066	00:37	
Air Blank		0.000	00:37	
Control Test		0.075	00:38	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 04/21/2022 19:53

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CAREY,ANDREW
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: RIDDLES

DATE/TIME OF TEST: 03/30/2022 02:51:25

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Control Test		0.081	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 04/21/2022 19:53

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESOUSA,CARLOS,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DESOUSA,CARLOS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/30/2022 20:18:14

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:16	
Control Test		0.081	20:16	
Air Blank		0.000	20:16	
Subject Sample	# 1		20:17	RFI Detect
Breath Volume		0.000	20:17	
Air Blank			20:18	RFI Detect
Air Blank			20:18	RFI Detect

Data Download Date/Time: 04/21/2022 19:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 03/30/2022 21:14:08

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.082	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.155	21:09
Breath Volume		2.937	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.152	21:12
Breath Volume		2.863	21:12
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 04/21/2022 19:54

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: ELENZ,TIMOTHY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: MOYE

DATE/TIME OF TEST: 03/31/2022 04:31:43

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.081	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.206	04:27
Breath Volume		5.070	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.207	04:30
Breath Volume		5.035	04:30
Air Blank		0.000	04:30
Control Test		0.079	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 04/21/2022 20:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/31/2022 15:06:55

Diagnostics Check	OK	15:06	
Air Blank	0.000	15:06	
		15:06	RFI Detect
Air Blank		15:06	RFI Detect

Data Download Date/Time: 04/21/2022 20:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIM,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 03/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1482048
DRY GAS EXPIRY DATE: 04/21/2024
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/31/2022 15:15:46

Diagnostics Check		OK	15:09
Air Blank		0.000	15:09
Control Test		0.081	15:10
Air Blank		0.000	15:10
Subject Sample	# 1	0.000	15:11
Breath Volume		4.445	15:11
Air Blank		0.000	15:11
Air Blank		0.000	15:13
Subject Sample	# 2	0.000	15:14
Breath Volume		4.136	15:14
Air Blank		0.000	15:14
Control Test		0.081	15:15
Air Blank		0.000	15:15
Diagnostics Check		OK	15:15

Data Download Date/Time: 04/21/2022 20:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: MADISE,GRAMBLING
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HARVEY FONTENOT

DATE/TIME OF TEST: 03/25/2022 22:36:04

Diagnostics Check		OK	22:25
Air Blank		0.000	22:26
Control Test		0.081	22:26
Air Blank		0.000	22:27
Subject Sample	# 1	0.223	22:28
Breath Volume		2.156	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.199	22:31
Breath Volume		1.417	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 3	0.202	22:34
Breath Volume		1.152	22:34
Air Blank		0.000	22:35
Control Test		0.082	22:35
Air Blank		0.000	22:35
Diagnostics Check		OK	22:36

Data Download Date/Time: 04/04/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SHAFFER,ROBIN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/26/2022 00:45:06

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.082	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.101	00:39
Breath Volume		1.945	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.102	00:43
Breath Volume		2.312	00:43
Air Blank		0.000	00:44
Control Test		0.082	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 04/04/2022 12:47

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIXON,WESLEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LIVINGSTON,KERRIA
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 02/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34520080A1
DRY GAS EXPIRY DATE: 02/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 03/05/2022 23:03:13

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.077	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.000	22:58
Breath Volume		2.808	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.000	23:01
Breath Volume		3.003	23:01
Air Blank		0.000	23:02
Control Test		0.077	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 03/10/2022 11:17

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCEWAN

DATE/TIME OF TEST: 03/04/2022 03:46:04

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.082	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.132	03:40
Breath Volume		1.195	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.128	03:44
Breath Volume		1.699	03:44
Air Blank		0.000	03:45
Control Test		0.081	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:46

Data Download Date/Time: 03/09/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BLAND,CURTIS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 03/04/2022 05:11:59

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.082	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.037	05:07
Breath Volume		3.605	05:07
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.036	05:10
Breath Volume		3.281	05:10
Air Blank		0.000	05:11
Control Test		0.082	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 03/09/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: STONE,KEVIN
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ROLAND

DATE/TIME OF TEST: 03/06/2022 23:17:04

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.081	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.176	23:12
Breath Volume		2.265	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:14
Subject Sample	# 2	0.177	23:15
Breath Volume		1.675	23:15
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:17

Data Download Date/Time: 03/09/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTER,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PORTER,ALEXANDER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/25/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: FORE

DATE/TIME OF TEST: 03/07/2022 01:47:00

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.098	01:42
Breath Volume		2.285	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:44
Subject Sample	# 2	0.101	01:45
Breath Volume		2.203	01:45
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:47

Data Download Date/Time: 03/09/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NEVES

DATE/TIME OF TEST: 03/11/2022 04:56:25

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.081	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.234	04:51
Breath Volume		2.054	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.245	04:54
Breath Volume		2.976	04:54
Air Blank		0.000	04:55
Control Test		0.080	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 04/25/2022 14:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RAUSCH,CLINT
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BUSTAMANTE

DATE/TIME OF TEST: 03/13/2022 06:32:08

Diagnostics Check		OK	06:20	
Air Blank		0.000	06:21	
Control Test		0.082	06:21	
Air Blank		0.000	06:21	
Subject Sample	# 1	0.129	06:24	Volume Not Met
Breath Volume		0.191	06:24	
Air Blank		0.000	06:25	
Air Blank		0.000	06:27	
Subject Sample	# 2	0.138	06:30	Volume Not Met
Breath Volume		0.656	06:30	
Air Blank		0.000	06:31	
Control Test		0.081	06:31	
Air Blank		0.000	06:32	
Diagnostics Check		OK	06:32	

Data Download Date/Time: 04/25/2022 14:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RAUSCH,CLINT
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: BUSTAMANTE

DATE/TIME OF TEST: 03/13/2022 06:50:26

Diagnostics Check		OK	06:40
Air Blank		0.000	06:40
Control Test		0.081	06:40
Air Blank		0.000	06:41
Subject Sample	# 1	0.137	06:43
Breath Volume		1.320	06:43
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.138	06:48
Breath Volume		1.597	06:48
Air Blank		0.000	06:49
Control Test		0.081	06:49
Air Blank		0.000	06:50
Diagnostics Check		OK	06:50

Data Download Date/Time: 04/25/2022 14:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BROSIUS,S
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:34

SUBJECT'S LAST NAME: IACOVINO

DATE/TIME OF TEST: 03/15/2022 12:01:15

Diagnostics Check		OK	11:55
Air Blank		0.000	11:55
Control Test		0.082	11:55
Air Blank		0.000	11:56
Subject Sample	# 1	0.000	11:56
Breath Volume		2.777	11:56
Air Blank		0.000	11:57
Air Blank		0.000	11:59
Subject Sample	# 2	0.000	11:59
Breath Volume		3.082	11:59
Air Blank		0.000	12:00
Control Test		0.081	12:00
Air Blank		0.000	12:01
Diagnostics Check		OK	12:01

Data Download Date/Time: 04/25/2022 14:22

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODALL,CORY,D
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WOODALL,CORY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 03/16/2022 02:50:50

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.081	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.195	02:46
Breath Volume		2.937	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.182	02:49
Breath Volume		1.699	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 04/25/2022 14:27

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHELL,GARY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: CICCARELLI,RAY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: STOTT

DATE/TIME OF TEST: 03/17/2022 05:35:03

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:26	
Control Test		0.082	05:26	
Air Blank		0.000	05:27	
Subject Sample	# 1	0.147	05:30	Volume Not Met
Breath Volume		0.054	05:30	
Air Blank		0.000	05:30	
Air Blank		0.000	05:32	
Subject Sample	# 2		05:33	Subject Test Refused
Breath Volume		0.085	05:33	
Air Blank		0.000	05:34	
Control Test		0.081	05:34	
Air Blank		0.000	05:34	
Diagnostics Check		OK	05:35	

Data Download Date/Time: 04/25/2022 14:27

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ROACHE,DUSTIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: ZING

DATE/TIME OF TEST: 03/17/2022 21:23:30

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.082	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.000	21:17
Breath Volume		1.808	21:17
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.000	21:22
Breath Volume		1.195	21:22
Air Blank		0.000	21:22
Control Test		0.081	21:22
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

Data Download Date/Time: 04/25/2022 14:27

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,FORREST,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALLEN,FORREST
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: COLEWELL

DATE/TIME OF TEST: 03/18/2022 05:45:07

Diagnostics Check		OK	05:38
Air Blank		0.000	05:39
Control Test		0.079	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.177	05:40
Breath Volume		1.507	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 2	0.180	05:43
Breath Volume		1.691	05:43
Air Blank		0.000	05:44
Control Test		0.081	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,FORREST,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALLEN,F
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 03/19/2022 02:12:36

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:01	
Control Test		0.081	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1	0.208	02:02	
Breath Volume		2.273	02:02	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2		02:08	No Sample Provided
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 3		02:11	Subject Test Refused
Breath Volume		0.000	02:11	
Air Blank		0.000	02:11	
Control Test		0.081	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HOUSE,STEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/19/2022 03:24:10

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.089	03:19
Breath Volume		1.742	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.099	03:22
Breath Volume		1.667	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DEXTER,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HALE,CHASE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/19/2022 13:22:29

Diagnostics Check	OK	13:22	
Air Blank		13:22	Improper Sample
Air Blank		13:22	Improper Sample

Data Download Date/Time: 04/25/2022 14:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DEXTER,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HALE,CHASE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/19/2022 13:45:45

Diagnostics Check	OK	13:44	
Air Blank		13:45	Ambient Fail
Air Blank	0.000	13:45	

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,DEXTER,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HALE,CHASE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/19/2022 14:21:21

Diagnostics Check		OK	14:14
Air Blank		0.000	14:15
Control Test		0.080	14:15
Air Blank		0.000	14:15
Subject Sample	# 1	0.239	14:16
Breath Volume		1.972	14:16
Air Blank		0.000	14:17
Air Blank		0.000	14:19
Subject Sample	# 2	0.250	14:19
Breath Volume		2.347	14:19
Air Blank		0.000	14:20
Control Test		0.080	14:20
Air Blank		0.000	14:21
Diagnostics Check		OK	14:21

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PAREDES RODRIGUEZ

DATE/TIME OF TEST: 03/20/2022 04:43:53

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.082	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.199	04:39
Breath Volume		1.316	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.207	04:42
Breath Volume		1.824	04:42
Air Blank		0.000	04:42
Control Test		0.081	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,FORREST,H
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ALLEN,FORREST
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SPRULL

DATE/TIME OF TEST: 03/25/2022 03:42:12

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.082	03:34	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:38	No Sample Provided
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:41	
Control Test		0.081	03:41	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 03/26/2022 21:46:44

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.081	21:42	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:46	No Sample Provided
Breath Volume		0.000	21:46	
Air Blank			21:46	Sequence Aborted
Air Blank		0.000	21:46	

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSE,STEVEN,D
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,CALE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/09/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1305520
DRY GAS EXPIRY DATE: 04/01/2023
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PEREZ FLORES

DATE/TIME OF TEST: 03/29/2022 01:16:25

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.082	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.217	01:12
Breath Volume		3.910	01:12
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.220	01:14
Breath Volume		3.023	01:14
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 04/25/2022 14:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,ISRAEL,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/02/2022 14:24:01

Diagnostics Check		OK	14:17
Air Blank		0.000	14:18
Control Test		0.080	14:18
Air Blank		0.000	14:19
Subject Sample	# 1	0.000	14:19
Breath Volume		2.519	14:19
Air Blank		0.000	14:20
Air Blank		0.000	14:22
Subject Sample	# 2	0.000	14:22
Breath Volume		2.656	14:22
Air Blank		0.000	14:23
Control Test		0.080	14:23
Air Blank		0.000	14:23
Diagnostics Check		OK	14:24

Data Download Date/Time: 03/02/2022 15:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: FRANCISCO,RAFAEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/25/2022 23:24:07

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.197	23:19
Breath Volume		1.605	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:21
Subject Sample	# 2	0.195	23:22
Breath Volume		1.921	23:22
Air Blank		0.000	23:23
Control Test		0.075	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 04/25/2022 14:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: BRANNON

DATE/TIME OF TEST: 03/26/2022 21:16:09

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.080	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.184	21:11
Breath Volume		1.820	21:11
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.185	21:14
Breath Volume		1.457	21:14
Air Blank		0.000	21:15
Control Test		0.078	21:15
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 04/25/2022 14:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: MARTIN,ERIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG121002
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MACKALL

DATE/TIME OF TEST: 03/27/2022 02:42:33

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.079	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:38	No Sample Provided
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.080	02:41	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 04/25/2022 14:42

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/09/2022 04:44:41

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.080	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.214	04:40
Breath Volume		3.074	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.207	04:43
Breath Volume		2.726	04:43
Air Blank		0.000	04:43
Control Test		0.079	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 03/21/2022 19:39

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PAUL,WILLIAM
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/12/2022 01:31:07

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.079	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.176	01:21	Volume Not Met
Breath Volume		0.359	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:24	
Subject Sample	# 2	0.176	01:26	
Breath Volume		1.601	01:26	
Air Blank		0.000	01:26	
Air Blank		0.000	01:28	
Subject Sample	# 3	0.173	01:29	
Breath Volume		1.738	01:29	
Air Blank		0.000	01:30	
Control Test		0.077	01:30	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 03/21/2022 19:39

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,STEVEN,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RODEN,MATTHEW
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 03/15/2022 00:29:34

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.077	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.290	00:24
Breath Volume		1.929	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.270	00:27
Breath Volume		1.359	00:27
Air Blank		0.000	00:28
Control Test		0.075	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 03/21/2022 19:39

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAISER,ANTHONY,J
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: KAISER,ANTHONY
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 03/18/2022 02:21:58

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.206	02:17
Breath Volume		3.324	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.205	02:20
Breath Volume		3.371	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 03/21/2022 19:39

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,CONNOR,S
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUNN,CONNOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 03/18/2022 02:30:18

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:26	
Control Test		0.078	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 03/21/2022 19:39

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOROS,CHUWAN,
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: SHELTON,M
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: GRAHM

DATE/TIME OF TEST: 03/18/2022 08:32:31

Diagnostics Check		OK	08:30	
Air Blank		0.000	08:30	
Control Test		0.079	08:30	
Air Blank		0.000	08:31	
Subject Sample	# 1		08:32	Sequence Aborted
Breath Volume		0.000	08:32	
Air Blank		0.000	08:32	

Data Download Date/Time: 03/21/2022 19:39

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KOUCHNIR,EUGENE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TRANQUILINO

DATE/TIME OF TEST: 03/21/2022 04:10:20

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.080	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.191	04:00	Volume Not Met
Breath Volume		0.332	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 2	0.204	04:03	
Breath Volume		1.832	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 3	0.198	04:08	
Breath Volume		1.539	04:08	
Air Blank		0.000	04:09	
Control Test		0.079	04:09	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 03/21/2022 19:40

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAISER,ANTHONY,J
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: RODEN,MATTHEW
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/24/2022 00:42:46

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.000	00:38
Breath Volume		2.746	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.000	00:41
Breath Volume		2.894	00:41
Air Blank		0.000	00:41
Control Test		0.078	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 04/26/2022 12:06

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/28/2022 03:17:28

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.188	03:11
Breath Volume		1.605	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.190	03:15
Breath Volume		1.835	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 04/26/2022 12:06

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PALMATEER,TRACY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 03/29/2022 14:17:32

Diagnostics Check		OK	14:11
Air Blank		0.000	14:12
Control Test		0.080	14:12
Air Blank		0.000	14:12
Subject Sample	# 1	0.018	14:13
Breath Volume		2.257	14:13
Air Blank		0.000	14:13
Air Blank		0.000	14:15
Subject Sample	# 2	0.015	14:16
Breath Volume		2.203	14:16
Air Blank		0.000	14:16
Control Test		0.079	14:16
Air Blank		0.000	14:17
Diagnostics Check		OK	14:17

Data Download Date/Time: 04/26/2022 12:06

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/21/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: WO32302427
DRY GAS EXPIRY DATE: 09/11/2024
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: KIRCHNER

DATE/TIME OF TEST: 03/31/2022 07:41:05

Diagnostics Check		OK	07:35
Air Blank		0.000	07:35
Control Test		0.080	07:36
Air Blank		0.000	07:36
Subject Sample	# 1	0.000	07:36
Breath Volume		3.199	07:36
Air Blank		0.000	07:37
Air Blank		0.000	07:39
Subject Sample	# 2	0.000	07:39
Breath Volume		3.203	07:39
Air Blank		0.000	07:40
Control Test		0.080	07:40
Air Blank		0.000	07:41
Diagnostics Check		OK	07:41

Data Download Date/Time: 04/26/2022 12:06

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/01/2022 01:02:28

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:54	
Control Test		0.080	00:54	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:58	No Sample Provided
Breath Volume		0.000	00:58	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 2		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.081	01:01	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: ROJAS EUGENIO

DATE/TIME OF TEST: 03/02/2022 17:55:06

Diagnostics Check		OK	17:48
Air Blank		0.000	17:49
Control Test		0.081	17:49
Air Blank		0.000	17:50
Subject Sample	# 1	0.129	17:50
Breath Volume		1.343	17:50
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 2	0.126	17:53
Breath Volume		3.402	17:53
Air Blank		0.000	17:54
Control Test		0.080	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: AYESH-STEVENS

DATE/TIME OF TEST: 03/03/2022 01:01:19

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:58	
Control Test		0.081	00:59	
Air Blank		0.000	00:59	
Subject Sample	# 1		00:59	Subject Test Refused
Breath Volume		0.000	00:59	
Air Blank		0.000	01:00	
Control Test		0.081	01:00	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 03/04/2022 00:44:54

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.027	00:40
Breath Volume		1.445	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.025	00:43
Breath Volume		1.457	00:43
Air Blank		0.000	00:43
Control Test		0.081	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: BUSWELL

DATE/TIME OF TEST: 03/06/2022 19:56:35

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.082	19:50
Air Blank		0.000	19:51
Subject Sample	# 1	0.181	19:52
Breath Volume		1.609	19:52
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.183	19:54
Breath Volume		1.332	19:54
Air Blank		0.000	19:55
Control Test		0.081	19:55
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: BUNTING,JESSICA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: DE HANN

DATE/TIME OF TEST: 03/06/2022 21:52:24

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.081	21:46
Air Blank		0.000	21:46
Subject Sample	# 1	0.000	21:47
Breath Volume		2.070	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.000	21:51
Breath Volume		2.000	21:51
Air Blank		0.000	21:51
Control Test		0.081	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO,MERCEDES,T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HELTON,TRAVIS
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 03/07/2022 00:18:09

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.081	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.000	00:13
Breath Volume		2.816	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.000	00:16
Breath Volume		2.910	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALCAGNO,MERCEDES,T
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: GONZALEZ,VICTOR
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: NAVA

DATE/TIME OF TEST: 03/07/2022 01:13:30

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.286	01:05
Breath Volume		2.871	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.258	01:08
Breath Volume		2.574	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 3	0.262	01:11
Breath Volume		3.472	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: TAGUJA-DORANTES

DATE/TIME OF TEST: 03/07/2022 21:52:13

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:49	
Control Test		0.081	21:50	
Air Blank		0.000	21:50	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:51	
Control Test		0.081	21:51	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BUCKNER

DATE/TIME OF TEST: 03/09/2022 02:15:58

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.081	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:14	Subject Test Refused
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 03/10/2022 00:54:07

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.109	00:49
Breath Volume		2.703	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.103	00:52
Breath Volume		1.503	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 03/11/2022 21:49:16

Diagnostics Check		OK	21:42
Air Blank		0.000	21:43
Control Test		0.082	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.000	21:44
Breath Volume		1.789	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.000	21:47
Breath Volume		1.250	21:47
Air Blank		0.000	21:48
Control Test		0.082	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/12/2022 03:26:30

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.081	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:25	Subject Test Refused
Breath Volume		0.000	03:25	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: LEON PEREZ

DATE/TIME OF TEST: 03/13/2022 03:03:21

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.312	01:58
Breath Volume		2.921	01:58
Air Blank		0.000	01:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.297	03:01
Breath Volume		1.628	03:01
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MEZA

DATE/TIME OF TEST: 03/17/2022 02:57:01

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.000	02:52
Breath Volume		1.835	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.000	02:55
Breath Volume		2.257	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:57

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: REYEZ

DATE/TIME OF TEST: 03/18/2022 00:43:37

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.081	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.117	00:39
Breath Volume		4.093	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.111	00:42
Breath Volume		1.636	00:42
Air Blank		0.000	00:42
Control Test		0.081	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 03/19/2022 18:58:39

Diagnostics Check		OK	18:50
Air Blank		0.000	18:51
Control Test		0.081	18:51
Air Blank		0.000	18:51
Subject Sample	# 1	0.123	18:52
Breath Volume		1.570	18:52
Air Blank		0.000	18:53
Air Blank		0.000	18:55
Subject Sample	# 2	0.134	18:57
Breath Volume		1.863	18:57
Air Blank		0.000	18:57
Control Test		0.080	18:58
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BUTTS

DATE/TIME OF TEST: 03/21/2022 23:53:10

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.082	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.000	23:48
Breath Volume		2.339	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.000	23:51
Breath Volume		1.429	23:51
Air Blank		0.000	23:52
Control Test		0.081	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 03/22/2022 09:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: ENFINGER

DATE/TIME OF TEST: 03/24/2022 20:56:54

Diagnostics Check		OK	20:52	
Air Blank		0.000	20:53	
Control Test		0.081	20:53	
Air Blank		0.000	20:54	
Subject Sample	# 1		20:55	Subject Test Refused
Breath Volume		0.000	20:55	
Air Blank		0.000	20:55	
Control Test		0.081	20:56	
Air Blank		0.000	20:56	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 04/12/2022 13:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: GAURONIAK,ERIC
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/24/2022 22:05:17

Diagnostics Check		OK	22:00	
Air Blank		0.000	22:01	
Control Test		0.080	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 1		22:03	Subject Test Refused
Breath Volume		0.000	22:03	
Air Blank		0.000	22:04	
Control Test		0.080	22:04	
Air Blank		0.000	22:05	
Diagnostics Check		OK	22:05	

Data Download Date/Time: 04/12/2022 13:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MUMFORD

DATE/TIME OF TEST: 03/25/2022 20:49:30

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.081	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.000	20:45
Breath Volume		1.703	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.000	20:48
Breath Volume		2.343	20:48
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 04/12/2022 13:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/25/2022 23:43:37

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.271	23:39
Breath Volume		2.085	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.266	23:42
Breath Volume		2.089	23:42
Air Blank		0.000	23:42
Control Test		0.078	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 04/12/2022 13:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 03/26/2022 20:39:11

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.139	20:34
Breath Volume		1.320	20:34
Air Blank		0.000	20:34
Air Blank		0.000	20:36
Subject Sample	# 2	0.135	20:37
Breath Volume		2.074	20:37
Air Blank		0.000	20:38
Control Test		0.079	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 04/12/2022 13:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 03/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 03/29/2022 03:37:00

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.000	03:32
Breath Volume		2.785	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.000	03:35
Breath Volume		2.808	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

Data Download Date/Time: 04/12/2022 13:01

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPLEBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RIPS

DATE/TIME OF TEST: 03/06/2022 04:22:19

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.140	04:17
Breath Volume		2.632	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.138	04:20
Breath Volume		2.847	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 03/29/2022 16:10

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONCALVES,JOAO,E
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: KRECIC,DYLAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 03/11/2022 18:53:24

Diagnostics Check		OK	18:46
Air Blank		0.000	18:47
Control Test		0.081	18:47
Air Blank		0.000	18:47
Subject Sample	# 1	0.208	18:49
Breath Volume		3.238	18:49
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.207	18:51
Breath Volume		2.230	18:51
Air Blank		0.000	18:52
Control Test		0.079	18:52
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 03/29/2022 16:10

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCARDI,A,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HERNANDEZ,Y
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CHANO

DATE/TIME OF TEST: 03/18/2022 03:01:00

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.223	02:55
Breath Volume		1.718	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.226	02:59
Breath Volume		2.480	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:01

Data Download Date/Time: 03/29/2022 16:10

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPLEBAUM,K
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 03/20/2022 04:59:20

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.071	04:54
Breath Volume		2.332	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.064	04:57
Breath Volume		1.726	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 03/29/2022 16:10

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: VARGUS,N
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 03/21/2022 02:55:07

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.080	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:53	Subject Test Refused
Breath Volume		0.000	02:53	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 03/29/2022 16:10

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODRUFF,MARK,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: VARGUS,NORB
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 03/21/2022 03:26:18

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.081	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.214	03:22
Breath Volume		1.101	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.215	03:24
Breath Volume		1.273	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 03/29/2022 16:10

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRNE,JOSEPH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: HARRIS,IAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A3
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: CAMBRON

DATE/TIME OF TEST: 03/24/2022 09:58:23

Diagnostics Check		OK	09:52
Air Blank		0.000	09:52
Control Test		0.080	09:53
Air Blank		0.000	09:53
Subject Sample	# 1	0.261	09:53
Breath Volume		1.453	09:53
Air Blank		0.000	09:54
Air Blank		0.000	09:56
Subject Sample	# 2	0.264	09:56
Breath Volume		1.203	09:56
Air Blank		0.000	09:57
Control Test		0.079	09:57
Air Blank		0.000	09:58
Diagnostics Check		OK	09:58

Data Download Date/Time: 03/29/2022 16:10

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MUNNI,ALEC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: GAIS

DATE/TIME OF TEST: 03/01/2022 16:37:27

Diagnostics Check		OK	16:30
Air Blank		0.000	16:31
Control Test		0.080	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.000	16:32
Breath Volume		3.082	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.000	16:36
Breath Volume		3.597	16:36
Air Blank		0.000	16:36
Control Test		0.080	16:36
Air Blank		0.000	16:37
Diagnostics Check		OK	16:37

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WICKERT,PAUL
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: COLANGELO

DATE/TIME OF TEST: 03/03/2022 23:13:28

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.080	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:12	Subject Test Refused
Breath Volume		0.000	23:12	
Air Blank		0.000	23:12	
Control Test		0.080	23:12	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: LIGHTSEY,DUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: HAND

DATE/TIME OF TEST: 03/04/2022 01:00:27

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.079	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.000	00:56
Breath Volume		2.250	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.000	00:59
Breath Volume		1.992	00:59
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RIFFEY

DATE/TIME OF TEST: 03/04/2022 20:27:40

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.080	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.163	20:23
Breath Volume		1.148	20:23
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.159	20:26
Breath Volume		1.277	20:26
Air Blank		0.000	20:26
Control Test		0.080	20:27
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERIC,JEAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FREDERIC,JEAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: KLUCKOWSKI

DATE/TIME OF TEST: 03/05/2022 22:28:32

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.000	22:23
Breath Volume		1.210	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.000	22:27
Breath Volume		2.917	22:27
Air Blank		0.000	22:27
Control Test		0.079	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: STERN

DATE/TIME OF TEST: 03/05/2022 23:09:01

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	23:00
Subject Sample	# 1	0.168	23:01
Breath Volume		1.359	23:01
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.191	23:04
Breath Volume		1.902	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:06
Subject Sample	# 3	0.200	23:07
Breath Volume		2.234	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:09

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PIZZUTI,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DIMARCO

DATE/TIME OF TEST: 03/06/2022 00:37:25

Diagnostics Check		OK	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:29
Subject Sample	# 1	0.197	00:31
Breath Volume		2.761	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.196	00:35
Breath Volume		2.496	00:35
Air Blank		0.000	00:36
Control Test		0.078	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEARSE,A,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 03/06/2022 04:49:14

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.171	04:44
Breath Volume		1.417	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.174	04:47
Breath Volume		2.781	04:47
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: ARAUJO

DATE/TIME OF TEST: 03/06/2022 23:00:52

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.219	22:56
Breath Volume		2.175	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.217	22:59
Breath Volume		2.406	22:59
Air Blank		0.000	22:59
Control Test		0.079	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 03/26/2022 17:18

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATSON,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: MATSON,ZACHARY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HAUGLAND

DATE/TIME OF TEST: 03/07/2022 02:17:26

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.079	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/08/2022 21:01:50

Diagnostics Check		OK	20:58	
Air Blank		0.000	20:59	
Control Test		0.080	20:59	
Air Blank		0.000	20:59	
Subject Sample	# 1		21:00	Subject Test Refused
Breath Volume		0.000	21:00	
Air Blank		0.000	21:00	
Control Test		0.080	21:01	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 03/10/2022 23:51:38

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:47	
Control Test		0.079	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:50	Subject Test Refused
Breath Volume		0.000	23:50	
Air Blank		0.000	23:50	
Control Test		0.079	23:51	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:51	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WITLES

DATE/TIME OF TEST: 03/11/2022 21:26:34

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.080	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:25	Subject Test Refused
Breath Volume		0.000	21:25	
Air Blank		0.000	21:25	
Control Test		0.080	21:25	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: BARRANCO

DATE/TIME OF TEST: 03/12/2022 05:05:14

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:02	
Control Test		0.079	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 1		05:03	Subject Test Refused
Breath Volume		0.000	05:03	
Air Blank		0.000	05:04	
Control Test		0.079	05:04	
Air Blank		0.000	05:05	
Diagnostics Check		OK	05:05	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROACHE,H
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ZORC

DATE/TIME OF TEST: 03/13/2022 01:38:52

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:34	
Control Test		0.081	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.081	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROACHE,H
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 03/13/2022 21:57:35

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:55	
Control Test		0.081	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 1		21:56	Subject Test Refused
Breath Volume		0.000	21:56	
Air Blank		0.000	21:56	
Control Test		0.080	21:56	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BRUGGER

DATE/TIME OF TEST: 03/16/2022 02:21:56

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.094	02:17
Breath Volume		3.019	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:19
Subject Sample	# 2	0.091	02:20
Breath Volume		4.125	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,BRANDON
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PINHASSIAN

DATE/TIME OF TEST: 03/18/2022 01:59:40

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:57	
Subject Sample	# 1		01:58	Subject Test Refused
Breath Volume		0.000	01:58	
Air Blank		0.000	01:58	
Control Test		0.079	01:59	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 03/26/2022 17:19

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HARTMAN,RYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: PHAMPHILE

DATE/TIME OF TEST: 03/18/2022 10:59:48

Diagnostics Check		OK	10:53
Air Blank		0.000	10:53
Control Test		0.080	10:53
Air Blank		0.000	10:54
Subject Sample	# 1	0.000	10:55
Breath Volume		2.359	10:55
Air Blank		0.000	10:55
Air Blank		0.000	10:57
Subject Sample	# 2	0.000	10:58
Breath Volume		1.203	10:58
Air Blank		0.000	10:58
Control Test		0.080	10:59
Air Blank		0.000	10:59
Diagnostics Check		OK	10:59

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,ERIC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: VOLLBRACHT

DATE/TIME OF TEST: 03/19/2022 18:11:09

Diagnostics Check		OK	18:04
Air Blank		0.000	18:05
Control Test		0.079	18:05
Air Blank		0.000	18:06
Subject Sample	# 1	0.025	18:06
Breath Volume		2.332	18:06
Air Blank		0.000	18:07
Air Blank		0.000	18:09
Subject Sample	# 2	0.024	18:09
Breath Volume		2.355	18:09
Air Blank		0.000	18:10
Control Test		0.079	18:10
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 03/19/2022 19:22:54

Diagnostics Check		OK	19:17	
Air Blank		0.000	19:18	
Control Test		0.079	19:18	
Air Blank		0.000	19:19	
Subject Sample	# 1		19:21	RFI Detect
Breath Volume		0.000	19:21	
Air Blank		0.000	19:21	
Control Test		0.080	19:22	
Air Blank		0.000	19:22	
Diagnostics Check		OK	19:22	

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 03/19/2022 19:25:00

Diagnostics Check	OK	19:24	
Air Blank		19:24	RFI Detect
Air Blank		19:25	RFI Detect

Data Download Date/Time: 03/26/2022 17:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 03/19/2022 19:27:13

Diagnostics Check	OK	19:26	
Air Blank		19:27	RFI Detect
Air Blank		19:27	RFI Detect

Data Download Date/Time: 03/26/2022 17:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 03/19/2022 19:35:33

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:30	
Control Test		0.080	19:30	
Air Blank		0.000	19:30	
Subject Sample	# 1		19:34	No Sample Provided
Breath Volume		0.000	19:34	
Air Blank			19:34	RFI Detect
Air Blank		0.000	19:34	
Control Test		0.079	19:34	
Air Blank		0.000	19:35	
Diagnostics Check		OK	19:35	

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 03/19/2022 19:42:36

Diagnostics Check		OK	19:39	
Air Blank		0.000	19:40	
Control Test		0.079	19:40	
Air Blank		0.000	19:41	
Subject Sample	# 1		19:41	Subject Test Refused
Breath Volume		0.000	19:41	
Air Blank		0.000	19:41	
Control Test		0.079	19:42	
Air Blank		0.000	19:42	
Diagnostics Check		OK	19:42	

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KESKINEN,D
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HANBIDGE

DATE/TIME OF TEST: 03/19/2022 22:53:53

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:51	
Control Test		0.079	22:51	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:52	Subject Test Refused
Breath Volume		0.000	22:52	
Air Blank		0.000	22:52	
Control Test		0.079	22:53	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,C,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEIGHAN,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 03/20/2022 03:14:44

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:03	
Control Test		0.079	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.224	03:07	Slope Not Level
Breath Volume		1.250	03:07	
Air Blank		0.000	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 2	0.224	03:10	
Breath Volume		1.761	03:10	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 3	0.218	03:13	
Breath Volume		1.355	03:13	
Air Blank			03:14	RFI Detect
Air Blank		0.000	03:14	

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: YORK,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 03/20/2022 19:37:31

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.079	19:32
Air Blank		0.000	19:32
Subject Sample	# 1	0.000	19:33
Breath Volume		2.769	19:33
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.000	19:36
Breath Volume		2.722	19:36
Air Blank		0.000	19:36
Control Test		0.079	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTMAN,KEVIN,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,N
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 03/24/2022 17:25:36

Diagnostics Check		OK	17:19
Air Blank		0.000	17:19
Control Test		0.079	17:19
Air Blank		0.000	17:20
Subject Sample	# 1	0.363	17:20
Breath Volume		1.699	17:20
Air Blank		0.000	17:21
Air Blank		0.000	17:23
Subject Sample	# 2	0.349	17:24
Breath Volume		1.410	17:24
Air Blank		0.000	17:24
Control Test		0.079	17:24
Air Blank		0.000	17:25
Diagnostics Check		OK	17:25

Data Download Date/Time: 03/26/2022 17:20

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: THOMAS,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 03/24/2022 21:40:41

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:38	
Control Test		0.079	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:39	Subject Test Refused
Breath Volume		0.000	21:39	
Air Blank		0.000	21:39	
Control Test		0.080	21:40	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 03/26/2022 17:21

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: AYALA,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GRAWE

DATE/TIME OF TEST: 03/25/2022 23:18:43

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.080	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:17	
Control Test		0.081	23:18	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 03/26/2022 17:21

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEACH,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 03/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 03/30/2022 00:50:32

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:38	
Subject Sample	# 1	0.082	00:42	Volume Not Met
Breath Volume		0.449	00:42	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2	0.125	00:45	
Breath Volume		1.527	00:45	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 3	0.119	00:48	
Breath Volume		2.351	00:48	
Air Blank		0.000	00:49	
Control Test		0.078	00:49	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:50	

Data Download Date/Time: 04/27/2022 06:34

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/01/2022 13:22:24

Diagnostics Check	OK	13:20	
Air Blank	0.000	13:21	
Control Test	0.000	13:21	Control Outside Tolerance
Air Blank	0.000	13:22	

Data Download Date/Time: 04/21/2022 11:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DOE,JON
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 03/01/2022 13:25:39

Diagnostics Check	OK	13:23	
Air Blank	0.000	13:24	
Control Test	0.000	13:25	Control Outside Tolerance
Air Blank	0.000	13:25	

Data Download Date/Time: 04/21/2022 11:09

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 03/08/2022 22:24:14

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.111	22:19
Breath Volume		2.097	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:21
Subject Sample	# 2	0.110	22:22
Breath Volume		1.523	22:22
Air Blank		0.000	22:23
Control Test		0.081	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 04/19/2022 09:31

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/12/2022 00:32:16

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.081	00:23	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:26	No Sample Provided
Breath Volume		0.000	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2		00:30	Subject Test Refused
Breath Volume		0.000	00:30	
Air Blank		0.000	00:31	
Control Test		0.081	00:31	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 04/19/2022 09:31

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KRONDAHL,CARL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/16/2022 20:02:24

Diagnostics Check		OK	19:52
Air Blank		0.000	19:53
Control Test		0.081	19:53
Air Blank		0.000	19:54
Subject Sample	# 1	0.233	19:56
Breath Volume		3.019	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.229	20:00
Breath Volume		3.000	20:00
Air Blank		0.000	20:01
Control Test		0.080	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 04/19/2022 09:31

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: COE,BARRY
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/20/2022 22:28:00

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.081	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.150	22:22
Breath Volume		1.316	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.150	22:26
Breath Volume		1.453	22:26
Air Blank		0.000	22:27
Control Test		0.079	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:28

Data Download Date/Time: 04/19/2022 09:31

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KRONDAHL,CARL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: AVINA

DATE/TIME OF TEST: 03/21/2022 00:57:14

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.000	00:52
Breath Volume		1.351	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.000	00:55
Breath Volume		1.972	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 04/19/2022 09:31

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KRONDAHL,CARL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: FLATT

DATE/TIME OF TEST: 03/21/2022 20:53:27

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:51	
Control Test		0.081	20:51	
Air Blank		0.000	20:51	
Subject Sample	# 1		20:52	Subject Test Refused
Breath Volume		0.000	20:52	
Air Blank		0.000	20:52	
Control Test		0.081	20:52	
Air Blank		0.000	20:53	
Diagnostics Check		OK	20:53	

Data Download Date/Time: 04/19/2022 09:32

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MCCLOSKEY

DATE/TIME OF TEST: 03/25/2022 21:20:54

Diagnostics Check		OK	21:17	
Air Blank		0.000	21:17	
Control Test		0.081	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1		21:19	Subject Test Refused
Breath Volume		0.000	21:19	
Air Blank		0.000	21:19	
Control Test		0.081	21:20	
Air Blank		0.000	21:20	
Diagnostics Check		OK	21:20	

Data Download Date/Time: 04/19/2022 09:32

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,WAYNE,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 211143
DRY GAS EXPIRY DATE: 07/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 03/26/2022 02:35:32

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:32	
Control Test		0.081	02:33	
Air Blank		0.000	02:33	
Subject Sample	# 1		02:34	Subject Test Refused
Breath Volume		0.000	02:34	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 04/19/2022 09:32

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TIRGRATH

DATE/TIME OF TEST: 03/02/2022 03:47:06

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.181	03:42
Breath Volume		4.089	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.180	03:45
Breath Volume		4.003	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BOUREY

DATE/TIME OF TEST: 03/03/2022 02:52:07

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.183	02:47
Breath Volume		2.101	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:49
Subject Sample	# 2	0.181	02:50
Breath Volume		1.671	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: BRATTON

DATE/TIME OF TEST: 03/04/2022 23:08:29

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.108	23:04
Breath Volume		1.789	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.108	23:07
Breath Volume		3.007	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 03/05/2022 02:31:04

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.178	02:26
Breath Volume		2.972	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.181	02:29
Breath Volume		3.148	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:31

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: ARNE

DATE/TIME OF TEST: 03/05/2022 05:13:53

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.080	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.099	05:09
Breath Volume		1.277	05:09
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.094	05:12
Breath Volume		1.375	05:12
Air Blank		0.000	05:12
Control Test		0.080	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LITHERLAND,J
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HANNAN

DATE/TIME OF TEST: 03/06/2022 02:02:24

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.081	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:55	No Sample Provided
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.083	02:00	Volume Not Met
Breath Volume		0.488	02:00	
Air Blank		0.000	02:01	
Control Test		0.080	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KARPOWECZ

DATE/TIME OF TEST: 03/06/2022 03:06:08

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.146	03:01
Breath Volume		2.972	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.140	03:04
Breath Volume		2.648	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIETRON,MICHAEL
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/07/2022 02:24:26

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.142	02:19
Breath Volume		1.203	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.144	02:22
Breath Volume		1.238	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: HUGAN

DATE/TIME OF TEST: 03/07/2022 21:39:02

Diagnostics Check		OK	21:32
Air Blank		0.000	21:33
Control Test		0.078	21:33
Air Blank		0.000	21:34
Subject Sample	# 1	0.167	21:34
Breath Volume		3.328	21:34
Air Blank		0.000	21:35
Air Blank		0.000	21:36
Subject Sample	# 2	0.167	21:37
Breath Volume		3.125	21:37
Air Blank		0.000	21:38
Control Test		0.078	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:39

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTUGUES,GABRIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PORTUGUES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: COFFEE

DATE/TIME OF TEST: 03/08/2022 02:06:09

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.155	02:01
Breath Volume		3.101	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.149	02:04
Breath Volume		2.062	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/08/2022 19:33:22

Diagnostics Check		OK	19:27
Air Blank		0.000	19:27
Control Test		0.079	19:28
Air Blank		0.000	19:28
Subject Sample	# 1	0.000	19:29
Breath Volume		2.281	19:29
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.000	19:32
Breath Volume		1.964	19:32
Air Blank		0.000	19:32
Control Test		0.080	19:32
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: COE

DATE/TIME OF TEST: 03/09/2022 20:57:45

Diagnostics Check		OK	20:50
Air Blank		0.000	20:51
Control Test		0.080	20:51
Air Blank		0.000	20:52
Subject Sample	# 1	0.186	20:52
Breath Volume		1.605	20:52
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.188	20:56
Breath Volume		1.679	20:56
Air Blank		0.000	20:56
Control Test		0.080	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KLEIN,BRETT
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 03/09/2022 21:22:24

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.080	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1	0.231	21:15	Volume Not Met
Breath Volume		0.000	21:15	
Air Blank		0.000	21:15	
Air Blank		0.000	21:17	
Subject Sample	# 2	0.228	21:20	Volume Not Met
Breath Volume		0.718	21:20	
Air Blank		0.000	21:21	
Control Test		0.078	21:21	
Air Blank		0.000	21:22	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WAKE

DATE/TIME OF TEST: 03/10/2022 02:52:05

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.087	02:47
Breath Volume		1.562	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.087	02:50
Breath Volume		1.929	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 03/10/2022 22:39:05

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.133	22:34
Breath Volume		3.339	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.128	22:37
Breath Volume		2.347	22:37
Air Blank		0.000	22:38
Control Test		0.080	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:39

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 03/11/2022 04:08:47

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.136	04:04
Breath Volume		2.878	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.132	04:07
Breath Volume		2.652	04:07
Air Blank		0.000	04:07
Control Test		0.077	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,JUAN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CANTWELL,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: LINDBLOM

DATE/TIME OF TEST: 03/11/2022 15:25:23

Diagnostics Check		OK	15:18
Air Blank		0.000	15:19
Control Test		0.079	15:19
Air Blank		0.000	15:19
Subject Sample	# 1	0.224	15:20
Breath Volume		1.921	15:20
Air Blank		0.000	15:21
Air Blank		0.000	15:23
Subject Sample	# 2	0.229	15:23
Breath Volume		1.808	15:23
Air Blank		0.000	15:24
Control Test		0.080	15:24
Air Blank		0.000	15:25
Diagnostics Check		OK	15:25

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: JIMINEZTAPIA

DATE/TIME OF TEST: 03/12/2022 01:37:00

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:30	No Sample Provided
Breath Volume		0.000	01:30	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2		01:35	No Sample Provided
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Control Test		0.080	01:36	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 03/12/2022 03:25:48

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:13	
Control Test		0.079	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 1	0.171	03:17	Slope Not Level
Breath Volume		2.695	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 2	0.173	03:20	
Breath Volume		2.382	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 3	0.178	03:24	
Breath Volume		2.632	03:24	
Air Blank		0.000	03:24	
Control Test		0.079	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILHARDT,DEP,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LANSING,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: LONGHURST

DATE/TIME OF TEST: 03/12/2022 06:17:20

Diagnostics Check		OK	06:11
Air Blank		0.000	06:11
Control Test		0.078	06:11
Air Blank		0.000	06:12
Subject Sample	# 1	0.141	06:12
Breath Volume		3.207	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.143	06:15
Breath Volume		3.359	06:15
Air Blank		0.000	06:16
Control Test		0.077	06:16
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/12/2022 23:05:52

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.081	23:00
Air Blank		0.000	23:01
Subject Sample	# 1	0.181	23:01
Breath Volume		1.484	23:01
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.178	23:04
Breath Volume		1.796	23:04
Air Blank		0.000	23:04
Control Test		0.081	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 03/13/2022 05:03:51

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.081	04:58
Air Blank		0.000	04:58
Subject Sample	# 1	0.147	04:59
Breath Volume		1.421	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.150	05:02
Breath Volume		1.410	05:02
Air Blank		0.000	05:02
Control Test		0.081	05:03
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SMYTH,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 03/13/2022 05:48:51

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.081	05:33	
Air Blank		0.000	05:33	
Subject Sample	# 1	0.127	05:37	Volume Not Met
Breath Volume		0.781	05:37	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 2	0.119	05:41	
Breath Volume		1.195	05:41	
Air Blank		0.000	05:42	
Air Blank		0.000	05:44	
Subject Sample	# 3	0.128	05:47	Volume Not Met
Breath Volume		0.058	05:47	
Air Blank		0.000	05:47	
Control Test		0.080	05:48	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 03/13/2022 11:35

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SKINNER,DAVID
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: WOLCOTT

DATE/TIME OF TEST: 03/14/2022 20:41:14

Diagnostics Check		OK	20:29	
Air Blank		0.000	20:30	
Control Test		0.081	20:30	
Air Blank		0.000	20:31	
Subject Sample	# 1		20:34	No Sample Provided
Breath Volume		0.000	20:34	
Air Blank		0.000	20:34	
Air Blank		0.000	20:36	
Subject Sample	# 2		20:39	No Sample Provided
Breath Volume		0.000	20:39	
Air Blank		0.000	20:40	
Control Test		0.081	20:40	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: TABURIAUX

DATE/TIME OF TEST: 03/14/2022 20:58:57

Diagnostics Check		OK	20:47	
Air Blank		0.000	20:48	
Control Test		0.081	20:48	
Air Blank		0.000	20:48	
Subject Sample	# 1	0.120	20:52	Volume Not Met
Breath Volume		0.050	20:52	
Air Blank		0.000	20:52	
Air Blank		0.000	20:54	
Subject Sample	# 2		20:57	No Sample Provided
Breath Volume		0.000	20:57	
Air Blank		0.000	20:58	
Control Test		0.080	20:58	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,STEFANIE,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SILVA,STEFANIE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 03/15/2022 02:25:16

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.208	02:20
Breath Volume		2.582	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.201	02:23
Breath Volume		2.066	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: GJOKA

DATE/TIME OF TEST: 03/15/2022 02:40:47

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.021	02:36
Breath Volume		2.835	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.019	02:39
Breath Volume		2.230	02:39
Air Blank		0.000	02:39
Control Test		0.081	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MULLET

DATE/TIME OF TEST: 03/15/2022 04:19:43

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.085	04:15
Breath Volume		1.500	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.088	04:18
Breath Volume		2.636	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FORTUNO,JAMES
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 03/15/2022 19:23:06

Diagnostics Check		OK	19:17
Air Blank		0.000	19:17
Control Test		0.080	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.000	19:18
Breath Volume		3.070	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.000	19:21
Breath Volume		3.179	19:21
Air Blank		0.000	19:22
Control Test		0.079	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ACHILLE

DATE/TIME OF TEST: 03/16/2022 00:40:17

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:36	
Control Test		0.080	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:38	Subject Test Refused
Breath Volume		0.000	00:38	
Air Blank		0.000	00:39	
Control Test		0.079	00:39	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,DEPUTY J.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 03/16/2022 06:43:24

Diagnostics Check		OK	06:37
Air Blank		0.000	06:37
Control Test		0.081	06:37
Air Blank		0.000	06:38
Subject Sample	# 1	0.129	06:38
Breath Volume		4.003	06:38
Air Blank		0.000	06:39
Air Blank		0.000	06:41
Subject Sample	# 2	0.125	06:41
Breath Volume		2.750	06:41
Air Blank		0.000	06:42
Control Test		0.080	06:42
Air Blank		0.000	06:43
Diagnostics Check		OK	06:43

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,DEP,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ROBERTS,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TUPPER

DATE/TIME OF TEST: 03/18/2022 04:01:48

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.189	03:57
Breath Volume		2.328	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.189	04:00
Breath Volume		2.726	04:00
Air Blank		0.000	04:00
Control Test		0.080	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: KURSCHNER

DATE/TIME OF TEST: 03/18/2022 05:19:46

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.082	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.132	05:15
Breath Volume		1.816	05:15
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.129	05:18
Breath Volume		2.136	05:18
Air Blank		0.000	05:18
Control Test		0.081	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NELSON,OFC
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: UHOUSE

DATE/TIME OF TEST: 03/18/2022 23:15:27

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.079	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.109	23:11
Breath Volume		1.191	23:11
Air Blank		0.000	23:11
Air Blank		0.000	23:13
Subject Sample	# 2	0.114	23:13
Breath Volume		1.570	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MITCHELL,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: DUMOND

DATE/TIME OF TEST: 03/20/2022 04:28:43

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.081	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.187	04:24
Breath Volume		3.347	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.179	04:27
Breath Volume		2.687	04:27
Air Blank		0.000	04:27
Control Test		0.079	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: ZUGAJ

DATE/TIME OF TEST: 03/21/2022 19:39:29

Diagnostics Check		OK	19:33
Air Blank		0.000	19:33
Control Test		0.080	19:33
Air Blank		0.000	19:34
Subject Sample	# 1	0.233	19:34
Breath Volume		1.941	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.228	19:38
Breath Volume		1.652	19:38
Air Blank		0.000	19:38
Control Test		0.078	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKLEY,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ACKLEY,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LOFTUS

DATE/TIME OF TEST: 03/22/2022 01:42:05

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.106	01:37
Breath Volume		1.824	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.109	01:40
Breath Volume		2.097	01:40
Air Blank		0.000	01:41
Control Test		0.078	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,J,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BARR,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 03/22/2022 18:16:00

Diagnostics Check		OK	18:09
Air Blank		0.000	18:10
Control Test		0.081	18:10
Air Blank		0.000	18:11
Subject Sample	# 1	0.000	18:11
Breath Volume		2.199	18:11
Air Blank		0.000	18:12
Air Blank		0.000	18:14
Subject Sample	# 2	0.000	18:14
Breath Volume		2.386	18:14
Air Blank		0.000	18:15
Control Test		0.080	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:16

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,KYLE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: TROTTER

DATE/TIME OF TEST: 03/22/2022 22:52:07

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.166	22:47
Breath Volume		1.605	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.172	22:50
Breath Volume		1.890	22:50
Air Blank		0.000	22:51
Control Test		0.080	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,D,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DALEY

DATE/TIME OF TEST: 03/25/2022 03:05:14

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.171	02:59
Breath Volume		2.148	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.188	03:03
Breath Volume		1.273	03:03
Air Blank		0.000	03:04
Control Test		0.081	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BILHARDT,DEP,K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GILE,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 03/30/2022 15:39:20

Diagnostics Check		OK	15:33
Air Blank		0.000	15:33
Control Test		0.081	15:34
Air Blank		0.000	15:34
Subject Sample	# 1	0.030	15:35
Breath Volume		1.660	15:35
Air Blank		0.000	15:35
Air Blank		0.000	15:37
Subject Sample	# 2	0.028	15:37
Breath Volume		1.992	15:37
Air Blank		0.000	15:38
Control Test		0.080	15:38
Air Blank		0.000	15:39
Diagnostics Check		OK	15:39

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15421080A1
DRY GAS EXPIRY DATE: 08/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BLANKENBILLER

DATE/TIME OF TEST: 03/30/2022 23:46:09

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.107	23:41
Breath Volume		1.816	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:43
Subject Sample	# 2	0.094	23:44
Breath Volume		1.816	23:44
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 04/28/2022 10:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DEPUTY D.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/13/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14020080A1
DRY GAS EXPIRY DATE: 07/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: BRODACZ

DATE/TIME OF TEST: 03/19/2022 19:55:39

Diagnostics Check		OK	19:49
Air Blank		0.000	19:50
Control Test		0.077	19:50
Air Blank		0.000	19:50
Subject Sample	# 1	0.137	19:51
Breath Volume		2.707	19:51
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.131	19:54
Breath Volume		2.746	19:54
Air Blank		0.000	19:54
Control Test		0.077	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 04/28/2022 10:07

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MESA NUNEZ,WALDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 03/03/2022 22:06:13

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.081	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:01
Breath Volume		2.667	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:04
Breath Volume		2.609	22:04
Air Blank		0.000	22:05
Control Test		0.081	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 04/04/2022 14:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,JAMIE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MENDEZ,JAMIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: BRIGIDO

DATE/TIME OF TEST: 03/07/2022 07:24:06

Diagnostics Check		OK	07:16
Air Blank		0.000	07:16
Control Test		0.081	07:17
Air Blank		0.000	07:17
Subject Sample	# 1	0.147	07:18
Breath Volume		1.105	07:18
Air Blank		0.000	07:19
Air Blank		0.000	07:21
Subject Sample	# 2	0.153	07:22
Breath Volume		1.695	07:22
Air Blank		0.000	07:23
Control Test		0.080	07:23
Air Blank		0.000	07:24
Diagnostics Check		OK	07:24

Data Download Date/Time: 04/04/2022 14:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 03/12/2022 06:49:48

Diagnostics Check		OK	06:44	
Air Blank		0.000	06:44	
Control Test		0.080	06:44	
Air Blank		0.000	06:45	
Subject Sample	# 1	0.150	06:48	Volume Not Met
Breath Volume		0.160	06:48	
Air Blank			06:49	Purge Fail
Air Blank		0.000	06:49	

Data Download Date/Time: 04/04/2022 14:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 03/12/2022 06:58:18

Diagnostics Check		OK	06:51
Air Blank		0.000	06:52
Control Test		0.079	06:52
Air Blank		0.000	06:53
Subject Sample	# 1	0.150	06:53
Breath Volume		1.546	06:53
Air Blank		0.000	06:54
Air Blank		0.000	06:56
Subject Sample	# 2	0.135	06:56
Breath Volume		1.398	06:56
Air Blank		0.000	06:57
Control Test		0.079	06:57
Air Blank		0.000	06:58
Diagnostics Check		OK	06:58

Data Download Date/Time: 04/04/2022 14:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NETTLES,JONATHAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NETTLES,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1175968
DRY GAS EXPIRY DATE: 07/30/2022
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 03/16/2022 01:49:01

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.081	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.000	01:43
Breath Volume		2.382	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.000	01:47
Breath Volume		3.277	01:47
Air Blank		0.000	01:48
Control Test		0.082	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:49

Data Download Date/Time: 04/04/2022 14:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEBO,ANDRES,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PERO,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 03/26/2022 21:40:36

Diagnostics Check		OK	21:33
Air Blank		0.000	21:34
Control Test		0.080	21:34
Air Blank		0.000	21:35
Subject Sample	# 1	0.198	21:36
Breath Volume		2.078	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.189	21:39
Breath Volume		1.546	21:39
Air Blank		0.000	21:39
Control Test		0.080	21:40
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 04/01/2022 18:11

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BAXMAN

DATE/TIME OF TEST: 03/05/2022 02:19:06

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.010	02:14
Breath Volume		2.546	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.007	02:17
Breath Volume		2.363	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 03/10/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HARRISON,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/06/2022 19:29:07

Diagnostics Check		OK	19:22
Air Blank		0.000	19:22
Control Test		0.079	19:23
Air Blank		0.000	19:23
Subject Sample	# 1	0.245	19:24
Breath Volume		2.011	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.237	19:27
Breath Volume		2.062	19:27
Air Blank		0.000	19:28
Control Test		0.078	19:28
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 03/10/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RIX,DAVID
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/07/2022 21:26:17

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.080	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:24	Subject Test Refused
Breath Volume		0.000	21:24	
Air Blank		0.000	21:25	
Control Test		0.080	21:25	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 03/10/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANUS,GLEN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HELTON,PATRICK
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/08/2022 03:20:01

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.082	03:15
Breath Volume		2.566	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:17
Subject Sample	# 2	0.082	03:18
Breath Volume		2.980	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:20

Data Download Date/Time: 03/10/2022 19:00

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JAKE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REEVES,DAVID
ARRESTING OFFICER AGENCY: Trenton PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: RHODEN

DATE/TIME OF TEST: 03/13/2022 20:25:48

Diagnostics Check	OK	20:25	
Air Blank		20:25	RFI Detect
Air Blank	0.000	20:25	

Data Download Date/Time: 05/02/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,JAKE,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: REEVES,DAVID
ARRESTING OFFICER AGENCY: Trenton PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: RHODEN

DATE/TIME OF TEST: 03/13/2022 20:36:24

Diagnostics Check		OK	20:27	
Air Blank		0.000	20:28	
Control Test		0.080	20:28	
Air Blank		0.000	20:28	
Subject Sample	# 1		20:32	No Sample Provided
Breath Volume		0.000	20:32	
Air Blank		0.000	20:32	
Air Blank		0.000	20:34	
Subject Sample	# 2		20:35	Subject Test Refused
Breath Volume		0.000	20:35	
Air Blank		0.000	20:35	
Control Test		0.080	20:35	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:36	

Data Download Date/Time: 05/02/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:58

SUBJECT'S LAST NAME: BIEZE

DATE/TIME OF TEST: 03/20/2022 00:02:42

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:59	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Control Test		0.080	00:02	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 05/02/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MACINTYRE,DEAN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CUMMING

DATE/TIME OF TEST: 03/20/2022 02:10:04

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:59	
Control Test		0.079	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:03	No Sample Provided
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2		02:08	No Sample Provided
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 05/02/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DAVIS,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TUCCIO

DATE/TIME OF TEST: 03/20/2022 19:58:36

Diagnostics Check		OK	19:48
Air Blank		0.000	19:49
Control Test		0.080	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.291	19:50
Breath Volume		3.355	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.253	19:53
Breath Volume		1.554	19:53
Air Blank		0.000	19:54
Air Blank		0.000	19:56
Subject Sample	# 3	0.290	19:56
Breath Volume		2.585	19:56
Air Blank		0.000	19:57
Control Test		0.077	19:57
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 05/02/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KEVIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MEYER,SAMUEL
ARRESTING OFFICER AGENCY: Trenton PD
LAST AGENCY INSPECTION: 03/04/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG106803
DRY GAS EXPIRY DATE: 03/09/2023
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/28/2022 23:39:53

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.080	23:34
Air Blank		0.000	23:34
Subject Sample	# 1	0.249	23:35
Breath Volume		3.246	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.252	23:38
Breath Volume		3.296	23:38
Air Blank		0.000	23:38
Control Test		0.078	23:39
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 05/02/2022 11:37

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MCKINNON

DATE/TIME OF TEST: 03/12/2022 04:04:56

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.109	04:00
Breath Volume		1.593	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.110	04:03
Breath Volume		1.687	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAZIME,S
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: TARTAGLIA

DATE/TIME OF TEST: 03/12/2022 21:44:55

Diagnostics Check		OK	21:42	
Air Blank		0.000	21:42	
Control Test		0.081	21:42	
Air Blank		0.000	21:43	
Subject Sample	# 1		21:43	Subject Test Refused
Breath Volume		0.000	21:43	
Air Blank		0.000	21:43	
Control Test		0.080	21:44	
Air Blank		0.000	21:44	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BOHRER

DATE/TIME OF TEST: 03/12/2022 23:35:02

Diagnostics Check		OK	23:28
Air Blank		0.000	23:29
Control Test		0.080	23:29
Air Blank		0.000	23:30
Subject Sample	# 1	0.098	23:30
Breath Volume		1.328	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.093	23:33
Breath Volume		1.593	23:33
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:35

Data Download Date/Time: 03/23/2022 16:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MARINAN

DATE/TIME OF TEST: 03/13/2022 01:15:28

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.081	01:13	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.081	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 03/23/2022 16:45

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: AHERN

DATE/TIME OF TEST: 03/13/2022 01:53:22

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:50	
Control Test		0.082	01:51	
Air Blank		0.000	01:51	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Control Test		0.081	01:52	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 03/23/2022 16:45

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG021605
DRY GAS EXPIRY DATE: 08/03/2022
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/13/2022 03:37:57

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:35	
Control Test		0.081	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 03/23/2022 16:45

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRANSUE,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: PRINCIPE

DATE/TIME OF TEST: 03/01/2022 16:59:54

Diagnostics Check		OK	16:52
Air Blank		0.000	16:52
Control Test		0.082	16:53
Air Blank		0.000	16:53
Subject Sample	# 1	0.232	16:54
Breath Volume		1.210	16:54
Air Blank		0.000	16:54
Air Blank		0.000	16:56
Subject Sample	# 2	0.234	16:58
Breath Volume		2.105	16:58
Air Blank		0.000	16:58
Control Test		0.083	16:59
Air Blank		0.000	16:59
Diagnostics Check		OK	16:59

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/02/2022 02:28:48

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.082	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.132	02:24
Breath Volume		1.589	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.120	02:27
Breath Volume		1.625	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KISSANE,BRIAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: LOFLIN

DATE/TIME OF TEST: 03/02/2022 20:48:56

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.082	20:43
Air Blank		0.000	20:43
Subject Sample	# 1	0.000	20:44
Breath Volume		2.320	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.000	20:47
Breath Volume		1.464	20:47
Air Blank		0.000	20:48
Control Test		0.082	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOUGHERTY,JOSEPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WEECH

DATE/TIME OF TEST: 03/03/2022 03:03:13

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.082	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.150	02:58
Breath Volume		3.109	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.151	03:01
Breath Volume		2.703	03:01
Air Blank		0.000	03:02
Control Test		0.082	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATKINS,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LEMUS

DATE/TIME OF TEST: 03/04/2022 01:34:28

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.082	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.153	01:30
Breath Volume		2.246	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.152	01:32
Breath Volume		2.371	01:32
Air Blank		0.000	01:33
Control Test		0.082	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ARROYO,ARIANA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 03/04/2022 02:33:15

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.082	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.152	02:27
Breath Volume		1.250	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.149	02:31
Breath Volume		1.191	02:31
Air Blank		0.000	02:32
Control Test		0.082	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: FRY

DATE/TIME OF TEST: 03/04/2022 12:29:40

Diagnostics Check		OK	12:20
Air Blank		0.000	12:20
Control Test		0.082	12:21
Air Blank		0.000	12:21
Subject Sample	# 1	0.175	12:22
Breath Volume		1.316	12:22
Air Blank		0.000	12:22
Air Blank		0.000	12:24
Subject Sample	# 2	0.198	12:25
Breath Volume		1.871	12:25
Air Blank		0.000	12:25
Air Blank		0.000	12:27
Subject Sample	# 3	0.195	12:28
Breath Volume		1.613	12:28
Air Blank		0.000	12:28
Control Test		0.081	12:29
Air Blank		0.000	12:29
Diagnostics Check		OK	12:29

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/05/2022 05:44:03

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:30	
Control Test		0.082	05:30	
Air Blank		0.000	05:31	
Subject Sample	# 1	0.197	05:31	No .020 agreement
Breath Volume		1.343	05:31	
Air Blank		0.000	05:32	
Air Blank		0.000	05:34	
Subject Sample	# 2	0.179	05:37	Volume Not Met
Breath Volume		0.363	05:37	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 3	0.172	05:42	No .020 agreement
Breath Volume		1.335	05:42	
Air Blank		0.000	05:43	
Control Test		0.081	05:43	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE,ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/05/2022 05:52:37

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.080	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.182	05:48
Breath Volume		1.550	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.173	05:51
Breath Volume		1.457	05:51
Air Blank		0.000	05:51
Control Test		0.081	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIBEIRO,NICOLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: JORDAHL

DATE/TIME OF TEST: 03/05/2022 22:19:41

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.082	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.000	22:14
Breath Volume		1.148	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.000	22:18
Breath Volume		1.164	22:18
Air Blank		0.000	22:18
Control Test		0.082	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 03/10/2022 12:49

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: KRZEWINSKI

DATE/TIME OF TEST: 03/06/2022 01:07:25

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.082	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.175	01:03
Breath Volume		1.703	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.172	01:05
Breath Volume		1.906	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 03/10/2022 12:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: URENA

DATE/TIME OF TEST: 03/06/2022 11:08:35

Diagnostics Check		OK	11:02
Air Blank		0.000	11:02
Control Test		0.082	11:03
Air Blank		0.000	11:03
Subject Sample	# 1	0.190	11:04
Breath Volume		2.203	11:04
Air Blank		0.000	11:04
Air Blank		0.000	11:06
Subject Sample	# 2	0.192	11:07
Breath Volume		3.242	11:07
Air Blank		0.000	11:07
Control Test		0.082	11:07
Air Blank		0.000	11:08
Diagnostics Check		OK	11:08

Data Download Date/Time: 03/10/2022 12:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRICKEY,ALEX
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 03/07/2022 01:45:54

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.082	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:38	No Sample Provided
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Air Blank		0.000	01:41	
Subject Sample	# 2		01:44	No Sample Provided
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.081	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 03/10/2022 12:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 03/08/2022 04:02:52

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.083	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.111	03:58
Breath Volume		2.140	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.106	04:01
Breath Volume		1.144	04:01
Air Blank		0.000	04:01
Control Test		0.082	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 03/10/2022 12:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,JEFFRIE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ANGUEIRA

DATE/TIME OF TEST: 03/09/2022 05:04:13

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:53	
Control Test		0.083	04:53	
Air Blank		0.000	04:53	
Subject Sample	# 1	0.071	04:57	Volume Not Met
Breath Volume		0.191	04:57	
Air Blank		0.000	04:57	
Air Blank		0.000	04:59	
Subject Sample	# 2	0.078	05:02	Volume Not Met
Breath Volume		0.359	05:02	
Air Blank		0.000	05:03	
Control Test		0.081	05:03	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 03/10/2022 12:50

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS,ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07220080A2
DRY GAS EXPIRY DATE: 05/05/2022
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GARCIA PAZ

DATE/TIME OF TEST: 03/11/2022 23:39:30

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.082	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.188	23:35
Breath Volume		2.859	23:35
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.174	23:37
Breath Volume		2.242	23:37
Air Blank		0.000	23:38
Control Test		0.082	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 04/07/2022 11:28

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYANT,BRANDEN,J
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: HANNON,R
ARRESTING OFFICER AGENCY: Interlachen PD
LAST AGENCY INSPECTION: 02/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 03/07/2022 22:57:05

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.078	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.187	22:52
Breath Volume		2.386	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:54
Subject Sample	# 2	0.184	22:55
Breath Volume		1.250	22:55
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 03/28/2022 12:54

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/22/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: OXENDINE

DATE/TIME OF TEST: 03/14/2022 18:49:47

Diagnostics Check		OK	18:41
Air Blank		0.000	18:41
Control Test		0.079	18:41
Air Blank		0.000	18:42
Subject Sample	# 1	0.043	18:42
Breath Volume		1.339	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.043	18:48
Breath Volume		1.406	18:48
Air Blank		0.000	18:48
Control Test		0.079	18:49
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 03/28/2022 12:54

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MICK,TYLER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MERCADO-TORRES

DATE/TIME OF TEST: 03/01/2022 22:05:58

Diagnostics Check		OK	21:59	
Air Blank		0.000	22:00	
Control Test		0.082	22:00	
Air Blank		0.000	22:01	
Subject Sample	# 1	0.238	22:04	Volume Not Met
Breath Volume		0.917	22:04	
Air Blank			22:05	Purge Fail
Air Blank		0.000	22:05	

Data Download Date/Time: 03/29/2022 03:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MICK,WILLIAM
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MERCADO-TORRES

DATE/TIME OF TEST: 03/01/2022 22:17:24

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.077	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1	0.225	22:15	Volume Not Met
Breath Volume		0.867	22:15	
Air Blank			22:16	Purge Fail
Air Blank		0.000	22:17	

Data Download Date/Time: 03/29/2022 03:56

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MICK,WILLIAM
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MERCADO-TORRES

DATE/TIME OF TEST: 03/01/2022 22:25:13

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.072	22:24	Control Outside Tolerance
Air Blank	0.000	22:25	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MICK,WILLIAM
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24620080A2
DRY GAS EXPIRY DATE: 11/05/2022
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MERCADO-TORRES

DATE/TIME OF TEST: 03/01/2022 22:36:37

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:30	
Control Test		0.078	22:31	
Air Blank		0.000	22:31	
Subject Sample	# 1	0.198	22:34	Volume Not Met
Breath Volume		0.292	22:34	
Air Blank			22:35	Purge Fail
Air Blank		0.000	22:36	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: BURKETT,RICHARD
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 03/04/2022 13:57:57

Diagnostics Check		OK	13:51
Air Blank		0.000	13:51
Control Test		0.082	13:51
Air Blank		0.000	13:52
Subject Sample	# 1	0.000	13:53
Breath Volume		3.222	13:53
Air Blank		0.000	13:53
Air Blank		0.000	13:55
Subject Sample	# 2	0.000	13:56
Breath Volume		1.753	13:56
Air Blank		0.000	13:57
Control Test		0.079	13:57
Air Blank		0.000	13:57
Diagnostics Check		OK	13:57

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: KRAEUTER,DARIAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: GOLLETTE

DATE/TIME OF TEST: 03/08/2022 17:56:11

Diagnostics Check		OK	17:47
Air Blank		0.000	17:48
Control Test		0.081	17:48
Air Blank		0.000	17:49
Subject Sample	# 1	0.124	17:51
Breath Volume		1.804	17:51
Air Blank		0.000	17:52
Air Blank		0.000	17:53
Subject Sample	# 2	0.114	17:54
Breath Volume		3.183	17:54
Air Blank		0.000	17:55
Control Test		0.078	17:55
Air Blank		0.000	17:56
Diagnostics Check		OK	17:56

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: KRAEUTER,DARIAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: GOLLETTE

DATE/TIME OF TEST: 03/08/2022 18:12:15

Diagnostics Check		OK	18:08	
Air Blank		0.000	18:08	
Control Test		0.079	18:09	
Air Blank		0.000	18:09	
Subject Sample	# 1	0.124	18:11	
Breath Volume		1.636	18:11	
Air Blank			18:11	Purge Fail
Air Blank		0.000	18:12	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: KRAEUTER,DARIAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: GOLLETTE

DATE/TIME OF TEST: 03/08/2022 18:21:17

Diagnostics Check		OK	18:14
Air Blank		0.000	18:14
Control Test		0.076	18:15
Air Blank		0.000	18:15
Subject Sample	# 1	0.111	18:16
Breath Volume		2.023	18:16
Air Blank		0.000	18:17
Air Blank		0.000	18:18
Subject Sample	# 2	0.119	18:19
Breath Volume		1.105	18:19
Air Blank		0.000	18:20
Control Test		0.077	18:20
Air Blank		0.000	18:21
Diagnostics Check		OK	18:21

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MICK,WILLIAM
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 03/19/2022 02:22:28

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.082	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.066	02:20	
Air Blank			02:21	Purge Fail
Air Blank		0.000	02:22	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MICK,WILLIAM
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 03/19/2022 02:26:21

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.072	02:25	Control Outside Tolerance
Air Blank	0.000	02:26	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MICK,WILLIAM
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 03/19/2022 02:37:53

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.080	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.188	02:36	
Breath Volume		1.300	02:36	
Air Blank			02:37	Purge Fail
Air Blank		0.000	02:37	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LOGAN,TYRONE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 03/22/2022 00:56:46

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.082	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1	0.197	00:43	
Breath Volume		1.562	00:43	
Air Blank		0.000	00:44	
Air Blank		0.000	00:46	
Subject Sample	# 2	0.158	00:49	Volume Not Met
Breath Volume		0.109	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 3		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:55	
Control Test		0.077	00:56	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JACOB,T
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: GREEN,KATELYN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 03/22/2022 01:26:47

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.079	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1	0.173	01:21	
Breath Volume		2.433	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 2	0.164	01:25	
Breath Volume		2.394	01:25	
Air Blank			01:26	Purge Fail
Air Blank		0.000	01:26	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: JONES,DAKOTA
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 02/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BUONO

DATE/TIME OF TEST: 03/25/2022 20:18:39

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:12	
Control Test		0.083	20:13	
Air Blank		0.000	20:13	
Subject Sample	# 1	0.190	20:16	Volume Not Met
Breath Volume		0.628	20:16	
Air Blank			20:17	Purge Fail
Air Blank		0.000	20:18	

Data Download Date/Time: 03/29/2022 03:57

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,URIEL,
BREATH TEST OPERATOR AGENCY: COLLIER COUNTY SO
ARRESTING OFFICER: TARAZONA,B
ARRESTING OFFICER AGENCY: COLLIER COUNTY SO
LAST AGENCY INSPECTION: 02/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11821080A2
DRY GAS EXPIRY DATE: 06/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ORDONEZ ORDONEZ

DATE/TIME OF TEST: 03/04/2022 03:51:56

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.207	03:47
Breath Volume		2.097	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.212	03:50
Breath Volume		2.687	03:50
Air Blank		0.000	03:50
Control Test		0.078	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 03/16/2022 13:44

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,HERVEY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROMERO,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: PINEDA CASTRO

DATE/TIME OF TEST: 03/03/2022 05:04:47

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:02	
Control Test		0.081	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 1		05:03	Subject Test Refused
Breath Volume		0.000	05:03	
Air Blank		0.000	05:03	
Control Test		0.081	05:04	
Air Blank		0.000	05:04	
Diagnostics Check		OK	05:04	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,HERVEY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROMERO,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: PINEDA CASTRO

DATE/TIME OF TEST: 03/03/2022 05:17:34

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:14	
Control Test		0.080	05:15	
Air Blank		0.000	05:15	
Subject Sample	# 1		05:16	Subject Test Refused
Breath Volume		0.000	05:16	
Air Blank		0.000	05:16	
Control Test		0.080	05:16	
Air Blank		0.000	05:17	
Diagnostics Check		OK	05:17	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/05/2022 01:19:47

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.082	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.283	01:13
Breath Volume		1.945	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.270	01:17
Breath Volume		1.394	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,AVERY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WHITE,AVERY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: BLANCO ESTRADA

DATE/TIME OF TEST: 03/06/2022 09:27:13

Diagnostics Check		OK	09:24	
Air Blank		0.000	09:24	
Control Test		0.082	09:25	
Air Blank		0.000	09:25	
Subject Sample	# 1		09:25	Subject Test Refused
Breath Volume		0.000	09:25	
Air Blank		0.000	09:26	
Control Test		0.081	09:26	
Air Blank		0.000	09:27	
Diagnostics Check		OK	09:27	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAMIAN,IULIA,V
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PACHECO,FRED
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: YANES

DATE/TIME OF TEST: 03/07/2022 05:13:59

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:01	
Control Test		0.081	05:02	
Air Blank		0.000	05:02	
Subject Sample	# 1	0.222	05:05	Volume Not Met
Breath Volume		0.371	05:05	
Air Blank		0.000	05:06	
Air Blank		0.000	05:08	
Subject Sample	# 2	0.217	05:09	
Breath Volume		1.261	05:09	
Air Blank		0.000	05:10	
Air Blank		0.000	05:11	
Subject Sample	# 3	0.232	05:12	
Breath Volume		1.519	05:12	
Air Blank		0.000	05:13	
Control Test		0.078	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINO,BRYAN,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANTANA,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 03/09/2022 23:08:26

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.083	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.246	23:03
Breath Volume		3.527	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.247	23:06
Breath Volume		3.070	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MARRERO,MARCOS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 09:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: RIVERO PACKETT

DATE/TIME OF TEST: 03/13/2022 10:11:41

Diagnostics Check		OK	10:05
Air Blank		0.000	10:05
Control Test		0.082	10:05
Air Blank		0.000	10:06
Subject Sample	# 1	0.216	10:07
Breath Volume		2.507	10:07
Air Blank		0.000	10:07
Air Blank		0.000	10:09
Subject Sample	# 2	0.215	10:10
Breath Volume		1.792	10:10
Air Blank		0.000	10:10
Control Test		0.080	10:11
Air Blank		0.000	10:11
Diagnostics Check		OK	10:11

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,DANELIA,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,FERNANDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SARABIA

DATE/TIME OF TEST: 03/13/2022 19:26:49

Diagnostics Check		OK	19:23	
Air Blank		0.000	19:24	
Control Test		0.081	19:24	
Air Blank		0.000	19:25	
Subject Sample	# 1		19:25	Subject Test Refused
Breath Volume		0.000	19:25	
Air Blank		0.000	19:25	
Control Test		0.081	19:26	
Air Blank		0.000	19:26	
Diagnostics Check		OK	19:26	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNADEZ,YUNIER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SARABIA

DATE/TIME OF TEST: 03/13/2022 20:25:38

Diagnostics Check		OK	20:17
Air Blank		0.000	20:18
Control Test		0.081	20:18
Air Blank		0.000	20:19
Subject Sample	# 1	0.125	20:20
Breath Volume		2.187	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:22
Subject Sample	# 2	0.125	20:24
Breath Volume		2.285	20:24
Air Blank		0.000	20:24
Control Test		0.081	20:25
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILES,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SILES,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TARRERO

DATE/TIME OF TEST: 03/17/2022 23:09:23

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.190	23:04
Breath Volume		1.566	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.204	23:07
Breath Volume		1.261	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUMBI,SELVIN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,CHIRLEDY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: VALLEJO

DATE/TIME OF TEST: 03/19/2022 02:00:50

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:58	
Control Test		0.082	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Control Test		0.081	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,JAMES,T
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARDNER,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 09:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: LINARES-OJEDA

DATE/TIME OF TEST: 03/19/2022 11:24:32

Diagnostics Check		OK	11:18	
Air Blank		0.000	11:18	
Control Test		0.082	11:19	
Air Blank		0.000	11:19	
Subject Sample	# 1	0.094	11:20	
Breath Volume		1.785	11:20	
Air Blank		0.000	11:21	
Air Blank		0.000	11:22	
Subject Sample	# 2	0.091	11:23	
Breath Volume		1.593	11:23	
Air Blank		0.000	11:23	
			11:24	RFI Detect
Air Blank		0.000	11:24	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,JAMES,T
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARDNER,STEPHANIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 09:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: LINARES-OJEDA

DATE/TIME OF TEST: 03/19/2022 11:45:48

Diagnostics Check		OK	11:38
Air Blank		0.000	11:38
Control Test		0.081	11:39
Air Blank		0.000	11:39
Subject Sample	# 1	0.081	11:40
Breath Volume		1.113	11:40
Air Blank		0.000	11:41
Air Blank		0.000	11:43
Subject Sample	# 2	0.078	11:44
Breath Volume		1.101	11:44
Air Blank		0.000	11:44
Control Test		0.080	11:45
Air Blank		0.000	11:45
Diagnostics Check		OK	11:45

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERCADO,KASSANDRA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WATSON,TYRONE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 03/20/2022 01:02:53

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 04/01/2022 12:03

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARAUZ,G
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 08:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/20/2022 09:16:33

Diagnostics Check		OK	09:13	
Air Blank		0.000	09:14	
Control Test		0.081	09:14	
Air Blank		0.000	09:15	
Subject Sample	# 1		09:15	Subject Test Refused
Breath Volume		0.000	09:15	
Air Blank		0.000	09:15	
Control Test		0.081	09:15	
Air Blank		0.000	09:16	
Diagnostics Check		OK	09:16	

Data Download Date/Time: 04/01/2022 12:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ALLENDE REYES

DATE/TIME OF TEST: 03/21/2022 02:24:06

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:21	
Control Test		0.082	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:23	
Control Test		0.080	02:23	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 04/01/2022 12:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ALLENDE REYES

DATE/TIME OF TEST: 03/21/2022 02:37:00

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.080	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 04/01/2022 12:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CAMILA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MORENO,PEDRO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: ZALDIVAR

DATE/TIME OF TEST: 03/21/2022 05:33:42

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:29	
Control Test		0.081	05:30	
Air Blank		0.000	05:30	
Subject Sample	# 1		05:32	Subject Test Refused
Breath Volume		0.000	05:32	
Air Blank		0.000	05:32	
Control Test		0.081	05:33	
Air Blank		0.000	05:33	
Diagnostics Check		OK	05:33	

Data Download Date/Time: 04/01/2022 12:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPSHUCK,CHRISTINA,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MONTALVAN,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/22/2022 00:42:10

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:36	
Control Test		0.081	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1	0.170	00:38	
Breath Volume		1.695	00:38	
Air Blank		0.000	00:39	
Air Blank			00:40	RFI Detect
Air Blank		0.000	00:41	
Control Test		0.081	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 04/01/2022 12:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPSHUCK,CHRISTINA,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MONTALVAN,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/22/2022 00:53:51

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.160	00:49
Breath Volume		2.570	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:51
Subject Sample	# 2	0.155	00:52
Breath Volume		1.523	00:52
Air Blank		0.000	00:52
Control Test		0.079	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 04/01/2022 12:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,AVERY,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MARTINEZ,JOSANDRE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: QUINTANA GOMEZ

DATE/TIME OF TEST: 03/27/2022 07:28:37

Diagnostics Check		OK	07:25	
Air Blank		0.000	07:26	
Control Test		0.082	07:26	
Air Blank		0.000	07:27	
Subject Sample	# 1		07:27	Subject Test Refused
Breath Volume		0.000	07:27	
Air Blank		0.000	07:27	
Control Test		0.081	07:28	
Air Blank		0.000	07:28	
Diagnostics Check		OK	07:28	

Data Download Date/Time: 04/01/2022 12:04

Printed: 12 May 2022

COBRA ©1997-2014 CMI, Inc.