

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 03/06/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A3
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 04/23/2022 05:42:37

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:39	
Control Test	0.081	05:39	
Air Blank	0.000	05:40	
Subject Sample	# 1 0.257	05:41	
Breath Volume	4.121	05:41	
Air Blank		05:41	Purge Fail
Air Blank		05:42	Purge Fail

Data Download Date/Time: 04/25/2022 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: IRIZARRY,H
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/25/2023
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GUADARRAMA

DATE/TIME OF TEST: 04/01/2022 03:05:57

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.078	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:05	Sequence Aborted
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	

Data Download Date/Time: 04/21/2022 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BENNETT,ROBERT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/28/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/25/2023
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 04/17/2022 22:51:37

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.078	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.253	22:46
Breath Volume	3.089	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.234	22:49
Breath Volume	3.167	22:49
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 04/21/2022 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEBRETON,Q
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/24/2022 04:32:40

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.178	04:26
Breath Volume	1.355	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.170	04:30
Breath Volume	1.796	04:30
Air Blank	0.000	04:31
Control Test	0.077	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 05/22/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LEWIS,BREEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GRIZZLE

DATE/TIME OF TEST: 04/26/2022 03:47:35

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.225	03:41
Breath Volume	1.367	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.226	03:45
Breath Volume	2.992	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 05/22/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/28/2022 04:07:37

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Subject Sample # 1	0.127	04:01	
Breath Volume	1.332	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample # 2	0.119	04:05	
Breath Volume	1.343	04:05	
Air Blank	0.000	04:06	
Control Test	0.074	04:06	Control Outside Tolerance
Air Blank	0.000	04:07	

Data Download Date/Time: 05/22/2022 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/28/2022 04:17:43

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.078	04:10	
Air Blank	0.000	04:10	
Subject Sample # 1	0.127	04:12	
Breath Volume	2.394	04:12	
Air Blank	0.000	04:13	
Air Blank	0.000	04:14	
Subject Sample # 2	0.117	04:15	
Breath Volume	1.582	04:15	
Air Blank	0.000	04:16	
Control Test	0.074	04:17	Control Outside Tolerance
Air Blank	0.000	04:17	

Data Download Date/Time: 05/22/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/28/2022 04:29:43

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:23	
Subject Sample # 1	0.128	04:24	
Breath Volume	2.472	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample # 2	0.118	04:28	
Breath Volume	1.210	04:28	
Air Blank	0.000	04:28	
Control Test	0.074	04:29	Control Outside Tolerance
Air Blank	0.000	04:29	

Data Download Date/Time: 05/22/2022 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTERIO,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: RADCLIFF,AARON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/21/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: CHATELAIN

DATE/TIME OF TEST: 04/30/2022 03:56:47

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.079	03:54	
Air Blank	0.000	03:54	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.079	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 05/22/2022 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLVERA,JESUS,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: AMBROSIO

DATE/TIME OF TEST: 04/30/2022 22:23:38

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.082	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.082	22:22	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 05/31/2022 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/05/2022 00:20:11

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.080	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	RFI Detect
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.081	00:19	RFI Detect
Air Blank	0.000	00:20	

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/05/2022 00:22:41

Diagnostics Check	OK	00:22	
Air Blank		00:22	RFI Detect
Air Blank	0.000	00:22	

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/05/2022 00:30:39

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.075	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.145	00:26
Breath Volume	4.683	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.136	00:28
Breath Volume	4.511	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: OSTERTAG,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: DEMYANOVICH

DATE/TIME OF TEST: 04/08/2022 21:00:40

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.082	21:00	Interferent Detect
Air Blank	0.000	21:00	

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: OSTERTAG,SHELBY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: DEMYANOVICH

DATE/TIME OF TEST: 04/08/2022 21:09:02

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.000	21:04
Breath Volume	1.851	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.000	21:07
Breath Volume	1.691	21:07
Air Blank	0.000	21:08
Control Test	0.082	21:08
Air Blank	0.000	21:08
Diagnostics Check	OK	21:09

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: AVILLANEDA

DATE/TIME OF TEST: 04/09/2022 02:55:12

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.081	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1 0.142	02:50	
Breath Volume	1.972	02:50	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2 0.141	02:53	
Breath Volume	1.875	02:53	
Air Blank	0.000	02:54	
Control Test	0.084	02:54	Interferent Detect
Air Blank	0.000	02:55	

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: AVILLANEDA

DATE/TIME OF TEST: 04/09/2022 03:04:23

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.146	02:59
Breath Volume	2.140	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.147	03:02
Breath Volume	2.539	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BOURIS

DATE/TIME OF TEST: 04/10/2022 00:46:08

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.081	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1 0.131	00:40	
Breath Volume	2.500	00:40	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample	# 2 0.127	00:43	
Breath Volume	1.480	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	Fail	00:45	Analytical Stability Test
Air Blank	0.000	00:46	

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,JACOB
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/23/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BOURIS

DATE/TIME OF TEST: 04/10/2022 00:56:19

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.082	00:48	
Air Blank	0.000	00:49	
Subject Sample # 1	0.132	00:50	
Breath Volume	1.828	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:52	
Subject Sample # 2	0.126	00:54	
Breath Volume	1.894	00:54	
Air Blank	0.000	00:54	
Control Test	0.082	00:55	
Air Blank	0.000	00:55	
Diagnostics Check	Fail	00:55	Analytical Stability Test
Air Blank	0.000	00:56	

Data Download Date/Time: 04/11/2022 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: BRUENNING

DATE/TIME OF TEST: 04/14/2022 16:50:32

Diagnostics Check	OK	16:44
Air Blank	0.000	16:44
Control Test	0.080	16:45
Air Blank	0.000	16:45
Subject Sample # 1	0.000	16:46
Breath Volume	3.628	16:46
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 2	0.000	16:49
Breath Volume	3.769	16:49
Air Blank	0.000	16:49
Control Test	0.080	16:49
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 05/09/2022 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERREO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CRILLEY

DATE/TIME OF TEST: 04/17/2022 01:05:15

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1 0.173	01:00	Volume Not Met
Breath Volume	0.757	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:03	Subject Test Refused
Breath Volume	0.058	01:03	
Air Blank	0.000	01:04	
Control Test	0.074	01:04	Control Outside Tolerance
Air Blank	0.000	01:05	

Data Download Date/Time: 05/09/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LAUGHLIN

DATE/TIME OF TEST: 04/20/2022 02:02:13

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.000	01:57
Breath Volume	1.460	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.000	02:00
Breath Volume	1.273	02:00
Air Blank	0.000	02:01
Control Test	0.076	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 05/09/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: VILLAVICENCIO MESA

DATE/TIME OF TEST: 04/23/2022 00:11:17

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.068	00:06
Breath Volume	2.246	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.069	00:09
Breath Volume	2.312	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 05/09/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,RICHARD,F
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SNYDER,RICHARD
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/11/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: RENCE

DATE/TIME OF TEST: 04/29/2022 19:56:15

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.125	19:51
Breath Volume	1.113	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.125	19:54
Breath Volume	1.714	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 05/09/2022 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILLIAMS,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/24/2022 17:00:50

Diagnostics Check	OK	16:49	
Air Blank	0.000	16:49	
Control Test	0.079	16:49	
Air Blank	0.000	16:50	
Subject Sample	# 1 0.211	16:53	Volume Not Met
Breath Volume	0.082	16:53	
Air Blank	0.000	16:54	
Air Blank	0.000	16:55	
Subject Sample	# 2 0.237	16:59	Volume Not Met
Breath Volume	0.246	16:59	
Air Blank	0.000	16:59	
Control Test	0.077	17:00	
Air Blank	0.000	17:00	
Diagnostics Check	OK	17:00	

Data Download Date/Time: 04/26/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILLIAMS,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 04/24/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22620080A2
DRY GAS EXPIRY DATE: 10/05/2022
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/24/2022 17:12:21

Diagnostics Check	OK	17:05
Air Blank	0.000	17:05
Control Test	0.077	17:05
Air Blank	0.000	17:06
Subject Sample # 1	0.255	17:07
Breath Volume	1.234	17:07
Air Blank	0.000	17:08
Air Blank	0.000	17:09
Subject Sample # 2	0.255	17:10
Breath Volume	1.492	17:10
Air Blank	0.000	17:11
Control Test	0.078	17:11
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 04/26/2022 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 04/23/2022 03:07:55

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.245	03:03
Breath Volume	1.179	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.242	03:06
Breath Volume	2.453	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEDRO JUAN

DATE/TIME OF TEST: 04/23/2022 04:24:12

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.175	04:19
Breath Volume	1.253	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.160	04:22
Breath Volume	1.312	04:22
Air Blank	0.000	04:23
Control Test	0.078	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SOARES DE MORAES

DATE/TIME OF TEST: 04/24/2022 03:10:59

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.206	03:06
Breath Volume	1.414	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.209	03:09
Breath Volume	2.417	03:09
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER, CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MLUMBI

DATE/TIME OF TEST: 04/26/2022 02:01:23

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1 0.129	01:53	Volume Not Met
Breath Volume	0.050	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2 0.150	01:56	
Breath Volume	1.769	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 3 0.155	01:59	
Breath Volume	1.679	01:59	
Air Blank	0.000	02:00	
Control Test	0.078	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/26/2022 23:23:00

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.182	23:18
Breath Volume	2.835	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.183	23:21
Breath Volume	3.031	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:23

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SPANGLER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: LENTINE

DATE/TIME OF TEST: 04/28/2022 00:39:21

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.212	00:35
Breath Volume	2.402	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.195	00:37
Breath Volume	1.304	00:37
Air Blank	0.000	00:38
Control Test	0.079	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LUBINSKI,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: STEVENSON JR

DATE/TIME OF TEST: 04/28/2022 08:55:16

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.079	08:50
Air Blank	0.000	08:50
Subject Sample # 1	0.000	08:51
Breath Volume	3.261	08:51
Air Blank	0.000	08:51
Air Blank	0.000	08:53
Subject Sample # 2	0.000	08:53
Breath Volume	2.621	08:53
Air Blank	0.000	08:54
Control Test	0.080	08:54
Air Blank	0.000	08:55
Diagnostics Check	OK	08:55

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: DOKA

DATE/TIME OF TEST: 04/29/2022 00:43:54

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.196	00:39
Breath Volume	2.027	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.199	00:42
Breath Volume	2.550	00:42
Air Blank	0.000	00:43
Control Test	0.077	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BORJA

DATE/TIME OF TEST: 04/29/2022 04:29:28

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.077	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.194	04:21
Breath Volume	1.789	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.171	04:24
Breath Volume	1.781	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 3	0.189	04:27
Breath Volume	2.734	04:27
Air Blank	0.000	04:28
Control Test	0.077	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VILLARIN, JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 04/30/2022 04:03:15

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.214	03:58
Breath Volume	1.535	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.218	04:01
Breath Volume	1.675	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: OWERS, PETER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 04/30/2022 05:10:51

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.140	05:06
Breath Volume	1.843	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.132	05:09
Breath Volume	1.113	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 05/13/2022 13:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRUDEAU,RYAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 04/01/2022 22:36:53

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:24	
Control Test	0.080	22:25	
Air Blank	0.000	22:25	
Subject Sample	# 1 0.174	22:28	Slope Not Level
Breath Volume	1.472	22:28	
Air Blank	0.000	22:29	
Air Blank	0.000	22:31	
Subject Sample	# 2 0.183	22:32	
Breath Volume	1.515	22:32	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 3 0.183	22:35	
Breath Volume	1.300	22:35	
Air Blank	0.000	22:35	
Control Test	0.078	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 04/08/2022 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: DOMINGO ORTIZ

DATE/TIME OF TEST: 04/01/2022 23:12:22

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.079	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:11	
Control Test	0.079	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 04/08/2022 14:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PEEZ VELASQUEZ

DATE/TIME OF TEST: 04/02/2022 00:49:17

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.250	00:44
Breath Volume	1.589	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.248	00:47
Breath Volume	1.343	00:47
Air Blank	0.000	00:48
Control Test	0.077	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 04/08/2022 14:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PADGETT,COVELLE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: WINGFIELD

DATE/TIME OF TEST: 04/05/2022 00:16:02

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.000	00:11
Breath Volume	1.753	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.000	00:14
Breath Volume	1.761	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:16

Data Download Date/Time: 04/08/2022 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: CARVALHO

DATE/TIME OF TEST: 04/05/2022 20:55:07

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.080	20:40	
Air Blank	0.000	20:40	
Subject Sample	# 1 0.154	20:43	Volume Not Met
Breath Volume	0.523	20:43	
Air Blank	0.000	20:44	
Air Blank	0.000	20:46	
Subject Sample	# 2 0.168	20:48	
Breath Volume	2.171	20:48	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 3 0.153	20:53	
Breath Volume	1.207	20:53	
Air Blank	0.000	20:54	
Control Test	0.078	20:54	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 04/08/2022 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOPPING,CHRISTOPHER
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 04/06/2022 04:26:08

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.152	04:21
Breath Volume	1.296	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.151	04:24
Breath Volume	2.812	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/08/2022 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 04/06/2022 19:53:49

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.080	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.160	19:48
Breath Volume	1.800	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.152	19:52
Breath Volume	1.945	19:52
Air Blank	0.000	19:52
Control Test	0.080	19:53
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 04/08/2022 14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MILLAN

DATE/TIME OF TEST: 04/08/2022 01:32:04

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.122	01:27
Breath Volume	1.382	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.126	01:30
Breath Volume	2.421	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:32

Data Download Date/Time: 04/08/2022 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GASPER

DATE/TIME OF TEST: 04/09/2022 02:40:10

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.151	02:35
Breath Volume	2.312	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.143	02:38
Breath Volume	1.957	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: COL POP

DATE/TIME OF TEST: 04/09/2022 23:04:10

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.207	22:59
Breath Volume	2.371	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.198	23:02
Breath Volume	2.218	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 04/10/2022 00:39:54

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.080	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.116	00:35
Breath Volume	2.664	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.116	00:38
Breath Volume	2.976	00:38
Air Blank	0.000	00:38
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ,MARQUIS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/10/2022 17:23:52

Diagnostics Check	OK	17:09	
Air Blank	0.000	17:09	
Control Test	0.080	17:10	
Air Blank	0.000	17:10	
Subject Sample	# 1 0.276	17:11	No .020 agreement
Breath Volume	2.339	17:11	
Air Blank	0.000	17:11	
Air Blank	0.000	17:13	
Subject Sample	# 2 0.247	17:16	No .020 agreement
Breath Volume	1.503	17:16	
Air Blank	0.000	17:17	
Air Blank	0.000	17:19	
Subject Sample	# 3 0.237	17:22	Volume Not Met
Breath Volume	1.070	17:22	
Air Blank	0.000	17:22	
Control Test	0.080	17:23	
Air Blank	0.000	17:23	
Diagnostics Check	OK	17:23	

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALES,MARQUIS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/10/2022 17:29:10

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:26	
Control Test	0.080	17:27	
Air Blank	0.000	17:27	
Subject Sample	# 1	17:27	Subject Test Refused
Breath Volume	0.000	17:27	
Air Blank	0.000	17:28	
Control Test	0.079	17:28	
Air Blank	0.000	17:29	
Diagnostics Check	OK	17:29	

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCLUSKEY,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: ROBLERO GUTIERREZ

DATE/TIME OF TEST: 04/10/2022 17:47:11

Diagnostics Check	OK	17:37	
Air Blank	0.000	17:37	
Control Test	0.080	17:37	
Air Blank	0.000	17:38	
Subject Sample	# 1 0.000	17:41	Volume Not Met
Breath Volume	0.015	17:41	
Air Blank	0.000	17:42	
Air Blank	0.000	17:43	
Subject Sample	# 2	17:45	Subject Test Refused
Breath Volume	0.070	17:45	
Air Blank	0.000	17:46	
Control Test	0.077	17:46	
Air Blank	0.000	17:47	
Diagnostics Check	OK	17:47	

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: COLEIII

DATE/TIME OF TEST: 04/10/2022 22:08:08

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.171	22:03
Breath Volume	2.390	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.175	22:06
Breath Volume	2.812	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS, THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: ARZENO

DATE/TIME OF TEST: 04/11/2022 05:06:05

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.000	05:01
Breath Volume	3.082	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.000	05:04
Breath Volume	3.460	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:06

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LERNER,EVERETT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: MISSANO

DATE/TIME OF TEST: 04/11/2022 18:18:54

Diagnostics Check	OK	18:12
Air Blank	0.000	18:13
Control Test	0.081	18:13
Air Blank	0.000	18:14
Subject Sample # 1	0.000	18:14
Breath Volume	2.007	18:14
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.000	18:17
Breath Volume	1.972	18:17
Air Blank	0.000	18:17
Control Test	0.080	18:18
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOHNSON,TONRICA
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DRUMMER

DATE/TIME OF TEST: 04/11/2022 22:07:49

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.385	22:03
Breath Volume	2.585	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.368	22:06
Breath Volume	2.285	22:06
Air Blank	0.000	22:06
Control Test	0.079	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORGAN,DAVID
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ISMAILI

DATE/TIME OF TEST: 04/12/2022 02:39:01

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.000	02:34
Breath Volume	2.679	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:37
Breath Volume	1.613	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAGER,MATTHEW
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HELICHER

DATE/TIME OF TEST: 04/12/2022 21:06:53

Diagnostics Check	OK	20:50	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:52	
Subject Sample	# 1 0.175	20:54	
Breath Volume	1.171	20:54	
Air Blank	0.000	20:54	
Air Blank	0.000	20:56	
Subject Sample	# 2 0.138	20:59	Volume Not Met
Breath Volume	0.121	20:59	
Air Blank	0.000	21:00	
Air Blank	0.000	21:02	
Subject Sample	# 3	21:05	Subject Test Refused
Breath Volume	0.246	21:05	
Air Blank	0.000	21:05	
Control Test	0.079	21:06	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: MUNDO

DATE/TIME OF TEST: 04/14/2022 02:16:48

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.137	02:12
Breath Volume	2.796	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.133	02:15
Breath Volume	2.425	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,REILY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ZULUAGA

DATE/TIME OF TEST: 04/15/2022 03:32:24

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.148	03:27
Breath Volume	1.359	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.154	03:30
Breath Volume	1.941	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROWE,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: ALVAREZ BONILLA

DATE/TIME OF TEST: 04/17/2022 17:06:02

Diagnostics Check	OK	16:59
Air Blank	0.000	16:59
Control Test	0.080	16:59
Air Blank	0.000	17:00
Subject Sample # 1	0.243	17:01
Breath Volume	1.179	17:01
Air Blank	0.000	17:01
Air Blank	0.000	17:03
Subject Sample # 2	0.245	17:04
Breath Volume	1.785	17:04
Air Blank	0.000	17:05
Control Test	0.080	17:05
Air Blank	0.000	17:05
Diagnostics Check	OK	17:06

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/19/2022 01:32:17

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.117	01:27
Breath Volume	2.320	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.114	01:30
Breath Volume	2.285	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLARKE,ROMAINE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PESKIN

DATE/TIME OF TEST: 04/19/2022 23:45:30

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.080	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.000	23:41
Breath Volume	2.476	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.000	23:44
Breath Volume	1.914	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STRONG,CHRISTINA
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: OLIVET

DATE/TIME OF TEST: 04/20/2022 20:49:55

Diagnostics Check	OK	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.000	20:45
Breath Volume	3.347	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:48
Subject Sample # 2	0.000	20:48
Breath Volume	3.757	20:48
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: LIVAR ORON

DATE/TIME OF TEST: 04/21/2022 05:38:48

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.081	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.140	05:34
Breath Volume	2.246	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.137	05:37
Breath Volume	2.335	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJEDA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ROSALES GARCIA

DATE/TIME OF TEST: 04/21/2022 23:23:35

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.199	23:18
Breath Volume	1.812	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.185	23:21
Breath Volume	2.375	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALVDORD,DAVID
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: OLIVET

DATE/TIME OF TEST: 04/22/2022 01:29:55

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.000	01:25
Breath Volume	3.042	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:28
Breath Volume	2.781	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GONZALEZ, MARQUIS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 04/22/2022 20:05:06

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.081	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.201	20:00
Breath Volume	2.886	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.202	20:03
Breath Volume	2.601	20:03
Air Blank	0.000	20:04
Control Test	0.079	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 04/22/2022 23:53:02

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.183	23:48
Breath Volume	2.742	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.184	23:51
Breath Volume	2.386	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:53

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: DISHON

DATE/TIME OF TEST: 04/23/2022 03:34:17

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.201	03:29
Breath Volume	3.628	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.201	03:32
Breath Volume	3.425	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MOURETO

DATE/TIME OF TEST: 04/24/2022 02:20:40

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:19	Slope Not Met
Breath Volume	1.488	02:19	
Air Blank	0.000	02:19	
Control Test	0.080	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MOURETO

DATE/TIME OF TEST: 04/24/2022 02:50:27

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.200	02:45
Breath Volume	3.539	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.196	02:48
Breath Volume	3.453	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALKER, JODI
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BOTTHBY

DATE/TIME OF TEST: 04/27/2022 03:09:08

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.223	03:03
Breath Volume	2.042	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.214	03:07
Breath Volume	1.777	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ELVES,EDUARDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/29/2022 04:25:27

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.212	04:20
Breath Volume	1.855	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.193	04:23
Breath Volume	2.058	04:23
Air Blank	0.000	04:24
Control Test	0.079	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 05/20/2022 11:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SIMMONS, TIMOTHY
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 04/30/2022 00:41:55

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.194	00:36
Breath Volume	2.964	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.176	00:40
Breath Volume	1.214	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 05/20/2022 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SOLOWAY,ALAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HUTCHINSON

DATE/TIME OF TEST: 04/30/2022 01:57:56

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:45	
Control Test	0.080	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1 0.175	01:49	Volume Not Met
Breath Volume	0.476	01:49	
Air Blank	0.000	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 2 0.188	01:52	
Breath Volume	1.914	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 3 0.184	01:56	
Breath Volume	1.503	01:56	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 05/20/2022 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/30/2022 02:28:32

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.196	02:23
Breath Volume	2.164	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.192	02:26
Breath Volume	1.980	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 05/20/2022 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MULLINGS

DATE/TIME OF TEST: 04/30/2022 05:07:04

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.102	05:01
Breath Volume	1.238	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.107	05:05
Breath Volume	2.421	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:07

Data Download Date/Time: 05/20/2022 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A4
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: MCNALLY

DATE/TIME OF TEST: 04/30/2022 22:21:12

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.351	22:13
Breath Volume	2.406	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.373	22:16
Breath Volume	2.550	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:18
Subject Sample # 3	0.342	22:19
Breath Volume	2.730	22:19
Air Blank	0.000	22:20
Control Test	0.079	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 05/20/2022 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: OSMAN

DATE/TIME OF TEST: 04/01/2022 02:32:11

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.097	02:27
Breath Volume	2.628	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.095	02:30
Breath Volume	3.214	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/27/2022 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: WIND

DATE/TIME OF TEST: 04/01/2022 22:05:12

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.000	22:00
Breath Volume	2.785	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.000	22:03
Breath Volume	2.906	22:03
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 04/27/2022 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 04/04/2022 23:43:03

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.193	23:38
Breath Volume	2.695	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.187	23:41
Breath Volume	2.277	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:43

Data Download Date/Time: 04/27/2022 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: DELAY,DEVIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BARTHOLEMEW

DATE/TIME OF TEST: 04/06/2022 00:34:44

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.161	00:30
Breath Volume	3.515	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.145	00:33
Breath Volume	1.519	00:33
Air Blank	0.000	00:33
Control Test	0.078	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 04/27/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEX
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/09/2022 00:32:44

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.161	00:28
Breath Volume	3.472	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.158	00:31
Breath Volume	2.878	00:31
Air Blank	0.000	00:31
Control Test	0.080	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 04/27/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HATTON

DATE/TIME OF TEST: 04/10/2022 00:33:54

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.028	00:29
Breath Volume	1.574	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.024	00:32
Breath Volume	1.570	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 04/27/2022 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: STROTHER

DATE/TIME OF TEST: 04/10/2022 02:47:56

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.082	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.068	02:43
Breath Volume	1.855	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.068	02:46
Breath Volume	1.250	02:46
Air Blank	0.000	02:47
Control Test	0.082	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 04/27/2022 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/14/2022 00:31:06

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.263	00:26
Breath Volume	3.628	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.244	00:29
Breath Volume	1.429	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 04/27/2022 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 04/24/2022 02:08:10

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.120	02:03
Breath Volume	3.574	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.117	02:06
Breath Volume	3.250	02:06
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 04/27/2022 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/26/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 202303
DRY GAS EXPIRY DATE: 03/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 04/29/2022 23:59:44

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:53	
Control Test	0.080	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1 0.011	23:55	
Breath Volume	1.578	23:55	
Air Blank	0.000	23:55	
Air Blank	0.000	23:57	
Subject Sample	# 2	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.079	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 05/23/2022 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/17/2022 20:09:20

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.079	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.000	20:05
Breath Volume	1.312	20:05
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.000	20:07
Breath Volume	1.582	20:07
Air Blank	0.000	20:08
Control Test	0.078	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 04/27/2022 20:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: EVELER,JONATHAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/03/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1206004
DRY GAS EXPIRY DATE: 09/20/2022
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 04/17/2022 20:25:21

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.078	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.158	20:20
Breath Volume	1.984	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.150	20:23
Breath Volume	1.976	20:23
Air Blank	0.000	20:24
Control Test	0.078	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 04/27/2022 20:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: THOMAS,JUSTIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 04/02/2022 05:53:03

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.079	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.222	05:48
Breath Volume	3.816	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:50
Subject Sample # 2	0.220	05:51
Breath Volume	3.472	05:51
Air Blank	0.000	05:52
Control Test	0.077	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:53

Data Download Date/Time: 04/28/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, MAURQUICE, D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PAYNE, ANDREW
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: VAZQUEZ BRIONES

DATE/TIME OF TEST: 04/04/2022 23:16:38

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.078	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:14	RFI Detect
Breath Volume	0.078	23:14	
Air Blank	0.000	23:15	
Control Test	0.078	23:15	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 04/28/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,MAURQUICE,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PAYNE,ANDREW
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: VAZQUEZ BRIONES

DATE/TIME OF TEST: 04/04/2022 23:28:26

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.077	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1 0.118	23:23	Volume Not Met
Breath Volume	0.382	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 2	23:27	Sequence Aborted
Breath Volume	0.699	23:27	
Air Blank	0.000	23:28	

Data Download Date/Time: 04/28/2022 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, MAURQUICE, D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: CREWS, KAGE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HOESCHELE

DATE/TIME OF TEST: 04/05/2022 02:06:58

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.256	02:02
Breath Volume	3.457	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:04
Subject Sample # 2	0.261	02:05
Breath Volume	3.089	02:05
Air Blank	0.000	02:06
Control Test	0.076	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 04/28/2022 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: DAVIS,MATTHEW
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: XENES

DATE/TIME OF TEST: 04/10/2022 06:22:23

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.078	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.155	06:16
Breath Volume	1.441	06:16
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.143	06:20
Breath Volume	1.566	06:20
Air Blank	0.000	06:21
Control Test	0.078	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 04/28/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ARANGO MERINO

DATE/TIME OF TEST: 04/17/2022 04:02:27

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.077	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample	# 2	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.075	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 04/28/2022 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 09:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: MARLOW

DATE/TIME OF TEST: 04/23/2022 10:38:54

Diagnostics Check	OK	10:25	
Air Blank	0.000	10:26	
Control Test	0.078	10:26	
Air Blank	0.000	10:26	
Subject Sample	# 1	10:30	No Sample Provided
Breath Volume	0.000	10:30	
Air Blank	0.000	10:30	
Air Blank	0.000	10:32	
Subject Sample	# 2	10:33	
Breath Volume	1.382	10:33	
Air Blank	0.000	10:34	
Air Blank	0.000	10:36	
Subject Sample	# 3	10:37	
Breath Volume	1.339	10:37	
Air Blank	0.000	10:37	
Control Test	0.078	10:38	
Air Blank	0.000	10:38	
Diagnostics Check	OK	10:38	

Data Download Date/Time: 04/28/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/24/2022 00:06:23

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.077	00:04	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	RFI Detect
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.077	00:05	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 04/28/2022 14:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 03/30/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/24/2022 00:15:16

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.077	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.163	00:10
Breath Volume	1.773	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.165	00:13
Breath Volume	1.437	00:13
Air Blank	0.000	00:14
Control Test	0.076	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/28/2022 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 04/27/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 04/30/2022 05:00:47

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.155	04:53
Breath Volume	2.640	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:55
Subject Sample # 2	0.176	04:56
Breath Volume	2.824	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 3	0.167	04:59
Breath Volume	3.320	04:59
Air Blank	0.000	04:59
Control Test	0.077	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 05/29/2022 21:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 04/02/2022 01:45:39

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:44	Subject Test Refused
Breath Volume	0.000	01:44	
Air Blank	0.000	01:44	
Control Test	0.081	01:45	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/03/2022 01:06:54

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.000	01:02
Breath Volume	1.417	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.000	01:05
Breath Volume	1.320	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: RODRIGUEZ ALFONSO

DATE/TIME OF TEST: 04/06/2022 20:01:53

Diagnostics Check	OK	19:58	
Air Blank	0.000	19:58	
Control Test	0.080	19:59	
Air Blank	0.000	19:59	
Subject Sample	# 1	20:00	Subject Test Refused
Breath Volume	0.000	20:00	
Air Blank	0.000	20:00	
Control Test	0.081	20:01	
Air Blank	0.000	20:01	
Diagnostics Check	OK	20:01	

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FOREMAN,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 04/13/2022 02:34:45

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.082	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.000	02:30
Breath Volume	2.125	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.000	02:33
Breath Volume	2.949	02:33
Air Blank	0.000	02:33
Control Test	0.082	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RIVERA,SAMUEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: STEADLEY

DATE/TIME OF TEST: 04/14/2022 00:38:07

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.152	00:33
Breath Volume	2.296	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:35
Subject Sample # 2	0.150	00:36
Breath Volume	2.031	00:36
Air Blank	0.000	00:37
Control Test	0.078	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: OLGUIN GONZALEZ

DATE/TIME OF TEST: 04/16/2022 22:30:25

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.221	22:25
Breath Volume	1.667	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.217	22:28
Breath Volume	1.519	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FOREMAN,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 04/17/2022 02:03:43

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.082	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: VILLARREAL

DATE/TIME OF TEST: 04/17/2022 03:31:21

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.082	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.134	03:26
Breath Volume	2.250	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.126	03:29
Breath Volume	1.734	03:29
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RIVERA,SAMUEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PELAEZ GOMEZ

DATE/TIME OF TEST: 04/23/2022 00:33:20

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.188	00:28
Breath Volume	2.058	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.188	00:31
Breath Volume	2.246	00:31
Air Blank	0.000	00:32
Control Test	0.077	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GEORGE,SHAQWAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/18/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/25/2022 00:20:05

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.179	00:15
Breath Volume	3.261	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.160	00:18
Breath Volume	1.437	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 05/31/2022 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,BENJAMIN,W
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: KENT,ROBERT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 04/01/2022 23:42:34

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.188	23:37
Breath Volume	3.496	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.190	23:40
Breath Volume	2.687	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 04/08/2022 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,BRENDAN,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MASSEY,SIDNEY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SQUAIRE

DATE/TIME OF TEST: 04/02/2022 02:10:47

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.081	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1 0.000	01:59	
Breath Volume	1.140	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2 0.000	02:04	Volume Not Met
Breath Volume	0.257	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 3 0.000	02:09	
Breath Volume	1.199	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 04/08/2022 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,KARISSA,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: ROEDEL,JUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/03/2022 05:15:46

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.289	05:06
Breath Volume	1.589	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.261	05:10
Breath Volume	1.589	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:12
Subject Sample # 3	0.284	05:14
Breath Volume	1.839	05:14
Air Blank	0.000	05:14
Control Test	0.076	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 04/08/2022 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMASEGOVICH,KODY,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SEGREE,SCOUT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/16/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 04/04/2022 23:53:29

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.081	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.159	23:48
Breath Volume	1.960	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.160	23:51
Breath Volume	1.546	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 04/08/2022 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGALADO,JORGE,I
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: REGALADO,JORGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 04/20/2022 09:20:48

Diagnostics Check	OK	09:11
Air Blank	0.000	09:12
Control Test	0.081	09:12
Air Blank	0.000	09:13
Subject Sample # 1	0.241	09:13
Breath Volume	1.652	09:13
Air Blank	0.000	09:14
Air Blank	0.000	09:16
Subject Sample # 2	0.232	09:18
Breath Volume	1.156	09:18
Air Blank	0.000	09:19
Control Test	0.080	09:20
Air Blank	0.000	09:20
Diagnostics Check	OK	09:20

Data Download Date/Time: 05/05/2022 08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HAITT, JORGE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/21/2022 02:46:22

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.158	02:41
Breath Volume	2.019	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.145	02:44
Breath Volume	1.160	02:44
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 05/05/2022 08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS, REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1408275
DRY GAS EXPIRY DATE: 11/19/2023
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/22/2022 03:37:35

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.224	03:32
Breath Volume	3.226	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.227	03:35
Breath Volume	1.980	03:35
Air Blank	0.000	03:36
Control Test	0.076	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 05/05/2022 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDON,JEFFERSON,W
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LLOYD,GEORGE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 04/23/2022 04:51:36

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.079	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.118	04:47
Breath Volume	3.496	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.118	04:50
Breath Volume	1.304	04:50
Air Blank	0.000	04:50
Control Test	0.078	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 05/05/2022 08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLGOOD,JEFFREY,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLGOOD,JEFFREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/08/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1283438
DRY GAS EXPIRY DATE: 02/25/2023
VIOLATION CODE: BUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 04/23/2022 20:33:30

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.079	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.080	20:32	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 05/05/2022 08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CARVER,JAMIE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 04/02/2022 12:04:06

Diagnostics Check	OK	11:57
Air Blank	0.000	11:58
Control Test	0.079	11:58
Air Blank	0.000	11:58
Subject Sample # 1	0.178	11:59
Breath Volume	2.355	11:59
Air Blank	0.000	12:00
Air Blank	0.000	12:01
Subject Sample # 2	0.180	12:02
Breath Volume	2.066	12:02
Air Blank	0.000	12:03
Control Test	0.079	12:03
Air Blank	0.000	12:04
Diagnostics Check	OK	12:04

Data Download Date/Time: 04/19/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,DAVID,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUNDT,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: QUIZON

DATE/TIME OF TEST: 04/04/2022 16:34:00

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.078	16:25
Air Blank	0.000	16:25
Subject Sample # 1	0.352	16:26
Breath Volume	1.796	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:28
Subject Sample # 2	0.387	16:29
Breath Volume	1.718	16:29
Air Blank	0.000	16:30
Air Blank	0.000	16:31
Subject Sample # 3	0.368	16:32
Breath Volume	2.187	16:32
Air Blank	0.000	16:33
Control Test	0.076	16:33
Air Blank	0.000	16:33
Diagnostics Check	OK	16:34

Data Download Date/Time: 04/19/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT, COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: BEACHNER

DATE/TIME OF TEST: 04/08/2022 18:45:46

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.080	18:40
Air Blank	0.000	18:40
Subject Sample # 1	0.286	18:41
Breath Volume	3.105	18:41
Air Blank	0.000	18:42
Air Blank	0.000	18:43
Subject Sample # 2	0.287	18:44
Breath Volume	3.703	18:44
Air Blank	0.000	18:44
Control Test	0.078	18:45
Air Blank	0.000	18:45
Diagnostics Check	OK	18:45

Data Download Date/Time: 04/19/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 04/09/2022 02:36:03

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.152	02:31
Breath Volume	1.566	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.165	02:34
Breath Volume	1.925	02:34
Air Blank	0.000	02:35
Control Test	0.076	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:36

Data Download Date/Time: 04/19/2022 12:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RHODEN, DELRICKUS, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RHODEN, DELRICKUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1346146
DRY GAS EXPIRY DATE: 07/07/2023
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 04/09/2022 22:02:55

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.197	21:53
Breath Volume	1.351	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:55
Subject Sample # 2	0.173	21:57
Breath Volume	1.628	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 3	0.179	22:01
Breath Volume	1.835	22:01
Air Blank	0.000	22:01
Control Test	0.077	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 04/19/2022 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MIRANDA JR

DATE/TIME OF TEST: 04/02/2022 04:34:04

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.119	04:29
Breath Volume	1.300	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.121	04:32
Breath Volume	1.277	04:32
Air Blank	0.000	04:33
Control Test	0.078	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:34

Data Download Date/Time: 04/07/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINOSO,PEDRO,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: REINOSO,PEDRO
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: DOMINGUEZ SANCHEZ

DATE/TIME OF TEST: 04/02/2022 13:33:03

Diagnostics Check	OK	13:30	
Air Blank	0.000	13:30	
Control Test	0.078	13:30	
Air Blank	0.000	13:31	
Subject Sample	# 1	13:31	Subject Test Refused
Breath Volume	0.000	13:31	
Air Blank	0.000	13:32	
Control Test	0.079	13:32	
Air Blank	0.000	13:32	
Diagnostics Check	OK	13:33	

Data Download Date/Time: 04/07/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 04/03/2022 03:57:53

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.262	03:52
Breath Volume	1.765	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.260	03:56
Breath Volume	1.167	03:56
Air Blank	0.000	03:56
Control Test	0.077	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/07/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HRADECKY,ALEXANDRIA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:34

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 04/04/2022 11:04:24

Diagnostics Check	OK	10:54
Air Blank	0.000	10:55
Control Test	0.078	10:55
Air Blank	0.000	10:56
Subject Sample # 1	0.033	10:58
Breath Volume	1.109	10:58
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.031	11:02
Breath Volume	1.167	11:02
Air Blank	0.000	11:03
Control Test	0.078	11:03
Air Blank	0.000	11:04
Diagnostics Check	OK	11:04

Data Download Date/Time: 04/07/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,CASEY,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORNELAS,ANTONIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: PILE

DATE/TIME OF TEST: 04/05/2022 18:28:34

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.078	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.160	18:23
Breath Volume	2.441	18:23
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.144	18:26
Breath Volume	1.960	18:26
Air Blank	0.000	18:27
Control Test	0.078	18:27
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 04/07/2022 14:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/10/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/06/2022 19:51:20

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.078	19:43	
Air Blank	0.000	19:43	
Subject Sample	# 1 0.159	19:46	Volume Not Met
Breath Volume	0.253	19:46	
Air Blank	0.000	19:47	
Air Blank	0.000	19:49	
Subject Sample	# 2	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:50	
Control Test	0.077	19:50	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

Data Download Date/Time: 04/07/2022 14:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: RENEAU-RICH

DATE/TIME OF TEST: 04/08/2022 01:11:58

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:11	
Control Test	0.078	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 05/12/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: MCNEW

DATE/TIME OF TEST: 04/08/2022 23:38:32

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.121	23:34
Breath Volume	1.691	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.126	23:36
Breath Volume	1.566	23:36
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 05/12/2022 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/10/2022 00:17:35

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:14	
Control Test	0.079	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 05/12/2022 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/12/2022 02:36:41

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.079	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.079	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/12/2022 21:28:02

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.078	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.079	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GIL,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GONZALEZ,GIL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: BRUNN

DATE/TIME OF TEST: 04/15/2022 16:03:27

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:01	
Control Test	0.078	16:01	
Air Blank	0.000	16:01	
Subject Sample	# 1	16:02	Subject Test Refused
Breath Volume	0.000	16:02	
Air Blank	0.000	16:02	
Control Test	0.078	16:02	
Air Blank	0.000	16:03	
Diagnostics Check	OK	16:03	

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: OLOIZIA

DATE/TIME OF TEST: 04/15/2022 23:56:38

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:55	
Control Test	0.079	23:56	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: JOSHI

DATE/TIME OF TEST: 04/16/2022 02:05:55

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.146	02:00
Breath Volume	1.167	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.153	02:04
Breath Volume	1.421	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MEDEROS

DATE/TIME OF TEST: 04/16/2022 02:27:31

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.198	02:23
Breath Volume	2.375	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.199	02:25
Breath Volume	2.304	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RAMIREZ,ANDRES
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/17/2022 04:30:15

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.078	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:28	Subject Test Refused
Breath Volume	0.000	04:28	
Air Blank	0.000	04:29	
Control Test	0.077	04:29	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JOEL,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 04/30/2022 02:36:53

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.206	02:31
Breath Volume	1.511	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.189	02:35
Breath Volume	1.195	02:35
Air Blank	0.000	02:35
Control Test	0.077	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RAMIREZ,ANDRAS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DIAZ HERNANDEZ

DATE/TIME OF TEST: 04/30/2022 03:36:59

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1 0.139	03:29	Volume Not Met
Breath Volume	0.156	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2 0.151	03:35	Volume Not Met
Breath Volume	0.410	03:35	
Air Blank	0.000	03:36	
Control Test	0.076	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RAMIREZ,ANDRES
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/07/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DAIZ HERNANDEZ

DATE/TIME OF TEST: 04/30/2022 03:49:31

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.144	03:44
Breath Volume	1.367	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.151	03:47
Breath Volume	1.609	03:47
Air Blank	0.000	03:48
Control Test	0.076	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 05/12/2022 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKLE,SALLY,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: HINKLE

DATE/TIME OF TEST: 04/07/2022 18:58:55

Diagnostics Check	OK	18:51	
Air Blank	0.000	18:52	
Control Test	0.079	18:52	
Air Blank	0.000	18:52	
Subject Sample	# 1 0.256	18:53	
Breath Volume	1.746	18:53	
Air Blank	0.000	18:54	
Air Blank	0.000	18:56	
Subject Sample	# 2 0.251	18:57	
Breath Volume	1.792	18:57	
Air Blank	0.000	18:57	
Control Test	0.074	18:58	Control Outside Tolerance
Air Blank	0.000	18:58	

Data Download Date/Time: 04/13/2022 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'DOR,ROBERT,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: O'DOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: HINKLE

DATE/TIME OF TEST: 04/07/2022 19:11:50

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.076	19:02
Air Blank	0.000	19:02
Subject Sample # 1	0.258	19:03
Breath Volume	1.894	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.237	19:06
Breath Volume	1.347	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 3	0.251	19:10
Breath Volume	2.109	19:10
Air Blank	0.000	19:10
Control Test	0.076	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 04/13/2022 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DULCIO

DATE/TIME OF TEST: 04/10/2022 02:40:06

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.078	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.150	02:34
Breath Volume	1.484	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.141	02:38
Breath Volume	1.585	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/13/2022 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/02/2022

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1157961
DRY GAS EXPIRY DATE: 06/17/2022
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 04/10/2022 04:39:03

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.178	04:33
Breath Volume	1.355	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.178	04:37
Breath Volume	2.140	04:37
Air Blank	0.000	04:38
Control Test	0.077	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:39

Data Download Date/Time: 04/13/2022 10:29